

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND,TROY,M  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: SMITH,SYDNEY  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: KEEFER

DATE/TIME OF TEST: 11/01/2020 19:53:11

---

Diagnostics Check	OK	19:46
Air Blank	0.000	19:47
Control Test	0.080	19:47
Air Blank	0.000	19:48
Subject Sample # 1	0.177	19:48
Breath Volume	4.378	19:48
Air Blank	0.000	19:49
Air Blank	0.000	19:51
Subject Sample # 2	0.179	19:51
Breath Volume	4.246	19:51
Air Blank	0.000	19:52
Control Test	0.080	19:52
Air Blank	0.000	19:53
Diagnostics Check	OK	19:53

---

Data Download Date/Time: 11/04/2020 22:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: MOORE,KIMBERLY  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: BRUCE

DATE/TIME OF TEST: 11/02/2020 01:52:57

---

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:48
Subject Sample # 1	0.000	01:48
Breath Volume	3.121	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.000	01:51
Breath Volume	3.257	01:51
Air Blank	0.000	01:52
Control Test	0.079	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

---

Data Download Date/Time: 11/04/2020 22:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: CANNADAY,ZACHARY  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MAHONY

DATE/TIME OF TEST: 11/02/2020 02:51:15

---

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.081	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.144	02:46
Breath Volume	1.757	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.142	02:49
Breath Volume	1.886	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

---

Data Download Date/Time: 11/04/2020 22:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: NORRIS,MARK  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: TRENT

DATE/TIME OF TEST: 11/02/2020 21:16:37

---

Diagnostics Check	OK	21:06
Air Blank	0.000	21:06
Control Test	0.080	21:07
Air Blank	0.000	21:07
Subject Sample # 1	0.164	21:08
Breath Volume	3.613	21:08
Air Blank	0.000	21:09
Air Blank	0.000	21:11
Subject Sample # 2	0.142	21:11
Breath Volume	2.097	21:11
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 3	0.139	21:15
Breath Volume	2.242	21:15
Air Blank	0.000	21:15
Control Test	0.081	21:16
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

---

Data Download Date/Time: 11/04/2020 22:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HILL,SEAN,  
BREATH TEST OPERATOR AGENCY: Sanford PD  
ARRESTING OFFICER: SALGADO,J  
ARRESTING OFFICER AGENCY: Sanford PD  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 11/20/2020 15:56:31

---

Diagnostics Check	OK	15:46
Air Blank	0.000	15:46
Control Test	0.080	15:46
Air Blank	0.000	15:47
Subject Sample # 1	0.278	15:48
Breath Volume	2.054	15:48
Air Blank	0.000	15:49
Air Blank	0.000	15:50
Subject Sample # 2	0.299	15:51
Breath Volume	2.406	15:51
Air Blank	0.000	15:52
Air Blank	0.000	15:53
Subject Sample # 3	0.289	15:55
Breath Volume	3.085	15:55
Air Blank	0.000	15:55
Control Test	0.080	15:55
Air Blank	0.000	15:56
Diagnostics Check	OK	15:56

---

Data Download Date/Time: 12/28/2020 22:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KRUPA,KYLE,T  
BREATH TEST OPERATOR AGENCY: Sanford PD  
ARRESTING OFFICER: HUMPHREY,JORDAN  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 11/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 11/28/2020 20:07:03

---

Diagnostics Check	OK	20:00
Air Blank	0.000	20:01
Control Test	0.080	20:01
Air Blank	0.000	20:02
Subject Sample # 1	0.110	20:02
Breath Volume	2.136	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.105	20:05
Breath Volume	3.093	20:05
Air Blank	0.000	20:06
Control Test	0.079	20:06
Air Blank	0.000	20:06
Diagnostics Check	OK	20:07

---

Data Download Date/Time: 12/28/2020 22:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOUSMAN,JOHN,D  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: HOUSMAN,JOHN  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 10/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1327124  
DRY GAS EXPIRY DATE: 05/14/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: WHALEY III

DATE/TIME OF TEST: 11/03/2020 05:06:09

---

Diagnostics Check		OK	04:59
Air Blank		0.000	05:00
Control Test		0.080	05:00
Air Blank		0.000	05:01
Subject Sample	# 1	0.098	05:01
Breath Volume		2.089	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.095	05:04
Breath Volume		1.472	05:04
Air Blank		0.000	05:05
Control Test		0.079	05:05
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

---

Data Download Date/Time: 11/20/2020 06:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOUSMAN,JOHN,D  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: HOUSMAN,JOHN  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 10/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1327124  
DRY GAS EXPIRY DATE: 05/14/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: MAYSE

DATE/TIME OF TEST: 11/13/2020 02:02:40

---

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.079	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.140	01:58
Breath Volume	1.445	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.142	02:01
Breath Volume	1.246	02:01
Air Blank	0.000	02:01
Control Test	0.079	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

---

Data Download Date/Time: 11/20/2020 06:32



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRIEBELL,DON,R  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: MODERSON,TUCKER  
ARRESTING OFFICER AGENCY: Indian Harbour Beach  
LAST AGENCY INSPECTION: 10/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1327124  
DRY GAS EXPIRY DATE: 05/14/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 11/14/2020 01:40:56

---

Diagnostics Check		OK	01:34
Air Blank		0.000	01:34
Control Test		0.079	01:35
Air Blank		0.000	01:35
Subject Sample	# 1	0.209	01:36
Breath Volume		1.816	01:36
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.200	01:39
Breath Volume		1.253	01:39
Air Blank		0.000	01:40
Control Test		0.079	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

---

Data Download Date/Time: 11/20/2020 06:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOUSMAN,JOHN,D  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: HOUSMAN,JOHN  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 11/20/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1327124  
DRY GAS EXPIRY DATE: 05/14/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 11/26/2020 01:10:13

---

Diagnostics Check		OK	01:03
Air Blank		0.000	01:04
Control Test		0.080	01:04
Air Blank		0.000	01:05
Subject Sample	# 1	0.055	01:05
Breath Volume		1.582	01:05
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.053	01:08
Breath Volume		1.843	01:08
Air Blank		0.000	01:09
Control Test		0.080	01:09
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

---

Data Download Date/Time: 12/21/2020 15:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOUSMAN,JOHN,D  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: HOUSMAN,JOHN  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 11/20/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1327124  
DRY GAS EXPIRY DATE: 05/14/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 11/28/2020 04:02:10

---

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.080	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.092	03:57
Breath Volume	3.074	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.089	04:00
Breath Volume	2.414	04:00
Air Blank	0.000	04:01
Control Test	0.078	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

---

Data Download Date/Time: 12/21/2020 15:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEILL, MICHAEL, A  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: WESTON, ANDREW  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 11/20/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1327124  
DRY GAS EXPIRY DATE: 05/14/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: LANGHAM

DATE/TIME OF TEST: 11/29/2020 01:36:27

---

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.000	01:32
Breath Volume	2.746	01:32
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.000	01:35
Breath Volume	3.781	01:35
Air Blank	0.000	01:35
Control Test	0.078	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

---

Data Download Date/Time: 12/21/2020 15:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEILL,MICHAEL,A  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: NEILL,MICHAEL  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 11/20/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1327124  
DRY GAS EXPIRY DATE: 05/14/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: DUFUR

DATE/TIME OF TEST: 11/29/2020 23:09:52

---

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.079	22:58
Air Blank	0.000	22:58
Subject Sample # 1	0.168	22:59
Breath Volume	1.261	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.190	23:03
Breath Volume	1.761	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 3	0.168	23:08
Breath Volume	2.453	23:08
Air Blank	0.000	23:08
Control Test	0.077	23:09
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

---

Data Download Date/Time: 12/21/2020 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CHRISS,WARREN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 10/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: FRITZ

DATE/TIME OF TEST: 11/02/2020 21:50:05

---

Diagnostics Check	OK	21:43
Air Blank	0.000	21:44
Control Test	0.080	21:44
Air Blank	0.000	21:45
Subject Sample # 1	0.070	21:45
Breath Volume	3.261	21:45
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample # 2	0.071	21:48
Breath Volume	3.257	21:48
Air Blank	0.000	21:49
Control Test	0.080	21:49
Air Blank	0.000	21:49
Diagnostics Check	OK	21:50

---

Data Download Date/Time: 11/05/2020 11:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: REIDY,TYLER  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 10/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: KEMMERER

DATE/TIME OF TEST: 11/03/2020 02:52:58

---

Diagnostics Check		OK	02:44	
Air Blank		0.000	02:44	
Control Test		0.080	02:45	
Air Blank		0.000	02:45	
Subject Sample	# 1	0.121	02:48	Volume Not Met
Breath Volume		0.031	02:48	
Air Blank		0.000	02:49	
Air Blank		0.000	02:51	
Subject Sample	# 2		02:51	Subject Test Refused
Breath Volume		0.000	02:51	
Air Blank		0.000	02:52	
Control Test		0.079	02:52	
Air Blank		0.000	02:52	
Diagnostics Check		OK	02:52	

---

Data Download Date/Time: 11/05/2020 11:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: THOMAS,ELISABETH  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 10/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 11/03/2020 21:01:46

---

Diagnostics Check	OK	20:55
Air Blank	0.000	20:56
Control Test	0.080	20:56
Air Blank	0.000	20:57
Subject Sample # 1	0.000	20:57
Breath Volume	1.984	20:57
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.000	21:00
Breath Volume	1.582	21:00
Air Blank	0.000	21:00
Control Test	0.081	21:01
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

---

Data Download Date/Time: 11/05/2020 11:39



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: TEMPLE,ELIZABETH  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: SHABAZZ

DATE/TIME OF TEST: 11/05/2020 03:19:26

---

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.081	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.000	03:15
Breath Volume	1.800	03:15
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.000	03:18
Breath Volume	1.687	03:18
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

---

Data Download Date/Time: 11/05/2020 11:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVIS  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: TEMPLE,ELIZABETH  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 11/05/2020 19:21:12

---

Diagnostics Check	OK	19:18	
Air Blank	0.000	19:18	
Control Test	0.080	19:19	
Air Blank	0.000	19:19	
Subject Sample	# 1	19:19	Subject Test Refused
Breath Volume	0.000	19:19	
Air Blank	0.000	19:20	
Control Test	0.081	19:20	
Air Blank	0.000	19:21	
Diagnostics Check	OK	19:21	

---

Data Download Date/Time: 12/05/2020 04:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: YEARGIN,GENESIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: OROZCO

DATE/TIME OF TEST: 11/06/2020 04:42:03

---

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.080	04:36
Air Blank	0.000	04:36
Subject Sample # 1	0.192	04:37
Breath Volume	1.386	04:37
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.206	04:40
Breath Volume	2.957	04:40
Air Blank	0.000	04:41
Control Test	0.081	04:41
Air Blank	0.000	04:41
Diagnostics Check	OK	04:42

---

Data Download Date/Time: 12/05/2020 04:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: TEMPLE,ELIZABETH  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: BOTKIN

DATE/TIME OF TEST: 11/07/2020 03:58:28

---

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.081	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.131	03:53
Breath Volume	1.273	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.142	03:56
Breath Volume	1.226	03:56
Air Blank	0.000	03:57
Control Test	0.082	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

---

Data Download Date/Time: 12/05/2020 04:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: TEMPLE,ELIZABETH  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: MORAHAN

DATE/TIME OF TEST: 11/07/2020 21:11:38

---

Diagnostics Check		OK	20:57	
Air Blank		0.000	20:58	
Control Test		0.080	20:58	
Air Blank		0.000	20:59	
Subject Sample	# 1	0.063	21:02	Volume Not Met
Breath Volume		1.066	21:02	
Air Blank		0.000	21:02	
Air Blank		0.000	21:04	
Subject Sample	# 2	0.064	21:06	
Breath Volume		1.199	21:06	
Air Blank		0.000	21:07	
Air Blank		0.000	21:09	
Subject Sample	# 3	0.063	21:10	
Breath Volume		1.171	21:10	
Air Blank		0.000	21:10	
Control Test		0.080	21:11	
Air Blank		0.000	21:11	
Diagnostics Check		OK	21:11	

---

Data Download Date/Time: 12/05/2020 04:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: TEMPLE,ELIZABETH  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: ECKELKAMP

DATE/TIME OF TEST: 11/08/2020 03:57:22

---

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:55	
Control Test	0.080	03:55	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank	0.000	03:56	
Control Test	0.080	03:56	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

---

Data Download Date/Time: 12/05/2020 04:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MULLINS,JOHN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: TYLER

DATE/TIME OF TEST: 11/10/2020 02:36:43

---

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.081	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.126	02:31
Breath Volume	1.675	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.127	02:35
Breath Volume	1.464	02:35
Air Blank	0.000	02:35
Control Test	0.080	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

---

Data Download Date/Time: 12/05/2020 04:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BARKLEY, DANIEL  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 13:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: SMITH III

DATE/TIME OF TEST: 11/10/2020 14:07:30

---

Diagnostics Check	OK	14:00
Air Blank	0.000	14:01
Control Test	0.081	14:01
Air Blank	0.000	14:01
Subject Sample # 1	0.000	14:03
Breath Volume	1.738	14:03
Air Blank	0.000	14:03
Air Blank	0.000	14:05
Subject Sample # 2	0.000	14:06
Breath Volume	1.171	14:06
Air Blank	0.000	14:06
Control Test	0.081	14:06
Air Blank	0.000	14:07
Diagnostics Check	OK	14:07

---

Data Download Date/Time: 12/05/2020 04:13



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVI,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAKE,LEVI  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 11/13/2020 01:21:12

---

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:16
Subject Sample # 1	0.000	01:16
Breath Volume	1.820	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.000	01:19
Breath Volume	1.835	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

---

Data Download Date/Time: 12/05/2020 04:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANEY,DAMON,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LANEY,DAMON  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MAGNO JR

DATE/TIME OF TEST: 11/13/2020 01:50:57

---

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.080	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.000	01:46
Breath Volume	1.964	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.000	01:49
Breath Volume	1.460	01:49
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

---

Data Download Date/Time: 12/05/2020 04:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MITCHELL,ALEXANDRA  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 11/16/2020 00:22:34

---

Diagnostics Check		OK	00:15
Air Blank		0.000	00:15
Control Test		0.080	00:16
Air Blank		0.000	00:16
Subject Sample	# 1	0.183	00:17
Breath Volume		1.738	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.198	00:21
Breath Volume		2.164	00:21
Air Blank		0.000	00:21
Control Test		0.080	00:21
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

---

Data Download Date/Time: 12/05/2020 04:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: VANGELL,JONATHAN  
ARRESTING OFFICER AGENCY: Treasure Island PD  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: FRIEDRICHS

DATE/TIME OF TEST: 11/16/2020 03:03:45

---

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:01	
Control Test	0.081	03:01	
Air Blank	0.000	03:01	
Subject Sample	# 1	03:02	Subject Test Refused
Breath Volume	0.000	03:02	
Air Blank	0.000	03:02	
Control Test	0.080	03:03	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:03	

---

Data Download Date/Time: 12/05/2020 04:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: SWANGLER, JOHN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 08:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:04

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 11/20/2020 09:31:06

---

Diagnostics Check	OK	09:25
Air Blank	0.000	09:25
Control Test	0.082	09:25
Air Blank	0.000	09:26
Subject Sample # 1	0.000	09:26
Breath Volume	2.484	09:26
Air Blank	0.000	09:27
Air Blank	0.000	09:29
Subject Sample # 2	0.000	09:29
Breath Volume	1.687	09:29
Air Blank	0.000	09:30
Control Test	0.081	09:30
Air Blank	0.000	09:31
Diagnostics Check	OK	09:31

---

Data Download Date/Time: 12/05/2020 04:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: HIRSHMAN,BRIAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: TUTTLE V

DATE/TIME OF TEST: 11/21/2020 04:42:40

---

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.080	04:37
Air Blank	0.000	04:37
Subject Sample # 1	0.126	04:38
Breath Volume	1.722	04:38
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.133	04:41
Breath Volume	1.769	04:41
Air Blank	0.000	04:41
Control Test	0.080	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

---

Data Download Date/Time: 12/05/2020 04:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: TEMPLE,ELIZABETH  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: FORTNER

DATE/TIME OF TEST: 11/21/2020 05:26:18

---

Diagnostics Check	OK	05:20
Air Blank	0.000	05:20
Control Test	0.081	05:20
Air Blank	0.000	05:21
Subject Sample # 1	0.000	05:21
Breath Volume	2.281	05:21
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 2	0.000	05:24
Breath Volume	2.460	05:24
Air Blank	0.000	05:25
Control Test	0.080	05:25
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

---

Data Download Date/Time: 12/05/2020 04:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,CRISTIAN,D  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: SNYDER,RYAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 11/22/2020 00:36:20

---

Diagnostics Check		OK	00:27
Air Blank		0.000	00:27
Control Test		0.081	00:27
Air Blank		0.000	00:28
Subject Sample	# 1	0.226	00:29
Breath Volume		2.070	00:29
Air Blank		0.000	00:30
Air Blank		0.000	00:32
Subject Sample	# 2	0.220	00:34
Breath Volume		1.750	00:34
Air Blank		0.000	00:35
Control Test		0.082	00:35
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

---

Data Download Date/Time: 12/05/2020 04:14



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: OSTROSKI,NICOLE  
ARRESTING OFFICER AGENCY: Treasure Island PD  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: RETTKA

DATE/TIME OF TEST: 11/22/2020 19:59:11

---

Diagnostics Check	OK	19:56	
Air Blank	0.000	19:56	
Control Test	0.081	19:57	
Air Blank	0.000	19:57	
Subject Sample	# 1	19:57	Subject Test Refused
Breath Volume	0.000	19:57	
Air Blank	0.000	19:58	
Control Test	0.081	19:58	
Air Blank	0.000	19:59	
Diagnostics Check	OK	19:59	

---

Data Download Date/Time: 12/05/2020 04:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MULLINS,JOHN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 11/25/2020 02:12:43

---

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:10	
Control Test	0.080	02:10	
Air Blank	0.000	02:11	
Subject Sample	# 1	02:11	Subject Test Refused
Breath Volume	0.000	02:11	
Air Blank	0.000	02:11	
Control Test	0.080	02:12	
Air Blank	0.000	02:12	
Diagnostics Check	OK	02:12	

---

Data Download Date/Time: 12/05/2020 04:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: HIRSHMAN,BRIAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 11/26/2020 05:00:00

---

Diagnostics Check	OK	04:57	
Air Blank	0.000	04:57	
Control Test	0.081	04:57	
Air Blank	0.000	04:58	
Subject Sample	# 1	04:58	Subject Test Refused
Breath Volume	0.000	04:58	
Air Blank	0.000	04:59	
Control Test	0.080	04:59	
Air Blank	0.000	04:59	
Diagnostics Check	OK	05:00	

---

Data Download Date/Time: 12/05/2020 04:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EASTTY,MARK,L  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: SPENCER,RYAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: BALDSCHUN

DATE/TIME OF TEST: 11/27/2020 05:06:05

---

Diagnostics Check	OK	04:54	
Air Blank	0.000	04:54	
Control Test	0.080	04:55	
Air Blank	0.000	04:55	
Subject Sample	# 1	04:58	No Sample Provided
Breath Volume	0.000	04:58	
Air Blank	0.000	04:59	
Air Blank	0.000	05:01	
Subject Sample	# 2	0.193	05:01
Breath Volume	1.500	05:01	
Air Blank	0.000	05:02	
Air Blank	0.000	05:04	
Subject Sample	# 3	0.189	05:04
Breath Volume	2.820	05:04	
Air Blank	0.000	05:05	
Control Test	0.080	05:05	
Air Blank	0.000	05:06	
Diagnostics Check	OK	05:06	

---

Data Download Date/Time: 12/05/2020 04:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MULLINS,JOHN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: MARTINS

DATE/TIME OF TEST: 11/28/2020 03:44:35

---

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.080	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.096	03:40
Breath Volume	2.589	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.083	03:43
Breath Volume	1.367	03:43
Air Blank	0.000	03:43
Control Test	0.081	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

---

Data Download Date/Time: 12/05/2020 04:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: ZSIDO,DANIELLE  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: BREUNIG

DATE/TIME OF TEST: 11/28/2020 04:44:55

---

Diagnostics Check	OK	04:38
Air Blank	0.000	04:39
Control Test	0.081	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.000	04:40
Breath Volume	3.308	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.000	04:43
Breath Volume	2.769	04:43
Air Blank	0.000	04:44
Control Test	0.080	04:44
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

---

Data Download Date/Time: 12/05/2020 04:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAREY,K,P  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CAREY,K  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 11/29/2020 03:23:55

---

Diagnostics Check		OK	03:17
Air Blank		0.000	03:17
Control Test		0.080	03:18
Air Blank		0.000	03:18
Subject Sample	# 1	0.051	03:19
Breath Volume		1.210	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:21
Subject Sample	# 2	0.050	03:22
Breath Volume		1.226	03:22
Air Blank		0.000	03:23
Control Test		0.080	03:23
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

---

Data Download Date/Time: 12/05/2020 04:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: HIRSHMAN,BRIAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: QUIZHPI

DATE/TIME OF TEST: 11/29/2020 21:57:19

---

Diagnostics Check	OK	21:54	
Air Blank	0.000	21:54	
Control Test	0.080	21:55	
Air Blank	0.000	21:55	
Subject Sample	# 1	21:55	Subject Test Refused
Breath Volume	0.000	21:55	
Air Blank	0.000	21:56	
Control Test	0.081	21:56	
Air Blank	0.000	21:57	
Diagnostics Check	OK	21:57	

---

Data Download Date/Time: 12/05/2020 04:14



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVIS  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: TEMPLE,ELIZABETH  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916501  
DRY GAS EXPIRY DATE: 06/14/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: BIGELOW

DATE/TIME OF TEST: 11/13/2020 18:04:51

---

Diagnostics Check	OK	18:00	
Air Blank	0.000	18:01	
Control Test	0.080	18:01	
Air Blank	0.000	18:02	
Subject Sample	# 1	18:03	Subject Test Refused
Breath Volume	0.000	18:03	
Air Blank	0.000	18:03	
Control Test	0.080	18:04	
Air Blank	0.000	18:04	
Diagnostics Check	OK	18:04	

---

Data Download Date/Time: 12/05/2020 04:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUZMAN,MARTIN,V  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: WILSON,GARY  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916501  
DRY GAS EXPIRY DATE: 06/14/2021  
VIOLATION CODE: DUI  
ARREST TIME: 12:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:29

SUBJECT'S LAST NAME: HERDMAN

DATE/TIME OF TEST: 11/16/2020 13:53:04

---

Diagnostics Check	OK	13:50	
Air Blank	0.000	13:50	
Control Test	0.080	13:51	
Air Blank	0.000	13:51	
Subject Sample	# 1	13:51	Subject Test Refused
Breath Volume	0.000	13:51	
Air Blank	0.000	13:52	
Control Test	0.079	13:52	
Air Blank	0.000	13:52	
Diagnostics Check	OK	13:53	

---

Data Download Date/Time: 12/05/2020 04:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: IFFT,GARRETT,T  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: IFFT,GARRETT  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916501  
DRY GAS EXPIRY DATE: 06/14/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 11/18/2020 02:57:00

---

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.081	02:45	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:48	No Sample Provided
Breath Volume	0.000	02:48	
Air Blank	0.000	02:49	
Air Blank	0.000	02:51	
Subject Sample	# 2	0.125	02:52
Breath Volume	1.445	02:52	
Air Blank	0.000	02:52	
Air Blank	0.000	02:54	
Subject Sample	# 3	0.127	02:55
Breath Volume	1.433	02:55	
Air Blank	0.000	02:56	
Control Test	0.081	02:56	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:57	

---

Data Download Date/Time: 12/05/2020 04:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOVEL,RYAN,M  
BREATH TEST OPERATOR AGENCY: Indialantic PD  
ARRESTING OFFICER: DOVEL,RYAN  
ARRESTING OFFICER AGENCY: Indialantic PD  
LAST AGENCY INSPECTION: 10/20/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A4  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: YELTON

DATE/TIME OF TEST: 11/16/2020 19:30:22

---

Diagnostics Check		OK	19:14	
Air Blank		0.000	19:14	
Control Test		0.080	19:14	
Air Blank		0.000	19:15	
Subject Sample	# 1	0.208	19:18	Volume Not Met
Breath Volume		0.316	19:18	
Air Blank		0.000	19:19	
Air Blank		0.000	19:21	
Subject Sample	# 2	0.216	19:23	
Breath Volume		1.750	19:23	
Air Blank		0.000	19:24	
Air Blank		0.000	19:25	
Subject Sample	# 3	0.207	19:29	
Breath Volume		1.265	19:29	
Air Blank			19:29	Ambient Fail
Air Blank		0.000	19:30	

---

Data Download Date/Time: 11/25/2020 11:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOVEL,RYAN,M  
BREATH TEST OPERATOR AGENCY: Indialantic PD  
ARRESTING OFFICER: SMITH,MATTHEW  
ARRESTING OFFICER AGENCY: Melbourne Beach PD  
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A4  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 11/24/2020 22:09:48

---

Diagnostics Check		OK	21:58	
Air Blank		0.000	21:58	
Control Test		0.080	21:58	
Air Blank		0.000	21:59	
Subject Sample	# 1	0.177	22:02	Volume Not Met
Breath Volume		0.335	22:02	
Air Blank		0.000	22:03	
Air Blank		0.000	22:05	
Subject Sample	# 2	0.182	22:08	Volume Not Met
Breath Volume		0.894	22:08	
Air Blank		0.000	22:08	
Control Test		0.077	22:09	
Air Blank		0.000	22:09	
Diagnostics Check		OK	22:09	

---

Data Download Date/Time: 11/25/2020 11:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: JACKSON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 10/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 09:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: HAYNES

DATE/TIME OF TEST: 11/09/2020 10:34:43

---

Diagnostics Check	OK	10:28
Air Blank	0.000	10:28
Control Test	0.082	10:29
Air Blank	0.000	10:29
Subject Sample # 1	0.000	10:30
Breath Volume	1.554	10:30
Air Blank	0.000	10:30
Air Blank	0.000	10:32
Subject Sample # 2	0.000	10:33
Breath Volume	1.429	10:33
Air Blank	0.000	10:33
Control Test	0.080	10:34
Air Blank	0.000	10:34
Diagnostics Check	OK	10:34

---

Data Download Date/Time: 12/29/2020 11:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: JACKSON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 10/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 14:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: JEAN-FRANCOIS

DATE/TIME OF TEST: 11/09/2020 15:36:25

---

Diagnostics Check	OK	15:30
Air Blank	0.000	15:30
Control Test	0.082	15:31
Air Blank	0.000	15:31
Subject Sample # 1	0.000	15:32
Breath Volume	2.855	15:32
Air Blank	0.000	15:32
Air Blank	0.000	15:34
Subject Sample # 2	0.000	15:35
Breath Volume	3.218	15:35
Air Blank	0.000	15:35
Control Test	0.081	15:35
Air Blank	0.000	15:36
Diagnostics Check	OK	15:36

---

Data Download Date/Time: 12/29/2020 11:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACSKON,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: JACKSON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 10/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: PROB/PAROL  
ARREST TIME: 14:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 11/09/2020 15:48:06

---

Diagnostics Check	OK	15:41
Air Blank	0.000	15:42
Control Test	0.082	15:42
Air Blank	0.000	15:43
Subject Sample # 1	0.000	15:43
Breath Volume	1.730	15:43
Air Blank	0.000	15:44
Air Blank	0.000	15:46
Subject Sample # 2	0.000	15:46
Breath Volume	1.609	15:46
Air Blank	0.000	15:47
Control Test	0.085	15:47
Air Blank	0.000	15:48
Diagnostics Check	OK	15:48

---

Data Download Date/Time: 12/29/2020 11:42



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: JACKSON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 10/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: PROB/PAROL  
ARREST TIME: 08:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: CREWS

DATE/TIME OF TEST: 11/10/2020 09:08:40

---

Diagnostics Check	OK	09:02
Air Blank	0.000	09:03
Control Test	0.082	09:03
Air Blank	0.000	09:03
Subject Sample # 1	0.000	09:04
Breath Volume	1.851	09:04
Air Blank	0.000	09:04
Air Blank	0.000	09:06
Subject Sample # 2	0.000	09:07
Breath Volume	1.832	09:07
Air Blank	0.000	09:07
Control Test	0.082	09:08
Air Blank	0.000	09:08
Diagnostics Check	OK	09:08

---

Data Download Date/Time: 12/29/2020 11:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEONARD,BRIAN,J  
BREATH TEST OPERATOR AGENCY: West Melbourne PD  
ARRESTING OFFICER: CROOKS,JUSTIN  
ARRESTING OFFICER AGENCY: West Melbourne PD  
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: JENSEN

DATE/TIME OF TEST: 11/03/2020 18:10:19

---

Diagnostics Check	OK	17:58	
Air Blank	0.000	17:58	
Control Test	0.081	17:59	
Air Blank	0.000	17:59	
Subject Sample	# 1	18:02	No Sample Provided
Breath Volume	0.000	18:02	
Air Blank	0.000	18:03	
Air Blank	0.000	18:05	
Subject Sample	# 2	0.000	18:05
Breath Volume	3.753	18:05	
Air Blank	0.000	18:06	
Air Blank	0.000	18:08	
Subject Sample	# 3	0.000	18:08
Breath Volume	3.242	18:08	
Air Blank	0.000	18:09	
Control Test	0.080	18:09	
Air Blank	0.000	18:10	
Diagnostics Check	OK	18:10	

---

Data Download Date/Time: 11/25/2020 14:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARRITY, THOMAS, A  
BREATH TEST OPERATOR AGENCY: West Melbourne PD  
ARRESTING OFFICER: GARRITY, THOMAS  
ARRESTING OFFICER AGENCY: West Melbourne PD  
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 07:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: MILLSAP

DATE/TIME OF TEST: 11/15/2020 09:24:52

---

Diagnostics Check	OK	09:17
Air Blank	0.000	09:18
Control Test	0.080	09:18
Air Blank	0.000	09:18
Subject Sample # 1	0.240	09:19
Breath Volume	1.648	09:19
Air Blank	0.000	09:20
Air Blank	0.000	09:22
Subject Sample # 2	0.221	09:23
Breath Volume	1.617	09:23
Air Blank	0.000	09:23
Control Test	0.080	09:24
Air Blank	0.000	09:24
Diagnostics Check	OK	09:24

---

Data Download Date/Time: 11/25/2020 14:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAU,BRITTNEY,D  
BREATH TEST OPERATOR AGENCY: West Melbourne PD  
ARRESTING OFFICER: WHEELER,JASON  
ARRESTING OFFICER AGENCY: West Melbourne PD  
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 13:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:22

SUBJECT'S LAST NAME: DONHOFF

DATE/TIME OF TEST: 11/17/2020 14:54:16

---

Diagnostics Check	OK	14:47
Air Blank	0.000	14:47
Control Test	0.081	14:48
Air Blank	0.000	14:48
Subject Sample # 1	0.219	14:49
Breath Volume	1.652	14:49
Air Blank	0.000	14:50
Air Blank	0.000	14:52
Subject Sample # 2	0.214	14:52
Breath Volume	1.152	14:52
Air Blank	0.000	14:53
Control Test	0.079	14:53
Air Blank	0.000	14:54
Diagnostics Check	OK	14:54

---

Data Download Date/Time: 11/25/2020 14:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POPA,SHANE,Z  
BREATH TEST OPERATOR AGENCY: West Melbourne PD  
ARRESTING OFFICER: BUTLER,ADAM  
ARRESTING OFFICER AGENCY: West Melbourne PD  
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:11

SUBJECT'S LAST NAME: SPORY

DATE/TIME OF TEST: 11/19/2020 17:47:03

---

Diagnostics Check	OK	17:40
Air Blank	0.000	17:41
Control Test	0.081	17:41
Air Blank	0.000	17:42
Subject Sample # 1	0.000	17:42
Breath Volume	3.140	17:42
Air Blank	0.000	17:43
Air Blank	0.000	17:45
Subject Sample # 2	0.000	17:45
Breath Volume	3.441	17:45
Air Blank	0.000	17:46
Control Test	0.081	17:46
Air Blank	0.000	17:46
Diagnostics Check	OK	17:47

---

Data Download Date/Time: 11/25/2020 14:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASEY,COREY,M  
BREATH TEST OPERATOR AGENCY: West Melbourne PD  
ARRESTING OFFICER: RADIG,TARAN  
ARRESTING OFFICER AGENCY: West Melbourne PD  
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: YAP

DATE/TIME OF TEST: 11/20/2020 03:37:04

---

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.080	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.213	03:32
Breath Volume		2.246	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.213	03:35
Breath Volume		1.949	03:35
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:37

---

Data Download Date/Time: 11/25/2020 14:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD, WILLIAM,  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: HOWARD, WILLIAM  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 10/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 11/24/2020 04:10:51

---

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.177	04:05
Breath Volume	1.246	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.174	04:09
Breath Volume	1.562	04:09
Air Blank	0.000	04:09
Control Test	0.080	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

---

Data Download Date/Time: 11/25/2020 14:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HICKS,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: West Melbourne PD  
ARRESTING OFFICER: JOHNSON,JAMES  
ARRESTING OFFICER AGENCY: West Melbourne PD  
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: MOLINA-SUAREZ

DATE/TIME OF TEST: 11/26/2020 03:37:47

---

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.160	03:33
Breath Volume	2.457	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:35
Subject Sample # 2	0.168	03:36
Breath Volume	1.937	03:36
Air Blank	0.000	03:36
Control Test	0.079	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

---

Data Download Date/Time: 01/05/2021 11:23



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASEY,COREY,M  
BREATH TEST OPERATOR AGENCY: West Melbourne PD  
ARRESTING OFFICER: DOLCI,PETER  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 11/29/2020 03:57:30

---

Diagnostics Check		OK	03:49
Air Blank		0.000	03:50
Control Test		0.081	03:50
Air Blank		0.000	03:51
Subject Sample	# 1	0.302	03:52
Breath Volume		1.316	03:52
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.305	03:55
Breath Volume		1.625	03:55
Air Blank		0.000	03:56
Control Test		0.081	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

---

Data Download Date/Time: 01/05/2021 11:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEILL,MICHAEL,A  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: NEILL,MICHAEL  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/29/2020 04:40:36

---

Diagnostics Check		OK	04:34
Air Blank		0.000	04:34
Control Test		0.081	04:34
Air Blank		0.000	04:35
Subject Sample	# 1	0.204	04:36
Breath Volume		3.078	04:36
Air Blank		0.000	04:36
Air Blank		0.000	04:38
Subject Sample	# 2	0.195	04:38
Breath Volume		3.304	04:38
Air Blank		0.000	04:39
Control Test		0.079	04:40
Air Blank		0.000	04:40
Diagnostics Check		OK	04:40

---

Data Download Date/Time: 01/05/2021 11:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLCROFT,AMANDA,L  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: HOLLCROFT,AMANDA  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 10/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: OGLE

DATE/TIME OF TEST: 11/06/2020 15:15:02

---

Diagnostics Check	OK	15:08
Air Blank	0.000	15:09
Control Test	0.080	15:09
Air Blank	0.000	15:10
Subject Sample # 1	0.000	15:10
Breath Volume	2.089	15:10
Air Blank	0.000	15:11
Air Blank	0.000	15:13
Subject Sample # 2	0.000	15:13
Breath Volume	2.183	15:13
Air Blank	0.000	15:14
Control Test	0.079	15:14
Air Blank	0.000	15:14
Diagnostics Check	OK	15:15

---

Data Download Date/Time: 12/29/2020 15:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,MARK,E  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: THOMPSON,MARK  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/06/2020 19:23:10

---

Diagnostics Check	OK	19:11	
Air Blank	0.000	19:12	
Control Test	0.080	19:12	
Air Blank	0.000	19:13	
Subject Sample	# 1	19:16	No Sample Provided
Breath Volume	0.000	19:16	
Air Blank	0.000	19:16	
Air Blank	0.000	19:18	
Subject Sample	# 2	19:21	No Sample Provided
Breath Volume	0.000	19:21	
Air Blank	0.000	19:22	
Control Test	0.081	19:22	
Air Blank	0.000	19:23	
Diagnostics Check	OK	19:23	

---

Data Download Date/Time: 11/30/2020 18:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIALOBRZESKI,M,  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: WESTON,A  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/07/2020 04:11:26

---

Diagnostics Check	OK	03:58	
Air Blank	0.000	03:58	
Control Test	0.081	03:59	
Air Blank	0.000	03:59	
Subject Sample	# 1 0.000	04:01	
Breath Volume	2.023	04:01	
Air Blank	0.000	04:01	
Air Blank	0.000	04:03	
Subject Sample	# 2 0.000	04:06	Volume Not Met
Breath Volume	0.015	04:06	
Air Blank	0.000	04:07	
Air Blank	0.000	04:09	
Subject Sample	# 3 0.000	04:10	
Breath Volume	1.324	04:10	
Air Blank	0.000	04:10	
Control Test	0.080	04:10	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:11	

---

Data Download Date/Time: 11/30/2020 18:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIALOBRZESKI,M,  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: BIALOBRZESKI,M  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: STEINHAUSER

DATE/TIME OF TEST: 11/07/2020 22:44:02

---

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.080	22:36
Air Blank	0.000	22:37
Subject Sample # 1	0.191	22:39
Breath Volume	1.906	22:39
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.205	22:42
Breath Volume	2.207	22:42
Air Blank	0.000	22:43
Control Test	0.080	22:43
Air Blank	0.000	22:43
Diagnostics Check	OK	22:44

---

Data Download Date/Time: 11/30/2020 18:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANG, MATTHEW, J  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: WALSH, AUSTIN  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: SHELTON

DATE/TIME OF TEST: 11/11/2020 04:33:25

---

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.081	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.240	04:28
Breath Volume	2.882	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.247	04:31
Breath Volume	2.605	04:31
Air Blank	0.000	04:32
Control Test	0.080	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

---

Data Download Date/Time: 11/30/2020 18:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,JUSTIN,A  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: MORILLO,RUANNY  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 11/15/2020 22:40:19

---

Diagnostics Check	OK	22:32
Air Blank	0.000	22:32
Control Test	0.081	22:33
Air Blank	0.000	22:33
Subject Sample # 1	0.144	22:35
Breath Volume	2.914	22:35
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.142	22:38
Breath Volume	3.570	22:38
Air Blank	0.000	22:39
Control Test	0.079	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

---

Data Download Date/Time: 11/30/2020 18:24



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DETRANO,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: DETRANO,JEFFREY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 11/04/2020 21:40:45

---

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:36	
Control Test	0.078	21:36	
Air Blank	0.000	21:36	
Subject Sample	# 1	21:39	Subject Test Refused
Breath Volume	0.000	21:39	
Air Blank	0.000	21:39	
Control Test	0.076	21:40	
Air Blank	0.000	21:40	
Diagnostics Check	OK	21:40	

---

Data Download Date/Time: 11/11/2020 09:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DETRANO,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: DETRANO,JEFFREY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: DOUGHERTY

DATE/TIME OF TEST: 11/05/2020 20:15:09

---

Diagnostics Check	OK	20:09
Air Blank	0.000	20:09
Control Test	0.078	20:10
Air Blank	0.000	20:10
Subject Sample # 1	0.143	20:10
Breath Volume	1.472	20:10
Air Blank	0.000	20:11
Air Blank	0.000	20:13
Subject Sample # 2	0.134	20:13
Breath Volume	2.210	20:13
Air Blank	0.000	20:14
Control Test	0.078	20:14
Air Blank	0.000	20:15
Diagnostics Check	OK	20:15

---

Data Download Date/Time: 11/11/2020 09:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUFF,ZACHARY,J  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: HUFF,ZACHARY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 11/06/2020 17:17:21

---

Diagnostics Check	OK	17:11
Air Blank	0.000	17:11
Control Test	0.078	17:11
Air Blank	0.000	17:12
Subject Sample # 1	0.018	17:12
Breath Volume	1.820	17:12
Air Blank	0.000	17:13
Air Blank	0.000	17:15
Subject Sample # 2	0.017	17:15
Breath Volume	1.156	17:15
Air Blank	0.000	17:16
Control Test	0.078	17:16
Air Blank	0.000	17:17
Diagnostics Check	OK	17:17

---

Data Download Date/Time: 11/11/2020 09:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DETRANO,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: DETRANO,JEFFREY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MERRILL

DATE/TIME OF TEST: 11/07/2020 02:39:08

---

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:36	
Control Test	0.078	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:37	Subject Test Refused
Breath Volume	0.000	02:37	
Air Blank	0.000	02:38	
Control Test	0.077	02:38	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:39	

---

Data Download Date/Time: 11/11/2020 09:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOLBICKA,LAWRENCE,R  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ELIAS,ELIAS  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 07:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: COLLIER

DATE/TIME OF TEST: 11/07/2020 09:01:25

---

Diagnostics Check	OK	08:55
Air Blank	0.000	08:55
Control Test	0.077	08:56
Air Blank	0.000	08:56
Subject Sample # 1	0.114	08:57
Breath Volume	2.734	08:57
Air Blank	0.000	08:57
Air Blank	0.000	08:59
Subject Sample # 2	0.115	08:59
Breath Volume	2.148	08:59
Air Blank	0.000	09:00
Control Test	0.079	09:00
Air Blank	0.000	09:01
Diagnostics Check	OK	09:01

---

Data Download Date/Time: 11/11/2020 09:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROLLESTON,JACOB,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ROLLESTON,JACOB  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/07/2020 23:08:01

---

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.077	22:59
Air Blank	0.000	23:00
Subject Sample # 1	0.211	23:03
Breath Volume	1.210	23:03
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.207	23:06
Breath Volume	1.601	23:06
Air Blank	0.000	23:07
Control Test	0.077	23:07
Air Blank	0.000	23:07
Diagnostics Check	OK	23:08

---

Data Download Date/Time: 11/11/2020 09:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROLLESTON,JACOB,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ROLLESTON,JACOB  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 11/08/2020 00:34:13

---

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:32	
Control Test	0.076	00:32	
Air Blank	0.000	00:32	
Subject Sample	# 1	00:33	Sequence Aborted
Breath Volume	0.000	00:33	
Air Blank	0.000	00:34	

---

Data Download Date/Time: 11/11/2020 09:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROLLESTON,JACOB,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ROLLESTON,JACOB  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 11/08/2020 01:01:35

---

Diagnostics Check	OK	00:55
Air Blank	0.000	00:55
Control Test	0.078	00:55
Air Blank	0.000	00:56
Subject Sample # 1	0.155	00:57
Breath Volume	1.570	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.142	01:00
Breath Volume	1.875	01:00
Air Blank	0.000	01:00
Control Test	0.077	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

---

Data Download Date/Time: 11/11/2020 09:08



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROLLESTON,JACOB,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: CALDWELL,GARY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 10:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:26

SUBJECT'S LAST NAME: AVERY

DATE/TIME OF TEST: 11/11/2020 10:47:54

---

Diagnostics Check	OK	10:46	
Air Blank	0.000	10:47	
Control Test	0.074	10:47	Control Outside Tolerance
Air Blank	0.000	10:47	

---

Data Download Date/Time: 12/16/2020 16:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROLLESTON,JACOB,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: CALDWELL,GARY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 10:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:26

SUBJECT'S LAST NAME: AVERY

DATE/TIME OF TEST: 11/11/2020 10:55:33

---

Diagnostics Check	OK	10:49
Air Blank	0.000	10:49
Control Test	0.075	10:49
Air Blank	0.000	10:50
Subject Sample # 1	0.229	10:51
Breath Volume	3.722	10:51
Air Blank	0.000	10:51
Air Blank	0.000	10:53
Subject Sample # 2	0.238	10:54
Breath Volume	4.007	10:54
Air Blank	0.000	10:54
Control Test	0.077	10:54
Air Blank	0.000	10:55
Diagnostics Check	OK	10:55

---

Data Download Date/Time: 12/16/2020 16:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMPKINS,BENJAMIN,F  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: HICKS,J  
ARRESTING OFFICER AGENCY: Indian Shores PD  
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: DUROSS

DATE/TIME OF TEST: 11/17/2020 02:43:04

---

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:35	
Control Test	0.079	02:36	
Air Blank	0.000	02:36	
Subject Sample	# 1 0.128	02:37	
Breath Volume	1.410	02:37	
Air Blank	0.000	02:38	
Air Blank	0.000	02:40	
Subject Sample	# 2 0.140	02:41	
Breath Volume	2.937	02:41	
Air Blank	0.000	02:42	
Control Test	0.081	02:42	Interferent Detect
Air Blank	0.000	02:43	

---

Data Download Date/Time: 12/16/2020 16:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMPKINS,BENJAMIN,F  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: HICKS,J  
ARRESTING OFFICER AGENCY: Indian Shores PD  
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: DUROSS

DATE/TIME OF TEST: 11/17/2020 02:51:30

---

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.077	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.143	02:46
Breath Volume	2.035	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.142	02:50
Breath Volume	2.828	02:50
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

---

Data Download Date/Time: 12/16/2020 16:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DETRANO,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: DETRANO,DETRANO  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: CHRISTIAN

DATE/TIME OF TEST: 11/19/2020 19:24:45

---

Diagnostics Check	OK	19:18
Air Blank	0.000	19:19
Control Test	0.078	19:19
Air Blank	0.000	19:19
Subject Sample # 1	0.000	19:20
Breath Volume	1.804	19:20
Air Blank	0.000	19:20
Air Blank	0.000	19:22
Subject Sample # 2	0.000	19:23
Breath Volume	2.410	19:23
Air Blank	0.000	19:23
Control Test	0.078	19:24
Air Blank	0.000	19:24
Diagnostics Check	OK	19:24

---

Data Download Date/Time: 12/16/2020 16:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROLLESTON,JACOB,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ROLLESTON,JACOB  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 11/21/2020 00:50:11

---

Diagnostics Check	OK	00:43
Air Blank	0.000	00:44
Control Test	0.077	00:44
Air Blank	0.000	00:44
Subject Sample # 1	0.105	00:45
Breath Volume	1.488	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.101	00:48
Breath Volume	1.640	00:48
Air Blank	0.000	00:49
Control Test	0.077	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

---

Data Download Date/Time: 12/16/2020 16:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAUGHETEE,THERESA,J  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: DAUGHETEE,THERESA  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/22/2020 04:11:26

---

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.078	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.113	04:07
Breath Volume	1.808	04:07
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.105	04:10
Breath Volume	2.878	04:10
Air Blank	0.000	04:10
Control Test	0.077	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

---

Data Download Date/Time: 12/16/2020 16:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUCHUK,J,A  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: MARCHIONE,B  
ARRESTING OFFICER AGENCY: Kenneth City PD  
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HILTEBEITEL

DATE/TIME OF TEST: 11/24/2020 04:51:42

---

Diagnostics Check	OK	04:48	
Air Blank	0.000	04:49	
Control Test	0.078	04:49	
Air Blank	0.000	04:50	
Subject Sample	# 1	04:50	Subject Test Refused
Breath Volume	0.000	04:50	
Air Blank	0.000	04:50	
Control Test	0.077	04:51	
Air Blank	0.000	04:51	
Diagnostics Check	OK	04:51	

---

Data Download Date/Time: 12/16/2020 16:05



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DETRANO,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: DETRANO,JEFFREY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: REY

DATE/TIME OF TEST: 11/26/2020 04:11:36

---

Diagnostics Check		OK	04:05
Air Blank		0.000	04:05
Control Test		0.077	04:05
Air Blank		0.000	04:06
Subject Sample	# 1	0.206	04:07
Breath Volume		2.535	04:07
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.214	04:10
Breath Volume		3.195	04:10
Air Blank		0.000	04:10
Control Test		0.078	04:11
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

---

Data Download Date/Time: 12/16/2020 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DETRANO,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: DETRANO,JEFFREY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: HUTCHINGS

DATE/TIME OF TEST: 11/28/2020 03:37:21

---

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:35	
Control Test	0.078	03:35	
Air Blank	0.000	03:35	
Subject Sample	# 1	03:36	Subject Test Refused
Breath Volume	0.000	03:36	
Air Blank	0.000	03:36	
Control Test	0.078	03:36	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

---

Data Download Date/Time: 12/16/2020 16:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,  
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf  
ARRESTING OFFICER: ESCOBAR,ALEJANDRO  
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf  
LAST AGENCY INSPECTION: 10/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094981  
DRY GAS EXPIRY DATE: 02/22/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 11/03/2020 21:32:57

---

Diagnostics Check	OK	21:26
Air Blank	0.000	21:27
Control Test	0.079	21:27
Air Blank	0.000	21:28
Subject Sample # 1	0.204	21:28
Breath Volume	2.375	21:28
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.197	21:31
Breath Volume	2.500	21:31
Air Blank	0.000	21:32
Control Test	0.079	21:32
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

---

Data Download Date/Time: 11/25/2020 13:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,  
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf  
ARRESTING OFFICER: ESCOBAR,ALEJANDRO  
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf  
LAST AGENCY INSPECTION: 10/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094981  
DRY GAS EXPIRY DATE: 02/22/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: DONNELLY

DATE/TIME OF TEST: 11/04/2020 01:09:48

---

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:07	
Control Test	0.080	01:07	
Air Blank	0.000	01:08	
Subject Sample	# 1	01:08	Subject Test Refused
Breath Volume	0.000	01:08	
Air Blank	0.000	01:08	
Control Test	0.080	01:09	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:09	

---

Data Download Date/Time: 11/25/2020 13:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY,GAVIN,O  
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf  
ARRESTING OFFICER: HENRY,GAVIN  
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf  
LAST AGENCY INSPECTION: 10/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094981  
DRY GAS EXPIRY DATE: 02/22/2022  
VIOLATION CODE: DUI  
ARREST TIME: 14:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:06

SUBJECT'S LAST NAME: CHUNG

DATE/TIME OF TEST: 11/04/2020 15:40:12

---

Diagnostics Check	OK	15:34
Air Blank	0.000	15:34
Control Test	0.077	15:34
Air Blank	0.000	15:35
Subject Sample # 1	0.247	15:35
Breath Volume	2.566	15:35
Air Blank	0.000	15:36
Air Blank	0.000	15:38
Subject Sample # 2	0.242	15:38
Breath Volume	1.722	15:38
Air Blank	0.000	15:39
Control Test	0.076	15:39
Air Blank	0.000	15:40
Diagnostics Check	OK	15:40

---

Data Download Date/Time: 11/25/2020 13:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,  
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf  
ARRESTING OFFICER: ESCOBAR,ALEJANDRO  
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf  
LAST AGENCY INSPECTION: 10/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094981  
DRY GAS EXPIRY DATE: 02/22/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: MOSLEY

DATE/TIME OF TEST: 11/11/2020 02:42:06

---

Diagnostics Check		OK	02:30	
Air Blank		0.000	02:31	
Control Test		0.079	02:31	
Air Blank		0.000	02:31	
Subject Sample	# 1	0.169	02:35	Volume Not Met
Breath Volume		0.832	02:35	
Air Blank		0.000	02:35	
Air Blank		0.000	02:37	
Subject Sample	# 2	0.139	02:40	Volume Not Met
Breath Volume		0.453	02:40	
Air Blank		0.000	02:41	
Control Test		0.079	02:41	
Air Blank		0.000	02:42	
Diagnostics Check		OK	02:42	

---

Data Download Date/Time: 11/25/2020 13:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,DONALD,P  
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf  
ARRESTING OFFICER: ESCOBAR,ALEJANDRO  
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf  
LAST AGENCY INSPECTION: 10/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094981  
DRY GAS EXPIRY DATE: 02/22/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: MOSLEY

DATE/TIME OF TEST: 11/11/2020 02:56:26

---

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.078	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.185	02:51
Breath Volume	1.171	02:51
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.189	02:54
Breath Volume	1.800	02:54
Air Blank	0.000	02:55
Control Test	0.079	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

---

Data Download Date/Time: 11/25/2020 13:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,  
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf  
ARRESTING OFFICER: ESCOBAR,ALEJANDRO  
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf  
LAST AGENCY INSPECTION: 10/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094981  
DRY GAS EXPIRY DATE: 02/22/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: HARROLD

DATE/TIME OF TEST: 11/12/2020 01:24:26

---

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:22	
Control Test	0.079	01:22	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:23	Subject Test Refused
Breath Volume	0.000	01:23	
Air Blank	0.000	01:23	
Control Test	0.079	01:23	
Air Blank	0.000	01:24	
Diagnostics Check	OK	01:24	

---

Data Download Date/Time: 11/25/2020 13:32



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,  
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf  
ARRESTING OFFICER: ESCOBAR,ALEJANDRO  
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf  
LAST AGENCY INSPECTION: 10/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094981  
DRY GAS EXPIRY DATE: 02/22/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: GAGER

DATE/TIME OF TEST: 11/16/2020 22:41:14

---

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.079	22:35
Air Blank	0.000	22:36
Subject Sample # 1	0.200	22:36
Breath Volume	2.132	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.208	22:39
Breath Volume	2.136	22:39
Air Blank	0.000	22:40
Control Test	0.079	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

---

Data Download Date/Time: 11/25/2020 13:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TUCKER,R,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LAWSON,N  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/17/2020 01:39:27

---

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.082	01:33
Air Blank		0.000	01:34
Subject Sample	# 1	0.078	01:34
Breath Volume		2.203	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.076	01:37
Breath Volume		2.488	01:37
Air Blank		0.000	01:38
Control Test		0.082	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

---

Data Download Date/Time: 11/17/2020 10:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,CHUCK,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: STANLEY,T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 11/18/2020 01:57:16

---

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:54	
Control Test	0.080	01:54	
Air Blank	0.000	01:55	
Subject Sample	# 1	01:55	Subject Test Refused
Breath Volume	0.000	01:55	
Air Blank	0.000	01:56	
Control Test	0.080	01:56	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

---

Data Download Date/Time: 12/09/2020 12:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,CHUCK,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LUGO,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: LACHANCE

DATE/TIME OF TEST: 11/19/2020 01:39:27

---

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.081	01:33
Air Blank	0.000	01:34
Subject Sample # 1	0.116	01:34
Breath Volume	1.519	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.112	01:37
Breath Volume	1.699	01:37
Air Blank	0.000	01:38
Control Test	0.081	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

---

Data Download Date/Time: 12/09/2020 12:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,CHUCK,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JORDAN,A  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: FIELD

DATE/TIME OF TEST: 11/19/2020 22:45:08

---

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:42	
Control Test	0.081	22:43	
Air Blank	0.000	22:43	
Subject Sample	# 1	22:43	Subject Test Refused
Breath Volume	0.000	22:43	
Air Blank	0.000	22:44	
Control Test	0.080	22:44	
Air Blank	0.000	22:45	
Diagnostics Check	OK	22:45	

---

Data Download Date/Time: 12/09/2020 12:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,CHUCK,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KEEN,C  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: MASHBURN

DATE/TIME OF TEST: 11/20/2020 03:21:55

---

Diagnostics Check		OK	03:15
Air Blank		0.000	03:15
Control Test		0.081	03:16
Air Blank		0.000	03:16
Subject Sample	# 1	0.122	03:17
Breath Volume		1.718	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:19
Subject Sample	# 2	0.128	03:20
Breath Volume		1.953	03:20
Air Blank		0.000	03:20
Control Test		0.080	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

---

Data Download Date/Time: 12/09/2020 12:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,CHUCK,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: ONDRIEZEK,B  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/20/2020 04:25:35

---

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.081	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.164	04:20
Breath Volume	1.121	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.155	04:23
Breath Volume	1.753	04:23
Air Blank	0.000	04:24
Control Test	0.080	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

---

Data Download Date/Time: 12/09/2020 12:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOVACS, TYLER, K  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HAWKINS, A  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 11/21/2020 01:20:34

---

Diagnostics Check	OK	01:13
Air Blank	0.000	01:14
Control Test	0.081	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.118	01:15
Breath Volume	3.039	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.118	01:18
Breath Volume	3.117	01:18
Air Blank	0.000	01:19
Control Test	0.081	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

---

Data Download Date/Time: 12/09/2020 12:03



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TUCKER,R,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 11/21/2020 23:51:27

---

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:48	
Control Test	0.082	23:48	
Air Blank	0.000	23:49	
Subject Sample	# 1 0.188	23:49	
Breath Volume	1.722	23:49	
Air Blank		23:50	RFI Detect
Air Blank	0.000	23:50	
Control Test	0.081	23:50	
Air Blank	0.000	23:51	
Diagnostics Check	OK	23:51	

---

Data Download Date/Time: 12/09/2020 12:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TUCKER,R,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 11/22/2020 00:06:37

---

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:54	
Control Test	0.080	23:54	
Air Blank	0.000	23:55	
Subject Sample	# 3 0.174	23:56	
Breath Volume	1.234	23:56	
Air Blank	0.000	23:56	
Air Blank	0.000	23:58	
Subject Sample	# 1 0.163	00:01	Slope Not Level
Breath Volume	1.171	00:01	
Air Blank	0.000	00:02	
Air Blank	0.000	00:04	
Subject Sample	# 2 0.173	00:05	
Breath Volume	1.132	00:05	
Air Blank	0.000	00:05	
Control Test	0.080	00:06	
Air Blank	0.000	00:06	
Diagnostics Check	OK	00:06	

---

Data Download Date/Time: 12/09/2020 12:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TUCKER,R,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: PEREZ,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 11/22/2020 03:54:43

---

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.081	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.148	03:50
Breath Volume	3.476	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.145	03:53
Breath Volume	3.347	03:53
Air Blank	0.000	03:53
Control Test	0.081	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

---

Data Download Date/Time: 12/09/2020 12:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TUCKER,R,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KEEN,C  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: ERBE

DATE/TIME OF TEST: 11/22/2020 05:04:59

---

Diagnostics Check	OK	05:02	
Air Blank	0.000	05:02	
Control Test	0.081	05:02	
Air Blank	0.000	05:03	
Subject Sample	# 1	05:03	Subject Test Refused
Breath Volume	0.000	05:03	
Air Blank	0.000	05:04	
Control Test	0.082	05:04	
Air Blank	0.000	05:04	
Diagnostics Check	OK	05:04	

---

Data Download Date/Time: 12/09/2020 12:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TUCKER,R,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/23/2020 03:26:46

---

Diagnostics Check	OK	03:19
Air Blank	0.000	03:20
Control Test	0.082	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.111	03:22
Breath Volume	2.953	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.117	03:25
Breath Volume	2.941	03:25
Air Blank	0.000	03:25
Control Test	0.080	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

---

Data Download Date/Time: 12/09/2020 12:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TUCKER,R,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNS,B  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:59

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/23/2020 06:25:35

---

Diagnostics Check	OK	06:19
Air Blank	0.000	06:20
Control Test	0.082	06:20
Air Blank	0.000	06:20
Subject Sample # 1	0.000	06:21
Breath Volume	2.519	06:21
Air Blank	0.000	06:21
Air Blank	0.000	06:23
Subject Sample # 2	0.000	06:24
Breath Volume	2.187	06:24
Air Blank	0.000	06:24
Control Test	0.081	06:24
Air Blank	0.000	06:25
Diagnostics Check	OK	06:25

---

Data Download Date/Time: 12/09/2020 12:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TUCKER,R,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNS,B  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 11/24/2020 01:35:52

---

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.082	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.000	01:31
Breath Volume	3.148	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.000	01:34
Breath Volume	3.832	01:34
Air Blank	0.000	01:34
Control Test	0.082	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

---

Data Download Date/Time: 12/09/2020 12:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOYOS,D,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: THOMAS,J  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:54

SUBJECT'S LAST NAME: KABWOGI

DATE/TIME OF TEST: 11/24/2020 16:18:23

---

Diagnostics Check	OK	16:15	
Air Blank	0.000	16:15	
Control Test	0.082	16:15	
Air Blank	0.000	16:16	
Subject Sample	# 1	16:17	Subject Test Refused
Breath Volume	0.000	16:17	
Air Blank	0.000	16:17	
Control Test	0.081	16:17	
Air Blank	0.000	16:18	
Diagnostics Check	OK	16:18	

---

Data Download Date/Time: 12/09/2020 12:04



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TUCKER,R,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JENSEN,D  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: PEEPLES

DATE/TIME OF TEST: 11/26/2020 02:02:04

---

Diagnostics Check	OK	01:59	
Air Blank	0.000	01:59	
Control Test	0.082	01:59	
Air Blank	0.000	02:00	
Subject Sample	# 1	02:00	Subject Test Refused
Breath Volume	0.000	02:00	
Air Blank	0.000	02:01	
Control Test	0.081	02:01	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:02	

---

Data Download Date/Time: 12/09/2020 12:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TUCKER,R,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MORGAN,Z  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: WILEY

DATE/TIME OF TEST: 11/26/2020 05:33:44

---

Diagnostics Check	OK	05:27
Air Blank	0.000	05:28
Control Test	0.081	05:28
Air Blank	0.000	05:28
Subject Sample # 1	0.111	05:29
Breath Volume	1.195	05:29
Air Blank	0.000	05:29
Air Blank	0.000	05:31
Subject Sample # 2	0.115	05:32
Breath Volume	2.019	05:32
Air Blank	0.000	05:32
Control Test	0.081	05:33
Air Blank	0.000	05:33
Diagnostics Check	OK	05:33

---

Data Download Date/Time: 12/09/2020 12:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAY,L,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DUBET,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 07:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:46

SUBJECT'S LAST NAME: HAGI

DATE/TIME OF TEST: 11/29/2020 09:12:35

---

Diagnostics Check		OK	09:06
Air Blank		0.000	09:06
Control Test		0.082	09:06
Air Blank		0.000	09:07
Subject Sample	# 1	0.157	09:08
Breath Volume		1.113	09:08
Air Blank		0.000	09:08
Air Blank		0.000	09:10
Subject Sample	# 2	0.153	09:10
Breath Volume		2.355	09:10
Air Blank		0.000	09:11
Control Test		0.081	09:11
Air Blank		0.000	09:12
Diagnostics Check		OK	09:12

---

Data Download Date/Time: 12/09/2020 12:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAYE, MAKINA, A  
BREATH TEST OPERATOR AGENCY: Coral Springs PD  
ARRESTING OFFICER: CLARKE, S  
ARRESTING OFFICER AGENCY: Coral Springs PD  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: ALT

DATE/TIME OF TEST: 11/07/2020 20:41:41

---

Diagnostics Check	OK	20:38	
Air Blank	0.000	20:39	
Control Test	0.082	20:39	
Air Blank	0.000	20:40	
Subject Sample	# 1	20:40	Subject Test Refused
Breath Volume	0.000	20:40	
Air Blank	0.000	20:40	
Control Test	0.081	20:41	
Air Blank	0.000	20:41	
Diagnostics Check	OK	20:41	

---

Data Download Date/Time: 11/27/2020 09:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAFFERTY,NAOMI,E  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: MCCULLOUGH,SEAN  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/01/2020 01:26:29

---

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.080	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.000	01:22
Breath Volume	2.851	01:22
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.000	01:25
Breath Volume	1.347	01:25
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

---

Data Download Date/Time: 11/25/2020 15:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUM,ZACKARY,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: HINES,PAUL  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 11/01/2020 03:36:08

---

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:28	
Control Test	0.079	03:28	
Air Blank	0.000	03:28	
Subject Sample	# 1	03:31	No Sample Provided
Breath Volume	0.000	03:31	
Air Blank	0.000	03:32	
Air Blank	0.000	03:34	
Subject Sample	# 2	03:34	Subject Test Refused
Breath Volume	0.000	03:34	
Air Blank	0.000	03:35	
Control Test	0.079	03:35	
Air Blank	0.000	03:36	
Diagnostics Check	OK	03:36	

---

Data Download Date/Time: 11/25/2020 15:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,EMILY,K  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: BRAGANO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: COOLEY

DATE/TIME OF TEST: 11/01/2020 20:27:26

---

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.079	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.190	20:22
Breath Volume	2.324	20:22
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.196	20:25
Breath Volume	2.277	20:25
Air Blank	0.000	20:26
Control Test	0.079	20:26
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

---

Data Download Date/Time: 11/25/2020 15:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAUSAM,MATTHEW,L  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: THOMAS,K.  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: ZELL

DATE/TIME OF TEST: 11/01/2020 23:57:21

---

Diagnostics Check	OK	23:54	
Air Blank	0.000	23:54	
Control Test	0.080	23:55	
Air Blank	0.000	23:55	
Subject Sample	# 1	23:55	Subject Test Refused
Breath Volume	0.000	23:55	
Air Blank	0.000	23:56	
Control Test	0.080	23:56	
Air Blank	0.000	23:57	
Diagnostics Check	OK	23:57	

---

Data Download Date/Time: 11/25/2020 15:50



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINES,PAUL,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: CRUM,ZACKARY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: BUTTICE

DATE/TIME OF TEST: 11/05/2020 01:51:17

---

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.163	01:46
Breath Volume	1.753	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.162	01:49
Breath Volume	1.375	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

---

Data Download Date/Time: 11/25/2020 15:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINES,PAUL,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: MCCULLOUGH,SEAN  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: MACHOLZ

DATE/TIME OF TEST: 11/06/2020 22:37:03

---

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:26	
Control Test	0.080	22:26	
Air Blank	0.000	22:26	
Subject Sample	# 1	22:30	No Sample Provided
Breath Volume	0.000	22:30	
Air Blank	0.000	22:30	
Air Blank	0.000	22:32	
Subject Sample	# 2	22:35	No Sample Provided
Breath Volume	0.000	22:35	
Air Blank	0.000	22:36	
Control Test	0.079	22:36	
Air Blank	0.000	22:36	
Diagnostics Check	OK	22:37	

---

Data Download Date/Time: 11/25/2020 15:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUM,ZACKARY,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: LAMOUR,Z  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 11/07/2020 00:44:56

---

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:40	
Control Test	0.080	00:40	
Air Blank	0.000	00:41	
Subject Sample	# 1	00:43	Subject Test Refused
Breath Volume	0.000	00:43	
Air Blank	0.000	00:44	
Control Test	0.079	00:44	
Air Blank	0.000	00:44	
Diagnostics Check	OK	00:44	

---

Data Download Date/Time: 11/25/2020 15:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBISON,CASEY,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: DEMARIA,,JAMES  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: INTILE

DATE/TIME OF TEST: 11/08/2020 00:33:26

---

Diagnostics Check		OK	00:21	
Air Blank		0.000	00:21	
Control Test		0.079	00:22	
Air Blank		0.000	00:22	
Subject Sample	# 1		00:25	No Sample Provided
Breath Volume		0.000	00:25	
Air Blank		0.000	00:26	
Air Blank		0.000	00:28	
Subject Sample	# 2	0.197	00:28	No .020 agreement
Breath Volume		1.292	00:28	
Air Blank		0.000	00:29	
Air Blank		0.000	00:31	
Subject Sample	# 3	0.224	00:31	No .020 agreement
Breath Volume		1.558	00:31	
Air Blank		0.000	00:32	
Control Test		0.077	00:32	
Air Blank		0.000	00:33	
Diagnostics Check		OK	00:33	

---

Data Download Date/Time: 11/25/2020 15:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBISON,CASEY,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: DEMARIA,,JAMES  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: INTILE

DATE/TIME OF TEST: 11/08/2020 01:17:29

---

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:06	
Control Test	0.079	01:06	
Air Blank	0.000	01:07	
Subject Sample	# 1	01:10	No Sample Provided
Breath Volume	0.000	01:10	
Air Blank	0.000	01:10	
Air Blank	0.000	01:12	
Subject Sample	# 2	01:16	No Sample Provided
Breath Volume	0.000	01:16	
Air Blank	0.000	01:16	
Control Test	0.080	01:16	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

---

Data Download Date/Time: 11/25/2020 15:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MIRANDA,RICHARD,  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: MARKUS,ROBERT  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: BIROS

DATE/TIME OF TEST: 11/08/2020 21:39:21

---

Diagnostics Check		OK	21:32
Air Blank		0.000	21:32
Control Test		0.080	21:32
Air Blank		0.000	21:33
Subject Sample	# 1	0.235	21:34
Breath Volume		2.210	21:34
Air Blank		0.000	21:35
Air Blank		0.000	21:37
Subject Sample	# 2	0.231	21:37
Breath Volume		2.589	21:37
Air Blank		0.000	21:38
Control Test		0.079	21:38
Air Blank		0.000	21:39
Diagnostics Check		OK	21:39

---

Data Download Date/Time: 11/25/2020 15:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLUTO,KEVIN,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: BUTTERFIED,KELLEY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: BLUTO

DATE/TIME OF TEST: 11/10/2020 19:57:28

---

Diagnostics Check	OK	19:56	
Air Blank		19:56	Sequence Aborted
Air Blank	0.000	19:57	

---

Data Download Date/Time: 11/25/2020 15:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: BUTTERFIELD,KELLEY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: BLUTO

DATE/TIME OF TEST: 11/10/2020 20:02:58

---

Diagnostics Check	OK	19:59	
Air Blank	0.000	19:59	
Control Test	0.079	20:00	
Air Blank	0.000	20:00	
Subject Sample	# 1	20:01	Subject Test Refused
Breath Volume	0.000	20:01	
Air Blank	0.000	20:02	
Control Test	0.078	20:02	
Air Blank	0.000	20:02	
Diagnostics Check	OK	20:02	

---

Data Download Date/Time: 11/25/2020 15:51



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUM,ZACKARY,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: HINES,PAUL  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: SHANKLIN

DATE/TIME OF TEST: 11/12/2020 21:19:57

---

Diagnostics Check	OK	21:17	
Air Blank	0.000	21:17	
Control Test	0.079	21:17	
Air Blank	0.000	21:18	
Subject Sample	# 1	21:18	Subject Test Refused
Breath Volume	0.000	21:18	
Air Blank	0.000	21:19	
Control Test	0.079	21:19	
Air Blank	0.000	21:19	
Diagnostics Check	OK	21:19	

---

Data Download Date/Time: 11/25/2020 15:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUM,ZACKARY,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: HINES,PAUL  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: GAY

DATE/TIME OF TEST: 11/13/2020 01:18:31

---

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:15	
Control Test	0.080	01:16	
Air Blank	0.000	01:16	
Subject Sample	# 1	01:17	Subject Test Refused
Breath Volume	0.000	01:17	
Air Blank	0.000	01:17	
Control Test	0.079	01:17	
Air Blank	0.000	01:18	
Diagnostics Check	OK	01:18	

---

Data Download Date/Time: 11/25/2020 15:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINES,PAUL,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: LAMOUR,ZACHARY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: CHUBB

DATE/TIME OF TEST: 11/13/2020 23:45:31

---

Diagnostics Check		OK	23:29	
Air Blank		0.000	23:29	
Control Test		0.079	23:30	
Air Blank		0.000	23:30	
Subject Sample	# 1		23:33	No Sample Provided
Breath Volume		0.000	23:33	
Air Blank		0.000	23:34	
Air Blank		0.000	23:36	
Subject Sample	# 2	0.262	23:38	
Breath Volume		1.636	23:38	
Air Blank		0.000	23:39	
Air Blank		0.000	23:40	
Subject Sample	# 3	0.230	23:43	Volume Not Met
Breath Volume		0.132	23:43	
Air Blank		0.000	23:44	
Control Test		0.078	23:44	
Air Blank		0.000	23:45	
Diagnostics Check		OK	23:45	

---

Data Download Date/Time: 11/25/2020 15:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINES,PAUL,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: LAMOUR,ZACHARY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: CHUBB

DATE/TIME OF TEST: 11/14/2020 00:01:41

---

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:47	
Control Test	0.078	23:47	
Air Blank	0.000	23:48	
Subject Sample # 2	0.272	23:49	
Breath Volume	2.597	23:49	
Air Blank	0.000	23:49	
Air Blank	0.000	23:51	
Subject Sample # 3	0.263	23:54	Volume Not Met
Breath Volume	0.273	23:54	
Air Blank	0.000	23:55	
Air Blank	0.000	23:57	
Subject Sample # 1	0.262	00:00	
Breath Volume	1.648	00:00	
Air Blank	0.000	00:00	
Control Test	0.077	00:01	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

---

Data Download Date/Time: 11/25/2020 15:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINES,PAUL,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: MCCULLOUGH,SEAN  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: BEAM

DATE/TIME OF TEST: 11/14/2020 03:25:22

---

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.078	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.133	03:20
Breath Volume		4.316	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.116	03:23
Breath Volume		1.867	03:23
Air Blank		0.000	03:24
Control Test		0.079	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

---

Data Download Date/Time: 11/25/2020 15:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINES,PAUL,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: LAMOUR,ZACHARY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 11/14/2020 23:31:40

---

Diagnostics Check	OK	23:28	
Air Blank	0.000	23:29	
Control Test	0.078	23:29	
Air Blank	0.000	23:29	
Subject Sample	# 1	23:30	Subject Test Refused
Breath Volume	0.000	23:30	
Air Blank	0.000	23:30	
Control Test	0.079	23:31	
Air Blank	0.000	23:31	
Diagnostics Check	OK	23:31	

---

Data Download Date/Time: 11/25/2020 15:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: MCCULLOUGH,SEAN  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: MACEACHERN

DATE/TIME OF TEST: 11/15/2020 01:30:47

---

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:27	
Control Test	0.079	01:28	
Air Blank	0.000	01:28	
Subject Sample	# 1	01:29	Subject Test Refused
Breath Volume	0.000	01:29	
Air Blank	0.000	01:29	
Control Test	0.080	01:30	
Air Blank	0.000	01:30	
Diagnostics Check	OK	01:30	

---

Data Download Date/Time: 11/25/2020 15:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAUSAM, MATTHEW, L  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: PEREZ, S.  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: ESPOSITO

DATE/TIME OF TEST: 11/17/2020 02:55:31

---

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:51	
Control Test	0.079	02:51	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:54	Subject Test Refused
Breath Volume	0.000	02:54	
Air Blank	0.000	02:54	
Control Test	0.079	02:54	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

---

Data Download Date/Time: 11/25/2020 15:51



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHWEIGER, JONATHAN, A  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: BISHOP, CHRISTOPHER  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: BRETT

DATE/TIME OF TEST: 11/17/2020 05:21:47

---

Diagnostics Check		OK	05:15
Air Blank		0.000	05:15
Control Test		0.080	05:15
Air Blank		0.000	05:16
Subject Sample	# 1	0.164	05:17
Breath Volume		2.257	05:17
Air Blank		0.000	05:17
Air Blank		0.000	05:19
Subject Sample	# 2	0.172	05:20
Breath Volume		2.832	05:20
Air Blank		0.000	05:20
Control Test		0.081	05:21
Air Blank		0.000	05:21
Diagnostics Check		OK	05:21

---

Data Download Date/Time: 11/25/2020 15:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK,BRENT,D  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: PIERCE,SCOTT  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 12:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: LAWSON

DATE/TIME OF TEST: 11/17/2020 13:32:41

---

Diagnostics Check	OK	13:25
Air Blank	0.000	13:26
Control Test	0.080	13:26
Air Blank	0.000	13:26
Subject Sample # 1	0.000	13:27
Breath Volume	2.089	13:27
Air Blank	0.000	13:28
Air Blank	0.000	13:30
Subject Sample # 2	0.000	13:31
Breath Volume	2.492	13:31
Air Blank	0.000	13:31
Control Test	0.080	13:32
Air Blank	0.000	13:32
Diagnostics Check	OK	13:32

---

Data Download Date/Time: 11/25/2020 15:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: HINES,PAUL  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: DORSEY

DATE/TIME OF TEST: 11/17/2020 18:49:07

---

Diagnostics Check	OK	18:42
Air Blank	0.000	18:43
Control Test	0.079	18:43
Air Blank	0.000	18:44
Subject Sample # 1	0.245	18:44
Breath Volume	2.503	18:44
Air Blank	0.000	18:45
Air Blank	0.000	18:47
Subject Sample # 2	0.246	18:47
Breath Volume	2.167	18:47
Air Blank	0.000	18:48
Control Test	0.078	18:48
Air Blank	0.000	18:49
Diagnostics Check	OK	18:49

---

Data Download Date/Time: 11/25/2020 15:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: HINES,PAUL  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: LOWERY

DATE/TIME OF TEST: 11/17/2020 20:46:21

---

Diagnostics Check	OK	20:43	
Air Blank	0.000	20:43	
Control Test	0.079	20:44	
Air Blank	0.000	20:44	
Subject Sample	# 1	20:44	Subject Test Refused
Breath Volume	0.000	20:44	
Air Blank	0.000	20:45	
Control Test	0.080	20:45	
Air Blank	0.000	20:46	
Diagnostics Check	OK	20:46	

---

Data Download Date/Time: 11/25/2020 15:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUM,ZACKARY,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: HINES,PAUL  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: PURSER

DATE/TIME OF TEST: 11/17/2020 22:59:10

---

Diagnostics Check		OK	22:52
Air Blank		0.000	22:52
Control Test		0.081	22:52
Air Blank		0.000	22:53
Subject Sample	# 1	0.047	22:54
Breath Volume		2.097	22:54
Air Blank		0.000	22:55
Air Blank		0.000	22:57
Subject Sample	# 2	0.044	22:57
Breath Volume		1.316	22:57
Air Blank		0.000	22:58
Control Test		0.080	22:58
Air Blank		0.000	22:59
Diagnostics Check		OK	22:59

---

Data Download Date/Time: 11/25/2020 15:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINES,PAUL,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: CRUM,ZACKARY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: JACKMAN

DATE/TIME OF TEST: 11/18/2020 22:52:38

---

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.080	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.091	22:48
Breath Volume	1.570	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.088	22:51
Breath Volume	1.113	22:51
Air Blank	0.000	22:51
Control Test	0.080	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

---

Data Download Date/Time: 11/25/2020 15:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBISON,CASEY,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: LAMOUR,ZACH  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: ATHANS

DATE/TIME OF TEST: 11/19/2020 02:40:38

---

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.079	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.155	02:35
Breath Volume	3.218	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.160	02:39
Breath Volume	3.050	02:39
Air Blank	0.000	02:39
Control Test	0.080	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

---

Data Download Date/Time: 11/25/2020 15:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBISON,CASEY,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: HINES,PAUL  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 11/19/2020 03:06:20

---

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.160	03:01
Breath Volume	2.253	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.158	03:04
Breath Volume	1.632	03:04
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

---

Data Download Date/Time: 11/25/2020 15:51



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBISON,CASEY,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: BISHOP,CHRISTOPHER  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: CHIADMI

DATE/TIME OF TEST: 11/19/2020 04:36:51

---

Diagnostics Check	OK	04:25	
Air Blank	0.000	04:25	
Control Test	0.080	04:26	
Air Blank	0.000	04:26	
Subject Sample	# 1	04:29	No Sample Provided
Breath Volume	0.000	04:29	
Air Blank	0.000	04:30	
Air Blank	0.000	04:32	
Subject Sample	# 2	04:35	No Sample Provided
Breath Volume	0.000	04:35	
Air Blank	0.000	04:35	
Control Test	0.080	04:36	
Air Blank	0.000	04:36	
Diagnostics Check	OK	04:36	

---

Data Download Date/Time: 11/25/2020 15:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SISTI,JORDAN,A  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: SAIA,M  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: KELLEY-CARTER

DATE/TIME OF TEST: 11/20/2020 04:04:11

---

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.080	03:56	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:59	No Sample Provided
Breath Volume	0.000	03:59	
Air Blank	0.000	04:00	
Air Blank	0.000	04:02	
Subject Sample	# 2	04:02	Subject Test Refused
Breath Volume	0.000	04:02	
Air Blank	0.000	04:03	
Control Test	0.080	04:03	
Air Blank	0.000	04:04	
Diagnostics Check	OK	04:04	

---

Data Download Date/Time: 11/25/2020 15:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: CLARK,BRENT  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 08:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:25

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/20/2020 08:52:10

---

Diagnostics Check	OK	08:45
Air Blank	0.000	08:46
Control Test	0.079	08:46
Air Blank	0.000	08:47
Subject Sample # 1	0.268	08:47
Breath Volume	2.734	08:47
Air Blank	0.000	08:48
Air Blank	0.000	08:50
Subject Sample # 2	0.270	08:50
Breath Volume	2.531	08:50
Air Blank	0.000	08:51
Control Test	0.078	08:51
Air Blank	0.000	08:52
Diagnostics Check	OK	08:52

---

Data Download Date/Time: 11/25/2020 15:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MIRANDA,RICHARD,  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: BUTTERFIELD,KELLEY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: BOLDEN

DATE/TIME OF TEST: 11/20/2020 17:54:17

---

Diagnostics Check	OK	17:47	
Air Blank	0.000	17:48	
Control Test	0.079	17:48	
Air Blank	0.000	17:49	
Subject Sample	# 1 0.373	17:49	
Breath Volume	3.558	17:49	
Air Blank	0.000	17:50	
Air Blank	0.000	17:52	
Subject Sample	# 2	17:52	Slope Not Met
Breath Volume	2.371	17:52	
Air Blank	0.000	17:53	
Control Test	0.078	17:53	
Air Blank	0.000	17:54	
Diagnostics Check	OK	17:54	

---

Data Download Date/Time: 11/25/2020 15:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MIRANDA,RICHARD,  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: BUTTERFIELD,KELLEY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: BOLDEN

DATE/TIME OF TEST: 11/20/2020 18:13:35

---

Diagnostics Check	OK	17:58	
Air Blank	0.000	17:58	
Control Test	0.079	17:58	
Air Blank	0.000	17:59	
Subject Sample	# 1 0.358	18:00	
Breath Volume	3.320	18:00	
Air Blank	0.000	18:01	
Air Blank	0.000	18:03	
Subject Sample	# 2	18:06	No Sample Provided
Breath Volume	0.000	18:06	
Air Blank	0.000	18:07	
Air Blank	0.000	18:09	
Subject Sample	# 3	18:12	No Sample Provided
Breath Volume	0.000	18:12	
Air Blank	0.000	18:12	
Control Test	0.079	18:12	
Air Blank	0.000	18:13	
Diagnostics Check	OK	18:13	

---

Data Download Date/Time: 11/25/2020 15:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUM,ZACKARY,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: HINES,P  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: GDALETA

DATE/TIME OF TEST: 11/21/2020 01:58:06

---

Diagnostics Check	OK	01:55	
Air Blank	0.000	01:55	
Control Test	0.080	01:55	
Air Blank	0.000	01:56	
Subject Sample	# 1	01:56	Subject Test Refused
Breath Volume	0.000	01:56	
Air Blank	0.000	01:57	
Control Test	0.079	01:57	
Air Blank	0.000	01:58	
Diagnostics Check	OK	01:58	

---

Data Download Date/Time: 11/25/2020 15:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINES,PAUL,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ROBISON,CASEY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: FENLEY

DATE/TIME OF TEST: 11/21/2020 03:02:47

---

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.095	02:58
Breath Volume		1.601	02:58
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.101	03:01
Breath Volume		1.230	03:01
Air Blank		0.000	03:01
Control Test		0.080	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

---

Data Download Date/Time: 11/25/2020 15:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAIA,MICHAEL,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: GODERICH,KRISTINA  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: BOLLMAN

DATE/TIME OF TEST: 11/22/2020 01:27:37

---

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:19	
Control Test	0.079	01:19	
Air Blank	0.000	01:20	
Subject Sample	# 1	01:23	No Sample Provided
Breath Volume	0.000	01:23	
Air Blank	0.000	01:23	
Air Blank	0.000	01:25	
Subject Sample	# 2	01:26	Subject Test Refused
Breath Volume	0.000	01:26	
Air Blank	0.000	01:26	
Control Test	0.079	01:27	
Air Blank	0.000	01:27	
Diagnostics Check	OK	01:27	

---

Data Download Date/Time: 11/25/2020 15:52



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBISON,CASEY,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ROBISON,CASEY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MULLIGAN

DATE/TIME OF TEST: 11/22/2020 01:43:39

---

Diagnostics Check		OK	01:36
Air Blank		0.000	01:37
Control Test		0.079	01:37
Air Blank		0.000	01:38
Subject Sample	# 1	0.322	01:39
Breath Volume		4.593	01:39
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.334	01:42
Breath Volume		4.460	01:42
Air Blank		0.000	01:42
Control Test		0.079	01:43
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

---

Data Download Date/Time: 11/25/2020 15:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBISON,CASEY,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: GODERICH,KRISTINA  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: GAETA

DATE/TIME OF TEST: 11/22/2020 04:32:13

---

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:21	
Control Test	0.080	04:21	
Air Blank	0.000	04:22	
Subject Sample	# 1	04:25	No Sample Provided
Breath Volume	0.000	04:25	
Air Blank	0.000	04:25	
Air Blank	0.000	04:27	
Subject Sample	# 2	04:30	No Sample Provided
Breath Volume	0.000	04:30	
Air Blank	0.000	04:31	
Control Test	0.080	04:31	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:32	

---

Data Download Date/Time: 11/25/2020 15:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHWEIGER, JONATHAN, A  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: THOMAS, KRISTEN  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 10/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: ZUBOR

DATE/TIME OF TEST: 11/22/2020 23:52:58

---

Diagnostics Check	OK	23:46
Air Blank	0.000	23:47
Control Test	0.079	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.157	23:48
Breath Volume	2.695	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:50
Subject Sample # 2	0.161	23:51
Breath Volume	2.890	23:51
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

---

Data Download Date/Time: 11/25/2020 15:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: LAFFERTY,NAOMI  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: HAWKSWORTH

DATE/TIME OF TEST: 11/25/2020 18:12:19

---

Diagnostics Check		OK	18:05
Air Blank		0.000	18:05
Control Test		0.079	18:06
Air Blank		0.000	18:06
Subject Sample	# 1	0.273	18:07
Breath Volume		2.023	18:07
Air Blank		0.000	18:08
Air Blank		0.000	18:10
Subject Sample	# 2	0.260	18:10
Breath Volume		1.632	18:10
Air Blank		0.000	18:11
Control Test		0.077	18:11
Air Blank		0.000	18:12
Diagnostics Check		OK	18:12

---

Data Download Date/Time: 12/29/2020 06:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBISON,CASEY,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: LAFFERTY,NAOMI  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: SPAFFORD

DATE/TIME OF TEST: 11/26/2020 01:29:38

---

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:18	
Control Test	0.079	01:18	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:22	No Sample Provided
Breath Volume	0.000	01:22	
Air Blank	0.000	01:23	
Air Blank	0.000	01:24	
Subject Sample	# 2	01:28	No Sample Provided
Breath Volume	0.000	01:28	
Air Blank	0.000	01:28	
Control Test	0.079	01:29	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

---

Data Download Date/Time: 12/29/2020 06:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KLINE,NICOLE,S  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: GODERICH,KRISTINA  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 11/27/2020 06:10:50

---

Diagnostics Check		OK	06:04
Air Blank		0.000	06:05
Control Test		0.079	06:05
Air Blank		0.000	06:05
Subject Sample	# 1	0.145	06:06
Breath Volume		1.113	06:06
Air Blank		0.000	06:06
Air Blank		0.000	06:08
Subject Sample	# 2	0.150	06:09
Breath Volume		2.808	06:09
Air Blank		0.000	06:09
Control Test		0.079	06:10
Air Blank		0.000	06:10
Diagnostics Check		OK	06:10

---

Data Download Date/Time: 12/29/2020 06:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: MCCULLOUGH,SEAN  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: CARDINAL

DATE/TIME OF TEST: 11/27/2020 22:02:11

---

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.078	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.122	21:57
Breath Volume	4.105	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.118	22:00
Breath Volume	3.312	22:00
Air Blank	0.000	22:01
Control Test	0.078	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

---

Data Download Date/Time: 12/29/2020 06:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: LAFFERTY,NAOMI  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: SARRA

DATE/TIME OF TEST: 11/28/2020 00:59:50

---

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:51	
Control Test	0.079	00:51	
Air Blank	0.000	00:52	
Subject Sample	# 1	00:55	No Sample Provided
Breath Volume	0.000	00:55	
Air Blank	0.000	00:56	
Air Blank	0.000	00:58	
Subject Sample	# 2	00:58	Subject Test Refused
Breath Volume	0.000	00:58	
Air Blank	0.000	00:58	
Control Test	0.079	00:59	
Air Blank	0.000	00:59	
Diagnostics Check	OK	00:59	

---

Data Download Date/Time: 12/29/2020 06:21



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: LAMOUR,ZACHARY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: MABREY

DATE/TIME OF TEST: 11/28/2020 04:11:09

---

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:03	
Control Test	0.080	04:03	
Air Blank	0.000	04:03	
Subject Sample	# 1	04:07	No Sample Provided
Breath Volume	0.000	04:07	
Air Blank	0.000	04:07	
Air Blank	0.000	04:09	
Subject Sample	# 2	04:09	Subject Test Refused
Breath Volume	0.000	04:09	
Air Blank	0.000	04:10	
Control Test	0.080	04:10	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:11	

---

Data Download Date/Time: 12/29/2020 06:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBISON,CASEY,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ROBISON,CASEY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: URIBE

DATE/TIME OF TEST: 11/28/2020 05:05:57

---

Diagnostics Check		OK	04:56
Air Blank		0.000	04:57
Control Test		0.080	04:57
Air Blank		0.000	04:57
Subject Sample	# 1	0.235	04:58
Breath Volume		2.589	04:58
Air Blank		0.000	04:59
Air Blank		0.000	05:00
Subject Sample	# 2	0.212	05:01
Breath Volume		1.480	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:03
Subject Sample	# 3	0.232	05:04
Breath Volume		2.707	05:04
Air Blank		0.000	05:05
Control Test		0.079	05:05
Air Blank		0.000	05:05
Diagnostics Check		OK	05:05

---

Data Download Date/Time: 12/29/2020 06:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAFFERTY,NAOMI,E  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: MCCULLOUGH,SEAN  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 11/28/2020 18:28:45

---

Diagnostics Check	OK	18:25	
Air Blank	0.000	18:26	
Control Test	0.080	18:26	
Air Blank	0.000	18:27	
Subject Sample	# 1	18:27	Subject Test Refused
Breath Volume	0.000	18:27	
Air Blank	0.000	18:27	
Control Test	0.079	18:28	
Air Blank	0.000	18:28	
Diagnostics Check	OK	18:28	

---

Data Download Date/Time: 12/29/2020 06:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MIRANDA,RICHARD,  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: HINES,PAUL  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: PICKETT

DATE/TIME OF TEST: 11/30/2020 19:30:51

---

Diagnostics Check	OK	19:27	
Air Blank	0.000	19:28	
Control Test	0.080	19:28	
Air Blank	0.000	19:29	
Subject Sample	# 1	19:29	Subject Test Refused
Breath Volume	0.000	19:29	
Air Blank	0.000	19:29	
Control Test	0.080	19:30	
Air Blank	0.000	19:30	
Diagnostics Check	OK	19:30	

---

Data Download Date/Time: 12/29/2020 06:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERON, WINSTON, G  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: WOOD, W  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 11/01/2020 04:19:25

---

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.082	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.205	04:14
Breath Volume	3.550	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.197	04:17
Breath Volume	3.183	04:17
Air Blank	0.000	04:18
Control Test	0.081	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

---

Data Download Date/Time: 12/02/2020 14:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HODGES,KAITLYN,M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: WOODS,W  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 11/02/2020 00:44:08

---

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.082	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.147	00:38
Breath Volume	1.875	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.146	00:42
Breath Volume	2.074	00:42
Air Blank	0.000	00:43
Control Test	0.081	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

---

Data Download Date/Time: 12/02/2020 14:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PICKETT,J,D  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: KNIGHT,G  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 11/03/2020 00:21:32

---

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.082	00:16
Air Blank	0.000	00:16
Subject Sample # 1	0.000	00:16
Breath Volume	3.316	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.000	00:20
Breath Volume	3.328	00:20
Air Blank	0.000	00:20
Control Test	0.082	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

---

Data Download Date/Time: 12/02/2020 14:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PICKETT,J,D  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HUNTER,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/06/2020 23:25:30

---

Diagnostics Check	OK	23:22	
Air Blank	0.000	23:23	
Control Test	0.081	23:23	
Air Blank	0.000	23:24	
Subject Sample	# 1	23:24	Subject Test Refused
Breath Volume	0.000	23:24	
Air Blank	0.000	23:24	
Control Test	0.081	23:24	
Air Blank	0.000	23:25	
Diagnostics Check	OK	23:25	

---

Data Download Date/Time: 12/02/2020 14:27



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PICKETT,J,D  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: KNIGHT,G  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 11/07/2020 01:13:50

---

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:11	
Control Test	0.082	01:11	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:12	
Control Test	0.081	01:13	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

---

Data Download Date/Time: 12/02/2020 14:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,J,  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HUNTER,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: HAMBRICK

DATE/TIME OF TEST: 11/07/2020 02:09:08

---

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.082	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.000	02:04
Breath Volume	2.308	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.000	02:07
Breath Volume	2.449	02:07
Air Blank	0.000	02:08
Control Test	0.082	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

---

Data Download Date/Time: 12/02/2020 14:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KUTZ,T,E  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HUNTER,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: BARTS

DATE/TIME OF TEST: 11/07/2020 21:02:22

---

Diagnostics Check	OK	20:59	
Air Blank	0.000	20:59	
Control Test	0.082	20:59	
Air Blank	0.000	21:00	
Subject Sample	# 1	21:01	Subject Test Refused
Breath Volume	0.000	21:01	
Air Blank	0.000	21:01	
Control Test	0.083	21:01	
Air Blank	0.000	21:02	
Diagnostics Check	OK	21:02	

---

Data Download Date/Time: 12/02/2020 14:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KUTZ,T,E  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HUNTER,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/07/2020 23:13:50

---

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:11	
Control Test	0.082	23:11	
Air Blank	0.000	23:12	
Subject Sample	# 1	23:12	Subject Test Refused
Breath Volume	0.000	23:12	
Air Blank	0.000	23:12	
Control Test	0.082	23:13	
Air Blank	0.000	23:13	
Diagnostics Check	OK	23:13	

---

Data Download Date/Time: 12/02/2020 14:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KUTZ,T,E  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: VAZQUEZ,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: SANDERS JR

DATE/TIME OF TEST: 11/08/2020 01:02:54

---

Diagnostics Check		OK	00:56
Air Blank		0.000	00:57
Control Test		0.082	00:57
Air Blank		0.000	00:57
Subject Sample	# 1	0.088	00:58
Breath Volume		1.335	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.084	01:01
Breath Volume		1.433	01:01
Air Blank		0.000	01:02
Control Test		0.081	01:02
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

---

Data Download Date/Time: 12/02/2020 14:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KUTZ,T,E  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HUNTER,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 11/08/2020 01:40:53

---

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.082	01:33
Air Blank	0.000	01:34
Subject Sample # 1	0.102	01:36
Breath Volume	1.113	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:38
Subject Sample # 2	0.104	01:39
Breath Volume	1.125	01:39
Air Blank	0.000	01:39
Control Test	0.081	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

---

Data Download Date/Time: 12/02/2020 14:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HODGES,K,M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HUNTER,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: LYNCKER

DATE/TIME OF TEST: 11/08/2020 03:46:48

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.080	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.142	03:42
Breath Volume	1.867	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:44
Subject Sample # 2	0.135	03:45
Breath Volume	1.113	03:45
Air Blank	0.000	03:45
Control Test	0.081	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

---

Data Download Date/Time: 12/02/2020 14:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMPTON,J,R  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: DUGGAN,L  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 11/11/2020 21:51:58

---

Diagnostics Check	OK	21:46
Air Blank	0.000	21:46
Control Test	0.082	21:46
Air Blank	0.000	21:47
Subject Sample # 1	0.000	21:47
Breath Volume	3.226	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 2	0.000	21:50
Breath Volume	3.410	21:50
Air Blank	0.000	21:51
Control Test	0.081	21:51
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

---

Data Download Date/Time: 12/02/2020 14:27



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PICKETT,J,D  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: VAZQUEZ,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: BARCARO

DATE/TIME OF TEST: 11/12/2020 21:50:38

---

Diagnostics Check	OK	21:47	
Air Blank	0.000	21:48	
Control Test	0.081	21:48	
Air Blank	0.000	21:49	
Subject Sample	# 1	21:49	Subject Test Refused
Breath Volume	0.000	21:49	
Air Blank	0.000	21:49	
Control Test	0.081	21:50	
Air Blank	0.000	21:50	
Diagnostics Check	OK	21:50	

---

Data Download Date/Time: 12/02/2020 14:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMPTON,J,R  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: WOODS,W  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: SUMMERFORD

DATE/TIME OF TEST: 11/14/2020 22:08:19

---

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:05	
Control Test	0.082	22:06	
Air Blank	0.000	22:06	
Subject Sample	# 1	22:06	Subject Test Refused
Breath Volume	0.000	22:06	
Air Blank	0.000	22:07	
Control Test	0.081	22:07	
Air Blank	0.000	22:08	
Diagnostics Check	OK	22:08	

---

Data Download Date/Time: 12/02/2020 14:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERON, WINSTON, G  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: BALLARD, J  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 11/15/2020 20:16:49

---

Diagnostics Check	OK	20:14	
Air Blank	0.000	20:14	
Control Test	0.082	20:14	
Air Blank	0.000	20:15	
Subject Sample	# 1	20:15	Subject Test Refused
Breath Volume	0.000	20:15	
Air Blank	0.000	20:15	
Control Test	0.081	20:16	
Air Blank	0.000	20:16	
Diagnostics Check	OK	20:16	

---

Data Download Date/Time: 12/02/2020 14:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KUTZ,T,K  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: TAYLOR,T  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: TOBIN

DATE/TIME OF TEST: 11/18/2020 02:34:42

---

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.082	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.155	02:29
Breath Volume	1.816	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.138	02:33
Breath Volume	2.121	02:33
Air Blank	0.000	02:33
Control Test	0.081	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

---

Data Download Date/Time: 12/02/2020 14:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TIPPINS,C,L  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: DELACRUZ,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: GAINNEY

DATE/TIME OF TEST: 11/18/2020 04:59:07

---

Diagnostics Check	OK	04:56	
Air Blank	0.000	04:56	
Control Test	0.082	04:57	
Air Blank	0.000	04:57	
Subject Sample	# 1	04:57	Subject Test Refused
Breath Volume	0.000	04:57	
Air Blank	0.000	04:58	
Control Test	0.081	04:58	
Air Blank	0.000	04:59	
Diagnostics Check	OK	04:59	

---

Data Download Date/Time: 12/02/2020 14:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIRES,J,L  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: DILL,T  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: WHITEN

DATE/TIME OF TEST: 11/21/2020 04:29:49

---

Diagnostics Check		OK	04:23
Air Blank		0.000	04:23
Control Test		0.081	04:23
Air Blank		0.000	04:24
Subject Sample	# 1	0.158	04:25
Breath Volume		1.796	04:25
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.161	04:28
Breath Volume		1.828	04:28
Air Blank		0.000	04:28
Control Test		0.080	04:29
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

---

Data Download Date/Time: 12/02/2020 14:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERON, WINSTON, G  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: MORGAN, J  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 11/21/2020 22:10:06

---

Diagnostics Check	OK	22:07	
Air Blank	0.000	22:07	
Control Test	0.081	22:08	
Air Blank	0.000	22:08	
Subject Sample	# 1	22:08	Subject Test Refused
Breath Volume	0.000	22:08	
Air Blank	0.000	22:09	
Control Test	0.082	22:09	
Air Blank	0.000	22:10	
Diagnostics Check	OK	22:10	

---

Data Download Date/Time: 12/02/2020 14:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMPTON,J,R  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: TIMOTHY,T  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: VINCENT

DATE/TIME OF TEST: 11/22/2020 23:41:50

---

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.082	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.000	23:37
Breath Volume	3.972	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.000	23:40
Breath Volume	3.851	23:40
Air Blank	0.000	23:40
Control Test	0.082	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

---

Data Download Date/Time: 12/02/2020 14:28



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PICKETT,J,D  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: TAYLOR,T  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 11/25/2020 20:28:50

---

Diagnostics Check	OK	20:22
Air Blank	0.000	20:22
Control Test	0.082	20:23
Air Blank	0.000	20:23
Subject Sample # 1	0.000	20:24
Breath Volume	4.199	20:24
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.000	20:27
Breath Volume	3.625	20:27
Air Blank	0.000	20:27
Control Test	0.081	20:28
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

---

Data Download Date/Time: 12/02/2020 14:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMPTON,J,R  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: DELACRUZ,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/27/2020 00:44:28

---

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.082	00:39
Air Blank	0.000	00:39
Subject Sample # 1	0.162	00:40
Breath Volume	3.167	00:40
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.165	00:43
Breath Volume	2.886	00:43
Air Blank	0.000	00:43
Control Test	0.083	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

---

Data Download Date/Time: 12/02/2020 14:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMPTON,J,R  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: WOODS,W  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 11/28/2020 03:45:48

---

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.082	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.124	03:41
Breath Volume	3.003	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.131	03:44
Breath Volume	4.503	03:44
Air Blank	0.000	03:44
Control Test	0.082	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

---

Data Download Date/Time: 12/02/2020 14:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERON, WINSTON, G  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: WOODS, W  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: CSORGO

DATE/TIME OF TEST: 11/29/2020 03:44:13

---

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.082	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.158	03:39
Breath Volume	3.375	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.144	03:42
Breath Volume	1.992	03:42
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

---

Data Download Date/Time: 12/02/2020 14:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERON, WINSTON, G  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: BALLARD, J  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: WHITLEY

DATE/TIME OF TEST: 11/29/2020 05:02:38

---

Diagnostics Check	OK	04:59	
Air Blank	0.000	05:00	
Control Test	0.082	05:00	
Air Blank	0.000	05:01	
Subject Sample	# 1	05:01	Subject Test Refused
Breath Volume	0.000	05:01	
Air Blank	0.000	05:01	
Control Test	0.082	05:02	
Air Blank	0.000	05:02	
Diagnostics Check	OK	05:02	

---

Data Download Date/Time: 12/02/2020 14:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,SUSAN,L  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: SMITH,SUSAN  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: COURT ORD  
ARREST TIME: 09:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:23

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 11/04/2020 09:51:47

---

Diagnostics Check	OK	09:45
Air Blank	0.000	09:46
Control Test	0.080	09:46
Air Blank	0.000	09:47
Subject Sample # 1	0.000	09:47
Breath Volume	2.187	09:47
Air Blank	0.000	09:48
Air Blank	0.000	09:50
Subject Sample # 2	0.000	09:50
Breath Volume	2.097	09:50
Air Blank	0.000	09:50
Control Test	0.080	09:51
Air Blank	0.000	09:51
Diagnostics Check	OK	09:51

---

Data Download Date/Time: 11/23/2020 10:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRISON,DUSTIN,J  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: CARRANZA,E  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: WORLEY

DATE/TIME OF TEST: 11/06/2020 23:05:41

---

Diagnostics Check	OK	22:58
Air Blank	0.000	22:59
Control Test	0.081	22:59
Air Blank	0.000	23:00
Subject Sample # 1	0.194	23:01
Breath Volume	1.421	23:01
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.199	23:04
Breath Volume	2.285	23:04
Air Blank	0.000	23:04
Control Test	0.081	23:05
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

---

Data Download Date/Time: 11/23/2020 10:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRISON,DUSTIN,J  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: OLSON,A  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: HERSTEIN

DATE/TIME OF TEST: 11/07/2020 21:36:59

---

Diagnostics Check	OK	21:30
Air Blank	0.000	21:31
Control Test	0.080	21:31
Air Blank	0.000	21:32
Subject Sample # 1	0.000	21:32
Breath Volume	1.546	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:35
Subject Sample # 2	0.000	21:35
Breath Volume	1.332	21:35
Air Blank	0.000	21:36
Control Test	0.080	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

---

Data Download Date/Time: 11/23/2020 10:03



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBLES,J,V  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: NEW,STEVE  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:04

SUBJECT'S LAST NAME: MORALES PALAFOX

DATE/TIME OF TEST: 11/08/2020 06:39:33

---

Diagnostics Check	OK	06:24	
Air Blank	0.000	06:24	
Control Test	0.080	06:25	
Air Blank	0.000	06:25	
Subject Sample	# 1 0.122	06:28	
Breath Volume	1.402	06:28	
Air Blank	0.000	06:29	
Air Blank	0.000	06:30	
Subject Sample	# 2 0.115	06:33	Volume Not Met
Breath Volume	1.003	06:33	
Air Blank	0.000	06:34	
Air Blank	0.000	06:36	
Subject Sample	# 3 0.119	06:38	
Breath Volume	1.347	06:38	
Air Blank	0.000	06:38	
Control Test	0.077	06:38	
Air Blank	0.000	06:39	
Diagnostics Check	OK	06:39	

---

Data Download Date/Time: 11/23/2020 10:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PECK,JOHN,R  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: GODINEZ,JORGE  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: TREJO ALEJO

DATE/TIME OF TEST: 11/11/2020 19:47:57

---

Diagnostics Check	OK	19:42	
Air Blank	0.000	19:43	
Control Test	0.080	19:43	
Air Blank	0.000	19:43	
Subject Sample	# 1	19:46	Subject Test Refused
Breath Volume	1.015	19:46	
Air Blank	0.000	19:47	
Control Test	0.083	19:47	
Air Blank	0.000	19:47	
Diagnostics Check	OK	19:47	

---

Data Download Date/Time: 11/23/2020 10:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PECK,JOHN,R  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: BEUFORD,S  
ARRESTING OFFICER AGENCY: Sebring PD  
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: TORRES MACIAS

DATE/TIME OF TEST: 11/12/2020 02:03:41

---

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:01	
Control Test	0.078	02:01	
Air Blank	0.000	02:02	
Subject Sample	# 1	02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:02	
Control Test	0.079	02:03	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

---

Data Download Date/Time: 11/23/2020 10:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROE,ROBERT,D  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: BECERRA,JOSE  
ARRESTING OFFICER AGENCY: Lake Placid PD  
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: REGAN

DATE/TIME OF TEST: 11/21/2020 23:56:46

---

Diagnostics Check		OK	23:45
Air Blank		0.000	23:46
Control Test		0.080	23:46
Air Blank		0.000	23:46
Subject Sample	# 1	0.277	23:47
Breath Volume		2.246	23:47
Air Blank		0.000	23:48
Air Blank		0.000	23:50
Subject Sample	# 2	0.250	23:51
Breath Volume		1.656	23:51
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 3	0.255	23:55
Breath Volume		2.156	23:55
Air Blank		0.000	23:55
Control Test		0.079	23:56
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

---

Data Download Date/Time: 11/23/2020 10:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROE,ROBERT,D  
BREATH TEST OPERATOR AGENCY: Highlands County SO  
ARRESTING OFFICER: GODINEZ,JORGE  
ARRESTING OFFICER AGENCY: Highlands County SO  
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 11/22/2020 01:18:19

---

Diagnostics Check		OK	01:12
Air Blank		0.000	01:12
Control Test		0.080	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.268	01:13
Breath Volume		2.367	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.273	01:16
Breath Volume		2.410	01:16
Air Blank		0.000	01:17
Control Test		0.080	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

---

Data Download Date/Time: 11/23/2020 10:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELCHER,ALEX,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BELCHER,ALEX  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 11/01/2020 00:03:47

---

Diagnostics Check	OK	23:54
Air Blank	0.000	23:55
Control Test	0.080	23:55
Air Blank	0.000	23:56
Subject Sample # 2	0.180	23:58
Breath Volume	1.480	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.175	00:02
Breath Volume	1.996	00:02
Air Blank	0.000	00:02
Control Test	0.079	00:03
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

---

Data Download Date/Time: 11/23/2020 11:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELCHER,ALEX,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BELCHER,ALEX  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: GUEVARA

DATE/TIME OF TEST: 11/01/2020 03:22:23

---

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.185	03:13
Breath Volume	1.136	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.209	03:17
Breath Volume	1.183	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 3	0.200	03:20
Breath Volume	1.136	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

---

Data Download Date/Time: 11/23/2020 11:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALLACE, CHARLES, A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WALLACE, CHARLES  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: STEPHENSON

DATE/TIME OF TEST: 11/04/2020 19:57:49

---

Diagnostics Check	OK	19:46	
Air Blank	0.000	19:46	
Control Test	0.080	19:47	
Air Blank	0.000	19:47	
Subject Sample	# 1	19:50	No Sample Provided
Breath Volume	0.000	19:50	
Air Blank	0.000	19:51	
Air Blank	0.000	19:53	
Subject Sample	# 2	19:56	No Sample Provided
Breath Volume	0.000	19:56	
Air Blank	0.000	19:56	
Control Test	0.080	19:57	
Air Blank	0.000	19:57	
Diagnostics Check	OK	19:57	

---

Data Download Date/Time: 11/23/2020 11:33



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELCHER,ALEX,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BELCHER,ALEX  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: ACKERLY

DATE/TIME OF TEST: 11/04/2020 23:45:38

---

Diagnostics Check	OK	23:34	
Air Blank	0.000	23:34	
Control Test	0.080	23:35	
Air Blank	0.000	23:35	
Subject Sample	# 1	23:38	No Sample Provided
Breath Volume	0.000	23:38	
Air Blank	0.000	23:39	
Air Blank	0.000	23:41	
Subject Sample	# 2	23:44	No Sample Provided
Breath Volume	0.000	23:44	
Air Blank	0.000	23:44	
Control Test	0.080	23:45	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:45	

---

Data Download Date/Time: 11/23/2020 11:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELCHER,ALEX,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PROPST,JOHN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/09/2020 18:16:35

---

Diagnostics Check	OK	18:06
Air Blank	0.000	18:07
Control Test	0.079	18:07
Air Blank	0.000	18:08
Subject Sample # 1	0.181	18:10
Breath Volume	1.718	18:10
Air Blank	0.000	18:10
Air Blank	0.000	18:12
Subject Sample # 2	0.171	18:15
Breath Volume	1.816	18:15
Air Blank	0.000	18:15
Control Test	0.079	18:15
Air Blank	0.000	18:16
Diagnostics Check	OK	18:16

---

Data Download Date/Time: 11/23/2020 11:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELCHER,ALEX,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BELCHER,ALEX  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: GERBER

DATE/TIME OF TEST: 11/09/2020 23:06:33

---

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.080	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.000	23:01
Breath Volume	3.328	23:01
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.000	23:05
Breath Volume	3.062	23:05
Air Blank	0.000	23:05
Control Test	0.080	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

---

Data Download Date/Time: 11/23/2020 11:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELCHER,ALEX,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WALLACE,CHARLES  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 11/10/2020 23:16:57

---

Diagnostics Check	OK	23:10
Air Blank	0.000	23:11
Control Test	0.078	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.000	23:12
Breath Volume	2.718	23:12
Air Blank	0.000	23:12
Air Blank	0.000	23:15
Subject Sample # 2	0.000	23:15
Breath Volume	2.199	23:15
Air Blank	0.000	23:16
Control Test	0.079	23:16
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

---

Data Download Date/Time: 11/23/2020 11:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELCHER,ALEX,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BELCHER,ALEX  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: RAO JR

DATE/TIME OF TEST: 11/11/2020 04:29:23

---

Diagnostics Check	OK	04:18	
Air Blank	0.000	04:18	
Control Test	0.080	04:18	
Air Blank	0.000	04:19	
Subject Sample	# 1	04:22	No Sample Provided
Breath Volume	0.000	04:22	
Air Blank	0.000	04:22	
Air Blank	0.000	04:24	
Subject Sample	# 2	04:28	No Sample Provided
Breath Volume	0.000	04:28	
Air Blank	0.000	04:28	
Control Test	0.080	04:28	
Air Blank	0.000	04:29	
Diagnostics Check	OK	04:29	

---

Data Download Date/Time: 11/23/2020 11:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELCHER,ALEX,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BELCHER,ALEX  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: CARNES

DATE/TIME OF TEST: 11/15/2020 02:10:26

---

Diagnostics Check	OK	02:09	
Air Blank		02:10	RFI Detect
Air Blank		02:10	RFI Detect

---

Data Download Date/Time: 11/23/2020 11:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELCHER,ALEX,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BELCHER,ALEX  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: CARNES

DATE/TIME OF TEST: 11/15/2020 02:24:55

---

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:14	
Control Test	0.080	02:14	
Air Blank	0.000	02:14	
Subject Sample	# 1	02:17	No Sample Provided
Breath Volume	0.000	02:17	
Air Blank	0.000	02:18	
Air Blank	0.000	02:20	
Subject Sample	# 2	02:23	No Sample Provided
Breath Volume	0.000	02:23	
Air Blank	0.000	02:24	
Control Test	0.079	02:24	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

---

Data Download Date/Time: 11/23/2020 11:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELCHER,ALEX,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PROPST,JOHN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: CADENA

DATE/TIME OF TEST: 11/15/2020 04:07:58

---

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:57	
Control Test	0.080	03:57	
Air Blank	0.000	03:57	
Subject Sample	# 1	04:01	No Sample Provided
Breath Volume	0.000	04:01	
Air Blank	0.000	04:01	
Air Blank	0.000	04:03	
Subject Sample	# 2	04:06	No Sample Provided
Breath Volume	0.000	04:06	
Air Blank	0.000	04:07	
Control Test	0.080	04:07	
Air Blank	0.000	04:07	
Diagnostics Check	OK	04:07	

---

Data Download Date/Time: 11/23/2020 11:34



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELCHER,ALEX,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BELCHER,ALEX  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 11/19/2020 04:04:24

---

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.080	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.000	03:58
Breath Volume	2.109	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.000	04:03
Breath Volume	1.886	04:03
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

---

Data Download Date/Time: 11/23/2020 11:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,ALBERT,J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WILLIAMS,ALBERT  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: DUARTE BUSTOS

DATE/TIME OF TEST: 11/24/2020 20:26:31

---

Diagnostics Check	OK	20:18
Air Blank	0.000	20:18
Control Test	0.080	20:19
Air Blank	0.000	20:19
Subject Sample # 1	0.000	20:21
Breath Volume	2.183	20:21
Air Blank	0.000	20:22
Air Blank	0.000	20:24
Subject Sample # 2	0.000	20:25
Breath Volume	1.726	20:25
Air Blank	0.000	20:25
Control Test	0.080	20:25
Air Blank	0.000	20:26
Diagnostics Check	OK	20:26

---

Data Download Date/Time: 12/18/2020 13:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MULKEARNS, MICHAEL, S  
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD  
ARRESTING OFFICER: SIEH, Z  
ARRESTING OFFICER AGENCY: Fernandina Beach PD  
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: LINDSEY

DATE/TIME OF TEST: 11/04/2020 23:01:30

---

Diagnostics Check	OK	22:54
Air Blank	0.000	22:55
Control Test	0.078	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.081	22:56
Breath Volume	2.835	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.083	23:00
Breath Volume	2.867	23:00
Air Blank	0.000	23:00
Control Test	0.079	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

---

Data Download Date/Time: 11/10/2020 10:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOTHARD, RONALD, S  
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD  
ARRESTING OFFICER: GOTHARD, RONALD  
ARRESTING OFFICER AGENCY: Fernandina Beach PD  
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 11/11/2020 01:39:19

---

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.077	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.168	01:34
Breath Volume	2.167	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.166	01:37
Breath Volume	1.933	01:37
Air Blank	0.000	01:38
Control Test	0.078	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

---

Data Download Date/Time: 12/11/2020 17:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,JORGE,L  
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD  
ARRESTING OFFICER: PAETSCH,JOSHUA  
ARRESTING OFFICER AGENCY: Fernandina Beach PD  
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:04

SUBJECT'S LAST NAME: SOUZA

DATE/TIME OF TEST: 11/11/2020 16:35:43

---

Diagnostics Check	OK	16:32	
Air Blank	0.000	16:32	
Control Test	0.078	16:33	
Air Blank	0.000	16:33	
Subject Sample	# 1	16:34	Subject Test Refused
Breath Volume	0.000	16:34	
Air Blank	0.000	16:34	
Control Test	0.077	16:35	
Air Blank	0.000	16:35	
Diagnostics Check	OK	16:35	

---

Data Download Date/Time: 12/11/2020 17:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAETSCH,JOSH,E  
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD  
ARRESTING OFFICER: PAETSCH,JOSH  
ARRESTING OFFICER AGENCY: Fernandina Beach PD  
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: CARDONA

DATE/TIME OF TEST: 11/22/2020 22:09:06

---

Diagnostics Check		OK	22:00
Air Blank		0.000	22:00
Control Test		0.078	22:00
Air Blank		0.000	22:01
Subject Sample	# 1	0.219	22:02
Breath Volume		1.437	22:02
Air Blank		0.000	22:03
Air Blank		0.000	22:05
Subject Sample	# 2	0.238	22:07
Breath Volume		1.296	22:07
Air Blank		0.000	22:08
Control Test		0.079	22:08
Air Blank		0.000	22:09
Diagnostics Check		OK	22:09

---

Data Download Date/Time: 12/11/2020 17:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOTHARD, RONALD, S  
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD  
ARRESTING OFFICER: GOTHARD, RONALD  
ARRESTING OFFICER AGENCY: Fernandina Beach PD  
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/28/2020 03:08:26

---

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:06	
Control Test	0.077	03:06	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:07	
Control Test	0.077	03:07	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

---

Data Download Date/Time: 12/11/2020 17:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAETSCH,JOSH,E  
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD  
ARRESTING OFFICER: PAETSCH,JOSH  
ARRESTING OFFICER AGENCY: Fernandina Beach PD  
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: MASSENGILL

DATE/TIME OF TEST: 11/29/2020 03:29:35

---

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:27	
Control Test	0.078	03:27	
Air Blank	0.000	03:27	
Subject Sample	# 1	03:28	Subject Test Refused
Breath Volume	0.000	03:28	
Air Blank	0.000	03:28	
Control Test	0.077	03:28	
Air Blank	0.000	03:29	
Diagnostics Check	OK	03:29	

---

Data Download Date/Time: 12/11/2020 17:29



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAETSCH,JOSH,E  
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD  
ARRESTING OFFICER: PAETSCH,JOSH  
ARRESTING OFFICER AGENCY: Fernandina Beach PD  
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: SPIEWAK

DATE/TIME OF TEST: 11/29/2020 20:49:01

---

Diagnostics Check	OK	20:42
Air Blank	0.000	20:43
Control Test	0.078	20:43
Air Blank	0.000	20:43
Subject Sample # 1	0.000	20:44
Breath Volume	1.550	20:44
Air Blank	0.000	20:44
Air Blank	0.000	20:47
Subject Sample # 2	0.000	20:47
Breath Volume	1.398	20:47
Air Blank	0.000	20:48
Control Test	0.079	20:48
Air Blank	0.000	20:48
Diagnostics Check	OK	20:49

---

Data Download Date/Time: 12/11/2020 17:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANDENBERG,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Gulfport PD  
ARRESTING OFFICER: VANDENBERG,MICHAEL  
ARRESTING OFFICER AGENCY: Gulfport PD  
LAST AGENCY INSPECTION: 10/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: FLORES GARCIA

DATE/TIME OF TEST: 11/12/2020 14:31:35

---

Diagnostics Check	OK	14:25
Air Blank	0.000	14:25
Control Test	0.080	14:26
Air Blank	0.000	14:26
Subject Sample # 1	0.290	14:27
Breath Volume	1.894	14:27
Air Blank	0.000	14:27
Air Blank	0.000	14:29
Subject Sample # 2	0.290	14:30
Breath Volume	1.597	14:30
Air Blank	0.000	14:30
Control Test	0.083	14:30
Air Blank	0.000	14:31
Diagnostics Check	OK	14:31

---

Data Download Date/Time: 11/28/2020 01:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ENGEL,JASON,S  
BREATH TEST OPERATOR AGENCY: New Port Richey PD  
ARRESTING OFFICER: CAMPANA,NICHOLAS  
ARRESTING OFFICER AGENCY: New Port Richey PD  
LAST AGENCY INSPECTION: 10/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A3  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: MAJTENYI

DATE/TIME OF TEST: 11/04/2020 03:02:53

---

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.079	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.165	02:57
Breath Volume	1.351	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.180	03:01
Breath Volume	1.953	03:01
Air Blank	0.000	03:01
Control Test	0.079	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

---

Data Download Date/Time: 11/26/2020 22:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ENGEL,JASON,S  
BREATH TEST OPERATOR AGENCY: New Port Richey PD  
ARRESTING OFFICER: GALLOWAY,GRANT  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A3  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: CORTES

DATE/TIME OF TEST: 11/07/2020 00:41:50

---

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.077	00:36
Air Blank	0.000	00:36
Subject Sample # 1	0.119	00:37
Breath Volume	1.468	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.117	00:40
Breath Volume	1.671	00:40
Air Blank	0.000	00:40
Control Test	0.077	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

---

Data Download Date/Time: 11/26/2020 22:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ECK,AARON,R  
BREATH TEST OPERATOR AGENCY: New Port Richey PD  
ARRESTING OFFICER: LEE,M  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A3  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 12:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PARDO DAVID

DATE/TIME OF TEST: 11/07/2020 01:57:05

---

Diagnostics Check		OK	01:50
Air Blank		0.000	01:51
Control Test		0.078	01:51
Air Blank		0.000	01:51
Subject Sample	# 1	0.129	01:52
Breath Volume		2.011	01:52
Air Blank		0.000	01:53
Air Blank		0.000	01:54
Subject Sample	# 2	0.130	01:55
Breath Volume		2.089	01:55
Air Blank		0.000	01:56
Control Test		0.079	01:56
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

---

Data Download Date/Time: 11/26/2020 22:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ENGEL,JASON,S  
BREATH TEST OPERATOR AGENCY: New Port Richey PD  
ARRESTING OFFICER: CAMPANA,NICHOLAS  
ARRESTING OFFICER AGENCY: New Port Richey PD  
LAST AGENCY INSPECTION: 10/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A3  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: DAILEY

DATE/TIME OF TEST: 11/09/2020 06:55:08

---

Diagnostics Check	OK	06:48
Air Blank	0.000	06:49
Control Test	0.078	06:49
Air Blank	0.000	06:50
Subject Sample # 1	0.051	06:50
Breath Volume	1.804	06:50
Air Blank	0.000	06:51
Air Blank	0.000	06:53
Subject Sample # 2	0.047	06:53
Breath Volume	2.691	06:53
Air Blank	0.000	06:54
Control Test	0.079	06:54
Air Blank	0.000	06:55
Diagnostics Check	OK	06:55

---

Data Download Date/Time: 11/26/2020 22:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ECK,AARON,R  
BREATH TEST OPERATOR AGENCY: New Port Richey PD  
ARRESTING OFFICER: TOLDO,M  
ARRESTING OFFICER AGENCY: New Port Richey PD  
LAST AGENCY INSPECTION: 11/26/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A3  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: LAM

DATE/TIME OF TEST: 11/28/2020 23:35:02

---

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.079	23:29
Air Blank	0.000	23:30
Subject Sample # 1	0.101	23:30
Breath Volume	3.066	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.096	23:33
Breath Volume	1.960	23:33
Air Blank	0.000	23:34
Control Test	0.079	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:35

---

Data Download Date/Time: 12/21/2020 02:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GIBSON,JOHN,A  
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD  
ARRESTING OFFICER: WOOD,N  
ARRESTING OFFICER AGENCY: Tarpon Springs PD  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: POSSO

DATE/TIME OF TEST: 11/01/2020 05:19:48

---

Diagnostics Check	OK	05:18	
Air Blank	0.000	05:18	
Control Test	0.074	05:19	Control Outside Tolerance
Air Blank	0.000	05:19	

---

Data Download Date/Time: 11/24/2020 12:10



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GIBSON,J,  
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD  
ARRESTING OFFICER: WOOD,N  
ARRESTING OFFICER AGENCY: Tarpon Springs PD  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: POSSO

DATE/TIME OF TEST: 11/01/2020 05:28:22

---

Diagnostics Check	OK	05:21	
Air Blank	0.000	05:22	
Control Test	0.075	05:22	
Air Blank	0.000	05:23	
Subject Sample	# 1 0.169	05:23	
Breath Volume	2.308	05:23	
Air Blank	0.000	05:24	
Air Blank	0.000	05:26	
Subject Sample	# 2 0.159	05:26	
Breath Volume	2.070	05:26	
Air Blank	0.000	05:27	
Control Test	0.074	05:27	Control Outside Tolerance
Air Blank	0.000	05:28	

---

Data Download Date/Time: 11/24/2020 12:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GIBSON,J,  
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD  
ARRESTING OFFICER: WOOD,N  
ARRESTING OFFICER AGENCY: Tarpon Springs PD  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: POSSO

DATE/TIME OF TEST: 11/01/2020 05:35:59

---

Diagnostics Check	OK	05:34	
Air Blank	0.000	05:35	
Control Test	0.074	05:35	Control Outside Tolerance
Air Blank	0.000	05:35	

---

Data Download Date/Time: 11/24/2020 12:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPATZ,JOHN,M  
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD  
ARRESTING OFFICER: PALMER,LEANN  
ARRESTING OFFICER AGENCY: Tarpon Springs PD  
LAST AGENCY INSPECTION: 10/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: OLSZEWSKI

DATE/TIME OF TEST: 11/01/2020 21:56:58

---

Diagnostics Check	OK	21:54	
Air Blank	0.000	21:54	
Control Test	0.077	21:54	
Air Blank	0.000	21:55	
Subject Sample	# 1	21:55	Subject Test Refused
Breath Volume	0.000	21:55	
Air Blank	0.000	21:56	
Control Test	0.077	21:56	
Air Blank	0.000	21:56	
Diagnostics Check	OK	21:56	

---

Data Download Date/Time: 11/24/2020 12:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: TEMPLE,ELIZABETH  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 10/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: CZEKIEL

DATE/TIME OF TEST: 11/01/2020 00:47:25

---

Diagnostics Check	OK	00:44	
Air Blank	0.000	00:44	
Control Test	0.082	00:45	
Air Blank	0.000	00:45	
Subject Sample	# 1	00:46	Subject Test Refused
Breath Volume	0.000	00:46	
Air Blank	0.000	00:46	
Control Test	0.082	00:46	
Air Blank	0.000	00:47	
Diagnostics Check	OK	00:47	

---

Data Download Date/Time: 11/05/2020 11:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MITCHELL,A  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 10/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BYERLY

DATE/TIME OF TEST: 11/01/2020 01:44:28

---

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:41	
Control Test	0.082	01:41	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:43	
Control Test	0.081	01:43	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

---

Data Download Date/Time: 11/05/2020 11:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: TORO,J  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 10/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 11/01/2020 03:06:11

---

Diagnostics Check	OK	03:03	
Air Blank	0.000	03:03	
Control Test	0.082	03:04	
Air Blank	0.000	03:04	
Subject Sample	# 1	03:04	Subject Test Refused
Breath Volume	0.000	03:04	
Air Blank	0.000	03:05	
Control Test	0.082	03:05	
Air Blank	0.000	03:06	
Diagnostics Check	OK	03:06	

---

Data Download Date/Time: 11/05/2020 11:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MULLINS,J  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 10/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 11/01/2020 04:59:18

---

Diagnostics Check	OK	04:50
Air Blank	0.000	04:51
Control Test	0.083	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.124	04:52
Breath Volume	1.335	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.122	04:57
Breath Volume	1.308	04:57
Air Blank	0.000	04:58
Control Test	0.082	04:58
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

---

Data Download Date/Time: 11/05/2020 11:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REEVES,MARK,O  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: INNES,JOSHUA  
ARRESTING OFFICER AGENCY: Treasure Island PD  
LAST AGENCY INSPECTION: 10/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: DEXTER

DATE/TIME OF TEST: 11/01/2020 19:11:56

---

Diagnostics Check	OK	19:04
Air Blank	0.000	19:05
Control Test	0.084	19:05
Air Blank	0.000	19:05
Subject Sample # 1	0.231	19:06
Breath Volume	2.230	19:06
Air Blank	0.000	19:07
Air Blank	0.000	19:09
Subject Sample # 2	0.215	19:10
Breath Volume	1.664	19:10
Air Blank	0.000	19:11
Control Test	0.081	19:11
Air Blank	0.000	19:11
Diagnostics Check	OK	19:11

---

Data Download Date/Time: 11/05/2020 11:35



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REFUSE, RONY,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: REFUSE, RONY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: PHAN

DATE/TIME OF TEST: 11/02/2020 00:44:39

---

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.083	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.054	00:39
Breath Volume	1.132	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.051	00:43
Breath Volume	1.203	00:43
Air Blank	0.000	00:43
Control Test	0.082	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

---

Data Download Date/Time: 11/05/2020 11:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REEVES,MARK,O  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: PETERSON,WILLIAM  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 10/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MARQUEZ RODRIGUEZ

DATE/TIME OF TEST: 11/02/2020 03:29:58

---

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.083	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.070	03:25
Breath Volume	2.941	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:27
Subject Sample # 2	0.066	03:28
Breath Volume	2.738	03:28
Air Blank	0.000	03:29
Control Test	0.082	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

---

Data Download Date/Time: 11/05/2020 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: IFFT,GARRETT,T  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: IFFT,GARRETT  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 12:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: DUPUIS

DATE/TIME OF TEST: 11/03/2020 01:38:51

---

Diagnostics Check		OK	01:32
Air Blank		0.000	01:33
Control Test		0.083	01:33
Air Blank		0.000	01:33
Subject Sample	# 1	0.206	01:34
Breath Volume		2.519	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:36
Subject Sample	# 2	0.200	01:37
Breath Volume		2.546	01:37
Air Blank		0.000	01:37
Control Test		0.082	01:38
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

---

Data Download Date/Time: 11/05/2020 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: DUNNING,RYAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 10/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: KASSENS

DATE/TIME OF TEST: 11/04/2020 03:06:52

---

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.083	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.000	03:02
Breath Volume	2.417	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.000	03:05
Breath Volume	2.207	03:05
Air Blank	0.000	03:05
Control Test	0.082	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

---

Data Download Date/Time: 11/05/2020 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNAVELY,KENT,J  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MILLER,DEPUTY  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 10/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 09:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: NOBLE

DATE/TIME OF TEST: 11/04/2020 11:04:17

---

Diagnostics Check	OK	10:55
Air Blank	0.000	10:56
Control Test	0.083	10:56
Air Blank	0.000	10:57
Subject Sample # 1	0.303	10:59
Breath Volume	1.183	10:59
Air Blank	0.000	10:59
Air Blank	0.000	11:01
Subject Sample # 2	0.322	11:02
Breath Volume	1.132	11:02
Air Blank	0.000	11:03
Control Test	0.080	11:03
Air Blank	0.000	11:04
Diagnostics Check	OK	11:04

---

Data Download Date/Time: 11/05/2020 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNAVELY,KENT,J  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: SNAVELY,KENT  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 10/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 11:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:35

SUBJECT'S LAST NAME: POOLE

DATE/TIME OF TEST: 11/04/2020 12:01:24

---

Diagnostics Check	OK	11:55
Air Blank	0.000	11:55
Control Test	0.081	11:55
Air Blank	0.000	11:56
Subject Sample # 1	0.000	11:57
Breath Volume	1.804	11:57
Air Blank	0.000	11:57
Air Blank	0.000	11:59
Subject Sample # 2	0.000	12:00
Breath Volume	2.093	12:00
Air Blank	0.000	12:00
Control Test	0.081	12:00
Air Blank	0.000	12:01
Diagnostics Check	OK	12:01

---

Data Download Date/Time: 11/05/2020 11:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVI,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAKE,LEVI  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: PIOTROWSKI

DATE/TIME OF TEST: 11/04/2020 22:10:37

---

Diagnostics Check	OK	22:07	
Air Blank	0.000	22:07	
Control Test	0.082	22:08	
Air Blank	0.000	22:08	
Subject Sample	# 1	22:09	Subject Test Refused
Breath Volume	0.000	22:09	
Air Blank	0.000	22:09	
Control Test	0.082	22:10	
Air Blank	0.000	22:10	
Diagnostics Check	OK	22:10	

---

Data Download Date/Time: 11/05/2020 11:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVIS  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: TORO,JARED  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: TROCONIS DE GUTIERRE

DATE/TIME OF TEST: 11/04/2020 23:25:23

---

Diagnostics Check	OK	23:15	
Air Blank	0.000	23:15	
Control Test	0.082	23:16	
Air Blank	0.000	23:16	
Subject Sample	# 1	23:19	No Sample Provided
Breath Volume	0.000	23:19	
Air Blank	0.000	23:20	
Air Blank	0.000	23:22	
Subject Sample	# 2	23:24	Subject Test Refused
Breath Volume	0.000	23:24	
Air Blank	0.000	23:24	
Control Test	0.082	23:24	
Air Blank	0.000	23:25	
Diagnostics Check	OK	23:25	

---

Data Download Date/Time: 11/05/2020 11:35



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANEY,DAMON,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LANEY,DAMON  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: KOWALSKI

DATE/TIME OF TEST: 11/05/2020 01:18:43

---

Diagnostics Check		OK	01:12
Air Blank		0.000	01:13
Control Test		0.082	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.095	01:14
Breath Volume		2.296	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:16
Subject Sample	# 2	0.100	01:17
Breath Volume		2.773	01:17
Air Blank		0.000	01:17
Control Test		0.082	01:18
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

---

Data Download Date/Time: 11/05/2020 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CONEBY,M  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 11/06/2020 21:51:53

---

Diagnostics Check	OK	21:45
Air Blank	0.000	21:45
Control Test	0.082	21:45
Air Blank	0.000	21:46
Subject Sample # 1	0.171	21:47
Breath Volume	1.453	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:49
Subject Sample # 2	0.160	21:50
Breath Volume	2.000	21:50
Air Blank	0.000	21:50
Control Test	0.081	21:51
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

---

Data Download Date/Time: 12/05/2020 04:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: JOHNSON,GABRIEL  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: GORDON-MARTIN

DATE/TIME OF TEST: 11/07/2020 05:41:00

---

Diagnostics Check	OK	05:34
Air Blank	0.000	05:35
Control Test	0.082	05:35
Air Blank	0.000	05:36
Subject Sample # 1	0.191	05:36
Breath Volume	2.132	05:36
Air Blank	0.000	05:37
Air Blank	0.000	05:38
Subject Sample # 2	0.192	05:39
Breath Volume	1.800	05:39
Air Blank	0.000	05:40
Control Test	0.082	05:40
Air Blank	0.000	05:40
Diagnostics Check	OK	05:41

---

Data Download Date/Time: 12/05/2020 04:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: JOHNSON,G  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: BURNS KEIRNAN

DATE/TIME OF TEST: 11/07/2020 21:45:06

---

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:42	
Control Test	0.082	21:42	
Air Blank	0.000	21:43	
Subject Sample	# 1	21:43	Subject Test Refused
Breath Volume	0.000	21:43	
Air Blank	0.000	21:44	
Control Test	0.081	21:44	
Air Blank	0.000	21:45	
Diagnostics Check	OK	21:45	

---

Data Download Date/Time: 12/05/2020 04:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOWATT,NATHAN,E  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MOWATT,NATHAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: PARTAIN

DATE/TIME OF TEST: 11/07/2020 23:00:20

---

Diagnostics Check	OK	22:53
Air Blank	0.000	22:54
Control Test	0.083	22:54
Air Blank	0.000	22:55
Subject Sample # 1	0.210	22:55
Breath Volume	1.812	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.209	22:58
Breath Volume	2.535	22:58
Air Blank	0.000	22:59
Control Test	0.081	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

---

Data Download Date/Time: 12/05/2020 04:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: SCHOMP,D  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: HUBLER

DATE/TIME OF TEST: 11/08/2020 00:02:50

---

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:00	
Control Test	0.081	00:00	
Air Blank	0.000	00:01	
Subject Sample	# 1	00:01	Subject Test Refused
Breath Volume	0.000	00:01	
Air Blank	0.000	00:01	
Control Test	0.082	00:02	
Air Blank	0.000	00:02	
Diagnostics Check	OK	00:02	

---

Data Download Date/Time: 12/05/2020 04:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CHRISS,W  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: HAYNES

DATE/TIME OF TEST: 11/08/2020 00:42:31

---

Diagnostics Check	OK	00:38	
Air Blank	0.000	00:38	
Control Test	0.081	00:39	
Air Blank	0.000	00:39	
Subject Sample	# 1	00:41	Subject Test Refused
Breath Volume	0.000	00:41	
Air Blank	0.000	00:41	
Control Test	0.082	00:41	
Air Blank	0.000	00:42	
Diagnostics Check	OK	00:42	

---

Data Download Date/Time: 12/05/2020 04:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REFUSE, RONY,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: REFUSE, RONY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 12:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 11/08/2020 01:34:22

---

Diagnostics Check	OK	01:27
Air Blank	0.000	01:28
Control Test	0.082	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.166	01:29
Breath Volume	2.230	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.167	01:32
Breath Volume	1.968	01:32
Air Blank	0.000	01:33
Control Test	0.081	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

---

Data Download Date/Time: 12/05/2020 04:10



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: REFUSE,RONY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: MATHEWS

DATE/TIME OF TEST: 11/08/2020 07:11:35

---

Diagnostics Check	OK	07:08	
Air Blank	0.000	07:09	
Control Test	0.082	07:09	
Air Blank	0.000	07:10	
Subject Sample	# 1	07:10	Subject Test Refused
Breath Volume	0.000	07:10	
Air Blank	0.000	07:10	
Control Test	0.082	07:10	
Air Blank	0.000	07:11	
Diagnostics Check	OK	07:11	

---

Data Download Date/Time: 12/05/2020 04:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOWATT,NATHAN,E  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MOWATT,NATHAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 11/09/2020 00:43:18

---

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:40	
Control Test	0.083	00:41	
Air Blank	0.000	00:41	
Subject Sample	# 1	00:41	Subject Test Refused
Breath Volume	0.000	00:41	
Air Blank	0.000	00:42	
Control Test	0.082	00:42	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

---

Data Download Date/Time: 12/05/2020 04:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: GRAVER,RYAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: HILDEBRAND

DATE/TIME OF TEST: 11/09/2020 02:28:49

---

Diagnostics Check		OK	02:21
Air Blank		0.000	02:22
Control Test		0.082	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.104	02:23
Breath Volume		2.070	02:23
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.105	02:27
Breath Volume		2.132	02:27
Air Blank		0.000	02:27
Control Test		0.081	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

---

Data Download Date/Time: 12/05/2020 04:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILTSE, CORPORAL W.,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LAMBORGHINI, DEPUTY P.  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: WOLF

DATE/TIME OF TEST: 11/09/2020 20:08:53

---

Diagnostics Check	OK	20:01
Air Blank	0.000	20:01
Control Test	0.082	20:01
Air Blank	0.000	20:02
Subject Sample # 1	0.255	20:03
Breath Volume	2.125	20:03
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.237	20:07
Breath Volume	1.992	20:07
Air Blank	0.000	20:07
Control Test	0.082	20:08
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

---

Data Download Date/Time: 12/05/2020 04:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNAVELY,KENT,J  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MILLER,HOWARD  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 11/10/2020 07:07:38

---

Diagnostics Check		OK	07:00
Air Blank		0.000	07:00
Control Test		0.082	07:01
Air Blank		0.000	07:01
Subject Sample	# 1	0.100	07:02
Breath Volume		2.792	07:02
Air Blank		0.000	07:03
Air Blank		0.000	07:05
Subject Sample	# 2	0.097	07:06
Breath Volume		1.496	07:06
Air Blank		0.000	07:06
Control Test		0.081	07:07
Air Blank		0.000	07:07
Diagnostics Check		OK	07:07

---

Data Download Date/Time: 12/05/2020 04:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: IFFT,GARRETT,T  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: IFFT,GARRETT  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 11:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: PANTOJA

DATE/TIME OF TEST: 11/11/2020 01:16:39

---

Diagnostics Check		OK	01:10
Air Blank		0.000	01:10
Control Test		0.083	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.000	01:12
Breath Volume		1.968	01:12
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.000	01:15
Breath Volume		2.492	01:15
Air Blank		0.000	01:15
Control Test		0.082	01:16
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

---

Data Download Date/Time: 12/05/2020 04:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, WILLIAM, T  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SMITH, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: HASSAN

DATE/TIME OF TEST: 11/11/2020 23:09:00

---

Diagnostics Check	OK	22:57	
Air Blank	0.000	22:57	
Control Test	0.082	22:58	
Air Blank	0.000	22:58	
Subject Sample	# 1 0.210	22:58	
Breath Volume	1.597	22:58	
Air Blank	0.000	22:59	
Air Blank	0.000	23:01	
Subject Sample	# 2 0.181	23:04	Volume Not Met
Breath Volume	0.417	23:04	
Air Blank	0.000	23:05	
Air Blank	0.000	23:07	
Subject Sample	# 3 0.196	23:07	
Breath Volume	1.730	23:07	
Air Blank	0.000	23:08	
Control Test	0.078	23:08	
Air Blank	0.000	23:08	
Diagnostics Check	OK	23:09	

---

Data Download Date/Time: 12/05/2020 04:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: DUNNING,R  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/12/2020 02:48:47

---

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:45	
Control Test	0.082	02:45	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:47	Subject Test Refused
Breath Volume	0.000	02:47	
Air Blank	0.000	02:47	
Control Test	0.082	02:48	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:48	

---

Data Download Date/Time: 12/05/2020 04:10



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MARTINI,ISABELLE  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: MARCHESE

DATE/TIME OF TEST: 11/12/2020 05:56:18

---

Diagnostics Check	OK	05:49
Air Blank	0.000	05:50
Control Test	0.082	05:50
Air Blank	0.000	05:51
Subject Sample # 1	0.160	05:51
Breath Volume	1.511	05:51
Air Blank	0.000	05:52
Air Blank	0.000	05:54
Subject Sample # 2	0.151	05:54
Breath Volume	2.035	05:54
Air Blank	0.000	05:55
Control Test	0.082	05:55
Air Blank	0.000	05:56
Diagnostics Check	OK	05:56

---

Data Download Date/Time: 12/05/2020 04:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CHURCHILL,C  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: HENEVELD

DATE/TIME OF TEST: 11/12/2020 23:30:42

---

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:28	
Control Test	0.083	23:28	
Air Blank	0.000	23:29	
Subject Sample	# 1	23:29	Subject Test Refused
Breath Volume	0.000	23:29	
Air Blank	0.000	23:29	
Control Test	0.081	23:30	
Air Blank	0.000	23:30	
Diagnostics Check	OK	23:30	

---

Data Download Date/Time: 12/05/2020 04:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: ZSIDO,D  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: MINIAK

DATE/TIME OF TEST: 11/13/2020 01:37:41

---

Diagnostics Check		OK	01:29
Air Blank		0.000	01:29
Control Test		0.082	01:30
Air Blank		0.000	01:30
Subject Sample	# 1	0.198	01:32
Breath Volume		2.273	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.193	01:36
Breath Volume		2.703	01:36
Air Blank		0.000	01:36
Control Test		0.081	01:37
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

---

Data Download Date/Time: 12/05/2020 04:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: GRILL,SHANNON  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: ADOUM

DATE/TIME OF TEST: 11/13/2020 06:07:04

---

Diagnostics Check	OK	06:00
Air Blank	0.000	06:01
Control Test	0.082	06:01
Air Blank	0.000	06:02
Subject Sample # 1	0.195	06:02
Breath Volume	3.378	06:02
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample # 2	0.193	06:05
Breath Volume	3.320	06:05
Air Blank	0.000	06:06
Control Test	0.081	06:06
Air Blank	0.000	06:06
Diagnostics Check	OK	06:07

---

Data Download Date/Time: 12/05/2020 04:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAREY,K,P  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CAREY,K  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: NAKAPAKORN

DATE/TIME OF TEST: 11/13/2020 23:00:20

---

Diagnostics Check	OK	22:53
Air Blank	0.000	22:54
Control Test	0.082	22:54
Air Blank	0.000	22:54
Subject Sample # 1	0.174	22:55
Breath Volume	2.632	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.172	22:58
Breath Volume	2.542	22:58
Air Blank	0.000	22:59
Control Test	0.081	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

---

Data Download Date/Time: 12/05/2020 04:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MITCHELL,A  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BARONCINI

DATE/TIME OF TEST: 11/14/2020 00:43:26

---

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:41	
Control Test	0.082	00:41	
Air Blank	0.000	00:41	
Subject Sample	# 1	00:42	Subject Test Refused
Breath Volume	0.000	00:42	
Air Blank	0.000	00:42	
Control Test	0.082	00:42	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

---

Data Download Date/Time: 12/05/2020 04:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MITCHELL,A  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: BRIDGES

DATE/TIME OF TEST: 11/14/2020 21:19:43

---

Diagnostics Check	OK	21:12
Air Blank	0.000	21:13
Control Test	0.083	21:13
Air Blank	0.000	21:14
Subject Sample # 1	0.130	21:15
Breath Volume	1.640	21:15
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.126	21:18
Breath Volume	1.691	21:18
Air Blank	0.000	21:18
Control Test	0.082	21:19
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

---

Data Download Date/Time: 12/05/2020 04:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: HAZIME,S  
ARRESTING OFFICER AGENCY: Tarpon Springs PD  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: LATTEA

DATE/TIME OF TEST: 11/15/2020 03:53:15

---

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.082	03:51	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:52	
Control Test	0.081	03:52	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

---

Data Download Date/Time: 12/05/2020 04:10



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILTSE,CORPORAL W.,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: WILTSE,CORPORAL W.  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: FERRER-PIZARRO

DATE/TIME OF TEST: 11/15/2020 04:42:15

---

Diagnostics Check		OK	04:35
Air Blank		0.000	04:35
Control Test		0.082	04:35
Air Blank		0.000	04:36
Subject Sample	# 1	0.200	04:37
Breath Volume		1.996	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:39
Subject Sample	# 2	0.193	04:40
Breath Volume		1.394	04:40
Air Blank		0.000	04:41
Control Test		0.082	04:41
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

---

Data Download Date/Time: 12/05/2020 04:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LUNDH,KAYLA  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 14:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:52

SUBJECT'S LAST NAME: MUDRA

DATE/TIME OF TEST: 11/17/2020 15:19:03

---

Diagnostics Check	OK	15:12
Air Blank	0.000	15:13
Control Test	0.082	15:13
Air Blank	0.000	15:14
Subject Sample # 1	0.000	15:14
Breath Volume	3.023	15:14
Air Blank	0.000	15:15
Air Blank	0.000	15:17
Subject Sample # 2	0.000	15:17
Breath Volume	2.933	15:17
Air Blank	0.000	15:18
Control Test	0.082	15:18
Air Blank	0.000	15:18
Diagnostics Check	OK	15:19

---

Data Download Date/Time: 12/05/2020 04:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CONNELL,CHASON  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: SAVOIE

DATE/TIME OF TEST: 11/17/2020 17:43:53

---

Diagnostics Check	OK	17:41	
Air Blank	0.000	17:41	
Control Test	0.082	17:41	
Air Blank	0.000	17:42	
Subject Sample	# 1	17:42	Subject Test Refused
Breath Volume	0.000	17:42	
Air Blank	0.000	17:42	
Control Test	0.082	17:43	
Air Blank	0.000	17:43	
Diagnostics Check	OK	17:43	

---

Data Download Date/Time: 12/05/2020 04:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CABASOS,ADRIAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: ALBITUS

DATE/TIME OF TEST: 11/18/2020 00:43:45

---

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.083	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.173	00:39
Breath Volume	2.890	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.183	00:42
Breath Volume	2.253	00:42
Air Blank	0.000	00:42
Control Test	0.082	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

---

Data Download Date/Time: 12/05/2020 04:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REFUSE, RONY,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: REFUSE, RONY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: WOODARD

DATE/TIME OF TEST: 11/18/2020 01:53:12

---

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:48	
Control Test	0.082	01:48	
Air Blank	0.000	01:49	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:52	
Control Test	0.082	01:52	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

---

Data Download Date/Time: 12/05/2020 04:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CHURCHILL,CAMERON  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: DICKMAN

DATE/TIME OF TEST: 11/18/2020 02:28:27

---

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:26	
Control Test	0.083	02:26	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:27	Subject Test Refused
Breath Volume	0.000	02:27	
Air Blank	0.000	02:27	
Control Test	0.082	02:27	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:28	

---

Data Download Date/Time: 12/05/2020 04:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BENTSCHNER,OLIVIA  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 11:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:14

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/18/2020 12:49:52

---

Diagnostics Check	OK	12:43
Air Blank	0.000	12:44
Control Test	0.083	12:44
Air Blank	0.000	12:44
Subject Sample # 1	0.000	12:45
Breath Volume	2.625	12:45
Air Blank	0.000	12:45
Air Blank	0.000	12:47
Subject Sample # 2	0.000	12:48
Breath Volume	2.179	12:48
Air Blank	0.000	12:48
Control Test	0.082	12:49
Air Blank	0.000	12:49
Diagnostics Check	OK	12:49

---

Data Download Date/Time: 12/05/2020 04:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LAMBORGHINI,P  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: VANDOREN

DATE/TIME OF TEST: 11/18/2020 22:12:57

---

Diagnostics Check	OK	22:06
Air Blank	0.000	22:06
Control Test	0.082	22:06
Air Blank	0.000	22:07
Subject Sample # 1	0.169	22:08
Breath Volume	1.871	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:10
Subject Sample # 2	0.160	22:11
Breath Volume	1.695	22:11
Air Blank	0.000	22:12
Control Test	0.081	22:12
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

---

Data Download Date/Time: 12/05/2020 04:10



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: ZSIDO,DANIELLE  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: SKUPIEN

DATE/TIME OF TEST: 11/20/2020 01:05:35

---

Diagnostics Check	OK	01:02	
Air Blank	0.000	01:03	
Control Test	0.082	01:03	
Air Blank	0.000	01:04	
Subject Sample	# 1	01:04	Subject Test Refused
Breath Volume	0.000	01:04	
Air Blank	0.000	01:04	
Control Test	0.081	01:04	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

---

Data Download Date/Time: 12/05/2020 04:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: DISHMEY,O  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 11/20/2020 20:59:49

---

Diagnostics Check	OK	20:53
Air Blank	0.000	20:53
Control Test	0.082	20:53
Air Blank	0.000	20:54
Subject Sample # 1	0.210	20:55
Breath Volume	1.582	20:55
Air Blank	0.000	20:56
Air Blank	0.000	20:57
Subject Sample # 2	0.208	20:58
Breath Volume	1.851	20:58
Air Blank	0.000	20:58
Control Test	0.081	20:59
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

---

Data Download Date/Time: 12/05/2020 04:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: COLMAN,M  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: BYRON

DATE/TIME OF TEST: 11/21/2020 00:51:00

---

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.082	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.000	00:46
Breath Volume	3.734	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.000	00:49
Breath Volume	2.812	00:49
Air Blank	0.000	00:50
Control Test	0.082	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:51

---

Data Download Date/Time: 12/05/2020 04:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BIGLIN,A  
ARRESTING OFFICER AGENCY: Treasure Island PD  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 11/21/2020 03:56:14

---

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.082	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.186	03:51
Breath Volume	1.972	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.190	03:54
Breath Volume	1.664	03:54
Air Blank	0.000	03:55
Control Test	0.079	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

---

Data Download Date/Time: 12/05/2020 04:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: JACOBSEN,STEPHEN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: STAFFORD

DATE/TIME OF TEST: 11/21/2020 05:06:37

---

Diagnostics Check	OK	05:03	
Air Blank	0.000	05:04	
Control Test	0.081	05:04	
Air Blank	0.000	05:05	
Subject Sample	# 1	05:05	Subject Test Refused
Breath Volume	0.000	05:05	
Air Blank	0.000	05:05	
Control Test	0.081	05:06	
Air Blank	0.000	05:06	
Diagnostics Check	OK	05:06	

---

Data Download Date/Time: 12/05/2020 04:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CHRISTIE,M  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 11/21/2020 21:13:50

---

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.082	21:08
Air Blank	0.000	21:08
Subject Sample # 1	0.000	21:09
Breath Volume	2.730	21:09
Air Blank	0.000	21:09
Air Blank	0.000	21:11
Subject Sample # 2	0.000	21:12
Breath Volume	3.527	21:12
Air Blank	0.000	21:12
Control Test	0.081	21:13
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

---

Data Download Date/Time: 12/05/2020 04:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: REIDY,T  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 11/21/2020 21:19:04

---

Diagnostics Check	OK	21:16	
Air Blank	0.000	21:16	
Control Test	0.081	21:17	
Air Blank	0.000	21:17	
Subject Sample	# 1	21:17	Subject Test Refused
Breath Volume	0.000	21:17	
Air Blank	0.000	21:18	
Control Test	0.081	21:18	
Air Blank	0.000	21:18	
Diagnostics Check	OK	21:19	

---

Data Download Date/Time: 12/05/2020 04:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: GARDNER,D  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GULLEDGE

DATE/TIME OF TEST: 11/22/2020 01:50:46

---

Diagnostics Check		OK	01:40
Air Blank		0.000	01:41
Control Test		0.083	01:41
Air Blank		0.000	01:42
Subject Sample	# 1	0.174	01:45
Breath Volume		2.277	01:45
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.182	01:49
Breath Volume		2.316	01:49
Air Blank		0.000	01:49
Control Test		0.080	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

---

Data Download Date/Time: 12/05/2020 04:10



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: JACOBSEN,STEPHEN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: HAPKE

DATE/TIME OF TEST: 11/22/2020 03:31:43

---

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.082	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.156	03:27
Breath Volume	2.847	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.163	03:30
Breath Volume	2.156	03:30
Air Blank	0.000	03:30
Control Test	0.081	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

---

Data Download Date/Time: 12/05/2020 04:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBY,LARRY,K  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MANGANARO,JESSICA  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: GADOMSKI

DATE/TIME OF TEST: 11/22/2020 05:34:34

---

Diagnostics Check	OK	05:28
Air Blank	0.000	05:28
Control Test	0.082	05:28
Air Blank	0.000	05:29
Subject Sample # 1	0.128	05:30
Breath Volume	2.640	05:30
Air Blank	0.000	05:30
Air Blank	0.000	05:32
Subject Sample # 2	0.131	05:33
Breath Volume	2.531	05:33
Air Blank	0.000	05:33
Control Test	0.082	05:33
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

---

Data Download Date/Time: 12/05/2020 04:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LANGE,KYLE  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: VSETECKA

DATE/TIME OF TEST: 11/22/2020 19:43:58

---

Diagnostics Check	OK	19:38
Air Blank	0.000	19:38
Control Test	0.082	19:38
Air Blank	0.000	19:39
Subject Sample # 1	0.013	19:39
Breath Volume	2.875	19:39
Air Blank	0.000	19:40
Air Blank	0.000	19:42
Subject Sample # 2	0.012	19:42
Breath Volume	2.019	19:42
Air Blank	0.000	19:43
Control Test	0.082	19:43
Air Blank	0.000	19:43
Diagnostics Check	OK	19:43

---

Data Download Date/Time: 12/05/2020 04:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CONEBY,MEGAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: MONEY

DATE/TIME OF TEST: 11/22/2020 23:01:57

---

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.082	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.000	22:57
Breath Volume	2.437	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.000	23:00
Breath Volume	1.800	23:00
Air Blank	0.000	23:01
Control Test	0.082	23:01
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

---

Data Download Date/Time: 12/05/2020 04:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: EDWARDS,PATRICK  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MAAFU

DATE/TIME OF TEST: 11/23/2020 02:18:36

---

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.082	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.154	02:14
Breath Volume	2.851	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.152	02:17
Breath Volume	1.953	02:17
Air Blank	0.000	02:17
Control Test	0.081	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

---

Data Download Date/Time: 12/05/2020 04:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARTWICK,MICHAEL,R  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: VANGELL,J  
ARRESTING OFFICER AGENCY: Treasure Island PD  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/23/2020 04:28:18

---

Diagnostics Check		OK	04:21
Air Blank		0.000	04:22
Control Test		0.081	04:22
Air Blank		0.000	04:23
Subject Sample	# 1	0.283	04:23
Breath Volume		2.566	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.289	04:26
Breath Volume		1.769	04:26
Air Blank		0.000	04:27
Control Test		0.079	04:27
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

---

Data Download Date/Time: 12/05/2020 04:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EASTTY,MARK,L  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: NIXON,AARON  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: PFISTER

DATE/TIME OF TEST: 11/25/2020 20:46:03

---

Diagnostics Check	OK	20:39
Air Blank	0.000	20:40
Control Test	0.082	20:40
Air Blank	0.000	20:41
Subject Sample # 1	0.296	20:41
Breath Volume	2.058	20:41
Air Blank	0.000	20:42
Air Blank	0.000	20:44
Subject Sample # 2	0.292	20:44
Breath Volume	1.742	20:44
Air Blank	0.000	20:45
Control Test	0.081	20:45
Air Blank	0.000	20:45
Diagnostics Check	OK	20:46

---

Data Download Date/Time: 12/05/2020 04:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EASTTY,MARK,L  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: ZSIDO,DANIELLE  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: SINDONI

DATE/TIME OF TEST: 11/25/2020 22:04:56

---

Diagnostics Check	OK	21:58
Air Blank	0.000	21:59
Control Test	0.082	21:59
Air Blank	0.000	21:59
Subject Sample # 1	0.212	22:00
Breath Volume	1.800	22:00
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.199	22:03
Breath Volume	1.539	22:03
Air Blank	0.000	22:04
Control Test	0.080	22:04
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

---

Data Download Date/Time: 12/05/2020 04:11



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: TEMPLE,ELIZABETH  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 11/26/2020 03:36:11

---

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:33	
Control Test	0.082	03:33	
Air Blank	0.000	03:34	
Subject Sample	# 1	03:34	Subject Test Refused
Breath Volume	0.000	03:34	
Air Blank	0.000	03:35	
Control Test	0.083	03:35	
Air Blank	0.000	03:36	
Diagnostics Check	OK	03:36	

---

Data Download Date/Time: 12/05/2020 04:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STYLES,LAUREN,N  
BREATH TEST OPERATOR AGENCY: Treasure Island PD  
ARRESTING OFFICER: STYLES,LAUREN  
ARRESTING OFFICER AGENCY: Treasure Island PD  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: SAVOIE

DATE/TIME OF TEST: 11/26/2020 04:46:35

---

Diagnostics Check		OK	04:40
Air Blank		0.000	04:40
Control Test		0.083	04:41
Air Blank		0.000	04:41
Subject Sample	# 1	0.183	04:42
Breath Volume		2.238	04:42
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.173	04:45
Breath Volume		2.265	04:45
Air Blank		0.000	04:45
Control Test		0.082	04:45
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

---

Data Download Date/Time: 12/05/2020 04:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: SCHOMP,DALTON  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 11/27/2020 01:43:19

---

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.082	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.189	01:38
Breath Volume	1.500	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:40
Subject Sample # 2	0.185	01:41
Breath Volume	1.605	01:41
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

---

Data Download Date/Time: 12/05/2020 04:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EASTTY,MARK,L  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: HIRSHMAN,BRIAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 11/27/2020 03:46:57

---

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:44	
Control Test	0.082	03:44	
Air Blank	0.000	03:45	
Subject Sample	# 1	03:45	Subject Test Refused
Breath Volume	0.000	03:45	
Air Blank	0.000	03:46	
Control Test	0.082	03:46	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

---

Data Download Date/Time: 12/05/2020 04:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EASTTY,MARK,L  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: ZSIDO,DANIELLE  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 11/27/2020 04:41:09

---

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.082	04:35
Air Blank	0.000	04:36
Subject Sample # 1	0.142	04:36
Breath Volume	1.558	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.129	04:39
Breath Volume	1.417	04:39
Air Blank	0.000	04:40
Control Test	0.081	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

---

Data Download Date/Time: 12/05/2020 04:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNAVELY,KENT,J  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: ARLINGTON,DEP  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: NOVAK

DATE/TIME OF TEST: 11/27/2020 16:43:37

---

Diagnostics Check	OK	16:30	
Air Blank	0.000	16:31	
Control Test	0.082	16:31	
Air Blank	0.000	16:32	
Subject Sample	# 1	16:35	No Sample Provided
Breath Volume	0.000	16:35	
Air Blank	0.000	16:35	
Air Blank	0.000	16:37	
Subject Sample	# 2	0.101	16:38
Breath Volume	1.843	16:38	
Air Blank	0.000	16:39	
Air Blank	0.000	16:41	
Subject Sample	# 3	0.096	16:42
Breath Volume	1.925	16:42	
Air Blank	0.000	16:42	
Control Test	0.082	16:43	
Air Blank	0.000	16:43	
Diagnostics Check	OK	16:43	

---

Data Download Date/Time: 12/05/2020 04:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANEY,DAMON,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LANEY,DAMON  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 11/28/2020 04:15:22

---

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:00	
Control Test	0.082	04:01	
Air Blank	0.000	04:01	
Subject Sample # 1	0.089	04:03	
Breath Volume	1.113	04:03	
Air Blank	0.000	04:03	
Air Blank	0.000	04:05	
Subject Sample # 2	0.085	04:08	Volume Not Met
Breath Volume	1.027	04:08	
Air Blank	0.000	04:09	
Air Blank	0.000	04:11	
Subject Sample # 3	0.090	04:13	
Breath Volume	1.277	04:13	
Air Blank	0.000	04:14	
Control Test	0.080	04:14	
Air Blank	0.000	04:15	
Diagnostics Check	OK	04:15	

---

Data Download Date/Time: 12/05/2020 04:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVIS  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LONGMIRE,BENJAMIN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 11/28/2020 05:01:31

---

Diagnostics Check	OK	04:55
Air Blank	0.000	04:55
Control Test	0.081	04:56
Air Blank	0.000	04:56
Subject Sample # 1	0.118	04:57
Breath Volume	2.441	04:57
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample # 2	0.112	05:00
Breath Volume	1.812	05:00
Air Blank	0.000	05:00
Control Test	0.082	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

---

Data Download Date/Time: 12/05/2020 04:11



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KASELAK,STEPHEN,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: FIGUEROA,CYNTHIA  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: RODRIGUEZ-PEREZ

DATE/TIME OF TEST: 11/28/2020 05:40:47

---

Diagnostics Check	OK	05:32
Air Blank	0.000	05:33
Control Test	0.082	05:33
Air Blank	0.000	05:33
Subject Sample # 1	0.104	05:36
Breath Volume	2.500	05:36
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.104	05:39
Breath Volume	1.957	05:39
Air Blank	0.000	05:39
Control Test	0.081	05:40
Air Blank	0.000	05:40
Diagnostics Check	OK	05:40

---

Data Download Date/Time: 12/05/2020 04:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVI,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAKE,LEVI  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 11/29/2020 03:08:23

---

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.083	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.034	03:04
Breath Volume	1.914	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.034	03:06
Breath Volume	2.160	03:06
Air Blank	0.000	03:07
Control Test	0.082	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

---

Data Download Date/Time: 12/05/2020 04:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: DUNNING,RYAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: HOVAN

DATE/TIME OF TEST: 11/30/2020 21:19:31

---

Diagnostics Check	OK	21:13
Air Blank	0.000	21:13
Control Test	0.080	21:13
Air Blank	0.000	21:14
Subject Sample # 1	0.309	21:14
Breath Volume	2.609	21:14
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.315	21:17
Breath Volume	2.726	21:17
Air Blank	0.000	21:18
Control Test	0.080	21:18
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

---

Data Download Date/Time: 12/05/2020 04:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOWATT,NATHAN,E  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MOWATT,NATHAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: BRECHTEL

DATE/TIME OF TEST: 11/30/2020 23:35:32

---

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.082	23:29
Air Blank	0.000	23:30
Subject Sample # 1	0.196	23:30
Breath Volume	1.308	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.189	23:33
Breath Volume	1.152	23:33
Air Blank	0.000	23:34
Control Test	0.081	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

---

Data Download Date/Time: 12/05/2020 04:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLEN,BRAD,A  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: ALLEN,BRAD  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 11/01/2020 05:41:39

---

Diagnostics Check	OK	05:34
Air Blank	0.000	05:35
Control Test	0.080	05:35
Air Blank	0.000	05:35
Subject Sample # 1	0.163	05:37
Breath Volume	2.972	05:37
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample # 2	0.164	05:40
Breath Volume	2.292	05:40
Air Blank	0.000	05:40
Control Test	0.080	05:41
Air Blank	0.000	05:41
Diagnostics Check	OK	05:41

---

Data Download Date/Time: 11/29/2020 00:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNEAU,OFFICER,  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: BRUNEAU,OFFICER  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 11/06/2020 00:12:40

---

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.079	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.177	00:08
Breath Volume	2.933	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.168	00:11
Breath Volume	2.730	00:11
Air Blank	0.000	00:11
Control Test	0.079	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

---

Data Download Date/Time: 11/29/2020 00:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNEAU,OFFICER,  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: BRUNEAU,OFFICE  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: RAMIREZ-VALADEZ

DATE/TIME OF TEST: 11/06/2020 23:27:36

---

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.078	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.228	23:22
Breath Volume	2.089	23:22
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.217	23:25
Breath Volume	3.027	23:25
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

---

Data Download Date/Time: 11/29/2020 00:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REED,CHAD,R  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: REED,CHAD  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MEDINA COLON

DATE/TIME OF TEST: 11/08/2020 01:57:27

---

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:46	
Control Test	0.079	01:46	
Air Blank	0.000	01:47	
Subject Sample	# 1	01:50	No Sample Provided
Breath Volume	0.000	01:50	
Air Blank	0.000	01:50	
Air Blank	0.000	01:52	
Subject Sample	# 2	01:56	No Sample Provided
Breath Volume	0.000	01:56	
Air Blank	0.000	01:56	
Control Test	0.079	01:56	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

---

Data Download Date/Time: 11/29/2020 00:12



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNEAU,OFFICER,  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: BRUNEAU,OFFICER  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 11/08/2020 03:16:39

---

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.078	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.197	03:12
Breath Volume	1.183	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.199	03:15
Breath Volume	1.589	03:15
Air Blank	0.000	03:15
Control Test	0.078	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

---

Data Download Date/Time: 11/29/2020 00:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EDMONDS,RICHARD,J  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: HUNT,ASHLEY  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/10/2020 03:37:21

---

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.079	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.126	03:32
Breath Volume	1.937	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.122	03:35
Breath Volume	3.171	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

---

Data Download Date/Time: 11/29/2020 00:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNEAU,OFFICER,  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: BRUNEAU,OFFICER  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: COOLEY

DATE/TIME OF TEST: 11/14/2020 22:47:33

---

Diagnostics Check	OK	22:40
Air Blank	0.000	22:41
Control Test	0.079	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.193	22:42
Breath Volume	3.734	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.192	22:45
Breath Volume	3.980	22:45
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

---

Data Download Date/Time: 11/29/2020 00:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: NEGERSMITH,DANIEL  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: CAMPANARO

DATE/TIME OF TEST: 11/14/2020 23:08:40

---

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.079	23:02
Air Blank	0.000	23:02
Subject Sample # 1	0.087	23:03
Breath Volume	3.562	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.082	23:07
Breath Volume	3.074	23:07
Air Blank	0.000	23:07
Control Test	0.079	23:08
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

---

Data Download Date/Time: 11/29/2020 00:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICHMOND, MATTHEW, D  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: RICHMOND, MATTHEW  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 08:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:53

SUBJECT'S LAST NAME: BARTRON

DATE/TIME OF TEST: 11/15/2020 09:29:01

---

Diagnostics Check	OK	09:23
Air Blank	0.000	09:23
Control Test	0.080	09:23
Air Blank	0.000	09:24
Subject Sample # 1	0.000	09:24
Breath Volume	4.121	09:24
Air Blank	0.000	09:25
Air Blank	0.000	09:27
Subject Sample # 2	0.000	09:27
Breath Volume	3.882	09:27
Air Blank	0.000	09:28
Control Test	0.082	09:28
Air Blank	0.000	09:28
Diagnostics Check	OK	09:29

---

Data Download Date/Time: 11/29/2020 00:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNEAU,OFFICER,  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: BRUNEAU,OFFICER  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: OVERSTREET

DATE/TIME OF TEST: 11/21/2020 21:14:23

---

Diagnostics Check		OK	21:06
Air Blank		0.000	21:06
Control Test		0.079	21:07
Air Blank		0.000	21:07
Subject Sample	# 1	0.255	21:09
Breath Volume		2.218	21:09
Air Blank		0.000	21:10
Air Blank		0.000	21:11
Subject Sample	# 2	0.258	21:12
Breath Volume		3.125	21:12
Air Blank		0.000	21:13
Control Test		0.079	21:13
Air Blank		0.000	21:14
Diagnostics Check		OK	21:14

---

Data Download Date/Time: 11/29/2020 00:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EDMONDS,RICHARD,J  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: HUNT,ASHLEY  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/24/2020 05:23:30

---

Diagnostics Check		OK	05:14
Air Blank		0.000	05:14
Control Test		0.080	05:15
Air Blank		0.000	05:15
Subject Sample	# 1	0.237	05:16
Breath Volume		2.242	05:16
Air Blank		0.000	05:16
Air Blank		0.000	05:18
Subject Sample	# 2	0.213	05:18
Breath Volume		1.636	05:18
Air Blank		0.000	05:19
Air Blank		0.000	05:21
Subject Sample	# 3	0.225	05:21
Breath Volume		1.949	05:21
Air Blank		0.000	05:22
Control Test		0.078	05:22
Air Blank		0.000	05:23
Diagnostics Check		OK	05:23

---

Data Download Date/Time: 11/29/2020 00:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUIS,JUSTIN,  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: BUIS,JUSTIN  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 11/24/2020 20:36:21

---

Diagnostics Check	OK	20:20	
Air Blank	0.000	20:21	
Control Test	0.079	20:21	
Air Blank	0.000	20:22	
Subject Sample # 1	0.222	20:23	
Breath Volume	2.496	20:23	
Air Blank	0.000	20:24	
Air Blank	0.000	20:26	
Subject Sample # 2		20:29	No Sample Provided
Breath Volume	0.000	20:29	
Air Blank	0.000	20:29	
Air Blank	0.000	20:31	
Subject Sample # 3		20:34	No Sample Provided
Breath Volume	0.000	20:34	
Air Blank	0.000	20:35	
Control Test	0.079	20:35	
Air Blank	0.000	20:36	
Diagnostics Check	OK	20:36	

---

Data Download Date/Time: 11/29/2020 00:12



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNEAU,OFFICER,  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: BRUNEAU,OFFICER  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 10/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: BUI  
ARREST TIME: 19:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: PORTILLO

DATE/TIME OF TEST: 11/28/2020 20:10:19

---

Diagnostics Check	OK	20:04	
Air Blank	0.000	20:04	
Control Test	0.080	20:05	
Air Blank	0.000	20:05	
Subject Sample # 1	0.096	20:08	
Breath Volume	1.984	20:08	
Air Blank		20:09	Purge Fail
Air Blank	0.000	20:10	

---

Data Download Date/Time: 11/29/2020 00:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,CHUCK,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FISETTE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 11/01/2020 01:06:52

---

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:04	
Control Test	0.082	01:04	
Air Blank	0.000	01:05	
Subject Sample	# 1	01:05	Subject Test Refused
Breath Volume	0.000	01:05	
Air Blank	0.000	01:05	
Control Test	0.082	01:06	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:06	

---

Data Download Date/Time: 11/17/2020 10:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,CHUCK,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: BLACKLIDGE

DATE/TIME OF TEST: 11/01/2020 02:47:29

---

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:45	
Control Test	0.082	02:45	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:46	Subject Test Refused
Breath Volume	0.000	02:46	
Air Blank	0.000	02:46	
Control Test	0.082	02:46	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

---

Data Download Date/Time: 11/17/2020 10:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,CHUCK,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LUGO,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: POPE

DATE/TIME OF TEST: 11/02/2020 23:53:36

---

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.083	23:48
Air Blank	0.000	23:48
Subject Sample # 1	0.030	23:49
Breath Volume	1.464	23:49
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.031	23:52
Breath Volume	1.792	23:52
Air Blank	0.000	23:52
Control Test	0.082	23:53
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

---

Data Download Date/Time: 11/17/2020 10:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TUCKER,R,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: ROBINSON,T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: MACMASTER

DATE/TIME OF TEST: 11/04/2020 01:10:13

---

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.083	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.000	01:05
Breath Volume	4.242	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.000	01:08
Breath Volume	4.687	01:08
Air Blank	0.000	01:09
Control Test	0.083	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

---

Data Download Date/Time: 11/17/2020 10:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,CHUCK,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 11/09/2020 19:39:47

---

Diagnostics Check	OK	19:33
Air Blank	0.000	19:34
Control Test	0.082	19:34
Air Blank	0.000	19:35
Subject Sample # 1	0.000	19:35
Breath Volume	2.281	19:35
Air Blank	0.000	19:36
Air Blank	0.000	19:38
Subject Sample # 2	0.000	19:38
Breath Volume	2.328	19:38
Air Blank	0.000	19:38
Control Test	0.082	19:39
Air Blank	0.000	19:39
Diagnostics Check	OK	19:39

---

Data Download Date/Time: 11/17/2020 10:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,CHUCK,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MOELLER,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 11/10/2020 01:20:39

---

Diagnostics Check	OK	01:14
Air Blank	0.000	01:15
Control Test	0.081	01:15
Air Blank	0.000	01:15
Subject Sample # 1	0.014	01:16
Breath Volume	2.449	01:16
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.014	01:19
Breath Volume	1.746	01:19
Air Blank	0.000	01:19
Control Test	0.081	01:20
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

---

Data Download Date/Time: 11/17/2020 10:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,CHUCK,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MOELLER,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 11/10/2020 19:07:13

---

Diagnostics Check	OK	19:00
Air Blank	0.000	19:01
Control Test	0.082	19:01
Air Blank	0.000	19:02
Subject Sample # 1	0.260	19:02
Breath Volume	2.855	19:02
Air Blank	0.000	19:03
Air Blank	0.000	19:05
Subject Sample # 2	0.256	19:05
Breath Volume	2.730	19:05
Air Blank	0.000	19:06
Control Test	0.081	19:06
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

---

Data Download Date/Time: 11/17/2020 10:06



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,CHUCK,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DURHAM,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 11/11/2020 02:45:52

---

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.082	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.157	02:41
Breath Volume	1.839	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.162	02:44
Breath Volume	2.082	02:44
Air Blank	0.000	02:44
Control Test	0.081	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

---

Data Download Date/Time: 11/17/2020 10:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,CHUCK,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DESOUZA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: MOTEN

DATE/TIME OF TEST: 11/12/2020 04:43:23

---

Diagnostics Check	OK	04:36
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:38
Subject Sample # 1	0.099	04:38
Breath Volume	2.058	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.094	04:41
Breath Volume	1.386	04:41
Air Blank	0.000	04:42
Control Test	0.082	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

---

Data Download Date/Time: 11/17/2020 10:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TUCKER,R,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KEEN,C  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 11/13/2020 04:08:13

---

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.082	04:02
Air Blank	0.000	04:03
Subject Sample # 1	0.112	04:03
Breath Volume	2.691	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.108	04:06
Breath Volume	2.796	04:06
Air Blank	0.000	04:07
Control Test	0.081	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

---

Data Download Date/Time: 11/17/2020 10:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TUCKER,R,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: OSBORNE,I  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: AWET

DATE/TIME OF TEST: 11/14/2020 02:58:55

---

Diagnostics Check		OK	02:52
Air Blank		0.000	02:52
Control Test		0.082	02:52
Air Blank		0.000	02:53
Subject Sample	# 1	0.263	02:54
Breath Volume		2.417	02:54
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.258	02:57
Breath Volume		1.140	02:57
Air Blank		0.000	02:57
Control Test		0.080	02:58
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

---

Data Download Date/Time: 11/17/2020 10:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOVACS, TYLER, K  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HILLIER, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 11/15/2020 03:33:42

---

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.082	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.000	03:29
Breath Volume	2.929	03:29
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.000	03:32
Breath Volume	2.277	03:32
Air Blank	0.000	03:32
Control Test	0.082	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

---

Data Download Date/Time: 11/17/2020 10:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TUCKER,R,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DURHAM,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: MOULTRIE

DATE/TIME OF TEST: 11/15/2020 20:23:46

---

Diagnostics Check	OK	20:17
Air Blank	0.000	20:17
Control Test	0.082	20:18
Air Blank	0.000	20:18
Subject Sample # 1	0.304	20:19
Breath Volume	1.781	20:19
Air Blank	0.000	20:20
Air Blank	0.000	20:21
Subject Sample # 2	0.306	20:22
Breath Volume	1.644	20:22
Air Blank	0.000	20:22
Control Test	0.081	20:23
Air Blank	0.000	20:23
Diagnostics Check	OK	20:23

---

Data Download Date/Time: 11/17/2020 10:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TUCKER,R,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: YOUNG,S  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: GATLUAK

DATE/TIME OF TEST: 11/16/2020 00:08:15

---

Diagnostics Check	OK	00:01
Air Blank	0.000	00:02
Control Test	0.082	00:02
Air Blank	0.000	00:03
Subject Sample # 1	0.320	00:03
Breath Volume	1.921	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.308	00:06
Breath Volume	1.851	00:06
Air Blank	0.000	00:07
Control Test	0.081	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

---

Data Download Date/Time: 11/17/2020 10:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TUCKER,R,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: OCONNELL,S  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: CONRAD

DATE/TIME OF TEST: 11/16/2020 02:24:07

---

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.082	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.159	02:19
Breath Volume	2.125	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:21
Subject Sample # 2	0.168	02:22
Breath Volume	3.015	02:22
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

---

Data Download Date/Time: 11/17/2020 10:07



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUTTON, BRADEN, J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LUGO, R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: SHIRLEY

DATE/TIME OF TEST: 11/19/2020 23:01:00

---

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.082	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.000	22:56
Breath Volume	3.710	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.000	22:59
Breath Volume	2.195	22:59
Air Blank	0.000	23:00
Control Test	0.081	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:01

---

Data Download Date/Time: 12/09/2020 11:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUTTON, BRADEN, J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HAWKINS, A  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: DANIELS JR

DATE/TIME OF TEST: 11/20/2020 01:21:29

---

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.081	01:15
Air Blank		0.000	01:16
Subject Sample	# 1	0.201	01:16
Breath Volume		2.031	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.197	01:19
Breath Volume		2.097	01:19
Air Blank		0.000	01:20
Control Test		0.081	01:20
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

---

Data Download Date/Time: 12/09/2020 11:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUTTON, BRADEN, J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LANGLOIS, J  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 11/20/2020 03:30:05

---

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:27	
Control Test	0.082	03:27	
Air Blank	0.000	03:28	
Subject Sample	# 1	03:28	Subject Test Refused
Breath Volume	0.000	03:28	
Air Blank	0.000	03:29	
Control Test	0.082	03:29	
Air Blank	0.000	03:29	
Diagnostics Check	OK	03:30	

---

Data Download Date/Time: 12/09/2020 11:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUTTON, BRADEN, J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DESOUZA, A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: CORTES-GONZALEZ

DATE/TIME OF TEST: 11/20/2020 04:37:25

---

Diagnostics Check	OK	04:34	
Air Blank	0.000	04:34	
Control Test	0.081	04:35	
Air Blank	0.000	04:35	
Subject Sample	# 1	04:36	Subject Test Refused
Breath Volume	0.000	04:36	
Air Blank	0.000	04:36	
Control Test	0.082	04:36	
Air Blank	0.000	04:37	
Diagnostics Check	OK	04:37	

---

Data Download Date/Time: 12/09/2020 11:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUTTON, BRADEN, J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FISETTE, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: GODWIN

DATE/TIME OF TEST: 11/21/2020 00:56:04

---

Diagnostics Check		OK	00:49
Air Blank		0.000	00:49
Control Test		0.082	00:50
Air Blank		0.000	00:50
Subject Sample	# 1	0.103	00:51
Breath Volume		2.062	00:51
Air Blank		0.000	00:52
Air Blank		0.000	00:53
Subject Sample	# 2	0.110	00:54
Breath Volume		2.660	00:54
Air Blank		0.000	00:55
Control Test		0.080	00:55
Air Blank		0.000	00:55
Diagnostics Check		OK	00:56

---

Data Download Date/Time: 12/09/2020 11:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUTTON, BRADEN, J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WATKINS, P  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: GAINNEY

DATE/TIME OF TEST: 11/21/2020 01:32:13

---

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.081	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.132	01:27
Breath Volume	3.457	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.121	01:30
Breath Volume	1.839	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

---

Data Download Date/Time: 12/09/2020 11:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUTTON, BRADEN, J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: RODRIGUEZ, M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: BAUMAN

DATE/TIME OF TEST: 11/21/2020 02:24:36

---

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.082	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.168	02:19
Breath Volume	1.343	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.171	02:22
Breath Volume	1.273	02:22
Air Blank	0.000	02:23
Control Test	0.082	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

---

Data Download Date/Time: 12/09/2020 11:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TUCKER,R,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HILLIER,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: LOWERY

DATE/TIME OF TEST: 11/22/2020 22:42:53

---

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:40	
Control Test	0.082	22:40	
Air Blank	0.000	22:41	
Subject Sample	# 1	22:41	Subject Test Refused
Breath Volume	0.000	22:41	
Air Blank	0.000	22:41	
Control Test	0.082	22:42	
Air Blank	0.000	22:42	
Diagnostics Check	OK	22:42	

---

Data Download Date/Time: 12/09/2020 11:59



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUTTON, BRADEN, J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MOELLER, R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: CHANCE

DATE/TIME OF TEST: 11/26/2020 23:04:35

---

Diagnostics Check	OK	23:01	
Air Blank	0.000	23:02	
Control Test	0.083	23:02	
Air Blank	0.000	23:02	
Subject Sample	# 1	23:03	Subject Test Refused
Breath Volume	0.000	23:03	
Air Blank	0.000	23:03	
Control Test	0.083	23:03	
Air Blank	0.000	23:04	
Diagnostics Check	OK	23:04	

---

Data Download Date/Time: 12/09/2020 11:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G  
BREATH TEST OPERATOR AGENCY: Dade City PD  
ARRESTING OFFICER: AMATRUDA,CHRIS  
ARRESTING OFFICER AGENCY: Dade City PD  
LAST AGENCY INSPECTION: 10/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: OSORIO-SANCHEZ

DATE/TIME OF TEST: 11/01/2020 01:34:10

---

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.080	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.170	01:26
Breath Volume	1.773	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:28
Subject Sample # 2	0.196	01:29
Breath Volume	2.085	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 3	0.182	01:32
Breath Volume	2.242	01:32
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

---

Data Download Date/Time: 11/23/2020 09:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OCASIO,DEBBIE,L  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: JONES,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 10/14/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: HOPKINS

DATE/TIME OF TEST: 11/06/2020 00:58:44

---

Diagnostics Check	OK	00:49
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:50
Subject Sample # 1	0.198	00:52
Breath Volume	2.449	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.186	00:56
Breath Volume	1.492	00:56
Air Blank	0.000	00:57
Control Test	0.079	00:58
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

---

Data Download Date/Time: 11/23/2020 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHUCK,SCOTT,M  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: FITTS,DEREK  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 11:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:43

SUBJECT'S LAST NAME: SANTO

DATE/TIME OF TEST: 11/26/2020 12:15:20

---

Diagnostics Check	OK	12:08
Air Blank	0.000	12:08
Control Test	0.081	12:09
Air Blank	0.000	12:09
Subject Sample # 1	0.195	12:10
Breath Volume	3.593	12:10
Air Blank	0.000	12:11
Air Blank	0.000	12:13
Subject Sample # 2	0.192	12:13
Breath Volume	3.816	12:13
Air Blank	0.000	12:14
Control Test	0.080	12:14
Air Blank	0.000	12:15
Diagnostics Check	OK	12:15

---

Data Download Date/Time: 12/18/2020 19:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHUCK,SCOTT,M  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: YANZER,NOAH  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 08:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:36

SUBJECT'S LAST NAME: KUNTZ

DATE/TIME OF TEST: 11/30/2020 09:06:59

---

Diagnostics Check		OK	09:00
Air Blank		0.000	09:00
Control Test		0.082	09:01
Air Blank		0.000	09:01
Subject Sample	# 1	0.000	09:02
Breath Volume		4.691	09:02
Air Blank		0.000	09:03
Air Blank		0.000	09:05
Subject Sample	# 2	0.000	09:05
Breath Volume		3.210	09:05
Air Blank		0.000	09:06
Control Test		0.081	09:06
Air Blank		0.000	09:06
Diagnostics Check		OK	09:06

---

Data Download Date/Time: 12/18/2020 19:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH-LONG,JEFFERSON,D  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: COMSTOCK,MICHAEL  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 11/30/2020 19:02:39

---

Diagnostics Check	OK	18:56
Air Blank	0.000	18:57
Control Test	0.081	18:57
Air Blank	0.000	18:57
Subject Sample # 1	0.198	18:58
Breath Volume	2.062	18:58
Air Blank	0.000	18:58
Air Blank	0.000	19:00
Subject Sample # 2	0.189	19:01
Breath Volume	3.652	19:01
Air Blank	0.000	19:01
Control Test	0.080	19:02
Air Blank	0.000	19:02
Diagnostics Check	OK	19:02

---

Data Download Date/Time: 12/18/2020 19:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALBRITTON,GARRETT,M  
BREATH TEST OPERATOR AGENCY: Hardee County SO  
ARRESTING OFFICER: BRANDEBERRY,JEREMY  
ARRESTING OFFICER AGENCY: Hardee County SO  
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1076482  
DRY GAS EXPIRY DATE: 01/15/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:24

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 11/08/2020 06:52:09

---

Diagnostics Check	OK	06:44
Air Blank	0.000	06:45
Control Test	0.078	06:45
Air Blank	0.000	06:46
Subject Sample # 1	0.095	06:47
Breath Volume	1.500	06:47
Air Blank	0.000	06:48
Air Blank	0.000	06:49
Subject Sample # 2	0.094	06:50
Breath Volume	1.835	06:50
Air Blank	0.000	06:51
Control Test	0.076	06:51
Air Blank	0.000	06:52
Diagnostics Check	OK	06:52

---

Data Download Date/Time: 12/03/2020 09:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,RAYNALDO,  
BREATH TEST OPERATOR AGENCY: Hardee County SO  
ARRESTING OFFICER: BROWN,JASON  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1076482  
DRY GAS EXPIRY DATE: 01/15/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 11/16/2020 22:46:50

---

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.077	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.249	22:42
Breath Volume	2.753	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:44
Subject Sample # 2	0.245	22:45
Breath Volume	2.957	22:45
Air Blank	0.000	22:45
Control Test	0.077	22:46
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

---

Data Download Date/Time: 12/03/2020 09:07



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BANDY,JAMIE,L  
BREATH TEST OPERATOR AGENCY: Hardee County SO  
ARRESTING OFFICER: TYLER,ETHAN  
ARRESTING OFFICER AGENCY: Hardee County SO  
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1076482  
DRY GAS EXPIRY DATE: 01/15/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: MORENO CARCAMO

DATE/TIME OF TEST: 11/25/2020 21:56:46

---

Diagnostics Check		OK	21:44	
Air Blank		0.000	21:44	
Control Test		0.078	21:45	
Air Blank		0.000	21:45	
Subject Sample	# 1	0.174	21:48	Volume Not Met
Breath Volume		0.390	21:48	
Air Blank		0.000	21:49	
Air Blank		0.000	21:51	
Subject Sample	# 2	0.177	21:51	
Breath Volume		2.046	21:51	
Air Blank		0.000	21:52	
Air Blank		0.000	21:54	
Subject Sample	# 3	0.178	21:55	
Breath Volume		1.785	21:55	
Air Blank		0.000	21:55	
Control Test		0.077	21:56	
Air Blank		0.000	21:56	
Diagnostics Check		OK	21:56	

---

Data Download Date/Time: 12/03/2020 09:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, RAYNALDO,  
BREATH TEST OPERATOR AGENCY: Hardee County SO  
ARRESTING OFFICER: ADLER, JAMES  
ARRESTING OFFICER AGENCY: Hardee County SO  
LAST AGENCY INSPECTION: 10/19/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1076482  
DRY GAS EXPIRY DATE: 01/15/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: PATINO

DATE/TIME OF TEST: 11/26/2020 04:41:26

---

Diagnostics Check	OK	04:37	
Air Blank	0.000	04:37	
Control Test	0.078	04:37	
Air Blank	0.000	04:38	
Subject Sample	# 1	04:40	Subject Test Refused
Breath Volume	0.000	04:40	
Air Blank	0.000	04:40	
Control Test	0.078	04:40	
Air Blank	0.000	04:41	
Diagnostics Check	OK	04:41	

---

Data Download Date/Time: 12/03/2020 09:07



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 10/05/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22419080A3

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 23:31

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: PEAK

DATE/TIME OF TEST: 11/01/2020 01:00:49

---

Diagnostics Check	OK	00:49	
Air Blank	0.000	00:49	
Control Test	0.080	00:50	
Air Blank	0.000	00:50	
Subject Sample	# 1	00:53	No Sample Provided
Breath Volume	0.000	00:53	
Air Blank	0.000	00:54	
Air Blank	0.000	00:56	
Subject Sample	# 2	00:59	No Sample Provided
Breath Volume	0.000	00:59	
Air Blank	0.000	00:59	
Control Test	0.081	01:00	
Air Blank	0.000	01:00	
Diagnostics Check	OK	01:00	

---

Data Download Date/Time: 11/12/2020 10:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUICK,SHANE,C  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: QUICK,SHANE  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: JEFFREY

DATE/TIME OF TEST: 11/01/2020 02:54:00

---

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:51	
Control Test	0.081	02:51	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Control Test	0.081	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:54	

---

Data Download Date/Time: 11/12/2020 10:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: MAJOR

DATE/TIME OF TEST: 11/01/2020 04:34:49

---

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.080	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.124	04:29
Breath Volume	1.585	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.130	04:33
Breath Volume	2.378	04:33
Air Blank	0.000	04:33
Control Test	0.079	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

---

Data Download Date/Time: 11/12/2020 10:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 10/05/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22419080A3

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 23:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: NGO

DATE/TIME OF TEST: 11/04/2020 00:02:27

---

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.082	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.192	23:58
Breath Volume	2.238	23:58
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.188	00:00
Breath Volume	2.617	00:00
Air Blank	0.000	00:01
Control Test	0.081	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

---

Data Download Date/Time: 11/12/2020 10:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: RIVERO

DATE/TIME OF TEST: 11/06/2020 05:02:39

---

Diagnostics Check	OK	04:56
Air Blank	0.000	04:56
Control Test	0.081	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.136	04:58
Breath Volume	2.957	04:58
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.133	05:01
Breath Volume	2.925	05:01
Air Blank	0.000	05:01
Control Test	0.080	05:02
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

---

Data Download Date/Time: 11/12/2020 10:08



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 11/06/2020 22:10:55

---

Diagnostics Check	OK	22:02
Air Blank	0.000	22:03
Control Test	0.081	22:03
Air Blank	0.000	22:04
Subject Sample # 1	0.000	22:06
Breath Volume	3.035	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.000	22:09
Breath Volume	3.183	22:09
Air Blank	0.000	22:10
Control Test	0.080	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

---

Data Download Date/Time: 11/12/2020 10:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER, MATTHEW, D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: RARICK, FREDERICK

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 10/05/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22419080A3

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 04:31

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: SHIFFLETT DUGGAN

DATE/TIME OF TEST: 11/07/2020 04:55:46

---

Diagnostics Check	OK	04:52	
Air Blank	0.000	04:52	
Control Test	0.080	04:52	
Air Blank	0.000	04:53	
Subject Sample	# 1	04:54	Subject Test Refused
Breath Volume	0.000	04:54	
Air Blank	0.000	04:54	
Control Test	0.081	04:55	
Air Blank	0.000	04:55	
Diagnostics Check	OK	04:55	

---

Data Download Date/Time: 11/12/2020 10:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARCOUX,HOPE,M  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: MARCOUX,HOPE  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 07:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: BUEXJUAJI

DATE/TIME OF TEST: 11/07/2020 07:56:32

---

Diagnostics Check	OK	07:50
Air Blank	0.000	07:50
Control Test	0.082	07:51
Air Blank	0.000	07:51
Subject Sample # 1	0.126	07:52
Breath Volume	1.156	07:52
Air Blank	0.000	07:52
Air Blank	0.000	07:54
Subject Sample # 2	0.135	07:55
Breath Volume	2.371	07:55
Air Blank	0.000	07:55
Control Test	0.081	07:55
Air Blank	0.000	07:56
Diagnostics Check	OK	07:56

---

Data Download Date/Time: 11/12/2020 10:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: BAUTISTE

DATE/TIME OF TEST: 11/08/2020 01:08:52

---

Diagnostics Check		OK	01:02
Air Blank		0.000	01:02
Control Test		0.082	01:03
Air Blank		0.000	01:03
Subject Sample	# 1	0.205	01:04
Breath Volume		1.972	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:06
Subject Sample	# 2	0.193	01:07
Breath Volume		2.148	01:07
Air Blank		0.000	01:07
Control Test		0.081	01:08
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

---

Data Download Date/Time: 11/12/2020 10:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALAMEDA,JORGE,L  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: ROBERTS,D  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:08

SUBJECT'S LAST NAME: CHAMPLIN

DATE/TIME OF TEST: 11/08/2020 17:35:46

---

Diagnostics Check	OK	17:29
Air Blank	0.000	17:29
Control Test	0.081	17:30
Air Blank	0.000	17:30
Subject Sample # 1	0.274	17:31
Breath Volume	2.867	17:31
Air Blank	0.000	17:31
Air Blank	0.000	17:33
Subject Sample # 2	0.292	17:34
Breath Volume	2.640	17:34
Air Blank	0.000	17:34
Control Test	0.080	17:35
Air Blank	0.000	17:35
Diagnostics Check	OK	17:35

---

Data Download Date/Time: 11/12/2020 10:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALAMEDA,JORGE,L  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: MORALES,RICARDO  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 11/10/2020 18:20:24

---

Diagnostics Check	OK	18:14
Air Blank	0.000	18:14
Control Test	0.081	18:14
Air Blank	0.000	18:15
Subject Sample # 1	0.161	18:15
Breath Volume	1.882	18:15
Air Blank	0.000	18:16
Air Blank	0.000	18:18
Subject Sample # 2	0.155	18:18
Breath Volume	1.289	18:18
Air Blank	0.000	18:19
Control Test	0.080	18:19
Air Blank	0.000	18:20
Diagnostics Check	OK	18:20

---

Data Download Date/Time: 11/12/2020 10:08



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUICK,SHANE,C  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: QUICK,SHANE  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 11/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: QUICK

DATE/TIME OF TEST: 11/12/2020 20:37:59

---

Diagnostics Check	OK	20:29	
Air Blank	0.000	20:29	
Control Test	0.082	20:30	
Air Blank	0.000	20:30	
Subject Sample	# 1	20:33	No Sample Provided
Breath Volume	0.000	20:33	
Air Blank	0.000	20:34	
Air Blank	0.000	20:36	
Subject Sample	# 2	20:36	Subject Test Refused
Breath Volume	0.000	20:36	
Air Blank	0.000	20:37	
Control Test	0.081	20:37	
Air Blank	0.000	20:37	
Diagnostics Check	OK	20:37	

---

Data Download Date/Time: 12/07/2020 13:52



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUICK,SHANE,C  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: QUICK,SHANE  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 11/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: TRIPLETT

DATE/TIME OF TEST: 11/12/2020 20:44:41

---

Diagnostics Check	OK	20:41	
Air Blank	0.000	20:42	
Control Test	0.081	20:42	
Air Blank	0.000	20:43	
Subject Sample	# 1	20:43	Subject Test Refused
Breath Volume	0.000	20:43	
Air Blank	0.000	20:43	
Control Test	0.081	20:44	
Air Blank	0.000	20:44	
Diagnostics Check	OK	20:44	

---

Data Download Date/Time: 12/07/2020 13:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 11/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: HARLAN

DATE/TIME OF TEST: 11/13/2020 04:22:49

---

Diagnostics Check		OK	04:10	
Air Blank		0.000	04:10	
Control Test		0.080	04:11	
Air Blank		0.000	04:11	
Subject Sample	# 1	0.169	04:14	Volume Not Met
Breath Volume		0.839	04:14	
Air Blank		0.000	04:15	
Air Blank		0.000	04:17	
Subject Sample	# 2	0.190	04:17	
Breath Volume		2.628	04:17	
Air Blank		0.000	04:18	
Air Blank		0.000	04:20	
Subject Sample	# 3	0.187	04:21	
Breath Volume		2.500	04:21	
Air Blank		0.000	04:21	
Control Test		0.080	04:22	
Air Blank		0.000	04:22	
Diagnostics Check		OK	04:22	

---

Data Download Date/Time: 12/07/2020 13:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUICK,SHANE,C  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: QUICK,SHANE  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 11/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: KIXMILLER

DATE/TIME OF TEST: 11/13/2020 04:58:48

---

Diagnostics Check	OK	04:52
Air Blank	0.000	04:52
Control Test	0.081	04:53
Air Blank	0.000	04:53
Subject Sample # 1	0.256	04:54
Breath Volume	3.386	04:54
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample # 2	0.251	04:57
Breath Volume	3.332	04:57
Air Blank	0.000	04:57
Control Test	0.080	04:58
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

---

Data Download Date/Time: 12/07/2020 13:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 11/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: RAMABADRAN

DATE/TIME OF TEST: 11/14/2020 04:00:10

---

Diagnostics Check	OK	03:48		
Air Blank	0.000	03:49		
Control Test	0.081	03:49		
Air Blank	0.000	03:49		
Subject Sample	# 1	03:53	No Sample Provided	
Breath Volume	0.000	03:53		
Air Blank	0.000	03:53		
Air Blank	0.000	03:55		
Subject Sample	# 2	0.134	03:58	Volume Not Met
Breath Volume	0.574	03:58		
Air Blank	0.000	03:59		
Control Test	0.081	03:59		
Air Blank	0.000	04:00		
Diagnostics Check	OK	04:00		

---

Data Download Date/Time: 12/07/2020 13:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 11/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: RAMABADRAN

DATE/TIME OF TEST: 11/14/2020 04:08:10

---

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.081	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.147	04:03
Breath Volume	2.535	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:05
Subject Sample # 2	0.145	04:06
Breath Volume	1.714	04:06
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

---

Data Download Date/Time: 12/07/2020 13:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KATZ,J,T  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: KATZ,J  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 11/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: ROSADO

DATE/TIME OF TEST: 11/15/2020 19:51:13

---

Diagnostics Check	OK	19:47	
Air Blank	0.000	19:48	
Control Test	0.081	19:48	
Air Blank	0.000	19:48	
Subject Sample	# 1	19:49	Subject Test Refused
Breath Volume	0.000	19:49	
Air Blank	0.000	19:50	
Control Test	0.081	19:50	
Air Blank	0.000	19:51	
Diagnostics Check	OK	19:51	

---

Data Download Date/Time: 12/07/2020 13:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 11/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: BRIDGES

DATE/TIME OF TEST: 11/19/2020 19:55:53

---

Diagnostics Check	OK	19:49
Air Blank	0.000	19:49
Control Test	0.081	19:50
Air Blank	0.000	19:50
Subject Sample # 1	0.000	19:51
Breath Volume	2.476	19:51
Air Blank	0.000	19:51
Air Blank	0.000	19:53
Subject Sample # 2	0.000	19:54
Breath Volume	1.734	19:54
Air Blank	0.000	19:54
Control Test	0.080	19:55
Air Blank	0.000	19:55
Diagnostics Check	OK	19:55

---

Data Download Date/Time: 12/07/2020 13:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 11/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/20/2020 02:31:59

---

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:17	
Control Test	0.081	02:18	
Air Blank	0.000	02:18	
Subject Sample # 1	0.193	02:19	
Breath Volume	2.390	02:19	
Air Blank	0.000	02:19	
Air Blank	0.000	02:21	
Subject Sample # 2		02:24	No Sample Provided
Breath Volume	0.000	02:24	
Air Blank	0.000	02:25	
Air Blank	0.000	02:27	
Subject Sample # 3		02:30	No Sample Provided
Breath Volume	0.000	02:30	
Air Blank	0.000	02:31	
Control Test	0.080	02:31	
Air Blank	0.000	02:31	
Diagnostics Check	OK	02:31	

---

Data Download Date/Time: 12/07/2020 13:52



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/12/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22419080A3

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 01:26

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/20/2020 02:54:10

---

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.081	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.193	02:49
Breath Volume	2.804	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.188	02:52
Breath Volume	2.597	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

---

Data Download Date/Time: 12/07/2020 13:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSIER, MATTHEW, D  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: RARICK, FREDERICK  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 11/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 11/21/2020 04:12:07

---

Diagnostics Check		OK	04:00	
Air Blank		0.000	04:00	
Control Test		0.080	04:01	
Air Blank		0.000	04:01	
Subject Sample	# 1	0.159	04:04	Volume Not Met
Breath Volume		0.097	04:04	
Air Blank		0.000	04:05	
Air Blank		0.000	04:07	
Subject Sample	# 2	0.159	04:07	
Breath Volume		1.640	04:07	
Air Blank		0.000	04:08	
Air Blank		0.000	04:10	
Subject Sample	# 3	0.146	04:10	
Breath Volume		1.546	04:10	
Air Blank		0.000	04:11	
Control Test		0.080	04:11	
Air Blank		0.000	04:12	
Diagnostics Check		OK	04:12	

---

Data Download Date/Time: 12/07/2020 13:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRNKA,L,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: FISHER,TROOPER  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 11/21/2020 17:17:24

---

Diagnostics Check		OK	17:07	
Air Blank		0.000	17:07	
Control Test		0.081	17:07	
Air Blank		0.000	17:08	
Subject Sample	# 1	0.397	17:09	No .020 agreement
Breath Volume		2.578	17:09	
Air Blank		0.000	17:09	
Air Blank		0.000	17:11	
Subject Sample	# 2	0.337	17:12	No .020 agreement
Breath Volume		1.515	17:12	
Air Blank		0.000	17:13	
Air Blank		0.000	17:15	
Subject Sample	# 3	0.372	17:15	No .020 agreement
Breath Volume		1.171	17:15	
Air Blank		0.000	17:16	
Control Test		0.078	17:16	
Air Blank		0.000	17:17	
Diagnostics Check		OK	17:17	

---

Data Download Date/Time: 12/07/2020 13:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRNKA,L,T  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: OTTO,N  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 11/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: GERBER

DATE/TIME OF TEST: 11/21/2020 20:06:06

---

Diagnostics Check	OK	20:03	
Air Blank	0.000	20:03	
Control Test	0.081	20:03	
Air Blank	0.000	20:04	
Subject Sample	# 1	20:04	Subject Test Refused
Breath Volume	0.000	20:04	
Air Blank	0.000	20:05	
Control Test	0.081	20:05	
Air Blank	0.000	20:06	
Diagnostics Check	OK	20:06	

---

Data Download Date/Time: 12/07/2020 13:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALAMEDA,JORGE,L  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: OTTO,NAYA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 11/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: LIPP

DATE/TIME OF TEST: 11/23/2020 20:29:03

---

Diagnostics Check	OK	20:24	
Air Blank	0.000	20:25	
Control Test	0.081	20:25	
Air Blank	0.000	20:26	
Subject Sample	# 1	20:27	Subject Test Refused
Breath Volume	0.000	20:27	
Air Blank	0.000	20:28	
Control Test	0.081	20:28	
Air Blank	0.000	20:28	
Diagnostics Check	OK	20:29	

---

Data Download Date/Time: 12/07/2020 13:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: URBANIAK,JEFFREY,P  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: ROCHA,ALEX  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 11/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: LEJMAN

DATE/TIME OF TEST: 11/26/2020 04:58:33

---

Diagnostics Check	OK	04:51
Air Blank	0.000	04:52
Control Test	0.082	04:52
Air Blank	0.000	04:53
Subject Sample # 1	0.000	04:54
Breath Volume	3.277	04:54
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample # 2	0.000	04:57
Breath Volume	2.820	04:57
Air Blank	0.000	04:57
Control Test	0.080	04:57
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

---

Data Download Date/Time: 12/07/2020 13:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 11/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: CASTRO MELERO

DATE/TIME OF TEST: 11/28/2020 04:03:53

---

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.102	03:59
Breath Volume	2.566	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.100	04:02
Breath Volume	2.234	04:02
Air Blank	0.000	04:02
Control Test	0.081	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

---

Data Download Date/Time: 12/07/2020 13:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PROPST,JOHN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 11/01/2020 03:58:41

---

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:47	
Control Test	0.078	03:48	
Air Blank	0.000	03:48	
Subject Sample	# 1	03:51	No Sample Provided
Breath Volume	0.000	03:51	
Air Blank	0.000	03:52	
Air Blank	0.000	03:54	
Subject Sample	# 2	03:57	No Sample Provided
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Control Test	0.079	03:58	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

---

Data Download Date/Time: 11/23/2020 12:15



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELANEY,AMY,L  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PREYSING,VON  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 07:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:14

SUBJECT'S LAST NAME: SALVATIERRA MORA

DATE/TIME OF TEST: 11/01/2020 09:45:06

---

Diagnostics Check	OK	09:42	
Air Blank	0.000	09:42	
Control Test	0.079	09:43	
Air Blank	0.000	09:43	
Subject Sample	# 1	09:43	Subject Test Refused
Breath Volume	0.000	09:43	
Air Blank	0.000	09:44	
Control Test	0.078	09:44	
Air Blank	0.000	09:45	
Diagnostics Check	OK	09:45	

---

Data Download Date/Time: 11/23/2020 12:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: TAYLOR,GABRIELLE  
ARRESTING OFFICER AGENCY: Winter Haven PD  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 11/02/2020 00:03:10

---

Diagnostics Check	OK	23:59	
Air Blank	0.000	23:59	
Control Test	0.078	00:00	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:01	Subject Test Refused
Breath Volume	0.000	00:01	
Air Blank	0.000	00:02	
Control Test	0.078	00:02	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

---

Data Download Date/Time: 11/23/2020 12:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: JERONIMO BAIL

DATE/TIME OF TEST: 11/02/2020 21:36:49

---

Diagnostics Check		OK	21:30
Air Blank		0.000	21:31
Control Test		0.078	21:31
Air Blank		0.000	21:31
Subject Sample	# 1	0.218	21:32
Breath Volume		2.242	21:32
Air Blank		0.000	21:33
Air Blank		0.000	21:34
Subject Sample	# 2	0.225	21:35
Breath Volume		2.601	21:35
Air Blank		0.000	21:35
Control Test		0.077	21:36
Air Blank		0.000	21:36
Diagnostics Check		OK	21:36

---

Data Download Date/Time: 11/23/2020 12:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PICHE,RUSSELL,G  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PICHE,RUSSELL  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: CAMERON

DATE/TIME OF TEST: 11/03/2020 23:00:06

---

Diagnostics Check	OK	22:57	
Air Blank	0.000	22:57	
Control Test	0.077	22:57	
Air Blank	0.000	22:58	
Subject Sample	# 1	22:58	Subject Test Refused
Breath Volume	0.000	22:58	
Air Blank	0.000	22:59	
Control Test	0.077	22:59	
Air Blank	0.000	23:00	
Diagnostics Check	OK	23:00	

---

Data Download Date/Time: 11/23/2020 12:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: OCASIO RIVERA

DATE/TIME OF TEST: 11/03/2020 23:44:39

---

Diagnostics Check	OK	23:38
Air Blank	0.000	23:39
Control Test	0.078	23:39
Air Blank	0.000	23:39
Subject Sample # 1	0.000	23:40
Breath Volume	2.550	23:40
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.000	23:43
Breath Volume	2.265	23:43
Air Blank	0.000	23:43
Control Test	0.079	23:44
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

---

Data Download Date/Time: 11/23/2020 12:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: SAFIEDDINE,JAMAL  
ARRESTING OFFICER AGENCY: Davenport PD  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: AYALA-CABRERA

DATE/TIME OF TEST: 11/06/2020 02:13:01

---

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.079	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.072	02:08
Breath Volume	1.777	02:08
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.074	02:11
Breath Volume	2.250	02:11
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:13

---

Data Download Date/Time: 11/23/2020 12:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: PERAZA

DATE/TIME OF TEST: 11/06/2020 19:33:36

---

Diagnostics Check	OK	19:25
Air Blank	0.000	19:26
Control Test	0.078	19:26
Air Blank	0.000	19:27
Subject Sample # 1	0.273	19:27
Breath Volume	1.382	19:27
Air Blank	0.000	19:28
Air Blank	0.000	19:30
Subject Sample # 2	0.260	19:31
Breath Volume	1.285	19:31
Air Blank	0.000	19:32
Control Test	0.077	19:32
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

---

Data Download Date/Time: 11/23/2020 12:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SYKES,STEVEN,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: SYKES,STEVEN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 11/06/2020 21:43:10

---

Diagnostics Check	OK	21:36
Air Blank	0.000	21:36
Control Test	0.078	21:37
Air Blank	0.000	21:37
Subject Sample # 1	0.030	21:38
Breath Volume	1.371	21:38
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.029	21:41
Breath Volume	1.574	21:41
Air Blank	0.000	21:42
Control Test	0.079	21:42
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

---

Data Download Date/Time: 11/23/2020 12:16



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: VELA

DATE/TIME OF TEST: 11/06/2020 23:13:36

---

Diagnostics Check	OK	23:06	
Air Blank	0.000	23:07	
Control Test	0.078	23:07	
Air Blank	0.000	23:08	
Subject Sample	# 1 0.179	23:09	
Breath Volume	1.804	23:09	
Air Blank	0.000	23:09	
Air Blank	0.000	23:11	
Subject Sample	# 2 0.166	23:12	
Breath Volume	1.699	23:12	
Air Blank		23:13	RFI Detect
Air Blank		23:13	RFI Detect

---

Data Download Date/Time: 11/23/2020 12:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: VELA

DATE/TIME OF TEST: 11/06/2020 23:22:54

---

Diagnostics Check	OK	23:16	
Air Blank	0.000	23:17	
Control Test	0.076	23:17	
Air Blank	0.000	23:17	
Subject Sample	# 1 0.180	23:18	
Breath Volume	1.703	23:18	
Air Blank	0.000	23:18	
Air Blank	0.000	23:20	
Subject Sample	# 2 0.172	23:21	
Breath Volume	1.648	23:21	
Air Blank	0.000	23:21	
Control Test	0.074	23:22	Control Outside Tolerance
Air Blank	0.000	23:22	

---

Data Download Date/Time: 11/23/2020 12:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: VELA

DATE/TIME OF TEST: 11/06/2020 23:34:24

---

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.078	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.162	23:28
Breath Volume	1.820	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.160	23:32
Breath Volume	1.832	23:32
Air Blank	0.000	23:33
Control Test	0.077	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

---

Data Download Date/Time: 11/23/2020 12:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SYKES,STEVEN,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WALSH,MICHAEL  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: GOODMAN

DATE/TIME OF TEST: 11/07/2020 02:05:21

---

Diagnostics Check		OK	01:56
Air Blank		0.000	01:57
Control Test		0.078	01:57
Air Blank		0.000	01:58
Subject Sample	# 1	0.120	01:59
Breath Volume		1.710	01:59
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.104	02:03
Breath Volume		1.960	02:03
Air Blank		0.000	02:04
Control Test		0.078	02:04
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

---

Data Download Date/Time: 11/23/2020 12:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,STEVEN,J  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: EVANS,STEVEN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: COLON CANALES

DATE/TIME OF TEST: 11/07/2020 04:26:08

---

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:20	
Control Test	0.078	04:20	
Air Blank	0.000	04:20	
Subject Sample # 1	0.144	04:21	
Breath Volume	2.160	04:21	
Air Blank	0.000	04:22	
Air Blank	0.000	04:23	
Subject Sample # 2	0.136	04:24	
Breath Volume	2.160	04:24	
Air Blank	0.000	04:25	
Control Test	0.073	04:25	Control Outside Tolerance
Air Blank	0.000	04:26	

---

Data Download Date/Time: 11/23/2020 12:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:39

SUBJECT'S LAST NAME: CEJOUR

DATE/TIME OF TEST: 11/07/2020 07:24:42

---

Diagnostics Check		OK	07:13	
Air Blank		0.000	07:13	
Control Test		0.078	07:14	
Air Blank		0.000	07:14	
Subject Sample	# 1	0.105	07:17	Volume Not Met
Breath Volume		0.054	07:17	
Air Blank		0.000	07:18	
Air Blank		0.000	07:20	
Subject Sample	# 2		07:23	No Sample Provided
Breath Volume		0.000	07:23	
Air Blank		0.000	07:23	
Control Test		0.077	07:24	
Air Blank		0.000	07:24	
Diagnostics Check		OK	07:24	

---

Data Download Date/Time: 11/23/2020 12:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SYKES,STEVEN,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: SYKES,STEVEN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:49

SUBJECT'S LAST NAME: HERNANDEZ HERNANDEZ

DATE/TIME OF TEST: 11/07/2020 19:16:51

---

Diagnostics Check	OK	19:10
Air Blank	0.000	19:11
Control Test	0.078	19:11
Air Blank	0.000	19:12
Subject Sample # 1	0.093	19:12
Breath Volume	2.535	19:12
Air Blank	0.000	19:13
Air Blank	0.000	19:14
Subject Sample # 2	0.094	19:15
Breath Volume	2.171	19:15
Air Blank	0.000	19:15
Control Test	0.077	19:16
Air Blank	0.000	19:16
Diagnostics Check	OK	19:16

---

Data Download Date/Time: 11/23/2020 12:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,STEVEN,J  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: EVANS,STEVEN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 11/07/2020 20:38:26

---

Diagnostics Check	OK	20:37	
Air Blank		20:37	RFI Detect
Air Blank	0.000	20:38	

---

Data Download Date/Time: 11/23/2020 12:16



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,STEVEN,J  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: EVANS,STEVEN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 11/07/2020 20:46:50

---

Diagnostics Check	OK	20:40
Air Blank	0.000	20:41
Control Test	0.078	20:41
Air Blank	0.000	20:41
Subject Sample # 1	0.051	20:42
Breath Volume	2.468	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:44
Subject Sample # 2	0.051	20:45
Breath Volume	2.757	20:45
Air Blank	0.000	20:45
Control Test	0.077	20:46
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

---

Data Download Date/Time: 11/23/2020 12:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHURCH,JOHN,M  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: CHURCH,JOHN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: HILL III

DATE/TIME OF TEST: 11/07/2020 22:11:42

---

Diagnostics Check	OK	22:08	
Air Blank	0.000	22:09	
Control Test	0.078	22:09	
Air Blank	0.000	22:10	
Subject Sample	# 1	22:10	Subject Test Refused
Breath Volume	0.000	22:10	
Air Blank	0.000	22:10	
Control Test	0.080	22:11	
Air Blank	0.000	22:11	
Diagnostics Check	OK	22:11	

---

Data Download Date/Time: 11/23/2020 12:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,STEVEN,J  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: SANTOS,CARLOS  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: VENEGAS GARCIA

DATE/TIME OF TEST: 11/08/2020 00:25:08

---

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.078	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.159	00:20
Breath Volume	1.648	00:20
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.156	00:23
Breath Volume	2.289	00:23
Air Blank	0.000	00:24
Control Test	0.077	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

---

Data Download Date/Time: 11/23/2020 12:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WILSON,KALEB  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: BARRON

DATE/TIME OF TEST: 11/08/2020 00:40:10

---

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.079	00:33
Air Blank	0.000	00:34
Subject Sample # 1	0.177	00:35
Breath Volume	1.296	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:37
Subject Sample # 2	0.170	00:38
Breath Volume	1.386	00:38
Air Blank	0.000	00:39
Control Test	0.078	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

---

Data Download Date/Time: 11/23/2020 12:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 11/08/2020 00:58:23

---

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:47	
Control Test	0.078	00:47	
Air Blank	0.000	00:48	
Subject Sample	# 1	00:51	No Sample Provided
Breath Volume	0.000	00:51	
Air Blank	0.000	00:51	
Air Blank	0.000	00:53	
Subject Sample	# 2	00:57	No Sample Provided
Breath Volume	0.000	00:57	
Air Blank	0.000	00:57	
Control Test	0.078	00:57	
Air Blank	0.000	00:58	
Diagnostics Check	OK	00:58	

---

Data Download Date/Time: 11/23/2020 12:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SYKES,STEVEN,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: SYKES,STEVEN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: LANDAVERDE ESPINOZA

DATE/TIME OF TEST: 11/08/2020 03:21:08

---

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.078	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.128	03:15
Breath Volume	1.382	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.130	03:19
Breath Volume	1.285	03:19
Air Blank	0.000	03:20
Control Test	0.078	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

---

Data Download Date/Time: 11/23/2020 12:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALSTON,SONDRA,L  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: HERMELBRACHT,DONALD  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 10:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:42

SUBJECT'S LAST NAME: HERNANDEZ-MUNOZ

DATE/TIME OF TEST: 11/08/2020 11:52:01

---

Diagnostics Check	OK	11:45
Air Blank	0.000	11:45
Control Test	0.078	11:46
Air Blank	0.000	11:46
Subject Sample # 1	0.102	11:47
Breath Volume	1.589	11:47
Air Blank	0.000	11:48
Air Blank	0.000	11:49
Subject Sample # 2	0.097	11:50
Breath Volume	1.828	11:50
Air Blank	0.000	11:51
Control Test	0.078	11:51
Air Blank	0.000	11:51
Diagnostics Check	OK	11:52

---

Data Download Date/Time: 11/23/2020 12:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COBB,JANIKQUA,L  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: VONPREYSING,RICHARD  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:47

SUBJECT'S LAST NAME: COBB

DATE/TIME OF TEST: 11/08/2020 15:00:47

---

Diagnostics Check		OK	14:50
Air Blank		0.000	14:51
Control Test		0.078	14:51
Air Blank		0.000	14:52
Subject Sample	# 1	0.275	14:52
Breath Volume		2.425	14:52
Air Blank		0.000	14:53
Air Blank		0.000	14:55
Subject Sample	# 2	0.254	14:56
Breath Volume		1.332	14:56
Air Blank		0.000	14:56
Air Blank		0.000	14:58
Subject Sample	# 3	0.255	14:59
Breath Volume		1.277	14:59
Air Blank		0.000	14:59
Control Test		0.078	15:00
Air Blank		0.000	15:00
Diagnostics Check		OK	15:00

---

Data Download Date/Time: 11/23/2020 12:16



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FLETCHER, MELISSA, A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: CRUZ, EDGARDO  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 11/08/2020 20:14:15

---

Diagnostics Check	OK	20:07
Air Blank	0.000	20:07
Control Test	0.078	20:07
Air Blank	0.000	20:08
Subject Sample # 1	0.090	20:09
Breath Volume	1.250	20:09
Air Blank	0.000	20:10
Air Blank	0.000	20:12
Subject Sample # 2	0.094	20:12
Breath Volume	1.218	20:12
Air Blank	0.000	20:13
Control Test	0.077	20:13
Air Blank	0.000	20:14
Diagnostics Check	OK	20:14

---

Data Download Date/Time: 11/23/2020 12:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PICHE,RUSSELL,G  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PICHE,RUSSELL  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: PAIGE

DATE/TIME OF TEST: 11/08/2020 23:08:19

---

Diagnostics Check		OK	23:01
Air Blank		0.000	23:02
Control Test		0.078	23:02
Air Blank		0.000	23:02
Subject Sample	# 1	0.226	23:03
Breath Volume		1.976	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.231	23:06
Breath Volume		1.917	23:06
Air Blank		0.000	23:07
Control Test		0.077	23:07
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

---

Data Download Date/Time: 11/23/2020 12:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: LEDFORD,ANN  
ARRESTING OFFICER AGENCY: Bartow PD  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: LOHART

DATE/TIME OF TEST: 11/10/2020 00:09:41

---

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:07	
Control Test	0.079	00:07	
Air Blank	0.000	00:08	
Subject Sample	# 1	00:08	Subject Test Refused
Breath Volume	0.000	00:08	
Air Blank	0.000	00:08	
Control Test	0.079	00:09	
Air Blank	0.000	00:09	
Diagnostics Check	OK	00:09	

---

Data Download Date/Time: 11/23/2020 12:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PROPST,JOHN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 11/11/2020 01:59:49

---

Diagnostics Check		OK	01:53
Air Blank		0.000	01:53
Control Test		0.078	01:54
Air Blank		0.000	01:54
Subject Sample	# 1	0.128	01:55
Breath Volume		2.351	01:55
Air Blank		0.000	01:55
Air Blank		0.000	01:57
Subject Sample	# 2	0.129	01:58
Breath Volume		2.605	01:58
Air Blank		0.000	01:58
Control Test		0.078	01:59
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

---

Data Download Date/Time: 11/23/2020 12:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRADLEY,KEVIN,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BRADLEY,KEVIN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 12:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: ORTA

DATE/TIME OF TEST: 11/12/2020 14:41:27

---

Diagnostics Check	OK	14:35
Air Blank	0.000	14:35
Control Test	0.078	14:36
Air Blank	0.000	14:36
Subject Sample # 1	0.117	14:37
Breath Volume	1.878	14:37
Air Blank	0.000	14:37
Air Blank	0.000	14:39
Subject Sample # 2	0.111	14:39
Breath Volume	1.574	14:39
Air Blank	0.000	14:40
Control Test	0.078	14:40
Air Blank	0.000	14:41
Diagnostics Check	OK	14:41

---

Data Download Date/Time: 11/23/2020 12:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PICHE,RUSSELL,G  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PICHE,RUSSELL  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: PALMA

DATE/TIME OF TEST: 11/13/2020 02:07:18

---

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:04	
Control Test	0.079	02:05	
Air Blank	0.000	02:05	
Subject Sample	# 1	02:05	Subject Test Refused
Breath Volume	0.000	02:05	
Air Blank	0.000	02:06	
Control Test	0.078	02:06	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

---

Data Download Date/Time: 11/23/2020 12:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OGG,DAVID,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: OGG,DAVID  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:54

SUBJECT'S LAST NAME: LASH

DATE/TIME OF TEST: 11/13/2020 14:22:52

---

Diagnostics Check	OK	14:15
Air Blank	0.000	14:16
Control Test	0.078	14:16
Air Blank	0.000	14:16
Subject Sample # 1	0.000	14:18
Breath Volume	2.136	14:18
Air Blank	0.000	14:18
Air Blank	0.000	14:20
Subject Sample # 2	0.000	14:21
Breath Volume	2.093	14:21
Air Blank	0.000	14:21
Control Test	0.079	14:22
Air Blank	0.000	14:22
Diagnostics Check	OK	14:22

---

Data Download Date/Time: 11/23/2020 12:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PROPST,JOHN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: BAUER

DATE/TIME OF TEST: 11/13/2020 21:50:04

---

Diagnostics Check	OK	21:43
Air Blank	0.000	21:44
Control Test	0.078	21:44
Air Blank	0.000	21:44
Subject Sample # 1	0.000	21:45
Breath Volume	2.285	21:45
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample # 2	0.000	21:48
Breath Volume	2.296	21:48
Air Blank	0.000	21:49
Control Test	0.079	21:49
Air Blank	0.000	21:49
Diagnostics Check	OK	21:50

---

Data Download Date/Time: 11/23/2020 12:17



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHILLIPS,JORDAN,W  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: MASLAK,ROBERT  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 12:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:35

SUBJECT'S LAST NAME: NIX

DATE/TIME OF TEST: 11/15/2020 14:21:10

---

Diagnostics Check	OK	14:14
Air Blank	0.000	14:15
Control Test	0.079	14:15
Air Blank	0.000	14:16
Subject Sample # 1	0.177	14:16
Breath Volume	1.621	14:16
Air Blank	0.000	14:17
Air Blank	0.000	14:19
Subject Sample # 2	0.175	14:19
Breath Volume	1.269	14:19
Air Blank	0.000	14:20
Control Test	0.078	14:20
Air Blank	0.000	14:21
Diagnostics Check	OK	14:21

---

Data Download Date/Time: 11/23/2020 12:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: SYKES,STEPHEN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: HODGE

DATE/TIME OF TEST: 11/17/2020 21:22:33

---

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:20	
Control Test	0.078	21:21	
Air Blank	0.000	21:21	
Subject Sample	# 1	21:21	RFI Detect
Breath Volume	0.101	21:21	
Air Blank		21:22	RFI Detect
Air Blank	0.000	21:22	

---

Data Download Date/Time: 11/23/2020 12:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: SYKES,STEPHEN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: HODGE

DATE/TIME OF TEST: 11/17/2020 21:40:21

---

Diagnostics Check	OK	21:28	
Air Blank	0.000	21:29	
Control Test	0.078	21:29	
Air Blank	0.000	21:30	
Subject Sample	# 1	21:33	No Sample Provided
Breath Volume	0.000	21:33	
Air Blank	0.000	21:33	
Air Blank	0.000	21:35	
Subject Sample	# 2	21:38	No Sample Provided
Breath Volume	0.000	21:38	
Air Blank	0.000	21:39	
Control Test	0.079	21:39	
Air Blank	0.000	21:40	
Diagnostics Check	OK	21:40	

---

Data Download Date/Time: 11/23/2020 12:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SYKES,STEVEN,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: SYKES,STEVEN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: HARDCASTLE

DATE/TIME OF TEST: 11/17/2020 22:00:45

---

Diagnostics Check	OK	21:49	
Air Blank	0.000	21:49	
Control Test	0.079	21:50	
Air Blank	0.000	21:50	
Subject Sample	# 1	21:53	No Sample Provided
Breath Volume	0.000	21:53	
Air Blank	0.000	21:54	
Air Blank	0.000	21:56	
Subject Sample	# 2	21:59	No Sample Provided
Breath Volume	0.000	21:59	
Air Blank	0.000	21:59	
Control Test	0.078	22:00	
Air Blank	0.000	22:00	
Diagnostics Check	OK	22:00	

---

Data Download Date/Time: 11/23/2020 12:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: LESTER

DATE/TIME OF TEST: 11/20/2020 20:25:32

---

Diagnostics Check	OK	20:18
Air Blank	0.000	20:19
Control Test	0.078	20:19
Air Blank	0.000	20:20
Subject Sample # 1	0.097	20:21
Breath Volume	1.367	20:21
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample # 2	0.104	20:23
Breath Volume	1.656	20:23
Air Blank	0.000	20:24
Control Test	0.079	20:24
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

---

Data Download Date/Time: 11/23/2020 12:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SYKES,STEVEN,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: SYKES,STEVEN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: FUTCH

DATE/TIME OF TEST: 11/20/2020 20:34:55

---

Diagnostics Check	OK	20:28
Air Blank	0.000	20:28
Control Test	0.078	20:29
Air Blank	0.000	20:29
Subject Sample # 1	0.153	20:30
Breath Volume	2.695	20:30
Air Blank	0.000	20:30
Air Blank	0.000	20:32
Subject Sample # 2	0.149	20:33
Breath Volume	2.378	20:33
Air Blank	0.000	20:33
Control Test	0.077	20:34
Air Blank	0.000	20:34
Diagnostics Check	OK	20:34

---

Data Download Date/Time: 11/23/2020 12:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SYKES,STEVEN,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: SYKES,STEVEN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: SANTIBANEZ

DATE/TIME OF TEST: 11/20/2020 22:01:38

---

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.078	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.064	21:57
Breath Volume	1.488	21:57
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.062	22:00
Breath Volume	1.449	22:00
Air Blank	0.000	22:00
Control Test	0.078	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

---

Data Download Date/Time: 11/23/2020 12:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONCION,CHRISTIAN,L  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: BEHRENSHOUSER,RICHARD  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BORTON

DATE/TIME OF TEST: 11/22/2020 00:34:14

---

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.079	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.120	00:29
Breath Volume	1.621	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:31
Subject Sample # 2	0.111	00:32
Breath Volume	1.226	00:32
Air Blank	0.000	00:33
Control Test	0.076	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

---

Data Download Date/Time: 11/23/2020 12:17



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SYKES, STEVEN,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: SYKES, STEVEN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/22/2020 02:36:32

---

Diagnostics Check		OK	02:29
Air Blank		0.000	02:29
Control Test		0.078	02:30
Air Blank		0.000	02:30
Subject Sample	# 1	0.215	02:31
Breath Volume		1.808	02:31
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.199	02:34
Breath Volume		1.207	02:34
Air Blank		0.000	02:35
Control Test		0.079	02:35
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

---

Data Download Date/Time: 11/23/2020 12:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FLETCHER,MELISSA,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: SANCHEZ,JUAN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 11/22/2020 05:37:05

---

Diagnostics Check	OK	05:28	
Air Blank	0.000	05:28	
Control Test	0.079	05:28	
Air Blank	0.000	05:29	
Subject Sample	# 1	05:32	No Sample Provided
Breath Volume	0.000	05:32	
Air Blank	0.000	05:33	
Air Blank	0.000	05:35	
Subject Sample	# 2	05:35	Subject Test Refused
Breath Volume	0.000	05:35	
Air Blank	0.000	05:36	
Control Test	0.078	05:36	
Air Blank	0.000	05:37	
Diagnostics Check	OK	05:37	

---

Data Download Date/Time: 11/23/2020 12:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHURCH,JOHN,M  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: SELLERS,JAMES  
ARRESTING OFFICER AGENCY: Bartow PD  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: RAYL

DATE/TIME OF TEST: 11/22/2020 22:43:27

---

Diagnostics Check	OK	22:38	
Air Blank	0.000	22:38	
Control Test	0.079	22:39	
Air Blank	0.000	22:39	
Subject Sample	# 1	22:41	Subject Test Refused
Breath Volume	0.136	22:41	
Air Blank	0.000	22:42	
Control Test	0.080	22:42	
Air Blank	0.000	22:43	
Diagnostics Check	OK	22:43	

---

Data Download Date/Time: 11/23/2020 12:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: LEDFORD,ANN  
ARRESTING OFFICER AGENCY: Bartow PD  
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 11/24/2020 19:23:12

---

Diagnostics Check	OK	19:20	
Air Blank	0.000	19:20	
Control Test	0.078	19:21	
Air Blank	0.000	19:21	
Subject Sample	# 1	19:21	Subject Test Refused
Breath Volume	0.000	19:21	
Air Blank	0.000	19:22	
Control Test	0.079	19:22	
Air Blank	0.000	19:23	
Diagnostics Check	OK	19:23	

---

Data Download Date/Time: 12/18/2020 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SYKES,STEVEN,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: SYKES,STEVEN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 11/25/2020 19:37:44

---

Diagnostics Check	OK	19:31
Air Blank	0.000	19:31
Control Test	0.078	19:32
Air Blank	0.000	19:32
Subject Sample # 1	0.111	19:33
Breath Volume	1.523	19:33
Air Blank	0.000	19:33
Air Blank	0.000	19:35
Subject Sample # 2	0.110	19:36
Breath Volume	1.425	19:36
Air Blank	0.000	19:36
Control Test	0.077	19:37
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

---

Data Download Date/Time: 12/18/2020 13:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PICHE,RUSSELL,G  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PICHE,RUSSELL  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: KAYWORTH

DATE/TIME OF TEST: 11/26/2020 02:09:35

---

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:07	
Control Test	0.078	02:07	
Air Blank	0.000	02:07	
Subject Sample	# 1	02:08	Subject Test Refused
Breath Volume	0.000	02:08	
Air Blank	0.000	02:08	
Control Test	0.079	02:08	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

---

Data Download Date/Time: 12/18/2020 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARREN,BILLY,J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WARREN,BILLY  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: SHELGREN

DATE/TIME OF TEST: 11/26/2020 04:47:00

---

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.078	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.212	04:41
Breath Volume	1.238	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.214	04:45
Breath Volume	3.160	04:45
Air Blank	0.000	04:46
Control Test	0.076	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:47

---

Data Download Date/Time: 12/18/2020 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SYKES,STEVEN,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: SYKES,STEVEN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: PASCALL

DATE/TIME OF TEST: 11/26/2020 20:11:29

---

Diagnostics Check	OK	20:04
Air Blank	0.000	20:04
Control Test	0.078	20:05
Air Blank	0.000	20:05
Subject Sample # 1	0.208	20:06
Breath Volume	2.015	20:06
Air Blank	0.000	20:07
Air Blank	0.000	20:09
Subject Sample # 2	0.204	20:09
Breath Volume	1.660	20:09
Air Blank	0.000	20:10
Control Test	0.077	20:10
Air Blank	0.000	20:11
Diagnostics Check	OK	20:11

---

Data Download Date/Time: 12/18/2020 13:45



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PICHE,RUSSELL,G  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PICHE,RUSSELL  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 11/26/2020 23:26:56

---

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.079	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.151	23:22
Breath Volume	1.394	23:22
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.149	23:25
Breath Volume	1.539	23:25
Air Blank	0.000	23:26
Control Test	0.078	23:26
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

---

Data Download Date/Time: 12/18/2020 13:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,ALBERT,J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WILLIAMS,ALBERT  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 11/28/2020 01:04:23

---

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:50	
Control Test	0.078	00:51	
Air Blank	0.000	00:51	
Subject Sample	# 1	00:54	No Sample Provided
Breath Volume	0.000	00:54	
Air Blank	0.000	00:55	
Air Blank	0.000	00:57	
Subject Sample	# 2	0.160	00:59
Breath Volume	2.804	00:59	
Air Blank	0.000	01:00	
Air Blank	0.000	01:02	
Subject Sample	# 3	0.161	01:02
Breath Volume	1.613	01:02	
Air Blank	0.000	01:03	
Control Test	0.079	01:03	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

---

Data Download Date/Time: 12/18/2020 13:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHILLIPS,JORDAN,W  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: DOUGLAS,WILLIAM  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 11/28/2020 07:23:56

---

Diagnostics Check	OK	07:21	
Air Blank	0.000	07:21	
Control Test	0.078	07:21	
Air Blank	0.000	07:22	
Subject Sample	# 1	07:22	Subject Test Refused
Breath Volume	0.000	07:22	
Air Blank	0.000	07:23	
Control Test	0.080	07:23	
Air Blank	0.000	07:23	
Diagnostics Check	OK	07:23	

---

Data Download Date/Time: 12/18/2020 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALLACE, CHARLES, A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WALLACE, CHARLES  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: DINNALD JR.

DATE/TIME OF TEST: 11/28/2020 21:18:38

---

Diagnostics Check	OK	21:12
Air Blank	0.000	21:12
Control Test	0.078	21:13
Air Blank	0.000	21:13
Subject Sample # 1	0.187	21:14
Breath Volume	2.613	21:14
Air Blank	0.000	21:15
Air Blank	0.000	21:16
Subject Sample # 2	0.187	21:17
Breath Volume	1.964	21:17
Air Blank	0.000	21:17
Control Test	0.078	21:18
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

---

Data Download Date/Time: 12/18/2020 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,ALBERT,J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WILLIAMS,ALBERT  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: SANABRIA

DATE/TIME OF TEST: 11/29/2020 02:07:09

---

Diagnostics Check		OK	01:58
Air Blank		0.000	01:59
Control Test		0.079	01:59
Air Blank		0.000	01:59
Subject Sample	# 1	0.202	02:02
Breath Volume		2.238	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:04
Subject Sample	# 2	0.200	02:05
Breath Volume		3.222	02:05
Air Blank		0.000	02:06
Control Test		0.078	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

---

Data Download Date/Time: 12/18/2020 13:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALLACE, CHARLES, A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WALLACE, CHARLES  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: STRICKLEN

DATE/TIME OF TEST: 11/29/2020 03:22:08

---

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.078	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.201	03:16
Breath Volume	2.113	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.215	03:20
Breath Volume	1.289	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

---

Data Download Date/Time: 12/18/2020 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,ALBERT,J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: MAYNARD,GREG  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: DULCIO

DATE/TIME OF TEST: 11/29/2020 03:41:32

---

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:30	
Control Test	0.077	03:30	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:34	No Sample Provided
Breath Volume	0.000	03:34	
Air Blank	0.000	03:34	
Air Blank	0.000	03:36	
Subject Sample	# 2	03:39	No Sample Provided
Breath Volume	0.000	03:39	
Air Blank	0.000	03:40	
Control Test	0.077	03:40	
Air Blank	0.000	03:41	
Diagnostics Check	OK	03:41	

---

Data Download Date/Time: 12/18/2020 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: CARTER,JACOB  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: BUI  
ARREST TIME: 19:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 11/29/2020 22:26:56

---

Diagnostics Check	OK	22:24	
Air Blank	0.000	22:24	
Control Test	0.079	22:24	
Air Blank	0.000	22:25	
Subject Sample	# 1	22:25	Subject Test Refused
Breath Volume	0.000	22:25	
Air Blank	0.000	22:26	
Control Test	0.079	22:26	
Air Blank	0.000	22:26	
Diagnostics Check	OK	22:26	

---

Data Download Date/Time: 12/18/2020 13:45



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH-LONG,JEFFERSON,D  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: BROWN,JEFFREY  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: GIARDINA

DATE/TIME OF TEST: 11/06/2020 22:37:31

---

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:36	
Control Test	0.081	22:36	
Air Blank		22:37	RFI Detect
Air Blank	0.000	22:37	

---

Data Download Date/Time: 11/09/2020 12:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH-LONG,JEFFERSON,D  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: BROWN,JEFFREY  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: GIARDINA

DATE/TIME OF TEST: 11/06/2020 22:46:32

---

Diagnostics Check	OK	22:40
Air Blank	0.000	22:41
Control Test	0.082	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.000	22:42
Breath Volume	2.925	22:42
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.000	22:45
Breath Volume	2.562	22:45
Air Blank	0.000	22:45
Control Test	0.082	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

---

Data Download Date/Time: 11/09/2020 12:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: STANLEY, T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 11/02/2020 01:38:47

---

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.000	01:34
Breath Volume	2.804	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.000	01:37
Breath Volume	2.691	01:37
Air Blank	0.000	01:37
Control Test	0.080	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

---

Data Download Date/Time: 11/16/2020 11:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: STANLEY, T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: SHUMAN

DATE/TIME OF TEST: 11/02/2020 20:52:43

---

Diagnostics Check	OK	20:49	
Air Blank	0.000	20:50	
Control Test	0.080	20:50	
Air Blank	0.000	20:51	
Subject Sample	# 1	20:51	Subject Test Refused
Breath Volume	0.000	20:51	
Air Blank	0.000	20:51	
Control Test	0.080	20:52	
Air Blank	0.000	20:52	
Diagnostics Check	OK	20:52	

---

Data Download Date/Time: 11/16/2020 11:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: WEATHERS

DATE/TIME OF TEST: 11/03/2020 04:53:08

---

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.081	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.211	04:48
Breath Volume	1.230	04:48
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.209	04:51
Breath Volume	1.550	04:51
Air Blank	0.000	04:52
Control Test	0.081	04:52
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

---

Data Download Date/Time: 11/16/2020 11:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CAMACHO, K  
ARRESTING OFFICER AGENCY: Neptune Beach PD  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: EIZENHAFFER

DATE/TIME OF TEST: 11/04/2020 03:22:18

---

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:19	
Control Test	0.080	03:20	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:21	
Control Test	0.081	03:21	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

---

Data Download Date/Time: 11/16/2020 11:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JAMES,ALFONSO,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNS,B  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: WINDHAM

DATE/TIME OF TEST: 11/05/2020 04:18:11

---

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:15	
Control Test	0.080	04:16	
Air Blank	0.000	04:16	
Subject Sample	# 1	04:16	Subject Test Refused
Breath Volume	0.000	04:16	
Air Blank	0.000	04:17	
Control Test	0.080	04:17	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

---

Data Download Date/Time: 11/16/2020 11:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: PARRISH,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: TOMLINSON

DATE/TIME OF TEST: 11/06/2020 05:17:41

---

Diagnostics Check	OK	05:11
Air Blank	0.000	05:11
Control Test	0.080	05:12
Air Blank	0.000	05:12
Subject Sample # 1	0.000	05:13
Breath Volume	2.082	05:13
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.000	05:16
Breath Volume	1.937	05:16
Air Blank	0.000	05:16
Control Test	0.080	05:17
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

---

Data Download Date/Time: 11/16/2020 11:31



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: PEREZ,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 11/07/2020 01:56:43

---

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.164	01:50
Breath Volume	1.972	01:50
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.160	01:55
Breath Volume	2.382	01:55
Air Blank	0.000	01:55
Control Test	0.078	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

---

Data Download Date/Time: 11/16/2020 11:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: RODRIGUEZ, M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 11/07/2020 20:28:17

---

Diagnostics Check	OK	20:22
Air Blank	0.000	20:22
Control Test	0.080	20:23
Air Blank	0.000	20:23
Subject Sample # 1	0.000	20:23
Breath Volume	4.417	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.000	20:26
Breath Volume	4.035	20:26
Air Blank	0.000	20:27
Control Test	0.079	20:27
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

---

Data Download Date/Time: 11/16/2020 11:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LUGO, R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: RAGADIO

DATE/TIME OF TEST: 11/07/2020 21:49:30

---

Diagnostics Check	OK	21:40
Air Blank	0.000	21:40
Control Test	0.079	21:40
Air Blank	0.000	21:41
Subject Sample # 1	0.094	21:42
Breath Volume	1.347	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.094	21:47
Breath Volume	1.285	21:47
Air Blank	0.000	21:48
Control Test	0.079	21:48
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

---

Data Download Date/Time: 11/16/2020 11:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FISETTE, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: ASHBY

DATE/TIME OF TEST: 11/08/2020 01:10:52

---

Diagnostics Check	OK	01:04
Air Blank	0.000	01:05
Control Test	0.079	01:05
Air Blank	0.000	01:06
Subject Sample # 1	0.092	01:06
Breath Volume	1.691	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:08
Subject Sample # 2	0.090	01:09
Breath Volume	1.171	01:09
Air Blank	0.000	01:09
Control Test	0.079	01:10
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

---

Data Download Date/Time: 11/16/2020 11:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: TOLBERT

DATE/TIME OF TEST: 11/08/2020 04:40:16

---

Diagnostics Check	OK	04:37	
Air Blank	0.000	04:37	
Control Test	0.079	04:38	
Air Blank	0.000	04:38	
Subject Sample	# 1	04:38	Subject Test Refused
Breath Volume	0.000	04:38	
Air Blank	0.000	04:39	
Control Test	0.079	04:39	
Air Blank	0.000	04:40	
Diagnostics Check	OK	04:40	

---

Data Download Date/Time: 11/16/2020 11:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: BARNES, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: HERNANDEZ RODRIGUEZ

DATE/TIME OF TEST: 11/08/2020 05:12:31

---

Diagnostics Check	OK	05:06
Air Blank	0.000	05:06
Control Test	0.079	05:07
Air Blank	0.000	05:07
Subject Sample # 1	0.195	05:08
Breath Volume	1.390	05:08
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 2	0.203	05:10
Breath Volume	1.781	05:10
Air Blank	0.000	05:11
Control Test	0.078	05:11
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

---

Data Download Date/Time: 11/16/2020 11:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FISETTE, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: COCKERILL

DATE/TIME OF TEST: 11/08/2020 22:20:31

---

Diagnostics Check		OK	22:14
Air Blank		0.000	22:15
Control Test		0.080	22:15
Air Blank		0.000	22:15
Subject Sample	# 1	0.167	22:16
Breath Volume		2.417	22:16
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.175	22:18
Breath Volume		2.152	22:18
Air Blank		0.000	22:19
Control Test		0.080	22:19
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

---

Data Download Date/Time: 11/16/2020 11:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 11/09/2020 03:10:37

---

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.180	03:06
Breath Volume	2.804	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.173	03:09
Breath Volume	2.968	03:09
Air Blank	0.000	03:09
Control Test	0.078	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

---

Data Download Date/Time: 11/16/2020 11:31



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOHN,ANTHONY,W  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MOELLER,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: PATINO

DATE/TIME OF TEST: 11/09/2020 17:18:44

---

Diagnostics Check	OK	17:15	
Air Blank	0.000	17:16	
Control Test	0.080	17:16	
Air Blank	0.000	17:17	
Subject Sample	# 1	17:17	Subject Test Refused
Breath Volume	0.000	17:17	
Air Blank	0.000	17:17	
Control Test	0.079	17:18	
Air Blank	0.000	17:18	
Diagnostics Check	OK	17:18	

---

Data Download Date/Time: 11/16/2020 11:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FISETTE, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 11/09/2020 23:15:07

---

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:10
Subject Sample # 1	0.000	23:10
Breath Volume	2.179	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.000	23:13
Breath Volume	1.339	23:13
Air Blank	0.000	23:14
Control Test	0.080	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

---

Data Download Date/Time: 11/16/2020 11:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/10/2020 20:33:39

---

Diagnostics Check	OK	20:26
Air Blank	0.000	20:27
Control Test	0.079	20:27
Air Blank	0.000	20:27
Subject Sample # 1	0.173	20:28
Breath Volume	2.035	20:28
Air Blank	0.000	20:29
Air Blank	0.000	20:31
Subject Sample # 2	0.166	20:32
Breath Volume	1.605	20:32
Air Blank	0.000	20:32
Control Test	0.079	20:33
Air Blank	0.000	20:33
Diagnostics Check	OK	20:33

---

Data Download Date/Time: 11/16/2020 11:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: STANLEY, T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: GEORGE

DATE/TIME OF TEST: 11/11/2020 00:31:51

---

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:29	
Control Test	0.079	00:29	
Air Blank	0.000	00:30	
Subject Sample	# 1	00:30	Subject Test Refused
Breath Volume	0.000	00:30	
Air Blank	0.000	00:30	
Control Test	0.080	00:31	
Air Blank	0.000	00:31	
Diagnostics Check	OK	00:31	

---

Data Download Date/Time: 11/16/2020 11:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: STANLEY, T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: ZIHAL

DATE/TIME OF TEST: 11/11/2020 02:58:58

---

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.181	02:54
Breath Volume	2.925	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.184	02:57
Breath Volume	3.289	02:57
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

---

Data Download Date/Time: 11/16/2020 11:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 11/12/2020 01:45:21

---

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.079	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.132	01:40
Breath Volume	1.781	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.136	01:43
Breath Volume	1.371	01:43
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

---

Data Download Date/Time: 11/16/2020 11:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MANDILE, J  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: BISONO

DATE/TIME OF TEST: 11/12/2020 06:22:22

---

Diagnostics Check	OK	06:15
Air Blank	0.000	06:15
Control Test	0.080	06:15
Air Blank	0.000	06:16
Subject Sample # 1	0.063	06:17
Breath Volume	1.878	06:17
Air Blank	0.000	06:18
Air Blank	0.000	06:19
Subject Sample # 2	0.063	06:20
Breath Volume	2.269	06:20
Air Blank	0.000	06:21
Control Test	0.079	06:21
Air Blank	0.000	06:22
Diagnostics Check	OK	06:22

---

Data Download Date/Time: 11/16/2020 11:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HART,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: MONTGOMERY

DATE/TIME OF TEST: 11/14/2020 21:36:42

---

Diagnostics Check	OK	21:25
Air Blank	0.000	21:26
Control Test	0.080	21:26
Air Blank	0.000	21:27
Subject Sample # 1	0.254	21:27
Breath Volume	2.859	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.215	21:31
Breath Volume	1.363	21:31
Air Blank	0.000	21:31
Air Blank	0.000	21:33
Subject Sample # 3	0.214	21:35
Breath Volume	1.355	21:35
Air Blank	0.000	21:35
Control Test	0.079	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

---

Data Download Date/Time: 11/16/2020 11:32



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: FEHSE

DATE/TIME OF TEST: 11/14/2020 23:24:04

---

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.080	23:18
Air Blank	0.000	23:19
Subject Sample # 1	0.000	23:19
Breath Volume	3.781	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.000	23:22
Breath Volume	3.757	23:22
Air Blank	0.000	23:23
Control Test	0.079	23:23
Air Blank	0.000	23:23
Diagnostics Check	OK	23:24

---

Data Download Date/Time: 11/16/2020 11:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HIXSON,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: NOLASCO PEREZ

DATE/TIME OF TEST: 11/15/2020 01:24:00

---

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:21	
Control Test	0.079	01:21	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:23	
Control Test	0.079	01:23	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:24	

---

Data Download Date/Time: 11/16/2020 11:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SIMMONS,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: POLVERINE

DATE/TIME OF TEST: 11/15/2020 03:39:57

---

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.079	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.092	03:35
Breath Volume	1.574	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:37
Subject Sample # 2	0.093	03:38
Breath Volume	1.933	03:38
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

---

Data Download Date/Time: 11/16/2020 11:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: O'CONNELL,S  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: LIMA

DATE/TIME OF TEST: 11/15/2020 05:26:26

---

Diagnostics Check	OK	05:23	
Air Blank	0.000	05:23	
Control Test	0.079	05:24	
Air Blank	0.000	05:24	
Subject Sample	# 1	05:25	Subject Test Refused
Breath Volume	0.000	05:25	
Air Blank	0.000	05:25	
Control Test	0.080	05:25	
Air Blank	0.000	05:26	
Diagnostics Check	OK	05:26	

---

Data Download Date/Time: 11/16/2020 11:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLDEN,E,T  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 11/15/2020 19:11:15

---

Diagnostics Check	OK	19:08	
Air Blank	0.000	19:08	
Control Test	0.080	19:09	
Air Blank	0.000	19:09	
Subject Sample	# 1	19:09	Subject Test Refused
Breath Volume	0.000	19:09	
Air Blank	0.000	19:10	
Control Test	0.079	19:10	
Air Blank	0.000	19:11	
Diagnostics Check	OK	19:11	

---

Data Download Date/Time: 11/16/2020 11:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLDEN,E,T  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LAWSON,N  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: VIAFRANCO

DATE/TIME OF TEST: 11/15/2020 22:06:42

---

Diagnostics Check	OK	22:00
Air Blank	0.000	22:00
Control Test	0.080	22:01
Air Blank	0.000	22:01
Subject Sample # 1	0.127	22:02
Breath Volume	1.574	22:02
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.119	22:05
Breath Volume	1.574	22:05
Air Blank	0.000	22:05
Control Test	0.079	22:06
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

---

Data Download Date/Time: 11/16/2020 11:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FISETTE, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 11/18/2020 00:01:38

---

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:59	
Control Test	0.080	23:59	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:00	Subject Test Refused
Breath Volume	0.000	00:00	
Air Blank	0.000	00:00	
Control Test	0.081	00:01	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

---

Data Download Date/Time: 12/09/2020 13:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: WANISH

DATE/TIME OF TEST: 11/18/2020 01:33:04

---

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:30	
Control Test	0.080	01:31	
Air Blank	0.000	01:31	
Subject Sample	# 1	01:31	Subject Test Refused
Breath Volume	0.000	01:31	
Air Blank	0.000	01:32	
Control Test	0.080	01:32	
Air Blank	0.000	01:32	
Diagnostics Check	OK	01:33	

---

Data Download Date/Time: 12/09/2020 13:36



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MORRISON,M  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: ROSTAYEE

DATE/TIME OF TEST: 11/18/2020 04:36:15

---

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.080	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.000	04:31
Breath Volume	2.125	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.000	04:34
Breath Volume	2.621	04:34
Air Blank	0.000	04:35
Control Test	0.081	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

---

Data Download Date/Time: 12/09/2020 13:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: STANLEY, T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: MENARD

DATE/TIME OF TEST: 11/19/2020 00:10:40

---

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.081	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.129	00:06
Breath Volume	2.082	00:06
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.129	00:09
Breath Volume	1.910	00:09
Air Blank	0.000	00:09
Control Test	0.080	00:10
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

---

Data Download Date/Time: 12/09/2020 13:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: RODRIGUEZ, M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: SHUPE

DATE/TIME OF TEST: 11/20/2020 22:21:42

---

Diagnostics Check	OK	22:18	
Air Blank	0.000	22:19	
Control Test	0.080	22:19	
Air Blank	0.000	22:20	
Subject Sample	# 1	22:20	Subject Test Refused
Breath Volume	0.000	22:20	
Air Blank	0.000	22:20	
Control Test	0.079	22:21	
Air Blank	0.000	22:21	
Diagnostics Check	OK	22:21	

---

Data Download Date/Time: 12/09/2020 13:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LUGO, R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: DOMINGO

DATE/TIME OF TEST: 11/21/2020 01:08:33

---

Diagnostics Check		OK	01:02
Air Blank		0.000	01:02
Control Test		0.080	01:02
Air Blank		0.000	01:03
Subject Sample	# 1	0.158	01:04
Breath Volume		2.488	01:04
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.164	01:06
Breath Volume		2.636	01:06
Air Blank		0.000	01:07
Control Test		0.080	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

---

Data Download Date/Time: 12/09/2020 13:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNS, B  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: MCCALL

DATE/TIME OF TEST: 11/21/2020 02:20:49

---

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.079	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.050	02:16
Breath Volume	1.570	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.050	02:19
Breath Volume	1.605	02:19
Air Blank	0.000	02:19
Control Test	0.079	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

---

Data Download Date/Time: 12/09/2020 13:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LAWSON,N  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: WONDOLOWSKI

DATE/TIME OF TEST: 11/22/2020 00:02:33

---

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:59	
Control Test	0.080	23:59	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:01	Subject Test Refused
Breath Volume	0.000	00:01	
Air Blank	0.000	00:01	
Control Test	0.080	00:01	
Air Blank	0.000	00:02	
Diagnostics Check	OK	00:02	

---

Data Download Date/Time: 12/09/2020 13:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HILLIER,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: MOBLEY

DATE/TIME OF TEST: 11/22/2020 03:53:40

---

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.021	03:49
Breath Volume	2.617	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.021	03:52
Breath Volume	2.929	03:52
Air Blank	0.000	03:52
Control Test	0.079	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

---

Data Download Date/Time: 12/09/2020 13:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DAVIS,P  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: PATE

DATE/TIME OF TEST: 11/22/2020 05:10:06

---

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.080	05:04
Air Blank		0.000	05:04
Subject Sample	# 1	0.208	05:05
Breath Volume		2.480	05:05
Air Blank		0.000	05:06
Air Blank		0.000	05:07
Subject Sample	# 2	0.202	05:08
Breath Volume		2.226	05:08
Air Blank		0.000	05:09
Control Test		0.079	05:09
Air Blank		0.000	05:10
Diagnostics Check		OK	05:10

---

Data Download Date/Time: 12/09/2020 13:37



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNS,B  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: DALZIEL

DATE/TIME OF TEST: 11/22/2020 22:08:51

---

Diagnostics Check	OK	22:02
Air Blank	0.000	22:03
Control Test	0.080	22:03
Air Blank	0.000	22:03
Subject Sample # 1	0.000	22:04
Breath Volume	3.523	22:04
Air Blank	0.000	22:04
Air Blank	0.000	22:07
Subject Sample # 2	0.000	22:07
Breath Volume	3.777	22:07
Air Blank	0.000	22:07
Control Test	0.079	22:08
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

---

Data Download Date/Time: 12/09/2020 13:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRESCOTT,A,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LAWSON,N  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: MORALEZ

DATE/TIME OF TEST: 11/23/2020 00:06:50

---

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:03	
Control Test	0.079	00:03	
Air Blank	0.000	00:04	
Subject Sample	# 1	00:05	Subject Test Refused
Breath Volume	0.000	00:05	
Air Blank	0.000	00:05	
Control Test	0.080	00:06	
Air Blank	0.000	00:06	
Diagnostics Check	OK	00:06	

---

Data Download Date/Time: 12/09/2020 13:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRESCOTT,A,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LAWSON,N  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CAMERON

DATE/TIME OF TEST: 11/23/2020 02:37:36

---

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.079	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.109	02:32
Breath Volume		2.296	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.104	02:36
Breath Volume		1.996	02:36
Air Blank		0.000	02:36
Control Test		0.079	02:37
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

---

Data Download Date/Time: 12/09/2020 13:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRESCOTT,A,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: PARRISH,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 11/23/2020 04:27:36

---

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:25	
Control Test	0.080	04:25	
Air Blank	0.000	04:26	
Subject Sample	# 1	04:26	Subject Test Refused
Breath Volume	0.000	04:26	
Air Blank	0.000	04:26	
Control Test	0.078	04:27	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

---

Data Download Date/Time: 12/09/2020 13:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MUNOZ,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: BRECKENRIDGE

DATE/TIME OF TEST: 11/23/2020 23:06:15

---

Diagnostics Check	OK	23:02	
Air Blank	0.000	23:03	
Control Test	0.080	23:03	
Air Blank	0.000	23:03	
Subject Sample	# 1	23:04	Interferent Detect
Breath Volume	2.460	23:04	
Air Blank	0.000	23:05	
Control Test	0.080	23:05	
Air Blank	0.000	23:06	
Diagnostics Check	OK	23:06	

---

Data Download Date/Time: 12/09/2020 13:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MUNOZ,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: BRECKENRIDGE

DATE/TIME OF TEST: 11/23/2020 23:12:12

---

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:09	
Control Test	0.079	23:09	
Air Blank	0.000	23:10	
Subject Sample	# 1	23:10	Interferent Detect
Breath Volume	2.234	23:10	
Air Blank	0.000	23:11	
Control Test	0.080	23:11	
Air Blank	0.000	23:12	
Diagnostics Check	OK	23:12	

---

Data Download Date/Time: 12/09/2020 13:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRESCOTT,A,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JENSEN,D  
ARRESTING OFFICER AGENCY: Neptune Beach PD  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: KEMP

DATE/TIME OF TEST: 11/25/2020 00:11:04

---

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:08	
Control Test	0.080	00:08	
Air Blank	0.000	00:09	
Subject Sample	# 1	00:09	Subject Test Refused
Breath Volume	0.000	00:09	
Air Blank	0.000	00:10	
Control Test	0.081	00:10	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:11	

---

Data Download Date/Time: 12/09/2020 13:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNS,B  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: OLVERA PEREZ

DATE/TIME OF TEST: 11/26/2020 00:22:25

---

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.080	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.117	00:17
Breath Volume	2.261	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.115	00:20
Breath Volume	1.539	00:20
Air Blank	0.000	00:21
Control Test	0.079	00:21
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

---

Data Download Date/Time: 12/09/2020 13:37



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FILBERT,E  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 11/26/2020 05:06:37

---

Diagnostics Check	OK	05:00
Air Blank	0.000	05:00
Control Test	0.079	05:01
Air Blank	0.000	05:01
Subject Sample # 1	0.135	05:02
Breath Volume	3.046	05:02
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.128	05:05
Breath Volume	3.117	05:05
Air Blank	0.000	05:05
Control Test	0.079	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

---

Data Download Date/Time: 12/09/2020 13:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: RODRIGUEZ, M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: HOLDEN

DATE/TIME OF TEST: 11/26/2020 23:31:05

---

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:25
Subject Sample # 1	0.161	23:26
Breath Volume	1.441	23:26
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.175	23:29
Breath Volume	1.617	23:29
Air Blank	0.000	23:30
Control Test	0.080	23:30
Air Blank	0.000	23:30
Diagnostics Check	OK	23:31

---

Data Download Date/Time: 12/09/2020 13:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: HOPSON

DATE/TIME OF TEST: 11/27/2020 00:57:29

---

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:55	
Control Test	0.079	00:55	
Air Blank	0.000	00:55	
Subject Sample	# 1	00:56	Subject Test Refused
Breath Volume	0.000	00:56	
Air Blank	0.000	00:56	
Control Test	0.079	00:56	
Air Blank	0.000	00:57	
Diagnostics Check	OK	00:57	

---

Data Download Date/Time: 12/09/2020 13:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HAWKINS, A  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: AUGE

DATE/TIME OF TEST: 11/27/2020 02:47:22

---

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.136	02:42
Breath Volume	1.410	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.153	02:45
Breath Volume	1.308	02:45
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

---

Data Download Date/Time: 12/09/2020 13:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FILBERT, E  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: SPANN

DATE/TIME OF TEST: 11/27/2020 04:55:40

---

Diagnostics Check	OK	04:49
Air Blank	0.000	04:49
Control Test	0.079	04:50
Air Blank	0.000	04:50
Subject Sample # 1	0.139	04:51
Breath Volume	2.437	04:51
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.142	04:54
Breath Volume	2.421	04:54
Air Blank	0.000	04:54
Control Test	0.079	04:55
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

---

Data Download Date/Time: 12/09/2020 13:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNS,D,L  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KEEN,C  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: STOKES

DATE/TIME OF TEST: 11/27/2020 20:53:13

---

Diagnostics Check	OK	20:50	
Air Blank	0.000	20:50	
Control Test	0.080	20:51	
Air Blank	0.000	20:51	
Subject Sample	# 1	20:51	Subject Test Refused
Breath Volume	0.000	20:51	
Air Blank	0.000	20:52	
Control Test	0.079	20:52	
Air Blank	0.000	20:53	
Diagnostics Check	OK	20:53	

---

Data Download Date/Time: 12/09/2020 13:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNS,D,L  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HAWKINS,A  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: ATWELL

DATE/TIME OF TEST: 11/27/2020 22:06:30

---

Diagnostics Check	OK	22:00
Air Blank	0.000	22:00
Control Test	0.079	22:00
Air Blank	0.000	22:01
Subject Sample # 1	0.245	22:02
Breath Volume	3.210	22:02
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.251	22:04
Breath Volume	2.375	22:04
Air Blank	0.000	22:05
Control Test	0.079	22:05
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

---

Data Download Date/Time: 12/09/2020 13:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNS,D,L  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KEEN,C  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ZUMBA

DATE/TIME OF TEST: 11/28/2020 04:27:06

---

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.080	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.116	04:22
Breath Volume		2.261	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.116	04:25
Breath Volume		2.539	04:25
Air Blank		0.000	04:26
Control Test		0.080	04:26
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

---

Data Download Date/Time: 12/09/2020 13:37



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOHN,A,W  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HOBBS,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 07:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:52

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/28/2020 10:16:06

---

Diagnostics Check	OK	10:13	
Air Blank	0.000	10:13	
Control Test	0.079	10:14	
Air Blank	0.000	10:14	
Subject Sample	# 1	10:14	Subject Test Refused
Breath Volume	0.000	10:14	
Air Blank	0.000	10:15	
Control Test	0.080	10:15	
Air Blank	0.000	10:16	
Diagnostics Check	OK	10:16	

---

Data Download Date/Time: 12/09/2020 13:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOHN,A,W  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DIAZ,I  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 12:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:48

SUBJECT'S LAST NAME: LAYTON

DATE/TIME OF TEST: 11/28/2020 14:12:09

---

Diagnostics Check	OK	14:09	
Air Blank	0.000	14:09	
Control Test	0.080	14:10	
Air Blank	0.000	14:10	
Subject Sample	# 1	14:10	Subject Test Refused
Breath Volume	0.000	14:10	
Air Blank	0.000	14:11	
Control Test	0.079	14:11	
Air Blank	0.000	14:12	
Diagnostics Check	OK	14:12	

---

Data Download Date/Time: 12/09/2020 13:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOHN,A,W  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FISETTE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:09

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 11/28/2020 18:32:02

---

Diagnostics Check	OK	18:29	
Air Blank	0.000	18:29	
Control Test	0.080	18:30	
Air Blank	0.000	18:30	
Subject Sample	# 1	18:30	Subject Test Refused
Breath Volume	0.000	18:30	
Air Blank	0.000	18:31	
Control Test	0.079	18:31	
Air Blank	0.000	18:31	
Diagnostics Check	OK	18:32	

---

Data Download Date/Time: 12/09/2020 13:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNS,D,L  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DESOUZA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 11/29/2020 19:42:27

---

Diagnostics Check	OK	19:35
Air Blank	0.000	19:36
Control Test	0.080	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.187	19:37
Breath Volume	2.785	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.186	19:40
Breath Volume	2.113	19:40
Air Blank	0.000	19:41
Control Test	0.079	19:41
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

---

Data Download Date/Time: 12/09/2020 13:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNS,D,L  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LUGO,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A2  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 11/30/2020 04:28:21

---

Diagnostics Check		OK	04:21
Air Blank		0.000	04:22
Control Test		0.080	04:22
Air Blank		0.000	04:23
Subject Sample	# 1	0.203	04:23
Breath Volume		2.937	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.200	04:26
Breath Volume		2.980	04:26
Air Blank		0.000	04:27
Control Test		0.080	04:27
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

---

Data Download Date/Time: 12/09/2020 13:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MENDEZ,JAMIE,J  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: MENDEZ,JAMIE  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: RAHO

DATE/TIME OF TEST: 11/08/2020 18:13:09

---

Diagnostics Check	OK	18:06
Air Blank	0.000	18:07
Control Test	0.081	18:07
Air Blank	0.000	18:08
Subject Sample # 1	0.000	18:08
Breath Volume	3.382	18:08
Air Blank	0.000	18:09
Air Blank	0.000	18:11
Subject Sample # 2	0.000	18:11
Breath Volume	3.632	18:11
Air Blank	0.000	18:12
Control Test	0.082	18:12
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

---

Data Download Date/Time: 11/27/2020 01:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAIR,TIMOTHY,D  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: BLAIR,TIMOTHY  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: SCHWARTZ

DATE/TIME OF TEST: 11/10/2020 19:08:41

---

Diagnostics Check	OK	19:02
Air Blank	0.000	19:02
Control Test	0.080	19:02
Air Blank	0.000	19:03
Subject Sample # 1	0.203	19:04
Breath Volume	1.628	19:04
Air Blank	0.000	19:04
Air Blank	0.000	19:06
Subject Sample # 2	0.199	19:07
Breath Volume	1.144	19:07
Air Blank	0.000	19:07
Control Test	0.079	19:08
Air Blank	0.000	19:08
Diagnostics Check	OK	19:08

---

Data Download Date/Time: 11/27/2020 01:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MENDEZ,JAMIE,J  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: MENDEZ,JAMIE  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: BARTLETT

DATE/TIME OF TEST: 11/12/2020 23:52:06

---

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.081	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.121	23:47
Breath Volume	2.527	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.117	23:50
Breath Volume	2.531	23:50
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

---

Data Download Date/Time: 11/27/2020 01:41



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAYLOR,CHANNING,L  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: WHITFIELD,CIARA  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: PIERCY

DATE/TIME OF TEST: 11/15/2020 23:01:29

---

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.080	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.169	22:56
Breath Volume	1.765	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.168	22:59
Breath Volume	2.023	22:59
Air Blank	0.000	23:00
Control Test	0.081	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

---

Data Download Date/Time: 11/27/2020 01:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MACK,SHAWN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: MACK,SHAWN  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 11/20/2020 03:16:24

---

Diagnostics Check		OK	03:04	
Air Blank		0.000	03:04	
Control Test		0.080	03:05	
Air Blank		0.000	03:05	
Subject Sample	# 1	0.158	03:08	Volume Not Met
Breath Volume		0.250	03:08	
Air Blank		0.000	03:09	
Air Blank		0.000	03:11	
Subject Sample	# 2	0.190	03:11	
Breath Volume		1.714	03:11	
Air Blank		0.000	03:12	
Air Blank		0.000	03:14	
Subject Sample	# 3	0.191	03:14	
Breath Volume		1.925	03:14	
Air Blank		0.000	03:15	
Control Test		0.079	03:15	
Air Blank		0.000	03:16	
Diagnostics Check		OK	03:16	

---

Data Download Date/Time: 11/27/2020 01:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAIR,TIMOTHY,D  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: BLAIR,TIMOTHY  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 11/23/2020 17:54:04

---

Diagnostics Check	OK	17:40
Air Blank	0.000	17:40
Control Test	0.080	17:41
Air Blank	0.000	17:41
Subject Sample # 1	0.233	17:44
Breath Volume	1.554	17:44
Air Blank	0.000	17:44
Air Blank	0.000	17:46
Subject Sample # 2	0.278	17:47
Breath Volume	1.699	17:47
Air Blank	0.000	17:48
Air Blank	0.000	17:50
Subject Sample # 3	0.238	17:52
Breath Volume	1.347	17:52
Air Blank	0.000	17:53
Control Test	0.078	17:53
Air Blank	0.000	17:53
Diagnostics Check	OK	17:54

---

Data Download Date/Time: 11/27/2020 01:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRIFFIN,SUMMER,N  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GRIFFIN,SUMMER  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: HANOMAN

DATE/TIME OF TEST: 11/01/2020 23:11:43

---

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.080	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.000	23:07
Breath Volume	1.351	23:07
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.000	23:10
Breath Volume	1.851	23:10
Air Blank	0.000	23:10
Control Test	0.080	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

---

Data Download Date/Time: 11/09/2020 13:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRINIDAD,RICKY,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GALLOWAY,GRANT  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 11/04/2020 03:10:04

---

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.122	03:05
Breath Volume	1.835	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.120	03:08
Breath Volume	2.929	03:08
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:10

---

Data Download Date/Time: 11/09/2020 13:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GRIFFIN,RAEFORD  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BOGUE

DATE/TIME OF TEST: 11/05/2020 02:58:54

---

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.078	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.162	02:54
Breath Volume	1.734	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.154	02:57
Breath Volume	1.449	02:57
Air Blank	0.000	02:57
Control Test	0.079	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

---

Data Download Date/Time: 11/09/2020 13:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUNNER,JOSEPH,A  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: RUNNER,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 11/05/2020 22:27:57

---

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:22	
Control Test	0.079	22:22	
Air Blank	0.000	22:23	
Subject Sample	# 1 0.109	22:26	Volume Not Met
Breath Volume	0.339	22:26	
Air Blank		22:26	RFI Detect
Air Blank	0.000	22:27	
Control Test	0.079	22:27	
Air Blank	0.000	22:27	
Diagnostics Check	OK	22:27	

---

Data Download Date/Time: 11/09/2020 13:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUNNER,JOSEPH,A  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: RUNNER,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 11/05/2020 22:32:49

---

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:29	
Control Test	0.078	22:30	
Air Blank	0.000	22:30	
Subject Sample	# 1	22:32	Subject Test Refused
Breath Volume	0.371	22:32	
Air Blank		22:32	RFI Detect
Air Blank		22:32	RFI Detect

---

Data Download Date/Time: 11/09/2020 13:05



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUNNER,JOSEPH,A  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: RUNNER,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 11/05/2020 22:47:40

---

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:36	
Control Test	0.078	22:37	
Air Blank	0.000	22:37	
Subject Sample	# 1	22:40	No Sample Provided
Breath Volume	0.000	22:40	
Air Blank	0.000	22:41	
Air Blank	0.000	22:43	
Subject Sample	# 2	22:46	No Sample Provided
Breath Volume	0.000	22:46	
Air Blank	0.000	22:46	
Control Test	0.079	22:47	
Air Blank	0.000	22:47	
Diagnostics Check	OK	22:47	

---

Data Download Date/Time: 11/09/2020 13:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOTO,JOHN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: NOTO,JOHN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: MCCABE

DATE/TIME OF TEST: 11/07/2020 03:00:27

---

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:52	
Control Test	0.080	02:52	
Air Blank	0.000	02:53	
Subject Sample	# 1	02:56	No Sample Provided
Breath Volume	0.000	02:56	
Air Blank	0.000	02:56	
Air Blank	0.000	02:58	
Subject Sample	# 2	02:59	Subject Test Refused
Breath Volume	0.000	02:59	
Air Blank	0.000	02:59	
Control Test	0.080	02:59	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

---

Data Download Date/Time: 11/09/2020 13:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOTO,JOHN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GALLOWAY,GRANT  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: DUBOVSKY

DATE/TIME OF TEST: 11/07/2020 23:16:42

---

Diagnostics Check	OK	23:09
Air Blank	0.000	23:10
Control Test	0.080	23:10
Air Blank	0.000	23:10
Subject Sample # 1	0.123	23:12
Breath Volume	1.890	23:12
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.119	23:15
Breath Volume	1.742	23:15
Air Blank	0.000	23:15
Control Test	0.080	23:16
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

---

Data Download Date/Time: 11/09/2020 13:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRINIDAD,RICKY,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: TRINIDAD,RICKY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 11/08/2020 00:15:21

---

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.122	00:10
Breath Volume	2.332	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.128	00:13
Breath Volume	2.011	00:13
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

---

Data Download Date/Time: 11/09/2020 13:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRINIDAD,RICKY,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: LEE,MATTHEW  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14219080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 11/08/2020 00:28:17

---

Diagnostics Check		OK	00:21
Air Blank		0.000	00:22
Control Test		0.080	00:22
Air Blank		0.000	00:23
Subject Sample	# 1	0.212	00:23
Breath Volume		2.277	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.221	00:26
Breath Volume		2.300	00:26
Air Blank		0.000	00:27
Control Test		0.078	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

---

Data Download Date/Time: 11/09/2020 13:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIBBEY,BRADLEY,A  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: KIBBEY,BRADLEY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: DICLEMENTE

DATE/TIME OF TEST: 11/10/2020 22:47:20

---

Diagnostics Check	OK	22:40
Air Blank	0.000	22:41
Control Test	0.080	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.210	22:42
Breath Volume	1.984	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.204	22:45
Breath Volume	1.535	22:45
Air Blank	0.000	22:46
Control Test	0.079	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

---

Data Download Date/Time: 12/18/2020 18:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANNON, MATTHEW, F  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: LEE, MATTHEW  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: BACCHUS

DATE/TIME OF TEST: 11/11/2020 21:00:31

---

Diagnostics Check	OK	20:52
Air Blank	0.000	20:53
Control Test	0.080	20:53
Air Blank	0.000	20:53
Subject Sample # 1	0.138	20:55
Breath Volume	2.929	20:55
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample # 2	0.141	20:59
Breath Volume	1.976	20:59
Air Blank	0.000	20:59
Control Test	0.078	20:59
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

---

Data Download Date/Time: 12/18/2020 18:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PALESE, ANTHONY, W  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: DION, K  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 09:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:42

SUBJECT'S LAST NAME: FRANKENFIELD

DATE/TIME OF TEST: 11/12/2020 13:08:48

---

Diagnostics Check	OK	13:02
Air Blank	0.000	13:03
Control Test	0.080	13:03
Air Blank	0.000	13:04
Subject Sample # 1	0.000	13:04
Breath Volume	2.765	13:04
Air Blank	0.000	13:05
Air Blank	0.000	13:07
Subject Sample # 2	0.000	13:07
Breath Volume	3.054	13:07
Air Blank	0.000	13:07
Control Test	0.080	13:08
Air Blank	0.000	13:08
Diagnostics Check	OK	13:08

---

Data Download Date/Time: 12/18/2020 18:56



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PULASKI,DAMON,J  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: PETERS,JOHN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:22

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 11/12/2020 14:58:23

---

Diagnostics Check	OK	14:51
Air Blank	0.000	14:52
Control Test	0.079	14:52
Air Blank	0.000	14:53
Subject Sample # 1	0.228	14:53
Breath Volume	2.343	14:53
Air Blank	0.000	14:54
Air Blank	0.000	14:56
Subject Sample # 2	0.230	14:56
Breath Volume	2.929	14:56
Air Blank	0.000	14:57
Control Test	0.079	14:57
Air Blank	0.000	14:58
Diagnostics Check	OK	14:58

---

Data Download Date/Time: 12/18/2020 18:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUNNER,JOSEPH,A  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: RUNNER,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SMALL

DATE/TIME OF TEST: 11/14/2020 03:17:16

---

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.080	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.109	03:12
Breath Volume	1.785	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.106	03:15
Breath Volume	2.093	03:15
Air Blank	0.000	03:16
Control Test	0.078	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

---

Data Download Date/Time: 12/18/2020 18:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANE,MARC,T  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: RISING,PATRICK  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: HANSEL

DATE/TIME OF TEST: 11/14/2020 19:42:13

---

Diagnostics Check	OK	19:35
Air Blank	0.000	19:36
Control Test	0.081	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.000	19:37
Breath Volume	1.886	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.000	19:40
Breath Volume	1.359	19:40
Air Blank	0.000	19:41
Control Test	0.080	19:41
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

---

Data Download Date/Time: 12/18/2020 18:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FLINT,KILIAN,A  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: FLINT,KILIAN  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: ROBBINS

DATE/TIME OF TEST: 11/15/2020 02:19:20

---

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.081	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.000	02:14
Breath Volume	2.300	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.000	02:17
Breath Volume	2.121	02:17
Air Blank	0.000	02:18
Control Test	0.081	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

---

Data Download Date/Time: 12/18/2020 18:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUNNER,JOSEPH,A  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: RUNNER,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: VOURVOPOULOS

DATE/TIME OF TEST: 11/15/2020 03:10:33

---

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:02	
Control Test	0.080	03:02	
Air Blank	0.000	03:03	
Subject Sample	# 1	03:06	No Sample Provided
Breath Volume	0.000	03:06	
Air Blank	0.000	03:06	
Air Blank	0.000	03:08	
Subject Sample	# 2	03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:09	
Control Test	0.080	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

---

Data Download Date/Time: 12/18/2020 18:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANNON, MATTHEW, F  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HANNON, MATTHEW  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/15/2020 03:57:44

---

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:55	
Control Test	0.080	03:55	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank	0.000	03:56	
Control Test	0.080	03:57	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

---

Data Download Date/Time: 12/18/2020 18:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GRIFFIN,RAEFORD  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: PARTHEYMULLER

DATE/TIME OF TEST: 11/15/2020 05:46:02

---

Diagnostics Check	OK	05:38
Air Blank	0.000	05:38
Control Test	0.080	05:39
Air Blank	0.000	05:39
Subject Sample # 1	0.173	05:40
Breath Volume	1.246	05:40
Air Blank	0.000	05:41
Air Blank	0.000	05:43
Subject Sample # 2	0.170	05:44
Breath Volume	1.718	05:44
Air Blank	0.000	05:45
Control Test	0.079	05:45
Air Blank	0.000	05:45
Diagnostics Check	OK	05:46

---

Data Download Date/Time: 12/18/2020 18:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EDGE,SHAWN,P  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: ROBERTS,JORDAN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CRESPO GONZALEZ

DATE/TIME OF TEST: 11/16/2020 03:12:51

---

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.080	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.154	03:07
Breath Volume	1.234	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.163	03:11
Breath Volume	1.871	03:11
Air Blank	0.000	03:11
Control Test	0.079	03:12
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

---

Data Download Date/Time: 12/18/2020 18:57



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOTO,JOHN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: DION,KEVIN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: KOTLARZ

DATE/TIME OF TEST: 11/17/2020 23:24:23

---

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.081	23:17
Air Blank	0.000	23:18
Subject Sample # 1	0.120	23:19
Breath Volume	1.164	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.110	23:22
Breath Volume	1.761	23:22
Air Blank	0.000	23:23
Control Test	0.080	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

---

Data Download Date/Time: 12/18/2020 18:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GRIFFIN,RAEFORD  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ELY

DATE/TIME OF TEST: 11/19/2020 23:36:32

---

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.081	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.202	23:31
Breath Volume	1.347	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.192	23:34
Breath Volume	2.683	23:34
Air Blank	0.000	23:35
Control Test	0.080	23:35
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

---

Data Download Date/Time: 12/18/2020 18:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANNON, MATTHEW, F  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HANNON, MATTHEW  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: HERDT

DATE/TIME OF TEST: 11/21/2020 03:20:04

---

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:10	
Control Test	0.080	03:10	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:14	No Sample Provided
Breath Volume	0.000	03:14	
Air Blank	0.000	03:15	
Air Blank	0.000	03:17	
Subject Sample	# 2	03:18	Subject Test Refused
Breath Volume	0.000	03:18	
Air Blank	0.000	03:19	
Control Test	0.080	03:19	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:20	

---

Data Download Date/Time: 12/18/2020 18:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANNON, MATTHEW, F  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HANNON, MATTHEW  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: CAMPIONE

DATE/TIME OF TEST: 11/22/2020 02:48:18

---

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:38	
Control Test	0.080	02:38	
Air Blank	0.000	02:39	
Subject Sample	# 1	02:42	No Sample Provided
Breath Volume	0.000	02:42	
Air Blank	0.000	02:42	
Air Blank	0.000	02:44	
Subject Sample	# 2	02:46	Subject Test Refused
Breath Volume	0.000	02:46	
Air Blank	0.000	02:47	
Control Test	0.080	02:47	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:48	

---

Data Download Date/Time: 12/18/2020 18:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRINIDAD,RICKY,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GALLOWAY,GRANT  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: KOEHN

DATE/TIME OF TEST: 11/22/2020 04:14:05

---

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.078	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.167	04:09
Breath Volume	3.484	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:11
Subject Sample # 2	0.170	04:12
Breath Volume	2.695	04:12
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

---

Data Download Date/Time: 12/18/2020 18:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRINIDAD,RICKY,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: TRINIDAD,RICKY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: HALSE

DATE/TIME OF TEST: 11/22/2020 04:22:46

---

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.000	04:18
Breath Volume	1.355	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.000	04:21
Breath Volume	1.730	04:21
Air Blank	0.000	04:21
Control Test	0.080	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

---

Data Download Date/Time: 12/18/2020 18:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NIXON, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: NELSON, WILLIAM  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 08:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:23

SUBJECT'S LAST NAME: THEISEN

DATE/TIME OF TEST: 11/22/2020 09:50:45

---

Diagnostics Check	OK	09:44
Air Blank	0.000	09:44
Control Test	0.079	09:44
Air Blank	0.000	09:45
Subject Sample # 1	0.091	09:46
Breath Volume	6.578	09:46
Air Blank	0.000	09:46
Air Blank	0.000	09:48
Subject Sample # 2	0.094	09:49
Breath Volume	6.445	09:49
Air Blank	0.000	09:49
Control Test	0.079	09:50
Air Blank	0.000	09:50
Diagnostics Check	OK	09:50

---

Data Download Date/Time: 12/18/2020 18:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NIXON, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: DION, KEVIN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 09:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:38

SUBJECT'S LAST NAME: MARTI

DATE/TIME OF TEST: 11/22/2020 11:05:15

---

Diagnostics Check	OK	10:58
Air Blank	0.000	10:59
Control Test	0.080	10:59
Air Blank	0.000	11:00
Subject Sample # 1	0.000	11:00
Breath Volume	2.339	11:00
Air Blank	0.000	11:01
Air Blank	0.000	11:03
Subject Sample # 2	0.000	11:03
Breath Volume	2.429	11:03
Air Blank	0.000	11:04
Control Test	0.080	11:04
Air Blank	0.000	11:05
Diagnostics Check	OK	11:05

---

Data Download Date/Time: 12/18/2020 18:57



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GRIFFIN,RAEFORD  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 11/24/2020 20:48:53

---

Diagnostics Check		OK	20:42
Air Blank		0.000	20:42
Control Test		0.081	20:43
Air Blank		0.000	20:43
Subject Sample	# 1	0.326	20:44
Breath Volume		2.621	20:44
Air Blank		0.000	20:44
Air Blank		0.000	20:46
Subject Sample	# 2	0.329	20:47
Breath Volume		1.984	20:47
Air Blank		0.000	20:47
Control Test		0.080	20:48
Air Blank		0.000	20:48
Diagnostics Check		OK	20:48

---

Data Download Date/Time: 12/18/2020 18:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANNON, MATTHEW, F  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: LEE, MATTHEW  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: BARBUTO

DATE/TIME OF TEST: 11/25/2020 21:28:26

---

Diagnostics Check	OK	21:20
Air Blank	0.000	21:21
Control Test	0.080	21:21
Air Blank	0.000	21:22
Subject Sample # 1	0.128	21:22
Breath Volume	2.105	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.125	21:26
Breath Volume	2.050	21:26
Air Blank	0.000	21:27
Control Test	0.080	21:27
Air Blank	0.000	21:28
Diagnostics Check	OK	21:28

---

Data Download Date/Time: 12/18/2020 18:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANNON, MATTHEW, F  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HANNON, MATTHEW  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: CHRISTIANSEN

DATE/TIME OF TEST: 11/26/2020 01:43:42

---

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:35	
Control Test	0.080	01:36	
Air Blank	0.000	01:36	
Subject Sample	# 1	01:39	No Sample Provided
Breath Volume	0.000	01:39	
Air Blank	0.000	01:40	
Air Blank	0.000	01:42	
Subject Sample	# 2	01:42	Subject Test Refused
Breath Volume	0.000	01:42	
Air Blank	0.000	01:42	
Control Test	0.080	01:43	
Air Blank	0.000	01:43	
Diagnostics Check	OK	01:43	

---

Data Download Date/Time: 12/18/2020 18:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOTO,JOHN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HUDSON,EDWARD  
ARRESTING OFFICER AGENCY: Zephyrhills PD  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: GYNGARD

DATE/TIME OF TEST: 11/26/2020 04:15:40

---

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:09	
Control Test	0.079	04:09	
Air Blank	0.000	04:10	
Subject Sample # 1	0.208	04:11	
Breath Volume	3.207	04:11	
Air Blank	0.000	04:12	
Air Blank	0.000	04:13	
Subject Sample # 2	0.205	04:14	
Breath Volume	2.984	04:14	
Air Blank		04:15	RFI Detect
Air Blank	0.000	04:15	

---

Data Download Date/Time: 12/18/2020 18:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOTO,JOHN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HUDSON,EDWARD  
ARRESTING OFFICER AGENCY: Zephyrhills PD  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: GYNGARD

DATE/TIME OF TEST: 11/26/2020 04:38:37

---

Diagnostics Check		OK	04:31
Air Blank		0.000	04:31
Control Test		0.080	04:32
Air Blank		0.000	04:32
Subject Sample	# 1	0.207	04:33
Breath Volume		2.992	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.190	04:37
Breath Volume		2.058	04:37
Air Blank		0.000	04:37
Control Test		0.078	04:38
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

---

Data Download Date/Time: 12/18/2020 18:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NIXON, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: COFFEE, JASON  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 11/26/2020 14:09:25

---

Diagnostics Check	OK	14:06	
Air Blank	0.000	14:06	
Control Test	0.079	14:06	
Air Blank	0.000	14:07	
Subject Sample	# 1	14:07	Slope Not Met
Breath Volume	1.000	14:07	
Air Blank	0.000	14:08	
Control Test	0.080	14:08	
Air Blank	0.000	14:09	
Diagnostics Check	OK	14:09	

---

Data Download Date/Time: 12/18/2020 18:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NIXON, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: COFFEE, JASON  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 11/26/2020 14:23:16

---

Diagnostics Check		OK	14:12	
Air Blank		0.000	14:12	
Control Test		0.080	14:12	
Air Blank		0.000	14:13	
Subject Sample	# 1	0.343	14:14	No .020 agreement
Breath Volume		1.507	14:14	
Air Blank		0.000	14:15	
Air Blank		0.000	14:17	
Subject Sample	# 2	0.319	14:18	No .020 agreement
Breath Volume		1.468	14:18	
Air Blank		0.000	14:19	
Air Blank		0.000	14:21	
Subject Sample	# 3	0.297	14:21	No .020 agreement
Breath Volume		1.238	14:21	
Air Blank		0.000	14:22	
Control Test		0.077	14:22	
Air Blank		0.000	14:23	
Diagnostics Check		OK	14:23	

---

Data Download Date/Time: 12/18/2020 18:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRINIDAD,RICKY,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: TRINIDAD,RICKY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: GOMEZ GOMEZ

DATE/TIME OF TEST: 11/27/2020 01:34:52

---

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:30
Subject Sample # 1	0.122	01:30
Breath Volume	2.781	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:32
Subject Sample # 2	0.124	01:33
Breath Volume	2.523	01:33
Air Blank	0.000	01:33
Control Test	0.078	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

---

Data Download Date/Time: 12/18/2020 18:57



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOTO,JOHN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: NOTO,JOHN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: JIMENEZ GOMEZ

DATE/TIME OF TEST: 11/27/2020 01:43:35

---

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.078	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.107	01:39
Breath Volume	2.820	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.101	01:41
Breath Volume	2.816	01:41
Air Blank	0.000	01:42
Control Test	0.078	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

---

Data Download Date/Time: 12/18/2020 18:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANNON, MATTHEW, F  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: ALVARADO, ANGEL  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: GRISANTI

DATE/TIME OF TEST: 11/28/2020 01:44:20

---

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.081	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.063	01:39
Breath Volume	3.398	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.063	01:42
Breath Volume	3.898	01:42
Air Blank	0.000	01:43
Control Test	0.081	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

---

Data Download Date/Time: 12/18/2020 18:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRIFFIN,SUMMER,N  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GRIFFIN,SUMMER  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: RIGGENBACH

DATE/TIME OF TEST: 11/28/2020 02:59:32

---

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.081	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.092	02:55
Breath Volume	3.042	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.092	02:58
Breath Volume	3.578	02:58
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

---

Data Download Date/Time: 12/18/2020 18:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: ROBERTS,JORDAN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BEARDSLEY

DATE/TIME OF TEST: 11/29/2020 02:58:50

---

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.081	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.222	02:54
Breath Volume	2.128	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.222	02:57
Breath Volume	2.117	02:57
Air Blank	0.000	02:57
Control Test	0.079	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

---

Data Download Date/Time: 12/18/2020 18:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GRIFFIN,RAEFORD  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 11/29/2020 04:51:37

---

Diagnostics Check	OK	04:44
Air Blank	0.000	04:45
Control Test	0.079	04:45
Air Blank	0.000	04:46
Subject Sample # 1	0.161	04:46
Breath Volume	2.476	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.149	04:50
Breath Volume	4.367	04:50
Air Blank	0.000	04:50
Control Test	0.080	04:51
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

---

Data Download Date/Time: 12/18/2020 18:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RITTENOUR,COREY,T  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: BARACK,SEAN  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: SANDKAMP

DATE/TIME OF TEST: 11/10/2020 04:59:22

---

Diagnostics Check		OK	04:52
Air Blank		0.000	04:52
Control Test		0.079	04:53
Air Blank		0.000	04:53
Subject Sample	# 1	0.202	04:54
Breath Volume		2.234	04:54
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.219	04:57
Breath Volume		3.128	04:57
Air Blank		0.000	04:58
Control Test		0.078	04:58
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

---

Data Download Date/Time: 11/16/2020 15:20