

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SIEBER,ROBERT,M

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: ROBINSON,KYLE

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 12/01/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 05620080A2

DRY GAS EXPIRY DATE: 04/05/2022

VIOLATION CODE: DUI

ARREST TIME: 12:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 12:48

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 12/01/2020 13:36:31

Diagnostics Check	OK	13:29
Air Blank	0.000	13:29
Control Test	0.078	13:30
Air Blank	0.000	13:30
Subject Sample # 1	0.000	13:32
Breath Volume	3.355	13:32
Air Blank	0.000	13:32
Air Blank	0.000	13:34
Subject Sample # 2	0.000	13:35
Breath Volume	2.433	13:35
Air Blank	0.000	13:35
Control Test	0.078	13:35
Air Blank	0.000	13:36
Diagnostics Check	OK	13:36

Data Download Date/Time: 12/11/2020 13:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ISHALL,JALEN,L

BREATH TEST OPERATOR AGENCY: Sumter County SO

ARRESTING OFFICER: PARRISH,RICHARD

ARRESTING OFFICER AGENCY: Wildwood PD

LAST AGENCY INSPECTION: 12/02/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 181632

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 20:35

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: SKUTAK

DATE/TIME OF TEST: 12/04/2020 21:52:11

Diagnostics Check	OK	21:45
Air Blank	0.000	21:45
Control Test	0.080	21:46
Air Blank	0.000	21:46
Subject Sample # 1	0.203	21:47
Breath Volume	1.429	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:49
Subject Sample # 2	0.203	21:50
Breath Volume	1.785	21:50
Air Blank	0.000	21:51
Control Test	0.078	21:51
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 01/27/2021 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M

BREATH TEST OPERATOR AGENCY: Sumter County SO

ARRESTING OFFICER: GARTNER,ARTHOR

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 12/02/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 181632

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 00:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: SAI BHARATH

DATE/TIME OF TEST: 12/07/2020 01:57:33

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:55	
Control Test	0.081	01:55	
Air Blank	0.000	01:55	
Subject Sample	# 1	01:56	Subject Test Refused
Breath Volume	0.000	01:56	
Air Blank	0.000	01:56	
Control Test	0.081	01:56	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

Data Download Date/Time: 01/27/2021 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NOTTINGHAM,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:49

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 12/07/2020 15:15:52

Diagnostics Check	OK	15:09
Air Blank	0.000	15:09
Control Test	0.080	15:10
Air Blank	0.000	15:10
Subject Sample # 1	0.032	15:11
Breath Volume	1.312	15:11
Air Blank	0.000	15:11
Air Blank	0.000	15:13
Subject Sample # 2	0.031	15:14
Breath Volume	1.792	15:14
Air Blank	0.000	15:14
Control Test	0.081	15:15
Air Blank	0.000	15:15
Diagnostics Check	OK	15:15

Data Download Date/Time: 01/27/2021 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: OTERO,H
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:47

SUBJECT'S LAST NAME: HOOVER

DATE/TIME OF TEST: 12/07/2020 17:15:26

Diagnostics Check		OK	17:07
Air Blank		0.000	17:08
Control Test		0.081	17:08
Air Blank		0.000	17:09
Subject Sample	# 1	0.104	17:10
Breath Volume		1.867	17:10
Air Blank		0.000	17:11
Air Blank		0.000	17:12
Subject Sample	# 2	0.095	17:13
Breath Volume		1.382	17:13
Air Blank		0.000	17:14
Control Test		0.080	17:14
Air Blank		0.000	17:15
Diagnostics Check		OK	17:15

Data Download Date/Time: 01/27/2021 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,TRAVIS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: TAN,RICHARD
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: VILLANUEVA

DATE/TIME OF TEST: 12/15/2020 22:13:19

Diagnostics Check	OK	22:10	
Air Blank	0.000	22:10	
Control Test	0.080	22:11	
Air Blank	0.000	22:11	
Subject Sample	# 1	22:11	Subject Test Refused
Breath Volume	0.000	22:11	
Air Blank	0.000	22:12	
Control Test	0.080	22:12	
Air Blank	0.000	22:13	
Diagnostics Check	OK	22:13	

Data Download Date/Time: 01/27/2021 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAVIS, D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: WILSON, DAKOTA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: GARISTO

DATE/TIME OF TEST: 12/19/2020 05:34:52

Diagnostics Check	OK	05:26
Air Blank	0.000	05:26
Control Test	0.082	05:26
Air Blank	0.000	05:27
Subject Sample # 1	0.161	05:29
Breath Volume	1.300	05:29
Air Blank	0.000	05:29
Air Blank	0.000	05:31
Subject Sample # 2	0.167	05:33
Breath Volume	1.632	05:33
Air Blank	0.000	05:33
Control Test	0.080	05:34
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

Data Download Date/Time: 01/27/2021 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRACKETT,ALAN,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MCCONNELL,AUSTIN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:48

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/24/2020 14:15:24

Diagnostics Check	OK	14:09
Air Blank	0.000	14:09
Control Test	0.080	14:09
Air Blank	0.000	14:10
Subject Sample # 1	0.000	14:10
Breath Volume	2.394	14:10
Air Blank	0.000	14:11
Air Blank	0.000	14:13
Subject Sample # 2	0.000	14:14
Breath Volume	2.773	14:14
Air Blank	0.000	14:14
Control Test	0.080	14:14
Air Blank	0.000	14:15
Diagnostics Check	OK	14:15

Data Download Date/Time: 01/27/2021 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ISHALL,JALEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: TAN,RICHARD
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: CHANCELLOR

DATE/TIME OF TEST: 12/30/2020 00:33:47

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:31	
Control Test	0.081	00:31	
Air Blank	0.000	00:31	
Subject Sample	# 1	00:32	Subject Test Refused
Breath Volume	0.000	00:32	
Air Blank	0.000	00:32	
Control Test	0.081	00:33	
Air Blank	0.000	00:33	
Diagnostics Check	OK	00:33	

Data Download Date/Time: 01/27/2021 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLON,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: DILLON,JUSTIN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 12/05/2020 06:06:20

Diagnostics Check	OK	05:54	
Air Blank	0.000	05:55	
Control Test	0.080	05:55	
Air Blank	0.000	05:56	
Subject Sample	# 1	05:59	No Sample Provided
Breath Volume	0.000	05:59	
Air Blank	0.000	05:59	
Air Blank	0.000	06:01	
Subject Sample	# 2	06:04	No Sample Provided
Breath Volume	0.000	06:04	
Air Blank	0.000	06:05	
Control Test	0.082	06:05	
Air Blank	0.000	06:06	
Diagnostics Check	OK	06:06	

Data Download Date/Time: 12/15/2020 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTSCHER,REBECCA,L
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: BUTSCHER,REBECCA
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: HARTMAN

DATE/TIME OF TEST: 12/05/2020 22:37:51

Diagnostics Check	OK	22:34	
Air Blank	0.000	22:35	
Control Test	0.081	22:35	
Air Blank	0.000	22:36	
Subject Sample	# 1	22:36	Subject Test Refused
Breath Volume	0.000	22:36	
Air Blank	0.000	22:36	
Control Test	0.081	22:37	
Air Blank	0.000	22:37	
Diagnostics Check	OK	22:37	

Data Download Date/Time: 12/15/2020 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURINGTON,PATRICK,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: CREWS,MICHAEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: COELHO

DATE/TIME OF TEST: 12/07/2020 00:39:46

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:28	
Control Test	0.079	00:29	
Air Blank	0.000	00:29	
Subject Sample	# 1 0.235	00:32	Volume Not Met
Breath Volume	0.066	00:32	
Air Blank	0.000	00:33	
Air Blank	0.000	00:35	
Subject Sample	# 2	00:38	Subject Test Refused
Breath Volume	0.000	00:38	
Air Blank	0.000	00:38	
Control Test	0.078	00:39	
Air Blank	0.000	00:39	
Diagnostics Check	OK	00:39	

Data Download Date/Time: 12/15/2020 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,BENJAMIN,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MYERS,BENJAMIN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: CHO

DATE/TIME OF TEST: 12/20/2020 01:09:35

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:07	
Control Test	0.079	01:07	
Air Blank	0.000	01:07	
Subject Sample	# 1	01:08	Subject Test Refused
Breath Volume	0.000	01:08	
Air Blank	0.000	01:08	
Control Test	0.080	01:08	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 01/20/2021 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLON,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MILLER,HAYLEY
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 12/20/2020 18:55:25

Diagnostics Check	OK	18:48
Air Blank	0.000	18:49
Control Test	0.080	18:49
Air Blank	0.000	18:50
Subject Sample # 1	0.000	18:51
Breath Volume	2.324	18:51
Air Blank	0.000	18:51
Air Blank	0.000	18:53
Subject Sample # 2	0.000	18:54
Breath Volume	2.160	18:54
Air Blank	0.000	18:54
Control Test	0.079	18:54
Air Blank	0.000	18:55
Diagnostics Check	OK	18:55

Data Download Date/Time: 01/20/2021 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURINGTON,PATRICK,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PURINGTON,PATRICK
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: YOCHIM

DATE/TIME OF TEST: 12/24/2020 00:51:44

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:49	
Control Test	0.080	00:49	
Air Blank	0.000	00:50	
Subject Sample	# 1	00:50	Subject Test Refused
Breath Volume	0.000	00:50	
Air Blank	0.000	00:50	
Control Test	0.081	00:51	
Air Blank	0.000	00:51	
Diagnostics Check	OK	00:51	

Data Download Date/Time: 01/20/2021 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURINGTON,PATRICK,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MCVEIGH,TREVOR
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: DELVA

DATE/TIME OF TEST: 12/24/2020 02:06:21

Diagnostics Check		OK	01:50	
Air Blank		0.000	01:51	
Control Test		0.079	01:51	
Air Blank		0.000	01:52	
Subject Sample	# 1	0.278	01:55	Volume Not Met
Breath Volume		0.316	01:55	
Air Blank		0.000	01:56	
Air Blank		0.000	01:57	
Subject Sample	# 2	0.307	01:58	
Breath Volume		1.675	01:58	
Air Blank		0.000	01:59	
Air Blank		0.000	02:01	
Subject Sample	# 3	0.262	02:04	Volume Not Met
Breath Volume		0.746	02:04	
Air Blank		0.000	02:05	
Control Test		0.073	02:05	Control Outside Tolerance
Air Blank		0.000	02:06	

Data Download Date/Time: 01/20/2021 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARQUIS,NORMAND,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MARQUIS,NORMAND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 12/04/2020 23:26:16

Diagnostics Check	OK	23:19
Air Blank	0.000	23:20
Control Test	0.080	23:20
Air Blank	0.000	23:21
Subject Sample # 1	0.000	23:21
Breath Volume	2.871	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.000	23:24
Breath Volume	3.230	23:24
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 12/29/2020 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,RUSSELL,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: COOK,RUSSELL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: KRAGH

DATE/TIME OF TEST: 12/05/2020 00:19:29

Diagnostics Check		OK	00:09
Air Blank		0.000	00:10
Control Test		0.079	00:10
Air Blank		0.000	00:11
Subject Sample	# 1	0.148	00:11
Breath Volume		2.000	00:11
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.126	00:14
Breath Volume		1.246	00:14
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 3	0.136	00:17
Breath Volume		2.226	00:17
Air Blank		0.000	00:18
Control Test		0.078	00:18
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 12/29/2020 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STALEY,KRISTOPHER,H
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HANKINS,MEGAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: BOWEN

DATE/TIME OF TEST: 12/06/2020 06:07:43

Diagnostics Check	OK	06:01
Air Blank	0.000	06:01
Control Test	0.081	06:02
Air Blank	0.000	06:02
Subject Sample # 1	0.171	06:03
Breath Volume	2.019	06:03
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample # 2	0.173	06:06
Breath Volume	1.660	06:06
Air Blank	0.000	06:06
Control Test	0.080	06:07
Air Blank	0.000	06:07
Diagnostics Check	OK	06:07

Data Download Date/Time: 12/29/2020 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STALEY, KRISTOPHER, H
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BARNHILL, WESLEY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:36

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 12/06/2020 07:11:57

Diagnostics Check	OK	07:06
Air Blank	0.000	07:06
Control Test	0.080	07:06
Air Blank	0.000	07:07
Subject Sample # 1	0.000	07:07
Breath Volume	3.019	07:07
Air Blank	0.000	07:08
Air Blank	0.000	07:10
Subject Sample # 2	0.000	07:10
Breath Volume	3.187	07:10
Air Blank	0.000	07:11
Control Test	0.081	07:11
Air Blank	0.000	07:11
Diagnostics Check	OK	07:11

Data Download Date/Time: 12/29/2020 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LOPEZ,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 16:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:29

SUBJECT'S LAST NAME: LEMON

DATE/TIME OF TEST: 12/11/2020 16:57:39

Diagnostics Check	OK	16:50
Air Blank	0.000	16:51
Control Test	0.081	16:51
Air Blank	0.000	16:52
Subject Sample # 1	0.000	16:52
Breath Volume	1.382	16:52
Air Blank	0.000	16:53
Air Blank	0.000	16:55
Subject Sample # 2	0.000	16:56
Breath Volume	2.757	16:56
Air Blank	0.000	16:56
Control Test	0.080	16:57
Air Blank	0.000	16:57
Diagnostics Check	OK	16:57

Data Download Date/Time: 12/29/2020 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: SMILEY

DATE/TIME OF TEST: 12/12/2020 00:51:34

Diagnostics Check		OK	00:39	
Air Blank		0.000	00:40	
Control Test		0.080	00:40	
Air Blank		0.000	00:41	
Subject Sample	# 1	0.153	00:44	Volume Not Met
Breath Volume		0.050	00:44	
Air Blank		0.000	00:45	
Air Blank		0.000	00:46	
Subject Sample	# 2	0.167	00:49	Volume Not Met
Breath Volume		0.082	00:49	
Air Blank		0.000	00:50	
Control Test		0.080	00:50	
Air Blank		0.000	00:51	
Diagnostics Check		OK	00:51	

Data Download Date/Time: 12/29/2020 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALENTINE, RONALD, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VALENTINE, RONALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 08:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: JEAMES

DATE/TIME OF TEST: 12/12/2020 09:45:03

Diagnostics Check	OK	09:42	
Air Blank	0.000	09:42	
Control Test	0.079	09:42	
Air Blank	0.000	09:43	
Subject Sample	# 1	09:43	Subject Test Refused
Breath Volume	0.000	09:43	
Air Blank	0.000	09:44	
Control Test	0.079	09:44	
Air Blank	0.000	09:44	
Diagnostics Check	OK	09:45	

Data Download Date/Time: 12/29/2020 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,RUSSELL,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: COOK,RUSSELL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 12/14/2020 00:54:27

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:52	
Control Test	0.080	00:52	
Air Blank	0.000	00:52	
Subject Sample	# 1	00:53	Subject Test Refused
Breath Volume	0.000	00:53	
Air Blank	0.000	00:53	
Control Test	0.080	00:53	
Air Blank	0.000	00:54	
Diagnostics Check	OK	00:54	

Data Download Date/Time: 12/29/2020 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDAWAY, BRANDON, R
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: ELLRICK, JOSHUA
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: PUCKETT

DATE/TIME OF TEST: 12/15/2020 18:55:26

Diagnostics Check	OK	18:47
Air Blank	0.000	18:47
Control Test	0.080	18:47
Air Blank	0.000	18:48
Subject Sample # 1	0.203	18:50
Breath Volume	1.773	18:50
Air Blank	0.000	18:51
Air Blank	0.000	18:53
Subject Sample # 2	0.211	18:53
Breath Volume	2.898	18:53
Air Blank	0.000	18:54
Control Test	0.080	18:54
Air Blank	0.000	18:55
Diagnostics Check	OK	18:55

Data Download Date/Time: 12/29/2020 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDAWAY, BRANDON, R
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: HARDAWAY, BRANDON
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 12/20/2020 03:51:21

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.080	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.208	03:46
Breath Volume		1.753	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.216	03:49
Breath Volume		1.234	03:49
Air Blank		0.000	03:50
Control Test		0.080	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 12/29/2020 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDAWAY, BRANDON, R
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: ELLRICK, JOSHUA
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: EADY

DATE/TIME OF TEST: 12/20/2020 19:09:19

Diagnostics Check	OK	19:02
Air Blank	0.000	19:03
Control Test	0.080	19:03
Air Blank	0.000	19:04
Subject Sample # 1	0.000	19:05
Breath Volume	2.785	19:05
Air Blank	0.000	19:05
Air Blank	0.000	19:07
Subject Sample # 2	0.000	19:07
Breath Volume	2.324	19:07
Air Blank	0.000	19:08
Control Test	0.080	19:08
Air Blank	0.000	19:09
Diagnostics Check	OK	19:09

Data Download Date/Time: 12/29/2020 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,RUSSELL,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VALENTINE,RONALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 12:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 12/25/2020 13:38:00

Diagnostics Check	OK	13:34	
Air Blank	0.000	13:35	
Control Test	0.081	13:35	
Air Blank	0.000	13:36	
Subject Sample	# 1	13:36	Subject Test Refused
Breath Volume	0.000	13:36	
Air Blank	0.000	13:37	
Control Test	0.081	13:37	
Air Blank	0.000	13:37	
Diagnostics Check	OK	13:38	

Data Download Date/Time: 12/29/2020 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,RUSSELL,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: COOK,RUSSELL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: WATTS

DATE/TIME OF TEST: 12/27/2020 23:17:38

Diagnostics Check	OK	23:11
Air Blank	0.000	23:11
Control Test	0.081	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.226	23:13
Breath Volume	1.589	23:13
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.230	23:15
Breath Volume	1.476	23:15
Air Blank	0.000	23:16
Control Test	0.080	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 12/29/2020 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON, RAY, C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER, MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 12:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:34

SUBJECT'S LAST NAME: MONTFORD

DATE/TIME OF TEST: 12/31/2020 13:01:11

Diagnostics Check	OK	12:54
Air Blank	0.000	12:54
Control Test	0.080	12:54
Air Blank	0.000	12:55
Subject Sample # 1	0.000	12:56
Breath Volume	5.703	12:56
Air Blank	0.000	12:56
Air Blank	0.000	12:58
Subject Sample # 2	0.000	12:59
Breath Volume	3.156	12:59
Air Blank	0.000	13:00
Control Test	0.081	13:00
Air Blank	0.000	13:01
Diagnostics Check	OK	13:01

Data Download Date/Time: 01/28/2021 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD,DERRICK,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ALEXANDER,C
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 06:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: GIPSON

DATE/TIME OF TEST: 12/10/2020 20:15:32

Diagnostics Check	OK	20:12	
Air Blank	0.000	20:13	
Control Test	0.079	20:13	
Air Blank	0.000	20:13	
Subject Sample	# 1	20:14	Subject Test Refused
Breath Volume	0.000	20:14	
Air Blank	0.000	20:14	
Control Test	0.080	20:14	
Air Blank	0.000	20:15	
Diagnostics Check	OK	20:15	

Data Download Date/Time: 01/11/2021 17:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNETT,JAMES,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: BARNETT,JAMES
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: FELGER

DATE/TIME OF TEST: 12/12/2020 22:57:09

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:43	
Control Test	0.080	22:43	
Air Blank	0.000	22:44	
Subject Sample	# 1 0.156	22:45	
Breath Volume	2.156	22:45	
Air Blank	0.000	22:46	
Air Blank	0.000	22:48	
Subject Sample	# 2 0.104	22:51	Volume Not Met
Breath Volume	0.351	22:51	
Air Blank	0.000	22:51	
Air Blank	0.000	22:53	
Subject Sample	# 3 0.144	22:55	
Breath Volume	1.593	22:55	
Air Blank	0.000	22:56	
Control Test	0.080	22:56	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

Data Download Date/Time: 01/11/2021 17:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLANDA, RICARDO, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HOLANDA, RICARDO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 12/24/2020 19:23:28

Diagnostics Check	OK	19:16
Air Blank	0.000	19:16
Control Test	0.082	19:16
Air Blank	0.000	19:17
Subject Sample # 1	0.195	19:19
Breath Volume	3.328	19:19
Air Blank	0.000	19:19
Air Blank	0.000	19:21
Subject Sample # 2	0.195	19:22
Breath Volume	3.179	19:22
Air Blank	0.000	19:22
Control Test	0.080	19:22
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

Data Download Date/Time: 01/11/2021 17:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD,DERRICK,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CULLIVER,TAVARIS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 12/26/2020 04:45:31

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.080	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.195	04:41
Breath Volume	2.792	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.197	04:43
Breath Volume	3.148	04:43
Air Blank	0.000	04:44
Control Test	0.079	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 01/11/2021 17:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNETT,JAMES,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: BARNETT,JAMES
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/27/2020 01:33:14

Diagnostics Check		OK	01:26
Air Blank		0.000	01:26
Control Test		0.081	01:26
Air Blank		0.000	01:27
Subject Sample	# 1	0.190	01:28
Breath Volume		2.597	01:28
Air Blank		0.000	01:29
Air Blank		0.000	01:30
Subject Sample	# 2	0.197	01:31
Breath Volume		2.460	01:31
Air Blank		0.000	01:32
Control Test		0.079	01:32
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 01/11/2021 17:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,C,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CULLIVER,TAVARIS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MCCASTLER

DATE/TIME OF TEST: 12/27/2020 23:57:24

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.081	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.163	23:52
Breath Volume	1.207	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.163	23:55
Breath Volume	1.812	23:55
Air Blank	0.000	23:56
Control Test	0.080	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 01/11/2021 17:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORMLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GORMLEY, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/31/2020 19:58:39

Diagnostics Check	OK	19:51
Air Blank	0.000	19:51
Control Test	0.081	19:51
Air Blank	0.000	19:52
Subject Sample # 1	0.000	19:53
Breath Volume	1.476	19:53
Air Blank	0.000	19:53
Air Blank	0.000	19:55
Subject Sample # 2	0.000	19:57
Breath Volume	1.750	19:57
Air Blank	0.000	19:57
Control Test	0.082	19:58
Air Blank	0.000	19:58
Diagnostics Check	OK	19:58

Data Download Date/Time: 01/11/2021 17:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEITE,STEVEN,G
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HYLER,THOMAS
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 12/04/2020 03:26:28

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.084	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.175	03:22
Breath Volume		1.429	03:22
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.171	03:24
Breath Volume		1.406	03:24
Air Blank		0.000	03:25
Control Test		0.076	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 12/09/2020 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD,DERRICK,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN,DELRIKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 10:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:04

SUBJECT'S LAST NAME: DOVER

DATE/TIME OF TEST: 12/05/2020 12:36:33

Diagnostics Check		OK	12:24	
Air Blank		0.000	12:25	
Control Test		0.081	12:25	
Air Blank		0.000	12:25	
Subject Sample	# 1	0.213	12:29	Volume Not Met
Breath Volume		0.753	12:29	
Air Blank		0.000	12:29	
Air Blank		0.000	12:31	
Subject Sample	# 2	0.228	12:34	Volume Not Met
Breath Volume		0.039	12:34	
Air Blank		0.000	12:35	
Control Test		0.077	12:35	
Air Blank		0.000	12:36	
Diagnostics Check		OK	12:36	

Data Download Date/Time: 12/09/2020 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD,DERRICK,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN,DELRIKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 10:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:47

SUBJECT'S LAST NAME: DOVER

DATE/TIME OF TEST: 12/05/2020 13:19:12

Diagnostics Check	OK	13:07	
Air Blank	0.000	13:07	
Control Test	0.079	13:08	
Air Blank	0.000	13:08	
Subject Sample	# 1 0.231	13:11	Volume Not Met
Breath Volume	0.167	13:11	
Air Blank	0.000	13:12	
Air Blank	0.000	13:14	
Subject Sample	# 2 0.221	13:17	Volume Not Met
Breath Volume	0.660	13:17	
Air Blank	0.000	13:18	
Control Test	0.078	13:18	
Air Blank	0.000	13:19	
Diagnostics Check	OK	13:19	

Data Download Date/Time: 12/09/2020 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNCH,CHAD,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LYNCH,CHAD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: BODIFORD

DATE/TIME OF TEST: 12/13/2020 00:29:04

Diagnostics Check		OK	00:22
Air Blank		0.000	00:22
Control Test		0.082	00:22
Air Blank		0.000	00:23
Subject Sample	# 1	0.059	00:24
Breath Volume		1.949	00:24
Air Blank		0.000	00:25
Air Blank		0.000	00:26
Subject Sample	# 2	0.055	00:27
Breath Volume		1.285	00:27
Air Blank		0.000	00:28
Control Test		0.082	00:28
Air Blank		0.000	00:28
Diagnostics Check		OK	00:29

Data Download Date/Time: 01/11/2021 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLER,DANIEL,V
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: BAISE,KIRSTYN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 12:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:10

SUBJECT'S LAST NAME: KIMMONS

DATE/TIME OF TEST: 12/14/2020 13:01:21

Diagnostics Check	OK	12:55
Air Blank	0.000	12:55
Control Test	0.083	12:56
Air Blank	0.000	12:56
Subject Sample # 1	0.000	12:57
Breath Volume	3.425	12:57
Air Blank	0.000	12:57
Air Blank	0.000	12:59
Subject Sample # 2	0.000	12:59
Breath Volume	3.707	12:59
Air Blank	0.000	13:00
Control Test	0.081	13:00
Air Blank	0.000	13:01
Diagnostics Check	OK	13:01

Data Download Date/Time: 01/11/2021 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORELLANA,ADALBERTO,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ORELLANA,ADALBERTO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 12:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:05

SUBJECT'S LAST NAME: BLANCO

DATE/TIME OF TEST: 12/15/2020 12:49:15

Diagnostics Check	OK	12:39
Air Blank	0.000	12:39
Control Test	0.084	12:40
Air Blank	0.000	12:40
Subject Sample # 1	0.000	12:42
Breath Volume	1.609	12:42
Air Blank	0.000	12:43
Air Blank	0.000	12:45
Subject Sample # 2	0.000	12:47
Breath Volume	1.566	12:47
Air Blank	0.000	12:48
Control Test	0.082	12:48
Air Blank	0.000	12:49
Diagnostics Check	OK	12:49

Data Download Date/Time: 01/11/2021 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 16:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:23

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/25/2020 16:48:53

Diagnostics Check	OK	16:43	
Air Blank	0.000	16:43	
Control Test	0.084	16:44	
Air Blank	0.000	16:44	
Subject Sample	# 1 0.250	16:44	
Breath Volume	2.390	16:44	
Air Blank	0.000	16:45	
Air Blank	0.000	16:47	
Subject Sample	# 2 0.251	16:47	
Breath Volume	1.968	16:47	
Air Blank		16:48	RFI Detect
Air Blank	0.000	16:48	

Data Download Date/Time: 01/11/2021 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN PELT, COLTEN, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VAN PELT, COLTEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: BAXLEY

DATE/TIME OF TEST: 12/28/2020 22:25:11

Diagnostics Check	OK	22:18
Air Blank	0.000	22:19
Control Test	0.083	22:19
Air Blank	0.000	22:20
Subject Sample # 1	0.162	22:20
Breath Volume	2.656	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.171	22:23
Breath Volume	2.550	22:23
Air Blank	0.000	22:24
Control Test	0.082	22:24
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 01/11/2021 16:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,BRUCE,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: IVEY,BRUCE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: DAECEY

DATE/TIME OF TEST: 12/10/2020 00:32:54

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:21	
Control Test	0.082	00:22	
Air Blank	0.000	00:22	
Subject Sample	# 1 0.000	00:25	Volume Not Met
Breath Volume	0.097	00:25	
Air Blank	0.000	00:26	
Air Blank	0.000	00:28	
Subject Sample	# 2	00:31	No Sample Provided
Breath Volume	0.000	00:31	
Air Blank	0.000	00:31	
Control Test	0.082	00:32	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 01/25/2021 09:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFITH,DALTON,M
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: MERRITT,JACOB
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 12/05/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 18:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: ENFINGER

DATE/TIME OF TEST: 12/16/2020 19:54:42

Diagnostics Check	OK	19:43	
Air Blank	0.000	19:43	
Control Test	0.081	19:44	
Air Blank	0.000	19:44	
Subject Sample	# 1	19:47	No Sample Provided
Breath Volume	0.000	19:47	
Air Blank	0.000	19:48	
Air Blank	0.000	19:50	
Subject Sample	# 2	19:53	No Sample Provided
Breath Volume	0.000	19:53	
Air Blank	0.000	19:53	
Control Test	0.000	19:54	RFI Detect
Air Blank	0.000	19:54	

Data Download Date/Time: 01/25/2021 09:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFITH,DALTON,M
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: MERRITT,JACOB
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 12/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 18:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: ENFINGER

DATE/TIME OF TEST: 12/16/2020 20:00:46

Diagnostics Check	OK	19:57	
Air Blank	0.000	19:58	
Control Test	0.082	19:58	
Air Blank	0.000	19:59	
Subject Sample	# 1	19:59	Subject Test Refused
Breath Volume	0.000	19:59	
Air Blank	0.000	19:59	
Control Test	0.080	20:00	
Air Blank	0.000	20:00	
Diagnostics Check	OK	20:00	

Data Download Date/Time: 01/25/2021 09:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 12/16/2020 20:23:42

Diagnostics Check		OK	20:09	
Air Blank		0.000	20:10	
Control Test		0.081	20:10	
Air Blank		0.000	20:11	
Subject Sample	# 1	0.177	20:14	Volume Not Met
Breath Volume		0.265	20:14	
Air Blank		0.000	20:15	
Air Blank		0.000	20:16	
Subject Sample	# 2	0.184	20:18	
Breath Volume		2.210	20:18	
Air Blank		0.000	20:19	
Air Blank		0.000	20:21	
Subject Sample	# 3	0.191	20:22	
Breath Volume		1.988	20:22	
Air Blank		0.000	20:22	
Control Test		0.078	20:23	
Air Blank		0.000	20:23	
Diagnostics Check		OK	20:23	

Data Download Date/Time: 01/25/2021 09:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,WILLIAM,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 08:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:57

SUBJECT'S LAST NAME: DOWLING

DATE/TIME OF TEST: 12/18/2020 10:25:15

Diagnostics Check	OK	10:19
Air Blank	0.000	10:19
Control Test	0.080	10:19
Air Blank	0.000	10:20
Subject Sample # 1	0.186	10:20
Breath Volume	2.046	10:20
Air Blank	0.000	10:21
Air Blank	0.000	10:23
Subject Sample # 2	0.184	10:23
Breath Volume	2.226	10:23
Air Blank	0.000	10:24
Control Test	0.081	10:24
Air Blank	0.000	10:25
Diagnostics Check	OK	10:25

Data Download Date/Time: 01/25/2021 09:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLEMAN,AARON,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GOLEMAN,AARON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/12/2020 03:02:52

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:54	
Control Test	0.080	02:55	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:58	No Sample Provided
Breath Volume	0.000	02:58	
Air Blank	0.000	02:59	
Air Blank	0.000	03:01	
Subject Sample	# 2	03:01	Subject Test Refused
Breath Volume	0.000	03:01	
Air Blank	0.000	03:01	
Control Test	0.080	03:02	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 01/04/2021 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,RUSSELL,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: COOK,RUSSELL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 12/12/2020 03:28:04

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.129	03:23
Breath Volume	2.816	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.129	03:26
Breath Volume	1.812	03:26
Air Blank	0.000	03:27
Control Test	0.078	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:28

Data Download Date/Time: 01/04/2021 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JOSHUA,I
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SCHULTZ,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: DUNNE

DATE/TIME OF TEST: 12/13/2020 01:17:38

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.154	01:12
Breath Volume	3.035	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.143	01:16
Breath Volume	3.445	01:16
Air Blank	0.000	01:16
Control Test	0.078	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 01/04/2021 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNDT,DAVID,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BATSON,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: RINEHART

DATE/TIME OF TEST: 12/13/2020 03:04:22

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.123	02:59
Breath Volume	3.808	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.116	03:02
Breath Volume	3.234	03:02
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 01/04/2021 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: JAIMET

DATE/TIME OF TEST: 12/13/2020 20:39:35

Diagnostics Check	OK	20:33
Air Blank	0.000	20:33
Control Test	0.080	20:34
Air Blank	0.000	20:34
Subject Sample # 1	0.153	20:35
Breath Volume	2.433	20:35
Air Blank	0.000	20:35
Air Blank	0.000	20:37
Subject Sample # 2	0.151	20:38
Breath Volume	2.003	20:38
Air Blank	0.000	20:38
Control Test	0.080	20:38
Air Blank	0.000	20:39
Diagnostics Check	OK	20:39

Data Download Date/Time: 01/04/2021 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNDT,DAVID,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BATSON,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: VAN WIE

DATE/TIME OF TEST: 12/19/2020 03:21:06

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:15	
Control Test	0.082	03:16	
Air Blank	0.000	03:16	
Subject Sample	# 1	03:19	Subject Test Refused
Breath Volume	0.000	03:19	
Air Blank	0.000	03:20	
Control Test	0.082	03:20	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 01/04/2021 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENNINGS,JOHN,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: JENNINGS,JOHN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 12/19/2020 22:08:53

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.081	22:03
Air Blank	0.000	22:03
Subject Sample # 1	0.154	22:04
Breath Volume	2.382	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:06
Subject Sample # 2	0.142	22:07
Breath Volume	1.820	22:07
Air Blank	0.000	22:07
Control Test	0.080	22:08
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 01/04/2021 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORENE,AIDAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NORENE,AIDAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: GEYER

DATE/TIME OF TEST: 12/20/2020 01:13:08

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.080	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.155	01:08
Breath Volume	2.554	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:10
Subject Sample # 2	0.157	01:11
Breath Volume	4.984	01:11
Air Blank	0.000	01:12
Control Test	0.080	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 01/04/2021 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVIER,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LEVIER,MATTHEW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 09:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:41

SUBJECT'S LAST NAME: CARLSON

DATE/TIME OF TEST: 12/20/2020 10:07:11

Diagnostics Check	OK	10:01	
Air Blank	0.000	10:02	
Control Test	0.080	10:02	
Air Blank	0.000	10:02	
Subject Sample	# 1	10:05	RFI Detect
Breath Volume	0.000	10:05	
Air Blank	0.000	10:06	
Control Test	0.080	10:06	
Air Blank	0.000	10:07	
Diagnostics Check	OK	10:07	

Data Download Date/Time: 01/04/2021 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNN,JAMISON,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LYNN,JAMISON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 16:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: ACKENHAUSEN

DATE/TIME OF TEST: 12/21/2020 17:33:10

Diagnostics Check	OK	17:21	
Air Blank	0.000	17:22	
Control Test	0.081	17:22	
Air Blank	0.000	17:23	
Subject Sample	# 1	17:26	No Sample Provided
Breath Volume	0.000	17:26	
Air Blank	0.000	17:26	
Air Blank	0.000	17:28	
Subject Sample	# 2	17:31	No Sample Provided
Breath Volume	0.000	17:31	
Air Blank	0.000	17:32	
Control Test	0.082	17:32	
Air Blank	0.000	17:33	
Diagnostics Check	OK	17:33	

Data Download Date/Time: 01/04/2021 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,DANIEL,K
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DAVIS,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: ZALESKI

DATE/TIME OF TEST: 12/23/2020 02:27:21

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:25	
Control Test	0.081	02:25	
Air Blank	0.000	02:25	
Subject Sample	# 1	02:25	Subject Test Refused
Breath Volume	0.000	02:25	
Air Blank	0.000	02:26	
Control Test	0.080	02:26	
Air Blank	0.000	02:27	
Diagnostics Check	OK	02:27	

Data Download Date/Time: 01/04/2021 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURR,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GUERRY,CHAD
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: GALLUP

DATE/TIME OF TEST: 12/05/2020 00:08:10

Diagnostics Check	OK	00:01
Air Blank	0.000	00:02
Control Test	0.079	00:02
Air Blank	0.000	00:03
Subject Sample # 1	0.146	00:03
Breath Volume	4.820	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:05
Subject Sample # 2	0.140	00:06
Breath Volume	3.886	00:06
Air Blank	0.000	00:07
Control Test	0.076	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 12/30/2020 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: PARISI,PARISI
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: WILEY

DATE/TIME OF TEST: 12/10/2020 19:41:50

Diagnostics Check	OK	19:35
Air Blank	0.000	19:35
Control Test	0.079	19:36
Air Blank	0.000	19:36
Subject Sample # 1	0.214	19:37
Breath Volume	1.242	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:39
Subject Sample # 2	0.217	19:40
Breath Volume	1.488	19:40
Air Blank	0.000	19:40
Control Test	0.077	19:41
Air Blank	0.000	19:41
Diagnostics Check	OK	19:41

Data Download Date/Time: 12/30/2020 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DEPUTY,D
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DEPUTY
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 12/12/2020 03:49:47

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:41	
Control Test	0.078	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:45	No Sample Provided
Breath Volume	0.000	03:45	
Air Blank	0.000	03:45	
Air Blank	0.000	03:47	
Subject Sample	# 2	03:48	Subject Test Refused
Breath Volume	0.000	03:48	
Air Blank	0.000	03:48	
Control Test	0.079	03:49	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 12/30/2020 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATES,BRIAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: STEPHENS,MAURICE
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: STINSON

DATE/TIME OF TEST: 12/12/2020 21:41:01

Diagnostics Check	OK	21:36	
Air Blank	0.000	21:36	
Control Test	0.078	21:36	
Air Blank	0.000	21:37	
Subject Sample	# 1	21:39	Subject Test Refused
Breath Volume	0.000	21:39	
Air Blank	0.000	21:40	
Control Test	0.079	21:40	
Air Blank	0.000	21:40	
Diagnostics Check	OK	21:41	

Data Download Date/Time: 12/30/2020 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLIGAN,CONNIE,A
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: GUERRY,CHAD
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: BOONE

DATE/TIME OF TEST: 12/13/2020 03:03:04

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.079	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.229	02:58
Breath Volume	3.195	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:00
Subject Sample # 2	0.227	03:01
Breath Volume	3.000	03:01
Air Blank	0.000	03:02
Control Test	0.076	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:03

Data Download Date/Time: 12/30/2020 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATES,BRIAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENTON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: TEAGUE

DATE/TIME OF TEST: 12/13/2020 04:08:56

Diagnostics Check		OK	03:55	
Air Blank		0.000	03:56	
Control Test		0.078	03:56	
Air Blank		0.000	03:57	
Subject Sample	# 1	0.087	04:00	Volume Not Met
Breath Volume		0.890	04:00	
Air Blank		0.000	04:01	
Air Blank		0.000	04:02	
Subject Sample	# 2	0.096	04:03	
Breath Volume		2.238	04:03	
Air Blank		0.000	04:04	
Air Blank		0.000	04:05	
Subject Sample	# 3	0.087	04:07	
Breath Volume		1.285	04:07	
Air Blank		0.000	04:08	
Control Test		0.075	04:08	
Air Blank		0.000	04:08	
Diagnostics Check		OK	04:08	

Data Download Date/Time: 12/30/2020 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: GURREY,C
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: M200104838370

DATE/TIME OF TEST: 12/19/2020 03:35:52

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:24	
Control Test		0.080	03:24	
Air Blank		0.000	03:25	
Subject Sample	# 1		03:28	No Sample Provided
Breath Volume		0.000	03:28	
Air Blank		0.000	03:28	
Air Blank		0.000	03:30	
Subject Sample	# 2	0.148	03:31	
Breath Volume		1.671	03:31	
Air Blank		0.000	03:32	
Air Blank		0.000	03:33	
Subject Sample	# 3	0.142	03:34	
Breath Volume		2.062	03:34	
Air Blank		0.000	03:34	
Control Test		0.075	03:35	
Air Blank		0.000	03:35	
Diagnostics Check		OK	03:35	

Data Download Date/Time: 12/30/2020 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: GURREY,C
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 12/20/2020 01:18:46

Diagnostics Check	OK	01:12
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.124	01:14
Breath Volume	1.644	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:16
Subject Sample # 2	0.112	01:17
Breath Volume	2.050	01:17
Air Blank	0.000	01:17
Control Test	0.076	01:18
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 12/30/2020 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: GURREY,C
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: BENTON

DATE/TIME OF TEST: 12/20/2020 03:35:07

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.079	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.262	03:30
Breath Volume	3.343	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:32
Subject Sample # 2	0.247	03:33
Breath Volume	2.179	03:33
Air Blank	0.000	03:34
Control Test	0.075	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 12/30/2020 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DEPUTY,D
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DEPUTY
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: DYAL

DATE/TIME OF TEST: 12/20/2020 20:18:30

Diagnostics Check	OK	20:09	
Air Blank	0.000	20:09	
Control Test	0.078	20:10	
Air Blank	0.000	20:10	
Subject Sample	# 1	20:13	No Sample Provided
Breath Volume	0.000	20:13	
Air Blank	0.000	20:14	
Air Blank	0.000	20:16	
Subject Sample	# 2	20:17	Subject Test Refused
Breath Volume	0.000	20:17	
Air Blank	0.000	20:17	
Control Test	0.078	20:17	
Air Blank	0.000	20:18	
Diagnostics Check	OK	20:18	

Data Download Date/Time: 12/30/2020 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: GUERRY,C
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/23/2020 21:23:27

Diagnostics Check	OK	21:17
Air Blank	0.000	21:17
Control Test	0.079	21:17
Air Blank	0.000	21:18
Subject Sample # 1	0.320	21:18
Breath Volume	1.976	21:18
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample # 2	0.311	21:21
Breath Volume	1.921	21:21
Air Blank	0.000	21:22
Control Test	0.076	21:22
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

Data Download Date/Time: 12/30/2020 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: GUERRY,C
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/24/2020 02:25:06

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.079	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.157	02:20
Breath Volume	2.199	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample # 2	0.157	02:23
Breath Volume	2.816	02:23
Air Blank	0.000	02:24
Control Test	0.077	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 12/30/2020 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRISON,JUSTIN,N
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GARRISON,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: HIRES JR

DATE/TIME OF TEST: 12/24/2020 23:10:23

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.079	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.138	23:05
Breath Volume	2.335	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.132	23:08
Breath Volume	4.269	23:08
Air Blank	0.000	23:09
Control Test	0.078	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 12/30/2020 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: GUERRY,C
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: CHILDRESS

DATE/TIME OF TEST: 12/24/2020 23:37:04

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:25	
Control Test	0.079	23:25	
Air Blank	0.000	23:26	
Subject Sample	# 1	23:29	No Sample Provided
Breath Volume	0.000	23:29	
Air Blank	0.000	23:30	
Air Blank	0.000	23:32	
Subject Sample	# 2	0.229	23:32
Breath Volume	4.562	23:32	
Air Blank	0.000	23:33	
Air Blank	0.000	23:34	
Subject Sample	# 3	0.211	23:35
Breath Volume	3.906	23:35	
Air Blank	0.000	23:36	
Control Test	0.075	23:36	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:37	

Data Download Date/Time: 12/30/2020 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DEPUTY,D
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DEPUTY
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: COMBEE

DATE/TIME OF TEST: 12/26/2020 03:48:31

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:40	
Control Test	0.080	03:40	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:44	No Sample Provided
Breath Volume	0.000	03:44	
Air Blank	0.000	03:44	
Air Blank	0.000	03:46	
Subject Sample	# 2	03:47	Subject Test Refused
Breath Volume	0.000	03:47	
Air Blank	0.000	03:47	
Control Test	0.080	03:47	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 12/30/2020 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DEPUTY,D
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DEPUTY
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 12/27/2020 02:13:25

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.086	02:08
Breath Volume	1.812	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.089	02:11
Breath Volume	1.976	02:11
Air Blank	0.000	02:12
Control Test	0.077	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 12/30/2020 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DEPUTY,D
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DEPUTY
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 12/27/2020 04:51:57

Diagnostics Check	OK	04:43	
Air Blank	0.000	04:44	
Control Test	0.080	04:44	
Air Blank	0.000	04:45	
Subject Sample # 1	0.116	04:47	
Breath Volume	3.046	04:47	
Air Blank	0.000	04:48	
Air Blank	0.000	04:49	
Subject Sample # 2	0.098	04:50	
Breath Volume	1.281	04:50	
Air Blank	0.000	04:51	
Control Test	0.074	04:51	Control Outside Tolerance
Air Blank	0.000	04:51	

Data Download Date/Time: 12/30/2020 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: OSTEN

DATE/TIME OF TEST: 12/27/2020 05:18:20

Diagnostics Check	OK	05:11	
Air Blank	0.000	05:11	
Control Test	0.079	05:12	
Air Blank	0.000	05:12	
Subject Sample	# 1 0.221	05:13	
Breath Volume	2.445	05:13	
Air Blank	0.000	05:14	
Air Blank	0.000	05:15	
Subject Sample	# 2 0.214	05:16	
Breath Volume	2.218	05:16	
Air Blank	0.000	05:17	
Control Test	0.074	05:17	Control Outside Tolerance
Air Blank	0.000	05:18	

Data Download Date/Time: 12/30/2020 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: OSTEN

DATE/TIME OF TEST: 12/27/2020 05:30:07

Diagnostics Check	OK	05:23	
Air Blank	0.000	05:23	
Control Test	0.078	05:24	
Air Blank	0.000	05:24	
Subject Sample # 1	0.210	05:25	
Breath Volume	3.296	05:25	
Air Blank	0.000	05:26	
Air Blank	0.000	05:27	
Subject Sample # 2	0.212	05:28	
Breath Volume	2.390	05:28	
Air Blank	0.000	05:29	
Control Test	0.072	05:29	Control Outside Tolerance
Air Blank	0.000	05:30	

Data Download Date/Time: 12/30/2020 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAURES, BRANDYN, K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DRYDEN, JOHN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: PARK

DATE/TIME OF TEST: 12/27/2020 07:17:14

Diagnostics Check	OK	07:10
Air Blank	0.000	07:11
Control Test	0.079	07:11
Air Blank	0.000	07:12
Subject Sample # 1	0.114	07:12
Breath Volume	3.105	07:12
Air Blank	0.000	07:13
Air Blank	0.000	07:15
Subject Sample # 2	0.113	07:15
Breath Volume	3.058	07:15
Air Blank	0.000	07:16
Control Test	0.079	07:16
Air Blank	0.000	07:17
Diagnostics Check	OK	07:17

Data Download Date/Time: 12/30/2020 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: GUERRY,C
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: HAMIL

DATE/TIME OF TEST: 12/28/2020 21:09:16

Diagnostics Check	OK	21:02
Air Blank	0.000	21:03
Control Test	0.079	21:03
Air Blank	0.000	21:04
Subject Sample # 1	0.246	21:04
Breath Volume	1.492	21:04
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.250	21:07
Breath Volume	1.738	21:07
Air Blank	0.000	21:08
Control Test	0.075	21:08
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

Data Download Date/Time: 12/30/2020 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: GUERRY,C
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: LAWSON

DATE/TIME OF TEST: 12/29/2020 02:14:44

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:06	
Control Test	0.079	02:06	
Air Blank	0.000	02:07	
Subject Sample	# 1	02:10	No Sample Provided
Breath Volume	0.000	02:10	
Air Blank	0.000	02:11	
Air Blank	0.000	02:13	
Subject Sample	# 2	02:13	Subject Test Refused
Breath Volume	0.000	02:13	
Air Blank	0.000	02:13	
Control Test	0.080	02:14	
Air Blank	0.000	02:14	
Diagnostics Check	OK	02:14	

Data Download Date/Time: 12/30/2020 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLIGAN,CONNIE,A
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: HENDREITH

DATE/TIME OF TEST: 12/30/2020 22:32:36

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.079	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.236	22:27
Breath Volume	4.550	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.227	22:30
Breath Volume	3.203	22:30
Air Blank	0.000	22:31
Control Test	0.076	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 02/02/2021 15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP TROOP A
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LERNER, MATTHEW, H
BREATH TEST OPERATOR AGENCY: Washington County SO
ARRESTING OFFICER: FOSTER, LOGAN
ARRESTING OFFICER AGENCY: Washington County SO
LAST AGENCY INSPECTION: 12/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 12/15/2020 22:12:40

Diagnostics Check	OK	22:05
Air Blank	0.000	22:06
Control Test	0.079	22:06
Air Blank	0.000	22:07
Subject Sample # 1	0.227	22:07
Breath Volume	1.214	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.224	22:11
Breath Volume	2.984	22:11
Air Blank	0.000	22:11
Control Test	0.078	22:12
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 01/19/2021 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITAKER,JOHNNY,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WHITAKER,JOHNNY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 12:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: PIKE

DATE/TIME OF TEST: 12/05/2020 01:28:32

Diagnostics Check	OK	01:18
Air Blank	0.000	01:19
Control Test	0.083	01:19
Air Blank	0.000	01:20
Subject Sample # 1	0.187	01:20
Breath Volume	3.406	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.164	01:23
Breath Volume	1.800	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 3	0.169	01:26
Breath Volume	1.687	01:26
Air Blank	0.000	01:27
Control Test	0.079	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 01/20/2021 14:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEITZ,DANIEL,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WHITAKER,JOHNNY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: FEITZ

DATE/TIME OF TEST: 12/12/2020 00:28:07

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:25	
Control Test	0.082	00:25	
Air Blank	0.000	00:26	
Subject Sample	# 1	00:26	Subject Test Refused
Breath Volume	0.000	00:26	
Air Blank	0.000	00:27	
Control Test	0.081	00:27	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:28	

Data Download Date/Time: 01/20/2021 14:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENSON,JAY,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENSON,JAY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:59

SUBJECT'S LAST NAME: SANDERSON

DATE/TIME OF TEST: 12/13/2020 17:30:28

Diagnostics Check		OK	17:19	
Air Blank		0.000	17:20	
Control Test		0.082	17:20	
Air Blank		0.000	17:21	
Subject Sample	# 1	0.282	17:21	No .020 agreement
Breath Volume		3.640	17:21	
Air Blank		0.000	17:22	
Air Blank		0.000	17:24	
Subject Sample	# 2	0.234	17:25	No .020 agreement
Breath Volume		1.906	17:25	
Air Blank		0.000	17:26	
Air Blank		0.000	17:27	
Subject Sample	# 3	0.257	17:28	No .020 agreement
Breath Volume		3.339	17:28	
Air Blank		0.000	17:29	
Control Test		0.076	17:29	
Air Blank		0.000	17:30	
Diagnostics Check		OK	17:30	

Data Download Date/Time: 01/20/2021 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHERS,WILLIAM,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MATHERS,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:13

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 12/19/2020 06:40:15

Diagnostics Check	OK	06:33
Air Blank	0.000	06:34
Control Test	0.082	06:34
Air Blank	0.000	06:35
Subject Sample # 1	0.192	06:35
Breath Volume	1.839	06:35
Air Blank	0.000	06:36
Air Blank	0.000	06:37
Subject Sample # 2	0.185	06:38
Breath Volume	2.234	06:38
Air Blank	0.000	06:39
Control Test	0.081	06:39
Air Blank	0.000	06:40
Diagnostics Check	OK	06:40

Data Download Date/Time: 01/20/2021 14:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHERS,WILLIAM,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MATHERS,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: AUTON

DATE/TIME OF TEST: 12/22/2020 04:06:04

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.083	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.000	04:01
Breath Volume	1.824	04:01
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.000	04:04
Breath Volume	1.761	04:04
Air Blank	0.000	04:05
Control Test	0.082	04:05
Air Blank	0.000	04:05
Diagnostics Check	OK	04:06

Data Download Date/Time: 01/20/2021 14:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAWYER,DARRYL,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SAWYER,DARRYL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: SAVERING

DATE/TIME OF TEST: 12/25/2020 04:07:01

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:02	
Control Test	0.082	04:02	
Air Blank	0.000	04:02	
Subject Sample	# 1	04:05	Subject Test Refused
Breath Volume	0.000	04:05	
Air Blank	0.000	04:06	
Control Test	0.083	04:06	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:07	

Data Download Date/Time: 01/20/2021 14:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITTLE,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LITTLE,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: TROUT

DATE/TIME OF TEST: 12/27/2020 22:06:55

Diagnostics Check	OK	21:56
Air Blank	0.000	21:57
Control Test	0.083	21:57
Air Blank	0.000	21:58
Subject Sample # 1	0.276	21:58
Breath Volume	1.593	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.251	22:01
Breath Volume	1.847	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 3	0.255	22:04
Breath Volume	2.308	22:04
Air Blank	0.000	22:05
Control Test	0.079	22:06
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 01/20/2021 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMS,KYLE,N
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SIMMS,KYLE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: BRAMAN

DATE/TIME OF TEST: 12/12/2020 00:44:38

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.079	00:38
Air Blank	0.000	00:39
Subject Sample # 1	0.196	00:39
Breath Volume	2.121	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.194	00:43
Breath Volume	1.386	00:43
Air Blank	0.000	00:43
Control Test	0.080	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 01/04/2021 13:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MERCURIO

DATE/TIME OF TEST: 12/02/2020 01:27:59

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.082	01:22
Air Blank	0.000	01:22
Subject Sample # 1	0.000	01:23
Breath Volume	1.289	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.000	01:26
Breath Volume	1.750	01:26
Air Blank	0.000	01:27
Control Test	0.081	01:27
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 12/23/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRIQUEZ ZUNIGA,MARCIO,E
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: HENRIQUEZ ZUNIGA,MARCIO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: M635550670840

DATE/TIME OF TEST: 12/04/2020 21:48:31

Diagnostics Check	OK	21:41
Air Blank	0.000	21:42
Control Test	0.081	21:42
Air Blank	0.000	21:42
Subject Sample # 1	0.277	21:44
Breath Volume	3.089	21:44
Air Blank	0.000	21:44
Air Blank	0.000	21:46
Subject Sample # 2	0.274	21:46
Breath Volume	2.703	21:46
Air Blank	0.000	21:47
Control Test	0.079	21:47
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

Data Download Date/Time: 12/23/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOUINARD,JEFFREY,M
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: CHOUINARD,JEFFREY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:32

SUBJECT'S LAST NAME: GRILLO SANTANA

DATE/TIME OF TEST: 12/07/2020 17:58:18

Diagnostics Check	OK	17:52	
Air Blank	0.000	17:53	
Control Test	0.081	17:53	
Air Blank	0.000	17:54	
Subject Sample	# 1	17:56	Subject Test Refused
Breath Volume	0.000	17:56	
Air Blank	0.000	17:57	
Control Test	0.080	17:57	
Air Blank	0.000	17:58	
Diagnostics Check	OK	17:58	

Data Download Date/Time: 12/23/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: ANTON,NICHOLAS
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/09/2020 02:42:51

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.081	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.127	02:38
Breath Volume	2.574	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:40
Subject Sample # 2	0.126	02:41
Breath Volume	1.929	02:41
Air Blank	0.000	02:41
Control Test	0.081	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 12/23/2020 15:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: SUMRA

DATE/TIME OF TEST: 12/10/2020 00:38:25

Diagnostics Check		OK	00:32
Air Blank		0.000	00:32
Control Test		0.081	00:33
Air Blank		0.000	00:33
Subject Sample	# 1	0.203	00:34
Breath Volume		1.281	00:34
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.199	00:36
Breath Volume		1.417	00:36
Air Blank		0.000	00:37
Control Test		0.081	00:37
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 12/23/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: BALL

DATE/TIME OF TEST: 12/12/2020 21:36:57

Diagnostics Check		OK	21:25	
Air Blank		0.000	21:25	
Control Test		0.081	21:26	
Air Blank		0.000	21:26	
Subject Sample	# 1	0.112	21:29	Volume Not Met
Breath Volume		0.089	21:29	
Air Blank		0.000	21:30	
Air Blank		0.000	21:32	
Subject Sample	# 2	0.114	21:35	Volume Not Met
Breath Volume		0.402	21:35	
Air Blank		0.000	21:36	
Control Test		0.078	21:36	
Air Blank		0.000	21:36	
Diagnostics Check		OK	21:36	

Data Download Date/Time: 12/23/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: BALL

DATE/TIME OF TEST: 12/12/2020 21:51:25

Diagnostics Check		OK	21:38	
Air Blank		0.000	21:38	
Control Test		0.079	21:38	
Air Blank		0.000	21:39	
Subject Sample	# 1	0.117	21:42	Volume Not Met
Breath Volume		0.046	21:42	
Air Blank		0.000	21:43	
Air Blank		0.000	21:44	
Subject Sample	# 2	0.103	21:46	
Breath Volume		1.375	21:46	
Air Blank		0.000	21:47	
Air Blank		0.000	21:49	
Subject Sample	# 3	0.106	21:49	
Breath Volume		1.441	21:49	
Air Blank		0.000	21:50	
Control Test		0.078	21:50	
Air Blank		0.000	21:51	
Diagnostics Check		OK	21:51	

Data Download Date/Time: 12/23/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAFIQUE,BILAL,M
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: SHAFIQUE,BILAL
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: RODRIGUE

DATE/TIME OF TEST: 12/13/2020 18:31:25

Diagnostics Check	OK	18:25	
Air Blank	0.000	18:26	
Control Test	0.080	18:26	
Air Blank	0.000	18:27	
Subject Sample	# 1 0.210	18:27	
Breath Volume	3.312	18:27	
Air Blank	0.000	18:28	
Air Blank	0.000	18:30	
Subject Sample	# 2 0.218	18:30	
Breath Volume	3.890	18:30	
Air Blank		18:30	Improper Sample
Air Blank	0.000	18:31	

Data Download Date/Time: 12/23/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAFIQUE,BILAL,M
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: SHAFIQUE,BILAL
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: RODRIGUE

DATE/TIME OF TEST: 12/13/2020 18:46:14

Diagnostics Check	OK	18:39
Air Blank	0.000	18:40
Control Test	0.079	18:40
Air Blank	0.000	18:41
Subject Sample # 1	0.213	18:41
Breath Volume	2.714	18:41
Air Blank	0.000	18:42
Air Blank	0.000	18:44
Subject Sample # 2	0.213	18:44
Breath Volume	2.644	18:44
Air Blank	0.000	18:45
Control Test	0.079	18:45
Air Blank	0.000	18:46
Diagnostics Check	OK	18:46

Data Download Date/Time: 12/23/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERIC,JEAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FLINT,ERIC
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: LEITL

DATE/TIME OF TEST: 12/13/2020 21:58:07

Diagnostics Check	OK	21:54	
Air Blank	0.000	21:54	
Control Test	0.080	21:55	
Air Blank	0.000	21:55	
Subject Sample	# 1	21:56	Subject Test Refused
Breath Volume	0.000	21:56	
Air Blank	0.000	21:57	
Control Test	0.080	21:57	
Air Blank	0.000	21:58	
Diagnostics Check	OK	21:58	

Data Download Date/Time: 12/23/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: KEEFE,RANDY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: BYERS

DATE/TIME OF TEST: 12/16/2020 00:27:18

Diagnostics Check		OK	00:14	
Air Blank		0.000	00:14	
Control Test		0.080	00:14	
Air Blank		0.000	00:15	
Subject Sample	# 1	0.152	00:18	Volume Not Met
Breath Volume		0.937	00:18	
Air Blank		0.000	00:19	
Air Blank		0.000	00:21	
Subject Sample	# 2	0.185	00:22	
Breath Volume		2.183	00:22	
Air Blank		0.000	00:23	
Air Blank		0.000	00:25	
Subject Sample	# 3	0.174	00:25	
Breath Volume		2.304	00:25	
Air Blank		0.000	00:26	
Control Test		0.077	00:26	
Air Blank		0.000	00:27	
Diagnostics Check		OK	00:27	

Data Download Date/Time: 12/23/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRIQUEZ ZUNIGA,MARCIO,E
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: HENRIQUEZ ZUNIGA,MARCIO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: SPEAR

DATE/TIME OF TEST: 12/19/2020 15:53:17

Diagnostics Check	OK	15:47
Air Blank	0.000	15:47
Control Test	0.081	15:48
Air Blank	0.000	15:48
Subject Sample # 1	0.000	15:49
Breath Volume	2.019	15:49
Air Blank	0.000	15:49
Air Blank	0.000	15:51
Subject Sample # 2	0.000	15:51
Breath Volume	1.527	15:51
Air Blank	0.000	15:52
Control Test	0.080	15:52
Air Blank	0.000	15:53
Diagnostics Check	OK	15:53

Data Download Date/Time: 12/23/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REID,MICHAEL,G
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: REID,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: MCKINNEY

DATE/TIME OF TEST: 12/19/2020 17:04:34

Diagnostics Check	OK	16:58
Air Blank	0.000	16:58
Control Test	0.080	16:59
Air Blank	0.000	16:59
Subject Sample # 1	0.000	17:00
Breath Volume	3.640	17:00
Air Blank	0.000	17:00
Air Blank	0.000	17:02
Subject Sample # 2	0.000	17:03
Breath Volume	3.898	17:03
Air Blank	0.000	17:03
Control Test	0.079	17:03
Air Blank	0.000	17:04
Diagnostics Check	OK	17:04

Data Download Date/Time: 12/23/2020 15:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HALL,RICHARD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: MURILLO RODRIGUEZ

DATE/TIME OF TEST: 12/20/2020 01:44:19

Diagnostics Check		OK	01:36
Air Blank		0.000	01:37
Control Test		0.081	01:37
Air Blank		0.000	01:37
Subject Sample	# 1	0.208	01:39
Breath Volume		1.820	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.207	01:42
Breath Volume		1.863	01:42
Air Blank		0.000	01:43
Control Test		0.078	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 12/23/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: REARDON

DATE/TIME OF TEST: 12/22/2020 15:07:33

Diagnostics Check	OK	15:00
Air Blank	0.000	15:01
Control Test	0.081	15:01
Air Blank	0.000	15:02
Subject Sample # 1	0.000	15:03
Breath Volume	2.222	15:03
Air Blank	0.000	15:03
Air Blank	0.000	15:05
Subject Sample # 2	0.000	15:06
Breath Volume	2.339	15:06
Air Blank	0.000	15:06
Control Test	0.081	15:06
Air Blank	0.000	15:07
Diagnostics Check	OK	15:07

Data Download Date/Time: 12/23/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,CODY,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: BROWN,CODY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 12/23/2020 21:20:30

Diagnostics Check		OK	21:14
Air Blank		0.000	21:14
Control Test		0.081	21:15
Air Blank		0.000	21:15
Subject Sample	# 1	0.223	21:16
Breath Volume		1.480	21:16
Air Blank		0.000	21:16
Air Blank		0.000	21:18
Subject Sample	# 2	0.217	21:18
Breath Volume		1.660	21:18
Air Blank		0.000	21:19
Control Test		0.078	21:19
Air Blank		0.000	21:20
Diagnostics Check		OK	21:20

Data Download Date/Time: 01/28/2021 07:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: NICOLAS,COLANGELO
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MADDALINA

DATE/TIME OF TEST: 12/25/2020 00:50:44

Diagnostics Check	OK	00:43
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:44
Subject Sample # 1	0.079	00:46
Breath Volume	1.621	00:46
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.072	00:49
Breath Volume	1.253	00:49
Air Blank	0.000	00:49
Control Test	0.079	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 01/28/2021 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HALL,RICHARD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 12/25/2020 03:42:17

Diagnostics Check		OK	03:34
Air Blank		0.000	03:35
Control Test		0.080	03:35
Air Blank		0.000	03:36
Subject Sample	# 1	0.201	03:37
Breath Volume		1.410	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:39
Subject Sample	# 2	0.214	03:40
Breath Volume		1.269	03:40
Air Blank		0.000	03:41
Control Test		0.078	03:41
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 01/28/2021 07:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODRUFF,MARK,D
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: COLANGELO,CODY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: KAISER

DATE/TIME OF TEST: 12/25/2020 22:33:47

Diagnostics Check	OK	22:27
Air Blank	0.000	22:27
Control Test	0.081	22:28
Air Blank	0.000	22:28
Subject Sample # 1	0.280	22:29
Breath Volume	3.953	22:29
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.293	22:32
Breath Volume	3.875	22:32
Air Blank	0.000	22:32
Control Test	0.080	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 01/28/2021 07:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: LARSEN

DATE/TIME OF TEST: 12/29/2020 18:52:41

Diagnostics Check	OK	18:46
Air Blank	0.000	18:46
Control Test	0.081	18:47
Air Blank	0.000	18:47
Subject Sample # 1	0.192	18:48
Breath Volume	5.664	18:48
Air Blank	0.000	18:48
Air Blank	0.000	18:50
Subject Sample # 2	0.189	18:51
Breath Volume	5.613	18:51
Air Blank	0.000	18:51
Control Test	0.080	18:52
Air Blank	0.000	18:52
Diagnostics Check	OK	18:52

Data Download Date/Time: 01/28/2021 07:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIBBEY,BRADLEY,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIBBEY,BRADLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: MONTERO YGLESIAS

DATE/TIME OF TEST: 12/02/2020 21:04:46

Diagnostics Check	OK	21:00	
Air Blank	0.000	21:01	
Control Test	0.081	21:01	
Air Blank	0.000	21:02	
Subject Sample	# 1	21:03	Subject Test Refused
Breath Volume	0.000	21:03	
Air Blank	0.000	21:03	
Control Test	0.082	21:04	
Air Blank	0.000	21:04	
Diagnostics Check	OK	21:04	

Data Download Date/Time: 12/12/2020 23:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELM,DEREK,W
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HELM,DEREK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: HOWE

DATE/TIME OF TEST: 12/03/2020 02:25:49

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:16	
Control Test		0.081	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 1	0.100	02:20	Volume Not Met
Breath Volume		0.550	02:20	
Air Blank		0.000	02:21	
Air Blank		0.000	02:22	
Subject Sample	# 2		02:24	Subject Test Refused
Breath Volume		0.457	02:24	
Air Blank		0.000	02:24	
Control Test		0.080	02:25	
Air Blank		0.000	02:25	
Diagnostics Check		OK	02:25	

Data Download Date/Time: 12/12/2020 23:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,IVAN,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MOONEY,JAMES
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: GOETZ

DATE/TIME OF TEST: 12/05/2020 01:35:12

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:32	
Control Test	0.081	01:32	
Air Blank	0.000	01:33	
Subject Sample	# 1	01:33	Subject Test Refused
Breath Volume	0.000	01:33	
Air Blank	0.000	01:34	
Control Test	0.081	01:34	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 12/12/2020 23:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,IVAN,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RIZZO,HALEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: BECKER

DATE/TIME OF TEST: 12/05/2020 03:57:14

Diagnostics Check		OK	03:43	
Air Blank		0.000	03:43	
Control Test		0.081	03:43	
Air Blank		0.000	03:44	
Subject Sample	# 1	0.083	03:47	Volume Not Met
Breath Volume		0.175	03:47	
Air Blank		0.000	03:48	
Air Blank		0.000	03:49	
Subject Sample	# 2	0.080	03:51	
Breath Volume		1.253	03:51	
Air Blank		0.000	03:52	
Air Blank		0.000	03:54	
Subject Sample	# 3	0.082	03:55	
Breath Volume		1.496	03:55	
Air Blank		0.000	03:56	
Control Test		0.079	03:56	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 12/12/2020 23:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZABRISKIE,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MATHIS,JONATHAN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:25

SUBJECT'S LAST NAME: ZABRISKIE

DATE/TIME OF TEST: 12/06/2020 11:49:08

Diagnostics Check	OK	11:45	
Air Blank	0.000	11:46	
Control Test	0.081	11:46	
Air Blank	0.000	11:47	
Subject Sample	# 1	11:47	Subject Test Refused
Breath Volume	0.000	11:47	
Air Blank	0.000	11:48	
Control Test	0.081	11:48	
Air Blank	0.000	11:49	
Diagnostics Check	OK	11:49	

Data Download Date/Time: 12/12/2020 23:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY, WESLEY, R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KELLY, WESLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 12/09/2020 16:33:59

Diagnostics Check	OK	16:26
Air Blank	0.000	16:26
Control Test	0.082	16:27
Air Blank	0.000	16:27
Subject Sample # 1	0.145	16:28
Breath Volume	2.480	16:28
Air Blank	0.000	16:29
Air Blank	0.000	16:31
Subject Sample # 2	0.141	16:32
Breath Volume	2.496	16:32
Air Blank	0.000	16:33
Control Test	0.081	16:33
Air Blank	0.000	16:33
Diagnostics Check	OK	16:33

Data Download Date/Time: 12/12/2020 23:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATIF, ANTHONY, R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RIZZO, HAYLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: RUCH

DATE/TIME OF TEST: 12/09/2020 22:46:54

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:37	
Control Test	0.081	22:38	
Air Blank	0.000	22:38	
Subject Sample	# 1	22:41	No Sample Provided
Breath Volume	0.000	22:41	
Air Blank	0.000	22:42	
Air Blank	0.000	22:44	
Subject Sample	# 2	22:45	Subject Test Refused
Breath Volume	0.000	22:45	
Air Blank	0.000	22:45	
Control Test	0.077	22:46	
Air Blank	0.000	22:46	
Diagnostics Check	OK	22:46	

Data Download Date/Time: 12/12/2020 23:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HECKMAN,DANIEL,O
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MOONKEY,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: YUNGMANN

DATE/TIME OF TEST: 12/10/2020 23:09:58

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.080	23:03
Air Blank	0.000	23:03
Subject Sample # 1	0.103	23:04
Breath Volume	2.023	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.101	23:08
Breath Volume	1.992	23:08
Air Blank	0.000	23:09
Control Test	0.079	23:09
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 12/12/2020 23:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,IVAN,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RIZZO,HAYLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: BEST

DATE/TIME OF TEST: 12/14/2020 22:58:53

Diagnostics Check	OK	22:51
Air Blank	0.000	22:52
Control Test	0.081	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.125	22:54
Breath Volume	1.589	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:56
Subject Sample # 2	0.114	22:57
Breath Volume	1.539	22:57
Air Blank	0.000	22:57
Control Test	0.079	22:58
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 01/20/2021 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: DELANEY

DATE/TIME OF TEST: 12/16/2020 19:23:16

Diagnostics Check	OK	19:16
Air Blank	0.000	19:17
Control Test	0.081	19:17
Air Blank	0.000	19:18
Subject Sample # 1	0.114	19:18
Breath Volume	2.914	19:18
Air Blank	0.000	19:19
Air Blank	0.000	19:21
Subject Sample # 2	0.108	19:21
Breath Volume	2.605	19:21
Air Blank	0.000	19:22
Control Test	0.079	19:22
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

Data Download Date/Time: 01/20/2021 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,JONATHON,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MOONEY,JAMES
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/19/2020 23:27:05

Diagnostics Check	OK	23:22	
Air Blank	0.000	23:23	
Control Test	0.081	23:23	
Air Blank	0.000	23:24	
Subject Sample	# 1	23:25	Subject Test Refused
Breath Volume	0.000	23:25	
Air Blank	0.000	23:26	
Control Test	0.080	23:26	
Air Blank	0.000	23:27	
Diagnostics Check	OK	23:27	

Data Download Date/Time: 01/20/2021 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,JONATHON,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MOONEY,JAMES
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/19/2020 23:42:53

Diagnostics Check	OK	23:35
Air Blank	0.000	23:35
Control Test	0.081	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.149	23:37
Breath Volume	1.578	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.147	23:41
Breath Volume	2.183	23:41
Air Blank	0.000	23:41
Control Test	0.080	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 01/20/2021 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: KJEKSTAD,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:46

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 12/20/2020 12:13:07

Diagnostics Check		OK	12:06
Air Blank		0.000	12:06
Control Test		0.081	12:07
Air Blank		0.000	12:07
Subject Sample	# 1	0.234	12:08
Breath Volume		1.214	12:08
Air Blank		0.000	12:08
Air Blank		0.000	12:10
Subject Sample	# 2	0.237	12:11
Breath Volume		1.691	12:11
Air Blank		0.000	12:12
Control Test		0.080	12:12
Air Blank		0.000	12:13
Diagnostics Check		OK	12:13

Data Download Date/Time: 01/20/2021 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/21/2020 19:27:16

Diagnostics Check	OK	19:20
Air Blank	0.000	19:20
Control Test	0.081	19:21
Air Blank	0.000	19:21
Subject Sample # 1	0.000	19:22
Breath Volume	2.121	19:22
Air Blank	0.000	19:23
Air Blank	0.000	19:25
Subject Sample # 2	0.000	19:25
Breath Volume	1.867	19:25
Air Blank	0.000	19:26
Control Test	0.079	19:26
Air Blank	0.000	19:27
Diagnostics Check	OK	19:27

Data Download Date/Time: 01/20/2021 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIBBEY,BRADLEY,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIBBEY,BRADLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 12/23/2020 03:29:10

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.081	03:23
Air Blank	0.000	03:23
Subject Sample # 1	0.193	03:24
Breath Volume	1.636	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:26
Subject Sample # 2	0.174	03:27
Breath Volume	1.863	03:27
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 01/20/2021 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO, IVAN,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MOONEY, JAMES
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: ROSADO

DATE/TIME OF TEST: 12/23/2020 21:50:16

Diagnostics Check	OK	21:42
Air Blank	0.000	21:43
Control Test	0.081	21:43
Air Blank	0.000	21:43
Subject Sample # 1	0.110	21:44
Breath Volume	1.464	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.115	21:48
Breath Volume	2.050	21:48
Air Blank	0.000	21:49
Control Test	0.080	21:49
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 01/20/2021 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,IVAN,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SCHMIDT,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 12/24/2020 22:49:03

Diagnostics Check	OK	22:48	
Air Blank		22:48	RFI Detect
Air Blank		22:49	RFI Detect

Data Download Date/Time: 01/20/2021 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO, IVAN,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SCHMIDT, ANDREW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 12/24/2020 23:16:24

Diagnostics Check	OK	23:09
Air Blank	0.000	23:10
Control Test	0.082	23:10
Air Blank	0.000	23:11
Subject Sample # 1	0.150	23:11
Breath Volume	2.644	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.148	23:14
Breath Volume	2.570	23:14
Air Blank	0.000	23:15
Control Test	0.080	23:15
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 01/20/2021 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: CASE

DATE/TIME OF TEST: 12/27/2020 02:32:26

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.081	02:21	
Air Blank	0.000	02:22	
Subject Sample	# 1	02:25	No Sample Provided
Breath Volume	0.000	02:25	
Air Blank	0.000	02:25	
Air Blank	0.000	02:27	
Subject Sample	# 2	02:31	No Sample Provided
Breath Volume	0.000	02:31	
Air Blank	0.000	02:31	
Control Test	0.082	02:31	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 01/20/2021 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 12/27/2020 20:22:31

Diagnostics Check	OK	20:16
Air Blank	0.000	20:16
Control Test	0.082	20:16
Air Blank	0.000	20:17
Subject Sample # 1	0.104	20:17
Breath Volume	2.988	20:17
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample # 2	0.107	20:20
Breath Volume	3.496	20:20
Air Blank	0.000	20:21
Control Test	0.081	20:21
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

Data Download Date/Time: 01/20/2021 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEVORAK,MATTHEW,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: VELAZQUEZ,C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 12/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 12/06/2020 03:40:39

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:37	
Control Test	0.077	03:38	
Air Blank	0.000	03:38	
Subject Sample	# 1	03:39	Subject Test Refused
Breath Volume	0.000	03:39	
Air Blank	0.000	03:39	
Control Test	0.078	03:40	
Air Blank	0.000	03:40	
Diagnostics Check	OK	03:40	

Data Download Date/Time: 02/17/2021 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASON,CHRISTOPHER,N
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: MCCULLOUGH,X
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: DAY

DATE/TIME OF TEST: 12/12/2020 23:04:19

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.078	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.111	22:58
Breath Volume	2.257	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.108	23:02
Breath Volume	2.062	23:02
Air Blank	0.000	23:03
Control Test	0.079	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 02/17/2021 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASON,CHRISTOPHER,N
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,CCSO
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: COTHERN

DATE/TIME OF TEST: 12/12/2020 23:49:21

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:45	
Control Test	0.078	23:46	
Air Blank	0.000	23:46	
Subject Sample	# 1	23:47	Subject Test Refused
Breath Volume	0.000	23:47	
Air Blank	0.000	23:48	
Control Test	0.079	23:48	
Air Blank	0.000	23:49	
Diagnostics Check	OK	23:49	

Data Download Date/Time: 02/17/2021 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASON,CHRISTOPHER,N
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: MCCULLOUGH,X
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: SMOLENSKY

DATE/TIME OF TEST: 12/14/2020 00:04:01

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.078	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.203	23:58
Breath Volume	1.902	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.213	00:02
Breath Volume	2.222	00:02
Air Blank	0.000	00:03
Control Test	0.078	00:03
Air Blank	0.000	00:03
Diagnostics Check	OK	00:04

Data Download Date/Time: 02/17/2021 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLK, BRANDON, A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BAKER, J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 12/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: WENGER

DATE/TIME OF TEST: 12/19/2020 02:05:42

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:03	
Control Test	0.078	02:03	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:04	Subject Test Refused
Breath Volume	0.000	02:04	
Air Blank	0.000	02:04	
Control Test	0.078	02:05	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 02/17/2021 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASON,CHRISTOPHER,N
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: FERRER,F
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 12/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/23/2020 00:57:04

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:53	
Control Test	0.078	00:54	
Air Blank	0.000	00:54	
Subject Sample	# 1	00:55	Slope Not Met
Breath Volume	1.648	00:55	
Air Blank	0.000	00:56	
Control Test	0.078	00:56	
Air Blank	0.000	00:56	
Diagnostics Check	OK	00:57	

Data Download Date/Time: 02/17/2021 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASON,CHRISTOPHER,N
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: FERRER,F
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 12/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/23/2020 01:06:23

Diagnostics Check		OK	00:59
Air Blank		0.000	00:59
Control Test		0.077	00:59
Air Blank		0.000	01:00
Subject Sample	# 1	0.262	01:01
Breath Volume		1.578	01:01
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.251	01:04
Breath Volume		1.417	01:04
Air Blank		0.000	01:05
Control Test		0.077	01:05
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 02/17/2021 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BLOTZ,G
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 12/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 12/24/2020 17:13:58

Diagnostics Check	OK	17:11	
Air Blank	0.000	17:11	
Control Test	0.078	17:11	
Air Blank	0.000	17:12	
Subject Sample	# 1	17:12	Subject Test Refused
Breath Volume	0.000	17:12	
Air Blank	0.000	17:13	
Control Test	0.078	17:13	
Air Blank	0.000	17:13	
Diagnostics Check	OK	17:13	

Data Download Date/Time: 02/17/2021 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CHASE,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: JOHNSON,D
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 12/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: SIDES

DATE/TIME OF TEST: 12/24/2020 22:00:12

Diagnostics Check	OK	21:53
Air Blank	0.000	21:54
Control Test	0.078	21:54
Air Blank	0.000	21:55
Subject Sample # 1	0.000	21:55
Breath Volume	2.964	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.000	21:58
Breath Volume	2.406	21:58
Air Blank	0.000	21:59
Control Test	0.078	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 02/17/2021 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASON,CHRISTOPHER,N
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROBINSON,K
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 12/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: MAZZARESE

DATE/TIME OF TEST: 12/28/2020 19:38:38

Diagnostics Check	OK	19:33	
Air Blank	0.000	19:34	
Control Test	0.078	19:34	
Air Blank	0.000	19:34	
Subject Sample	# 1	19:37	Subject Test Refused
Breath Volume	0.000	19:37	
Air Blank	0.000	19:37	
Control Test	0.078	19:38	
Air Blank	0.000	19:38	
Diagnostics Check	OK	19:38	

Data Download Date/Time: 02/17/2021 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DAVENPORT,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: BRESSAN STACEY

DATE/TIME OF TEST: 12/29/2020 20:09:49

Diagnostics Check	OK	20:03
Air Blank	0.000	20:04
Control Test	0.077	20:04
Air Blank	0.000	20:05
Subject Sample # 1	0.000	20:05
Breath Volume	2.273	20:05
Air Blank	0.000	20:06
Air Blank	0.000	20:08
Subject Sample # 2	0.000	20:08
Breath Volume	1.921	20:08
Air Blank	0.000	20:08
Control Test	0.077	20:09
Air Blank	0.000	20:09
Diagnostics Check	OK	20:09

Data Download Date/Time: 02/17/2021 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASON,CHRISTOPHER,N
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 12/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/31/2020 22:51:09

Diagnostics Check	OK	22:48	
Air Blank	0.000	22:48	
Control Test	0.078	22:48	
Air Blank	0.000	22:49	
Subject Sample	# 1	22:49	Subject Test Refused
Breath Volume	0.000	22:49	
Air Blank	0.000	22:50	
Control Test	0.078	22:50	
Air Blank	0.000	22:51	
Diagnostics Check	OK	22:51	

Data Download Date/Time: 02/17/2021 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,RODNEY,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: CARTER,BENJAMIN
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: D'HONDT

DATE/TIME OF TEST: 12/01/2020 02:47:32

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.082	02:45	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:46	Subject Test Refused
Breath Volume	0.000	02:46	
Air Blank	0.000	02:46	
Control Test	0.083	02:46	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 12/16/2020 22:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURCHINSON, RONNIE,
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: STEWART, JUSTIN
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: CURTICE

DATE/TIME OF TEST: 12/02/2020 01:38:14

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:33	
Control Test	0.083	01:33	
Air Blank	0.000	01:33	
Subject Sample	# 1	01:36	Subject Test Refused
Breath Volume	0.000	01:36	
Air Blank	0.000	01:37	
Control Test	0.083	01:37	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 12/16/2020 22:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURCHINSON, RONNIE,
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: BARNES, MONIQUE
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 12/02/2020 02:23:11

Diagnostics Check	OK	02:11		
Air Blank	0.000	02:11		
Control Test	0.084	02:12		
Air Blank	0.000	02:12		
Subject Sample	# 1	02:15	No Sample Provided	
Breath Volume	0.000	02:15		
Air Blank	0.000	02:16		
Air Blank	0.000	02:18		
Subject Sample	# 2	0.150	02:21	Volume Not Met
Breath Volume	0.382	02:21		
Air Blank	0.000	02:22		
Control Test	0.084	02:22		
Air Blank	0.000	02:23		
Diagnostics Check	OK	02:23		

Data Download Date/Time: 12/16/2020 22:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURCHINSON, RONNIE,
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: BARNES, MONIQUE
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 12/02/2020 02:33:00

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.082	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.179	02:27
Breath Volume	1.511	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.175	02:31
Breath Volume	2.238	02:31
Air Blank	0.000	02:32
Control Test	0.082	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:33

Data Download Date/Time: 12/16/2020 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,C,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: WILLIAMS,Z
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: BURKHARDT

DATE/TIME OF TEST: 12/04/2020 02:56:44

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:54	
Control Test	0.082	02:54	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:55	Subject Test Refused
Breath Volume	0.000	02:55	
Air Blank	0.000	02:55	
Control Test	0.082	02:56	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

Data Download Date/Time: 12/16/2020 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIEVES, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: GRIEVES, MATTHEW
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: BOWERS

DATE/TIME OF TEST: 12/04/2020 17:00:39

Diagnostics Check	OK	16:55	
Air Blank	0.000	16:56	
Control Test	0.082	16:56	
Air Blank	0.000	16:57	
Subject Sample	# 1	16:59	Subject Test Refused
Breath Volume	0.000	16:59	
Air Blank	0.000	16:59	
Control Test	0.082	17:00	
Air Blank	0.000	17:00	
Diagnostics Check	OK	17:00	

Data Download Date/Time: 12/16/2020 22:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALSH,SHANNON,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: CHESSHER,S
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: SAMPLES

DATE/TIME OF TEST: 12/04/2020 20:45:30

Diagnostics Check	OK	20:39
Air Blank	0.000	20:39
Control Test	0.082	20:40
Air Blank	0.000	20:40
Subject Sample # 1	0.000	20:41
Breath Volume	2.609	20:41
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.000	20:44
Breath Volume	1.968	20:44
Air Blank	0.000	20:44
Control Test	0.082	20:44
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 12/16/2020 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIJAGIC,SAMIR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ALIJAGIC,SAMIR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 12/04/2020 22:29:36

Diagnostics Check	OK	22:22
Air Blank	0.000	22:23
Control Test	0.081	22:23
Air Blank	0.000	22:24
Subject Sample # 1	0.153	22:24
Breath Volume	2.546	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.161	22:27
Breath Volume	2.007	22:27
Air Blank	0.000	22:28
Control Test	0.079	22:28
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 12/16/2020 22:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADE,C,L
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: SULLIVAN,Z
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 07:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:13

SUBJECT'S LAST NAME: ORTIZ SERRANO

DATE/TIME OF TEST: 12/05/2020 08:42:32

Diagnostics Check	OK	08:34	
Air Blank	0.000	08:34	
Control Test	0.082	08:34	
Air Blank	0.000	08:35	
Subject Sample	# 1	08:38	No Sample Provided
Breath Volume	0.000	08:38	
Air Blank	0.000	08:38	
Air Blank	0.000	08:40	
Subject Sample	# 2	08:41	Subject Test Refused
Breath Volume	0.000	08:41	
Air Blank	0.000	08:41	
Control Test	0.082	08:41	
Air Blank	0.000	08:42	
Diagnostics Check	OK	08:42	

Data Download Date/Time: 12/16/2020 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,BENJAMIN,A
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: BAUDERS,SEAN
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/05/2020 23:15:26

Diagnostics Check	OK	23:12	
Air Blank	0.000	23:13	
Control Test	0.083	23:13	
Air Blank	0.000	23:13	
Subject Sample	# 1	23:14	Subject Test Refused
Breath Volume	0.000	23:14	
Air Blank	0.000	23:14	
Control Test	0.083	23:14	
Air Blank	0.000	23:15	
Diagnostics Check	OK	23:15	

Data Download Date/Time: 12/16/2020 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: BANNICK,COURTNEY
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: ^?

DATE/TIME OF TEST: 12/06/2020 01:58:18

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.083	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.122	01:53
Breath Volume	2.835	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.113	01:56
Breath Volume	1.941	01:56
Air Blank	0.000	01:57
Control Test	0.082	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 12/16/2020 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, RONALD,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: RODRIGUEZ, RONALD
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 16:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 12/06/2020 18:42:51

Diagnostics Check	OK	18:39	
Air Blank	0.000	18:40	
Control Test	0.083	18:40	
Air Blank	0.000	18:41	
Subject Sample	# 1	18:41	Subject Test Refused
Breath Volume	0.000	18:41	
Air Blank	0.000	18:41	
Control Test	0.083	18:42	
Air Blank	0.000	18:42	
Diagnostics Check	OK	18:42	

Data Download Date/Time: 12/16/2020 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: RODEN,ANDREW
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: EICHSTADT

DATE/TIME OF TEST: 12/06/2020 20:10:24

Diagnostics Check	OK	20:07	
Air Blank	0.000	20:08	
Control Test	0.083	20:08	
Air Blank	0.000	20:08	
Subject Sample	# 1	20:09	Subject Test Refused
Breath Volume	0.000	20:09	
Air Blank	0.000	20:09	
Control Test	0.083	20:09	
Air Blank	0.000	20:10	
Diagnostics Check	OK	20:10	

Data Download Date/Time: 12/16/2020 22:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORFEL,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WORFEL,KYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: MONTERO

DATE/TIME OF TEST: 12/09/2020 05:15:53

Diagnostics Check	OK	05:10	
Air Blank	0.000	05:11	
Control Test	0.083	05:11	
Air Blank	0.000	05:12	
Subject Sample	# 1	05:14	Subject Test Refused
Breath Volume	0.000	05:14	
Air Blank	0.000	05:14	
Control Test	0.084	05:15	
Air Blank	0.000	05:15	
Diagnostics Check	OK	05:15	

Data Download Date/Time: 12/16/2020 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: BANNICK,COURTNEY
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 12/11/2020 00:34:37

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:32	
Control Test	0.083	00:32	
Air Blank	0.000	00:33	
Subject Sample	# 1	00:33	Subject Test Refused
Breath Volume	0.000	00:33	
Air Blank	0.000	00:33	
Control Test	0.083	00:34	
Air Blank	0.000	00:34	
Diagnostics Check	OK	00:34	

Data Download Date/Time: 12/16/2020 22:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PARRISH,C
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 12/12/2020 03:41:04

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.083	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.153	03:36
Breath Volume	2.507	03:36
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.149	03:39
Breath Volume	2.296	03:39
Air Blank	0.000	03:40
Control Test	0.083	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:41

Data Download Date/Time: 12/16/2020 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,R,S
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: PENA,R
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/12/2020 05:20:12

Diagnostics Check	OK	05:19	
Air Blank		05:19	RFI Detect
Air Blank	0.000	05:20	

Data Download Date/Time: 12/16/2020 22:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,R,S
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: PENA,R
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/12/2020 05:28:57

Diagnostics Check	OK	05:21
Air Blank	0.000	05:22
Control Test	0.082	05:22
Air Blank	0.000	05:23
Subject Sample # 1	0.080	05:24
Breath Volume	2.171	05:24
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.082	05:27
Breath Volume	1.179	05:27
Air Blank	0.000	05:28
Control Test	0.082	05:28
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 12/16/2020 22:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BELL,ETHAN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: QUILES

DATE/TIME OF TEST: 12/12/2020 21:36:37

Diagnostics Check	OK	21:29
Air Blank	0.000	21:30
Control Test	0.082	21:30
Air Blank	0.000	21:31
Subject Sample # 1	0.195	21:32
Breath Volume	1.824	21:32
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.193	21:35
Breath Volume	2.179	21:35
Air Blank	0.000	21:35
Control Test	0.082	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 12/16/2020 22:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: TODD,MICHAEL
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: SKAGGS

DATE/TIME OF TEST: 12/15/2020 03:40:27

Diagnostics Check		OK	03:33
Air Blank		0.000	03:34
Control Test		0.083	03:34
Air Blank		0.000	03:35
Subject Sample	# 1	0.245	03:35
Breath Volume		1.773	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.248	03:38
Breath Volume		1.640	03:38
Air Blank		0.000	03:39
Control Test		0.082	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 12/16/2020 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,ANDREW,J
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BARNES,STEVEN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 11/11/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 10:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:23

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 12/15/2020 11:51:03

Diagnostics Check	OK	11:44
Air Blank	0.000	11:44
Control Test	0.082	11:45
Air Blank	0.000	11:45
Subject Sample # 1	0.313	11:46
Breath Volume	1.468	11:46
Air Blank	0.000	11:47
Air Blank	0.000	11:48
Subject Sample # 2	0.319	11:49
Breath Volume	1.292	11:49
Air Blank	0.000	11:50
Control Test	0.082	11:50
Air Blank	0.000	11:50
Diagnostics Check	OK	11:51

Data Download Date/Time: 12/16/2020 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORFEL,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WORFEL,KYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:15

SUBJECT'S LAST NAME: PAGLIALONGA

DATE/TIME OF TEST: 12/17/2020 09:08:14

Diagnostics Check	OK	09:05	
Air Blank	0.000	09:05	
Control Test	0.082	09:05	
Air Blank	0.000	09:06	
Subject Sample	# 1 0.136	09:07	
Breath Volume	1.195	09:07	
Air Blank		09:07	RFI Detect
Air Blank		09:07	RFI Detect
Air Blank	0.000	09:08	

Data Download Date/Time: 01/13/2021 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORFEL,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WORFEL,KYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:15

SUBJECT'S LAST NAME: PAGLIALONGA

DATE/TIME OF TEST: 12/17/2020 09:15:56

Diagnostics Check	OK	09:09
Air Blank	0.000	09:10
Control Test	0.082	09:10
Air Blank	0.000	09:11
Subject Sample # 1	0.146	09:11
Breath Volume	1.894	09:11
Air Blank	0.000	09:12
Air Blank	0.000	09:13
Subject Sample # 2	0.149	09:14
Breath Volume	1.691	09:14
Air Blank	0.000	09:15
Control Test	0.082	09:15
Air Blank	0.000	09:15
Diagnostics Check	OK	09:15

Data Download Date/Time: 01/13/2021 16:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,C,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: KESSINGER,C
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 12/18/2020 04:36:24

Diagnostics Check	OK	04:33	
Air Blank	0.000	04:34	
Control Test	0.083	04:34	
Air Blank	0.000	04:34	
Subject Sample	# 1	04:35	Subject Test Refused
Breath Volume	0.000	04:35	
Air Blank	0.000	04:35	
Control Test	0.084	04:35	
Air Blank	0.000	04:36	
Diagnostics Check	OK	04:36	

Data Download Date/Time: 01/13/2021 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: COON

DATE/TIME OF TEST: 12/19/2020 03:24:16

Diagnostics Check		OK	03:17
Air Blank		0.000	03:17
Control Test		0.083	03:18
Air Blank		0.000	03:18
Subject Sample	# 1	0.177	03:19
Breath Volume		1.460	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:21
Subject Sample	# 2	0.173	03:22
Breath Volume		1.574	03:22
Air Blank		0.000	03:23
Control Test		0.083	03:23
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 01/13/2021 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,ROBERT,L
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: COLSTON BARNES,MONIQUE
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/20/2020 00:54:41

Diagnostics Check	OK	00:46
Air Blank	0.000	00:47
Control Test	0.082	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.185	00:49
Breath Volume	1.238	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.174	00:53
Breath Volume	1.402	00:53
Air Blank	0.000	00:53
Control Test	0.082	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 01/13/2021 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,ROBERT,L
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: FREITAS,MARCIA
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: CONTERAS

DATE/TIME OF TEST: 12/20/2020 01:16:30

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.082	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.208	01:09
Breath Volume	1.265	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.217	01:14
Breath Volume	1.894	01:14
Air Blank	0.000	01:15
Control Test	0.082	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 01/13/2021 16:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,ROBERT,L
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: RODEN,ANDREW
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: CASTILLO SANCHEZ

DATE/TIME OF TEST: 12/20/2020 01:26:36

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:23	
Control Test	0.081	01:23	
Air Blank	0.000	01:23	
Subject Sample	# 1	01:25	Subject Test Refused
Breath Volume	0.000	01:25	
Air Blank	0.000	01:25	
Control Test	0.082	01:26	
Air Blank	0.000	01:26	
Diagnostics Check	OK	01:26	

Data Download Date/Time: 01/13/2021 16:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORFEL,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WORFEL,KYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: APPLEWHITE

DATE/TIME OF TEST: 12/20/2020 02:32:54

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.082	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.218	02:28
Breath Volume	2.191	02:28
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.221	02:31
Breath Volume	1.722	02:31
Air Blank	0.000	02:31
Control Test	0.081	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 01/13/2021 16:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURCHINSON, RONNIE,
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: MURCHINSON, RONNIE
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: LOPEZ LOPEZ

DATE/TIME OF TEST: 12/25/2020 03:36:06

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:31	
Control Test	0.083	03:32	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:34	Subject Test Refused
Breath Volume	0.000	03:34	
Air Blank	0.000	03:35	
Control Test	0.083	03:35	
Air Blank	0.000	03:36	
Diagnostics Check	OK	03:36	

Data Download Date/Time: 01/13/2021 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTER,ALEC,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RITTER,ALEC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: DESIMONE

DATE/TIME OF TEST: 12/27/2020 00:49:45

Diagnostics Check		OK	00:42
Air Blank		0.000	00:43
Control Test		0.083	00:43
Air Blank		0.000	00:44
Subject Sample	# 1	0.131	00:45
Breath Volume		1.179	00:45
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.134	00:48
Breath Volume		1.730	00:48
Air Blank		0.000	00:48
Control Test		0.083	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 01/13/2021 16:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,C,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BROWNSBERGER,B
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: BOSTWICK

DATE/TIME OF TEST: 12/27/2020 03:58:53

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:56	
Control Test	0.084	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 1	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Control Test	0.084	03:58	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 01/13/2021 16:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTER,ALEC,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RITTER,ALEC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: GABINO

DATE/TIME OF TEST: 12/27/2020 04:58:28

Diagnostics Check	OK	04:51
Air Blank	0.000	04:51
Control Test	0.083	04:51
Air Blank	0.000	04:52
Subject Sample # 1	0.150	04:53
Breath Volume	2.386	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:55
Subject Sample # 2	0.151	04:56
Breath Volume	2.746	04:56
Air Blank	0.000	04:57
Control Test	0.083	04:57
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 01/13/2021 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: D'AGOSTINO,STEFANO
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: KESSLER

DATE/TIME OF TEST: 12/29/2020 20:45:21

Diagnostics Check	OK	20:38
Air Blank	0.000	20:39
Control Test	0.083	20:39
Air Blank	0.000	20:39
Subject Sample # 1	0.239	20:40
Breath Volume	2.113	20:40
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.228	20:43
Breath Volume	2.214	20:43
Air Blank	0.000	20:44
Control Test	0.082	20:44
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 01/13/2021 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: DONAKOWSKI,J
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: CENTENO

DATE/TIME OF TEST: 12/30/2020 22:59:23

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.082	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.247	22:54
Breath Volume	2.421	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.252	22:57
Breath Volume	2.394	22:57
Air Blank	0.000	22:58
Control Test	0.082	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 01/13/2021 16:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: RICKETTS,J
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 12/31/2020 00:01:50

Diagnostics Check		OK	23:50	
Air Blank		0.000	23:50	
Control Test		0.081	23:50	
Air Blank		0.000	23:51	
Subject Sample	# 2	0.091	23:54	Volume Not Met
Breath Volume		0.054	23:54	
Air Blank		0.000	23:55	
Air Blank		0.000	23:57	
Subject Sample	# 1	0.103	00:00	Volume Not Met
Breath Volume		0.101	00:00	
Air Blank		0.000	00:00	
Control Test		0.080	00:01	
Air Blank		0.000	00:01	
Diagnostics Check		OK	00:01	

Data Download Date/Time: 01/13/2021 16:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: RICKETTS,J
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 12/31/2020 00:15:38

Diagnostics Check		OK	00:03	
Air Blank		0.000	00:03	
Control Test		0.080	00:04	
Air Blank		0.000	00:04	
Subject Sample	# 1	0.103	00:05	
Breath Volume		1.210	00:05	
Air Blank		0.000	00:06	
Air Blank		0.000	00:07	
Subject Sample	# 2	0.095	00:11	Volume Not Met
Breath Volume		0.398	00:11	
Air Blank		0.000	00:11	
Air Blank		0.000	00:13	
Subject Sample	# 3	0.109	00:14	
Breath Volume		1.250	00:14	
Air Blank		0.000	00:14	
Control Test		0.080	00:15	
Air Blank		0.000	00:15	
Diagnostics Check		OK	00:15	

Data Download Date/Time: 01/13/2021 16:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: RICKETTS,J
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: GALLAGHER

DATE/TIME OF TEST: 12/31/2020 22:41:02

Diagnostics Check	OK	22:33	
Air Blank	0.000	22:34	
Control Test	0.082	22:34	
Air Blank	0.000	22:34	
Subject Sample	# 1 0.176	22:36	
Breath Volume	2.089	22:36	
Air Blank	0.000	22:37	
Air Blank	0.000	22:38	
Subject Sample	# 2 0.178	22:39	
Breath Volume	2.378	22:39	
Air Blank	0.000	22:40	
Control Test	0.087	22:40	Control Outside Tolerance
Air Blank	0.000	22:41	

Data Download Date/Time: 01/13/2021 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: RICKETTS,J
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: GALLAGHER

DATE/TIME OF TEST: 12/31/2020 22:49:21

Diagnostics Check	OK	22:43
Air Blank	0.000	22:43
Control Test	0.082	22:43
Air Blank	0.000	22:44
Subject Sample # 1	0.170	22:44
Breath Volume	2.097	22:44
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.170	22:47
Breath Volume	2.042	22:47
Air Blank	0.000	22:48
Control Test	0.082	22:48
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 01/13/2021 16:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WALSH,SHANNON,L

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: WALSH,SHANNON

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30219080A3

DRY GAS EXPIRY DATE: 01/05/2022

VIOLATION CODE: DUI

ARREST TIME: 23:36

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: SHIFFLETT

DATE/TIME OF TEST: 12/06/2020 00:32:42

Diagnostics Check		OK	00:17	
Air Blank		0.000	00:18	
Control Test		0.081	00:18	
Air Blank		0.000	00:19	
Subject Sample	# 1	0.175	00:22	Volume Not Met
Breath Volume		0.066	00:22	
Air Blank		0.000	00:22	
Air Blank		0.000	00:24	
Subject Sample	# 2	0.192	00:27	No .020 agreement
Breath Volume		2.167	00:27	
Air Blank		0.000	00:27	
Air Blank		0.000	00:29	
Subject Sample	# 3	0.170	00:31	No .020 agreement
Breath Volume		2.046	00:31	
Air Blank		0.000	00:31	
Control Test		0.080	00:32	
Air Blank		0.000	00:32	
Diagnostics Check		OK	00:32	

Data Download Date/Time: 01/20/2021 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINDLEY,CARLTONJANE,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: LAMOREAUX,ERIC
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: OLAVARRIA

DATE/TIME OF TEST: 12/13/2020 03:23:41

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.133	03:19
Breath Volume	2.140	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.131	03:22
Breath Volume	2.968	03:22
Air Blank	0.000	03:22
Control Test	0.080	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 01/20/2021 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KIP,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PETERSON,KIP
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: FEAGLE

DATE/TIME OF TEST: 12/09/2020 03:30:03

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.125	03:25
Breath Volume	2.390	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.113	03:28
Breath Volume	1.289	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:30

Data Download Date/Time: 12/21/2020 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KIP,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MOORE,M
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: ESPOSITO

DATE/TIME OF TEST: 12/12/2020 23:26:39

Diagnostics Check		OK	23:10	
Air Blank		0.000	23:11	
Control Test		0.079	23:11	
Air Blank		0.000	23:11	
Subject Sample	# 1	0.355	23:14	No .020 agreement
Breath Volume		1.386	23:14	
Air Blank		0.000	23:15	
Air Blank		0.000	23:16	
Subject Sample	# 2	0.346	23:19	Volume Not Met
Breath Volume		0.632	23:19	
Air Blank		0.000	23:20	
Air Blank		0.000	23:22	
Subject Sample	# 3	0.330	23:24	No .020 agreement
Breath Volume		1.417	23:24	
Air Blank		0.000	23:25	
Control Test		0.075	23:26	
Air Blank		0.000	23:26	
Diagnostics Check		OK	23:26	

Data Download Date/Time: 12/21/2020 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KIP,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MOORE,MICAH
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: ESPOSITO

DATE/TIME OF TEST: 12/12/2020 23:38:49

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.077	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.335	23:32
Breath Volume	1.648	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.338	23:37
Breath Volume	1.332	23:37
Air Blank	0.000	23:37
Control Test	0.077	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 12/21/2020 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,JAMES,W
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BURGESS,JAMES
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 12/13/2020 00:07:07

Diagnostics Check	OK	23:59
Air Blank	0.000	00:00
Control Test	0.076	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.126	00:02
Breath Volume	2.484	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:04
Subject Sample # 2	0.126	00:05
Breath Volume	2.488	00:05
Air Blank	0.000	00:06
Control Test	0.077	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 12/21/2020 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KIP,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PETERSON,KIP
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 12/21/2020 17:53:50

Diagnostics Check	OK	17:47
Air Blank	0.000	17:48
Control Test	0.080	17:48
Air Blank	0.000	17:49
Subject Sample # 1	0.000	17:49
Breath Volume	1.535	17:49
Air Blank	0.000	17:50
Air Blank	0.000	17:52
Subject Sample # 2	0.000	17:52
Breath Volume	2.347	17:52
Air Blank	0.000	17:52
Control Test	0.079	17:53
Air Blank	0.000	17:53
Diagnostics Check	OK	17:53

Data Download Date/Time: 12/23/2020 08:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: NAVARRO

DATE/TIME OF TEST: 12/13/2020 01:57:28

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.080	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.170	01:52
Breath Volume	1.699	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.182	01:55
Breath Volume	1.828	01:55
Air Blank	0.000	01:56
Control Test	0.078	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 01/11/2021 18:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: RODRIGUEZ HERNANDEZ

DATE/TIME OF TEST: 12/13/2020 20:49:37

Diagnostics Check	OK	20:43
Air Blank	0.000	20:43
Control Test	0.080	20:43
Air Blank	0.000	20:44
Subject Sample # 1	0.182	20:45
Breath Volume	1.914	20:45
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.190	20:48
Breath Volume	2.054	20:48
Air Blank	0.000	20:48
Control Test	0.079	20:49
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 01/11/2021 18:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 12/18/2020 23:00:59

Diagnostics Check	OK	22:54
Air Blank	0.000	22:55
Control Test	0.081	22:55
Air Blank	0.000	22:55
Subject Sample # 1	0.194	22:56
Breath Volume	2.441	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.198	22:59
Breath Volume	2.777	22:59
Air Blank	0.000	23:00
Control Test	0.081	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 01/11/2021 18:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: MALDONADO ZAPATA

DATE/TIME OF TEST: 12/20/2020 02:19:22

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.079	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.139	02:14
Breath Volume	2.074	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.119	02:17
Breath Volume	1.429	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 01/11/2021 18:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: AZAN

DATE/TIME OF TEST: 12/02/2020 00:17:27

Diagnostics Check		OK	00:10
Air Blank		0.000	00:10
Control Test		0.080	00:11
Air Blank		0.000	00:11
Subject Sample	# 1	0.305	00:12
Breath Volume		3.144	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:14
Subject Sample	# 2	0.316	00:15
Breath Volume		5.062	00:15
Air Blank		0.000	00:16
Control Test		0.077	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 01/28/2021 07:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/03/2020 22:03:01

Diagnostics Check		OK	21:56
Air Blank		0.000	21:56
Control Test		0.079	21:56
Air Blank		0.000	21:57
Subject Sample	# 1	0.154	21:58
Breath Volume		2.707	21:58
Air Blank		0.000	21:59
Air Blank		0.000	22:00
Subject Sample	# 2	0.159	22:01
Breath Volume		2.441	22:01
Air Blank		0.000	22:02
Control Test		0.077	22:02
Air Blank		0.000	22:02
Diagnostics Check		OK	22:03

Data Download Date/Time: 01/28/2021 07:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 12/03/2020 22:33:22

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.078	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.139	22:28
Breath Volume	2.542	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:30
Subject Sample # 2	0.142	22:31
Breath Volume	3.144	22:31
Air Blank	0.000	22:32
Control Test	0.078	22:32
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 01/28/2021 07:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: JOAQUIN JOAQUIN

DATE/TIME OF TEST: 12/05/2020 01:24:09

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.079	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.219	01:19
Breath Volume	3.886	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:21
Subject Sample # 2	0.214	01:22
Breath Volume	2.714	01:22
Air Blank	0.000	01:23
Control Test	0.078	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 01/28/2021 07:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: WILLIAMS III

DATE/TIME OF TEST: 12/05/2020 05:35:40

Diagnostics Check	OK	05:29
Air Blank	0.000	05:29
Control Test	0.079	05:29
Air Blank	0.000	05:30
Subject Sample # 1	0.000	05:31
Breath Volume	5.929	05:31
Air Blank	0.000	05:31
Air Blank	0.000	05:33
Subject Sample # 2	0.000	05:34
Breath Volume	2.558	05:34
Air Blank	0.000	05:34
Control Test	0.079	05:35
Air Blank	0.000	05:35
Diagnostics Check	OK	05:35

Data Download Date/Time: 01/28/2021 07:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/06/2020 06:07:18

Diagnostics Check	OK	05:59	
Air Blank	0.000	06:00	
Control Test	0.079	06:00	
Air Blank	0.000	06:01	
Subject Sample	# 1 0.117	06:02	
Breath Volume	1.707	06:02	
Air Blank	0.000	06:02	
Air Blank	0.000	06:04	
Subject Sample	# 2	06:05	Slope Not Met
Breath Volume	0.828	06:05	
Air Blank	0.000	06:06	
Control Test	0.079	06:06	
Air Blank	0.000	06:07	
Diagnostics Check	OK	06:07	

Data Download Date/Time: 01/28/2021 07:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/06/2020 06:16:57

Diagnostics Check		OK	06:10
Air Blank		0.000	06:10
Control Test		0.079	06:10
Air Blank		0.000	06:11
Subject Sample	# 1	0.108	06:12
Breath Volume		1.312	06:12
Air Blank		0.000	06:12
Air Blank		0.000	06:14
Subject Sample	# 2	0.119	06:15
Breath Volume		1.464	06:15
Air Blank		0.000	06:16
Control Test		0.079	06:16
Air Blank		0.000	06:16
Diagnostics Check		OK	06:16

Data Download Date/Time: 01/28/2021 07:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: BELSER

DATE/TIME OF TEST: 12/07/2020 04:57:40

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.079	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.309	04:52
Breath Volume	2.550	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.309	04:55
Breath Volume	2.894	04:55
Air Blank	0.000	04:56
Control Test	0.078	04:57
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 01/28/2021 07:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: ROACH

DATE/TIME OF TEST: 12/08/2020 02:36:47

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:33	
Control Test	0.079	02:33	
Air Blank	0.000	02:33	
Subject Sample	# 1	02:35	Slope Not Met
Breath Volume	1.058	02:35	
Air Blank	0.000	02:35	
Control Test	0.079	02:36	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

Data Download Date/Time: 01/28/2021 07:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: ROACH

DATE/TIME OF TEST: 12/08/2020 03:02:45

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.079	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.174	02:57
Breath Volume		4.214	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.179	03:01
Breath Volume		4.410	03:01
Air Blank		0.000	03:01
Control Test		0.078	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 01/28/2021 07:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VANTREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: KENYON

DATE/TIME OF TEST: 12/09/2020 02:11:51

Diagnostics Check	OK	02:08	
Air Blank	0.000	02:08	
Control Test	0.079	02:08	
Air Blank	0.000	02:09	
Subject Sample	# 1	02:10	Subject Test Refused
Breath Volume	0.000	02:10	
Air Blank	0.000	02:10	
Control Test	0.079	02:11	
Air Blank	0.000	02:11	
Diagnostics Check	OK	02:11	

Data Download Date/Time: 01/28/2021 07:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDI
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:58

SUBJECT'S LAST NAME: DYAL

DATE/TIME OF TEST: 12/31/2020 17:44:02

Diagnostics Check	OK	17:40	
Air Blank	0.000	17:41	
Control Test	0.078	17:41	
Air Blank	0.000	17:42	
Subject Sample	# 1	17:42	Subject Test Refused
Breath Volume	0.000	17:42	
Air Blank	0.000	17:43	
Control Test	0.078	17:43	
Air Blank	0.000	17:43	
Diagnostics Check	OK	17:44	

Data Download Date/Time: 01/08/2021 21:55