Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,GREG,C
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: FERRELL,RYAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: COLLAZO

DATE/TIME OF TEST: 12/03/2020 15:29:21

Diagnostics Check		OK	15:22
Air Blank		0.000	15:23
Control Test		0.081	15:23
Air Blank		0.000	15:24
Subject Sample	# 1	0.000	15:25
Breath Volume		1.906	15:25
Air Blank		0.000	15:25
Air Blank		0.000	15:27
Subject Sample	# 2	0.000	15:27
Breath Volume		1.824	15:27
Air Blank		0.000	15:28
Control Test		0.080	15:28
Air Blank		0.000	15:29
Diagnostics Check		OK	15:29

Data Download Date/Time: 12/29/2020 17:11

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,COLE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MCDONALD,COLE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: RAUBER-MUSCARNERE

DATE/TIME OF TEST: 12/04/2020 01:05:15

Diagnostics Check		OK	00:58
Air Blank		0.000	00:59
Control Test		0.080	00:59
Air Blank		0.000	01:00
Subject Sample	# 1	0.134	01:00
Breath Volume		4.117	01:00
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.134	01:03
Breath Volume		4.035	01:03
Air Blank		0.000	01:04
Control Test		0.080	01:04
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 12/29/2020 17:11

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,COLE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MCDONALD,COLE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/05/2020 02:12:22

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.080	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.115	02:07
Breath Volume		3.136	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.109	02:10
Breath Volume		2.808	02:10
Air Blank		0.000	02:11
Control Test		0.080	02:11
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 12/29/2020 17:11

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBOLI,MELANIE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: TRIMBOLI,MELANIE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 12/10/2020 04:45:36

Diagnostics Check		OK	04:39
Air Blank		0.000	04:39
Control Test		0.081	04:40
Air Blank		0.000	04:40
Subject Sample	# 1	0.162	04:41
Breath Volume		3.312	04:41
Air Blank		0.000	04:41
Air Blank		0.000	04:43
Subject Sample	# 2	0.159	04:44
Breath Volume		2.710	04:44
Air Blank		0.000	04:44
Control Test		0.080	04:45
Air Blank		0.000	04:45
Diagnostics Check		OK	04:45

Data Download Date/Time: 12/29/2020 17:12

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,COLE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MCDONALD,COLE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: RUBIN

DATE/TIME OF TEST: 12/12/2020 04:07:48

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.080	04:02
Air Blank		0.000	04:02
Subject Sample	# 1	0.198	04:03
Breath Volume		1.710	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:05
Subject Sample	# 2	0.193	04:06
Breath Volume		1.156	04:06
Air Blank		0.000	04:06
Control Test		0.080	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 12/29/2020 17:12

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,COLE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MCDONALD,COLE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: MILLARD

DATE/TIME OF TEST: 12/21/2020 22:08:40

Diagnostics Check		OK	22:02
Air Blank		0.000	22:03
Control Test		0.080	22:03
Air Blank		0.000	22:03
Subject Sample	# 1	0.184	22:04
Breath Volume		1.964	22:04
Air Blank		0.000	22:04
Air Blank		0.000	22:06
Subject Sample	# 2	0.195	22:07
Breath Volume		1.597	22:07
Air Blank		0.000	22:07
Control Test		0.080	22:08
Air Blank		0.000	22:08
Diagnostics Check		OK	22:08

Data Download Date/Time: 12/29/2020 17:12

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,COLE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MCDONALD,COLE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/24/2020 23:23:49

Diagnostics Check		OK	23:17
Air Blank		0.000	23:17
Control Test		0.080	23:18
Air Blank		0.000	23:18
Subject Sample	# 1	0.167	23:19
Breath Volume		2.359	23:19
Air Blank		0.000	23:20
Air Blank		0.000	23:21
Subject Sample	# 2	0.169	23:22
Breath Volume		2.250	23:22
Air Blank		0.000	23:22
Control Test		0.080	23:23
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

Data Download Date/Time: 12/29/2020 17:12

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLCROFT,AMANDA,L
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HOLLCROFT,AMANDA
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: MODEEN

DATE/TIME OF TEST: 12/26/2020 12:22:32

Diagnostics Check		OK	12:16
Air Blank		0.000	12:16
Control Test		0.081	12:17
Air Blank		0.000	12:17
Subject Sample	# 1	0.000	12:18
Breath Volume		1.203	12:18
Air Blank		0.000	12:18
Air Blank		0.000	12:20
Subject Sample	# 2	0.000	12:21
Breath Volume		1.394	12:21
Air Blank		0.000	12:21
Control Test		0.082	12:21
Air Blank		0.000	12:22
Diagnostics Check		OK	12:22

Data Download Date/Time: 12/29/2020 17:12

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBOLI,MELANIE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: TRIMBOLI,MELANIE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 12/26/2020 21:23:35

Diagnostics Check		OK	21:17	
Air Blank		0.000	21:18	
Control Test		0.081	21:18	
Air Blank		0.000	21:19	
Subject Sample	# 1	0.000	21:19	
Breath Volume		2.316	21:19	
Air Blank		0.000	21:20	
Air Blank		0.000	21:22	
Subject Sample	# 2	0.000	21:22	
Breath Volume		2.125	21:22	
Air Blank			21:23	RFI Detect
Air Blank		0.000	21:23	

Data Download Date/Time: 12/29/2020 17:12

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBOLI,MELANIE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: TRIMBOLI,MELANIE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 12/26/2020 21:31:59

Diagnostics Check		OK	21:25
Air Blank		0.000	21:26
Control Test		0.080	21:26
Air Blank		0.000	21:27
Subject Sample	# 1	0.000	21:27
Breath Volume		2.261	21:27
Air Blank		0.000	21:28
Air Blank		0.000	21:30
Subject Sample	# 2	0.000	21:30
Breath Volume		2.722	21:30
Air Blank		0.000	21:31
Control Test		0.080	21:31
Air Blank		0.000	21:31
Diagnostics Check		OK	21:31

Data Download Date/Time: 12/29/2020 17:12

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LOGAN,JONATHAN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: KNAPP

DATE/TIME OF TEST: 12/27/2020 04:41:24

Diagnostics Check		OK	04:33
Air Blank		0.000	04:34
Control Test		0.082	04:34
Air Blank		0.000	04:34
Subject Sample	# 1	0.185	04:36
Breath Volume		1.621	04:36
Air Blank		0.000	04:37
Air Blank		0.000	04:39
Subject Sample	# 2	0.182	04:39
Breath Volume		1.355	04:39
Air Blank		0.000	04:40
Control Test		0.080	04:40
Air Blank		0.000	04:41
Diagnostics Check		OK	04:41

Data Download Date/Time: 12/29/2020 17:12

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,COLE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MCDONALD,COLE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: ROHR

DATE/TIME OF TEST: 12/29/2020 02:09:21

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.081	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.176	02:04
Breath Volume		1.500	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.173	02:07
Breath Volume		1.464	02:07
Air Blank		0.000	02:08
Control Test		0.080	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 12/29/2020 17:12

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT,KELLY,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: WITT,KELLY
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: DELVECCHIO

DATE/TIME OF TEST: 12/01/2020 23:42:08

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.080	23:36
Air Blank		0.000	23:36
Subject Sample	# 1	0.173	23:37
Breath Volume		5.210	23:37
Air Blank		0.000	23:38
Air Blank		0.000	23:39
Subject Sample	# 2	0.169	23:40
Breath Volume		3.808	23:40
Air Blank		0.000	23:41
Control Test		0.080	23:41
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 12/16/2020 14:17

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: MEIERS

DATE/TIME OF TEST: 12/02/2020 20:24:48

Diagnostics Check		OK	20:18
Air Blank		0.000	20:18
Control Test		0.080	20:19
Air Blank		0.000	20:19
Subject Sample	# 1	0.220	20:20
Breath Volume		3.378	20:20
Air Blank		0.000	20:20
Air Blank		0.000	20:22
Subject Sample	# 2	0.225	20:23
Breath Volume		3.421	20:23
Air Blank		0.000	20:23
Control Test		0.080	20:24
Air Blank		0.000	20:24
Diagnostics Check		OK	20:24

Data Download Date/Time: 12/16/2020 14:17

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANPELT,J,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HOYOS,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/04/2020 20:45:35

Diagnostics Check		OK	20:38
Air Blank		0.000	20:38
Control Test		0.080	20:39
Air Blank		0.000	20:39
Subject Sample	# 1	0.027	20:40
Breath Volume		1.621	20:40
Air Blank		0.000	20:41
Air Blank		0.000	20:43
Subject Sample	# 2	0.023	20:44
Breath Volume		1.585	20:44
Air Blank		0.000	20:44
Control Test		0.079	20:44
Air Blank		0.000	20:45
Diagnostics Check		OK	20:45

Data Download Date/Time: 12/16/2020 14:17

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,JOSEPH,T
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: WAY,JOSEPH
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: COPPOLA

DATE/TIME OF TEST: 12/05/2020 15:46:16

Diagnostics Check		OK	15:38
Air Blank		0.000	15:38
Control Test		0.081	15:38
Air Blank		0.000	15:39
Subject Sample	# 1	0.000	15:41
Breath Volume		1.511	15:41
Air Blank		0.000	15:41
Air Blank		0.000	15:43
Subject Sample	# 2	0.000	15:44
Breath Volume		1.496	15:44
Air Blank		0.000	15:45
Control Test		0.080	15:45
Air Blank		0.000	15:46
Diagnostics Check		OK	15:46

Data Download Date/Time: 12/16/2020 14:17

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT,KELLY,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: REYNOLDS,CHRISTIAN
ARRESTING OFFICER AGENCY: Lee Port Authority
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:13

SUBJECT'S LAST NAME: PETKO

DATE/TIME OF TEST: 12/06/2020 17:43:35

Diagnostics Check		OK	17:36
Air Blank		0.000	17:37
Control Test		0.080	17:37
Air Blank		0.000	17:37
Subject Sample	# 1	0.251	17:38
Breath Volume		2.410	17:38
Air Blank		0.000	17:39
Air Blank		0.000	17:41
Subject Sample	# 2	0.254	17:41
Breath Volume		2.320	17:41
Air Blank		0.000	17:42
Control Test		0.079	17:42
Air Blank		0.000	17:43
Diagnostics Check		OK	17:43

Data Download Date/Time: 12/16/2020 14:17

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DZUGAN,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: DZUGAN,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: CROWE

DATE/TIME OF TEST: 12/06/2020 22:25:24

Diagnostics Check		OK	22:18
Air Blank		0.000	22:19
Control Test		0.080	22:19
Air Blank		0.000	22:20
Subject Sample	# 1	0.232	22:20
Breath Volume		4.257	22:20
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.226	22:23
Breath Volume		3.621	22:23
Air Blank		0.000	22:24
Control Test		0.077	22:24
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 12/16/2020 14:17

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HARGIS,KYLE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: FERREIRA

DATE/TIME OF TEST: 12/08/2020 03:49:14

Diagnostics Check		OK	03:41
Air Blank		0.000	03:42
Control Test		0.080	03:42
Air Blank		0.000	03:43
Subject Sample	# 1	0.022	03:44
Breath Volume		1.546	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.021	03:47
Breath Volume		1.906	03:47
Air Blank		0.000	03:48
Control Test		0.080	03:48
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 12/16/2020 14:17

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: SCHROEDER

DATE/TIME OF TEST: 12/09/2020 01:07:46

Diagnostics Check		OK	01:00
Air Blank		0.000	01:00
Control Test		0.080	01:00
Air Blank		0.000	01:01
Subject Sample	# 1	0.209	01:02
Breath Volume		2.378	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.217	01:06
Breath Volume		2.898	01:06
Air Blank		0.000	01:06
Control Test		0.078	01:07
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 12/16/2020 14:17

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,LISA,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SINGLETON,C
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: MCGOWAN

DATE/TIME OF TEST: 12/11/2020 06:48:50

Diagnostics Check		OK	06:41
Air Blank		0.000	06:41
Control Test		0.080	06:42
Air Blank		0.000	06:42
Subject Sample	# 1	0.000	06:44
Breath Volume		1.410	06:44
Air Blank		0.000	06:44
Air Blank		0.000	06:46
Subject Sample	# 2	0.000	06:47
Breath Volume		1.625	06:47
Air Blank		0.000	06:47
Control Test		0.081	06:48
Air Blank		0.000	06:48
Diagnostics Check		OK	06:48

Data Download Date/Time: 12/16/2020 14:18

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: DEMARTINO

DATE/TIME OF TEST: 12/11/2020 23:27:52

Diagnostics Check		OK	23:21
Air Blank		0.000	23:21
Control Test		0.080	23:21
Air Blank		0.000	23:22
Subject Sample	# 1	0.152	23:23
Breath Volume		5.160	23:23
Air Blank		0.000	23:23
Air Blank		0.000	23:25
Subject Sample	# 2	0.146	23:26
Breath Volume		5.960	23:26
Air Blank		0.000	23:26
Control Test		0.079	23:27
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 12/16/2020 14:18

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 12/12/2020 16:23:25

Diagnostics Check		OK	16:20	
Air Blank		0.000	16:20	
Control Test		0.079	16:21	
Air Blank		0.000	16:21	
Subject Sample	# 1		16:22	Subject Test Refused
Breath Volume		0.000	16:22	
Air Blank		0.000	16:22	
Control Test		0.080	16:22	
Air Blank		0.000	16:23	
Diagnostics Check		OK	16:23	

Data Download Date/Time: 12/16/2020 14:18

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEMEN,LOGAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEMEN,LOGAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: TOLEDO RAMIREZ

DATE/TIME OF TEST: 12/13/2020 00:03:56

Diagnostics Check		OK	23:57
Air Blank		0.000	23:57
Control Test		0.080	23:57
Air Blank		0.000	23:58
Subject Sample	# 2	0.188	23:59
Breath Volume		2.367	23:59
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.182	00:02
Breath Volume		1.855	00:02
Air Blank		0.000	00:03
Control Test		0.079	00:03
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 12/16/2020 14:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEINER,RYAN,T
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: BEINER,RYAN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: GIAMBRA

DATE/TIME OF TEST: 12/13/2020 22:06:22

Diagnostics Check		OK	21:59
Air Blank		0.000	22:00
Control Test		0.080	22:00
Air Blank		0.000	22:01
Subject Sample	# 1	0.237	22:01
Breath Volume		1.210	22:01
Air Blank		0.000	22:02
Air Blank		0.000	22:03
Subject Sample	# 2	0.233	22:04
Breath Volume		1.566	22:04
Air Blank		0.000	22:05
Control Test		0.079	22:05
Air Blank		0.000	22:06
Diagnostics Check		OK	22:06

Data Download Date/Time: 12/16/2020 14:18

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAK,KYLE,F
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LEE,JASMINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 12/14/2020 20:21:46

Diagnostics Check		OK	20:13	
Air Blank		0.000	20:13	
Control Test		0.080	20:14	
Air Blank		0.000	20:14	
Subject Sample	# 1		20:17	No Sample Provided
Breath Volume		0.000	20:17	
Air Blank		0.000	20:18	
Air Blank		0.000	20:20	
Subject Sample	# 2		20:20	Subject Test Refused
Breath Volume		0.000	20:20	
Air Blank		0.000	20:20	
Control Test		0.079	20:21	
Air Blank		0.000	20:21	
Diagnostics Check		OK	20:21	

Data Download Date/Time: 12/16/2020 14:18

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONSALL,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUBOIS,MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: HANNAH

DATE/TIME OF TEST: 12/15/2020 21:47:42

Diagnostics Check		OK	21:41
Air Blank		0.000	21:41
Control Test		0.080	21:42
Air Blank		0.000	21:42
Subject Sample	# 1	0.214	21:43
Breath Volume		2.179	21:43
Air Blank		0.000	21:43
Air Blank		0.000	21:45
Subject Sample	# 2	0.202	21:46
Breath Volume		1.847	21:46
Air Blank		0.000	21:46
Control Test		0.080	21:47
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

Data Download Date/Time: 12/16/2020 14:18

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONSALL,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BONSALL,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: PORTAL GIMENEZ

DATE/TIME OF TEST: 12/16/2020 00:30:59

Diagnostics Check		OK	00:21
Air Blank		0.000	00:22
Control Test		0.080	00:22
Air Blank		0.000	00:22
Subject Sample	# 1	0.161	00:23
Breath Volume		2.261	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.136	00:26
Breath Volume		1.820	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:28
Subject Sample	# 3	0.137	00:29
Breath Volume		1.882	00:29
Air Blank		0.000	00:30
Control Test		0.079	00:30
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 12/16/2020 14:18

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALLUCCA,MARIO,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FALLUCCA,MARIO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 12/14/2020 21:53:53

Diagnostics Check		OK	21:47
Air Blank		0.000	21:47
Control Test		0.077	21:47
Air Blank		0.000	21:48
Subject Sample	# 1	0.213	21:49
Breath Volume		3.781	21:49
Air Blank		0.000	21:49
Air Blank		0.000	21:51
Subject Sample	# 2	0.215	21:52
Breath Volume		3.722	21:52
Air Blank		0.000	21:52
Control Test		0.075	21:53
Air Blank		0.000	21:53
Diagnostics Check		OK	21:53

Data Download Date/Time: 02/17/2021 09:37

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHT,JAMES,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: VAUGHT,JAMES
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: LOPEZ LEON

DATE/TIME OF TEST: 12/15/2020 00:11:16

Diagnostics Check		OK	00:04
Air Blank		0.000	00:05
Control Test		0.077	00:05
Air Blank		0.000	00:06
Subject Sample	# 1	0.169	00:06
Breath Volume		2.589	00:06
Air Blank		0.000	00:07
Air Blank		0.000	00:08
Subject Sample	# 2	0.174	00:09
Breath Volume		3.468	00:09
Air Blank		0.000	00:10
Control Test		0.075	00:10
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 02/17/2021 09:37

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,RYAN,
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: VAUGHN,RYAN
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 10:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:53

SUBJECT'S LAST NAME: KANE

DATE/TIME OF TEST: 12/17/2020 11:20:49

Diagnostics Check		OK	11:14
Air Blank		0.000	11:15
Control Test		0.078	11:15
Air Blank		0.000	11:15
Subject Sample	# 1	0.000	11:16
Breath Volume		5.035	11:16
Air Blank		0.000	11:16
Air Blank		0.000	11:18
Subject Sample	# 2	0.000	11:19
Breath Volume		4.664	11:19
Air Blank		0.000	11:19
Control Test		0.077	11:20
Air Blank		0.000	11:20
Diagnostics Check		OK	11:20

Data Download Date/Time: 02/17/2021 09:37

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: DAVIS,DONALD
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/17/2020 22:41:52

Diagnostics Check		OK	22:38	
Air Blank		0.000	22:38	
Control Test		0.078	22:38	
Air Blank		0.000	22:39	
Subject Sample	# 1		22:40	Subject Test Refused
Breath Volume		0.000	22:40	
Air Blank		0.000	22:40	
Control Test		0.078	22:41	
Air Blank		0.000	22:41	
Diagnostics Check		OK	22:41	

Data Download Date/Time: 02/17/2021 09:37

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JIMENEZ,WILLIAM,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JIMENEZ,WILIAM
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: WILKINSON

DATE/TIME OF TEST: 12/18/2020 22:48:30

Diagnostics Check		OK	22:41
Air Blank		0.000	22:42
Control Test		0.078	22:42
Air Blank		0.000	22:42
Subject Sample	# 1	0.124	22:43
Breath Volume		2.167	22:43
Air Blank		0.000	22:44
Air Blank		0.000	22:45
Subject Sample	# 2	0.120	22:46
Breath Volume		2.015	22:46
Air Blank		0.000	22:47
Control Test		0.075	22:47
Air Blank		0.000	22:48
Diagnostics Check		OK	22:48

Data Download Date/Time: 02/17/2021 09:37

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JASON,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BROWN,JASON
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 12:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: DANGIOLILLO

DATE/TIME OF TEST: 12/19/2020 04:12:52

Diagnostics Check		OK	04:06
Air Blank		0.000	04:06
Control Test		0.079	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.000	04:08
Breath Volume		3.027	04:08
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.000	04:11
Breath Volume		1.343	04:11
Air Blank		0.000	04:11
Control Test		0.079	04:12
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 02/17/2021 09:38

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALLUCCA,MARIO,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FALLUCCA,MARIO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: MARSH

DATE/TIME OF TEST: 12/19/2020 21:23:07

Diagnostics Check		OK	21:16
Air Blank		0.000	21:17
Control Test		0.078	21:17
Air Blank		0.000	21:18
Subject Sample	# 1	0.000	21:18
Breath Volume		4.941	21:18
Air Blank		0.000	21:19
Air Blank		0.000	21:21
Subject Sample	# 2	0.000	21:21
Breath Volume		4.804	21:21
Air Blank		0.000	21:22
Control Test		0.076	21:22
Air Blank		0.000	21:23
Diagnostics Check		OK	21:23

Data Download Date/Time: 02/17/2021 09:38

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,MELISSA,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GONZALEZ,JOEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 12/19/2020 21:43:49

Diagnostics Check	OK	21:43	
Air Blank		21:43	RFI Detect
Air Blank		21:43	RFI Detect

Data Download Date/Time: 02/17/2021 09:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALLUCCA,MARIO,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FALLUCCA,MARIO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 12/19/2020 21:52:47

Diagnostics Check		OK	21:45
Air Blank		0.000	21:46
Control Test		0.078	21:46
Air Blank		0.000	21:47
Subject Sample	# 1	0.162	21:47
Breath Volume		1.730	21:47
Air Blank		0.000	21:48
Air Blank		0.000	21:50
Subject Sample	# 2	0.167	21:51
Breath Volume		1.640	21:51
Air Blank		0.000	21:51
Control Test		0.077	21:52
Air Blank		0.000	21:52
Diagnostics Check		OK	21:52

Data Download Date/Time: 02/17/2021 09:38

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMBERLIN,KEVIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TOMBERLIN,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 12/21/2020 15:48:55

Diagnostics Check		OK	15:45	
Air Blank		0.000	15:45	
Control Test		0.078	15:46	
Air Blank		0.000	15:46	
Subject Sample	# 1	0.295	15:47	
Breath Volume		5.125	15:47	
Air Blank			15:47	RFI Detect
Air Blank		0.000	15:47	
Control Test		0.077	15:48	
Air Blank		0.000	15:48	
Diagnostics Check		OK	15:48	

Data Download Date/Time: 02/17/2021 09:38

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMBERLIN,KEVIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TOMBERLIN,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 12/21/2020 15:57:00

Diagnostics Check		OK	15:50
Air Blank		0.000	15:51
Control Test		0.076	15:51
Air Blank		0.000	15:51
Subject Sample	# 1	0.284	15:52
Breath Volume		3.812	15:52
Air Blank		0.000	15:53
Air Blank		0.000	15:54
Subject Sample	# 2	0.286	15:55
Breath Volume		3.679	15:55
Air Blank		0.000	15:56
Control Test		0.075	15:56
Air Blank		0.000	15:56
Diagnostics Check		OK	15:57

Data Download Date/Time: 02/17/2021 09:38

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 12/23/2020 04:41:46

Diagnostics Check		OK	04:35
Air Blank		0.000	04:35
Control Test		0.078	04:36
Air Blank		0.000	04:36
Subject Sample	# 1	0.141	04:37
Breath Volume		1.839	04:37
Air Blank		0.000	04:37
Air Blank		0.000	04:39
Subject Sample	# 2	0.136	04:40
Breath Volume		1.203	04:40
Air Blank		0.000	04:40
Control Test		0.078	04:41
Air Blank		0.000	04:41
Diagnostics Check		OK	04:41

Data Download Date/Time: 02/17/2021 09:38

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CHASTAIN,KEVIN
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 12:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:58

SUBJECT'S LAST NAME: MORALES VALASQUEZ

DATE/TIME OF TEST: 12/25/2020 14:41:53

Diagnostics Check		OK	14:39	
Air Blank		0.000	14:39	
Control Test		0.078	14:39	
Air Blank		0.000	14:40	
Subject Sample	# 1		14:40	Subject Test Refused
Breath Volume		0.000	14:40	
Air Blank		0.000	14:40	
Control Test		0.079	14:41	
Air Blank		0.000	14:41	
Diagnostics Check		OK	14:41	

Data Download Date/Time: 02/17/2021 09:38

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/26/2020 02:13:33

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:02	
Control Test		0.078	02:02	
Air Blank		0.000	02:03	
Subject Sample	# 1	0.151	02:06	Volume Not Met
Breath Volume		0.070	02:06	
Air Blank		0.000	02:06	
Air Blank		0.000	02:08	
Subject Sample	# 2	0.123	02:11	Volume Not Met
Breath Volume		0.500	02:11	
Air Blank		0.000	02:12	
Control Test		0.077	02:12	
Air Blank		0.000	02:13	
Diagnostics Check		OK	02:13	

Data Download Date/Time: 02/17/2021 09:38

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/26/2020 02:27:53

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:16	
Control Test		0.077	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 1	0.141	02:20	Volume Not Met
Breath Volume		0.441	02:20	
Air Blank		0.000	02:21	
Air Blank		0.000	02:22	
Subject Sample	# 2	0.137	02:26	Volume Not Met
Breath Volume		0.523	02:26	
Air Blank		0.000	02:26	
Control Test		0.075	02:27	
Air Blank		0.000	02:27	
Diagnostics Check		OK	02:27	

Data Download Date/Time: 02/17/2021 09:38

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MASON,SAM
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: WARSHEFSKI

DATE/TIME OF TEST: 12/26/2020 03:59:47

Diagnostics Check		OK	03:53
Air Blank		0.000	03:53
Control Test		0.078	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.221	03:55
Breath Volume		1.523	03:55
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.223	03:58
Breath Volume		4.066	03:58
Air Blank		0.000	03:58
Control Test		0.078	03:59
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 02/17/2021 09:38

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORRAO,MICHAEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 11:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: LUVIANO

DATE/TIME OF TEST: 12/27/2020 00:13:41

Diagnostics Check		OK	00:07
Air Blank		0.000	00:07
Control Test		0.078	00:08
Air Blank		0.000	00:08
Subject Sample	# 1	0.133	00:09
Breath Volume		2.335	00:09
Air Blank		0.000	00:09
Air Blank		0.000	00:11
Subject Sample	# 2	0.137	00:12
Breath Volume		2.222	00:12
Air Blank		0.000	00:12
Control Test		0.078	00:13
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 02/17/2021 09:38

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUERO,JOYRICK,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BACON,DEDDRICK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: DE LEON CHUN

DATE/TIME OF TEST: 12/27/2020 04:37:26

Diagnostics Check		OK	04:30
Air Blank		0.000	04:31
Control Test		0.078	04:31
Air Blank		0.000	04:32
Subject Sample	# 1	0.287	04:32
Breath Volume		2.824	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.279	04:35
Breath Volume		1.828	04:35
Air Blank		0.000	04:36
Control Test		0.075	04:36
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 02/17/2021 09:38

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: ALVARADO MENDOZA

DATE/TIME OF TEST: 12/31/2020 01:16:06

Diagnostics Check		OK	01:08
Air Blank		0.000	01:09
Control Test		0.077	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.131	01:11
Breath Volume		1.843	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:13
Subject Sample	# 2	0.133	01:14
Breath Volume		2.433	01:14
Air Blank		0.000	01:15
Control Test		0.077	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 02/17/2021 09:38

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSBY,C,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A2
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/02/2020 19:40:11

Diagnostics Check		OK	19:32
Air Blank		0.000	19:32
Control Test		0.079	19:33
Air Blank		0.000	19:33
Subject Sample	# 1	0.207	19:34
Breath Volume		1.687	19:34
Air Blank		0.000	19:35
Air Blank		0.000	19:37
Subject Sample	# 2	0.209	19:38
Breath Volume		1.562	19:38
Air Blank		0.000	19:39
Control Test		0.079	19:39
Air Blank		0.000	19:40
Diagnostics Check		OK	19:40

Data Download Date/Time: 12/09/2020 11:49

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSBY,C,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A2
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 12/04/2020 03:17:54

Diagnostics Check		OK	03:11
Air Blank		0.000	03:12
Control Test		0.080	03:12
Air Blank		0.000	03:13
Subject Sample	# 1	0.119	03:13
Breath Volume		2.589	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:15
Subject Sample	# 2	0.120	03:16
Breath Volume		2.222	03:16
Air Blank		0.000	03:16
Control Test		0.079	03:17
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 12/09/2020 11:49

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSBY,C,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A2
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: WICKMAN

DATE/TIME OF TEST: 12/05/2020 01:13:59

Diagnostics Check		OK	01:07
Air Blank		0.000	01:08
Control Test		0.079	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.014	01:09
Breath Volume		2.566	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.014	01:12
Breath Volume		2.746	01:12
Air Blank		0.000	01:13
Control Test		0.079	01:13
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 12/09/2020 11:49

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSBY,C,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A2
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 12/05/2020 04:47:37

Diagnostics Check		OK	04:44	
Air Blank		0.000	04:45	
Control Test		0.080	04:45	
Air Blank		0.000	04:46	
Subject Sample	# 1		04:46	Subject Test Refused
Breath Volume		0.000	04:46	
Air Blank		0.000	04:46	
Control Test		0.078	04:46	
Air Blank		0.000	04:47	
Diagnostics Check		OK	04:47	

Data Download Date/Time: 12/09/2020 11:49

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A2
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MAFFEI

DATE/TIME OF TEST: 12/06/2020 02:51:19

Diagnostics Check		OK	02:45
Air Blank		0.000	02:45
Control Test		0.080	02:46
Air Blank		0.000	02:46
Subject Sample	# 1	0.103	02:46
Breath Volume		2.523	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.098	02:49
Breath Volume		3.621	02:49
Air Blank		0.000	02:50
Control Test		0.080	02:50
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 12/09/2020 11:49

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOROZOV,S
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A2
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: OBRIEN

DATE/TIME OF TEST: 12/06/2020 05:34:47

Diagnostics Check		OK	05:28
Air Blank		0.000	05:29
Control Test		0.080	05:29
Air Blank		0.000	05:29
Subject Sample	# 1	0.154	05:30
Breath Volume		2.496	05:30
Air Blank		0.000	05:30
Air Blank		0.000	05:32
Subject Sample	# 2	0.142	05:33
Breath Volume		1.394	05:33
Air Blank		0.000	05:33
Control Test		0.080	05:34
Air Blank		0.000	05:34
Diagnostics Check		OK	05:34

Data Download Date/Time: 12/09/2020 11:49

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A2
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 12/06/2020 20:52:52

Diagnostics Check		OK	20:45
Air Blank		0.000	20:45
Control Test		0.080	20:46
Air Blank		0.000	20:46
Subject Sample	# 1	0.199	20:47
Breath Volume		1.777	20:47
Air Blank		0.000	20:48
Air Blank		0.000	20:49
Subject Sample	# 2	0.206	20:51
Breath Volume		2.152	20:51
Air Blank		0.000	20:51
Control Test		0.079	20:52
Air Blank		0.000	20:52
Diagnostics Check		OK	20:52

Data Download Date/Time: 12/09/2020 11:49

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A2
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: BARNES JR

DATE/TIME OF TEST: 12/06/2020 21:32:58

Diagnostics Check		OK	21:27
Air Blank		0.000	21:27
Control Test		0.080	21:27
Air Blank		0.000	21:28
Subject Sample	# 1	0.000	21:28
Breath Volume		1.535	21:28
Air Blank		0.000	21:29
Air Blank		0.000	21:31
Subject Sample	# 2	0.000	21:31
Breath Volume		1.679	21:31
Air Blank		0.000	21:32
Control Test		0.080	21:32
Air Blank		0.000	21:32
Diagnostics Check		OK	21:32

Data Download Date/Time: 12/09/2020 11:49

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A2
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: PEROTTI

DATE/TIME OF TEST: 12/07/2020 01:47:13

Diagnostics Check		OK	01:44	
Air Blank		0.000	01:44	
Control Test		0.080	01:45	
Air Blank		0.000	01:45	
Subject Sample	# 1		01:45	Subject Test Refused
Breath Volume		0.000	01:45	
Air Blank		0.000	01:46	
Control Test		0.080	01:46	
Air Blank		0.000	01:47	
Diagnostics Check		OK	01:47	

Data Download Date/Time: 12/09/2020 11:50

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A2
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: WHITTINGTON

DATE/TIME OF TEST: 12/07/2020 05:09:31

Diagnostics Check		OK	05:06	
Air Blank		0.000	05:07	
Control Test		0.080	05:07	
Air Blank		0.000	05:07	
Subject Sample	# 1		05:08	Subject Test Refused
Breath Volume		0.000	05:08	
Air Blank		0.000	05:08	
Control Test		0.079	05:08	
Air Blank		0.000	05:09	
Diagnostics Check		OK	05:09	

Data Download Date/Time: 12/09/2020 11:50

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JONES,A
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A2
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 12/07/2020 20:16:52

Diagnostics Check		OK	20:09
Air Blank		0.000	20:09
Control Test		0.080	20:10
Air Blank		0.000	20:10
Subject Sample	# 1	0.197	20:11
Breath Volume		2.320	20:11
Air Blank		0.000	20:12
Air Blank		0.000	20:14
Subject Sample	# 2	0.202	20:15
Breath Volume		2.527	20:15
Air Blank		0.000	20:15
Control Test		0.079	20:16
Air Blank		0.000	20:16
Diagnostics Check		OK	20:16

Data Download Date/Time: 12/09/2020 11:50

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A2
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/07/2020 23:50:54

Diagnostics Check		OK	23:47	
Air Blank		0.000	23:48	
Control Test		0.079	23:48	
Air Blank		0.000	23:49	
Subject Sample	# 1		23:49	Subject Test Refused
Breath Volume		0.000	23:49	
Air Blank		0.000	23:49	
Control Test		0.080	23:50	
Air Blank		0.000	23:50	
Diagnostics Check		OK	23:50	

Data Download Date/Time: 12/09/2020 11:56

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A2
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: RUTHERFORD

DATE/TIME OF TEST: 12/09/2020 02:35:15

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.080	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.203	02:30
Breath Volume		3.031	02:30
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.193	02:33
Breath Volume		3.183	02:33
Air Blank		0.000	02:34
Control Test		0.080	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 12/09/2020 11:56

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A2
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: LLANES PEREZ

DATE/TIME OF TEST: 12/09/2020 04:09:16

Diagnostics Check		OK	04:02
Air Blank		0.000	04:02
Control Test		0.080	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.125	04:04
Breath Volume		2.136	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.120	04:07
Breath Volume		2.164	04:07
Air Blank		0.000	04:08
Control Test		0.080	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 12/09/2020 11:56

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROJAS,GABRIEL,L
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: SALAZAR,LAURENT
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 12/12/2020 00:12:23

Diagnostics Check		OK	00:06
Air Blank		0.000	00:06
Control Test		0.079	00:07
Air Blank		0.000	00:07
Subject Sample	# 1	0.011	00:08
Breath Volume		2.292	00:08
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.010	00:10
Breath Volume		2.285	00:10
Air Blank		0.000	00:11
Control Test		0.080	00:11
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 01/06/2021 01:39

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARREA,S,
BREATH TEST OPERATOR AGENCY: HIALEAH PD
ARRESTING OFFICER: LARREA,S
ARRESTING OFFICER AGENCY: HIALEAH PD
LAST AGENCY INSPECTION: 11/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 11:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: CUELLO

DATE/TIME OF TEST: 12/24/2020 12:11:47

Diagnostics Check		OK	12:08	
Air Blank		0.000	12:08	
Control Test		0.077	12:09	
Air Blank		0.000	12:09	
Subject Sample	# 1		12:10	Subject Test Refused
Breath Volume		0.000	12:10	
Air Blank		0.000	12:10	
Control Test		0.077	12:11	
Air Blank		0.000	12:11	
Diagnostics Check		OK	12:11	

Data Download Date/Time: 12/31/2020 14:05

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANCO,BRYAN,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: ROMERO,CARLOS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 12/12/2020 22:49:01

Diagnostics Check		OK	22:40
Air Blank		0.000	22:41
Control Test		0.078	22:41
Air Blank		0.000	22:41
Subject Sample	# 1	0.102	22:43
Breath Volume		1.707	22:43
Air Blank		0.000	22:43
Air Blank		0.000	22:45
Subject Sample	# 2	0.094	22:47
Breath Volume		1.519	22:47
Air Blank		0.000	22:48
Control Test		0.077	22:48
Air Blank		0.000	22:48
Diagnostics Check		OK	22:49

Data Download Date/Time: 12/29/2020 11:43

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANCO,BRYAN,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: ROMERO,CARLOS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: ROSALES RODRIGUEZ

DATE/TIME OF TEST: 12/27/2020 23:42:43

Diagnostics Check		OK	23:30	
Air Blank		0.000	23:31	
Control Test		0.079	23:31	
Air Blank		0.000	23:32	
Subject Sample	# 1	0.000	23:35	Volume Not Met
Breath Volume		0.000	23:35	
Air Blank		0.000	23:35	
Air Blank		0.000	23:37	
Subject Sample	# 2	0.187	23:40	Volume Not Met
Breath Volume		0.105	23:40	
Air Blank		0.000	23:41	
Control Test		0.079	23:42	
Air Blank		0.000	23:42	
Diagnostics Check		OK	23:42	

Data Download Date/Time: 12/29/2020 11:43

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANCO,BRYAN,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: ROMERO,CARLOS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: ROSALES RODRIGUEZ

DATE/TIME OF TEST: 12/27/2020 23:58:35

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:45	
Control Test		0.077	23:45	
Air Blank		0.000	23:46	
Subject Sample	# 1	0.190	23:49	Volume Not Met
Breath Volume		0.378	23:49	
Air Blank		0.000	23:50	
Air Blank		0.000	23:52	
Subject Sample	# 2	0.183	23:53	
Breath Volume		1.390	23:53	
Air Blank		0.000	23:54	
Air Blank		0.000	23:56	
Subject Sample	# 3	0.183	23:56	
Breath Volume		1.617	23:56	
Air Blank		0.000	23:57	
Control Test		0.074	23:57	Control Outside Tolerance
Air Blank		0.000	23:58	

Data Download Date/Time: 12/29/2020 11:43

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANCO,BRYAN,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: ROMERO,CARLOS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: ROSALES RODRIGUEZ

DATE/TIME OF TEST: 12/28/2020 00:24:23

Diagnostics Check		OK	00:15
Air Blank		0.000	00:16
Control Test		0.079	00:16
Air Blank		0.000	00:17
Subject Sample	# 1	0.200	00:17
Breath Volume		1.378	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:19
Subject Sample	# 2	0.183	00:22
Breath Volume		1.863	00:22
Air Blank		0.000	00:23
Control Test		0.078	00:23
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 12/29/2020 11:43

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTO,JOHN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTO,JOHN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/09/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MATOS

DATE/TIME OF TEST: 12/10/2020 01:46:11

Diagnostics Check		OK	01:36	
Air Blank		0.000	01:36	
Control Test		0.081	01:37	
Air Blank		0.000	01:37	
Subject Sample	# 1		01:40	No Sample Provided
Breath Volume		0.000	01:40	
Air Blank		0.000	01:41	
Air Blank		0.000	01:43	
Subject Sample	# 2		01:44	Subject Test Refused
Breath Volume		0.000	01:44	
Air Blank		0.000	01:45	
Control Test		0.080	01:45	
Air Blank		0.000	01:46	
Diagnostics Check		OK	01:46	

Data Download Date/Time: 12/18/2020 19:00

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTO,JOHN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTO,JOHN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: ALTMAN

DATE/TIME OF TEST: 12/20/2020 00:09:50

Diagnostics Check	OK	00:09	
Air Blank		00:09	RFI Detect
Air Blank	0.000	00:09	

Data Download Date/Time: 01/19/2021 18:54

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTO,JOHN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTO,JOHN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: ALTMAN

DATE/TIME OF TEST: 12/20/2020 00:15:30

Diagnostics Check		OK	00:11	
Air Blank		0.000	00:12	
Control Test		0.081	00:12	
Air Blank		0.000	00:13	
Subject Sample	# 1		00:14	RFI Detect
Breath Volume		0.000	00:14	
Air Blank			00:15	RFI Detect
Air Blank		0.000	00:15	

Data Download Date/Time: 01/19/2021 18:54

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 12/10/2020 01:59:12

Diagnostics Check		OK	01:51
Air Blank		0.000	01:52
Control Test		0.081	01:52
Air Blank		0.000	01:52
Subject Sample	# 1	0.000	01:54
Breath Volume		2.382	01:54
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.000	01:57
Breath Volume		1.386	01:57
Air Blank		0.000	01:58
Control Test		0.081	01:58
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: VAZQUEZ GARCIA

DATE/TIME OF TEST: 12/10/2020 05:35:33

Diagnostics Check		OK	05:29
Air Blank		0.000	05:29
Control Test		0.081	05:29
Air Blank		0.000	05:30
Subject Sample	# 1	0.133	05:31
Breath Volume		1.335	05:31
Air Blank		0.000	05:31
Air Blank		0.000	05:33
Subject Sample	# 2	0.129	05:34
Breath Volume		1.121	05:34
Air Blank		0.000	05:34
Control Test		0.080	05:34
Air Blank		0.000	05:35
Diagnostics Check		OK	05:35

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: BUTSCHER

DATE/TIME OF TEST: 12/10/2020 21:33:58

Diagnostics Check		OK	21:27
Air Blank		0.000	21:28
Control Test		0.081	21:28
Air Blank		0.000	21:29
Subject Sample	# 1	0.091	21:29
Breath Volume		1.535	21:29
Air Blank		0.000	21:30
Air Blank		0.000	21:32
Subject Sample	# 2	0.084	21:32
Breath Volume		1.804	21:32
Air Blank		0.000	21:33
Control Test		0.080	21:33
Air Blank		0.000	21:33
Diagnostics Check		OK	21:33

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FERGUSON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: FLORES ORTIS

DATE/TIME OF TEST: 12/11/2020 00:25:20

Diagnostics Check		OK	00:08	
Air Blank		0.000	00:08	
Control Test		0.080	00:09	
Air Blank		0.000	00:09	
Subject Sample	# 1	0.096	00:12	Volume Not Met
Breath Volume		0.171	00:12	
Air Blank		0.000	00:13	
Air Blank		0.000	00:15	
Subject Sample	# 2	0.100	00:18	
Breath Volume		1.425	00:18	
Air Blank		0.000	00:18	
Air Blank		0.000	00:20	
Subject Sample	# 3	0.095	00:23	Volume Not Met
Breath Volume		0.183	00:23	
Air Blank		0.000	00:24	
Control Test		0.080	00:24	
Air Blank		0.000	00:25	
Diagnostics Check		OK	00:25	

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: OCCIDE

DATE/TIME OF TEST: 12/11/2020 01:21:21

Diagnostics Check		OK	01:14
Air Blank		0.000	01:15
Control Test		0.079	01:15
Air Blank		0.000	01:16
Subject Sample	# 1	0.042	01:17
Breath Volume		1.539	01:17
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.041	01:19
Breath Volume		1.527	01:19
Air Blank		0.000	01:20
Control Test		0.080	01:20
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 12/11/2020 23:32:51

Diagnostics Check		OK	23:26
Air Blank		0.000	23:27
Control Test		0.080	23:27
Air Blank		0.000	23:28
Subject Sample	# 1	0.104	23:28
Breath Volume		1.667	23:28
Air Blank		0.000	23:29
Air Blank		0.000	23:30
Subject Sample	# 2	0.111	23:31
Breath Volume		2.238	23:31
Air Blank		0.000	23:31
Control Test		0.080	23:32
Air Blank		0.000	23:32
Diagnostics Check		OK	23:32

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: MOHAMED

DATE/TIME OF TEST: 12/12/2020 02:27:35

Diagnostics Check		OK	02:22	
Air Blank		0.000	02:22	
Control Test		0.080	02:22	
Air Blank		0.000	02:23	
Subject Sample	# 1		02:26	Subject Test Refused
Breath Volume		0.000	02:26	
Air Blank		0.000	02:26	
Control Test		0.081	02:26	
Air Blank		0.000	02:27	
Diagnostics Check		OK	02:27	

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SULEYMANOV

DATE/TIME OF TEST: 12/12/2020 03:46:59

Diagnostics Check		OK	03:37
Air Blank		0.000	03:37
Control Test		0.080	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.135	03:40
Breath Volume		1.398	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:42
Subject Sample	# 2	0.128	03:45
Breath Volume		1.382	03:45
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DAVID,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,DAVID
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:27

SUBJECT'S LAST NAME: RUSNAK

DATE/TIME OF TEST: 12/12/2020 06:55:00

Diagnostics Check		OK	06:47
Air Blank		0.000	06:48
Control Test		0.080	06:48
Air Blank		0.000	06:49
Subject Sample	# 1	0.153	06:49
Breath Volume		3.898	06:49
Air Blank		0.000	06:50
Air Blank		0.000	06:52
Subject Sample	# 2	0.158	06:53
Breath Volume		4.136	06:53
Air Blank		0.000	06:54
Control Test		0.079	06:54
Air Blank		0.000	06:54
Diagnostics Check		OK	06:55

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: FAKESS

DATE/TIME OF TEST: 12/12/2020 20:14:49

Diagnostics Check		OK	20:07
Air Blank		0.000	20:08
Control Test		0.080	20:08
Air Blank		0.000	20:09
Subject Sample	# 1	0.013	20:09
Breath Volume		2.417	20:09
Air Blank		0.000	20:10
Air Blank		0.000	20:12
Subject Sample	# 2	0.013	20:13
Breath Volume		1.230	20:13
Air Blank		0.000	20:13
Control Test		0.080	20:14
Air Blank		0.000	20:14
Diagnostics Check		OK	20:14

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: CORDON

DATE/TIME OF TEST: 12/13/2020 00:05:43

Diagnostics Check		OK	23:59
Air Blank		0.000	23:59
Control Test		0.080	00:00
Air Blank		0.000	00:00
Subject Sample	# 1	0.131	00:01
Breath Volume		2.890	00:01
Air Blank		0.000	00:01
Air Blank		0.000	00:03
Subject Sample	# 2	0.125	00:04
Breath Volume		2.285	00:04
Air Blank		0.000	00:04
Control Test		0.080	00:05
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: GEIGER

DATE/TIME OF TEST: 12/13/2020 01:40:26

Diagnostics Check		OK	01:33
Air Blank		0.000	01:34
Control Test		0.080	01:34
Air Blank		0.000	01:35
Subject Sample	# 1	0.104	01:35
Breath Volume		3.484	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.107	01:38
Breath Volume		3.238	01:38
Air Blank		0.000	01:39
Control Test		0.079	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 12/13/2020 04:07:06

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.080	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.118	04:02
Breath Volume		1.414	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.119	04:05
Breath Volume		1.238	04:05
Air Blank		0.000	04:06
Control Test		0.079	04:06
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GALICIA,A
ARRESTING OFFICER AGENCY: Macdill Air Force Ba
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: ZAVALA SALAS

DATE/TIME OF TEST: 12/13/2020 07:27:45

Diagnostics Check		OK	07:21
Air Blank		0.000	07:21
Control Test		0.080	07:22
Air Blank		0.000	07:22
Subject Sample	# 1	0.188	07:22
Breath Volume		3.339	07:22
Air Blank		0.000	07:23
Air Blank		0.000	07:25
Subject Sample	# 2	0.187	07:26
Breath Volume		3.394	07:26
Air Blank		0.000	07:26
Control Test		0.080	07:27
Air Blank		0.000	07:27
Diagnostics Check		OK	07:27

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: TORRES GARCIA

DATE/TIME OF TEST: 12/13/2020 22:33:31

Diagnostics Check		OK	22:26
Air Blank		0.000	22:26
Control Test		0.080	22:27
Air Blank		0.000	22:27
Subject Sample	# 1	0.194	22:28
Breath Volume		3.378	22:28
Air Blank		0.000	22:29
Air Blank		0.000	22:31
Subject Sample	# 2	0.205	22:31
Breath Volume		3.519	22:31
Air Blank		0.000	22:32
Control Test		0.080	22:32
Air Blank		0.000	22:33
Diagnostics Check		OK	22:33

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 12/14/2020 03:32:34

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.080	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.195	03:27
Breath Volume		3.140	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.195	03:31
Breath Volume		2.386	03:31
Air Blank		0.000	03:31
Control Test		0.079	03:31
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: GUMMERMAN

DATE/TIME OF TEST: 12/14/2020 07:31:24

Diagnostics Check		OK	07:25
Air Blank		0.000	07:25
Control Test		0.080	07:25
Air Blank		0.000	07:26
Subject Sample	# 1	0.152	07:27
Breath Volume		3.554	07:27
Air Blank		0.000	07:27
Air Blank		0.000	07:29
Subject Sample	# 2	0.154	07:29
Breath Volume		2.308	07:29
Air Blank		0.000	07:30
Control Test		0.080	07:30
Air Blank		0.000	07:31
Diagnostics Check		OK	07:31

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FERGUSON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: GERARDO

DATE/TIME OF TEST: 12/15/2020 04:13:58

Diagnostics Check		OK	04:06
Air Blank		0.000	04:07
Control Test		0.080	04:07
Air Blank		0.000	04:07
Subject Sample	# 1	0.179	04:09
Breath Volume		1.664	04:09
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.185	04:12
Breath Volume		1.750	04:12
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: GRIMALDI III

DATE/TIME OF TEST: 12/16/2020 23:22:59

Diagnostics Check		OK	23:17
Air Blank		0.000	23:17
Control Test		0.080	23:17
Air Blank		0.000	23:18
Subject Sample	# 1	0.000	23:18
Breath Volume		2.839	23:18
Air Blank		0.000	23:19
Air Blank		0.000	23:21
Subject Sample	# 2	0.000	23:21
Breath Volume		3.011	23:21
Air Blank		0.000	23:22
Control Test		0.081	23:22
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: SALGADO

DATE/TIME OF TEST: 12/17/2020 00:26:19

Diagnostics Check		OK	00:19
Air Blank		0.000	00:20
Control Test		0.080	00:20
Air Blank		0.000	00:20
Subject Sample	# 1	0.175	00:21
Breath Volume		2.921	00:21
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.173	00:24
Breath Volume		2.562	00:24
Air Blank		0.000	00:25
Control Test		0.080	00:25
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: ACKERMAN

DATE/TIME OF TEST: 12/17/2020 02:54:41

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.080	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.059	02:50
Breath Volume		3.054	02:50
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.054	02:53
Breath Volume		1.414	02:53
Air Blank		0.000	02:53
Control Test		0.080	02:54
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: PAYSENO

DATE/TIME OF TEST: 12/18/2020 02:54:56

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:41	
Control Test		0.081	02:41	
Air Blank		0.000	02:41	
Subject Sample	# 1	0.253	02:43	
Breath Volume		1.238	02:43	
Air Blank		0.000	02:44	
Air Blank		0.000	02:46	
Subject Sample	# 2	0.240	02:49	Slope Not Level
Breath Volume		1.242	02:49	
Air Blank		0.000	02:50	
Air Blank		0.000	02:51	
Subject Sample	# 3	0.241	02:53	
Breath Volume		1.265	02:53	
Air Blank		0.000	02:53	
Control Test		0.078	02:54	
Air Blank		0.000	02:54	
Diagnostics Check		OK	02:54	

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: MEEK

DATE/TIME OF TEST: 12/18/2020 03:55:01

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.080	03:47
Air Blank		0.000	03:48
Subject Sample	# 1	0.271	03:50
Breath Volume		4.406	03:50
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.271	03:53
Breath Volume		3.894	03:53
Air Blank		0.000	03:54
Control Test		0.079	03:54
Air Blank		0.000	03:54
Diagnostics Check		OK	03:55

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HARRIS,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: TAZI

DATE/TIME OF TEST: 12/18/2020 21:13:15

Diagnostics Check		OK	21:06
Air Blank		0.000	21:06
Control Test		0.081	21:07
Air Blank		0.000	21:07
Subject Sample	# 1	0.217	21:08
Breath Volume		2.527	21:08
Air Blank		0.000	21:09
Air Blank		0.000	21:11
Subject Sample	# 2	0.222	21:11
Breath Volume		2.582	21:11
Air Blank		0.000	21:12
Control Test		0.081	21:12
Air Blank		0.000	21:13
Diagnostics Check		OK	21:13

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: BERGLUND

DATE/TIME OF TEST: 12/18/2020 22:08:31

Diagnostics Check		OK	22:00
Air Blank		0.000	22:00
Control Test		0.081	22:00
Air Blank		0.000	22:01
Subject Sample	# 1	0.177	22:03
Breath Volume		1.343	22:03
Air Blank		0.000	22:04
Air Blank		0.000	22:06
Subject Sample	# 2	0.191	22:06
Breath Volume		1.644	22:06
Air Blank		0.000	22:07
Control Test		0.079	22:07
Air Blank		0.000	22:08
Diagnostics Check		OK	22:08

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RICHBOURG,K
ARRESTING OFFICER AGENCY: Tampa Airport PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: MAURO

DATE/TIME OF TEST: 12/19/2020 01:55:25

Diagnostics Check		OK	01:40	
Air Blank		0.000	01:40	
Control Test		0.081	01:41	
Air Blank		0.000	01:41	
Subject Sample	# 1		01:44	No Sample Provided
Breath Volume		0.000	01:44	
Air Blank		0.000	01:45	
Air Blank		0.000	01:47	
Subject Sample	# 2	0.123	01:49	No .020 agreement
Breath Volume		1.347	01:49	
Air Blank		0.000	01:50	
Air Blank		0.000	01:52	
Subject Sample	# 3	0.151	01:53	No .020 agreement
Breath Volume		2.402	01:53	
Air Blank		0.000	01:54	
Control Test		0.080	01:54	
Air Blank		0.000	01:55	
Diagnostics Check		OK	01:55	

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PELL,H
ARRESTING OFFICER AGENCY: Tampa Airport PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: MAURO

DATE/TIME OF TEST: 12/19/2020 02:06:37

Diagnostics Check		OK	01:58
Air Blank		0.000	01:58
Control Test		0.080	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.145	02:01
Breath Volume		1.714	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:03
Subject Sample	# 2	0.148	02:05
Breath Volume		2.480	02:05
Air Blank		0.000	02:05
Control Test		0.080	02:06
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VEAZQUEZ,B
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 12/19/2020 05:56:16

Diagnostics Check		OK	05:49
Air Blank		0.000	05:49
Control Test		0.080	05:50
Air Blank		0.000	05:50
Subject Sample	# 1	0.137	05:51
Breath Volume		2.156	05:51
Air Blank		0.000	05:52
Air Blank		0.000	05:54
Subject Sample	# 2	0.141	05:54
Breath Volume		1.824	05:54
Air Blank		0.000	05:55
Control Test		0.080	05:55
Air Blank		0.000	05:56
Diagnostics Check		OK	05:56

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:52

SUBJECT'S LAST NAME: BALDWIN

DATE/TIME OF TEST: 12/19/2020 07:20:57

Diagnostics Check		OK	07:12
Air Blank		0.000	07:12
Control Test		0.080	07:13
Air Blank		0.000	07:13
Subject Sample	# 1	0.151	07:14
Breath Volume		1.875	07:14
Air Blank		0.000	07:15
Air Blank		0.000	07:17
Subject Sample	# 2	0.155	07:19
Breath Volume		1.371	07:19
Air Blank		0.000	07:20
Control Test		0.080	07:20
Air Blank		0.000	07:20
Diagnostics Check		OK	07:20

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:02

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 12/19/2020 07:31:09

Diagnostics Check		OK	07:24
Air Blank		0.000	07:25
Control Test		0.080	07:25
Air Blank		0.000	07:25
Subject Sample	# 1	0.022	07:26
Breath Volume		1.824	07:26
Air Blank		0.000	07:27
Air Blank		0.000	07:29
Subject Sample	# 2	0.021	07:29
Breath Volume		2.847	07:29
Air Blank		0.000	07:30
Control Test		0.080	07:30
Air Blank		0.000	07:31
Diagnostics Check		OK	07:31

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: NAJERA SANTIAGO

DATE/TIME OF TEST: 12/19/2020 21:11:12

Diagnostics Check		OK	21:05	
Air Blank		0.000	21:05	
Control Test		0.080	21:05	
Air Blank		0.000	21:06	
Subject Sample	# 1	0.250	21:07	
Breath Volume		3.492	21:07	
Air Blank		0.000	21:08	
Air Blank			21:09	RFI Detect
Air Blank		0.000	21:10	
Control Test		0.081	21:10	
Air Blank		0.000	21:11	
Diagnostics Check		OK	21:11	

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: NAJERA SANTIAGO

DATE/TIME OF TEST: 12/19/2020 21:20:45

Diagnostics Check		OK	21:14
Air Blank		0.000	21:14
Control Test		0.080	21:15
Air Blank		0.000	21:15
Subject Sample	# 1	0.251	21:16
Breath Volume		3.019	21:16
Air Blank		0.000	21:16
Air Blank		0.000	21:18
Subject Sample	# 2	0.253	21:19
Breath Volume		3.187	21:19
Air Blank		0.000	21:19
Control Test		0.080	21:20
Air Blank		0.000	21:20
Diagnostics Check		OK	21:20

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/19/2020 23:24:49

Diagnostics Check		OK	23:11	
Air Blank		0.000	23:12	
Control Test		0.079	23:12	
Air Blank		0.000	23:13	
Subject Sample	# 1	0.150	23:16	Volume Not Met
Breath Volume		0.367	23:16	
Air Blank		0.000	23:16	
Air Blank		0.000	23:18	
Subject Sample	# 2	0.169	23:19	
Breath Volume		1.796	23:19	
Air Blank		0.000	23:19	
Air Blank		0.000	23:21	
Subject Sample	# 3	0.166	23:23	
Breath Volume		2.027	23:23	
Air Blank		0.000	23:23	
Control Test		0.080	23:24	
Air Blank		0.000	23:24	
Diagnostics Check		OK	23:24	

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FERGUSON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 12/20/2020 01:34:54

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:21	
Control Test		0.080	01:22	
Air Blank		0.000	01:22	
Subject Sample	# 1	0.127	01:25	Volume Not Met
Breath Volume		0.742	01:25	
Air Blank		0.000	01:26	
Air Blank		0.000	01:28	
Subject Sample	# 2	0.126	01:29	
Breath Volume		1.589	01:29	
Air Blank		0.000	01:29	
Air Blank		0.000	01:31	
Subject Sample	# 3	0.124	01:33	
Breath Volume		1.593	01:33	
Air Blank		0.000	01:33	
Control Test		0.079	01:34	
Air Blank		0.000	01:34	
Diagnostics Check		OK	01:34	

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 12/20/2020 02:50:49

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.080	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.132	02:46
Breath Volume		3.191	02:46
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.136	02:49
Breath Volume		2.925	02:49
Air Blank		0.000	02:49
Control Test		0.079	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FERGUSON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: MARTINEZ MADRID

DATE/TIME OF TEST: 12/20/2020 05:10:01

Diagnostics Check		OK	05:00
Air Blank		0.000	05:01
Control Test		0.079	05:01
Air Blank		0.000	05:02
Subject Sample	# 1	0.173	05:04
Breath Volume		1.906	05:04
Air Blank		0.000	05:04
Air Blank		0.000	05:06
Subject Sample	# 2	0.168	05:08
Breath Volume		1.828	05:08
Air Blank		0.000	05:09
Control Test		0.079	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:10

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KLENOFSKY,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: CATES

DATE/TIME OF TEST: 12/20/2020 05:35:31

Diagnostics Check		OK	05:28
Air Blank		0.000	05:29
Control Test		0.079	05:29
Air Blank		0.000	05:29
Subject Sample	# 1	0.126	05:30
Breath Volume		1.117	05:30
Air Blank		0.000	05:31
Air Blank		0.000	05:33
Subject Sample	# 2	0.126	05:33
Breath Volume		1.347	05:33
Air Blank		0.000	05:34
Control Test		0.079	05:34
Air Blank		0.000	05:35
Diagnostics Check		OK	05:35

Data Download Date/Time: 01/11/2021 19:15

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: BAZZI

DATE/TIME OF TEST: 12/20/2020 05:50:05

Diagnostics Check		OK	05:42
Air Blank		0.000	05:43
Control Test		0.079	05:43
Air Blank		0.000	05:44
Subject Sample	# 1	0.126	05:45
Breath Volume		1.468	05:45
Air Blank		0.000	05:46
Air Blank		0.000	05:47
Subject Sample	# 2	0.128	05:48
Breath Volume		3.042	05:48
Air Blank		0.000	05:49
Control Test		0.079	05:49
Air Blank		0.000	05:50
Diagnostics Check		OK	05:50

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: MATTSON

DATE/TIME OF TEST: 12/20/2020 19:40:39

Diagnostics Check		OK	19:31
Air Blank		0.000	19:31
Control Test		0.079	19:32
Air Blank		0.000	19:32
Subject Sample	# 1	0.175	19:33
Breath Volume		1.515	19:33
Air Blank		0.000	19:34
Air Blank		0.000	19:36
Subject Sample	# 2	0.176	19:39
Breath Volume		1.781	19:39
Air Blank		0.000	19:39
Control Test		0.080	19:40
Air Blank		0.000	19:40
Diagnostics Check		OK	19:40

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KLENOFSKY,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: TORRES MUNOZ

DATE/TIME OF TEST: 12/20/2020 23:55:13

Diagnostics Check		OK	23:48
Air Blank		0.000	23:49
Control Test		0.080	23:49
Air Blank		0.000	23:50
Subject Sample	# 1	0.174	23:50
Breath Volume		2.664	23:50
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 2	0.163	23:53
Breath Volume		1.957	23:53
Air Blank		0.000	23:54
Control Test		0.079	23:54
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: ACEVEDO FELIX

DATE/TIME OF TEST: 12/21/2020 06:23:28

Diagnostics Check		OK	06:16
Air Blank		0.000	06:16
Control Test		0.080	06:16
Air Blank		0.000	06:17
Subject Sample	# 1	0.129	06:19
Breath Volume		1.394	06:19
Air Blank		0.000	06:19
Air Blank		0.000	06:21
Subject Sample	# 2	0.136	06:21
Breath Volume		1.796	06:21
Air Blank		0.000	06:22
Control Test		0.080	06:22
Air Blank		0.000	06:23
Diagnostics Check		OK	06:23

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 12/22/2020 03:21:25

Diagnostics Check		OK	03:15
Air Blank		0.000	03:15
Control Test		0.081	03:15
Air Blank		0.000	03:16
Subject Sample	# 1	0.000	03:17
Breath Volume		2.960	03:17
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.000	03:20
Breath Volume		3.457	03:20
Air Blank		0.000	03:20
Control Test		0.081	03:20
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: FARLOW

DATE/TIME OF TEST: 12/22/2020 04:37:51

Diagnostics Check		OK	04:30
Air Blank		0.000	04:30
Control Test		0.080	04:31
Air Blank		0.000	04:31
Subject Sample	# 1	0.214	04:32
Breath Volume		1.855	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:34
Subject Sample	# 2	0.214	04:36
Breath Volume		2.296	04:36
Air Blank		0.000	04:36
Control Test		0.079	04:37
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: O'DONNELL,T 154
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:26

SUBJECT'S LAST NAME: COARD WINGO

DATE/TIME OF TEST: 12/22/2020 07:55:20

Diagnostics Check		OK	07:48
Air Blank		0.000	07:49
Control Test		0.081	07:49
Air Blank		0.000	07:50
Subject Sample	# 1	0.016	07:50
Breath Volume		1.781	07:50
Air Blank		0.000	07:51
Air Blank		0.000	07:53
Subject Sample	# 2	0.015	07:53
Breath Volume		1.539	07:53
Air Blank		0.000	07:54
Control Test		0.080	07:54
Air Blank		0.000	07:55
Diagnostics Check		OK	07:55

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: SERRANO ISIDRE

DATE/TIME OF TEST: 12/23/2020 04:21:45

Diagnostics Check		OK	04:18	
Air Blank		0.000	04:18	
Control Test		0.081	04:18	
Air Blank		0.000	04:19	
Subject Sample	# 1		04:20	Subject Test Refused
Breath Volume		0.000	04:20	
Air Blank		0.000	04:20	
Control Test		0.081	04:21	
Air Blank		0.000	04:21	
Diagnostics Check		OK	04:21	

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: DUDLEY

DATE/TIME OF TEST: 12/24/2020 00:06:51

Diagnostics Check		OK	00:02	
Air Blank		0.000	00:03	
Control Test		0.080	00:03	
Air Blank		0.000	00:03	
Subject Sample	# 1		00:05	Subject Test Refused
Breath Volume		0.000	00:05	
Air Blank		0.000	00:05	
Control Test		0.082	00:06	
Air Blank		0.000	00:06	
Diagnostics Check		OK	00:06	

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 12/24/2020 05:33:31

Diagnostics Check		OK	05:26
Air Blank		0.000	05:26
Control Test		0.080	05:27
Air Blank		0.000	05:27
Subject Sample	# 1	0.092	05:28
Breath Volume		2.800	05:28
Air Blank		0.000	05:29
Air Blank		0.000	05:31
Subject Sample	# 2	0.091	05:31
Breath Volume		3.277	05:31
Air Blank		0.000	05:32
Control Test		0.080	05:32
Air Blank		0.000	05:33
Diagnostics Check		OK	05:33

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HAIRSTON,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: GARCIA TRUJILLO

DATE/TIME OF TEST: 12/24/2020 06:48:12

Diagnostics Check		OK	06:41
Air Blank		0.000	06:41
Control Test		0.081	06:41
Air Blank		0.000	06:42
Subject Sample	# 1	0.000	06:43
Breath Volume		1.566	06:43
Air Blank		0.000	06:43
Air Blank		0.000	06:45
Subject Sample	# 2	0.000	06:46
Breath Volume		1.722	06:46
Air Blank		0.000	06:47
Control Test		0.080	06:47
Air Blank		0.000	06:48
Diagnostics Check		OK	06:48

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 12/24/2020 21:25:15

Diagnostics Check		OK	21:18
Air Blank		0.000	21:18
Control Test		0.080	21:19
Air Blank		0.000	21:19
Subject Sample	# 1	0.000	21:20
Breath Volume		1.238	21:20
Air Blank		0.000	21:21
Air Blank		0.000	21:23
Subject Sample	# 2	0.000	21:23
Breath Volume		1.273	21:23
Air Blank		0.000	21:24
Control Test		0.080	21:24
Air Blank		0.000	21:25
Diagnostics Check		OK	21:25

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: QUINTERO TAMAYO

DATE/TIME OF TEST: 12/24/2020 22:09:00

Diagnostics Check		OK	22:02
Air Blank		0.000	22:03
Control Test		0.080	22:03
Air Blank		0.000	22:03
Subject Sample	# 1	0.069	22:04
Breath Volume		1.785	22:04
Air Blank		0.000	22:05
Air Blank		0.000	22:07
Subject Sample	# 2	0.066	22:07
Breath Volume		1.421	22:07
Air Blank		0.000	22:08
Control Test		0.080	22:08
Air Blank		0.000	22:08
Diagnostics Check		OK	22:09

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GILLOTTE,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/24/2020 22:28:26

Diagnostics Check		OK	22:21
Air Blank		0.000	22:21
Control Test		0.080	22:22
Air Blank		0.000	22:22
Subject Sample	# 1	0.149	22:24
Breath Volume		2.046	22:24
Air Blank		0.000	22:24
Air Blank		0.000	22:26
Subject Sample	# 2	0.142	22:26
Breath Volume		1.503	22:26
Air Blank		0.000	22:27
Control Test		0.079	22:27
Air Blank		0.000	22:28
Diagnostics Check		OK	22:28

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 12/25/2020 01:00:39

Diagnostics Check		OK	00:54
Air Blank		0.000	00:55
Control Test		0.080	00:55
Air Blank		0.000	00:55
Subject Sample	# 1	0.026	00:56
Breath Volume		2.687	00:56
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.024	00:59
Breath Volume		2.246	00:59
Air Blank		0.000	00:59
Control Test		0.080	01:00
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: JUZTZUY

DATE/TIME OF TEST: 12/25/2020 04:44:04

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.081	04:38
Air Blank		0.000	04:38
Subject Sample	# 1	0.218	04:39
Breath Volume		1.871	04:39
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.212	04:42
Breath Volume		1.613	04:42
Air Blank		0.000	04:43
Control Test		0.079	04:43
Air Blank		0.000	04:43
Diagnostics Check		OK	04:44

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: DESPAW

DATE/TIME OF TEST: 12/25/2020 20:20:17

Diagnostics Check		OK	20:13
Air Blank		0.000	20:14
Control Test		0.080	20:14
Air Blank		0.000	20:15
Subject Sample	# 1	0.206	20:15
Breath Volume		2.847	20:15
Air Blank		0.000	20:16
Air Blank		0.000	20:18
Subject Sample	# 2	0.204	20:18
Breath Volume		2.164	20:18
Air Blank		0.000	20:19
Control Test		0.081	20:19
Air Blank		0.000	20:20
Diagnostics Check		OK	20:20

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: AVILA

DATE/TIME OF TEST: 12/25/2020 23:28:20

Diagnostics Check		OK	23:21
Air Blank		0.000	23:22
Control Test		0.082	23:22
Air Blank		0.000	23:23
Subject Sample	# 1	0.190	23:23
Breath Volume		1.324	23:23
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.178	23:26
Breath Volume		1.414	23:26
Air Blank		0.000	23:27
Control Test		0.081	23:27
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: DE LEON MEJIA

DATE/TIME OF TEST: 12/25/2020 23:52:11

Diagnostics Check		OK	23:45
Air Blank		0.000	23:46
Control Test		0.080	23:46
Air Blank		0.000	23:46
Subject Sample	# 1	0.167	23:47
Breath Volume		2.046	23:47
Air Blank		0.000	23:48
Air Blank		0.000	23:50
Subject Sample	# 2	0.168	23:50
Breath Volume		1.664	23:50
Air Blank		0.000	23:51
Control Test		0.083	23:51
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: BALLANCE

DATE/TIME OF TEST: 12/26/2020 15:02:43

Diagnostics Check		OK	14:56
Air Blank		0.000	14:56
Control Test		0.081	14:57
Air Blank		0.000	14:57
Subject Sample	# 1	0.224	14:58
Breath Volume		2.332	14:58
Air Blank		0.000	14:58
Air Blank		0.000	15:00
Subject Sample	# 2	0.219	15:01
Breath Volume		3.085	15:01
Air Blank		0.000	15:01
Control Test		0.081	15:02
Air Blank		0.000	15:02
Diagnostics Check		OK	15:02

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: PORTILLO VALLE

DATE/TIME OF TEST: 12/26/2020 22:15:59

Diagnostics Check		OK	22:07
Air Blank		0.000	22:07
Control Test		0.082	22:08
Air Blank		0.000	22:08
Subject Sample	# 1	0.144	22:11
Breath Volume		1.843	22:11
Air Blank		0.000	22:12
Air Blank		0.000	22:13
Subject Sample	# 2	0.136	22:14
Breath Volume		1.554	22:14
Air Blank		0.000	22:15
Control Test		0.081	22:15
Air Blank		0.000	22:15
Diagnostics Check		OK	22:15

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 12/27/2020 19:36:51

Diagnostics Check		OK	19:29
Air Blank		0.000	19:30
Control Test		0.081	19:30
Air Blank		0.000	19:31
Subject Sample	# 1	0.229	19:32
Breath Volume		1.386	19:32
Air Blank		0.000	19:32
Air Blank		0.000	19:34
Subject Sample	# 2	0.230	19:35
Breath Volume		2.117	19:35
Air Blank		0.000	19:35
Control Test		0.080	19:36
Air Blank		0.000	19:36
Diagnostics Check		OK	19:36

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MAILEY

DATE/TIME OF TEST: 12/27/2020 22:45:06

Diagnostics Check		OK	22:38
Air Blank		0.000	22:38
Control Test		0.081	22:39
Air Blank		0.000	22:39
Subject Sample	# 1	0.194	22:40
Breath Volume		1.808	22:40
Air Blank		0.000	22:41
Air Blank		0.000	22:43
Subject Sample	# 2	0.188	22:43
Breath Volume		1.527	22:43
Air Blank		0.000	22:44
Control Test		0.081	22:44
Air Blank		0.000	22:45
Diagnostics Check		OK	22:45

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KRYGIER,B
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: RODRIGUEZ GONZALEZ

DATE/TIME OF TEST: 12/28/2020 00:23:45

Diagnostics Check		OK	00:19	
Air Blank		0.000	00:20	
Control Test		0.081	00:20	
Air Blank		0.000	00:20	
Subject Sample	# 1		00:22	Subject Test Refused
Breath Volume		0.000	00:22	
Air Blank		0.000	00:22	
Control Test		0.081	00:23	
Air Blank		0.000	00:23	
Diagnostics Check		OK	00:23	

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: SALAZAR-SANCHEZ

DATE/TIME OF TEST: 12/29/2020 22:36:27

Diagnostics Check		OK	22:29
Air Blank		0.000	22:29
Control Test		0.080	22:30
Air Blank		0.000	22:30
Subject Sample	# 1	0.137	22:31
Breath Volume		2.398	22:31
Air Blank		0.000	22:32
Air Blank		0.000	22:34
Subject Sample	# 2	0.143	22:34
Breath Volume		2.125	22:34
Air Blank		0.000	22:35
Control Test		0.079	22:35
Air Blank		0.000	22:36
Diagnostics Check		OK	22:36

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: TAMAYO

DATE/TIME OF TEST: 12/30/2020 04:32:01

Diagnostics Check		OK	04:28	
Air Blank		0.000	04:29	
Control Test		0.081	04:29	
Air Blank		0.000	04:30	
Subject Sample	# 1		04:30	Subject Test Refused
Breath Volume		0.000	04:30	
Air Blank		0.000	04:31	
Control Test		0.081	04:31	
Air Blank		0.000	04:31	
Diagnostics Check		OK	04:32	

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: MORANCHEN

DATE/TIME OF TEST: 12/30/2020 06:11:18

Diagnostics Check		OK	06:04
Air Blank		0.000	06:05
Control Test		0.080	06:05
Air Blank		0.000	06:06
Subject Sample	# 1	0.236	06:06
Breath Volume		1.609	06:06
Air Blank		0.000	06:07
Air Blank		0.000	06:09
Subject Sample	# 2	0.238	06:09
Breath Volume		2.410	06:09
Air Blank		0.000	06:10
Control Test		0.080	06:10
Air Blank		0.000	06:11
Diagnostics Check		OK	06:11

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: RICO ROZAS

DATE/TIME OF TEST: 12/30/2020 19:21:29

Diagnostics Check		OK	19:14
Air Blank		0.000	19:15
Control Test		0.080	19:15
Air Blank		0.000	19:16
Subject Sample	# 1	0.000	19:17
Breath Volume		1.554	19:17
Air Blank		0.000	19:17
Air Blank		0.000	19:19
Subject Sample	# 2	0.000	19:20
Breath Volume		1.492	19:20
Air Blank		0.000	19:20
Control Test		0.080	19:20
Air Blank		0.000	19:21
Diagnostics Check		OK	19:21

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 12/30/2020 21:13:29

Diagnostics Check		OK	21:07
Air Blank		0.000	21:07
Control Test		0.081	21:08
Air Blank		0.000	21:08
Subject Sample	# 1	0.000	21:09
Breath Volume		2.429	21:09
Air Blank		0.000	21:09
Air Blank		0.000	21:11
Subject Sample	# 2	0.000	21:12
Breath Volume		2.375	21:12
Air Blank		0.000	21:12
Control Test		0.080	21:12
Air Blank		0.000	21:13
Diagnostics Check		OK	21:13

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: NAVA MARCOS

DATE/TIME OF TEST: 12/31/2020 01:05:25

Diagnostics Check		OK	00:58
Air Blank		0.000	00:59
Control Test		0.080	00:59
Air Blank		0.000	00:59
Subject Sample	# 1	0.194	01:00
Breath Volume		2.722	01:00
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.185	01:03
Breath Volume		1.183	01:03
Air Blank		0.000	01:04
Control Test		0.079	01:04
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: GONZALEZ-PEREZ

DATE/TIME OF TEST: 12/31/2020 19:09:10

Diagnostics Check		OK	19:00	
Air Blank		0.000	19:01	
Control Test		0.080	19:01	
Air Blank		0.000	19:02	
Subject Sample	# 1		19:05	No Sample Provided
Breath Volume		0.000	19:05	
Air Blank		0.000	19:05	
Air Blank		0.000	19:07	
Subject Sample	# 2		19:07	Subject Test Refused
Breath Volume		0.000	19:07	
Air Blank		0.000	19:08	
Control Test		0.080	19:08	
Air Blank		0.000	19:09	
Diagnostics Check		OK	19:09	

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 12/31/2020 20:00:38

Diagnostics Check		OK	19:53
Air Blank		0.000	19:53
Control Test		0.080	19:54
Air Blank		0.000	19:54
Subject Sample	# 1	0.184	19:55
Breath Volume		2.062	19:55
Air Blank		0.000	19:56
Air Blank		0.000	19:58
Subject Sample	# 2	0.183	19:59
Breath Volume		1.917	19:59
Air Blank		0.000	19:59
Control Test		0.079	20:00
Air Blank		0.000	20:00
Diagnostics Check		OK	20:00

Data Download Date/Time: 01/11/2021 19:16

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: PROCHASKA

DATE/TIME OF TEST: 12/04/2020 03:50:22

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.080	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.189	03:45
Breath Volume		1.632	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.194	03:48
Breath Volume		1.273	03:48
Air Blank		0.000	03:49
Control Test		0.080	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 12/09/2020 18:47

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: DE VASTO JR

DATE/TIME OF TEST: 12/06/2020 02:36:02

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.081	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.112	02:31
Breath Volume		1.210	02:31
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.101	02:34
Breath Volume		2.121	02:34
Air Blank		0.000	02:35
Control Test		0.080	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:36

Data Download Date/Time: 12/09/2020 18:47

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: COTTO JR

DATE/TIME OF TEST: 12/10/2020 02:30:09

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.081	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.169	02:25
Breath Volume		3.296	02:25
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.169	02:28
Breath Volume		3.941	02:28
Air Blank		0.000	02:29
Control Test		0.080	02:29
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 01/28/2021 08:07

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: TOMAS

DATE/TIME OF TEST: 12/11/2020 02:08:42

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.080	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.159	02:03
Breath Volume		2.234	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.161	02:07
Breath Volume		2.757	02:07
Air Blank		0.000	02:07
Control Test		0.079	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 01/28/2021 08:07

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MCKENZIE

DATE/TIME OF TEST: 12/11/2020 02:24:22

Diagnostics Check		OK	02:16
Air Blank		0.000	02:17
Control Test		0.079	02:17
Air Blank		0.000	02:18
Subject Sample	# 1	0.182	02:19
Breath Volume		1.402	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.202	02:22
Breath Volume		2.390	02:22
Air Blank		0.000	02:23
Control Test		0.079	02:23
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 01/28/2021 08:07

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 12/12/2020 05:17:04

Diagnostics Check		OK	05:10
Air Blank		0.000	05:11
Control Test		0.081	05:11
Air Blank		0.000	05:11
Subject Sample	# 1	0.122	05:12
Breath Volume		1.179	05:12
Air Blank		0.000	05:13
Air Blank		0.000	05:14
Subject Sample	# 2	0.114	05:15
Breath Volume		1.187	05:15
Air Blank		0.000	05:16
Control Test		0.080	05:16
Air Blank		0.000	05:16
Diagnostics Check		OK	05:17

Data Download Date/Time: 01/28/2021 08:07

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: CABEZAS

DATE/TIME OF TEST: 12/13/2020 00:36:45

Diagnostics Check		OK	00:29
Air Blank		0.000	00:30
Control Test		0.080	00:30
Air Blank		0.000	00:31
Subject Sample	# 1	0.124	00:32
Breath Volume		3.191	00:32
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.127	00:35
Breath Volume		3.699	00:35
Air Blank		0.000	00:35
Control Test		0.080	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 01/28/2021 08:07

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 12/13/2020 03:27:49

Diagnostics Check		OK	03:20
Air Blank		0.000	03:20
Control Test		0.080	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.181	03:22
Breath Volume		1.433	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.170	03:26
Breath Volume		1.195	03:26
Air Blank		0.000	03:26
Control Test		0.080	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 01/28/2021 08:07

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:06

SUBJECT'S LAST NAME: RIOS QUESADA

DATE/TIME OF TEST: 12/13/2020 07:53:55

Diagnostics Check		OK	07:42
Air Blank		0.000	07:43
Control Test		0.081	07:43
Air Blank		0.000	07:44
Subject Sample	# 1	0.188	07:46
Breath Volume		2.824	07:46
Air Blank		0.000	07:47
Air Blank		0.000	07:48
Subject Sample	# 2	0.165	07:49
Breath Volume		1.375	07:49
Air Blank		0.000	07:49
Air Blank		0.000	07:51
Subject Sample	# 3	0.182	07:52
Breath Volume		2.441	07:52
Air Blank		0.000	07:53
Control Test		0.079	07:53
Air Blank		0.000	07:53
Diagnostics Check		OK	07:53

Data Download Date/Time: 01/28/2021 08:07

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GIRON,D
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:54

SUBJECT'S LAST NAME: DOMINGO MATIAS

DATE/TIME OF TEST: 12/13/2020 11:27:48

Diagnostics Check		OK	11:14	
Air Blank		0.000	11:15	
Control Test		0.080	11:15	
Air Blank		0.000	11:15	
Subject Sample	# 1	0.236	11:19	Volume Not Met
Breath Volume		0.843	11:19	
Air Blank		0.000	11:19	
Air Blank		0.000	11:21	
Subject Sample	# 2	0.261	11:22	
Breath Volume		2.242	11:22	
Air Blank		0.000	11:23	
Air Blank		0.000	11:25	
Subject Sample	# 3	0.241	11:26	
Breath Volume		1.187	11:26	
Air Blank		0.000	11:26	
Control Test		0.078	11:27	
Air Blank		0.000	11:27	
Diagnostics Check		OK	11:27	

Data Download Date/Time: 01/28/2021 08:07

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: MOYAO

DATE/TIME OF TEST: 12/13/2020 23:29:24

Diagnostics Check		OK	23:22
Air Blank		0.000	23:23
Control Test		0.080	23:23
Air Blank		0.000	23:23
Subject Sample	# 1	0.210	23:24
Breath Volume		3.664	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:27
Subject Sample	# 2	0.211	23:27
Breath Volume		3.312	23:27
Air Blank		0.000	23:28
Control Test		0.078	23:28
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 01/28/2021 08:08

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 12/14/2020 19:59:08

Diagnostics Check		OK	19:52
Air Blank		0.000	19:53
Control Test		0.080	19:53
Air Blank		0.000	19:53
Subject Sample	# 1	0.073	19:54
Breath Volume		1.664	19:54
Air Blank		0.000	19:54
Air Blank		0.000	19:56
Subject Sample	# 2	0.081	19:57
Breath Volume		2.242	19:57
Air Blank		0.000	19:58
Control Test		0.083	19:58
Air Blank		0.000	19:59
Diagnostics Check		OK	19:59

Data Download Date/Time: 01/28/2021 08:08

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 12/17/2020 04:16:32

Diagnostics Check		OK	04:08
Air Blank		0.000	04:08
Control Test		0.081	04:09
Air Blank		0.000	04:09
Subject Sample	# 1	0.159	04:12
Breath Volume		2.492	04:12
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.166	04:14
Breath Volume		2.769	04:14
Air Blank		0.000	04:15
Control Test		0.081	04:15
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 01/28/2021 08:08

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: REYNA

DATE/TIME OF TEST: 12/17/2020 21:24:17

Diagnostics Check		OK	21:17
Air Blank		0.000	21:18
Control Test		0.080	21:18
Air Blank		0.000	21:19
Subject Sample	# 1	0.156	21:19
Breath Volume		1.511	21:19
Air Blank		0.000	21:20
Air Blank		0.000	21:22
Subject Sample	# 2	0.151	21:22
Breath Volume		1.734	21:22
Air Blank		0.000	21:23
Control Test		0.080	21:23
Air Blank		0.000	21:24
Diagnostics Check		OK	21:24

Data Download Date/Time: 01/28/2021 08:08

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: VILLALOBOS

DATE/TIME OF TEST: 12/19/2020 02:36:22

Diagnostics Check		OK	02:29
Air Blank		0.000	02:29
Control Test		0.082	02:29
Air Blank		0.000	02:30
Subject Sample	# 1	0.148	02:31
Breath Volume		1.480	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.146	02:34
Breath Volume		2.496	02:34
Air Blank		0.000	02:35
Control Test		0.081	02:35
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

Data Download Date/Time: 01/28/2021 08:08

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: VARGAS ESPINAL

DATE/TIME OF TEST: 12/19/2020 04:43:15

Diagnostics Check		OK	04:36
Air Blank		0.000	04:36
Control Test		0.081	04:36
Air Blank		0.000	04:37
Subject Sample	# 1	0.101	04:38
Breath Volume		1.382	04:38
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.105	04:41
Breath Volume		1.300	04:41
Air Blank		0.000	04:42
Control Test		0.081	04:42
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 01/28/2021 08:08

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KLENOFSKY,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: FAUSTINO

DATE/TIME OF TEST: 12/20/2020 00:21:50

Diagnostics Check		OK	00:14
Air Blank		0.000	00:15
Control Test		0.081	00:15
Air Blank		0.000	00:16
Subject Sample	# 1	0.126	00:17
Breath Volume		1.644	00:17
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 2	0.126	00:20
Breath Volume		2.082	00:20
Air Blank		0.000	00:20
Control Test		0.080	00:21
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 01/28/2021 08:08

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/20/2020 02:38:14

Diagnostics Check		OK	02:31
Air Blank		0.000	02:32
Control Test		0.080	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.097	02:33
Breath Volume		2.105	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.094	02:36
Breath Volume		2.402	02:36
Air Blank		0.000	02:37
Control Test		0.080	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 01/28/2021 08:08

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: MIRANDA III

DATE/TIME OF TEST: 12/20/2020 04:24:13

Diagnostics Check		OK	04:17
Air Blank		0.000	04:18
Control Test		0.080	04:18
Air Blank		0.000	04:19
Subject Sample	# 1	0.182	04:19
Breath Volume		2.558	04:19
Air Blank		0.000	04:20
Air Blank		0.000	04:22
Subject Sample	# 2	0.183	04:22
Breath Volume		1.324	04:22
Air Blank		0.000	04:23
Control Test		0.079	04:23
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 01/28/2021 08:08

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: NOBALBO DE LA ROSA

DATE/TIME OF TEST: 12/23/2020 03:14:25

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.081	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.077	03:09
Breath Volume		1.453	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:11
Subject Sample	# 2	0.084	03:12
Breath Volume		2.203	03:12
Air Blank		0.000	03:13
Control Test		0.079	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 01/28/2021 08:08

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: TIPALDI

DATE/TIME OF TEST: 12/25/2020 03:29:28

Diagnostics Check		OK	03:21
Air Blank		0.000	03:22
Control Test		0.081	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.000	03:24
Breath Volume		1.164	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.000	03:28
Breath Volume		1.246	03:28
Air Blank		0.000	03:28
Control Test		0.080	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 01/28/2021 08:09

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 12/27/2020 04:21:21

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.082	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.131	04:16
Breath Volume		1.460	04:16
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.120	04:19
Breath Volume		1.515	04:19
Air Blank		0.000	04:20
Control Test		0.079	04:20
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 01/28/2021 08:09

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: COURSEN

DATE/TIME OF TEST: 12/27/2020 21:47:05

Diagnostics Check		OK	21:40
Air Blank		0.000	21:40
Control Test		0.082	21:40
Air Blank		0.000	21:41
Subject Sample	# 1	0.266	21:42
Breath Volume		3.507	21:42
Air Blank		0.000	21:43
Air Blank		0.000	21:44
Subject Sample	# 2	0.267	21:45
Breath Volume		3.273	21:45
Air Blank		0.000	21:46
Control Test		0.081	21:46
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

Data Download Date/Time: 01/28/2021 08:09

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: ROCKHOLD

DATE/TIME OF TEST: 12/31/2020 01:14:36

Diagnostics Check		OK	01:06
Air Blank		0.000	01:06
Control Test		0.080	01:07
Air Blank		0.000	01:07
Subject Sample	# 1	0.265	01:09
Breath Volume		1.746	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.256	01:13
Breath Volume		1.226	01:13
Air Blank		0.000	01:13
Control Test		0.078	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 01/28/2021 08:09

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RAMOS,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: PATRICIO PATRICIO

DATE/TIME OF TEST: 12/31/2020 18:58:13

Diagnostics Check		OK	18:51
Air Blank		0.000	18:52
Control Test		0.080	18:52
Air Blank		0.000	18:52
Subject Sample	# 1	0.340	18:53
Breath Volume		2.328	18:53
Air Blank		0.000	18:54
Air Blank		0.000	18:56
Subject Sample	# 2	0.351	18:56
Breath Volume		2.312	18:56
Air Blank		0.000	18:57
Control Test		0.079	18:57
Air Blank		0.000	18:58
Diagnostics Check		OK	18:58

Data Download Date/Time: 01/28/2021 08:09

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

INSTRUMENT REGISTERED TO: FHP TROOP K
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,DAVID,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: COX,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1327124
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 15:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: FARREN

DATE/TIME OF TEST: 12/01/2020 15:53:40

Diagnostics Check		OK	15:50	
Air Blank		0.000	15:50	
Control Test		0.081	15:50	
Air Blank		0.000	15:51	
Subject Sample	# 1		15:52	Subject Test Refused
Breath Volume		0.000	15:52	
Air Blank		0.000	15:52	
Control Test		0.081	15:53	
Air Blank		0.000	15:53	
Diagnostics Check		OK	15:53	

Data Download Date/Time: 12/10/2020 17:35

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

INSTRUMENT REGISTERED TO: FHP TROOP K
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LERNER,MATTHEW,H
BREATH TEST OPERATOR AGENCY: Washington County SO
ARRESTING OFFICER: PASLEY,JACOB
ARRESTING OFFICER AGENCY: Chipley PD
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1327124
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: LANG

DATE/TIME OF TEST: 12/01/2020 23:11:20

Diagnostics Check		OK	23:05	
Air Blank		0.000	23:05	
Control Test		0.081	23:06	
Air Blank		0.000	23:06	
Subject Sample	# 1	0.187	23:07	
Breath Volume		3.007	23:07	
Air Blank		0.000	23:08	
Air Blank		0.000	23:09	
			23:09	RFI Detect
Air Blank		0.000	23:10	
Control Test		0.082	23:10	
Air Blank		0.000	23:11	
Diagnostics Check		OK	23:11	

Data Download Date/Time: 12/10/2020 17:35

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

INSTRUMENT REGISTERED TO: FHP TROOP K
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LERNERM,MATTHEW,H
BREATH TEST OPERATOR AGENCY: Washington County SO
ARRESTING OFFICER: PASLEY,JACON
ARRESTING OFFICER AGENCY: Chipley PD
LAST AGENCY INSPECTION: 11/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1327124
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: LANG

DATE/TIME OF TEST: 12/01/2020 23:20:14

Diagnostics Check		OK	23:13
Air Blank		0.000	23:14
Control Test		0.080	23:14
Air Blank		0.000	23:15
Subject Sample	# 1	0.184	23:15
Breath Volume		2.726	23:15
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.181	23:18
Breath Volume		1.664	23:18
Air Blank		0.000	23:19
Control Test		0.080	23:19
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 12/10/2020 17:35

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: CARRICHI-ORTIZ

DATE/TIME OF TEST: 12/16/2020 23:11:28

Diagnostics Check		OK	23:04
Air Blank		0.000	23:04
Control Test		0.080	23:05
Air Blank		0.000	23:05
Subject Sample	# 1	0.054	23:06
Breath Volume		1.433	23:06
Air Blank		0.000	23:07
Air Blank		0.000	23:09
Subject Sample	# 2	0.053	23:09
Breath Volume		2.445	23:09
Air Blank		0.000	23:10
Control Test		0.079	23:10
Air Blank		0.000	23:11
Diagnostics Check		OK	23:11

Data Download Date/Time: 01/19/2021 12:10

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: FLINT

DATE/TIME OF TEST: 12/22/2020 02:00:49

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:47	
Control Test		0.081	01:47	
Air Blank		0.000	01:48	
Subject Sample	# 1	0.175	01:51	Volume Not Met
Breath Volume		1.082	01:51	
Air Blank		0.000	01:51	
Air Blank		0.000	01:53	
Subject Sample	# 2	0.193	01:55	
Breath Volume		1.382	01:55	
Air Blank		0.000	01:56	
Air Blank		0.000	01:57	
Subject Sample	# 3	0.210	01:59	
Breath Volume		2.761	01:59	
Air Blank			02:00	Ambient Fail
Air Blank			02:00	Purge Fail

Data Download Date/Time: 01/19/2021 12:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWF0RD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 12/02/2020 00:31:54

Diagnostics Check		OK	00:25
Air Blank		0.000	00:26
Control Test		0.082	00:26
Air Blank		0.000	00:27
Subject Sample	# 1	0.151	00:27
Breath Volume		1.390	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.159	00:30
Breath Volume		1.855	00:30
Air Blank		0.000	00:31
Control Test		0.081	00:31
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 12/15/2020 09:48

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: CORDERO MUNIZ

DATE/TIME OF TEST: 12/05/2020 21:21:29

Diagnostics Check		OK	21:16	
Air Blank		0.000	21:16	
Control Test		0.081	21:17	
Air Blank		0.000	21:17	
Subject Sample	# 1		21:19	RFI Detect
Breath Volume		0.046	21:19	
Air Blank		0.000	21:20	
Control Test		0.081	21:20	
Air Blank		0.000	21:21	
Diagnostics Check		OK	21:21	

Data Download Date/Time: 12/15/2020 09:48

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 11/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: CORDERO MUNIZ

DATE/TIME OF TEST: 12/05/2020 21:31:42

Diagnostics Check		OK	21:23	
Air Blank		0.000	21:23	
Control Test		0.081	21:24	
Air Blank		0.000	21:24	
Subject Sample	# 1	0.104	21:27	Volume Not Met
Breath Volume		0.253	21:27	
Air Blank		0.000	21:28	
Air Blank		0.000	21:30	
Subject Sample	# 2		21:30	Subject Test Refused
Breath Volume		0.000	21:30	
Air Blank		0.000	21:30	
Control Test		0.078	21:31	
Air Blank		0.000	21:31	
Diagnostics Check		OK	21:31	

Data Download Date/Time: 12/15/2020 09:48

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLANCK,ARIEL,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: KRITENBRINK

DATE/TIME OF TEST: 12/03/2020 00:34:37

Diagnostics Check		OK	00:31	
Air Blank		0.000	00:32	
Control Test		0.081	00:32	
Air Blank		0.000	00:33	
Subject Sample	# 1		00:33	Subject Test Refused
Breath Volume		0.000	00:33	
Air Blank		0.000	00:33	
Control Test		0.082	00:34	
Air Blank		0.000	00:34	
Diagnostics Check		OK	00:34	

Data Download Date/Time: 12/04/2020 10:52

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEENEY,ERIC,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FEENEY,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: SHELEK

DATE/TIME OF TEST: 12/03/2020 22:24:38

Diagnostics Check		OK	22:18
Air Blank		0.000	22:18
Control Test		0.080	22:19
Air Blank		0.000	22:19
Subject Sample	# 1	0.281	22:19
Breath Volume		2.289	22:19
Air Blank		0.000	22:20
Air Blank		0.000	22:22
Subject Sample	# 2	0.279	22:23
Breath Volume		3.035	22:23
Air Blank		0.000	22:23
Control Test		0.080	22:24
Air Blank		0.000	22:24
Diagnostics Check		OK	22:24

Data Download Date/Time: 12/04/2020 10:52

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLANCK,ARIEL,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: TURNER,JACOB
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: SIEGRIST

DATE/TIME OF TEST: 12/04/2020 05:41:44

Diagnostics Check		OK	05:35
Air Blank		0.000	05:35
Control Test		0.081	05:35
Air Blank		0.000	05:36
Subject Sample	# 1	0.054	05:36
Breath Volume		1.457	05:36
Air Blank		0.000	05:37
Air Blank		0.000	05:39
Subject Sample	# 2	0.053	05:40
Breath Volume		1.320	05:40
Air Blank		0.000	05:40
Control Test		0.080	05:41
Air Blank		0.000	05:41
Diagnostics Check		OK	05:41

Data Download Date/Time: 12/04/2020 10:52

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,SHELBY,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: CHRESTMAN

DATE/TIME OF TEST: 12/05/2020 03:30:53

Diagnostics Check		OK	03:28	
Air Blank		0.000	03:28	
Control Test		0.081	03:28	
Air Blank		0.000	03:29	
Subject Sample	# 1		03:29	Subject Test Refused
Breath Volume		0.000	03:29	
Air Blank		0.000	03:29	
Control Test		0.081	03:30	
Air Blank		0.000	03:30	
Diagnostics Check		OK	03:30	

Data Download Date/Time: 01/15/2021 11:46

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MLYNARSKI,NIKLAS,E
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MLYNARSKI,NIKLAS
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: O'BRIEN

DATE/TIME OF TEST: 12/05/2020 04:10:26

Diagnostics Check		OK	04:06	
Air Blank		0.000	04:07	
Control Test		0.081	04:07	
Air Blank		0.000	04:08	
Subject Sample	# 1		04:09	Subject Test Refused
Breath Volume		0.000	04:09	
Air Blank		0.000	04:09	
Control Test		0.081	04:09	
Air Blank		0.000	04:10	
Diagnostics Check		OK	04:10	

Data Download Date/Time: 01/15/2021 11:46

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON,ERIC,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STINSON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 12/05/2020 04:21:12

Diagnostics Check		OK	04:18	
Air Blank		0.000	04:18	
Control Test		0.081	04:19	
Air Blank		0.000	04:19	
Subject Sample	# 1		04:19	Subject Test Refused
Breath Volume		0.000	04:19	
Air Blank		0.000	04:20	
Control Test		0.081	04:20	
Air Blank		0.000	04:21	
Diagnostics Check		OK	04:21	

Data Download Date/Time: 01/15/2021 11:46

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,GABRIEL,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ROBERSON,GABRIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: DARDEN

DATE/TIME OF TEST: 12/05/2020 04:53:08

Diagnostics Check		OK	04:50	
Air Blank		0.000	04:50	
Control Test		0.082	04:51	
Air Blank		0.000	04:51	
Subject Sample	# 1		04:51	Subject Test Refused
Breath Volume		0.000	04:51	
Air Blank		0.000	04:52	
Control Test		0.081	04:52	
Air Blank		0.000	04:53	
Diagnostics Check		OK	04:53	

Data Download Date/Time: 01/15/2021 11:46

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK,TAVISH,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK,TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 12/05/2020 06:25:33

Diagnostics Check		OK	06:19
Air Blank		0.000	06:20
Control Test		0.081	06:20
Air Blank		0.000	06:20
Subject Sample	# 1	0.000	06:21
Breath Volume		1.156	06:21
Air Blank		0.000	06:21
Air Blank		0.000	06:23
Subject Sample	# 2	0.000	06:24
Breath Volume		1.656	06:24
Air Blank		0.000	06:24
Control Test		0.080	06:24
Air Blank		0.000	06:25
Diagnostics Check		OK	06:25

Data Download Date/Time: 01/15/2021 11:46

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HART,JOSEPH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: LEMASTER

DATE/TIME OF TEST: 12/05/2020 22:51:01

Diagnostics Check		OK	22:44
Air Blank		0.000	22:44
Control Test		0.081	22:45
Air Blank		0.000	22:45
Subject Sample	# 1	0.176	22:46
Breath Volume		1.152	22:46
Air Blank		0.000	22:46
Air Blank		0.000	22:48
Subject Sample	# 2	0.184	22:49
Breath Volume		2.027	22:49
Air Blank		0.000	22:50
Control Test		0.080	22:50
Air Blank		0.000	22:50
Diagnostics Check		OK	22:51

Data Download Date/Time: 01/15/2021 11:46

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON,ERIC,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQUEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: AMMONS

DATE/TIME OF TEST: 12/06/2020 03:31:58

Diagnostics Check		OK	03:25
Air Blank		0.000	03:26
Control Test		0.081	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.161	03:27
Breath Volume		1.808	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:29
Subject Sample	# 2	0.143	03:30
Breath Volume		1.886	03:30
Air Blank		0.000	03:31
Control Test		0.079	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 01/15/2021 11:46

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,GABRIEL,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ROBERSON,GABRIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: GABRIEL MENDOZA

DATE/TIME OF TEST: 12/06/2020 06:00:25

Diagnostics Check		OK	05:54
Air Blank		0.000	05:54
Control Test		0.080	05:55
Air Blank		0.000	05:55
Subject Sample	# 1	0.188	05:56
Breath Volume		1.750	05:56
Air Blank		0.000	05:56
Air Blank		0.000	05:58
Subject Sample	# 2	0.185	05:58
Breath Volume		1.937	05:58
Air Blank		0.000	05:59
Control Test		0.079	05:59
Air Blank		0.000	06:00
Diagnostics Check		OK	06:00

Data Download Date/Time: 01/15/2021 11:46

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDE,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BLOCK,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: ICALCUCUL

DATE/TIME OF TEST: 12/06/2020 06:09:52

Diagnostics Check		OK	06:06	
Air Blank		0.000	06:06	
Control Test		0.081	06:07	
Air Blank		0.000	06:07	
Subject Sample	# 1		06:08	Subject Test Refused
Breath Volume		0.000	06:08	
Air Blank		0.000	06:08	
Control Test		0.081	06:09	
Air Blank		0.000	06:09	
Diagnostics Check		OK	06:09	

Data Download Date/Time: 01/15/2021 11:46

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: LUTHER

DATE/TIME OF TEST: 12/06/2020 19:57:12

Diagnostics Check		OK	19:54	
Air Blank		0.000	19:54	
Control Test		0.081	19:55	
Air Blank		0.000	19:55	
Subject Sample	# 1		19:55	Subject Test Refused
Breath Volume		0.000	19:55	
Air Blank		0.000	19:56	
Control Test		0.081	19:56	
Air Blank		0.000	19:57	
Diagnostics Check		OK	19:57	

Data Download Date/Time: 01/15/2021 11:46

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON,ERIC,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STINSON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/06/2020 21:01:19

Diagnostics Check		OK	20:55
Air Blank		0.000	20:55
Control Test		0.081	20:55
Air Blank		0.000	20:56
Subject Sample	# 1	0.162	20:56
Breath Volume		2.785	20:56
Air Blank		0.000	20:57
Air Blank		0.000	20:59
Subject Sample	# 2	0.153	20:59
Breath Volume		2.082	20:59
Air Blank		0.000	21:00
Control Test		0.080	21:00
Air Blank		0.000	21:01
Diagnostics Check		OK	21:01

Data Download Date/Time: 01/15/2021 11:46

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEENEY,ERIC,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FEENEY,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 12:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:27

SUBJECT'S LAST NAME: HADERSBECK

DATE/TIME OF TEST: 12/07/2020 13:53:19

Diagnostics Check		OK	13:50	
Air Blank		0.000	13:50	
Control Test		0.081	13:51	
Air Blank		0.000	13:51	
Subject Sample	# 1		13:51	Subject Test Refused
Breath Volume		0.000	13:51	
Air Blank		0.000	13:52	
Control Test		0.081	13:52	
Air Blank		0.000	13:53	
Diagnostics Check		OK	13:53	

Data Download Date/Time: 01/15/2021 11:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLANCK,ARIEL,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PLANCK,ARIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 12/08/2020 00:59:20

Diagnostics Check		OK	00:56	
Air Blank		0.000	00:56	
Control Test		0.081	00:57	
Air Blank		0.000	00:57	
Subject Sample	# 1		00:57	Subject Test Refused
Breath Volume		0.000	00:57	
Air Blank		0.000	00:58	
Control Test		0.080	00:58	
Air Blank		0.000	00:59	
Diagnostics Check		OK	00:59	

Data Download Date/Time: 01/15/2021 11:46

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEENEY,ERIC,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ZELLARS,TRENT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: DUGAN

DATE/TIME OF TEST: 12/11/2020 20:57:15

Diagnostics Check		OK	20:54	
Air Blank		0.000	20:54	
Control Test		0.080	20:55	
Air Blank		0.000	20:55	
Subject Sample	# 1		20:55	Subject Test Refused
Breath Volume		0.000	20:55	
Air Blank		0.000	20:56	
Control Test		0.080	20:56	
Air Blank		0.000	20:57	
Diagnostics Check		OK	20:57	

Data Download Date/Time: 01/15/2021 11:46

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEENEY,ERIC,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,K
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: GAUTHIER

DATE/TIME OF TEST: 12/12/2020 02:17:35

Diagnostics Check		OK	02:10
Air Blank		0.000	02:11
Control Test		0.080	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.206	02:12
Breath Volume		1.718	02:12
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.202	02:16
Breath Volume		2.105	02:16
Air Blank		0.000	02:16
Control Test		0.079	02:16
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 01/15/2021 11:47

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ANDREW,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: PATISHTAN DELACRUZ

DATE/TIME OF TEST: 12/18/2020 20:09:02

Diagnostics Check		OK	20:02
Air Blank		0.000	20:03
Control Test		0.082	20:03
Air Blank		0.000	20:04
Subject Sample	# 1	0.176	20:04
Breath Volume		2.398	20:04
Air Blank		0.000	20:05
Air Blank		0.000	20:07
Subject Sample	# 2	0.178	20:07
Breath Volume		2.070	20:07
Air Blank		0.000	20:08
Control Test		0.080	20:08
Air Blank		0.000	20:08
Diagnostics Check		OK	20:09

Data Download Date/Time: 01/15/2021 11:47

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: FUGON-CUBAS

DATE/TIME OF TEST: 12/20/2020 04:19:21

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.081	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.242	04:12
Breath Volume		3.011	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.231	04:17
Breath Volume		1.812	04:17
Air Blank		0.000	04:18
Control Test		0.080	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 01/15/2021 11:47

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: TATE

DATE/TIME OF TEST: 12/21/2020 02:23:29

Diagnostics Check		OK	02:17
Air Blank		0.000	02:17
Control Test		0.081	02:17
Air Blank		0.000	02:18
Subject Sample	# 1	0.000	02:19
Breath Volume		2.566	02:19
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.000	02:22
Breath Volume		2.304	02:22
Air Blank		0.000	02:22
Control Test		0.080	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 01/15/2021 11:47

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELF,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SELF,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: MITTLESTEADT

DATE/TIME OF TEST: 12/22/2020 02:17:34

Diagnostics Check		OK	02:09
Air Blank		0.000	02:09
Control Test		0.081	02:09
Air Blank		0.000	02:10
Subject Sample	# 1	0.213	02:13
Breath Volume		2.226	02:13
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.206	02:15
Breath Volume		1.863	02:15
Air Blank		0.000	02:16
Control Test		0.078	02:16
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 01/15/2021 11:47

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORTON,JAMES,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: NORENE,AIDAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: LOPEZ-MAZARIEGOS

DATE/TIME OF TEST: 12/23/2020 00:33:44

Diagnostics Check		OK	00:22	
Air Blank		0.000	00:22	
Control Test		0.081	00:23	
Air Blank		0.000	00:23	
Subject Sample	# 1		00:26	No Sample Provided
Breath Volume		0.000	00:26	
Air Blank		0.000	00:27	
Air Blank		0.000	00:29	
Subject Sample	# 2		00:32	No Sample Provided
Breath Volume		0.000	00:32	
Air Blank		0.000	00:32	
Control Test		0.081	00:33	
Air Blank		0.000	00:33	
Diagnostics Check		OK	00:33	

Data Download Date/Time: 01/15/2021 11:47

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMALL,BRANNON,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SMALL,BRANNON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 14:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: PLATE

DATE/TIME OF TEST: 12/23/2020 15:41:38

Diagnostics Check		OK	15:35
Air Blank		0.000	15:35
Control Test		0.081	15:36
Air Blank		0.000	15:36
Subject Sample	# 1	0.000	15:37
Breath Volume		1.710	15:37
Air Blank		0.000	15:37
Air Blank		0.000	15:39
Subject Sample	# 2	0.000	15:40
Breath Volume		1.589	15:40
Air Blank		0.000	15:40
Control Test		0.079	15:41
Air Blank		0.000	15:41
Diagnostics Check		OK	15:41

Data Download Date/Time: 01/15/2021 11:47

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDE,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQUEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 12/24/2020 02:08:21

Diagnostics Check		OK	02:04	
Air Blank		0.000	02:05	
Control Test		0.081	02:05	
Air Blank		0.000	02:06	
Subject Sample	# 1		02:06	Subject Test Refused
Breath Volume		0.000	02:06	
Air Blank		0.000	02:07	
Control Test		0.081	02:07	
Air Blank		0.000	02:08	
Diagnostics Check		OK	02:08	

Data Download Date/Time: 01/15/2021 11:47

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDE,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILHELM,TONY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 12/24/2020 02:54:04

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:51	
Control Test		0.080	02:52	
Air Blank		0.000	02:52	
Subject Sample	# 1		02:52	Subject Test Refused
Breath Volume		0.000	02:52	
Air Blank		0.000	02:53	
Control Test		0.081	02:53	
Air Blank		0.000	02:53	
Diagnostics Check		OK	02:54	

Data Download Date/Time: 01/15/2021 11:47

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: FLACHOFSKY

DATE/TIME OF TEST: 12/24/2020 19:45:04

Diagnostics Check		OK	19:38
Air Blank		0.000	19:39
Control Test		0.082	19:39
Air Blank		0.000	19:39
Subject Sample	# 1	0.000	19:40
Breath Volume		2.132	19:40
Air Blank		0.000	19:41
Air Blank		0.000	19:43
Subject Sample	# 2	0.000	19:43
Breath Volume		1.386	19:43
Air Blank		0.000	19:44
Control Test		0.081	19:44
Air Blank		0.000	19:44
Diagnostics Check		OK	19:45

Data Download Date/Time: 01/15/2021 11:47

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILHELM,TONY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 12:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: MEJIA BOLLATE

DATE/TIME OF TEST: 12/25/2020 00:59:37

Diagnostics Check		OK	00:48	
Air Blank		0.000	00:48	
Control Test		0.082	00:49	
Air Blank		0.000	00:49	
Subject Sample	# 1		00:52	No Sample Provided
Breath Volume		0.000	00:52	
Air Blank		0.000	00:53	
Air Blank		0.000	00:55	
Subject Sample	# 2		00:58	No Sample Provided
Breath Volume		0.000	00:58	
Air Blank		0.000	00:58	
Control Test		0.082	00:59	
Air Blank		0.000	00:59	
Diagnostics Check		OK	00:59	

Data Download Date/Time: 01/15/2021 11:47

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILHELM,TONY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MEJIA BOLLATE

DATE/TIME OF TEST: 12/25/2020 01:16:07

Diagnostics Check		OK	01:11	
Air Blank		0.000	01:11	
Control Test		0.082	01:12	
Air Blank		0.000	01:12	
Subject Sample	# 1		01:14	Subject Test Refused
Breath Volume		0.109	01:14	
Air Blank		0.000	01:15	
Control Test		0.082	01:15	
Air Blank		0.000	01:16	
Diagnostics Check		OK	01:16	

Data Download Date/Time: 01/15/2021 11:47

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDE,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BLOCK,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/25/2020 01:43:36

Diagnostics Check		OK	01:33	
Air Blank		0.000	01:33	
Control Test		0.081	01:34	
Air Blank		0.000	01:34	
Subject Sample	# 1	0.197	01:37	Volume Not Met
Breath Volume		0.535	01:37	
Air Blank		0.000	01:38	
Air Blank		0.000	01:40	
Subject Sample	# 2		01:42	Subject Test Refused
Breath Volume		0.019	01:42	
Air Blank		0.000	01:42	
Control Test		0.080	01:42	
Air Blank		0.000	01:43	
Diagnostics Check		OK	01:43	

Data Download Date/Time: 01/15/2021 11:47

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELF,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GANDER,PHILLIP
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: BALTES

DATE/TIME OF TEST: 12/27/2020 22:46:22

Diagnostics Check		OK	22:43	
Air Blank		0.000	22:44	
Control Test		0.083	22:44	
Air Blank		0.000	22:44	
Subject Sample	# 1		22:45	Subject Test Refused
Breath Volume		0.000	22:45	
Air Blank		0.000	22:45	
Control Test		0.082	22:45	
Air Blank		0.000	22:46	
Diagnostics Check		OK	22:46	

Data Download Date/Time: 01/15/2021 11:47

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: BALLARD

DATE/TIME OF TEST: 12/28/2020 21:47:54

Diagnostics Check		OK	21:41
Air Blank		0.000	21:42
Control Test		0.082	21:42
Air Blank		0.000	21:42
Subject Sample	# 1	0.150	21:43
Breath Volume		1.992	21:43
Air Blank		0.000	21:44
Air Blank		0.000	21:45
Subject Sample	# 2	0.142	21:46
Breath Volume		1.230	21:46
Air Blank		0.000	21:47
Control Test		0.081	21:47
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

Data Download Date/Time: 01/15/2021 11:47

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK,TAVISH,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK,TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: COPPERSMITH

DATE/TIME OF TEST: 12/29/2020 02:09:06

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.081	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.235	02:04
Breath Volume		1.460	02:04
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.235	02:07
Breath Volume		1.734	02:07
Air Blank		0.000	02:08
Control Test		0.081	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 01/15/2021 11:47

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRANE,RICHARD,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DRANE,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: TYLER

DATE/TIME OF TEST: 12/29/2020 22:29:41

Diagnostics Check		OK	22:22	
Air Blank		0.000	22:23	
Control Test		0.081	22:23	
Air Blank		0.000	22:23	
Subject Sample	# 1	0.210	22:24	
Breath Volume		1.296	22:24	
Air Blank		0.000	22:25	
Air Blank		0.000	22:27	
Subject Sample	# 2	0.198	22:28	
Breath Volume		1.632	22:28	
Air Blank			22:29	RFI Detect
Air Blank		0.000	22:29	

Data Download Date/Time: 01/15/2021 11:47

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORTON,JAMES,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SELF,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/30/2020 22:35:31

Diagnostics Check		OK	22:23	
Air Blank		0.000	22:24	
Control Test		0.080	22:24	
Air Blank		0.000	22:24	
Subject Sample	# 1	0.209	22:28	Volume Not Met
Breath Volume		0.031	22:28	
Air Blank		0.000	22:28	
Air Blank		0.000	22:30	
Subject Sample	# 2	0.190	22:33	Volume Not Met
Breath Volume		0.628	22:33	
Air Blank		0.000	22:34	
Control Test		0.077	22:34	
Air Blank		0.000	22:35	
Diagnostics Check		OK	22:35	

Data Download Date/Time: 01/15/2021 11:47

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: CARTWRIGHT

DATE/TIME OF TEST: 12/31/2020 23:42:33

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.081	23:35
Air Blank		0.000	23:36
Subject Sample	# 1	0.181	23:37
Breath Volume		2.148	23:37
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.176	23:40
Breath Volume		2.246	23:40
Air Blank		0.000	23:41
Control Test		0.079	23:41
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 01/15/2021 11:47

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLAP,GABRIEL,C
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: DUNLAP,GABRIEL
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 11/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 12/12/2020 01:51:08

Diagnostics Check		OK	01:39	
Air Blank		0.000	01:39	
Control Test		0.080	01:40	
Air Blank		0.000	01:40	
Subject Sample	# 1	0.196	01:43	Volume Not Met
Breath Volume		0.367	01:43	
Air Blank		0.000	01:44	
Air Blank		0.000	01:46	
Subject Sample	# 2	0.172	01:49	Volume Not Met
Breath Volume		0.082	01:49	
Air Blank		0.000	01:50	
Control Test		0.078	01:50	
Air Blank		0.000	01:51	
Diagnostics Check		OK	01:51	

Data Download Date/Time: 12/31/2020 01:10

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLAP,GABRIEL,C
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: DUNLAP,GABRIEL
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 11/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: GRUBE

DATE/TIME OF TEST: 12/16/2020 21:36:02

Diagnostics Check		OK	21:33	
Air Blank		0.000	21:33	
Control Test		0.080	21:33	
Air Blank		0.000	21:34	
Subject Sample	# 1		21:34	Subject Test Refused
Breath Volume		0.000	21:34	
Air Blank		0.000	21:35	
Control Test		0.080	21:35	
Air Blank		0.000	21:35	
Diagnostics Check		OK	21:36	

Data Download Date/Time: 12/31/2020 01:10

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLAP,GABRIEL,C
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: DUNLAP,GABRIEL
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 11/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 12/21/2020 20:47:55

Diagnostics Check		OK	20:45	
Air Blank		0.000	20:45	
Control Test		0.080	20:45	
Air Blank		0.000	20:46	
Subject Sample	# 1		20:46	RFI Detect
Breath Volume		0.000	20:46	
Air Blank		0.000	20:47	
Control Test		0.081	20:47	
Air Blank		0.000	20:47	
Diagnostics Check		OK	20:47	

Data Download Date/Time: 12/31/2020 01:10

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLAP,GABRIEL,C
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: DUNLAP,GABRIEL
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 11/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 12/21/2020 20:53:18

Diagnostics Check		OK	20:50	
Air Blank		0.000	20:50	
Control Test		0.080	20:51	
Air Blank		0.000	20:51	
Subject Sample	# 1		20:51	Subject Test Refused
Breath Volume		0.000	20:51	
Air Blank		0.000	20:52	
Control Test		0.080	20:52	
Air Blank		0.000	20:53	
Diagnostics Check		OK	20:53	

Data Download Date/Time: 12/31/2020 01:10

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLAP,GABRIEL,C
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: WHITE,TIMOTHY
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 11/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: SHEPPARD

DATE/TIME OF TEST: 12/26/2020 20:27:48

Diagnostics Check		OK	20:21
Air Blank		0.000	20:21
Control Test		0.080	20:22
Air Blank		0.000	20:22
Subject Sample	# 1	0.334	20:23
Breath Volume		2.429	20:23
Air Blank		0.000	20:24
Air Blank		0.000	20:25
Subject Sample	# 2	0.330	20:26
Breath Volume		2.812	20:26
Air Blank		0.000	20:26
Control Test		0.079	20:27
Air Blank		0.000	20:27
Diagnostics Check		OK	20:27

Data Download Date/Time: 12/31/2020 01:11

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLAP,GABRIEL,C
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: WHITE,TIMOTHY
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 11/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: SHEPPARD

DATE/TIME OF TEST: 12/26/2020 20:57:31

Diagnostics Check		OK	20:54	
Air Blank		0.000	20:55	
Control Test		0.080	20:55	
Air Blank		0.000	20:55	
Subject Sample	# 1		20:56	RFI Detect
Breath Volume		0.000	20:56	
Air Blank		0.000	20:56	
Control Test		0.081	20:56	
Air Blank		0.000	20:57	
Diagnostics Check		OK	20:57	

Data Download Date/Time: 12/31/2020 01:11

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLAP,GABRIEL,C
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: WHITE,TIMOTHY
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 11/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: SHEPPARD

DATE/TIME OF TEST: 12/26/2020 21:06:00

Diagnostics Check		OK	20:59	
Air Blank		0.000	21:00	
Control Test		0.081	21:00	
Air Blank		0.000	21:01	
Subject Sample	# 1	0.309	21:01	
Breath Volume		1.320	21:01	
Air Blank		0.000	21:02	
Air Blank		0.000	21:04	
Subject Sample	# 2		21:04	Subject Test Refused
Breath Volume		0.000	21:04	
Air Blank		0.000	21:05	
Control Test		0.080	21:05	
Air Blank		0.000	21:05	
Diagnostics Check		OK	21:06	

Data Download Date/Time: 12/31/2020 01:11

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: GARRIDOCRUZ

DATE/TIME OF TEST: 12/01/2020 16:21:26

Diagnostics Check		OK	16:15
Air Blank		0.000	16:15
Control Test		0.080	16:16
Air Blank		0.000	16:16
Subject Sample	# 1	0.157	16:17
Breath Volume		2.265	16:17
Air Blank		0.000	16:17
Air Blank		0.000	16:19
Subject Sample	# 2	0.161	16:19
Breath Volume		3.050	16:19
Air Blank		0.000	16:20
Control Test		0.079	16:20
Air Blank		0.000	16:21
Diagnostics Check		OK	16:21

Data Download Date/Time: 12/09/2020 08:54

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: AVILA SANTOS

DATE/TIME OF TEST: 12/02/2020 03:52:16

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:47
Subject Sample	# 1	0.100	03:47
Breath Volume		2.406	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.091	03:50
Breath Volume		2.839	03:50
Air Blank		0.000	03:51
Control Test		0.080	03:51
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 12/09/2020 08:54

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: MAKOWSKI

DATE/TIME OF TEST: 12/04/2020 23:09:12

Diagnostics Check		OK	23:04	
Air Blank		0.000	23:04	
Control Test		0.079	23:05	
Air Blank		0.000	23:05	
Subject Sample	# 1		23:07	Subject Test Refused
Breath Volume		0.000	23:07	
Air Blank		0.000	23:08	
Control Test		0.079	23:08	
Air Blank		0.000	23:09	
Diagnostics Check		OK	23:09	

Data Download Date/Time: 12/09/2020 08:54

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,M
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: NOVIELLI

DATE/TIME OF TEST: 12/05/2020 03:15:52

Diagnostics Check		OK	03:04	
Air Blank		0.000	03:04	
Control Test		0.079	03:05	
Air Blank		0.000	03:05	
Subject Sample	# 1		03:08	No Sample Provided
Breath Volume		0.000	03:08	
Air Blank		0.000	03:09	
Air Blank		0.000	03:11	
Subject Sample	# 2		03:14	No Sample Provided
Breath Volume		0.000	03:14	
Air Blank		0.000	03:14	
Control Test		0.079	03:15	
Air Blank		0.000	03:15	
Diagnostics Check		OK	03:15	

Data Download Date/Time: 12/09/2020 08:54

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BOETGER,OFC
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: RHOADS

DATE/TIME OF TEST: 12/05/2020 20:06:16

Diagnostics Check		OK	20:03	
Air Blank		0.000	20:03	
Control Test		0.080	20:04	
Air Blank		0.000	20:04	
Subject Sample	# 1	0.257	20:05	
Breath Volume		2.683	20:05	
Air Blank			20:05	RFI Detect
Air Blank			20:05	RFI Detect
Air Blank		0.000	20:06	

Data Download Date/Time: 12/09/2020 08:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BOETGER,OFC
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: RHOADS

DATE/TIME OF TEST: 12/05/2020 20:14:45

Diagnostics Check		OK	20:08
Air Blank		0.000	20:08
Control Test		0.080	20:09
Air Blank		0.000	20:09
Subject Sample	# 1	0.247	20:10
Breath Volume		2.984	20:10
Air Blank		0.000	20:10
Air Blank		0.000	20:12
Subject Sample	# 2	0.241	20:13
Breath Volume		2.558	20:13
Air Blank		0.000	20:13
Control Test		0.080	20:14
Air Blank		0.000	20:14
Diagnostics Check		OK	20:14

Data Download Date/Time: 12/09/2020 08:54

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: SHERMAN

DATE/TIME OF TEST: 12/06/2020 01:41:46

Diagnostics Check		OK	01:26	
Air Blank		0.000	01:26	
Control Test		0.080	01:26	
Air Blank		0.000	01:27	
Subject Sample	# 1	0.140	01:30	Volume Not Met
Breath Volume		0.183	01:30	
Air Blank		0.000	01:31	
Air Blank		0.000	01:32	
Subject Sample	# 2	0.150	01:34	
Breath Volume		1.632	01:34	
Air Blank		0.000	01:35	
Air Blank		0.000	01:37	
Subject Sample	# 3	0.113	01:40	Volume Not Met
Breath Volume		0.542	01:40	
Air Blank		0.000	01:40	
Control Test		0.078	01:41	
Air Blank		0.000	01:41	
Diagnostics Check		OK	01:41	

Data Download Date/Time: 12/09/2020 08:54

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,THERESA,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: SHERMAN

DATE/TIME OF TEST: 12/06/2020 01:56:56

Diagnostics Check		OK	01:45	
Air Blank		0.000	01:45	
Control Test		0.077	01:46	
Air Blank		0.000	01:46	
Subject Sample	# 1	0.152	01:49	Volume Not Met
Breath Volume		0.039	01:49	
Air Blank		0.000	01:50	
Air Blank		0.000	01:52	
Subject Sample	# 2	0.122	01:55	Volume Not Met
Breath Volume		0.726	01:55	
Air Blank		0.000	01:55	
Control Test		0.078	01:56	
Air Blank		0.000	01:56	
Diagnostics Check		OK	01:56	

Data Download Date/Time: 12/09/2020 08:54

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: SHERMAN

DATE/TIME OF TEST: 12/06/2020 02:08:18

Diagnostics Check		OK	01:59	
Air Blank		0.000	02:00	
Control Test		0.077	02:00	
Air Blank		0.000	02:01	
Subject Sample	# 1	0.138	02:04	
Breath Volume		1.746	02:04	
Air Blank		0.000	02:04	
Air Blank		0.000	02:06	
Subject Sample	# 2		02:06	Subject Test Refused
Breath Volume		0.000	02:06	
Air Blank		0.000	02:07	
Control Test		0.076	02:07	
Air Blank		0.000	02:08	
Diagnostics Check		OK	02:08	

Data Download Date/Time: 12/09/2020 08:54

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NELSON,OFFICER
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 12/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: OELENSCHLAGER

DATE/TIME OF TEST: 12/16/2020 00:47:01

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.080	00:41
Air Blank		0.000	00:42
Subject Sample	# 1	0.026	00:42
Breath Volume		2.894	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.025	00:45
Breath Volume		3.000	00:45
Air Blank		0.000	00:46
Control Test		0.079	00:46
Air Blank		0.000	00:46
Diagnostics Check		OK	00:47

Data Download Date/Time: 01/29/2021 11:09

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: RUFFIN

DATE/TIME OF TEST: 12/16/2020 03:22:00

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.079	03:16
Air Blank		0.000	03:16
Subject Sample	# 1	0.164	03:17
Breath Volume		3.960	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.173	03:20
Breath Volume		4.136	03:20
Air Blank		0.000	03:21
Control Test		0.078	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:22

Data Download Date/Time: 01/29/2021 11:09

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:37

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/18/2020 16:06:43

Diagnostics Check		OK	15:57
Air Blank		0.000	15:57
Control Test		0.080	15:58
Air Blank		0.000	15:58
Subject Sample	# 1	0.335	15:59
Breath Volume		2.492	15:59
Air Blank		0.000	16:00
Air Blank		0.000	16:01
Subject Sample	# 2	0.312	16:02
Breath Volume		1.828	16:02
Air Blank		0.000	16:02
Air Blank		0.000	16:04
Subject Sample	# 3	0.312	16:05
Breath Volume		2.226	16:05
Air Blank		0.000	16:05
Control Test		0.080	16:06
Air Blank		0.000	16:06
Diagnostics Check		OK	16:06

Data Download Date/Time: 01/29/2021 11:09

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: LIEBENBERG

DATE/TIME OF TEST: 12/18/2020 23:40:54

Diagnostics Check		OK	23:34
Air Blank		0.000	23:35
Control Test		0.080	23:35
Air Blank		0.000	23:36
Subject Sample	# 1	0.150	23:36
Breath Volume		3.515	23:36
Air Blank		0.000	23:37
Air Blank		0.000	23:38
Subject Sample	# 2	0.145	23:39
Breath Volume		2.855	23:39
Air Blank		0.000	23:40
Control Test		0.080	23:40
Air Blank		0.000	23:40
Diagnostics Check		OK	23:40

Data Download Date/Time: 01/29/2021 11:09

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NELSON,OFC
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 12/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: SCHWIND

DATE/TIME OF TEST: 12/19/2020 03:24:56

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.213	03:20
Breath Volume		2.394	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:22
Subject Sample	# 2	0.210	03:23
Breath Volume		2.546	03:23
Air Blank		0.000	03:24
Control Test		0.079	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 01/29/2021 11:09

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,M
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: MIGASHKIN

DATE/TIME OF TEST: 12/19/2020 23:50:13

Diagnostics Check		OK	23:43
Air Blank		0.000	23:44
Control Test		0.079	23:44
Air Blank		0.000	23:44
Subject Sample	# 1	0.246	23:45
Breath Volume		3.105	23:45
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.253	23:48
Breath Volume		3.644	23:48
Air Blank		0.000	23:49
Control Test		0.078	23:49
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 01/29/2021 11:09

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: PECH

DATE/TIME OF TEST: 12/24/2020 23:47:35

Diagnostics Check		OK	23:41
Air Blank		0.000	23:41
Control Test		0.079	23:41
Air Blank		0.000	23:42
Subject Sample	# 1	0.227	23:43
Breath Volume		2.394	23:43
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.220	23:46
Breath Volume		1.437	23:46
Air Blank		0.000	23:46
Control Test		0.078	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 01/29/2021 11:09

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: ROSARIO HERNANDEZ

DATE/TIME OF TEST: 12/01/2020 00:18:10

Diagnostics Check		OK	00:09
Air Blank		0.000	00:10
Control Test		0.081	00:10
Air Blank		0.000	00:11
Subject Sample	# 1	0.230	00:12
Breath Volume		1.640	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.226	00:16
Breath Volume		1.453	00:16
Air Blank		0.000	00:17
Control Test		0.080	00:17
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 12/11/2020 08:37

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 12/03/2020 02:34:46

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.082	02:29
Air Blank		0.000	02:30
Subject Sample	# 1	0.078	02:30
Breath Volume		1.527	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:32
Subject Sample	# 2	0.077	02:33
Breath Volume		2.941	02:33
Air Blank		0.000	02:33
Control Test		0.082	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 12/11/2020 08:37

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: MONTELONGO

DATE/TIME OF TEST: 12/04/2020 05:04:54

Diagnostics Check		OK	04:51	
Air Blank		0.000	04:51	
Control Test		0.081	04:52	
Air Blank		0.000	04:52	
Subject Sample	# 1	0.000	04:55	Volume Not Met
Breath Volume		0.062	04:55	
Air Blank		0.000	04:56	
Air Blank		0.000	04:58	
Subject Sample	# 2	0.207	04:59	
Breath Volume		1.714	04:59	
Air Blank		0.000	04:59	
Air Blank		0.000	05:01	
Subject Sample	# 3	0.196	05:03	
Breath Volume		2.214	05:03	
Air Blank		0.000	05:03	
Control Test		0.081	05:04	
Air Blank		0.000	05:04	
Diagnostics Check		OK	05:04	

Data Download Date/Time: 12/11/2020 08:37

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:43

SUBJECT'S LAST NAME: IGHILE

DATE/TIME OF TEST: 12/04/2020 06:11:23

Diagnostics Check		OK	06:04
Air Blank		0.000	06:04
Control Test		0.081	06:05
Air Blank		0.000	06:05
Subject Sample	# 1	0.151	06:06
Breath Volume		1.101	06:06
Air Blank		0.000	06:07
Air Blank		0.000	06:08
Subject Sample	# 2	0.157	06:09
Breath Volume		2.656	06:09
Air Blank		0.000	06:10
Control Test		0.081	06:10
Air Blank		0.000	06:11
Diagnostics Check		OK	06:11

Data Download Date/Time: 12/11/2020 08:37

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JOHNSON,B
ARRESTING OFFICER AGENCY: Macdill Air Force Ba
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 12/05/2020 01:17:31

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.081	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.284	01:13
Breath Volume		2.097	01:13
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.292	01:15
Breath Volume		2.222	01:15
Air Blank		0.000	01:16
Control Test		0.081	01:16
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 12/11/2020 08:37

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/05/2020 04:01:11

Diagnostics Check		OK	03:44	
Air Blank		0.000	03:44	
Control Test		0.081	03:44	
Air Blank		0.000	03:45	
Subject Sample	# 1	0.063	03:48	Volume Not Met
Breath Volume		0.019	03:48	
Air Blank		0.000	03:49	
Air Blank		0.000	03:51	
Subject Sample	# 2	0.069	03:53	
Breath Volume		1.441	03:53	
Air Blank		0.000	03:54	
Air Blank		0.000	03:56	
Subject Sample	# 3	0.076	03:59	Volume Not Met
Breath Volume		0.550	03:59	
Air Blank		0.000	04:00	
Control Test		0.082	04:00	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 12/11/2020 08:37

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/05/2020 04:13:59

Diagnostics Check		OK	04:05
Air Blank		0.000	04:05
Control Test		0.081	04:05
Air Blank		0.000	04:06
Subject Sample	# 1	0.078	04:08
Breath Volume		1.753	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:10
Subject Sample	# 2	0.072	04:12
Breath Volume		1.558	04:12
Air Blank		0.000	04:13
Control Test		0.081	04:13
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 12/11/2020 08:37

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: LOVING

DATE/TIME OF TEST: 12/05/2020 20:24:42

Diagnostics Check		OK	20:17
Air Blank		0.000	20:17
Control Test		0.081	20:17
Air Blank		0.000	20:18
Subject Sample	# 1	0.207	20:19
Breath Volume		2.652	20:19
Air Blank		0.000	20:20
Air Blank		0.000	20:22
Subject Sample	# 2	0.208	20:23
Breath Volume		2.132	20:23
Air Blank		0.000	20:23
Control Test		0.081	20:24
Air Blank		0.000	20:24
Diagnostics Check		OK	20:24

Data Download Date/Time: 12/11/2020 08:37

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: SANER

DATE/TIME OF TEST: 12/05/2020 23:53:41

Diagnostics Check		OK	23:47
Air Blank		0.000	23:47
Control Test		0.081	23:48
Air Blank		0.000	23:48
Subject Sample	# 1	0.183	23:49
Breath Volume		1.628	23:49
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.175	23:52
Breath Volume		1.597	23:52
Air Blank		0.000	23:52
Control Test		0.081	23:53
Air Blank		0.000	23:53
Diagnostics Check		OK	23:53

Data Download Date/Time: 12/11/2020 08:37

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: OYOLA

DATE/TIME OF TEST: 12/06/2020 19:36:08

Diagnostics Check		OK	19:29
Air Blank		0.000	19:30
Control Test		0.081	19:30
Air Blank		0.000	19:31
Subject Sample	# 1	0.000	19:31
Breath Volume		2.824	19:31
Air Blank		0.000	19:32
Air Blank		0.000	19:34
Subject Sample	# 2	0.000	19:34
Breath Volume		2.746	19:34
Air Blank		0.000	19:35
Control Test		0.081	19:35
Air Blank		0.000	19:36
Diagnostics Check		OK	19:36

Data Download Date/Time: 12/11/2020 08:37

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: YEBRA

DATE/TIME OF TEST: 12/08/2020 03:57:46

Diagnostics Check		OK	03:48	
Air Blank		0.000	03:48	
Control Test		0.082	03:49	
Air Blank		0.000	03:49	
Subject Sample	# 1		03:52	No Sample Provided
Breath Volume		0.000	03:52	
Air Blank		0.000	03:53	
Air Blank		0.000	03:55	
Subject Sample	# 2		03:56	Subject Test Refused
Breath Volume		0.000	03:56	
Air Blank		0.000	03:56	
Control Test		0.082	03:57	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 12/11/2020 08:37

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: CHRISTO

DATE/TIME OF TEST: 12/09/2020 05:27:42

Diagnostics Check		OK	05:20
Air Blank		0.000	05:20
Control Test		0.082	05:21
Air Blank		0.000	05:21
Subject Sample	# 1	0.000	05:23
Breath Volume		3.000	05:23
Air Blank		0.000	05:23
Air Blank		0.000	05:25
Subject Sample	# 2	0.000	05:26
Breath Volume		1.972	05:26
Air Blank		0.000	05:26
Control Test		0.082	05:27
Air Blank		0.000	05:27
Diagnostics Check		OK	05:27

Data Download Date/Time: 12/11/2020 08:37

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARCIA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 12/09/2020 16:49:47

Diagnostics Check		OK	16:42
Air Blank		0.000	16:42
Control Test		0.082	16:43
Air Blank		0.000	16:43
Subject Sample	# 1	0.260	16:44
Breath Volume		2.269	16:44
Air Blank		0.000	16:45
Air Blank		0.000	16:47
Subject Sample	# 2	0.257	16:48
Breath Volume		1.945	16:48
Air Blank		0.000	16:48
Control Test		0.081	16:49
Air Blank		0.000	16:49
Diagnostics Check		OK	16:49

Data Download Date/Time: 12/11/2020 08:37

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: HOURIHAN

DATE/TIME OF TEST: 12/31/2020 00:03:55

Diagnostics Check		OK	00:01	
Air Blank		0.000	00:01	
Control Test		0.079	00:01	
Air Blank		0.000	00:02	
Subject Sample	# 1		00:02	Subject Test Refused
Breath Volume		0.000	00:02	
Air Blank		0.000	00:02	
Control Test		0.079	00:03	
Air Blank		0.000	00:03	
Diagnostics Check		OK	00:03	

Data Download Date/Time: 01/23/2021 08:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 12/31/2020 00:38:18

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:35	
Control Test		0.079	00:36	
Air Blank		0.000	00:36	
Subject Sample	# 1		00:36	Subject Test Refused
Breath Volume		0.000	00:36	
Air Blank		0.000	00:37	
Control Test		0.079	00:37	
Air Blank		0.000	00:38	
Diagnostics Check		OK	00:38	

Data Download Date/Time: 01/23/2021 08:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: ZIGA

DATE/TIME OF TEST: 12/31/2020 01:51:03

Diagnostics Check		OK	01:43
Air Blank		0.000	01:44
Control Test		0.079	01:44
Air Blank		0.000	01:45
Subject Sample	# 1	0.123	01:46
Breath Volume		1.371	01:46
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.118	01:49
Breath Volume		2.121	01:49
Air Blank		0.000	01:50
Control Test		0.077	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:51

Data Download Date/Time: 01/23/2021 08:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: GOLDSTEIN

DATE/TIME OF TEST: 12/31/2020 22:15:44

Diagnostics Check		OK	22:12	
Air Blank		0.000	22:13	
Control Test		0.079	22:13	
Air Blank		0.000	22:14	
Subject Sample	# 1		22:14	Subject Test Refused
Breath Volume		0.000	22:14	
Air Blank		0.000	22:14	
Control Test		0.079	22:15	
Air Blank		0.000	22:15	
Diagnostics Check		OK	22:15	

Data Download Date/Time: 01/23/2021 08:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICKETSON,JAY,R
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: WAYCHOFF,COLLIN
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 12/09/2020 16:06:44

Diagnostics Check		OK	16:03	
Air Blank		0.000	16:04	
Control Test		0.079	16:04	
Air Blank		0.000	16:05	
Subject Sample	# 1		16:06	Sequence Aborted
Breath Volume		0.000	16:06	
Air Blank		0.000	16:06	

Data Download Date/Time: 12/22/2020 10:33

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICKETSON,JAY,R
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: WAYCHOFF,COLLIN
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 12/09/2020 16:23:10

Diagnostics Check		OK	16:16
Air Blank		0.000	16:16
Control Test		0.078	16:16
Air Blank		0.000	16:17
Subject Sample	# 1	0.259	16:18
Breath Volume		3.320	16:18
Air Blank		0.000	16:19
Air Blank		0.000	16:21
Subject Sample	# 2	0.253	16:21
Breath Volume		3.234	16:21
Air Blank		0.000	16:22
Control Test		0.075	16:22
Air Blank		0.000	16:23
Diagnostics Check		OK	16:23

Data Download Date/Time: 12/22/2020 10:33

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,CHARLES,J
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: THIGPEN,JORDAN
ARRESTING OFFICER AGENCY: Taylor County SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 12/15/2020 18:17:01

Diagnostics Check		OK	18:06	
Air Blank		0.000	18:06	
Control Test		0.079	18:07	
Air Blank		0.000	18:07	
Subject Sample	# 1		18:10	No Sample Provided
Breath Volume		0.000	18:10	
Air Blank		0.000	18:11	
Air Blank		0.000	18:13	
Subject Sample	# 2		18:15	Subject Test Refused
Breath Volume		0.019	18:15	
Air Blank		0.000	18:15	
Control Test		0.079	18:16	
Air Blank			18:16	RFI Detect
Air Blank		0.000	18:17	

Data Download Date/Time: 12/22/2020 10:33

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELPS,KOVEN,
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: PHELPS,KOVEN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: KEEN

DATE/TIME OF TEST: 12/19/2020 00:37:56

Diagnostics Check		OK	00:24	
Air Blank		0.000	00:25	
Control Test		0.081	00:25	
Air Blank		0.000	00:26	
Subject Sample	# 1		00:29	No Sample Provided
Breath Volume		0.000	00:29	
Air Blank		0.000	00:29	
Air Blank		0.000	00:31	
Subject Sample	# 2	0.255	00:32	
Breath Volume		1.863	00:32	
Air Blank		0.000	00:33	
Air Blank		0.000	00:35	
Subject Sample	# 3	0.232	00:36	
Breath Volume		2.093	00:36	
Air Blank		0.000	00:37	
Control Test		0.072	00:37	Control Outside Tolerance
Air Blank		0.000	00:37	

Data Download Date/Time: 12/22/2020 10:33

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELPS,KOVEN,
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: PHELPS,KOVEN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: KEEN

DATE/TIME OF TEST: 12/19/2020 00:46:30

Diagnostics Check		OK	00:42	
Air Blank		0.000	00:42	
Control Test		0.077	00:42	
Air Blank		0.000	00:43	
Subject Sample	# 1	0.249	00:44	
Breath Volume		3.597	00:44	
Air Blank			00:45	Purge Fail
Air Blank			00:46	Purge Fail

Data Download Date/Time: 12/22/2020 10:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,MATTHEW,M
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: SMITH,MATTHEW
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 12/21/2020 19:45:14

Diagnostics Check	OK	19:43	
Air Blank	0.000	19:44	
Control Test	0.079	19:44	
Air Blank		19:44	Improper Sample
Air Blank	0.000	19:45	

Data Download Date/Time: 12/22/2020 10:33

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,MATTHEW,M
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: SMITH,MATTHEW
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 12/21/2020 19:53:16

Diagnostics Check		OK	19:47
Air Blank		0.000	19:47
Control Test		0.080	19:47
Air Blank		0.000	19:48
Subject Sample	# 1	0.000	19:49
Breath Volume		2.402	19:49
Air Blank		0.000	19:49
Air Blank		0.000	19:51
Subject Sample	# 2	0.000	19:51
Breath Volume		2.554	19:51
Air Blank		0.000	19:52
Control Test		0.078	19:52
Air Blank		0.000	19:53
Diagnostics Check		OK	19:53

Data Download Date/Time: 12/22/2020 10:33

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,MATTHEW,M
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: SMITH,MATTHEW
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: ROMANO

DATE/TIME OF TEST: 12/21/2020 22:22:31

Diagnostics Check		OK	22:15	
Air Blank		0.000	22:15	
Control Test		0.080	22:15	
Air Blank		0.000	22:16	
Subject Sample	# 1	0.093	22:17	
Breath Volume		2.546	22:17	
Air Blank		0.000	22:17	
Air Blank		0.000	22:19	
Subject Sample	# 2	0.094	22:20	
Breath Volume		2.957	22:20	
Air Blank		0.000	22:21	
Control Test		0.079	22:21	
Air Blank			22:22	RFI Detect
Air Blank		0.000	22:22	

Data Download Date/Time: 12/22/2020 10:34

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,MATTHEW,M
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: SMITH,MATTHEW
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 12/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 12/22/2020 23:09:05

Diagnostics Check		OK	23:04	
Air Blank		0.000	23:05	
Control Test		0.082	23:05	
Air Blank		0.000	23:06	
Subject Sample	# 1		23:07	Slope Not Met
Breath Volume		0.476	23:07	
Air Blank		0.000	23:08	
Control Test		0.081	23:08	
Air Blank		0.000	23:08	
Diagnostics Check		OK	23:09	

Data Download Date/Time: 01/26/2021 10:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,MATTHEW,M
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: SMITH,MATTHEW
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 12/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 12/22/2020 23:40:46

Diagnostics Check		OK	23:28	
Air Blank		0.000	23:28	
Control Test		0.079	23:28	
Air Blank		0.000	23:29	
Subject Sample	# 1	0.159	23:30	
Breath Volume		1.312	23:30	
Air Blank		0.000	23:30	
Air Blank		0.000	23:32	
Subject Sample	# 2	0.147	23:35	Volume Not Met
Breath Volume		1.035	23:35	
Air Blank		0.000	23:36	
Air Blank		0.000	23:38	
Subject Sample	# 3	0.169	23:39	
Breath Volume		2.957	23:39	
Air Blank		0.000	23:39	
Control Test		0.076	23:40	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 01/26/2021 10:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINS,KYLE,L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: MCCURRY,KELLEY
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/18/2020 04:40:51

Diagnostics Check		OK	04:33
Air Blank		0.000	04:34
Control Test		0.082	04:34
Air Blank		0.000	04:35
Subject Sample	# 1	0.192	04:35
Breath Volume		1.582	04:35
Air Blank		0.000	04:36
Air Blank		0.000	04:38
Subject Sample	# 2	0.190	04:38
Breath Volume		1.371	04:38
Air Blank		0.000	04:39
Control Test		0.079	04:40
Air Blank		0.000	04:40
Diagnostics Check		OK	04:40

Data Download Date/Time: 01/12/2021 11:41

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,DANIEL,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DAVIS,DANIEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: ORREGO GOMEZ

DATE/TIME OF TEST: 12/18/2020 21:48:59

Diagnostics Check		OK	21:42
Air Blank		0.000	21:42
Control Test		0.082	21:43
Air Blank		0.000	21:43
Subject Sample	# 1	0.200	21:44
Breath Volume		2.207	21:44
Air Blank		0.000	21:45
Air Blank		0.000	21:46
Subject Sample	# 2	0.201	21:47
Breath Volume		2.398	21:47
Air Blank		0.000	21:48
Control Test		0.078	21:48
Air Blank		0.000	21:48
Diagnostics Check		OK	21:48

Data Download Date/Time: 01/12/2021 11:41

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINS,KYLE,L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: RODRIGUEZ,EDDIE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 12/28/2020 04:18:52

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.081	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.047	04:13
Breath Volume		1.332	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.041	04:17
Breath Volume		1.519	04:17
Air Blank		0.000	04:17
Control Test		0.082	04:18
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 01/12/2021 11:41

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT,KELLY,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: JOHNSTON,TORI
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: KALSOW

DATE/TIME OF TEST: 12/29/2020 01:42:13

Diagnostics Check		OK	01:34
Air Blank		0.000	01:34
Control Test		0.081	01:35
Air Blank		0.000	01:35
Subject Sample	# 1	0.147	01:36
Breath Volume		1.773	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.147	01:40
Breath Volume		2.800	01:40
Air Blank		0.000	01:41
Control Test		0.079	01:41
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 01/12/2021 11:42

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT,KELLY,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH,TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: FERRARO

DATE/TIME OF TEST: 12/29/2020 01:54:52

Diagnostics Check		OK	01:47
Air Blank		0.000	01:47
Control Test		0.079	01:48
Air Blank		0.000	01:48
Subject Sample	# 1	0.202	01:49
Breath Volume		2.062	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.205	01:52
Breath Volume		2.308	01:52
Air Blank		0.000	01:53
Control Test		0.079	01:54
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 01/12/2021 11:42

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERONI,KRISTI,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: JOHNSTON,T
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: CIRER

DATE/TIME OF TEST: 12/31/2020 22:10:07

Diagnostics Check		OK	22:06	
Air Blank		0.000	22:06	
Control Test		0.081	22:06	
Air Blank		0.000	22:07	
Subject Sample	# 1		22:08	Slope Not Met
Breath Volume		0.953	22:08	
Air Blank		0.000	22:09	
Control Test		0.081	22:09	
Air Blank		0.000	22:10	
Diagnostics Check		OK	22:10	

Data Download Date/Time: 01/12/2021 11:42

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERONI,KRISTI,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: JOHNSTON,TORI
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: CIRER

DATE/TIME OF TEST: 12/31/2020 22:20:44

Diagnostics Check		OK	22:12
Air Blank		0.000	22:13
Control Test		0.080	22:13
Air Blank		0.000	22:14
Subject Sample	# 1	0.256	22:15
Breath Volume		1.886	22:15
Air Blank		0.000	22:16
Air Blank		0.000	22:17
Subject Sample	# 2	0.241	22:18
Breath Volume		1.460	22:18
Air Blank		0.000	22:19
Control Test		0.079	22:20
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 01/12/2021 11:42

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,MARK
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: GORDILS

DATE/TIME OF TEST: 12/04/2020 02:00:32

Diagnostics Check		OK	01:57	
Air Blank		0.000	01:58	
Control Test		0.082	01:58	
Air Blank		0.000	01:58	
Subject Sample	# 1		01:59	Subject Test Refused
Breath Volume		0.000	01:59	
Air Blank		0.000	01:59	
Control Test		0.083	01:59	
Air Blank		0.000	02:00	
Diagnostics Check		OK	02:00	

Data Download Date/Time: 12/19/2020 02:33

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: TELLEZ-AVILA

DATE/TIME OF TEST: 12/06/2020 06:17:19

Diagnostics Check		OK	06:08	
Air Blank		0.000	06:09	
Control Test		0.081	06:09	
Air Blank		0.000	06:10	
Subject Sample	# 1		06:13	No Sample Provided
Breath Volume		0.000	06:13	
Air Blank		0.000	06:13	
Air Blank		0.000	06:15	
Subject Sample	# 2		06:15	Subject Test Refused
Breath Volume		0.000	06:15	
Air Blank		0.000	06:16	
Control Test		0.081	06:16	
Air Blank		0.000	06:17	
Diagnostics Check		OK	06:17	

Data Download Date/Time: 12/19/2020 02:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: PEREZ GONZALEZ

DATE/TIME OF TEST: 12/11/2020 05:46:23

Diagnostics Check		OK	05:39
Air Blank		0.000	05:40
Control Test		0.082	05:40
Air Blank		0.000	05:40
Subject Sample	# 1	0.138	05:41
Breath Volume		1.277	05:41
Air Blank		0.000	05:42
Air Blank		0.000	05:44
Subject Sample	# 2	0.150	05:44
Breath Volume		2.015	05:44
Air Blank		0.000	05:45
Control Test		0.083	05:45
Air Blank		0.000	05:46
Diagnostics Check		OK	05:46

Data Download Date/Time: 12/19/2020 02:33

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: LEGUISAMON JR

DATE/TIME OF TEST: 12/13/2020 03:24:51

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.082	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.215	03:20
Breath Volume		2.648	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:22
Subject Sample	# 2	0.209	03:23
Breath Volume		2.027	03:23
Air Blank		0.000	03:23
Control Test		0.082	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 12/19/2020 02:33

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: SANCHEZ-RIVERA

DATE/TIME OF TEST: 12/17/2020 01:42:35

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.081	01:33
Air Blank		0.000	01:34
Subject Sample	# 1	0.098	01:36
Breath Volume		1.523	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.103	01:40
Breath Volume		1.292	01:40
Air Blank		0.000	01:41
Control Test		0.080	01:41
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 12/19/2020 02:33

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: MENDOZA-ROBLES

DATE/TIME OF TEST: 12/31/2020 04:34:05

Diagnostics Check		OK	04:27
Air Blank		0.000	04:28
Control Test		0.082	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.138	04:29
Breath Volume		1.589	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.141	04:32
Breath Volume		1.546	04:32
Air Blank		0.000	04:33
Control Test		0.081	04:33
Air Blank		0.000	04:33
Diagnostics Check		OK	04:34

Data Download Date/Time: 01/26/2021 11:32

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: TROCHE

DATE/TIME OF TEST: 12/05/2020 05:03:41

Diagnostics Check	OK	05:02	
Air Blank		05:02	Ambient Fail
Air Blank		05:03	Purge Fail

Data Download Date/Time: 12/19/2020 02:39

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: TROCHE

DATE/TIME OF TEST: 12/05/2020 05:07:44

Diagnostics Check	OK	05:06	
Air Blank		05:07	Ambient Fail
Air Blank	0.000	05:07	

Data Download Date/Time: 12/19/2020 02:39

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: TROCHE

DATE/TIME OF TEST: 12/05/2020 05:23:48

Diagnostics Check		OK	05:17
Air Blank		0.000	05:18
Control Test		0.079	05:18
Air Blank		0.000	05:19
Subject Sample	# 1	0.049	05:19
Breath Volume		2.480	05:19
Air Blank		0.000	05:20
Air Blank		0.000	05:21
Subject Sample	# 2	0.050	05:22
Breath Volume		2.742	05:22
Air Blank		0.000	05:22
Control Test		0.079	05:23
Air Blank		0.000	05:23
Diagnostics Check		OK	05:23

Data Download Date/Time: 12/19/2020 02:39

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: RAWLUK,S
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/25/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: BENTLEY

DATE/TIME OF TEST: 12/06/2020 02:40:54

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.078	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.168	02:36
Breath Volume		1.757	02:36
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.169	02:39
Breath Volume		1.664	02:39
Air Blank		0.000	02:39
Control Test		0.077	02:40
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 12/19/2020 02:39

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT,KELLY,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: WITT,KELLY
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: DELVECCHIO

DATE/TIME OF TEST: 12/06/2020 17:32:12

Diagnostics Check		OK	17:28	
Air Blank		0.000	17:29	
Control Test		0.081	17:29	
Air Blank		0.000	17:30	
Subject Sample	# 1		17:30	Subject Test Refused
Breath Volume		0.000	17:30	
Air Blank		0.000	17:31	
Control Test		0.079	17:31	
Air Blank		0.000	17:32	
Diagnostics Check		OK	17:32	

Data Download Date/Time: 12/16/2020 13:09

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROBEL,RAYMOND,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: STROBEL,RAYMOND
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: WHITNEY

DATE/TIME OF TEST: 12/08/2020 13:48:11

Diagnostics Check		OK	13:41
Air Blank		0.000	13:42
Control Test		0.080	13:42
Air Blank		0.000	13:42
Subject Sample	# 1	0.000	13:43
Breath Volume		1.871	13:43
Air Blank		0.000	13:44
Air Blank		0.000	13:46
Subject Sample	# 2	0.000	13:46
Breath Volume		2.074	13:46
Air Blank		0.000	13:47
Control Test		0.079	13:47
Air Blank		0.000	13:48
Diagnostics Check		OK	13:48

Data Download Date/Time: 12/16/2020 13:09

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYRNIOS,GEORGE,T
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMYRNIOS,GEORGE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: MADIGAN

DATE/TIME OF TEST: 12/10/2020 16:51:54

Diagnostics Check		OK	16:45
Air Blank		0.000	16:46
Control Test		0.080	16:46
Air Blank		0.000	16:47
Subject Sample	# 1	0.000	16:47
Breath Volume		3.441	16:47
Air Blank		0.000	16:47
Air Blank		0.000	16:49
Subject Sample	# 2	0.000	16:50
Breath Volume		3.363	16:50
Air Blank		0.000	16:50
Control Test		0.079	16:51
Air Blank		0.000	16:51
Diagnostics Check		OK	16:51

Data Download Date/Time: 12/16/2020 13:09

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT,KELLY,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: WITT,KELLY
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: LANG

DATE/TIME OF TEST: 12/11/2020 22:15:55

Diagnostics Check		OK	22:09
Air Blank		0.000	22:09
Control Test		0.080	22:10
Air Blank		0.000	22:10
Subject Sample	# 1	0.150	22:11
Breath Volume		1.925	22:11
Air Blank		0.000	22:12
Air Blank		0.000	22:13
Subject Sample	# 2	0.146	22:14
Breath Volume		2.101	22:14
Air Blank		0.000	22:14
Control Test		0.078	22:15
Air Blank		0.000	22:15
Diagnostics Check		OK	22:15

Data Download Date/Time: 12/16/2020 13:10

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAK,KYLE,F
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LEE,JASMINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 12/14/2020 20:25:48

Diagnostics Check		OK	20:20	
Air Blank		0.000	20:20	
Control Test		0.079	20:20	
Air Blank		0.000	20:21	
Subject Sample	# 1	0.150	20:24	Volume Not Met
Breath Volume		0.070	20:24	
Air Blank			20:25	Purge Fail
Air Blank		0.000	20:25	

Data Download Date/Time: 12/16/2020 13:10

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAK,KYLE,F
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LEE,JASMINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 12/14/2020 20:31:29

Diagnostics Check		OK	20:28	
Air Blank		0.000	20:29	
Control Test		0.077	20:29	
Air Blank		0.000	20:29	
Subject Sample	# 1		20:30	Subject Test Refused
Breath Volume		0.000	20:30	
Air Blank		0.000	20:30	
Control Test		0.078	20:30	
Air Blank		0.000	20:31	
Diagnostics Check		OK	20:31	

Data Download Date/Time: 12/16/2020 13:10

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,DAVID,L
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: FERNANDEZ,SHELSY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: ROONEY

DATE/TIME OF TEST: 12/03/2020 03:42:05

Diagnostics Check		OK	03:39	
Air Blank		0.000	03:40	
Control Test		0.081	03:40	
Air Blank		0.000	03:40	
Subject Sample	# 1		03:41	Subject Test Refused
Breath Volume		0.000	03:41	
Air Blank			03:41	RFI Detect
Air Blank		0.000	03:42	

Data Download Date/Time: 12/03/2020 10:09

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: DAVIS-HICKS

DATE/TIME OF TEST: 12/05/2020 03:11:28

Diagnostics Check		OK	03:04
Air Blank		0.000	03:05
Control Test		0.080	03:05
Air Blank		0.000	03:06
Subject Sample	# 1	0.210	03:06
Breath Volume		1.105	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.206	03:09
Breath Volume		1.917	03:09
Air Blank		0.000	03:10
Control Test		0.078	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 12/28/2020 15:48

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: ROJAS-MENDEZ

DATE/TIME OF TEST: 12/06/2020 03:08:00

Diagnostics Check		OK	03:01
Air Blank		0.000	03:02
Control Test		0.081	03:02
Air Blank		0.000	03:02
Subject Sample	# 1	0.151	03:03
Breath Volume		2.914	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:05
Subject Sample	# 2	0.151	03:06
Breath Volume		1.433	03:06
Air Blank		0.000	03:07
Control Test		0.080	03:07
Air Blank		0.000	03:07
Diagnostics Check		OK	03:08

Data Download Date/Time: 12/28/2020 15:49

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANT,ANTHONY,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 11/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 12/16/2020 02:43:27

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.080	02:36
Air Blank		0.000	02:36
Subject Sample	# 1	0.146	02:38
Breath Volume		1.652	02:38
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.140	02:41
Breath Volume		2.218	02:41
Air Blank		0.000	02:42
Control Test		0.078	02:42
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 12/28/2020 15:49

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HANSELL,ROBERT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/27/2020 22:47:09

Diagnostics Check	OK	22:46	
Air Blank		22:46	RFI Detect
Air Blank	0.000	22:47	

Data Download Date/Time: 01/20/2021 07:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HANSELL,ROBERT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/27/2020 23:02:56

Diagnostics Check		OK	22:55
Air Blank		0.000	22:55
Control Test		0.081	22:55
Air Blank		0.000	22:56
Subject Sample	# 1	0.117	22:58
Breath Volume		1.687	22:58
Air Blank		0.000	22:59
Air Blank		0.000	23:00
Subject Sample	# 2	0.113	23:01
Breath Volume		1.730	23:01
Air Blank		0.000	23:02
Control Test		0.078	23:02
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 01/20/2021 07:07

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ANTONETTI,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 11/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MUNOZ NEVAREZ

DATE/TIME OF TEST: 12/01/2020 01:58:49

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.080	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.190	01:54
Breath Volume		1.738	01:54
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.189	01:57
Breath Volume		2.636	01:57
Air Blank		0.000	01:57
Control Test		0.078	01:58
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 12/12/2020 10:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALKER,JODI
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: ILVES

DATE/TIME OF TEST: 12/01/2020 02:49:19

Diagnostics Check		OK	02:42
Air Blank		0.000	02:43
Control Test		0.081	02:43
Air Blank		0.000	02:44
Subject Sample	# 1	0.159	02:44
Breath Volume		2.417	02:44
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.156	02:47
Breath Volume		1.937	02:47
Air Blank		0.000	02:48
Control Test		0.079	02:48
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 12/12/2020 10:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KEEL,SHELDON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 12/01/2020 23:54:12

Diagnostics Check		OK	23:50	
Air Blank		0.000	23:50	
Control Test		0.081	23:51	
Air Blank		0.000	23:51	
Subject Sample	# 1		23:52	Subject Test Refused
Breath Volume		0.066	23:52	
Air Blank		0.000	23:53	
Control Test		0.081	23:53	
Air Blank		0.000	23:54	
Diagnostics Check		OK	23:54	

Data Download Date/Time: 12/12/2020 10:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 12/03/2020 00:42:49

Diagnostics Check		OK	00:30	
Air Blank		0.000	00:30	
Control Test		0.081	00:31	
Air Blank		0.000	00:31	
Subject Sample	# 1	0.177	00:32	
Breath Volume		1.195	00:32	
Air Blank		0.000	00:32	
Air Blank		0.000	00:34	
Subject Sample	# 2	0.158	00:37	Volume Not Met
Breath Volume		0.437	00:37	
Air Blank		0.000	00:38	
Air Blank		0.000	00:40	
Subject Sample	# 3	0.173	00:41	
Breath Volume		1.269	00:41	
Air Blank		0.000	00:41	
Control Test		0.077	00:42	
Air Blank		0.000	00:42	
Diagnostics Check		OK	00:42	

Data Download Date/Time: 12/12/2020 10:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCGILLICUDDY,STEVEN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 11/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: GABRIEL

DATE/TIME OF TEST: 12/04/2020 00:54:06

Diagnostics Check		OK	00:46
Air Blank		0.000	00:47
Control Test		0.081	00:47
Air Blank		0.000	00:48
Subject Sample	# 1	0.216	00:49
Breath Volume		3.113	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:51
Subject Sample	# 2	0.199	00:52
Breath Volume		2.558	00:52
Air Blank		0.000	00:53
Control Test		0.077	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 12/12/2020 10:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HARVEY,ANTHONY
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 11/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: O'DONNELL

DATE/TIME OF TEST: 12/04/2020 02:00:13

Diagnostics Check		OK	01:52
Air Blank		0.000	01:53
Control Test		0.080	01:53
Air Blank		0.000	01:54
Subject Sample	# 1	0.184	01:55
Breath Volume		3.046	01:55
Air Blank		0.000	01:55
Air Blank		0.000	01:57
Subject Sample	# 2	0.183	01:58
Breath Volume		2.738	01:58
Air Blank		0.000	01:59
Control Test		0.078	01:59
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

Data Download Date/Time: 12/12/2020 10:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FERNANDEZ,ALAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 12/04/2020 03:56:28

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.080	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.139	03:51
Breath Volume		1.218	03:51
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.139	03:54
Breath Volume		1.957	03:54
Air Blank		0.000	03:55
Control Test		0.079	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 12/12/2020 10:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: GELLERMANN

DATE/TIME OF TEST: 12/05/2020 03:41:30

Diagnostics Check		OK	03:35
Air Blank		0.000	03:35
Control Test		0.081	03:35
Air Blank		0.000	03:36
Subject Sample	# 1	0.154	03:36
Breath Volume		1.894	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.156	03:39
Breath Volume		2.351	03:39
Air Blank		0.000	03:40
Control Test		0.080	03:40
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 12/12/2020 10:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SMITH,MARK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: GOTAY

DATE/TIME OF TEST: 12/05/2020 04:07:02

Diagnostics Check		OK	03:59
Air Blank		0.000	04:00
Control Test		0.081	04:00
Air Blank		0.000	04:00
Subject Sample	# 1	0.164	04:02
Breath Volume		3.167	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:04
Subject Sample	# 2	0.160	04:05
Breath Volume		2.863	04:05
Air Blank		0.000	04:06
Control Test		0.078	04:06
Air Blank		0.000	04:06
Diagnostics Check		OK	04:07

Data Download Date/Time: 12/12/2020 10:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: THOMAS,MICAH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: BORGES

DATE/TIME OF TEST: 12/05/2020 14:16:10

Diagnostics Check		OK	14:09
Air Blank		0.000	14:09
Control Test		0.081	14:10
Air Blank		0.000	14:10
Subject Sample	# 1	0.000	14:11
Breath Volume		1.898	14:11
Air Blank		0.000	14:11
Air Blank		0.000	14:13
Subject Sample	# 2	0.000	14:14
Breath Volume		2.566	14:14
Air Blank		0.000	14:15
Control Test		0.080	14:15
Air Blank		0.000	14:16
Diagnostics Check		OK	14:16

Data Download Date/Time: 12/12/2020 10:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TATE,COREY
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 11/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: CALDERON RAMOS

DATE/TIME OF TEST: 12/05/2020 20:21:00

Diagnostics Check		OK	20:14
Air Blank		0.000	20:14
Control Test		0.080	20:15
Air Blank		0.000	20:15
Subject Sample	# 1	0.080	20:16
Breath Volume		3.445	20:16
Air Blank		0.000	20:17
Air Blank		0.000	20:18
Subject Sample	# 2	0.079	20:19
Breath Volume		3.363	20:19
Air Blank		0.000	20:20
Control Test		0.080	20:20
Air Blank		0.000	20:20
Diagnostics Check		OK	20:21

Data Download Date/Time: 12/12/2020 10:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCGILLICUDDY,STEPHEN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 11/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: STEELE

DATE/TIME OF TEST: 12/05/2020 21:16:43

Diagnostics Check		OK	21:07
Air Blank		0.000	21:08
Control Test		0.081	21:08
Air Blank		0.000	21:08
Subject Sample	# 1	0.100	21:11
Breath Volume		1.394	21:11
Air Blank		0.000	21:12
Air Blank		0.000	21:13
Subject Sample	# 2	0.099	21:15
Breath Volume		1.339	21:15
Air Blank		0.000	21:15
Control Test		0.080	21:16
Air Blank		0.000	21:16
Diagnostics Check		OK	21:16

Data Download Date/Time: 12/12/2020 10:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER,CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: DAYAN

DATE/TIME OF TEST: 12/05/2020 21:44:52

Diagnostics Check		OK	21:38
Air Blank		0.000	21:38
Control Test		0.080	21:39
Air Blank		0.000	21:39
Subject Sample	# 1	0.070	21:40
Breath Volume		3.097	21:40
Air Blank		0.000	21:40
Air Blank		0.000	21:42
Subject Sample	# 2	0.068	21:43
Breath Volume		3.050	21:43
Air Blank		0.000	21:43
Control Test		0.079	21:44
Air Blank		0.000	21:44
Diagnostics Check		OK	21:44

Data Download Date/Time: 12/12/2020 10:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SOLOWAY,ALAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: ALTINE

DATE/TIME OF TEST: 12/06/2020 02:11:16

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.081	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.226	02:05
Breath Volume		1.386	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.222	02:09
Breath Volume		1.214	02:09
Air Blank		0.000	02:10
Control Test		0.078	02:10
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 12/12/2020 10:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PEREZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: North Palm Beach
LAST AGENCY INSPECTION: 11/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: DE LEON

DATE/TIME OF TEST: 12/06/2020 03:32:14

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.080	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.210	03:27
Breath Volume		2.328	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.211	03:30
Breath Volume		2.394	03:30
Air Blank		0.000	03:31
Control Test		0.082	03:31
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 12/12/2020 10:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 12/06/2020 04:38:56

Diagnostics Check		OK	04:32
Air Blank		0.000	04:32
Control Test		0.079	04:32
Air Blank		0.000	04:33
Subject Sample	# 1	0.212	04:34
Breath Volume		3.027	04:34
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.211	04:37
Breath Volume		2.976	04:37
Air Blank		0.000	04:38
Control Test		0.078	04:38
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 12/12/2020 10:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TRABAL,MILTON
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 11/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:09

SUBJECT'S LAST NAME: CHERISMA

DATE/TIME OF TEST: 12/09/2020 09:41:05

Diagnostics Check		OK	09:34
Air Blank		0.000	09:35
Control Test		0.081	09:35
Air Blank		0.000	09:35
Subject Sample	# 1	0.150	09:36
Breath Volume		1.835	09:36
Air Blank		0.000	09:37
Air Blank		0.000	09:38
Subject Sample	# 2	0.145	09:39
Breath Volume		1.894	09:39
Air Blank		0.000	09:40
Control Test		0.079	09:40
Air Blank		0.000	09:41
Diagnostics Check		OK	09:41

Data Download Date/Time: 12/12/2020 10:36

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: THOMAS,MICAH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 12/10/2020 17:45:04

Diagnostics Check		OK	17:38
Air Blank		0.000	17:39
Control Test		0.080	17:39
Air Blank		0.000	17:39
Subject Sample	# 1	0.079	17:40
Breath Volume		2.484	17:40
Air Blank		0.000	17:41
Air Blank		0.000	17:42
Subject Sample	# 2	0.084	17:43
Breath Volume		2.167	17:43
Air Blank		0.000	17:44
Control Test		0.080	17:44
Air Blank		0.000	17:44
Diagnostics Check		OK	17:45

Data Download Date/Time: 12/12/2020 10:37

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HOWARD,EDWARD
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 12/11/2020 00:06:09

Diagnostics Check		OK	23:59
Air Blank		0.000	23:59
Control Test		0.080	00:00
Air Blank		0.000	00:00
Subject Sample	# 1	0.176	00:01
Breath Volume		3.179	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:03
Subject Sample	# 2	0.179	00:04
Breath Volume		2.667	00:04
Air Blank		0.000	00:05
Control Test		0.079	00:05
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

Data Download Date/Time: 12/12/2020 10:37

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR,NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 11/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: VICK

DATE/TIME OF TEST: 12/11/2020 01:36:52

Diagnostics Check		OK	01:29
Air Blank		0.000	01:30
Control Test		0.080	01:30
Air Blank		0.000	01:30
Subject Sample	# 1	0.205	01:31
Breath Volume		2.355	01:31
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 2	0.213	01:35
Breath Volume		2.757	01:35
Air Blank		0.000	01:35
Control Test		0.079	01:36
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 12/12/2020 10:37

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: GILLETT

DATE/TIME OF TEST: 12/31/2020 21:26:42

Diagnostics Check		OK	21:22	
Air Blank		0.000	21:23	
Control Test		0.078	21:23	
Air Blank		0.000	21:23	
Subject Sample	# 1		21:25	Subject Test Refused
Breath Volume		0.000	21:25	
Air Blank		0.000	21:25	
Control Test		0.079	21:26	
Air Blank		0.000	21:26	
Diagnostics Check		OK	21:26	

Data Download Date/Time: 02/02/2021 12:43

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTAYA,GUILLERMO,
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: PRESSER,JAMES
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: BUI
ARREST TIME: 17:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: BERTELE

DATE/TIME OF TEST: 12/12/2020 18:41:15

Diagnostics Check		OK	18:34
Air Blank		0.000	18:34
Control Test		0.080	18:35
Air Blank		0.000	18:35
Subject Sample	# 1	0.134	18:36
Breath Volume		1.671	18:36
Air Blank		0.000	18:37
Air Blank		0.000	18:39
Subject Sample	# 2	0.148	18:39
Breath Volume		1.738	18:39
Air Blank		0.000	18:40
Control Test		0.078	18:40
Air Blank		0.000	18:41
Diagnostics Check		OK	18:41

Data Download Date/Time: 12/21/2020 12:17

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTAYA,GUILLERMO,
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: CARTAYA,GUILLERMO
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 11/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: BUI
ARREST TIME: 15:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/13/2020 16:14:57

Diagnostics Check		OK	16:11	
Air Blank		0.000	16:12	
Control Test		0.080	16:12	
Air Blank		0.000	16:12	
Subject Sample	# 1		16:13	Subject Test Refused
Breath Volume		0.000	16:13	
Air Blank		0.000	16:14	
Control Test		0.080	16:14	
Air Blank		0.000	16:14	
Diagnostics Check		OK	16:14	

Data Download Date/Time: 12/21/2020 12:17

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 11/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: STASHENKO

DATE/TIME OF TEST: 12/02/2020 00:29:07

Diagnostics Check		OK	00:22	
Air Blank		0.000	00:22	
Control Test		0.080	00:22	
Air Blank		0.000	00:23	
Subject Sample	# 1	0.291	00:24	
Breath Volume		2.949	00:24	
Air Blank		0.000	00:25	
Air Blank		0.000	00:27	
Subject Sample	# 2	0.292	00:27	
Breath Volume		4.097	00:27	
Air Blank			00:28	Purge Fail
Air Blank		0.000	00:29	

Data Download Date/Time: 12/04/2020 00:23

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BAMACA-DOMINGO

DATE/TIME OF TEST: 12/06/2020 00:48:47

Diagnostics Check		OK	00:40
Air Blank		0.000	00:40
Control Test		0.079	00:41
Air Blank		0.000	00:41
Subject Sample	# 1	0.182	00:44
Breath Volume		1.796	00:44
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.186	00:47
Breath Volume		2.902	00:47
Air Blank		0.000	00:47
Control Test		0.078	00:48
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 01/29/2021 09:13

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: HYPPOLITE

DATE/TIME OF TEST: 12/07/2020 00:01:09

Diagnostics Check		OK	23:47	
Air Blank		0.000	23:47	
Control Test		0.079	23:47	
Air Blank		0.000	23:48	
Subject Sample	# 1	0.151	23:51	Volume Not Met
Breath Volume		0.113	23:51	
Air Blank		0.000	23:52	
Air Blank		0.000	23:53	
Subject Sample	# 2	0.159	23:55	
Breath Volume		1.792	23:55	
Air Blank		0.000	23:56	
Air Blank		0.000	23:57	
Subject Sample	# 3	0.159	23:59	
Breath Volume		1.917	23:59	
Air Blank		0.000	00:00	
Control Test		0.079	00:00	
Air Blank		0.000	00:01	
Diagnostics Check		OK	00:01	

Data Download Date/Time: 01/29/2021 09:13

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: DELANCY

DATE/TIME OF TEST: 12/09/2020 19:53:05

Diagnostics Check		OK	19:45	
Air Blank		0.000	19:46	
Control Test		0.082	19:46	
Air Blank		0.000	19:46	
Subject Sample	# 1	0.114	19:47	
Breath Volume		2.367	19:47	
Air Blank		0.000	19:48	
Air Blank		0.000	19:50	
Subject Sample	# 2	0.113	19:50	
Breath Volume		2.132	19:50	
Air Blank		0.000	19:51	
Control Test		0.080	19:51	
Air Blank		0.000	19:52	
Diagnostics Check		Fail	19:52	Temperature Regulation Test
Air Blank		0.000	19:53	

Data Download Date/Time: 01/29/2021 09:13

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: VERONESI

DATE/TIME OF TEST: 12/10/2020 02:28:43

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:16	
Control Test		0.082	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 1		02:20	No Sample Provided
Breath Volume		0.000	02:20	
Air Blank		0.000	02:20	
Air Blank		0.000	02:22	
Subject Sample	# 2	0.252	02:23	
Breath Volume		3.757	02:23	
Air Blank		0.000	02:24	
Air Blank		0.000	02:25	
Subject Sample	# 3	0.259	02:26	
Breath Volume		3.402	02:26	
Air Blank		0.000	02:27	
Control Test		0.080	02:27	
Air Blank		0.000	02:28	
Diagnostics Check		Fail	02:28	Temperature Regulation Test
Air Blank		0.000	02:28	

Data Download Date/Time: 01/29/2021 09:13

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: COMIZA

DATE/TIME OF TEST: 12/13/2020 20:11:48

Diagnostics Check	OK	20:10	
Air Blank	0.000	20:10	
Control Test	0.080	20:10	
Air Blank		20:11	RFI Detect
Air Blank	0.000	20:11	

Data Download Date/Time: 01/29/2021 09:13

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: COMIZA

DATE/TIME OF TEST: 12/13/2020 20:21:28

Diagnostics Check		OK	20:14
Air Blank		0.000	20:14
Control Test		0.079	20:15
Air Blank		0.000	20:15
Subject Sample	# 1	0.208	20:16
Breath Volume		3.730	20:16
Air Blank		0.000	20:17
Air Blank		0.000	20:19
Subject Sample	# 2	0.211	20:19
Breath Volume		3.835	20:19
Air Blank		0.000	20:20
Control Test		0.080	20:20
Air Blank		0.000	20:21
Diagnostics Check		OK	20:21

Data Download Date/Time: 01/29/2021 09:13

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/14/2020 21:27:17

Diagnostics Check		OK	21:19
Air Blank		0.000	21:19
Control Test		0.079	21:19
Air Blank		0.000	21:20
Subject Sample	# 1	0.338	21:21
Breath Volume		1.550	21:21
Air Blank		0.000	21:22
Air Blank		0.000	21:24
Subject Sample	# 2	0.332	21:25
Breath Volume		1.378	21:25
Air Blank		0.000	21:26
Control Test		0.078	21:26
Air Blank		0.000	21:27
Diagnostics Check		OK	21:27

Data Download Date/Time: 01/29/2021 09:13

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/16/2020 02:58:44

Diagnostics Check		OK	02:45	
Air Blank		0.000	02:45	
Control Test		0.079	02:46	
Air Blank		0.000	02:46	
Subject Sample	# 1	0.325	02:48	
Breath Volume		3.000	02:48	
Air Blank		0.000	02:48	
Air Blank		0.000	02:50	
Subject Sample	# 2		02:53	No Sample Provided
Breath Volume		0.000	02:53	
Air Blank		0.000	02:54	
Air Blank		0.000	02:56	
Subject Sample	# 3	0.328	02:56	
Breath Volume		2.800	02:56	
Air Blank		0.000	02:57	
Control Test		0.079	02:58	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:58	

Data Download Date/Time: 01/29/2021 09:13

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 12/19/2020 20:48:22

Diagnostics Check		OK	20:41
Air Blank		0.000	20:41
Control Test		0.080	20:42
Air Blank		0.000	20:42
Subject Sample	# 1	0.080	20:43
Breath Volume		3.804	20:43
Air Blank		0.000	20:44
Air Blank		0.000	20:46
Subject Sample	# 2	0.076	20:46
Breath Volume		3.847	20:46
Air Blank		0.000	20:47
Control Test		0.079	20:47
Air Blank		0.000	20:48
Diagnostics Check		OK	20:48

Data Download Date/Time: 01/29/2021 09:13

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: GARCIA FABIAN

DATE/TIME OF TEST: 12/31/2020 23:41:36

Diagnostics Check		OK	23:34
Air Blank		0.000	23:34
Control Test		0.079	23:35
Air Blank		0.000	23:35
Subject Sample	# 1	0.153	23:36
Breath Volume		1.667	23:36
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.154	23:39
Breath Volume		2.582	23:39
Air Blank		0.000	23:40
Control Test		0.079	23:40
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 01/29/2021 09:13

Printed: 08 March 2021

COBRA ©1997-2014 CMI, Inc.