

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAWLINS,DUSTIN,P  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: JELKS,RANDY  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:01

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 02/01/2020 07:37:23

---

Diagnostics Check		OK	07:30
Air Blank		0.000	07:31
Control Test		0.080	07:31
Air Blank		0.000	07:32
Subject Sample	# 1	0.207	07:32
Breath Volume		2.757	07:32
Air Blank		0.000	07:33
Air Blank		0.000	07:35
Subject Sample	# 2	0.199	07:35
Breath Volume		2.742	07:35
Air Blank		0.000	07:36
Control Test		0.080	07:36
Air Blank		0.000	07:37
Diagnostics Check		OK	07:37

---

Data Download Date/Time: 02/05/2020 11:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD 18096,A,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: HOYOS 19022,J  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: CELESTE

DATE/TIME OF TEST: 02/06/2020 03:16:28

---

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.081	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.220	03:11
Breath Volume		1.484	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.217	03:14
Breath Volume		1.636	03:14
Air Blank		0.000	03:15
Control Test		0.078	03:15
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

---

Data Download Date/Time: 03/04/2020 13:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD 18096,A,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LOPEZ-BURNSIDE 13099,C  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: MARSHALL-SIMS

DATE/TIME OF TEST: 02/06/2020 22:40:54

---

Diagnostics Check	OK	22:32
Air Blank	0.000	22:32
Control Test	0.080	22:32
Air Blank	0.000	22:33
Subject Sample # 1	0.156	22:36
Breath Volume	1.433	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:38
Subject Sample # 2	0.150	22:39
Breath Volume	1.253	22:39
Air Blank	0.000	22:39
Control Test	0.078	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

---

Data Download Date/Time: 03/04/2020 13:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD 18096,A,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SIN 17096,P  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: QUINTANA ALVEZ

DATE/TIME OF TEST: 02/06/2020 22:54:05

---

Diagnostics Check	OK	22:47
Air Blank	0.000	22:48
Control Test	0.078	22:48
Air Blank	0.000	22:48
Subject Sample # 1	0.128	22:49
Breath Volume	3.070	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:51
Subject Sample # 2	0.136	22:52
Breath Volume	4.554	22:52
Air Blank	0.000	22:53
Control Test	0.079	22:53
Air Blank	0.000	22:53
Diagnostics Check	OK	22:54

---

Data Download Date/Time: 03/04/2020 13:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD 18096,A,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: CAFFRONI 616,I  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: COSTIGAN

DATE/TIME OF TEST: 02/07/2020 01:48:34

---

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.079	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.000	01:43
Breath Volume	3.328	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.000	01:47
Breath Volume	3.660	01:47
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

---

Data Download Date/Time: 03/04/2020 13:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD 18096,A,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SOMMERS 13120,E  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: MARTINEZ BARRERA

DATE/TIME OF TEST: 02/07/2020 03:09:53

---

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.080	03:03
Air Blank		0.000	03:04
Subject Sample	# 1	0.216	03:05
Breath Volume		2.074	03:05
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.203	03:08
Breath Volume		2.335	03:08
Air Blank		0.000	03:08
Control Test		0.079	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

---

Data Download Date/Time: 03/04/2020 13:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SERRANO,ROMAN,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: PAHLKE,M  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MAZZA

DATE/TIME OF TEST: 02/07/2020 04:36:27

---

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.080	04:30
Air Blank	0.000	04:30
Subject Sample # 1	0.000	04:31
Breath Volume	2.074	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.000	04:35
Breath Volume	2.574	04:35
Air Blank	0.000	04:35
Control Test	0.079	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

---

Data Download Date/Time: 03/04/2020 13:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROMAN,A,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: TOMPKINS,K  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: GOSSET

DATE/TIME OF TEST: 02/07/2020 05:28:06

---

Diagnostics Check	OK	05:21
Air Blank	0.000	05:21
Control Test	0.080	05:22
Air Blank	0.000	05:22
Subject Sample # 1	0.142	05:23
Breath Volume	2.246	05:23
Air Blank	0.000	05:24
Air Blank	0.000	05:25
Subject Sample # 2	0.142	05:26
Breath Volume	2.117	05:26
Air Blank	0.000	05:27
Control Test	0.079	05:27
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

---

Data Download Date/Time: 03/04/2020 13:09



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, JOSIAH, R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SMITH, JOSIAH  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: FULBRIGHT

DATE/TIME OF TEST: 02/08/2020 04:10:42

---

Diagnostics Check		OK	04:04
Air Blank		0.000	04:05
Control Test		0.080	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.237	04:06
Breath Volume		1.640	04:06
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.237	04:09
Breath Volume		1.921	04:09
Air Blank		0.000	04:09
Control Test		0.079	04:10
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

---

Data Download Date/Time: 03/04/2020 13:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROMAN,A,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: CAILVANO,S  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: RODRIGUEZ MARTINEZ

DATE/TIME OF TEST: 02/09/2020 05:48:36

---

Diagnostics Check	OK	05:42
Air Blank	0.000	05:42
Control Test	0.081	05:42
Air Blank	0.000	05:43
Subject Sample # 1	0.204	05:44
Breath Volume	1.750	05:44
Air Blank	0.000	05:44
Air Blank	0.000	05:46
Subject Sample # 2	0.207	05:46
Breath Volume	2.054	05:46
Air Blank	0.000	05:47
Control Test	0.080	05:47
Air Blank	0.000	05:48
Diagnostics Check	OK	05:48

---

Data Download Date/Time: 03/04/2020 13:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINS,KYLE,L  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: BROWN,TERRELL  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 02/10/2020 03:44:10

---

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:39
Subject Sample # 1	0.180	03:39
Breath Volume	3.820	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.175	03:42
Breath Volume	3.480	03:42
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

---

Data Download Date/Time: 03/04/2020 13:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVENDANO,JUAN,M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: PANTOJA,ABRAHAM  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: MATIAS

DATE/TIME OF TEST: 02/15/2020 19:47:32

---

Diagnostics Check	OK	19:40	
Air Blank	0.000	19:41	
Control Test	0.079	19:41	
Air Blank	0.000	19:41	
Subject Sample # 1	0.241	19:42	
Breath Volume	2.441	19:42	
Air Blank	0.000	19:43	
Air Blank	0.000	19:45	
Subject Sample # 2		19:45	Slope Not Met
Breath Volume	2.394	19:45	
Air Blank	0.000	19:46	
Control Test	0.079	19:46	
Air Blank	0.000	19:47	
Diagnostics Check	OK	19:47	

---

Data Download Date/Time: 03/04/2020 13:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVENDANO,JUAN,M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: PANTOJA,ABRAHAM  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: MATIAS

DATE/TIME OF TEST: 02/15/2020 19:58:14

---

Diagnostics Check	OK	19:51
Air Blank	0.000	19:52
Control Test	0.079	19:52
Air Blank	0.000	19:53
Subject Sample # 1	0.234	19:53
Breath Volume	3.062	19:53
Air Blank	0.000	19:54
Air Blank	0.000	19:56
Subject Sample # 2	0.241	19:56
Breath Volume	2.968	19:56
Air Blank	0.000	19:57
Control Test	0.079	19:57
Air Blank	0.000	19:58
Diagnostics Check	OK	19:58

---

Data Download Date/Time: 03/04/2020 13:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVENDANO,JUAN,M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SHOEMAKER,JACLYN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: GORMAN

DATE/TIME OF TEST: 02/16/2020 03:43:41

---

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.172	03:38
Breath Volume	1.894	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.172	03:42
Breath Volume	1.761	03:42
Air Blank	0.000	03:42
Control Test	0.079	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

---

Data Download Date/Time: 03/04/2020 13:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TISDALE,KATIE,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: CONLIN,EVAN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: BARTREM

DATE/TIME OF TEST: 02/23/2020 00:01:47

---

Diagnostics Check	OK	23:54
Air Blank	0.000	23:55
Control Test	0.080	23:55
Air Blank	0.000	23:56
Subject Sample # 2	0.089	23:56
Breath Volume	1.613	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 1	0.086	00:00
Breath Volume	1.406	00:00
Air Blank	0.000	00:00
Control Test	0.080	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

---

Data Download Date/Time: 03/04/2020 13:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TISDALE,KATIE,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: VALCIN,JEFFREY  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: PIGOTT

DATE/TIME OF TEST: 02/23/2020 00:22:02

---

Diagnostics Check		OK	00:15
Air Blank		0.000	00:15
Control Test		0.080	00:15
Air Blank		0.000	00:16
Subject Sample	# 1	0.216	00:17
Breath Volume		1.671	00:17
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 2	0.208	00:20
Breath Volume		1.921	00:20
Air Blank		0.000	00:21
Control Test		0.079	00:21
Air Blank		0.000	00:21
Diagnostics Check		OK	00:22

---

Data Download Date/Time: 03/04/2020 13:10



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEINER,RYAN,T  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: BEINER,RYAN  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: ALVAREZ LAITANO

DATE/TIME OF TEST: 02/27/2020 02:30:58

---

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.080	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.135	02:26
Breath Volume	3.289	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:28
Subject Sample # 2	0.129	02:29
Breath Volume	2.644	02:29
Air Blank	0.000	02:30
Control Test	0.078	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

---

Data Download Date/Time: 03/04/2020 13:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANDA,CHELSY,  
BREATH TEST OPERATOR AGENCY: SOUTH MIAMI PD  
ARRESTING OFFICER: MORUA,E  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: AG827002  
DRY GAS EXPIRY DATE: 09/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 02/02/2020 01:26:25

---

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:22	
Control Test	0.079	01:23	
Air Blank	0.000	01:23	
Subject Sample	# 1	01:25	Subject Test Refused
Breath Volume	0.000	01:25	
Air Blank	0.000	01:25	
Control Test	0.079	01:25	
Air Blank	0.000	01:26	
Diagnostics Check	OK	01:26	

---

Data Download Date/Time: 02/26/2020 13:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: IGLESIAS,YIMY,  
BREATH TEST OPERATOR AGENCY: FL STATE ATTORNEY  
ARRESTING OFFICER: RAWLUK,SHANE  
ARRESTING OFFICER AGENCY: FL STATE ATTORNEY  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG827002  
DRY GAS EXPIRY DATE: 09/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 05:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 02/02/2020 06:31:32

---

Diagnostics Check	OK	06:28	
Air Blank	0.000	06:29	
Control Test	0.078	06:29	
Air Blank	0.000	06:29	
Subject Sample	# 1	06:30	Subject Test Refused
Breath Volume	0.000	06:30	
Air Blank	0.000	06:30	
Control Test	0.078	06:30	
Air Blank	0.000	06:31	
Diagnostics Check	OK	06:31	

---

Data Download Date/Time: 02/26/2020 13:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNOZ,ROBERT,A  
BREATH TEST OPERATOR AGENCY: PINECREST PD  
ARRESTING OFFICER: SPENCE,GARY  
ARRESTING OFFICER AGENCY: PINECREST PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG827002  
DRY GAS EXPIRY DATE: 09/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 05:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:58

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 02/04/2020 07:30:12

---

Diagnostics Check		OK	07:18	
Air Blank		0.000	07:19	
Control Test		0.079	07:19	
Air Blank		0.000	07:20	
Subject Sample	# 1	0.119	07:23	Volume Not Met
Breath Volume		0.128	07:23	
Air Blank		0.000	07:24	
Air Blank		0.000	07:25	
Subject Sample	# 2	0.125	07:28	Volume Not Met
Breath Volume		0.312	07:28	
Air Blank			07:29	Purge Fail
Air Blank		0.000	07:30	

---

Data Download Date/Time: 02/26/2020 13:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNOZ,ROBERT,A  
BREATH TEST OPERATOR AGENCY: PINECREST PD  
ARRESTING OFFICER: SPENCE,GARY  
ARRESTING OFFICER AGENCY: PINECREST PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG827002  
DRY GAS EXPIRY DATE: 09/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 05:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:58

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 02/04/2020 07:43:57

---

Diagnostics Check	OK	07:42	
Air Blank	0.000	07:42	
Control Test	0.065	07:43	Control Outside Tolerance
Air Blank	0.000	07:43	

---

Data Download Date/Time: 02/26/2020 13:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNOZ,ROBERT,A  
BREATH TEST OPERATOR AGENCY: PINECREST PD  
ARRESTING OFFICER: SPENCE,GARY  
ARRESTING OFFICER AGENCY: PINECREST PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG827002  
DRY GAS EXPIRY DATE: 09/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 05:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:58

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 02/04/2020 07:47:14

---

Diagnostics Check	OK	07:45	
Air Blank	0.000	07:46	
Control Test	0.063	07:46	Control Outside Tolerance
Air Blank	0.000	07:47	

---

Data Download Date/Time: 02/26/2020 13:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNOZ,ROBERT,A  
BREATH TEST OPERATOR AGENCY: PINECREST PD  
ARRESTING OFFICER: SPENCE,GARY  
ARRESTING OFFICER AGENCY: PINECREST PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG827002  
DRY GAS EXPIRY DATE: 09/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 05:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:58

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 02/04/2020 08:01:26

---

Diagnostics Check	OK	08:00	
Air Blank	0.000	08:00	
Control Test	0.067	08:00	Control Outside Tolerance
Air Blank	0.000	08:01	

---

Data Download Date/Time: 02/26/2020 13:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,DARIEN,  
BREATH TEST OPERATOR AGENCY: SOUTH MIAMI PD  
ARRESTING OFFICER: RODRIGUEZ,DARIEN  
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG827002  
DRY GAS EXPIRY DATE: 09/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: OCASIO MORALES

DATE/TIME OF TEST: 02/06/2020 04:30:43

---

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:19	
Control Test	0.079	04:20	
Air Blank	0.000	04:20	
Subject Sample	# 1	04:23	No Sample Provided
Breath Volume	0.000	04:23	
Air Blank	0.000	04:24	
Air Blank	0.000	04:26	
Subject Sample	# 2	04:29	No Sample Provided
Breath Volume	0.000	04:29	
Air Blank	0.000	04:29	
Control Test	0.079	04:30	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

---

Data Download Date/Time: 02/26/2020 13:53



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,DARIEN,  
BREATH TEST OPERATOR AGENCY: SOUTH MIAMI PD  
ARRESTING OFFICER: RODRIGUEZ,DARIEN  
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG827002  
DRY GAS EXPIRY DATE: 09/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: OCASIO MORALES

DATE/TIME OF TEST: 02/06/2020 04:47:18

---

Diagnostics Check	OK	04:44	
Air Blank	0.000	04:44	
Control Test	0.079	04:45	
Air Blank	0.000	04:45	
Subject Sample	# 1	04:45	Subject Test Refused
Breath Volume	0.000	04:45	
Air Blank	0.000	04:46	
Control Test	0.077	04:46	
Air Blank	0.000	04:47	
Diagnostics Check	OK	04:47	

---

Data Download Date/Time: 02/26/2020 13:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,DARIEN,  
BREATH TEST OPERATOR AGENCY: SOUTH MIAMI PD  
ARRESTING OFFICER: RODRIGUEZ,DARIEN  
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG827002  
DRY GAS EXPIRY DATE: 09/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: BASHIRIMOGHADDAM

DATE/TIME OF TEST: 02/13/2020 03:23:28

---

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:21	
Control Test	0.079	03:21	
Air Blank	0.000	03:21	
Subject Sample	# 1	03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:22	
Control Test	0.079	03:22	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

---

Data Download Date/Time: 02/26/2020 13:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,DARIEN,  
BREATH TEST OPERATOR AGENCY: SOUTH MIAMI PD  
ARRESTING OFFICER: RODRIGUEZ,DARIEN  
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG827002  
DRY GAS EXPIRY DATE: 09/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 02/16/2020 07:20:57

---

Diagnostics Check		OK	07:10
Air Blank		0.000	07:11
Control Test		0.079	07:11
Air Blank		0.000	07:12
Subject Sample	# 1	0.202	07:12
Breath Volume		1.507	07:12
Air Blank		0.000	07:13
Air Blank		0.000	07:15
Subject Sample	# 2	0.167	07:16
Breath Volume		1.429	07:16
Air Blank		0.000	07:17
Air Blank		0.000	07:18
Subject Sample	# 3	0.173	07:19
Breath Volume		1.429	07:19
Air Blank		0.000	07:19
Control Test		0.076	07:20
Air Blank		0.000	07:20
Diagnostics Check		OK	07:20

---

Data Download Date/Time: 02/26/2020 13:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANDA,CHELSY,  
BREATH TEST OPERATOR AGENCY: SOUTH MIAMI PD  
ARRESTING OFFICER: LANDA,CHELSY  
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG000606  
DRY GAS EXPIRY DATE: 01/06/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: SEIBERT

DATE/TIME OF TEST: 02/23/2020 05:32:42

---

Diagnostics Check	OK	05:21	
Air Blank	0.000	05:22	
Control Test	0.078	05:22	
Air Blank	0.000	05:23	
Subject Sample	# 1	05:26	No Sample Provided
Breath Volume	0.000	05:26	
Air Blank	0.000	05:26	
Air Blank	0.000	05:28	
Subject Sample	# 2	05:31	Subject Test Refused
Breath Volume	0.000	05:31	
Air Blank	0.000	05:31	
Control Test	0.071	05:32	Control Outside Tolerance
Air Blank	0.000	05:32	

---

Data Download Date/Time: 02/26/2020 13:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

INSTRUMENT REGISTERED TO: SOUTH MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANDA,CHELSY,  
BREATH TEST OPERATOR AGENCY: SOUTH MIAMI PD  
ARRESTING OFFICER: LANDA,CHELSY  
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG000606  
DRY GAS EXPIRY DATE: 01/06/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: SEIBERT

DATE/TIME OF TEST: 02/23/2020 05:57:18

---

Diagnostics Check	OK	05:51	
Air Blank	0.000	05:52	
Control Test	0.078	05:52	
Air Blank	0.000	05:52	
Subject Sample	# 1 0.121	05:56	Volume Not Met
Breath Volume	0.226	05:56	
Air Blank		05:56	Purge Fail
Air Blank	0.000	05:57	

---

Data Download Date/Time: 02/26/2020 13:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZAMORA,DEP,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: ALLEN,DEP  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 02/02/2020 23:14:27

---

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.080	23:08
Air Blank	0.000	23:09
Subject Sample # 1	0.149	23:10
Breath Volume	2.113	23:10
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.147	23:12
Breath Volume	1.664	23:12
Air Blank	0.000	23:13
Control Test	0.080	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

---

Data Download Date/Time: 02/25/2020 00:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZAMORA,DEP,  
BREATH TEST OPERATOR AGENCY: Punta Gorda PD  
ARRESTING OFFICER: MEDDAUGH,OFF  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MCGINNIS

DATE/TIME OF TEST: 02/03/2020 01:25:48

---

Diagnostics Check	OK	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:20
Subject Sample # 1	0.164	01:21
Breath Volume	1.863	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:23
Subject Sample # 2	0.160	01:24
Breath Volume	1.218	01:24
Air Blank	0.000	01:24
Control Test	0.081	01:25
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

---

Data Download Date/Time: 02/25/2020 00:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: HERNANDEZ VELASQUEZ

DATE/TIME OF TEST: 02/16/2020 02:02:27

---

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.135	01:57
Breath Volume	2.460	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.132	02:00
Breath Volume	1.894	02:00
Air Blank	0.000	02:01
Control Test	0.077	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

---

Data Download Date/Time: 03/25/2020 09:58



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: ENLOW

DATE/TIME OF TEST: 02/29/2020 00:10:21

---

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.080	00:04
Air Blank	0.000	00:04
Subject Sample # 1	0.178	00:05
Breath Volume	1.492	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:07
Subject Sample # 2	0.198	00:08
Breath Volume	1.968	00:08
Air Blank	0.000	00:09
Control Test	0.079	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

---

Data Download Date/Time: 03/25/2020 09:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOSTON,D,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DUKIC,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: ALDACO

DATE/TIME OF TEST: 02/01/2020 09:16:50

---

Diagnostics Check		OK	09:10
Air Blank		0.000	09:10
Control Test		0.079	09:11
Air Blank		0.000	09:11
Subject Sample	# 1	0.177	09:12
Breath Volume		2.207	09:12
Air Blank		0.000	09:12
Air Blank		0.000	09:14
Subject Sample	# 2	0.166	09:15
Breath Volume		1.582	09:15
Air Blank		0.000	09:15
Control Test		0.080	09:16
Air Blank		0.000	09:16
Diagnostics Check		OK	09:16

---

Data Download Date/Time: 02/12/2020 14:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT,J,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: TURPIN,D  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: KUBIK

DATE/TIME OF TEST: 02/01/2020 16:36:34

---

Diagnostics Check	OK	16:33	
Air Blank	0.000	16:33	
Control Test	0.080	16:34	
Air Blank	0.000	16:34	
Subject Sample	# 1	16:35	Subject Test Refused
Breath Volume	0.000	16:35	
Air Blank	0.000	16:35	
Control Test	0.080	16:35	
Air Blank	0.000	16:36	
Diagnostics Check	OK	16:36	

---

Data Download Date/Time: 02/12/2020 14:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DOUGHFMAN, D  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 02/01/2020 21:10:33

---

Diagnostics Check	OK	21:07	
Air Blank	0.000	21:08	
Control Test	0.079	21:08	
Air Blank	0.000	21:08	
Subject Sample	# 1	21:09	Subject Test Refused
Breath Volume	0.000	21:09	
Air Blank	0.000	21:09	
Control Test	0.079	21:09	
Air Blank	0.000	21:10	
Diagnostics Check	OK	21:10	

---

Data Download Date/Time: 02/12/2020 14:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KEEN, C  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: KIEBACH

DATE/TIME OF TEST: 02/02/2020 01:22:20

---

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:19	
Control Test	0.079	01:20	
Air Blank	0.000	01:20	
Subject Sample	# 1	01:20	Subject Test Refused
Breath Volume	0.000	01:20	
Air Blank	0.000	01:21	
Control Test	0.080	01:21	
Air Blank	0.000	01:22	
Diagnostics Check	OK	01:22	

---

Data Download Date/Time: 02/12/2020 14:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUTHE,SAMUEL,  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: CHOWDHURY,P  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:01

SUBJECT'S LAST NAME: YOUTHE

DATE/TIME OF TEST: 02/02/2020 08:30:19

---

Diagnostics Check	OK	08:24
Air Blank	0.000	08:24
Control Test	0.079	08:24
Air Blank	0.000	08:25
Subject Sample # 1	0.102	08:25
Breath Volume	2.847	08:25
Air Blank	0.000	08:26
Air Blank	0.000	08:28
Subject Sample # 2	0.100	08:28
Breath Volume	2.675	08:28
Air Blank	0.000	08:29
Control Test	0.080	08:29
Air Blank	0.000	08:30
Diagnostics Check	OK	08:30

---

Data Download Date/Time: 02/12/2020 14:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT,J,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: MALBASIC

DATE/TIME OF TEST: 02/02/2020 23:33:04

---

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.301	23:28
Breath Volume	1.644	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.290	23:31
Breath Volume	2.230	23:31
Air Blank	0.000	23:32
Control Test	0.079	23:32
Air Blank	0.000	23:32
Diagnostics Check	OK	23:33

---

Data Download Date/Time: 02/12/2020 14:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DEW, P  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 02/03/2020 00:43:09

---

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:40	
Control Test	0.080	00:40	
Air Blank	0.000	00:41	
Subject Sample	# 1	00:41	Subject Test Refused
Breath Volume	0.000	00:41	
Air Blank	0.000	00:42	
Control Test	0.080	00:42	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

---

Data Download Date/Time: 02/12/2020 14:33



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT,J,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LUGO,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: BOSTICK

DATE/TIME OF TEST: 02/03/2020 01:35:58

---

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:33	
Control Test	0.080	01:33	
Air Blank	0.000	01:34	
Subject Sample	# 1	01:34	Subject Test Refused
Breath Volume	0.000	01:34	
Air Blank	0.000	01:35	
Control Test	0.079	01:35	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

---

Data Download Date/Time: 02/12/2020 14:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE, J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: SHIELDS

DATE/TIME OF TEST: 02/03/2020 02:40:23

---

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.079	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.120	02:35
Breath Volume	2.019	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.117	02:38
Breath Volume	1.976	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

---

Data Download Date/Time: 02/12/2020 14:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: ULYSSE, S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 02/04/2020 02:02:55

---

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.098	01:58
Breath Volume	2.535	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.098	02:01
Breath Volume	3.039	02:01
Air Blank	0.000	02:01
Control Test	0.079	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

---

Data Download Date/Time: 02/12/2020 14:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUTTON, BRADEN, J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DOUGHFMAN, D  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: SINGLETARY

DATE/TIME OF TEST: 02/05/2020 03:15:56

---

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.079	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.000	03:11
Breath Volume	1.902	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.000	03:14
Breath Volume	2.863	03:14
Air Blank	0.000	03:15
Control Test	0.079	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

---

Data Download Date/Time: 02/12/2020 14:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MOELLER, R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: MCCAMMACK

DATE/TIME OF TEST: 02/05/2020 04:05:16

---

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.000	04:00
Breath Volume	2.394	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.000	04:03
Breath Volume	3.007	04:03
Air Blank	0.000	04:04
Control Test	0.080	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

---

Data Download Date/Time: 02/12/2020 14:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FARLEY, J  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 02/06/2020 00:17:05

---

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.079	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.179	00:11
Breath Volume	2.511	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.186	00:15
Breath Volume	2.550	00:15
Air Blank	0.000	00:16
Control Test	0.079	00:16
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

---

Data Download Date/Time: 02/12/2020 14:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: RODRIGUEZ, M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: ANGUS

DATE/TIME OF TEST: 02/06/2020 04:04:07

---

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.021	03:59
Breath Volume	2.593	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.018	04:02
Breath Volume	2.468	04:02
Air Blank	0.000	04:03
Control Test	0.079	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

---

Data Download Date/Time: 02/12/2020 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUTTON, BRADEN, J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DOUGHFMAN, D  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 02/11/2020 20:42:11

---

Diagnostics Check	OK	20:39	
Air Blank	0.000	20:39	
Control Test	0.079	20:40	
Air Blank	0.000	20:40	
Subject Sample	# 1	20:40	Subject Test Refused
Breath Volume	0.000	20:40	
Air Blank	0.000	20:41	
Control Test	0.078	20:41	
Air Blank	0.000	20:42	
Diagnostics Check	OK	20:42	

---

Data Download Date/Time: 02/12/2020 14:34



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUTTON, BRADEN, J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LUGO, R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 02/12/2020 01:43:28

---

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:41	
Control Test	0.079	01:41	
Air Blank	0.000	01:41	
Subject Sample	# 1	01:42	Subject Test Refused
Breath Volume	0.000	01:42	
Air Blank	0.000	01:42	
Control Test	0.080	01:42	
Air Blank	0.000	01:43	
Diagnostics Check	OK	01:43	

---

Data Download Date/Time: 02/12/2020 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,ALEXIS,  
BREATH TEST OPERATOR AGENCY: Hialeah PD  
ARRESTING OFFICER: PELAEZ,MICHAEL  
ARRESTING OFFICER AGENCY: Hialeah PD  
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1224982  
DRY GAS EXPIRY DATE: 10/31/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ANEIROS

DATE/TIME OF TEST: 02/05/2020 23:36:28

---

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.080	23:30
Air Blank	0.000	23:31
Subject Sample # 1	0.221	23:31
Breath Volume	2.632	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.225	23:34
Breath Volume	3.152	23:34
Air Blank	0.000	23:35
Control Test	0.079	23:35
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

---

Data Download Date/Time: 02/26/2020 11:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ARROJAS,GABRIEL,L

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: HERNANDEZ,OSDANY

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1224982

DRY GAS EXPIRY DATE: 10/31/2022

VIOLATION CODE: DUI

ARREST TIME: 23:30

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 02/14/2020 00:52:59

---

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:50	
Control Test	0.081	00:50	
Air Blank	0.000	00:51	
Subject Sample	# 1	00:51	Subject Test Refused
Breath Volume	0.000	00:51	
Air Blank	0.000	00:52	
Control Test	0.081	00:52	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

---

Data Download Date/Time: 02/26/2020 11:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANO,ADRIEL,  
BREATH TEST OPERATOR AGENCY: Hialeah PD  
ARRESTING OFFICER: CAMARGO,A  
ARRESTING OFFICER AGENCY: Hialeah PD  
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1224982  
DRY GAS EXPIRY DATE: 10/31/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 02/23/2020 00:20:48

---

Diagnostics Check	OK	00:15	
Air Blank	0.000	00:15	
Control Test	0.081	00:16	
Air Blank	0.000	00:16	
Subject Sample	# 1	00:19	Subject Test Refused
Breath Volume	0.000	00:19	
Air Blank	0.000	00:19	
Control Test	0.077	00:20	
Air Blank	0.000	00:20	
Diagnostics Check	OK	00:20	

---

Data Download Date/Time: 02/26/2020 11:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARROJAS,GABRIEL,L  
BREATH TEST OPERATOR AGENCY: HIALEAH PD  
ARRESTING OFFICER: ALVEREZ,REINALDO  
ARRESTING OFFICER AGENCY: HIALEAH PD  
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1224982  
DRY GAS EXPIRY DATE: 10/31/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/09/2020 02:43:01

---

Diagnostics Check		OK	02:30	
Air Blank		0.000	02:30	
Control Test		0.081	02:30	
Air Blank		0.000	02:31	
Subject Sample	# 1	0.124	02:34	Volume Not Met
Breath Volume		0.429	02:34	
Air Blank		0.000	02:35	
Air Blank		0.000	02:36	
Subject Sample	# 2	0.142	02:37	
Breath Volume		1.281	02:37	
Air Blank		0.000	02:38	
Air Blank		0.000	02:40	
Subject Sample	# 3	0.132	02:41	
Breath Volume		1.996	02:41	
Air Blank		0.000	02:42	
Control Test		0.079	02:42	
Air Blank		0.000	02:42	
Diagnostics Check		OK	02:43	

---

Data Download Date/Time: 02/26/2020 11:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ARROJAS,GABRIEL,L

BREATH TEST OPERATOR AGENCY: HIALEAH PD

ARRESTING OFFICER: ARROJAS,GABRIEL

ARRESTING OFFICER AGENCY: HIALEAH PD

LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1224982

DRY GAS EXPIRY DATE: 10/31/2022

VIOLATION CODE: DUI

ARREST TIME: 23:10

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: VILLEGAS

DATE/TIME OF TEST: 02/10/2020 00:10:43

---

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:00	
Control Test	0.081	00:00	
Air Blank	0.000	00:01	
Subject Sample	# 1	00:04	No Sample Provided
Breath Volume	0.000	00:04	
Air Blank	0.000	00:04	
Air Blank	0.000	00:06	
Subject Sample	# 2	00:09	Subject Test Refused
Breath Volume	0.000	00:09	
Air Blank	0.000	00:09	
Control Test	0.081	00:10	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

---

Data Download Date/Time: 02/26/2020 11:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002566

INSTRUMENT REGISTERED TO: SURFSIDE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNSCHY,PATRICK,R  
BREATH TEST OPERATOR AGENCY: Bay Harbor Island PD  
ARRESTING OFFICER: BLANCO,ERICA  
ARRESTING OFFICER AGENCY: Bal Harbour Vil PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 980558  
DRY GAS EXPIRY DATE: 07/09/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 02/09/2020 04:34:06

---

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:31	
Control Test	0.077	04:31	
Air Blank	0.000	04:31	
Subject Sample	# 1	04:32	Subject Test Refused
Breath Volume	0.000	04:32	
Air Blank	0.000	04:33	
Control Test	0.077	04:33	
Air Blank	0.000	04:34	
Diagnostics Check	OK	04:34	

---

Data Download Date/Time: 02/27/2020 13:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002566

INSTRUMENT REGISTERED TO: SURFSIDE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNSCHY,PATRICK,R  
BREATH TEST OPERATOR AGENCY: Bay Harbor Island PD  
ARRESTING OFFICER: BLANCO,ERICA  
ARRESTING OFFICER AGENCY: Bal Harbour Vil PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 980558  
DRY GAS EXPIRY DATE: 07/09/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: RICCA

DATE/TIME OF TEST: 02/18/2020 01:33:40

---

Diagnostics Check	OK	01:26	
Air Blank	0.000	01:26	
Control Test	0.078	01:27	
Air Blank	0.000	01:27	
Subject Sample	# 1 0.256	01:29	
Breath Volume	1.269	01:29	
Air Blank	0.000	01:29	
Air Blank	0.000	01:31	
Subject Sample	# 2 0.246	01:32	
Breath Volume	3.105	01:32	
Air Blank	0.000	01:32	
Control Test	0.074	01:33	Control Outside Tolerance
Air Blank	0.000	01:33	

---

Data Download Date/Time: 02/27/2020 13:02



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTANO,JAIVER,D  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: CASTANO,JAIVER  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: VISAGE

DATE/TIME OF TEST: 02/22/2020 05:39:06

---

Diagnostics Check	OK	05:23	
Air Blank	0.000	05:24	
Control Test	0.079	05:24	
Air Blank	0.000	05:25	
Subject Sample	# 1 0.178	05:26	
Breath Volume	2.210	05:26	
Air Blank	0.000	05:26	
Air Blank	0.000	05:28	
Subject Sample	# 2 0.147	05:31	Volume Not Met
Breath Volume	0.550	05:31	
Air Blank	0.000	05:32	
Air Blank	0.000	05:34	
Subject Sample	# 3 0.183	05:37	
Breath Volume	3.820	05:37	
Air Blank	0.000	05:38	
Control Test	0.077	05:38	
Air Blank	0.000	05:39	
Diagnostics Check	OK	05:39	

---

Data Download Date/Time: 02/27/2020 10:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HODGES,TITUS,R  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: HODGES,TITUS  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 02/23/2020 04:46:47

---

Diagnostics Check	OK	04:38
Air Blank	0.000	04:39
Control Test	0.081	04:39
Air Blank	0.000	04:39
Subject Sample # 1	0.290	04:41
Breath Volume	1.277	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.307	04:45
Breath Volume	1.945	04:45
Air Blank	0.000	04:45
Control Test	0.080	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

---

Data Download Date/Time: 02/27/2020 10:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAMAYO,KEVIN,X  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: ST. LOUIS,IFRATHIEU  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 05/11/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 02/25/2020 00:29:05

---

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.082	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:27	Subject Test Refused
Breath Volume	0.000	00:27	
Air Blank	0.000	00:28	
Control Test	0.080	00:28	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:29	

---

Data Download Date/Time: 02/27/2020 10:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROY,GARY,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: ROY,GARY  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: POUPARD

DATE/TIME OF TEST: 02/08/2020 21:47:58

---

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:46	
Control Test	0.078	21:46	
Air Blank		21:47	Ambient Fail
Air Blank	0.000	21:47	

---

Data Download Date/Time: 03/30/2020 14:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROY,GARY,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: ROY,GARY  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: POUPARD

DATE/TIME OF TEST: 02/08/2020 21:50:42

---

Diagnostics Check	OK	21:49	
Air Blank		21:50	Ambient Fail
Air Blank	0.000	21:50	

---

Data Download Date/Time: 03/30/2020 14:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROY,GARY,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: ROY,GARY  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MONEY

DATE/TIME OF TEST: 02/13/2020 00:01:40

---

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:00	
Control Test	0.082	00:00	Interferent Detect
Air Blank	0.000	00:01	

---

Data Download Date/Time: 03/30/2020 14:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROY,GARY,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: ROY,GARY  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MONEY

DATE/TIME OF TEST: 02/13/2020 00:04:29

---

Diagnostics Check	OK	00:03	
Air Blank		00:04	Ambient Fail
Air Blank	0.000	00:04	

---

Data Download Date/Time: 03/30/2020 14:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VISSER,A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: GOHIL

DATE/TIME OF TEST: 02/01/2020 04:02:49

---

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.081	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.094	03:58
Breath Volume	2.585	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.096	04:01
Breath Volume	2.371	04:01
Air Blank	0.000	04:01
Control Test	0.080	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

---

Data Download Date/Time: 02/05/2020 18:51



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: THORNE,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: RICH

DATE/TIME OF TEST: 02/01/2020 20:20:07

---

Diagnostics Check	OK	20:13
Air Blank	0.000	20:14
Control Test	0.081	20:14
Air Blank	0.000	20:14
Subject Sample # 1	0.151	20:15
Breath Volume	2.726	20:15
Air Blank	0.000	20:16
Air Blank	0.000	20:18
Subject Sample # 2	0.147	20:18
Breath Volume	1.757	20:18
Air Blank	0.000	20:19
Control Test	0.081	20:19
Air Blank	0.000	20:20
Diagnostics Check	OK	20:20

---

Data Download Date/Time: 02/05/2020 18:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VISSER,A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: MARDY

DATE/TIME OF TEST: 02/02/2020 03:18:40

---

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.081	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.080	03:14
Breath Volume	3.152	03:14
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.078	03:17
Breath Volume	3.132	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

---

Data Download Date/Time: 02/05/2020 18:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LYON,M  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: REYES SARABIA

DATE/TIME OF TEST: 02/02/2020 03:37:10

---

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.081	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.056	03:32
Breath Volume		2.855	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.055	03:35
Breath Volume		2.238	03:35
Air Blank		0.000	03:36
Control Test		0.081	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

---

Data Download Date/Time: 02/05/2020 18:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SIMS,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: BUIE

DATE/TIME OF TEST: 02/03/2020 02:11:59

---

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.081	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.222	02:06
Breath Volume	1.980	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.217	02:10
Breath Volume	1.679	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

---

Data Download Date/Time: 02/05/2020 18:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: BLACKMORE

DATE/TIME OF TEST: 02/04/2020 01:14:27

---

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.082	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.098	01:09
Breath Volume	3.152	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.099	01:12
Breath Volume	3.441	01:12
Air Blank	0.000	01:13
Control Test	0.082	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

---

Data Download Date/Time: 02/05/2020 18:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RODRIGUEZ,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: ZAVALA

DATE/TIME OF TEST: 02/05/2020 05:34:01

---

Diagnostics Check	OK	05:26
Air Blank	0.000	05:26
Control Test	0.081	05:26
Air Blank	0.000	05:27
Subject Sample # 1	0.198	05:28
Breath Volume	1.234	05:28
Air Blank	0.000	05:29
Air Blank	0.000	05:31
Subject Sample # 2	0.205	05:32
Breath Volume	2.519	05:32
Air Blank	0.000	05:33
Control Test	0.080	05:33
Air Blank	0.000	05:33
Diagnostics Check	OK	05:34

---

Data Download Date/Time: 02/05/2020 18:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURKE,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 02/06/2020 00:00:00

---

Diagnostics Check		OK	23:52
Air Blank		0.000	23:52
Control Test		0.081	23:53
Air Blank		0.000	23:53
Subject Sample	# 1	0.208	23:55
Breath Volume		1.433	23:55
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.208	23:58
Breath Volume		1.417	23:58
Air Blank		0.000	23:59
Control Test		0.080	23:59
Air Blank		0.000	23:59
Diagnostics Check		OK	00:00

---

Data Download Date/Time: 03/09/2020 16:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/06/2020 03:04:35

---

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.080	02:59	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:03	Subject Test Refused
Breath Volume	0.000	03:03	
Air Blank	0.000	03:03	
Control Test	0.080	03:04	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:04	

---

Data Download Date/Time: 03/09/2020 16:22



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ANDREWS,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: BURGOS

DATE/TIME OF TEST: 02/06/2020 05:57:29

---

Diagnostics Check	OK	05:51
Air Blank	0.000	05:51
Control Test	0.081	05:52
Air Blank	0.000	05:52
Subject Sample # 1	0.000	05:53
Breath Volume	3.402	05:53
Air Blank	0.000	05:53
Air Blank	0.000	05:55
Subject Sample # 2	0.000	05:56
Breath Volume	3.394	05:56
Air Blank	0.000	05:56
Control Test	0.081	05:56
Air Blank	0.000	05:57
Diagnostics Check	OK	05:57

---

Data Download Date/Time: 03/09/2020 16:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKIPPER,ROGER,G  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KENNEDY,D 129956  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 09:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:01

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 02/06/2020 10:28:28

---

Diagnostics Check	OK	10:22
Air Blank	0.000	10:22
Control Test	0.080	10:22
Air Blank	0.000	10:23
Subject Sample # 1	0.310	10:24
Breath Volume	3.246	10:24
Air Blank	0.000	10:24
Air Blank	0.000	10:26
Subject Sample # 2	0.297	10:26
Breath Volume	2.097	10:26
Air Blank	0.000	10:27
Control Test	0.080	10:27
Air Blank	0.000	10:28
Diagnostics Check	OK	10:28

---

Data Download Date/Time: 03/09/2020 16:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DRISCOLL,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 02/07/2020 20:31:03

---

Diagnostics Check	OK	20:23
Air Blank	0.000	20:24
Control Test	0.081	20:24
Air Blank	0.000	20:24
Subject Sample # 1	0.247	20:26
Breath Volume	3.875	20:26
Air Blank	0.000	20:27
Air Blank	0.000	20:28
Subject Sample # 2	0.246	20:29
Breath Volume	3.601	20:29
Air Blank	0.000	20:30
Control Test	0.080	20:30
Air Blank	0.000	20:30
Diagnostics Check	OK	20:31

---

Data Download Date/Time: 03/09/2020 16:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SULAIMAN,R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: MEADER

DATE/TIME OF TEST: 02/08/2020 02:40:50

---

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.081	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.127	02:36
Breath Volume	1.679	02:36
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.131	02:39
Breath Volume	3.820	02:39
Air Blank	0.000	02:39
Control Test	0.081	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

---

Data Download Date/Time: 03/09/2020 16:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: DEVELLO

DATE/TIME OF TEST: 02/08/2020 04:58:22

---

Diagnostics Check	OK	04:49	
Air Blank	0.000	04:50	
Control Test	0.082	04:50	
Air Blank	0.000	04:51	
Subject Sample	# 1	04:54	No Sample Provided
Breath Volume	0.000	04:54	
Air Blank	0.000	04:54	
Air Blank	0.000	04:56	
Subject Sample	# 2	04:57	Subject Test Refused
Breath Volume	0.000	04:57	
Air Blank	0.000	04:57	
Control Test	0.082	04:57	
Air Blank	0.000	04:58	
Diagnostics Check	OK	04:58	

---

Data Download Date/Time: 03/09/2020 16:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WASHINGTON,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: REYES MALDONADO

DATE/TIME OF TEST: 02/08/2020 23:17:42

---

Diagnostics Check	OK	23:10
Air Blank	0.000	23:11
Control Test	0.082	23:11
Air Blank	0.000	23:11
Subject Sample # 1	0.196	23:13
Breath Volume	2.847	23:13
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.197	23:16
Breath Volume	2.632	23:16
Air Blank	0.000	23:16
Control Test	0.082	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

---

Data Download Date/Time: 03/09/2020 16:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RENDON,R  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: SANCHEZ DIMAS

DATE/TIME OF TEST: 02/09/2020 05:41:18

---

Diagnostics Check	OK	05:35
Air Blank	0.000	05:35
Control Test	0.082	05:35
Air Blank	0.000	05:36
Subject Sample # 1	0.180	05:36
Breath Volume	2.437	05:36
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample # 2	0.177	05:39
Breath Volume	1.886	05:39
Air Blank	0.000	05:40
Control Test	0.082	05:40
Air Blank	0.000	05:41
Diagnostics Check	OK	05:41

---

Data Download Date/Time: 03/09/2020 16:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SCHAEFER,J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 02/09/2020 05:57:57

---

Diagnostics Check	OK	05:51
Air Blank	0.000	05:52
Control Test	0.082	05:52
Air Blank	0.000	05:53
Subject Sample # 1	0.188	05:53
Breath Volume	3.839	05:53
Air Blank	0.000	05:54
Air Blank	0.000	05:55
Subject Sample # 2	0.186	05:56
Breath Volume	3.406	05:56
Air Blank	0.000	05:57
Control Test	0.080	05:57
Air Blank	0.000	05:57
Diagnostics Check	OK	05:57

---

Data Download Date/Time: 03/09/2020 16:23



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR,BARBARA,H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BERG,T  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: BUI  
ARREST TIME: 13:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:27

SUBJECT'S LAST NAME: RUPERD

DATE/TIME OF TEST: 02/09/2020 14:55:56

---

Diagnostics Check	OK	14:49
Air Blank	0.000	14:50
Control Test	0.082	14:50
Air Blank	0.000	14:51
Subject Sample # 1	0.093	14:51
Breath Volume	4.355	14:51
Air Blank	0.000	14:52
Air Blank	0.000	14:54
Subject Sample # 2	0.096	14:54
Breath Volume	4.480	14:54
Air Blank	0.000	14:55
Control Test	0.081	14:55
Air Blank	0.000	14:55
Diagnostics Check	OK	14:55

---

Data Download Date/Time: 03/09/2020 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SCHAEFER,J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: JAKSEC

DATE/TIME OF TEST: 02/09/2020 21:35:36

---

Diagnostics Check	OK	21:28
Air Blank	0.000	21:28
Control Test	0.080	21:29
Air Blank	0.000	21:29
Subject Sample # 1	0.126	21:31
Breath Volume	1.988	21:31
Air Blank	0.000	21:31
Air Blank	0.000	21:33
Subject Sample # 2	0.125	21:34
Breath Volume	1.968	21:34
Air Blank	0.000	21:34
Control Test	0.081	21:35
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

---

Data Download Date/Time: 03/09/2020 16:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: JOHNSON III

DATE/TIME OF TEST: 02/10/2020 05:34:30

---

Diagnostics Check		OK	05:28
Air Blank		0.000	05:28
Control Test		0.081	05:28
Air Blank		0.000	05:29
Subject Sample	# 1	0.207	05:30
Breath Volume		2.367	05:30
Air Blank		0.000	05:30
Air Blank		0.000	05:32
Subject Sample	# 2	0.216	05:32
Breath Volume		2.675	05:32
Air Blank		0.000	05:33
Control Test		0.081	05:33
Air Blank		0.000	05:34
Diagnostics Check		OK	05:34

---

Data Download Date/Time: 03/09/2020 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAAR,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: FREEMAN JR

DATE/TIME OF TEST: 02/10/2020 21:39:58

---

Diagnostics Check		OK	21:28	
Air Blank		0.000	21:28	
Control Test		0.080	21:29	
Air Blank		0.000	21:29	
Subject Sample	# 1	0.174	21:32	Volume Not Met
Breath Volume		0.878	21:32	
Air Blank		0.000	21:33	
Air Blank		0.000	21:35	
Subject Sample	# 2	0.179	21:38	Volume Not Met
Breath Volume		0.980	21:38	
Air Blank		0.000	21:39	
Control Test		0.079	21:39	
Air Blank		0.000	21:39	
Diagnostics Check		OK	21:39	

---

Data Download Date/Time: 03/09/2020 16:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAAR,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: FREEMAN JR

DATE/TIME OF TEST: 02/10/2020 21:57:30

---

Diagnostics Check		OK	21:42	
Air Blank		0.000	21:43	
Control Test		0.080	21:43	
Air Blank		0.000	21:44	
Subject Sample	# 1	0.180	21:47	Volume Not Met
Breath Volume		0.031	21:47	
Air Blank		0.000	21:47	
Air Blank		0.000	21:49	
Subject Sample	# 2	0.192	21:50	
Breath Volume		1.343	21:50	
Air Blank		0.000	21:50	
Air Blank		0.000	21:52	
Subject Sample	# 3	0.171	21:55	Volume Not Met
Breath Volume		1.042	21:55	
Air Blank		0.000	21:56	
Control Test		0.079	21:56	
Air Blank		0.000	21:57	
Diagnostics Check		OK	21:57	

---

Data Download Date/Time: 03/09/2020 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILLIAMS,R  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 02/11/2020 19:24:44

---

Diagnostics Check		OK	19:17
Air Blank		0.000	19:18
Control Test		0.081	19:18
Air Blank		0.000	19:19
Subject Sample	# 1	0.267	19:19
Breath Volume		1.300	19:19
Air Blank		0.000	19:20
Air Blank		0.000	19:22
Subject Sample	# 2	0.261	19:23
Breath Volume		2.171	19:23
Air Blank		0.000	19:23
Control Test		0.079	19:24
Air Blank		0.000	19:24
Diagnostics Check		OK	19:24

---

Data Download Date/Time: 03/09/2020 16:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: THORNE,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: RAMOS-MANDELL

DATE/TIME OF TEST: 02/11/2020 22:28:57

---

Diagnostics Check	OK	22:22
Air Blank	0.000	22:23
Control Test	0.081	22:23
Air Blank	0.000	22:24
Subject Sample # 1	0.000	22:24
Breath Volume	1.304	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.000	22:27
Breath Volume	1.378	22:27
Air Blank	0.000	22:28
Control Test	0.081	22:28
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

---

Data Download Date/Time: 03/09/2020 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: CARLYLE

DATE/TIME OF TEST: 02/13/2020 02:43:55

---

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.082	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.112	02:37
Breath Volume	1.726	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.108	02:42
Breath Volume	1.886	02:42
Air Blank	0.000	02:43
Control Test	0.081	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

---

Data Download Date/Time: 03/09/2020 16:23



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DELORENZO,D  
ARRESTING OFFICER AGENCY: Temple Terrace PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: KNAPP JR

DATE/TIME OF TEST: 02/14/2020 01:04:35

---

Diagnostics Check	OK	00:58
Air Blank	0.000	00:59
Control Test	0.081	00:59
Air Blank	0.000	01:00
Subject Sample # 1	0.120	01:00
Breath Volume	1.570	01:00
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.114	01:03
Breath Volume	3.226	01:03
Air Blank	0.000	01:03
Control Test	0.079	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

---

Data Download Date/Time: 03/09/2020 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SCHAEFER,J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 02/14/2020 01:44:55

---

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.081	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.148	01:40
Breath Volume	1.503	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:42
Subject Sample # 2	0.155	01:43
Breath Volume	3.046	01:43
Air Blank	0.000	01:44
Control Test	0.081	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

---

Data Download Date/Time: 03/09/2020 16:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FERGUSON,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MORAD

DATE/TIME OF TEST: 02/15/2020 02:58:40

---

Diagnostics Check		OK	02:52
Air Blank		0.000	02:53
Control Test		0.082	02:53
Air Blank		0.000	02:53
Subject Sample	# 1	0.106	02:54
Breath Volume		2.167	02:54
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.103	02:57
Breath Volume		2.230	02:57
Air Blank		0.000	02:57
Control Test		0.081	02:58
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

---

Data Download Date/Time: 03/09/2020 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURKE,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/15/2020 03:51:08

---

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.080	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.120	03:46
Breath Volume	4.109	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.112	03:49
Breath Volume	2.347	03:49
Air Blank	0.000	03:50
Control Test	0.082	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

---

Data Download Date/Time: 03/09/2020 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: SOTOMAYOR PEREZ

DATE/TIME OF TEST: 02/15/2020 04:57:16

---

Diagnostics Check		OK	04:49
Air Blank		0.000	04:50
Control Test		0.081	04:50
Air Blank		0.000	04:51
Subject Sample	# 1	0.078	04:52
Breath Volume		1.363	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:54
Subject Sample	# 2	0.078	04:55
Breath Volume		1.265	04:55
Air Blank		0.000	04:56
Control Test		0.081	04:56
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

---

Data Download Date/Time: 03/09/2020 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FERGUSON,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 02/16/2020 01:14:36

---

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.081	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.190	01:09
Breath Volume	3.367	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.194	01:13
Breath Volume	3.535	01:13
Air Blank	0.000	01:13
Control Test	0.082	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

---

Data Download Date/Time: 03/09/2020 16:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TIBURCIO,I  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 02/16/2020 03:20:54

---

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.081	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.093	03:16
Breath Volume	3.191	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.090	03:19
Breath Volume	3.871	03:19
Air Blank	0.000	03:20
Control Test	0.081	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

---

Data Download Date/Time: 03/09/2020 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HALL,L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 02/16/2020 04:35:37

---

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.081	04:29
Air Blank	0.000	04:30
Subject Sample # 1	0.194	04:31
Breath Volume	1.812	04:31
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.195	04:34
Breath Volume	2.867	04:34
Air Blank	0.000	04:34
Control Test	0.081	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

---

Data Download Date/Time: 03/09/2020 16:23



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HOWARD,D  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 02/16/2020 05:09:38

---

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.079	05:03
Air Blank		0.000	05:04
Subject Sample	# 1	0.168	05:05
Breath Volume		3.433	05:05
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.169	05:08
Breath Volume		3.187	05:08
Air Blank		0.000	05:08
Control Test		0.080	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

---

Data Download Date/Time: 03/09/2020 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER,MELONIE,E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FISHER,T  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:08

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 02/16/2020 07:35:27

---

Diagnostics Check		OK	07:28
Air Blank		0.000	07:29
Control Test		0.081	07:29
Air Blank		0.000	07:30
Subject Sample	# 1	0.127	07:30
Breath Volume		1.871	07:30
Air Blank		0.000	07:31
Air Blank		0.000	07:33
Subject Sample	# 2	0.124	07:33
Breath Volume		1.394	07:33
Air Blank		0.000	07:34
Control Test		0.081	07:34
Air Blank		0.000	07:35
Diagnostics Check		OK	07:35

---

Data Download Date/Time: 03/09/2020 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HARRIS,S  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: APONTE

DATE/TIME OF TEST: 02/16/2020 22:37:27

---

Diagnostics Check		OK	22:30
Air Blank		0.000	22:31
Control Test		0.081	22:31
Air Blank		0.000	22:31
Subject Sample	# 1	0.164	22:32
Breath Volume		2.214	22:32
Air Blank		0.000	22:33
Air Blank		0.000	22:35
Subject Sample	# 2	0.155	22:35
Breath Volume		1.621	22:35
Air Blank		0.000	22:36
Control Test		0.081	22:36
Air Blank		0.000	22:37
Diagnostics Check		OK	22:37

---

Data Download Date/Time: 03/09/2020 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WASHINGTON,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: MULROY

DATE/TIME OF TEST: 02/17/2020 22:24:37

---

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.081	22:18
Air Blank	0.000	22:19
Subject Sample # 1	0.000	22:20
Breath Volume	3.156	22:20
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.000	22:23
Breath Volume	2.710	22:23
Air Blank	0.000	22:23
Control Test	0.082	22:24
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

---

Data Download Date/Time: 03/09/2020 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RENDON,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: AMBROSIO JOSE

DATE/TIME OF TEST: 02/18/2020 22:15:26

---

Diagnostics Check		OK	22:03	
Air Blank		0.000	22:03	
Control Test		0.081	22:04	
Air Blank		0.000	22:04	
Subject Sample	# 1	0.292	22:07	Slope Not Level
Breath Volume		1.664	22:07	
Air Blank		0.000	22:08	
Air Blank		0.000	22:10	
Subject Sample	# 2	0.308	22:10	
Breath Volume		1.683	22:10	
Air Blank		0.000	22:11	
Air Blank		0.000	22:13	
Subject Sample	# 3	0.313	22:13	
Breath Volume		1.566	22:13	
Air Blank		0.000	22:14	
Control Test		0.079	22:14	
Air Blank		0.000	22:15	
Diagnostics Check		OK	22:15	

---

Data Download Date/Time: 03/09/2020 16:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKIPPER,ROGER,G  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TAVARES,N 450  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 07:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:53

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/19/2020 09:22:17

---

Diagnostics Check	OK	09:16
Air Blank	0.000	09:16
Control Test	0.080	09:16
Air Blank	0.000	09:17
Subject Sample # 1	0.183	09:17
Breath Volume	4.160	09:17
Air Blank	0.000	09:18
Air Blank	0.000	09:20
Subject Sample # 2	0.182	09:20
Breath Volume	3.710	09:20
Air Blank	0.000	09:21
Control Test	0.081	09:21
Air Blank	0.000	09:22
Diagnostics Check	OK	09:22

---

Data Download Date/Time: 03/09/2020 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FERGUSON,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/20/2020 00:03:56

---

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.080	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.118	23:59
Breath Volume	2.035	23:59
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.120	00:02
Breath Volume	1.253	00:02
Air Blank	0.000	00:03
Control Test	0.079	00:03
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

---

Data Download Date/Time: 03/09/2020 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURKE,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: MENDEZ-IBARRA

DATE/TIME OF TEST: 02/20/2020 01:01:09

---

Diagnostics Check		OK	00:53
Air Blank		0.000	00:53
Control Test		0.080	00:54
Air Blank		0.000	00:54
Subject Sample	# 1	0.034	00:55
Breath Volume		1.261	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.032	00:59
Breath Volume		1.503	00:59
Air Blank		0.000	01:00
Control Test		0.081	01:00
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

---

Data Download Date/Time: 03/09/2020 16:23



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURKE,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MEDINA TORRES

DATE/TIME OF TEST: 02/21/2020 01:21:59

---

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.081	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.049	01:17
Breath Volume	2.164	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.047	01:20
Breath Volume	1.371	01:20
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

---

Data Download Date/Time: 03/09/2020 16:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: THORNE,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: MONTIEL MORENO

DATE/TIME OF TEST: 02/21/2020 02:03:04

---

Diagnostics Check		OK	01:47	
Air Blank		0.000	01:48	
Control Test		0.081	01:48	
Air Blank		0.000	01:49	
Subject Sample	# 1	0.119	01:52	Volume Not Met
Breath Volume		0.621	01:52	
Air Blank		0.000	01:53	
Air Blank		0.000	01:54	
Subject Sample	# 2	0.139	01:57	
Breath Volume		1.718	01:57	
Air Blank		0.000	01:57	
Air Blank		0.000	01:59	
Subject Sample	# 3	0.139	02:01	
Breath Volume		1.734	02:01	
Air Blank		0.000	02:02	
Control Test		0.079	02:02	
Air Blank		0.000	02:02	
Diagnostics Check		OK	02:03	

---

Data Download Date/Time: 03/09/2020 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: ALSTON

DATE/TIME OF TEST: 02/21/2020 21:20:51

---

Diagnostics Check	OK	21:13
Air Blank	0.000	21:14
Control Test	0.081	21:14
Air Blank	0.000	21:14
Subject Sample # 1	0.172	21:16
Breath Volume	1.906	21:16
Air Blank	0.000	21:16
Air Blank	0.000	21:18
Subject Sample # 2	0.178	21:19
Breath Volume	2.558	21:19
Air Blank	0.000	21:19
Control Test	0.081	21:20
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

---

Data Download Date/Time: 03/09/2020 16:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LYON,M  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: GIRARD

DATE/TIME OF TEST: 02/22/2020 01:49:57

---

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:38	
Control Test	0.081	01:38	
Air Blank	0.000	01:39	
Subject Sample	# 1 0.193	01:41	
Breath Volume	1.246	01:41	
Air Blank	0.000	01:41	
Air Blank	0.000	01:43	
Subject Sample	# 2 0.216	01:45	
Breath Volume	2.742	01:45	
Air Blank	0.000	01:45	
Air Blank	0.000	01:47	
Subject Sample	# 3	01:48	Subject Test Refused
Breath Volume	0.000	01:48	
Air Blank	0.000	01:49	
Control Test	0.080	01:49	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

---

Data Download Date/Time: 03/09/2020 16:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SULAIMAN,R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: SIRICO

DATE/TIME OF TEST: 02/22/2020 02:47:09

---

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.081	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.000	02:42
Breath Volume		2.660	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.000	02:45
Breath Volume		2.617	02:45
Air Blank		0.000	02:46
Control Test		0.081	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

---

Data Download Date/Time: 03/09/2020 16:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ARCHAMBEAU,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: MARCOS

DATE/TIME OF TEST: 02/22/2020 04:16:25

---

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.081	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.172	04:11
Breath Volume	1.230	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.162	04:14
Breath Volume	1.976	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

---

Data Download Date/Time: 03/09/2020 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DRISCOLL,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: HERNANDEZ CIVARDI

DATE/TIME OF TEST: 02/22/2020 04:49:22

---

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:41	
Control Test	0.081	04:41	
Air Blank	0.000	04:42	
Subject Sample	# 1 0.146	04:45	Volume Not Met
Breath Volume	0.085	04:45	
Air Blank	0.000	04:45	
Air Blank	0.000	04:47	
Subject Sample	# 2	04:48	Subject Test Refused
Breath Volume	0.000	04:48	
Air Blank	0.000	04:48	
Control Test	0.080	04:48	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:49	

---

Data Download Date/Time: 03/09/2020 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: LAMONDRA

DATE/TIME OF TEST: 02/22/2020 05:02:51

---

Diagnostics Check	OK	04:55
Air Blank	0.000	04:56
Control Test	0.080	04:56
Air Blank	0.000	04:57
Subject Sample # 1	0.167	04:58
Breath Volume	1.609	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:00
Subject Sample # 2	0.159	05:01
Breath Volume	1.113	05:01
Air Blank	0.000	05:01
Control Test	0.081	05:02
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

---

Data Download Date/Time: 03/09/2020 16:23



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS,ANGELA,V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CARSON,J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 09:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:56

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 02/22/2020 10:22:42

---

Diagnostics Check	OK	10:16
Air Blank	0.000	10:16
Control Test	0.080	10:17
Air Blank	0.000	10:17
Subject Sample # 1	0.309	10:18
Breath Volume	1.378	10:18
Air Blank	0.000	10:18
Air Blank	0.000	10:20
Subject Sample # 2	0.303	10:21
Breath Volume	2.203	10:21
Air Blank	0.000	10:21
Control Test	0.079	10:22
Air Blank	0.000	10:22
Diagnostics Check	OK	10:22

---

Data Download Date/Time: 03/09/2020 16:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ROSALES,F  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: CASTILLO CRUZ

DATE/TIME OF TEST: 02/22/2020 22:11:36

---

Diagnostics Check		OK	22:04
Air Blank		0.000	22:05
Control Test		0.081	22:05
Air Blank		0.000	22:06
Subject Sample	# 1	0.252	22:07
Breath Volume		2.417	22:07
Air Blank		0.000	22:07
Air Blank		0.000	22:09
Subject Sample	# 2	0.252	22:10
Breath Volume		1.882	22:10
Air Blank		0.000	22:10
Control Test		0.081	22:11
Air Blank		0.000	22:11
Diagnostics Check		OK	22:11

---

Data Download Date/Time: 03/09/2020 16:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ANDREWS,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: BLAND II

DATE/TIME OF TEST: 02/23/2020 01:00:49

---

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.080	00:55
Air Blank	0.000	00:56
Subject Sample # 1	0.135	00:56
Breath Volume	1.546	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:58
Subject Sample # 2	0.137	00:59
Breath Volume	1.816	00:59
Air Blank	0.000	00:59
Control Test	0.079	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

---

Data Download Date/Time: 03/09/2020 16:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ARCHAMBEAU,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 02/23/2020 04:16:08

---

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.082	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.125	04:11
Breath Volume	1.136	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.129	04:14
Breath Volume	1.636	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

---

Data Download Date/Time: 03/09/2020 16:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: ZAYAS JR

DATE/TIME OF TEST: 02/23/2020 04:33:29

---

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.080	04:28
Air Blank	0.000	04:28
Subject Sample # 1	0.137	04:28
Breath Volume	1.500	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.131	04:31
Breath Volume	1.675	04:31
Air Blank	0.000	04:32
Control Test	0.080	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

---

Data Download Date/Time: 03/09/2020 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: BEDFORD III

DATE/TIME OF TEST: 02/23/2020 05:11:24

---

Diagnostics Check	OK	05:04
Air Blank	0.000	05:05
Control Test	0.080	05:05
Air Blank	0.000	05:05
Subject Sample # 1	0.128	05:06
Breath Volume	1.804	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.125	05:09
Breath Volume	4.160	05:09
Air Blank	0.000	05:10
Control Test	0.080	05:10
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

---

Data Download Date/Time: 03/09/2020 16:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BERG,T  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: BLEVINS

DATE/TIME OF TEST: 02/23/2020 21:34:05

---

Diagnostics Check	OK	21:27
Air Blank	0.000	21:28
Control Test	0.081	21:28
Air Blank	0.000	21:29
Subject Sample # 1	0.096	21:29
Breath Volume	4.046	21:29
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample # 2	0.096	21:32
Breath Volume	3.910	21:32
Air Blank	0.000	21:33
Control Test	0.080	21:33
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

---

Data Download Date/Time: 03/09/2020 16:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/24/2020 03:35:38

---

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.080	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.090	03:31
Breath Volume	1.335	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.086	03:34
Breath Volume	1.175	03:34
Air Blank	0.000	03:34
Control Test	0.079	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

---

Data Download Date/Time: 03/09/2020 16:24



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FISHER,T  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: KALES

DATE/TIME OF TEST: 02/25/2020 00:24:51

---

Diagnostics Check		OK	00:18
Air Blank		0.000	00:19
Control Test		0.081	00:19
Air Blank		0.000	00:20
Subject Sample	# 1	0.110	00:20
Breath Volume		1.593	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.115	00:23
Breath Volume		1.703	00:23
Air Blank		0.000	00:23
Control Test		0.081	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

---

Data Download Date/Time: 03/09/2020 16:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: THORNE,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 02/25/2020 20:42:30

---

Diagnostics Check		OK	20:31	
Air Blank		0.000	20:31	
Control Test		0.081	20:31	
Air Blank		0.000	20:32	
Subject Sample	# 1	0.101	20:35	Volume Not Met
Breath Volume		0.089	20:35	
Air Blank		0.000	20:36	
Air Blank		0.000	20:37	
Subject Sample	# 2	0.093	20:41	Volume Not Met
Breath Volume		0.214	20:41	
Air Blank		0.000	20:41	
Control Test		0.080	20:41	
Air Blank		0.000	20:42	
Diagnostics Check		OK	20:42	

---

Data Download Date/Time: 03/09/2020 16:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAAR,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: DEEN JR

DATE/TIME OF TEST: 02/25/2020 23:24:16

---

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.081	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.125	23:19
Breath Volume	2.843	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.129	23:22
Breath Volume	2.402	23:22
Air Blank	0.000	23:23
Control Test	0.081	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

---

Data Download Date/Time: 03/09/2020 16:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR,BARBARA,H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PERALTA,A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:31

SUBJECT'S LAST NAME: SCHAFFER

DATE/TIME OF TEST: 02/27/2020 16:01:50

---

Diagnostics Check	OK	15:54
Air Blank	0.000	15:54
Control Test	0.081	15:55
Air Blank	0.000	15:55
Subject Sample # 1	0.078	15:57
Breath Volume	1.300	15:57
Air Blank	0.000	15:57
Air Blank	0.000	15:59
Subject Sample # 2	0.077	16:00
Breath Volume	1.203	16:00
Air Blank	0.000	16:00
Control Test	0.080	16:01
Air Blank	0.000	16:01
Diagnostics Check	OK	16:01

---

Data Download Date/Time: 03/09/2020 16:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 02/27/2020 20:30:16

---

Diagnostics Check	OK	20:18	
Air Blank	0.000	20:19	
Control Test	0.081	20:19	
Air Blank	0.000	20:20	
Subject Sample	# 1	20:23	No Sample Provided
Breath Volume	0.000	20:23	
Air Blank	0.000	20:23	
Air Blank	0.000	20:25	
Subject Sample	# 2	20:28	No Sample Provided
Breath Volume	0.000	20:28	
Air Blank	0.000	20:29	
Control Test	0.080	20:29	
Air Blank	0.000	20:30	
Diagnostics Check	OK	20:30	

---

Data Download Date/Time: 03/09/2020 16:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS,ANGELA,V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BYERS,JAMES  
ARRESTING OFFICER AGENCY: Macdill Air Force Ba  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: ADMIN  
ARREST TIME: 11:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:06

SUBJECT'S LAST NAME: ESTOCK

DATE/TIME OF TEST: 02/28/2020 11:40:43

---

Diagnostics Check	OK	11:34
Air Blank	0.000	11:34
Control Test	0.081	11:35
Air Blank	0.000	11:35
Subject Sample # 1	0.047	11:36
Breath Volume	1.648	11:36
Air Blank	0.000	11:36
Air Blank	0.000	11:38
Subject Sample # 2	0.048	11:39
Breath Volume	1.675	11:39
Air Blank	0.000	11:39
Control Test	0.082	11:40
Air Blank	0.000	11:40
Diagnostics Check	OK	11:40

---

Data Download Date/Time: 03/09/2020 16:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HUTCHESON,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: LAPINE

DATE/TIME OF TEST: 02/28/2020 21:31:24

---

Diagnostics Check		OK	21:24
Air Blank		0.000	21:24
Control Test		0.081	21:25
Air Blank		0.000	21:25
Subject Sample	# 1	0.172	21:26
Breath Volume		4.796	21:26
Air Blank		0.000	21:27
Air Blank		0.000	21:29
Subject Sample	# 2	0.170	21:29
Breath Volume		4.503	21:29
Air Blank		0.000	21:30
Control Test		0.082	21:30
Air Blank		0.000	21:31
Diagnostics Check		OK	21:31

---

Data Download Date/Time: 03/09/2020 16:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VANTREESE,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: CHRISTIE

DATE/TIME OF TEST: 02/29/2020 01:41:57

---

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.081	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.176	01:37
Breath Volume	2.386	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.182	01:40
Breath Volume	1.691	01:40
Air Blank	0.000	01:41
Control Test	0.081	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

---

Data Download Date/Time: 03/09/2020 16:24



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: THORNE,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: DE GRAFF

DATE/TIME OF TEST: 02/29/2020 02:33:07

---

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.081	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.092	02:28
Breath Volume	1.527	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:30
Subject Sample # 2	0.097	02:31
Breath Volume	1.691	02:31
Air Blank	0.000	02:32
Control Test	0.081	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

---

Data Download Date/Time: 03/09/2020 16:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: REYES SANTIAGO

DATE/TIME OF TEST: 02/29/2020 04:36:10

---

Diagnostics Check		OK	04:30
Air Blank		0.000	04:30
Control Test		0.080	04:31
Air Blank		0.000	04:31
Subject Sample	# 1	0.077	04:31
Breath Volume		2.410	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.080	04:34
Breath Volume		1.886	04:34
Air Blank		0.000	04:35
Control Test		0.081	04:35
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

---

Data Download Date/Time: 03/09/2020 16:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: THORNE,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: MARTIN PEREZ

DATE/TIME OF TEST: 02/29/2020 04:59:35

---

Diagnostics Check		OK	04:53
Air Blank		0.000	04:53
Control Test		0.080	04:53
Air Blank		0.000	04:54
Subject Sample	# 1	0.106	04:54
Breath Volume		2.527	04:54
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.108	04:58
Breath Volume		1.367	04:58
Air Blank		0.000	04:58
Control Test		0.079	04:59
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

---

Data Download Date/Time: 03/09/2020 16:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: LUMPKIN

DATE/TIME OF TEST: 02/29/2020 06:47:18

---

Diagnostics Check	OK	06:35	
Air Blank	0.000	06:36	
Control Test	0.081	06:36	
Air Blank	0.000	06:37	
Subject Sample	# 1	06:40	No Sample Provided
Breath Volume	0.000	06:40	
Air Blank	0.000	06:40	
Air Blank	0.000	06:42	
Subject Sample	# 2	06:45	No Sample Provided
Breath Volume	0.000	06:45	
Air Blank	0.000	06:46	
Control Test	0.082	06:46	
Air Blank	0.000	06:47	
Diagnostics Check	OK	06:47	

---

Data Download Date/Time: 03/09/2020 16:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:42

SUBJECT'S LAST NAME: LUMPKIN

DATE/TIME OF TEST: 02/29/2020 07:09:48

---

Diagnostics Check		OK	07:02
Air Blank		0.000	07:03
Control Test		0.081	07:03
Air Blank		0.000	07:04
Subject Sample	# 1	0.154	07:05
Breath Volume		1.667	07:05
Air Blank		0.000	07:06
Air Blank		0.000	07:07
Subject Sample	# 2	0.152	07:08
Breath Volume		1.703	07:08
Air Blank		0.000	07:08
Control Test		0.080	07:09
Air Blank		0.000	07:09
Diagnostics Check		OK	07:09

---

Data Download Date/Time: 03/09/2020 16:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VISSER,A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: KOTLYAR

DATE/TIME OF TEST: 02/29/2020 22:30:26

---

Diagnostics Check	OK	22:24
Air Blank	0.000	22:24
Control Test	0.082	22:24
Air Blank	0.000	22:25
Subject Sample # 1	0.007	22:26
Breath Volume	1.335	22:26
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.007	22:29
Breath Volume	1.488	22:29
Air Blank	0.000	22:29
Control Test	0.082	22:29
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

---

Data Download Date/Time: 03/09/2020 16:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOBOSIEWICZ,ANTHONY,N  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: DOBOSIEWICZ,ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: MCEACHIN

DATE/TIME OF TEST: 02/07/2020 21:22:06

---

Diagnostics Check	OK	21:13	
Air Blank	0.000	21:14	
Control Test	0.081	21:14	
Air Blank	0.000	21:14	
Subject Sample	# 1	21:18	No Sample Provided
Breath Volume	0.000	21:18	
Air Blank	0.000	21:18	
Air Blank	0.000	21:20	
Subject Sample	# 2	21:20	Subject Test Refused
Breath Volume	0.000	21:20	
Air Blank	0.000	21:21	
Control Test	0.081	21:21	
Air Blank	0.000	21:22	
Diagnostics Check	OK	21:22	

---

Data Download Date/Time: 02/11/2020 10:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANO,BRANDON,T  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: CANO,BRANDON  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1185592  
DRY GAS EXPIRY DATE: 08/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:54

SUBJECT'S LAST NAME: AMBROSE

DATE/TIME OF TEST: 02/17/2020 17:27:29

---

Diagnostics Check	OK	17:21
Air Blank	0.000	17:21
Control Test	0.081	17:22
Air Blank	0.000	17:22
Subject Sample # 1	0.011	17:23
Breath Volume	1.511	17:23
Air Blank	0.000	17:23
Air Blank	0.000	17:25
Subject Sample # 2	0.012	17:26
Breath Volume	1.480	17:26
Air Blank	0.000	17:26
Control Test	0.083	17:26
Air Blank	0.000	17:27
Diagnostics Check	OK	17:27

---

Data Download Date/Time: 02/25/2020 08:36



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JATEFF, TIFFANY, J  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: CANO, BRANDON  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1185592  
DRY GAS EXPIRY DATE: 08/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:13

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 02/23/2020 17:34:16

---

Diagnostics Check	OK	17:33	
Air Blank		17:33	RFI Detect
Air Blank	0.000	17:34	

---

Data Download Date/Time: 02/25/2020 08:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANO, BRANDON, T  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: CANO, BRANDON  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1185592  
DRY GAS EXPIRY DATE: 08/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:13

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 02/23/2020 17:43:04

---

Diagnostics Check	OK	17:37	
Air Blank	0.000	17:37	
Control Test	0.082	17:38	
Air Blank	0.000	17:38	
Subject Sample # 1	0.246	17:41	Volume Not Met
Breath Volume	0.746	17:41	
Air Blank		17:42	Purge Fail
Air Blank	0.000	17:43	

---

Data Download Date/Time: 02/25/2020 08:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANO,BRANDON,T  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: CANO,BRANDON  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1185592  
DRY GAS EXPIRY DATE: 08/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:13

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 02/23/2020 18:00:35

---

Diagnostics Check		OK	17:45	
Air Blank		0.000	17:46	
Control Test		0.080	17:46	
Air Blank		0.000	17:47	
Subject Sample	# 1	0.243	17:50	Volume Not Met
Breath Volume		0.906	17:50	
Air Blank		0.000	17:51	
Air Blank		0.000	17:52	
Subject Sample	# 2	0.252	17:56	
Breath Volume		1.593	17:56	
Air Blank		0.000	17:56	
Air Blank		0.000	17:58	
Subject Sample	# 3	0.252	17:59	
Breath Volume		1.292	17:59	
Air Blank		0.000	17:59	
Control Test		0.079	18:00	
Air Blank		0.000	18:00	
Diagnostics Check		OK	18:00	

---

Data Download Date/Time: 02/25/2020 08:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAVALLE,DANIEL,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: LAVALLE,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 01/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1101850  
DRY GAS EXPIRY DATE: 03/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: DAMM

DATE/TIME OF TEST: 02/05/2020 01:15:31

---

Diagnostics Check		OK	01:09
Air Blank		0.000	01:09
Control Test		0.079	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.211	01:10
Breath Volume		2.300	01:10
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.204	01:13
Breath Volume		1.183	01:13
Air Blank		0.000	01:14
Control Test		0.079	01:14
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

---

Data Download Date/Time: 02/28/2020 17:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HELM,DEREK,W  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HELM,DEREK  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 01/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1101850  
DRY GAS EXPIRY DATE: 03/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/07/2020 22:46:15

---

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.081	22:39
Air Blank	0.000	22:39
Subject Sample # 1	0.146	22:40
Breath Volume	1.488	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.164	22:44
Breath Volume	1.546	22:44
Air Blank	0.000	22:45
Control Test	0.081	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

---

Data Download Date/Time: 02/28/2020 17:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAVALLE,DANIEL,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: LAVALLE,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 01/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1101850  
DRY GAS EXPIRY DATE: 03/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: PROFFER

DATE/TIME OF TEST: 02/15/2020 01:57:25

---

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.173	01:52
Breath Volume	4.191	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:54
Subject Sample # 2	0.164	01:55
Breath Volume	3.656	01:55
Air Blank	0.000	01:56
Control Test	0.078	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

---

Data Download Date/Time: 02/28/2020 17:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAVALLE,DANIEL,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: LAVALLE,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 01/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1101850  
DRY GAS EXPIRY DATE: 03/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: KUNOW

DATE/TIME OF TEST: 02/21/2020 02:01:55

---

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:01	
Control Test	0.079	02:01	
Air Blank		02:01	RFI Detect
Air Blank		02:01	RFI Detect

---

Data Download Date/Time: 02/28/2020 17:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAVALLE,DANIEL,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: LAVALLE,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 01/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1101850  
DRY GAS EXPIRY DATE: 03/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: KUNOW

DATE/TIME OF TEST: 02/21/2020 02:11:32

---

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.080	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.172	02:06
Breath Volume	1.113	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:08
Subject Sample # 2	0.186	02:09
Breath Volume	1.859	02:09
Air Blank	0.000	02:10
Control Test	0.080	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

---

Data Download Date/Time: 02/28/2020 17:53



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOOD,DERRICK,A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: TIMMONS,ALLAN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 01/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: HAHN

DATE/TIME OF TEST: 02/09/2020 21:46:33

---

Diagnostics Check	OK	21:46	
Air Blank		21:46	RFI Detect
Air Blank		21:46	RFI Detect

---

Data Download Date/Time: 02/26/2020 11:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOOD,DERRICK,A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: TIMMONS,ALLAN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 01/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: HAHN

DATE/TIME OF TEST: 02/09/2020 21:50:11

---

Diagnostics Check	OK	21:49	
Air Blank		21:49	RFI Detect
Air Blank	0.000	21:50	

---

Data Download Date/Time: 02/26/2020 11:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOOD,DERRICK,A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: TIMMONS,ALLAN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 01/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: HAHN

DATE/TIME OF TEST: 02/09/2020 21:57:58

---

Diagnostics Check	OK	21:51	
Air Blank	0.000	21:52	
Control Test	0.082	21:52	
Air Blank	0.000	21:52	
Subject Sample # 1	0.196	21:53	
Breath Volume	1.890	21:53	
Air Blank	0.000	21:54	
Air Blank	0.000	21:56	
Subject Sample # 2		21:56	Subject Test Refused
Breath Volume	0.000	21:56	
Air Blank	0.000	21:57	
Control Test	0.080	21:57	
Air Blank	0.000	21:57	
Diagnostics Check	OK	21:57	

---

Data Download Date/Time: 02/26/2020 11:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYERS,MICHELLE,H  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: SAYERS,MICHELLE  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: RODRIGUEZ-RAMIREZ

DATE/TIME OF TEST: 02/19/2020 00:57:57

---

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:50	
Control Test	0.078	00:51	
Air Blank	0.000	00:51	
Subject Sample # 1	0.168	00:54	Volume Not Met
Breath Volume	0.046	00:54	
Air Blank	0.000	00:55	
Air Blank	0.000	00:57	
		00:57	Sequence Aborted
Air Blank	0.000	00:57	

---

Data Download Date/Time: 02/26/2020 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYERS,MICHELLE,H  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: SAYERS,MICHELLE  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: RODRIGUEZ-RAMIREZ

DATE/TIME OF TEST: 02/19/2020 01:27:14

---

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:16	
Control Test	0.077	01:16	
Air Blank	0.000	01:16	
Subject Sample	# 1	01:20	No Sample Provided
Breath Volume	0.000	01:20	
Air Blank	0.000	01:20	
Air Blank	0.000	01:22	
Subject Sample	# 2	01:25	No Sample Provided
Breath Volume	0.000	01:25	
Air Blank	0.000	01:26	
Control Test	0.077	01:26	
Air Blank	0.000	01:27	
Diagnostics Check	OK	01:27	

---

Data Download Date/Time: 02/26/2020 15:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYERS,MICHELLE,H  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: SAYERS,MICHELLE  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PINA-OSORIO

DATE/TIME OF TEST: 02/22/2020 02:12:30

---

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.080	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.071	02:08
Breath Volume	2.464	02:08
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.068	02:10
Breath Volume	3.203	02:10
Air Blank	0.000	02:11
Control Test	0.081	02:11
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

---

Data Download Date/Time: 02/26/2020 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYERS,MICHELLE,H  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: SAYERS,MICHELLE  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/22/2020 21:55:35

---

Diagnostics Check	OK	21:46
Air Blank	0.000	21:47
Control Test	0.080	21:47
Air Blank	0.000	21:48
Subject Sample # 1	0.187	21:50
Breath Volume	1.734	21:50
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.172	21:53
Breath Volume	2.023	21:53
Air Blank	0.000	21:54
Control Test	0.079	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

---

Data Download Date/Time: 02/26/2020 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYERS,MICHELLE,H  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: SAYERS,MICHELLE  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: APONTE

DATE/TIME OF TEST: 02/23/2020 04:10:26

---

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.073	04:04
Breath Volume	1.695	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.074	04:08
Breath Volume	1.941	04:08
Air Blank	0.000	04:09
Control Test	0.079	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

---

Data Download Date/Time: 02/26/2020 15:08



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 01/27/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 02:11

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: MACON STRICKLAND

DATE/TIME OF TEST: 02/08/2020 02:37:20

---

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.081	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.165	02:32
Breath Volume	1.593	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.166	02:35
Breath Volume	1.875	02:35
Air Blank	0.000	02:36
Control Test	0.081	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

---

Data Download Date/Time: 02/11/2020 10:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MULDERRIG,SEAN,F  
BREATH TEST OPERATOR AGENCY: Lakeland PD  
ARRESTING OFFICER: MULDERRIG,SEAN  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: SPEARS

DATE/TIME OF TEST: 02/22/2020 19:24:09

---

Diagnostics Check	OK	19:17
Air Blank	0.000	19:18
Control Test	0.081	19:18
Air Blank	0.000	19:18
Subject Sample # 1	0.176	19:19
Breath Volume	2.300	19:19
Air Blank	0.000	19:20
Air Blank	0.000	19:22
Subject Sample # 2	0.189	19:22
Breath Volume	2.488	19:22
Air Blank	0.000	19:23
Control Test	0.081	19:23
Air Blank	0.000	19:24
Diagnostics Check	OK	19:24

---

Data Download Date/Time: 03/03/2020 10:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAWFORD,DAVID,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: CRAWFORD,DAVID  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: FONSECA RODRIGUEZ

DATE/TIME OF TEST: 02/06/2020 03:38:31

---

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:33	
Control Test	0.078	03:33	
Air Blank	0.000	03:34	
Subject Sample	# 1 0.189	03:34	
Breath Volume	2.550	03:34	
Air Blank	0.000	03:35	
Air Blank	0.000	03:37	
Subject Sample	# 2	03:38	Sequence Aborted
Breath Volume	0.000	03:38	
Air Blank	0.000	03:38	

---

Data Download Date/Time: 02/07/2020 19:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAWFORD,DAVID,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: CRAWFORD,DAVID  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: FONSECA RODRIGUEZ

DATE/TIME OF TEST: 02/06/2020 04:04:47

---

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.078	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.188	04:00
Breath Volume	3.023	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:02
Subject Sample # 2	0.188	04:03
Breath Volume	2.964	04:03
Air Blank	0.000	04:03
Control Test	0.077	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

---

Data Download Date/Time: 02/07/2020 19:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAWFORD,DAVID,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: WHOBREY,SPENCER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: UMANA-PRUDENCIO

DATE/TIME OF TEST: 02/21/2020 00:31:20

---

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:30	
Control Test	0.069	00:30	Control Outside Tolerance
Air Blank	0.000	00:31	

---

Data Download Date/Time: 02/26/2020 12:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAWFORD,DAVID,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: WHOBREY,SPENCER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: UMANA-PRUDENCIO

DATE/TIME OF TEST: 02/21/2020 00:35:15

---

Diagnostics Check	OK	00:33	
Air Blank	0.000	00:34	
Control Test	0.095	00:34	Interferent Detect
Air Blank	0.000	00:35	

---

Data Download Date/Time: 02/26/2020 12:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAWFORD,DAVID,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: WHOBREY,SPENCER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: UMANA-PRUDENCIO

DATE/TIME OF TEST: 02/21/2020 00:37:44

---

Diagnostics Check	OK	00:36	
Air Blank		00:37	Ambient Fail
Air Blank	0.000	00:37	

---

Data Download Date/Time: 02/26/2020 12:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAWFORD,DAVID,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: WHOBREY,SPENCER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: UMANA-PRUDENCIO

DATE/TIME OF TEST: 02/21/2020 00:51:06

---

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:47	
Control Test	0.077	00:47	
Air Blank	0.000	00:48	
Subject Sample	# 1	00:48	Interferent Detect
Breath Volume	1.929	00:48	
Air Blank	0.000	00:49	
Control Test	0.079	00:50	
Air Blank		00:50	Ambient Fail
Air Blank	0.000	00:51	

---

Data Download Date/Time: 02/26/2020 12:53



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NABULSI,T,  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: NABULSI,T  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 754250  
DRY GAS EXPIRY DATE: 04/03/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: SCHIMMER

DATE/TIME OF TEST: 02/28/2020 22:56:51

---

Diagnostics Check		OK	22:43	
Air Blank		0.000	22:44	
Control Test		0.079	22:44	
Air Blank		0.000	22:45	
Subject Sample	# 1	0.207	22:48	Volume Not Met
Breath Volume		1.074	22:48	
Air Blank		0.000	22:49	
Air Blank		0.000	22:50	
Subject Sample	# 2	0.219	22:52	
Breath Volume		1.882	22:52	
Air Blank		0.000	22:53	
Air Blank		0.000	22:54	
Subject Sample	# 3	0.226	22:55	
Breath Volume		1.574	22:55	
Air Blank		0.000	22:55	
Control Test		0.076	22:56	
Air Blank		0.000	22:56	
Diagnostics Check		OK	22:56	

---

Data Download Date/Time: 03/19/2020 13:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELDON,ERIC,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SHELDON,ERIC  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 754250  
DRY GAS EXPIRY DATE: 04/03/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: DOUGLASS

DATE/TIME OF TEST: 02/29/2020 01:54:04

---

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.079	01:52	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:53	
Control Test	0.079	01:53	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:54	

---

Data Download Date/Time: 03/19/2020 13:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STINSON,ERIC,S  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: STINSON,ERIC  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 754250  
DRY GAS EXPIRY DATE: 04/03/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: BATUNGBAKAL

DATE/TIME OF TEST: 02/29/2020 03:29:47

---

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:17	
Control Test	0.079	03:17	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:21	No Sample Provided
Breath Volume	0.000	03:21	
Air Blank	0.000	03:21	
Air Blank	0.000	03:23	
Subject Sample	# 2	0.149	03:24
Breath Volume	1.554	03:24	
Air Blank	0.000	03:25	
Air Blank	0.000	03:27	
Subject Sample	# 3	0.140	03:28
Breath Volume	2.714	03:28	
Air Blank	0.000	03:28	
Control Test	0.077	03:29	
Air Blank	0.000	03:29	
Diagnostics Check	OK	03:29	

---

Data Download Date/Time: 03/19/2020 13:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELDON,ERIC,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: COSTANTINI,TONY  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 754250  
DRY GAS EXPIRY DATE: 04/03/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: BOYKIN

DATE/TIME OF TEST: 02/29/2020 04:53:20

---

Diagnostics Check	OK	04:46
Air Blank	0.000	04:47
Control Test	0.079	04:47
Air Blank	0.000	04:48
Subject Sample # 1	0.123	04:48
Breath Volume	2.933	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.112	04:51
Breath Volume	2.523	04:51
Air Blank	0.000	04:52
Control Test	0.079	04:52
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

---

Data Download Date/Time: 03/19/2020 13:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMALL,BRANNON,A  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SMALL,BRANNON  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: SUCHIL

DATE/TIME OF TEST: 02/29/2020 04:40:54

---

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.079	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.104	04:35
Breath Volume	1.632	04:35
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.114	04:39
Breath Volume	2.085	04:39
Air Blank	0.000	04:39
Control Test	0.079	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

---

Data Download Date/Time: 03/19/2020 13:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NABULSI,T,  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: NABULSI,T  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 754250  
DRY GAS EXPIRY DATE: 04/03/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: ALICEA

DATE/TIME OF TEST: 02/29/2020 19:32:25

---

Diagnostics Check	OK	19:25
Air Blank	0.000	19:25
Control Test	0.079	19:26
Air Blank	0.000	19:26
Subject Sample # 1	0.227	19:27
Breath Volume	1.785	19:27
Air Blank	0.000	19:28
Air Blank	0.000	19:30
Subject Sample # 2	0.209	19:30
Breath Volume	1.398	19:30
Air Blank	0.000	19:31
Control Test	0.078	19:31
Air Blank	0.000	19:32
Diagnostics Check	OK	19:32

---

Data Download Date/Time: 03/19/2020 13:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMALL,BRANNON,A  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: ELSASSER,ETHAN  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 02/29/2020 20:18:30

---

Diagnostics Check	OK	20:07	
Air Blank	0.000	20:08	
Control Test	0.078	20:08	
Air Blank	0.000	20:08	
Subject Sample	# 1	20:12	No Sample Provided
Breath Volume	0.000	20:12	
Air Blank	0.000	20:12	
Air Blank	0.000	20:14	
Subject Sample	# 2	20:17	Subject Test Refused
Breath Volume	0.000	20:17	
Air Blank	0.000	20:17	
Control Test	0.079	20:17	
Air Blank	0.000	20:18	
Diagnostics Check	OK	20:18	

---

Data Download Date/Time: 03/19/2020 13:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 02/01/2020 03:10:55

---

Diagnostics Check		OK	03:01
Air Blank		0.000	03:02
Control Test		0.081	03:02
Air Blank		0.000	03:03
Subject Sample	# 1	0.129	03:04
Breath Volume		1.289	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.120	03:09
Breath Volume		1.167	03:09
Air Blank		0.000	03:09
Control Test		0.083	03:10
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

---

Data Download Date/Time: 02/05/2020 18:46



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURKE,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 02/01/2020 04:05:41

---

Diagnostics Check		OK	03:59
Air Blank		0.000	03:59
Control Test		0.081	04:00
Air Blank		0.000	04:00
Subject Sample	# 1	0.160	04:01
Breath Volume		2.214	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:03
Subject Sample	# 2	0.161	04:04
Breath Volume		2.257	04:04
Air Blank		0.000	04:04
Control Test		0.081	04:05
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

---

Data Download Date/Time: 02/05/2020 18:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: ALTAF

DATE/TIME OF TEST: 02/01/2020 04:57:42

---

Diagnostics Check		OK	04:51
Air Blank		0.000	04:51
Control Test		0.081	04:51
Air Blank		0.000	04:52
Subject Sample	# 1	0.138	04:53
Breath Volume		1.437	04:53
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.135	04:56
Breath Volume		1.371	04:56
Air Blank		0.000	04:56
Control Test		0.081	04:57
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

---

Data Download Date/Time: 02/05/2020 18:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: WARPEHA

DATE/TIME OF TEST: 02/01/2020 20:28:50

---

Diagnostics Check	OK	20:22
Air Blank	0.000	20:23
Control Test	0.080	20:23
Air Blank	0.000	20:24
Subject Sample # 1	0.247	20:24
Breath Volume	1.417	20:24
Air Blank	0.000	20:25
Air Blank	0.000	20:26
Subject Sample # 2	0.236	20:27
Breath Volume	1.289	20:27
Air Blank	0.000	20:27
Control Test	0.081	20:28
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

---

Data Download Date/Time: 02/05/2020 18:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ANDREWS,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: MARIE

DATE/TIME OF TEST: 02/02/2020 02:09:53

---

Diagnostics Check		OK	02:02
Air Blank		0.000	02:03
Control Test		0.081	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.158	02:05
Breath Volume		1.648	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:07
Subject Sample	# 2	0.160	02:08
Breath Volume		2.433	02:08
Air Blank		0.000	02:08
Control Test		0.081	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

---

Data Download Date/Time: 02/05/2020 18:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: THORNE,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: MUNIZ

DATE/TIME OF TEST: 02/02/2020 03:05:17

---

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.081	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.103	03:00
Breath Volume	3.765	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.101	03:03
Breath Volume	3.781	03:03
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

---

Data Download Date/Time: 02/05/2020 18:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILCHER,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: JAQUEZ

DATE/TIME OF TEST: 02/02/2020 03:52:30

---

Diagnostics Check		OK	03:37	
Air Blank		0.000	03:37	
Control Test		0.081	03:38	
Air Blank		0.000	03:38	
Subject Sample	# 1	0.135	03:41	Volume Not Met
Breath Volume		0.558	03:41	
Air Blank		0.000	03:42	
Air Blank		0.000	03:44	
Subject Sample	# 2	0.165	03:45	
Breath Volume		1.640	03:45	
Air Blank		0.000	03:45	
Air Blank		0.000	03:47	
Subject Sample	# 3	0.142	03:50	Volume Not Met
Breath Volume		0.175	03:50	
Air Blank		0.000	03:51	
Control Test		0.083	03:51	
Air Blank		0.000	03:52	
Diagnostics Check		OK	03:52	

---

Data Download Date/Time: 02/05/2020 18:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILCHER,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: JAQUEZ

DATE/TIME OF TEST: 02/02/2020 04:10:45

---

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.080	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.166	04:06
Breath Volume	1.746	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.159	04:09
Breath Volume	2.542	04:09
Air Blank	0.000	04:09
Control Test	0.080	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

---

Data Download Date/Time: 02/05/2020 18:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VANTREESE,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: READ JR

DATE/TIME OF TEST: 02/02/2020 22:00:48

---

Diagnostics Check	OK	21:54
Air Blank	0.000	21:54
Control Test	0.081	21:55
Air Blank	0.000	21:55
Subject Sample # 1	0.218	21:56
Breath Volume	2.859	21:56
Air Blank	0.000	21:57
Air Blank	0.000	21:58
Subject Sample # 2	0.221	21:59
Breath Volume	2.703	21:59
Air Blank	0.000	21:59
Control Test	0.081	22:00
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

---

Data Download Date/Time: 02/05/2020 18:47



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: EWING

DATE/TIME OF TEST: 02/05/2020 00:27:29

---

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.081	00:22
Air Blank	0.000	00:22
Subject Sample # 1	0.115	00:22
Breath Volume	1.746	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.116	00:25
Breath Volume	1.816	00:25
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

---

Data Download Date/Time: 02/05/2020 18:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RENDON,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: WILLIAMS JR

DATE/TIME OF TEST: 02/05/2020 03:28:55

---

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.000	03:24
Breath Volume	1.640	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.000	03:27
Breath Volume	2.421	03:27
Air Blank	0.000	03:28
Control Test	0.081	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

---

Data Download Date/Time: 02/05/2020 18:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR,BARBARA,H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SCHAEFER,J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 05:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:31

SUBJECT'S LAST NAME: PEREZ-HERNANDEZ

DATE/TIME OF TEST: 02/05/2020 07:03:41

---

Diagnostics Check	OK	06:56
Air Blank	0.000	06:57
Control Test	0.081	06:57
Air Blank	0.000	06:58
Subject Sample # 1	0.201	06:59
Breath Volume	1.199	06:59
Air Blank	0.000	06:59
Air Blank	0.000	07:01
Subject Sample # 2	0.217	07:02
Breath Volume	2.062	07:02
Air Blank	0.000	07:02
Control Test	0.080	07:03
Air Blank	0.000	07:03
Diagnostics Check	OK	07:03

---

Data Download Date/Time: 02/05/2020 18:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FISHER,T  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: SCHULZE

DATE/TIME OF TEST: 02/05/2020 23:30:03

---

Diagnostics Check		OK	23:23
Air Blank		0.000	23:23
Control Test		0.082	23:24
Air Blank		0.000	23:24
Subject Sample	# 1	0.207	23:24
Breath Volume		1.722	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:27
Subject Sample	# 2	0.202	23:28
Breath Volume		1.605	23:28
Air Blank		0.000	23:29
Control Test		0.081	23:29
Air Blank		0.000	23:29
Diagnostics Check		OK	23:30

---

Data Download Date/Time: 03/09/2020 18:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: SOARES

DATE/TIME OF TEST: 02/07/2020 03:34:46

---

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.081	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.069	03:30
Breath Volume	1.925	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.070	03:33
Breath Volume	2.636	03:33
Air Blank	0.000	03:33
Control Test	0.081	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

---

Data Download Date/Time: 03/09/2020 18:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LYON,M  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: STERMAN

DATE/TIME OF TEST: 02/08/2020 01:32:41

---

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.082	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.000	01:28
Breath Volume	4.000	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.000	01:31
Breath Volume	4.074	01:31
Air Blank	0.000	01:31
Control Test	0.082	01:32
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

---

Data Download Date/Time: 03/09/2020 18:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 02/08/2020 02:29:55

---

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.082	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.236	02:25
Breath Volume	3.109	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.238	02:28
Breath Volume	3.332	02:28
Air Blank	0.000	02:28
Control Test	0.081	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

---

Data Download Date/Time: 03/09/2020 18:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VANTREESE,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: BANDA

DATE/TIME OF TEST: 02/08/2020 23:34:12

---

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.081	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.034	23:29
Breath Volume	2.847	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.034	23:32
Breath Volume	2.585	23:32
Air Blank	0.000	23:33
Control Test	0.081	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

---

Data Download Date/Time: 03/09/2020 18:27



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: EDER

DATE/TIME OF TEST: 02/09/2020 02:36:21

---

Diagnostics Check		OK	02:24
Air Blank		0.000	02:24
Control Test		0.081	02:25
Air Blank		0.000	02:25
Subject Sample	# 1	0.181	02:28
Breath Volume		1.753	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.147	02:31
Breath Volume		1.554	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 3	0.153	02:34
Breath Volume		2.230	02:34
Air Blank		0.000	02:35
Control Test		0.079	02:35
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

---

Data Download Date/Time: 03/09/2020 18:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TIBURCIO,I  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: KURIAN

DATE/TIME OF TEST: 02/09/2020 23:03:12

---

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.081	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.095	22:58
Breath Volume	1.324	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.092	23:01
Breath Volume	1.375	23:01
Air Blank	0.000	23:02
Control Test	0.081	23:02
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

---

Data Download Date/Time: 03/09/2020 18:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RENDON,A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: GARCIA GARCIA

DATE/TIME OF TEST: 02/10/2020 03:32:58

---

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.081	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.134	03:28
Breath Volume	2.171	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.126	03:31
Breath Volume	2.605	03:31
Air Blank	0.000	03:32
Control Test	0.081	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

---

Data Download Date/Time: 03/09/2020 18:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FISHER,T  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: HUGHEY

DATE/TIME OF TEST: 02/11/2020 04:10:19

---

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.082	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.137	04:05
Breath Volume	1.929	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.134	04:08
Breath Volume	2.519	04:08
Air Blank	0.000	04:09
Control Test	0.081	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

---

Data Download Date/Time: 03/09/2020 18:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURKE,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: SCOTCH

DATE/TIME OF TEST: 02/11/2020 22:22:10

---

Diagnostics Check	OK	22:14
Air Blank	0.000	22:14
Control Test	0.081	22:15
Air Blank	0.000	22:15
Subject Sample # 1	0.187	22:16
Breath Volume	1.945	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.188	22:20
Breath Volume	2.351	22:20
Air Blank	0.000	22:21
Control Test	0.081	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

---

Data Download Date/Time: 03/09/2020 18:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DEJESUS,V  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: MAUSER

DATE/TIME OF TEST: 02/12/2020 00:20:47

---

Diagnostics Check	OK	00:14
Air Blank	0.000	00:14
Control Test	0.081	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.175	00:16
Breath Volume	2.058	00:16
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.176	00:19
Breath Volume	1.476	00:19
Air Blank	0.000	00:19
Control Test	0.081	00:20
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

---

Data Download Date/Time: 03/09/2020 18:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FISHER,T  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: RIGSBY

DATE/TIME OF TEST: 02/12/2020 03:39:33

---

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.081	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.234	03:35
Breath Volume		1.605	03:35
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.254	03:37
Breath Volume		2.019	03:37
Air Blank		0.000	03:38
Control Test		0.081	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

---

Data Download Date/Time: 03/09/2020 18:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER,MELONIE,E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CHU,A  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 08:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 02/12/2020 10:12:40

---

Diagnostics Check	OK	10:08	
Air Blank	0.000	10:08	
Control Test	0.081	10:08	
Air Blank	0.000	10:09	
Subject Sample	# 1	10:11	Subject Test Refused
Breath Volume	0.015	10:11	
Air Blank	0.000	10:11	
Control Test	0.082	10:12	
Air Blank	0.000	10:12	
Diagnostics Check	OK	10:12	

---

Data Download Date/Time: 03/09/2020 18:28



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RENDON,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: PEREZ-GARCIA

DATE/TIME OF TEST: 02/13/2020 22:52:21

---

Diagnostics Check		OK	22:40	
Air Blank		0.000	22:41	
Control Test		0.081	22:41	
Air Blank		0.000	22:42	
Subject Sample	# 1	0.135	22:45	Volume Not Met
Breath Volume		0.191	22:45	
Air Blank		0.000	22:45	
Air Blank		0.000	22:47	
Subject Sample	# 2	0.132	22:50	Volume Not Met
Breath Volume		0.101	22:50	
Air Blank		0.000	22:51	
Control Test		0.081	22:51	
Air Blank		0.000	22:52	
Diagnostics Check		OK	22:52	

---

Data Download Date/Time: 03/09/2020 18:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: BERGFELD

DATE/TIME OF TEST: 02/14/2020 03:43:41

---

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.081	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.181	03:39
Breath Volume	2.859	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.183	03:42
Breath Volume	2.933	03:42
Air Blank	0.000	03:42
Control Test	0.080	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

---

Data Download Date/Time: 03/09/2020 18:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PASS,J  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: BUI  
ARREST TIME: 21:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: BALL

DATE/TIME OF TEST: 02/15/2020 22:23:33

---

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.082	22:17
Air Blank	0.000	22:17
Subject Sample # 1	0.121	22:18
Breath Volume	2.175	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.119	22:22
Breath Volume	2.691	22:22
Air Blank	0.000	22:22
Control Test	0.082	22:22
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

---

Data Download Date/Time: 03/09/2020 18:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: TAGLIAFERRO NOALE

DATE/TIME OF TEST: 02/16/2020 02:35:03

---

Diagnostics Check		OK	02:27
Air Blank		0.000	02:28
Control Test		0.081	02:28
Air Blank		0.000	02:28
Subject Sample	# 1	0.202	02:29
Breath Volume		2.046	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.211	02:33
Breath Volume		1.687	02:33
Air Blank		0.000	02:34
Control Test		0.080	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:35

---

Data Download Date/Time: 03/09/2020 18:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: REID,W  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 02/16/2020 05:27:29

---

Diagnostics Check		OK	05:20
Air Blank		0.000	05:21
Control Test		0.080	05:21
Air Blank		0.000	05:22
Subject Sample	# 1	0.163	05:22
Breath Volume		2.007	05:22
Air Blank		0.000	05:23
Air Blank		0.000	05:25
Subject Sample	# 2	0.159	05:25
Breath Volume		1.746	05:25
Air Blank		0.000	05:26
Control Test		0.080	05:26
Air Blank		0.000	05:27
Diagnostics Check		OK	05:27

---

Data Download Date/Time: 03/09/2020 18:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RIVERA,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: TANG

DATE/TIME OF TEST: 02/17/2020 02:59:30

---

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.081	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.081	02:54
Breath Volume	1.421	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.082	02:57
Breath Volume	1.281	02:57
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

---

Data Download Date/Time: 03/09/2020 18:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ANDREWS,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 02/17/2020 03:26:15

---

Diagnostics Check		OK	03:19
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.215	03:21
Breath Volume		2.605	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.221	03:24
Breath Volume		2.726	03:24
Air Blank		0.000	03:25
Control Test		0.080	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

---

Data Download Date/Time: 03/09/2020 18:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FERGUSON,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: MOTT

DATE/TIME OF TEST: 02/17/2020 05:51:35

---

Diagnostics Check		OK	05:44
Air Blank		0.000	05:45
Control Test		0.080	05:45
Air Blank		0.000	05:46
Subject Sample	# 1	0.266	05:47
Breath Volume		2.007	05:47
Air Blank		0.000	05:47
Air Blank		0.000	05:49
Subject Sample	# 2	0.265	05:50
Breath Volume		3.042	05:50
Air Blank		0.000	05:50
Control Test		0.080	05:50
Air Blank		0.000	05:51
Diagnostics Check		OK	05:51

---

Data Download Date/Time: 03/09/2020 18:28



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAAR,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: ESCOBAR

DATE/TIME OF TEST: 02/18/2020 04:42:59

---

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:32	
Control Test	0.081	04:32	
Air Blank	0.000	04:32	
Subject Sample	# 1 0.000	04:36	Volume Not Met
Breath Volume	0.156	04:36	
Air Blank	0.000	04:36	
Air Blank	0.000	04:38	
Subject Sample	# 2	04:41	No Sample Provided
Breath Volume	0.000	04:41	
Air Blank	0.000	04:42	
Control Test	0.079	04:42	
Air Blank	0.000	04:42	
Diagnostics Check	OK	04:42	

---

Data Download Date/Time: 03/09/2020 18:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: URSO

DATE/TIME OF TEST: 02/20/2020 02:13:47

---

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.080	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.112	02:08
Breath Volume	1.871	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.113	02:12
Breath Volume	1.757	02:12
Air Blank	0.000	02:12
Control Test	0.081	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

---

Data Download Date/Time: 03/09/2020 18:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LYON,M  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 02/21/2020 03:29:23

---

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.081	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.196	03:24
Breath Volume	1.128	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.214	03:27
Breath Volume	3.316	03:27
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

---

Data Download Date/Time: 03/09/2020 18:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ANDREWS,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: LISS

DATE/TIME OF TEST: 02/21/2020 23:05:41

---

Diagnostics Check	OK	22:59
Air Blank	0.000	23:00
Control Test	0.081	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.110	23:01
Breath Volume	1.246	23:01
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.118	23:04
Breath Volume	1.917	23:04
Air Blank	0.000	23:04
Control Test	0.081	23:05
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

---

Data Download Date/Time: 03/09/2020 18:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SULAIMAN,R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: KLECKLEY

DATE/TIME OF TEST: 02/22/2020 00:38:42

---

Diagnostics Check	OK	00:29
Air Blank	0.000	00:29
Control Test	0.081	00:30
Air Blank	0.000	00:30
Subject Sample # 1	0.172	00:31
Breath Volume	1.726	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.173	00:37
Breath Volume	1.925	00:37
Air Blank	0.000	00:37
Control Test	0.081	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

---

Data Download Date/Time: 03/09/2020 18:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ANDREWS,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 02/22/2020 03:16:30

---

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.081	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.000	03:12
Breath Volume	3.160	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.000	03:15
Breath Volume	3.542	03:15
Air Blank	0.000	03:15
Control Test	0.081	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

---

Data Download Date/Time: 03/09/2020 18:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: TERANCIEL

DATE/TIME OF TEST: 02/22/2020 03:35:29

---

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:24	
Control Test		0.081	03:24	
Air Blank		0.000	03:25	
Subject Sample	# 1	0.118	03:28	Volume Not Met
Breath Volume		0.085	03:28	
Air Blank		0.000	03:28	
Air Blank		0.000	03:30	
Subject Sample	# 2	0.106	03:33	Volume Not Met
Breath Volume		0.605	03:33	
Air Blank		0.000	03:34	
Control Test		0.081	03:34	
Air Blank		0.000	03:35	
Diagnostics Check		OK	03:35	

---

Data Download Date/Time: 03/09/2020 18:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 02/22/2020 03:57:19

---

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.080	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.075	03:52
Breath Volume		1.496	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.073	03:55
Breath Volume		1.644	03:55
Air Blank		0.000	03:56
Control Test		0.081	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

---

Data Download Date/Time: 03/09/2020 18:28



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: TERANCIEL

DATE/TIME OF TEST: 02/22/2020 04:49:42

---

Diagnostics Check		OK	04:36	
Air Blank		0.000	04:36	
Control Test		0.081	04:37	
Air Blank		0.000	04:37	
Subject Sample	# 1	0.118	04:40	Volume Not Met
Breath Volume		0.406	04:40	
Air Blank		0.000	04:41	
Air Blank		0.000	04:43	
Subject Sample	# 2	0.156	04:44	
Breath Volume		1.691	04:44	
Air Blank		0.000	04:45	
Air Blank		0.000	04:46	
Subject Sample	# 3	0.147	04:48	
Breath Volume		1.312	04:48	
Air Blank		0.000	04:48	
Control Test		0.080	04:49	
Air Blank		0.000	04:49	
Diagnostics Check		OK	04:49	

---

Data Download Date/Time: 03/09/2020 18:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HALL,L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 02/22/2020 05:20:39

---

Diagnostics Check	OK	05:13
Air Blank	0.000	05:14
Control Test	0.080	05:14
Air Blank	0.000	05:14
Subject Sample # 1	0.110	05:15
Breath Volume	2.351	05:15
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.112	05:19
Breath Volume	2.851	05:19
Air Blank	0.000	05:19
Control Test	0.080	05:20
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

---

Data Download Date/Time: 03/09/2020 18:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ARCHAMBEAU,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/22/2020 23:01:27

---

Diagnostics Check	OK	22:52
Air Blank	0.000	22:53
Control Test	0.081	22:53
Air Blank	0.000	22:53
Subject Sample # 1	0.214	22:56
Breath Volume	1.714	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.213	22:59
Breath Volume	2.625	22:59
Air Blank	0.000	23:00
Control Test	0.081	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

---

Data Download Date/Time: 03/09/2020 18:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: NOPAL PEREZ

DATE/TIME OF TEST: 02/22/2020 23:46:34

---

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.081	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.136	23:42
Breath Volume	1.480	23:42
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.137	23:45
Breath Volume	1.910	23:45
Air Blank	0.000	23:45
Control Test	0.081	23:45
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

---

Data Download Date/Time: 03/09/2020 18:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 02/23/2020 01:47:27

---

Diagnostics Check		OK	01:34	
Air Blank		0.000	01:35	
Control Test		0.081	01:35	
Air Blank		0.000	01:35	
Subject Sample	# 1	0.149	01:39	Volume Not Met
Breath Volume		0.109	01:39	
Air Blank		0.000	01:39	
Air Blank		0.000	01:41	
Subject Sample	# 2	0.150	01:42	
Breath Volume		1.371	01:42	
Air Blank		0.000	01:43	
Air Blank		0.000	01:45	
Subject Sample	# 3	0.149	01:45	
Breath Volume		1.175	01:45	
Air Blank		0.000	01:46	
Control Test		0.081	01:46	
Air Blank		0.000	01:47	
Diagnostics Check		OK	01:47	

---

Data Download Date/Time: 03/09/2020 18:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: COSTA,A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: RODRIGUEZ JR

DATE/TIME OF TEST: 02/23/2020 05:21:09

---

Diagnostics Check		OK	05:14
Air Blank		0.000	05:15
Control Test		0.081	05:15
Air Blank		0.000	05:15
Subject Sample	# 1	0.206	05:16
Breath Volume		2.015	05:16
Air Blank		0.000	05:17
Air Blank		0.000	05:19
Subject Sample	# 2	0.196	05:19
Breath Volume		1.906	05:19
Air Blank		0.000	05:20
Control Test		0.081	05:20
Air Blank		0.000	05:21
Diagnostics Check		OK	05:21

---

Data Download Date/Time: 03/09/2020 18:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RENDON,R  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: GONZALEZ MORALES

DATE/TIME OF TEST: 02/24/2020 00:35:41

---

Diagnostics Check	OK	00:29
Air Blank	0.000	00:29
Control Test	0.080	00:29
Air Blank	0.000	00:30
Subject Sample # 1	0.206	00:31
Breath Volume	3.742	00:31
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.219	00:34
Breath Volume	3.152	00:34
Air Blank	0.000	00:34
Control Test	0.080	00:35
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

---

Data Download Date/Time: 03/09/2020 18:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HARRIS,S  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/25/2020 02:15:25

---

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.081	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.000	02:11
Breath Volume	2.187	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.000	02:14
Breath Volume	1.984	02:14
Air Blank	0.000	02:14
Control Test	0.081	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

---

Data Download Date/Time: 03/09/2020 18:28



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURKE,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: TABOADA

DATE/TIME OF TEST: 02/26/2020 05:20:46

---

Diagnostics Check	OK	05:13
Air Blank	0.000	05:14
Control Test	0.081	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.154	05:16
Breath Volume	1.472	05:16
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.158	05:19
Breath Volume	1.562	05:19
Air Blank	0.000	05:19
Control Test	0.080	05:20
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

---

Data Download Date/Time: 03/09/2020 18:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DRISCOLL,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 02/28/2020 02:22:11

---

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.081	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.000	02:17
Breath Volume	1.269	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.000	02:20
Breath Volume	2.828	02:20
Air Blank	0.000	02:21
Control Test	0.081	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

---

Data Download Date/Time: 03/09/2020 18:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SULAIMAN,R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/28/2020 03:35:51

---

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.081	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.130	03:31
Breath Volume	1.632	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:33
Subject Sample # 2	0.133	03:34
Breath Volume	1.914	03:34
Air Blank	0.000	03:34
Control Test	0.081	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

---

Data Download Date/Time: 03/09/2020 18:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FISHER,T  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 02/29/2020 00:39:01

---

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.081	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.141	00:34
Breath Volume	3.082	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.147	00:37
Breath Volume	2.218	00:37
Air Blank	0.000	00:38
Control Test	0.081	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:39

---

Data Download Date/Time: 03/09/2020 18:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: BERNARD JR

DATE/TIME OF TEST: 02/29/2020 05:05:16

---

Diagnostics Check		OK	04:57
Air Blank		0.000	04:58
Control Test		0.081	04:58
Air Blank		0.000	04:59
Subject Sample	# 1	0.054	05:00
Breath Volume		1.226	05:00
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.058	05:03
Breath Volume		1.398	05:03
Air Blank		0.000	05:04
Control Test		0.082	05:04
Air Blank		0.000	05:05
Diagnostics Check		OK	05:05

---

Data Download Date/Time: 03/09/2020 18:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SULLIVAN,T  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 02/29/2020 05:37:54

---

Diagnostics Check	OK	05:30
Air Blank	0.000	05:31
Control Test	0.081	05:31
Air Blank	0.000	05:31
Subject Sample # 1	0.119	05:33
Breath Volume	2.394	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:35
Subject Sample # 2	0.118	05:36
Breath Volume	1.785	05:36
Air Blank	0.000	05:36
Control Test	0.079	05:37
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

---

Data Download Date/Time: 03/09/2020 18:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: KENT,N  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 01/20/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:34

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 02/03/2020 18:00:41

---

Diagnostics Check	OK	17:54
Air Blank	0.000	17:55
Control Test	0.082	17:55
Air Blank	0.000	17:55
Subject Sample # 1	0.000	17:56
Breath Volume	3.769	17:56
Air Blank	0.000	17:56
Air Blank	0.000	17:58
Subject Sample # 2	0.000	17:59
Breath Volume	1.613	17:59
Air Blank	0.000	17:59
Control Test	0.082	18:00
Air Blank	0.000	18:00
Diagnostics Check	OK	18:00

---

Data Download Date/Time: 02/17/2020 12:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KJEKSTAD,ERIC,C  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: KJEKSTAD,ERIC  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 01/20/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: DUDOCKHIN

DATE/TIME OF TEST: 02/06/2020 20:27:37

---

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.081	20:22
Air Blank	0.000	20:22
Subject Sample # 1	0.026	20:23
Breath Volume	2.285	20:23
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.022	20:26
Breath Volume	1.722	20:26
Air Blank	0.000	20:26
Control Test	0.079	20:27
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

---

Data Download Date/Time: 02/17/2020 12:53



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RITCHIE,FRANCIS,V  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: RITCHIE,FRANCIS  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 01/20/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 02/15/2020 22:09:05

---

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.081	22:02
Air Blank	0.000	22:02
Subject Sample # 1	0.116	22:04
Breath Volume	1.472	22:04
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.112	22:07
Breath Volume	1.656	22:07
Air Blank	0.000	22:08
Control Test	0.080	22:08
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

---

Data Download Date/Time: 02/17/2020 12:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CROFT,BRETT,  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: CROFT,BRETT  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 01/20/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: SIEVERS

DATE/TIME OF TEST: 02/17/2020 02:59:23

---

Diagnostics Check		OK	02:52
Air Blank		0.000	02:52
Control Test		0.082	02:52
Air Blank		0.000	02:53
Subject Sample	# 1	0.106	02:54
Breath Volume		1.136	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.103	02:57
Breath Volume		1.730	02:57
Air Blank		0.000	02:58
Control Test		0.080	02:58
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

---

Data Download Date/Time: 02/17/2020 12:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORTIZ,BILLY,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: ORTIZ,BILLY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 01/20/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: WALL

DATE/TIME OF TEST: 02/17/2020 03:43:36

---

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:41	
Control Test	0.081	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:42	
Control Test	0.081	03:43	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

---

Data Download Date/Time: 02/17/2020 12:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EASTTY,MARK,L  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CRUM,ZACK  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG831804  
DRY GAS EXPIRY DATE: 11/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: NARDONTONIA

DATE/TIME OF TEST: 02/22/2020 23:14:44

---

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.080	23:09
Air Blank	0.000	23:09
Subject Sample # 1	0.053	23:10
Breath Volume	3.062	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:12
Subject Sample # 2	0.052	23:13
Breath Volume	3.886	23:13
Air Blank	0.000	23:13
Control Test	0.080	23:14
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

---

Data Download Date/Time: 03/06/2020 21:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EASTTY,MARK,L  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LAMOUR,ZACHARY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG831804  
DRY GAS EXPIRY DATE: 11/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: ALDANA-PENA

DATE/TIME OF TEST: 02/23/2020 01:41:44

---

Diagnostics Check	OK	01:41	
Air Blank		01:41	RFI Detect
Air Blank		01:41	RFI Detect

---

Data Download Date/Time: 03/06/2020 21:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EASTTY,MARK,L  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LAMOUR,ZACHARY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG831804  
DRY GAS EXPIRY DATE: 11/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: ALDANA-PENA

DATE/TIME OF TEST: 02/23/2020 01:43:27

---

Diagnostics Check	OK	01:43	
Air Blank		01:43	RFI Detect
Air Blank		01:43	RFI Detect

---

Data Download Date/Time: 03/06/2020 21:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EASTTY,MARK,L  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LAMOUR,ZACHARY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG831804  
DRY GAS EXPIRY DATE: 11/14/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: ALDANA-PENA

DATE/TIME OF TEST: 02/23/2020 01:45:46

---

Diagnostics Check	OK	01:45	
Air Blank		01:45	RFI Detect
Air Blank		01:45	RFI Detect

---

Data Download Date/Time: 03/06/2020 21:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLAFLIN,DAVID,S  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: TESTER,T  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28818080A1  
DRY GAS EXPIRY DATE: 12/05/2020  
VIOLATION CODE: OTHER  
ARREST TIME: 18:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 02/18/2020 19:38:39

---

Diagnostics Check	OK	19:31
Air Blank	0.000	19:32
Control Test	0.081	19:32
Air Blank	0.000	19:33
Subject Sample # 1	0.000	19:33
Breath Volume	2.757	19:33
Air Blank	0.000	19:34
Air Blank	0.000	19:36
Subject Sample # 2	0.000	19:37
Breath Volume	2.128	19:37
Air Blank	0.000	19:37
Control Test	0.080	19:37
Air Blank	0.000	19:38
Diagnostics Check	OK	19:38

---

Data Download Date/Time: 03/04/2020 10:24



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENYON,KYLE,A  
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD  
ARRESTING OFFICER: KENYON,KYLE  
ARRESTING OFFICER AGENCY: Fernandina Beach PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: MCCONATHA

DATE/TIME OF TEST: 02/02/2020 03:22:36

---

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.100	03:17
Breath Volume		1.726	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.091	03:20
Breath Volume		1.730	03:20
Air Blank		0.000	03:21
Control Test		0.078	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

---

Data Download Date/Time: 02/11/2020 11:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENYON,KYLE,A  
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD  
ARRESTING OFFICER: KENYON,KYLE  
ARRESTING OFFICER AGENCY: Fernandina Beach PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 02/03/2020 03:42:38

---

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:37	
Control Test	0.081	03:37	
Air Blank	0.000	03:38	
Subject Sample	# 1 0.135	03:41	
Breath Volume	2.015	03:41	
Air Blank		03:42	Purge Fail
Air Blank	0.000	03:42	

---

Data Download Date/Time: 02/11/2020 11:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENYON,KYLE,A  
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD  
ARRESTING OFFICER: KENYON,KYLE  
ARRESTING OFFICER AGENCY: Fernandina Beach PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 02/03/2020 03:55:40

---

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.078	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.127	03:51
Breath Volume	2.003	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.135	03:54
Breath Volume	1.546	03:54
Air Blank	0.000	03:54
Control Test	0.076	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

---

Data Download Date/Time: 02/11/2020 11:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENYON,KYLE,A  
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD  
ARRESTING OFFICER: KENYON,KYLE  
ARRESTING OFFICER AGENCY: Fernandina Beach PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: FIRSTER

DATE/TIME OF TEST: 02/08/2020 23:04:57

---

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.172	22:59
Breath Volume	1.488	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.168	23:03
Breath Volume	2.007	23:03
Air Blank	0.000	23:04
Control Test	0.078	23:04
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

---

Data Download Date/Time: 02/11/2020 11:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAREY,J,  
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD  
ARRESTING OFFICER: MITCHELL,J  
ARRESTING OFFICER AGENCY: Fernandina Beach PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 02/09/2020 15:30:47

---

Diagnostics Check	OK	15:24
Air Blank	0.000	15:25
Control Test	0.081	15:25
Air Blank	0.000	15:25
Subject Sample # 1	0.000	15:26
Breath Volume	4.500	15:26
Air Blank	0.000	15:26
Air Blank	0.000	15:28
Subject Sample # 2	0.000	15:29
Breath Volume	5.085	15:29
Air Blank	0.000	15:29
Control Test	0.080	15:30
Air Blank	0.000	15:30
Diagnostics Check	OK	15:30

---

Data Download Date/Time: 02/11/2020 11:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENYON,KYLE,A  
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD  
ARRESTING OFFICER: PAETSCH,J  
ARRESTING OFFICER AGENCY: Fernandina Beach PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 02/10/2020 01:41:44

---

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:34	
Control Test	0.080	01:34	
Air Blank	0.000	01:35	
Subject Sample # 1	0.079	01:37	
Breath Volume	1.351	01:37	
Air Blank	0.000	01:38	
Air Blank	0.000	01:39	
Subject Sample # 2	0.073	01:40	
Breath Volume	1.433	01:40	
Air Blank	0.000	01:40	
Control Test	0.074	01:41	Control Outside Tolerance
Air Blank	0.000	01:41	

---

Data Download Date/Time: 02/11/2020 11:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENYON,KYLE,A  
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD  
ARRESTING OFFICER: PAETSCH,JOSH  
ARRESTING OFFICER AGENCY: Fernandina Beach PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 02/10/2020 01:59:22

---

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.079	01:53
Air Blank	0.000	01:54
Subject Sample # 1	0.084	01:54
Breath Volume	1.542	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.082	01:57
Breath Volume	2.394	01:57
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

---

Data Download Date/Time: 02/11/2020 11:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: CLOSIUS,RICHARD  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34917080A4  
DRY GAS EXPIRY DATE: 02/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: RAMENTOL

DATE/TIME OF TEST: 02/01/2020 05:06:19

---

Diagnostics Check		OK	05:00
Air Blank		0.000	05:00
Control Test		0.078	05:01
Air Blank		0.000	05:01
Subject Sample	# 1	0.129	05:01
Breath Volume		1.550	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.126	05:04
Breath Volume		1.242	05:04
Air Blank		0.000	05:05
Control Test		0.078	05:05
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

---

Data Download Date/Time: 02/26/2020 14:47



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: CLOSIUS,RICHARD  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 02/08/2020 04:04:00

---

Diagnostics Check		OK	03:56
Air Blank		0.000	03:56
Control Test		0.080	03:57
Air Blank		0.000	03:57
Subject Sample	# 1	0.048	03:59
Breath Volume		2.085	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:01
Subject Sample	# 2	0.045	04:02
Breath Volume		2.007	04:02
Air Blank		0.000	04:03
Control Test		0.079	04:03
Air Blank		0.000	04:03
Diagnostics Check		OK	04:04

---

Data Download Date/Time: 02/26/2020 14:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: CLOSIUS,RICHARD  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: CRUZ-VIXTHA

DATE/TIME OF TEST: 02/13/2020 03:15:29

---

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:13	
Control Test	0.078	03:13	
Air Blank	0.000	03:13	
Subject Sample	# 1	03:14	Subject Test Refused
Breath Volume	0.000	03:14	
Air Blank	0.000	03:14	
Control Test	0.078	03:14	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:15	

---

Data Download Date/Time: 02/26/2020 14:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: CLOSIUS,RICHARD  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: HOZ

DATE/TIME OF TEST: 02/15/2020 05:24:13

---

Diagnostics Check	OK	05:17
Air Blank	0.000	05:18
Control Test	0.079	05:18
Air Blank	0.000	05:18
Subject Sample # 1	0.012	05:19
Breath Volume	1.742	05:19
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample # 2	0.011	05:22
Breath Volume	2.011	05:22
Air Blank	0.000	05:23
Control Test	0.078	05:23
Air Blank	0.000	05:24
Diagnostics Check	OK	05:24

---

Data Download Date/Time: 02/26/2020 14:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: HAYES,BRITTNEY  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: FUENTES MEJIAS

DATE/TIME OF TEST: 02/20/2020 00:40:16

---

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:37	
Control Test	0.078	00:38	
Air Blank	0.000	00:38	
Subject Sample	# 1	00:38	Subject Test Refused
Breath Volume	0.000	00:38	
Air Blank	0.000	00:39	
Control Test	0.078	00:39	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:40	

---

Data Download Date/Time: 02/26/2020 14:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: QUINTANA,,JORGE  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: TAPIA

DATE/TIME OF TEST: 02/20/2020 03:09:11

---

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.077	03:07	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:08	
Control Test	0.078	03:08	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

---

Data Download Date/Time: 02/26/2020 14:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: VARGAS,DAEL  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 02/22/2020 01:20:16

---

Diagnostics Check	OK	01:13
Air Blank	0.000	01:14
Control Test	0.080	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.156	01:15
Breath Volume	1.195	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.168	01:18
Breath Volume	1.808	01:18
Air Blank	0.000	01:19
Control Test	0.079	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

---

Data Download Date/Time: 02/26/2020 14:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: QUINTANA,,JORGE  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/23/2020 01:18:22

---

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.079	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.125	01:13
Breath Volume		2.460	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.133	01:16
Breath Volume		1.667	01:16
Air Blank		0.000	01:17
Control Test		0.079	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

---

Data Download Date/Time: 02/26/2020 14:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: CLOSIUS,RICHARD  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 02/23/2020 04:15:43

---

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:13	
Control Test	0.080	04:13	
Air Blank	0.000	04:14	
Subject Sample	# 1	04:14	Subject Test Refused
Breath Volume	0.000	04:14	
Air Blank	0.000	04:14	
Control Test	0.079	04:15	
Air Blank	0.000	04:15	
Diagnostics Check	OK	04:15	

---

Data Download Date/Time: 02/26/2020 14:48



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: CLOSIUS,RICHARD  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 02/27/2020 02:46:04

---

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.091	02:41
Breath Volume	1.859	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.094	02:44
Breath Volume	3.820	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:46

---

Data Download Date/Time: 03/25/2020 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: CLOSIUS,RICHARD  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: SENRA

DATE/TIME OF TEST: 02/29/2020 05:33:15

---

Diagnostics Check	OK	05:26
Air Blank	0.000	05:27
Control Test	0.081	05:27
Air Blank	0.000	05:28
Subject Sample # 1	0.059	05:28
Breath Volume	1.886	05:28
Air Blank	0.000	05:29
Air Blank	0.000	05:31
Subject Sample # 2	0.060	05:31
Breath Volume	2.671	05:31
Air Blank	0.000	05:32
Control Test	0.080	05:32
Air Blank	0.000	05:33
Diagnostics Check	OK	05:33

---

Data Download Date/Time: 03/25/2020 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZRAK,JOSEPH,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: AZRAK,JOSEPH  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: BLANDON-PEREZ

DATE/TIME OF TEST: 02/07/2020 04:50:11

---

Diagnostics Check	OK	04:43
Air Blank	0.000	04:44
Control Test	0.080	04:44
Air Blank	0.000	04:45
Subject Sample # 1	0.143	04:45
Breath Volume	2.972	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.140	04:48
Breath Volume	3.011	04:48
Air Blank	0.000	04:49
Control Test	0.079	04:49
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

---

Data Download Date/Time: 02/26/2020 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZRAK,JOSEPH,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: AZRAK,JOSEPH  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: QUINTERO-CASTRO

DATE/TIME OF TEST: 02/08/2020 01:33:59

---

Diagnostics Check	OK	01:27
Air Blank	0.000	01:28
Control Test	0.082	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.142	01:29
Breath Volume	1.261	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:31
Subject Sample # 2	0.144	01:32
Breath Volume	2.355	01:32
Air Blank	0.000	01:33
Control Test	0.081	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

---

Data Download Date/Time: 02/26/2020 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZRAK,JOSEPH,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: BIANCHI,M  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: CACERS-HERNANDEZ

DATE/TIME OF TEST: 02/09/2020 02:56:43

---

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.081	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.156	02:52
Breath Volume	1.492	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.155	02:55
Breath Volume	2.156	02:55
Air Blank	0.000	02:55
Control Test	0.081	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

---

Data Download Date/Time: 02/26/2020 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZRAK,JOSEPH,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: AZRAK,JOSEPH  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: GAYLE

DATE/TIME OF TEST: 02/14/2020 02:58:31

---

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.079	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.153	02:54
Breath Volume	2.339	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.152	02:56
Breath Volume	2.269	02:56
Air Blank	0.000	02:57
Control Test	0.078	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

---

Data Download Date/Time: 02/26/2020 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZRAK,JOSEPH,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: AZRAK,JOSEPH  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: LUANGO

DATE/TIME OF TEST: 02/15/2020 03:38:56

---

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.079	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.174	03:34
Breath Volume	2.750	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:36
Subject Sample # 2	0.185	03:37
Breath Volume	3.085	03:37
Air Blank	0.000	03:37
Control Test	0.080	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

---

Data Download Date/Time: 02/26/2020 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZRAK,JOSEPH,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: AZRAK,JOSEPH  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: WILLS JR

DATE/TIME OF TEST: 02/19/2020 01:11:11

---

Diagnostics Check		OK	01:04
Air Blank		0.000	01:05
Control Test		0.079	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.225	01:06
Breath Volume		1.902	01:06
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.217	01:09
Breath Volume		1.828	01:09
Air Blank		0.000	01:10
Control Test		0.078	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

---

Data Download Date/Time: 02/26/2020 13:24



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZRAK,JOSEPH,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: AZRAK,JOSEPH  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: GALEANO-CASTELLON

DATE/TIME OF TEST: 02/20/2020 03:20:46

---

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.078	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.175	03:16
Breath Volume	1.910	03:16
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.164	03:19
Breath Volume	1.804	03:19
Air Blank	0.000	03:19
Control Test	0.078	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

---

Data Download Date/Time: 02/26/2020 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZRAK,JOSEPH,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: AZRAK,JOSEPH  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: CENTENO

DATE/TIME OF TEST: 02/22/2020 00:57:11

---

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.169	00:52
Breath Volume	2.144	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.162	00:55
Breath Volume	2.281	00:55
Air Blank	0.000	00:56
Control Test	0.081	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

---

Data Download Date/Time: 02/26/2020 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: SIMON,JIM

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 01/27/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 02:41

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/08/2020 04:30:53

---

Diagnostics Check		OK	04:23
Air Blank		0.000	04:23
Control Test		0.080	04:24
Air Blank		0.000	04:24
Subject Sample	# 1	0.169	04:26
Breath Volume		1.304	04:26
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.170	04:29
Breath Volume		1.167	04:29
Air Blank		0.000	04:29
Control Test		0.078	04:30
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

---

Data Download Date/Time: 02/11/2020 10:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUACKENBUSH,BRET,C  
BREATH TEST OPERATOR AGENCY: Lakeland PD  
ARRESTING OFFICER: DOOLIN,EMILY  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 02/12/2020 21:31:26

---

Diagnostics Check	OK	21:24
Air Blank	0.000	21:24
Control Test	0.081	21:25
Air Blank	0.000	21:25
Subject Sample # 1	0.159	21:26
Breath Volume	1.867	21:26
Air Blank	0.000	21:27
Air Blank	0.000	21:29
Subject Sample # 2	0.162	21:29
Breath Volume	1.492	21:29
Air Blank	0.000	21:30
Control Test	0.079	21:30
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

---

Data Download Date/Time: 03/03/2020 09:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 22:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/15/2020 23:56:00

---

Diagnostics Check		OK	23:48
Air Blank		0.000	23:49
Control Test		0.080	23:49
Air Blank		0.000	23:50
Subject Sample	# 1	0.115	23:51
Breath Volume		1.207	23:51
Air Blank		0.000	23:52
Air Blank		0.000	23:53
Subject Sample	# 2	0.133	23:54
Breath Volume		1.714	23:54
Air Blank		0.000	23:55
Control Test		0.079	23:55
Air Blank		0.000	23:55
Diagnostics Check		OK	23:56

---

Data Download Date/Time: 03/03/2020 09:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 00:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: BALA-ORTIZ

DATE/TIME OF TEST: 02/16/2020 01:19:26

---

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:14
Subject Sample # 1	0.297	01:14
Breath Volume	2.964	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.299	01:17
Breath Volume	3.359	01:17
Air Blank	0.000	01:18
Control Test	0.078	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

---

Data Download Date/Time: 03/03/2020 09:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELCHER,ALEX,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BELCHER,ALEX  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 02/17/2020 22:29:05

---

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.082	22:23
Air Blank	0.000	22:23
Subject Sample # 1	0.098	22:24
Breath Volume	1.406	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.101	22:27
Breath Volume	1.777	22:27
Air Blank	0.000	22:28
Control Test	0.080	22:28
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

---

Data Download Date/Time: 03/03/2020 09:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYERS,MICHELLE,H  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: SAYERS,MICHELLE  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MARTINEZ-LOPEZ

DATE/TIME OF TEST: 02/28/2020 03:30:50

---

Diagnostics Check		OK	03:20
Air Blank		0.000	03:20
Control Test		0.081	03:20
Air Blank		0.000	03:21
Subject Sample	# 1	0.253	03:22
Breath Volume		2.386	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.223	03:25
Breath Volume		1.917	03:25
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 3	0.234	03:29
Breath Volume		2.683	03:29
Air Blank		0.000	03:29
Control Test		0.080	03:30
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

---

Data Download Date/Time: 03/31/2020 11:53



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARENT,GARY,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SCHNEIDER,DAVID  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22817080A5  
DRY GAS EXPIRY DATE: 10/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: NAPOLI

DATE/TIME OF TEST: 02/08/2020 02:19:38

---

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.153	02:15
Breath Volume	1.781	02:15
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.158	02:18
Breath Volume	2.703	02:18
Air Blank	0.000	02:18
Control Test	0.080	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

---

Data Download Date/Time: 02/11/2020 14:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRIGLIA,NICHOLAS,S  
BREATH TEST OPERATOR AGENCY: Stuart PD  
ARRESTING OFFICER: BRIGLIA,NICHOLAS  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06717080A7  
DRY GAS EXPIRY DATE: 05/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 16:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: BENDEL

DATE/TIME OF TEST: 02/14/2020 17:23:51

---

Diagnostics Check	OK	17:16	
Air Blank	0.000	17:17	
Control Test	0.079	17:17	
Air Blank	0.000	17:18	
Subject Sample	# 1 0.169	17:19	
Breath Volume	1.101	17:19	
Air Blank	0.000	17:20	
Air Blank	0.000	17:21	
Subject Sample	# 2 0.183	17:22	
Breath Volume	1.117	17:22	
Air Blank	0.000	17:22	
Control Test	0.073	17:23	Control Outside Tolerance
Air Blank	0.000	17:23	

---

Data Download Date/Time: 03/06/2020 18:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRIGLIA,NICHOLAS,S  
BREATH TEST OPERATOR AGENCY: Stuart PD  
ARRESTING OFFICER: BRIGLIA,NICHOLAS  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06717080A7  
DRY GAS EXPIRY DATE: 05/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 14:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: BENDEL

DATE/TIME OF TEST: 02/14/2020 18:03:00

---

Diagnostics Check	OK	17:55
Air Blank	0.000	17:56
Control Test	0.078	17:56
Air Blank	0.000	17:57
Subject Sample # 1	0.172	17:57
Breath Volume	3.140	17:57
Air Blank	0.000	17:58
Air Blank	0.000	18:00
Subject Sample # 2	0.172	18:01
Breath Volume	3.011	18:01
Air Blank	0.000	18:02
Control Test	0.075	18:02
Air Blank	0.000	18:02
Diagnostics Check	OK	18:03

---

Data Download Date/Time: 03/06/2020 18:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: THOMAS,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 01/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: WELLMAKER

DATE/TIME OF TEST: 02/01/2020 04:43:27

---

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:41	
Control Test	0.081	04:41	
Air Blank	0.000	04:41	
Subject Sample	# 1	04:42	Subject Test Refused
Breath Volume	0.000	04:42	
Air Blank	0.000	04:42	
Control Test	0.077	04:42	
Air Blank	0.000	04:43	
Diagnostics Check	OK	04:43	

---

Data Download Date/Time: 02/07/2020 17:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: THOMAS,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 01/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 02/04/2020 01:28:46

---

Diagnostics Check		OK	01:22
Air Blank		0.000	01:22
Control Test		0.080	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.150	01:24
Breath Volume		3.179	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:26
Subject Sample	# 2	0.140	01:27
Breath Volume		2.058	01:27
Air Blank		0.000	01:27
Control Test		0.080	01:28
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

---

Data Download Date/Time: 02/07/2020 17:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: THOMAS,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 01/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: GRIFFITH

DATE/TIME OF TEST: 02/05/2020 00:38:17

---

Diagnostics Check		OK	00:26	
Air Blank		0.000	00:27	
Control Test		0.083	00:27	
Air Blank		0.000	00:27	
Subject Sample	# 1	0.137	00:31	Volume Not Met
Breath Volume		0.132	00:31	
Air Blank		0.000	00:31	
Air Blank		0.000	00:33	
Subject Sample	# 2	0.126	00:36	Volume Not Met
Breath Volume		0.085	00:36	
Air Blank		0.000	00:37	
Control Test		0.080	00:37	
Air Blank		0.000	00:38	
Diagnostics Check		OK	00:38	

---

Data Download Date/Time: 02/07/2020 17:39