

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH,DEVIN,A

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: TAYLOR,T

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 13518080A5

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 17:50

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: HEATH

DATE/TIME OF TEST: 02/03/2020 19:19:06

---

Diagnostics Check	OK	19:10	
Air Blank	0.000	19:10	
Control Test	0.077	19:10	
Air Blank	0.000	19:11	
Subject Sample	# 1	19:14	No Sample Provided
Breath Volume	0.000	19:14	
Air Blank	0.000	19:14	
Air Blank	0.000	19:16	
Subject Sample	# 2	19:17	Subject Test Refused
Breath Volume	0.000	19:17	
Air Blank	0.000	19:18	
Control Test	0.076	19:18	
Air Blank	0.000	19:19	
Diagnostics Check	OK	19:19	

---

Data Download Date/Time: 03/10/2020 17:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CASON,CHRISTOPHER,N

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: WIGGLESWORTH,C

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 02/08/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13518080A5

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 17:35

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: TERHUNE

DATE/TIME OF TEST: 02/08/2020 18:47:26

---

Diagnostics Check	OK	18:43	
Air Blank	0.000	18:43	
Control Test	0.077	18:44	
Air Blank	0.000	18:44	
Subject Sample	# 1	18:46	Subject Test Refused
Breath Volume	0.000	18:46	
Air Blank	0.000	18:46	
Control Test	0.076	18:46	
Air Blank	0.000	18:47	
Diagnostics Check	OK	18:47	

---

Data Download Date/Time: 03/10/2020 17:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH,DEVIN,A

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: ERDMAN,N

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 02/08/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13518080A5

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 19:17

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 02/08/2020 20:03:40

---

Diagnostics Check	OK	20:00	
Air Blank	0.000	20:00	
Control Test	0.076	20:00	
Air Blank	0.000	20:01	
Subject Sample	# 1	20:02	Subject Test Refused
Breath Volume	0.000	20:02	
Air Blank	0.000	20:02	
Control Test	0.077	20:03	
Air Blank	0.000	20:03	
Diagnostics Check	OK	20:03	

---

Data Download Date/Time: 03/10/2020 17:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CASON,CHRISTOPHER,N

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: WIGGLSEWORTH,C

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13518080A5

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 00:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/14/2020 01:23:10

---

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:20	
Control Test	0.077	01:21	
Air Blank	0.000	01:21	
Subject Sample	# 1	01:21	Subject Test Refused
Breath Volume	0.000	01:21	
Air Blank	0.000	01:22	
Control Test	0.077	01:22	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

---

Data Download Date/Time: 03/10/2020 17:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVENPORT, JOSHUA, J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: CLOUD, TODD  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: SCHILLING

DATE/TIME OF TEST: 02/14/2020 19:00:31

---

Diagnostics Check	OK	18:54
Air Blank	0.000	18:54
Control Test	0.077	18:55
Air Blank	0.000	18:55
Subject Sample # 1	0.193	18:56
Breath Volume	2.078	18:56
Air Blank	0.000	18:56
Air Blank	0.000	18:58
Subject Sample # 2	0.174	18:59
Breath Volume	2.234	18:59
Air Blank	0.000	18:59
Control Test	0.077	18:59
Air Blank	0.000	19:00
Diagnostics Check	OK	19:00

---

Data Download Date/Time: 03/10/2020 17:05

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE SOFTWARE VERSION: 8100.27 BREATH TEST OPERATOR: MARTIN,JAMES,E BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD ARRESTING OFFICER: CHAPIN,NATHAN ARRESTING OFFICER AGENCY: Ft. Walton Beach PD LAST AGENCY INSPECTION: 01/31/2020	Date or Time Changed: No DRY GAS STANDARD LOT #: 181692 DRY GAS EXPIRY DATE: 02/05/2021 VIOLATION CODE: DUI ARREST TIME: 21:25 GENDER: Male OBSERVATION PERIOD BEGAN: 22:05
--	---

SUBJECT'S LAST NAME: HILL    DATE/TIME OF TEST: 02/08/2020                      22:37:53

---

Diagnostics Check	OK	22:26	
Air Blank	0.000	22:26	
Control Test	0.078	22:27	
Air Blank	0.000	22:27	
Subject Sample	# 1 0.212	22:30	Volume Not Met
Breath Volume	0.609	22:30	
Air Blank	0.000	22:31	
Air Blank	0.000	22:33	
Subject Sample	# 2	22:35	Subject Test Refused
Breath Volume	0.132	22:35	
Air Blank	0.000	22:36	
Control Test	0.078	22:37	
Air Blank	0.000	22:37	
Diagnostics Check	OK	22:37	

---

Data Download Date/Time: 03/04/2020 12:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOLL,MICHAEL,  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: WOLL,MICHAEL  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: RAGSDALE

DATE/TIME OF TEST: 02/15/2020 03:16:58

---

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:14	
Control Test	0.081	03:14	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:16	
Control Test	0.081	03:16	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

---

Data Download Date/Time: 03/04/2020 12:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE  
 SOFTWARE VERSION: 8100.27  
 BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D  
 BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
 ARRESTING OFFICER: CRAWFORD,BENJAMIN  
 ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
 LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
 DRY GAS STANDARD LOT #: 181692  
 DRY GAS EXPIRY DATE: 02/05/2021  
 VIOLATION CODE: DUI  
 ARREST TIME: 19:31  
 GENDER: Female  
 OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 02/18/2020 20:34:00

Diagnostics Check	OK	20:27
Air Blank	0.000	20:28
Control Test	0.080	20:28
Air Blank	0.000	20:29
Subject Sample # 1	0.110	20:29
Breath Volume	2.359	20:29
Air Blank	0.000	20:30
Air Blank	0.000	20:31
Subject Sample # 2	0.104	20:32
Breath Volume	2.546	20:32
Air Blank	0.000	20:33
Control Test	0.080	20:33
Air Blank	0.000	20:33
Diagnostics Check	OK	20:34

Data Download Date/Time: 03/04/2020 12:38



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIRMANS,ABNER,K  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: SIRMANS,ABNER  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: SCAFIDI

DATE/TIME OF TEST: 02/21/2020 22:34:37

---

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:26	
Control Test	0.081	22:26	
Air Blank	0.000	22:27	
Subject Sample	# 1	22:30	No Sample Provided
Breath Volume	0.000	22:30	
Air Blank	0.000	22:30	
Air Blank	0.000	22:32	
Subject Sample	# 2	22:33	Subject Test Refused
Breath Volume	0.000	22:33	
Air Blank	0.000	22:33	
Control Test	0.082	22:34	
Air Blank	0.000	22:34	
Diagnostics Check	OK	22:34	

---

Data Download Date/Time: 03/04/2020 12:38

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIRMANS, ABNER, K  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: SIRMANS, ABNER  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: SCAFIDI

DATE/TIME OF TEST: 02/21/2020 22:45:24

---

Diagnostics Check	OK	22:40	
Air Blank	0.000	22:41	
Control Test	0.081	22:41	
Air Blank	0.000	22:42	
Subject Sample # 1	0.167	22:43	
Breath Volume	1.199	22:43	
Air Blank		22:43	RFI Detect
Air Blank	0.000	22:44	
Control Test	0.081	22:44	
Air Blank	0.000	22:45	
Diagnostics Check	OK	22:45	

---

Data Download Date/Time: 03/04/2020 12:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIRMANS, ABNER, K  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: SIRMANS, ABNER  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: SCAFIDI

DATE/TIME OF TEST: 02/21/2020 22:57:43

---

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.079	22:51
Air Blank	0.000	22:51
Subject Sample # 1	0.165	22:52
Breath Volume	1.734	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.169	22:56
Breath Volume	1.613	22:56
Air Blank	0.000	22:56
Control Test	0.078	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

---

Data Download Date/Time: 03/04/2020 12:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOLL, MICHAEL,  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: WOLL, MICHAEL  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: IVEY

DATE/TIME OF TEST: 02/22/2020 00:10:34

---

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.080	00:04
Air Blank	0.000	00:04
Subject Sample # 1	0.174	00:05
Breath Volume	2.964	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.177	00:08
Breath Volume	1.445	00:08
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

---

Data Download Date/Time: 03/04/2020 12:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOLL,MICHAEL,  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: WOLL,MICHAEL  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HAMBLEY

DATE/TIME OF TEST: 02/23/2020 01:26:33

---

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:24	
Control Test	0.080	01:24	
Air Blank	0.000	01:24	
Subject Sample	# 1	01:25	Subject Test Refused
Breath Volume	0.000	01:25	
Air Blank	0.000	01:25	
Control Test	0.080	01:25	
Air Blank	0.000	01:26	
Diagnostics Check	OK	01:26	

---

Data Download Date/Time: 03/04/2020 12:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: CHAPIN,NATHAN  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BRYAN

DATE/TIME OF TEST: 02/27/2020 02:08:38

---

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:06	
Control Test	0.082	02:06	
Air Blank	0.000	02:06	
Subject Sample	# 1	02:07	Subject Test Refused
Breath Volume	0.000	02:07	
Air Blank	0.000	02:07	
Control Test	0.082	02:08	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

---

Data Download Date/Time: 03/04/2020 12:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CABRAL,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: CABRAL,JONATHAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: VAN BUREN

DATE/TIME OF TEST: 02/21/2020 02:57:56

---

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:52	
Control Test	0.080	02:53	
Air Blank	0.000	02:53	
Subject Sample	# 1	02:56	Subject Test Refused
Breath Volume	0.000	02:56	
Air Blank	0.000	02:57	
Control Test	0.081	02:57	
Air Blank	0.000	02:57	
Diagnostics Check	OK	02:57	

---

Data Download Date/Time: 03/10/2020 14:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LITZKOW,CARLTON,D  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: LITZKOW,CARLTON  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 02/24/2020 01:36:53

---

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:25	
Control Test	0.079	01:26	
Air Blank	0.000	01:26	
Subject Sample	# 1	01:29	No Sample Provided
Breath Volume	0.000	01:29	
Air Blank	0.000	01:30	
Air Blank	0.000	01:32	
Subject Sample	# 2	01:35	No Sample Provided
Breath Volume	0.000	01:35	
Air Blank	0.000	01:35	
Control Test	0.081	01:36	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

---

Data Download Date/Time: 03/10/2020 14:31



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELDON,ERIC,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SHELDON,ERIC  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: CARLOS

DATE/TIME OF TEST: 02/01/2020 05:17:00

---

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:14	
Control Test	0.081	05:14	
Air Blank	0.000	05:15	
Subject Sample	# 1	05:15	Subject Test Refused
Breath Volume	0.000	05:15	
Air Blank	0.000	05:16	
Control Test	0.081	05:16	
Air Blank	0.000	05:16	
Diagnostics Check	OK	05:17	

---

Data Download Date/Time: 02/28/2020 11:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELDON,ERIC,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SHELDON,ERIC  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: SAVOY

DATE/TIME OF TEST: 02/03/2020 01:31:07

---

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.082	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.163	01:26
Breath Volume	2.714	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.158	01:29
Breath Volume	2.730	01:29
Air Blank	0.000	01:30
Control Test	0.081	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

---

Data Download Date/Time: 02/28/2020 11:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILHELM,TONY,J  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: WILHELM,TONY  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: GAFFORD

DATE/TIME OF TEST: 02/03/2020 01:55:16

---

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.081	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.094	01:50
Breath Volume		1.816	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.086	01:53
Breath Volume		1.929	01:53
Air Blank		0.000	01:54
Control Test		0.080	01:54
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

---

Data Download Date/Time: 02/28/2020 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELDON,ERIC,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: CARROLL,SHELBY  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 02/03/2020 02:21:39

---

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:17	
Control Test	0.081	02:17	
Air Blank	0.000	02:18	
Subject Sample	# 1	02:20	Subject Test Refused
Breath Volume	0.000	02:20	
Air Blank	0.000	02:20	
Control Test	0.082	02:21	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

---

Data Download Date/Time: 02/28/2020 11:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,RAYMOND,S  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SMITH,JACKSON  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/03/2020 23:37:56

---

Diagnostics Check	OK	23:34	
Air Blank	0.000	23:34	
Control Test	0.081	23:34	
Air Blank	0.000	23:35	
Subject Sample	# 1	23:36	Subject Test Refused
Breath Volume	0.000	23:36	
Air Blank	0.000	23:37	
Control Test	0.081	23:37	
Air Blank	0.000	23:37	
Diagnostics Check	OK	23:37	

---

Data Download Date/Time: 02/28/2020 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: ABEL,ANTHONY  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: GILES

DATE/TIME OF TEST: 02/07/2020 01:48:59

---

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:46	
Control Test	0.081	01:46	
Air Blank	0.000	01:47	
Subject Sample	# 1	01:47	Subject Test Refused
Breath Volume	0.000	01:47	
Air Blank	0.000	01:48	
Control Test	0.081	01:48	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

---

Data Download Date/Time: 02/28/2020 11:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELDON,ERIC,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SHELDON,ERIC  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: LANGSTON

DATE/TIME OF TEST: 02/08/2020 03:04:04

---

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.082	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.151	02:59
Breath Volume	2.714	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:01
Subject Sample # 2	0.149	03:02
Breath Volume	2.152	03:02
Air Blank	0.000	03:03
Control Test	0.082	03:03
Air Blank	0.000	03:03
Diagnostics Check	OK	03:04

---

Data Download Date/Time: 02/28/2020 11:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELDON,ERIC,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SHELDON,ERIC  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/09/2020 00:29:14

---

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:23
Subject Sample # 1	0.128	00:24
Breath Volume	1.578	00:24
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.129	00:27
Breath Volume	1.558	00:27
Air Blank	0.000	00:28
Control Test	0.080	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

---

Data Download Date/Time: 02/28/2020 11:13



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELDON,ERIC,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SHELDON,ERIC  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 02/09/2020 05:04:54

---

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.081	04:59
Air Blank	0.000	04:59
Subject Sample # 1	0.057	05:00
Breath Volume	1.421	05:00
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.058	05:03
Breath Volume	2.750	05:03
Air Blank	0.000	05:03
Control Test	0.077	05:04
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

---

Data Download Date/Time: 02/28/2020 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,LANCE,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: MONARCH,BRIAN  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: TURNPAUGH

DATE/TIME OF TEST: 02/09/2020 21:28:55

---

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:20	
Control Test	0.081	21:21	
Air Blank	0.000	21:21	
Subject Sample	# 1	21:24	No Sample Provided
Breath Volume	0.000	21:24	
Air Blank	0.000	21:25	
Air Blank	0.000	21:27	
Subject Sample	# 2	21:27	Subject Test Refused
Breath Volume	0.000	21:27	
Air Blank	0.000	21:28	
Control Test	0.081	21:28	
Air Blank	0.000	21:28	
Diagnostics Check	OK	21:28	

---

Data Download Date/Time: 02/28/2020 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELDON,ERIC,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SHELDON,ERIC  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: FINLEY

DATE/TIME OF TEST: 02/12/2020 21:14:08

---

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:11	
Control Test	0.081	21:11	
Air Blank	0.000	21:12	
Subject Sample	# 1	21:12	Subject Test Refused
Breath Volume	0.000	21:12	
Air Blank	0.000	21:13	
Control Test	0.080	21:13	
Air Blank	0.000	21:14	
Diagnostics Check	OK	21:14	

---

Data Download Date/Time: 02/28/2020 11:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELDON,ERIC,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SHELDON,ERIC  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: MOLTER

DATE/TIME OF TEST: 02/13/2020 02:28:50

---

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.080	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.019	02:24
Breath Volume	2.386	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.018	02:27
Breath Volume	2.375	02:27
Air Blank	0.000	02:27
Control Test	0.081	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

---

Data Download Date/Time: 02/28/2020 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMSON, TYLER, H  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: WILLIAMSON, TYLER  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: BARREIRA

DATE/TIME OF TEST: 02/14/2020 02:18:50

---

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.082	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.256	02:14
Breath Volume	2.574	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:16
Subject Sample # 2	0.241	02:17
Breath Volume	2.164	02:17
Air Blank	0.000	02:17
Control Test	0.081	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

---

Data Download Date/Time: 02/28/2020 11:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELDON,ERIC,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: ABEL,ANTHONY  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SHOLLER

DATE/TIME OF TEST: 02/15/2020 03:32:04

---

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.082	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.205	03:27
Breath Volume	3.078	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:29
Subject Sample # 2	0.196	03:30
Breath Volume	2.320	03:30
Air Blank	0.000	03:31
Control Test	0.081	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:32

---

Data Download Date/Time: 02/28/2020 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FREDERICK, TAVISH, M  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: FREDERICK, TAVISH  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: BONILLA-OCHOA

DATE/TIME OF TEST: 02/16/2020 05:11:25

---

Diagnostics Check		OK	05:00	
Air Blank		0.000	05:00	
Control Test		0.082	05:00	
Air Blank		0.000	05:01	
Subject Sample	# 1	0.141	05:04	Volume Not Met
Breath Volume		0.246	05:04	
Air Blank		0.000	05:05	
Air Blank		0.000	05:06	
Subject Sample	# 2		05:10	No Sample Provided
Breath Volume		0.000	05:10	
Air Blank		0.000	05:10	
Control Test		0.080	05:10	
Air Blank		0.000	05:11	
Diagnostics Check		OK	05:11	

---

Data Download Date/Time: 02/28/2020 11:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DREW,MAREIKE,M  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: DREW,MAREIKE  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: SALVADOR

DATE/TIME OF TEST: 02/16/2020 05:45:15

---

Diagnostics Check	OK	05:38
Air Blank	0.000	05:38
Control Test	0.081	05:38
Air Blank	0.000	05:39
Subject Sample # 1	0.192	05:39
Breath Volume	1.402	05:39
Air Blank	0.000	05:40
Air Blank	0.000	05:42
Subject Sample # 2	0.206	05:43
Breath Volume	4.621	05:43
Air Blank	0.000	05:44
Control Test	0.081	05:44
Air Blank	0.000	05:45
Diagnostics Check	OK	05:45

---

Data Download Date/Time: 02/28/2020 11:13



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELDON,ERIC,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SHELDON,ERIC  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: SEALE

DATE/TIME OF TEST: 02/16/2020 06:20:14

---

Diagnostics Check	OK	06:17	
Air Blank	0.000	06:17	
Control Test	0.081	06:18	
Air Blank	0.000	06:18	
Subject Sample	# 1	06:18	Subject Test Refused
Breath Volume	0.000	06:18	
Air Blank	0.000	06:19	
Control Test	0.080	06:19	
Air Blank	0.000	06:20	
Diagnostics Check	OK	06:20	

---

Data Download Date/Time: 02/28/2020 11:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FREDERICK, TAVISH, M  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: FREDERICK, TAVISH  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BALDER

DATE/TIME OF TEST: 02/17/2020 04:33:17

---

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:22	
Control Test	0.080	04:22	
Air Blank	0.000	04:23	
Subject Sample	# 1	04:26	No Sample Provided
Breath Volume	0.000	04:26	
Air Blank	0.000	04:26	
Air Blank	0.000	04:28	
Subject Sample	# 2	04:31	No Sample Provided
Breath Volume	0.000	04:31	
Air Blank	0.000	04:32	
Control Test	0.081	04:32	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

---

Data Download Date/Time: 02/28/2020 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, RAYMOND, S  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SMITH, JACKSON  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: BALTAZAR CATALAN

DATE/TIME OF TEST: 02/19/2020 04:30:49

---

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.148	04:25
Breath Volume	3.078	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.146	04:29
Breath Volume	2.804	04:29
Air Blank	0.000	04:29
Control Test	0.079	04:30
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

---

Data Download Date/Time: 02/28/2020 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARSHALL,ALEXANDRIA,R  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: PAPPAS,D  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:34

SUBJECT'S LAST NAME: STATON

DATE/TIME OF TEST: 02/19/2020 07:03:02

---

Diagnostics Check	OK	06:54	
Air Blank	0.000	06:54	
Control Test	0.080	06:55	
Air Blank	0.000	06:55	
Subject Sample	# 1	06:58	No Sample Provided
Breath Volume	0.000	06:58	
Air Blank	0.000	06:59	
Air Blank	0.000	07:01	
Subject Sample	# 2	07:01	Subject Test Refused
Breath Volume	0.000	07:01	
Air Blank	0.000	07:02	
Control Test	0.080	07:02	
Air Blank	0.000	07:02	
Diagnostics Check	OK	07:03	

---

Data Download Date/Time: 02/28/2020 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DREW,MAREIKE,M  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: DREW,MAREIKE  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 02/20/2020 02:41:20

---

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.080	02:30	
Air Blank	0.000	02:31	
Subject Sample	# 1	02:34	No Sample Provided
Breath Volume	0.000	02:34	
Air Blank	0.000	02:34	
Air Blank	0.000	02:36	
Subject Sample	# 2	02:39	No Sample Provided
Breath Volume	0.000	02:39	
Air Blank	0.000	02:40	
Control Test	0.081	02:40	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

---

Data Download Date/Time: 02/28/2020 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FREDERICK, TAVISH, M  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: FREDERICK, TAVISH  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: LOPEZ CADENA

DATE/TIME OF TEST: 02/20/2020 02:47:01

---

Diagnostics Check	OK	02:46	
Air Blank		02:46	RFI Detect
Air Blank	0.000	02:47	

---

Data Download Date/Time: 02/28/2020 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FREDERICK, TAVISH, M  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: FREDERICK, TAVISH  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: LOPEZ CADENA

DATE/TIME OF TEST: 02/20/2020 03:05:28

---

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:49	
Control Test	0.080	02:49	
Air Blank	0.000	02:50	
Subject Sample # 1	0.171	02:52	
Breath Volume	1.144	02:52	
Air Blank	0.000	02:53	
Air Blank	0.000	02:55	
Subject Sample # 2	0.155	02:58	Volume Not Met
Breath Volume	0.535	02:58	
Air Blank	0.000	02:59	
Air Blank	0.000	03:00	
Subject Sample # 3		03:04	No Sample Provided
Breath Volume	0.000	03:04	
Air Blank	0.000	03:04	
Control Test	0.073	03:04	Control Outside Tolerance
Air Blank	0.000	03:05	

---

Data Download Date/Time: 02/28/2020 11:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMALL,BRANNON,A  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: COOPER,TERRI  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 11:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:24

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 02/20/2020 12:51:13

---

Diagnostics Check	OK	12:44
Air Blank	0.000	12:45
Control Test	0.079	12:45
Air Blank	0.000	12:46
Subject Sample # 1	0.363	12:46
Breath Volume	2.000	12:46
Air Blank	0.000	12:47
Air Blank	0.000	12:49
Subject Sample # 2	0.351	12:49
Breath Volume	2.554	12:49
Air Blank	0.000	12:50
Control Test	0.075	12:50
Air Blank	0.000	12:51
Diagnostics Check	OK	12:51

---

Data Download Date/Time: 02/28/2020 11:13



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: ESNEUL,MICHAEL  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: SCHUBERT

DATE/TIME OF TEST: 02/21/2020 05:32:29

---

Diagnostics Check	OK	05:28	
Air Blank	0.000	05:29	
Control Test	0.081	05:29	
Air Blank	0.000	05:30	
Subject Sample	# 1	05:31	Subject Test Refused
Breath Volume	0.000	05:31	
Air Blank	0.000	05:31	
Control Test	0.081	05:31	
Air Blank	0.000	05:32	
Diagnostics Check	OK	05:32	

---

Data Download Date/Time: 02/28/2020 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINCOLN,A,  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SMITH,JACKSON  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: TRAHAN

DATE/TIME OF TEST: 02/22/2020 05:04:57

---

Diagnostics Check	OK	05:02	
Air Blank	0.000	05:02	
Control Test	0.082	05:02	
Air Blank	0.000	05:03	
Subject Sample	# 1	05:03	Subject Test Refused
Breath Volume	0.000	05:03	
Air Blank	0.000	05:04	
Control Test	0.082	05:04	
Air Blank	0.000	05:04	
Diagnostics Check	OK	05:04	

---

Data Download Date/Time: 02/28/2020 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRISTOL,JARRET,W  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BRISTOL,JARRET  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 10:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: BREEDEN

DATE/TIME OF TEST: 02/22/2020 11:19:03

---

Diagnostics Check	OK	11:11
Air Blank	0.000	11:12
Control Test	0.083	11:12
Air Blank	0.000	11:13
Subject Sample # 1	0.120	11:14
Breath Volume	2.378	11:14
Air Blank	0.000	11:14
Air Blank	0.000	11:16
Subject Sample # 2	0.130	11:17
Breath Volume	2.863	11:17
Air Blank	0.000	11:18
Control Test	0.082	11:18
Air Blank	0.000	11:18
Diagnostics Check	OK	11:19

---

Data Download Date/Time: 02/28/2020 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMILTON, RAY, C  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BEMENT, JESSICA  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 02/22/2020 23:54:31

---

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.081	23:46
Air Blank	0.000	23:47
Subject Sample # 1	0.207	23:49
Breath Volume	1.320	23:49
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.192	23:52
Breath Volume	2.902	23:52
Air Blank	0.000	23:53
Control Test	0.083	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

---

Data Download Date/Time: 02/28/2020 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELDON,ERIC,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SHELDON,ERIC  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 02/23/2020 01:24:16

---

Diagnostics Check	OK	01:17
Air Blank	0.000	01:18
Control Test	0.082	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.199	01:19
Breath Volume	4.503	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.198	01:22
Breath Volume	4.027	01:22
Air Blank	0.000	01:23
Control Test	0.082	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

---

Data Download Date/Time: 02/28/2020 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARRINGTON,IAN,M  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: ARRINGTON,IAN  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: DUNBAR

DATE/TIME OF TEST: 02/23/2020 19:29:38

---

Diagnostics Check	OK	19:26	
Air Blank	0.000	19:26	
Control Test	0.081	19:27	
Air Blank	0.000	19:27	
Subject Sample	# 1	19:28	Subject Test Refused
Breath Volume	0.000	19:28	
Air Blank	0.000	19:28	
Control Test	0.081	19:29	
Air Blank	0.000	19:29	
Diagnostics Check	OK	19:29	

---

Data Download Date/Time: 02/28/2020 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMILTON, RAY, C  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BEMENT, JESSICA  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: DEARING

DATE/TIME OF TEST: 02/23/2020 21:23:00

---

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:19	
Control Test	0.081	21:19	
Air Blank	0.000	21:20	
Subject Sample	# 1	21:21	Subject Test Refused
Breath Volume	0.000	21:21	
Air Blank	0.000	21:22	
Control Test	0.081	21:22	
Air Blank	0.000	21:22	
Diagnostics Check	OK	21:23	

---

Data Download Date/Time: 02/28/2020 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,LANCE,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: WILLIAMS,LANCE  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: ESPARANZA

DATE/TIME OF TEST: 02/23/2020 22:29:53

---

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:26	
Control Test	0.081	22:26	
Air Blank	0.000	22:26	
Subject Sample	# 1	22:28	Subject Test Refused
Breath Volume	0.000	22:28	
Air Blank	0.000	22:28	
Control Test	0.081	22:29	
Air Blank	0.000	22:29	
Diagnostics Check	OK	22:29	

---

Data Download Date/Time: 02/28/2020 11:13



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNHAM,J,D  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: PAPPAS,DANIEL  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: HENN

DATE/TIME OF TEST: 02/24/2020 01:46:21

---

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.081	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.199	01:41
Breath Volume	1.292	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.211	01:44
Breath Volume	1.343	01:44
Air Blank	0.000	01:45
Control Test	0.080	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

---

Data Download Date/Time: 02/28/2020 11:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMALL,BRANNON,A  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: STEPHENSON,DEBRA  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 02/25/2020 02:11:36

---

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:03	
Control Test	0.080	02:03	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:07	No Sample Provided
Breath Volume	0.000	02:07	
Air Blank	0.000	02:08	
Air Blank	0.000	02:10	
Subject Sample	# 2	02:10	Subject Test Refused
Breath Volume	0.000	02:10	
Air Blank	0.000	02:10	
Control Test	0.080	02:10	
Air Blank	0.000	02:11	
Diagnostics Check	OK	02:11	

---

Data Download Date/Time: 02/28/2020 11:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNHAM,J,D  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BURNHAM,J  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: ZERO TOL  
ARREST TIME: 22:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: LAIN

DATE/TIME OF TEST: 02/27/2020 23:38:31

---

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.083	23:32
Air Blank	0.000	23:33
Subject Sample # 1	0.115	23:33
Breath Volume	1.777	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.127	23:36
Breath Volume	2.132	23:36
Air Blank	0.000	23:37
Control Test	0.081	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

---

Data Download Date/Time: 02/28/2020 11:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURNHAM,J,D  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BURNHAM,J  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: SINGLETARY

DATE/TIME OF TEST: 02/28/2020 02:10:37

---

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.083	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.196	02:05
Breath Volume	1.296	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.193	02:09
Breath Volume	1.179	02:09
Air Blank	0.000	02:09
Control Test	0.082	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

---

Data Download Date/Time: 02/28/2020 11:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,ALISON,C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: HEPBURN,ANDREW  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 792976  
DRY GAS EXPIRY DATE: 06/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 08:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/03/2020 10:17:12

---

Diagnostics Check	OK	10:05	
Air Blank	0.000	10:06	
Control Test	0.081	10:06	
Air Blank	0.000	10:07	
Subject Sample	# 1	10:10	No Sample Provided
Breath Volume	0.000	10:10	
Air Blank	0.000	10:10	
Air Blank	0.000	10:12	
Subject Sample	# 2	10:15	No Sample Provided
Breath Volume	0.000	10:15	
Air Blank	0.000	10:16	
Control Test	0.082	10:16	
Air Blank	0.000	10:17	
Diagnostics Check	OK	10:17	

---

Data Download Date/Time: 02/14/2020 22:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANNING,CARL,T  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: ALEXANDER,C  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 792976  
DRY GAS EXPIRY DATE: 06/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:47

SUBJECT'S LAST NAME: CANFIELD

DATE/TIME OF TEST: 02/10/2020 15:44:55

---

Diagnostics Check	OK	15:38
Air Blank	0.000	15:38
Control Test	0.080	15:38
Air Blank	0.000	15:39
Subject Sample # 1	0.328	15:40
Breath Volume	2.429	15:40
Air Blank	0.000	15:40
Air Blank	0.000	15:42
Subject Sample # 2	0.332	15:43
Breath Volume	1.843	15:43
Air Blank	0.000	15:43
Control Test	0.078	15:44
Air Blank	0.000	15:44
Diagnostics Check	OK	15:44

---

Data Download Date/Time: 02/14/2020 22:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEEKS, JEREMIAH, S  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: MEEKS, JEREMIAH  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 792976  
DRY GAS EXPIRY DATE: 06/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: DELORGE

DATE/TIME OF TEST: 02/11/2020 02:51:13

---

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.081	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.168	02:46
Breath Volume	2.703	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.161	02:49
Breath Volume	2.359	02:49
Air Blank	0.000	02:50
Control Test	0.078	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

---

Data Download Date/Time: 02/14/2020 22:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEEKS, JEREMIAH, S  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: MEEKS, JEREMIAH  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 792976  
DRY GAS EXPIRY DATE: 06/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: GILES JR

DATE/TIME OF TEST: 02/11/2020 20:31:59

---

Diagnostics Check	OK	20:25
Air Blank	0.000	20:26
Control Test	0.081	20:26
Air Blank	0.000	20:27
Subject Sample # 1	0.000	20:27
Breath Volume	4.218	20:27
Air Blank	0.000	20:28
Air Blank	0.000	20:30
Subject Sample # 2	0.000	20:30
Breath Volume	4.417	20:30
Air Blank	0.000	20:31
Control Test	0.080	20:31
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

---

Data Download Date/Time: 02/14/2020 22:15



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ACKER,AARON,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: ACKER,AARON  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 792976  
DRY GAS EXPIRY DATE: 06/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ROSADO

DATE/TIME OF TEST: 02/12/2020 03:40:22

---

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.082	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.168	03:33
Breath Volume	1.183	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.171	03:38
Breath Volume	1.394	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

---

Data Download Date/Time: 02/14/2020 22:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: SHELLEY,SAMUEL  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: OLSHEFSKIE

DATE/TIME OF TEST: 02/15/2020 01:13:35

---

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.082	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.095	01:08
Breath Volume	1.531	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:10
Subject Sample # 2	0.110	01:12
Breath Volume	2.031	01:12
Air Blank	0.000	01:12
Control Test	0.081	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

---

Data Download Date/Time: 03/06/2020 22:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEITE,STEVEN,G  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: LEITE,STEVEN  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: ZBYDENIEWSKI

DATE/TIME OF TEST: 02/16/2020 02:28:00

---

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:17	
Control Test	0.082	02:17	
Air Blank	0.000	02:17	
Subject Sample	# 1	02:21	No Sample Provided
Breath Volume	0.000	02:21	
Air Blank	0.000	02:21	
Air Blank	0.000	02:23	
Subject Sample	# 2	02:26	No Sample Provided
Breath Volume	0.000	02:26	
Air Blank	0.000	02:27	
Control Test	0.082	02:27	
Air Blank	0.000	02:27	
Diagnostics Check	OK	02:28	

---

Data Download Date/Time: 03/06/2020 22:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEITE,STEVEN,G  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: LEITE,STEVEN  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: WEBSTER

DATE/TIME OF TEST: 02/22/2020 02:44:24

---

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.083	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.133	02:39
Breath Volume	1.468	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:41
Subject Sample # 2	0.137	02:42
Breath Volume	1.601	02:42
Air Blank	0.000	02:43
Control Test	0.082	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

---

Data Download Date/Time: 03/06/2020 22:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENLEY, WILLIAM, C  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: SELLERS, DONALD  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: DANSBY

DATE/TIME OF TEST: 02/29/2020 23:47:21

---

Diagnostics Check	OK	23:37
Air Blank	0.000	23:38
Control Test	0.083	23:38
Air Blank	0.000	23:39
Subject Sample # 1	0.252	23:39
Breath Volume	1.613	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.277	23:42
Breath Volume	1.519	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 3	0.271	23:45
Breath Volume	1.246	23:45
Air Blank	0.000	23:46
Control Test	0.081	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

---

Data Download Date/Time: 03/06/2020 22:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COWAN,RICHARD,T  
BREATH TEST OPERATOR AGENCY: Jackson County SO  
ARRESTING OFFICER: RAFFERTY,JOSHUA  
ARRESTING OFFICER AGENCY: Jackson County SO  
LAST AGENCY INSPECTION: 01/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1076482  
DRY GAS EXPIRY DATE: 01/15/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 02/01/2020 19:00:35

---

Diagnostics Check	OK	18:53
Air Blank	0.000	18:53
Control Test	0.082	18:54
Air Blank	0.000	18:54
Subject Sample # 1	0.249	18:55
Breath Volume	1.242	18:55
Air Blank	0.000	18:55
Air Blank	0.000	18:57
Subject Sample # 2	0.248	18:58
Breath Volume	1.378	18:58
Air Blank	0.000	18:59
Control Test	0.079	18:59
Air Blank	0.000	19:00
Diagnostics Check	OK	19:00

---

Data Download Date/Time: 02/14/2020 19:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBLES,MIGUEL,A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: ROBLES,MIGUEL  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 02/03/2020 23:37:13

---

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.079	23:30
Air Blank	0.000	23:31
Subject Sample # 1	0.168	23:32
Breath Volume	3.363	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:34
Subject Sample # 2	0.164	23:35
Breath Volume	4.207	23:35
Air Blank	0.000	23:36
Control Test	0.078	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

---

Data Download Date/Time: 02/18/2020 14:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LITZKOW,CLAYTON,P  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: RALEY,BRITT  
ARRESTING OFFICER AGENCY: Alachua PD  
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 02/04/2020 01:51:04

---

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.079	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.167	01:46
Breath Volume	1.683	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.160	01:49
Breath Volume	1.488	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:51

---

Data Download Date/Time: 02/18/2020 14:18



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUGHES,DAVID,D  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: HUGHES,DAVID  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 02/09/2020 01:02:14

---

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:59	
Control Test	0.081	00:59	
Air Blank	0.000	00:59	
Subject Sample	# 1	01:00	Subject Test Refused
Breath Volume	0.000	01:00	
Air Blank	0.000	01:01	
Control Test	0.081	01:01	
Air Blank		01:01	RFI Detect
Air Blank	0.000	01:02	

---

Data Download Date/Time: 02/18/2020 14:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DILLON,JUSTIN,P  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: DILLON,JUSTIN  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: GLINERT

DATE/TIME OF TEST: 02/11/2020 00:11:06

---

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.079	00:02
Air Blank	0.000	00:03
Subject Sample # 1	0.203	00:06
Breath Volume	2.105	00:06
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.219	00:09
Breath Volume	1.851	00:09
Air Blank	0.000	00:10
Control Test	0.078	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

---

Data Download Date/Time: 02/18/2020 14:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITTMAN,BRETT,T  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: PITTMAN,BRETT  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 02/13/2020 01:54:49

---

Diagnostics Check	OK	01:54	
Air Blank		01:54	RFI Detect
Air Blank	0.000	01:54	

---

Data Download Date/Time: 02/18/2020 14:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITTMAN,BRETT,T  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: PITTMAN,BRETT  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 02/13/2020 02:03:26

---

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.080	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.105	01:58
Breath Volume	1.667	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.105	02:01
Breath Volume	2.437	02:01
Air Blank	0.000	02:02
Control Test	0.078	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

---

Data Download Date/Time: 02/18/2020 14:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUTSCHER,REBECCA,L  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: MYERS,BEN  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:21

SUBJECT'S LAST NAME: MOYER

DATE/TIME OF TEST: 02/15/2020 06:55:21

---

Diagnostics Check		OK	06:43	
Air Blank		0.000	06:43	
Control Test		0.081	06:44	
Air Blank		0.000	06:44	
Subject Sample	# 1	0.155	06:47	Volume Not Met
Breath Volume		0.792	06:47	
Air Blank		0.000	06:48	
Air Blank		0.000	06:50	
Subject Sample	# 2	0.201	06:50	
Breath Volume		2.179	06:50	
Air Blank		0.000	06:51	
Air Blank		0.000	06:53	
Subject Sample	# 3	0.202	06:53	
Breath Volume		1.765	06:53	
Air Blank		0.000	06:54	
Control Test		0.078	06:54	
Air Blank		0.000	06:55	
Diagnostics Check		OK	06:55	

---

Data Download Date/Time: 02/18/2020 14:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREIRA,JOSE,L  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: PEREIRA,JOSE  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 754250  
DRY GAS EXPIRY DATE: 04/03/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 02/01/2020 00:32:22

---

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.176	00:27
Breath Volume	3.269	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.174	00:30
Breath Volume	2.710	00:30
Air Blank	0.000	00:31
Control Test	0.080	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

---

Data Download Date/Time: 02/27/2020 11:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPRAGUE,JONATHON,D  
BREATH TEST OPERATOR AGENCY: Crestview PD  
ARRESTING OFFICER: SPRAGUE,JONATHON  
ARRESTING OFFICER AGENCY: Crestview PD  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 754250  
DRY GAS EXPIRY DATE: 04/03/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: KUMFER

DATE/TIME OF TEST: 02/03/2020 03:13:27

---

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.082	03:08
Breath Volume	4.289	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.080	03:11
Breath Volume	3.343	03:11
Air Blank	0.000	03:12
Control Test	0.080	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

---

Data Download Date/Time: 02/27/2020 11:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STEPHENSON,PATRICK,W  
BREATH TEST OPERATOR AGENCY: Crestview PD  
ARRESTING OFFICER: STEPHENSON,PATRICK  
ARRESTING OFFICER AGENCY: Crestview PD  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 754250  
DRY GAS EXPIRY DATE: 04/03/2020  
VIOLATION CODE: DUI  
ARREST TIME: 15:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:47

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 02/05/2020 16:16:34

---

Diagnostics Check	OK	16:08
Air Blank	0.000	16:09
Control Test	0.080	16:09
Air Blank	0.000	16:09
Subject Sample # 1	0.177	16:11
Breath Volume	1.359	16:11
Air Blank	0.000	16:12
Air Blank	0.000	16:14
Subject Sample # 2	0.185	16:14
Breath Volume	1.343	16:14
Air Blank	0.000	16:15
Control Test	0.078	16:15
Air Blank	0.000	16:16
Diagnostics Check	OK	16:16

---

Data Download Date/Time: 02/27/2020 11:56



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,JODY,W  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: WALLEN,JUSTIN  
ARRESTING OFFICER AGENCY: Crestview PD  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 754250  
DRY GAS EXPIRY DATE: 04/03/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: TOLES

DATE/TIME OF TEST: 02/08/2020 01:28:53

---

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:26	
Control Test	0.081	01:26	
Air Blank	0.000	01:27	
Subject Sample	# 1	01:27	Subject Test Refused
Breath Volume	0.000	01:27	
Air Blank	0.000	01:27	
Control Test	0.080	01:28	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

---

Data Download Date/Time: 02/27/2020 11:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,JODY,W  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: BURGESS,JOSH  
ARRESTING OFFICER AGENCY: Crestview PD  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 754250  
DRY GAS EXPIRY DATE: 04/03/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: HUTCHERSON

DATE/TIME OF TEST: 02/08/2020 03:46:10

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.081	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.120	03:41
Breath Volume	1.570	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.118	03:44
Breath Volume	1.710	03:44
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

---

Data Download Date/Time: 02/27/2020 11:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,JODY,W  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: HOWARD,JODY  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 754250  
DRY GAS EXPIRY DATE: 04/03/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 02/09/2020 02:02:25

---

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.080	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.214	01:57
Breath Volume	2.230	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.227	02:00
Breath Volume	2.433	02:00
Air Blank	0.000	02:01
Control Test	0.078	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

---

Data Download Date/Time: 02/27/2020 11:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOORE,JACOB,A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: MOORE,JACOB  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 754250  
DRY GAS EXPIRY DATE: 04/03/2020  
VIOLATION CODE: DUI  
ARREST TIME: 13:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:38

SUBJECT'S LAST NAME: MALAMOS

DATE/TIME OF TEST: 02/11/2020 15:05:39

---

Diagnostics Check	OK	14:58
Air Blank	0.000	14:59
Control Test	0.081	14:59
Air Blank	0.000	15:00
Subject Sample # 1	0.053	15:01
Breath Volume	1.445	15:01
Air Blank	0.000	15:01
Air Blank	0.000	15:03
Subject Sample # 2	0.054	15:04
Breath Volume	2.023	15:04
Air Blank	0.000	15:04
Control Test	0.080	15:05
Air Blank	0.000	15:05
Diagnostics Check	OK	15:05

---

Data Download Date/Time: 02/27/2020 11:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GENRICH,KRISTEN,R  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: LANGER,MATTHEW  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 754250  
DRY GAS EXPIRY DATE: 04/03/2020  
VIOLATION CODE: DUI  
ARREST TIME: 13:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: WESTBROOK

DATE/TIME OF TEST: 02/17/2020 14:22:29

---

Diagnostics Check	OK	14:15
Air Blank	0.000	14:16
Control Test	0.079	14:16
Air Blank	0.000	14:17
Subject Sample # 1	0.191	14:17
Breath Volume	4.824	14:17
Air Blank	0.000	14:18
Air Blank	0.000	14:20
Subject Sample # 2	0.180	14:20
Breath Volume	4.898	14:20
Air Blank	0.000	14:21
Control Test	0.079	14:21
Air Blank	0.000	14:22
Diagnostics Check	OK	14:22

---

Data Download Date/Time: 02/27/2020 11:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAIN,DAVID,A  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SPRAGUE,JONATHON  
ARRESTING OFFICER AGENCY: Crestview PD  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 02/22/2020 01:32:40

---

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:30	
Control Test	0.081	01:30	
Air Blank	0.000	01:31	
Subject Sample	# 1	01:31	Subject Test Refused
Breath Volume	0.000	01:31	
Air Blank	0.000	01:31	
Control Test	0.081	01:32	
Air Blank	0.000	01:32	
Diagnostics Check	OK	01:32	

---

Data Download Date/Time: 02/27/2020 11:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURGESS,MASON,W  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: JONES,KEITH  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/22/2020 20:47:30

---

Diagnostics Check	OK	20:38
Air Blank	0.000	20:39
Control Test	0.081	20:39
Air Blank	0.000	20:40
Subject Sample # 1	0.315	20:40
Breath Volume	2.207	20:40
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.325	20:45
Breath Volume	1.343	20:45
Air Blank	0.000	20:46
Control Test	0.078	20:46
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

---

Data Download Date/Time: 02/27/2020 11:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GORDON,JON,D  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: HURST,CARY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 02/23/2020 02:57:47

---

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.082	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.125	02:53
Breath Volume	3.464	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:55
Subject Sample # 2	0.129	02:56
Breath Volume	2.496	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

---

Data Download Date/Time: 02/27/2020 11:56



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MACFADDEN,RYAN,E  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MACFADDEN,RYAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: NOLAND

DATE/TIME OF TEST: 02/06/2020 03:59:36

---

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.077	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.120	03:55
Breath Volume	2.101	03:55
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.121	03:57
Breath Volume	1.953	03:57
Air Blank	0.000	03:58
Control Test	0.077	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

---

Data Download Date/Time: 02/18/2020 10:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,DANIEL,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MCFARLAND,MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 02/08/2020 18:28:05

---

Diagnostics Check	OK	18:25	
Air Blank	0.000	18:25	
Control Test	0.079	18:25	
Air Blank	0.000	18:26	
Subject Sample	# 1	18:26	Subject Test Refused
Breath Volume	0.000	18:26	
Air Blank	0.000	18:27	
Control Test	0.079	18:27	
Air Blank	0.000	18:28	
Diagnostics Check	OK	18:28	

---

Data Download Date/Time: 02/18/2020 10:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: BINES,DYLAN  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1175968  
DRY GAS EXPIRY DATE: 07/30/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: HOLMAN

DATE/TIME OF TEST: 02/11/2020 23:05:08

---

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.079	22:58
Air Blank	0.000	22:59
Subject Sample # 1	0.245	23:00
Breath Volume	2.390	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:02
Subject Sample # 2	0.247	23:03
Breath Volume	2.984	23:03
Air Blank	0.000	23:04
Control Test	0.077	23:04
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

---

Data Download Date/Time: 02/18/2020 10:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BATES,BRIAN,C  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: STEPHENS,JAKE  
ARRESTING OFFICER AGENCY: Lake City PD  
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/13/2020 20:47:38

---

Diagnostics Check	OK	20:40
Air Blank	0.000	20:41
Control Test	0.078	20:41
Air Blank	0.000	20:42
Subject Sample # 1	0.068	20:43
Breath Volume	2.101	20:43
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.060	20:46
Breath Volume	2.093	20:46
Air Blank	0.000	20:46
Control Test	0.076	20:47
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

---

Data Download Date/Time: 02/18/2020 10:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NATIVIDAD,DANIEL,A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: SISTRUNK,TRENT  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 02/14/2020 00:48:08

---

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:45	
Control Test	0.079	00:46	
Air Blank	0.000	00:46	
Subject Sample	# 1	00:46	Subject Test Refused
Breath Volume	0.000	00:46	
Air Blank	0.000	00:47	
Control Test	0.079	00:47	
Air Blank	0.000	00:48	
Diagnostics Check	OK	00:48	

---

Data Download Date/Time: 02/18/2020 10:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARVEY,DAVID,A  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: HARVEY,DAVID  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: REGISTER

DATE/TIME OF TEST: 02/14/2020 21:56:49

---

Diagnostics Check	OK	21:50
Air Blank	0.000	21:50
Control Test	0.078	21:51
Air Blank	0.000	21:51
Subject Sample # 1	0.230	21:52
Breath Volume	1.410	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:54
Subject Sample # 2	0.244	21:54
Breath Volume	1.468	21:54
Air Blank	0.000	21:55
Control Test	0.077	21:56
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

---

Data Download Date/Time: 02/18/2020 10:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARVEY,DAVID,A  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: MONTEMURNO,MATTHEW  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: MUNCY

DATE/TIME OF TEST: 02/15/2020 01:33:12

---

Diagnostics Check	OK	01:24	
Air Blank	0.000	01:24	
Control Test	0.079	01:25	
Air Blank	0.000	01:25	
Subject Sample	# 1	01:28	No Sample Provided
Breath Volume	0.000	01:28	
Air Blank	0.000	01:29	
Air Blank	0.000	01:31	
Subject Sample	# 2	01:31	Subject Test Refused
Breath Volume	0.000	01:31	
Air Blank	0.000	01:32	
Control Test	0.080	01:32	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

---

Data Download Date/Time: 02/18/2020 10:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARVEY,DAVID,A  
BREATH TEST OPERATOR AGENCY: Lake City PD  
ARRESTING OFFICER: PLEMMONS,LAQUINTA  
ARRESTING OFFICER AGENCY: Lake City PD  
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HUTCHESON

DATE/TIME OF TEST: 02/15/2020 02:28:02

---

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.151	02:23
Breath Volume		2.292	02:23
Air Blank		0.000	02:24
Air Blank		0.000	02:25
Subject Sample	# 2	0.153	02:26
Breath Volume		2.761	02:26
Air Blank		0.000	02:27
Control Test		0.077	02:27
Air Blank		0.000	02:27
Diagnostics Check		OK	02:28

---

Data Download Date/Time: 02/18/2020 10:30



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARVEY,DAVID,A  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: HARVEY,DAVID  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: BREEN

DATE/TIME OF TEST: 02/15/2020 20:11:03

---

Diagnostics Check	OK	20:04	
Air Blank	0.000	20:05	
Control Test	0.079	20:05	
Air Blank	0.000	20:06	
Subject Sample # 1	0.033	20:07	
Breath Volume	3.976	20:07	
Air Blank	0.000	20:07	
Air Blank	0.000	20:09	
Subject Sample # 2		20:09	Subject Test Refused
Breath Volume	0.000	20:09	
Air Blank	0.000	20:10	
Control Test	0.077	20:10	
Air Blank	0.000	20:10	
Diagnostics Check	OK	20:11	

---

Data Download Date/Time: 02/18/2020 10:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARVEY,DAVID,A  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: MONTEMURNO,MATTHEW  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 02/16/2020 03:00:21

---

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:55
Subject Sample # 1	0.101	02:55
Breath Volume	2.644	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.096	02:58
Breath Volume	2.582	02:58
Air Blank	0.000	02:59
Control Test	0.078	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

---

Data Download Date/Time: 02/18/2020 10:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUCKLES,SCHIEFER,S  
BREATH TEST OPERATOR AGENCY: FL Dept of AG  
ARRESTING OFFICER: BUCKLES,SCHIEFER  
ARRESTING OFFICER AGENCY: FL Dept of AG  
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1175968  
DRY GAS EXPIRY DATE: 07/30/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SERNA

DATE/TIME OF TEST: 02/16/2020 03:57:18

---

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.126	03:52
Breath Volume		1.308	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.122	03:55
Breath Volume		2.789	03:55
Air Blank		0.000	03:56
Control Test		0.076	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

---

Data Download Date/Time: 02/18/2020 10:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARVEY,DAVID,A  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: HARVEY,DAVID  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 02/17/2020 00:31:12

---

Diagnostics Check	OK	00:24
Air Blank	0.000	00:25
Control Test	0.079	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.103	00:26
Breath Volume	3.910	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:28
Subject Sample # 2	0.099	00:29
Breath Volume	3.976	00:29
Air Blank	0.000	00:30
Control Test	0.076	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

---

Data Download Date/Time: 02/18/2020 10:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARVEY,DAVID,A  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: MONTEMURNO,MATTHEW  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: BOUSQUET

DATE/TIME OF TEST: 02/17/2020 01:35:32

---

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.078	01:29
Air Blank	0.000	01:30
Subject Sample # 1	0.113	01:30
Breath Volume	3.500	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.114	01:33
Breath Volume	4.343	01:33
Air Blank	0.000	01:34
Control Test	0.077	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

---

Data Download Date/Time: 02/18/2020 10:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WORTHINGTON,J,J  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: DUPREE,A  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 02/17/2020 18:45:35

---

Diagnostics Check	OK	18:38
Air Blank	0.000	18:39
Control Test	0.078	18:39
Air Blank	0.000	18:40
Subject Sample # 1	0.191	18:40
Breath Volume	3.179	18:40
Air Blank	0.000	18:41
Air Blank	0.000	18:43
Subject Sample # 2	0.189	18:43
Breath Volume	3.894	18:43
Air Blank	0.000	18:44
Control Test	0.075	18:44
Air Blank	0.000	18:45
Diagnostics Check	OK	18:45

---

Data Download Date/Time: 02/18/2020 10:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SISTRUNK,TRENT,D  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: SISTRUNK,TRENT  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:11

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 02/26/2020 16:38:15

---

Diagnostics Check	OK	16:32
Air Blank	0.000	16:32
Control Test	0.079	16:33
Air Blank	0.000	16:33
Subject Sample # 1	0.000	16:34
Breath Volume	2.000	16:34
Air Blank	0.000	16:34
Air Blank	0.000	16:36
Subject Sample # 2	0.000	16:36
Breath Volume	1.968	16:36
Air Blank	0.000	16:37
Control Test	0.077	16:37
Air Blank	0.000	16:38
Diagnostics Check	OK	16:38

---

Data Download Date/Time: 04/03/2020 09:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,DANIEL,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MARTIN,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: KIRKLAND

DATE/TIME OF TEST: 02/26/2020 22:24:10

---

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:21	
Control Test	0.079	22:22	
Air Blank	0.000	22:22	
Subject Sample	# 1	22:22	Subject Test Refused
Breath Volume	0.000	22:22	
Air Blank	0.000	22:23	
Control Test	0.079	22:23	
Air Blank	0.000	22:24	
Diagnostics Check	OK	22:24	

---

Data Download Date/Time: 04/03/2020 09:25



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BALDREE,TASHAHNDA,J  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BALDREE,TASHAHNDA  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: STIRN

DATE/TIME OF TEST: 02/28/2020 19:04:46

---

Diagnostics Check	OK	19:00	
Air Blank	0.000	19:00	
Control Test	0.081	19:01	
Air Blank	0.000	19:01	
Subject Sample	# 1 0.330	19:03	
Breath Volume	1.925	19:03	
Air Blank		19:04	Purge Fail
Air Blank	0.000	19:04	

---

Data Download Date/Time: 04/03/2020 09:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BALDREE,TASHAHNDA,J  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BALDREE,TASHAHNDA  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: STIRN

DATE/TIME OF TEST: 02/28/2020 19:24:09

---

Diagnostics Check	OK	19:15
Air Blank	0.000	19:15
Control Test	0.079	19:15
Air Blank	0.000	19:16
Subject Sample # 1	0.319	19:16
Breath Volume	1.468	19:16
Air Blank	0.000	19:17
Air Blank	0.000	19:19
Subject Sample # 2	0.328	19:22
Breath Volume	2.101	19:22
Air Blank	0.000	19:23
Control Test	0.078	19:23
Air Blank	0.000	19:24
Diagnostics Check	OK	19:24

---

Data Download Date/Time: 04/03/2020 09:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOLIN,ALBERT,L  
BREATH TEST OPERATOR AGENCY: FL Dept of AG  
ARRESTING OFFICER: QUATER,ANTHONY  
ARRESTING OFFICER AGENCY: FL Dept of AG  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: COLEMAN MURILLO

DATE/TIME OF TEST: 02/01/2020 01:12:03

---

Diagnostics Check	OK	01:11	
Air Blank		01:11	RFI Detect
Air Blank		01:12	RFI Detect

---

Data Download Date/Time: 02/13/2020 11:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOLIN,ALBERT,L  
BREATH TEST OPERATOR AGENCY: FL Dept of AG  
ARRESTING OFFICER: QUATER,ANTHONY  
ARRESTING OFFICER AGENCY: FL Dept of AG  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: COLEMAN MURILLO

DATE/TIME OF TEST: 02/01/2020 01:21:39

---

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:16	
Control Test	0.078	01:16	
Air Blank	0.000	01:17	
Subject Sample	# 1	01:19	RFI Detect
Breath Volume	0.695	01:19	
Air Blank	0.000	01:20	
Control Test	0.079	01:21	
Air Blank	0.000	01:21	
Diagnostics Check	OK	01:21	

---

Data Download Date/Time: 02/13/2020 11:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOLIN,ALBERT,L  
BREATH TEST OPERATOR AGENCY: FL Dept of AG  
ARRESTING OFFICER: QUATER,ANTHONY  
ARRESTING OFFICER AGENCY: FL Dept of AG  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: COLEMAN MURILLO

DATE/TIME OF TEST: 02/01/2020 01:31:30

---

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.076	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.123	01:25
Breath Volume	3.410	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:27
Subject Sample # 2	0.111	01:29
Breath Volume	2.363	01:29
Air Blank	0.000	01:30
Control Test	0.076	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

---

Data Download Date/Time: 02/13/2020 11:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITE, RAYMOND, C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: WHITE, RAYMOND  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 02/12/2020 20:02:54

---

Diagnostics Check	OK	19:55
Air Blank	0.000	19:56
Control Test	0.080	19:56
Air Blank	0.000	19:57
Subject Sample # 1	0.226	19:58
Breath Volume	3.652	19:58
Air Blank	0.000	19:58
Air Blank	0.000	20:00
Subject Sample # 2	0.206	20:01
Breath Volume	1.812	20:01
Air Blank	0.000	20:01
Control Test	0.078	20:02
Air Blank	0.000	20:02
Diagnostics Check	OK	20:02

---

Data Download Date/Time: 02/13/2020 11:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,DAVID,H  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: SMITH,DAVID  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: BERNARD

DATE/TIME OF TEST: 02/19/2020 22:51:46

---

Diagnostics Check		OK	22:43	
Air Blank		0.000	22:44	
Control Test		0.079	22:44	
Air Blank		0.000	22:44	
Subject Sample	# 1	0.170	22:48	Volume Not Met
Breath Volume		0.531	22:48	
Air Blank		0.000	22:48	
Air Blank		0.000	22:50	
Subject Sample	# 2		22:51	Sequence Aborted
Breath Volume		0.000	22:51	
Air Blank		0.000	22:51	

---

Data Download Date/Time: 03/10/2020 12:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP TROOP A  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OPRY, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: OPRY, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 01/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 02/08/2020 02:41:08

---

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.079	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.101	02:36
Breath Volume	1.167	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.098	02:39
Breath Volume	1.878	02:39
Air Blank	0.000	02:40
Control Test	0.078	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

---

Data Download Date/Time: 02/14/2020 19:12



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP TROOP A

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: OPRY, WILLIAM, J

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: OPRY, WILLIAM

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 01/01/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 770784

DRY GAS EXPIRY DATE: 04/27/2020

VIOLATION CODE: DUI

ARREST TIME: 01:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 02/08/2020 03:01:07

---

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:58	
Control Test	0.078	02:58	
Air Blank	0.000	02:59	
Subject Sample	# 1	02:59	Subject Test Refused
Breath Volume	0.000	02:59	
Air Blank	0.000	03:00	
Control Test	0.078	03:00	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

---

Data Download Date/Time: 02/14/2020 19:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP TROOP A  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOLOMON, JR,  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: SOLOMON, JR  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 02/17/2020 21:46:44

---

Diagnostics Check	OK	21:40
Air Blank	0.000	21:41
Control Test	0.079	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.000	21:42
Breath Volume	1.617	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.000	21:45
Breath Volume	2.132	21:45
Air Blank	0.000	21:45
Control Test	0.079	21:46
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

---

Data Download Date/Time: 03/05/2020 15:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GENTRY, JONATHAN, P  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: GENTRY, JONATHAN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 01/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/07/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: TURNEY

DATE/TIME OF TEST: 02/11/2020 18:59:49

---

Diagnostics Check	OK	18:51
Air Blank	0.000	18:51
Control Test	0.084	18:51
Air Blank	0.000	18:52
Subject Sample # 1	0.192	18:54
Breath Volume	1.843	18:54
Air Blank	0.000	18:55
Air Blank	0.000	18:56
Subject Sample # 2	0.188	18:57
Breath Volume	1.648	18:57
Air Blank	0.000	18:58
Control Test	0.079	18:59
Air Blank	0.000	18:59
Diagnostics Check	OK	18:59

---

Data Download Date/Time: 02/12/2020 02:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LITTLE,NICHOLAS,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: LITTLE,NICHOLAS  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/07/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:13

SUBJECT'S LAST NAME: COTTLE

DATE/TIME OF TEST: 02/14/2020 07:40:42

---

Diagnostics Check	OK	07:33
Air Blank	0.000	07:34
Control Test	0.083	07:34
Air Blank	0.000	07:34
Subject Sample # 1	0.228	07:35
Breath Volume	4.050	07:35
Air Blank	0.000	07:36
Air Blank	0.000	07:38
Subject Sample # 2	0.230	07:38
Breath Volume	3.929	07:38
Air Blank	0.000	07:39
Control Test	0.078	07:40
Air Blank	0.000	07:40
Diagnostics Check	OK	07:40

---

Data Download Date/Time: 03/02/2020 03:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOWMAN,JOANNA,N  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: BOWMAN,JOANNA  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/07/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: CEYDELI

DATE/TIME OF TEST: 02/22/2020 02:26:02

---

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:23	
Control Test	0.084	02:23	
Air Blank	0.000	02:24	
Subject Sample	# 1	02:24	Subject Test Refused
Breath Volume	0.000	02:24	
Air Blank	0.000	02:25	
Control Test	0.084	02:25	
Air Blank	0.000	02:25	
Diagnostics Check	OK	02:26	

---

Data Download Date/Time: 03/02/2020 03:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: LEWIS,TIMOTHY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/07/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: BAXLEY

DATE/TIME OF TEST: 02/23/2020 20:53:52

---

Diagnostics Check	OK	20:45
Air Blank	0.000	20:46
Control Test	0.083	20:46
Air Blank	0.000	20:47
Subject Sample # 1	0.099	20:47
Breath Volume	1.230	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 2	0.104	20:52
Breath Volume	1.589	20:52
Air Blank	0.000	20:52
Control Test	0.082	20:53
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

---

Data Download Date/Time: 03/02/2020 03:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GENTRY, JONATHAN, P  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: GENTRY, JONATHAN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/07/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: FORD II

DATE/TIME OF TEST: 02/28/2020 18:47:47

---

Diagnostics Check	OK	18:40
Air Blank	0.000	18:40
Control Test	0.084	18:41
Air Blank	0.000	18:41
Subject Sample # 1	0.335	18:42
Breath Volume	4.187	18:42
Air Blank	0.000	18:43
Air Blank	0.000	18:45
Subject Sample # 2	0.324	18:45
Breath Volume	3.414	18:45
Air Blank	0.000	18:46
Control Test	0.080	18:47
Air Blank	0.000	18:47
Diagnostics Check	OK	18:47

---

Data Download Date/Time: 03/02/2020 03:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINCOLN,TREY,B  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: LINCOLN,TREY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/07/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/29/2020 04:18:30

---

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.085	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.179	04:13
Breath Volume	1.523	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.178	04:16
Breath Volume	1.480	04:16
Air Blank	0.000	04:17
Control Test	0.081	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

---

Data Download Date/Time: 03/02/2020 03:11



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENSON,JAY,C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: HENSON,JAY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/07/2020  
VIOLATION CODE: DUI  
ARREST TIME: 15:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:34

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 02/29/2020 17:02:51

---

Diagnostics Check	OK	16:55	
Air Blank	0.000	16:55	
Control Test	0.083	16:56	
Air Blank	0.000	16:56	
Subject Sample # 1	0.266	16:57	
Breath Volume	1.191	16:57	
Air Blank	0.000	16:58	
Air Blank	0.000	17:00	
Subject Sample # 2	0.253	17:01	
Breath Volume	1.488	17:01	
Air Blank	0.000	17:01	
Control Test	0.080	17:02	Interferent Detect
Air Blank	0.000	17:02	

---

Data Download Date/Time: 03/02/2020 03:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TATRO,KERSTAN,J  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: TATRO,KERSTAN  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: MCLEOD

DATE/TIME OF TEST: 02/07/2020 22:44:59

---

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.080	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.084	22:40
Breath Volume	1.367	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:42
Subject Sample # 2	0.086	22:43
Breath Volume	2.289	22:43
Air Blank	0.000	22:44
Control Test	0.080	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

---

Data Download Date/Time: 02/10/2020 02:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERSON,THOMAS,C  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: WEIERS,BLAINE  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: CONDREAY

DATE/TIME OF TEST: 02/09/2020 04:03:53

---

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.079	03:58
Air Blank		0.000	03:58
Subject Sample	# 1	0.245	03:59
Breath Volume		2.027	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:01
Subject Sample	# 2	0.254	04:02
Breath Volume		2.296	04:02
Air Blank		0.000	04:02
Control Test		0.080	04:03
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

---

Data Download Date/Time: 02/10/2020 02:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERSON,THOMAS,C  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: WHITE,TRAVIS  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: FRILOUX

DATE/TIME OF TEST: 02/09/2020 04:55:35

---

Diagnostics Check	OK	04:49
Air Blank	0.000	04:49
Control Test	0.079	04:50
Air Blank	0.000	04:50
Subject Sample # 1	0.048	04:51
Breath Volume	2.300	04:51
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.046	04:54
Breath Volume	2.679	04:54
Air Blank	0.000	04:54
Control Test	0.079	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

---

Data Download Date/Time: 02/10/2020 02:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERSON,THOMAS,C  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: WEIERS,BLAINE  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 02/10/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 02/13/2020 04:28:28

---

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:23	
Control Test	0.079	04:23	
Air Blank	0.000	04:23	
Subject Sample # 1	0.163	04:27	Volume Not Met
Breath Volume	0.992	04:27	
Air Blank		04:27	Purge Fail
Air Blank	0.000	04:28	

---

Data Download Date/Time: 03/03/2020 00:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERSON,THOMAS,C  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: WEIERS,BLAINE  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 02/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 02/13/2020 04:35:59

---

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:31	
Control Test	0.078	04:32	
Air Blank	0.000	04:32	
Subject Sample	# 1	04:35	Subject Test Refused
Breath Volume	1.167	04:35	
Air Blank		04:35	RFI Detect
Air Blank	0.000	04:35	

---

Data Download Date/Time: 03/03/2020 00:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMS,KYLE,N  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: SIMMS,KYLE  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 02/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 02/28/2020 19:11:26

---

Diagnostics Check	OK	19:05
Air Blank	0.000	19:05
Control Test	0.080	19:06
Air Blank	0.000	19:06
Subject Sample # 1	0.000	19:07
Breath Volume	4.183	19:07
Air Blank	0.000	19:07
Air Blank	0.000	19:09
Subject Sample # 2	0.000	19:10
Breath Volume	3.355	19:10
Air Blank	0.000	19:10
Control Test	0.081	19:10
Air Blank	0.000	19:11
Diagnostics Check	OK	19:11

---

Data Download Date/Time: 03/03/2020 00:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMB,HUNTER,  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LIMB,HUNTER  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: LOSEY

DATE/TIME OF TEST: 02/03/2020 23:52:13

---

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:47
Subject Sample # 1	0.138	23:47
Breath Volume	1.558	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.135	23:50
Breath Volume	1.824	23:50
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

---

Data Download Date/Time: 02/04/2020 06:22



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: RAYA RAYA

DATE/TIME OF TEST: 02/04/2020 21:32:32

---

Diagnostics Check	OK	21:26
Air Blank	0.000	21:26
Control Test	0.080	21:27
Air Blank	0.000	21:27
Subject Sample # 1	0.109	21:28
Breath Volume	2.925	21:28
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.111	21:31
Breath Volume	3.035	21:31
Air Blank	0.000	21:31
Control Test	0.080	21:31
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

---

Data Download Date/Time: 02/26/2020 11:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOODRUFF,MARK,D  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: ECKRODE,ALBERT  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: SICES

DATE/TIME OF TEST: 02/05/2020 05:10:49

---

Diagnostics Check		OK	05:02	
Air Blank		0.000	05:03	
Control Test		0.080	05:03	
Air Blank		0.000	05:04	
Subject Sample	# 1	0.288	05:06	
Breath Volume		1.902	05:06	
Air Blank		0.000	05:07	
Air Blank		0.000	05:08	
Subject Sample	# 2		05:09	Slope Not Met
Breath Volume		0.625	05:09	
Air Blank		0.000	05:09	
Control Test		0.079	05:10	
Air Blank		0.000	05:10	
Diagnostics Check		OK	05:10	

---

Data Download Date/Time: 02/26/2020 11:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SELPH,JASON,H  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SELPH,JASON  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 02/06/2020 15:52:20

---

Diagnostics Check	OK	15:46
Air Blank	0.000	15:46
Control Test	0.079	15:46
Air Blank	0.000	15:47
Subject Sample # 1	0.000	15:48
Breath Volume	2.816	15:48
Air Blank	0.000	15:48
Air Blank	0.000	15:50
Subject Sample # 2	0.000	15:50
Breath Volume	2.746	15:50
Air Blank	0.000	15:51
Control Test	0.079	15:51
Air Blank	0.000	15:52
Diagnostics Check	OK	15:52

---

Data Download Date/Time: 02/26/2020 11:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUILES,ALEXANDER,  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: QUILES,ALEXANDER  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: LEROY

DATE/TIME OF TEST: 02/12/2020 03:19:25

---

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.079	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.143	03:14
Breath Volume	2.445	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.146	03:17
Breath Volume	1.632	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

---

Data Download Date/Time: 02/26/2020 11:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 02/13/2020 19:19:55

---

Diagnostics Check	OK	19:13
Air Blank	0.000	19:14
Control Test	0.079	19:14
Air Blank	0.000	19:15
Subject Sample # 1	0.000	19:15
Breath Volume	2.500	19:15
Air Blank	0.000	19:16
Air Blank	0.000	19:18
Subject Sample # 2	0.000	19:18
Breath Volume	2.625	19:18
Air Blank	0.000	19:19
Control Test	0.080	19:19
Air Blank	0.000	19:19
Diagnostics Check	OK	19:19

---

Data Download Date/Time: 02/26/2020 11:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,RICKY,  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: PEREZ,RICKY  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/14/2020 21:14:13

---

Diagnostics Check	OK	21:06
Air Blank	0.000	21:06
Control Test	0.080	21:07
Air Blank	0.000	21:07
Subject Sample # 1	0.190	21:09
Breath Volume	1.480	21:09
Air Blank	0.000	21:09
Air Blank	0.000	21:11
Subject Sample # 2	0.199	21:12
Breath Volume	1.777	21:12
Air Blank	0.000	21:13
Control Test	0.080	21:13
Air Blank	0.000	21:14
Diagnostics Check	OK	21:14

---

Data Download Date/Time: 02/26/2020 11:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRAWDY,LAINE,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SUHLING,TASHA  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: SOMMER

DATE/TIME OF TEST: 02/15/2020 01:40:13

---

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.094	01:35
Breath Volume	2.835	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.097	01:38
Breath Volume	3.066	01:38
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

---

Data Download Date/Time: 02/26/2020 11:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUILES,ALEXANDER,  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: QUILES,ALEXANDER  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: PINE

DATE/TIME OF TEST: 02/16/2020 18:42:57

---

Diagnostics Check	OK	18:36
Air Blank	0.000	18:37
Control Test	0.079	18:37
Air Blank	0.000	18:37
Subject Sample # 1	0.169	18:38
Breath Volume	3.714	18:38
Air Blank	0.000	18:39
Air Blank	0.000	18:40
Subject Sample # 2	0.159	18:41
Breath Volume	2.843	18:41
Air Blank	0.000	18:42
Control Test	0.079	18:42
Air Blank	0.000	18:42
Diagnostics Check	OK	18:42

---

Data Download Date/Time: 02/26/2020 11:53



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRAWDY,LAINE,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: DRAWDY,LAINE  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 02/17/2020 00:57:16

---

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.000	00:52
Breath Volume	2.390	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.000	00:55
Breath Volume	2.507	00:55
Air Blank	0.000	00:56
Control Test	0.080	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

---

Data Download Date/Time: 02/26/2020 11:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 02/19/2020 01:26:27

---

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.053	01:21
Breath Volume	2.476	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.052	01:24
Breath Volume	2.628	01:24
Air Blank	0.000	01:25
Control Test	0.080	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

---

Data Download Date/Time: 02/26/2020 11:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUILES,ALEXANDER,  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: QUILES,ALEXANDER  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: GRIMSLEY

DATE/TIME OF TEST: 02/20/2020 21:59:24

---

Diagnostics Check	OK	21:56	
Air Blank	0.000	21:57	
Control Test	0.079	21:57	
Air Blank	0.000	21:58	
Subject Sample	# 1	21:58	Sequence Aborted
Breath Volume	0.000	21:58	
Air Blank	0.000	21:59	

---

Data Download Date/Time: 02/26/2020 11:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: RIVERA AQUINO

DATE/TIME OF TEST: 02/21/2020 23:18:38

---

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.080	23:12
Air Blank	0.000	23:12
Subject Sample # 1	0.056	23:13
Breath Volume	1.609	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.058	23:17
Breath Volume	1.644	23:17
Air Blank	0.000	23:17
Control Test	0.080	23:18
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

---

Data Download Date/Time: 02/26/2020 11:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: MAGANA ANDRADE

DATE/TIME OF TEST: 02/23/2020 00:43:47

---

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:40	
Control Test	0.080	00:40	
Air Blank	0.000	00:40	
Subject Sample	# 1	00:42	Subject Test Refused
Breath Volume	0.000	00:42	
Air Blank	0.000	00:42	
Control Test	0.080	00:43	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

---

Data Download Date/Time: 02/26/2020 11:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASSERLY,STEPHEN,R  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: CASSERLY,STEPHEN  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 02/24/2020 21:35:51

---

Diagnostics Check	OK	21:29
Air Blank	0.000	21:30
Control Test	0.080	21:30
Air Blank	0.000	21:30
Subject Sample # 1	0.173	21:31
Breath Volume	1.621	21:31
Air Blank	0.000	21:32
Air Blank	0.000	21:33
Subject Sample # 2	0.155	21:34
Breath Volume	1.687	21:34
Air Blank	0.000	21:34
Control Test	0.078	21:35
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

---

Data Download Date/Time: 02/26/2020 11:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 02/26/2020 20:23:40

---

Diagnostics Check	OK	20:17
Air Blank	0.000	20:17
Control Test	0.079	20:18
Air Blank	0.000	20:18
Subject Sample # 1	0.295	20:19
Breath Volume	1.871	20:19
Air Blank	0.000	20:19
Air Blank	0.000	20:21
Subject Sample # 2	0.284	20:22
Breath Volume	1.839	20:22
Air Blank	0.000	20:22
Control Test	0.079	20:23
Air Blank	0.000	20:23
Diagnostics Check	OK	20:23

---

Data Download Date/Time: 03/13/2020 14:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHOUINARD,JEFFREY,M  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: CHOUINARD,JEFFREY  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: RAYAAZ

DATE/TIME OF TEST: 02/27/2020 03:30:18

---

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.215	03:25
Breath Volume	2.203	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.205	03:28
Breath Volume	1.835	03:28
Air Blank	0.000	03:29
Control Test	0.077	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

---

Data Download Date/Time: 03/13/2020 14:24



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MACFADDEN,RYAN,E  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MACFADDEN,RYAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 02/02/2020 04:03:14

---

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.199	03:58
Breath Volume	1.726	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:00
Subject Sample # 2	0.186	04:01
Breath Volume	2.375	04:01
Air Blank	0.000	04:02
Control Test	0.078	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

---

Data Download Date/Time: 02/03/2020 22:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CZABAN,NICHOLAS,W  
BREATH TEST OPERATOR AGENCY: FL Dept of AG  
ARRESTING OFFICER: HOWELL,DANIEL  
ARRESTING OFFICER AGENCY: FL Dept of AG  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:02

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 02/02/2020 14:37:58

---

Diagnostics Check	OK	14:30
Air Blank	0.000	14:31
Control Test	0.078	14:31
Air Blank	0.000	14:31
Subject Sample # 1	0.225	14:33
Breath Volume	2.988	14:33
Air Blank	0.000	14:33
Air Blank	0.000	14:35
Subject Sample # 2	0.212	14:36
Breath Volume	2.480	14:36
Air Blank	0.000	14:37
Control Test	0.076	14:37
Air Blank	0.000	14:37
Diagnostics Check	OK	14:37

---

Data Download Date/Time: 02/03/2020 22:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BALDREE,TASHAHNDA,J  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BALDREE,TASHAHNDA  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 09:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 02/08/2020 10:17:10

---

Diagnostics Check	OK	10:10
Air Blank	0.000	10:11
Control Test	0.079	10:11
Air Blank	0.000	10:12
Subject Sample # 1	0.108	10:12
Breath Volume	2.250	10:12
Air Blank	0.000	10:13
Air Blank	0.000	10:15
Subject Sample # 2	0.106	10:15
Breath Volume	2.050	10:15
Air Blank	0.000	10:16
Control Test	0.079	10:16
Air Blank	0.000	10:17
Diagnostics Check	OK	10:17

---

Data Download Date/Time: 02/27/2020 18:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WADFORD,CHARLES,T  
BREATH TEST OPERATOR AGENCY: Hamilton County SO  
ARRESTING OFFICER: WADFORD,CHARLES  
ARRESTING OFFICER AGENCY: Hamilton County SO  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GREATHOUSE

DATE/TIME OF TEST: 02/22/2020 01:55:15

---

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.190	01:50
Breath Volume	1.808	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.188	01:53
Breath Volume	1.265	01:53
Air Blank	0.000	01:54
Control Test	0.079	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

---

Data Download Date/Time: 02/27/2020 18:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CURRY, WILLIAM, C  
BREATH TEST OPERATOR AGENCY: Hamilton County SO  
ARRESTING OFFICER: TOWNSEND, CLYDE  
ARRESTING OFFICER AGENCY: FL Dept of AG  
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: AVINS

DATE/TIME OF TEST: 02/23/2020 03:49:41

---

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:47	
Control Test	0.080	03:47	
Air Blank	0.000	03:47	
Subject Sample	# 1	03:48	Subject Test Refused
Breath Volume	0.000	03:48	
Air Blank	0.000	03:48	
Control Test	0.080	03:49	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

---

Data Download Date/Time: 02/27/2020 18:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIBBEY,BRADLEY,A  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HELM,DEREK  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: HIMES-RUETH

DATE/TIME OF TEST: 02/18/2020 01:41:37

---

Diagnostics Check		OK	01:34
Air Blank		0.000	01:35
Control Test		0.078	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.244	01:36
Breath Volume		1.945	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.252	01:40
Breath Volume		1.710	01:40
Air Blank		0.000	01:40
Control Test		0.077	01:41
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

---

Data Download Date/Time: 03/31/2020 09:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CROFT,BRETT,  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: CROFT,BRETT  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 02/20/2020 01:25:57

---

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:20	
Control Test	0.077	01:21	
Air Blank	0.000	01:21	
Subject Sample	# 1	01:24	Subject Test Refused
Breath Volume	0.000	01:24	
Air Blank	0.000	01:25	
Control Test	0.077	01:25	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

---

Data Download Date/Time: 03/31/2020 09:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPENCER, COLLEEN, E  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: CASTIGLIONE, SALVATORE  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: HILDEBRAND

DATE/TIME OF TEST: 02/22/2020 00:47:51

---

Diagnostics Check		OK	00:40
Air Blank		0.000	00:41
Control Test		0.078	00:41
Air Blank		0.000	00:41
Subject Sample	# 1	0.132	00:43
Breath Volume		1.464	00:43
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.139	00:46
Breath Volume		1.949	00:46
Air Blank		0.000	00:46
Control Test		0.079	00:47
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

---

Data Download Date/Time: 03/31/2020 09:01



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: HELM,D  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 02/22/2020 04:16:47

---

Diagnostics Check	OK	04:13	
Air Blank	0.000	04:14	
Control Test	0.079	04:14	
Air Blank	0.000	04:14	
Subject Sample	# 1	04:15	Subject Test Refused
Breath Volume	0.000	04:15	
Air Blank	0.000	04:15	
Control Test	0.079	04:16	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

---

Data Download Date/Time: 03/31/2020 09:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASCITELLI,V,D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: MASCITELLI,V  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: GARIEPY

DATE/TIME OF TEST: 02/22/2020 22:39:20

---

Diagnostics Check	OK	22:33
Air Blank	0.000	22:33
Control Test	0.079	22:33
Air Blank	0.000	22:34
Subject Sample # 1	0.076	22:34
Breath Volume	2.734	22:34
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.070	22:37
Breath Volume	2.191	22:37
Air Blank	0.000	22:38
Control Test	0.079	22:38
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

---

Data Download Date/Time: 03/31/2020 09:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: SCHMIDT,D  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 02/22/2020 23:55:11

---

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:51	
Control Test	0.077	23:51	
Air Blank	0.000	23:51	
Subject Sample	# 1	23:53	Subject Test Refused
Breath Volume	0.000	23:53	
Air Blank	0.000	23:54	
Control Test	0.078	23:54	
Air Blank	0.000	23:55	
Diagnostics Check	OK	23:55	

---

Data Download Date/Time: 03/31/2020 09:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HECKMAN,DANIEL,O  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: MASCITELLI,VINCENT  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: SYKES

DATE/TIME OF TEST: 02/27/2020 00:45:34

---

Diagnostics Check	OK	00:34
Air Blank	0.000	00:35
Control Test	0.077	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.157	00:38
Breath Volume	1.164	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:40
Subject Sample # 2	0.156	00:43
Breath Volume	1.757	00:43
Air Blank	0.000	00:44
Control Test	0.077	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

---

Data Download Date/Time: 03/31/2020 09:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARTNER,ARTHUR,F  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GARTNER,ARTHUR  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 12:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: SPENCE

DATE/TIME OF TEST: 02/28/2020 01:42:56

---

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.029	01:38
Breath Volume	1.242	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.029	01:41
Breath Volume	1.437	01:41
Air Blank	0.000	01:42
Control Test	0.079	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

---

Data Download Date/Time: 03/31/2020 09:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CROFT,BRETT,  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: CROFT,BRETT  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: SCROGGINS

DATE/TIME OF TEST: 02/29/2020 01:57:51

---

Diagnostics Check	OK	01:53	
Air Blank	0.000	01:53	
Control Test	0.078	01:54	
Air Blank	0.000	01:54	
Subject Sample	# 1	01:56	Subject Test Refused
Breath Volume	0.000	01:56	
Air Blank	0.000	01:56	
Control Test	0.079	01:57	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

---

Data Download Date/Time: 03/31/2020 09:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLIS,TERENCE,L  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: WELLMAN,A  
ARRESTING OFFICER AGENCY: Wildwood PD  
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: WRAY

DATE/TIME OF TEST: 02/18/2020 02:51:11

---

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:48	
Control Test	0.078	02:49	
Air Blank	0.000	02:49	
Subject Sample	# 1	02:49	Subject Test Refused
Breath Volume	0.000	02:49	
Air Blank	0.000	02:50	
Control Test	0.077	02:50	
Air Blank	0.000	02:51	
Diagnostics Check	OK	02:51	

---

Data Download Date/Time: 03/14/2020 19:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAUDLE,PAUL,A  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: MCCONNELL,AUSTIN  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: GOMAZ MENDEZ

DATE/TIME OF TEST: 02/19/2020 19:21:16

---

Diagnostics Check	OK	19:18	
Air Blank	0.000	19:18	
Control Test	0.077	19:19	
Air Blank	0.000	19:19	
Subject Sample	# 1	19:19	Subject Test Refused
Breath Volume	0.000	19:19	
Air Blank	0.000	19:20	
Control Test	0.077	19:20	
Air Blank	0.000	19:21	
Diagnostics Check	OK	19:21	

---

Data Download Date/Time: 03/14/2020 19:24



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAUDLE,PAUL,A  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: MEYERS,ERIC  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: ZEEMAN

DATE/TIME OF TEST: 02/20/2020 23:09:29

---

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.077	23:02
Air Blank		0.000	23:03
Subject Sample	# 1	0.281	23:04
Breath Volume		1.292	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:06
Subject Sample	# 2	0.282	23:07
Breath Volume		3.480	23:07
Air Blank		0.000	23:08
Control Test		0.075	23:08
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

---

Data Download Date/Time: 03/14/2020 19:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRACKETT,ALAN,M  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: LOMBARDI,ANTHONY  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MATTIE

DATE/TIME OF TEST: 02/21/2020 04:27:33

---

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.077	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.159	04:22
Breath Volume	3.433	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.163	04:25
Breath Volume	3.335	04:25
Air Blank	0.000	04:26
Control Test	0.077	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

---

Data Download Date/Time: 03/14/2020 19:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLIS,TERENCE,L  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: GUINANE,J  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 15:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:17

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 02/22/2020 18:49:08

---

Diagnostics Check	OK	18:37
Air Blank	0.000	18:38
Control Test	0.078	18:38
Air Blank	0.000	18:39
Subject Sample # 1	0.291	18:40
Breath Volume	1.964	18:40
Air Blank	0.000	18:41
Air Blank	0.000	18:42
Subject Sample # 2	0.269	18:43
Breath Volume	1.160	18:43
Air Blank	0.000	18:44
Air Blank	0.000	18:46
Subject Sample # 3	0.283	18:47
Breath Volume	1.789	18:47
Air Blank	0.000	18:48
Control Test	0.078	18:48
Air Blank	0.000	18:49
Diagnostics Check	OK	18:49

---

Data Download Date/Time: 03/14/2020 19:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLIS,TERENCE,L  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: OTERO,H  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: BADAL

DATE/TIME OF TEST: 02/23/2020 00:43:25

---

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.079	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.115	00:38
Breath Volume	1.941	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.122	00:41
Breath Volume	2.164	00:41
Air Blank	0.000	00:42
Control Test	0.078	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

---

Data Download Date/Time: 03/14/2020 19:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLIS,TERENCE,L  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: COLLINS,M  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: MCMILLEN

DATE/TIME OF TEST: 02/24/2020 01:58:40

---

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.078	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.107	01:54
Breath Volume		2.960	01:54
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.105	01:57
Breath Volume		1.980	01:57
Air Blank		0.000	01:57
Control Test		0.078	01:58
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

---

Data Download Date/Time: 03/14/2020 19:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONG,ERIC,W  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: MCCONNELL,AUSTIN  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: KUJAVA

DATE/TIME OF TEST: 02/01/2020 23:06:14

---

Diagnostics Check	OK	23:03	
Air Blank	0.000	23:03	
Control Test	0.079	23:04	
Air Blank	0.000	23:04	
Subject Sample	# 1	23:04	Subject Test Refused
Breath Volume	0.000	23:04	
Air Blank	0.000	23:05	
Control Test	0.079	23:05	
Air Blank	0.000	23:06	
Diagnostics Check	OK	23:06	

---

Data Download Date/Time: 02/15/2020 09:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAUDLE,PAUL,A  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: SCHMIDT,ANDREW  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: KLEPACKI

DATE/TIME OF TEST: 02/02/2020 01:18:15

---

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.079	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.211	01:13
Breath Volume		2.296	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.194	01:16
Breath Volume		1.933	01:16
Air Blank		0.000	01:17
Control Test		0.080	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

---

Data Download Date/Time: 02/15/2020 09:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: KELLY,WESLEY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:13

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 02/06/2020 13:39:28

---

Diagnostics Check	OK	13:33
Air Blank	0.000	13:33
Control Test	0.078	13:33
Air Blank	0.000	13:34
Subject Sample # 1	0.081	13:34
Breath Volume	2.148	13:34
Air Blank	0.000	13:35
Air Blank	0.000	13:37
Subject Sample # 2	0.076	13:37
Breath Volume	1.320	13:37
Air Blank	0.000	13:38
Control Test	0.078	13:38
Air Blank	0.000	13:39
Diagnostics Check	OK	13:39

---

Data Download Date/Time: 02/15/2020 09:09



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON,BRENDA,R  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: FLYNN,PATRICK  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: DYAL

DATE/TIME OF TEST: 02/08/2020 21:12:43

---

Diagnostics Check	OK	21:05
Air Blank	0.000	21:06
Control Test	0.081	21:06
Air Blank	0.000	21:07
Subject Sample # 1	0.071	21:08
Breath Volume	2.496	21:08
Air Blank	0.000	21:09
Air Blank	0.000	21:10
Subject Sample # 2	0.072	21:11
Breath Volume	1.785	21:11
Air Blank	0.000	21:11
Control Test	0.080	21:12
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

---

Data Download Date/Time: 02/15/2020 09:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLIS,TERENCE,L  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: MATA,R  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/09/2020 22:20:27

---

Diagnostics Check	OK	22:12
Air Blank	0.000	22:12
Control Test	0.080	22:12
Air Blank	0.000	22:13
Subject Sample # 1	0.211	22:15
Breath Volume	1.980	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.213	22:18
Breath Volume	1.480	22:18
Air Blank	0.000	22:19
Control Test	0.079	22:19
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

---

Data Download Date/Time: 02/15/2020 09:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLIS,TERENCE,L  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: POWERS,C  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 02/12/2020 19:33:50

---

Diagnostics Check	OK	19:30	
Air Blank	0.000	19:31	
Control Test	0.079	19:31	
Air Blank	0.000	19:32	
Subject Sample	# 1	19:32	Subject Test Refused
Breath Volume	0.000	19:32	
Air Blank	0.000	19:32	
Control Test	0.080	19:33	
Air Blank	0.000	19:33	
Diagnostics Check	OK	19:33	

---

Data Download Date/Time: 02/15/2020 09:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAUDLE,PAUL,A  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: GILBERT,JASON  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: LUDWIG

DATE/TIME OF TEST: 02/15/2020 00:58:34

---

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:55	
Control Test	0.080	00:56	
Air Blank	0.000	00:56	
Subject Sample	# 1	00:57	Subject Test Refused
Breath Volume	0.000	00:57	
Air Blank	0.000	00:57	
Control Test	0.080	00:57	
Air Blank	0.000	00:58	
Diagnostics Check	OK	00:58	

---

Data Download Date/Time: 02/15/2020 09:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAUDLE,PAUL,A  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: NELSON,GARRETT  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/15/2020 03:36:53

---

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.000	03:32
Breath Volume	2.484	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.000	03:35
Breath Volume	2.093	03:35
Air Blank	0.000	03:35
Control Test	0.079	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

---

Data Download Date/Time: 02/15/2020 09:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: SCHMIET,ANDREW  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: GALARZA

DATE/TIME OF TEST: 02/15/2020 06:35:40

---

Diagnostics Check	OK	06:32	
Air Blank	0.000	06:33	
Control Test	0.080	06:33	
Air Blank	0.000	06:34	
Subject Sample	# 1	06:34	Subject Test Refused
Breath Volume	0.000	06:34	
Air Blank	0.000	06:34	
Control Test	0.080	06:35	
Air Blank	0.000	06:35	
Diagnostics Check	OK	06:35	

---

Data Download Date/Time: 02/15/2020 09:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOODS,JUSTICE,J  
BREATH TEST OPERATOR AGENCY: Citrus County SO  
ARRESTING OFFICER: ELIAS,DANIEL  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 02/20/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: HERNANDEZ GUERRERO

DATE/TIME OF TEST: 02/21/2020 01:19:59

---

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.076	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.172	01:15
Breath Volume	1.503	01:15
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.163	01:18
Breath Volume	1.796	01:18
Air Blank	0.000	01:19
Control Test	0.075	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

---

Data Download Date/Time: 12/11/2020 13:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,DEVIN,A  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: BOUDREAU,J  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 02/20/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: HULL

DATE/TIME OF TEST: 02/23/2020 01:02:10

---

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:59	
Control Test	0.077	00:59	
Air Blank	0.000	01:00	
Subject Sample	# 1	01:00	Subject Test Refused
Breath Volume	0.000	01:00	
Air Blank	0.000	01:01	
Control Test	0.077	01:01	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

---

Data Download Date/Time: 12/11/2020 13:36



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TURO,JAMES,M  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: STRICKLAND,D  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 11:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: LABARRE

DATE/TIME OF TEST: 02/28/2020 13:53:17

---

Diagnostics Check	OK	13:51	
Air Blank	0.000	13:52	
Control Test	0.078	13:52	
Air Blank		13:52	Improper Sample
Air Blank	0.000	13:53	

---

Data Download Date/Time: 12/11/2020 13:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TURO,JAMES,M  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: STRICKLAND,D  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 11:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: LABARRE

DATE/TIME OF TEST: 02/28/2020 13:57:29

---

Diagnostics Check	OK	13:56	
Air Blank		13:57	Ambient Fail
Air Blank	0.000	13:57	

---

Data Download Date/Time: 12/11/2020 13:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TURO,J,  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: STRICKLAND,D  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 11:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: LEBARRE

DATE/TIME OF TEST: 02/28/2020 14:06:26

---

Diagnostics Check	OK	13:59
Air Blank	0.000	14:00
Control Test	0.077	14:00
Air Blank	0.000	14:01
Subject Sample # 1	0.237	14:01
Breath Volume	1.792	14:01
Air Blank	0.000	14:02
Air Blank	0.000	14:04
Subject Sample # 2	0.228	14:04
Breath Volume	3.234	14:04
Air Blank	0.000	14:05
Control Test	0.076	14:05
Air Blank	0.000	14:06
Diagnostics Check	OK	14:06

---

Data Download Date/Time: 12/11/2020 13:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHMIDT,ANDREW,D  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SCHMIDT,ANDREW  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: SKEENS

DATE/TIME OF TEST: 02/29/2020 04:20:43

---

Diagnostics Check	OK	04:17	
Air Blank	0.000	04:18	
Control Test	0.078	04:18	
Air Blank	0.000	04:18	
Subject Sample	# 1	04:19	Subject Test Refused
Breath Volume	0.000	04:19	
Air Blank	0.000	04:19	
Control Test	0.078	04:20	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:20	

---

Data Download Date/Time: 12/11/2020 13:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVENPORT,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: DAVENPORT,JOSHUA  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: HOLOVACKO

DATE/TIME OF TEST: 02/29/2020 04:49:19

---

Diagnostics Check		OK	04:42
Air Blank		0.000	04:42
Control Test		0.078	04:43
Air Blank		0.000	04:43
Subject Sample	# 1	0.138	04:44
Breath Volume		1.707	04:44
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.144	04:47
Breath Volume		1.574	04:47
Air Blank		0.000	04:48
Control Test		0.077	04:48
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

---

Data Download Date/Time: 12/11/2020 13:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITCHER,DARRYL,P  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: BINDER,BRONSON  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 02/14/2020 00:43:50

---

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:41	
Control Test	0.079	00:41	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:42	Subject Test Refused
Breath Volume	0.000	00:42	
Air Blank	0.000	00:42	
Control Test	0.079	00:43	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

---

Data Download Date/Time: 02/21/2020 02:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITCHER,DARRYL,P  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: PITCHER,DARRYL  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: THEISEN

DATE/TIME OF TEST: 02/15/2020 00:17:53

---

Diagnostics Check	OK	00:05	
Air Blank	0.000	00:06	
Control Test	0.080	00:06	
Air Blank	0.000	00:07	
Subject Sample # 1	0.211	00:10	Volume Not Met
Breath Volume	1.046	00:10	
Air Blank	0.000	00:10	
Air Blank	0.000	00:12	
Subject Sample # 2	0.225	00:15	Slope Not Level
Breath Volume	1.289	00:15	
Air Blank	0.000	00:16	
Control Test	0.080	00:16	
Air Blank	0.000	00:17	
Diagnostics Check	Fail	00:17	Analytical Stability Test
Air Blank	0.000	00:17	

---

Data Download Date/Time: 02/21/2020 02:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WALSH,SHANNON,L

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: WALSH,SHANNON

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 01/19/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30219080A3

DRY GAS EXPIRY DATE: 01/05/2022

VIOLATION CODE: DUI

ARREST TIME: 21:58

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: WINDER

DATE/TIME OF TEST: 02/04/2020 22:46:27

---

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.078	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.119	22:41
Breath Volume	1.683	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.103	22:45
Breath Volume	1.562	22:45
Air Blank	0.000	22:45
Control Test	0.079	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

---

Data Download Date/Time: 02/25/2020 00:17



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WALSH,SHANNON,L

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: WALSH,SHANNON

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 01/19/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30219080A3

DRY GAS EXPIRY DATE: 01/05/2022

VIOLATION CODE: DUI

ARREST TIME: 19:27

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 02/16/2020 20:12:36

---

Diagnostics Check	OK	20:05
Air Blank	0.000	20:05
Control Test	0.080	20:06
Air Blank	0.000	20:06
Subject Sample # 1	0.287	20:07
Breath Volume	1.968	20:07
Air Blank	0.000	20:08
Air Blank	0.000	20:09
Subject Sample # 2	0.267	20:11
Breath Volume	2.492	20:11
Air Blank	0.000	20:11
Control Test	0.080	20:12
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

---

Data Download Date/Time: 02/25/2020 00:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,JONATHAN,B  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: RICHARD,LATASHA  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 01/19/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 02/23/2020 02:55:10

---

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.081	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.178	02:50
Breath Volume	1.253	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.163	02:53
Breath Volume	1.804	02:53
Air Blank	0.000	02:54
Control Test	0.079	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

---

Data Download Date/Time: 02/25/2020 00:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOMMERSDORF,JOHN,G  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: DAGOSTINO,STEFANO  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: MORELLI

DATE/TIME OF TEST: 02/26/2020 00:07:56

---

Diagnostics Check	OK	00:05	
Air Blank	0.000	00:05	
Control Test	0.081	00:05	
Air Blank	0.000	00:06	
Subject Sample	# 1	00:06	Subject Test Refused
Breath Volume	0.000	00:06	
Air Blank	0.000	00:07	
Control Test	0.081	00:07	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:07	

---

Data Download Date/Time: 03/19/2020 00:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: SIMS,T  
ARRESTING OFFICER AGENCY: TAMPA PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 02/02/2020 01:41:16

---

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:33	
Control Test	0.079	01:33	
Air Blank	0.000	01:34	
Subject Sample	# 1	01:37	No Sample Provided
Breath Volume	0.000	01:37	
Air Blank	0.000	01:37	
Air Blank	0.000	01:39	
Subject Sample	# 2	01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:40	
Control Test	0.080	01:40	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

---

Data Download Date/Time: 02/05/2020 18:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: TAMPA PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: STRATTON

DATE/TIME OF TEST: 02/02/2020 02:26:49

---

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.062	02:22
Breath Volume	2.238	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:24
Subject Sample # 2	0.062	02:25
Breath Volume	1.453	02:25
Air Blank	0.000	02:25
Control Test	0.080	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

---

Data Download Date/Time: 02/05/2020 18:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: BURKE,B  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PRYSHCHEP

DATE/TIME OF TEST: 02/02/2020 04:28:06

---

Diagnostics Check		OK	04:16	
Air Blank		0.000	04:16	
Control Test		0.080	04:17	
Air Blank		0.000	04:17	
Subject Sample	# 1	0.097	04:20	Volume Not Met
Breath Volume		0.351	04:20	
Air Blank		0.000	04:21	
Air Blank		0.000	04:23	
Subject Sample	# 2	0.116	04:23	
Breath Volume		1.441	04:23	
Air Blank		0.000	04:24	
Air Blank		0.000	04:26	
Subject Sample	# 3	0.117	04:26	
Breath Volume		1.785	04:26	
Air Blank		0.000	04:27	
Control Test		0.079	04:27	
Air Blank		0.000	04:28	
Diagnostics Check		OK	04:28	

---

Data Download Date/Time: 02/05/2020 18:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: HALL,K  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 02/02/2020 04:47:57

---

Diagnostics Check	OK	04:35	
Air Blank	0.000	04:35	
Control Test	0.079	04:35	
Air Blank	0.000	04:36	
Subject Sample	# 1	04:39	No Sample Provided
Breath Volume	0.000	04:39	
Air Blank	0.000	04:40	
Air Blank	0.000	04:42	
Subject Sample	# 2	0.154	04:42
Breath Volume	3.582	04:42	
Air Blank	0.000	04:43	
Air Blank	0.000	04:45	
Subject Sample	# 3	0.159	04:46
Breath Volume	4.308	04:46	
Air Blank	0.000	04:47	
Control Test	0.079	04:47	
Air Blank	0.000	04:47	
Diagnostics Check	OK	04:47	

---

Data Download Date/Time: 02/05/2020 18:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: RIVERA,M  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: HOLCOMB

DATE/TIME OF TEST: 02/02/2020 21:03:01

---

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.080	20:57
Air Blank	0.000	20:57
Subject Sample # 1	0.184	20:58
Breath Volume	1.867	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:00
Subject Sample # 2	0.196	21:01
Breath Volume	1.832	21:01
Air Blank	0.000	21:02
Control Test	0.080	21:02
Air Blank	0.000	21:02
Diagnostics Check	OK	21:03

---

Data Download Date/Time: 02/05/2020 18:41



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: SULAIMAN,R  
ARRESTING OFFICER AGENCY: FHP TROOP C  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: RODRIGUEZ RIVERA

DATE/TIME OF TEST: 02/03/2020 22:26:33

---

Diagnostics Check	OK	22:22	
Air Blank	0.000	22:23	
Control Test	0.081	22:23	
Air Blank	0.000	22:24	
Subject Sample # 1	0.197	22:25	
Breath Volume	1.621	22:25	
Air Blank		22:25	RFI Detect
Air Blank	0.000	22:25	
Control Test	0.079	22:25	
Air Blank	0.000	22:26	
Diagnostics Check	OK	22:26	

---

Data Download Date/Time: 02/05/2020 18:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: SULAIMAN,R  
ARRESTING OFFICER AGENCY: FHP TROOP C  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: RODRIGUEZ RIVERA

DATE/TIME OF TEST: 02/03/2020 22:34:45

---

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.080	22:28
Air Blank	0.000	22:28
Subject Sample # 1	0.206	22:30
Breath Volume	1.761	22:30
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.207	22:33
Breath Volume	2.105	22:33
Air Blank	0.000	22:33
Control Test	0.080	22:34
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

---

Data Download Date/Time: 02/05/2020 18:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: SULAIMAN,R  
ARRESTING OFFICER AGENCY: FHP TROOP C  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: DUBOIS

DATE/TIME OF TEST: 02/04/2020 03:55:38

---

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.081	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.023	03:51
Breath Volume	1.347	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.022	03:54
Breath Volume	2.007	03:54
Air Blank	0.000	03:54
Control Test	0.080	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

---

Data Download Date/Time: 02/05/2020 18:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: TAMPA PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: WRENN

DATE/TIME OF TEST: 02/04/2020 21:31:15

---

Diagnostics Check	OK	21:22	
Air Blank	0.000	21:23	
Control Test	0.079	21:23	
Air Blank	0.000	21:24	
Subject Sample	# 1	21:27	No Sample Provided
Breath Volume	0.000	21:27	
Air Blank	0.000	21:27	
Air Blank	0.000	21:29	
Subject Sample	# 2	21:29	Subject Test Refused
Breath Volume	0.000	21:29	
Air Blank	0.000	21:30	
Control Test	0.080	21:30	
Air Blank	0.000	21:31	
Diagnostics Check	OK	21:31	

---

Data Download Date/Time: 02/05/2020 18:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: VISSER,A  
ARRESTING OFFICER AGENCY: TAMPA PD  
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 02/05/2020 00:49:07

---

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.079	00:42
Air Blank		0.000	00:43
Subject Sample	# 1	0.200	00:44
Breath Volume		1.382	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.195	00:47
Breath Volume		1.417	00:47
Air Blank		0.000	00:48
Control Test		0.078	00:48
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

---

Data Download Date/Time: 02/05/2020 18:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: CAPLAN,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GASCA

DATE/TIME OF TEST: 02/19/2020 01:30:58

---

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:22	
Control Test	0.077	01:23	
Air Blank	0.000	01:23	
Subject Sample	# 1 0.135	01:26	Volume Not Met
Breath Volume	0.507	01:26	
Air Blank	0.000	01:27	
Air Blank	0.000	01:29	
Subject Sample	# 2	01:29	Subject Test Refused
Breath Volume	0.000	01:29	
Air Blank	0.000	01:30	
Control Test	0.075	01:30	
Air Blank	0.000	01:30	
Diagnostics Check	OK	01:30	

---

Data Download Date/Time: 03/06/2020 15:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRIGLIA,NICHOLAS,S  
BREATH TEST OPERATOR AGENCY: Stuart PD  
ARRESTING OFFICER: BRIGLIA,NICHOLAS  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 16:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 02/19/2020 17:33:23

---

Diagnostics Check	OK	17:25	
Air Blank	0.000	17:25	
Control Test	0.077	17:26	
Air Blank	0.000	17:26	
Subject Sample	# 1 0.187	17:28	
Breath Volume	2.019	17:28	
Air Blank	0.000	17:29	
Air Blank	0.000	17:31	
Subject Sample	# 2 0.181	17:31	
Breath Volume	1.742	17:31	
Air Blank	0.000	17:32	
Control Test	0.073	17:32	Control Outside Tolerance
Air Blank	0.000	17:33	

---

Data Download Date/Time: 03/06/2020 15:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRIGLIA,NICHOLAS,S  
BREATH TEST OPERATOR AGENCY: Stuart PD  
ARRESTING OFFICER: BRIGLIA,NICHOLAS  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 16:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 02/19/2020 17:43:32

---

Diagnostics Check	OK	17:35
Air Blank	0.000	17:35
Control Test	0.076	17:36
Air Blank	0.000	17:36
Subject Sample # 1	0.200	17:38
Breath Volume	2.449	17:38
Air Blank	0.000	17:39
Air Blank	0.000	17:41
Subject Sample # 2	0.200	17:41
Breath Volume	2.402	17:41
Air Blank	0.000	17:42
Control Test	0.075	17:42
Air Blank	0.000	17:43
Diagnostics Check	OK	17:43

---

Data Download Date/Time: 03/06/2020 15:32



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: CAPLAN,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: BUSHBY

DATE/TIME OF TEST: 02/21/2020 23:53:30

---

Diagnostics Check	OK	23:46
Air Blank	0.000	23:47
Control Test	0.077	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.146	23:48
Breath Volume	2.621	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.140	23:51
Breath Volume	2.171	23:51
Air Blank	0.000	23:52
Control Test	0.076	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

---

Data Download Date/Time: 03/06/2020 15:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAPLAN,CHRISOPHR,M  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: CAPLAN,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ROWLING

DATE/TIME OF TEST: 02/22/2020 22:02:10

---

Diagnostics Check		OK	21:52
Air Blank		0.000	21:52
Control Test		0.077	21:52
Air Blank		0.000	21:53
Subject Sample	# 1	0.204	21:54
Breath Volume		1.425	21:54
Air Blank		0.000	21:54
Air Blank		0.000	21:56
Subject Sample	# 2	0.183	21:57
Breath Volume		1.121	21:57
Air Blank		0.000	21:58
Air Blank		0.000	21:59
Subject Sample	# 3	0.184	22:00
Breath Volume		1.371	22:00
Air Blank		0.000	22:01
Control Test		0.077	22:01
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

---

Data Download Date/Time: 03/06/2020 15:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORRIS,KEVIN,E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: MORRIS,KEVIN  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:26

SUBJECT'S LAST NAME: LOPEZ-LOPEZ

DATE/TIME OF TEST: 02/23/2020 06:53:41

---

Diagnostics Check	OK	06:47
Air Blank	0.000	06:47
Control Test	0.077	06:47
Air Blank	0.000	06:48
Subject Sample # 1	0.137	06:48
Breath Volume	3.648	06:48
Air Blank	0.000	06:49
Air Blank	0.000	06:51
Subject Sample # 2	0.130	06:51
Breath Volume	3.300	06:51
Air Blank	0.000	06:52
Control Test	0.075	06:53
Air Blank	0.000	06:53
Diagnostics Check	OK	06:53

---

Data Download Date/Time: 03/06/2020 15:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OSINGA,KYLE,L  
BREATH TEST OPERATOR AGENCY: Stuart PD  
ARRESTING OFFICER: OSINGA,KYLE  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 16:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 02/23/2020 17:39:55

---

Diagnostics Check	OK	17:32	
Air Blank	0.000	17:33	
Control Test	0.076	17:33	
Air Blank	0.000	17:34	
Subject Sample	# 1 0.202	17:34	
Breath Volume	2.472	17:34	
Air Blank	0.000	17:35	
Air Blank	0.000	17:37	
Subject Sample	# 2 0.195	17:38	
Breath Volume	2.695	17:38	
Air Blank	0.000	17:38	
Control Test	0.074	17:39	Control Outside Tolerance
Air Blank		17:39	Ambient Fail

---

Data Download Date/Time: 03/06/2020 15:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUBOIS,BROOKS,E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: DUBOIS,BROOKS  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: ANDERSEN

DATE/TIME OF TEST: 02/24/2020 19:31:09

---

Diagnostics Check	OK	19:19	
Air Blank	0.000	19:20	
Control Test	0.077	19:20	
Air Blank	0.000	19:21	
Subject Sample	# 1	19:24	No Sample Provided
Breath Volume	0.000	19:24	
Air Blank	0.000	19:24	
Air Blank	0.000	19:26	
Subject Sample	# 2	19:29	No Sample Provided
Breath Volume	0.000	19:29	
Air Blank	0.000	19:30	
Control Test	0.078	19:30	Interferent Detect
Air Blank	0.000	19:31	

---

Data Download Date/Time: 03/06/2020 15:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER,BRADFORD,B  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: BAKER,BRADFORD  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: GROSSMAN

DATE/TIME OF TEST: 02/25/2020 03:56:38

---

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:45	
Control Test	0.078	03:45	
Air Blank	0.000	03:46	
Subject Sample	# 1 0.154	03:46	
Breath Volume	2.589	03:46	
Air Blank	0.000	03:47	
Air Blank	0.000	03:49	
Subject Sample	# 2	03:52	No Sample Provided
Breath Volume	0.000	03:52	
Air Blank	0.000	03:53	
Air Blank	0.000	03:55	
Subject Sample	# 3	03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:55	
Control Test	0.077	03:56	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

---

Data Download Date/Time: 03/06/2020 15:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUBOIS,BROOKS,E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: DUBOIS,BROOKS  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 16:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: HAYSLIP

DATE/TIME OF TEST: 02/26/2020 17:18:44

---

Diagnostics Check	OK	17:14	
Air Blank	0.000	17:15	
Control Test	0.076	17:15	
Air Blank	0.000	17:16	
Subject Sample	# 1	17:17	Interferent Detect
Breath Volume	3.394	17:17	
Air Blank	0.000	17:17	
Control Test	0.075	17:18	Interferent Detect
Air Blank	0.000	17:18	

---

Data Download Date/Time: 03/06/2020 15:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUBOIS,BROOKS,E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: DUBOIS,BROOKS  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 16:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: HAYSLIP

DATE/TIME OF TEST: 02/26/2020 17:28:00

---

Diagnostics Check	OK	17:20
Air Blank	0.000	17:21
Control Test	0.076	17:21
Air Blank	0.000	17:22
Subject Sample # 1	0.314	17:23
Breath Volume	2.765	17:23
Air Blank	0.000	17:23
Air Blank	0.000	17:25
Subject Sample # 2	0.314	17:26
Breath Volume	2.824	17:26
Air Blank	0.000	17:27
Control Test	0.076	17:27
Air Blank	0.000	17:27
Diagnostics Check	OK	17:28

---

Data Download Date/Time: 03/06/2020 15:32



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: CAPLAN,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: GAAR

DATE/TIME OF TEST: 02/27/2020 01:45:54

---

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.078	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.166	01:41
Breath Volume	1.863	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.162	01:44
Breath Volume	1.425	01:44
Air Blank	0.000	01:44
Control Test	0.077	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

---

Data Download Date/Time: 03/06/2020 15:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORRIS,KEVIN,E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: MORRIS,KEVIN  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 08:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:12

SUBJECT'S LAST NAME: KLINO

DATE/TIME OF TEST: 02/27/2020 09:45:24

---

Diagnostics Check	OK	09:40	
Air Blank	0.000	09:40	
Control Test	0.077	09:41	
Air Blank	0.000	09:41	
Subject Sample	# 1	09:44	Subject Test Refused
Breath Volume	0.000	09:44	
Air Blank	0.000	09:44	
Control Test	0.077	09:44	
Air Blank	0.000	09:45	
Diagnostics Check	OK	09:45	

---

Data Download Date/Time: 03/06/2020 15:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSKO,DAVID,J  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: ROSKO,DAVID  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 07:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:32

SUBJECT'S LAST NAME: KIRCHNER

DATE/TIME OF TEST: 02/28/2020 07:55:12

---

Diagnostics Check	OK	07:54	
Air Blank		07:54	RFI Detect
Air Blank	0.000	07:55	

---

Data Download Date/Time: 03/06/2020 15:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSKO,DAVID,J  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: ROSKO,DAVID  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 07:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:32

SUBJECT'S LAST NAME: KIRCHNER

DATE/TIME OF TEST: 02/28/2020 07:58:40

---

Diagnostics Check	OK	07:56	
Air Blank	0.000	07:57	
Control Test	0.078	07:57	
Air Blank		07:58	Ambient Fail
Air Blank	0.000	07:58	

---

Data Download Date/Time: 03/06/2020 15:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSKO,DAVID,J  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: ROSKO,DAVID  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 07:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:32

SUBJECT'S LAST NAME: KIRCHNER

DATE/TIME OF TEST: 02/28/2020 08:03:48

---

Diagnostics Check	OK	08:00	
Air Blank	0.000	08:00	
Control Test	0.078	08:01	
Air Blank	0.000	08:01	
Subject Sample	# 1	08:02	Interferent Detect
Breath Volume	1.785	08:02	
Air Blank	0.000	08:02	
Control Test	0.077	08:03	
Air Blank	0.000	08:03	
Diagnostics Check	OK	08:03	

---

Data Download Date/Time: 03/06/2020 15:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSKO,DAVID,J  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: ROSKO,DAVID  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 07:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:32

SUBJECT'S LAST NAME: KIRCHNER

DATE/TIME OF TEST: 02/28/2020 08:12:51

---

Diagnostics Check	OK	08:06
Air Blank	0.000	08:06
Control Test	0.077	08:06
Air Blank	0.000	08:07
Subject Sample # 1	0.000	08:08
Breath Volume	2.394	08:08
Air Blank	0.000	08:08
Air Blank	0.000	08:10
Subject Sample # 2	0.000	08:11
Breath Volume	2.000	08:11
Air Blank	0.000	08:11
Control Test	0.077	08:12
Air Blank	0.000	08:12
Diagnostics Check	OK	08:12

---

Data Download Date/Time: 03/06/2020 15:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRADY,REBECCA,A  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: BRADY,REBECCA  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:18

SUBJECT'S LAST NAME: JACINTO

DATE/TIME OF TEST: 02/29/2020 07:48:12

---

Diagnostics Check	OK	07:41
Air Blank	0.000	07:41
Control Test	0.077	07:41
Air Blank	0.000	07:42
Subject Sample # 1	0.155	07:43
Breath Volume	3.363	07:43
Air Blank	0.000	07:44
Air Blank	0.000	07:45
Subject Sample # 2	0.152	07:46
Breath Volume	3.800	07:46
Air Blank	0.000	07:47
Control Test	0.076	07:47
Air Blank	0.000	07:48
Diagnostics Check	OK	07:48

---

Data Download Date/Time: 03/06/2020 15:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: CAPLAN,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 01/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: PAGUADA MEZA

DATE/TIME OF TEST: 02/05/2020 01:31:11

---

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.078	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.000	01:26
Breath Volume	4.296	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.000	01:29
Breath Volume	4.550	01:29
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

---

Data Download Date/Time: 02/07/2020 18:15



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: CAPLAN,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 01/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: LOPEZ VICENTE

DATE/TIME OF TEST: 02/06/2020 02:03:28

---

Diagnostics Check		OK	01:53	
Air Blank		0.000	01:54	
Control Test		0.079	01:54	
Air Blank		0.000	01:54	
Subject Sample	# 1	0.167	01:58	Volume Not Met
Breath Volume		0.132	01:58	
Air Blank		0.000	01:58	
Air Blank		0.000	02:00	
Subject Sample	# 2		02:02	Subject Test Refused
Breath Volume		0.000	02:02	
Air Blank		0.000	02:02	
Control Test		0.078	02:02	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:03	

---

Data Download Date/Time: 02/07/2020 18:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: CAPLAN,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: WHEATON

DATE/TIME OF TEST: 02/08/2020 00:00:55

---

Diagnostics Check	OK	23:49	
Air Blank	0.000	23:50	
Control Test	0.079	23:50	
Air Blank	0.000	23:51	
Subject Sample	# 1 0.148	23:54	Volume Not Met
Breath Volume	0.871	23:54	
Air Blank	0.000	23:54	
Air Blank	0.000	23:56	
Subject Sample	# 2	23:59	Subject Test Refused
Breath Volume	1.289	23:59	
Air Blank	0.000	23:59	
Control Test	0.078	00:00	
Air Blank	0.000	00:00	
Diagnostics Check	OK	00:00	

---

Data Download Date/Time: 03/27/2020 15:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER,BRADFORD,B  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: BAKER,BRADFORD  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 02/08/2020 03:45:49

---

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:36	
Control Test	0.078	03:37	
Air Blank	0.000	03:37	
Subject Sample	# 1 0.170	03:39	
Breath Volume	1.132	03:39	
Air Blank	0.000	03:40	
Air Blank	0.000	03:41	
Subject Sample	# 2 0.148	03:44	
Breath Volume	2.152	03:44	
Air Blank		03:45	Purge Fail
Air Blank	0.000	03:45	

---

Data Download Date/Time: 03/27/2020 15:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER,BRADFORD,B  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: BAKER,BRADFORD  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 02/08/2020 04:01:33

---

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.077	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:54	No Sample Provided
Breath Volume	0.000	03:54	
Air Blank	0.000	03:55	
Air Blank	0.000	03:57	
Subject Sample	# 2	04:00	No Sample Provided
Breath Volume	0.000	04:00	
Air Blank	0.000	04:00	
Control Test	0.078	04:00	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

---

Data Download Date/Time: 03/27/2020 15:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HEITFELD,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Stuart PD  
ARRESTING OFFICER: HEITFELD,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 15:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: DAWES III

DATE/TIME OF TEST: 02/08/2020 17:41:32

---

Diagnostics Check	OK	17:33
Air Blank	0.000	17:33
Control Test	0.079	17:33
Air Blank	0.000	17:34
Subject Sample # 1	0.326	17:36
Breath Volume	2.140	17:36
Air Blank	0.000	17:37
Air Blank	0.000	17:39
Subject Sample # 2	0.334	17:39
Breath Volume	1.984	17:39
Air Blank	0.000	17:40
Control Test	0.075	17:40
Air Blank	0.000	17:41
Diagnostics Check	OK	17:41

---

Data Download Date/Time: 03/27/2020 15:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,WILLIE,V  
BREATH TEST OPERATOR AGENCY: FHP Troop K  
ARRESTING OFFICER: ANDRE,GAETGENS  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/09/2020 07:47:25

---

Diagnostics Check	OK	07:45	
Air Blank	0.000	07:46	
Control Test	0.079	07:46	
Air Blank		07:46	Improper Sample
Air Blank	0.000	07:47	

---

Data Download Date/Time: 03/27/2020 15:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,WILLIE,V  
BREATH TEST OPERATOR AGENCY: FHP Troop K  
ARRESTING OFFICER: ANDRE,GAETGENS  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/09/2020 07:56:33

---

Diagnostics Check	OK	07:50
Air Blank	0.000	07:50
Control Test	0.078	07:51
Air Blank	0.000	07:51
Subject Sample # 1	0.107	07:52
Breath Volume	3.199	07:52
Air Blank	0.000	07:52
Air Blank	0.000	07:54
Subject Sample # 2	0.105	07:54
Breath Volume	2.894	07:54
Air Blank	0.000	07:55
Control Test	0.077	07:55
Air Blank	0.000	07:56
Diagnostics Check	OK	07:56

---

Data Download Date/Time: 03/27/2020 15:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLMES,DARIUS,J  
BREATH TEST OPERATOR AGENCY: Stuart PD  
ARRESTING OFFICER: HOLMES,DARIUS  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: DAWES

DATE/TIME OF TEST: 02/09/2020 20:03:16

---

Diagnostics Check	OK	19:57
Air Blank	0.000	19:57
Control Test	0.079	19:57
Air Blank	0.000	19:58
Subject Sample # 1	0.354	19:58
Breath Volume	2.136	19:58
Air Blank	0.000	19:59
Air Blank	0.000	20:01
Subject Sample # 2	0.335	20:01
Breath Volume	1.421	20:01
Air Blank	0.000	20:02
Control Test	0.078	20:02
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

---

Data Download Date/Time: 03/27/2020 15:37



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLASZYK,JACOB,A  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: BLASZYK,JACOB  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05218080A4  
DRY GAS EXPIRY DATE: 05/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: TAYON HERRERA

DATE/TIME OF TEST: 02/10/2020 04:07:02

---

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.079	04:00
Air Blank		0.000	04:01
Subject Sample	# 1	0.251	04:02
Breath Volume		1.769	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:04
Subject Sample	# 2	0.252	04:05
Breath Volume		3.781	04:05
Air Blank		0.000	04:06
Control Test		0.079	04:06
Air Blank		0.000	04:06
Diagnostics Check		OK	04:07

---

Data Download Date/Time: 03/27/2020 15:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: CHESSHER,SANDRA  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: LINTON

DATE/TIME OF TEST: 02/08/2020 00:55:44

---

Diagnostics Check		OK	00:48
Air Blank		0.000	00:49
Control Test		0.079	00:49
Air Blank		0.000	00:50
Subject Sample	# 1	0.220	00:50
Breath Volume		1.726	00:50
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.222	00:54
Breath Volume		1.378	00:54
Air Blank		0.000	00:54
Control Test		0.080	00:55
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

---

Data Download Date/Time: 02/21/2020 02:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: PITCHER,DARRYL  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 02/13/2020 01:44:49

---

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.079	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.164	01:40
Breath Volume	1.828	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.158	01:43
Breath Volume	1.582	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

---

Data Download Date/Time: 02/21/2020 02:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: BINDER,BRONSON  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: TRAVIS

DATE/TIME OF TEST: 02/15/2020 01:48:08

---

Diagnostics Check		OK	01:35	
Air Blank		0.000	01:36	
Control Test		0.080	01:36	
Air Blank		0.000	01:37	
Subject Sample	# 1	0.054	01:40	Volume Not Met
Breath Volume		0.570	01:40	
Air Blank		0.000	01:41	
Air Blank		0.000	01:42	
Subject Sample	# 2	0.072	01:43	
Breath Volume		3.328	01:43	
Air Blank		0.000	01:44	
Air Blank		0.000	01:45	
Subject Sample	# 3	0.060	01:46	
Breath Volume		1.207	01:46	
Air Blank		0.000	01:47	
Control Test		0.077	01:47	
Air Blank		0.000	01:48	
Diagnostics Check		OK	01:48	

---

Data Download Date/Time: 02/21/2020 02:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: BINDER,BRONSON  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: UNDERWOOD

DATE/TIME OF TEST: 02/16/2020 02:22:13

---

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.079	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.148	02:17
Breath Volume		2.265	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.151	02:20
Breath Volume		1.949	02:20
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

---

Data Download Date/Time: 02/21/2020 02:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,ETHAN,C  
BREATH TEST OPERATOR AGENCY: Clermont PD  
ARRESTING OFFICER: BELL,ETHAN  
ARRESTING OFFICER AGENCY: Clermont PD  
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: PAQUETTE

DATE/TIME OF TEST: 02/16/2020 03:17:53

---

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.078	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.201	03:12
Breath Volume		1.210	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.217	03:16
Breath Volume		1.648	03:16
Air Blank		0.000	03:16
Control Test		0.080	03:17
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

---

Data Download Date/Time: 02/21/2020 02:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUFFY,J,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: KANE,J  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:48

SUBJECT'S LAST NAME: NARDINI-FREYRE

DATE/TIME OF TEST: 02/01/2020 16:15:17

---

Diagnostics Check		OK	16:09
Air Blank		0.000	16:09
Control Test		0.080	16:09
Air Blank		0.000	16:10
Subject Sample	# 1	0.260	16:10
Breath Volume		2.652	16:10
Air Blank		0.000	16:11
Air Blank		0.000	16:13
Subject Sample	# 2	0.253	16:13
Breath Volume		2.636	16:13
Air Blank		0.000	16:14
Control Test		0.079	16:14
Air Blank		0.000	16:15
Diagnostics Check		OK	16:15

---

Data Download Date/Time: 02/05/2020 08:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,DANNY,W  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: GRAY,DANNY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: PALAZOLO

DATE/TIME OF TEST: 02/01/2020 22:24:07

---

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.081	22:17
Air Blank	0.000	22:17
Subject Sample # 1	0.106	22:18
Breath Volume	2.265	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.102	22:22
Breath Volume	2.800	22:22
Air Blank	0.000	22:23
Control Test	0.081	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

---

Data Download Date/Time: 02/05/2020 08:56



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: AMBRIOSO

DATE/TIME OF TEST: 02/01/2020 23:59:57

---

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.176	23:55
Breath Volume	3.195	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.184	23:58
Breath Volume	3.878	23:58
Air Blank	0.000	23:59
Control Test	0.080	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

---

Data Download Date/Time: 02/05/2020 08:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: REYES DE JESUS

DATE/TIME OF TEST: 02/02/2020 03:47:07

---

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.080	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.125	03:42
Breath Volume		2.484	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.127	03:45
Breath Volume		3.082	03:45
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

---

Data Download Date/Time: 02/05/2020 08:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: LOWE,ALEXANDER  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: MCCORMICK

DATE/TIME OF TEST: 02/06/2020 20:49:41

---

Diagnostics Check	OK	20:42
Air Blank	0.000	20:43
Control Test	0.080	20:43
Air Blank	0.000	20:43
Subject Sample # 1	0.147	20:44
Breath Volume	2.589	20:44
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.147	20:48
Breath Volume	2.781	20:48
Air Blank	0.000	20:48
Control Test	0.080	20:49
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

---

Data Download Date/Time: 03/05/2020 08:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: CATANIA

DATE/TIME OF TEST: 02/07/2020 02:15:26

---

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.161	02:10
Breath Volume	2.710	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.165	02:13
Breath Volume	2.664	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

---

Data Download Date/Time: 03/05/2020 08:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINI,RYAN,H  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: FRONK,MATTHEW  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 09:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:55

SUBJECT'S LAST NAME: WALDRUP

DATE/TIME OF TEST: 02/07/2020 10:23:58

---

Diagnostics Check	OK	10:17
Air Blank	0.000	10:18
Control Test	0.080	10:18
Air Blank	0.000	10:18
Subject Sample # 1	0.000	10:19
Breath Volume	2.863	10:19
Air Blank	0.000	10:20
Air Blank	0.000	10:22
Subject Sample # 2	0.000	10:22
Breath Volume	2.667	10:22
Air Blank	0.000	10:23
Control Test	0.079	10:23
Air Blank	0.000	10:23
Diagnostics Check	OK	10:23

---

Data Download Date/Time: 03/05/2020 08:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINI,RYAN,H  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MARTINI,RYAN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 12:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:11

SUBJECT'S LAST NAME: TOMPKINS

DATE/TIME OF TEST: 02/08/2020 13:40:51

---

Diagnostics Check	OK	13:34
Air Blank	0.000	13:35
Control Test	0.081	13:35
Air Blank	0.000	13:35
Subject Sample # 1	0.099	13:36
Breath Volume	2.539	13:36
Air Blank	0.000	13:37
Air Blank	0.000	13:38
Subject Sample # 2	0.094	13:39
Breath Volume	1.386	13:39
Air Blank	0.000	13:39
Control Test	0.080	13:40
Air Blank	0.000	13:40
Diagnostics Check	OK	13:40

---

Data Download Date/Time: 03/05/2020 08:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUFFY,J,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: CHINKEL,M  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:46

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 02/11/2020 14:21:38

---

Diagnostics Check	OK	14:15
Air Blank	0.000	14:15
Control Test	0.080	14:15
Air Blank	0.000	14:16
Subject Sample # 1	0.131	14:17
Breath Volume	1.761	14:17
Air Blank	0.000	14:17
Air Blank	0.000	14:19
Subject Sample # 2	0.129	14:20
Breath Volume	1.839	14:20
Air Blank	0.000	14:20
Control Test	0.080	14:21
Air Blank	0.000	14:21
Diagnostics Check	OK	14:21

---

Data Download Date/Time: 03/05/2020 08:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 02/14/2020 22:02:31

---

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.189	21:57
Breath Volume	3.589	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.189	22:01
Breath Volume	3.238	22:01
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

---

Data Download Date/Time: 03/05/2020 08:12



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: SHIPPY,DYLAN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: DEWEY

DATE/TIME OF TEST: 02/15/2020 02:56:56

---

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.080	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.173	02:52
Breath Volume	1.703	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.163	02:55
Breath Volume	2.375	02:55
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

---

Data Download Date/Time: 03/05/2020 08:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZARRILLO,N,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: OCHS,D  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: ZIOLKOWSKI

DATE/TIME OF TEST: 02/15/2020 18:59:10

---

Diagnostics Check	OK	18:53
Air Blank	0.000	18:53
Control Test	0.080	18:53
Air Blank	0.000	18:54
Subject Sample # 1	0.000	18:54
Breath Volume	4.937	18:54
Air Blank	0.000	18:55
Air Blank	0.000	18:57
Subject Sample # 2	0.000	18:57
Breath Volume	4.941	18:57
Air Blank	0.000	18:58
Control Test	0.081	18:58
Air Blank	0.000	18:59
Diagnostics Check	OK	18:59

---

Data Download Date/Time: 03/05/2020 08:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 02/17/2020 01:03:12

---

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.079	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.103	00:58
Breath Volume	1.496	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.107	01:01
Breath Volume	1.300	01:01
Air Blank	0.000	01:02
Control Test	0.080	01:02
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

---

Data Download Date/Time: 03/05/2020 08:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: LAGOW

DATE/TIME OF TEST: 02/17/2020 04:50:04

---

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.080	04:44
Air Blank	0.000	04:44
Subject Sample # 1	0.158	04:45
Breath Volume	2.535	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.144	04:48
Breath Volume	2.761	04:48
Air Blank	0.000	04:49
Control Test	0.079	04:49
Air Blank	0.000	04:49
Diagnostics Check	OK	04:50

---

Data Download Date/Time: 03/05/2020 08:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FUGIEL,ADRIAN,M  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: SCHWIGK,GEORGE  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: JAMES JR

DATE/TIME OF TEST: 02/18/2020 22:56:22

---

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.080	22:50
Air Blank	0.000	22:51
Subject Sample # 1	0.000	22:51
Breath Volume	1.648	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.000	22:55
Breath Volume	1.101	22:55
Air Blank	0.000	22:55
Control Test	0.081	22:55
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

---

Data Download Date/Time: 03/05/2020 08:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZARRILLO,N,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: LOWE,A  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SURA

DATE/TIME OF TEST: 02/25/2020 01:40:51

---

Diagnostics Check	OK	01:40	
Air Blank		01:40	RFI Detect
Air Blank	0.000	01:40	

---

Data Download Date/Time: 03/05/2020 08:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZARRILLO,N,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: LOWE,A  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SURA

DATE/TIME OF TEST: 02/25/2020 01:43:36

---

Diagnostics Check	OK	01:43	
Air Blank		01:43	RFI Detect
Air Blank	0.000	01:43	

---

Data Download Date/Time: 03/05/2020 08:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZARRILLO,N,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: LOWE,A  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SURA

DATE/TIME OF TEST: 02/25/2020 01:47:33

---

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:45	
Control Test	0.080	01:46	
Air Blank	0.000	01:46	
Subject Sample	# 1 0.193	01:47	
Breath Volume	2.382	01:47	
Air Blank		01:47	RFI Detect
Air Blank		01:47	RFI Detect
Air Blank		01:47	RFI Detect

---

Data Download Date/Time: 03/05/2020 08:12



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZARRILLO,N,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: LOWE,A  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SURA

DATE/TIME OF TEST: 02/25/2020 01:57:05

---

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.178	01:52
Breath Volume	1.539	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.189	01:55
Breath Volume	1.781	01:55
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

---

Data Download Date/Time: 03/05/2020 08:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FILLAK, MICHAEL, O  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: PRIETO PEREZ, CARLOS  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: SWEET

DATE/TIME OF TEST: 02/27/2020 01:09:16

---

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.080	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.170	01:04
Breath Volume	1.914	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.167	01:07
Breath Volume	2.277	01:07
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

---

Data Download Date/Time: 03/05/2020 08:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSS,CHEYANNE,K  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: ROSS,CHEYANNE  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: CAPIZZI

DATE/TIME OF TEST: 02/27/2020 23:22:41

---

Diagnostics Check	OK	23:17	
Air Blank	0.000	23:17	
Control Test	0.080	23:18	
Air Blank	0.000	23:18	
Subject Sample	# 1	23:21	Subject Test Refused
Breath Volume	0.019	23:21	
Air Blank	0.000	23:21	
Control Test	0.080	23:22	
Air Blank	0.000	23:22	
Diagnostics Check	OK	23:22	

---

Data Download Date/Time: 03/05/2020 08:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,ALFONSO,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: RODRIGUEZ,ALFONSO  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 02/29/2020 00:58:55

---

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.080	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.237	00:54
Breath Volume		1.906	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:56
Subject Sample	# 2	0.235	00:57
Breath Volume		3.019	00:57
Air Blank		0.000	00:58
Control Test		0.080	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

---

Data Download Date/Time: 03/05/2020 08:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: BARKLEY,JEFFREY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: WETMORE

DATE/TIME OF TEST: 02/29/2020 23:45:08

---

Diagnostics Check	OK	23:38
Air Blank	0.000	23:39
Control Test	0.081	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.063	23:40
Breath Volume	4.839	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.063	23:43
Breath Volume	5.425	23:43
Air Blank	0.000	23:44
Control Test	0.081	23:44
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

---

Data Download Date/Time: 03/05/2020 08:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUEHLER,DEP,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BUEHLER,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: DIETER

DATE/TIME OF TEST: 02/05/2020 22:50:10

---

Diagnostics Check	OK	22:43
Air Blank	0.000	22:44
Control Test	0.080	22:44
Air Blank	0.000	22:44
Subject Sample # 1	0.032	22:45
Breath Volume	4.070	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.031	22:48
Breath Volume	3.640	22:48
Air Blank	0.000	22:49
Control Test	0.080	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

---

Data Download Date/Time: 02/26/2020 08:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUEHLER,DEP,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MONTANEZ,OFFICER  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: PENTTILA

DATE/TIME OF TEST: 02/29/2020 21:11:12

---

Diagnostics Check	OK	21:04
Air Blank	0.000	21:04
Control Test	0.081	21:05
Air Blank	0.000	21:05
Subject Sample # 1	0.182	21:06
Breath Volume	5.191	21:06
Air Blank	0.000	21:07
Air Blank	0.000	21:08
Subject Sample # 2	0.175	21:09
Breath Volume	3.664	21:09
Air Blank	0.000	21:10
Control Test	0.081	21:10
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

---

Data Download Date/Time: 03/26/2020 08:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEMPLE,DARRYL,D  
BREATH TEST OPERATOR AGENCY: Calhoun County SO  
ARRESTING OFFICER: RODGERS,WILL  
ARRESTING OFFICER AGENCY: Calhoun County SO  
LAST AGENCY INSPECTION: 01/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/11/2020 18:43:39

---

Diagnostics Check	OK	18:37
Air Blank	0.000	18:38
Control Test	0.083	18:38
Air Blank	0.000	18:39
Subject Sample # 1	0.000	18:39
Breath Volume	1.878	18:39
Air Blank	0.000	18:39
Air Blank	0.000	18:41
Subject Sample # 2	0.000	18:42
Breath Volume	1.898	18:42
Air Blank	0.000	18:42
Control Test	0.082	18:43
Air Blank	0.000	18:43
Diagnostics Check	OK	18:43

---

Data Download Date/Time: 02/12/2020 03:41



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHEETZ,JOHN,W  
BREATH TEST OPERATOR AGENCY: Calhoun County SO  
ARRESTING OFFICER: DILLARD,PAUL  
ARRESTING OFFICER AGENCY: Calhoun County SO  
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/27/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:22

SUBJECT'S LAST NAME: MONTFORD

DATE/TIME OF TEST: 02/21/2020 07:56:25

---

Diagnostics Check		OK	07:49
Air Blank		0.000	07:50
Control Test		0.083	07:50
Air Blank		0.000	07:51
Subject Sample	# 1	0.265	07:51
Breath Volume		2.042	07:51
Air Blank		0.000	07:52
Air Blank		0.000	07:54
Subject Sample	# 2	0.253	07:54
Breath Volume		2.332	07:54
Air Blank		0.000	07:55
Control Test		0.083	07:55
Air Blank		0.000	07:56
Diagnostics Check		OK	07:56

---

Data Download Date/Time: 03/02/2020 03:07