

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,ANDREW,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: MAIZO,BRYAN
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MORUARTE JIMENEZ

DATE/TIME OF TEST: 02/12/2020 03:54:00

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:51	
Control Test	0.081	03:51	
Air Blank	0.000	03:52	
Subject Sample	# 1	03:52	Subject Test Refused
Breath Volume	0.000	03:52	
Air Blank	0.000	03:53	
Control Test	0.081	03:53	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 02/27/2020 09:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,CRISTINO,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: TILLERY,JORDAN
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: KLIEGLE

DATE/TIME OF TEST: 02/20/2020 18:43:46

Diagnostics Check	OK	18:40	
Air Blank	0.000	18:41	
Control Test	0.082	18:41	
Air Blank	0.000	18:41	
Subject Sample	# 1	18:42	Subject Test Refused
Breath Volume	0.000	18:42	
Air Blank	0.000	18:42	
Control Test	0.082	18:43	
Air Blank	0.000	18:43	
Diagnostics Check	OK	18:43	

Data Download Date/Time: 02/27/2020 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKER,ALEXA,C
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: SHERAN,ASHLEY
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: CENDAN

DATE/TIME OF TEST: 02/22/2020 02:47:45

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.082	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.187	02:42
Breath Volume	1.148	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.185	02:46
Breath Volume	1.234	02:46
Air Blank	0.000	02:46
Control Test	0.080	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 02/27/2020 09:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMON, VALENTI,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: RODRIGUEZ, ALEXANDER
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:14

SUBJECT'S LAST NAME: RAMON

DATE/TIME OF TEST: 02/22/2020 06:42:47

Diagnostics Check	OK	06:39	
Air Blank	0.000	06:40	
Control Test	0.081	06:40	
Air Blank	0.000	06:41	
Subject Sample	# 1	06:41	Subject Test Refused
Breath Volume	0.000	06:41	
Air Blank	0.000	06:41	
Control Test	0.082	06:42	
Air Blank	0.000	06:42	
Diagnostics Check	OK	06:42	

Data Download Date/Time: 02/27/2020 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKER,ALEXA,C
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: ROSAS,ANYELINNA
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:03

SUBJECT'S LAST NAME: BUENAVENTURA

DATE/TIME OF TEST: 02/29/2020 08:30:26

Diagnostics Check	OK	08:23
Air Blank	0.000	08:24
Control Test	0.082	08:24
Air Blank	0.000	08:24
Subject Sample # 1	0.060	08:25
Breath Volume	1.421	08:25
Air Blank	0.000	08:26
Air Blank	0.000	08:28
Subject Sample # 2	0.060	08:28
Breath Volume	3.062	08:28
Air Blank	0.000	08:29
Control Test	0.081	08:29
Air Blank	0.000	08:30
Diagnostics Check	OK	08:30

Data Download Date/Time: 03/26/2020 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOCARRAS,PEDRO,A
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: SMITH,DABIR
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 11:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:03

SUBJECT'S LAST NAME: GARZA

DATE/TIME OF TEST: 02/01/2020 12:31:58

Diagnostics Check	OK	12:25
Air Blank	0.000	12:26
Control Test	0.081	12:26
Air Blank	0.000	12:27
Subject Sample # 1	0.297	12:27
Breath Volume	1.625	12:27
Air Blank	0.000	12:28
Air Blank	0.000	12:30
Subject Sample # 2	0.296	12:30
Breath Volume	1.476	12:30
Air Blank	0.000	12:31
Control Test	0.081	12:31
Air Blank	0.000	12:31
Diagnostics Check	OK	12:31

Data Download Date/Time: 02/27/2020 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 02/03/2020 01:30:37

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.081	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.184	01:26
Breath Volume	2.722	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.179	01:29
Breath Volume	3.781	01:29
Air Blank	0.000	01:29
Control Test	0.082	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 02/27/2020 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON,REGINALD,B
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: HORTON,REGINALD
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: GURDJIAN

DATE/TIME OF TEST: 02/07/2020 04:16:56

Diagnostics Check	OK	04:13	
Air Blank	0.000	04:13	
Control Test	0.082	04:14	
Air Blank	0.000	04:14	
Subject Sample	# 1	04:15	Subject Test Refused
Breath Volume	0.000	04:15	
Air Blank	0.000	04:16	
Control Test	0.081	04:16	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 02/27/2020 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN-
ROSQUETE,STEPHANNE,J
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: BROWN-
ROSQUETE,STEPHANNE
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022

VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male

OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: KUCHEN

DATE/TIME OF TEST: 02/08/2020 01:36:04

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.082	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.295	01:31
Breath Volume	2.222	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.290	01:34
Breath Volume	2.027	01:34
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:36

Data Download Date/Time: 02/27/2020 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,FREDERICK,A
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: RIVERA,BRIAN
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 07:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:06

SUBJECT'S LAST NAME: TORRES PEREZ

DATE/TIME OF TEST: 02/09/2020 08:46:06

Diagnostics Check		OK	08:39
Air Blank		0.000	08:39
Control Test		0.081	08:40
Air Blank		0.000	08:40
Subject Sample	# 1	0.200	08:41
Breath Volume		3.082	08:41
Air Blank		0.000	08:41
Air Blank		0.000	08:43
Subject Sample	# 2	0.203	08:44
Breath Volume		3.363	08:44
Air Blank		0.000	08:45
Control Test		0.082	08:45
Air Blank		0.000	08:46
Diagnostics Check		OK	08:46

Data Download Date/Time: 02/27/2020 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELENDEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: MELENDEZ,DANIEL
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: ESCANDELL

DATE/TIME OF TEST: 02/10/2020 03:23:08

Diagnostics Check		OK	03:15
Air Blank		0.000	03:15
Control Test		0.082	03:15
Air Blank		0.000	03:16
Subject Sample	# 1	0.230	03:18
Breath Volume		1.242	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.245	03:21
Breath Volume		1.347	03:21
Air Blank		0.000	03:22
Control Test		0.082	03:22
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 02/27/2020 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELENDEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: MELENDEZ,DANIEL
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: CARVAJAL

DATE/TIME OF TEST: 02/14/2020 02:45:09

Diagnostics Check		OK	02:39
Air Blank		0.000	02:39
Control Test		0.082	02:39
Air Blank		0.000	02:40
Subject Sample	# 1	0.212	02:40
Breath Volume		2.300	02:40
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.211	02:43
Breath Volume		1.953	02:43
Air Blank		0.000	02:44
Control Test		0.081	02:44
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 02/27/2020 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELENDEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: DIONNE,MARTIN
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 02/15/2020 06:54:02

Diagnostics Check	OK	06:51	
Air Blank	0.000	06:51	
Control Test	0.082	06:51	
Air Blank	0.000	06:52	
Subject Sample	# 1	06:52	Subject Test Refused
Breath Volume	0.000	06:52	
Air Blank	0.000	06:53	
Control Test	0.082	06:53	
Air Blank	0.000	06:53	
Diagnostics Check	OK	06:54	

Data Download Date/Time: 02/27/2020 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN-
ROSQUETE,STEPHANNE,J
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: BROWN-
ROSQUETE,STEPHANNE
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022

VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Male

OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: CUELLAR

DATE/TIME OF TEST: 02/16/2020 06:09:07

Diagnostics Check	OK	06:06	
Air Blank	0.000	06:06	
Control Test	0.082	06:07	
Air Blank	0.000	06:07	
Subject Sample	# 1	06:07	Subject Test Refused
Breath Volume	0.000	06:07	
Air Blank	0.000	06:08	
Control Test	0.082	06:08	
Air Blank	0.000	06:09	
Diagnostics Check	OK	06:09	

Data Download Date/Time: 02/27/2020 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTLE,SHELANDRA,P
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: CHATFIELD,LATASHA
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 02/17/2020 18:36:40

Diagnostics Check	OK	18:33	
Air Blank	0.000	18:34	
Control Test	0.081	18:34	
Air Blank	0.000	18:34	
Subject Sample	# 1	18:35	Subject Test Refused
Breath Volume	0.000	18:35	
Air Blank	0.000	18:35	
Control Test	0.081	18:36	
Air Blank	0.000	18:36	
Diagnostics Check	OK	18:36	

Data Download Date/Time: 02/27/2020 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUAREZ, RONALD,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: GONZALEZ, PABLO
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 02/19/2020 19:02:27

Diagnostics Check	OK	18:56
Air Blank	0.000	18:56
Control Test	0.082	18:57
Air Blank	0.000	18:57
Subject Sample # 1	0.243	18:58
Breath Volume	2.746	18:58
Air Blank	0.000	18:58
Air Blank	0.000	19:00
Subject Sample # 2	0.242	19:00
Breath Volume	2.230	19:00
Air Blank	0.000	19:01
Control Test	0.082	19:01
Air Blank	0.000	19:02
Diagnostics Check	OK	19:02

Data Download Date/Time: 02/27/2020 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELENDEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: MELENDEZ,DANIEL
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: MINA

DATE/TIME OF TEST: 02/22/2020 01:52:56

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:48	
Control Test	0.081	01:49	
Air Blank	0.000	01:49	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:52	
Control Test	0.082	01:52	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 02/27/2020 13:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN-
ROSQUETE,STEPHANNE,J
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: BROWN-
ROSQUETE,STEPHANNE
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022

VIOLATION CODE: DUI
ARREST TIME: 05:24
GENDER: Male

OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: JAIME

DATE/TIME OF TEST: 02/22/2020 06:06:10

Diagnostics Check		OK	05:59
Air Blank		0.000	06:00
Control Test		0.082	06:00
Air Blank		0.000	06:01
Subject Sample	# 1	0.197	06:01
Breath Volume		1.351	06:01
Air Blank		0.000	06:02
Air Blank		0.000	06:04
Subject Sample	# 2	0.199	06:04
Breath Volume		1.187	06:04
Air Blank		0.000	06:05
Control Test		0.081	06:05
Air Blank		0.000	06:06
Diagnostics Check		OK	06:06

Data Download Date/Time: 02/27/2020 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELENDEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: RIVERA,BRIAN
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: TSIOTSIAS

DATE/TIME OF TEST: 02/23/2020 03:13:08

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:10	
Control Test	0.082	03:11	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:11	Subject Test Refused
Breath Volume	0.000	03:11	
Air Blank	0.000	03:12	
Control Test	0.082	03:12	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 02/27/2020 13:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUMPHRY,KERRY,R
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: COSNER,STEVE
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 05:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: BOWEN

DATE/TIME OF TEST: 02/27/2020 06:35:52

Diagnostics Check	OK	06:33	
Air Blank	0.000	06:33	
Control Test	0.080	06:33	
Air Blank	0.000	06:34	
Subject Sample	# 1	06:34	Subject Test Refused
Breath Volume	0.000	06:34	
Air Blank	0.000	06:34	
Control Test	0.082	06:35	
Air Blank	0.000	06:35	
Diagnostics Check	OK	06:35	

Data Download Date/Time: 02/27/2020 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DARLING,RENARD,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DARLING,RENARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 02/01/2020 07:20:00

Diagnostics Check	OK	07:12
Air Blank	0.000	07:13
Control Test	0.079	07:13
Air Blank	0.000	07:14
Subject Sample # 1	0.141	07:15
Breath Volume	1.558	07:15
Air Blank	0.000	07:15
Air Blank	0.000	07:17
Subject Sample # 2	0.136	07:18
Breath Volume	2.292	07:18
Air Blank	0.000	07:19
Control Test	0.077	07:19
Air Blank	0.000	07:19
Diagnostics Check	OK	07:20

Data Download Date/Time: 02/26/2020 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELFNER,ISAAC,S
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: PEREZ,EMILO
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: MAZUR

DATE/TIME OF TEST: 02/19/2020 15:31:15

Diagnostics Check		OK	15:25
Air Blank		0.000	15:25
Control Test		0.077	15:26
Air Blank		0.000	15:26
Subject Sample	# 1	0.238	15:26
Breath Volume		2.257	15:26
Air Blank		0.000	15:27
Air Blank		0.000	15:29
Subject Sample	# 2	0.236	15:29
Breath Volume		2.113	15:29
Air Blank		0.000	15:30
Control Test		0.077	15:30
Air Blank		0.000	15:31
Diagnostics Check		OK	15:31

Data Download Date/Time: 02/26/2020 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCWILLIAMS,TERRENCE,
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: MCWILLIAMS,TERRENCE
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: PUTNAM

DATE/TIME OF TEST: 02/23/2020 03:06:18

Diagnostics Check	OK	03:03	
Air Blank	0.000	03:03	
Control Test	0.080	03:04	
Air Blank	0.000	03:04	
Subject Sample	# 1	03:04	Subject Test Refused
Breath Volume	0.000	03:04	
Air Blank	0.000	03:05	
Control Test	0.081	03:05	
Air Blank	0.000	03:06	
Diagnostics Check	OK	03:06	

Data Download Date/Time: 02/26/2020 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORIVAL,PHANEL,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: ESTEVEZ,MIGUEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: DIEU

DATE/TIME OF TEST: 02/28/2020 03:46:05

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.146	03:41
Breath Volume	3.035	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.139	03:44
Breath Volume	3.886	03:44
Air Blank	0.000	03:45
Control Test	0.079	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:46

Data Download Date/Time: 03/25/2020 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: POSADA

DATE/TIME OF TEST: 02/28/2020 04:13:59

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:11	
Control Test	0.080	04:11	
Air Blank	0.000	04:12	
Subject Sample	# 1	04:12	Subject Test Refused
Breath Volume	0.000	04:12	
Air Blank	0.000	04:13	
Control Test	0.080	04:13	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 03/25/2020 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GATTORNO,ALEXANDER,F
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FULLANA,ADAM
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 02/29/2020 10:04:18

Diagnostics Check	OK	09:57
Air Blank	0.000	09:58
Control Test	0.078	09:58
Air Blank	0.000	09:59
Subject Sample # 1	0.176	09:59
Breath Volume	3.414	09:59
Air Blank	0.000	10:00
Air Blank	0.000	10:02
Subject Sample # 2	0.176	10:02
Breath Volume	3.082	10:02
Air Blank	0.000	10:03
Control Test	0.079	10:03
Air Blank	0.000	10:04
Diagnostics Check	OK	10:04

Data Download Date/Time: 03/25/2020 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DARLING,RENARD,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DARLING,RENARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:17

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 02/29/2020 10:15:28

Diagnostics Check	OK	10:09
Air Blank	0.000	10:09
Control Test	0.079	10:09
Air Blank	0.000	10:10
Subject Sample # 1	0.085	10:10
Breath Volume	1.371	10:10
Air Blank	0.000	10:11
Air Blank	0.000	10:13
Subject Sample # 2	0.086	10:13
Breath Volume	2.574	10:13
Air Blank	0.000	10:14
Control Test	0.080	10:14
Air Blank	0.000	10:15
Diagnostics Check	OK	10:15

Data Download Date/Time: 03/25/2020 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CHIRINO

DATE/TIME OF TEST: 02/20/2020 23:38:39

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:36	
Control Test	0.080	23:36	
Air Blank	0.000	23:37	
Subject Sample	# 1	23:37	Subject Test Refused
Breath Volume	0.000	23:37	
Air Blank	0.000	23:37	
Control Test	0.079	23:38	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

Data Download Date/Time: 02/26/2020 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,ROBERTO,C
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MENDOZA,ROBERTO
ARRESTING OFFICER AGENCY: MIAMI-DADE SCHOOL PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 02/25/2020 16:29:28

Diagnostics Check	OK	16:26	
Air Blank	0.000	16:26	
Control Test	0.079	16:26	
Air Blank	0.000	16:27	
Subject Sample	# 1	16:28	Subject Test Refused
Breath Volume	0.000	16:28	
Air Blank	0.000	16:28	
Control Test	0.078	16:28	
Air Blank	0.000	16:29	
Diagnostics Check	OK	16:29	

Data Download Date/Time: 02/26/2020 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: CAVENDISH

DATE/TIME OF TEST: 02/29/2020 05:22:43

Diagnostics Check	OK	05:19	
Air Blank	0.000	05:20	
Control Test	0.081	05:20	
Air Blank	0.000	05:21	
Subject Sample	# 1	05:21	Subject Test Refused
Breath Volume	0.000	05:21	
Air Blank	0.000	05:21	
Control Test	0.080	05:22	
Air Blank	0.000	05:22	
Diagnostics Check	OK	05:22	

Data Download Date/Time: 03/25/2020 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CASTILLO,JOSUET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: CUEVA

DATE/TIME OF TEST: 02/01/2020 04:57:07

Diagnostics Check	OK	04:54	
Air Blank	0.000	04:54	
Control Test	0.080	04:54	
Air Blank	0.000	04:55	
Subject Sample	# 1	04:55	Subject Test Refused
Breath Volume	0.000	04:55	
Air Blank	0.000	04:56	
Control Test	0.080	04:56	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:57	

Data Download Date/Time: 02/26/2020 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SAVINON,CARLOS
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: SOLOMON

DATE/TIME OF TEST: 02/04/2020 03:49:03

Diagnostics Check	OK	03:42
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.240	03:44
Breath Volume	1.574	03:44
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.238	03:47
Breath Volume	1.558	03:47
Air Blank	0.000	03:48
Control Test	0.081	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:49

Data Download Date/Time: 02/26/2020 13:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LOUBRIEL,YARIXALYNE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: CARVAJAL

DATE/TIME OF TEST: 02/04/2020 23:31:08

Diagnostics Check	OK	23:28	
Air Blank	0.000	23:28	
Control Test	0.079	23:29	
Air Blank	0.000	23:29	
Subject Sample	# 1	23:29	Subject Test Refused
Breath Volume	0.000	23:29	
Air Blank	0.000	23:30	
Control Test	0.080	23:30	
Air Blank	0.000	23:31	
Diagnostics Check	OK	23:31	

Data Download Date/Time: 02/26/2020 13:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,JOSUET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CASTILLO,JOSUET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 02/07/2020 21:19:20

Diagnostics Check	OK	21:15	
Air Blank	0.000	21:16	
Control Test	0.080	21:16	
Air Blank	0.000	21:17	
Subject Sample	# 1	21:17	Subject Test Refused
Breath Volume	0.000	21:17	
Air Blank	0.000	21:18	
Control Test	0.080	21:18	
Air Blank	0.000	21:19	
Diagnostics Check	OK	21:19	

Data Download Date/Time: 02/26/2020 13:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GUTIERREZ,KRISTINA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: JARQUIN

DATE/TIME OF TEST: 02/22/2020 08:13:32

Diagnostics Check	OK	08:10	
Air Blank	0.000	08:11	
Control Test	0.081	08:11	
Air Blank	0.000	08:11	
Subject Sample	# 1	08:12	Subject Test Refused
Breath Volume	0.000	08:12	
Air Blank	0.000	08:12	
Control Test	0.080	08:12	
Air Blank	0.000	08:13	
Diagnostics Check	OK	08:13	

Data Download Date/Time: 02/26/2020 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SAVINON,CARLOS
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: SALADO RODRIGUEZ

DATE/TIME OF TEST: 02/23/2020 05:01:16

Diagnostics Check	OK	04:52	
Air Blank	0.000	04:52	
Control Test	0.081	04:53	
Air Blank	0.000	04:53	
Subject Sample	# 1 0.129	04:56	Volume Not Met
Breath Volume	0.089	04:56	
Air Blank	0.000	04:57	
Air Blank	0.000	04:59	
Subject Sample	# 2	04:59	Subject Test Refused
Breath Volume	0.000	04:59	
Air Blank	0.000	05:00	
Control Test	0.080	05:00	
Air Blank	0.000	05:01	
Diagnostics Check	OK	05:01	

Data Download Date/Time: 02/26/2020 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEANDRE,GERALD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LEANDRE,GERALD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 02/01/2020 04:47:25

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:39	
Control Test	0.079	04:40	
Air Blank	0.000	04:40	
Subject Sample # 1	0.237	04:42	
Breath Volume	1.339	04:42	
Air Blank	0.000	04:43	
Air Blank	0.000	04:44	
Subject Sample # 2	0.240	04:46	
Breath Volume	1.445	04:46	
Air Blank		04:46	Purge Fail
Air Blank	0.000	04:47	

Data Download Date/Time: 02/26/2020 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEANDRE,GERALD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LEANDRE,GERALD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 02/01/2020 04:54:47

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:51	
Control Test	0.075	04:51	
Air Blank	0.000	04:52	
Subject Sample # 1	0.275	04:53	
Breath Volume	1.355	04:53	
Air Blank		04:54	Purge Fail
Air Blank	0.000	04:54	

Data Download Date/Time: 02/26/2020 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEANDRE,GERALD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LEANDRE,GERALD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 02/01/2020 05:01:55

Diagnostics Check	OK	05:00	
Air Blank	0.000	05:01	
Control Test	0.073	05:01	Control Outside Tolerance
Air Blank	0.000	05:01	

Data Download Date/Time: 02/26/2020 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEANDRE,GERALD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LEANDRE,GERALD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 02/01/2020 05:19:34

Diagnostics Check	OK	05:14	
Air Blank	0.000	05:15	
Control Test	0.077	05:15	
Air Blank	0.000	05:16	
Subject Sample	# 1 0.251	05:17	
Breath Volume	1.136	05:17	
Air Blank		05:18	Purge Fail
Air Blank	0.000	05:19	

Data Download Date/Time: 02/26/2020 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,JOSE,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BONILLA,JOSE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: OCASIO MOLINA

DATE/TIME OF TEST: 02/04/2020 19:29:55

Diagnostics Check	OK	19:29	
Air Blank		19:29	RFI Detect
Air Blank	0.000	19:29	

Data Download Date/Time: 02/26/2020 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,JOSE,I
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BONILLA,JOSE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: OCASIO MOLINA

DATE/TIME OF TEST: 02/04/2020 19:38:01

Diagnostics Check		OK	19:31
Air Blank		0.000	19:32
Control Test		0.077	19:32
Air Blank		0.000	19:33
Subject Sample	# 1	0.283	19:33
Breath Volume		1.703	19:33
Air Blank		0.000	19:34
Air Blank		0.000	19:35
Subject Sample	# 2	0.292	19:36
Breath Volume		2.183	19:36
Air Blank		0.000	19:37
Control Test		0.077	19:37
Air Blank		0.000	19:37
Diagnostics Check		OK	19:38

Data Download Date/Time: 02/26/2020 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,FRANCISCO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MORENO,FRANCISCO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: BASTOS

DATE/TIME OF TEST: 02/11/2020 13:35:56

Diagnostics Check	OK	13:33	
Air Blank	0.000	13:33	
Control Test	0.081	13:33	
Air Blank	0.000	13:34	
Subject Sample	# 1	13:34	Subject Test Refused
Breath Volume	0.000	13:34	
Air Blank	0.000	13:35	
Control Test	0.080	13:35	
Air Blank	0.000	13:35	
Diagnostics Check	OK	13:35	

Data Download Date/Time: 02/26/2020 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,SANDRA,P
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CASTRO,JUAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: A223720673770

DATE/TIME OF TEST: 02/22/2020 21:11:25

Diagnostics Check	OK	21:02	
Air Blank	0.000	21:03	
Control Test	0.081	21:03	
Air Blank	0.000	21:04	
Subject Sample	# 1	21:07	No Sample Provided
Breath Volume	0.000	21:07	
Air Blank	0.000	21:07	
Air Blank	0.000	21:09	
Subject Sample	# 2	21:10	Subject Test Refused
Breath Volume	0.000	21:10	
Air Blank	0.000	21:10	
Control Test	0.081	21:10	
Air Blank	0.000	21:11	
Diagnostics Check	OK	21:11	

Data Download Date/Time: 02/26/2020 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,SANDRA,P
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CASTRO,JUAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: ACOSTA ZAMORA

DATE/TIME OF TEST: 02/22/2020 21:26:21

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:21	
Control Test	0.080	21:21	
Air Blank	0.000	21:21	
Subject Sample	# 1	21:25	Subject Test Refused
Breath Volume	0.000	21:25	
Air Blank	0.000	21:25	
Control Test	0.081	21:25	
Air Blank	0.000	21:26	
Diagnostics Check	OK	21:26	

Data Download Date/Time: 02/26/2020 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,DARREN,B
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DIAZ,DARREN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: RIETTIE

DATE/TIME OF TEST: 02/23/2020 04:00:36

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:55	
Control Test	0.081	03:55	
Air Blank	0.000	03:56	
Subject Sample	# 1 0.174	03:59	Volume Not Met
Breath Volume	0.906	03:59	
Air Blank		03:59	Purge Fail
Air Blank	0.000	04:00	

Data Download Date/Time: 02/26/2020 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,DARREN,B
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DIAZ,DARREN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: RIETTIE

DATE/TIME OF TEST: 02/23/2020 04:09:04

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:03
Subject Sample # 1	0.178	04:04
Breath Volume	1.535	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.179	04:07
Breath Volume	1.269	04:07
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:09

Data Download Date/Time: 02/26/2020 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,DARREN,B
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DIAZ,DARREN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: RIETTIE

DATE/TIME OF TEST: 02/23/2020 04:35:52

Diagnostics Check	OK	04:35	
Air Blank		04:35	RFI Detect
Air Blank		04:35	RFI Detect

Data Download Date/Time: 02/26/2020 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,DARREN,B
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DIAZ,DARREN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: RIETTIE

DATE/TIME OF TEST: 02/23/2020 04:44:38

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.080	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.157	04:39
Breath Volume	1.730	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.160	04:43
Breath Volume	2.023	04:43
Air Blank	0.000	04:43
Control Test	0.079	04:44
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 02/26/2020 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OSTROV,VANESSA,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: HAQ,F
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 01/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 02/03/2020 00:17:19

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:13	
Control Test	0.078	00:13	
Air Blank	0.000	00:14	
Subject Sample	# 1	00:15	Subject Test Refused
Breath Volume	0.000	00:15	
Air Blank	0.000	00:16	
Control Test	0.078	00:16	
Air Blank	0.000	00:17	
Diagnostics Check	OK	00:17	

Data Download Date/Time: 02/15/2020 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: BALDWIN

DATE/TIME OF TEST: 02/01/2020 00:41:29

Diagnostics Check	OK	00:33
Air Blank	0.000	00:34
Control Test	0.079	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.235	00:35
Breath Volume	1.136	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.238	00:39
Breath Volume	1.441	00:39
Air Blank	0.000	00:40
Control Test	0.078	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 02/15/2020 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: BRIGHT

DATE/TIME OF TEST: 02/01/2020 03:59:31

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:57	
Control Test	0.080	03:57	
Air Blank	0.000	03:57	
Subject Sample	# 1	03:58	Subject Test Refused
Breath Volume	0.000	03:58	
Air Blank	0.000	03:58	
Control Test	0.080	03:58	
Air Blank	0.000	03:59	
Diagnostics Check	OK	03:59	

Data Download Date/Time: 02/15/2020 22:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: NADER

DATE/TIME OF TEST: 02/02/2020 02:27:02

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.135	02:22
Breath Volume	2.304	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:24
Subject Sample # 2	0.136	02:25
Breath Volume	1.808	02:25
Air Blank	0.000	02:26
Control Test	0.081	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:27

Data Download Date/Time: 02/15/2020 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRIFFIN,KYLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: JESPERSEN

DATE/TIME OF TEST: 02/03/2020 01:04:14

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:54	
Control Test	0.080	00:54	
Air Blank	0.000	00:54	
Subject Sample	# 1	00:58	No Sample Provided
Breath Volume	0.000	00:58	
Air Blank	0.000	00:58	
Air Blank	0.000	01:00	
Subject Sample	# 2	01:02	Subject Test Refused
Breath Volume	0.000	01:02	
Air Blank	0.000	01:03	
Control Test	0.081	01:03	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

Data Download Date/Time: 02/15/2020 22:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHETEE,THERESA,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DAUGHETEE,THERESA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 02/03/2020 02:47:20

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.081	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.173	02:42
Breath Volume	1.308	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.178	02:45
Breath Volume	1.523	02:45
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 02/15/2020 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTHINGTON,JUSTIN,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WORTHINGTON,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: SCAFIDI

DATE/TIME OF TEST: 02/04/2020 03:26:15

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:22	
Control Test	0.081	03:23	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:24	Subject Test Refused
Breath Volume	0.000	03:24	
Air Blank	0.000	03:25	
Control Test	0.080	03:25	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 02/15/2020 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCINTYRE,KEN
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 06:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:22

SUBJECT'S LAST NAME: SORENSON

DATE/TIME OF TEST: 02/04/2020 07:06:53

Diagnostics Check	OK	07:04	
Air Blank	0.000	07:04	
Control Test	0.080	07:04	
Air Blank	0.000	07:05	
Subject Sample	# 1	07:05	Subject Test Refused
Breath Volume	0.000	07:05	
Air Blank	0.000	07:05	
Control Test	0.081	07:06	
Air Blank	0.000	07:06	
Diagnostics Check	OK	07:06	

Data Download Date/Time: 02/15/2020 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: HARTE

DATE/TIME OF TEST: 02/05/2020 00:00:35

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:55
Subject Sample # 1	0.166	23:55
Breath Volume	2.296	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.166	23:58
Breath Volume	2.351	23:58
Air Blank	0.000	23:59
Control Test	0.080	23:59
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 02/15/2020 22:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 02/06/2020 04:46:24

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.079	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.177	04:41
Breath Volume	1.792	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.180	04:44
Breath Volume	2.343	04:44
Air Blank	0.000	04:45
Control Test	0.080	04:45
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 02/15/2020 22:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON,GREG,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HUFF,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 13:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:48

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 02/06/2020 14:15:16

Diagnostics Check	OK	14:08
Air Blank	0.000	14:09
Control Test	0.080	14:09
Air Blank	0.000	14:09
Subject Sample # 1	0.000	14:10
Breath Volume	1.929	14:10
Air Blank	0.000	14:11
Air Blank	0.000	14:13
Subject Sample # 2	0.000	14:13
Breath Volume	1.308	14:13
Air Blank	0.000	14:14
Control Test	0.080	14:14
Air Blank	0.000	14:15
Diagnostics Check	OK	14:15

Data Download Date/Time: 02/15/2020 22:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 02/08/2020 02:26:27

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.114	02:21
Breath Volume	2.355	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.113	02:24
Breath Volume	1.871	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 02/15/2020 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANWORP,JUDSON,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VANWORP,JUDSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: TAYSER

DATE/TIME OF TEST: 02/08/2020 03:16:07

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.142	03:11
Breath Volume	2.351	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.142	03:14
Breath Volume	1.687	03:14
Air Blank	0.000	03:15
Control Test	0.080	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 02/15/2020 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: ROMARY

DATE/TIME OF TEST: 02/08/2020 04:57:25

Diagnostics Check	OK	04:52	
Air Blank	0.000	04:52	
Control Test	0.080	04:53	
Air Blank	0.000	04:53	
Subject Sample	# 1	04:56	Subject Test Refused
Breath Volume	0.000	04:56	
Air Blank	0.000	04:56	
Control Test	0.081	04:56	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:57	

Data Download Date/Time: 02/15/2020 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RODRIGUEZ,NELSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: VARGAS-MORALES

DATE/TIME OF TEST: 02/08/2020 19:35:49

Diagnostics Check	OK	19:32	
Air Blank	0.000	19:33	
Control Test	0.080	19:33	
Air Blank	0.000	19:34	
Subject Sample	# 1	19:34	Subject Test Refused
Breath Volume	0.000	19:34	
Air Blank	0.000	19:34	
Control Test	0.080	19:35	
Air Blank	0.000	19:35	
Diagnostics Check	OK	19:35	

Data Download Date/Time: 02/15/2020 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: DLUGOZIMA

DATE/TIME OF TEST: 02/09/2020 01:13:24

Diagnostics Check	OK	01:09	
Air Blank	0.000	01:09	
Control Test	0.080	01:09	
Air Blank	0.000	01:10	
Subject Sample	# 1	01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:12	
Control Test	0.081	01:12	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 02/15/2020 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MULLINS

DATE/TIME OF TEST: 02/09/2020 03:52:41

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:41	
Control Test		0.080	03:41	
Air Blank		0.000	03:41	
Subject Sample	# 1	0.155	03:45	Volume Not Met
Breath Volume		0.148	03:45	
Air Blank		0.000	03:45	
Air Blank		0.000	03:47	
Subject Sample	# 2	0.189	03:47	
Breath Volume		1.679	03:47	
Air Blank		0.000	03:48	
Air Blank		0.000	03:50	
Subject Sample	# 3	0.201	03:51	
Breath Volume		1.601	03:51	
Air Blank		0.000	03:51	
Control Test		0.076	03:52	
Air Blank		0.000	03:52	
Diagnostics Check		OK	03:52	

Data Download Date/Time: 02/15/2020 22:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: FUTCH

DATE/TIME OF TEST: 02/09/2020 04:41:07

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.079	04:35
Air Blank	0.000	04:36
Subject Sample # 1	0.073	04:36
Breath Volume	2.062	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.074	04:39
Breath Volume	1.683	04:39
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 02/15/2020 22:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: CROKER

DATE/TIME OF TEST: 02/11/2020 03:52:59

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.081	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.000	03:48
Breath Volume	1.820	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.000	03:51
Breath Volume	2.253	03:51
Air Blank	0.000	03:52
Control Test	0.081	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 02/15/2020 22:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: GIGLIONE

DATE/TIME OF TEST: 02/12/2020 00:02:21

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.081	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.040	23:58
Breath Volume	1.535	23:58
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.039	00:00
Breath Volume	1.593	00:00
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 02/15/2020 22:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: GANT

DATE/TIME OF TEST: 02/03/2020 00:51:37

Diagnostics Check	OK	00:38	
Air Blank	0.000	00:39	
Control Test	0.080	00:39	
Air Blank	0.000	00:39	
Subject Sample	# 1 0.151	00:41	
Breath Volume	3.085	00:41	
Air Blank	0.000	00:42	
Air Blank	0.000	00:43	
Subject Sample	# 2 0.139	00:47	Volume Not Met
Breath Volume	0.714	00:47	
Air Blank	0.000	00:47	
Air Blank	0.000	00:49	
Subject Sample	# 3 0.144	00:50	
Breath Volume	3.035	00:50	
Air Blank	0.000	00:50	
Control Test	0.079	00:51	
Air Blank	0.000	00:51	
Diagnostics Check	OK	00:51	

Data Download Date/Time: 02/24/2020 09:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: BENEDICT

DATE/TIME OF TEST: 02/06/2020 21:57:16

Diagnostics Check	OK	21:51
Air Blank	0.000	21:51
Control Test	0.078	21:51
Air Blank	0.000	21:52
Subject Sample # 1	0.102	21:52
Breath Volume	1.375	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.107	21:55
Breath Volume	1.636	21:55
Air Blank	0.000	21:56
Control Test	0.077	21:56
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

Data Download Date/Time: 02/24/2020 09:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICKEY, WILLIAM, Q
BREATH TEST OPERATOR AGENCY: PORT ST. JOE PD
ARRESTING OFFICER: ALDAY, S
ARRESTING OFFICER AGENCY: GULF COUNTY SO
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: KITE

DATE/TIME OF TEST: 02/08/2020 05:17:18

Diagnostics Check	OK	05:14	
Air Blank	0.000	05:14	
Control Test	0.079	05:15	
Air Blank	0.000	05:15	
Subject Sample	# 1	05:15	Subject Test Refused
Breath Volume	0.000	05:15	
Air Blank	0.000	05:16	
Control Test	0.079	05:16	
Air Blank	0.000	05:17	
Diagnostics Check	OK	05:17	

Data Download Date/Time: 02/09/2020 20:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICKEY, WILLIAM, Q
BREATH TEST OPERATOR AGENCY: PORT ST. JOE PD
ARRESTING OFFICER: UITTS, C.P.
ARRESTING OFFICER AGENCY: PORT ST. JOE PD
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: BRONDUM

DATE/TIME OF TEST: 02/09/2020 00:28:59

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.078	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:27	Subject Test Refused
Breath Volume	0.000	00:27	
Air Blank	0.000	00:28	
Control Test	0.078	00:28	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:28	

Data Download Date/Time: 02/09/2020 20:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,PAUL,G
BREATH TEST OPERATOR AGENCY: GULF COUNTY SO
ARRESTING OFFICER: YOUNG,PAUL
ARRESTING OFFICER AGENCY: GULF COUNTY SO
LAST AGENCY INSPECTION: 02/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 02/23/2020 02:00:58

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.187	01:56
Breath Volume	1.605	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.175	01:59
Breath Volume	1.910	01:59
Air Blank	0.000	02:00
Control Test	0.080	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 03/17/2020 16:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINDHOLM,JOHN,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINDHOLM,JOHN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: SERRANO LOPEZ

DATE/TIME OF TEST: 02/03/2020 02:41:45

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.081	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.190	02:37
Breath Volume	2.222	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.194	02:40
Breath Volume	2.273	02:40
Air Blank	0.000	02:40
Control Test	0.081	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 02/05/2020 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,DANIEL,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DAVIS,DANIEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BEARD

DATE/TIME OF TEST: 02/07/2020 02:01:54

Diagnostics Check		OK	01:54
Air Blank		0.000	01:54
Control Test		0.080	01:55
Air Blank		0.000	01:55
Subject Sample	# 1	0.237	01:57
Breath Volume		3.621	01:57
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.240	02:00
Breath Volume		4.445	02:00
Air Blank		0.000	02:00
Control Test		0.080	02:01
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 03/04/2020 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SCHEY,TYLER
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 02/08/2020 03:26:41

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.080	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.189	03:21
Breath Volume	4.730	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.181	03:25
Breath Volume	3.945	03:25
Air Blank	0.000	03:25
Control Test	0.080	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 03/04/2020 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HARGIS,KYLE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: SANTIAGO VEGA

DATE/TIME OF TEST: 02/09/2020 02:25:42

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.081	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.116	02:20
Breath Volume	2.058	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample # 2	0.121	02:24
Breath Volume	2.179	02:24
Air Blank	0.000	02:24
Control Test	0.080	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 03/04/2020 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: PAHLKE 17045,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: QUINT

DATE/TIME OF TEST: 02/14/2020 21:55:18

Diagnostics Check	OK	21:48
Air Blank	0.000	21:49
Control Test	0.080	21:49
Air Blank	0.000	21:49
Subject Sample # 1	0.183	21:50
Breath Volume	3.097	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:52
Subject Sample # 2	0.184	21:53
Breath Volume	2.585	21:53
Air Blank	0.000	21:54
Control Test	0.081	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 03/04/2020 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 02/15/2020 00:46:37

Diagnostics Check	OK	00:40
Air Blank	0.000	00:40
Control Test	0.080	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.060	00:42
Breath Volume	1.343	00:42
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.060	00:45
Breath Volume	1.214	00:45
Air Blank	0.000	00:45
Control Test	0.080	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 03/04/2020 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: JOHNSTON 582,T
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MUNOZ HERNADEZ

DATE/TIME OF TEST: 02/15/2020 03:07:22

Diagnostics Check		OK	02:58
Air Blank		0.000	02:59
Control Test		0.080	02:59
Air Blank		0.000	02:59
Subject Sample	# 1	0.211	03:02
Breath Volume		3.230	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:04
Subject Sample	# 2	0.210	03:05
Breath Volume		3.324	03:05
Air Blank		0.000	03:06
Control Test		0.079	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 03/04/2020 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MELIDOR 556,B
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 02/16/2020 23:34:40

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.080	23:29
Air Blank	0.000	23:29
Subject Sample # 1	0.069	23:30
Breath Volume	1.683	23:30
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.065	23:33
Breath Volume	2.917	23:33
Air Blank	0.000	23:33
Control Test	0.081	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 03/04/2020 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: JOHNSTON 582,T
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 02/17/2020 02:32:49

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:27	
Control Test	0.080	02:27	
Air Blank	0.000	02:28	
Subject Sample	# 1 0.276	02:28	
Breath Volume	2.242	02:28	
Air Blank	0.000	02:29	
Air Blank	0.000	02:30	
Subject Sample	# 2	02:31	Subject Test Refused
Breath Volume	0.000	02:31	
Air Blank	0.000	02:31	
Control Test	0.079	02:32	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 03/04/2020 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HOYOS 19022,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: HERNANDEZ MARIN

DATE/TIME OF TEST: 02/17/2020 04:43:54

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.080	04:36
Air Blank	0.000	04:36
Subject Sample # 1	0.199	04:39
Breath Volume	2.066	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:41
Subject Sample # 2	0.201	04:42
Breath Volume	1.667	04:42
Air Blank	0.000	04:42
Control Test	0.079	04:43
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 03/04/2020 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: TAVERNA,D
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: COLVIN

DATE/TIME OF TEST: 02/17/2020 20:12:21

Diagnostics Check	OK	20:05
Air Blank	0.000	20:06
Control Test	0.081	20:06
Air Blank	0.000	20:07
Subject Sample # 1	0.145	20:07
Breath Volume	2.984	20:07
Air Blank	0.000	20:08
Air Blank	0.000	20:10
Subject Sample # 2	0.128	20:10
Breath Volume	1.585	20:10
Air Blank	0.000	20:11
Control Test	0.080	20:11
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

Data Download Date/Time: 03/04/2020 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEINER,RYAN,T
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SALTER,E
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/18/2020 02:48:59

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.183	02:44
Breath Volume	2.378	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:46
Subject Sample # 2	0.181	02:47
Breath Volume	2.296	02:47
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 03/04/2020 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINS,KYLE,L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: LEON,BRIAN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 02/18/2020 04:29:32

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:24	
Control Test	0.080	04:24	
Air Blank	0.000	04:25	
Subject Sample	# 1 0.193	04:27	
Breath Volume	2.539	04:27	
Air Blank		04:28	Purge Fail
Air Blank		04:29	Purge Fail

Data Download Date/Time: 03/04/2020 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINS,KYLE,L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: LEON,BRIAN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 02/18/2020 04:37:36

Diagnostics Check		OK	04:30
Air Blank		0.000	04:31
Control Test		0.079	04:31
Air Blank		0.000	04:32
Subject Sample	# 1	0.206	04:32
Breath Volume		2.277	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.199	04:35
Breath Volume		1.980	04:35
Air Blank		0.000	04:36
Control Test		0.079	04:36
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 03/04/2020 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINS,KYLE,L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: LUNA,MICHAEL
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: MARTINEZ ZARATE

DATE/TIME OF TEST: 02/23/2020 05:46:37

Diagnostics Check	OK	05:40
Air Blank	0.000	05:40
Control Test	0.081	05:41
Air Blank	0.000	05:41
Subject Sample # 1	0.166	05:42
Breath Volume	3.621	05:42
Air Blank	0.000	05:42
Air Blank	0.000	05:44
Subject Sample # 2	0.167	05:45
Breath Volume	3.066	05:45
Air Blank	0.000	05:45
Control Test	0.081	05:46
Air Blank	0.000	05:46
Diagnostics Check	OK	05:46

Data Download Date/Time: 03/04/2020 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,KATIE,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KIRKBY,STEPHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 02/23/2020 22:19:23

Diagnostics Check		OK	22:10	
Air Blank		0.000	22:11	
Control Test		0.081	22:11	
Air Blank		0.000	22:12	
Subject Sample	# 1	0.089	22:15	Volume Not Met
Breath Volume		0.187	22:15	
Air Blank		0.000	22:16	
Air Blank		0.000	22:17	
Subject Sample	# 2		22:18	Subject Test Refused
Breath Volume		0.000	22:18	
Air Blank		0.000	22:18	
Control Test		0.079	22:18	
Air Blank		0.000	22:19	
Diagnostics Check		OK	22:19	

Data Download Date/Time: 03/04/2020 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,KATIE,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KIRKBY,STEPHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CRAWFORD JR

DATE/TIME OF TEST: 02/23/2020 22:48:07

Diagnostics Check	OK	22:40
Air Blank	0.000	22:41
Control Test	0.081	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.115	22:43
Breath Volume	1.421	22:43
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.117	22:46
Breath Volume	1.699	22:46
Air Blank	0.000	22:47
Control Test	0.079	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 03/04/2020 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDIN,JEREMY,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SALTERS,ERIC
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: BENNINGHAUS

DATE/TIME OF TEST: 02/24/2020 05:02:46

Diagnostics Check	OK	04:56
Air Blank	0.000	04:57
Control Test	0.080	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.000	04:58
Breath Volume	2.601	04:58
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.000	05:01
Breath Volume	2.847	05:01
Air Blank	0.000	05:01
Control Test	0.079	05:02
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 03/04/2020 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,DANIEL,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DAVIS,D
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: TIETZ

DATE/TIME OF TEST: 02/24/2020 21:00:14

Diagnostics Check	OK	20:53
Air Blank	0.000	20:53
Control Test	0.080	20:54
Air Blank	0.000	20:54
Subject Sample # 1	0.248	20:55
Breath Volume	3.304	20:55
Air Blank	0.000	20:56
Air Blank	0.000	20:57
Subject Sample # 2	0.243	20:58
Breath Volume	3.281	20:58
Air Blank	0.000	20:59
Control Test	0.079	20:59
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time: 03/04/2020 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,KATIE,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HERZFELD,STEPHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: PALKO

DATE/TIME OF TEST: 02/27/2020 02:43:15

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.242	02:38
Breath Volume	1.414	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.246	02:41
Breath Volume	1.804	02:41
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 03/04/2020 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SHOEMAKER 15030,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 02/28/2020 23:13:10

Diagnostics Check		OK	23:00	
Air Blank		0.000	23:01	
Control Test		0.080	23:01	
Air Blank		0.000	23:02	
Subject Sample	# 1	0.160	23:05	Volume Not Met
Breath Volume		0.015	23:05	
Air Blank		0.000	23:05	
Air Blank		0.000	23:07	
Subject Sample	# 2	0.174	23:08	
Breath Volume		1.230	23:08	
Air Blank		0.000	23:08	
Air Blank		0.000	23:10	
Subject Sample	# 3	0.182	23:11	
Breath Volume		2.015	23:11	
Air Blank		0.000	23:12	
Control Test		0.080	23:12	
Air Blank		0.000	23:13	
Diagnostics Check		OK	23:13	

Data Download Date/Time: 03/04/2020 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: PAHLKE 17045,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: JASPERS

DATE/TIME OF TEST: 02/29/2020 00:16:40

Diagnostics Check		OK	00:09
Air Blank		0.000	00:09
Control Test		0.079	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.214	00:11
Breath Volume		4.632	00:11
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.213	00:15
Breath Volume		2.480	00:15
Air Blank		0.000	00:15
Control Test		0.079	00:16
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 03/04/2020 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIED 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: JOHNSTON 582,T
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: LOPEZ ITZEP

DATE/TIME OF TEST: 02/29/2020 03:56:11

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.081	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.128	03:51
Breath Volume		3.539	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:53
Subject Sample	# 2	0.129	03:54
Breath Volume		3.289	03:54
Air Blank		0.000	03:55
Control Test		0.081	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 03/04/2020 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANPELT,J,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: EUBANKS,A
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: ROWLETT

DATE/TIME OF TEST: 02/29/2020 20:20:49

Diagnostics Check	OK	20:14
Air Blank	0.000	20:14
Control Test	0.080	20:15
Air Blank	0.000	20:15
Subject Sample # 1	0.093	20:16
Breath Volume	3.601	20:16
Air Blank	0.000	20:16
Air Blank	0.000	20:18
Subject Sample # 2	0.085	20:19
Breath Volume	2.625	20:19
Air Blank	0.000	20:19
Control Test	0.080	20:20
Air Blank	0.000	20:20
Diagnostics Check	OK	20:20

Data Download Date/Time: 03/04/2020 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANG, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: VILLACRES, NICHOLAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: MAHER

DATE/TIME OF TEST: 02/01/2020 01:24:21

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:13	
Control Test	0.079	01:13	
Air Blank	0.000	01:14	
Subject Sample	# 1	01:17	No Sample Provided
Breath Volume	0.000	01:17	
Air Blank	0.000	01:17	
Air Blank	0.000	01:19	
Subject Sample	# 2	01:22	Volume Not Met
Breath Volume	0.171	01:22	
Air Blank		01:23	Purge Fail
Air Blank		01:24	Purge Fail

Data Download Date/Time: 02/28/2020 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,JASON,M
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: ROSA,EMILIO
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: DURDEN

DATE/TIME OF TEST: 02/08/2020 03:32:14

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.099	03:27
Breath Volume	1.984	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.097	03:30
Breath Volume	1.753	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 02/28/2020 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAOUNI,K,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DIXON,D
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: GAVIRIA BARRERA

DATE/TIME OF TEST: 02/08/2020 04:17:22

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.078	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.129	04:12
Breath Volume	3.445	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.126	04:15
Breath Volume	3.828	04:15
Air Blank	0.000	04:16
Control Test	0.078	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 02/28/2020 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEL DUCA, ANTHONY, D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DEL DUCA, ANTHONY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 02/11/2020 23:06:17

Diagnostics Check	OK	22:57	
Air Blank	0.000	22:57	
Control Test	0.079	22:57	
Air Blank	0.000	22:58	
Subject Sample	# 1 0.321	22:58	
Breath Volume	1.160	22:58	
Air Blank	0.000	22:59	
Air Blank	0.000	23:01	
Subject Sample	# 2 0.276	23:01	
Breath Volume	1.394	23:01	
Air Blank	0.000	23:02	
Air Blank	0.000	23:04	
Subject Sample	# 3 0.318	23:05	
Breath Volume	1.707	23:05	
Air Blank		23:05	Purge Fail
Air Blank	0.000	23:06	

Data Download Date/Time: 02/28/2020 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,J,
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: DERISE,V
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: CHIRCO

DATE/TIME OF TEST: 02/15/2020 20:50:53

Diagnostics Check	OK	20:40
Air Blank	0.000	20:41
Control Test	0.079	20:41
Air Blank	0.000	20:42
Subject Sample # 1	0.195	20:44
Breath Volume	2.007	20:44
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.178	20:49
Breath Volume	2.156	20:49
Air Blank	0.000	20:49
Control Test	0.078	20:50
Air Blank	0.000	20:50
Diagnostics Check	OK	20:50

Data Download Date/Time: 02/28/2020 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WINSTEAD,JUSTIN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: WIECZOREK

DATE/TIME OF TEST: 02/17/2020 03:04:16

Diagnostics Check		OK	02:57
Air Blank		0.000	02:57
Control Test		0.078	02:58
Air Blank		0.000	02:58
Subject Sample	# 1	0.193	02:59
Breath Volume		1.707	02:59
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.187	03:02
Breath Volume		1.449	03:02
Air Blank		0.000	03:03
Control Test		0.077	03:03
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 02/28/2020 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,NICHOLAS,W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: PATTERSON,T
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: MORAGNE

DATE/TIME OF TEST: 02/21/2020 22:40:59

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:37	
Control Test	0.079	22:37	
Air Blank	0.000	22:37	
Subject Sample	# 1	22:39	Subject Test Refused
Breath Volume	0.000	22:39	
Air Blank	0.000	22:40	
Control Test	0.079	22:40	
Air Blank	0.000	22:40	
Diagnostics Check	OK	22:40	

Data Download Date/Time: 02/28/2020 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,JASON,M
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: SANCHEZ,JASON
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: CARMICHAEL

DATE/TIME OF TEST: 02/22/2020 00:18:03

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:12	
Control Test	0.078	00:13	
Air Blank	0.000	00:13	
Subject Sample	# 1 0.313	00:16	
Breath Volume	2.375	00:16	
Air Blank		00:17	Purge Fail
Air Blank		00:18	Purge Fail

Data Download Date/Time: 02/28/2020 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,JASON,M
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: SANCHEZ,JASON
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: CARMICHAEL

DATE/TIME OF TEST: 02/22/2020 00:27:43

Diagnostics Check		OK	00:20
Air Blank		0.000	00:21
Control Test		0.077	00:21
Air Blank		0.000	00:22
Subject Sample	# 1	0.306	00:23
Breath Volume		1.371	00:23
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.312	00:26
Breath Volume		1.117	00:26
Air Blank		0.000	00:26
Control Test		0.076	00:27
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 02/28/2020 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRATT,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PRATT,NICHOLAS
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: MCDONOUGH

DATE/TIME OF TEST: 02/07/2020 00:39:22

Diagnostics Check		OK	00:33
Air Blank		0.000	00:33
Control Test		0.079	00:33
Air Blank		0.000	00:34
Subject Sample	# 1	0.254	00:34
Breath Volume		2.023	00:34
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.261	00:37
Breath Volume		2.531	00:37
Air Blank		0.000	00:38
Control Test		0.079	00:38
Air Blank		0.000	00:39
Diagnostics Check		OK	00:39

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUONPANE,V,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: POUNDS,R
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: RODRIGUEZGONZALEZ

DATE/TIME OF TEST: 02/08/2020 03:15:44

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.079	03:11
Breath Volume	1.574	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.074	03:14
Breath Volume	1.281	03:14
Air Blank	0.000	03:14
Control Test	0.080	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,S,J
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GASH,C
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: SCHAFFER

DATE/TIME OF TEST: 02/08/2020 23:15:45

Diagnostics Check	OK	23:09
Air Blank	0.000	23:09
Control Test	0.080	23:10
Air Blank	0.000	23:10
Subject Sample # 1	0.131	23:11
Breath Volume	1.726	23:11
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.129	23:14
Breath Volume	2.140	23:14
Air Blank	0.000	23:14
Control Test	0.079	23:15
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,THOMAS,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: TOBLER,JARAH
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: EDMUNDSON

DATE/TIME OF TEST: 02/11/2020 23:20:31

Diagnostics Check	OK	23:13
Air Blank	0.000	23:14
Control Test	0.079	23:14
Air Blank	0.000	23:14
Subject Sample # 1	0.221	23:15
Breath Volume	2.722	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.228	23:18
Breath Volume	3.777	23:18
Air Blank	0.000	23:19
Control Test	0.080	23:19
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARR,L,J
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MOLCAN,M
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: MAGERA

DATE/TIME OF TEST: 02/12/2020 00:07:07

Diagnostics Check		OK	00:00
Air Blank		0.000	00:01
Control Test		0.080	00:01
Air Blank		0.000	00:02
Subject Sample	# 1	0.204	00:02
Breath Volume		1.386	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.212	00:05
Breath Volume		1.742	00:05
Air Blank		0.000	00:06
Control Test		0.080	00:06
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THROESCH,PATRICK,J
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: POUNDS,R
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ALICE

DATE/TIME OF TEST: 02/13/2020 03:15:20

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.080	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.132	03:10
Breath Volume		4.250	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.131	03:13
Breath Volume		4.542	03:13
Air Blank		0.000	03:14
Control Test		0.079	03:14
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVAS,W,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: COHEN,B
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: PONCE

DATE/TIME OF TEST: 02/14/2020 23:21:24

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.080	23:16
Air Blank	0.000	23:16
Subject Sample # 1	0.136	23:17
Breath Volume	2.402	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.136	23:19
Breath Volume	3.300	23:19
Air Blank	0.000	23:20
Control Test	0.080	23:20
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KAZMIERCZAK,B
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: TROPIANO

DATE/TIME OF TEST: 02/17/2020 00:44:42

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.079	00:38
Air Blank	0.000	00:39
Subject Sample # 1	0.170	00:40
Breath Volume	1.312	00:40
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.176	00:43
Breath Volume	1.644	00:43
Air Blank	0.000	00:43
Control Test	0.080	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONSEES,D,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ZYNDA,T
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 02/18/2020 00:31:31

Diagnostics Check		OK	00:21
Air Blank		0.000	00:21
Control Test		0.080	00:22
Air Blank		0.000	00:22
Subject Sample	# 1	0.275	00:24
Breath Volume		2.703	00:24
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.249	00:27
Breath Volume		2.191	00:27
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 3	0.241	00:30
Breath Volume		1.417	00:30
Air Blank		0.000	00:30
Control Test		0.080	00:30
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLMSTEAQD,J,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MCCLOSKEY,J
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: STARZEC

DATE/TIME OF TEST: 02/18/2020 04:02:06

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.216	03:57
Breath Volume	3.636	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.223	04:00
Breath Volume	2.218	04:00
Air Blank	0.000	04:01
Control Test	0.078	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,D,S
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: TANGUMA

DATE/TIME OF TEST: 02/18/2020 12:58:33

Diagnostics Check	OK	12:51
Air Blank	0.000	12:52
Control Test	0.080	12:52
Air Blank	0.000	12:52
Subject Sample # 1	0.170	12:54
Breath Volume	1.789	12:54
Air Blank	0.000	12:54
Air Blank	0.000	12:56
Subject Sample # 2	0.174	12:57
Breath Volume	2.441	12:57
Air Blank	0.000	12:57
Control Test	0.080	12:57
Air Blank	0.000	12:58
Diagnostics Check	OK	12:58

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINT-FORT,PAUL,W
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: SCHAARE,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:09

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/20/2020 15:35:46

Diagnostics Check	OK	15:34	
Air Blank	0.000	15:35	
		15:35	RFI Detect
Air Blank	0.000	15:35	

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINT-FORT,PAUL,W
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: SCHAARE,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:09

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/20/2020 15:49:49

Diagnostics Check	OK	15:38	
Air Blank	0.000	15:38	
Control Test	0.080	15:39	
Air Blank	0.000	15:39	
Subject Sample	# 1	15:42	No Sample Provided
Breath Volume	0.000	15:42	
Air Blank	0.000	15:43	
Air Blank	0.000	15:45	
Subject Sample	# 2	15:48	No Sample Provided
Breath Volume	0.000	15:48	
Air Blank	0.000	15:48	
Control Test	0.080	15:49	
Air Blank	0.000	15:49	
Diagnostics Check	OK	15:49	

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,S,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ANTHONY,R
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: HAMLIN

DATE/TIME OF TEST: 02/21/2020 03:35:42

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.079	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.228	03:30
Breath Volume	1.351	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:32
Subject Sample # 2	0.219	03:34
Breath Volume	1.164	03:34
Air Blank	0.000	03:34
Control Test	0.080	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARR,L,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: LEZAMA,J
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: LACCETTI

DATE/TIME OF TEST: 02/21/2020 03:52:08

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.079	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.132	03:47
Breath Volume	2.281	03:47
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.124	03:50
Breath Volume	1.933	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS, MATTHEW, L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MCCLOSKEY, J
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/21/2020 23:18:44

Diagnostics Check		OK	23:10
Air Blank		0.000	23:11
Control Test		0.081	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.204	23:13
Breath Volume		2.742	23:13
Air Blank		0.000	23:14
Air Blank		0.000	23:16
Subject Sample	# 2	0.196	23:17
Breath Volume		1.656	23:17
Air Blank		0.000	23:17
Control Test		0.080	23:18
Air Blank		0.000	23:18
Diagnostics Check		OK	23:18

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAFFEE,C,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: DIAZ,S
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BATCHER

DATE/TIME OF TEST: 02/23/2020 03:21:38

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.080	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.012	03:17
Breath Volume	3.324	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.012	03:20
Breath Volume	3.179	03:20
Air Blank	0.000	03:20
Control Test	0.081	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRATT,N,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PRATT,N
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: DRAVECKY

DATE/TIME OF TEST: 02/24/2020 23:52:37

Diagnostics Check		OK	23:46
Air Blank		0.000	23:46
Control Test		0.079	23:47
Air Blank		0.000	23:47
Subject Sample	# 1	0.216	23:48
Breath Volume		2.007	23:48
Air Blank		0.000	23:48
Air Blank		0.000	23:50
Subject Sample	# 2	0.211	23:51
Breath Volume		2.683	23:51
Air Blank		0.000	23:51
Control Test		0.079	23:52
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERMAN,BRENT,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: WHITE,T
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: GILLILAND

DATE/TIME OF TEST: 02/27/2020 01:00:07

Diagnostics Check		OK	00:53
Air Blank		0.000	00:54
Control Test		0.079	00:54
Air Blank		0.000	00:55
Subject Sample	# 1	0.260	00:55
Breath Volume		2.464	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.260	00:58
Breath Volume		1.957	00:58
Air Blank		0.000	00:59
Control Test		0.079	00:59
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DECKER,DAVID,P
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BORCHERS,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: LOCKHART

DATE/TIME OF TEST: 02/27/2020 17:57:23

Diagnostics Check	OK	17:46
Air Blank	0.000	17:47
Control Test	0.080	17:47
Air Blank	0.000	17:47
Subject Sample # 1	0.182	17:50
Breath Volume	1.203	17:50
Air Blank	0.000	17:51
Air Blank	0.000	17:52
Subject Sample # 2	0.179	17:55
Breath Volume	1.531	17:55
Air Blank	0.000	17:56
Control Test	0.081	17:56
Air Blank	0.000	17:57
Diagnostics Check	OK	17:57

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,PATRICK,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: WOODS,H
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: RADENKOVIC

DATE/TIME OF TEST: 02/27/2020 22:27:13

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.080	22:18
Air Blank	0.000	22:19
Subject Sample # 1	0.199	22:22
Breath Volume	3.617	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:24
Subject Sample # 2	0.203	22:25
Breath Volume	2.164	22:25
Air Blank	0.000	22:26
Control Test	0.079	22:26
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRICE,MARCUS,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ZYNDA,T
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: CEPEDA PRIETO

DATE/TIME OF TEST: 02/28/2020 04:39:05

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.080	04:33
Air Blank	0.000	04:33
Subject Sample # 1	0.131	04:34
Breath Volume	1.371	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.121	04:37
Breath Volume	1.660	04:37
Air Blank	0.000	04:38
Control Test	0.081	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVAS,W,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: STEVENSON,C
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: ESCALANTE

DATE/TIME OF TEST: 02/29/2020 03:06:12

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.080	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.173	03:01
Breath Volume	3.175	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.163	03:04
Breath Volume	2.496	03:04
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA, JOSUE, D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: HERNANDEZ, JOEL
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 02/03/2020 12:46:07

Diagnostics Check	OK	12:40
Air Blank	0.000	12:40
Control Test	0.078	12:40
Air Blank	0.000	12:41
Subject Sample # 1	0.000	12:41
Breath Volume	1.992	12:41
Air Blank	0.000	12:42
Air Blank	0.000	12:44
Subject Sample # 2	0.000	12:44
Breath Volume	3.183	12:44
Air Blank	0.000	12:45
Control Test	0.076	12:45
Air Blank	0.000	12:46
Diagnostics Check	OK	12:46

Data Download Date/Time: 02/26/2020 03:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVINGSTON,JESSICA,L
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: YOUNGMAN,TALON
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: HARLOW

DATE/TIME OF TEST: 02/12/2020 01:04:55

Diagnostics Check	OK	00:56	
Air Blank	0.000	00:57	
Control Test	0.076	00:57	
Air Blank	0.000	00:58	
Subject Sample # 1	0.203	00:59	
Breath Volume	1.605	00:59	
Air Blank	0.000	01:00	
Air Blank	0.000	01:02	
Subject Sample # 2	0.208	01:03	
Breath Volume	2.078	01:03	
Air Blank	0.000	01:04	
Control Test	0.074	01:04	Control Outside Tolerance
Air Blank	0.000	01:04	

Data Download Date/Time: 02/26/2020 03:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA, JOSUE, D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: YOUNGMAN, T
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: HARLOW

DATE/TIME OF TEST: 02/12/2020 12:12:33

Diagnostics Check	OK	12:06
Air Blank	0.000	12:06
Control Test	0.076	12:07
Air Blank	0.000	12:07
Subject Sample # 1	0.015	12:08
Breath Volume	1.707	12:08
Air Blank	0.000	12:08
Air Blank	0.000	12:10
Subject Sample # 2	0.015	12:11
Breath Volume	1.820	12:11
Air Blank	0.000	12:11
Control Test	0.075	12:11
Air Blank	0.000	12:12
Diagnostics Check	OK	12:12

Data Download Date/Time: 02/26/2020 03:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA, JOSUE, D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: YOUNGMAN, T
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: HARLOW

DATE/TIME OF TEST: 02/12/2020 14:01:58

Diagnostics Check	OK	13:56
Air Blank	0.000	13:56
Control Test	0.076	13:56
Air Blank	0.000	13:57
Subject Sample # 1	0.000	13:57
Breath Volume	1.558	13:57
Air Blank	0.000	13:58
Air Blank	0.000	14:00
Subject Sample # 2	0.000	14:00
Breath Volume	1.800	14:00
Air Blank	0.000	14:01
Control Test	0.075	14:01
Air Blank	0.000	14:01
Diagnostics Check	OK	14:01

Data Download Date/Time: 02/26/2020 03:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDLE, CHARES, M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE, CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:08

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/03/2020 07:33:38

Diagnostics Check	OK	07:28	
Air Blank	0.000	07:29	
Control Test	0.080	07:29	
Air Blank	0.000	07:29	
Subject Sample # 1	0.000	07:30	
Breath Volume	1.328	07:30	
Air Blank	0.000	07:31	
Air Blank		07:32	RFI Detect
Air Blank		07:33	RFI Detect
Air Blank	0.000	07:33	

Data Download Date/Time: 02/17/2020 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDLE, CHARLES,
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE, CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:08

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/03/2020 07:40:10

Diagnostics Check	OK	07:35	
Air Blank	0.000	07:35	
Control Test	0.079	07:36	
Air Blank	0.000	07:36	
Subject Sample	# 1	07:38	Subject Test Refused
Breath Volume	0.558	07:38	
Air Blank	0.000	07:39	
Control Test	0.079	07:39	
Air Blank	0.000	07:40	
Diagnostics Check	OK	07:40	

Data Download Date/Time: 02/17/2020 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDLE, CHARLES,
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE, CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:18

SUBJECT'S LAST NAME: VANN

DATE/TIME OF TEST: 02/03/2020 15:41:42

Diagnostics Check	OK	15:38	
Air Blank	0.000	15:38	
Control Test	0.078	15:39	
Air Blank	0.000	15:39	
Subject Sample	# 1	15:40	Interferent Detect
Breath Volume	2.921	15:40	
Air Blank	0.000	15:40	
Control Test	0.079	15:41	
Air Blank	0.000	15:41	
Diagnostics Check	OK	15:41	

Data Download Date/Time: 02/17/2020 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDLE, CHARLES,
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE, CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:18

SUBJECT'S LAST NAME: VANN

DATE/TIME OF TEST: 02/03/2020 15:47:06

Diagnostics Check	OK	15:43	
Air Blank	0.000	15:44	
Control Test	0.079	15:44	
Air Blank	0.000	15:44	
Subject Sample	# 1	15:45	Interferent Detect
Breath Volume	1.921	15:45	
Air Blank	0.000	15:46	
Control Test	0.078	15:46	
Air Blank	0.000	15:47	
Diagnostics Check	OK	15:47	

Data Download Date/Time: 02/17/2020 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN, CRAIG, W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSN, CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 02/16/2020 01:01:58

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:00	
Control Test	0.080	01:00	
Air Blank	0.000	01:01	
		01:01	Improper Sample
Air Blank	0.000	01:01	

Data Download Date/Time: 02/17/2020 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 02/16/2020 01:11:55

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:05	
Control Test	0.079	01:05	
Air Blank	0.000	01:06	
Subject Sample # 1	0.051	01:06	
Breath Volume	1.148	01:06	
Air Blank	0.000	01:07	
Air Blank	0.000	01:09	
Subject Sample # 2	0.053	01:10	
Breath Volume	2.429	01:10	
Air Blank	0.000	01:11	
Control Test	0.080	01:11	Interferent Detect
Air Blank	0.000	01:11	

Data Download Date/Time: 02/17/2020 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN, CRAIG, W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN, CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 02/16/2020 01:20:58

Diagnostics Check	OK	01:13
Air Blank	0.000	01:14
Control Test	0.079	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.050	01:16
Breath Volume	1.605	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:18
Subject Sample # 2	0.052	01:19
Breath Volume	2.210	01:19
Air Blank	0.000	01:20
Control Test	0.078	01:20
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 02/17/2020 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PILKANIS,MARY,C
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: WOOD,DARREN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 242
DRY GAS EXPIRY DATE: 01/01/2023
VIOLATION CODE: DUI
ARREST TIME: 08:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:36

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 02/08/2020 09:59:55

Diagnostics Check	OK	09:57	
Air Blank	0.000	09:57	
Control Test	0.080	09:57	
Air Blank	0.000	09:58	
Subject Sample	# 1	09:58	Subject Test Refused
Breath Volume	0.000	09:58	
Air Blank	0.000	09:59	
Control Test	0.080	09:59	
Air Blank	0.000	09:59	
Diagnostics Check	OK	09:59	

Data Download Date/Time: 03/04/2020 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: VANN,T
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: REVERON

DATE/TIME OF TEST: 02/09/2020 03:37:23

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.168	03:32
Breath Volume	3.136	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.163	03:35
Breath Volume	1.597	03:35
Air Blank	0.000	03:36
Control Test	0.081	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 03/04/2020 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,A.,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: GONZALEZ,A.
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 242
DRY GAS EXPIRY DATE: 01/01/2023
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: ZABALA

DATE/TIME OF TEST: 02/09/2020 04:50:26

Diagnostics Check	OK	04:43
Air Blank	0.000	04:44
Control Test	0.079	04:44
Air Blank	0.000	04:45
Subject Sample # 1	0.122	04:45
Breath Volume	1.800	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.117	04:48
Breath Volume	1.703	04:48
Air Blank	0.000	04:49
Control Test	0.082	04:49
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 03/04/2020 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,ROBERT,A
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WESTMAN,PATRICK
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 02/10/2020 03:49:20

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.080	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.197	03:44
Breath Volume	2.539	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:46
Subject Sample # 2	0.199	03:47
Breath Volume	3.738	03:47
Air Blank	0.000	03:48
Control Test	0.082	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 03/04/2020 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: MONEY

DATE/TIME OF TEST: 02/13/2020 00:56:02

Diagnostics Check	OK	00:49
Air Blank	0.000	00:49
Control Test	0.079	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.052	00:51
Breath Volume	2.726	00:51
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.051	00:54
Breath Volume	3.429	00:54
Air Blank	0.000	00:55
Control Test	0.080	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:56

Data Download Date/Time: 03/04/2020 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,ROBERT,A
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WESTMAN,PATRICK
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: QUALLO

DATE/TIME OF TEST: 02/13/2020 03:54:58

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.079	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.000	03:50
Breath Volume	4.777	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.000	03:53
Breath Volume	3.171	03:53
Air Blank	0.000	03:54
Control Test	0.081	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 03/04/2020 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 02/13/2020 21:20:58

Diagnostics Check	OK	21:15
Air Blank	0.000	21:15
Control Test	0.079	21:15
Air Blank	0.000	21:16
Subject Sample # 1	0.161	21:16
Breath Volume	1.812	21:16
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample # 2	0.167	21:19
Breath Volume	2.636	21:19
Air Blank	0.000	21:20
Control Test	0.079	21:20
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

Data Download Date/Time: 03/04/2020 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: VANN,T
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 02/15/2020 02:59:34

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.080	02:54
Air Blank		0.000	02:54
Subject Sample	# 1	0.228	02:55
Breath Volume		2.039	02:55
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.238	02:58
Breath Volume		2.257	02:58
Air Blank		0.000	02:58
Control Test		0.081	02:58
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 03/04/2020 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,ROBERT,A
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WESTMAN,PATRICK
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: WILKINS

DATE/TIME OF TEST: 02/17/2020 23:10:00

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:04	
Control Test	0.079	23:04	
Air Blank	0.000	23:05	
Subject Sample	# 1 0.176	23:08	Volume Not Met
Breath Volume	0.523	23:08	
Air Blank		23:09	Purge Fail
Air Blank	0.000	23:10	

Data Download Date/Time: 03/04/2020 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,ROBERT,A
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WESTMAN,PATRICK
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: WILKINS

DATE/TIME OF TEST: 02/17/2020 23:20:11

Diagnostics Check	OK	23:12
Air Blank	0.000	23:13
Control Test	0.077	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.199	23:14
Breath Volume	2.050	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.205	23:18
Breath Volume	2.101	23:18
Air Blank	0.000	23:19
Control Test	0.079	23:19
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 03/04/2020 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 02/18/2020 22:56:39

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.080	22:51
Air Blank	0.000	22:51
Subject Sample # 1	0.100	22:52
Breath Volume	2.156	22:52
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.102	22:55
Breath Volume	2.949	22:55
Air Blank	0.000	22:55
Control Test	0.081	22:56
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 03/04/2020 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,ROBERT,A
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: KISTNER,CRAIG
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 02/21/2020 22:11:33

Diagnostics Check	OK	22:04
Air Blank	0.000	22:05
Control Test	0.081	22:05
Air Blank	0.000	22:05
Subject Sample # 1	0.215	22:06
Breath Volume	2.593	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.215	22:09
Breath Volume	3.765	22:09
Air Blank	0.000	22:10
Control Test	0.083	22:10
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 03/04/2020 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 02/22/2020 02:27:52

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.081	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.073	02:23
Breath Volume	2.625	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.067	02:26
Breath Volume	2.726	02:26
Air Blank	0.000	02:26
Control Test	0.082	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 03/04/2020 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: IVERY

DATE/TIME OF TEST: 02/23/2020 01:07:45

Diagnostics Check	OK	01:01
Air Blank	0.000	01:02
Control Test	0.081	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.047	01:03
Breath Volume	2.523	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.046	01:06
Breath Volume	2.507	01:06
Air Blank	0.000	01:06
Control Test	0.083	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 03/04/2020 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: MCFEE

DATE/TIME OF TEST: 02/23/2020 04:25:09

Diagnostics Check		OK	04:16
Air Blank		0.000	04:16
Control Test		0.081	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.185	04:17
Breath Volume		3.023	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.164	04:20
Breath Volume		2.019	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 3	0.184	04:23
Breath Volume		3.609	04:23
Air Blank		0.000	04:24
Control Test		0.082	04:24
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 03/04/2020 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/26/2020 01:38:49

Diagnostics Check		OK	01:32
Air Blank		0.000	01:33
Control Test		0.080	01:33
Air Blank		0.000	01:34
Subject Sample	# 1	0.172	01:34
Breath Volume		1.765	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:36
Subject Sample	# 2	0.157	01:37
Breath Volume		1.609	01:37
Air Blank		0.000	01:37
Control Test		0.080	01:38
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 03/04/2020 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,ANDREW,L
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: HEDGES,MATTHEW
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: SCHLUER

DATE/TIME OF TEST: 02/26/2020 03:54:50

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:52	
Control Test	0.080	03:52	
Air Blank	0.000	03:53	
Subject Sample	# 1	03:53	Subject Test Refused
Breath Volume	0.000	03:53	
Air Blank	0.000	03:53	
Control Test	0.080	03:54	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 03/04/2020 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: KINDER

DATE/TIME OF TEST: 02/28/2020 01:00:21

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:55
Subject Sample # 1	0.000	00:56
Breath Volume	2.207	00:56
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.000	00:58
Breath Volume	1.812	00:58
Air Blank	0.000	00:59
Control Test	0.082	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 03/04/2020 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,ROBERT,A
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: TUCKER,CEDRICK
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 02/28/2020 02:18:14

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.080	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.219	02:12
Breath Volume	2.535	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.223	02:16
Breath Volume	3.863	02:16
Air Blank	0.000	02:17
Control Test	0.083	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 03/04/2020 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 02/29/2020 03:55:53

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.081	03:50
Air Blank		0.000	03:50
Subject Sample	# 1	0.184	03:51
Breath Volume		1.468	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:53
Subject Sample	# 2	0.173	03:54
Breath Volume		1.191	03:54
Air Blank		0.000	03:54
Control Test		0.081	03:55
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 03/04/2020 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIDSON,JUSTIN,T
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: DAVIDSON,JUSTIN
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 02/05/2020 20:40:07

Diagnostics Check		OK	20:33
Air Blank		0.000	20:34
Control Test		0.079	20:34
Air Blank		0.000	20:34
Subject Sample	# 1	0.284	20:35
Breath Volume		1.300	20:35
Air Blank		0.000	20:36
Air Blank		0.000	20:37
Subject Sample	# 2	0.275	20:38
Breath Volume		1.250	20:38
Air Blank		0.000	20:39
Control Test		0.078	20:39
Air Blank		0.000	20:40
Diagnostics Check		OK	20:40

Data Download Date/Time: 03/04/2020 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 02/08/2020 03:04:34

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.189	02:59
Breath Volume	1.304	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.182	03:03
Breath Volume	3.746	03:03
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 03/04/2020 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: SOMMER

DATE/TIME OF TEST: 02/08/2020 03:58:13

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.081	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.170	03:53
Breath Volume	2.156	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.171	03:56
Breath Volume	1.390	03:56
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 03/04/2020 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: POUPARD

DATE/TIME OF TEST: 02/08/2020 22:30:01

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.080	22:24
Air Blank	0.000	22:24
Subject Sample # 1	0.223	22:25
Breath Volume	1.246	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:27
Subject Sample # 2	0.211	22:28
Breath Volume	1.480	22:28
Air Blank	0.000	22:29
Control Test	0.080	22:29
Air Blank	0.000	22:29
Diagnostics Check	OK	22:30

Data Download Date/Time: 03/04/2020 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A1
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 02/15/2020 02:18:17

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.143	02:11
Breath Volume	1.695	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.134	02:16
Breath Volume	1.175	02:16
Air Blank	0.000	02:17
Control Test	0.079	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 03/04/2020 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000960

INSTRUMENT REGISTERED TO: LIBERTY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROHECKER,KYNDAL,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: STROHECKER,KYNDAL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GOFF

DATE/TIME OF TEST: 02/29/2020 04:07:50

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.081	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.124	04:03
Breath Volume	1.671	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.127	04:06
Breath Volume	1.472	04:06
Air Blank	0.000	04:06
Control Test	0.080	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 03/17/2020 15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,VICKI,N
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: OWENS,JOHN
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 02/17/2020 01:57:09

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.078	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.125	01:52
Breath Volume	1.753	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.126	01:55
Breath Volume	3.074	01:55
Air Blank	0.000	01:56
Control Test	0.077	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 02/26/2020 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,RONDRICKA,N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: THOMPSON,RONDRICKA
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: LEMATTEY

DATE/TIME OF TEST: 02/19/2020 01:51:30

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:47	
Control Test	0.078	01:47	
Air Blank	0.000	01:48	
Subject Sample	# 1	01:50	Subject Test Refused
Breath Volume	0.000	01:50	
Air Blank	0.000	01:50	
Control Test	0.078	01:50	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 02/26/2020 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: OROZCOAGUILAR

DATE/TIME OF TEST: 02/01/2020 00:04:33

Diagnostics Check	OK	23:58
Air Blank	0.000	23:59
Control Test	0.082	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.142	00:00
Breath Volume	3.328	00:00
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 2	0.140	00:03
Breath Volume	3.664	00:03
Air Blank	0.000	00:03
Control Test	0.081	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 02/06/2020 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: FENELUS

DATE/TIME OF TEST: 02/01/2020 02:09:26

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.082	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.169	02:05
Breath Volume	1.570	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.166	02:07
Breath Volume	2.218	02:07
Air Blank	0.000	02:08
Control Test	0.081	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 02/06/2020 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HANSTINE,AUSTIN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: FERMIN

DATE/TIME OF TEST: 02/01/2020 02:28:47

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.168	02:24
Breath Volume	3.625	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.162	02:27
Breath Volume	2.320	02:27
Air Blank	0.000	02:27
Control Test	0.081	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 02/06/2020 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HOPKINS,JEFFREY
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: CEPEDES

DATE/TIME OF TEST: 02/01/2020 04:01:46

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.082	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.028	03:57
Breath Volume	3.359	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.026	04:00
Breath Volume	3.199	04:00
Air Blank	0.000	04:00
Control Test	0.082	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 02/06/2020 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BINET,MARLON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: SOLA

DATE/TIME OF TEST: 02/02/2020 03:53:58

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.082	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.193	03:49
Breath Volume	1.488	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.190	03:52
Breath Volume	1.578	03:52
Air Blank	0.000	03:53
Control Test	0.081	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 02/06/2020 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HOLIDAY,LUCAS
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/02/2020 05:19:04

Diagnostics Check	OK	05:12
Air Blank	0.000	05:13
Control Test	0.082	05:13
Air Blank	0.000	05:13
Subject Sample # 1	0.086	05:14
Breath Volume	1.492	05:14
Air Blank	0.000	05:15
Air Blank	0.000	05:17
Subject Sample # 2	0.085	05:17
Breath Volume	2.089	05:17
Air Blank	0.000	05:18
Control Test	0.080	05:18
Air Blank	0.000	05:18
Diagnostics Check	OK	05:19

Data Download Date/Time: 02/06/2020 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRAZEE,COLE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 02/03/2020 00:20:30

Diagnostics Check	OK	00:14
Air Blank	0.000	00:14
Control Test	0.082	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.000	00:16
Breath Volume	3.238	00:16
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.000	00:19
Breath Volume	2.617	00:19
Air Blank	0.000	00:19
Control Test	0.082	00:19
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 02/06/2020 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BAKSH,SAEED
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: PINKSTON

DATE/TIME OF TEST: 02/03/2020 02:33:53

Diagnostics Check	OK	02:27
Air Blank	0.000	02:28
Control Test	0.081	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.229	02:29
Breath Volume	2.757	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:31
Subject Sample # 2	0.215	02:32
Breath Volume	1.171	02:32
Air Blank	0.000	02:32
Control Test	0.081	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 02/06/2020 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ADAMS,JAMES
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: DIEULIFETE

DATE/TIME OF TEST: 02/05/2020 03:54:23

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.082	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.145	03:50
Breath Volume	1.496	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.144	03:52
Breath Volume	1.593	03:52
Air Blank	0.000	03:53
Control Test	0.082	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 02/06/2020 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRAZEE,COLE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: NOPPINGER

DATE/TIME OF TEST: 02/05/2020 20:28:02

Diagnostics Check	OK	20:24	
Air Blank	0.000	20:24	
Control Test	0.082	20:24	
Air Blank	0.000	20:25	
Subject Sample	# 1	20:26	Interferent Detect
Breath Volume	2.253	20:26	
Air Blank	0.000	20:27	
Control Test	0.082	20:27	
Air Blank	0.000	20:27	
Diagnostics Check	OK	20:28	

Data Download Date/Time: 02/06/2020 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRAZEE,COLE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: NOPPINGER

DATE/TIME OF TEST: 02/05/2020 20:32:45

Diagnostics Check	OK	20:29	
Air Blank	0.000	20:30	
Control Test	0.081	20:30	
Air Blank	0.000	20:30	
Subject Sample	# 1	20:31	Interferent Detect
Breath Volume	2.656	20:31	
Air Blank	0.000	20:31	
Control Test	0.081	20:32	
Air Blank	0.000	20:32	
Diagnostics Check	OK	20:32	

Data Download Date/Time: 02/06/2020 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GILZEN,MARCUS
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: SANTULLO

DATE/TIME OF TEST: 02/07/2020 00:12:17

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.082	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.199	00:07
Breath Volume	1.429	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.194	00:10
Breath Volume	1.773	00:10
Air Blank	0.000	00:11
Control Test	0.080	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 02/11/2020 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BANOS,FRANK
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: KILIAN

DATE/TIME OF TEST: 02/08/2020 06:21:08

Diagnostics Check	OK	06:14
Air Blank	0.000	06:15
Control Test	0.083	06:15
Air Blank	0.000	06:16
Subject Sample # 1	0.213	06:16
Breath Volume	3.601	06:16
Air Blank	0.000	06:17
Air Blank	0.000	06:19
Subject Sample # 2	0.197	06:19
Breath Volume	2.195	06:19
Air Blank	0.000	06:20
Control Test	0.081	06:20
Air Blank	0.000	06:21
Diagnostics Check	OK	06:21

Data Download Date/Time: 02/11/2020 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SLACK,JOHNIE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: BRENT

DATE/TIME OF TEST: 02/09/2020 01:18:39

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.083	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.172	01:14
Breath Volume	1.457	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.171	01:17
Breath Volume	2.152	01:17
Air Blank	0.000	01:17
Control Test	0.081	01:18
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 02/11/2020 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOMINGUEZ,JAYMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 02/09/2020 03:38:53

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.083	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.113	03:34
Breath Volume	1.429	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.117	03:37
Breath Volume	1.191	03:37
Air Blank	0.000	03:37
Control Test	0.081	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 02/11/2020 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JACKSON,MELISSA
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: PEREZREYES

DATE/TIME OF TEST: 02/09/2020 05:27:32

Diagnostics Check		OK	05:20
Air Blank		0.000	05:21
Control Test		0.082	05:21
Air Blank		0.000	05:22
Subject Sample	# 1	0.241	05:23
Breath Volume		2.039	05:23
Air Blank		0.000	05:23
Air Blank		0.000	05:25
Subject Sample	# 2	0.230	05:26
Breath Volume		1.464	05:26
Air Blank		0.000	05:26
Control Test		0.081	05:26
Air Blank		0.000	05:27
Diagnostics Check		OK	05:27

Data Download Date/Time: 02/11/2020 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ABREW, JUAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:41

SUBJECT'S LAST NAME: GUSTAFSON

DATE/TIME OF TEST: 02/09/2020 06:11:33

Diagnostics Check	OK	06:05	
Air Blank	0.000	06:05	
Control Test	0.082	06:05	
Air Blank	0.000	06:06	
Subject Sample # 1	0.129	06:07	
Breath Volume	4.644	06:07	
Air Blank	0.000	06:07	
Air Blank	0.000	06:09	
Subject Sample # 2	0.130	06:10	
Breath Volume	4.367	06:10	
Air Blank	0.000	06:10	
Control Test	0.082	06:10	
Air Blank		06:11	RFI Detect
Air Blank	0.000	06:11	

Data Download Date/Time: 02/11/2020 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ABREW, JUAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:41

SUBJECT'S LAST NAME: GUSTAFSON

DATE/TIME OF TEST: 02/09/2020 06:21:18

Diagnostics Check	OK	06:14
Air Blank	0.000	06:15
Control Test	0.082	06:15
Air Blank	0.000	06:16
Subject Sample # 1	0.128	06:16
Breath Volume	3.875	06:16
Air Blank	0.000	06:17
Air Blank	0.000	06:19
Subject Sample # 2	0.125	06:19
Breath Volume	3.578	06:19
Air Blank	0.000	06:20
Control Test	0.081	06:20
Air Blank	0.000	06:21
Diagnostics Check	OK	06:21

Data Download Date/Time: 02/11/2020 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DELLOVA, JOSEPH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:59

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 02/09/2020 07:00:45

Diagnostics Check	OK	06:54
Air Blank	0.000	06:54
Control Test	0.082	06:54
Air Blank	0.000	06:55
Subject Sample # 1	0.075	06:56
Breath Volume	1.632	06:56
Air Blank	0.000	06:56
Air Blank	0.000	06:58
Subject Sample # 2	0.078	06:59
Breath Volume	2.261	06:59
Air Blank	0.000	06:59
Control Test	0.081	07:00
Air Blank	0.000	07:00
Diagnostics Check	OK	07:00

Data Download Date/Time: 02/11/2020 15:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMEY,MICHAEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 02/12/2020 03:07:50

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.082	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.204	03:03
Breath Volume		1.250	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.199	03:06
Breath Volume		1.140	03:06
Air Blank		0.000	03:06
Control Test		0.081	03:07
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 02/20/2020 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JUBER,KATE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: BONILLACORTEZ

DATE/TIME OF TEST: 02/15/2020 02:09:16

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.082	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.180	02:04
Breath Volume	2.910	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.174	02:07
Breath Volume	3.234	02:07
Air Blank	0.000	02:08
Control Test	0.081	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 02/20/2020 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRAZEE,COLE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: BORREROVELEZ

DATE/TIME OF TEST: 02/15/2020 04:32:15

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.082	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.124	04:27
Breath Volume	1.507	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.118	04:30
Breath Volume	1.265	04:30
Air Blank	0.000	04:31
Control Test	0.081	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 02/20/2020 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRAZEE,COLE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: LIGHTSEY

DATE/TIME OF TEST: 02/16/2020 04:46:02

Diagnostics Check	OK	04:38
Air Blank	0.000	04:38
Control Test	0.082	04:39
Air Blank	0.000	04:39
Subject Sample # 1	0.064	04:41
Breath Volume	1.464	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:43
Subject Sample # 2	0.064	04:44
Breath Volume	1.554	04:44
Air Blank	0.000	04:45
Control Test	0.082	04:45
Air Blank	0.000	04:45
Diagnostics Check	OK	04:46

Data Download Date/Time: 02/20/2020 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROCCHILD, MIGUEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: ROSARIO ANDINO

DATE/TIME OF TEST: 02/16/2020 22:50:29

Diagnostics Check	OK	22:44
Air Blank	0.000	22:44
Control Test	0.081	22:44
Air Blank	0.000	22:45
Subject Sample # 1	0.159	22:45
Breath Volume	2.285	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.158	22:48
Breath Volume	2.777	22:48
Air Blank	0.000	22:49
Control Test	0.081	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 02/20/2020 16:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARMAN, RICHARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: OROZCO MEJIA

DATE/TIME OF TEST: 02/17/2020 05:06:55

Diagnostics Check		OK	04:49	
Air Blank		0.000	04:50	
Control Test		0.083	04:50	
Air Blank		0.000	04:50	
Subject Sample	# 1	0.116	04:54	Volume Not Met
Breath Volume		0.292	04:54	
Air Blank		0.000	04:54	
Air Blank		0.000	04:56	
Subject Sample	# 2	0.123	04:59	
Breath Volume		1.535	04:59	
Air Blank		0.000	05:00	
Air Blank		0.000	05:02	
Subject Sample	# 3	0.121	05:05	Volume Not Met
Breath Volume		0.378	05:05	
Air Blank		0.000	05:06	
Control Test		0.081	05:06	
Air Blank		0.000	05:06	
Diagnostics Check		OK	05:06	

Data Download Date/Time: 02/20/2020 16:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HARMAN, RICHARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: OROZCO MEJIA

DATE/TIME OF TEST: 02/17/2020 05:16:35

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.078	05:10
Air Blank	0.000	05:10
Subject Sample # 1	0.130	05:11
Breath Volume	1.222	05:11
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.119	05:15
Breath Volume	1.421	05:15
Air Blank	0.000	05:15
Control Test	0.081	05:15
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 02/20/2020 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DELLOVA, JOSEPH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 02/18/2020 02:47:52

Diagnostics Check		OK	02:36	
Air Blank		0.000	02:36	
Control Test		0.082	02:37	
Air Blank		0.000	02:37	
Subject Sample	# 1	0.093	02:40	Volume Not Met
Breath Volume		0.269	02:40	
Air Blank		0.000	02:41	
Air Blank		0.000	02:43	
Subject Sample	# 2	0.099	02:46	Volume Not Met
Breath Volume		0.785	02:46	
Air Blank		0.000	02:46	
Control Test		0.082	02:47	
Air Blank		0.000	02:47	
Diagnostics Check		OK	02:47	

Data Download Date/Time: 02/20/2020 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DELLOVA, JOSEPH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 02/18/2020 02:58:56

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.081	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.108	02:53
Breath Volume	1.425	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:55
Subject Sample # 2	0.100	02:57
Breath Volume	1.214	02:57
Air Blank	0.000	02:58
Control Test	0.081	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 02/20/2020 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BODE, JOSHUA
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: RAMSAROOP

DATE/TIME OF TEST: 02/18/2020 22:52:02

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:42	
Control Test	0.080	22:42	
Air Blank	0.000	22:43	
Subject Sample	# 1	22:46	No Sample Provided
Breath Volume	0.000	22:46	
Air Blank	0.000	22:47	
Air Blank	0.000	22:49	
Subject Sample	# 2	22:50	Subject Test Refused
Breath Volume	0.000	22:50	
Air Blank	0.000	22:51	
Control Test	0.082	22:51	
Air Blank	0.000	22:51	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 02/20/2020 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BINET,MARLO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 02/20/2020 02:21:20

Diagnostics Check		OK	02:13
Air Blank		0.000	02:14
Control Test		0.081	02:14
Air Blank		0.000	02:14
Subject Sample	# 1	0.223	02:16
Breath Volume		2.515	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.230	02:19
Breath Volume		3.585	02:19
Air Blank		0.000	02:20
Control Test		0.081	02:20
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 02/20/2020 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOYE, JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: SIMONEAU

DATE/TIME OF TEST: 02/23/2020 01:12:24

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.082	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.083	01:07
Breath Volume	3.218	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.084	01:10
Breath Volume	3.410	01:10
Air Blank	0.000	01:11
Control Test	0.082	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 02/27/2020 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DAVIS, DWAYNE
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: GARCIAPULIDO

DATE/TIME OF TEST: 02/26/2020 05:03:53

Diagnostics Check	OK	04:56
Air Blank	0.000	04:57
Control Test	0.082	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.110	04:58
Breath Volume	1.589	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.111	05:02
Breath Volume	1.925	05:02
Air Blank	0.000	05:02
Control Test	0.082	05:03
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 02/27/2020 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KISSANE,BRIAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: RAM

DATE/TIME OF TEST: 02/27/2020 00:13:51

Diagnostics Check		OK	00:02	
Air Blank		0.000	00:02	
Control Test		0.083	00:03	
Air Blank		0.000	00:03	
Subject Sample	# 1	0.012	00:06	Volume Not Met
Breath Volume		0.359	00:06	
Air Blank		0.000	00:07	
Air Blank		0.000	00:09	
Subject Sample	# 2		00:12	No Sample Provided
Breath Volume		0.000	00:12	
Air Blank		0.000	00:12	
Control Test		0.082	00:13	
Air Blank		0.000	00:13	
Diagnostics Check		OK	00:13	

Data Download Date/Time: 02/27/2020 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SAYLER, ROBERT
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: DURANGO

DATE/TIME OF TEST: 02/27/2020 02:07:51

Diagnostics Check	OK	02:01
Air Blank	0.000	02:02
Control Test	0.082	02:02
Air Blank	0.000	02:03
Subject Sample # 1	0.127	02:03
Breath Volume	2.074	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.126	02:06
Breath Volume	2.585	02:06
Air Blank	0.000	02:06
Control Test	0.082	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 02/27/2020 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: STARNES,JACOB
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: STEWARD

DATE/TIME OF TEST: 02/27/2020 21:10:46

Diagnostics Check	OK	21:04
Air Blank	0.000	21:05
Control Test	0.082	21:05
Air Blank	0.000	21:06
Subject Sample # 1	0.206	21:06
Breath Volume	2.277	21:06
Air Blank	0.000	21:07
Air Blank	0.000	21:08
Subject Sample # 2	0.199	21:09
Breath Volume	1.851	21:09
Air Blank	0.000	21:09
Control Test	0.081	21:10
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

Data Download Date/Time: 03/05/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMPSON,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: VALENCIA

DATE/TIME OF TEST: 02/28/2020 21:21:56

Diagnostics Check		OK	21:15
Air Blank		0.000	21:16
Control Test		0.082	21:16
Air Blank		0.000	21:17
Subject Sample	# 1	0.144	21:17
Breath Volume		2.097	21:17
Air Blank		0.000	21:18
Air Blank		0.000	21:20
Subject Sample	# 2	0.144	21:20
Breath Volume		2.464	21:20
Air Blank		0.000	21:21
Control Test		0.082	21:21
Air Blank		0.000	21:21
Diagnostics Check		OK	21:21

Data Download Date/Time: 03/05/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MONTGOMERY,JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 02/29/2020 02:56:04

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.082	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.232	02:51
Breath Volume	3.769	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.234	02:54
Breath Volume	3.433	02:54
Air Blank	0.000	02:55
Control Test	0.079	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:56

Data Download Date/Time: 03/05/2020 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: PINEDA

DATE/TIME OF TEST: 02/09/2020 02:39:38

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.079	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.150	02:34
Breath Volume	1.667	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.150	02:38
Breath Volume	1.234	02:38
Air Blank	0.000	02:38
Control Test	0.079	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 02/26/2020 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOUVRES,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: DOUVRES,MICHAEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: BAXTER RIOS

DATE/TIME OF TEST: 02/09/2020 03:56:46

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.159	03:52
Breath Volume	3.207	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.165	03:55
Breath Volume	2.679	03:55
Air Blank	0.000	03:55
Control Test	0.078	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 02/26/2020 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: FLECK

DATE/TIME OF TEST: 02/09/2020 23:37:50

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.079	23:32
Air Blank	0.000	23:32
Subject Sample # 1	0.000	23:33
Breath Volume	3.882	23:33
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.000	23:36
Breath Volume	2.972	23:36
Air Blank	0.000	23:36
Control Test	0.077	23:37
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 02/26/2020 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 05:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: HASSELL

DATE/TIME OF TEST: 02/10/2020 06:35:33

Diagnostics Check	OK	06:29
Air Blank	0.000	06:29
Control Test	0.080	06:30
Air Blank	0.000	06:30
Subject Sample # 1	0.157	06:31
Breath Volume	1.605	06:31
Air Blank	0.000	06:31
Air Blank	0.000	06:33
Subject Sample # 2	0.152	06:34
Breath Volume	1.433	06:34
Air Blank	0.000	06:34
Control Test	0.079	06:34
Air Blank	0.000	06:35
Diagnostics Check	OK	06:35

Data Download Date/Time: 02/26/2020 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRISWOLD,DAVID,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WHOBREY,SPENCER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 02/16/2020 02:48:02

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.079	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.148	02:43
Breath Volume	2.113	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:45
Subject Sample # 2	0.147	02:46
Breath Volume	2.402	02:46
Air Blank	0.000	02:47
Control Test	0.080	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:48

Data Download Date/Time: 02/26/2020 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,DUSTIN,E
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HENDERSON,DUSTIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 11:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: COLEY

DATE/TIME OF TEST: 02/18/2020 01:26:27

Diagnostics Check		OK	01:14	
Air Blank		0.000	01:15	
Control Test		0.080	01:15	
Air Blank		0.000	01:16	
Subject Sample	# 1	0.164	01:19	Volume Not Met
Breath Volume		0.097	01:19	
Air Blank		0.000	01:19	
Air Blank		0.000	01:21	
Subject Sample	# 2	0.165	01:24	Volume Not Met
Breath Volume		0.035	01:24	
Air Blank		0.000	01:25	
Control Test		0.080	01:25	
Air Blank		0.000	01:26	
Diagnostics Check		OK	01:26	

Data Download Date/Time: 02/26/2020 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,DUSTIN,E
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HENDERSON,DUSTIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 11:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: COLEY

DATE/TIME OF TEST: 02/18/2020 01:38:52

Diagnostics Check		OK	01:30	
Air Blank		0.000	01:30	
Control Test		0.079	01:31	
Air Blank		0.000	01:31	
Subject Sample	# 1	0.157	01:34	Volume Not Met
Breath Volume		0.207	01:34	
Air Blank		0.000	01:35	
Air Blank		0.000	01:37	
Subject Sample	# 2		01:37	Subject Test Refused
Breath Volume		0.000	01:37	
Air Blank		0.000	01:37	
Control Test		0.077	01:38	
Air Blank		0.000	01:38	
Diagnostics Check		OK	01:38	

Data Download Date/Time: 02/26/2020 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: FORBES

DATE/TIME OF TEST: 02/20/2020 03:33:20

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.078	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.027	03:28
Breath Volume	1.601	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.027	03:31
Breath Volume	2.121	03:31
Air Blank	0.000	03:32
Control Test	0.077	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 02/26/2020 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WHOBREY,SPENCER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: UMANA-PRUDENCIO

DATE/TIME OF TEST: 02/21/2020 01:46:41

Diagnostics Check		OK	01:40
Air Blank		0.000	01:40
Control Test		0.077	01:41
Air Blank		0.000	01:41
Subject Sample	# 1	0.198	01:42
Breath Volume		1.898	01:42
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.200	01:45
Breath Volume		2.039	01:45
Air Blank		0.000	01:45
Control Test		0.077	01:46
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 02/26/2020 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: FISHER,JORDAN
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: SANCHEZ-ARRIAGAS

DATE/TIME OF TEST: 02/23/2020 05:41:51

Diagnostics Check	OK	05:35
Air Blank	0.000	05:35
Control Test	0.079	05:36
Air Blank	0.000	05:36
Subject Sample # 1	0.098	05:37
Breath Volume	2.441	05:37
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample # 2	0.098	05:40
Breath Volume	2.195	05:40
Air Blank	0.000	05:40
Control Test	0.079	05:41
Air Blank	0.000	05:41
Diagnostics Check	OK	05:41

Data Download Date/Time: 02/26/2020 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES, RONNIE,
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: BARBERI, FRANCIS
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 02/29/2020 22:43:43

Diagnostics Check	OK	22:37
Air Blank	0.000	22:37
Control Test	0.080	22:37
Air Blank	0.000	22:38
Subject Sample # 1	0.000	22:39
Breath Volume	5.621	22:39
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.000	22:42
Breath Volume	5.703	22:42
Air Blank	0.000	22:42
Control Test	0.081	22:43
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time: 03/31/2020 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BROOKS,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 13:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:12

SUBJECT'S LAST NAME: BALDWIN

DATE/TIME OF TEST: 02/15/2020 14:47:56

Diagnostics Check	OK	14:41
Air Blank	0.000	14:41
Control Test	0.080	14:42
Air Blank	0.000	14:42
Subject Sample # 1	0.216	14:43
Breath Volume	1.335	14:43
Air Blank	0.000	14:44
Air Blank	0.000	14:45
Subject Sample # 2	0.226	14:46
Breath Volume	1.371	14:46
Air Blank	0.000	14:47
Control Test	0.079	14:47
Air Blank	0.000	14:47
Diagnostics Check	OK	14:47

Data Download Date/Time: 02/26/2020 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,DAVID,
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: ORTIZ,DAVID
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 02/03/2020 02:42:17

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.080	02:36
Air Blank		0.000	02:37
Subject Sample	# 1	0.158	02:37
Breath Volume		2.015	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.170	02:40
Breath Volume		2.679	02:40
Air Blank		0.000	02:41
Control Test		0.078	02:41
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 02/25/2020 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,DAVID,
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: EASTMAN,MATTHEW
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: LYTWYN

DATE/TIME OF TEST: 02/06/2020 04:33:50

Diagnostics Check	OK	04:27
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.122	04:29
Breath Volume	2.308	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:31
Subject Sample # 2	0.123	04:32
Breath Volume	2.527	04:32
Air Blank	0.000	04:32
Control Test	0.080	04:33
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 02/25/2020 10:00