

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKER,ALEXA,C
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: VAZQUEZ,RANDY
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: VALLE

DATE/TIME OF TEST: 03/05/2020 04:12:14

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.082	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.144	04:07
Breath Volume		3.703	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.148	04:10
Breath Volume		3.867	04:10
Air Blank		0.000	04:11
Control Test		0.080	04:11
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 03/26/2020 12:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,JUSTIN,J
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: CORTES,RAUL
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:54

SUBJECT'S LAST NAME: GOENAGA

DATE/TIME OF TEST: 03/07/2020 18:20:28

Diagnostics Check	OK	18:16	
Air Blank	0.000	18:17	
Control Test	0.081	18:17	
Air Blank	0.000	18:18	
Subject Sample	# 1	18:19	Subject Test Refused
Breath Volume	0.000	18:19	
Air Blank	0.000	18:19	
Control Test	0.081	18:19	
Air Blank	0.000	18:20	
Diagnostics Check	OK	18:20	

Data Download Date/Time: 03/26/2020 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA MILERA, KARLA,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: VILLAMIL, NICHOLAS
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: ROBBINS

DATE/TIME OF TEST: 03/09/2020 11:44:38

Diagnostics Check	OK	11:41	
Air Blank	0.000	11:42	
Control Test	0.082	11:42	
Air Blank	0.000	11:43	
Subject Sample	# 1	11:43	Subject Test Refused
Breath Volume	0.000	11:43	
Air Blank	0.000	11:43	
Control Test	0.081	11:44	
Air Blank	0.000	11:44	
Diagnostics Check	OK	11:44	

Data Download Date/Time: 03/26/2020 12:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKER,ALEXA,C
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: VAZQUEZ,RANDY
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 03/11/2020 03:03:42

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:01	
Control Test	0.082	03:01	
Air Blank	0.000	03:02	
Subject Sample	# 1	03:02	Subject Test Refused
Breath Volume	0.000	03:02	
Air Blank	0.000	03:02	
Control Test	0.082	03:03	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:03	

Data Download Date/Time: 03/26/2020 12:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VELASQUEZ,A.,D
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: VELASQUEZ,A.
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/13/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: WILSON CORDERO

DATE/TIME OF TEST: 03/15/2020 03:18:27

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:16	
Control Test	0.080	03:16	
Air Blank	0.000	03:16	
Subject Sample	# 1	03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:17	
Control Test	0.081	03:17	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

Data Download Date/Time: 05/05/2020 06:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYEGH,MICHELLE,M
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: MONTESINOS,ROBERT
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 03/09/2020 00:24:24

Diagnostics Check		OK	00:11	
Air Blank		0.000	00:11	
Control Test		0.081	00:12	
Air Blank		0.000	00:12	
Subject Sample	# 1	0.119	00:15	Volume Not Met
Breath Volume		0.535	00:15	
Air Blank		0.000	00:16	
Air Blank		0.000	00:18	
Subject Sample	# 2	0.118	00:19	
Breath Volume		1.808	00:19	
Air Blank		0.000	00:20	
Air Blank		0.000	00:22	
Subject Sample	# 3	0.115	00:22	
Breath Volume		1.304	00:22	
Air Blank		0.000	00:23	
Control Test		0.080	00:23	
Air Blank		0.000	00:24	
Diagnostics Check		OK	00:24	

Data Download Date/Time: 03/18/2020 12:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON,REGINALD,B
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: HORTON,REGINALD
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: HERDER

DATE/TIME OF TEST: 03/12/2020 01:21:51

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:20	
Control Test	0.147	01:21	Control Outside Tolerance
Air Blank	0.000	01:21	

Data Download Date/Time: 03/18/2020 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON,REGINALD,B
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: HORTON,REGINALD
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: HERDER

DATE/TIME OF TEST: 03/12/2020 01:29:57

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.081	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.184	01:25
Breath Volume	1.722	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.185	01:28
Breath Volume	1.433	01:28
Air Blank	0.000	01:29
Control Test	0.081	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 03/18/2020 12:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOCARRAS,PEDRO,A
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: DELGADO,EDWARD
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: BRONER

DATE/TIME OF TEST: 03/13/2020 21:35:58

Diagnostics Check	OK	21:30	
Air Blank	0.000	21:31	
Control Test	0.081	21:31	
Air Blank	0.000	21:32	
Subject Sample	# 1	21:34	Subject Test Refused
Breath Volume	0.000	21:34	
Air Blank	0.000	21:35	
Control Test	0.081	21:35	
Air Blank	0.000	21:35	
Diagnostics Check	OK	21:35	

Data Download Date/Time: 03/18/2020 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: SARETTE

DATE/TIME OF TEST: 03/11/2020 03:58:16

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.333	03:53
Breath Volume	2.859	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.337	03:56
Breath Volume	2.441	03:56
Air Blank	0.000	03:57
Control Test	0.078	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 03/25/2020 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,MIGUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/14/2020 05:12:04

Diagnostics Check	OK	05:05
Air Blank	0.000	05:05
Control Test	0.079	05:05
Air Blank	0.000	05:06
Subject Sample # 1	0.154	05:06
Breath Volume	2.082	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.155	05:10
Breath Volume	3.445	05:10
Air Blank	0.000	05:11
Control Test	0.081	05:11
Air Blank	0.000	05:11
Diagnostics Check	OK	05:12

Data Download Date/Time: 03/25/2020 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DARLING,R,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DARLING,R
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 03/20/2020 07:57:08

Diagnostics Check		OK	07:50
Air Blank		0.000	07:50
Control Test		0.080	07:50
Air Blank		0.000	07:51
Subject Sample	# 1	0.206	07:51
Breath Volume		1.128	07:51
Air Blank		0.000	07:52
Air Blank		0.000	07:54
Subject Sample	# 2	0.194	07:55
Breath Volume		1.492	07:55
Air Blank		0.000	07:56
Control Test		0.078	07:56
Air Blank		0.000	07:57
Diagnostics Check		OK	07:57

Data Download Date/Time: 03/25/2020 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEJERA,MARTHA,M
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: TEJERA,MARTHA
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 03/22/2020 19:46:24

Diagnostics Check	OK	19:39
Air Blank	0.000	19:40
Control Test	0.080	19:40
Air Blank	0.000	19:41
Subject Sample # 1	0.217	19:41
Breath Volume	2.304	19:41
Air Blank	0.000	19:42
Air Blank	0.000	19:44
Subject Sample # 2	0.197	19:44
Breath Volume	2.175	19:44
Air Blank	0.000	19:45
Control Test	0.079	19:45
Air Blank	0.000	19:46
Diagnostics Check	OK	19:46

Data Download Date/Time: 03/25/2020 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,MIGUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: HADLEY JR

DATE/TIME OF TEST: 03/23/2020 03:16:46

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.080	03:14	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:15	
Control Test	0.078	03:16	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 03/25/2020 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,JOSUET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CASTILLO,JOSUET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CASTRO DUQUE

DATE/TIME OF TEST: 03/02/2020 04:30:50

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:27	
Control Test	0.082	04:28	
Air Blank	0.000	04:28	
Subject Sample	# 1	04:29	Subject Test Refused
Breath Volume	0.000	04:29	
Air Blank	0.000	04:29	
Control Test	0.081	04:30	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

Data Download Date/Time: 03/25/2020 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ALBERTO,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RIVERA,ALBERTO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:09

SUBJECT'S LAST NAME: FUNTIEO

DATE/TIME OF TEST: 03/01/2020 17:42:51

Diagnostics Check	OK	17:36
Air Blank	0.000	17:36
Control Test	0.080	17:37
Air Blank	0.000	17:37
Subject Sample # 1	0.173	17:38
Breath Volume	2.082	17:38
Air Blank	0.000	17:38
Air Blank	0.000	17:40
Subject Sample # 2	0.168	17:41
Breath Volume	1.667	17:41
Air Blank	0.000	17:41
Control Test	0.080	17:42
Air Blank	0.000	17:42
Diagnostics Check	OK	17:42

Data Download Date/Time: 03/25/2020 11:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ALBERTO,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ROMERO,JANETTE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: RESENDIZ

DATE/TIME OF TEST: 03/08/2020 18:05:52

Diagnostics Check	OK	18:03	
Air Blank	0.000	18:03	
Control Test	0.081	18:03	
Air Blank	0.000	18:04	
Subject Sample	# 1	18:04	Subject Test Refused
Breath Volume	0.000	18:04	
Air Blank	0.000	18:04	
Control Test	0.080	18:05	
Air Blank	0.000	18:05	
Diagnostics Check	OK	18:05	

Data Download Date/Time: 03/25/2020 11:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUADRA,MARVIN,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: POLA,ANTONIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: DAVILA

DATE/TIME OF TEST: 03/29/2020 21:16:35

Diagnostics Check	OK	21:12	
Air Blank	0.000	21:13	
Control Test	0.081	21:13	
Air Blank	0.000	21:13	
Subject Sample	# 1	21:15	Subject Test Refused
Breath Volume	0.000	21:15	
Air Blank	0.000	21:15	
Control Test	0.082	21:16	
Air Blank	0.000	21:16	
Diagnostics Check	OK	21:16	

Data Download Date/Time: 04/29/2020 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HUDDLESTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/11/2020 02:39:00

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.081	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.126	02:34
Breath Volume	3.171	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.141	02:37
Breath Volume	3.117	02:37
Air Blank	0.000	02:38
Control Test	0.081	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:39

Data Download Date/Time: 04/23/2020 20:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REQUENA,CARLOS,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: PULSIFER,MARK
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/11/2020 19:20:06

Diagnostics Check	OK	19:10
Air Blank	0.000	19:11
Control Test	0.081	19:11
Air Blank	0.000	19:12
Subject Sample # 1	0.185	19:12
Breath Volume	2.609	19:12
Air Blank	0.000	19:13
Air Blank	0.000	19:15
Subject Sample # 2	0.213	19:15
Breath Volume	2.261	19:15
Air Blank	0.000	19:16
Air Blank	0.000	19:18
Subject Sample # 3	0.204	19:18
Breath Volume	1.707	19:18
Air Blank	0.000	19:19
Control Test	0.080	19:19
Air Blank	0.000	19:20
Diagnostics Check	OK	19:20

Data Download Date/Time: 04/23/2020 20:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GUY,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: STRANG

DATE/TIME OF TEST: 03/11/2020 20:54:00

Diagnostics Check	OK	20:51	
Air Blank	0.000	20:51	
Control Test	0.081	20:51	
Air Blank	0.000	20:52	
Subject Sample	# 1	20:52	Subject Test Refused
Breath Volume	0.000	20:52	
Air Blank	0.000	20:53	
Control Test	0.080	20:53	
Air Blank	0.000	20:53	
Diagnostics Check	OK	20:54	

Data Download Date/Time: 04/23/2020 20:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: COSTI

DATE/TIME OF TEST: 03/13/2020 02:25:11

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.081	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.000	02:21
Breath Volume	1.847	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.000	02:23
Breath Volume	1.867	02:23
Air Blank	0.000	02:24
Control Test	0.080	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 04/23/2020 20:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: COLUMBUS

DATE/TIME OF TEST: 03/14/2020 00:12:34

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.081	00:07
Air Blank	0.000	00:07
Subject Sample # 1	0.035	00:07
Breath Volume	1.531	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.033	00:11
Breath Volume	1.371	00:11
Air Blank	0.000	00:11
Control Test	0.081	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 04/23/2020 20:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: LANKFORD

DATE/TIME OF TEST: 03/14/2020 06:01:52

Diagnostics Check	OK	05:55
Air Blank	0.000	05:55
Control Test	0.081	05:56
Air Blank	0.000	05:56
Subject Sample # 1	0.199	05:57
Breath Volume	2.128	05:57
Air Blank	0.000	05:57
Air Blank	0.000	05:59
Subject Sample # 2	0.199	06:00
Breath Volume	2.046	06:00
Air Blank	0.000	06:00
Control Test	0.081	06:01
Air Blank	0.000	06:01
Diagnostics Check	OK	06:01

Data Download Date/Time: 04/23/2020 20:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEHR,GEORGE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: DUDENKO

DATE/TIME OF TEST: 03/16/2020 19:00:58

Diagnostics Check	OK	18:57	
Air Blank	0.000	18:58	
Control Test	0.081	18:58	
Air Blank	0.000	18:59	
Subject Sample	# 1	18:59	Subject Test Refused
Breath Volume	0.000	18:59	
Air Blank	0.000	19:00	
Control Test	0.080	19:00	
Air Blank	0.000	19:00	
Diagnostics Check	OK	19:00	

Data Download Date/Time: 04/23/2020 20:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,CORPORAL WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,CORPORAL WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/17/2020 00:13:22

Diagnostics Check	OK	00:07
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:08
Subject Sample # 1	0.117	00:08
Breath Volume	1.964	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.115	00:11
Breath Volume	1.597	00:11
Air Blank	0.000	00:12
Control Test	0.081	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 04/23/2020 20:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAMBEL,ZOE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: CIONI

DATE/TIME OF TEST: 03/17/2020 04:16:10

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.081	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.259	04:11
Breath Volume	1.898	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:13
Subject Sample # 2	0.255	04:14
Breath Volume	1.949	04:14
Air Blank	0.000	04:15
Control Test	0.081	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 04/23/2020 20:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARBANAS,SHANE,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ARBANAS,SHANE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: CATALDI

DATE/TIME OF TEST: 03/19/2020 21:40:38

Diagnostics Check	OK	21:37	
Air Blank	0.000	21:38	
Control Test	0.081	21:38	
Air Blank	0.000	21:38	
Subject Sample	# 1	21:39	Subject Test Refused
Breath Volume	0.000	21:39	
Air Blank	0.000	21:39	
Control Test	0.081	21:40	
Air Blank	0.000	21:40	
Diagnostics Check	OK	21:40	

Data Download Date/Time: 04/23/2020 20:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TIMMY,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: DECAMP

DATE/TIME OF TEST: 03/19/2020 22:24:45

Diagnostics Check	OK	22:24	
Air Blank		22:24	RFI Detect
Air Blank		22:24	RFI Detect

Data Download Date/Time: 04/23/2020 20:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TIMMY,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: DECAMP

DATE/TIME OF TEST: 03/19/2020 22:27:55

Diagnostics Check	OK	22:27	
Air Blank		22:27	RFI Detect
Air Blank		22:27	RFI Detect

Data Download Date/Time: 04/23/2020 20:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TIMMY,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: DECAMP

DATE/TIME OF TEST: 03/19/2020 22:30:41

Diagnostics Check	OK	22:30	
Air Blank		22:30	RFI Detect
Air Blank		22:30	RFI Detect

Data Download Date/Time: 04/23/2020 20:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TIMMY,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: DECAMP

DATE/TIME OF TEST: 03/19/2020 22:34:20

Diagnostics Check	OK	22:34	
Air Blank		22:34	RFI Detect
Air Blank		22:34	RFI Detect

Data Download Date/Time: 04/23/2020 20:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TIMMY,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: DECAMP

DATE/TIME OF TEST: 03/19/2020 22:41:54

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:39	
Control Test	0.081	22:39	
Air Blank	0.000	22:40	
Subject Sample	# 1	22:40	Subject Test Refused
Breath Volume	0.000	22:40	
Air Blank	0.000	22:40	
Control Test	0.081	22:41	
Air Blank	0.000	22:41	
Diagnostics Check	OK	22:41	

Data Download Date/Time: 04/23/2020 20:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARBANAS,SHANE,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ARBANAS,SHANE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: WILLINGHAM

DATE/TIME OF TEST: 03/21/2020 04:27:56

Diagnostics Check		OK	04:21
Air Blank		0.000	04:22
Control Test		0.081	04:22
Air Blank		0.000	04:22
Subject Sample	# 1	0.088	04:23
Breath Volume		1.761	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:25
Subject Sample	# 2	0.086	04:26
Breath Volume		1.578	04:26
Air Blank		0.000	04:27
Control Test		0.080	04:27
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 04/23/2020 20:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 03/21/2020 06:08:15

Diagnostics Check	OK	06:05	
Air Blank	0.000	06:05	
Control Test	0.081	06:06	
Air Blank	0.000	06:06	
Subject Sample	# 1	06:06	Subject Test Refused
Breath Volume	0.000	06:06	
Air Blank	0.000	06:07	
Control Test	0.081	06:07	
Air Blank	0.000	06:08	
Diagnostics Check	OK	06:08	

Data Download Date/Time: 04/23/2020 20:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO-FELIX,DEP,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUGO-FELIX,DEP
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:03

SUBJECT'S LAST NAME: LIVERNOIS

DATE/TIME OF TEST: 03/22/2020 06:33:08

Diagnostics Check	OK	06:25
Air Blank	0.000	06:26
Control Test	0.082	06:26
Air Blank	0.000	06:27
Subject Sample # 1	0.092	06:28
Breath Volume	1.394	06:28
Air Blank	0.000	06:29
Air Blank	0.000	06:30
Subject Sample # 2	0.096	06:31
Breath Volume	2.019	06:31
Air Blank	0.000	06:32
Control Test	0.081	06:32
Air Blank	0.000	06:33
Diagnostics Check	OK	06:33

Data Download Date/Time: 04/23/2020 20:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SNAVELY,KENT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 06:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 03/22/2020 07:28:27

Diagnostics Check	OK	07:21
Air Blank	0.000	07:22
Control Test	0.081	07:22
Air Blank	0.000	07:23
Subject Sample # 1	0.000	07:23
Breath Volume	1.289	07:23
Air Blank	0.000	07:24
Air Blank	0.000	07:26
Subject Sample # 2	0.000	07:27
Breath Volume	1.210	07:27
Air Blank	0.000	07:27
Control Test	0.081	07:27
Air Blank	0.000	07:28
Diagnostics Check	OK	07:28

Data Download Date/Time: 04/23/2020 20:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: LANDIS

DATE/TIME OF TEST: 03/25/2020 22:14:34

Diagnostics Check	OK	22:10	
Air Blank	0.000	22:10	
Control Test	0.081	22:10	
Air Blank	0.000	22:11	
Subject Sample	# 1	22:13	Subject Test Refused
Breath Volume	0.000	22:13	
Air Blank	0.000	22:13	
Control Test	0.080	22:13	
Air Blank	0.000	22:14	
Diagnostics Check	OK	22:14	

Data Download Date/Time: 04/23/2020 20:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARBANAS,SHANE,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ARBANAS,SHANE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: MCPHILAMY

DATE/TIME OF TEST: 03/26/2020 01:30:54

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.080	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.171	01:26
Breath Volume	1.851	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:28
Subject Sample # 2	0.174	01:29
Breath Volume	1.824	01:29
Air Blank	0.000	01:29
Control Test	0.080	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 04/23/2020 20:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: HECKER

DATE/TIME OF TEST: 03/26/2020 21:57:28

Diagnostics Check	OK	21:51
Air Blank	0.000	21:51
Control Test	0.081	21:52
Air Blank	0.000	21:52
Subject Sample # 1	0.110	21:53
Breath Volume	1.437	21:53
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.121	21:55
Breath Volume	2.253	21:55
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

Data Download Date/Time: 04/23/2020 20:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BUSCH,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: TOWNER

DATE/TIME OF TEST: 03/31/2020 16:23:04

Diagnostics Check		OK	16:14
Air Blank		0.000	16:15
Control Test		0.080	16:15
Air Blank		0.000	16:15
Subject Sample	# 1	0.252	16:17
Breath Volume		1.402	16:17
Air Blank		0.000	16:17
Air Blank		0.000	16:19
Subject Sample	# 2	0.250	16:21
Breath Volume		1.816	16:21
Air Blank		0.000	16:22
Control Test		0.080	16:22
Air Blank		0.000	16:22
Diagnostics Check		OK	16:23

Data Download Date/Time: 04/23/2020 20:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1060911
BREATH TEST OPERATOR: CORBIN,BRITTON,L	DRY GAS EXPIRY DATE: 12/11/2021
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: CORBIN,BRITTON	ARREST TIME: 17:45
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 02/20/2020	OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: MILLER	DATE/TIME OF TEST: 03/26/2020	18:35:03
-----------------------------	-------------------------------	----------

Diagnostics Check	OK	18:28
Air Blank	0.000	18:29
Control Test	0.080	18:29
Air Blank	0.000	18:29
Subject Sample # 1	0.128	18:30
Breath Volume	3.550	18:30
Air Blank	0.000	18:31
Air Blank	0.000	18:32
Subject Sample # 2	0.126	18:33
Breath Volume	3.628	18:33
Air Blank	0.000	18:34
Control Test	0.080	18:34
Air Blank	0.000	18:34
Diagnostics Check	OK	18:35

Data Download Date/Time: 03/28/2020 18:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD,NICOLE,M
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BASFORD,NICOLE
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 02/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/21/2021
VIOLATION CODE: BUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 03/13/2020 19:02:33

Diagnostics Check	OK	18:58	
Air Blank	0.000	18:59	
Control Test	0.080	18:59	
Air Blank	0.000	18:59	
Subject Sample	# 1	19:01	Subject Test Refused
Breath Volume	0.000	19:01	
Air Blank	0.000	19:01	
Control Test	0.078	19:01	
Air Blank	0.000	19:02	
Diagnostics Check	OK	19:02	

Data Download Date/Time: 03/28/2020 18:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRASER, THOMAS, J
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: RODRIGUEZ, FRANK
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BOREK

DATE/TIME OF TEST: 03/04/2020 04:12:00

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.079	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.339	04:07
Breath Volume	2.152	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.350	04:10
Breath Volume	2.695	04:10
Air Blank	0.000	04:11
Control Test	0.078	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:12

Data Download Date/Time: 03/26/2020 12:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,PAUL,G
BREATH TEST OPERATOR AGENCY: GULF COUNTY SO
ARRESTING OFFICER: PAGE,JESSICA
ARRESTING OFFICER AGENCY: GULF COUNTY SO
LAST AGENCY INSPECTION: 02/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: HOOD

DATE/TIME OF TEST: 03/07/2020 00:58:21

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.079	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.126	00:53
Breath Volume	2.781	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.128	00:56
Breath Volume	1.457	00:56
Air Blank	0.000	00:57
Control Test	0.079	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 03/17/2020 16:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HEERDEN, TY, C
BREATH TEST OPERATOR AGENCY: PORT ST. JOE PD
ARRESTING OFFICER: RICHARDS, JAKE
ARRESTING OFFICER AGENCY: PORT ST. JOE PD
LAST AGENCY INSPECTION: 03/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: HOLLIS

DATE/TIME OF TEST: 03/29/2020 20:56:18

Diagnostics Check		OK	20:48
Air Blank		0.000	20:49
Control Test		0.078	20:49
Air Blank		0.000	20:50
Subject Sample	# 1	0.194	20:50
Breath Volume		1.160	20:50
Air Blank		0.000	20:50
Air Blank		0.000	20:52
Subject Sample	# 2	0.188	20:54
Breath Volume		1.328	20:54
Air Blank		0.000	20:55
Control Test		0.079	20:55
Air Blank		0.000	20:56
Diagnostics Check		OK	20:56

Data Download Date/Time: 04/28/2020 18:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARRILLO,N,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SUTPHIN,S
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 03/05/2020 00:17:32

Diagnostics Check		OK	00:10
Air Blank		0.000	00:10
Control Test		0.080	00:11
Air Blank		0.000	00:11
Subject Sample	# 1	0.245	00:12
Breath Volume		1.191	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.249	00:15
Breath Volume		1.320	00:15
Air Blank		0.000	00:16
Control Test		0.078	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 04/08/2020 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KIRKBY,S
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: SAULTZ

DATE/TIME OF TEST: 03/08/2020 01:50:15

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.135	01:45
Breath Volume	2.878	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.130	01:48
Breath Volume	3.367	01:48
Air Blank	0.000	01:49
Control Test	0.081	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 04/08/2020 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HIDEG,JARED
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: TURQUIA

DATE/TIME OF TEST: 03/09/2020 00:57:39

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.080	00:52
Air Blank	0.000	00:52
Subject Sample # 1	0.000	00:53
Breath Volume	3.277	00:53
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.000	00:56
Breath Volume	2.843	00:56
Air Blank	0.000	00:56
Control Test	0.081	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 04/08/2020 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSEPH,G
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BELL,JOSEPH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 03/12/2020 22:10:56

Diagnostics Check		OK	22:04
Air Blank		0.000	22:04
Control Test		0.081	22:04
Air Blank		0.000	22:05
Subject Sample	# 1	0.106	22:06
Breath Volume		1.605	22:06
Air Blank		0.000	22:07
Air Blank		0.000	22:08
Subject Sample	# 2	0.103	22:09
Breath Volume		1.710	22:09
Air Blank		0.000	22:10
Control Test		0.079	22:10
Air Blank		0.000	22:10
Diagnostics Check		OK	22:10

Data Download Date/Time: 04/08/2020 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONIECZKI,D,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KONIECZKI,D
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/13/2020 16:17:38

Diagnostics Check	OK	16:11
Air Blank	0.000	16:11
Control Test	0.080	16:11
Air Blank	0.000	16:12
Subject Sample # 1	0.142	16:13
Breath Volume	3.437	16:13
Air Blank	0.000	16:13
Air Blank	0.000	16:15
Subject Sample # 2	0.138	16:16
Breath Volume	1.707	16:16
Air Blank	0.000	16:16
Control Test	0.080	16:17
Air Blank	0.000	16:17
Diagnostics Check	OK	16:17

Data Download Date/Time: 04/08/2020 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONIECZKI,D,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KONIECZKI,D
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: CUPP

DATE/TIME OF TEST: 03/14/2020 01:19:59

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.081	01:14
Air Blank	0.000	01:14
Subject Sample # 1	0.146	01:15
Breath Volume	1.347	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:17
Subject Sample # 2	0.150	01:18
Breath Volume	1.687	01:18
Air Blank	0.000	01:19
Control Test	0.079	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 04/08/2020 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICOLAS,BENCHY,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: NICOLAS,BENCHY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: LINDSLY

DATE/TIME OF TEST: 03/18/2020 03:46:21

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:45	
Control Test	0.080	03:45	
Air Blank		03:45	Improper Sample
Air Blank		03:46	Improper Sample

Data Download Date/Time: 04/08/2020 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICOLAS,B,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: NICOLAS,BENCHY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: LINDLEY

DATE/TIME OF TEST: 03/18/2020 03:49:41

Diagnostics Check	OK	03:49	
Air Blank		03:49	Improper Sample
Air Blank		03:49	Improper Sample

Data Download Date/Time: 04/08/2020 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSEPH,G
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: NICOLAS,BENCHY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: LINDSLY

DATE/TIME OF TEST: 03/18/2020 03:56:15

Diagnostics Check	OK	03:55	
Air Blank		03:55	Ambient Fail
Air Blank	0.000	03:56	

Data Download Date/Time: 04/08/2020 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRANO,ROMAN,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SERRANO,ROMAN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 03/24/2020 15:57:01

Diagnostics Check	OK	15:51
Air Blank	0.000	15:51
Control Test	0.081	15:52
Air Blank	0.000	15:52
Subject Sample # 1	0.000	15:52
Breath Volume	1.878	15:52
Air Blank	0.000	15:53
Air Blank	0.000	15:55
Subject Sample # 2	0.000	15:55
Breath Volume	2.765	15:55
Air Blank	0.000	15:56
Control Test	0.079	15:56
Air Blank	0.000	15:56
Diagnostics Check	OK	15:57

Data Download Date/Time: 04/08/2020 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAEFER,KENDALL,A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SCHAEFER,KENDALL
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:09

SUBJECT'S LAST NAME: MATEO

DATE/TIME OF TEST: 03/28/2020 06:46:18

Diagnostics Check	OK	06:38
Air Blank	0.000	06:39
Control Test	0.080	06:39
Air Blank	0.000	06:40
Subject Sample # 1	0.147	06:41
Breath Volume	2.308	06:41
Air Blank	0.000	06:41
Air Blank	0.000	06:43
Subject Sample # 2	0.144	06:44
Breath Volume	2.710	06:44
Air Blank	0.000	06:45
Control Test	0.079	06:45
Air Blank	0.000	06:46
Diagnostics Check	OK	06:46

Data Download Date/Time: 04/08/2020 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINDHOLM,JOHN,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINDHOLM,JOHN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 03/29/2020 05:42:39

Diagnostics Check		OK	05:35
Air Blank		0.000	05:35
Control Test		0.080	05:35
Air Blank		0.000	05:36
Subject Sample	# 1	0.138	05:37
Breath Volume		1.300	05:37
Air Blank		0.000	05:38
Air Blank		0.000	05:39
Subject Sample	# 2	0.119	05:40
Breath Volume		1.691	05:40
Air Blank		0.000	05:41
Control Test		0.079	05:42
Air Blank		0.000	05:42
Diagnostics Check		OK	05:42

Data Download Date/Time: 04/08/2020 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SELLERS 11-101,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: PHAM

DATE/TIME OF TEST: 03/30/2020 02:38:46

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:23	
Control Test	0.081	02:24	
Air Blank	0.000	02:24	
Subject Sample # 1	0.186	02:27	
Breath Volume	1.761	02:27	
Air Blank	0.000	02:28	
Air Blank	0.000	02:30	
Subject Sample # 2	0.153	02:33	Volume Not Met
Breath Volume	0.218	02:33	
Air Blank	0.000	02:34	
Air Blank	0.000	02:35	
Subject Sample # 3		02:37	Subject Test Refused
Breath Volume	0.359	02:37	
Air Blank	0.000	02:37	
Control Test	0.078	02:38	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 04/08/2020 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURTSINGER,C,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: AUSTIN,B
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: TAPANES

DATE/TIME OF TEST: 03/01/2020 17:54:49

Diagnostics Check	OK	17:48
Air Blank	0.000	17:49
Control Test	0.081	17:49
Air Blank	0.000	17:49
Subject Sample # 1	0.000	17:50
Breath Volume	1.855	17:50
Air Blank	0.000	17:51
Air Blank	0.000	17:53
Subject Sample # 2	0.000	17:53
Breath Volume	1.812	17:53
Air Blank	0.000	17:53
Control Test	0.081	17:54
Air Blank	0.000	17:54
Diagnostics Check	OK	17:54

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRATT,N,J
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PRATT,N
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/06/2020 04:12:28

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.180	04:07
Breath Volume	1.492	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:09
Subject Sample # 2	0.179	04:10
Breath Volume	2.625	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: DWELLE

DATE/TIME OF TEST: 03/07/2020 02:55:54

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.080	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.064	02:50
Breath Volume	1.132	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.058	02:54
Breath Volume	1.214	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,P,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MILLER,K
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 03/07/2020 04:38:33

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.080	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.149	04:33
Breath Volume	1.625	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.135	04:36
Breath Volume	1.359	04:36
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUONPANE,V,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: DIAZ,S
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: CARRILLO

DATE/TIME OF TEST: 03/07/2020 05:02:36

Diagnostics Check	OK	04:55
Air Blank	0.000	04:56
Control Test	0.080	04:56
Air Blank	0.000	04:57
Subject Sample # 1	0.091	04:58
Breath Volume	1.894	04:58
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.092	05:01
Breath Volume	1.714	05:01
Air Blank	0.000	05:01
Control Test	0.080	05:02
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DECKER,STEPHEN,J
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: SOOSTEN,V
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: LLOYD

DATE/TIME OF TEST: 03/07/2020 16:03:23

Diagnostics Check	OK	15:53	
Air Blank	0.000	15:54	
Control Test	0.080	15:54	
Air Blank	0.000	15:54	
Subject Sample	# 1	15:57	No Sample Provided
Breath Volume	0.000	15:57	
Air Blank	0.000	15:58	
Air Blank	0.000	16:00	
Subject Sample	# 2	16:02	Subject Test Refused
Breath Volume	0.000	16:02	
Air Blank	0.000	16:02	
Control Test	0.081	16:02	
Air Blank	0.000	16:03	
Diagnostics Check	OK	16:03	

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARVIN,ANGELINA,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GROENDYK,R
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 03/07/2020 16:30:27

Diagnostics Check	OK	16:22
Air Blank	0.000	16:22
Control Test	0.080	16:22
Air Blank	0.000	16:23
Subject Sample # 1	0.258	16:25
Breath Volume	1.437	16:25
Air Blank	0.000	16:26
Air Blank	0.000	16:28
Subject Sample # 2	0.264	16:28
Breath Volume	2.000	16:28
Air Blank	0.000	16:29
Control Test	0.080	16:29
Air Blank	0.000	16:30
Diagnostics Check	OK	16:30

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THROESCH,PATRICK,J
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: WOODS,HUNTER
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: SHANNON

DATE/TIME OF TEST: 03/08/2020 03:48:56

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.191	03:44
Breath Volume	2.289	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:46
Subject Sample # 2	0.207	03:47
Breath Volume	2.019	03:47
Air Blank	0.000	03:48
Control Test	0.079	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,S,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: RAMOS,Y
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: MARTINEZ-VILLASANA

DATE/TIME OF TEST: 03/08/2020 04:48:21

Diagnostics Check	OK	04:41
Air Blank	0.000	04:42
Control Test	0.080	04:42
Air Blank	0.000	04:42
Subject Sample # 1	0.228	04:43
Breath Volume	2.386	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.214	04:46
Breath Volume	1.644	04:46
Air Blank	0.000	04:47
Control Test	0.079	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLMSTEAD,J,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: WOOD,H
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: MULLEN

DATE/TIME OF TEST: 03/09/2020 02:11:25

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.246	02:06
Breath Volume	1.855	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.233	02:09
Breath Volume	2.265	02:09
Air Blank	0.000	02:10
Control Test	0.080	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOEL,Y,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: NOVA,M
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: ALONSO

DATE/TIME OF TEST: 03/09/2020 19:58:58

Diagnostics Check	OK	19:55	
Air Blank	0.000	19:56	
Control Test	0.079	19:56	
Air Blank	0.000	19:57	
Subject Sample	# 1	19:57	Subject Test Refused
Breath Volume	0.000	19:57	
Air Blank	0.000	19:58	
Control Test	0.080	19:58	
Air Blank	0.000	19:58	
Diagnostics Check	OK	19:58	

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,THOMAS,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: DARREN,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 03/09/2020 21:30:05

Diagnostics Check	OK	21:15	
Air Blank	0.000	21:15	
Control Test	0.080	21:15	
Air Blank	0.000	21:16	
Subject Sample	# 1 0.193	21:17	
Breath Volume	1.128	21:17	
Air Blank	0.000	21:17	
Air Blank	0.000	21:19	
Subject Sample	# 2 0.171	21:22	Volume Not Met
Breath Volume	0.699	21:22	
Air Blank	0.000	21:23	
Air Blank	0.000	21:25	
Subject Sample	# 3 0.188	21:28	Volume Not Met
Breath Volume	0.832	21:28	
Air Blank	0.000	21:29	
Control Test	0.078	21:29	
Air Blank	0.000	21:30	
Diagnostics Check	OK	21:30	

Data Download Date/Time: 03/12/2020 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRILLO,XAVIER,R
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CARRILLO,XAVIER
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 03/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: MUZO

DATE/TIME OF TEST: 03/16/2020 22:53:14

Diagnostics Check	OK	22:45
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.221	22:48
Breath Volume	1.238	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.225	22:51
Breath Volume	1.171	22:51
Air Blank	0.000	22:52
Control Test	0.079	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 04/08/2020 11:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATSON,ANDRAE,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PEREZ,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 03/16/2020 03:03:08

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.082	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.326	02:55
Breath Volume		1.269	02:55
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.269	02:58
Breath Volume		1.199	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 3	0.317	03:01
Breath Volume		1.683	03:01
Air Blank		0.000	03:02
Control Test		0.080	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 04/13/2020 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONSEES,D,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HEMDERSON,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: SOLER

DATE/TIME OF TEST: 03/16/2020 22:37:41

Diagnostics Check	OK	22:31
Air Blank	0.000	22:31
Control Test	0.082	22:31
Air Blank	0.000	22:32
Subject Sample # 1	0.051	22:32
Breath Volume	1.492	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.050	22:36
Breath Volume	1.324	22:36
Air Blank	0.000	22:36
Control Test	0.082	22:37
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 04/13/2020 15:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,S,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ALUSMA,J
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: KUNCHAL

DATE/TIME OF TEST: 03/19/2020 02:44:22

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.082	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.185	02:40
Breath Volume	1.339	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.195	02:42
Breath Volume	1.273	02:42
Air Blank	0.000	02:43
Control Test	0.081	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 04/13/2020 15:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PEREZ,M
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BADHAM

DATE/TIME OF TEST: 03/20/2020 00:29:54

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.081	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.037	00:25
Breath Volume	1.765	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:27
Subject Sample # 2	0.035	00:28
Breath Volume	2.832	00:28
Air Blank	0.000	00:29
Control Test	0.082	00:29
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 04/13/2020 15:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: RAMOS, Y
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: HUERTA

DATE/TIME OF TEST: 03/21/2020 04:03:17

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.082	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.108	03:58
Breath Volume	2.730	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.103	04:01
Breath Volume	1.835	04:01
Air Blank	0.000	04:02
Control Test	0.082	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 04/13/2020 15:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: RAMOSRAMOS

DATE/TIME OF TEST: 03/22/2020 03:11:42

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.082	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.105	03:07
Breath Volume	1.949	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.107	03:10
Breath Volume	1.941	03:10
Air Blank	0.000	03:10
Control Test	0.082	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 04/13/2020 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAFFEE,C,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WARGO,T
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: PIZICAS

DATE/TIME OF TEST: 03/22/2020 19:09:58

Diagnostics Check	OK	19:00
Air Blank	0.000	19:01
Control Test	0.081	19:01
Air Blank	0.000	19:02
Subject Sample # 1	0.139	19:04
Breath Volume	1.140	19:04
Air Blank	0.000	19:04
Air Blank	0.000	19:06
Subject Sample # 2	0.131	19:08
Breath Volume	2.035	19:08
Air Blank	0.000	19:09
Control Test	0.082	19:09
Air Blank	0.000	19:09
Diagnostics Check	OK	19:09

Data Download Date/Time: 04/13/2020 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERMAN,BRENT,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: FORD,L
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: DIEGOJUAN

DATE/TIME OF TEST: 03/22/2020 21:07:29

Diagnostics Check	OK	21:01
Air Blank	0.000	21:01
Control Test	0.081	21:02
Air Blank	0.000	21:02
Subject Sample # 1	0.225	21:03
Breath Volume	2.398	21:03
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.225	21:05
Breath Volume	2.605	21:05
Air Blank	0.000	21:06
Control Test	0.081	21:06
Air Blank	0.000	21:07
Diagnostics Check	OK	21:07

Data Download Date/Time: 04/13/2020 15:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JOSE,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MASSEY,REBECCA
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: SENAT

DATE/TIME OF TEST: 03/23/2020 22:55:53

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.081	22:50
Air Blank	0.000	22:50
Subject Sample # 1	0.223	22:51
Breath Volume	1.917	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:53
Subject Sample # 2	0.236	22:54
Breath Volume	2.425	22:54
Air Blank	0.000	22:54
Control Test	0.082	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 04/13/2020 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DECKER,DAVID,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: COMBS,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:49

SUBJECT'S LAST NAME: MCAULIFFE

DATE/TIME OF TEST: 03/25/2020 16:20:18

Diagnostics Check		OK	16:13
Air Blank		0.000	16:14
Control Test		0.081	16:14
Air Blank		0.000	16:14
Subject Sample	# 1	0.255	16:15
Breath Volume		2.070	16:15
Air Blank		0.000	16:16
Air Blank		0.000	16:17
Subject Sample	# 2	0.247	16:18
Breath Volume		1.976	16:18
Air Blank		0.000	16:19
Control Test		0.081	16:19
Air Blank		0.000	16:20
Diagnostics Check		OK	16:20

Data Download Date/Time: 04/13/2020 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JOSE,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MASSEY,REBECCA
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: POSADA

DATE/TIME OF TEST: 03/28/2020 21:54:07

Diagnostics Check	OK	21:47
Air Blank	0.000	21:48
Control Test	0.081	21:48
Air Blank	0.000	21:49
Subject Sample # 1	0.145	21:49
Breath Volume	2.527	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.143	21:52
Breath Volume	3.285	21:52
Air Blank	0.000	21:53
Control Test	0.081	21:53
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

Data Download Date/Time: 04/13/2020 15:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,P,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ALTER,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: KRIEGER

DATE/TIME OF TEST: 03/30/2020 23:19:45

Diagnostics Check	OK	23:12
Air Blank	0.000	23:12
Control Test	0.082	23:12
Air Blank	0.000	23:13
Subject Sample # 1	0.252	23:14
Breath Volume	3.703	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.250	23:18
Breath Volume	3.105	23:18
Air Blank	0.000	23:18
Control Test	0.081	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 04/13/2020 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADRUGO,FELICIA,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CHARLES,CLEEFORD
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 02/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: KIGER

DATE/TIME OF TEST: 03/03/2020 00:31:52

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.076	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.311	00:27
Breath Volume	1.195	00:27
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.316	00:30
Breath Volume	1.445	00:30
Air Blank	0.000	00:30
Control Test	0.076	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 03/30/2020 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA, JOSUE, D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CHARLES, C
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 02/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: KIGER

DATE/TIME OF TEST: 03/03/2020 11:21:49

Diagnostics Check		OK	11:15
Air Blank		0.000	11:16
Control Test		0.076	11:16
Air Blank		0.000	11:17
Subject Sample	# 1	0.135	11:17
Breath Volume		1.300	11:17
Air Blank		0.000	11:18
Air Blank		0.000	11:19
Subject Sample	# 2	0.141	11:20
Breath Volume		1.359	11:20
Air Blank		0.000	11:20
Control Test		0.076	11:21
Air Blank		0.000	11:21
Diagnostics Check		OK	11:21

Data Download Date/Time: 03/30/2020 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA, JOSUE, D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CHARLES, C
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 02/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: KIGER

DATE/TIME OF TEST: 03/03/2020 15:28:31

Diagnostics Check	OK	15:22
Air Blank	0.000	15:23
Control Test	0.076	15:23
Air Blank	0.000	15:23
Subject Sample # 1	0.070	15:24
Breath Volume	1.796	15:24
Air Blank	0.000	15:24
Air Blank	0.000	15:26
Subject Sample # 2	0.066	15:26
Breath Volume	1.601	15:26
Air Blank	0.000	15:27
Control Test	0.076	15:27
Air Blank	0.000	15:28
Diagnostics Check	OK	15:28

Data Download Date/Time: 03/30/2020 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WESTON,LARRY,L
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: YOUNGMAN,TALON
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 02/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: B400210994080

DATE/TIME OF TEST: 03/15/2020 16:16:00

Diagnostics Check	OK	16:09
Air Blank	0.000	16:09
Control Test	0.076	16:09
Air Blank	0.000	16:10
Subject Sample # 1	0.000	16:11
Breath Volume	1.515	16:11
Air Blank	0.000	16:11
Air Blank	0.000	16:13
Subject Sample # 2	0.000	16:14
Breath Volume	1.296	16:14
Air Blank	0.000	16:15
Control Test	0.077	16:15
Air Blank	0.000	16:15
Diagnostics Check	OK	16:16

Data Download Date/Time: 03/30/2020 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENING,FRED,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: LEVESQUE

DATE/TIME OF TEST: 03/07/2020 20:26:50

Diagnostics Check	OK	20:20
Air Blank	0.000	20:20
Control Test	0.081	20:21
Air Blank	0.000	20:21
Subject Sample # 1	0.180	20:22
Breath Volume	1.601	20:22
Air Blank	0.000	20:23
Air Blank	0.000	20:24
Subject Sample # 2	0.174	20:25
Breath Volume	1.429	20:25
Air Blank	0.000	20:25
Control Test	0.079	20:26
Air Blank	0.000	20:26
Diagnostics Check	OK	20:26

Data Download Date/Time: 04/01/2020 14:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENING,FRED,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HENING,FRED
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: WOLLSCHLAGER

DATE/TIME OF TEST: 03/07/2020 20:20:58

Diagnostics Check	OK	20:18	
Air Blank	0.000	20:18	
Control Test	0.079	20:18	
Air Blank	0.000	20:19	
Subject Sample	# 1	20:19	Subject Test Refused
Breath Volume	0.000	20:19	
Air Blank	0.000	20:20	
Control Test	0.079	20:20	
Air Blank	0.000	20:20	
Diagnostics Check	OK	20:20	

Data Download Date/Time: 03/19/2020 13:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,RONDRICKA,N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: THOMPSON,RONDRICKA
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 03/18/2020 03:18:00

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:15	
Control Test	0.079	03:15	
Air Blank	0.000	03:16	
Subject Sample	# 1	03:16	Subject Test Refused
Breath Volume	0.000	03:16	
Air Blank	0.000	03:17	
Control Test	0.080	03:17	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:18	

Data Download Date/Time: 03/19/2020 13:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,FAUSTO,A
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: ROSARIO,FAUSTO
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 03/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: DURAN GARIN

DATE/TIME OF TEST: 03/21/2020 22:10:14

Diagnostics Check	OK	22:08	
Air Blank	0.000	22:09	
Control Test	0.000	22:09	Control Outside Tolerance
Air Blank	0.000	22:10	

Data Download Date/Time: 04/07/2020 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,FAUSTO,A
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: ROSARIO,FAUSTO
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 03/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: DURAN GARIN

DATE/TIME OF TEST: 03/21/2020 22:13:44

Diagnostics Check	OK	22:12	
Air Blank	0.000	22:12	
Control Test	0.000	22:13	Control Outside Tolerance
Air Blank	0.000	22:13	

Data Download Date/Time: 04/07/2020 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,FAUSTO,A
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: ROSARIO,FAUSTO
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 03/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: DURAN GARIN

DATE/TIME OF TEST: 03/21/2020 22:21:25

Diagnostics Check	OK	22:19	
Air Blank	0.000	22:20	
Control Test	0.000	22:20	Control Outside Tolerance
Air Blank	0.000	22:21	

Data Download Date/Time: 04/07/2020 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,FAUSTO,A
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: SHAW,ROBERT
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 03/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: DURAN GARIN

DATE/TIME OF TEST: 03/21/2020 22:33:04

Diagnostics Check	OK	22:31	
Air Blank	0.000	22:31	
Control Test	0.000	22:32	Control Outside Tolerance
Air Blank	0.000	22:33	

Data Download Date/Time: 04/07/2020 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,FAUSTO,A
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: SHAW,ROBERT
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 03/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: DURAN GARIN

DATE/TIME OF TEST: 03/21/2020 22:38:40

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:37	
Control Test	0.000	22:38	Control Outside Tolerance
Air Blank	0.000	22:38	

Data Download Date/Time: 04/07/2020 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,FAUSTO,A
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: SHAW,ROBERT
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 03/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: DURAN GARIN

DATE/TIME OF TEST: 03/21/2020 22:49:35

Diagnostics Check	OK	22:47	
Air Blank	0.000	22:48	
Control Test	0.000	22:49	Control Outside Tolerance
Air Blank	0.000	22:49	

Data Download Date/Time: 04/07/2020 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,FAUSTO,A
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: SHAW,ROBERT
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 03/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: DURAN GARIN

DATE/TIME OF TEST: 03/21/2020 22:55:38

Diagnostics Check	OK	22:53	
Air Blank	0.000	22:54	
Control Test	0.000	22:55	Control Outside Tolerance
Air Blank	0.000	22:55	

Data Download Date/Time: 04/07/2020 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,FAUSTO,A
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: SHAW,ROBERT
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 03/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: DURAN GARIN

DATE/TIME OF TEST: 03/21/2020 23:00:20

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:59	
Control Test	0.000	22:59	Control Outside Tolerance
Air Blank	0.000	23:00	

Data Download Date/Time: 04/07/2020 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,FAUSTO,A
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: ROSARIO,FAUSTO
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 03/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: DURAN GARIN

DATE/TIME OF TEST: 03/21/2020 23:13:20

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:12	
Control Test	0.000	23:12	Control Outside Tolerance
Air Blank	0.000	23:13	

Data Download Date/Time: 04/07/2020 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,FAUSTO,A
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: ROSARIO,FAUSTO
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 03/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 03/23/2020 06:23:24

Diagnostics Check	OK	06:21	
Air Blank	0.000	06:22	
Control Test	0.000	06:22	Control Outside Tolerance
Air Blank	0.000	06:23	

Data Download Date/Time: 04/07/2020 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,ROBERT,A
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WESTMAN,PATRICK
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/03/2020 03:20:23

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.199	03:15
Breath Volume	2.054	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.180	03:18
Breath Volume	1.707	03:18
Air Blank	0.000	03:19
Control Test	0.082	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 03/04/2020 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: THOMPSON,JEFFREY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: SUNDAY

DATE/TIME OF TEST: 03/04/2020 01:07:20

Diagnostics Check		OK	01:00
Air Blank		0.000	01:00
Control Test		0.081	01:00
Air Blank		0.000	01:01
Subject Sample	# 1	0.156	01:02
Breath Volume		1.312	01:02
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.166	01:05
Breath Volume		1.332	01:05
Air Blank		0.000	01:06
Control Test		0.082	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 03/04/2020 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: POPPELL

DATE/TIME OF TEST: 03/04/2020 22:10:46

Diagnostics Check	OK	22:04
Air Blank	0.000	22:04
Control Test	0.080	22:04
Air Blank	0.000	22:05
Subject Sample # 1	0.047	22:06
Breath Volume	1.406	22:06
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.054	22:09
Breath Volume	2.140	22:09
Air Blank	0.000	22:09
Control Test	0.081	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 04/01/2020 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISCHER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: FISCHER,ANDREW
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: GODWIN

DATE/TIME OF TEST: 03/06/2020 01:34:23

Diagnostics Check		OK	01:27
Air Blank		0.000	01:28
Control Test		0.081	01:28
Air Blank		0.000	01:29
Subject Sample	# 1	0.175	01:29
Breath Volume		4.101	01:29
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.170	01:32
Breath Volume		4.144	01:32
Air Blank		0.000	01:33
Control Test		0.082	01:33
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 04/01/2020 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: LENSCH

DATE/TIME OF TEST: 03/06/2020 03:03:59

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.081	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.176	02:59
Breath Volume	1.398	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.190	03:02
Breath Volume	1.804	03:02
Air Blank	0.000	03:03
Control Test	0.082	03:03
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 04/01/2020 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,R,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: STEELE,R
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 03/07/2020 03:18:03

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.081	03:12
Air Blank		0.000	03:12
Subject Sample	# 1	0.138	03:13
Breath Volume		4.191	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:15
Subject Sample	# 2	0.138	03:16
Breath Volume		4.679	03:16
Air Blank		0.000	03:17
Control Test		0.083	03:17
Air Blank		0.000	03:17
Diagnostics Check		OK	03:18

Data Download Date/Time: 04/01/2020 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARTHY,TED,G
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: MCCARTHY,TED
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:19

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/07/2020 06:59:01

Diagnostics Check	OK	06:56	
Air Blank	0.000	06:56	
Control Test	0.081	06:56	
Air Blank	0.000	06:57	
Subject Sample	# 1	06:57	Subject Test Refused
Breath Volume	0.000	06:57	
Air Blank	0.000	06:58	
Control Test	0.081	06:58	
Air Blank	0.000	06:58	
Diagnostics Check	OK	06:59	

Data Download Date/Time: 04/01/2020 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITAKER,H,H
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: WHITAKER,H
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 03/08/2020 05:18:55

Diagnostics Check		OK	05:12
Air Blank		0.000	05:13
Control Test		0.081	05:13
Air Blank		0.000	05:13
Subject Sample	# 1	0.268	05:14
Breath Volume		2.550	05:14
Air Blank		0.000	05:15
Air Blank		0.000	05:16
Subject Sample	# 2	0.262	05:17
Breath Volume		2.363	05:17
Air Blank		0.000	05:18
Control Test		0.082	05:18
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

Data Download Date/Time: 04/01/2020 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITAKER,H,H
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: DEMING,W
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:09

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 03/08/2020 06:50:24

Diagnostics Check	OK	06:44
Air Blank	0.000	06:44
Control Test	0.081	06:44
Air Blank	0.000	06:45
Subject Sample # 1	0.178	06:45
Breath Volume	2.488	06:45
Air Blank	0.000	06:46
Air Blank	0.000	06:48
Subject Sample # 2	0.181	06:48
Breath Volume	2.691	06:48
Air Blank	0.000	06:49
Control Test	0.083	06:49
Air Blank	0.000	06:50
Diagnostics Check	OK	06:50

Data Download Date/Time: 04/01/2020 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PILKANIS,MARY,C
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: PILKANIS,MARY
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:16

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 03/08/2020 06:59:50

Diagnostics Check	OK	06:57	
Air Blank	0.000	06:57	
Control Test	0.082	06:57	
Air Blank	0.000	06:58	
Subject Sample	# 1	06:58	Subject Test Refused
Breath Volume	0.000	06:58	
Air Blank	0.000	06:58	
Control Test	0.082	06:59	
Air Blank	0.000	06:59	
Diagnostics Check	OK	06:59	

Data Download Date/Time: 04/01/2020 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,ROBERT,A
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WESTMAN,PATRICK
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: LAMADRIZ

DATE/TIME OF TEST: 03/09/2020 03:45:32

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.082	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.063	03:39
Breath Volume	2.082	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.069	03:44
Breath Volume	4.527	03:44
Air Blank	0.000	03:44
Control Test	0.085	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 04/01/2020 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD,LEE,M
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: SPEARS, TIFFANY
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 03/12/2020 03:39:58

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:37	
Control Test	0.081	03:37	
Air Blank	0.000	03:38	
Subject Sample	# 1	03:38	Subject Test Refused
Breath Volume	0.000	03:38	
Air Blank	0.000	03:39	
Control Test	0.081	03:39	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 04/01/2020 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,MAX,J
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: ANDRADE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: STEELE

DATE/TIME OF TEST: 03/16/2020 20:23:22

Diagnostics Check		OK	20:17
Air Blank		0.000	20:17
Control Test		0.080	20:17
Air Blank		0.000	20:18
Subject Sample	# 1	0.288	20:18
Breath Volume		1.988	20:18
Air Blank		0.000	20:19
Air Blank		0.000	20:21
Subject Sample	# 2	0.279	20:21
Breath Volume		1.988	20:21
Air Blank		0.000	20:22
Control Test		0.080	20:22
Air Blank		0.000	20:23
Diagnostics Check		OK	20:23

Data Download Date/Time: 04/01/2020 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: GAY

DATE/TIME OF TEST: 03/18/2020 03:11:44

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.081	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.000	03:07
Breath Volume	2.941	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.000	03:10
Breath Volume	2.378	03:10
Air Blank	0.000	03:10
Control Test	0.082	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 04/01/2020 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISCHER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: FISCHER,ANDREW
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: CONNER

DATE/TIME OF TEST: 03/18/2020 19:00:43

Diagnostics Check	OK	18:58	
Air Blank	0.000	18:58	
Control Test	0.081	18:58	
Air Blank	0.000	18:59	
Subject Sample	# 1	19:00	Sequence Aborted
Breath Volume	0.000	19:00	
Air Blank	0.000	19:00	

Data Download Date/Time: 04/01/2020 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/21/2020 03:29:26

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.081	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.101	03:25
Breath Volume	2.914	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.097	03:27
Breath Volume	3.046	03:27
Air Blank	0.000	03:28
Control Test	0.081	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 04/01/2020 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: BELLEW

DATE/TIME OF TEST: 03/21/2020 21:01:35

Diagnostics Check	OK	20:55
Air Blank	0.000	20:55
Control Test	0.080	20:56
Air Blank	0.000	20:56
Subject Sample # 1	0.181	20:57
Breath Volume	3.476	20:57
Air Blank	0.000	20:57
Air Blank	0.000	20:59
Subject Sample # 2	0.177	21:00
Breath Volume	2.460	21:00
Air Blank	0.000	21:00
Control Test	0.081	21:01
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

Data Download Date/Time: 04/01/2020 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARROW, JESHUA, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MARTINA, RYAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: RIVENBARK

DATE/TIME OF TEST: 03/22/2020 00:40:45

Diagnostics Check	Fail	00:40	Voltage/Current Test
Air Blank	0.000	00:40	

Data Download Date/Time: 04/01/2020 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARROW, JESHUA, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MARTINA, RYAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: RIVENBARK

DATE/TIME OF TEST: 03/22/2020 00:48:15

Diagnostics Check	OK	00:43	
Air Blank	0.000	00:44	
Control Test	0.081	00:44	
Air Blank	0.000	00:45	
Subject Sample	# 1	00:46	Subject Test Refused
Breath Volume	0.000	00:46	
Air Blank	0.000	00:47	
Control Test	0.081	00:47	
Air Blank	0.000	00:48	
Diagnostics Check	OK	00:48	

Data Download Date/Time: 04/01/2020 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARROW, JESHUA, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MIDDLEBROOKS, ALLEN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: HEATH

DATE/TIME OF TEST: 03/22/2020 20:15:05

Diagnostics Check		OK	20:00	
Air Blank		0.000	20:01	
Control Test		0.080	20:01	
Air Blank		0.000	20:02	
Subject Sample	# 1	0.269	20:05	Volume Not Met
Breath Volume		0.066	20:05	
Air Blank		0.000	20:06	
Air Blank		0.000	20:07	
Subject Sample	# 2	0.316	20:09	
Breath Volume		2.652	20:09	
Air Blank		0.000	20:10	
Air Blank		0.000	20:12	
Subject Sample	# 3	0.333	20:13	
Breath Volume		2.656	20:13	
Air Blank		0.000	20:14	
Control Test		0.078	20:14	
Air Blank		0.000	20:14	
Diagnostics Check		OK	20:15	

Data Download Date/Time: 04/01/2020 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: COLLINS,T
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 03/24/2020 23:32:09

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:26
Subject Sample # 1	0.203	23:27
Breath Volume	2.988	23:27
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.207	23:30
Breath Volume	1.152	23:30
Air Blank	0.000	23:31
Control Test	0.082	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 04/01/2020 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,ANDREW,L
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: TACZAK,DAVID
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: HOLLOMAN

DATE/TIME OF TEST: 03/25/2020 02:15:24

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.219	02:10
Breath Volume	1.519	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.218	02:13
Breath Volume	2.121	02:13
Air Blank	0.000	02:14
Control Test	0.082	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 04/01/2020 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: INNES

DATE/TIME OF TEST: 03/27/2020 23:19:40

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.080	23:14
Air Blank	0.000	23:14
Subject Sample # 1	0.146	23:15
Breath Volume	3.367	23:15
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.133	23:18
Breath Volume	2.277	23:18
Air Blank	0.000	23:18
Control Test	0.081	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 04/01/2020 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISCHER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: FISCHER,ANDREW
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 03/17/2020 23:38:07

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.080	23:32
Air Blank	0.000	23:32
Subject Sample # 1	0.244	23:33
Breath Volume	1.511	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:35
Subject Sample # 2	0.231	23:36
Breath Volume	1.195	23:36
Air Blank	0.000	23:37
Control Test	0.079	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 04/01/2020 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL, VICKI, N
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: WEBB, EDWARD
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: SARTIN

DATE/TIME OF TEST: 03/09/2020 17:30:04

Diagnostics Check	OK	17:27	
Air Blank	0.000	17:27	
Control Test	0.082	17:28	
Air Blank	0.000	17:28	
Subject Sample	# 1	17:28	Subject Test Refused
Breath Volume	0.000	17:28	
Air Blank	0.000	17:29	
Control Test	0.080	17:29	
Air Blank	0.000	17:29	
Diagnostics Check	OK	17:30	

Data Download Date/Time: 03/19/2020 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,RONDRICKA,N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: TALLMAN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: TORGERSON

DATE/TIME OF TEST: 03/12/2020 19:36:03

Diagnostics Check	OK	19:33	
Air Blank	0.000	19:34	
Control Test	0.081	19:34	
Air Blank	0.000	19:34	
Subject Sample	# 1	19:35	RFI Detect
Breath Volume	0.000	19:35	
Air Blank		19:35	RFI Detect
Air Blank	0.000	19:36	

Data Download Date/Time: 03/19/2020 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,RONDRICKA,N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: TALLMAN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: TORGERSON

DATE/TIME OF TEST: 03/12/2020 19:43:49

Diagnostics Check	OK	19:38	
Air Blank	0.000	19:38	
Control Test	0.080	19:38	
Air Blank	0.000	19:39	
Subject Sample	# 1 0.087	19:40	
Breath Volume	1.742	19:40	
Air Blank	0.000	19:40	
Air Blank	0.000	19:42	
Subject Sample	# 2 0.076	19:43	
Breath Volume	2.386	19:43	
Air Blank		19:43	RFI Detect
Air Blank	0.000	19:43	

Data Download Date/Time: 03/19/2020 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAUFFMAN, MITCHELL, P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: RUSSELL, NATHANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: STOCKTON

DATE/TIME OF TEST: 03/25/2020 19:38:09

Diagnostics Check	OK	19:31
Air Blank	0.000	19:31
Control Test	0.079	19:32
Air Blank	0.000	19:32
Subject Sample # 1	0.229	19:33
Breath Volume	2.191	19:33
Air Blank	0.000	19:34
Air Blank	0.000	19:35
Subject Sample # 2	0.238	19:36
Breath Volume	4.234	19:36
Air Blank	0.000	19:37
Control Test	0.078	19:37
Air Blank	0.000	19:38
Diagnostics Check	OK	19:38

Data Download Date/Time: 04/15/2020 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REVELL,DENNIS,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: REVELL,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: PILAND

DATE/TIME OF TEST: 03/28/2020 04:00:33

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.217	03:55
Breath Volume	3.375	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.217	03:59
Breath Volume	4.328	03:59
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 04/15/2020 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BANOS,FRANK
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: BERNARDI

DATE/TIME OF TEST: 03/02/2020 01:32:31

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.082	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.135	01:28
Breath Volume	1.550	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.145	01:31
Breath Volume	2.367	01:31
Air Blank	0.000	01:31
Control Test	0.081	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 03/05/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VENTURA,RONNIE
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: CARTAGENA ROMAN

DATE/TIME OF TEST: 03/03/2020 04:34:19

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.082	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.159	04:29
Breath Volume	1.570	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.151	04:32
Breath Volume	1.871	04:32
Air Blank	0.000	04:33
Control Test	0.081	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 03/05/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOMINGUEZ,JAYMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: SOTOPAJARO

DATE/TIME OF TEST: 03/07/2020 03:51:59

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.082	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.190	03:47
Breath Volume	1.152	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.195	03:50
Breath Volume	2.757	03:50
Air Blank	0.000	03:51
Control Test	0.082	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 03/11/2020 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DENIJS,SAMIRA
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 03/07/2020 04:57:36

Diagnostics Check	OK	04:50
Air Blank	0.000	04:51
Control Test	0.082	04:51
Air Blank	0.000	04:52
Subject Sample # 1	0.169	04:53
Breath Volume	3.304	04:53
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.170	04:56
Breath Volume	3.015	04:56
Air Blank	0.000	04:56
Control Test	0.082	04:57
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 03/11/2020 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IMPARATO,FRANK
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 03/07/2020 06:36:53

Diagnostics Check	OK	06:30
Air Blank	0.000	06:31
Control Test	0.082	06:31
Air Blank	0.000	06:31
Subject Sample # 1	0.120	06:32
Breath Volume	1.351	06:32
Air Blank	0.000	06:32
Air Blank	0.000	06:34
Subject Sample # 2	0.123	06:35
Breath Volume	2.843	06:35
Air Blank	0.000	06:35
Control Test	0.082	06:36
Air Blank	0.000	06:36
Diagnostics Check	OK	06:36

Data Download Date/Time: 03/11/2020 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SALAZAR,JUAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: HERNANDEZGALLOZA

DATE/TIME OF TEST: 03/08/2020 05:31:36

Diagnostics Check	OK	05:22
Air Blank	0.000	05:23
Control Test	0.082	05:23
Air Blank	0.000	05:24
Subject Sample # 1	0.099	05:27
Breath Volume	2.828	05:27
Air Blank	0.000	05:27
Air Blank	0.000	05:29
Subject Sample # 2	0.104	05:30
Breath Volume	3.492	05:30
Air Blank	0.000	05:30
Control Test	0.079	05:31
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

Data Download Date/Time: 03/11/2020 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRISON,ROBBIE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: RONEY

DATE/TIME OF TEST: 03/09/2020 01:59:38

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.081	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.153	01:55
Breath Volume	1.605	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.159	01:58
Breath Volume	1.335	01:58
Air Blank	0.000	01:58
Control Test	0.080	01:59
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 03/11/2020 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FACEY,ANDRE
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: SEPE

DATE/TIME OF TEST: 03/10/2020 02:23:52

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.082	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.043	02:19
Breath Volume	1.628	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.042	02:22
Breath Volume	1.718	02:22
Air Blank	0.000	02:22
Control Test	0.081	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 03/11/2020 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KELEMEN, JANOS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 03/11/2020 02:38:38

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.082	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.097	02:34
Breath Volume	2.765	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.098	02:37
Breath Volume	2.652	02:37
Air Blank	0.000	02:37
Control Test	0.081	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 03/11/2020 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOMINGUEZ,JAYMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 03/13/2020 22:37:27

Diagnostics Check	OK	22:30
Air Blank	0.000	22:31
Control Test	0.082	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.254	22:32
Breath Volume	3.523	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.264	22:35
Breath Volume	1.796	22:35
Air Blank	0.000	22:36
Control Test	0.081	22:36
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 03/19/2020 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOONEY,JAMES
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: CRUZ CAMPOS

DATE/TIME OF TEST: 03/14/2020 00:45:46

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.082	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.000	00:41
Breath Volume	1.425	00:41
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.000	00:44
Breath Volume	1.425	00:44
Air Blank	0.000	00:44
Control Test	0.081	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 03/19/2020 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOHNS,MARK
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: GURIERREZ

DATE/TIME OF TEST: 03/14/2020 02:37:40

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.082	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.113	02:33
Breath Volume	1.589	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.105	02:36
Breath Volume	2.679	02:36
Air Blank	0.000	02:36
Control Test	0.081	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 03/19/2020 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KUZMA,TIM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ROSADO

DATE/TIME OF TEST: 03/15/2020 04:29:00

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.082	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.072	04:24
Breath Volume	2.261	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.069	04:27
Breath Volume	2.230	04:27
Air Blank	0.000	04:28
Control Test	0.081	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:29

Data Download Date/Time: 03/19/2020 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: DOYEN,ALEXIA
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 12:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BAQUERO GARZON

DATE/TIME OF TEST: 03/01/2020 01:42:12

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.080	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.200	01:37
Breath Volume	3.285	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.199	01:40
Breath Volume	3.285	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 03/31/2020 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WHOBREY,SPENCER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: DI PILATO

DATE/TIME OF TEST: 03/01/2020 23:25:28

Diagnostics Check	OK	23:18
Air Blank	0.000	23:19
Control Test	0.080	23:19
Air Blank	0.000	23:19
Subject Sample # 1	0.118	23:20
Breath Volume	2.042	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:22
Subject Sample # 2	0.124	23:23
Breath Volume	2.640	23:23
Air Blank	0.000	23:24
Control Test	0.081	23:24
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 03/31/2020 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,C,B
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: GUZMAN,V
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 08:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:44

SUBJECT'S LAST NAME: D42306097030

DATE/TIME OF TEST: 03/06/2020 09:10:58

Diagnostics Check	OK	09:04
Air Blank	0.000	09:04
Control Test	0.079	09:05
Air Blank	0.000	09:05
Subject Sample # 1	0.131	09:06
Breath Volume	1.617	09:06
Air Blank	0.000	09:07
Air Blank	0.000	09:09
Subject Sample # 2	0.128	09:09
Breath Volume	1.296	09:09
Air Blank	0.000	09:10
Control Test	0.078	09:10
Air Blank	0.000	09:10
Diagnostics Check	OK	09:10

Data Download Date/Time: 03/31/2020 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CARTY,JAMIE
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: BREEDEN

DATE/TIME OF TEST: 03/09/2020 01:07:39

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.080	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.000	01:03
Breath Volume	1.285	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.000	01:06
Breath Volume	1.531	01:06
Air Blank	0.000	01:06
Control Test	0.080	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 03/31/2020 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOUVRES,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: LANZO,MIGUEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: GATES

DATE/TIME OF TEST: 03/09/2020 01:37:50

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.180	01:33
Breath Volume	1.925	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.183	01:36
Breath Volume	1.597	01:36
Air Blank	0.000	01:36
Control Test	0.079	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 03/31/2020 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BREWER,DAVID
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: PENNINGTON

DATE/TIME OF TEST: 03/09/2020 03:53:37

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.078	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.142	03:48
Breath Volume	2.687	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.143	03:52
Breath Volume	2.445	03:52
Air Blank	0.000	03:52
Control Test	0.078	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 03/31/2020 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES, RONNIE,
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: AURIN, MICHAEL
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: SHENUSKI JR

DATE/TIME OF TEST: 03/09/2020 23:22:08

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.080	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.081	23:17
Breath Volume	3.808	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.083	23:20
Breath Volume	2.859	23:20
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 03/31/2020 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: AURIN,MICHAEL
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: KNOX

DATE/TIME OF TEST: 03/14/2020 02:38:45

Diagnostics Check		OK	02:31
Air Blank		0.000	02:32
Control Test		0.080	02:32
Air Blank		0.000	02:33
Subject Sample	# 1	0.186	02:34
Breath Volume		2.257	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.168	02:37
Breath Volume		1.238	02:37
Air Blank		0.000	02:37
Control Test		0.077	02:38
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 03/31/2020 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,CHARLES,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCCONNELL,CHARLES
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: ESPINOZA DIAZ

DATE/TIME OF TEST: 03/14/2020 03:59:06

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:46	
Control Test	0.077	03:46	
Air Blank	0.000	03:47	
Subject Sample # 1	0.125	03:48	
Breath Volume	1.128	03:48	
Air Blank	0.000	03:49	
Air Blank	0.000	03:50	
Subject Sample # 2	0.093	03:54	Volume Not Met
Breath Volume	0.445	03:54	
Air Blank	0.000	03:55	
Air Blank	0.000	03:56	
Subject Sample # 3	0.120	03:57	
Breath Volume	3.710	03:57	
Air Blank	0.000	03:58	
Control Test	0.068	03:58	Control Outside Tolerance
Air Blank	0.000	03:59	

Data Download Date/Time: 03/31/2020 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,CHARLES,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCCONNELL,CHARLES
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: ESPINOZA DIAZ

DATE/TIME OF TEST: 03/14/2020 04:03:52

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:02	
Control Test	0.072	04:03	Control Outside Tolerance
Air Blank	0.000	04:03	

Data Download Date/Time: 03/31/2020 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,CHARLES,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCCONNELL,CHARLES
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 12:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:21

SUBJECT'S LAST NAME: ROCKWELL

DATE/TIME OF TEST: 03/26/2020 13:57:37

Diagnostics Check	OK	13:56	
Air Blank	0.000	13:56	
Control Test	0.074	13:57	Control Outside Tolerance
Air Blank	0.000	13:57	

Data Download Date/Time: 03/31/2020 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,CHARLES,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCCONNELL,CHARLES
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 12:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:21

SUBJECT'S LAST NAME: ROCKWELL

DATE/TIME OF TEST: 03/26/2020 14:12:56

Diagnostics Check	OK	14:05
Air Blank	0.000	14:05
Control Test	0.075	14:05
Air Blank	0.000	14:06
Subject Sample # 1	0.236	14:07
Breath Volume	2.355	14:07
Air Blank	0.000	14:08
Air Blank	0.000	14:09
Subject Sample # 2	0.232	14:11
Breath Volume	1.851	14:11
Air Blank	0.000	14:12
Control Test	0.075	14:12
Air Blank	0.000	14:12
Diagnostics Check	OK	14:12

Data Download Date/Time: 03/31/2020 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES, RONNIE,
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: FERTIC, THOMAS
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/29/2020 21:50:27

Diagnostics Check	OK	21:40
Air Blank	0.000	21:41
Control Test	0.080	21:41
Air Blank	0.000	21:41
Subject Sample # 1	0.216	21:44
Breath Volume	1.433	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.208	21:48
Breath Volume	1.449	21:48
Air Blank	0.000	21:49
Control Test	0.078	21:49
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 03/31/2020 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUSLER,E,D
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HAUSLER,E
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: TERRY

DATE/TIME OF TEST: 03/26/2020 20:01:58

Diagnostics Check		OK	19:50	
Air Blank		0.000	19:50	
Control Test		0.079	19:51	
Air Blank		0.000	19:51	
Subject Sample	# 1	0.096	19:54	Slope Not Level
Breath Volume		1.410	19:54	
Air Blank		0.000	19:55	
Air Blank		0.000	19:57	
Subject Sample	# 2	0.099	20:00	Volume Not Met
Breath Volume		0.453	20:00	
Air Blank		0.000	20:01	
Control Test		0.078	20:01	
Air Blank		0.000	20:01	
Diagnostics Check		OK	20:01	

Data Download Date/Time: 05/22/2020 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCHBACH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HAUSLER,ETHAN
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: TERRY

DATE/TIME OF TEST: 03/26/2020 20:22:05

Diagnostics Check		OK	20:10	
Air Blank		0.000	20:11	
Control Test		0.078	20:11	
Air Blank		0.000	20:11	
Subject Sample	# 1	0.082	20:15	Volume Not Met
Breath Volume		0.355	20:15	
Air Blank		0.000	20:15	
Air Blank		0.000	20:17	
Subject Sample	# 2	0.091	20:20	Slope Not Level
Breath Volume		1.179	20:20	
Air Blank		0.000	20:21	
Control Test		0.078	20:21	
Air Blank		0.000	20:22	
Diagnostics Check		OK	20:22	

Data Download Date/Time: 05/22/2020 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,JACOB,T
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: WEAVER,JACOB
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: BRASHEAR

DATE/TIME OF TEST: 03/16/2020 00:32:57

Diagnostics Check		OK	00:25
Air Blank		0.000	00:25
Control Test		0.081	00:26
Air Blank		0.000	00:26
Subject Sample	# 1	0.010	00:28
Breath Volume		1.207	00:28
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.010	00:31
Breath Volume		2.414	00:31
Air Blank		0.000	00:32
Control Test		0.081	00:32
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 04/10/2020 18:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,JASON,M
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: ROSA,EMILIO
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: SPANGLER

DATE/TIME OF TEST: 03/22/2020 22:04:10

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.079	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.153	21:58
Breath Volume	1.375	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.165	22:02
Breath Volume	1.449	22:02
Air Blank	0.000	22:03
Control Test	0.077	22:03
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 04/10/2020 18:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURDEN,SHAWNTRALE,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: SOTO,BRANDON
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/27/2020 16:43:38

Diagnostics Check	OK	16:37
Air Blank	0.000	16:37
Control Test	0.081	16:38
Air Blank	0.000	16:38
Subject Sample # 1	0.292	16:39
Breath Volume	2.167	16:39
Air Blank	0.000	16:39
Air Blank	0.000	16:41
Subject Sample # 2	0.279	16:42
Breath Volume	2.015	16:42
Air Blank	0.000	16:42
Control Test	0.080	16:43
Air Blank	0.000	16:43
Diagnostics Check	OK	16:43

Data Download Date/Time: 04/10/2020 18:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURDEN,SHAWNTRALE,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: QUICK,CHAD
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 03/31/2020 12:52:58

Diagnostics Check		OK	12:46
Air Blank		0.000	12:46
Control Test		0.080	12:47
Air Blank		0.000	12:47
Subject Sample	# 1	0.248	12:48
Breath Volume		3.710	12:48
Air Blank		0.000	12:49
Air Blank		0.000	12:50
Subject Sample	# 2	0.236	12:51
Breath Volume		4.519	12:51
Air Blank		0.000	12:52
Control Test		0.078	12:52
Air Blank		0.000	12:52
Diagnostics Check		OK	12:52

Data Download Date/Time: 04/10/2020 18:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURDEN,SHAWNTRALE,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: DURDEN,SHAWNTRALE
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 03/31/2020 21:13:29

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.081	21:07
Air Blank	0.000	21:08
Subject Sample # 1	0.185	21:09
Breath Volume	4.945	21:09
Air Blank	0.000	21:09
Air Blank	0.000	21:11
Subject Sample # 2	0.168	21:11
Breath Volume	5.000	21:11
Air Blank	0.000	21:12
Control Test	0.080	21:12
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

Data Download Date/Time: 04/10/2020 18:36