

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRESENZI,G,
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: CRESENZI,G
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 03/30/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 04/02/2020 22:21:43

Diagnostics Check	OK	22:13	
Air Blank	0.000	22:13	
Control Test	0.079	22:14	
Air Blank	0.000	22:14	
Subject Sample	# 1 0.223	22:17	Volume Not Met
Breath Volume	0.199	22:17	
Air Blank	0.000	22:18	
Air Blank	0.000	22:20	
Subject Sample	# 2	22:20	Subject Test Refused
Breath Volume	0.000	22:20	
Air Blank	0.000	22:20	
Control Test	0.078	22:21	
Air Blank	0.000	22:21	
Diagnostics Check	OK	22:21	

Data Download Date/Time: 04/20/2020 14:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATSCH,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PATSCH,JOSHUA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 04/04/2020 03:48:29

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.163	03:43
Breath Volume	2.281	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.158	03:47
Breath Volume	2.535	03:47
Air Blank	0.000	03:47
Control Test	0.079	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 04/20/2020 14:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRESENZI,G,
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: BERGERON,D
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 06:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:27

SUBJECT'S LAST NAME: MARTEL

DATE/TIME OF TEST: 04/16/2020 06:56:02

Diagnostics Check	OK	06:49
Air Blank	0.000	06:50
Control Test	0.079	06:50
Air Blank	0.000	06:50
Subject Sample # 1	0.130	06:51
Breath Volume	4.632	06:51
Air Blank	0.000	06:52
Air Blank	0.000	06:54
Subject Sample # 2	0.114	06:54
Breath Volume	1.699	06:54
Air Blank	0.000	06:55
Control Test	0.078	06:55
Air Blank	0.000	06:55
Diagnostics Check	OK	06:56

Data Download Date/Time: 04/20/2020 14:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRESENZI,G,
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: SPURLOCK,A
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 03/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 04/17/2020 00:11:35

Diagnostics Check		OK	00:04
Air Blank		0.000	00:05
Control Test		0.080	00:05
Air Blank		0.000	00:06
Subject Sample	# 1	0.209	00:06
Breath Volume		1.839	00:06
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.204	00:10
Breath Volume		1.300	00:10
Air Blank		0.000	00:10
Control Test		0.080	00:10
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 04/20/2020 14:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGERON,BRADY,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: POTTER,KYLE
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 04/20/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 13:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:04

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 04/29/2020 14:08:20

Diagnostics Check	OK	14:07	
Air Blank		14:07	RFI Detect
Air Blank	0.000	14:08	

Data Download Date/Time: 05/28/2020 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGERON,BRADY,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: POTTER,KYLE
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 04/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 13:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:04

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 04/29/2020 14:20:09

Diagnostics Check	OK	14:11	
Air Blank	0.000	14:12	
Control Test	0.077	14:12	
Air Blank	0.000	14:12	
Subject Sample	# 1	14:16	No Sample Provided
Breath Volume	0.000	14:16	
Air Blank	0.000	14:16	
Air Blank	0.000	14:18	
Subject Sample	# 2	14:18	Subject Test Refused
Breath Volume	0.000	14:18	
Air Blank	0.000	14:19	
Control Test	0.077	14:19	
Air Blank	0.000	14:20	
Diagnostics Check	OK	14:20	

Data Download Date/Time: 05/28/2020 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGERON,BRADY,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: POTTER,KYLE
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 04/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 13:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:04

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 04/29/2020 14:30:21

Diagnostics Check	OK	14:24
Air Blank	0.000	14:24
Control Test	0.077	14:25
Air Blank	0.000	14:25
Subject Sample # 1	0.229	14:25
Breath Volume	1.386	14:25
Air Blank	0.000	14:26
Air Blank	0.000	14:28
Subject Sample # 2	0.230	14:28
Breath Volume	1.507	14:28
Air Blank	0.000	14:29
Control Test	0.077	14:29
Air Blank	0.000	14:30
Diagnostics Check	OK	14:30

Data Download Date/Time: 05/28/2020 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HUGHES,JARON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: ONEIL

DATE/TIME OF TEST: 04/04/2020 00:47:24

Diagnostics Check	OK	00:36
Air Blank	0.000	00:37
Control Test	0.080	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.182	00:38
Breath Volume	1.492	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.206	00:43
Breath Volume	2.179	00:43
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 3	0.187	00:45
Breath Volume	1.363	00:45
Air Blank	0.000	00:46
Control Test	0.078	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 04/28/2020 09:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,AMANDA,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: LANE,AMANDA
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 14:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: RAIMONDO

DATE/TIME OF TEST: 04/16/2020 15:48:07

Diagnostics Check	OK	15:40
Air Blank	0.000	15:40
Control Test	0.080	15:41
Air Blank	0.000	15:41
Subject Sample # 1	0.000	15:43
Breath Volume	2.539	15:43
Air Blank	0.000	15:43
Air Blank	0.000	15:45
Subject Sample # 2	0.000	15:46
Breath Volume	3.386	15:46
Air Blank	0.000	15:47
Control Test	0.081	15:47
Air Blank	0.000	15:48
Diagnostics Check	OK	15:48

Data Download Date/Time: 04/28/2020 09:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,SCOTT,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CLINE,BRUCE
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 03/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 04/17/2020 00:21:33

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:19	
Control Test	0.080	00:19	
Air Blank	0.000	00:19	
Subject Sample	# 1	00:20	Subject Test Refused
Breath Volume	0.000	00:20	
Air Blank	0.000	00:20	
Control Test	0.080	00:20	
Air Blank	0.000	00:21	
Diagnostics Check	OK	00:21	

Data Download Date/Time: 04/28/2020 09:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORALUZZO,MICHAEL,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CORALUZZO,MICHAEL
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/21/2020 18:10:43

Diagnostics Check	OK	18:03
Air Blank	0.000	18:04
Control Test	0.081	18:04
Air Blank	0.000	18:05
Subject Sample # 1	0.000	18:05
Breath Volume	1.859	18:05
Air Blank	0.000	18:06
Air Blank	0.000	18:08
Subject Sample # 2	0.000	18:09
Breath Volume	1.808	18:09
Air Blank	0.000	18:09
Control Test	0.078	18:10
Air Blank	0.000	18:10
Diagnostics Check	OK	18:10

Data Download Date/Time: 04/28/2020 09:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CAMDEN,DENNIS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/24/2020 03:20:50

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.081	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.024	03:16
Breath Volume	2.164	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:18
Subject Sample # 2	0.022	03:19
Breath Volume	1.855	03:19
Air Blank	0.000	03:19
Control Test	0.081	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 04/28/2020 09:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,JOEL,A
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: JENKINS,JOEL
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 03/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: BRANDT

DATE/TIME OF TEST: 04/24/2020 19:30:48

Diagnostics Check	OK	19:24
Air Blank	0.000	19:25
Control Test	0.080	19:25
Air Blank	0.000	19:25
Subject Sample # 1	0.263	19:26
Breath Volume	1.605	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:28
Subject Sample # 2	0.254	19:29
Breath Volume	2.183	19:29
Air Blank	0.000	19:29
Control Test	0.079	19:30
Air Blank	0.000	19:30
Diagnostics Check	OK	19:30

Data Download Date/Time: 04/28/2020 09:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DA SILVA,ANDRE,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: JENSEN,D
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 03/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: BIEL

DATE/TIME OF TEST: 04/25/2020 02:49:44

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.080	02:45	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:48	Subject Test Refused
Breath Volume	0.000	02:48	
Air Blank	0.000	02:48	
Control Test	0.080	02:49	
Air Blank	0.000	02:49	
Diagnostics Check	OK	02:49	

Data Download Date/Time: 04/28/2020 09:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 06:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:54

SUBJECT'S LAST NAME: TREADWELL

DATE/TIME OF TEST: 04/26/2020 07:21:29

Diagnostics Check	OK	07:15
Air Blank	0.000	07:15
Control Test	0.080	07:15
Air Blank	0.000	07:16
Subject Sample # 1	0.116	07:17
Breath Volume	2.761	07:17
Air Blank	0.000	07:17
Air Blank	0.000	07:19
Subject Sample # 2	0.117	07:19
Breath Volume	2.042	07:19
Air Blank	0.000	07:20
Control Test	0.080	07:20
Air Blank	0.000	07:21
Diagnostics Check	OK	07:21

Data Download Date/Time: 04/28/2020 09:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DA SILVA,A,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HARVEY,C
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: EXFORD

DATE/TIME OF TEST: 04/27/2020 03:42:33

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.080	03:34
Air Blank		0.000	03:35
Subject Sample	# 1	0.081	03:37
Breath Volume		1.492	03:37
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.078	03:40
Breath Volume		1.675	03:40
Air Blank		0.000	03:41
Control Test		0.081	03:41
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 04/28/2020 09:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHROCK,L,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: WALLACE,K
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 14:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 04/27/2020 15:37:10

Diagnostics Check	OK	15:30
Air Blank	0.000	15:31
Control Test	0.080	15:31
Air Blank	0.000	15:32
Subject Sample # 1	0.261	15:32
Breath Volume	3.171	15:32
Air Blank	0.000	15:33
Air Blank	0.000	15:35
Subject Sample # 2	0.278	15:35
Breath Volume	2.878	15:35
Air Blank	0.000	15:36
Control Test	0.078	15:36
Air Blank	0.000	15:37
Diagnostics Check	OK	15:37

Data Download Date/Time: 04/28/2020 09:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BODKIN,MICHEAL,W
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: FOREMAN,KEITH
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 03/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/08/2022
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 04/01/2020 01:56:56

Diagnostics Check		OK	01:50
Air Blank		0.000	01:50
Control Test		0.082	01:51
Air Blank		0.000	01:51
Subject Sample	# 1	0.345	01:52
Breath Volume		1.542	01:52
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.329	01:55
Breath Volume		1.562	01:55
Air Blank		0.000	01:56
Control Test		0.081	01:56
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 04/14/2020 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001149

INSTRUMENT REGISTERED TO: LAKE HELEN P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSHWORTH,JOSEPH,D
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: GRAY,TREVOR
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/25/2020 20:02:58

Diagnostics Check	OK	19:56
Air Blank	0.000	19:57
Control Test	0.080	19:57
Air Blank	0.000	19:58
Subject Sample # 1	0.146	19:58
Breath Volume	2.148	19:58
Air Blank	0.000	19:59
Air Blank	0.000	20:00
Subject Sample # 2	0.144	20:01
Breath Volume	2.601	20:01
Air Blank	0.000	20:02
Control Test	0.079	20:02
Air Blank	0.000	20:02
Diagnostics Check	OK	20:02

Data Download Date/Time: 05/26/2020 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUBLINO,NICK,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: SAFIEDDINE,JAMAL
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 04/20/2020 01:17:59

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.083	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.000	01:13
Breath Volume	1.808	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.000	01:16
Breath Volume	1.738	01:16
Air Blank	0.000	01:17
Control Test	0.080	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 04/24/2020 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINZ,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: HUNNEFELD,RALPH
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 03/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:11

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 04/20/2020 10:38:26

Diagnostics Check	OK	10:32
Air Blank	0.000	10:32
Control Test	0.078	10:32
Air Blank	0.000	10:33
Subject Sample # 1	0.000	10:33
Breath Volume	2.796	10:33
Air Blank	0.000	10:34
Air Blank	0.000	10:36
Subject Sample # 2	0.000	10:37
Breath Volume	3.417	10:37
Air Blank	0.000	10:37
Control Test	0.078	10:37
Air Blank	0.000	10:38
Diagnostics Check	OK	10:38

Data Download Date/Time: 05/01/2020 05:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,KYLE,T
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: POTTER,KYLE
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: SCHAAB

DATE/TIME OF TEST: 04/30/2020 09:07:48

Diagnostics Check	OK	09:01
Air Blank	0.000	09:01
Control Test	0.078	09:02
Air Blank	0.000	09:02
Subject Sample # 1	0.184	09:03
Breath Volume	3.566	09:03
Air Blank	0.000	09:03
Air Blank	0.000	09:05
Subject Sample # 2	0.180	09:06
Breath Volume	3.160	09:06
Air Blank	0.000	09:06
Control Test	0.078	09:07
Air Blank	0.000	09:07
Diagnostics Check	OK	09:07

Data Download Date/Time: 05/01/2020 05:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,RICHARD,L
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: OGLE,SHAWN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 03/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 04/20/2020 00:05:34

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:02	
Control Test	0.081	00:03	
Air Blank	0.000	00:03	
Subject Sample	# 1	00:04	Subject Test Refused
Breath Volume	0.000	00:04	
Air Blank	0.000	00:04	
Control Test	0.081	00:04	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:05	

Data Download Date/Time: 04/28/2020 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOFFMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: PENA,ENMANUEL
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/20/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 04/23/2020 21:38:53

Diagnostics Check	OK	21:36	
Air Blank	0.000	21:36	
Control Test	0.079	21:36	
Air Blank	0.000	21:37	
Subject Sample	# 1	21:37	Subject Test Refused
Breath Volume	0.000	21:37	
Air Blank	0.000	21:37	
Control Test	0.079	21:38	
Air Blank	0.000	21:38	
Diagnostics Check	OK	21:38	

Data Download Date/Time: 05/30/2020 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: DORAN

DATE/TIME OF TEST: 04/02/2020 02:27:16

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.089	02:22
Breath Volume	1.359	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.080	02:25
Breath Volume	2.109	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 04/09/2020 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAY,ERICK,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DAY,ERICK
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:17

SUBJECT'S LAST NAME: SHEPHERD

DATE/TIME OF TEST: 04/03/2020 13:05:23

Diagnostics Check		OK	12:52	
Air Blank		0.000	12:53	
Control Test		0.081	12:53	
Air Blank		0.000	12:54	
Subject Sample	# 1	0.217	12:57	Slope Not Level
Breath Volume		1.132	12:57	
Air Blank		0.000	12:57	
Air Blank		0.000	12:59	
Subject Sample	# 2	0.237	13:00	
Breath Volume		1.390	13:00	
Air Blank		0.000	13:00	
Air Blank		0.000	13:02	
Subject Sample	# 3	0.228	13:03	
Breath Volume		1.136	13:03	
Air Blank		0.000	13:04	
Control Test		0.078	13:04	
Air Blank		0.000	13:05	
Diagnostics Check		OK	13:05	

Data Download Date/Time: 04/09/2020 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: PEDRO-LUCAS

DATE/TIME OF TEST: 04/04/2020 01:07:48

Diagnostics Check	OK	01:01
Air Blank	0.000	01:02
Control Test	0.080	01:02
Air Blank	0.000	01:03
Subject Sample # 1	0.231	01:03
Breath Volume	4.046	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:05
Subject Sample # 2	0.220	01:06
Breath Volume	3.000	01:06
Air Blank	0.000	01:06
Control Test	0.080	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 04/09/2020 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BURKHARDT

DATE/TIME OF TEST: 04/07/2020 01:30:50

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.081	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.141	01:26
Breath Volume	1.152	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.142	01:29
Breath Volume	1.710	01:29
Air Blank	0.000	01:29
Control Test	0.080	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 04/09/2020 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUBOIS,BROOKS,E
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUBOIS,BROOKS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: MEIER

DATE/TIME OF TEST: 04/08/2020 16:40:21

Diagnostics Check	OK	16:28	
Air Blank	0.000	16:29	
Control Test	0.080	16:29	
Air Blank	0.000	16:30	
Subject Sample	# 1	16:33	No Sample Provided
Breath Volume	0.000	16:33	
Air Blank	0.000	16:33	
Air Blank	0.000	16:35	
Subject Sample	# 2	16:38	No Sample Provided
Breath Volume	0.000	16:38	
Air Blank	0.000	16:39	
Control Test	0.080	16:39	
Air Blank	0.000	16:40	
Diagnostics Check	OK	16:40	

Data Download Date/Time: 04/09/2020 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: GIGNAC

DATE/TIME OF TEST: 04/11/2020 22:37:50

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.082	22:32
Air Blank	0.000	22:32
Subject Sample # 1	0.000	22:33
Breath Volume	2.218	22:33
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.000	22:36
Breath Volume	2.328	22:36
Air Blank	0.000	22:36
Control Test	0.082	22:37
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 05/08/2020 16:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,REBECCA,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRADY,REBECCA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: TRAN

DATE/TIME OF TEST: 04/11/2020 23:15:45

Diagnostics Check	OK	23:02
Air Blank	0.000	23:03
Control Test	0.082	23:03
Air Blank	0.000	23:04
Subject Sample # 1	0.275	23:07
Breath Volume	1.390	23:07
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.246	23:10
Breath Volume	1.136	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 3	0.260	23:14
Breath Volume	2.554	23:14
Air Blank	0.000	23:14
Control Test	0.078	23:15
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 05/08/2020 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CAPLAN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: BAGBY

DATE/TIME OF TEST: 04/15/2020 01:32:19

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.174	01:27
Breath Volume	3.851	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.172	01:30
Breath Volume	2.707	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 05/08/2020 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 04/25/2020 22:27:13

Diagnostics Check	OK	22:20
Air Blank	0.000	22:21
Control Test	0.082	22:21
Air Blank	0.000	22:21
Subject Sample # 1	0.056	22:22
Breath Volume	2.339	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.055	22:25
Breath Volume	2.761	22:25
Air Blank	0.000	22:26
Control Test	0.082	22:26
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 05/08/2020 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: TOMPOROWSKI

DATE/TIME OF TEST: 04/26/2020 02:01:49

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.082	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.194	01:57
Breath Volume	3.128	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.197	02:00
Breath Volume	2.859	02:00
Air Blank	0.000	02:00
Control Test	0.081	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 05/08/2020 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CAPLAN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: ANTONIO

DATE/TIME OF TEST: 04/28/2020 00:26:28

Diagnostics Check	OK	00:19
Air Blank	0.000	00:19
Control Test	0.081	00:20
Air Blank	0.000	00:20
Subject Sample # 1	0.181	00:21
Breath Volume	2.980	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.177	00:24
Breath Volume	2.117	00:24
Air Blank	0.000	00:25
Control Test	0.081	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 05/08/2020 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINN,CAMERON,B
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HOFFMAN,JORDAN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: BERNAT

DATE/TIME OF TEST: 04/03/2020 21:44:52

Diagnostics Check	OK	21:38
Air Blank	0.000	21:39
Control Test	0.079	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.210	21:40
Breath Volume	3.011	21:40
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.207	21:43
Breath Volume	2.570	21:43
Air Blank	0.000	21:43
Control Test	0.079	21:44
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 04/30/2020 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIRKS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: PATTERSON,DUSTIN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: SMALL JR

DATE/TIME OF TEST: 04/03/2020 23:51:25

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.079	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.140	23:47
Breath Volume	2.156	23:47
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.134	23:49
Breath Volume	1.191	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 04/30/2020 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAUGA,JOHN,E
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MCNEIL,RAYMOND
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 15:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: GRIMM

DATE/TIME OF TEST: 04/05/2020 16:11:00

Diagnostics Check	OK	16:05
Air Blank	0.000	16:05
Control Test	0.079	16:05
Air Blank	0.000	16:06
Subject Sample # 1	0.000	16:06
Breath Volume	1.652	16:06
Air Blank	0.000	16:07
Air Blank	0.000	16:09
Subject Sample # 2	0.000	16:09
Breath Volume	1.535	16:09
Air Blank	0.000	16:10
Control Test	0.078	16:10
Air Blank	0.000	16:10
Diagnostics Check	OK	16:11

Data Download Date/Time: 04/30/2020 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIRKS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: TAYLOR,CHRIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 04/05/2020 19:00:57

Diagnostics Check	OK	18:54
Air Blank	0.000	18:55
Control Test	0.078	18:55
Air Blank	0.000	18:55
Subject Sample # 1	0.216	18:56
Breath Volume	1.210	18:56
Air Blank	0.000	18:57
Air Blank	0.000	18:58
Subject Sample # 2	0.220	18:59
Breath Volume	2.019	18:59
Air Blank	0.000	19:00
Control Test	0.078	19:00
Air Blank	0.000	19:00
Diagnostics Check	OK	19:00

Data Download Date/Time: 04/30/2020 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORE,RICHARD,A
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BROWN,DUSTIN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: CROSS

DATE/TIME OF TEST: 04/07/2020 22:12:08

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.079	22:05
Air Blank	0.000	22:06
Subject Sample # 1	0.000	22:07
Breath Volume	1.871	22:07
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.000	22:10
Breath Volume	2.593	22:10
Air Blank	0.000	22:11
Control Test	0.078	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 04/30/2020 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: QUIGLEY

DATE/TIME OF TEST: 04/12/2020 02:32:49

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:29	
Control Test	0.080	02:30	
Air Blank	0.000	02:30	
Subject Sample # 1	0.000	02:31	
Breath Volume	2.867	02:31	
Air Blank		02:31	Improper Sample
Air Blank	0.000	02:31	
Control Test	0.079	02:32	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 04/30/2020 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: QUIGLEY

DATE/TIME OF TEST: 04/12/2020 02:42:57

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.079	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.000	02:38
Breath Volume	2.640	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.000	02:41
Breath Volume	1.902	02:41
Air Blank	0.000	02:42
Control Test	0.079	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 04/30/2020 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMSEY,JOHNATHAN,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: RAMSEY,J
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: SMILEY

DATE/TIME OF TEST: 04/20/2020 23:55:15

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.080	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.000	23:48
Breath Volume	1.214	23:48
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.000	23:53
Breath Volume	1.945	23:53
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 04/30/2020 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLEN,PRESTON
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: SHIVER

DATE/TIME OF TEST: 04/30/2020 02:46:08

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:40	
Control Test	0.079	02:41	
Air Blank	0.000	02:41	
Subject Sample # 1	0.109	02:42	
Breath Volume	2.417	02:42	
Air Blank	0.000	02:42	
Air Blank	0.000	02:44	
Subject Sample # 2	0.112	02:44	
Breath Volume	2.480	02:44	
Air Blank	0.000	02:45	
		02:45	RFI Detect
Air Blank	0.000	02:46	

Data Download Date/Time: 04/30/2020 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARDO, ADALBERTO,
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: SCELFO, ZACHARY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/12/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: PEREZ-SOLANO

DATE/TIME OF TEST: 04/12/2020 03:34:41

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:31	
Control Test	0.082	03:31	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:33	Subject Test Refused
Breath Volume	0.000	03:33	
Air Blank	0.000	03:33	
Control Test	0.083	03:34	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 04/13/2020 19:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,KATIE,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LOPEZ,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: HUNDLEY

DATE/TIME OF TEST: 04/09/2020 01:19:18

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:16	
Control Test	0.080	01:16	
Air Blank	0.000	01:17	
Subject Sample	# 1	01:17	Subject Test Refused
Breath Volume	0.000	01:17	
Air Blank	0.000	01:18	
Control Test	0.080	01:18	
Air Blank	0.000	01:19	
Diagnostics Check	OK	01:19	

Data Download Date/Time: 05/07/2020 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMAN,A,C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KHAN,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: BERNTHON

DATE/TIME OF TEST: 04/09/2020 12:53:28

Diagnostics Check	OK	12:47
Air Blank	0.000	12:47
Control Test	0.080	12:48
Air Blank	0.000	12:48
Subject Sample # 1	0.000	12:49
Breath Volume	3.335	12:49
Air Blank	0.000	12:49
Air Blank	0.000	12:51
Subject Sample # 2	0.000	12:52
Breath Volume	1.746	12:52
Air Blank	0.000	12:52
Control Test	0.080	12:52
Air Blank	0.000	12:53
Diagnostics Check	OK	12:53

Data Download Date/Time: 05/07/2020 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNY, MARK, A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HODAKOSKI, SARAH
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:27

SUBJECT'S LAST NAME: KENNY

DATE/TIME OF TEST: 04/10/2020 12:16:02

Diagnostics Check	OK	12:07
Air Blank	0.000	12:07
Control Test	0.080	12:08
Air Blank	0.000	12:08
Subject Sample # 1	0.000	12:09
Breath Volume	1.867	12:09
Air Blank	0.000	12:10
Air Blank	0.000	12:12
Subject Sample # 2	0.000	12:14
Breath Volume	3.472	12:14
Air Blank	0.000	12:15
Control Test	0.080	12:15
Air Blank	0.000	12:15
Diagnostics Check	OK	12:16

Data Download Date/Time: 05/07/2020 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROBEL, RAYMOND, R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: STROBEL, RAYMOND
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: BILLINGHAM

DATE/TIME OF TEST: 04/10/2020 12:58:16

Diagnostics Check	OK	12:51
Air Blank	0.000	12:52
Control Test	0.079	12:52
Air Blank	0.000	12:53
Subject Sample # 1	0.297	12:53
Breath Volume	1.437	12:53
Air Blank	0.000	12:54
Air Blank	0.000	12:56
Subject Sample # 2	0.289	12:56
Breath Volume	1.355	12:56
Air Blank	0.000	12:57
Control Test	0.079	12:57
Air Blank	0.000	12:58
Diagnostics Check	OK	12:58

Data Download Date/Time: 05/07/2020 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: VERBIEST 2369,R
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: CASTRO GAITAN

DATE/TIME OF TEST: 04/13/2020 01:06:01

Diagnostics Check		OK	00:59
Air Blank		0.000	00:59
Control Test		0.080	00:59
Air Blank		0.000	01:00
Subject Sample	# 1	0.260	01:01
Breath Volume		1.968	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.261	01:04
Breath Volume		1.257	01:04
Air Blank		0.000	01:05
Control Test		0.079	01:05
Air Blank		0.000	01:05
Diagnostics Check		OK	01:06

Data Download Date/Time: 05/07/2020 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:31

SUBJECT'S LAST NAME: LAWRENCE-MCNULTY

DATE/TIME OF TEST: 04/13/2020 18:00:52

Diagnostics Check	OK	17:54
Air Blank	0.000	17:54
Control Test	0.080	17:55
Air Blank	0.000	17:55
Subject Sample # 1	0.272	17:56
Breath Volume	2.464	17:56
Air Blank	0.000	17:56
Air Blank	0.000	17:58
Subject Sample # 2	0.276	17:59
Breath Volume	2.621	17:59
Air Blank	0.000	17:59
Control Test	0.079	18:00
Air Blank	0.000	18:00
Diagnostics Check	OK	18:00

Data Download Date/Time: 05/07/2020 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 04/15/2020 23:58:11

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.080	23:53
Air Blank	0.000	23:53
Subject Sample # 1	0.000	23:53
Breath Volume	1.578	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.000	23:56
Breath Volume	1.488	23:56
Air Blank	0.000	23:57
Control Test	0.080	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 05/07/2020 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: PAHLKE 17045,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: NICOL

DATE/TIME OF TEST: 04/16/2020 22:08:24

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.080	22:02
Air Blank	0.000	22:03
Subject Sample # 1	0.000	22:03
Breath Volume	3.531	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.000	22:07
Breath Volume	2.546	22:07
Air Blank	0.000	22:07
Control Test	0.080	22:07
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 05/07/2020 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HOYOS 19022,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: JUAN FRANCISCO

DATE/TIME OF TEST: 04/16/2020 23:58:33

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.212	23:53
Breath Volume	2.011	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:55
Subject Sample # 2	0.210	23:57
Breath Volume	2.019	23:57
Air Blank	0.000	23:57
Control Test	0.078	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 05/07/2020 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,KATIE,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CARRILLO,LUIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: TECK

DATE/TIME OF TEST: 04/18/2020 00:08:57

Diagnostics Check		OK	00:02
Air Blank		0.000	00:02
Control Test		0.080	00:03
Air Blank		0.000	00:03
Subject Sample	# 1	0.183	00:04
Breath Volume		2.054	00:04
Air Blank		0.000	00:04
Air Blank		0.000	00:06
Subject Sample	# 2	0.183	00:07
Breath Volume		2.015	00:07
Air Blank		0.000	00:08
Control Test		0.079	00:08
Air Blank		0.000	00:08
Diagnostics Check		OK	00:08

Data Download Date/Time: 05/07/2020 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,KATIE,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: RIDGEWAY,CAITLIN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: US IXCOTOYAC

DATE/TIME OF TEST: 04/18/2020 01:05:49

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.080	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.178	01:01
Breath Volume	2.363	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.181	01:04
Breath Volume	2.578	01:04
Air Blank	0.000	01:04
Control Test	0.079	01:05
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 05/07/2020 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,KATIE,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LOPEZ,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 04/18/2020 01:44:43

Diagnostics Check		OK	01:37
Air Blank		0.000	01:38
Control Test		0.080	01:38
Air Blank		0.000	01:38
Subject Sample	# 1	0.095	01:40
Breath Volume		2.609	01:40
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.097	01:43
Breath Volume		2.296	01:43
Air Blank		0.000	01:43
Control Test		0.079	01:44
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 05/07/2020 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: MEJIA TZUNUX

DATE/TIME OF TEST: 04/18/2020 22:24:38

Diagnostics Check		OK	22:18
Air Blank		0.000	22:18
Control Test		0.080	22:18
Air Blank		0.000	22:19
Subject Sample	# 1	0.275	22:20
Breath Volume		1.476	22:20
Air Blank		0.000	22:20
Air Blank		0.000	22:22
Subject Sample	# 2	0.277	22:23
Breath Volume		1.398	22:23
Air Blank		0.000	22:23
Control Test		0.080	22:24
Air Blank		0.000	22:24
Diagnostics Check		OK	22:24

Data Download Date/Time: 05/07/2020 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,KATIE,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: JENSEN,BRADLEY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/19/2020 23:07:25

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.080	23:02
Air Blank	0.000	23:02
Subject Sample # 1	0.241	23:02
Breath Volume	3.410	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.239	23:05
Breath Volume	2.894	23:05
Air Blank	0.000	23:06
Control Test	0.079	23:06
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 05/07/2020 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAK,KYLE,F
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MCCLOUGHEN,KORY
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:07

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 04/20/2020 08:43:06

Diagnostics Check	OK	08:36
Air Blank	0.000	08:36
Control Test	0.080	08:36
Air Blank	0.000	08:37
Subject Sample # 1	0.107	08:38
Breath Volume	2.429	08:38
Air Blank	0.000	08:39
Air Blank	0.000	08:41
Subject Sample # 2	0.105	08:41
Breath Volume	2.156	08:41
Air Blank	0.000	08:42
Control Test	0.079	08:42
Air Blank	0.000	08:43
Diagnostics Check	OK	08:43

Data Download Date/Time: 05/07/2020 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONSALL,M,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BONSALL,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: PONZINI

DATE/TIME OF TEST: 04/22/2020 00:16:03

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.081	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.107	00:11
Breath Volume	1.687	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.103	00:14
Breath Volume	1.699	00:14
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:16

Data Download Date/Time: 05/07/2020 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEMEN,LOGAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEMEN,LOGAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: COULTER

DATE/TIME OF TEST: 04/23/2020 01:11:53

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.080	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.012	01:07
Breath Volume	2.894	01:07
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.011	01:10
Breath Volume	3.453	01:10
Air Blank	0.000	01:10
Control Test	0.079	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 05/07/2020 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,KATIE,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LEON,BRIAN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 04/23/2020 22:05:55

Diagnostics Check		OK	21:54	
Air Blank		0.000	21:54	
Control Test		0.081	21:54	
Air Blank		0.000	21:55	
Subject Sample	# 1	0.077	21:58	Volume Not Met
Breath Volume		0.121	21:58	
Air Blank		0.000	21:59	
Air Blank		0.000	22:00	
Subject Sample	# 2	0.083	22:01	
Breath Volume		2.003	22:01	
Air Blank		0.000	22:02	
Air Blank		0.000	22:04	
Subject Sample	# 3	0.084	22:04	
Breath Volume		2.226	22:04	
Air Blank		0.000	22:05	
Control Test		0.079	22:05	
Air Blank		0.000	22:05	
Diagnostics Check		OK	22:05	

Data Download Date/Time: 05/07/2020 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: WOLF

DATE/TIME OF TEST: 04/24/2020 00:02:46

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.081	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.210	23:57
Breath Volume	2.582	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.214	00:01
Breath Volume	2.230	00:01
Air Blank	0.000	00:01
Control Test	0.080	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 05/07/2020 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORES,ERIK,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: HODAKOSKI,SARAH
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:03

SUBJECT'S LAST NAME: AMPARO

DATE/TIME OF TEST: 04/25/2020 16:34:48

Diagnostics Check	OK	16:28
Air Blank	0.000	16:29
Control Test	0.081	16:29
Air Blank	0.000	16:30
Subject Sample # 1	0.000	16:30
Breath Volume	2.121	16:30
Air Blank	0.000	16:31
Air Blank	0.000	16:33
Subject Sample # 2	0.000	16:33
Breath Volume	2.351	16:33
Air Blank	0.000	16:33
Control Test	0.079	16:34
Air Blank	0.000	16:34
Diagnostics Check	OK	16:34

Data Download Date/Time: 05/07/2020 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: PANTOJA 603,A
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 04/26/2020 01:14:56

Diagnostics Check	OK	01:08
Air Blank	0.000	01:09
Control Test	0.080	01:09
Air Blank	0.000	01:09
Subject Sample # 1	0.088	01:10
Breath Volume	1.523	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:12
Subject Sample # 2	0.083	01:13
Breath Volume	1.660	01:13
Air Blank	0.000	01:14
Control Test	0.080	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 05/07/2020 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,T,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SELLERS,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: ROSADO CRUZ

DATE/TIME OF TEST: 04/26/2020 19:25:13

Diagnostics Check	OK	19:17
Air Blank	0.000	19:17
Control Test	0.080	19:18
Air Blank	0.000	19:18
Subject Sample # 1	0.166	19:20
Breath Volume	1.421	19:20
Air Blank	0.000	19:20
Air Blank	0.000	19:22
Subject Sample # 2	0.158	19:23
Breath Volume	1.371	19:23
Air Blank	0.000	19:24
Control Test	0.078	19:24
Air Blank	0.000	19:25
Diagnostics Check	OK	19:25

Data Download Date/Time: 05/07/2020 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DZUGAN 1773,M
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 04/27/2020 03:24:13

Diagnostics Check		OK	03:18
Air Blank		0.000	03:18
Control Test		0.080	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.000	03:19
Breath Volume		4.691	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.000	03:22
Breath Volume		4.792	03:22
Air Blank		0.000	03:23
Control Test		0.080	03:23
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 05/07/2020 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,T,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SELLERS,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: MCGEE

DATE/TIME OF TEST: 04/30/2020 19:13:52

Diagnostics Check	OK	19:07
Air Blank	0.000	19:08
Control Test	0.080	19:08
Air Blank	0.000	19:08
Subject Sample # 1	0.000	19:09
Breath Volume	2.507	19:09
Air Blank	0.000	19:09
Air Blank	0.000	19:11
Subject Sample # 2	0.000	19:12
Breath Volume	1.968	19:12
Air Blank	0.000	19:12
Control Test	0.080	19:13
Air Blank	0.000	19:13
Diagnostics Check	OK	19:13

Data Download Date/Time: 05/07/2020 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MICAH,L
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: THOMAS,MICAH
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 03/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 04/05/2020 19:46:19

Diagnostics Check	OK	19:32	
Air Blank	0.000	19:32	
Control Test	0.080	19:33	
Air Blank	0.000	19:33	
Subject Sample	# 1 0.125	19:35	
Breath Volume	1.500	19:35	
Air Blank	0.000	19:36	
Air Blank	0.000	19:37	
Subject Sample	# 2 0.111	19:41	Slope Not Level
Breath Volume	1.144	19:41	
Air Blank	0.000	19:41	
Air Blank	0.000	19:43	
Subject Sample	# 3 0.126	19:44	
Breath Volume	1.140	19:44	
Air Blank	0.000	19:45	
Control Test	0.080	19:45	
Air Blank	0.000	19:46	
Diagnostics Check	OK	19:46	

Data Download Date/Time: 05/01/2020 08:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MICAH,L
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: THOMAS,MICAH
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 03/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 04/13/2020 15:47:31

Diagnostics Check	OK	15:41
Air Blank	0.000	15:41
Control Test	0.079	15:42
Air Blank	0.000	15:42
Subject Sample # 1	0.281	15:43
Breath Volume	1.175	15:43
Air Blank	0.000	15:43
Air Blank	0.000	15:45
Subject Sample # 2	0.270	15:45
Breath Volume	1.421	15:45
Air Blank	0.000	15:46
Control Test	0.080	15:46
Air Blank	0.000	15:47
Diagnostics Check	OK	15:47

Data Download Date/Time: 05/01/2020 08:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MICAH,L
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: THOMAS,MICAH
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 03/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: BUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: SARZUELA

DATE/TIME OF TEST: 04/22/2020 21:37:48

Diagnostics Check	OK	21:31
Air Blank	0.000	21:32
Control Test	0.079	21:32
Air Blank	0.000	21:33
Subject Sample # 1	0.000	21:33
Breath Volume	3.554	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.000	21:36
Breath Volume	2.886	21:36
Air Blank	0.000	21:36
Control Test	0.079	21:37
Air Blank	0.000	21:37
Diagnostics Check	OK	21:37

Data Download Date/Time: 05/01/2020 08:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MICAH,L
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: THOMAS,MICAH
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 03/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: GUARCHAJ TAMBRIZ

DATE/TIME OF TEST: 04/24/2020 13:01:56

Diagnostics Check	OK	12:51
Air Blank	0.000	12:52
Control Test	0.080	12:52
Air Blank	0.000	12:53
Subject Sample # 1	0.235	12:53
Breath Volume	3.042	12:53
Air Blank	0.000	12:54
Air Blank	0.000	12:55
Subject Sample # 2	0.209	12:57
Breath Volume	2.160	12:57
Air Blank	0.000	12:58
Air Blank	0.000	12:59
Subject Sample # 3	0.234	13:00
Breath Volume	2.078	13:00
Air Blank	0.000	13:01
Control Test	0.079	13:01
Air Blank	0.000	13:01
Diagnostics Check	OK	13:01

Data Download Date/Time: 05/01/2020 08:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SECORD,RYAN,T
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: DILLARD,DANIEL
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 03/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: KRIEG

DATE/TIME OF TEST: 04/25/2020 15:22:32

Diagnostics Check	OK	15:16
Air Blank	0.000	15:17
Control Test	0.080	15:17
Air Blank	0.000	15:17
Subject Sample # 1	0.000	15:18
Breath Volume	2.789	15:18
Air Blank	0.000	15:18
Air Blank	0.000	15:20
Subject Sample # 2	0.000	15:21
Breath Volume	2.304	15:21
Air Blank	0.000	15:21
Control Test	0.080	15:21
Air Blank	0.000	15:22
Diagnostics Check	OK	15:22

Data Download Date/Time: 05/01/2020 08:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,RICHARD,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OGG,DAVID
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 04/30/2020 02:17:59

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.078	02:12
Air Blank		0.000	02:12
Subject Sample	# 1	0.218	02:13
Breath Volume		2.875	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:15
Subject Sample	# 2	0.213	02:16
Breath Volume		2.433	02:16
Air Blank		0.000	02:17
Control Test		0.078	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 05/27/2020 07:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: RAMON III

DATE/TIME OF TEST: 04/30/2020 22:32:46

Diagnostics Check		OK	22:26
Air Blank		0.000	22:27
Control Test		0.079	22:27
Air Blank		0.000	22:27
Subject Sample	# 1	0.174	22:28
Breath Volume		1.421	22:28
Air Blank		0.000	22:29
Air Blank		0.000	22:30
Subject Sample	# 2	0.181	22:31
Breath Volume		1.371	22:31
Air Blank		0.000	22:31
Control Test		0.078	22:32
Air Blank		0.000	22:32
Diagnostics Check		OK	22:32

Data Download Date/Time: 05/27/2020 07:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUBI,UZIEL,Y
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: DUBI,UZIEL
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 03/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MERCADO

DATE/TIME OF TEST: 04/12/2020 03:01:37

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:59	
Control Test	0.081	02:59	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:00	
Control Test	0.081	03:01	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

Data Download Date/Time: 04/13/2020 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CODY,DEONTA,M
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: CODY,DEONTA
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 04/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 04/17/2020 22:28:54

Diagnostics Check		OK	22:15	
Air Blank		0.000	22:16	
Control Test		0.082	22:16	
Air Blank		0.000	22:17	
Subject Sample	# 1	0.161	22:20	Volume Not Met
Breath Volume		0.105	22:20	
Air Blank		0.000	22:21	
Air Blank		0.000	22:22	
Subject Sample	# 2	0.164	22:23	
Breath Volume		1.847	22:23	
Air Blank		0.000	22:24	
Air Blank		0.000	22:26	
Subject Sample	# 3	0.159	22:27	
Breath Volume		2.339	22:27	
Air Blank		0.000	22:27	
Control Test		0.082	22:28	
Air Blank		0.000	22:28	
Diagnostics Check		OK	22:28	

Data Download Date/Time: 05/05/2020 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE, JOSHUA,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: CLARK, K
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 04/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: JUBITHANA

DATE/TIME OF TEST: 04/22/2020 00:05:19

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.082	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.141	00:00
Breath Volume	1.835	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:02
Subject Sample # 2	0.146	00:03
Breath Volume	2.546	00:03
Air Blank	0.000	00:04
Control Test	0.082	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 05/05/2020 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALSH,SHANNON,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: SMITH,S
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 04/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 12:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: PUENTES

DATE/TIME OF TEST: 04/22/2020 14:11:48

Diagnostics Check	OK	14:05
Air Blank	0.000	14:05
Control Test	0.083	14:06
Air Blank	0.000	14:06
Subject Sample # 1	0.000	14:07
Breath Volume	1.664	14:07
Air Blank	0.000	14:07
Air Blank	0.000	14:09
Subject Sample # 2	0.000	14:10
Breath Volume	2.148	14:10
Air Blank	0.000	14:10
Control Test	0.082	14:11
Air Blank	0.000	14:11
Diagnostics Check	OK	14:11

Data Download Date/Time: 05/05/2020 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,DOUG,J
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: BAUDERS,SEAN
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 04/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 04/25/2020 06:02:15

Diagnostics Check	OK	05:54
Air Blank	0.000	05:54
Control Test	0.081	05:55
Air Blank	0.000	05:55
Subject Sample # 1	0.127	05:57
Breath Volume	1.121	05:57
Air Blank	0.000	05:58
Air Blank	0.000	06:00
Subject Sample # 2	0.134	06:00
Breath Volume	2.507	06:00
Air Blank	0.000	06:01
Control Test	0.081	06:01
Air Blank	0.000	06:02
Diagnostics Check	OK	06:02

Data Download Date/Time: 05/05/2020 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,NEAL,E
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: ANDERSON,N
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 04/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 11:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:39

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 04/25/2020 12:05:51

Diagnostics Check	OK	12:02	
Air Blank	0.000	12:03	
Control Test	0.082	12:03	
Air Blank	0.000	12:04	
Subject Sample	# 1	12:04	Subject Test Refused
Breath Volume	0.000	12:04	
Air Blank	0.000	12:04	
Control Test	0.080	12:05	
Air Blank	0.000	12:05	
Diagnostics Check	OK	12:05	

Data Download Date/Time: 05/05/2020 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALSH,SHANNON,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: POWELL,K
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 04/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ZUCCHINI

DATE/TIME OF TEST: 04/26/2020 02:37:20

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.082	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.000	02:32
Breath Volume	4.675	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.000	02:35
Breath Volume	2.863	02:35
Air Blank	0.000	02:36
Control Test	0.082	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 05/05/2020 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE, JOSHUA,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: BAUDERS, S
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 04/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: BLAKE

DATE/TIME OF TEST: 04/29/2020 21:05:38

Diagnostics Check		OK	20:52	
Air Blank		0.000	20:52	
Control Test		0.082	20:53	
Air Blank		0.000	20:53	
Subject Sample	# 1	0.179	20:56	Volume Not Met
Breath Volume		0.691	20:56	
Air Blank		0.000	20:57	
Air Blank		0.000	20:59	
Subject Sample	# 2	0.189	21:00	
Breath Volume		1.523	21:00	
Air Blank		0.000	21:01	
Air Blank		0.000	21:03	
Subject Sample	# 3	0.188	21:04	
Breath Volume		1.320	21:04	
Air Blank		0.000	21:04	
Control Test		0.080	21:05	
Air Blank		0.000	21:05	
Diagnostics Check		OK	21:05	

Data Download Date/Time: 05/05/2020 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRENNAN, MATTHEW, W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHRISTIE, KEVIN
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 04/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 15:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 04/30/2020 19:00:32

Diagnostics Check	OK	18:53
Air Blank	0.000	18:54
Control Test	0.082	18:54
Air Blank	0.000	18:55
Subject Sample # 1	0.239	18:55
Breath Volume	2.773	18:55
Air Blank	0.000	18:56
Air Blank	0.000	18:58
Subject Sample # 2	0.251	18:58
Breath Volume	2.507	18:58
Air Blank	0.000	18:59
Control Test	0.082	18:59
Air Blank	0.000	19:00
Diagnostics Check	OK	19:00

Data Download Date/Time: 05/05/2020 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAROCHE, CHARLES, W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: SHEHAN, KARL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: ARNEY

DATE/TIME OF TEST: 04/04/2020 22:39:53

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:25	
Control Test	0.081	22:26	
Air Blank	0.000	22:26	
Subject Sample # 1	0.242	22:27	
Breath Volume	1.570	22:27	
Air Blank	0.000	22:28	
Air Blank	0.000	22:30	
Subject Sample # 2	0.164	22:33	Volume Not Met
Breath Volume	0.261	22:33	
Air Blank	0.000	22:34	
Air Blank	0.000	22:35	
Subject Sample # 3		22:38	Subject Test Refused
Breath Volume	0.464	22:38	
Air Blank	0.000	22:38	
Control Test	0.078	22:39	
Air Blank	0.000	22:39	
Diagnostics Check	OK	22:39	

Data Download Date/Time: 04/14/2020 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PEERS,MICHAEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: SANCHEZBORJAS

DATE/TIME OF TEST: 04/02/2020 04:50:28

Diagnostics Check	OK	04:44
Air Blank	0.000	04:45
Control Test	0.082	04:45
Air Blank	0.000	04:45
Subject Sample # 1	0.000	04:46
Breath Volume	2.273	04:46
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.000	04:49
Breath Volume	1.574	04:49
Air Blank	0.000	04:49
Control Test	0.080	04:49
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 04/02/2020 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,CHRIS
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:51

SUBJECT'S LAST NAME: CHILDRESS

DATE/TIME OF TEST: 04/02/2020 16:22:55

Diagnostics Check	OK	16:16
Air Blank	0.000	16:17
Control Test	0.081	16:17
Air Blank	0.000	16:18
Subject Sample # 1	0.000	16:18
Breath Volume	4.371	16:18
Air Blank	0.000	16:19
Air Blank	0.000	16:21
Subject Sample # 2	0.000	16:21
Breath Volume	2.777	16:21
Air Blank	0.000	16:22
Control Test	0.079	16:22
Air Blank	0.000	16:22
Diagnostics Check	OK	16:22

Data Download Date/Time: 04/09/2020 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MITCHELL, CEDRIC
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: GABAUD

DATE/TIME OF TEST: 04/04/2020 03:28:39

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.081	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.231	03:23
Breath Volume	1.148	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.219	03:26
Breath Volume	1.765	03:26
Air Blank	0.000	03:27
Control Test	0.079	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 04/09/2020 15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SLACK,JOHNIE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 04/06/2020 00:41:46

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.081	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.189	00:37
Breath Volume	2.078	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.181	00:40
Breath Volume	1.191	00:40
Air Blank	0.000	00:40
Control Test	0.078	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 04/09/2020 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MICHALAK,TIMOTHY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: TURNBULL

DATE/TIME OF TEST: 04/11/2020 21:50:05

Diagnostics Check	OK	21:43
Air Blank	0.000	21:43
Control Test	0.081	21:44
Air Blank	0.000	21:44
Subject Sample # 1	0.215	21:45
Breath Volume	1.894	21:45
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample # 2	0.205	21:48
Breath Volume	1.226	21:48
Air Blank	0.000	21:49
Control Test	0.080	21:49
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 04/23/2020 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER, OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH, STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: DUROCHER

DATE/TIME OF TEST: 04/14/2020 10:19:59

Diagnostics Check	OK	10:13
Air Blank	0.000	10:14
Control Test	0.082	10:14
Air Blank	0.000	10:15
Subject Sample # 1	0.000	10:15
Breath Volume	2.800	10:15
Air Blank	0.000	10:16
Air Blank	0.000	10:18
Subject Sample # 2	0.000	10:18
Breath Volume	1.894	10:18
Air Blank	0.000	10:19
Control Test	0.081	10:19
Air Blank	0.000	10:19
Diagnostics Check	OK	10:19

Data Download Date/Time: 04/23/2020 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SAYLER,ROBERT
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:48

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 04/14/2020 16:17:12

Diagnostics Check		OK	16:10
Air Blank		0.000	16:11
Control Test		0.082	16:11
Air Blank		0.000	16:12
Subject Sample	# 1	0.215	16:12
Breath Volume		1.472	16:12
Air Blank		0.000	16:13
Air Blank		0.000	16:15
Subject Sample	# 2	0.215	16:15
Breath Volume		1.328	16:15
Air Blank		0.000	16:16
Control Test		0.081	16:16
Air Blank		0.000	16:17
Diagnostics Check		OK	16:17

Data Download Date/Time: 04/23/2020 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BLACKWELL, ROBERT
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: PEREZ BANCO

DATE/TIME OF TEST: 04/17/2020 02:04:35

Diagnostics Check		OK	01:58
Air Blank		0.000	01:58
Control Test		0.083	01:59
Air Blank		0.000	01:59
Subject Sample	# 1	0.059	02:00
Breath Volume		1.796	02:00
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.059	02:02
Breath Volume		3.777	02:02
Air Blank		0.000	02:03
Control Test		0.081	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 04/23/2020 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS, TABITHA, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LEE, MIKEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/18/2020 23:32:56

Diagnostics Check		OK	23:26
Air Blank		0.000	23:27
Control Test		0.082	23:27
Air Blank		0.000	23:27
Subject Sample	# 1	0.158	23:28
Breath Volume		2.699	23:28
Air Blank		0.000	23:29
Air Blank		0.000	23:30
Subject Sample	# 2	0.158	23:31
Breath Volume		2.656	23:31
Air Blank		0.000	23:32
Control Test		0.081	23:32
Air Blank		0.000	23:32
Diagnostics Check		OK	23:32

Data Download Date/Time: 04/23/2020 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS, TABITHA, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ADAMS, JAMES
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: CREEK

DATE/TIME OF TEST: 04/21/2020 01:20:56

Diagnostics Check		OK	01:08	
Air Blank		0.000	01:08	
Control Test		0.082	01:09	
Air Blank		0.000	01:09	
Subject Sample	# 1	0.167	01:12	Slope Not Level
Breath Volume		1.675	01:12	
Air Blank		0.000	01:13	
Air Blank		0.000	01:15	
Subject Sample	# 2	0.161	01:16	
Breath Volume		2.292	01:16	
Air Blank		0.000	01:17	
Air Blank		0.000	01:18	
Subject Sample	# 3	0.159	01:19	
Breath Volume		1.667	01:19	
Air Blank		0.000	01:20	
Control Test		0.080	01:20	
Air Blank		0.000	01:20	
Diagnostics Check		OK	01:20	

Data Download Date/Time: 04/23/2020 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: OSENKARSKI, MAX
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: CARRERO BRACERO

DATE/TIME OF TEST: 04/22/2020 23:09:40

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.082	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.000	23:05
Breath Volume	1.484	23:05
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.000	23:08
Breath Volume	1.960	23:08
Air Blank	0.000	23:08
Control Test	0.082	23:09
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 04/23/2020 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KISSANE,BRIAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 04/23/2020 01:52:20

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.082	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.278	01:47
Breath Volume	2.523	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.271	01:50
Breath Volume	1.976	01:50
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 04/23/2020 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PEREZ,JUAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: SHIFRIN

DATE/TIME OF TEST: 04/23/2020 22:07:10

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.082	22:01
Air Blank	0.000	22:02
Subject Sample # 1	0.000	22:02
Breath Volume	2.859	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.000	22:05
Breath Volume	3.859	22:05
Air Blank	0.000	22:06
Control Test	0.080	22:06
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 04/30/2020 17:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CASTELLANOPADRON, BRYAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: ROMANKO

DATE/TIME OF TEST: 04/24/2020 17:37:10

Diagnostics Check	OK	17:30
Air Blank	0.000	17:31
Control Test	0.081	17:31
Air Blank	0.000	17:32
Subject Sample # 1	0.286	17:32
Breath Volume	3.843	17:32
Air Blank	0.000	17:33
Air Blank	0.000	17:34
Subject Sample # 2	0.279	17:35
Breath Volume	3.265	17:35
Air Blank	0.000	17:36
Control Test	0.082	17:36
Air Blank	0.000	17:37
Diagnostics Check	OK	17:37

Data Download Date/Time: 04/30/2020 17:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: BLANDONHERNANDEZ

DATE/TIME OF TEST: 04/25/2020 00:59:35

Diagnostics Check		OK	00:47	
Air Blank		0.000	00:48	
Control Test		0.082	00:48	
Air Blank		0.000	00:49	
Subject Sample	# 1	0.200	00:52	Volume Not Met
Breath Volume		0.121	00:52	
Air Blank		0.000	00:53	
Air Blank		0.000	00:54	
Subject Sample	# 2	0.202	00:57	Volume Not Met
Breath Volume		0.105	00:57	
Air Blank		0.000	00:58	
Control Test		0.079	00:58	
Air Blank		0.000	00:59	
Diagnostics Check		OK	00:59	

Data Download Date/Time: 04/30/2020 17:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: BLANDONHERNANDEZ

DATE/TIME OF TEST: 04/25/2020 01:14:32

Diagnostics Check	OK	01:02	
Air Blank	0.000	01:02	
Control Test	0.079	01:02	
Air Blank	0.000	01:03	
Subject Sample # 1	0.196	01:04	
Breath Volume	1.265	01:04	
Air Blank	0.000	01:04	
Air Blank	0.000	01:06	
Subject Sample # 2	0.163	01:09	Volume Not Met
Breath Volume	0.433	01:09	
Air Blank	0.000	01:10	
Air Blank	0.000	01:12	
Subject Sample # 3	0.202	01:12	
Breath Volume	1.679	01:12	
Air Blank	0.000	01:13	
Control Test	0.078	01:13	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

Data Download Date/Time: 04/30/2020 17:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERN,ETHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 04/25/2020 23:20:48

Diagnostics Check	OK	23:09	
Air Blank	0.000	23:09	
Control Test	0.081	23:10	
Air Blank	0.000	23:10	
Subject Sample	# 1	23:13	No Sample Provided
Breath Volume	0.000	23:13	
Air Blank	0.000	23:14	
Air Blank	0.000	23:16	
Subject Sample	# 2	23:19	No Sample Provided
Breath Volume	0.000	23:19	
Air Blank	0.000	23:19	
Control Test	0.082	23:20	
Air Blank	0.000	23:20	
Diagnostics Check	OK	23:20	

Data Download Date/Time: 04/30/2020 17:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRAZEE,COLE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: DESHAY

DATE/TIME OF TEST: 04/26/2020 03:40:19

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.082	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.081	03:35
Breath Volume	2.175	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.081	03:38
Breath Volume	1.531	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 04/30/2020 17:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 04/27/2020 02:40:23

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.082	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.173	02:35
Breath Volume	1.507	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.174	02:38
Breath Volume	1.226	02:38
Air Blank	0.000	02:39
Control Test	0.081	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 04/30/2020 17:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DODD,COLIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: JEAN LOUIS

DATE/TIME OF TEST: 04/30/2020 04:06:58

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.081	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.182	04:02
Breath Volume	2.800	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:04
Subject Sample # 2	0.183	04:05
Breath Volume	3.050	04:05
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 04/30/2020 17:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRIGGS,DERIENNE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BLAIR JR

DATE/TIME OF TEST: 04/08/2020 00:26:30

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.080	00:21
Air Blank		0.000	00:21
Subject Sample	# 1	0.195	00:21
Breath Volume		2.406	00:21
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.189	00:24
Breath Volume		2.871	00:24
Air Blank		0.000	00:25
Control Test		0.080	00:25
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 04/09/2020 16:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HUSS,JOHN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: BOLVITO

DATE/TIME OF TEST: 04/09/2020 23:31:48

Diagnostics Check	OK	23:24
Air Blank	0.000	23:24
Control Test	0.080	23:24
Air Blank	0.000	23:25
Subject Sample # 1	0.239	23:26
Breath Volume	2.601	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:28
Subject Sample # 2	0.243	23:30
Breath Volume	2.250	23:30
Air Blank	0.000	23:30
Control Test	0.079	23:31
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 04/23/2020 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDIS,BENJAMIN,M
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: LEITCH,RICKFORD
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 03/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:33

SUBJECT'S LAST NAME: MEEKER

DATE/TIME OF TEST: 04/01/2020 08:13:16

Diagnostics Check		OK	08:07
Air Blank		0.000	08:07
Control Test		0.078	08:07
Air Blank		0.000	08:08
Subject Sample	# 1	0.196	08:08
Breath Volume		1.367	08:08
Air Blank		0.000	08:09
Air Blank		0.000	08:11
Subject Sample	# 2	0.186	08:11
Breath Volume		1.835	08:11
Air Blank		0.000	08:12
Control Test		0.076	08:12
Air Blank		0.000	08:13
Diagnostics Check		OK	08:13

Data Download Date/Time: 04/23/2020 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,JAMES,P
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: MOREHEAD,AVERY
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 03/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/22/2022
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: BABENDREIER

DATE/TIME OF TEST: 04/02/2020 23:47:03

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:44	
Control Test	0.079	23:45	
Air Blank	0.000	23:45	
Subject Sample	# 1	23:45	Subject Test Refused
Breath Volume	0.000	23:45	
Air Blank	0.000	23:46	
Control Test	0.079	23:46	
Air Blank	0.000	23:46	
Diagnostics Check	OK	23:47	

Data Download Date/Time: 05/06/2020 04:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, TYRELL, J
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: LORAIN, CHARLES
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 03/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/22/2022
VIOLATION CODE: DUI
ARREST TIME: 14:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: KORECKY

DATE/TIME OF TEST: 04/21/2020 16:10:00

Diagnostics Check	OK	16:06	
Air Blank	0.000	16:07	
Control Test	0.078	16:07	
Air Blank	0.000	16:07	
Subject Sample	# 1	16:08	Subject Test Refused
Breath Volume	0.000	16:08	
Air Blank	0.000	16:09	
Control Test	0.078	16:09	
Air Blank	0.000	16:09	
Diagnostics Check	OK	16:10	

Data Download Date/Time: 05/06/2020 04:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS,HEATHER,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIVINS,HEATHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: UBINAS COBALLES

DATE/TIME OF TEST: 04/06/2020 23:55:25

Diagnostics Check		OK	23:49
Air Blank		0.000	23:49
Control Test		0.081	23:49
Air Blank		0.000	23:50
Subject Sample	# 1	0.201	23:51
Breath Volume		1.406	23:51
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 2	0.200	23:53
Breath Volume		2.015	23:53
Air Blank		0.000	23:54
Control Test		0.079	23:54
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 05/04/2020 08:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: BLAZEWICZ,A
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 03/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 04/21/2020 00:21:45

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.082	00:16
Air Blank	0.000	00:16
Subject Sample # 1	0.000	00:17
Breath Volume	1.960	00:17
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.000	00:20
Breath Volume	2.277	00:20
Air Blank	0.000	00:20
Control Test	0.080	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 05/04/2020 08:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 04/22/2020 19:56:05

Diagnostics Check	OK	19:49
Air Blank	0.000	19:50
Control Test	0.082	19:50
Air Blank	0.000	19:51
Subject Sample # 1	0.173	19:51
Breath Volume	1.386	19:51
Air Blank	0.000	19:52
Air Blank	0.000	19:54
Subject Sample # 2	0.164	19:54
Breath Volume	1.191	19:54
Air Blank	0.000	19:55
Control Test	0.081	19:55
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

Data Download Date/Time: 05/04/2020 08:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVINGSTON,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: BOLAND,K
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/30/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 11:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:55

SUBJECT'S LAST NAME: WOERTENDUKE

DATE/TIME OF TEST: 04/24/2020 11:42:24

Diagnostics Check		OK	11:35
Air Blank		0.000	11:36
Control Test		0.081	11:36
Air Blank		0.000	11:37
Subject Sample	# 1	0.144	11:37
Breath Volume		2.066	11:37
Air Blank		0.000	11:38
Air Blank		0.000	11:40
Subject Sample	# 2	0.140	11:40
Breath Volume		2.437	11:40
Air Blank		0.000	11:41
Control Test		0.082	11:41
Air Blank		0.000	11:42
Diagnostics Check		OK	11:42

Data Download Date/Time: 05/04/2020 08:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: COCCHIOLA,AUSTIN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 03/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 16:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 04/26/2020 18:03:13

Diagnostics Check	OK	17:56
Air Blank	0.000	17:57
Control Test	0.082	17:57
Air Blank	0.000	17:58
Subject Sample # 1	0.000	17:58
Breath Volume	2.902	17:58
Air Blank	0.000	17:59
Air Blank	0.000	18:01
Subject Sample # 2	0.000	18:01
Breath Volume	2.894	18:01
Air Blank	0.000	18:02
Control Test	0.081	18:02
Air Blank	0.000	18:03
Diagnostics Check	OK	18:03

Data Download Date/Time: 05/04/2020 08:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS, ANTHONY, M
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: VAZQUEZ, A
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 03/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: MARINO

DATE/TIME OF TEST: 04/26/2020 20:36:34

Diagnostics Check	OK	20:30
Air Blank	0.000	20:30
Control Test	0.081	20:30
Air Blank	0.000	20:31
Subject Sample # 1	0.000	20:32
Breath Volume	1.812	20:32
Air Blank	0.000	20:32
Air Blank	0.000	20:34
Subject Sample # 2	0.000	20:35
Breath Volume	1.402	20:35
Air Blank	0.000	20:35
Control Test	0.080	20:35
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

Data Download Date/Time: 05/04/2020 08:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: COOPER,E
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 03/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: MICHAUD

DATE/TIME OF TEST: 04/26/2020 22:22:34

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.081	22:17
Air Blank	0.000	22:17
Subject Sample # 1	0.000	22:18
Breath Volume	2.843	22:18
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.000	22:21
Breath Volume	2.781	22:21
Air Blank	0.000	22:21
Control Test	0.081	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 05/04/2020 08:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,ALEXIS,M
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: CRAMER,LISA
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 03/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 10:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: MUNICH

DATE/TIME OF TEST: 04/27/2020 11:31:04

Diagnostics Check	OK	11:22
Air Blank	0.000	11:22
Control Test	0.081	11:22
Air Blank	0.000	11:23
Subject Sample # 1	0.000	11:26
Breath Volume	4.250	11:26
Air Blank	0.000	11:26
Air Blank	0.000	11:28
Subject Sample # 2	0.000	11:29
Breath Volume	3.148	11:29
Air Blank	0.000	11:30
Control Test	0.080	11:30
Air Blank	0.000	11:30
Diagnostics Check	OK	11:31

Data Download Date/Time: 05/04/2020 08:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CAMPBELL,JACOB
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 04/02/2020 02:56:55

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:45	
Control Test	0.078	02:46	
Air Blank	0.000	02:46	
Subject Sample	# 1 0.159	02:49	Volume Not Met
Breath Volume	0.265	02:49	
Air Blank	0.000	02:50	
Air Blank	0.000	02:52	
Subject Sample	# 2	02:55	Subject Test Refused
Breath Volume	0.207	02:55	
Air Blank	0.000	02:55	
Control Test	0.076	02:56	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

Data Download Date/Time: 04/14/2020 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: KEARNEY,BRANDON
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 03/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: MAKOWSKI

DATE/TIME OF TEST: 04/06/2020 21:42:01

Diagnostics Check	OK	21:39	
Air Blank	0.000	21:39	
Control Test	0.078	21:39	
Air Blank	0.000	21:40	
Subject Sample	# 1	21:40	Subject Test Refused
Breath Volume	0.000	21:40	
Air Blank	0.000	21:41	
Control Test	0.077	21:41	
Air Blank	0.000	21:41	
Diagnostics Check	OK	21:42	

Data Download Date/Time: 04/14/2020 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAMUELS,ANDREW,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: SAMUELS,ANDREW
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 03/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: KORBUT

DATE/TIME OF TEST: 04/10/2020 02:13:33

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.078	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.115	02:08
Breath Volume	2.355	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.124	02:12
Breath Volume	1.488	02:12
Air Blank	0.000	02:12
Control Test	0.078	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 04/14/2020 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MARK,E
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: THOMPSON,MARK
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: COWELL

DATE/TIME OF TEST: 04/12/2020 08:25:02

Diagnostics Check	OK	08:17
Air Blank	0.000	08:18
Control Test	0.078	08:18
Air Blank	0.000	08:18
Subject Sample # 1	0.184	08:20
Breath Volume	3.136	08:20
Air Blank	0.000	08:21
Air Blank	0.000	08:22
Subject Sample # 2	0.185	08:23
Breath Volume	1.937	08:23
Air Blank	0.000	08:24
Control Test	0.077	08:24
Air Blank	0.000	08:24
Diagnostics Check	OK	08:25

Data Download Date/Time: 04/14/2020 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAMUELS,A,G
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: SAMUELS,A
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: CADE-BOYNTON

DATE/TIME OF TEST: 04/21/2020 03:57:16

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.078	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.000	03:52
Breath Volume	1.628	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.000	03:55
Breath Volume	2.417	03:55
Air Blank	0.000	03:56
Control Test	0.078	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 05/17/2020 20:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWELL,NATHANAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LEWIS,JEREMY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:34

SUBJECT'S LAST NAME: PERRUSQUIA

DATE/TIME OF TEST: 04/25/2020 08:49:58

Diagnostics Check	OK	08:43
Air Blank	0.000	08:44
Control Test	0.077	08:44
Air Blank	0.000	08:45
Subject Sample # 1	0.152	08:45
Breath Volume	2.691	08:45
Air Blank	0.000	08:46
Air Blank	0.000	08:47
Subject Sample # 2	0.148	08:48
Breath Volume	3.417	08:48
Air Blank	0.000	08:49
Control Test	0.077	08:49
Air Blank	0.000	08:49
Diagnostics Check	OK	08:49

Data Download Date/Time: 05/17/2020 20:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWISHER, EDWARD, T
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SWISHER, EDWARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: KIRN

DATE/TIME OF TEST: 04/25/2020 21:23:17

Diagnostics Check	OK	21:15	
Air Blank	0.000	21:15	
Control Test	0.077	21:16	
Air Blank	0.000	21:16	
Subject Sample	# 1 0.289	21:19	
Breath Volume	2.531	21:19	
Air Blank	0.000	21:19	
Air Blank	0.000	21:21	
Subject Sample	# 2	21:21	RFI Detect
Breath Volume	0.000	21:21	
Air Blank	0.000	21:22	
Control Test	0.077	21:22	
Air Blank	0.000	21:23	
Diagnostics Check	OK	21:23	

Data Download Date/Time: 05/17/2020 20:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWISHER, EDWARD, T
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SWISHER, EDWARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: KIRN

DATE/TIME OF TEST: 04/25/2020 21:28:55

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:25	
Control Test	0.077	21:26	
Air Blank	0.000	21:26	
Subject Sample	# 1	21:27	Slope Not Met
Breath Volume	1.660	21:27	
Air Blank	0.000	21:28	
Control Test	0.077	21:28	
Air Blank	0.000	21:28	
Diagnostics Check	OK	21:28	

Data Download Date/Time: 05/17/2020 20:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWISHER, EDWARD, T
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SWISHER, EDWARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: KIRN

DATE/TIME OF TEST: 04/25/2020 21:36:54

Diagnostics Check	OK	21:30
Air Blank	0.000	21:31
Control Test	0.077	21:31
Air Blank	0.000	21:31
Subject Sample # 1	0.281	21:32
Breath Volume	1.718	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:34
Subject Sample # 2	0.261	21:35
Breath Volume	1.375	21:35
Air Blank	0.000	21:36
Control Test	0.077	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 05/17/2020 20:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORJAS,NOE,A
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: BORJAS,NOE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: BEAUCHAMP

DATE/TIME OF TEST: 04/27/2020 21:29:12

Diagnostics Check	OK	21:22
Air Blank	0.000	21:22
Control Test	0.078	21:23
Air Blank	0.000	21:23
Subject Sample # 1	0.000	21:24
Breath Volume	4.132	21:24
Air Blank	0.000	21:24
Air Blank	0.000	21:26
Subject Sample # 2	0.000	21:27
Breath Volume	3.597	21:27
Air Blank	0.000	21:28
Control Test	0.077	21:28
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 05/17/2020 20:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LOGAN,JONATHAN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: BURDETT

DATE/TIME OF TEST: 04/30/2020 21:12:01

Diagnostics Check	OK	21:05
Air Blank	0.000	21:06
Control Test	0.078	21:06
Air Blank	0.000	21:06
Subject Sample # 1	0.196	21:07
Breath Volume	2.781	21:07
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample # 2	0.193	21:10
Breath Volume	2.132	21:10
Air Blank	0.000	21:11
Control Test	0.076	21:11
Air Blank	0.000	21:11
Diagnostics Check	OK	21:12

Data Download Date/Time: 05/17/2020 20:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 03/28/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28719080A1

DRY GAS EXPIRY DATE: 12/05/2021

VIOLATION CODE: DUI

ARREST TIME: 21:14

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: SAINVIL

DATE/TIME OF TEST: 04/12/2020 21:33:55

Diagnostics Check		OK	21:18	
Air Blank		0.000	21:18	
Control Test		0.081	21:19	
Air Blank		0.000	21:19	
Subject Sample	# 1	0.279	21:22	Volume Not Met
Breath Volume		0.148	21:22	
Air Blank		0.000	21:23	
Air Blank		0.000	21:25	
Subject Sample	# 2	0.323	21:26	
Breath Volume		2.738	21:26	
Air Blank		0.000	21:27	
Air Blank		0.000	21:29	
Subject Sample	# 3	0.309	21:32	Volume Not Met
Breath Volume		0.023	21:32	
Air Blank		0.000	21:32	
Control Test		0.080	21:33	
Air Blank		0.000	21:33	
Diagnostics Check		OK	21:33	

Data Download Date/Time: 04/29/2020 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 03/28/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28719080A1

DRY GAS EXPIRY DATE: 12/05/2021

VIOLATION CODE: DUI

ARREST TIME: 21:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: HARTLEY

DATE/TIME OF TEST: 04/17/2020 21:25:59

Diagnostics Check	OK	21:21	
Air Blank	0.000	21:21	
Control Test	0.081	21:22	
Air Blank	0.000	21:22	
Subject Sample	# 1	21:24	Subject Test Refused
Breath Volume	0.000	21:24	
Air Blank	0.000	21:25	
Control Test	0.080	21:25	
Air Blank	0.000	21:25	
Diagnostics Check	OK	21:25	

Data Download Date/Time: 04/29/2020 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 03/28/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28719080A1

DRY GAS EXPIRY DATE: 12/05/2021

VIOLATION CODE: DUI

ARREST TIME: 04:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 04/18/2020 04:26:57

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.081	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.000	04:22
Breath Volume		1.753	04:22
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.000	04:25
Breath Volume		1.300	04:25
Air Blank		0.000	04:26
Control Test		0.080	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 04/29/2020 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,COLE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MCDONALD,COLE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 03/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: JOLEENE

DATE/TIME OF TEST: 04/24/2020 23:01:30

Diagnostics Check	OK	22:53
Air Blank	0.000	22:54
Control Test	0.081	22:54
Air Blank	0.000	22:55
Subject Sample # 1	0.148	22:56
Breath Volume	2.589	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.159	22:59
Breath Volume	2.304	22:59
Air Blank	0.000	23:00
Control Test	0.080	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 04/29/2020 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLCROFT,AMANDA,L
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MEEHAN,ROBERT
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 03/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: LUMSDON

DATE/TIME OF TEST: 04/04/2020 17:26:23

Diagnostics Check	OK	17:20
Air Blank	0.000	17:20
Control Test	0.081	17:21
Air Blank	0.000	17:21
Subject Sample # 1	0.158	17:22
Breath Volume	2.054	17:22
Air Blank	0.000	17:22
Air Blank	0.000	17:24
Subject Sample # 2	0.163	17:24
Breath Volume	1.628	17:24
Air Blank	0.000	17:25
Control Test	0.080	17:25
Air Blank	0.000	17:26
Diagnostics Check	OK	17:26

Data Download Date/Time: 04/29/2020 12:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: STANLEY,JOANNE
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 03/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: BROWNE

DATE/TIME OF TEST: 04/19/2020 02:17:57

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.082	02:12
Air Blank		0.000	02:12
Subject Sample	# 1	0.018	02:13
Breath Volume		2.238	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:15
Subject Sample	# 2	0.015	02:16
Breath Volume		1.175	02:16
Air Blank		0.000	02:17
Control Test		0.081	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 04/19/2020 22:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 15:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:06

SUBJECT'S LAST NAME: BURGOYNE

DATE/TIME OF TEST: 04/01/2020 16:32:46

Diagnostics Check	OK	16:26
Air Blank	0.000	16:26
Control Test	0.079	16:27
Air Blank	0.000	16:27
Subject Sample # 1	0.000	16:28
Breath Volume	2.335	16:28
Air Blank	0.000	16:28
Air Blank	0.000	16:30
Subject Sample # 2	0.000	16:31
Breath Volume	3.035	16:31
Air Blank	0.000	16:31
Control Test	0.079	16:32
Air Blank	0.000	16:32
Diagnostics Check	OK	16:32

Data Download Date/Time: 04/23/2020 21:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DELLECHIAIE,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BEDMARSKI

DATE/TIME OF TEST: 04/03/2020 00:56:08

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:52	
Control Test	0.079	00:53	
Air Blank	0.000	00:53	
Subject Sample # 1	0.309	00:54	
Breath Volume	1.792	00:54	
Air Blank		00:55	Purge Fail
Air Blank	0.000	00:56	

Data Download Date/Time: 04/23/2020 21:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANWORP,JUDSON,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: EALY,STEVEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 12:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:48

SUBJECT'S LAST NAME: NESSMITH

DATE/TIME OF TEST: 04/04/2020 13:13:35

Diagnostics Check	OK	13:10	
Air Blank	0.000	13:10	
Control Test	0.079	13:11	
Air Blank	0.000	13:11	
Subject Sample	# 1	13:12	Subject Test Refused
Breath Volume	0.000	13:12	
Air Blank	0.000	13:12	
Control Test	0.079	13:12	
Air Blank	0.000	13:13	
Diagnostics Check	OK	13:13	

Data Download Date/Time: 04/23/2020 21:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: ENZOR

DATE/TIME OF TEST: 04/07/2020 00:45:25

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:42	
Control Test	0.079	00:43	
Air Blank	0.000	00:43	
Subject Sample	# 1	00:43	Subject Test Refused
Breath Volume	0.000	00:43	
Air Blank	0.000	00:44	
Control Test	0.080	00:44	
Air Blank	0.000	00:45	
Diagnostics Check	OK	00:45	

Data Download Date/Time: 04/23/2020 21:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GUY, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 04/08/2020 22:52:47

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:45	
Control Test	0.080	22:46	
Air Blank	0.000	22:46	
Subject Sample # 1	0.085	22:48	
Breath Volume	1.109	22:48	
Air Blank	0.000	22:48	
Air Blank	0.000	22:50	
Subject Sample # 2	0.073	22:50	
Breath Volume	1.367	22:50	
Air Blank	0.000	22:51	
Control Test	0.073	22:52	Control Outside Tolerance
Air Blank	0.000	22:52	

Data Download Date/Time: 04/23/2020 21:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SURITA,CHRISTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SURITA,CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:21

SUBJECT'S LAST NAME: BOWERMAN

DATE/TIME OF TEST: 04/09/2020 14:53:41

Diagnostics Check	OK	14:45	
Air Blank	0.000	14:45	
Control Test	0.079	14:45	
Air Blank	0.000	14:46	
Subject Sample # 1	0.048	14:48	
Breath Volume	1.417	14:48	
Air Blank	0.000	14:49	
Air Blank	0.000	14:50	
Subject Sample # 2	0.042	14:51	
Breath Volume	1.226	14:51	
Air Blank	0.000	14:52	
Control Test	0.073	14:53	Control Outside Tolerance
Air Blank	0.000	14:53	

Data Download Date/Time: 04/23/2020 21:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: FERRETTI

DATE/TIME OF TEST: 04/24/2020 21:27:43

Diagnostics Check	OK	21:24	
Air Blank	0.000	21:25	
Control Test	0.080	21:25	
Air Blank	0.000	21:26	
Subject Sample	# 1	21:26	Subject Test Refused
Breath Volume	0.000	21:26	
Air Blank	0.000	21:26	
Control Test	0.080	21:27	
Air Blank	0.000	21:27	
Diagnostics Check	OK	21:27	

Data Download Date/Time: 05/14/2020 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO-FELIX,DEP,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUGO-FELIX,DEP
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HANSON

DATE/TIME OF TEST: 04/28/2020 00:28:04

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.079	00:21
Air Blank	0.000	00:22
Subject Sample # 1	0.202	00:23
Breath Volume	2.820	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.205	00:26
Breath Volume	2.421	00:26
Air Blank	0.000	00:27
Control Test	0.080	00:27
Air Blank	0.000	00:27
Diagnostics Check	OK	00:28

Data Download Date/Time: 05/14/2020 14:11