

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA,J,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: GARCIA,J  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 03/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/25/2020 15:14:15

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 15:11 |                      |
| Air Blank         | 0.000 | 15:11 |                      |
| Control Test      | 0.080 | 15:12 |                      |
| Air Blank         | 0.000 | 15:12 |                      |
| Subject Sample    | # 1   | 15:12 | Subject Test Refused |
| Breath Volume     | 0.000 | 15:12 |                      |
| Air Blank         | 0.000 | 15:13 |                      |
| Control Test      | 0.079 | 15:13 |                      |
| Air Blank         | 0.000 | 15:14 |                      |
| Diagnostics Check | OK    | 15:14 |                      |

---

Data Download Date/Time: 05/05/2020 06:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER,KENNETH,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: WINCHESTER,A  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 03/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 10:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: AUBRY

DATE/TIME OF TEST: 04/14/2020 11:46:16

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 11:42 |                      |
| Air Blank         | 0.000 | 11:43 |                      |
| Control Test      | 0.078 | 11:43 |                      |
| Air Blank         | 0.000 | 11:44 |                      |
| Subject Sample    | # 1   | 11:44 | Subject Test Refused |
| Breath Volume     | 0.000 | 11:44 |                      |
| Air Blank         | 0.000 | 11:45 |                      |
| Control Test      | 0.078 | 11:45 |                      |
| Air Blank         | 0.000 | 11:46 |                      |
| Diagnostics Check | OK    | 11:46 |                      |

---

Data Download Date/Time: 04/15/2020 18:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOSTALY, STEVENS,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: OSTROV, VANESSA  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 04/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: SHEVELEV

DATE/TIME OF TEST: 04/19/2020 00:36:12

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:33 |                      |
| Air Blank         | 0.000 | 00:33 |                      |
| Control Test      | 0.079 | 00:34 |                      |
| Air Blank         | 0.000 | 00:34 |                      |
| Subject Sample    | # 1   | 00:34 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:34 |                      |
| Air Blank         | 0.000 | 00:35 |                      |
| Control Test      | 0.078 | 00:35 |                      |
| Air Blank         | 0.000 | 00:36 |                      |
| Diagnostics Check | OK    | 00:36 |                      |

---

Data Download Date/Time: 05/27/2020 21:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MITCHELL,ALEXANDERA  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916303  
DRY GAS EXPIRY DATE: 06/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: RUDY

DATE/TIME OF TEST: 04/01/2020 04:13:40

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:07 |                      |
| Air Blank         | 0.000 | 04:08 |                      |
| Control Test      | 0.081 | 04:08 |                      |
| Air Blank         | 0.000 | 04:09 |                      |
| Subject Sample    | # 1   | 04:12 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:12 |                      |
| Air Blank         | 0.000 | 04:12 |                      |
| Control Test      | 0.080 | 04:13 |                      |
| Air Blank         | 0.000 | 04:13 |                      |
| Diagnostics Check | OK    | 04:13 |                      |

---

Data Download Date/Time: 04/23/2020 20:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REFUSE, RONY,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: REFUSE, RONY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916303  
DRY GAS EXPIRY DATE: 06/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: MACDONALD

DATE/TIME OF TEST: 04/04/2020 21:25:33

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:22 |                      |
| Air Blank         | 0.000 | 21:23 |                      |
| Control Test      | 0.084 | 21:23 |                      |
| Air Blank         | 0.000 | 21:23 |                      |
| Subject Sample    | # 1   | 21:24 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:24 |                      |
| Air Blank         | 0.000 | 21:24 |                      |
| Control Test      | 0.083 | 21:24 |                      |
| Air Blank         | 0.000 | 21:25 |                      |
| Diagnostics Check | OK    | 21:25 |                      |

---

Data Download Date/Time: 04/23/2020 20:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MULLINS,JOHN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916303  
DRY GAS EXPIRY DATE: 06/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: STIBBINS

DATE/TIME OF TEST: 04/05/2020 22:07:36

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:03 |                      |
| Air Blank         | 0.000 | 22:04 |                      |
| Control Test      | 0.084 | 22:04 |                      |
| Air Blank         | 0.000 | 22:04 |                      |
| Subject Sample    | # 1   | 22:06 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:06 |                      |
| Air Blank         | 0.000 | 22:06 |                      |
| Control Test      | 0.083 | 22:06 |                      |
| Air Blank         | 0.000 | 22:07 |                      |
| Diagnostics Check | OK    | 22:07 |                      |

---

Data Download Date/Time: 04/23/2020 20:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LONGMIRE,BENJAMIN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916303  
DRY GAS EXPIRY DATE: 06/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: LIVESAY

DATE/TIME OF TEST: 04/06/2020 03:54:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:48 |
| Air Blank          | 0.000 | 03:48 |
| Control Test       | 0.082 | 03:48 |
| Air Blank          | 0.000 | 03:49 |
| Subject Sample # 1 | 0.177 | 03:50 |
| Breath Volume      | 1.835 | 03:50 |
| Air Blank          | 0.000 | 03:51 |
| Air Blank          | 0.000 | 03:52 |
| Subject Sample # 2 | 0.193 | 03:53 |
| Breath Volume      | 1.636 | 03:53 |
| Air Blank          | 0.000 | 03:54 |
| Control Test       | 0.081 | 03:54 |
| Air Blank          | 0.000 | 03:54 |
| Diagnostics Check  | OK    | 03:54 |

---

Data Download Date/Time: 04/23/2020 20:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCE,SCOTT,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MILLER,HOWARD  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916303  
DRY GAS EXPIRY DATE: 06/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 07:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: MCCLELLAN

DATE/TIME OF TEST: 04/08/2020 08:19:02

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 08:15 |                      |
| Air Blank         | 0.000 | 08:15 |                      |
| Control Test      | 0.083 | 08:16 |                      |
| Air Blank         | 0.000 | 08:16 |                      |
| Subject Sample    | # 1   | 08:17 | Subject Test Refused |
| Breath Volume     | 0.000 | 08:17 |                      |
| Air Blank         | 0.000 | 08:18 |                      |
| Control Test      | 0.082 | 08:18 |                      |
| Air Blank         | 0.000 | 08:18 |                      |
| Diagnostics Check | OK    | 08:19 |                      |

---

Data Download Date/Time: 04/23/2020 20:17



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNAVELY,KENT,J  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BUSCH,TYLER  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916303  
DRY GAS EXPIRY DATE: 06/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/08/2020 17:27:09

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 17:23 |                      |
| Air Blank         | 0.000 | 17:23 |                      |
| Control Test      | 0.081 | 17:24 |                      |
| Air Blank         | 0.000 | 17:24 |                      |
| Subject Sample    | # 1   | 17:25 | Subject Test Refused |
| Breath Volume     | 0.000 | 17:25 |                      |
| Air Blank         | 0.000 | 17:26 |                      |
| Control Test      | 0.081 | 17:26 |                      |
| Air Blank         | 0.000 | 17:27 |                      |
| Diagnostics Check | OK    | 17:27 |                      |

---

Data Download Date/Time: 04/23/2020 20:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SURITA,CHRISTIAN,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SURITA,CHRISTIAN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916303  
DRY GAS EXPIRY DATE: 06/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:21

SUBJECT'S LAST NAME: BOWERMAN

DATE/TIME OF TEST: 04/09/2020 15:04:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:57 |
| Air Blank          | 0.000 | 14:58 |
| Control Test       | 0.082 | 14:58 |
| Air Blank          | 0.000 | 14:59 |
| Subject Sample # 1 | 0.050 | 15:00 |
| Breath Volume      | 2.378 | 15:00 |
| Air Blank          | 0.000 | 15:00 |
| Air Blank          | 0.000 | 15:02 |
| Subject Sample # 2 | 0.052 | 15:03 |
| Breath Volume      | 3.386 | 15:03 |
| Air Blank          | 0.000 | 15:03 |
| Control Test       | 0.082 | 15:04 |
| Air Blank          | 0.000 | 15:04 |
| Diagnostics Check  | OK    | 15:04 |

---

Data Download Date/Time: 04/23/2020 20:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILTSE,CORPORAL W.,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: WILTSE,CORPORAL W.  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916303  
DRY GAS EXPIRY DATE: 06/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: MEAD

DATE/TIME OF TEST: 04/15/2020 01:03:15

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 00:56 |
| Air Blank         |     | 0.000 | 00:56 |
| Control Test      |     | 0.082 | 00:57 |
| Air Blank         |     | 0.000 | 00:57 |
| Subject Sample    | # 1 | 0.170 | 00:58 |
| Breath Volume     |     | 3.156 | 00:58 |
| Air Blank         |     | 0.000 | 00:59 |
| Air Blank         |     | 0.000 | 01:01 |
| Subject Sample    | # 2 | 0.175 | 01:01 |
| Breath Volume     |     | 3.191 | 01:01 |
| Air Blank         |     | 0.000 | 01:02 |
| Control Test      |     | 0.081 | 01:02 |
| Air Blank         |     | 0.000 | 01:03 |
| Diagnostics Check |     | OK    | 01:03 |

---

Data Download Date/Time: 04/23/2020 20:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILTSE,CORPORAL W.,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: SKALCO,DEPUTY R.  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916303  
DRY GAS EXPIRY DATE: 06/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 04/15/2020 02:08:58

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 01:59 |
| Air Blank         |     | 0.000 | 01:59 |
| Control Test      |     | 0.080 | 02:00 |
| Air Blank         |     | 0.000 | 02:00 |
| Subject Sample    | # 1 | 0.223 | 02:04 |
| Breath Volume     |     | 2.570 | 02:04 |
| Air Blank         |     | 0.000 | 02:04 |
| Air Blank         |     | 0.000 | 02:06 |
| Subject Sample    | # 2 | 0.210 | 02:07 |
| Breath Volume     |     | 2.589 | 02:07 |
| Air Blank         |     | 0.000 | 02:08 |
| Control Test      |     | 0.080 | 02:08 |
| Air Blank         |     | 0.000 | 02:08 |
| Diagnostics Check |     | OK    | 02:08 |

---

Data Download Date/Time: 04/23/2020 20:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: IFFT,GARRETT,T  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SMITH,WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: AG916303  
DRY GAS EXPIRY DATE: 06/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: DIEGO

DATE/TIME OF TEST: 04/17/2020 21:57:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:51 |
| Air Blank          | 0.000 | 21:51 |
| Control Test       | 0.082 | 21:51 |
| Air Blank          | 0.000 | 21:52 |
| Subject Sample # 1 | 0.155 | 21:52 |
| Breath Volume      | 3.371 | 21:52 |
| Air Blank          | 0.000 | 21:53 |
| Air Blank          | 0.000 | 21:55 |
| Subject Sample # 2 | 0.160 | 21:55 |
| Breath Volume      | 3.406 | 21:55 |
| Air Blank          | 0.000 | 21:56 |
| Control Test       | 0.081 | 21:56 |
| Air Blank          | 0.000 | 21:57 |
| Diagnostics Check  | OK    | 21:57 |

---

Data Download Date/Time: 04/23/2020 20:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REFUSE, RONY,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: REFUSE, RONY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916303  
DRY GAS EXPIRY DATE: 06/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: TREXLER

DATE/TIME OF TEST: 04/18/2020 03:16:22

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:11 |                      |
| Air Blank         | 0.000 | 03:12 |                      |
| Control Test      | 0.081 | 03:12 |                      |
| Air Blank         | 0.000 | 03:13 |                      |
| Subject Sample    | # 1   | 03:14 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:14 |                      |
| Air Blank         | 0.000 | 03:15 |                      |
| Control Test      | 0.081 | 03:15 |                      |
| Air Blank         | 0.000 | 03:16 |                      |
| Diagnostics Check | OK    | 03:16 |                      |

---

Data Download Date/Time: 04/23/2020 20:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNAVELY,KENT,J  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LANG,LYDIA  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916303  
DRY GAS EXPIRY DATE: 06/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: MCELWANEY

DATE/TIME OF TEST: 04/18/2020 15:07:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:00 |
| Air Blank          | 0.000 | 15:01 |
| Control Test       | 0.081 | 15:01 |
| Air Blank          | 0.000 | 15:02 |
| Subject Sample # 1 | 0.298 | 15:02 |
| Breath Volume      | 1.515 | 15:02 |
| Air Blank          | 0.000 | 15:03 |
| Air Blank          | 0.000 | 15:05 |
| Subject Sample # 2 | 0.289 | 15:06 |
| Breath Volume      | 1.535 | 15:06 |
| Air Blank          | 0.000 | 15:06 |
| Control Test       | 0.081 | 15:07 |
| Air Blank          | 0.000 | 15:07 |
| Diagnostics Check  | OK    | 15:07 |

---

Data Download Date/Time: 04/23/2020 20:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOLNAR,JARROD,T  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: BROWN,JASON  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 03/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: BUI  
ARREST TIME: 15:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 04/11/2020 15:56:04

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 15:53 |                      |
| Air Blank         | 0.000 | 15:53 |                      |
| Control Test      | 0.080 | 15:54 |                      |
| Air Blank         | 0.000 | 15:54 |                      |
| Subject Sample    | # 1   | 15:54 | Subject Test Refused |
| Breath Volume     | 0.000 | 15:54 |                      |
| Air Blank         | 0.000 | 15:55 |                      |
| Control Test      | 0.080 | 15:55 |                      |
| Air Blank         | 0.000 | 15:55 |                      |
| Diagnostics Check | OK    | 15:56 |                      |

---

Data Download Date/Time: 04/28/2020 17:05



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAHR,DANIEL,E  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: HAHR,DANIEL  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 03/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: BUI  
ARREST TIME: 16:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: CORONA MORALES

DATE/TIME OF TEST: 04/14/2020 17:46:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:40 |
| Air Blank          | 0.000 | 17:40 |
| Control Test       | 0.080 | 17:41 |
| Air Blank          | 0.000 | 17:41 |
| Subject Sample # 1 | 0.179 | 17:42 |
| Breath Volume      | 3.250 | 17:42 |
| Air Blank          | 0.000 | 17:42 |
| Air Blank          | 0.000 | 17:44 |
| Subject Sample # 2 | 0.177 | 17:45 |
| Breath Volume      | 3.218 | 17:45 |
| Air Blank          | 0.000 | 17:46 |
| Control Test       | 0.079 | 17:46 |
| Air Blank          | 0.000 | 17:46 |
| Diagnostics Check  | OK    | 17:46 |

---

Data Download Date/Time: 04/28/2020 17:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORBIN,BRITTON,L  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: CORBIN,BRITTON  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 03/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: BUI  
ARREST TIME: 19:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/20/2020 20:43:35

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 20:36 |
| Air Blank         |     | 0.000 | 20:37 |
| Control Test      |     | 0.079 | 20:37 |
| Air Blank         |     | 0.000 | 20:37 |
| Subject Sample    | # 1 | 0.157 | 20:38 |
| Breath Volume     |     | 1.605 | 20:38 |
| Air Blank         |     | 0.000 | 20:39 |
| Air Blank         |     | 0.000 | 20:41 |
| Subject Sample    | # 2 | 0.164 | 20:42 |
| Breath Volume     |     | 1.714 | 20:42 |
| Air Blank         |     | 0.000 | 20:42 |
| Control Test      |     | 0.080 | 20:42 |
| Air Blank         |     | 0.000 | 20:43 |
| Diagnostics Check |     | OK    | 20:43 |

---

Data Download Date/Time: 04/28/2020 17:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARTLETT,SHELTON,T  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: HICKS,ROBERT  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 03/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: BUI  
ARREST TIME: 19:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: POOLER

DATE/TIME OF TEST: 04/25/2020 20:04:51

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:01 |                      |
| Air Blank         | 0.000 | 20:02 |                      |
| Control Test      | 0.079 | 20:02 |                      |
| Air Blank         | 0.000 | 20:03 |                      |
| Subject Sample    | # 1   | 20:03 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:03 |                      |
| Air Blank         | 0.000 | 20:03 |                      |
| Control Test      | 0.079 | 20:04 |                      |
| Air Blank         | 0.000 | 20:04 |                      |
| Diagnostics Check | OK    | 20:04 |                      |

---

Data Download Date/Time: 04/28/2020 17:05

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

|  |                                 |
|--|---------------------------------|
| INSTRUMENT REGISTERED TO: FWC                    | Date or Time Changed: No        |
| SOFTWARE VERSION: 8100.27                        | DRY GAS STANDARD LOT #: 1060911 |
| BREATH TEST OPERATOR: BASFORD,NICOLE,M           | DRY GAS EXPIRY DATE: 12/21/2021 |
| BREATH TEST OPERATOR AGENCY: FFWCC (Panama City) | VIOLATION CODE: BUI             |
| ARRESTING OFFICER: WICKER,STEPHEN                | ARREST TIME: 20:10              |
| ARRESTING OFFICER AGENCY: FFWCC (Panama City)    | GENDER: Male                    |
| LAST AGENCY INSPECTION: 04/28/2020               | OBSERVATION PERIOD BEGAN: 20:15 |

|                              |                               |          |
|------------------------------|-------------------------------|----------|
| SUBJECT'S LAST NAME: MESSAGE | DATE/TIME OF TEST: 04/28/2020 | 20:46:35 |
|------------------------------|-------------------------------|----------|

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:40 |
| Air Blank          | 0.000 | 20:40 |
| Control Test       | 0.080 | 20:40 |
| Air Blank          | 0.000 | 20:41 |
| Subject Sample # 1 | 0.140 | 20:41 |
| Breath Volume      | 2.847 | 20:41 |
| Air Blank          | 0.000 | 20:42 |
| Air Blank          | 0.000 | 20:44 |
| Subject Sample # 2 | 0.133 | 20:45 |
| Breath Volume      | 1.855 | 20:45 |
| Air Blank          | 0.000 | 20:45 |
| Control Test       | 0.079 | 20:45 |
| Air Blank          | 0.000 | 20:46 |
| Diagnostics Check  | OK    | 20:46 |

---

Data Download Date/Time: 05/28/2020 20:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TOMBERLIN,KEVIN,W  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: TOMBERLIN,KEVIN  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 04/19/2020 00:13:39

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:10 |                      |
| Air Blank         | 0.000 | 00:11 |                      |
| Control Test      | 0.079 | 00:11 |                      |
| Air Blank         | 0.000 | 00:12 |                      |
| Subject Sample    | # 1   | 00:12 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:12 |                      |
| Air Blank         | 0.000 | 00:12 |                      |
| Control Test      | 0.079 | 00:13 |                      |
| Air Blank         | 0.000 | 00:13 |                      |
| Diagnostics Check | OK    | 00:13 |                      |

---

Data Download Date/Time: 05/29/2020 09:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRERO,FRAICOR,  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: TERRERO,FRAICOR  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: FRANCISCO

DATE/TIME OF TEST: 04/19/2020 03:44:15

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 03:37 |
| Air Blank         |     | 0.000 | 03:38 |
| Control Test      |     | 0.078 | 03:38 |
| Air Blank         |     | 0.000 | 03:39 |
| Subject Sample    | # 1 | 0.246 | 03:39 |
| Breath Volume     |     | 2.972 | 03:39 |
| Air Blank         |     | 0.000 | 03:40 |
| Air Blank         |     | 0.000 | 03:42 |
| Subject Sample    | # 2 | 0.247 | 03:42 |
| Breath Volume     |     | 3.074 | 03:42 |
| Air Blank         |     | 0.000 | 03:43 |
| Control Test      |     | 0.078 | 03:43 |
| Air Blank         |     | 0.000 | 03:44 |
| Diagnostics Check |     | OK    | 03:44 |

---

Data Download Date/Time: 05/29/2020 09:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JIMENEZ, WILLIAM,  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: JIMENEZ, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 10:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 04/19/2020 11:14:24

---

|                    |       |       |            |
|--------------------|-------|-------|------------|
| Diagnostics Check  | OK    | 11:08 |            |
| Air Blank          | 0.000 | 11:08 |            |
| Control Test       | 0.078 | 11:08 |            |
| Air Blank          | 0.000 | 11:09 |            |
| Subject Sample # 1 | 0.096 | 11:09 |            |
| Breath Volume      | 2.875 | 11:09 |            |
| Air Blank          | 0.000 | 11:10 |            |
| Air Blank          | 0.000 | 11:12 |            |
| Subject Sample # 2 | 0.098 | 11:12 |            |
| Breath Volume      | 2.433 | 11:12 |            |
| Air Blank          | 0.000 | 11:13 |            |
| Control Test       | 0.078 | 11:13 |            |
| Air Blank          |       | 11:13 | RFI Detect |
| Air Blank          | 0.000 | 11:14 |            |

---

Data Download Date/Time: 05/29/2020 09:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRERO,FRAICOR,  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: TERRERO,FRAICOR  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: POYTHRESS

DATE/TIME OF TEST: 04/20/2020 01:48:34

---

|                   |           |       |               |
|-------------------|-----------|-------|---------------|
| Diagnostics Check | OK        | 01:41 |               |
| Air Blank         | 0.000     | 01:42 |               |
| Control Test      | 0.078     | 01:42 |               |
| Air Blank         | 0.000     | 01:42 |               |
| Subject Sample    | # 1 0.169 | 01:43 |               |
| Breath Volume     | 2.734     | 01:43 |               |
| Air Blank         | 0.000     | 01:44 |               |
| Air Blank         | 0.000     | 01:46 |               |
| Subject Sample    | # 2       | 01:47 | Slope Not Met |
| Breath Volume     | 1.417     | 01:47 |               |
| Air Blank         | 0.000     | 01:47 |               |
| Control Test      | 0.078     | 01:47 |               |
| Air Blank         | 0.000     | 01:48 |               |
| Diagnostics Check | OK        | 01:48 |               |

---

Data Download Date/Time: 05/29/2020 09:48



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRERO,FRAICOR,  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: TERRERO,FRAICOR  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: POYTHRESS

DATE/TIME OF TEST: 04/20/2020 01:57:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:51 |
| Air Blank          | 0.000 | 01:51 |
| Control Test       | 0.078 | 01:51 |
| Air Blank          | 0.000 | 01:52 |
| Subject Sample # 1 | 0.164 | 01:52 |
| Breath Volume      | 1.925 | 01:52 |
| Air Blank          | 0.000 | 01:53 |
| Air Blank          | 0.000 | 01:55 |
| Subject Sample # 2 | 0.160 | 01:55 |
| Breath Volume      | 1.718 | 01:55 |
| Air Blank          | 0.000 | 01:56 |
| Control Test       | 0.079 | 01:56 |
| Air Blank          | 0.000 | 01:57 |
| Diagnostics Check  | OK    | 01:57 |

---

Data Download Date/Time: 05/29/2020 09:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLE,CALVIN,K  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: HOLLE,CALVIN  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: MCFARLING

DATE/TIME OF TEST: 04/25/2020 20:32:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:25 |
| Air Blank          | 0.000 | 20:25 |
| Control Test       | 0.079 | 20:26 |
| Air Blank          | 0.000 | 20:26 |
| Subject Sample # 1 | 0.229 | 20:27 |
| Breath Volume      | 1.980 | 20:27 |
| Air Blank          | 0.000 | 20:27 |
| Air Blank          | 0.000 | 20:29 |
| Subject Sample # 2 | 0.213 | 20:30 |
| Breath Volume      | 2.070 | 20:30 |
| Air Blank          | 0.000 | 20:31 |
| Control Test       | 0.077 | 20:31 |
| Air Blank          | 0.000 | 20:32 |
| Diagnostics Check  | OK    | 20:32 |

---

Data Download Date/Time: 05/29/2020 09:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PETERSON, RONALD, E  
BREATH TEST OPERATOR AGENCY: Bradenton PD  
ARRESTING OFFICER: MOORE, ADAM  
ARRESTING OFFICER AGENCY: Bradenton PD  
LAST AGENCY INSPECTION: 04/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: FAISON

DATE/TIME OF TEST: 04/28/2020 05:15:04

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 05:02 |                |
| Air Blank         |     | 0.000 | 05:02 |                |
| Control Test      |     | 0.080 | 05:02 |                |
| Air Blank         |     | 0.000 | 05:03 |                |
| Subject Sample    | # 1 | 0.116 | 05:06 | Volume Not Met |
| Breath Volume     |     | 0.589 | 05:06 |                |
| Air Blank         |     | 0.000 | 05:07 |                |
| Air Blank         |     | 0.000 | 05:08 |                |
| Subject Sample    | # 2 | 0.114 | 05:10 |                |
| Breath Volume     |     | 1.406 | 05:10 |                |
| Air Blank         |     | 0.000 | 05:11 |                |
| Air Blank         |     | 0.000 | 05:13 |                |
| Subject Sample    | # 3 | 0.109 | 05:13 |                |
| Breath Volume     |     | 1.582 | 05:13 |                |
| Air Blank         |     | 0.000 | 05:14 |                |
| Control Test      |     | 0.079 | 05:14 |                |
| Air Blank         |     | 0.000 | 05:14 |                |
| Diagnostics Check |     | OK    | 05:15 |                |

---

Data Download Date/Time: 05/29/2020 09:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PODHYSKI,PHILIP,A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: PODHYSKI,PHILIP  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: CEDILLO TORRES

DATE/TIME OF TEST: 04/29/2020 01:15:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:06 |
| Air Blank          | 0.000 | 01:07 |
| Control Test       | 0.080 | 01:07 |
| Air Blank          | 0.000 | 01:07 |
| Subject Sample # 1 | 0.026 | 01:10 |
| Breath Volume      | 2.539 | 01:10 |
| Air Blank          | 0.000 | 01:10 |
| Air Blank          | 0.000 | 01:12 |
| Subject Sample # 2 | 0.023 | 01:13 |
| Breath Volume      | 2.878 | 01:13 |
| Air Blank          | 0.000 | 01:14 |
| Control Test       | 0.079 | 01:14 |
| Air Blank          | 0.000 | 01:15 |
| Diagnostics Check  | OK    | 01:15 |

---

Data Download Date/Time: 05/29/2020 09:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 867139  
DRY GAS EXPIRY DATE: 11/15/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HAGAN

DATE/TIME OF TEST: 04/30/2020 02:28:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:21 |
| Air Blank          | 0.000 | 02:22 |
| Control Test       | 0.079 | 02:22 |
| Air Blank          | 0.000 | 02:23 |
| Subject Sample # 1 | 0.166 | 02:23 |
| Breath Volume      | 3.691 | 02:23 |
| Air Blank          | 0.000 | 02:24 |
| Air Blank          | 0.000 | 02:26 |
| Subject Sample # 2 | 0.164 | 02:27 |
| Breath Volume      | 3.585 | 02:27 |
| Air Blank          | 0.000 | 02:27 |
| Control Test       | 0.078 | 02:27 |
| Air Blank          | 0.000 | 02:28 |
| Diagnostics Check  | OK    | 02:28 |

---

Data Download Date/Time: 05/29/2020 09:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROMAN,A,C  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BARRERO,J  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: CHIROY CAC

DATE/TIME OF TEST: 04/04/2020 16:31:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:25 |
| Air Blank          | 0.000 | 16:25 |
| Control Test       | 0.080 | 16:26 |
| Air Blank          | 0.000 | 16:26 |
| Subject Sample # 1 | 0.224 | 16:27 |
| Breath Volume      | 1.449 | 16:27 |
| Air Blank          | 0.000 | 16:27 |
| Air Blank          | 0.000 | 16:29 |
| Subject Sample # 2 | 0.227 | 16:29 |
| Breath Volume      | 1.371 | 16:29 |
| Air Blank          | 0.000 | 16:30 |
| Control Test       | 0.080 | 16:31 |
| Air Blank          | 0.000 | 16:31 |
| Diagnostics Check  | OK    | 16:31 |

---

Data Download Date/Time: 04/08/2020 10:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOULD,LEONARD,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: ENGREN,GEORGE  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 03/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:23

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/06/2020 15:53:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:47 |
| Air Blank          | 0.000 | 15:48 |
| Control Test       | 0.080 | 15:48 |
| Air Blank          | 0.000 | 15:48 |
| Subject Sample # 1 | 0.000 | 15:49 |
| Breath Volume      | 1.855 | 15:49 |
| Air Blank          | 0.000 | 15:49 |
| Air Blank          | 0.000 | 15:51 |
| Subject Sample # 2 | 0.000 | 15:52 |
| Breath Volume      | 1.257 | 15:52 |
| Air Blank          | 0.000 | 15:52 |
| Control Test       | 0.079 | 15:53 |
| Air Blank          | 0.000 | 15:53 |
| Diagnostics Check  | OK    | 15:53 |

---

Data Download Date/Time: 04/08/2020 10:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EMANS,ALEX,R  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: PEREZ,M  
ARRESTING OFFICER AGENCY: Naples PD  
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: PURCELL

DATE/TIME OF TEST: 04/03/2020 23:37:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:30 |
| Air Blank          | 0.000 | 23:31 |
| Control Test       | 0.081 | 23:31 |
| Air Blank          | 0.000 | 23:32 |
| Subject Sample # 1 | 0.151 | 23:32 |
| Breath Volume      | 1.273 | 23:32 |
| Air Blank          | 0.000 | 23:33 |
| Air Blank          | 0.000 | 23:35 |
| Subject Sample # 2 | 0.155 | 23:35 |
| Breath Volume      | 1.183 | 23:35 |
| Air Blank          | 0.000 | 23:36 |
| Control Test       | 0.081 | 23:36 |
| Air Blank          | 0.000 | 23:37 |
| Diagnostics Check  | OK    | 23:37 |

---

Data Download Date/Time: 04/13/2020 15:39



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GIBBONS,SEAN,L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: MCCLOSKEY,J  
ARRESTING OFFICER AGENCY: Naples PD  
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: LAPINSKI

DATE/TIME OF TEST: 04/05/2020 01:44:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:37 |
| Air Blank          | 0.000 | 01:37 |
| Control Test       | 0.082 | 01:38 |
| Air Blank          | 0.000 | 01:38 |
| Subject Sample # 1 | 0.000 | 01:39 |
| Breath Volume      | 2.445 | 01:39 |
| Air Blank          | 0.000 | 01:40 |
| Air Blank          | 0.000 | 01:42 |
| Subject Sample # 2 | 0.000 | 01:42 |
| Breath Volume      | 1.539 | 01:42 |
| Air Blank          | 0.000 | 01:43 |
| Control Test       | 0.081 | 01:43 |
| Air Blank          | 0.000 | 01:43 |
| Diagnostics Check  | OK    | 01:44 |

---

Data Download Date/Time: 04/13/2020 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PATTERSON,D,S  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: ZEGERS,BRYAN  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 07:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: VAZQUEZPEREZ

DATE/TIME OF TEST: 04/09/2020 09:44:51

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 09:43 |            |
| Air Blank         | 0.000 | 09:43 |            |
| Control Test      | 0.082 | 09:43 |            |
| Air Blank         |       | 09:44 | RFI Detect |
| Air Blank         | 0.000 | 09:44 |            |

---

Data Download Date/Time: 04/13/2020 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PATTERSON,D,S  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: ZEGERS,BRYAN  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 07:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: VAZQUEZPEREZ

DATE/TIME OF TEST: 04/09/2020 09:54:22

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 09:48 |
| Air Blank         |     | 0.000 | 09:48 |
| Control Test      |     | 0.082 | 09:49 |
| Air Blank         |     | 0.000 | 09:49 |
| Subject Sample    | # 1 | 0.000 | 09:50 |
| Breath Volume     |     | 1.667 | 09:50 |
| Air Blank         |     | 0.000 | 09:50 |
| Air Blank         |     | 0.000 | 09:52 |
| Subject Sample    | # 2 | 0.000 | 09:53 |
| Breath Volume     |     | 1.363 | 09:53 |
| Air Blank         |     | 0.000 | 09:53 |
| Control Test      |     | 0.083 | 09:53 |
| Air Blank         |     | 0.000 | 09:54 |
| Diagnostics Check |     | OK    | 09:54 |

---

Data Download Date/Time: 04/13/2020 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENNEDY,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KENNEDY,DANIEL  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: SKELLY

DATE/TIME OF TEST: 04/10/2020 00:26:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:19 |
| Air Blank          | 0.000 | 00:20 |
| Control Test       | 0.082 | 00:20 |
| Air Blank          | 0.000 | 00:21 |
| Subject Sample # 1 | 0.119 | 00:21 |
| Breath Volume      | 3.121 | 00:21 |
| Air Blank          | 0.000 | 00:22 |
| Air Blank          | 0.000 | 00:24 |
| Subject Sample # 2 | 0.116 | 00:24 |
| Breath Volume      | 3.074 | 00:24 |
| Air Blank          | 0.000 | 00:25 |
| Control Test       | 0.082 | 00:25 |
| Air Blank          | 0.000 | 00:26 |
| Diagnostics Check  | OK    | 00:26 |

---

Data Download Date/Time: 04/13/2020 15:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOEL,Y,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: MILLER,K  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 04/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: LEVENS

DATE/TIME OF TEST: 04/21/2020 22:39:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:31 |
| Air Blank          | 0.000 | 22:32 |
| Control Test       | 0.081 | 22:32 |
| Air Blank          | 0.000 | 22:33 |
| Subject Sample # 1 | 0.118 | 22:33 |
| Breath Volume      | 2.203 | 22:33 |
| Air Blank          | 0.000 | 22:34 |
| Air Blank          | 0.000 | 22:36 |
| Subject Sample # 2 | 0.104 | 22:37 |
| Breath Volume      | 1.632 | 22:37 |
| Air Blank          | 0.000 | 22:38 |
| Control Test       | 0.082 | 22:38 |
| Air Blank          | 0.000 | 22:39 |
| Diagnostics Check  | OK    | 22:39 |

---

Data Download Date/Time: 05/07/2020 12:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAYLOR,E,W  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: TERUZ,G  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: VAUPEL

DATE/TIME OF TEST: 04/22/2020 23:16:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:10 |
| Air Blank          | 0.000 | 23:11 |
| Control Test       | 0.081 | 23:11 |
| Air Blank          | 0.000 | 23:12 |
| Subject Sample # 1 | 0.000 | 23:12 |
| Breath Volume      | 2.847 | 23:12 |
| Air Blank          | 0.000 | 23:12 |
| Air Blank          | 0.000 | 23:14 |
| Subject Sample # 2 | 0.000 | 23:15 |
| Breath Volume      | 2.835 | 23:15 |
| Air Blank          | 0.000 | 23:15 |
| Control Test       | 0.081 | 23:16 |
| Air Blank          | 0.000 | 23:16 |
| Diagnostics Check  | OK    | 23:16 |

---

Data Download Date/Time: 05/07/2020 12:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENNEDY,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KENNEDY,DANIEL  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 04/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: MATHESON

DATE/TIME OF TEST: 04/25/2020 22:19:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:11 |
| Air Blank          | 0.000 | 22:11 |
| Control Test       | 0.080 | 22:12 |
| Air Blank          | 0.000 | 22:12 |
| Subject Sample # 1 | 0.286 | 22:14 |
| Breath Volume      | 2.648 | 22:14 |
| Air Blank          | 0.000 | 22:15 |
| Air Blank          | 0.000 | 22:17 |
| Subject Sample # 2 | 0.286 | 22:18 |
| Breath Volume      | 2.398 | 22:18 |
| Air Blank          | 0.000 | 22:18 |
| Control Test       | 0.080 | 22:19 |
| Air Blank          | 0.000 | 22:19 |
| Diagnostics Check  | OK    | 22:19 |

---

Data Download Date/Time: 05/07/2020 12:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUIGLEY,M,R  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: REILLY,S  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:22

SUBJECT'S LAST NAME: ADARIO

DATE/TIME OF TEST: 04/28/2020 07:54:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:44 |
| Air Blank          | 0.000 | 07:44 |
| Control Test       | 0.081 | 07:45 |
| Air Blank          | 0.000 | 07:45 |
| Subject Sample # 1 | 0.155 | 07:48 |
| Breath Volume      | 1.382 | 07:48 |
| Air Blank          | 0.000 | 07:49 |
| Air Blank          | 0.000 | 07:50 |
| Subject Sample # 2 | 0.139 | 07:52 |
| Breath Volume      | 1.359 | 07:52 |
| Air Blank          | 0.000 | 07:53 |
| Control Test       | 0.079 | 07:53 |
| Air Blank          | 0.000 | 07:54 |
| Diagnostics Check  | OK    | 07:54 |

---

Data Download Date/Time: 05/07/2020 12:33



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARRILLO,XAVIER,R  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: PUKA,M  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 04/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HERNESTO

DATE/TIME OF TEST: 04/29/2020 02:31:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:25 |
| Air Blank          | 0.000 | 02:25 |
| Control Test       | 0.080 | 02:26 |
| Air Blank          | 0.000 | 02:26 |
| Subject Sample # 1 | 0.177 | 02:27 |
| Breath Volume      | 1.582 | 02:27 |
| Air Blank          | 0.000 | 02:27 |
| Air Blank          | 0.000 | 02:29 |
| Subject Sample # 2 | 0.171 | 02:30 |
| Breath Volume      | 2.109 | 02:30 |
| Air Blank          | 0.000 | 02:31 |
| Control Test       | 0.080 | 02:31 |
| Air Blank          | 0.000 | 02:31 |
| Diagnostics Check  | OK    | 02:31 |

---

Data Download Date/Time: 05/07/2020 12:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORRIS,THOMAS,L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: PEREZ,M  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 04/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: FOGEL

DATE/TIME OF TEST: 04/30/2020 02:52:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:46 |
| Air Blank          | 0.000 | 02:46 |
| Control Test       | 0.081 | 02:47 |
| Air Blank          | 0.000 | 02:47 |
| Subject Sample # 1 | 0.166 | 02:48 |
| Breath Volume      | 2.128 | 02:48 |
| Air Blank          | 0.000 | 02:48 |
| Air Blank          | 0.000 | 02:50 |
| Subject Sample # 2 | 0.165 | 02:51 |
| Breath Volume      | 2.289 | 02:51 |
| Air Blank          | 0.000 | 02:52 |
| Control Test       | 0.081 | 02:52 |
| Air Blank          | 0.000 | 02:52 |
| Diagnostics Check  | OK    | 02:52 |

---

Data Download Date/Time: 05/07/2020 12:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENNEDY,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KENNEDY,DANIEL  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 04/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A1  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 04/30/2020 23:56:04

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:52 |                      |
| Air Blank         | 0.000 | 23:52 |                      |
| Control Test      | 0.081 | 23:53 |                      |
| Air Blank         | 0.000 | 23:53 |                      |
| Subject Sample    | # 1   | 23:54 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:54 |                      |
| Air Blank         | 0.000 | 23:55 |                      |
| Control Test      | 0.081 | 23:55 |                      |
| Air Blank         | 0.000 | 23:55 |                      |
| Diagnostics Check | OK    | 23:56 |                      |

---

Data Download Date/Time: 05/07/2020 12:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WRIGHT, JOSHUA, N  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: DIXON, WESLEY  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: VALAZQUEZ RODAS

DATE/TIME OF TEST: 04/08/2020 19:44:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:38 |
| Air Blank          | 0.000 | 19:38 |
| Control Test       | 0.077 | 19:39 |
| Air Blank          | 0.000 | 19:39 |
| Subject Sample # 1 | 0.000 | 19:40 |
| Breath Volume      | 4.417 | 19:40 |
| Air Blank          | 0.000 | 19:40 |
| Air Blank          | 0.000 | 19:42 |
| Subject Sample # 2 | 0.000 | 19:43 |
| Breath Volume      | 4.394 | 19:43 |
| Air Blank          | 0.000 | 19:43 |
| Control Test       | 0.077 | 19:44 |
| Air Blank          | 0.000 | 19:44 |
| Diagnostics Check  | OK    | 19:44 |

---

Data Download Date/Time: 04/14/2020 12:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO,FAUSTO,A  
BREATH TEST OPERATOR AGENCY: Clewiston PD  
ARRESTING OFFICER: ROSARIO,FAUSTO  
ARRESTING OFFICER AGENCY: Clewiston PD  
LAST AGENCY INSPECTION: 03/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35019080A3  
DRY GAS EXPIRY DATE: 02/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 10:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 04/06/2020 10:50:16

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 10:48 |                           |
| Air Blank         | 0.000 | 10:49 |                           |
| Control Test      | 0.000 | 10:49 | Control Outside Tolerance |
| Air Blank         | 0.000 | 10:50 |                           |

---

Data Download Date/Time: 04/07/2020 13:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBB,THOMAS,J  
BREATH TEST OPERATOR AGENCY: Franklin County SO  
ARRESTING OFFICER: JAMES,STEVE  
ARRESTING OFFICER AGENCY: Franklin County SO  
LAST AGENCY INSPECTION: 04/15/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: KIMBRELL

DATE/TIME OF TEST: 04/17/2020 18:03:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:55 |
| Air Blank          | 0.000 | 17:55 |
| Control Test       | 0.079 | 17:55 |
| Air Blank          | 0.000 | 17:56 |
| Subject Sample # 1 | 0.415 | 17:57 |
| Breath Volume      | 1.175 | 17:57 |
| Air Blank          | 0.000 | 17:58 |
| Air Blank          | 0.000 | 17:59 |
| Subject Sample # 2 | 0.406 | 18:01 |
| Breath Volume      | 1.574 | 18:01 |
| Air Blank          | 0.000 | 18:02 |
| Control Test       | 0.076 | 18:02 |
| Air Blank          | 0.000 | 18:02 |
| Diagnostics Check  | OK    | 18:03 |

---

Data Download Date/Time: 05/21/2020 14:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUNDAY,JESSIE,M  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: FOOTE,B.  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 04/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: LOPEZ MAIRENA

DATE/TIME OF TEST: 04/22/2020 02:26:40

---

|                   |       |       |                  |
|-------------------|-------|-------|------------------|
| Diagnostics Check | OK    | 02:24 |                  |
| Air Blank         | 0.000 | 02:24 |                  |
| Control Test      | 0.081 | 02:24 |                  |
| Air Blank         | 0.000 | 02:25 |                  |
| Subject Sample    | # 1   | 02:26 | Sequence Aborted |
| Breath Volume     | 0.000 | 02:26 |                  |
| Air Blank         | 0.000 | 02:26 |                  |

---

Data Download Date/Time: 05/05/2020 10:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUNDAY,JESSIE,M  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: FOOTE,B.  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 04/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: LOPEZ MAIRENA

DATE/TIME OF TEST: 04/22/2020 02:35:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:28 |
| Air Blank          | 0.000 | 02:29 |
| Control Test       | 0.080 | 02:29 |
| Air Blank          | 0.000 | 02:29 |
| Subject Sample # 1 | 0.156 | 02:30 |
| Breath Volume      | 1.812 | 02:30 |
| Air Blank          | 0.000 | 02:31 |
| Air Blank          | 0.000 | 02:33 |
| Subject Sample # 2 | 0.158 | 02:34 |
| Breath Volume      | 1.222 | 02:34 |
| Air Blank          | 0.000 | 02:34 |
| Control Test       | 0.081 | 02:34 |
| Air Blank          | 0.000 | 02:35 |
| Diagnostics Check  | OK    | 02:35 |

---

Data Download Date/Time: 05/05/2020 10:41



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUNDAY, JESSIE, M  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: FOOTE, B.  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 04/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: LOPEZ MAIRENA

DATE/TIME OF TEST: 04/22/2020 02:48:56

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 02:42 |
| Air Blank         |     | 0.000 | 02:43 |
| Control Test      |     | 0.079 | 02:43 |
| Air Blank         |     | 0.000 | 02:43 |
| Subject Sample    | # 1 | 0.164 | 02:44 |
| Breath Volume     |     | 1.265 | 02:44 |
| Air Blank         |     | 0.000 | 02:45 |
| Air Blank         |     | 0.000 | 02:46 |
| Subject Sample    | # 2 | 0.165 | 02:47 |
| Breath Volume     |     | 1.511 | 02:47 |
| Air Blank         |     | 0.000 | 02:48 |
| Control Test      |     | 0.081 | 02:48 |
| Air Blank         |     | 0.000 | 02:48 |
| Diagnostics Check |     | OK    | 02:48 |

---

Data Download Date/Time: 05/05/2020 10:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: NORTHWAY,GEORGIA  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 04/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:58

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 04/25/2020 07:28:01

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 07:20 |
| Air Blank         |     | 0.000 | 07:21 |
| Control Test      |     | 0.081 | 07:21 |
| Air Blank         |     | 0.000 | 07:22 |
| Subject Sample    | # 1 | 0.157 | 07:22 |
| Breath Volume     |     | 1.554 | 07:22 |
| Air Blank         |     | 0.000 | 07:23 |
| Air Blank         |     | 0.000 | 07:25 |
| Subject Sample    | # 2 | 0.154 | 07:26 |
| Breath Volume     |     | 2.789 | 07:26 |
| Air Blank         |     | 0.000 | 07:27 |
| Control Test      |     | 0.082 | 07:27 |
| Air Blank         |     | 0.000 | 07:27 |
| Diagnostics Check |     | OK    | 07:28 |

---

Data Download Date/Time: 05/05/2020 10:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: NORTHWAY,GEORGIA  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 04/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: PERSONETT

DATE/TIME OF TEST: 04/29/2020 00:26:36

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:23 |                      |
| Air Blank         | 0.000 | 00:23 |                      |
| Control Test      | 0.081 | 00:24 |                      |
| Air Blank         | 0.000 | 00:24 |                      |
| Subject Sample    | # 1   | 00:25 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:25 |                      |
| Air Blank         | 0.000 | 00:25 |                      |
| Control Test      | 0.081 | 00:26 |                      |
| Air Blank         | 0.000 | 00:26 |                      |
| Diagnostics Check | OK    | 00:26 |                      |

---

Data Download Date/Time: 05/05/2020 10:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000960

INSTRUMENT REGISTERED TO: LIBERTY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MOSCHETTO, ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 03/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 04/05/2020 22:51:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:45 |
| Air Blank          | 0.000 | 22:45 |
| Control Test       | 0.079 | 22:45 |
| Air Blank          | 0.000 | 22:46 |
| Subject Sample # 1 | 0.071 | 22:47 |
| Breath Volume      | 3.875 | 22:47 |
| Air Blank          | 0.000 | 22:47 |
| Air Blank          | 0.000 | 22:49 |
| Subject Sample # 2 | 0.066 | 22:49 |
| Breath Volume      | 3.812 | 22:49 |
| Air Blank          | 0.000 | 22:50 |
| Control Test       | 0.079 | 22:50 |
| Air Blank          | 0.000 | 22:51 |
| Diagnostics Check  | OK    | 22:51 |

---

Data Download Date/Time: 04/13/2020 13:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLEN,NATHAN,R  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: OWENS,JONATHAN  
ARRESTING OFFICER AGENCY: Wakulla County SO  
LAST AGENCY INSPECTION: 04/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 04/20/2020 23:56:12

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 23:50 |                |
| Air Blank         | 0.000     | 23:50 |                |
| Control Test      | 0.080     | 23:51 |                |
| Air Blank         | 0.000     | 23:51 |                |
| Subject Sample    | # 1 0.172 | 23:54 | Volume Not Met |
| Breath Volume     | 0.042     | 23:54 |                |
| Air Blank         |           | 23:55 | Purge Fail     |
| Air Blank         | 0.000     | 23:56 |                |

---

Data Download Date/Time: 05/21/2020 10:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLEN,NATHAN,R  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: OWNES,JONATHAN  
ARRESTING OFFICER AGENCY: Wakulla County SO  
LAST AGENCY INSPECTION: 04/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 04/21/2020 00:06:30

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 23:58 |                      |
| Air Blank         | 0.000     | 23:58 |                      |
| Control Test      | 0.079     | 23:59 |                      |
| Air Blank         | 0.000     | 23:59 |                      |
| Subject Sample    | # 1 0.197 | 00:00 |                      |
| Breath Volume     | 1.480     | 00:00 |                      |
| Air Blank         | 0.000     | 00:01 |                      |
| Air Blank         | 0.000     | 00:03 |                      |
| Subject Sample    | # 2       | 00:05 | Subject Test Refused |
| Breath Volume     | 0.000     | 00:05 |                      |
| Air Blank         | 0.000     | 00:05 |                      |
| Control Test      | 0.079     | 00:05 |                      |
| Air Blank         | 0.000     | 00:06 |                      |
| Diagnostics Check | OK        | 00:06 |                      |

---

Data Download Date/Time: 05/21/2020 10:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TALLMAN,JOHN,R  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MIDDLEBROOKS,ALAN  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 04/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: GLENN

DATE/TIME OF TEST: 04/22/2020 17:52:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:46 |
| Air Blank          | 0.000 | 17:46 |
| Control Test       | 0.081 | 17:47 |
| Air Blank          | 0.000 | 17:47 |
| Subject Sample # 1 | 0.082 | 17:48 |
| Breath Volume      | 1.453 | 17:48 |
| Air Blank          | 0.000 | 17:49 |
| Air Blank          | 0.000 | 17:51 |
| Subject Sample # 2 | 0.084 | 17:51 |
| Breath Volume      | 2.472 | 17:51 |
| Air Blank          | 0.000 | 17:52 |
| Control Test       | 0.078 | 17:52 |
| Air Blank          | 0.000 | 17:52 |
| Diagnostics Check  | OK    | 17:52 |

---

Data Download Date/Time: 05/21/2020 10:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORALES,BOBBY,M  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: MORALES,BOBBY  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 04/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: BUI  
ARREST TIME: 17:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: PEARCE

DATE/TIME OF TEST: 04/25/2020 18:41:21

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 18:38 |                      |
| Air Blank         | 0.000 | 18:38 |                      |
| Control Test      | 0.080 | 18:39 |                      |
| Air Blank         | 0.000 | 18:39 |                      |
| Subject Sample    | # 1   | 18:39 | Subject Test Refused |
| Breath Volume     | 0.000 | 18:39 |                      |
| Air Blank         | 0.000 | 18:40 |                      |
| Control Test      | 0.081 | 18:40 |                      |
| Air Blank         | 0.000 | 18:41 |                      |
| Diagnostics Check | OK    | 18:41 |                      |

---

Data Download Date/Time: 05/21/2020 10:55



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MALLOW, CHARLES, S  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: MALLOW, CHARLES  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 04/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: BUI  
ARREST TIME: 20:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: PEPPER

DATE/TIME OF TEST: 04/25/2020 21:42:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:36 |
| Air Blank          | 0.000 | 21:36 |
| Control Test       | 0.081 | 21:36 |
| Air Blank          | 0.000 | 21:37 |
| Subject Sample # 1 | 0.117 | 21:37 |
| Breath Volume      | 2.164 | 21:37 |
| Air Blank          | 0.000 | 21:38 |
| Air Blank          | 0.000 | 21:40 |
| Subject Sample # 2 | 0.111 | 21:40 |
| Breath Volume      | 1.480 | 21:40 |
| Air Blank          | 0.000 | 21:41 |
| Control Test       | 0.079 | 21:41 |
| Air Blank          | 0.000 | 21:42 |
| Diagnostics Check  | OK    | 21:42 |

---

Data Download Date/Time: 05/21/2020 10:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WILBENS,MARCELLON  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: HENRI

DATE/TIME OF TEST: 04/04/2020 00:16:28

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 00:10 |
| Air Blank         |     | 0.000 | 00:10 |
| Control Test      |     | 0.082 | 00:11 |
| Air Blank         |     | 0.000 | 00:11 |
| Subject Sample    | # 1 | 0.181 | 00:12 |
| Breath Volume     |     | 2.960 | 00:12 |
| Air Blank         |     | 0.000 | 00:12 |
| Air Blank         |     | 0.000 | 00:14 |
| Subject Sample    | # 2 | 0.186 | 00:14 |
| Breath Volume     |     | 2.484 | 00:14 |
| Air Blank         |     | 0.000 | 00:15 |
| Control Test      |     | 0.081 | 00:15 |
| Air Blank         |     | 0.000 | 00:16 |
| Diagnostics Check |     | OK    | 00:16 |

---

Data Download Date/Time: 04/09/2020 15:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: PARRISH,KAYLA  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: LOUIMANN

DATE/TIME OF TEST: 04/04/2020 22:45:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:38 |
| Air Blank          | 0.000 | 22:39 |
| Control Test       | 0.081 | 22:39 |
| Air Blank          | 0.000 | 22:39 |
| Subject Sample # 1 | 0.089 | 22:41 |
| Breath Volume      | 2.015 | 22:41 |
| Air Blank          | 0.000 | 22:41 |
| Air Blank          | 0.000 | 22:43 |
| Subject Sample # 2 | 0.090 | 22:44 |
| Breath Volume      | 2.015 | 22:44 |
| Air Blank          | 0.000 | 22:44 |
| Control Test       | 0.081 | 22:44 |
| Air Blank          | 0.000 | 22:45 |
| Diagnostics Check  | OK    | 22:45 |

---

Data Download Date/Time: 04/09/2020 15:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: PENA,MARLON  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MATIASITZEP

DATE/TIME OF TEST: 04/08/2020 02:16:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:10 |
| Air Blank          | 0.000 | 02:10 |
| Control Test       | 0.082 | 02:10 |
| Air Blank          | 0.000 | 02:11 |
| Subject Sample # 1 | 0.153 | 02:12 |
| Breath Volume      | 2.855 | 02:12 |
| Air Blank          | 0.000 | 02:13 |
| Air Blank          | 0.000 | 02:14 |
| Subject Sample # 2 | 0.155 | 02:15 |
| Breath Volume      | 2.890 | 02:15 |
| Air Blank          | 0.000 | 02:15 |
| Control Test       | 0.080 | 02:16 |
| Air Blank          | 0.000 | 02:16 |
| Diagnostics Check  | OK    | 02:16 |

---

Data Download Date/Time: 04/09/2020 15:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: AZEVEDO, FABIO  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: MALLOY

DATE/TIME OF TEST: 04/08/2020 07:29:53

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 07:18 |                |
| Air Blank         |     | 0.000 | 07:18 |                |
| Control Test      |     | 0.081 | 07:18 |                |
| Air Blank         |     | 0.000 | 07:19 |                |
| Subject Sample    | # 1 | 0.199 | 07:22 | Volume Not Met |
| Breath Volume     |     | 0.539 | 07:22 |                |
| Air Blank         |     | 0.000 | 07:23 |                |
| Air Blank         |     | 0.000 | 07:24 |                |
| Subject Sample    | # 2 | 0.234 | 07:25 |                |
| Breath Volume     |     | 2.035 | 07:25 |                |
| Air Blank         |     | 0.000 | 07:26 |                |
| Air Blank         |     | 0.000 | 07:27 |                |
| Subject Sample    | # 3 | 0.231 | 07:28 |                |
| Breath Volume     |     | 1.757 | 07:28 |                |
| Air Blank         |     | 0.000 | 07:28 |                |
| Control Test      |     | 0.078 | 07:29 |                |
| Air Blank         |     | 0.000 | 07:29 |                |
| Diagnostics Check |     | OK    | 07:29 |                |

---

Data Download Date/Time: 04/09/2020 15:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: TRENDAFILOV, TREN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:53

SUBJECT'S LAST NAME: MULLEN

DATE/TIME OF TEST: 04/20/2020 14:30:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:24 |
| Air Blank          | 0.000 | 14:25 |
| Control Test       | 0.082 | 14:25 |
| Air Blank          | 0.000 | 14:26 |
| Subject Sample # 1 | 0.000 | 14:26 |
| Breath Volume      | 4.042 | 14:26 |
| Air Blank          | 0.000 | 14:27 |
| Air Blank          | 0.000 | 14:29 |
| Subject Sample # 2 | 0.000 | 14:29 |
| Breath Volume      | 3.207 | 14:29 |
| Air Blank          | 0.000 | 14:30 |
| Control Test       | 0.081 | 14:30 |
| Air Blank          | 0.000 | 14:30 |
| Diagnostics Check  | OK    | 14:30 |

---

Data Download Date/Time: 04/23/2020 10:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: PRICE,FORREST  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: LORDMENDEZ

DATE/TIME OF TEST: 04/25/2020 01:10:44

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 01:03 |
| Air Blank         |     | 0.000 | 01:03 |
| Control Test      |     | 0.082 | 01:03 |
| Air Blank         |     | 0.000 | 01:04 |
| Subject Sample    | # 1 | 0.171 | 01:04 |
| Breath Volume     |     | 1.492 | 01:04 |
| Air Blank         |     | 0.000 | 01:05 |
| Air Blank         |     | 0.000 | 01:07 |
| Subject Sample    | # 2 | 0.189 | 01:09 |
| Breath Volume     |     | 3.890 | 01:09 |
| Air Blank         |     | 0.000 | 01:09 |
| Control Test      |     | 0.081 | 01:10 |
| Air Blank         |     | 0.000 | 01:10 |
| Diagnostics Check |     | OK    | 01:10 |

---

Data Download Date/Time: 04/30/2020 16:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: VENTURA,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: GOMEZRESTREPO

DATE/TIME OF TEST: 04/27/2020 06:03:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:55 |
| Air Blank          | 0.000 | 05:56 |
| Control Test       | 0.083 | 05:56 |
| Air Blank          | 0.000 | 05:56 |
| Subject Sample # 1 | 0.210 | 05:57 |
| Breath Volume      | 1.898 | 05:57 |
| Air Blank          | 0.000 | 05:58 |
| Air Blank          | 0.000 | 05:59 |
| Subject Sample # 2 | 0.212 | 06:02 |
| Breath Volume      | 1.500 | 06:02 |
| Air Blank          | 0.000 | 06:02 |
| Control Test       | 0.080 | 06:03 |
| Air Blank          | 0.000 | 06:03 |
| Diagnostics Check  | OK    | 06:03 |

---

Data Download Date/Time: 04/30/2020 16:20



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ SUERO,JOSE,M  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: PEREZ SUERO,JOSE  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: BREault

DATE/TIME OF TEST: 04/06/2020 19:55:00

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 19:52 |            |
| Air Blank         | 0.000 | 19:53 |            |
| Control Test      | 0.080 | 19:53 |            |
| Air Blank         | 0.000 | 19:53 |            |
| Subject Sample    | # 1   | 19:54 | RFI Detect |
| Breath Volume     | 0.000 | 19:54 |            |
| Air Blank         |       | 19:54 | RFI Detect |
| Air Blank         | 0.000 | 19:55 |            |

---

Data Download Date/Time: 04/21/2020 12:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ SUERO,JOSE,M  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: PEREZ SUERO,JOSE  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: BREault

DATE/TIME OF TEST: 04/06/2020 20:04:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:58 |
| Air Blank          | 0.000 | 19:58 |
| Control Test       | 0.080 | 19:59 |
| Air Blank          | 0.000 | 19:59 |
| Subject Sample # 1 | 0.109 | 20:00 |
| Breath Volume      | 2.085 | 20:00 |
| Air Blank          | 0.000 | 20:00 |
| Air Blank          | 0.000 | 20:02 |
| Subject Sample # 2 | 0.108 | 20:03 |
| Breath Volume      | 2.152 | 20:03 |
| Air Blank          | 0.000 | 20:03 |
| Control Test       | 0.080 | 20:04 |
| Air Blank          | 0.000 | 20:04 |
| Diagnostics Check  | OK    | 20:04 |

---

Data Download Date/Time: 04/21/2020 12:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORALES, RONNIE,  
BREATH TEST OPERATOR AGENCY: St. Cloud PD  
ARRESTING OFFICER: MORALES, RONNIE  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 03/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: SPIGLE

DATE/TIME OF TEST: 04/08/2020 01:20:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:13 |
| Air Blank          | 0.000 | 01:14 |
| Control Test       | 0.080 | 01:14 |
| Air Blank          | 0.000 | 01:15 |
| Subject Sample # 1 | 0.000 | 01:15 |
| Breath Volume      | 3.328 | 01:15 |
| Air Blank          | 0.000 | 01:15 |
| Air Blank          | 0.000 | 01:17 |
| Subject Sample # 2 | 0.000 | 01:18 |
| Breath Volume      | 2.917 | 01:18 |
| Air Blank          | 0.000 | 01:19 |
| Control Test       | 0.080 | 01:19 |
| Air Blank          | 0.000 | 01:20 |
| Diagnostics Check  | OK    | 01:20 |

---

Data Download Date/Time: 04/21/2020 12:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOPEZ GOLZALES,BAUDILIO,  
BREATH TEST OPERATOR AGENCY: FHP Troop K  
ARRESTING OFFICER: ESCOBAR,RIGOBERTO  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 03/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: LOPEZ GONZELES

DATE/TIME OF TEST: 04/13/2020 20:35:33

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:32 |                      |
| Air Blank         | 0.000 | 20:33 |                      |
| Control Test      | 0.080 | 20:33 |                      |
| Air Blank         | 0.000 | 20:33 |                      |
| Subject Sample    | # 1   | 20:34 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:34 |                      |
| Air Blank         | 0.000 | 20:34 |                      |
| Control Test      | 0.080 | 20:34 |                      |
| Air Blank         | 0.000 | 20:35 |                      |
| Diagnostics Check | OK    | 20:35 |                      |

---

Data Download Date/Time: 04/21/2020 12:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHOBREY,SPENCER,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: OLKA,MATTHEW  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 03/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: CULLER

DATE/TIME OF TEST: 04/16/2020 00:02:47

---

|                   |       |       |                  |
|-------------------|-------|-------|------------------|
| Diagnostics Check | OK    | 00:00 |                  |
| Air Blank         | 0.000 | 00:00 |                  |
| Control Test      | 0.080 | 00:01 |                  |
| Air Blank         | 0.000 | 00:01 |                  |
| Subject Sample    | # 1   | 00:02 | Sequence Aborted |
| Breath Volume     | 0.000 | 00:02 |                  |
| Air Blank         | 0.000 | 00:02 |                  |

---

Data Download Date/Time: 04/21/2020 12:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHOBREY,SPENCER,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: OLKA,MATTHEW  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 03/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: CULLER

DATE/TIME OF TEST: 04/16/2020 00:20:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:13 |
| Air Blank          | 0.000 | 00:14 |
| Control Test       | 0.081 | 00:14 |
| Air Blank          | 0.000 | 00:15 |
| Subject Sample # 1 | 0.192 | 00:16 |
| Breath Volume      | 1.414 | 00:16 |
| Air Blank          | 0.000 | 00:16 |
| Air Blank          | 0.000 | 00:18 |
| Subject Sample # 2 | 0.202 | 00:19 |
| Breath Volume      | 1.265 | 00:19 |
| Air Blank          | 0.000 | 00:19 |
| Control Test       | 0.080 | 00:20 |
| Air Blank          | 0.000 | 00:20 |
| Diagnostics Check  | OK    | 00:20 |

---

Data Download Date/Time: 04/21/2020 12:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYERS,MICHELLE,H  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: SAYERS,MICHELLE  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 03/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 04/17/2020 19:38:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:32 |
| Air Blank          | 0.000 | 19:32 |
| Control Test       | 0.080 | 19:33 |
| Air Blank          | 0.000 | 19:33 |
| Subject Sample # 1 | 0.000 | 19:34 |
| Breath Volume      | 2.453 | 19:34 |
| Air Blank          | 0.000 | 19:34 |
| Air Blank          | 0.000 | 19:36 |
| Subject Sample # 2 | 0.000 | 19:37 |
| Breath Volume      | 2.480 | 19:37 |
| Air Blank          | 0.000 | 19:37 |
| Control Test       | 0.079 | 19:38 |
| Air Blank          | 0.000 | 19:38 |
| Diagnostics Check  | OK    | 19:38 |

---

Data Download Date/Time: 04/21/2020 12:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAYERS,MICHELLE,H  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: SAYERS,MICHELLE  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 03/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: KLEIN

DATE/TIME OF TEST: 04/19/2020 06:13:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:06 |
| Air Blank          | 0.000 | 06:07 |
| Control Test       | 0.080 | 06:07 |
| Air Blank          | 0.000 | 06:08 |
| Subject Sample # 1 | 0.000 | 06:08 |
| Breath Volume      | 2.292 | 06:08 |
| Air Blank          | 0.000 | 06:09 |
| Air Blank          | 0.000 | 06:11 |
| Subject Sample # 2 | 0.000 | 06:11 |
| Breath Volume      | 2.617 | 06:11 |
| Air Blank          | 0.000 | 06:12 |
| Control Test       | 0.079 | 06:12 |
| Air Blank          | 0.000 | 06:13 |
| Diagnostics Check  | OK    | 06:13 |

---

Data Download Date/Time: 04/21/2020 12:14



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORALES, RONNIE,  
BREATH TEST OPERATOR AGENCY: St. Cloud PD  
ARRESTING OFFICER: MORALES, RONNIE  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 03/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 04/20/2020 20:53:49

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:51 |                      |
| Air Blank         | 0.000 | 20:51 |                      |
| Control Test      | 0.080 | 20:51 |                      |
| Air Blank         | 0.000 | 20:52 |                      |
| Subject Sample    | # 1   | 20:52 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:52 |                      |
| Air Blank         | 0.000 | 20:52 |                      |
| Control Test      | 0.079 | 20:53 |                      |
| Air Blank         | 0.000 | 20:53 |                      |
| Diagnostics Check | OK    | 20:53 |                      |

---

Data Download Date/Time: 04/21/2020 12:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHOBREY,SPENCER,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: SOLIS,GUSTAVO  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 04/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: AYERS

DATE/TIME OF TEST: 04/21/2020 22:09:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:02 |
| Air Blank          | 0.000 | 22:02 |
| Control Test       | 0.080 | 22:02 |
| Air Blank          | 0.000 | 22:03 |
| Subject Sample # 1 | 0.307 | 22:04 |
| Breath Volume      | 3.207 | 22:04 |
| Air Blank          | 0.000 | 22:05 |
| Air Blank          | 0.000 | 22:06 |
| Subject Sample # 2 | 0.294 | 22:07 |
| Breath Volume      | 2.226 | 22:07 |
| Air Blank          | 0.000 | 22:08 |
| Control Test       | 0.079 | 22:08 |
| Air Blank          | 0.000 | 22:09 |
| Diagnostics Check  | OK    | 22:09 |

---

Data Download Date/Time: 05/01/2020 10:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAWFORD,DAVID,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: CRAWFORD,DAVID  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 04/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 04/25/2020 02:32:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:25 |
| Air Blank          | 0.000 | 02:26 |
| Control Test       | 0.079 | 02:26 |
| Air Blank          | 0.000 | 02:26 |
| Subject Sample # 1 | 0.116 | 02:27 |
| Breath Volume      | 1.964 | 02:27 |
| Air Blank          | 0.000 | 02:28 |
| Air Blank          | 0.000 | 02:30 |
| Subject Sample # 2 | 0.115 | 02:31 |
| Breath Volume      | 2.140 | 02:31 |
| Air Blank          | 0.000 | 02:31 |
| Control Test       | 0.080 | 02:32 |
| Air Blank          | 0.000 | 02:32 |
| Diagnostics Check  | OK    | 02:32 |

---

Data Download Date/Time: 05/01/2020 10:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHOBREY,SPENCER,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: SOLIS,GUSTAVO  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 04/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: PENA LOPEZ

DATE/TIME OF TEST: 04/27/2020 01:45:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:39 |
| Air Blank          | 0.000 | 01:39 |
| Control Test       | 0.079 | 01:39 |
| Air Blank          | 0.000 | 01:40 |
| Subject Sample # 1 | 0.182 | 01:41 |
| Breath Volume      | 3.019 | 01:41 |
| Air Blank          | 0.000 | 01:41 |
| Air Blank          | 0.000 | 01:43 |
| Subject Sample # 2 | 0.188 | 01:44 |
| Breath Volume      | 2.656 | 01:44 |
| Air Blank          | 0.000 | 01:45 |
| Control Test       | 0.080 | 01:45 |
| Air Blank          | 0.000 | 01:45 |
| Diagnostics Check  | OK    | 01:45 |

---

Data Download Date/Time: 05/01/2020 10:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHOBREY,SPENCER,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: WHOBREY,SPENCER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 04/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:08

SUBJECT'S LAST NAME: BOURN

DATE/TIME OF TEST: 04/30/2020 06:36:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:29 |
| Air Blank          | 0.000 | 06:29 |
| Control Test       | 0.080 | 06:29 |
| Air Blank          | 0.000 | 06:30 |
| Subject Sample # 1 | 0.175 | 06:31 |
| Breath Volume      | 1.296 | 06:31 |
| Air Blank          | 0.000 | 06:32 |
| Air Blank          | 0.000 | 06:34 |
| Subject Sample # 2 | 0.155 | 06:34 |
| Breath Volume      | 1.265 | 06:34 |
| Air Blank          | 0.000 | 06:35 |
| Control Test       | 0.079 | 06:35 |
| Air Blank          | 0.000 | 06:36 |
| Diagnostics Check  | OK    | 06:36 |

---

Data Download Date/Time: 05/01/2020 10:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,J,  
BREATH TEST OPERATOR AGENCY: Titusville PD  
ARRESTING OFFICER: DERISE,V  
ARRESTING OFFICER AGENCY: Titusville PD  
LAST AGENCY INSPECTION: 04/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: FUENTES OSORTO

DATE/TIME OF TEST: 04/13/2020 04:34:56

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 04:28 |
| Air Blank         |     | 0.000 | 04:28 |
| Control Test      |     | 0.081 | 04:28 |
| Air Blank         |     | 0.000 | 04:29 |
| Subject Sample    | # 1 | 0.124 | 04:30 |
| Breath Volume     |     | 1.117 | 04:30 |
| Air Blank         |     | 0.000 | 04:31 |
| Air Blank         |     | 0.000 | 04:32 |
| Subject Sample    | # 2 | 0.117 | 04:33 |
| Breath Volume     |     | 2.281 | 04:33 |
| Air Blank         |     | 0.000 | 04:34 |
| Control Test      |     | 0.080 | 04:34 |
| Air Blank         |     | 0.000 | 04:34 |
| Diagnostics Check |     | OK    | 04:34 |

---

Data Download Date/Time: 05/20/2020 19:31

**Florida Department of Law Enforcement**

**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,J,  
BREATH TEST OPERATOR AGENCY: Titusville PD  
ARRESTING OFFICER: ROSA,E  
ARRESTING OFFICER AGENCY: Titusville PD  
LAST AGENCY INSPECTION: 04/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: ROSA

DATE/TIME OF TEST: 04/24/2020 00:51:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:43 |
| Air Blank          | 0.000 | 00:44 |
| Control Test       | 0.081 | 00:44 |
| Air Blank          | 0.000 | 00:45 |
| Subject Sample # 1 | 0.102 | 00:46 |
| Breath Volume      | 1.238 | 00:46 |
| Air Blank          | 0.000 | 00:47 |
| Air Blank          | 0.000 | 00:48 |
| Subject Sample # 2 | 0.105 | 00:49 |
| Breath Volume      | 2.324 | 00:49 |
| Air Blank          | 0.000 | 00:50 |
| Control Test       | 0.079 | 00:50 |
| Air Blank          | 0.000 | 00:51 |
| Diagnostics Check  | OK    | 00:51 |

---

Data Download Date/Time: 05/20/2020 19:31