

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ECHEVARRIA,LISA,V  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: TAYLOR,T  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 04/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:36

SUBJECT'S LAST NAME: DRUMMOND

DATE/TIME OF TEST: 05/03/2020 15:06:12

---

Diagnostics Check	OK	14:59
Air Blank	0.000	15:00
Control Test	0.076	15:00
Air Blank	0.000	15:01
Subject Sample # 1	0.261	15:01
Breath Volume	1.679	15:01
Air Blank	0.000	15:02
Air Blank	0.000	15:04
Subject Sample # 2	0.261	15:04
Breath Volume	2.289	15:04
Air Blank	0.000	15:05
Control Test	0.076	15:05
Air Blank	0.000	15:06
Diagnostics Check	OK	15:06

---

Data Download Date/Time: 06/03/2020 23:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEVORAK,MATTHEW,J  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: HOLTZHOUSE,JOHN  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 04/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: BYMASTER

DATE/TIME OF TEST: 05/05/2020 22:38:51

---

Diagnostics Check	OK	22:33	
Air Blank	0.000	22:34	
Control Test	0.076	22:34	
Air Blank	0.000	22:35	
Subject Sample	# 1	22:37	Subject Test Refused
Breath Volume	0.000	22:37	
Air Blank	0.000	22:37	
Control Test	0.075	22:38	
Air Blank	0.000	22:38	
Diagnostics Check	OK	22:38	

---

Data Download Date/Time: 06/03/2020 23:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEVORAK,MATTHEW,J  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: HOLTZHOUSE,JOHN  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 04/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: BYMASTER

DATE/TIME OF TEST: 05/05/2020 22:44:57

---

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:42	
Control Test	0.076	22:42	
Air Blank	0.000	22:43	
Subject Sample	# 1	22:43	Subject Test Refused
Breath Volume	0.000	22:43	
Air Blank	0.000	22:44	
Control Test	0.076	22:44	
Air Blank	0.000	22:44	
Diagnostics Check	OK	22:44	

---

Data Download Date/Time: 06/03/2020 23:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEVORAK, MATTHEW, J  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: ROBINSON, KYLE  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 04/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: KIMPEL

DATE/TIME OF TEST: 05/06/2020 00:40:44

---

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:36	
Control Test	0.076	00:36	
Air Blank	0.000	00:37	
Subject Sample	# 1	00:39	Subject Test Refused
Breath Volume	0.214	00:39	
Air Blank	0.000	00:39	
Control Test	0.077	00:40	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:40	

---

Data Download Date/Time: 06/03/2020 23:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HELM,DEREK,W  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HELM,DEREK  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: FAMA

DATE/TIME OF TEST: 05/08/2020 02:32:03

---

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.076	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.147	02:27
Breath Volume	2.121	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.159	02:30
Breath Volume	1.359	02:30
Air Blank	0.000	02:31
Control Test	0.076	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:32

---

Data Download Date/Time: 06/03/2020 23:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVENPORT,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: DAVENPORT,JOSHUA  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: HARRINGTON

DATE/TIME OF TEST: 05/10/2020 00:37:44

---

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.076	00:32
Air Blank		0.000	00:32
Subject Sample	# 1	0.186	00:33
Breath Volume		1.722	00:33
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.190	00:36
Breath Volume		1.679	00:36
Air Blank		0.000	00:36
Control Test		0.077	00:37
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

---

Data Download Date/Time: 06/03/2020 23:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASON,CHRISTOPHER,N  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: HOWARD,H  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 04/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: HAMMACK

DATE/TIME OF TEST: 05/11/2020 21:29:45

---

Diagnostics Check	OK	21:26	
Air Blank	0.000	21:27	
Control Test	0.076	21:27	
Air Blank	0.000	21:27	
Subject Sample	# 1	21:28	Subject Test Refused
Breath Volume	0.000	21:28	
Air Blank	0.000	21:28	
Control Test	0.076	21:29	
Air Blank	0.000	21:29	
Diagnostics Check	OK	21:29	

---

Data Download Date/Time: 06/03/2020 23:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FLOYD,DANIEL,Z

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: BLOTZ,GEOFFREY

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 04/07/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13518080A5

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 18:08

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/13/2020 19:16:54

---

Diagnostics Check	OK	19:10
Air Blank	0.000	19:10
Control Test	0.076	19:11
Air Blank	0.000	19:11
Subject Sample # 1	0.000	19:12
Breath Volume	2.117	19:12
Air Blank	0.000	19:12
Air Blank	0.000	19:14
Subject Sample # 2	0.000	19:15
Breath Volume	2.050	19:15
Air Blank	0.000	19:16
Control Test	0.076	19:16
Air Blank	0.000	19:16
Diagnostics Check	OK	19:16

---

Data Download Date/Time: 06/03/2020 23:42



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ECHEVARRIA,LISA,V  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: HOWARD,R  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 04/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 12:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:46

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 05/16/2020 13:12:42

---

Diagnostics Check	OK	13:06
Air Blank	0.000	13:06
Control Test	0.076	13:07
Air Blank	0.000	13:07
Subject Sample # 1	0.066	13:08
Breath Volume	2.343	13:08
Air Blank	0.000	13:08
Air Blank	0.000	13:10
Subject Sample # 2	0.061	13:11
Breath Volume	1.187	13:11
Air Blank	0.000	13:11
Control Test	0.076	13:12
Air Blank	0.000	13:12
Diagnostics Check	OK	13:12

---

Data Download Date/Time: 06/03/2020 23:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ECHEVARRIA,LISA,V  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: TAYLOR,T  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 04/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 05/16/2020 18:02:25

---

Diagnostics Check	OK	17:59	
Air Blank	0.000	17:59	
Control Test	0.075	17:59	
Air Blank	0.000	18:00	
Subject Sample	# 1	18:01	Subject Test Refused
Breath Volume	0.000	18:01	
Air Blank	0.000	18:01	
Control Test	0.076	18:01	
Air Blank	0.000	18:02	
Diagnostics Check	OK	18:02	

---

Data Download Date/Time: 06/03/2020 23:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH,DEVIN,A

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: ANSTEAD,J

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 04/07/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13518080A5

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 20:10

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: HULLENDER

DATE/TIME OF TEST: 05/16/2020 22:10:09

---

Diagnostics Check	OK	22:03	
Air Blank	0.000	22:03	
Control Test	0.076	22:03	
Air Blank	0.000	22:04	
Subject Sample # 1	0.137	22:05	
Breath Volume	2.636	22:05	
Air Blank	0.000	22:06	
Air Blank	0.000	22:07	
Subject Sample # 2	0.133	22:08	
Breath Volume	1.863	22:08	
Air Blank	0.000	22:09	
Control Test	0.074	22:09	Control Outside Tolerance
Air Blank	0.000	22:10	

---

Data Download Date/Time: 06/03/2020 23:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ECHEVARRIA,LISA,V  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: HOWARD,R  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 04/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 15:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 05/21/2020 16:19:32

---

Diagnostics Check	OK	16:16	
Air Blank	0.000	16:17	
Control Test	0.076	16:17	
Air Blank	0.000	16:18	
Subject Sample	# 1	16:18	Subject Test Refused
Breath Volume	0.000	16:18	
Air Blank	0.000	16:18	
Control Test	0.076	16:18	
Air Blank	0.000	16:19	
Diagnostics Check	OK	16:19	

---

Data Download Date/Time: 06/03/2020 23:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASON,CHRISTOPHER,N  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: SUMPTER,S  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 04/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: PURINTON

DATE/TIME OF TEST: 05/22/2020 21:20:04

---

Diagnostics Check	OK	21:13
Air Blank	0.000	21:13
Control Test	0.076	21:14
Air Blank	0.000	21:14
Subject Sample # 1	0.022	21:15
Breath Volume	2.757	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:17
Subject Sample # 2	0.021	21:18
Breath Volume	2.039	21:18
Air Blank	0.000	21:19
Control Test	0.076	21:19
Air Blank	0.000	21:19
Diagnostics Check	OK	21:20

---

Data Download Date/Time: 06/03/2020 23:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVENPORT,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: DAVENPORT,JOSHUA  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: JACOBS J

DATE/TIME OF TEST: 05/23/2020 00:17:36

---

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.076	00:11
Air Blank	0.000	00:12
Subject Sample # 1	0.156	00:12
Breath Volume	1.492	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.149	00:16
Breath Volume	1.585	00:16
Air Blank	0.000	00:16
Control Test	0.077	00:17
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

---

Data Download Date/Time: 06/03/2020 23:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASON,CHRISTOPHER,N  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: THOMAS,A  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 04/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 05/25/2020 23:09:33

---

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.077	23:01
Air Blank	0.000	23:01
Subject Sample # 1	0.146	23:03
Breath Volume	2.230	23:03
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.134	23:07
Breath Volume	1.964	23:07
Air Blank	0.000	23:08
Control Test	0.078	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

---

Data Download Date/Time: 06/03/2020 23:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTIGLIONE,SALVATORE,M  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: CASTIGLIONE,SALVATORE  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: HINES

DATE/TIME OF TEST: 05/26/2020 22:26:28

---

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.077	22:21
Air Blank	0.000	22:21
Subject Sample # 1	0.213	22:22
Breath Volume	3.664	22:22
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.193	22:24
Breath Volume	2.937	22:24
Air Blank	0.000	22:25
Control Test	0.077	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

---

Data Download Date/Time: 06/03/2020 23:42



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTIGLIONE,SALVATORE,M  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: CASTIGLIONE',SALVATORE  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: COAKLEY

DATE/TIME OF TEST: 05/30/2020 21:07:50

---

Diagnostics Check	OK	21:00	
Air Blank	0.000	21:01	
Control Test	0.076	21:01	
Air Blank	0.000	21:01	
Subject Sample # 1	0.289	21:03	
Breath Volume	1.320	21:03	
Air Blank	0.000	21:04	
Air Blank	0.000	21:05	
Subject Sample # 2	0.272	21:06	
Breath Volume	1.933	21:06	
Air Blank	0.000	21:06	
Control Test	0.074	21:07	Control Outside Tolerance
Air Blank	0.000	21:07	

---

Data Download Date/Time: 06/03/2020 23:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURINGTON,PATRICK,A  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: BAKER,WILLIAM  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: FRANCWAY

DATE/TIME OF TEST: 05/14/2020 21:14:04

---

Diagnostics Check	OK	21:12	
Air Blank	0.000	21:13	
Control Test	0.000	21:13	Control Outside Tolerance
Air Blank	0.000	21:14	

---

Data Download Date/Time: 05/27/2020 16:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURINGTON,PATRICK,A  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: BAKER,WILLIAM  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 08:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: FRANCWAY

DATE/TIME OF TEST: 05/14/2020 21:17:57

---

Diagnostics Check	OK	21:16	
Air Blank	0.000	21:17	
Control Test	0.000	21:17	Control Outside Tolerance
Air Blank	0.000	21:17	

---

Data Download Date/Time: 05/27/2020 16:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CABRAL,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: CABRAL,JONATHAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 05/28/2020 05:04:12

---

Diagnostics Check	OK	04:57
Air Blank	0.000	04:58
Control Test	0.079	04:58
Air Blank	0.000	04:58
Subject Sample # 1	0.123	04:59
Breath Volume	1.703	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:01
Subject Sample # 2	0.116	05:02
Breath Volume	2.328	05:02
Air Blank	0.000	05:03
Control Test	0.080	05:03
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

---

Data Download Date/Time: 06/29/2020 10:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TONEY,CRAIG,P  
BREATH TEST OPERATOR AGENCY: Valparaiso PD  
ARRESTING OFFICER: HENNIS,TIMOTHY  
ARRESTING OFFICER AGENCY: Eglin AFB PD  
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 21:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/16/2020 21:48:55

---

Diagnostics Check	OK	21:42
Air Blank	0.000	21:42
Control Test	0.080	21:43
Air Blank	0.000	21:43
Subject Sample # 1	0.000	21:44
Breath Volume	2.246	21:44
Air Blank	0.000	21:44
Air Blank	0.000	21:46
Subject Sample # 2	0.000	21:47
Breath Volume	1.714	21:47
Air Blank	0.000	21:48
Control Test	0.079	21:48
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

---

Data Download Date/Time: 05/19/2020 14:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TONEY,CRAIG,P  
BREATH TEST OPERATOR AGENCY: Valparaiso PD  
ARRESTING OFFICER: HENNIS,TIMOTHY  
ARRESTING OFFICER AGENCY: Eglin AFB PD  
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 21:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: HENRIQUEZ PENA

DATE/TIME OF TEST: 05/16/2020 22:00:19

---

Diagnostics Check	OK	21:52
Air Blank	0.000	21:53
Control Test	0.080	21:53
Air Blank	0.000	21:54
Subject Sample # 1	0.028	21:55
Breath Volume	1.175	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.024	21:58
Breath Volume	1.582	21:58
Air Blank	0.000	21:59
Control Test	0.078	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

---

Data Download Date/Time: 05/19/2020 14:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAMIRES, DAVID,  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: LEITE, STEVEN  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 792976  
DRY GAS EXPIRY DATE: 06/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: GONZALEZ CANO

DATE/TIME OF TEST: 05/01/2020 02:35:09

---

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.082	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.180	02:30
Breath Volume	2.339	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.173	02:33
Breath Volume	2.546	02:33
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

---

Data Download Date/Time: 05/18/2020 10:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARDNER,JEFF,A  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: GARDNER,JEFF  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 792976  
DRY GAS EXPIRY DATE: 06/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: VILAR

DATE/TIME OF TEST: 05/05/2020 23:47:10

---

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.081	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.119	23:42
Breath Volume	4.855	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.107	23:45
Breath Volume	3.972	23:45
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

---

Data Download Date/Time: 05/18/2020 10:02



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORELLANA, ADALBERTO, L  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: CLARK, JOEL  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 792976  
DRY GAS EXPIRY DATE: 06/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 05:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/20/2020 06:50:38

---

Diagnostics Check	OK	06:43
Air Blank	0.000	06:43
Control Test	0.082	06:44
Air Blank	0.000	06:44
Subject Sample # 1	0.178	06:45
Breath Volume	2.871	06:45
Air Blank	0.000	06:46
Air Blank	0.000	06:48
Subject Sample # 2	0.178	06:48
Breath Volume	3.437	06:48
Air Blank	0.000	06:49
Control Test	0.081	06:50
Air Blank	0.000	06:50
Diagnostics Check	OK	06:50

---

Data Download Date/Time: 06/19/2020 16:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VAN PELT, COLTEN, C  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: VAN PELT, COLTEN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 792976  
DRY GAS EXPIRY DATE: 06/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 05/22/2020 19:44:11

---

Diagnostics Check	OK	19:37
Air Blank	0.000	19:38
Control Test	0.082	19:38
Air Blank	0.000	19:39
Subject Sample # 1	0.000	19:39
Breath Volume	1.890	19:39
Air Blank	0.000	19:40
Air Blank	0.000	19:42
Subject Sample # 2	0.000	19:42
Breath Volume	1.371	19:42
Air Blank	0.000	19:43
Control Test	0.081	19:43
Air Blank	0.000	19:44
Diagnostics Check	OK	19:44

---

Data Download Date/Time: 06/19/2020 16:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBINSON,ANDREW,J  
BREATH TEST OPERATOR AGENCY: OTHER Fed  
ARRESTING OFFICER: HUNTER,JOHN  
ARRESTING OFFICER AGENCY: OTHER Fed  
LAST AGENCY INSPECTION: 05/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 792976  
DRY GAS EXPIRY DATE: 06/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: BONE

DATE/TIME OF TEST: 05/23/2020 21:12:28

---

Diagnostics Check	OK	21:05
Air Blank	0.000	21:06
Control Test	0.081	21:06
Air Blank	0.000	21:07
Subject Sample # 1	0.102	21:07
Breath Volume	3.093	21:07
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample # 2	0.107	21:10
Breath Volume	4.203	21:10
Air Blank	0.000	21:11
Control Test	0.081	21:11
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

---

Data Download Date/Time: 06/19/2020 16:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAMBERT,BRIAN,K  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: LAMBERT,BRIAN  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 792976  
DRY GAS EXPIRY DATE: 06/06/2020  
VIOLATION CODE: BUI  
ARREST TIME: 19:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 05/24/2020 20:25:59

---

Diagnostics Check	OK	20:22	
Air Blank	0.000	20:23	
Control Test	0.081	20:23	
Air Blank	0.000	20:24	
Subject Sample	# 1	20:24	Subject Test Refused
Breath Volume	0.000	20:24	
Air Blank	0.000	20:25	
Control Test	0.081	20:25	
Air Blank	0.000	20:25	
Diagnostics Check	OK	20:25	

---

Data Download Date/Time: 06/19/2020 16:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANNING,CARL,T  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: MANNING,CARL  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 792976  
DRY GAS EXPIRY DATE: 06/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 10:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 05/28/2020 11:05:10

---

Diagnostics Check	OK	10:58
Air Blank	0.000	10:58
Control Test	0.082	10:59
Air Blank	0.000	10:59
Subject Sample # 1	0.000	11:00
Breath Volume	2.089	11:00
Air Blank	0.000	11:00
Air Blank	0.000	11:02
Subject Sample # 2	0.000	11:03
Breath Volume	1.941	11:03
Air Blank	0.000	11:04
Control Test	0.080	11:04
Air Blank	0.000	11:05
Diagnostics Check	OK	11:05

---

Data Download Date/Time: 06/19/2020 16:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARCUM,JOSEPH,K  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: MARCUM,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 792976  
DRY GAS EXPIRY DATE: 06/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 12:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:41

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 05/29/2020 13:08:08

---

Diagnostics Check	OK	13:02
Air Blank	0.000	13:02
Control Test	0.082	13:02
Air Blank	0.000	13:03
Subject Sample # 1	0.000	13:03
Breath Volume	2.332	13:03
Air Blank	0.000	13:04
Air Blank	0.000	13:06
Subject Sample # 2	0.000	13:06
Breath Volume	2.992	13:06
Air Blank	0.000	13:07
Control Test	0.081	13:07
Air Blank	0.000	13:08
Diagnostics Check	OK	13:08

---

Data Download Date/Time: 06/19/2020 16:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORELLANA, ADALBERTO, L  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: ORELLANA, ADALBERTO  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1045940  
DRY GAS EXPIRY DATE: 11/13/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 05/03/2020 16:47:24

---

Diagnostics Check	OK	16:40
Air Blank	0.000	16:40
Control Test	0.084	16:41
Air Blank	0.000	16:41
Subject Sample # 1	0.000	16:43
Breath Volume	3.867	16:43
Air Blank	0.000	16:43
Air Blank	0.000	16:45
Subject Sample # 2	0.000	16:46
Breath Volume	3.867	16:46
Air Blank	0.000	16:46
Control Test	0.083	16:46
Air Blank	0.000	16:47
Diagnostics Check	OK	16:47

---

Data Download Date/Time: 05/18/2020 10:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNETT,JAMES,C  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: LAND,MATTHEW  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1045940  
DRY GAS EXPIRY DATE: 11/13/2021  
VIOLATION CODE: BUI  
ARREST TIME: 18:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 05/05/2020 19:33:55

---

Diagnostics Check	OK	19:27
Air Blank	0.000	19:27
Control Test	0.081	19:28
Air Blank	0.000	19:28
Subject Sample # 1	0.159	19:29
Breath Volume	3.785	19:29
Air Blank	0.000	19:29
Air Blank	0.000	19:31
Subject Sample # 2	0.151	19:32
Breath Volume	2.191	19:32
Air Blank	0.000	19:33
Control Test	0.082	19:33
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

---

Data Download Date/Time: 05/18/2020 10:38



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAVES,TIMOTHY,F  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: MARQUIS,NORMAND  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1045940  
DRY GAS EXPIRY DATE: 11/13/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 05/08/2020 20:04:54

---

Diagnostics Check	OK	19:58
Air Blank	0.000	19:59
Control Test	0.082	19:59
Air Blank	0.000	19:59
Subject Sample # 1	0.152	20:00
Breath Volume	3.867	20:00
Air Blank	0.000	20:01
Air Blank	0.000	20:02
Subject Sample # 2	0.158	20:03
Breath Volume	3.820	20:03
Air Blank	0.000	20:04
Control Test	0.082	20:04
Air Blank	0.000	20:04
Diagnostics Check	OK	20:04

---

Data Download Date/Time: 05/18/2020 10:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARCUM,JOSEPH,K  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: MARCUM,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1045940  
DRY GAS EXPIRY DATE: 11/13/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: EDDINS

DATE/TIME OF TEST: 05/09/2020 19:28:48

---

Diagnostics Check	OK	19:24	
Air Blank	0.000	19:25	
Control Test	0.083	19:25	
Air Blank	0.000	19:26	
Subject Sample	# 1	19:27	Subject Test Refused
Breath Volume	0.000	19:27	
Air Blank	0.000	19:27	
Control Test	0.082	19:28	
Air Blank	0.000	19:28	
Diagnostics Check	OK	19:28	

---

Data Download Date/Time: 05/18/2020 10:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYNCH,CHAD,J  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: LYNCH,CHAD  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1045940  
DRY GAS EXPIRY DATE: 11/13/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: WEEKS

DATE/TIME OF TEST: 05/15/2020 16:21:35

---

Diagnostics Check	OK	16:15
Air Blank	0.000	16:15
Control Test	0.084	16:16
Air Blank	0.000	16:16
Subject Sample # 1	0.098	16:17
Breath Volume	1.652	16:17
Air Blank	0.000	16:17
Air Blank	0.000	16:19
Subject Sample # 2	0.098	16:20
Breath Volume	1.527	16:20
Air Blank	0.000	16:20
Control Test	0.083	16:20
Air Blank	0.000	16:21
Diagnostics Check	OK	16:21

---

Data Download Date/Time: 05/18/2020 10:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COX,DAVID,S  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: COX,DAVID  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1076482  
DRY GAS EXPIRY DATE: 01/15/2022  
VIOLATION CODE: DUI  
ARREST TIME: 13:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:08

SUBJECT'S LAST NAME: FANT

DATE/TIME OF TEST: 05/21/2020 14:41:22

---

Diagnostics Check	OK	14:34
Air Blank	0.000	14:35
Control Test	0.081	14:35
Air Blank	0.000	14:36
Subject Sample # 1	0.364	14:36
Breath Volume	2.398	14:36
Air Blank	0.000	14:37
Air Blank	0.000	14:39
Subject Sample # 2	0.370	14:39
Breath Volume	1.984	14:39
Air Blank	0.000	14:40
Control Test	0.079	14:40
Air Blank	0.000	14:41
Diagnostics Check	OK	14:41

---

Data Download Date/Time: 06/01/2020 16:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALVAREZ,ROBERTO,A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: ALVAREZ,ROBERTO  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 05/01/2020 22:02:45

---

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.079	21:57
Air Blank	0.000	21:57
Subject Sample # 1	0.175	21:58
Breath Volume	1.320	21:58
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.188	22:01
Breath Volume	1.820	22:01
Air Blank	0.000	22:01
Control Test	0.078	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

---

Data Download Date/Time: 05/28/2020 07:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BENNETT, JONATHON, K  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BENNETT, JONATHON  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:57

SUBJECT'S LAST NAME: HOLLANDER

DATE/TIME OF TEST: 05/09/2020 15:36:11

---

Diagnostics Check	OK	15:29
Air Blank	0.000	15:29
Control Test	0.079	15:30
Air Blank	0.000	15:30
Subject Sample # 1	0.000	15:31
Breath Volume	3.722	15:31
Air Blank	0.000	15:31
Air Blank	0.000	15:33
Subject Sample # 2	0.000	15:34
Breath Volume	3.093	15:34
Air Blank	0.000	15:35
Control Test	0.079	15:35
Air Blank	0.000	15:36
Diagnostics Check	OK	15:36

---

Data Download Date/Time: 05/28/2020 07:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURINGTON,PATRICK,A  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: BAKER,WILLIAM  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: FRANCWAY

DATE/TIME OF TEST: 05/14/2020 22:21:19

---

Diagnostics Check	OK	22:15
Air Blank	0.000	22:15
Control Test	0.079	22:16
Air Blank	0.000	22:16
Subject Sample # 1	0.145	22:16
Breath Volume	2.070	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.146	22:19
Breath Volume	2.097	22:19
Air Blank	0.000	22:20
Control Test	0.077	22:20
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

---

Data Download Date/Time: 05/28/2020 07:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUGHES,DAVID,D  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: HUGHES,DAVID  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: BAUER

DATE/TIME OF TEST: 05/17/2020 03:22:12

---

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.109	03:17
Breath Volume	1.808	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.121	03:20
Breath Volume	2.355	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

---

Data Download Date/Time: 05/28/2020 07:57



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURINGTON,PATRICK,A  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: PURINGTON,PATRICK  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: FARNSWORTH

DATE/TIME OF TEST: 05/22/2020 22:50:26

---

Diagnostics Check	OK	22:47	
Air Blank	0.000	22:47	
Control Test	0.078	22:48	
Air Blank	0.000	22:48	
Subject Sample	# 1	22:49	Subject Test Refused
Breath Volume	0.000	22:49	
Air Blank	0.000	22:49	
Control Test	0.079	22:49	
Air Blank	0.000	22:50	
Diagnostics Check	OK	22:50	

---

Data Download Date/Time: 05/28/2020 07:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DILLON,JUSTIN,P  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: MYERS,BENJAMIN  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 05/23/2020 02:13:19

---

Diagnostics Check	OK	02:07	
Air Blank	0.000	02:08	
Control Test	0.078	02:08	
Air Blank	0.000	02:09	
Subject Sample	# 1	02:11	Subject Test Refused
Breath Volume	0.000	02:11	
Air Blank	0.000	02:12	
Control Test	0.080	02:12	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

---

Data Download Date/Time: 05/28/2020 07:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MURPHY,JAVON,D  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MURPHY,JAVON  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/26/2020 02:31:37

---

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.079	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.150	02:27
Breath Volume	1.609	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.152	02:30
Breath Volume	1.410	02:30
Air Blank	0.000	02:30
Control Test	0.079	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

---

Data Download Date/Time: 05/28/2020 07:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GENRICH,KRISTEN,R  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: LANGER,MATTHEW  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: GASKIN

DATE/TIME OF TEST: 05/01/2020 22:33:34

---

Diagnostics Check	OK	22:27
Air Blank	0.000	22:27
Control Test	0.081	22:28
Air Blank	0.000	22:28
Subject Sample # 1	0.000	22:29
Breath Volume	2.386	22:29
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.000	22:32
Breath Volume	2.074	22:32
Air Blank	0.000	22:32
Control Test	0.079	22:32
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

---

Data Download Date/Time: 05/29/2020 13:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PALMATEER,TRACY,D  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: PALMATEER,TRACY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 11:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: CAMBELL

DATE/TIME OF TEST: 05/02/2020 23:44:01

---

Diagnostics Check	OK	23:37
Air Blank	0.000	23:38
Control Test	0.082	23:38
Air Blank	0.000	23:38
Subject Sample # 1	0.167	23:39
Breath Volume	2.753	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.174	23:42
Breath Volume	2.519	23:42
Air Blank	0.000	23:43
Control Test	0.081	23:43
Air Blank	0.000	23:43
Diagnostics Check	OK	23:44

---

Data Download Date/Time: 05/29/2020 13:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUS,CHRISTINA,J  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: DUS,CHRISTINA  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: KILCREASE

DATE/TIME OF TEST: 05/05/2020 02:47:35

---

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.081	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.168	02:43
Breath Volume	3.144	02:43
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.162	02:46
Breath Volume	3.128	02:46
Air Blank	0.000	02:46
Control Test	0.079	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

---

Data Download Date/Time: 05/29/2020 13:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIMBRIL,JOSHUA,B  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: KIMBRIL,JOSHUA  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 05/07/2020 23:08:15

---

Diagnostics Check	OK	23:01
Air Blank	0.000	23:02
Control Test	0.081	23:02
Air Blank	0.000	23:02
Subject Sample # 1	0.128	23:03
Breath Volume	3.343	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.120	23:06
Breath Volume	2.171	23:06
Air Blank	0.000	23:07
Control Test	0.080	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

---

Data Download Date/Time: 05/29/2020 13:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIMBRIL,JOSHUA,B  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: RURYK,ANN  
ARRESTING OFFICER AGENCY: Crestview PD  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 05/08/2020 21:23:13

---

Diagnostics Check	OK	21:17
Air Blank	0.000	21:17
Control Test	0.081	21:18
Air Blank	0.000	21:18
Subject Sample # 1	0.000	21:18
Breath Volume	2.859	21:18
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample # 2	0.000	21:21
Breath Volume	2.699	21:21
Air Blank	0.000	21:22
Control Test	0.079	21:22
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

---

Data Download Date/Time: 05/29/2020 13:34



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIMBRIL,JOSHUA,B  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: KIMBRIL,JOSHUA  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 05/12/2020 23:11:12

---

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.081	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.000	23:06
Breath Volume	3.003	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.000	23:09
Breath Volume	3.121	23:09
Air Blank	0.000	23:10
Control Test	0.081	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

---

Data Download Date/Time: 05/29/2020 13:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COOK,RUSSELL,L  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: COOK,RUSSELL  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: TROMBLEY

DATE/TIME OF TEST: 05/15/2020 21:43:11

---

Diagnostics Check	OK	21:40	
Air Blank	0.000	21:40	
Control Test	0.081	21:41	
Air Blank	0.000	21:41	
Subject Sample	# 1	21:41	Subject Test Refused
Breath Volume	0.000	21:41	
Air Blank	0.000	21:42	
Control Test	0.081	21:42	
Air Blank	0.000	21:43	
Diagnostics Check	OK	21:43	

---

Data Download Date/Time: 05/29/2020 13:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIMBRIL,JOSHUA,B  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: KIMBRIL,JOSHUA  
ARRESTING OFFICER AGENCY: Madison County SO  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: FEE

DATE/TIME OF TEST: 05/16/2020 03:06:47

---

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.081	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.000	03:02
Breath Volume	3.625	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.000	03:05
Breath Volume	2.949	03:05
Air Blank	0.000	03:05
Control Test	0.079	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

---

Data Download Date/Time: 05/29/2020 13:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAHR,DANIEL,E  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: HAHR,DANIEL  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: BUI  
ARREST TIME: 18:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 05/22/2020 20:02:38

---

Diagnostics Check	OK	19:59	
Air Blank	0.000	20:00	
Control Test	0.081	20:00	
Air Blank	0.000	20:01	
Subject Sample	# 1	20:01	Subject Test Refused
Breath Volume	0.000	20:01	
Air Blank	0.000	20:01	
Control Test	0.081	20:02	
Air Blank	0.000	20:02	
Diagnostics Check	OK	20:02	

---

Data Download Date/Time: 05/29/2020 13:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLE,DANDRA,M  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: HARVEY,DAVID  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 12:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 05/02/2020 01:20:59

---

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:11	
Control Test	0.079	01:11	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:15	No Sample Provided
Breath Volume	0.000	01:15	
Air Blank	0.000	01:16	
Air Blank	0.000	01:18	
Subject Sample	# 2	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:20	
Control Test	0.079	01:20	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

---

Data Download Date/Time: 05/03/2020 02:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARVEY,DAVID,A  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: HARVEY,DAVID  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 05/02/2020 23:30:57

---

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:25	
Control Test	0.079	23:26	
Air Blank	0.000	23:26	
Subject Sample	# 1 0.120	23:26	
Breath Volume	2.195	23:26	
Air Blank	0.000	23:27	
Air Blank	0.000	23:29	
Subject Sample	# 2 0.111	23:29	
Breath Volume	1.996	23:29	
Air Blank		23:30	RFI Detect
Air Blank	0.000	23:30	

---

Data Download Date/Time: 05/03/2020 02:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHELAN, MATTHEW, S  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: PHELAN, MATTHEW  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: TURNSPLENTY

DATE/TIME OF TEST: 05/03/2020 21:40:28

---

Diagnostics Check	OK	21:37	
Air Blank	0.000	21:37	
Control Test	0.078	21:38	
Air Blank	0.000	21:38	
Subject Sample	# 1	21:39	Subject Test Refused
Breath Volume	0.000	21:39	
Air Blank	0.000	21:39	
Control Test	0.079	21:39	
Air Blank	0.000	21:40	
Diagnostics Check	OK	21:40	

---

Data Download Date/Time: 05/29/2020 18:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLIGAN,CONNIE,A  
BREATH TEST OPERATOR AGENCY: Lake City PD  
ARRESTING OFFICER: POTEAT,RICHARD  
ARRESTING OFFICER AGENCY: Lake City PD  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: FULTON

DATE/TIME OF TEST: 05/04/2020 03:10:01

---

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.078	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.000	03:05
Breath Volume	1.957	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.000	03:08
Breath Volume	2.468	03:08
Air Blank	0.000	03:09
Control Test	0.076	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:10

---

Data Download Date/Time: 05/29/2020 18:44



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BALDREE,TASHAHNDA,J  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: COODY,STEPHEN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: KARGO

DATE/TIME OF TEST: 05/10/2020 20:10:07

---

Diagnostics Check	OK	20:00
Air Blank	0.000	20:00
Control Test	0.078	20:01
Air Blank	0.000	20:01
Subject Sample # 1	0.310	20:02
Breath Volume	3.441	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:04
Subject Sample # 2	0.287	20:05
Breath Volume	2.152	20:05
Air Blank	0.000	20:06
Air Blank	0.000	20:07
Subject Sample # 3	0.307	20:08
Breath Volume	4.203	20:08
Air Blank	0.000	20:09
Control Test	0.077	20:09
Air Blank	0.000	20:10
Diagnostics Check	OK	20:10

---

Data Download Date/Time: 05/29/2020 18:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CZABAN,NICHOLAS,W  
BREATH TEST OPERATOR AGENCY: FL Dept of AG  
ARRESTING OFFICER: SANDS,WILLIAM  
ARRESTING OFFICER AGENCY: FL Dept of AG  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BILINA

DATE/TIME OF TEST: 05/15/2020 03:32:36

---

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.078	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.000	03:28
Breath Volume	2.800	03:28
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.000	03:31
Breath Volume	2.449	03:31
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

---

Data Download Date/Time: 05/29/2020 18:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARVEY,DAVID,A  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: MONTEMURNO,MATTHEW  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: EDMUNDS

DATE/TIME OF TEST: 05/16/2020 01:34:09

---

Diagnostics Check	OK	01:27
Air Blank	0.000	01:28
Control Test	0.078	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.015	01:29
Breath Volume	3.355	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.015	01:32
Breath Volume	3.582	01:32
Air Blank	0.000	01:33
Control Test	0.077	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

---

Data Download Date/Time: 05/29/2020 18:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SISTRUNK,TRENT,D  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: SISTRUNK,TRENT  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: DOWDY

DATE/TIME OF TEST: 05/16/2020 04:59:32

---

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:58	
Control Test	0.073	04:58	Control Outside Tolerance
Air Blank	0.000	04:59	

---

Data Download Date/Time: 05/29/2020 18:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SISTRUK,TRENT,D  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: SISTRUNK,TRENT  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: DOWDY

DATE/TIME OF TEST: 05/16/2020 05:07:53

---

Diagnostics Check	OK	05:01	
Air Blank	0.000	05:02	
Control Test	0.077	05:02	
Air Blank	0.000	05:02	
Subject Sample # 1	0.226	05:03	
Breath Volume	1.746	05:03	
Air Blank	0.000	05:04	
Air Blank	0.000	05:05	
Subject Sample # 2	0.238	05:06	
Breath Volume	3.394	05:06	
Air Blank	0.000	05:06	
Control Test	0.074	05:07	Control Outside Tolerance
Air Blank	0.000	05:07	

---

Data Download Date/Time: 05/29/2020 18:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLIGAN,CONNIE,A  
BREATH TEST OPERATOR AGENCY: Lake City PD  
ARRESTING OFFICER: POTEAT,RICHARD  
ARRESTING OFFICER AGENCY: Lake City PD  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 05/16/2020 21:11:48

---

Diagnostics Check	OK	21:05
Air Blank	0.000	21:06
Control Test	0.078	21:06
Air Blank	0.000	21:06
Subject Sample # 1	0.190	21:07
Breath Volume	2.375	21:07
Air Blank	0.000	21:08
Air Blank	0.000	21:09
Subject Sample # 2	0.188	21:10
Breath Volume	2.640	21:10
Air Blank	0.000	21:10
Control Test	0.077	21:11
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

---

Data Download Date/Time: 05/29/2020 18:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAREY,DAVID,A  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: MONTEMURNO,MATTHEW  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: HENDON

DATE/TIME OF TEST: 05/17/2020 00:36:15

---

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.078	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.122	00:31
Breath Volume	1.953	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.108	00:34
Breath Volume	2.914	00:34
Air Blank	0.000	00:35
Control Test	0.076	00:35
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

---

Data Download Date/Time: 05/29/2020 18:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARVEY,DAVID,T  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: HARVEY,DAVID  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: WYCHE

DATE/TIME OF TEST: 05/17/2020 02:12:11

---

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:03	
Control Test	0.078	02:04	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:07	No Sample Provided
Breath Volume	0.000	02:07	
Air Blank	0.000	02:08	
Air Blank	0.000	02:10	
Subject Sample	# 2	02:10	Subject Test Refused
Breath Volume	0.000	02:10	
Air Blank	0.000	02:11	
Control Test	0.078	02:11	
Air Blank		02:11	RFI Detect
Air Blank	0.000	02:12	

---

Data Download Date/Time: 05/29/2020 18:45



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARVEY,DAVID,A  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: MONTEMURNO,MATTHEW  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 05/17/2020 05:05:19

---

Diagnostics Check	OK	04:56
Air Blank	0.000	04:56
Control Test	0.077	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.079	04:58
Breath Volume	1.691	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.077	05:03
Breath Volume	2.132	05:03
Air Blank	0.000	05:04
Control Test	0.075	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

---

Data Download Date/Time: 05/29/2020 18:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SULLIVAN,BRETT,L  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: MONTEMURNO,MATTHEW  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 05/23/2020 00:05:08

---

Diagnostics Check	OK	23:58
Air Blank	0.000	23:59
Control Test	0.076	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.000	00:00
Breath Volume	3.398	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.000	00:03
Breath Volume	3.421	00:03
Air Blank	0.000	00:04
Control Test	0.076	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

---

Data Download Date/Time: 05/29/2020 18:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SULLIVAN,BRETT,L  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: SULLIVAN,BRETT  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: BUCHANAN

DATE/TIME OF TEST: 05/24/2020 22:00:56

---

Diagnostics Check	OK	21:54
Air Blank	0.000	21:54
Control Test	0.079	21:55
Air Blank	0.000	21:55
Subject Sample # 1	0.151	21:56
Breath Volume	2.632	21:56
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.152	21:59
Breath Volume	2.941	21:59
Air Blank	0.000	22:00
Control Test	0.079	22:00
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

---

Data Download Date/Time: 05/29/2020 18:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARVEY,DAVID,A  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: HARVEY,DAVID  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 05/24/2020 23:42:00

---

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:38	
Control Test	0.079	23:38	
Air Blank	0.000	23:39	
Subject Sample # 1	0.044	23:40	
Breath Volume	1.445	23:40	
Air Blank		23:40	RFI Detect
Air Blank	0.000	23:41	
Control Test	0.077	23:41	
Air Blank	0.000	23:41	
Diagnostics Check	OK	23:42	

---

Data Download Date/Time: 05/29/2020 18:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SULLIVAN,BRETT,L  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: ELLIOTT,NIGEL  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 05/25/2020 02:13:48

---

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.077	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.089	02:09
Breath Volume	3.199	02:09
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.086	02:12
Breath Volume	3.250	02:12
Air Blank	0.000	02:12
Control Test	0.076	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

---

Data Download Date/Time: 05/29/2020 18:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,DANIEL,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MARTIN,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 09:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 05/26/2020 10:11:23

---

Diagnostics Check	OK	10:05
Air Blank	0.000	10:05
Control Test	0.078	10:05
Air Blank	0.000	10:06
Subject Sample # 1	0.000	10:07
Breath Volume	1.851	10:07
Air Blank	0.000	10:07
Air Blank	0.000	10:09
Subject Sample # 2	0.000	10:10
Breath Volume	1.246	10:10
Air Blank	0.000	10:10
Control Test	0.077	10:10
Air Blank	0.000	10:11
Diagnostics Check	OK	10:11

---

Data Download Date/Time: 05/29/2020 18:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARVEY,DAVID,A  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: HARVEY,DAVID  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: FORSYTHE

DATE/TIME OF TEST: 05/28/2020 18:29:44

---

Diagnostics Check	OK	18:23
Air Blank	0.000	18:23
Control Test	0.078	18:24
Air Blank	0.000	18:24
Subject Sample # 1	0.053	18:25
Breath Volume	1.250	18:25
Air Blank	0.000	18:25
Air Blank	0.000	18:27
Subject Sample # 2	0.050	18:28
Breath Volume	1.976	18:28
Air Blank	0.000	18:28
Control Test	0.077	18:29
Air Blank	0.000	18:29
Diagnostics Check	OK	18:29

---

Data Download Date/Time: 05/29/2020 18:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASON,JAMES,R  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: LIPELIS,RUSLAN  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: SHERYL

DATE/TIME OF TEST: 05/31/2020 17:43:47

---

Diagnostics Check	OK	17:38	
Air Blank	0.000	17:39	
Control Test	0.078	17:39	
Air Blank	0.000	17:40	
Subject Sample	# 1 0.202	17:42	
Breath Volume	3.031	17:42	
Air Blank		17:43	Purge Fail
Air Blank	0.000	17:43	

---

Data Download Date/Time: 06/30/2020 10:01



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASON,JAMES,R  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: LIPELIS,RUSLAN  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: SHERYL

DATE/TIME OF TEST: 05/31/2020 18:07:37

---

Diagnostics Check	OK	17:56
Air Blank	0.000	17:56
Control Test	0.076	17:57
Air Blank	0.000	17:57
Subject Sample # 1	0.199	17:58
Breath Volume	2.503	17:58
Air Blank	0.000	17:59
Air Blank	0.000	18:00
Subject Sample # 2	0.178	18:02
Breath Volume	1.777	18:02
Air Blank	0.000	18:03
Air Blank	0.000	18:05
Subject Sample # 3	0.187	18:05
Breath Volume	2.214	18:05
Air Blank	0.000	18:06
Control Test	0.075	18:06
Air Blank	0.000	18:07
Diagnostics Check	OK	18:07

---

Data Download Date/Time: 06/30/2020 10:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARREN,BRIAN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: WARREN,BRIAN  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 05/22/2020 22:59:10

---

Diagnostics Check	OK	22:56	
Air Blank	0.000	22:56	
Control Test	0.080	22:57	
Air Blank	0.000	22:57	
Subject Sample	# 1	22:57	Subject Test Refused
Breath Volume	0.000	22:57	
Air Blank	0.000	22:58	
Control Test	0.078	22:58	
Air Blank	0.000	22:59	
Diagnostics Check	OK	22:59	

---

Data Download Date/Time: 06/09/2020 10:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOREY,FREDERICK,J  
BREATH TEST OPERATOR AGENCY: Madison PD  
ARRESTING OFFICER: SMITH,LUCAS  
ARRESTING OFFICER AGENCY: Madison PD  
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 05/24/2020 01:57:43

---

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.091	01:53
Breath Volume	2.632	01:53
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.084	01:56
Breath Volume	1.273	01:56
Air Blank	0.000	01:56
Control Test	0.077	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

---

Data Download Date/Time: 06/09/2020 10:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAURES,BRANDYN,K  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BAURES,BRANDYN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 05/16/2020 18:05:39

---

Diagnostics Check	OK	17:59
Air Blank	0.000	17:59
Control Test	0.083	18:00
Air Blank	0.000	18:00
Subject Sample # 1	0.212	18:01
Breath Volume	2.050	18:01
Air Blank	0.000	18:01
Air Blank	0.000	18:03
Subject Sample # 2	0.214	18:03
Breath Volume	1.441	18:03
Air Blank	0.000	18:04
Control Test	0.083	18:05
Air Blank	0.000	18:05
Diagnostics Check	OK	18:05

---

Data Download Date/Time: 05/17/2020 23:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MACFADDEN,RYAN,E  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MACFADDEN,RYAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: GULIERREZ

DATE/TIME OF TEST: 05/25/2020 04:49:33

---

Diagnostics Check	OK	04:42
Air Blank	0.000	04:43
Control Test	0.081	04:43
Air Blank	0.000	04:44
Subject Sample # 1	0.152	04:44
Breath Volume	2.050	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.152	04:47
Breath Volume	1.164	04:47
Air Blank	0.000	04:48
Control Test	0.081	04:48
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

---

Data Download Date/Time: 07/01/2020 23:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MACFADDEN,RYAN,E  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MACFADDEN,RYAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: FAHRNY

DATE/TIME OF TEST: 05/29/2020 02:56:13

---

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:53	
Control Test	0.082	02:54	
Air Blank	0.000	02:54	
Subject Sample	# 1	02:54	Subject Test Refused
Breath Volume	0.000	02:54	
Air Blank	0.000	02:55	
Control Test	0.082	02:55	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

---

Data Download Date/Time: 07/01/2020 23:51

**Florida Department of Law Enforcement**

**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARGE,SUSAN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: LINCOLN,TREY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: KOZMA

DATE/TIME OF TEST: 05/10/2020 13:55:16

---

Diagnostics Check		OK	13:42	
Air Blank		0.000	13:43	
Control Test		0.080	13:43	
Air Blank		0.000	13:44	
Subject Sample	# 1		13:47	No Sample Provided
Breath Volume		0.000	13:47	
Air Blank		0.000	13:47	
Air Blank		0.000	13:49	
Subject Sample	# 2	0.173	13:50	No .020 agreement
Breath Volume		3.285	13:50	
Air Blank		0.000	13:51	
Air Blank		0.000	13:53	
Subject Sample	# 3	0.148	13:53	No .020 agreement
Breath Volume		1.339	13:53	
Air Blank		0.000	13:54	
Control Test		0.079	13:54	
Air Blank		0.000	13:55	
Diagnostics Check		OK	13:55	

---

Data Download Date/Time: 06/01/2020 16:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARGE,SUSAN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: LINCOLN,TREY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: KOZMA

DATE/TIME OF TEST: 05/10/2020 14:04:15

---

Diagnostics Check	OK	13:57
Air Blank	0.000	13:57
Control Test	0.079	13:58
Air Blank	0.000	13:58
Subject Sample # 1	0.167	13:59
Breath Volume	2.023	13:59
Air Blank	0.000	14:00
Air Blank	0.000	14:02
Subject Sample # 2	0.153	14:02
Breath Volume	1.515	14:02
Air Blank	0.000	14:03
Control Test	0.078	14:03
Air Blank	0.000	14:04
Diagnostics Check	OK	14:04

---

Data Download Date/Time: 06/01/2020 16:27



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LITTLE,NICHLAS,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: LITTLE,NICHOLAS  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:26

SUBJECT'S LAST NAME: HAGEMAN

DATE/TIME OF TEST: 05/15/2020 16:54:11

---

Diagnostics Check	OK	16:47
Air Blank	0.000	16:47
Control Test	0.080	16:48
Air Blank	0.000	16:48
Subject Sample # 1	0.259	16:49
Breath Volume	1.925	16:49
Air Blank	0.000	16:50
Air Blank	0.000	16:52
Subject Sample # 2	0.279	16:52
Breath Volume	1.734	16:52
Air Blank	0.000	16:53
Control Test	0.077	16:53
Air Blank	0.000	16:54
Diagnostics Check	OK	16:54

---

Data Download Date/Time: 06/01/2020 16:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAWYER,DARRYL,B  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: LEWIS,TIMOTHY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 05/23/2020 06:24:31

---

Diagnostics Check	OK	06:17
Air Blank	0.000	06:18
Control Test	0.081	06:18
Air Blank	0.000	06:19
Subject Sample # 1	0.151	06:19
Breath Volume	1.554	06:19
Air Blank	0.000	06:20
Air Blank	0.000	06:22
Subject Sample # 2	0.144	06:22
Breath Volume	1.375	06:22
Air Blank	0.000	06:23
Control Test	0.081	06:23
Air Blank	0.000	06:24
Diagnostics Check	OK	06:24

---

Data Download Date/Time: 06/01/2020 16:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAWYER,DARRYL,B  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: HENSON,JAY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 05/23/2020 23:49:34

---

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.080	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.161	23:44
Breath Volume	1.257	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.164	23:48
Breath Volume	1.761	23:48
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

---

Data Download Date/Time: 06/01/2020 16:27



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP TROOP A  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUBEL,NEIL,F  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: RUBEL,NEIL  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 05/28/2020 06:16:24

---

Diagnostics Check	OK	06:10
Air Blank	0.000	06:10
Control Test	0.079	06:10
Air Blank	0.000	06:11
Subject Sample # 1	0.135	06:11
Breath Volume	3.609	06:11
Air Blank	0.000	06:12
Air Blank	0.000	06:14
Subject Sample # 2	0.125	06:14
Breath Volume	2.292	06:14
Air Blank	0.000	06:15
Control Test	0.079	06:15
Air Blank	0.000	06:16
Diagnostics Check	OK	06:16

---

Data Download Date/Time: 06/01/2020 15:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STROHECKER,KYNDAL,C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: STROHECKER,KYNDAL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: RANKIN

DATE/TIME OF TEST: 05/13/2020 22:01:55

---

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.081	21:55
Air Blank	0.000	21:56
Subject Sample # 1	0.242	21:57
Breath Volume	1.656	21:57
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.253	22:00
Breath Volume	1.445	22:00
Air Blank	0.000	22:00
Control Test	0.080	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

---

Data Download Date/Time: 06/01/2020 16:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COX,DAVID,S  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: COX,DAVID  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:07

SUBJECT'S LAST NAME: MOSLEY

DATE/TIME OF TEST: 05/20/2020 15:36:22

---

Diagnostics Check	OK	15:29
Air Blank	0.000	15:30
Control Test	0.081	15:30
Air Blank	0.000	15:31
Subject Sample # 1	0.137	15:31
Breath Volume	2.476	15:31
Air Blank	0.000	15:32
Air Blank	0.000	15:34
Subject Sample # 2	0.141	15:34
Breath Volume	1.945	15:34
Air Blank	0.000	15:35
Control Test	0.079	15:35
Air Blank	0.000	15:36
Diagnostics Check	OK	15:36

---

Data Download Date/Time: 06/01/2020 16:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BANKS,SEAN,C  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: BANKS,SEAN  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 05/05/2020 03:25:14

---

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.078	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.156	03:20
Breath Volume	3.753	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.159	03:23
Breath Volume	3.332	03:23
Air Blank	0.000	03:24
Control Test	0.078	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

---

Data Download Date/Time: 06/04/2020 11:04



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBINSON,ANDREW,J  
BREATH TEST OPERATOR AGENCY: OTHER Fed  
ARRESTING OFFICER: ROBINSON,ANDREW  
ARRESTING OFFICER AGENCY: OTHER Fed  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: STICKEN

DATE/TIME OF TEST: 05/06/2020 20:16:53

---

Diagnostics Check	OK	20:10
Air Blank	0.000	20:11
Control Test	0.077	20:11
Air Blank	0.000	20:12
Subject Sample # 1	0.120	20:12
Breath Volume	1.394	20:12
Air Blank	0.000	20:13
Air Blank	0.000	20:14
Subject Sample # 2	0.115	20:15
Breath Volume	3.332	20:15
Air Blank	0.000	20:15
Control Test	0.080	20:16
Air Blank	0.000	20:16
Diagnostics Check	OK	20:16

---

Data Download Date/Time: 06/04/2020 11:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BANKS,SEAN,C  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: BANKS,SEAN  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 05/10/2020 22:58:35

---

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.078	22:53
Air Blank	0.000	22:53
Subject Sample # 1	0.167	22:54
Breath Volume	3.167	22:54
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.171	22:57
Breath Volume	3.449	22:57
Air Blank	0.000	22:57
Control Test	0.079	22:58
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

---

Data Download Date/Time: 06/04/2020 11:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMS,KYLE,N  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: SIMMS,KYLE  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: CHAMBERS

DATE/TIME OF TEST: 05/11/2020 22:45:58

---

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.079	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.101	22:41
Breath Volume	4.015	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:43
Subject Sample # 2	0.098	22:44
Breath Volume	2.929	22:44
Air Blank	0.000	22:45
Control Test	0.079	22:45
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

---

Data Download Date/Time: 06/04/2020 11:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMS,KYLE,N  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: SIMMS,KYLE  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: CLAUDIO

DATE/TIME OF TEST: 05/17/2020 02:58:29

---

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.078	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.164	02:54
Breath Volume	2.003	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.166	02:56
Breath Volume	3.687	02:56
Air Blank	0.000	02:57
Control Test	0.079	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

---

Data Download Date/Time: 06/04/2020 11:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBINSON,ANDREW,J  
BREATH TEST OPERATOR AGENCY: OTHER Fed  
ARRESTING OFFICER: HUNTER,JOHN  
ARRESTING OFFICER AGENCY: OTHER Fed  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: GOBBLE

DATE/TIME OF TEST: 05/17/2020 17:46:57

---

Diagnostics Check	OK	17:40
Air Blank	0.000	17:41
Control Test	0.077	17:41
Air Blank	0.000	17:41
Subject Sample # 1	0.196	17:42
Breath Volume	1.953	17:42
Air Blank	0.000	17:43
Air Blank	0.000	17:44
Subject Sample # 2	0.182	17:45
Breath Volume	1.191	17:45
Air Blank	0.000	17:46
Control Test	0.077	17:46
Air Blank	0.000	17:46
Diagnostics Check	OK	17:46

---

Data Download Date/Time: 06/04/2020 11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: YOUNGREN,DENNIS  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: LYNN

DATE/TIME OF TEST: 05/31/2020 19:50:24

---

Diagnostics Check	OK	19:40
Air Blank	0.000	19:40
Control Test	0.078	19:41
Air Blank	0.000	19:41
Subject Sample # 1	0.198	19:42
Breath Volume	1.648	19:42
Air Blank	0.000	19:43
Air Blank	0.000	19:44
Subject Sample # 2	0.219	19:45
Breath Volume	1.226	19:45
Air Blank	0.000	19:46
Air Blank	0.000	19:48
Subject Sample # 3	0.207	19:48
Breath Volume	1.843	19:48
Air Blank	0.000	19:49
Control Test	0.078	19:49
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

---

Data Download Date/Time: 06/04/2020 11:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMB,HUNTER,  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LIMB,HUNTER  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/01/2020 22:14:33

---

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.080	22:09
Air Blank	0.000	22:09
Subject Sample # 1	0.160	22:09
Breath Volume	3.628	22:09
Air Blank	0.000	22:10
Air Blank	0.000	22:12
Subject Sample # 2	0.160	22:13
Breath Volume	2.679	22:13
Air Blank	0.000	22:13
Control Test	0.080	22:13
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

---

Data Download Date/Time: 05/21/2020 14:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 05/02/2020 19:38:02

---

Diagnostics Check	OK	19:25	
Air Blank	0.000	19:26	
Control Test	0.080	19:26	
Air Blank	0.000	19:26	
Subject Sample	# 1	19:29	No Sample Provided
Breath Volume	0.000	19:29	
Air Blank	0.000	19:30	
Air Blank	0.000	19:32	
Subject Sample	# 2	0.136	19:33
Breath Volume	2.406	19:33	
Air Blank	0.000	19:34	
Air Blank	0.000	19:36	
Subject Sample	# 3	0.132	19:36
Breath Volume	2.160	19:36	
Air Blank	0.000	19:37	
Control Test	0.079	19:37	
Air Blank	0.000	19:37	
Diagnostics Check	OK	19:38	

---

Data Download Date/Time: 05/21/2020 14:03



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMB,HUNTER,  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LIMB,HUNTER  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 05/03/2020 00:33:23

---

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.081	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.051	00:28
Breath Volume	1.812	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.052	00:31
Breath Volume	2.753	00:31
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

---

Data Download Date/Time: 05/21/2020 14:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: BENTON

DATE/TIME OF TEST: 05/06/2020 20:46:47

---

Diagnostics Check	OK	20:40
Air Blank	0.000	20:41
Control Test	0.081	20:41
Air Blank	0.000	20:41
Subject Sample # 1	0.206	20:42
Breath Volume	1.800	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:44
Subject Sample # 2	0.199	20:45
Breath Volume	1.175	20:45
Air Blank	0.000	20:45
Control Test	0.080	20:46
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

---

Data Download Date/Time: 05/21/2020 14:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUILES,ALEXANDER,  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: GELIN,KYLE  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: DI SALVO

DATE/TIME OF TEST: 05/06/2020 21:34:47

---

Diagnostics Check	OK	21:28
Air Blank	0.000	21:29
Control Test	0.079	21:29
Air Blank	0.000	21:29
Subject Sample # 1	0.205	21:30
Breath Volume	3.460	21:30
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample # 2	0.196	21:33
Breath Volume	2.703	21:33
Air Blank	0.000	21:33
Control Test	0.079	21:34
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

---

Data Download Date/Time: 05/21/2020 14:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 05/07/2020 22:33:13

---

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.080	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.173	22:28
Breath Volume	1.375	22:28
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.175	22:31
Breath Volume	1.250	22:31
Air Blank	0.000	22:32
Control Test	0.080	22:32
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

---

Data Download Date/Time: 05/21/2020 14:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMB,HUNTER,  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LIMB,HUNTER  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: WORTH

DATE/TIME OF TEST: 05/07/2020 23:15:30

---

Diagnostics Check	OK	23:09
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:10
Subject Sample # 1	0.115	23:10
Breath Volume	3.277	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.112	23:13
Breath Volume	3.250	23:13
Air Blank	0.000	23:14
Control Test	0.080	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

---

Data Download Date/Time: 05/21/2020 14:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRAWDY,LAINE,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: DRAWDY,LAINE  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: MALINCONICO

DATE/TIME OF TEST: 05/09/2020 00:49:43

---

Diagnostics Check	OK	00:42
Air Blank	0.000	00:43
Control Test	0.080	00:43
Air Blank	0.000	00:43
Subject Sample # 1	0.000	00:45
Breath Volume	1.253	00:45
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.000	00:48
Breath Volume	2.312	00:48
Air Blank	0.000	00:48
Control Test	0.080	00:49
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

---

Data Download Date/Time: 05/21/2020 14:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMB,HUNTER,  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LIMB,HUNTER  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: FRANCO

DATE/TIME OF TEST: 05/10/2020 03:09:23

---

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.080	03:03
Air Blank		0.000	03:04
Subject Sample	# 1	0.145	03:04
Breath Volume		2.003	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.145	03:07
Breath Volume		3.246	03:07
Air Blank		0.000	03:08
Control Test		0.079	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

---

Data Download Date/Time: 05/21/2020 14:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SHEELAR,JOHNATHAN  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: GRAVINA

DATE/TIME OF TEST: 05/11/2020 19:28:55

---

Diagnostics Check	OK	19:21
Air Blank	0.000	19:21
Control Test	0.080	19:21
Air Blank	0.000	19:22
Subject Sample # 1	0.103	19:24
Breath Volume	1.953	19:24
Air Blank	0.000	19:24
Air Blank	0.000	19:26
Subject Sample # 2	0.102	19:27
Breath Volume	1.597	19:27
Air Blank	0.000	19:28
Control Test	0.080	19:28
Air Blank	0.000	19:28
Diagnostics Check	OK	19:28

---

Data Download Date/Time: 05/21/2020 14:03



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: OSTEEN,CLAYTON  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: STOUDEMIRE

DATE/TIME OF TEST: 05/12/2020 00:34:16

---

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.080	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.000	00:30
Breath Volume	2.441	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.000	00:32
Breath Volume	2.468	00:32
Air Blank	0.000	00:33
Control Test	0.081	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

---

Data Download Date/Time: 05/21/2020 14:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: SEPEDA

DATE/TIME OF TEST: 05/12/2020 18:41:59

---

Diagnostics Check	OK	18:34
Air Blank	0.000	18:34
Control Test	0.080	18:34
Air Blank	0.000	18:35
Subject Sample # 1	0.214	18:36
Breath Volume	1.144	18:36
Air Blank	0.000	18:37
Air Blank	0.000	18:38
Subject Sample # 2	0.217	18:40
Breath Volume	1.195	18:40
Air Blank	0.000	18:41
Control Test	0.080	18:41
Air Blank	0.000	18:41
Diagnostics Check	OK	18:41

---

Data Download Date/Time: 05/21/2020 14:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRAWDY,LAINE,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: DRAWDY,LAINE  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: ORGAN

DATE/TIME OF TEST: 05/14/2020 20:16:24

---

Diagnostics Check	OK	20:15	
Air Blank	0.000	20:15	
		20:15	RFI Detect
Air Blank	0.000	20:16	

---

Data Download Date/Time: 05/21/2020 14:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRAWDY,LAINE,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: DRAWDY,LAINE  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: ORGAN

DATE/TIME OF TEST: 05/14/2020 20:24:01

---

Diagnostics Check	OK	20:17
Air Blank	0.000	20:18
Control Test	0.081	20:18
Air Blank	0.000	20:18
Subject Sample # 1	0.000	20:19
Breath Volume	3.039	20:19
Air Blank	0.000	20:20
Air Blank	0.000	20:22
Subject Sample # 2	0.000	20:22
Breath Volume	3.359	20:22
Air Blank	0.000	20:23
Control Test	0.080	20:23
Air Blank	0.000	20:23
Diagnostics Check	OK	20:24

---

Data Download Date/Time: 05/21/2020 14:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMB,HUNTER,  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LIMB,HUNTER  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 05/15/2020 01:53:40

---

Diagnostics Check	OK	01:46
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.163	01:49
Breath Volume	1.507	01:49
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.161	01:52
Breath Volume	1.976	01:52
Air Blank	0.000	01:52
Control Test	0.079	01:53
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

---

Data Download Date/Time: 05/21/2020 14:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: COLANGELO,CODY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 05/16/2020 01:26:02

---

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:23	
Control Test	0.080	01:23	
Air Blank	0.000	01:24	
Subject Sample	# 1	01:24	Subject Test Refused
Breath Volume	0.000	01:24	
Air Blank	0.000	01:25	
Control Test	0.079	01:25	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:26	

---

Data Download Date/Time: 05/21/2020 14:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: SALDANA PACHECO

DATE/TIME OF TEST: 05/17/2020 04:21:20

---

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.079	04:15
Air Blank	0.000	04:16
Subject Sample # 1	0.100	04:16
Breath Volume	1.585	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.095	04:19
Breath Volume	1.625	04:19
Air Blank	0.000	04:20
Control Test	0.079	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

---

Data Download Date/Time: 05/21/2020 14:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRAWDY,LAINE,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: DRAWDY,LAINE  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: HOYT

DATE/TIME OF TEST: 05/20/2020 04:43:48

---

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.080	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.246	04:39
Breath Volume	2.113	04:39
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.243	04:42
Breath Volume	2.246	04:42
Air Blank	0.000	04:42
Control Test	0.079	04:43
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

---

Data Download Date/Time: 05/21/2020 14:03



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: FRIEDMAN

DATE/TIME OF TEST: 05/20/2020 21:30:48

---

Diagnostics Check	OK	21:25
Air Blank	0.000	21:25
Control Test	0.080	21:25
Air Blank	0.000	21:26
Subject Sample # 1	0.000	21:26
Breath Volume	2.269	21:26
Air Blank	0.000	21:27
Air Blank	0.000	21:29
Subject Sample # 2	0.000	21:29
Breath Volume	2.722	21:29
Air Blank	0.000	21:29
Control Test	0.080	21:30
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

---

Data Download Date/Time: 05/21/2020 14:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMB,HUNTER,  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LIMB,HUNTER  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: DEL HERBE

DATE/TIME OF TEST: 05/20/2020 23:33:41

---

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.000	23:29
Breath Volume	3.472	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.000	23:32
Breath Volume	3.214	23:32
Air Blank	0.000	23:32
Control Test	0.079	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

---

Data Download Date/Time: 05/21/2020 14:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMB,HUNTER,  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LIMB,HUNTER  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 05/22/2020 21:56:50

---

Diagnostics Check	OK	21:49
Air Blank	0.000	21:50
Control Test	0.080	21:50
Air Blank	0.000	21:51
Subject Sample # 1	0.149	21:51
Breath Volume	2.031	21:51
Air Blank	0.000	21:52
Air Blank	0.000	21:53
Subject Sample # 2	0.149	21:55
Breath Volume	1.851	21:55
Air Blank	0.000	21:55
Control Test	0.080	21:56
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

---

Data Download Date/Time: 06/24/2020 15:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRAWDY,LAINE,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: DRAWDY,LAINE  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: HAESELIN

DATE/TIME OF TEST: 05/23/2020 00:17:07

---

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.080	00:11
Air Blank	0.000	00:12
Subject Sample # 1	0.000	00:12
Breath Volume	4.542	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.000	00:15
Breath Volume	3.750	00:15
Air Blank	0.000	00:16
Control Test	0.080	00:16
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

---

Data Download Date/Time: 06/24/2020 15:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRAWDY,LAINE,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: WASHINGTON,DENZEL  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 05/23/2020 02:43:40

---

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:39	
Control Test	0.080	02:40	
Air Blank	0.000	02:40	
Subject Sample	# 1	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:42	
Control Test	0.080	02:43	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

---

Data Download Date/Time: 06/24/2020 15:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICCARDI,A,  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: RICCARDI,A  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ENGLERT

DATE/TIME OF TEST: 05/23/2020 03:10:32

---

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.080	03:08	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:09	
Control Test	0.081	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

---

Data Download Date/Time: 06/24/2020 15:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARNOLD,J,  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: TURNER,W  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: BUI  
ARREST TIME: 18:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: TRIANA

DATE/TIME OF TEST: 05/23/2020 19:43:52

---

Diagnostics Check	OK	19:41	
Air Blank	0.000	19:41	
Control Test	0.081	19:41	
Air Blank	0.000	19:42	
Subject Sample	# 1	19:42	Subject Test Refused
Breath Volume	0.000	19:42	
Air Blank	0.000	19:42	
Control Test	0.080	19:43	
Air Blank	0.000	19:43	
Diagnostics Check	OK	19:43	

---

Data Download Date/Time: 06/24/2020 15:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMB,HUNTER,  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LIMB,5C10  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 05/23/2020 23:25:45

---

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.080	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.218	23:20
Breath Volume	3.003	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.202	23:24
Breath Volume	2.609	23:24
Air Blank	0.000	23:24
Control Test	0.079	23:25
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

---

Data Download Date/Time: 06/24/2020 15:15



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRAWDY,LAINE,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: DRAWDY,LAINE  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 05/24/2020 00:08:07

---

Diagnostics Check	OK	23:59
Air Blank	0.000	23:59
Control Test	0.080	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.179	00:03
Breath Volume	2.843	00:03
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.187	00:06
Breath Volume	1.382	00:06
Air Blank	0.000	00:07
Control Test	0.078	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

---

Data Download Date/Time: 06/24/2020 15:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRAWDY,LAINE,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: DRAWDY,LAINE  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 05/24/2020 04:49:43

---

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.080	04:43
Air Blank	0.000	04:44
Subject Sample # 1	0.122	04:45
Breath Volume	2.621	04:45
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.124	04:48
Breath Volume	2.398	04:48
Air Blank	0.000	04:48
Control Test	0.080	04:49
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

---

Data Download Date/Time: 06/24/2020 15:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRAWDY,LAINE,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: MONDRAGON,LIONEL  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MCNULTY

DATE/TIME OF TEST: 05/25/2020 04:19:03

---

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.000	04:14
Breath Volume	3.644	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.000	04:17
Breath Volume	3.664	04:17
Air Blank	0.000	04:18
Control Test	0.079	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:19

---

Data Download Date/Time: 06/24/2020 15:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STOKES,SHAYNE,R  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: SOBON,STEVEN  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 08:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:23

SUBJECT'S LAST NAME: MADDALINA

DATE/TIME OF TEST: 05/26/2020 08:46:43

---

Diagnostics Check	OK	08:43	
Air Blank	0.000	08:44	
Control Test	0.079	08:44	
Air Blank	0.000	08:45	
Subject Sample	# 1	08:45	Subject Test Refused
Breath Volume	0.000	08:45	
Air Blank	0.000	08:45	
Control Test	0.079	08:46	
Air Blank	0.000	08:46	
Diagnostics Check	OK	08:46	

---

Data Download Date/Time: 06/24/2020 15:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SELPH,JASON,H  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: GILKEY,JOSHUA  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:03

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/26/2020 16:47:59

---

Diagnostics Check	OK	16:41
Air Blank	0.000	16:41
Control Test	0.080	16:42
Air Blank	0.000	16:42
Subject Sample # 1	0.224	16:43
Breath Volume	1.980	16:43
Air Blank	0.000	16:43
Air Blank	0.000	16:45
Subject Sample # 2	0.224	16:46
Breath Volume	2.175	16:46
Air Blank	0.000	16:47
Control Test	0.080	16:47
Air Blank	0.000	16:47
Diagnostics Check	OK	16:47

---

Data Download Date/Time: 06/24/2020 15:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 05/26/2020 23:31:54

---

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.000	23:27
Breath Volume	3.953	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.000	23:30
Breath Volume	4.023	23:30
Air Blank	0.000	23:30
Control Test	0.079	23:31
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

---

Data Download Date/Time: 06/24/2020 15:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMB,HUNTER,  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LIMB,HUNTER  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: KOHLER

DATE/TIME OF TEST: 05/29/2020 02:12:38

---

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.080	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.249	02:08
Breath Volume	2.730	02:08
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.254	02:11
Breath Volume	2.191	02:11
Air Blank	0.000	02:11
Control Test	0.080	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

---

Data Download Date/Time: 06/24/2020 15:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MC ENERNEY, MELANIE, A  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: HATT, CODY  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: MC ENERNEY

DATE/TIME OF TEST: 05/29/2020 20:58:23

---

Diagnostics Check	OK	20:49	
Air Blank	0.000	20:50	
Control Test	0.080	20:50	
Air Blank	0.000	20:51	
Subject Sample	# 1	20:54	No Sample Provided
Breath Volume	0.000	20:54	
Air Blank	0.000	20:54	
Air Blank	0.000	20:56	
Subject Sample	# 2	20:57	Subject Test Refused
Breath Volume	0.000	20:57	
Air Blank	0.000	20:57	
Control Test	0.080	20:57	
Air Blank	0.000	20:58	
Diagnostics Check	OK	20:58	

---

Data Download Date/Time: 06/24/2020 15:16



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 05/30/2020 03:22:34

---

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.124	03:18
Breath Volume	2.113	03:18
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.126	03:21
Breath Volume	2.714	03:21
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

---

Data Download Date/Time: 06/24/2020 15:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,CODY,L  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: BROWN,CODY  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 11:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 05/31/2020 13:05:55

---

Diagnostics Check	OK	12:59
Air Blank	0.000	13:00
Control Test	0.080	13:00
Air Blank	0.000	13:01
Subject Sample # 1	0.055	13:01
Breath Volume	2.851	13:01
Air Blank	0.000	13:02
Air Blank	0.000	13:04
Subject Sample # 2	0.053	13:04
Breath Volume	2.664	13:04
Air Blank	0.000	13:05
Control Test	0.079	13:05
Air Blank	0.000	13:05
Diagnostics Check	OK	13:05

---

Data Download Date/Time: 06/24/2020 15:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAYNE,JAMES,B  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: PAYNE,JAMES  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: BUI  
ARREST TIME: 15:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:06

SUBJECT'S LAST NAME: SKALA

DATE/TIME OF TEST: 05/31/2020 16:32:50

---

Diagnostics Check	OK	16:26
Air Blank	0.000	16:27
Control Test	0.080	16:27
Air Blank	0.000	16:28
Subject Sample # 1	0.170	16:28
Breath Volume	1.816	16:28
Air Blank	0.000	16:29
Air Blank	0.000	16:30
Subject Sample # 2	0.169	16:31
Breath Volume	1.265	16:31
Air Blank	0.000	16:31
Control Test	0.079	16:32
Air Blank	0.000	16:32
Diagnostics Check	OK	16:32

---

Data Download Date/Time: 06/24/2020 15:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PATTERSON,KYLE,L  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: PATTERSON,KYLE  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: BUI  
ARREST TIME: 18:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: SCHROPP

DATE/TIME OF TEST: 05/31/2020 19:29:11

---

Diagnostics Check	OK	19:23
Air Blank	0.000	19:23
Control Test	0.080	19:23
Air Blank	0.000	19:24
Subject Sample # 1	0.070	19:24
Breath Volume	1.468	19:24
Air Blank	0.000	19:25
Air Blank	0.000	19:27
Subject Sample # 2	0.066	19:27
Breath Volume	2.128	19:27
Air Blank	0.000	19:28
Control Test	0.080	19:28
Air Blank	0.000	19:29
Diagnostics Check	OK	19:29

---

Data Download Date/Time: 06/24/2020 15:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAIG,SHAWN,T  
BREATH TEST OPERATOR AGENCY: Live Oak PD  
ARRESTING OFFICER: CRAIG,SHAWN  
ARRESTING OFFICER AGENCY: Live Oak PD  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 05/13/2020 00:31:33

---

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.077	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.111	00:27
Breath Volume	1.613	00:27
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.106	00:30
Breath Volume	1.316	00:30
Air Blank	0.000	00:30
Control Test	0.077	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

---

Data Download Date/Time: 05/31/2020 17:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CZABAN,NICHOLAS,W  
BREATH TEST OPERATOR AGENCY: FL Dept of AG  
ARRESTING OFFICER: DILL,JOHN  
ARRESTING OFFICER AGENCY: FL Dept of AG  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 05/24/2020 18:18:33

---

Diagnostics Check	OK	18:10	
Air Blank	0.000	18:10	
Control Test	0.078	18:11	
Air Blank	0.000	18:11	
Subject Sample # 1	0.131	18:12	
Breath Volume	1.312	18:12	
Air Blank	0.000	18:13	
Air Blank	0.000	18:15	
Subject Sample # 2	0.130	18:16	
Breath Volume	1.210	18:16	
Air Blank	0.000	18:17	
Control Test	0.075	18:17	
Air Blank		18:18	Ambient Fail
Air Blank	0.000	18:18	

---

Data Download Date/Time: 05/31/2020 17:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERDER,BYRON,E  
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)  
ARRESTING OFFICER: BURNSED III,DAVID  
ARRESTING OFFICER AGENCY: FFWCC (Lake City)  
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: BUI  
ARREST TIME: 19:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: HERDER

DATE/TIME OF TEST: 05/24/2020 20:05:09

---

Diagnostics Check	OK	19:53	
Air Blank	0.000	19:54	
Control Test	0.077	19:54	
Air Blank	0.000	19:55	
Subject Sample	# 1	19:58	No Sample Provided
Breath Volume	0.000	19:58	
Air Blank	0.000	19:58	
Air Blank	0.000	20:00	
Subject Sample	# 2	20:03	No Sample Provided
Breath Volume	0.000	20:03	
Air Blank	0.000	20:04	
Control Test	0.078	20:04	
Air Blank	0.000	20:05	
Diagnostics Check	OK	20:05	

---

Data Download Date/Time: 05/31/2020 17:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAYLOR,SHANE,G  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: TAYLOR,SHANE  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 12:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:29

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 05/10/2020 14:17:18

---

Diagnostics Check	OK	14:10
Air Blank	0.000	14:11
Control Test	0.080	14:11
Air Blank	0.000	14:12
Subject Sample # 1	0.269	14:12
Breath Volume	2.667	14:12
Air Blank	0.000	14:13
Air Blank	0.000	14:15
Subject Sample # 2	0.277	14:15
Breath Volume	3.878	14:15
Air Blank	0.000	14:16
Control Test	0.080	14:16
Air Blank	0.000	14:17
Diagnostics Check	OK	14:17

---

Data Download Date/Time: 06/10/2020 12:06



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FUTCH,JAMES,M  
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)  
ARRESTING OFFICER: FUTCH,JAMES  
ARRESTING OFFICER AGENCY: FFWCC (Lake City)  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: BUI  
ARREST TIME: 19:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 05/23/2020 20:31:51

---

Diagnostics Check	OK	20:25
Air Blank	0.000	20:26
Control Test	0.079	20:26
Air Blank	0.000	20:27
Subject Sample # 1	0.070	20:27
Breath Volume	1.519	20:27
Air Blank	0.000	20:28
Air Blank	0.000	20:30
Subject Sample # 2	0.068	20:30
Breath Volume	2.058	20:30
Air Blank	0.000	20:30
Control Test	0.079	20:31
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

---

Data Download Date/Time: 06/10/2020 12:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASH,EARL,C  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MASH,EARL  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 05/26/2020 16:38:47

---

Diagnostics Check	OK	16:35	
Air Blank	0.000	16:35	
Control Test	0.079	16:36	
Air Blank	0.000	16:36	
Subject Sample	# 1	16:37	Subject Test Refused
Breath Volume	0.000	16:37	
Air Blank	0.000	16:37	
Control Test	0.079	16:38	
Air Blank	0.000	16:38	
Diagnostics Check	OK	16:38	

---

Data Download Date/Time: 06/10/2020 12:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,MATT,M  
BREATH TEST OPERATOR AGENCY: Perry PD  
ARRESTING OFFICER: SMITH,MATT  
ARRESTING OFFICER AGENCY: Perry PD  
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: STALANS

DATE/TIME OF TEST: 05/12/2020 20:08:57

---

Diagnostics Check		OK	19:59
Air Blank		0.000	20:00
Control Test		0.081	20:00
Air Blank		0.000	20:00
Subject Sample	# 1	0.035	20:02
Breath Volume		1.417	20:02
Air Blank		0.000	20:03
Air Blank		0.000	20:05
Subject Sample	# 2	0.033	20:07
Breath Volume		1.234	20:07
Air Blank		0.000	20:08
Control Test		0.079	20:08
Air Blank		0.000	20:08
Diagnostics Check		OK	20:08

---

Data Download Date/Time: 05/27/2020 11:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CEPHUS,JAMIN,I  
BREATH TEST OPERATOR AGENCY: Perry PD  
ARRESTING OFFICER: CEPHUS,JAMIN  
ARRESTING OFFICER AGENCY: Perry PD  
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: ZERO TOL  
ARREST TIME: 12:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 05/14/2020 13:20:16

---

Diagnostics Check	OK	13:14
Air Blank	0.000	13:14
Control Test	0.079	13:15
Air Blank	0.000	13:15
Subject Sample # 1	0.000	13:15
Breath Volume	2.863	13:15
Air Blank	0.000	13:16
Air Blank	0.000	13:18
Subject Sample # 2	0.000	13:18
Breath Volume	3.175	13:18
Air Blank	0.000	13:19
Control Test	0.079	13:19
Air Blank	0.000	13:20
Diagnostics Check	OK	13:20

---

Data Download Date/Time: 05/27/2020 11:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHAFFIN,TRACE,D  
BREATH TEST OPERATOR AGENCY: Perry PD  
ARRESTING OFFICER: RICE,JOHN  
ARRESTING OFFICER AGENCY: Perry PD  
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: FUDGE

DATE/TIME OF TEST: 05/21/2020 23:58:48

---

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:53	
Control Test	0.079	23:54	
Air Blank	0.000	23:54	
Subject Sample	# 1 0.136	23:55	
Breath Volume	1.539	23:55	
Air Blank	0.000	23:55	
Air Blank	0.000	23:57	
Subject Sample	# 2 0.128	23:58	
Breath Volume	1.441	23:58	
Air Blank		23:58	RFI Detect
Air Blank		23:58	RFI Detect

---

Data Download Date/Time: 05/27/2020 11:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICKETSON,JAY,R  
BREATH TEST OPERATOR AGENCY: Perry PD  
ARRESTING OFFICER: TOUCHTON,J  
ARRESTING OFFICER AGENCY: FFWCC (Lake City)  
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG815602  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 10:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: SPEES

DATE/TIME OF TEST: 05/24/2020 22:59:15

---

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.080	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.076	22:54
Breath Volume	3.718	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.077	22:57
Breath Volume	3.898	22:57
Air Blank	0.000	22:58
Control Test	0.080	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

---

Data Download Date/Time: 05/27/2020 11:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: SHEFFIELD,SHYKEITHA  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: HOBDY

DATE/TIME OF TEST: 05/01/2020 00:38:48

---

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.082	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.000	00:34
Breath Volume	2.351	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.000	00:37
Breath Volume	1.253	00:37
Air Blank	0.000	00:37
Control Test	0.082	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

---

Data Download Date/Time: 05/26/2020 21:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORCHARD,DUSTIN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: ORCHARD,DUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: MANTOVANI

DATE/TIME OF TEST: 05/01/2020 17:02:28

---

Diagnostics Check		OK	16:56
Air Blank		0.000	16:56
Control Test		0.080	16:57
Air Blank		0.000	16:57
Subject Sample	# 1	0.370	16:58
Breath Volume		1.656	16:58
Air Blank		0.000	16:58
Air Blank		0.000	17:00
Subject Sample	# 2	0.351	17:00
Breath Volume		2.156	17:00
Air Blank		0.000	17:01
Control Test		0.081	17:01
Air Blank		0.000	17:02
Diagnostics Check		OK	17:02

---

Data Download Date/Time: 05/26/2020 21:41



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESQUIVEL,CHRISTINA,T  
BREATH TEST OPERATOR AGENCY: Belleview PD  
ARRESTING OFFICER: ESQUIVEL,CHRISTINA  
ARRESTING OFFICER AGENCY: Belleview PD  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: NEVEAU

DATE/TIME OF TEST: 05/03/2020 22:35:23

---

Diagnostics Check	OK	22:28
Air Blank	0.000	22:29
Control Test	0.081	22:29
Air Blank	0.000	22:30
Subject Sample # 1	0.000	22:30
Breath Volume	2.781	22:30
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.000	22:34
Breath Volume	3.320	22:34
Air Blank	0.000	22:34
Control Test	0.081	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

---

Data Download Date/Time: 05/26/2020 21:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PETERSON,KIP,J  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: PETERSON,KIP  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 05/06/2020 20:49:55

---

Diagnostics Check	OK	20:43
Air Blank	0.000	20:43
Control Test	0.082	20:44
Air Blank	0.000	20:44
Subject Sample # 1	0.264	20:45
Breath Volume	1.437	20:45
Air Blank	0.000	20:46
Air Blank	0.000	20:48
Subject Sample # 2	0.261	20:48
Breath Volume	2.242	20:48
Air Blank	0.000	20:49
Control Test	0.079	20:49
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

---

Data Download Date/Time: 05/26/2020 21:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BATTS,C,  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: BATTS,C  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: N400118012830

DATE/TIME OF TEST: 05/07/2020 01:07:50

---

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.081	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.157	01:03
Breath Volume	1.601	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.153	01:06
Breath Volume	2.105	01:06
Air Blank	0.000	01:06
Control Test	0.081	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

---

Data Download Date/Time: 05/26/2020 21:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COON,AUSTIN,D  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: COON,AUSTIN  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 05/08/2020 22:43:03

---

Diagnostics Check	OK	22:40	
Air Blank	0.000	22:40	
Control Test	0.080	22:40	
Air Blank	0.000	22:41	
Subject Sample	# 1	22:41	Subject Test Refused
Breath Volume	0.000	22:41	
Air Blank	0.000	22:42	
Control Test	0.081	22:42	
Air Blank	0.000	22:42	
Diagnostics Check	OK	22:43	

---

Data Download Date/Time: 05/26/2020 21:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESQUIVEL,ANTHONY,L  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: ESQUIVEL,ANTHONY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: MAGAN

DATE/TIME OF TEST: 05/11/2020 22:46:54

---

Diagnostics Check	OK	22:40
Air Blank	0.000	22:41
Control Test	0.081	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.213	22:42
Breath Volume	3.121	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:44
Subject Sample # 2	0.220	22:45
Breath Volume	2.546	22:45
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

---

Data Download Date/Time: 05/26/2020 21:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PETERSON,KIP,J  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: PETERSON,KIP  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:13

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 05/12/2020 18:42:17

---

Diagnostics Check	OK	18:33
Air Blank	0.000	18:34
Control Test	0.081	18:34
Air Blank	0.000	18:35
Subject Sample # 1	0.235	18:36
Breath Volume	3.128	18:36
Air Blank	0.000	18:37
Air Blank	0.000	18:39
Subject Sample # 2	0.224	18:40
Breath Volume	2.226	18:40
Air Blank	0.000	18:41
Control Test	0.079	18:41
Air Blank	0.000	18:42
Diagnostics Check	OK	18:42

---

Data Download Date/Time: 05/26/2020 21:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BALL,CODY,W  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BALL,CODY  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/14/2020 00:36:05

---

Diagnostics Check	OK	00:28
Air Blank	0.000	00:29
Control Test	0.081	00:29
Air Blank	0.000	00:30
Subject Sample # 1	0.134	00:31
Breath Volume	1.976	00:31
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.135	00:34
Breath Volume	2.058	00:34
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:35
Diagnostics Check	OK	00:36

---

Data Download Date/Time: 05/26/2020 21:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PETERSON,KIP,J  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: PETERSON,KIP  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: OROZCO

DATE/TIME OF TEST: 05/14/2020 01:15:32

---

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.081	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.000	01:11
Breath Volume	1.578	01:11
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.000	01:14
Breath Volume	1.386	01:14
Air Blank	0.000	01:14
Control Test	0.081	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

---

Data Download Date/Time: 05/26/2020 21:41



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CURTIS,LAYNE,A  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: CURTIS,LAYNE  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 12:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:23

SUBJECT'S LAST NAME: HANSON

DATE/TIME OF TEST: 05/14/2020 12:56:51

---

Diagnostics Check	OK	12:53	
Air Blank	0.000	12:54	
Control Test	0.081	12:54	
Air Blank	0.000	12:54	
Subject Sample	# 1	12:55	Subject Test Refused
Breath Volume	0.000	12:55	
Air Blank	0.000	12:55	
Control Test	0.081	12:56	
Air Blank	0.000	12:56	
Diagnostics Check	OK	12:56	

---

Data Download Date/Time: 05/26/2020 21:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PULFORD,G,  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: PULFORD,G  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 05/16/2020 22:33:42

---

Diagnostics Check	OK	22:27
Air Blank	0.000	22:27
Control Test	0.081	22:28
Air Blank	0.000	22:28
Subject Sample # 1	0.133	22:29
Breath Volume	1.492	22:29
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.141	22:32
Breath Volume	1.582	22:32
Air Blank	0.000	22:32
Control Test	0.081	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

---

Data Download Date/Time: 05/26/2020 21:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BALL,CODY,W  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BALL,CODY  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: TIBBETTS

DATE/TIME OF TEST: 05/19/2020 00:28:08

---

Diagnostics Check		OK	00:18
Air Blank		0.000	00:19
Control Test		0.081	00:19
Air Blank		0.000	00:19
Subject Sample	# 1	0.168	00:22
Breath Volume		1.917	00:22
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.156	00:26
Breath Volume		1.707	00:26
Air Blank		0.000	00:27
Control Test		0.080	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

---

Data Download Date/Time: 05/26/2020 21:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 05/20/2020 04:55:36

---

Diagnostics Check		OK	04:43	
Air Blank		0.000	04:44	
Control Test		0.080	04:44	
Air Blank		0.000	04:45	
Subject Sample	# 1	0.187	04:48	Volume Not Met
Breath Volume		0.000	04:48	
Air Blank		0.000	04:49	
Air Blank		0.000	04:50	
Subject Sample	# 2	0.196	04:53	Volume Not Met
Breath Volume		0.253	04:53	
Air Blank		0.000	04:54	
Control Test		0.080	04:54	
Air Blank		0.000	04:55	
Diagnostics Check		OK	04:55	

---

Data Download Date/Time: 05/26/2020 21:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 05/20/2020 05:06:16

---

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.080	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.186	05:00
Breath Volume	1.953	05:00
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.190	05:04
Breath Volume	1.246	05:04
Air Blank	0.000	05:05
Control Test	0.080	05:05
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

---

Data Download Date/Time: 05/26/2020 21:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WORFEL,KYLE,A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: WORFEL,KYLE  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: LANGSTON

DATE/TIME OF TEST: 05/20/2020 05:39:05

---

Diagnostics Check	OK	05:31
Air Blank	0.000	05:31
Control Test	0.079	05:31
Air Blank	0.000	05:32
Subject Sample # 1	0.173	05:34
Breath Volume	2.066	05:34
Air Blank	0.000	05:35
Air Blank	0.000	05:36
Subject Sample # 2	0.175	05:37
Breath Volume	1.351	05:37
Air Blank	0.000	05:38
Control Test	0.080	05:38
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

---

Data Download Date/Time: 05/26/2020 21:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESQUIVEL,ANTHONY,L  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: BROWN,KYLE  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: WOHLER

DATE/TIME OF TEST: 05/21/2020 04:57:13

---

Diagnostics Check	OK	04:51
Air Blank	0.000	04:51
Control Test	0.082	04:51
Air Blank	0.000	04:52
Subject Sample # 1	0.104	04:52
Breath Volume	1.308	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.108	04:55
Breath Volume	2.031	04:55
Air Blank	0.000	04:56
Control Test	0.081	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

---

Data Download Date/Time: 05/26/2020 21:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POPE,CALVIN,W  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: POPE,CALVIN  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 05/21/2020 21:49:14

---

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:46	
Control Test	0.081	21:47	
Air Blank	0.000	21:47	
Subject Sample	# 1	21:47	Subject Test Refused
Breath Volume	0.000	21:47	
Air Blank	0.000	21:48	
Control Test	0.081	21:48	
Air Blank	0.000	21:49	
Diagnostics Check	OK	21:49	

---

Data Download Date/Time: 05/26/2020 21:41



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HALL,CATHERINE,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: HALL,CATHERINE  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: WORLO

DATE/TIME OF TEST: 05/23/2020 00:23:09

---

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:20	
Control Test	0.081	00:21	
Air Blank	0.000	00:21	
Subject Sample	# 1	00:21	Subject Test Refused
Breath Volume	0.000	00:21	
Air Blank	0.000	00:22	
Control Test	0.081	00:22	
Air Blank	0.000	00:23	
Diagnostics Check	OK	00:23	

---

Data Download Date/Time: 05/26/2020 21:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DONELY,JARRICK,S  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: DONELY,JARRICK  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 05/23/2020 21:07:53

---

Diagnostics Check	OK	20:56	
Air Blank	0.000	20:56	
Control Test	0.080	20:57	
Air Blank	0.000	20:57	
Subject Sample	# 1	21:00	No Sample Provided
Breath Volume	0.000	21:00	
Air Blank	0.000	21:01	
Air Blank	0.000	21:03	
Subject Sample	# 2	21:06	No Sample Provided
Breath Volume	0.000	21:06	
Air Blank	0.000	21:06	
Control Test	0.082	21:07	
Air Blank	0.000	21:07	
Diagnostics Check	OK	21:07	

---

Data Download Date/Time: 05/26/2020 21:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COON,AUSTIN,D  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: COON,AUSTIN  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: KERN

DATE/TIME OF TEST: 05/24/2020 21:41:55

---

Diagnostics Check	OK	21:35
Air Blank	0.000	21:35
Control Test	0.081	21:36
Air Blank	0.000	21:36
Subject Sample # 1	0.125	21:37
Breath Volume	1.558	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:39
Subject Sample # 2	0.120	21:40
Breath Volume	1.796	21:40
Air Blank	0.000	21:41
Control Test	0.081	21:41
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

---

Data Download Date/Time: 05/26/2020 21:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, TRAVIS, D  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: WILSON, DAKOTA  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 05/04/2020 22:36:47

---

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:21	
Control Test	0.079	22:21	
Air Blank	0.000	22:22	
Subject Sample # 1	0.013	22:24	
Breath Volume	1.589	22:24	
Air Blank	0.000	22:24	
Air Blank	0.000	22:26	
Subject Sample # 2	0.010	22:29	Volume Not Met
Breath Volume	0.554	22:29	
Air Blank	0.000	22:30	
Air Blank	0.000	22:32	
Subject Sample # 3	0.010	22:35	
Breath Volume	1.679	22:35	
Air Blank	0.000	22:35	
Control Test	0.078	22:36	
Air Blank	0.000	22:36	
Diagnostics Check	OK	22:36	

---

Data Download Date/Time: 05/23/2020 20:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: BALL,CODY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: GROGHAN

DATE/TIME OF TEST: 05/08/2020 22:16:03

---

Diagnostics Check	OK	22:05
Air Blank	0.000	22:06
Control Test	0.078	22:06
Air Blank	0.000	22:06
Subject Sample # 1	0.315	22:08
Breath Volume	1.332	22:08
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.336	22:11
Breath Volume	3.132	22:11
Air Blank	0.000	22:12
Air Blank	0.000	22:13
Subject Sample # 3	0.324	22:14
Breath Volume	2.285	22:14
Air Blank	0.000	22:15
Control Test	0.077	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:16

---

Data Download Date/Time: 05/23/2020 20:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,WYATT,M  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: MANDERS,M  
ARRESTING OFFICER AGENCY: Wildwood PD  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: DURANT

DATE/TIME OF TEST: 05/11/2020 23:03:06

---

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.079	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.173	22:58
Breath Volume	1.886	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.161	23:01
Breath Volume	1.675	23:01
Air Blank	0.000	23:02
Control Test	0.078	23:02
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

---

Data Download Date/Time: 05/23/2020 20:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, TRAVIS, D  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: COSSEBOOM, MICHAEL  
ARRESTING OFFICER AGENCY: Wildwood PD  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: ENGLISH

DATE/TIME OF TEST: 05/15/2020 00:42:38

---

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:40	
Control Test	0.079	00:40	
Air Blank	0.000	00:41	
Subject Sample	# 1	00:41	Subject Test Refused
Breath Volume	0.000	00:41	
Air Blank	0.000	00:41	
Control Test	0.078	00:42	
Air Blank	0.000	00:42	
Diagnostics Check	OK	00:42	

---

Data Download Date/Time: 05/23/2020 20:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ISHALL,JALEN,L  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: ADAMS,CLINT  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: MCINTYRE

DATE/TIME OF TEST: 05/15/2020 20:16:25

---

Diagnostics Check	OK	20:13	
Air Blank	0.000	20:13	
Control Test	0.079	20:14	
Air Blank	0.000	20:14	
Subject Sample	# 1	20:15	Subject Test Refused
Breath Volume	0.000	20:15	
Air Blank	0.000	20:15	
Control Test	0.078	20:15	
Air Blank	0.000	20:16	
Diagnostics Check	OK	20:16	

---

Data Download Date/Time: 05/23/2020 20:42



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LETSON,REBECCA,G  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: LETSON,REBECCA  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: SCHAEFFER

DATE/TIME OF TEST: 05/16/2020 02:04:09

---

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.078	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.121	01:59
Breath Volume	4.066	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.116	02:02
Breath Volume	2.761	02:02
Air Blank	0.000	02:03
Control Test	0.079	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

---

Data Download Date/Time: 05/23/2020 20:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARD,NICHOLAS,D  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: ELMORE,PATRICK  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 07:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: KENNETT

DATE/TIME OF TEST: 05/19/2020 08:17:22

---

Diagnostics Check	OK	08:10
Air Blank	0.000	08:11
Control Test	0.079	08:11
Air Blank	0.000	08:12
Subject Sample # 1	0.140	08:12
Breath Volume	3.703	08:12
Air Blank	0.000	08:13
Air Blank	0.000	08:15
Subject Sample # 2	0.142	08:15
Breath Volume	3.187	08:15
Air Blank	0.000	08:16
Control Test	0.079	08:16
Air Blank	0.000	08:17
Diagnostics Check	OK	08:17

---

Data Download Date/Time: 05/23/2020 20:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LETSON,REBECCA,G  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: LETSON,REBECCA  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 05/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 05/29/2020 23:32:15

---

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.078	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.114	23:27
Breath Volume	1.550	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.110	23:30
Breath Volume	1.765	23:30
Air Blank	0.000	23:31
Control Test	0.077	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

---

Data Download Date/Time: 06/21/2020 00:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ISHALL,JALEN,L  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: ALEXANDER,DENZEL  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 05/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: NAJI

DATE/TIME OF TEST: 05/31/2020 20:51:29

---

Diagnostics Check	OK	20:48	
Air Blank	0.000	20:49	
Control Test	0.078	20:49	
Air Blank	0.000	20:49	
Subject Sample	# 1	20:50	Subject Test Refused
Breath Volume	0.000	20:50	
Air Blank	0.000	20:50	
Control Test	0.078	20:50	
Air Blank	0.000	20:51	
Diagnostics Check	OK	20:51	

---

Data Download Date/Time: 06/21/2020 00:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HOBBS,KYLE,R

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: SILVERIO,A

ARRESTING OFFICER AGENCY: Mascotte PD

LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30219080A3

DRY GAS EXPIRY DATE: 01/05/2022

VIOLATION CODE: DUI

ARREST TIME: 20:23

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: NEGRETE

DATE/TIME OF TEST: 05/02/2020 22:08:33

---

Diagnostics Check		OK	21:59	
Air Blank		0.000	22:00	
Control Test		0.080	22:00	
Air Blank		0.000	22:01	
Subject Sample	# 1	0.247	22:04	Volume Not Met
Breath Volume		0.835	22:04	
Air Blank		0.000	22:05	
Air Blank		0.000	22:06	
Subject Sample	# 2		22:08	Sequence Aborted
Breath Volume		0.046	22:08	
Air Blank		0.000	22:08	

---

Data Download Date/Time: 05/22/2020 20:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HOBBS,KYLE,R

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: HOBBS,KYLE

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30219080A3

DRY GAS EXPIRY DATE: 01/05/2022

VIOLATION CODE: DUI

ARREST TIME: 21:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: CARDENAS

DATE/TIME OF TEST: 05/03/2020 22:16:15

---

Diagnostics Check		OK	22:03	
Air Blank		0.000	22:04	
Control Test		0.081	22:04	
Air Blank		0.000	22:05	
Subject Sample	# 1	0.212	22:08	Volume Not Met
Breath Volume		0.292	22:08	
Air Blank		0.000	22:08	
Air Blank		0.000	22:10	
Subject Sample	# 2	0.198	22:11	
Breath Volume		1.835	22:11	
Air Blank		0.000	22:12	
Air Blank		0.000	22:14	
Subject Sample	# 3	0.201	22:14	
Breath Volume		1.843	22:14	
Air Blank		0.000	22:15	
Control Test		0.076	22:15	
Air Blank		0.000	22:16	
Diagnostics Check		OK	22:16	

---

Data Download Date/Time: 05/22/2020 20:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: WASHINGTON,D  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 04/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 05/01/2020 22:16:18

---

Diagnostics Check	OK	22:04	
Air Blank	0.000	22:05	
Control Test	0.080	22:05	
Air Blank	0.000	22:06	
Subject Sample	# 1 0.187	22:09	Volume Not Met
Breath Volume	0.949	22:09	
Air Blank	0.000	22:09	
Air Blank	0.000	22:11	
Subject Sample	# 2	22:14	No Sample Provided
Breath Volume	0.000	22:14	
Air Blank	0.000	22:15	
Control Test	0.078	22:15	
Air Blank	0.000	22:16	
Diagnostics Check	OK	22:16	

---

Data Download Date/Time: 05/08/2020 08:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: WASHINGTON,D  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 04/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 05/01/2020 22:26:45

---

Diagnostics Check		OK	22:18	
Air Blank		0.000	22:18	
Control Test		0.079	22:18	
Air Blank		0.000	22:19	
Subject Sample	# 1	0.166	22:22	Volume Not Met
Breath Volume		0.242	22:22	
Air Blank		0.000	22:23	
Air Blank		0.000	22:24	
Subject Sample	# 2		22:25	Subject Test Refused
Breath Volume		0.000	22:25	
Air Blank		0.000	22:25	
Control Test		0.078	22:26	
Air Blank		0.000	22:26	
Diagnostics Check		OK	22:26	

---

Data Download Date/Time: 05/08/2020 08:30



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: RENDON,D  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 04/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: GOMEZ MORALES

DATE/TIME OF TEST: 05/03/2020 02:09:06

---

Diagnostics Check		OK	02:02
Air Blank		0.000	02:03
Control Test		0.079	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.188	02:04
Breath Volume		2.234	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.199	02:07
Breath Volume		2.453	02:07
Air Blank		0.000	02:08
Control Test		0.079	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

---

Data Download Date/Time: 05/08/2020 08:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: PATRON,A  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 04/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: BANDA

DATE/TIME OF TEST: 05/03/2020 02:18:52

---

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.162	02:14
Breath Volume	2.578	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.156	02:17
Breath Volume	2.015	02:17
Air Blank	0.000	02:17
Control Test	0.078	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

---

Data Download Date/Time: 05/08/2020 08:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR, BARBARA, H  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: CHU, A  
ARRESTING OFFICER AGENCY: FHP TROOP C  
LAST AGENCY INSPECTION: 04/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: LOVESTRAND

DATE/TIME OF TEST: 05/03/2020 18:11:53

---

Diagnostics Check	OK	18:05
Air Blank	0.000	18:06
Control Test	0.079	18:06
Air Blank	0.000	18:07
Subject Sample # 1	0.174	18:07
Breath Volume	3.156	18:07
Air Blank	0.000	18:08
Air Blank	0.000	18:09
Subject Sample # 2	0.173	18:10
Breath Volume	2.117	18:10
Air Blank	0.000	18:10
Control Test	0.079	18:11
Air Blank	0.000	18:11
Diagnostics Check	OK	18:11

---

Data Download Date/Time: 05/08/2020 08:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: VAN-TREESE,S  
ARRESTING OFFICER AGENCY: TAMPA PD  
LAST AGENCY INSPECTION: 04/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: KEMP

DATE/TIME OF TEST: 05/03/2020 23:17:25

---

Diagnostics Check	OK	23:10
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:11
Subject Sample # 1	0.188	23:12
Breath Volume	1.226	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.186	23:15
Breath Volume	2.421	23:15
Air Blank	0.000	23:16
Control Test	0.079	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

---

Data Download Date/Time: 05/08/2020 08:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS,ANGELA,V  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: BRIZENDINE,J  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 07:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:21

SUBJECT'S LAST NAME: MOTWANI

DATE/TIME OF TEST: 05/08/2020 09:50:01

---

Diagnostics Check	OK	09:42
Air Blank	0.000	09:42
Control Test	0.079	09:43
Air Blank	0.000	09:43
Subject Sample # 1	0.000	09:45
Breath Volume	1.949	09:45
Air Blank	0.000	09:46
Air Blank	0.000	09:48
Subject Sample # 2	0.000	09:48
Breath Volume	2.132	09:48
Air Blank	0.000	09:49
Control Test	0.079	09:49
Air Blank	0.000	09:49
Diagnostics Check	OK	09:50

---

Data Download Date/Time: 06/04/2020 14:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: TAMPA PD  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: FIELDS HALL

DATE/TIME OF TEST: 05/08/2020 23:58:37

---

Diagnostics Check		OK	23:47	
Air Blank		0.000	23:47	
Control Test		0.079	23:47	
Air Blank		0.000	23:48	
Subject Sample	# 1	0.187	23:51	Volume Not Met
Breath Volume		0.847	23:51	
Air Blank		0.000	23:52	
Air Blank		0.000	23:53	
Subject Sample	# 2	0.186	23:57	Volume Not Met
Breath Volume		0.750	23:57	
Air Blank		0.000	23:57	
Control Test		0.078	23:58	
Air Blank		0.000	23:58	
Diagnostics Check		OK	23:58	

---

Data Download Date/Time: 06/04/2020 14:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: SHEINBERG,T  
ARRESTING OFFICER AGENCY: FHP TROOP C  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 05/09/2020 01:46:13

---

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.105	01:41
Breath Volume	1.785	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:43
Subject Sample # 2	0.111	01:44
Breath Volume	1.746	01:44
Air Blank	0.000	01:45
Control Test	0.079	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

---

Data Download Date/Time: 06/04/2020 14:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: PATRON,A  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 05/09/2020 02:21:03

---

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:10	
Control Test	0.079	02:10	
Air Blank	0.000	02:10	
Subject Sample	# 1	02:14	No Sample Provided
Breath Volume	0.000	02:14	
Air Blank	0.000	02:14	
Air Blank	0.000	02:16	
Subject Sample	# 2	02:19	No Sample Provided
Breath Volume	0.000	02:19	
Air Blank	0.000	02:20	
Control Test	0.079	02:20	
Air Blank	0.000	02:20	
Diagnostics Check	OK	02:21	

---

Data Download Date/Time: 06/04/2020 14:13



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: PATRON,A  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 05/09/2020 02:34:38

---

Diagnostics Check	OK	02:23		
Air Blank	0.000	02:23		
Control Test	0.079	02:23		
Air Blank	0.000	02:24		
Subject Sample	# 1	02:27	No Sample Provided	
Breath Volume	0.000	02:27		
Air Blank	0.000	02:27		
Air Blank	0.000	02:29		
Subject Sample	# 2	0.114	02:33	Volume Not Met
Breath Volume	0.171	02:33		
Air Blank	0.000	02:33		
Control Test	0.080	02:34		
Air Blank	0.000	02:34		
Diagnostics Check	OK	02:34		

---

Data Download Date/Time: 06/04/2020 14:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: PATRON,A  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 05/09/2020 02:50:33

---

Diagnostics Check		OK	02:36	
Air Blank		0.000	02:36	
Control Test		0.078	02:37	
Air Blank		0.000	02:37	
Subject Sample	# 1	0.134	02:40	Volume Not Met
Breath Volume		0.257	02:40	
Air Blank		0.000	02:41	
Air Blank		0.000	02:43	
Subject Sample	# 2	0.123	02:45	
Breath Volume		1.750	02:45	
Air Blank		0.000	02:45	
Air Blank		0.000	02:47	
Subject Sample	# 3	0.126	02:49	
Breath Volume		1.425	02:49	
Air Blank		0.000	02:49	
Control Test		0.077	02:49	
Air Blank		0.000	02:50	
Diagnostics Check		OK	02:50	

---

Data Download Date/Time: 06/04/2020 14:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: WASHINGTON,D  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: RAMSINHG

DATE/TIME OF TEST: 05/10/2020 03:11:55

---

Diagnostics Check	OK	03:00		
Air Blank	0.000	03:00		
Control Test	0.080	03:01		
Air Blank	0.000	03:01		
Subject Sample	# 1	03:04	No Sample Provided	
Breath Volume	0.000	03:04		
Air Blank	0.000	03:05		
Air Blank	0.000	03:07		
Subject Sample	# 2	0.165	03:10	Volume Not Met
Breath Volume	0.191	03:10		
Air Blank	0.000	03:11		
Control Test	0.079	03:11		
Air Blank	0.000	03:11		
Diagnostics Check	OK	03:11		

---

Data Download Date/Time: 06/04/2020 14:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: HUTCHESON,S  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: RUIZ GARCIA

DATE/TIME OF TEST: 05/10/2020 23:29:52

---

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.079	23:23
Air Blank	0.000	23:24
Subject Sample # 1	0.168	23:25
Breath Volume	3.695	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:27
Subject Sample # 2	0.169	23:28
Breath Volume	3.457	23:28
Air Blank	0.000	23:28
Control Test	0.079	23:29
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

---

Data Download Date/Time: 06/04/2020 14:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: COSTA,A  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: DAVEY-BEAN

DATE/TIME OF TEST: 05/12/2020 02:45:50

---

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.079	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.154	02:41
Breath Volume	2.757	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.149	02:44
Breath Volume	1.476	02:44
Air Blank	0.000	02:44
Control Test	0.078	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

---

Data Download Date/Time: 06/04/2020 14:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARBARA,BARR,H  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: BARNA,T  
ARRESTING OFFICER AGENCY: FHP TROOP C  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:04

SUBJECT'S LAST NAME: MILLIKAN

DATE/TIME OF TEST: 05/12/2020 17:31:40

---

Diagnostics Check	OK	17:25
Air Blank	0.000	17:25
Control Test	0.079	17:26
Air Blank	0.000	17:26
Subject Sample # 1	0.232	17:27
Breath Volume	2.503	17:27
Air Blank	0.000	17:27
Air Blank	0.000	17:29
Subject Sample # 2	0.221	17:30
Breath Volume	1.863	17:30
Air Blank	0.000	17:30
Control Test	0.079	17:31
Air Blank	0.000	17:31
Diagnostics Check	OK	17:31

---

Data Download Date/Time: 06/04/2020 14:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: KLENOFSKY,A  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: REEDY

DATE/TIME OF TEST: 05/15/2020 00:14:48

---

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:08
Subject Sample # 1	0.132	00:09
Breath Volume	2.976	00:09
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.136	00:13
Breath Volume	3.148	00:13
Air Blank	0.000	00:13
Control Test	0.080	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

---

Data Download Date/Time: 06/04/2020 14:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: RODRIGUEZ,D  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/15/2020 23:24:35

---

Diagnostics Check		OK	23:10	
Air Blank		0.000	23:10	
Control Test		0.080	23:10	
Air Blank		0.000	23:11	
Subject Sample	# 1	0.000	23:14	Volume Not Met
Breath Volume		0.136	23:14	
Air Blank		0.000	23:15	
Air Blank		0.000	23:17	
Subject Sample	# 2	0.000	23:20	
Breath Volume		1.390	23:20	
Air Blank		0.000	23:20	
Air Blank		0.000	23:22	
Subject Sample	# 3	0.000	23:23	
Breath Volume		1.476	23:23	
Air Blank		0.000	23:23	
Control Test		0.080	23:23	
Air Blank		0.000	23:24	
Diagnostics Check		OK	23:24	

---

Data Download Date/Time: 06/04/2020 14:13



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: PATRON,A  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CHIARENZA

DATE/TIME OF TEST: 05/16/2020 04:07:28

---

Diagnostics Check		OK	04:00
Air Blank		0.000	04:01
Control Test		0.080	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.138	04:02
Breath Volume		1.292	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.142	04:05
Breath Volume		1.636	04:05
Air Blank		0.000	04:06
Control Test		0.079	04:06
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

---

Data Download Date/Time: 06/04/2020 14:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: VISSER,A  
ARRESTING OFFICER AGENCY: TAMPA PD  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 05/17/2020 00:44:43

---

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:41	
Control Test	0.080	00:41	
Air Blank	0.000	00:41	
Subject Sample	# 1	00:43	Subject Test Refused
Breath Volume	0.000	00:43	
Air Blank	0.000	00:43	
Control Test	0.079	00:44	
Air Blank	0.000	00:44	
Diagnostics Check	OK	00:44	

---

Data Download Date/Time: 06/04/2020 14:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: COSTA,A  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: BRANDON

DATE/TIME OF TEST: 05/17/2020 01:56:20

---

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.080	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.036	01:51
Breath Volume	2.093	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.035	01:54
Breath Volume	1.988	01:54
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

---

Data Download Date/Time: 06/04/2020 14:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: PATRON,A  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 05/17/2020 04:13:20

---

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.163	04:09
Breath Volume	1.554	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.157	04:11
Breath Volume	1.562	04:11
Air Blank	0.000	04:12
Control Test	0.079	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

---

Data Download Date/Time: 06/04/2020 14:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: PATRON,A  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: UMBERGER

DATE/TIME OF TEST: 05/18/2020 01:25:23

---

Diagnostics Check		OK	01:18
Air Blank		0.000	01:19
Control Test		0.080	01:19
Air Blank		0.000	01:20
Subject Sample	# 1	0.281	01:21
Breath Volume		1.500	01:21
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.277	01:23
Breath Volume		2.097	01:23
Air Blank		0.000	01:24
Control Test		0.080	01:24
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

---

Data Download Date/Time: 06/04/2020 14:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: SULAIMAN,R  
ARRESTING OFFICER AGENCY: FHP TROOP C  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 05/21/2020 04:00:52

---

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.080	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.093	03:56
Breath Volume	2.246	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:58
Subject Sample # 2	0.087	03:59
Breath Volume	1.257	03:59
Air Blank	0.000	03:59
Control Test	0.079	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

---

Data Download Date/Time: 06/04/2020 14:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: HALL,L  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 05/21/2020 23:34:08

---

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.080	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.242	23:29
Breath Volume	1.875	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.231	23:32
Breath Volume	1.996	23:32
Air Blank	0.000	23:33
Control Test	0.079	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

---

Data Download Date/Time: 06/04/2020 14:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JENIFER,  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: PATRON,A  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: SUSENBACH

DATE/TIME OF TEST: 05/22/2020 03:40:30

---

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.129	03:36
Breath Volume	2.398	03:36
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.129	03:39
Breath Volume	2.457	03:39
Air Blank	0.000	03:39
Control Test	0.078	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

---

Data Download Date/Time: 06/04/2020 14:13



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: FISHER,T  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: FRITZ

DATE/TIME OF TEST: 05/22/2020 20:50:25

---

Diagnostics Check	OK	20:44
Air Blank	0.000	20:44
Control Test	0.080	20:44
Air Blank	0.000	20:45
Subject Sample # 1	0.096	20:45
Breath Volume	2.191	20:45
Air Blank	0.000	20:46
Air Blank	0.000	20:48
Subject Sample # 2	0.098	20:48
Breath Volume	2.433	20:48
Air Blank	0.000	20:49
Control Test	0.079	20:49
Air Blank	0.000	20:50
Diagnostics Check	OK	20:50

---

Data Download Date/Time: 06/04/2020 14:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: CLARK,I  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/23/2020 01:46:22

---

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.067	01:42
Breath Volume	2.972	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.066	01:44
Breath Volume	2.640	01:44
Air Blank	0.000	01:45
Control Test	0.079	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

---

Data Download Date/Time: 06/04/2020 14:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: TAMPA PD  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: LEGON ABREU

DATE/TIME OF TEST: 05/24/2020 01:35:39

---

Diagnostics Check		OK	01:29
Air Blank		0.000	01:29
Control Test		0.079	01:30
Air Blank		0.000	01:30
Subject Sample	# 1	0.000	01:31
Breath Volume		1.437	01:31
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.000	01:34
Breath Volume		1.613	01:34
Air Blank		0.000	01:34
Control Test		0.079	01:35
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

---

Data Download Date/Time: 06/04/2020 14:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: BURKE,B  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: DIAZ PEREZ

DATE/TIME OF TEST: 05/25/2020 23:18:13

---

Diagnostics Check		OK	23:04	
Air Blank		0.000	23:04	
Control Test		0.081	23:04	
Air Blank		0.000	23:05	
Subject Sample	# 1	0.149	23:08	Volume Not Met
Breath Volume		0.328	23:08	
Air Blank		0.000	23:09	
Air Blank		0.000	23:11	
Subject Sample	# 2	0.186	23:13	
Breath Volume		2.003	23:13	
Air Blank		0.000	23:14	
Air Blank		0.000	23:15	
Subject Sample	# 3	0.182	23:16	
Breath Volume		1.960	23:16	
Air Blank		0.000	23:17	
Control Test		0.077	23:17	
Air Blank		0.000	23:18	
Diagnostics Check		OK	23:18	

---

Data Download Date/Time: 06/04/2020 14:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: TAMPA PD  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: ARIZA QUINTANA

DATE/TIME OF TEST: 05/28/2020 03:36:43

---

Diagnostics Check		OK	03:28
Air Blank		0.000	03:29
Control Test		0.080	03:29
Air Blank		0.000	03:30
Subject Sample	# 1	0.089	03:32
Breath Volume		1.265	03:32
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.093	03:35
Breath Volume		1.707	03:35
Air Blank		0.000	03:35
Control Test		0.079	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

---

Data Download Date/Time: 06/04/2020 14:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: RODRIGUEZ,R  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: XOC PACAY

DATE/TIME OF TEST: 05/30/2020 00:06:07

---

Diagnostics Check		OK	00:00
Air Blank		0.000	00:00
Control Test		0.079	00:00
Air Blank		0.000	00:01
Subject Sample	# 1	0.175	00:01
Breath Volume		2.343	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.187	00:04
Breath Volume		3.082	00:04
Air Blank		0.000	00:05
Control Test		0.079	00:05
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

---

Data Download Date/Time: 06/04/2020 14:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: HALL,L  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: KUMAR

DATE/TIME OF TEST: 05/31/2020 05:15:40

---

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.080	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.174	05:11
Breath Volume	1.441	05:11
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.180	05:14
Breath Volume	2.253	05:14
Air Blank	0.000	05:14
Control Test	0.080	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

---

Data Download Date/Time: 06/04/2020 14:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: BINDER,BRONSON  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: BUI  
ARREST TIME: 16:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:43

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 05/23/2020 17:29:00

---

Diagnostics Check	OK	17:25	
Air Blank	0.000	17:26	
Control Test	0.078	17:26	
Air Blank	0.000	17:27	
Subject Sample	# 1	17:27	Subject Test Refused
Breath Volume	0.000	17:27	
Air Blank	0.000	17:28	
Control Test	0.078	17:28	
Air Blank	0.000	17:28	
Diagnostics Check	OK	17:29	

---

Data Download Date/Time: 06/24/2020 18:17



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WURST,JOSHUA,D  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: WURST,JOSHUA  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: BELLANDE

DATE/TIME OF TEST: 05/05/2020 21:36:23

---

Diagnostics Check	OK	21:29
Air Blank	0.000	21:30
Control Test	0.080	21:30
Air Blank	0.000	21:31
Subject Sample # 1	0.000	21:32
Breath Volume	2.531	21:32
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.000	21:35
Breath Volume	2.570	21:35
Air Blank	0.000	21:35
Control Test	0.080	21:35
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

---

Data Download Date/Time: 06/01/2020 20:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: WEBER,JEFFREY  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 05/16/2020 02:11:31

---

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.161	02:06
Breath Volume	2.546	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.161	02:10
Breath Volume	2.441	02:10
Air Blank	0.000	02:10
Control Test	0.080	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

---

Data Download Date/Time: 06/01/2020 20:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHERI,DIANE,M  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: JOHNSON,ROBERT  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 05/18/2020 22:31:31

---

Diagnostics Check	OK	22:23
Air Blank	0.000	22:23
Control Test	0.080	22:24
Air Blank	0.000	22:24
Subject Sample # 1	0.000	22:26
Breath Volume	1.808	22:26
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.000	22:30
Breath Volume	1.558	22:30
Air Blank	0.000	22:30
Control Test	0.081	22:30
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

---

Data Download Date/Time: 06/01/2020 20:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHERI,DIANE,M  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MCKEEBY,WADE  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: WILDER

DATE/TIME OF TEST: 05/20/2020 00:08:03

---

Diagnostics Check	OK	00:05	
Air Blank	0.000	00:05	
Control Test	0.079	00:06	
Air Blank	0.000	00:06	
Subject Sample	# 1	00:06	Subject Test Refused
Breath Volume	0.000	00:06	
Air Blank	0.000	00:07	
Control Test	0.080	00:07	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:08	

---

Data Download Date/Time: 06/01/2020 20:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSS,ROBERT,R  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MOSS,ROBERT  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 05/23/2020 05:50:09

---

Diagnostics Check	OK	05:43
Air Blank	0.000	05:44
Control Test	0.080	05:44
Air Blank	0.000	05:45
Subject Sample # 1	0.024	05:45
Breath Volume	1.960	05:45
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample # 2	0.024	05:48
Breath Volume	1.843	05:48
Air Blank	0.000	05:49
Control Test	0.079	05:49
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

---

Data Download Date/Time: 06/01/2020 20:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHERI,DIANE,M  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MCKEEBY,WADE  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: BUTTS

DATE/TIME OF TEST: 05/24/2020 01:07:15

---

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.079	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.091	01:02
Breath Volume	1.707	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.090	01:05
Breath Volume	1.710	01:05
Air Blank	0.000	01:06
Control Test	0.079	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

---

Data Download Date/Time: 06/01/2020 20:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WURST,JOSHUA,D  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: WURST,JOSHUA  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: POWERS-EATON

DATE/TIME OF TEST: 05/24/2020 16:30:31

---

Diagnostics Check	OK	16:24
Air Blank	0.000	16:24
Control Test	0.080	16:24
Air Blank	0.000	16:25
Subject Sample # 1	0.127	16:25
Breath Volume	1.746	16:25
Air Blank	0.000	16:26
Air Blank	0.000	16:28
Subject Sample # 2	0.132	16:29
Breath Volume	3.171	16:29
Air Blank	0.000	16:29
Control Test	0.080	16:29
Air Blank	0.000	16:30
Diagnostics Check	OK	16:30

---

Data Download Date/Time: 06/01/2020 20:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHERI,DIANE,M  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MACIAS,NICHOLAS  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 05/26/2020 03:38:41

---

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:36	
Control Test	0.080	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 1	03:37	Subject Test Refused
Breath Volume	0.000	03:37	
Air Blank	0.000	03:37	
Control Test	0.080	03:38	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

---

Data Download Date/Time: 06/01/2020 20:20



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSS,ROBERT,R  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MOSS,ROBERT  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 11:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:48

SUBJECT'S LAST NAME: JACKS

DATE/TIME OF TEST: 05/27/2020 12:14:56

---

Diagnostics Check	OK	12:08
Air Blank	0.000	12:09
Control Test	0.080	12:09
Air Blank	0.000	12:09
Subject Sample # 1	0.000	12:10
Breath Volume	2.101	12:10
Air Blank	0.000	12:11
Air Blank	0.000	12:13
Subject Sample # 2	0.000	12:13
Breath Volume	2.019	12:13
Air Blank	0.000	12:14
Control Test	0.080	12:14
Air Blank	0.000	12:14
Diagnostics Check	OK	12:14

---

Data Download Date/Time: 06/01/2020 20:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHERI,DIANE,M  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: CHERI,DIANE  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: DEVILLIER

DATE/TIME OF TEST: 05/28/2020 03:29:30

---

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.081	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.147	03:23
Breath Volume	2.718	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.148	03:28
Breath Volume	2.574	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

---

Data Download Date/Time: 06/01/2020 20:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSS,ROBERT,R  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MOSS,ROBERT  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:41

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 05/28/2020 06:09:53

---

Diagnostics Check	OK	06:01
Air Blank	0.000	06:02
Control Test	0.080	06:02
Air Blank	0.000	06:03
Subject Sample # 1	0.192	06:05
Breath Volume	1.414	06:05
Air Blank	0.000	06:05
Air Blank	0.000	06:07
Subject Sample # 2	0.193	06:08
Breath Volume	1.359	06:08
Air Blank	0.000	06:08
Control Test	0.080	06:09
Air Blank	0.000	06:09
Diagnostics Check	OK	06:09

---

Data Download Date/Time: 06/01/2020 20:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUSH,DANIEL,C  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: LOWE,ALEXANDER  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 05/04/2020 20:21:53

---

Diagnostics Check	OK	20:15
Air Blank	0.000	20:16
Control Test	0.080	20:16
Air Blank	0.000	20:16
Subject Sample # 1	0.000	20:17
Breath Volume	2.007	20:17
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample # 2	0.000	20:20
Breath Volume	1.546	20:20
Air Blank	0.000	20:21
Control Test	0.079	20:21
Air Blank	0.000	20:21
Diagnostics Check	OK	20:21

---

Data Download Date/Time: 05/07/2020 13:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLS,CHARLES,S  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MILLS,CHARLES  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: CAMERON

DATE/TIME OF TEST: 05/05/2020 21:50:45

---

Diagnostics Check	OK	21:42
Air Blank	0.000	21:43
Control Test	0.080	21:43
Air Blank	0.000	21:43
Subject Sample # 1	0.307	21:45
Breath Volume	1.453	21:45
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample # 2	0.290	21:49
Breath Volume	1.164	21:49
Air Blank	0.000	21:49
Control Test	0.079	21:50
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

---

Data Download Date/Time: 05/07/2020 13:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLS,CHARLES,S  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: LASALANDRA,MICHAEL  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/06/2020 01:25:36

---

Diagnostics Check	OK	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:20
Subject Sample # 1	0.008	01:21
Breath Volume	1.984	01:21
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.000	01:24
Breath Volume	1.324	01:24
Air Blank	0.000	01:24
Control Test	0.080	01:25
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

---

Data Download Date/Time: 05/07/2020 13:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OHERN,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: OHERN,ANDREW  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: LOBIANCO

DATE/TIME OF TEST: 05/07/2020 20:28:58

---

Diagnostics Check	OK	20:22
Air Blank	0.000	20:23
Control Test	0.080	20:23
Air Blank	0.000	20:23
Subject Sample # 1	0.000	20:24
Breath Volume	2.648	20:24
Air Blank	0.000	20:25
Air Blank	0.000	20:27
Subject Sample # 2	0.000	20:27
Breath Volume	2.910	20:27
Air Blank	0.000	20:28
Control Test	0.080	20:28
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

---

Data Download Date/Time: 06/03/2020 11:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MADURO,VICTOR,A  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MADURO,VICTOR  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 05/07/2020 22:14:48

---

Diagnostics Check	OK	22:07
Air Blank	0.000	22:07
Control Test	0.078	22:08
Air Blank	0.000	22:08
Subject Sample # 1	0.199	22:10
Breath Volume	2.171	22:10
Air Blank	0.000	22:10
Air Blank	0.000	22:12
Subject Sample # 2	0.199	22:13
Breath Volume	2.128	22:13
Air Blank	0.000	22:13
Control Test	0.079	22:14
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

---

Data Download Date/Time: 06/03/2020 11:12



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUSKEVICH,CURT,E  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: SHIPPY,D  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 05/08/2020 22:06:51

---

Diagnostics Check	OK	21:59
Air Blank	0.000	22:00
Control Test	0.080	22:00
Air Blank	0.000	22:01
Subject Sample # 1	0.000	22:02
Breath Volume	2.039	22:02
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.000	22:05
Breath Volume	2.535	22:05
Air Blank	0.000	22:05
Control Test	0.080	22:06
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

---

Data Download Date/Time: 06/03/2020 11:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: RAMOS-REDDEN

DATE/TIME OF TEST: 05/09/2020 03:46:18

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.000	03:41
Breath Volume	2.144	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.000	03:44
Breath Volume	1.199	03:44
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

---

Data Download Date/Time: 06/03/2020 11:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUSH,DANIEL,C  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: VELAZQUEZ,RICARDO  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: TRACEY

DATE/TIME OF TEST: 05/10/2020 00:05:53

---

Diagnostics Check		OK	23:59
Air Blank		0.000	23:59
Control Test		0.081	23:59
Air Blank		0.000	00:00
Subject Sample	# 1	0.201	00:01
Breath Volume		2.601	00:01
Air Blank		0.000	00:01
Air Blank		0.000	00:03
Subject Sample	# 2	0.204	00:04
Breath Volume		2.304	00:04
Air Blank		0.000	00:04
Control Test		0.078	00:05
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

---

Data Download Date/Time: 06/03/2020 11:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUSH,DANIEL,C  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: SHIPPY,DYLAN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MILAM

DATE/TIME OF TEST: 05/10/2020 00:47:20

---

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.080	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.000	00:43
Breath Volume	1.644	00:43
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.000	00:45
Breath Volume	4.199	00:45
Air Blank	0.000	00:46
Control Test	0.079	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

---

Data Download Date/Time: 06/03/2020 11:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SOUZA

DATE/TIME OF TEST: 05/10/2020 02:37:58

---

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.079	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.121	02:33
Breath Volume	1.910	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:35
Subject Sample # 2	0.115	02:36
Breath Volume	2.652	02:36
Air Blank	0.000	02:37
Control Test	0.079	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

---

Data Download Date/Time: 06/03/2020 11:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: GARLIN

DATE/TIME OF TEST: 05/11/2020 02:51:06

---

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.000	02:46
Breath Volume	1.312	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.000	02:49
Breath Volume	1.414	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

---

Data Download Date/Time: 06/03/2020 11:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: HOWES,JOHN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: BOWMAN

DATE/TIME OF TEST: 05/17/2020 21:29:07

---

Diagnostics Check	OK	21:22
Air Blank	0.000	21:23
Control Test	0.081	21:23
Air Blank	0.000	21:23
Subject Sample # 1	0.060	21:24
Breath Volume	1.625	21:24
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.059	21:27
Breath Volume	2.101	21:27
Air Blank	0.000	21:28
Control Test	0.080	21:28
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

---

Data Download Date/Time: 06/03/2020 11:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MEADOWS,KELSEY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: VELDE

DATE/TIME OF TEST: 05/17/2020 21:45:16

---

Diagnostics Check	OK	21:38
Air Blank	0.000	21:38
Control Test	0.080	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.369	21:40
Breath Volume	2.132	21:40
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.366	21:43
Breath Volume	1.902	21:43
Air Blank	0.000	21:44
Control Test	0.079	21:44
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

---

Data Download Date/Time: 06/03/2020 11:12



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUSKEVICH,CURT,E  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: PERIAM,C  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: TITUS

DATE/TIME OF TEST: 05/20/2020 04:54:08

---

Diagnostics Check	OK	04:47
Air Blank	0.000	04:48
Control Test	0.080	04:48
Air Blank	0.000	04:48
Subject Sample # 1	0.147	04:49
Breath Volume	1.902	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:51
Subject Sample # 2	0.128	04:52
Breath Volume	1.613	04:52
Air Blank	0.000	04:53
Control Test	0.080	04:53
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

---

Data Download Date/Time: 06/03/2020 11:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ENGH,WILLIAM,C  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: ROSS,C  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 05/21/2020 03:11:48

---

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:05
Subject Sample # 1	0.000	03:07
Breath Volume	2.890	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.000	03:10
Breath Volume	3.054	03:10
Air Blank	0.000	03:10
Control Test	0.080	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

---

Data Download Date/Time: 06/03/2020 11:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,DANNY,W  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: COOPER,NICHOLAS  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 14:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:49

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 05/22/2020 16:16:14

---

Diagnostics Check	OK	16:10
Air Blank	0.000	16:10
Control Test	0.080	16:10
Air Blank	0.000	16:11
Subject Sample # 1	0.000	16:11
Breath Volume	2.296	16:11
Air Blank	0.000	16:12
Air Blank	0.000	16:14
Subject Sample # 2	0.000	16:14
Breath Volume	1.871	16:14
Air Blank	0.000	16:15
Control Test	0.080	16:15
Air Blank	0.000	16:16
Diagnostics Check	OK	16:16

---

Data Download Date/Time: 06/03/2020 11:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUSH,DANIEL,C  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: LOWE,A  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: FENNELL

DATE/TIME OF TEST: 05/22/2020 23:53:11

---

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.080	23:48
Air Blank	0.000	23:48
Subject Sample # 1	0.009	23:49
Breath Volume	2.558	23:49
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.008	23:51
Breath Volume	2.695	23:51
Air Blank	0.000	23:52
Control Test	0.080	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

---

Data Download Date/Time: 06/03/2020 11:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUFFY,J,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: SWATSW,W  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 07:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: PENDER

DATE/TIME OF TEST: 05/24/2020 08:29:39

---

Diagnostics Check	OK	08:20
Air Blank	0.000	08:21
Control Test	0.080	08:21
Air Blank	0.000	08:22
Subject Sample # 1	0.082	08:24
Breath Volume	1.667	08:24
Air Blank	0.000	08:25
Air Blank	0.000	08:27
Subject Sample # 2	0.081	08:28
Breath Volume	3.207	08:28
Air Blank	0.000	08:28
Control Test	0.080	08:29
Air Blank	0.000	08:29
Diagnostics Check	OK	08:29

---

Data Download Date/Time: 06/03/2020 11:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: BEVELLE

DATE/TIME OF TEST: 05/24/2020 23:35:33

---

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.080	23:29
Air Blank	0.000	23:30
Subject Sample # 1	0.265	23:30
Breath Volume	2.953	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.255	23:34
Breath Volume	3.050	23:34
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

---

Data Download Date/Time: 06/03/2020 11:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,ALFONSO,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: COLWELL,CHASE  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: VILLAGOMEZ ORTEGA

DATE/TIME OF TEST: 05/25/2020 02:26:48

---

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.166	02:22
Breath Volume	1.714	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.171	02:25
Breath Volume	1.527	02:25
Air Blank	0.000	02:25
Control Test	0.080	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

---

Data Download Date/Time: 06/03/2020 11:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINI,RYAN,H  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: DOTSON,WILLIAM  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: SPICCI

DATE/TIME OF TEST: 05/25/2020 16:37:20

---

Diagnostics Check	OK	16:31
Air Blank	0.000	16:31
Control Test	0.081	16:32
Air Blank	0.000	16:32
Subject Sample # 1	0.000	16:33
Breath Volume	3.191	16:33
Air Blank	0.000	16:33
Air Blank	0.000	16:35
Subject Sample # 2	0.000	16:35
Breath Volume	3.164	16:35
Air Blank	0.000	16:36
Control Test	0.080	16:36
Air Blank	0.000	16:37
Diagnostics Check	OK	16:37

---

Data Download Date/Time: 06/03/2020 11:12



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MADURO,VICTOR,A  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: CARVALHO,ASHLEY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 05/26/2020 01:14:17

---

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.081	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.193	01:09
Breath Volume	2.867	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.181	01:12
Breath Volume	2.492	01:12
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

---

Data Download Date/Time: 06/03/2020 11:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYES,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: J250783832120

DATE/TIME OF TEST: 05/27/2020 23:20:48

---

Diagnostics Check	OK	23:16	
Air Blank	0.000	23:16	
Control Test	0.079	23:16	
Air Blank	0.000	23:17	
Subject Sample	# 1	23:19	Subject Test Refused
Breath Volume	0.000	23:19	
Air Blank	0.000	23:19	
Control Test	0.079	23:20	
Air Blank	0.000	23:20	
Diagnostics Check	OK	23:20	

---

Data Download Date/Time: 06/03/2020 11:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FILLAK, MICHAEL, O  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: ENGH, WILLIAM  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: DURRETT

DATE/TIME OF TEST: 05/29/2020 23:33:47

---

Diagnostics Check		OK	23:22	
Air Blank		0.000	23:22	
Control Test		0.081	23:23	
Air Blank		0.000	23:23	
Subject Sample	# 1	0.174	23:26	Volume Not Met
Breath Volume		0.812	23:26	
Air Blank		0.000	23:27	
Air Blank		0.000	23:29	
Subject Sample	# 2	0.181	23:32	Volume Not Met
Breath Volume		0.578	23:32	
Air Blank		0.000	23:32	
Control Test		0.078	23:33	
Air Blank		0.000	23:33	
Diagnostics Check		OK	23:33	

---

Data Download Date/Time: 06/03/2020 11:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUEHLER,DEP,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BUEHLER,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: CRIMINS

DATE/TIME OF TEST: 05/05/2020 21:56:21

---

Diagnostics Check	OK	21:49
Air Blank	0.000	21:49
Control Test	0.079	21:50
Air Blank	0.000	21:50
Subject Sample # 1	0.243	21:51
Breath Volume	1.882	21:51
Air Blank	0.000	21:52
Air Blank	0.000	21:53
Subject Sample # 2	0.228	21:54
Breath Volume	1.605	21:54
Air Blank	0.000	21:55
Control Test	0.078	21:55
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

---

Data Download Date/Time: 05/12/2020 07:52