

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOSUE,
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: PHILLIPS,R
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: JACOBOWITZ

DATE/TIME OF TEST: 05/04/2020 05:17:52

Diagnostics Check	OK	05:10
Air Blank	0.000	05:11
Control Test	0.082	05:11
Air Blank	0.000	05:12
Subject Sample # 1	0.070	05:12
Breath Volume	2.132	05:12
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.069	05:16
Breath Volume	1.851	05:16
Air Blank	0.000	05:16
Control Test	0.082	05:17
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MCDONNELL,BRANDON
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: COFFEY

DATE/TIME OF TEST: 05/08/2020 22:12:07

Diagnostics Check	OK	22:05
Air Blank	0.000	22:06
Control Test	0.080	22:06
Air Blank	0.000	22:07
Subject Sample # 1	0.059	22:07
Breath Volume	2.687	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.052	22:10
Breath Volume	1.359	22:10
Air Blank	0.000	22:11
Control Test	0.081	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ARGENTINE,;BRITTANY
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: HEADLEY

DATE/TIME OF TEST: 05/09/2020 00:33:02

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.164	00:28
Breath Volume	2.093	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.168	00:31
Breath Volume	2.007	00:31
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:33

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CANNADAY,ZACH
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 05/09/2020 22:24:18

Diagnostics Check	OK	22:17
Air Blank	0.000	22:18
Control Test	0.080	22:18
Air Blank	0.000	22:18
Subject Sample # 1	0.167	22:19
Breath Volume	1.839	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.185	22:22
Breath Volume	1.312	22:22
Air Blank	0.000	22:23
Control Test	0.081	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,JACOB,A
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: ROSS,JACOB
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: CHILDS

DATE/TIME OF TEST: 05/10/2020 06:04:11

Diagnostics Check	OK	05:57
Air Blank	0.000	05:58
Control Test	0.079	05:58
Air Blank	0.000	05:59
Subject Sample # 1	0.118	05:59
Breath Volume	2.179	05:59
Air Blank	0.000	06:00
Air Blank	0.000	06:02
Subject Sample # 2	0.110	06:02
Breath Volume	3.539	06:02
Air Blank	0.000	06:03
Control Test	0.080	06:03
Air Blank	0.000	06:04
Diagnostics Check	OK	06:04

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLBURN-MUNN,MEGAN,E
BREATH TEST OPERATOR AGENCY: Casselberry PD
ARRESTING OFFICER: BURK,MATTHEW
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: GIRONDA

DATE/TIME OF TEST: 05/11/2020 04:29:03

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:23
Subject Sample # 1	0.133	04:24
Breath Volume	1.363	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.146	04:27
Breath Volume	1.398	04:27
Air Blank	0.000	04:28
Control Test	0.081	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:29

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOSUE,
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: MCDUFFIE

DATE/TIME OF TEST: 05/11/2020 20:58:09

Diagnostics Check	OK	20:52
Air Blank	0.000	20:52
Control Test	0.081	20:52
Air Blank	0.000	20:53
Subject Sample # 1	0.000	20:53
Breath Volume	2.433	20:53
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.000	20:56
Breath Volume	1.988	20:56
Air Blank	0.000	20:57
Control Test	0.081	20:57
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOSUE,
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: KINKOPF,TIM
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: FERRUFINO

DATE/TIME OF TEST: 05/12/2020 06:28:20

Diagnostics Check	OK	06:20
Air Blank	0.000	06:21
Control Test	0.081	06:21
Air Blank	0.000	06:22
Subject Sample # 1	0.152	06:23
Breath Volume	1.503	06:23
Air Blank	0.000	06:23
Air Blank	0.000	06:25
Subject Sample # 2	0.150	06:26
Breath Volume	1.675	06:26
Air Blank	0.000	06:27
Control Test	0.081	06:27
Air Blank	0.000	06:28
Diagnostics Check	OK	06:28

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIDDLE,JAMES,A
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: STUBBS,MATHEW
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 05/14/2020 17:30:30

Diagnostics Check	OK	17:21
Air Blank	0.000	17:22
Control Test	0.081	17:22
Air Blank	0.000	17:23
Subject Sample # 1	0.180	17:23
Breath Volume	1.894	17:23
Air Blank	0.000	17:24
Air Blank	0.000	17:26
Subject Sample # 2	0.185	17:29
Breath Volume	1.734	17:29
Air Blank	0.000	17:29
Control Test	0.083	17:29
Air Blank	0.000	17:30
Diagnostics Check	OK	17:30

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOSUE,
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 05/15/2020 20:11:59

Diagnostics Check	OK	20:05
Air Blank	0.000	20:06
Control Test	0.081	20:06
Air Blank	0.000	20:06
Subject Sample # 1	0.000	20:07
Breath Volume	1.933	20:07
Air Blank	0.000	20:08
Air Blank	0.000	20:10
Subject Sample # 2	0.000	20:10
Breath Volume	1.691	20:10
Air Blank	0.000	20:11
Control Test	0.081	20:11
Air Blank	0.000	20:11
Diagnostics Check	OK	20:11

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,,JOSUE,
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: ZHIGUISACA HUMALA

DATE/TIME OF TEST: 05/17/2020 02:28:05

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.081	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.186	02:23
Breath Volume	1.988	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.183	02:26
Breath Volume	2.343	02:26
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUPA,KYLE,T
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: BOUTELLE,ADAM
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 12:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:53

SUBJECT'S LAST NAME: STEPP

DATE/TIME OF TEST: 05/17/2020 15:18:44

Diagnostics Check	OK	15:15	
Air Blank	0.000	15:16	
Control Test	0.081	15:16	
Air Blank	0.000	15:17	
Subject Sample	# 1	15:17	Subject Test Refused
Breath Volume	0.000	15:17	
Air Blank	0.000	15:17	
Control Test	0.081	15:18	
Air Blank	0.000	15:18	
Diagnostics Check	OK	15:18	

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KRUPA,KYLE,T

BREATH TEST OPERATOR AGENCY: Sanford PD

ARRESTING OFFICER: CUSIMANO,GERALD

ARRESTING OFFICER AGENCY: Sanford PD

LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 915310

DRY GAS EXPIRY DATE: 03/07/2021

VIOLATION CODE: DUI

ARREST TIME: 16:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: BAMACA LOPEZ

DATE/TIME OF TEST: 05/17/2020 17:49:39

Diagnostics Check	OK	17:48	
Air Blank	0.000	17:48	
Control Test	0.081	17:49	
Air Blank		17:49	RFI Detect
Air Blank		17:49	RFI Detect

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KRUPAK,KYLE,T

BREATH TEST OPERATOR AGENCY: Sanford PD

ARRESTING OFFICER: CUSIMANO,GERALD

ARRESTING OFFICER AGENCY: Sanford PD

LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 915310

DRY GAS EXPIRY DATE: 03/07/2021

VIOLATION CODE: DUI

ARREST TIME: 16:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: BAMACA LOPEZ

DATE/TIME OF TEST: 05/17/2020 17:52:30

Diagnostics Check	OK	17:51	
Air Blank	0.000	17:51	
Control Test	0.080	17:52	
Air Blank		17:52	RFI Detect
Air Blank		17:52	RFI Detect

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUPA,KYLE,T
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: CUSIMANO,GERALD
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: BAMACA LOPEZ

DATE/TIME OF TEST: 05/17/2020 17:56:14

Diagnostics Check	OK	17:53	
Air Blank	0.000	17:54	
Control Test	0.080	17:54	
Air Blank	0.000	17:55	
Subject Sample	# 1	17:55	RFI Detect
Breath Volume	0.000	17:55	
Air Blank		17:55	RFI Detect
Air Blank	0.000	17:56	

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUPA,KYLE,T
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: CUSIMANO,GERALD
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: BAMACA LOPEZ

DATE/TIME OF TEST: 05/17/2020 18:03:20

Diagnostics Check	OK	17:57	
Air Blank	0.000	17:57	
Control Test	0.081	17:58	
Air Blank	0.000	17:58	
Subject Sample	# 1 0.194	18:01	Volume Not Met
Breath Volume	0.093	18:01	
Air Blank		18:01	Improper Sample
Air Blank	0.000	18:02	
Control Test	0.082	18:02	
Air Blank	0.000	18:03	
Diagnostics Check	OK	18:03	

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUPA,KYLE,T
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: CUSIMANO,GERALD
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: BAMACA LOPEZ

DATE/TIME OF TEST: 05/17/2020 18:08:50

Diagnostics Check	OK	18:04	
Air Blank	0.000	18:04	
Control Test	0.079	18:05	
Air Blank	0.000	18:05	
Subject Sample	# 1	18:07	Subject Test Refused
Breath Volume	0.164	18:07	
Air Blank	0.000	18:07	
Control Test	0.080	18:08	
Air Blank	0.000	18:08	
Diagnostics Check	OK	18:08	

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOSUE,
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: BROWN,JOEL
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: BELCHEN

DATE/TIME OF TEST: 05/18/2020 05:27:30

Diagnostics Check	OK	05:21
Air Blank	0.000	05:21
Control Test	0.081	05:21
Air Blank	0.000	05:22
Subject Sample # 1	0.173	05:22
Breath Volume	1.882	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.182	05:26
Breath Volume	1.769	05:26
Air Blank	0.000	05:26
Control Test	0.082	05:26
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,JASON,B
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: DAUZAT,RYAN
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 05/22/2020 02:06:22

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.079	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.225	02:01
Breath Volume	2.101	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:03
Subject Sample # 2	0.223	02:04
Breath Volume	1.972	02:04
Air Blank	0.000	02:05
Control Test	0.078	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RODSUDALAH,BROOKS
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 05/23/2020 02:34:49

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.080	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.206	02:30
Breath Volume	1.542	02:30
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.210	02:33
Breath Volume	1.968	02:33
Air Blank	0.000	02:33
Control Test	0.080	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DASQUE,DARREN
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: CASTRO FEBUS

DATE/TIME OF TEST: 05/24/2020 01:35:46

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.080	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.045	01:30
Breath Volume	2.156	01:30
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.038	01:34
Breath Volume	1.324	01:34
Air Blank	0.000	01:34
Control Test	0.080	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLBURN-MUNN,MEGAN,E
BREATH TEST OPERATOR AGENCY: Casselberry PD
ARRESTING OFFICER: HAMILL,ANDREW
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 05/24/2020 05:48:39

Diagnostics Check	OK	05:42
Air Blank	0.000	05:43
Control Test	0.081	05:43
Air Blank	0.000	05:43
Subject Sample # 1	0.082	05:44
Breath Volume	1.398	05:44
Air Blank	0.000	05:44
Air Blank	0.000	05:46
Subject Sample # 2	0.082	05:47
Breath Volume	1.152	05:47
Air Blank	0.000	05:47
Control Test	0.081	05:48
Air Blank	0.000	05:48
Diagnostics Check	OK	05:48

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,ABEL,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: QUINTANA,ABEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: BERNARDO RAMOS

DATE/TIME OF TEST: 05/30/2020 02:00:03

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:54
Subject Sample # 1	0.195	01:55
Breath Volume	1.179	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.205	01:58
Breath Volume	1.714	01:58
Air Blank	0.000	01:59
Control Test	0.080	01:59
Air Blank	0.000	01:59
Diagnostics Check	OK	02:00

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,ABEL,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CASSATO,KAYLA
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 05/30/2020 02:16:31

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.080	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.145	02:12
Breath Volume	2.230	02:12
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.152	02:15
Breath Volume	2.015	02:15
Air Blank	0.000	02:15
Control Test	0.081	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOSUE,
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: VALLANCOURT

DATE/TIME OF TEST: 05/31/2020 03:20:41

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.081	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.136	03:15
Breath Volume	1.464	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.144	03:19
Breath Volume	1.425	03:19
Air Blank	0.000	03:19
Control Test	0.081	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 07/01/2020 22:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUDTKE,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SMITH,JEREMY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 06:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:03

SUBJECT'S LAST NAME: H400164930840

DATE/TIME OF TEST: 05/05/2020 09:18:06

Diagnostics Check	OK	09:06	
Air Blank	0.000	09:06	
Control Test	0.079	09:07	
Air Blank	0.000	09:07	
Subject Sample	# 1 0.065	09:07	
Breath Volume	1.109	09:07	
Air Blank	0.000	09:08	
Air Blank	0.000	09:10	
Subject Sample	# 2 0.061	09:13	Volume Not Met
Breath Volume	0.921	09:13	
Air Blank	0.000	09:14	
Air Blank	0.000	09:16	
Subject Sample	# 3 0.071	09:16	
Breath Volume	1.375	09:16	
Air Blank	0.000	09:17	
Control Test	0.078	09:17	
Air Blank	0.000	09:18	
Diagnostics Check	OK	09:18	

Data Download Date/Time: 05/14/2020 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBSEN,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: MONG

DATE/TIME OF TEST: 05/09/2020 01:11:21

Diagnostics Check	OK	01:08	
Air Blank	0.000	01:09	
Control Test	0.080	01:09	
Air Blank	0.000	01:09	
Subject Sample	# 1	01:09	Subject Test Refused
Breath Volume	0.000	01:09	
Air Blank	0.000	01:10	
Control Test	0.080	01:10	
Air Blank	0.000	01:11	
Diagnostics Check	OK	01:11	

Data Download Date/Time: 05/14/2020 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANGANARO,JESSICA,H
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MANGANARO,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/10/2020 02:55:56

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.155	02:51
Breath Volume	12.56	02:51
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.152	02:54
Breath Volume	13.61	02:54
Air Blank	0.000	02:55
Control Test	0.081	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 05/14/2020 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DUNNING,RYAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: HENDERSHOTT

DATE/TIME OF TEST: 05/11/2020 05:21:35

Diagnostics Check	OK	05:15
Air Blank	0.000	05:16
Control Test	0.081	05:16
Air Blank	0.000	05:16
Subject Sample # 1	0.197	05:17
Breath Volume	4.593	05:17
Air Blank	0.000	05:17
Air Blank	0.000	05:19
Subject Sample # 2	0.186	05:20
Breath Volume	3.570	05:20
Air Blank	0.000	05:20
Control Test	0.079	05:20
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 05/14/2020 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REFUSE, RONY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE, RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 12:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:32

SUBJECT'S LAST NAME: SHERMAN

DATE/TIME OF TEST: 05/29/2020 13:59:02

Diagnostics Check	OK	13:52
Air Blank	0.000	13:53
Control Test	0.080	13:53
Air Blank	0.000	13:54
Subject Sample # 1	0.000	13:54
Breath Volume	2.492	13:54
Air Blank	0.000	13:54
Air Blank	0.000	13:56
Subject Sample # 2	0.000	13:57
Breath Volume	3.582	13:57
Air Blank	0.000	13:58
Control Test	0.080	13:58
Air Blank	0.000	13:58
Diagnostics Check	OK	13:59

Data Download Date/Time: 06/04/2020 20:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: REFUSE,RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: QUITERIO-GALICIA

DATE/TIME OF TEST: 05/29/2020 21:39:24

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.080	21:34
Air Blank	0.000	21:34
Subject Sample # 1	0.103	21:34
Breath Volume	3.441	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.112	21:37
Breath Volume	4.171	21:37
Air Blank	0.000	21:38
Control Test	0.079	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 06/04/2020 20:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSEPH,A
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: SMITH,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: REYES TORRES

DATE/TIME OF TEST: 05/17/2020 21:35:00

Diagnostics Check	OK	21:28
Air Blank	0.000	21:29
Control Test	0.078	21:29
Air Blank	0.000	21:30
Subject Sample # 1	0.066	21:30
Breath Volume	1.339	21:30
Air Blank	0.000	21:31
Air Blank	0.000	21:33
Subject Sample # 2	0.063	21:33
Breath Volume	1.136	21:33
Air Blank	0.000	21:34
Control Test	0.078	21:34
Air Blank	0.000	21:34
Diagnostics Check	OK	21:35

Data Download Date/Time: 06/23/2020 23:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIDNER,CODY,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CODDING,JOSHUA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 05/19/2020 21:05:59

Diagnostics Check	OK	20:55	
Air Blank	0.000	20:55	
Control Test	0.078	20:56	
Air Blank	0.000	20:56	
Subject Sample	# 1 0.160	20:59	Volume Not Met
Breath Volume	0.183	20:59	
Air Blank	0.000	21:00	
Air Blank	0.000	21:02	
Subject Sample	# 2	21:04	Subject Test Refused
Breath Volume	0.457	21:04	
Air Blank	0.000	21:05	
Control Test	0.076	21:05	
Air Blank	0.000	21:05	
Diagnostics Check	OK	21:05	

Data Download Date/Time: 06/23/2020 23:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIALOBRZESKI,M,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BROOKS,N
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: YANES

DATE/TIME OF TEST: 05/22/2020 23:47:30

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.078	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.153	23:42
Breath Volume	2.613	23:42
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.166	23:46
Breath Volume	2.968	23:46
Air Blank	0.000	23:46
Control Test	0.079	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 06/23/2020 23:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIDNER,CODY,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: IRIZARRY,HEIDEN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: MONE

DATE/TIME OF TEST: 05/25/2020 05:07:51

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:58	
Control Test	0.078	04:58	
Air Blank	0.000	04:59	
Subject Sample	# 1 0.222	05:02	
Breath Volume	1.253	05:02	
Air Blank	0.000	05:03	
Air Blank	0.000	05:04	
Subject Sample	# 2	05:06	Subject Test Refused
Breath Volume	0.218	05:06	
Air Blank	0.000	05:06	
Control Test	0.077	05:07	
Air Blank	0.000	05:07	
Diagnostics Check	OK	05:07	

Data Download Date/Time: 06/23/2020 23:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAN, TYLER, M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: THOMAN, TYLER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: BRINK

DATE/TIME OF TEST: 05/25/2020 19:23:04

Diagnostics Check	OK	19:16
Air Blank	0.000	19:17
Control Test	0.078	19:17
Air Blank	0.000	19:18
Subject Sample # 1	0.265	19:18
Breath Volume	2.003	19:18
Air Blank	0.000	19:19
Air Blank	0.000	19:21
Subject Sample # 2	0.264	19:21
Breath Volume	2.468	19:21
Air Blank	0.000	19:22
Control Test	0.076	19:22
Air Blank	0.000	19:22
Diagnostics Check	OK	19:23

Data Download Date/Time: 06/23/2020 23:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIDNER,CODY,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CODDING,JOSHUA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 05/27/2020 23:08:53

Diagnostics Check	OK	23:01
Air Blank	0.000	23:02
Control Test	0.078	23:02
Air Blank	0.000	23:02
Subject Sample # 1	0.216	23:04
Breath Volume	1.695	23:04
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.227	23:07
Breath Volume	1.429	23:07
Air Blank	0.000	23:07
Control Test	0.076	23:08
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 06/23/2020 23:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LLEWELLYN,A,
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: SHEHAN,K
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 05/30/2020 00:27:58

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.078	00:22
Air Blank	0.000	00:23
Subject Sample # 1	0.185	00:23
Breath Volume	2.406	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.181	00:26
Breath Volume	1.566	00:26
Air Blank	0.000	00:27
Control Test	0.079	00:27
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 06/23/2020 23:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LLEWELLYN,A,
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: GIVANS,G
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: DARR

DATE/TIME OF TEST: 05/30/2020 23:05:02

Diagnostics Check	OK	22:53	
Air Blank	0.000	22:53	
Control Test	0.078	22:53	
Air Blank	0.000	22:54	
Subject Sample	# 1 0.278	22:54	
Breath Volume	2.300	22:54	
Air Blank	0.000	22:55	
Air Blank	0.000	22:57	
Subject Sample	# 2 0.256	23:00	Volume Not Met
Breath Volume	0.156	23:00	
Air Blank	0.000	23:01	
Air Blank	0.000	23:03	
Subject Sample	# 3 0.264	23:03	
Breath Volume	2.066	23:03	
Air Blank	0.000	23:04	
Control Test	0.077	23:04	
Air Blank	0.000	23:04	
Diagnostics Check	OK	23:05	

Data Download Date/Time: 06/23/2020 23:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSEPH,A
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: ALEPRETE,JOSHUA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: GERKEN

DATE/TIME OF TEST: 05/31/2020 21:51:38

Diagnostics Check	OK	21:45
Air Blank	0.000	21:46
Control Test	0.079	21:46
Air Blank	0.000	21:46
Subject Sample # 1	0.192	21:47
Breath Volume	1.753	21:47
Air Blank	0.000	21:47
Air Blank	0.000	21:49
Subject Sample # 2	0.195	21:50
Breath Volume	1.132	21:50
Air Blank	0.000	21:50
Control Test	0.077	21:51
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

Data Download Date/Time: 06/23/2020 23:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: EASTTY,MARK
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: ALBEE

DATE/TIME OF TEST: 05/01/2020 23:40:17

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.078	23:32
Air Blank	0.000	23:32
Subject Sample # 1	0.173	23:32
Breath Volume	2.507	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.146	23:35
Breath Volume	2.234	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 3	0.151	23:38
Breath Volume	1.964	23:38
Air Blank	0.000	23:39
Control Test	0.079	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 05/14/2020 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANGANARO,JESSICA,H
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MANGANARO,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: BENJAMIN

DATE/TIME OF TEST: 05/02/2020 04:03:41

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:01	
Control Test	0.080	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1	04:02	Subject Test Refused
Breath Volume	0.000	04:02	
Air Blank	0.000	04:02	
Control Test	0.080	04:03	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 05/14/2020 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,K,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CAREY,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: DONAHUE

DATE/TIME OF TEST: 05/02/2020 22:04:08

Diagnostics Check	OK	22:01	
Air Blank	0.000	22:01	
Control Test	0.080	22:01	
Air Blank	0.000	22:02	
Subject Sample	# 1	22:02	Subject Test Refused
Breath Volume	0.000	22:02	
Air Blank	0.000	22:03	
Control Test	0.080	22:03	
Air Blank	0.000	22:04	
Diagnostics Check	OK	22:04	

Data Download Date/Time: 05/14/2020 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMIN,N
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/02/2020 23:06:01

Diagnostics Check	OK	23:03	
Air Blank	0.000	23:03	
Control Test	0.080	23:03	
Air Blank	0.000	23:04	
Subject Sample	# 1	23:04	Subject Test Refused
Breath Volume	0.000	23:04	
Air Blank	0.000	23:05	
Control Test	0.080	23:05	
Air Blank	0.000	23:05	
Diagnostics Check	OK	23:06	

Data Download Date/Time: 05/14/2020 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANGANARO,JESSICA,H
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MANGANARO,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CHILDREY

DATE/TIME OF TEST: 05/07/2020 04:36:42

Diagnostics Check	OK	04:33	
Air Blank	0.000	04:34	
Control Test	0.081	04:34	
Air Blank	0.000	04:35	
Subject Sample	# 1	04:35	Subject Test Refused
Breath Volume	0.000	04:35	
Air Blank	0.000	04:35	
Control Test	0.082	04:36	
Air Blank	0.000	04:36	
Diagnostics Check	OK	04:36	

Data Download Date/Time: 05/14/2020 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RODRIGUEZ,NELSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: THIERJUNG

DATE/TIME OF TEST: 05/07/2020 23:26:39

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.079	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.087	23:22
Breath Volume	1.199	23:22
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.104	23:25
Breath Volume	2.976	23:25
Air Blank	0.000	23:25
Control Test	0.080	23:26
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 05/14/2020 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: BAKER JR

DATE/TIME OF TEST: 05/08/2020 22:02:20

Diagnostics Check	OK	21:53
Air Blank	0.000	21:54
Control Test	0.080	21:54
Air Blank	0.000	21:55
Subject Sample # 1	0.206	21:56
Breath Volume	1.570	21:56
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.226	22:00
Breath Volume	1.804	22:00
Air Blank	0.000	22:01
Control Test	0.079	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 05/14/2020 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IFFT,GARRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: IFFT,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: DONES

DATE/TIME OF TEST: 05/08/2020 23:55:07

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:52	
Control Test	0.080	23:53	
Air Blank	0.000	23:53	
Subject Sample	# 1	23:53	Subject Test Refused
Breath Volume	0.000	23:53	
Air Blank	0.000	23:54	
Control Test	0.079	23:54	
Air Blank	0.000	23:55	
Diagnostics Check	OK	23:55	

Data Download Date/Time: 05/14/2020 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IFFT,GARRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: IFFT,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: BOROKHOVA

DATE/TIME OF TEST: 05/10/2020 01:20:00

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:17	
Control Test	0.081	01:17	
Air Blank	0.000	01:18	
Subject Sample	# 1	01:18	Subject Test Refused
Breath Volume	0.000	01:18	
Air Blank	0.000	01:19	
Control Test	0.080	01:19	
Air Blank	0.000	01:19	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 05/14/2020 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTHINGTON,JUSTIN,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WORTHINGTON,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: FAIR

DATE/TIME OF TEST: 05/13/2020 00:50:15

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.080	00:48	
Air Blank	0.000	00:48	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:49	
Control Test	0.080	00:49	
Air Blank	0.000	00:50	
Diagnostics Check	OK	00:50	

Data Download Date/Time: 05/14/2020 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RODRIGUEZ,NELSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: TRIANA-DORTA

DATE/TIME OF TEST: 05/17/2020 02:05:57

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.080	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.136	02:01
Breath Volume	1.179	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:03
Subject Sample # 2	0.145	02:04
Breath Volume	1.800	02:04
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 06/04/2020 01:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RODRIGUEZ,NELSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: ZAVALETA-ROMAN

DATE/TIME OF TEST: 05/18/2020 00:09:07

Diagnostics Check	OK	00:01
Air Blank	0.000	00:02
Control Test	0.080	00:02
Air Blank	0.000	00:03
Subject Sample # 1	0.204	00:03
Breath Volume	1.285	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.199	00:07
Breath Volume	1.464	00:07
Air Blank	0.000	00:08
Control Test	0.081	00:08
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

Data Download Date/Time: 06/04/2020 01:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE, CORPORAL WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE, CORPORAL WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MCCREARY

DATE/TIME OF TEST: 05/18/2020 04:07:57

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.114	04:03
Breath Volume	3.515	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.109	04:06
Breath Volume	3.453	04:06
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 06/04/2020 01:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 12:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:14

SUBJECT'S LAST NAME: EVERETT

DATE/TIME OF TEST: 05/26/2020 13:51:20

Diagnostics Check	OK	13:45
Air Blank	0.000	13:45
Control Test	0.080	13:46
Air Blank	0.000	13:46
Subject Sample # 1	0.140	13:46
Breath Volume	2.500	13:46
Air Blank	0.000	13:47
Air Blank	0.000	13:49
Subject Sample # 2	0.138	13:49
Breath Volume	1.933	13:49
Air Blank	0.000	13:50
Control Test	0.080	13:50
Air Blank	0.000	13:51
Diagnostics Check	OK	13:51

Data Download Date/Time: 06/04/2020 01:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 15:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: SWAIN

DATE/TIME OF TEST: 05/26/2020 16:23:58

Diagnostics Check	OK	16:17
Air Blank	0.000	16:17
Control Test	0.080	16:17
Air Blank	0.000	16:18
Subject Sample # 1	0.000	16:19
Breath Volume	1.949	16:19
Air Blank	0.000	16:20
Air Blank	0.000	16:22
Subject Sample # 2	0.000	16:22
Breath Volume	1.667	16:22
Air Blank	0.000	16:23
Control Test	0.080	16:23
Air Blank	0.000	16:23
Diagnostics Check	OK	16:23

Data Download Date/Time: 06/04/2020 01:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE, CORPORAL W.,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE, CORPORAL W.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: HAHN

DATE/TIME OF TEST: 05/27/2020 00:19:27

Diagnostics Check	OK	00:12
Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:13
Subject Sample # 1	0.094	00:14
Breath Volume	1.332	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.098	00:17
Breath Volume	1.503	00:17
Air Blank	0.000	00:18
Control Test	0.081	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 06/04/2020 01:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARBANAS,SHANE,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ARBNAS,SHANE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: CATALDI

DATE/TIME OF TEST: 05/27/2020 21:42:35

Diagnostics Check	OK	21:35
Air Blank	0.000	21:36
Control Test	0.080	21:36
Air Blank	0.000	21:37
Subject Sample # 1	0.120	21:38
Breath Volume	1.410	21:38
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.122	21:41
Breath Volume	2.093	21:41
Air Blank	0.000	21:41
Control Test	0.080	21:41
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time: 06/04/2020 01:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOVEL,RYAN,M
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: DOVEL,RYAN
ARRESTING OFFICER AGENCY: Indialantic PD
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: KLINGER

DATE/TIME OF TEST: 05/05/2020 23:09:50

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.080	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.195	23:05
Breath Volume	2.679	23:05
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.184	23:08
Breath Volume	4.867	23:08
Air Blank	0.000	23:08
Control Test	0.078	23:09
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 05/30/2020 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: MILLAN DEL RIO

DATE/TIME OF TEST: 05/03/2020 22:49:02

Diagnostics Check	OK	22:42
Air Blank	0.000	22:43
Control Test	0.082	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.175	22:44
Breath Volume	1.523	22:44
Air Blank	0.000	22:45
Air Blank	0.000	22:46
Subject Sample # 2	0.169	22:47
Breath Volume	1.386	22:47
Air Blank	0.000	22:48
Control Test	0.084	22:48
Air Blank	0.000	22:48
Diagnostics Check	OK	22:49

Data Download Date/Time: 05/29/2020 07:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINET,RICKY,H
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BINET,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 15:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:34

SUBJECT'S LAST NAME: SELBY

DATE/TIME OF TEST: 05/11/2020 16:21:21

Diagnostics Check	OK	16:12
Air Blank	0.000	16:12
Control Test	0.083	16:12
Air Blank	0.000	16:13
Subject Sample # 1	0.346	16:13
Breath Volume	2.351	16:13
Air Blank	0.000	16:14
Air Blank	0.000	16:16
Subject Sample # 2	0.325	16:16
Breath Volume	2.316	16:16
Air Blank	0.000	16:17
Air Blank	0.000	16:19
Subject Sample # 3	0.345	16:19
Breath Volume	2.792	16:19
Air Blank	0.000	16:20
Control Test	0.076	16:20
Air Blank	0.000	16:21
Diagnostics Check	OK	16:21

Data Download Date/Time: 05/29/2020 07:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBOLI,MELANIE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: TRIMBOLI,MELANIE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: TEBEDO

DATE/TIME OF TEST: 05/04/2020 20:24:04

Diagnostics Check	OK	20:17
Air Blank	0.000	20:17
Control Test	0.080	20:18
Air Blank	0.000	20:18
Subject Sample # 1	0.000	20:19
Breath Volume	2.484	20:19
Air Blank	0.000	20:20
Air Blank	0.000	20:22
Subject Sample # 2	0.000	20:22
Breath Volume	1.671	20:22
Air Blank	0.000	20:23
Control Test	0.079	20:23
Air Blank	0.000	20:23
Diagnostics Check	OK	20:24

Data Download Date/Time: 05/28/2020 16:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HOLLCROFT,AMANDA,L

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: MEEHAN,ROBERT

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 07:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 06:24

SUBJECT'S LAST NAME: PORTWOOD

DATE/TIME OF TEST: 05/05/2020 08:23:48

Diagnostics Check	OK	08:17
Air Blank	0.000	08:17
Control Test	0.080	08:18
Air Blank	0.000	08:18
Subject Sample # 1	0.000	08:19
Breath Volume	1.273	08:19
Air Blank	0.000	08:19
Air Blank	0.000	08:21
Subject Sample # 2	0.000	08:22
Breath Volume	1.261	08:22
Air Blank	0.000	08:22
Control Test	0.080	08:23
Air Blank	0.000	08:23
Diagnostics Check	OK	08:23

Data Download Date/Time: 05/28/2020 16:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BORJAS,NOE,A

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: BERRY,VICTORIA

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 21:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: ARBAIZA

DATE/TIME OF TEST: 05/15/2020 22:49:31

Diagnostics Check	OK	22:42
Air Blank	0.000	22:43
Control Test	0.080	22:43
Air Blank	0.000	22:44
Subject Sample # 1	0.161	22:45
Breath Volume	2.144	22:45
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.153	22:47
Breath Volume	2.355	22:47
Air Blank	0.000	22:48
Control Test	0.079	22:48
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 05/28/2020 16:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCDONALD,COLE,M

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: MCDONALD,COLE

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 23:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 05/20/2020 00:27:51

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.080	00:21
Air Blank	0.000	00:22
Subject Sample # 1	0.239	00:23
Breath Volume	1.503	00:23
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.248	00:26
Breath Volume	1.667	00:26
Air Blank	0.000	00:26
Control Test	0.080	00:27
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 05/28/2020 16:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAMUELS,ANDREW,G
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: SAMUELS,ANDREW
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: HAMMOND

DATE/TIME OF TEST: 05/22/2020 21:43:35

Diagnostics Check	OK	21:39	
Air Blank	0.000	21:40	
Control Test	0.080	21:40	
Air Blank	0.000	21:40	
Subject Sample	# 1	21:42	Subject Test Refused
Breath Volume	0.000	21:42	
Air Blank	0.000	21:42	
Control Test	0.080	21:42	
Air Blank	0.000	21:43	
Diagnostics Check	OK	21:43	

Data Download Date/Time: 05/28/2020 16:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BORJAS,NOE,A

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: BERRY,VICTORIA

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 21:08

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MATTE

DATE/TIME OF TEST: 05/30/2020 21:56:52

Diagnostics Check	OK	21:50
Air Blank	0.000	21:50
Control Test	0.080	21:51
Air Blank	0.000	21:51
Subject Sample # 1	0.129	21:52
Breath Volume	1.308	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:54
Subject Sample # 2	0.145	21:55
Breath Volume	2.738	21:55
Air Blank	0.000	21:55
Control Test	0.080	21:56
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 06/28/2020 18:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/01/2020 00:05:24

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:02	
Control Test	0.079	00:03	
Air Blank	0.000	00:03	
Subject Sample	# 1	00:03	Subject Test Refused
Breath Volume	0.000	00:03	
Air Blank	0.000	00:04	
Control Test	0.079	00:04	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:05	

Data Download Date/Time: 05/14/2020 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON, C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 05/02/2020 02:37:51

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.078	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.168	02:33
Breath Volume	1.199	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.169	02:36
Breath Volume	1.906	02:36
Air Blank	0.000	02:36
Control Test	0.077	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 05/14/2020 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROE,B
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: VENTRE

DATE/TIME OF TEST: 05/02/2020 04:11:57

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:09	
Control Test	0.078	04:09	
Air Blank	0.000	04:10	
Subject Sample	# 1	04:10	Subject Test Refused
Breath Volume	0.000	04:10	
Air Blank	0.000	04:11	
Control Test	0.078	04:11	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:11	

Data Download Date/Time: 05/14/2020 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: NUSZKOWSKI

DATE/TIME OF TEST: 05/02/2020 20:13:30

Diagnostics Check	OK	20:10	
Air Blank	0.000	20:11	
Control Test	0.077	20:11	
Air Blank	0.000	20:11	
Subject Sample	# 1	20:12	Subject Test Refused
Breath Volume	0.000	20:12	
Air Blank	0.000	20:12	
Control Test	0.078	20:12	
Air Blank	0.000	20:13	
Diagnostics Check	OK	20:13	

Data Download Date/Time: 05/14/2020 08:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: NEWBERN

DATE/TIME OF TEST: 05/03/2020 01:45:49

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:43	
Control Test	0.078	01:43	
Air Blank	0.000	01:44	
Subject Sample	# 1	01:44	Subject Test Refused
Breath Volume	0.000	01:44	
Air Blank	0.000	01:44	
Control Test	0.078	01:45	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 05/14/2020 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PATTERSON, K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: PAGE

DATE/TIME OF TEST: 05/03/2020 05:29:15

Diagnostics Check	OK	05:22
Air Blank	0.000	05:23
Control Test	0.078	05:23
Air Blank	0.000	05:24
Subject Sample # 1	0.101	05:24
Breath Volume	3.343	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.094	05:27
Breath Volume	3.191	05:27
Air Blank	0.000	05:28
Control Test	0.078	05:28
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 05/14/2020 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOWELL,H
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: NORTHCUT

DATE/TIME OF TEST: 05/04/2020 02:52:50

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.078	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.173	02:48
Breath Volume	2.570	02:48
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.174	02:51
Breath Volume	2.558	02:51
Air Blank	0.000	02:51
Control Test	0.078	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 05/14/2020 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: GARAY UMANA

DATE/TIME OF TEST: 05/05/2020 03:26:41

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.078	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.000	03:22
Breath Volume	2.355	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.000	03:25
Breath Volume	1.175	03:25
Air Blank	0.000	03:25
Control Test	0.078	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 05/14/2020 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 05/05/2020 20:27:53

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.078	20:22
Air Blank	0.000	20:22
Subject Sample # 1	0.152	20:23
Breath Volume	2.214	20:23
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.148	20:26
Breath Volume	1.257	20:26
Air Blank	0.000	20:26
Control Test	0.077	20:27
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 05/14/2020 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: GADD

DATE/TIME OF TEST: 05/06/2020 05:43:57

Diagnostics Check	OK	05:40	
Air Blank	0.000	05:41	
Control Test	0.078	05:41	
Air Blank	0.000	05:42	
Subject Sample	# 1	05:42	Subject Test Refused
Breath Volume	0.000	05:42	
Air Blank	0.000	05:42	
Control Test	0.078	05:43	
Air Blank	0.000	05:43	
Diagnostics Check	OK	05:43	

Data Download Date/Time: 05/14/2020 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: BOUCHARD

DATE/TIME OF TEST: 05/06/2020 23:49:32

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:47	
Control Test	0.079	23:47	
Air Blank	0.000	23:47	
Subject Sample	# 1	23:48	Subject Test Refused
Breath Volume	0.000	23:48	
Air Blank	0.000	23:48	
Control Test	0.078	23:48	
Air Blank	0.000	23:49	
Diagnostics Check	OK	23:49	

Data Download Date/Time: 05/14/2020 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROE, B
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: LANDER

DATE/TIME OF TEST: 05/07/2020 03:27:54

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.079	03:23
Breath Volume	1.683	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.079	03:26
Breath Volume	2.914	03:26
Air Blank	0.000	03:26
Control Test	0.079	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 05/14/2020 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 05/08/2020 00:36:31

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.079	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.046	00:31
Breath Volume	2.355	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.043	00:34
Breath Volume	2.015	00:34
Air Blank	0.000	00:35
Control Test	0.079	00:35
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 05/14/2020 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 05/09/2020 20:46:53

Diagnostics Check	OK	20:44	
Air Blank	0.000	20:44	
Control Test	0.078	20:44	
Air Blank	0.000	20:45	
Subject Sample	# 1	20:45	Subject Test Refused
Breath Volume	0.000	20:45	
Air Blank	0.000	20:45	
Control Test	0.079	20:46	
Air Blank	0.000	20:46	
Diagnostics Check	OK	20:46	

Data Download Date/Time: 05/14/2020 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: LAWLER

DATE/TIME OF TEST: 05/10/2020 00:12:09

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.078	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.193	00:07
Breath Volume	1.613	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.182	00:10
Breath Volume	1.804	00:10
Air Blank	0.000	00:11
Control Test	0.078	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 05/14/2020 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON, C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CONARD

DATE/TIME OF TEST: 05/10/2020 02:16:51

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.079	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.000	02:12
Breath Volume	3.636	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.000	02:15
Breath Volume	3.937	02:15
Air Blank	0.000	02:15
Control Test	0.079	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 05/14/2020 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/11/2020 01:30:35

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.079	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.135	01:26
Breath Volume	1.644	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.136	01:28
Breath Volume	1.679	01:28
Air Blank	0.000	01:29
Control Test	0.078	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 05/14/2020 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: PARNELL

DATE/TIME OF TEST: 05/11/2020 04:25:21

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.078	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.118	04:20
Breath Volume	1.988	04:20
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.117	04:23
Breath Volume	1.949	04:23
Air Blank	0.000	04:24
Control Test	0.078	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 05/14/2020 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: KAMWANGA

DATE/TIME OF TEST: 05/11/2020 05:28:19

Diagnostics Check	OK	05:25	
Air Blank	0.000	05:25	
Control Test	0.079	05:25	
Air Blank	0.000	05:26	
Subject Sample	# 1	05:26	Subject Test Refused
Breath Volume	0.000	05:26	
Air Blank	0.000	05:27	
Control Test	0.078	05:27	
Air Blank		05:27	RFI Detect
Air Blank	0.000	05:28	

Data Download Date/Time: 05/14/2020 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: KAMWANGA

DATE/TIME OF TEST: 05/11/2020 05:33:18

Diagnostics Check	OK	05:30	
Air Blank	0.000	05:30	
Control Test	0.079	05:31	
Air Blank	0.000	05:31	
Subject Sample	# 1	05:31	Subject Test Refused
Breath Volume	0.000	05:31	
Air Blank	0.000	05:32	
Control Test	0.078	05:32	
Air Blank	0.000	05:33	
Diagnostics Check	OK	05:33	

Data Download Date/Time: 05/14/2020 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: REESOR

DATE/TIME OF TEST: 05/12/2020 02:48:49

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.000	02:44
Breath Volume	2.437	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.000	02:47
Breath Volume	2.527	02:47
Air Blank	0.000	02:47
Control Test	0.079	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 05/14/2020 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: AMADOR

DATE/TIME OF TEST: 05/13/2020 01:52:39

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.130	01:47
Breath Volume	3.578	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.121	01:50
Breath Volume	3.046	01:50
Air Blank	0.000	01:51
Control Test	0.079	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 05/14/2020 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WILLIAMS, E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: COURSEY

DATE/TIME OF TEST: 05/14/2020 01:50:38

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.079	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.136	01:46
Breath Volume	1.289	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.135	01:48
Breath Volume	2.101	01:48
Air Blank	0.000	01:49
Control Test	0.078	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 05/14/2020 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: TANNER

DATE/TIME OF TEST: 05/14/2020 05:43:53

Diagnostics Check	OK	05:38	
Air Blank	0.000	05:38	
Control Test	0.079	05:38	
Air Blank	0.000	05:39	
Subject Sample # 1	0.000	05:39	
Breath Volume	3.765	05:39	
Air Blank	0.000	05:40	
Air Blank	0.000	05:42	
		05:42	Improper Sample
Air Blank	0.000	05:42	
Control Test	0.079	05:43	
Air Blank	0.000	05:43	
Diagnostics Check	OK	05:43	

Data Download Date/Time: 05/14/2020 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: TANNER

DATE/TIME OF TEST: 05/14/2020 05:51:33

Diagnostics Check	OK	05:45
Air Blank	0.000	05:45
Control Test	0.079	05:46
Air Blank	0.000	05:46
Subject Sample # 1	0.000	05:47
Breath Volume	2.781	05:47
Air Blank	0.000	05:47
Air Blank	0.000	05:49
Subject Sample # 2	0.000	05:50
Breath Volume	3.167	05:50
Air Blank	0.000	05:50
Control Test	0.079	05:50
Air Blank	0.000	05:51
Diagnostics Check	OK	05:51

Data Download Date/Time: 05/14/2020 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUFF,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HUFF,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: COOL

DATE/TIME OF TEST: 05/06/2020 14:22:14

Diagnostics Check	OK	14:16
Air Blank	0.000	14:16
Control Test	0.078	14:16
Air Blank	0.000	14:17
Subject Sample # 1	0.000	14:17
Breath Volume	2.722	14:17
Air Blank	0.000	14:18
Air Blank	0.000	14:20
Subject Sample # 2	0.000	14:20
Breath Volume	1.902	14:20
Air Blank	0.000	14:21
Control Test	0.080	14:21
Air Blank	0.000	14:22
Diagnostics Check	OK	14:22

Data Download Date/Time: 06/24/2020 18:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: WHITMIRE,CHRISTAN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 05/10/2020 02:47:13

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.078	02:45	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:46	
Control Test	0.078	02:46	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 06/24/2020 18:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUCHENEY, WESLEY, A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: KOKAISEL, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:19

SUBJECT'S LAST NAME: HAMM

DATE/TIME OF TEST: 05/17/2020 13:45:52

Diagnostics Check	OK	13:39
Air Blank	0.000	13:40
Control Test	0.078	13:40
Air Blank	0.000	13:41
Subject Sample # 1	0.000	13:41
Breath Volume	3.675	13:41
Air Blank	0.000	13:42
Air Blank	0.000	13:44
Subject Sample # 2	0.000	13:44
Breath Volume	3.527	13:44
Air Blank	0.000	13:44
Control Test	0.079	13:45
Air Blank	0.000	13:45
Diagnostics Check	OK	13:45

Data Download Date/Time: 06/24/2020 18:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUFF,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HUFF,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:32

SUBJECT'S LAST NAME: KENDRICKS

DATE/TIME OF TEST: 05/25/2020 18:00:04

Diagnostics Check	OK	17:53
Air Blank	0.000	17:53
Control Test	0.078	17:54
Air Blank	0.000	17:54
Subject Sample # 1	0.175	17:55
Breath Volume	2.355	17:55
Air Blank	0.000	17:56
Air Blank	0.000	17:57
Subject Sample # 2	0.164	17:58
Breath Volume	1.812	17:58
Air Blank	0.000	17:59
Control Test	0.078	17:59
Air Blank	0.000	17:59
Diagnostics Check	OK	18:00

Data Download Date/Time: 06/24/2020 18:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGOODIAL,TRISTON,T
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: VREEKEN,RANDEN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: GABRUS

DATE/TIME OF TEST: 05/08/2020 18:51:55

Diagnostics Check	OK	18:45
Air Blank	0.000	18:45
Control Test	0.078	18:46
Air Blank	0.000	18:46
Subject Sample # 1	0.000	18:47
Breath Volume	4.015	18:47
Air Blank	0.000	18:47
Air Blank	0.000	18:49
Subject Sample # 2	0.000	18:50
Breath Volume	3.511	18:50
Air Blank	0.000	18:51
Control Test	0.078	18:51
Air Blank	0.000	18:51
Diagnostics Check	OK	18:51

Data Download Date/Time: 05/31/2020 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGOODIAL,TRISTON,T
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GODINEZ,JORGE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: DENTON

DATE/TIME OF TEST: 05/09/2020 03:47:09

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:44	
Control Test	0.079	03:45	
Air Blank	0.000	03:45	
Subject Sample	# 1	03:45	Subject Test Refused
Breath Volume	0.000	03:45	
Air Blank	0.000	03:46	
Control Test	0.079	03:46	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

Data Download Date/Time: 05/31/2020 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGOODIAL,TRISTON,T
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: DANLEY,DANIEL
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: GOODMAN

DATE/TIME OF TEST: 05/09/2020 18:32:58

Diagnostics Check	OK	18:29	
Air Blank	0.000	18:30	
Control Test	0.079	18:30	
Air Blank	0.000	18:30	
Subject Sample	# 1	18:31	Subject Test Refused
Breath Volume	0.000	18:31	
Air Blank	0.000	18:32	
Control Test	0.078	18:32	
Air Blank	0.000	18:32	
Diagnostics Check	OK	18:32	

Data Download Date/Time: 05/31/2020 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 05/02/2020 02:44:19

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.077	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.128	02:39
Breath Volume	1.289	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.127	02:42
Breath Volume	1.765	02:42
Air Blank	0.000	02:43
Control Test	0.077	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 05/14/2020 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 05/02/2020 04:44:32

Diagnostics Check	OK	04:41	
Air Blank	0.000	04:42	
Control Test	0.078	04:42	
Air Blank	0.000	04:42	
Subject Sample	# 1	04:43	Subject Test Refused
Breath Volume	0.000	04:43	
Air Blank	0.000	04:43	
Control Test	0.078	04:43	
Air Blank	0.000	04:44	
Diagnostics Check	OK	04:44	

Data Download Date/Time: 05/14/2020 11:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SANDIFORD JR,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: STEWARD

DATE/TIME OF TEST: 05/03/2020 01:12:30

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.078	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.108	01:07
Breath Volume	1.984	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.116	01:10
Breath Volume	1.472	01:10
Air Blank	0.000	01:11
Control Test	0.078	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 05/14/2020 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: WOOLEY

DATE/TIME OF TEST: 05/03/2020 03:09:44

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:07	
Control Test	0.077	03:07	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:08	
Control Test	0.077	03:09	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

Data Download Date/Time: 05/14/2020 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 05/05/2020 23:57:13

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.078	23:49
Air Blank	0.000	23:50
Subject Sample # 1	0.149	23:52
Breath Volume	2.210	23:52
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.132	23:55
Breath Volume	1.101	23:55
Air Blank	0.000	23:56
Control Test	0.076	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 05/14/2020 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS, TYLER, K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: CHEEK

DATE/TIME OF TEST: 05/14/2020 23:33:12

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.078	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.061	23:28
Breath Volume	1.296	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.063	23:31
Breath Volume	1.257	23:31
Air Blank	0.000	23:32
Control Test	0.078	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 05/27/2020 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS, TYLER, K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 05/16/2020 00:51:37

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.078	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.000	00:47
Breath Volume	1.496	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.000	00:50
Breath Volume	1.558	00:50
Air Blank	0.000	00:50
Control Test	0.077	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 05/27/2020 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS, TYLER, K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN, C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: MALONEY

DATE/TIME OF TEST: 05/16/2020 02:11:31

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.077	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.095	02:06
Breath Volume	2.218	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.098	02:09
Breath Volume	2.296	02:09
Air Blank	0.000	02:10
Control Test	0.077	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 05/27/2020 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS, TYLER, K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SORRELLS, M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: SALEM

DATE/TIME OF TEST: 05/17/2020 02:49:18

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.000	02:44
Breath Volume	2.792	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.000	02:47
Breath Volume	2.687	02:47
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 05/27/2020 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS, TYLER, K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HESS, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 05/17/2020 05:21:34

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:13	
Control Test	0.079	05:13	
Air Blank	0.000	05:14	
Subject Sample	# 1 0.201	05:17	
Breath Volume	1.164	05:17	
Air Blank	0.000	05:18	
Air Blank	0.000	05:19	
Subject Sample	# 2	05:20	Subject Test Refused
Breath Volume	0.000	05:20	
Air Blank	0.000	05:20	
Control Test	0.076	05:20	
Air Blank	0.000	05:21	
Diagnostics Check	OK	05:21	

Data Download Date/Time: 05/27/2020 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS, TYLER, K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 05/17/2020 22:02:51

Diagnostics Check	OK	21:59	
Air Blank	0.000	22:00	
Control Test	0.079	22:00	
Air Blank	0.000	22:01	
Subject Sample	# 1	22:01	Subject Test Refused
Breath Volume	0.000	22:01	
Air Blank	0.000	22:01	
Control Test	0.078	22:02	
Air Blank	0.000	22:02	
Diagnostics Check	OK	22:02	

Data Download Date/Time: 05/27/2020 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS, TYLER, K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 05/18/2020 02:48:25

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.078	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.161	02:43
Breath Volume	1.500	02:43
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.156	02:46
Breath Volume	1.191	02:46
Air Blank	0.000	02:47
Control Test	0.077	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 05/27/2020 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: LARGENT

DATE/TIME OF TEST: 05/18/2020 23:11:08

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.078	23:05
Air Blank	0.000	23:06
Subject Sample # 1	0.000	23:06
Breath Volume	3.082	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.000	23:09
Breath Volume	2.910	23:09
Air Blank	0.000	23:10
Control Test	0.078	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 05/27/2020 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 05/20/2020 01:47:16

Diagnostics Check	OK	01:44	
Air Blank	0.000	01:44	
Control Test	0.079	01:45	
Air Blank	0.000	01:45	
Subject Sample	# 1	01:45	Subject Test Refused
Breath Volume	0.000	01:45	
Air Blank	0.000	01:46	
Control Test	0.078	01:46	
Air Blank	0.000	01:47	
Diagnostics Check	OK	01:47	

Data Download Date/Time: 05/27/2020 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: NORRISH

DATE/TIME OF TEST: 05/21/2020 00:45:48

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.079	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.175	00:41
Breath Volume	1.511	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:43
Subject Sample # 2	0.176	00:44
Breath Volume	1.132	00:44
Air Blank	0.000	00:44
Control Test	0.078	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 05/27/2020 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 05/22/2020 01:43:28

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.078	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.088	01:38
Breath Volume	1.269	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:40
Subject Sample # 2	0.084	01:41
Breath Volume	1.207	01:41
Air Blank	0.000	01:42
Control Test	0.078	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 05/27/2020 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 05/23/2020 00:02:54

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.078	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.157	23:58
Breath Volume	1.226	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.162	00:01
Breath Volume	1.292	00:01
Air Blank	0.000	00:01
Control Test	0.076	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 05/27/2020 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 05/23/2020 05:24:05

Diagnostics Check	OK	05:12	
Air Blank	0.000	05:12	
Control Test	0.079	05:13	
Air Blank	0.000	05:13	
Subject Sample	# 1 0.014	05:16	Volume Not Met
Breath Volume	1.003	05:16	
Air Blank	0.000	05:17	
Air Blank	0.000	05:19	
Subject Sample	# 2 0.015	05:19	
Breath Volume	1.484	05:19	
Air Blank	0.000	05:20	
Air Blank	0.000	05:22	
Subject Sample	# 3 0.015	05:22	
Breath Volume	1.226	05:22	
Air Blank	0.000	05:23	
Control Test	0.078	05:23	
Air Blank	0.000	05:24	
Diagnostics Check	OK	05:24	

Data Download Date/Time: 05/27/2020 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS, TYLER, K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/23/2020 22:50:14

Diagnostics Check	OK	22:43
Air Blank	0.000	22:44
Control Test	0.078	22:44
Air Blank	0.000	22:45
Subject Sample # 1	0.049	22:45
Breath Volume	1.726	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.047	22:48
Breath Volume	1.945	22:48
Air Blank	0.000	22:49
Control Test	0.078	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 05/27/2020 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS, TYLER, K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 05/24/2020 01:30:18

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.078	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.185	01:25
Breath Volume	1.601	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:27
Subject Sample # 2	0.173	01:28
Breath Volume	1.507	01:28
Air Blank	0.000	01:29
Control Test	0.076	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 05/27/2020 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS, TYLER, K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ, M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: MESTER

DATE/TIME OF TEST: 05/24/2020 02:30:13

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.000	02:25
Breath Volume	3.617	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.000	02:28
Breath Volume	3.308	02:28
Air Blank	0.000	02:29
Control Test	0.078	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 05/27/2020 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS, TYLER, K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SNIDER, C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: KOTHMANN

DATE/TIME OF TEST: 05/24/2020 04:34:26

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:32	
Control Test	0.078	04:32	
Air Blank	0.000	04:32	
Subject Sample	# 1	04:33	Subject Test Refused
Breath Volume	0.000	04:33	
Air Blank	0.000	04:33	
Control Test	0.077	04:33	
Air Blank	0.000	04:34	
Diagnostics Check	OK	04:34	

Data Download Date/Time: 05/27/2020 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS, TYLER, K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SULLIVAN, C
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: SIMONS

DATE/TIME OF TEST: 05/24/2020 20:05:07

Diagnostics Check	OK	19:58
Air Blank	0.000	19:58
Control Test	0.078	19:59
Air Blank	0.000	19:59
Subject Sample # 1	0.084	20:00
Breath Volume	2.738	20:00
Air Blank	0.000	20:01
Air Blank	0.000	20:02
Subject Sample # 2	0.084	20:03
Breath Volume	2.527	20:03
Air Blank	0.000	20:04
Control Test	0.077	20:04
Air Blank	0.000	20:05
Diagnostics Check	OK	20:05

Data Download Date/Time: 05/27/2020 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS, TYLER, K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ, M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: VELASCO PEREZ

DATE/TIME OF TEST: 05/24/2020 23:01:11

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:58	
Control Test	0.077	22:59	
Air Blank	0.000	22:59	
Subject Sample	# 1	22:59	Subject Test Refused
Breath Volume	0.000	22:59	
Air Blank	0.000	23:00	
Control Test	0.077	23:00	
Air Blank	0.000	23:01	
Diagnostics Check	OK	23:01	

Data Download Date/Time: 05/27/2020 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SANDIFORD,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: MANICK

DATE/TIME OF TEST: 05/29/2020 06:06:00

Diagnostics Check	OK	06:03	
Air Blank	0.000	06:03	
Control Test	0.079	06:03	
Air Blank	0.000	06:04	
Subject Sample	# 1	06:04	Subject Test Refused
Breath Volume	0.000	06:04	
Air Blank	0.000	06:05	
Control Test	0.078	06:05	
Air Blank	0.000	06:05	
Diagnostics Check	OK	06:06	

Data Download Date/Time: 06/17/2020 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAYE, MAKINA, A
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: BURGOS, JULIAN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 035119080A
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: MADORMA

DATE/TIME OF TEST: 05/09/2020 23:19:56

Diagnostics Check	OK	23:12
Air Blank	0.000	23:12
Control Test	0.078	23:13
Air Blank	0.000	23:13
Subject Sample # 1	0.193	23:14
Breath Volume	1.644	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.205	23:18
Breath Volume	1.960	23:18
Air Blank	0.000	23:19
Control Test	0.077	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 05/28/2020 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,B,A
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: MOORE,A
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 035119080A
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:04

SUBJECT'S LAST NAME: PONTON

DATE/TIME OF TEST: 05/10/2020 09:41:13

Diagnostics Check	OK	09:34
Air Blank	0.000	09:34
Control Test	0.079	09:35
Air Blank	0.000	09:35
Subject Sample # 1	0.132	09:36
Breath Volume	2.953	09:36
Air Blank	0.000	09:37
Air Blank	0.000	09:39
Subject Sample # 2	0.134	09:39
Breath Volume	1.890	09:39
Air Blank	0.000	09:40
Control Test	0.077	09:40
Air Blank	0.000	09:41
Diagnostics Check	OK	09:41

Data Download Date/Time: 05/28/2020 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WETHERINGTON,STEPHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 035119080A
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: RUBIO-SALAZAR

DATE/TIME OF TEST: 05/29/2020 22:57:24

Diagnostics Check	OK	22:51	
Air Blank	0.000	22:51	
Control Test	0.078	22:51	
Air Blank	0.000	22:52	
Subject Sample # 1	0.201	22:53	
Breath Volume	2.839	22:53	
Air Blank	0.000	22:54	
Air Blank		22:56	RFI Detect
Air Blank	0.000	22:56	
Control Test	0.080	22:56	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

Data Download Date/Time: 06/29/2020 14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WETHERINGTON,STEPHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 035119080A
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: RUBIO-SALAZAR

DATE/TIME OF TEST: 05/29/2020 23:05:59

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.079	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.196	23:01
Breath Volume	2.695	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.191	23:04
Breath Volume	2.398	23:04
Air Blank	0.000	23:05
Control Test	0.079	23:05
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 06/29/2020 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERON,WINSTON,G
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BOWEN,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: GAHRIS

DATE/TIME OF TEST: 05/06/2020 21:18:12

Diagnostics Check	OK	21:11
Air Blank	0.000	21:11
Control Test	0.080	21:11
Air Blank	0.000	21:12
Subject Sample # 1	0.182	21:13
Breath Volume	2.644	21:13
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.189	21:16
Breath Volume	2.820	21:16
Air Blank	0.000	21:17
Control Test	0.081	21:17
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

Data Download Date/Time: 05/30/2020 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIPPINS,C,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: BATTEN

DATE/TIME OF TEST: 05/09/2020 19:36:08

Diagnostics Check	OK	19:33	
Air Blank	0.000	19:33	
Control Test	0.081	19:33	
Air Blank	0.000	19:34	
Subject Sample	# 1	19:34	Subject Test Refused
Breath Volume	0.000	19:34	
Air Blank	0.000	19:35	
Control Test	0.082	19:35	
Air Blank	0.000	19:36	
Diagnostics Check	OK	19:36	

Data Download Date/Time: 05/30/2020 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KUTZ,T,E

BREATH TEST OPERATOR AGENCY: Nassau County SO

ARRESTING OFFICER: KNIGHT,G

ARRESTING OFFICER AGENCY: Nassau County SO

LAST AGENCY INSPECTION: 04/27/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30219080A3

DRY GAS EXPIRY DATE: 01/05/2022

VIOLATION CODE: DUI

ARREST TIME: 01:53

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 05/11/2020 03:28:14

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.081	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.053	03:23
Breath Volume	1.710	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.049	03:26
Breath Volume	1.125	03:26
Air Blank	0.000	03:27
Control Test	0.081	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 05/30/2020 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,J,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KNIGHT,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: SOBOLESKI

DATE/TIME OF TEST: 05/13/2020 23:24:15

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:21	
Control Test	0.082	23:21	
Air Blank	0.000	23:21	
Subject Sample # 1	0.000	23:22	
Breath Volume	2.042	23:22	
Air Blank		23:22	Improper Sample
Air Blank	0.000	23:23	
Control Test	0.081	23:23	
Air Blank	0.000	23:24	
Diagnostics Check	OK	23:24	

Data Download Date/Time: 05/30/2020 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,J,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KNIGHT,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: SOBOLESKI

DATE/TIME OF TEST: 05/13/2020 23:38:29

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.081	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.000	23:34
Breath Volume	2.074	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.000	23:37
Breath Volume	3.218	23:37
Air Blank	0.000	23:37
Control Test	0.081	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 05/30/2020 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERON, WINSTON, G
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WOODS, W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: SWEAT

DATE/TIME OF TEST: 05/16/2020 23:50:37

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.081	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.173	23:45
Breath Volume	3.070	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.165	23:49
Breath Volume	3.089	23:49
Air Blank	0.000	23:49
Control Test	0.081	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 05/30/2020 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HERON,WINSTON,G

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: PATTERSON,K

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 04/27/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30219080A3

DRY GAS EXPIRY DATE: 01/05/2022

VIOLATION CODE: DUI

ARREST TIME: 00:58

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: ELCHERT

DATE/TIME OF TEST: 05/17/2020 02:39:49

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.081	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.114	02:35
Breath Volume	4.097	02:35
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.113	02:38
Breath Volume	3.429	02:38
Air Blank	0.000	02:38
Control Test	0.081	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 05/30/2020 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERON,WINSTON,G
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: ZITTROWER,T
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: BRADSHAW

DATE/TIME OF TEST: 05/17/2020 03:21:30

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:14	
Control Test	0.081	03:15	
Air Blank	0.000	03:15	
Subject Sample	# 1 0.122	03:16	
Breath Volume	2.046	03:16	
Air Blank	0.000	03:17	
Air Blank	0.000	03:19	
Subject Sample	# 2	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:20	
Control Test	0.081	03:20	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 05/30/2020 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,J,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KNIGHT,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: CLAYTON

DATE/TIME OF TEST: 05/18/2020 23:28:40

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:26	
Control Test	0.080	23:26	
Air Blank	0.000	23:27	
Subject Sample	# 1	23:27	Subject Test Refused
Breath Volume	0.000	23:27	
Air Blank	0.000	23:27	
Control Test	0.080	23:28	
Air Blank	0.000	23:28	
Diagnostics Check	OK	23:28	

Data Download Date/Time: 05/30/2020 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: VELTRI

DATE/TIME OF TEST: 05/19/2020 00:40:27

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:38	
Control Test	0.081	00:38	
Air Blank	0.000	00:38	
Subject Sample	# 1	00:39	Subject Test Refused
Breath Volume	0.000	00:39	
Air Blank	0.000	00:39	
Control Test	0.081	00:39	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:40	

Data Download Date/Time: 05/30/2020 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,D,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUFFMON,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/21/2020 04:01:59

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.081	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.222	03:57
Breath Volume	1.695	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.236	04:00
Breath Volume	1.742	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 05/30/2020 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,J,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 05/22/2020 23:29:11

Diagnostics Check	OK	23:22
Air Blank	0.000	23:23
Control Test	0.081	23:23
Air Blank	0.000	23:24
Subject Sample # 1	0.094	23:24
Breath Volume	4.261	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.095	23:27
Breath Volume	4.187	23:27
Air Blank	0.000	23:28
Control Test	0.081	23:28
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 05/30/2020 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,J,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KNIGHT,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 05/23/2020 01:56:57

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.127	01:52
Breath Volume	1.750	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.126	01:55
Breath Volume	2.207	01:55
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 05/30/2020 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: WHITE III

DATE/TIME OF TEST: 05/24/2020 03:12:24

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.082	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.183	03:07
Breath Volume	1.707	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:09
Subject Sample # 2	0.177	03:10
Breath Volume	1.171	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 05/30/2020 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: PADGETT III

DATE/TIME OF TEST: 05/24/2020 22:37:42

Diagnostics Check	OK	22:29
Air Blank	0.000	22:29
Control Test	0.082	22:30
Air Blank	0.000	22:30
Subject Sample # 1	0.106	22:32
Breath Volume	1.609	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.106	22:36
Breath Volume	1.238	22:36
Air Blank	0.000	22:36
Control Test	0.080	22:37
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 05/30/2020 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,J,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: JOHNSON,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 04/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/25/2020 00:38:35

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.081	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.212	00:34
Breath Volume	2.679	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.213	00:37
Breath Volume	3.375	00:37
Air Blank	0.000	00:37
Control Test	0.081	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 05/30/2020 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KUTZ,T,E

BREATH TEST OPERATOR AGENCY: Nassau County SO

ARRESTING OFFICER: KNIGHT,G

ARRESTING OFFICER AGENCY: Nassau County SO

LAST AGENCY INSPECTION: 04/27/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30219080A3

DRY GAS EXPIRY DATE: 01/05/2022

VIOLATION CODE: DUI

ARREST TIME: 18:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: EDGY

DATE/TIME OF TEST: 05/28/2020 19:48:56

Diagnostics Check	OK	19:43
Air Blank	0.000	19:43
Control Test	0.081	19:43
Air Blank	0.000	19:44
Subject Sample # 1	0.000	19:44
Breath Volume	1.753	19:44
Air Blank	0.000	19:45
Air Blank	0.000	19:47
Subject Sample # 2	0.000	19:47
Breath Volume	2.070	19:47
Air Blank	0.000	19:48
Control Test	0.081	19:48
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 05/30/2020 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERON,WINSTON,G
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: CENSON,T
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: VINES

DATE/TIME OF TEST: 05/29/2020 18:28:48

Diagnostics Check	OK	18:26	
Air Blank	0.000	18:26	
Control Test	0.081	18:26	
Air Blank	0.000	18:27	
Subject Sample	# 1	18:27	Subject Test Refused
Breath Volume	0.000	18:27	
Air Blank	0.000	18:27	
Control Test	0.081	18:28	
Air Blank	0.000	18:28	
Diagnostics Check	OK	18:28	

Data Download Date/Time: 05/30/2020 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERON,WINSTON,G
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MANDRICK,T
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: GALVEZ MANRIQUE

DATE/TIME OF TEST: 05/31/2020 02:29:12

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.155	02:24
Breath Volume	1.812	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.161	02:27
Breath Volume	2.679	02:27
Air Blank	0.000	02:28
Control Test	0.081	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 06/26/2020 08:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,DEANNE,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GOETHE,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/31/2020 03:07:57

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.080	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.161	03:03
Breath Volume	1.691	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:05
Subject Sample # 2	0.162	03:06
Breath Volume	1.261	03:06
Air Blank	0.000	03:07
Control Test	0.081	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 06/26/2020 08:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN, DALE, E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SCHNEIDER, OFFICER
ARRESTING OFFICER AGENCY: Lake Placid PD
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 05/01/2020 16:06:56

Diagnostics Check	OK	16:00
Air Blank	0.000	16:00
Control Test	0.081	16:01
Air Blank	0.000	16:01
Subject Sample # 1	0.138	16:02
Breath Volume	2.281	16:02
Air Blank	0.000	16:02
Air Blank	0.000	16:04
Subject Sample # 2	0.142	16:05
Breath Volume	2.546	16:05
Air Blank	0.000	16:06
Control Test	0.081	16:06
Air Blank	0.000	16:06
Diagnostics Check	OK	16:06

Data Download Date/Time: 05/31/2020 08:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BECK,EMILY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 05/01/2020 21:50:03

Diagnostics Check	OK	21:47	
Air Blank	0.000	21:47	
Control Test	0.080	21:47	
Air Blank	0.000	21:48	
Subject Sample	# 1	21:48	Subject Test Refused
Breath Volume	0.000	21:48	
Air Blank	0.000	21:49	
Control Test	0.080	21:49	
Air Blank	0.000	21:49	
Diagnostics Check	OK	21:50	

Data Download Date/Time: 05/31/2020 08:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: RITENOUR,STEVEN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: HELMS

DATE/TIME OF TEST: 05/02/2020 21:20:32

Diagnostics Check	OK	21:13
Air Blank	0.000	21:14
Control Test	0.080	21:14
Air Blank	0.000	21:15
Subject Sample # 1	0.127	21:16
Breath Volume	1.511	21:16
Air Blank	0.000	21:16
Air Blank	0.000	21:18
Subject Sample # 2	0.129	21:19
Breath Volume	1.539	21:19
Air Blank	0.000	21:19
Control Test	0.079	21:19
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

Data Download Date/Time: 05/31/2020 08:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PERALTA,ACEVEDO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/03/2020 00:21:57

Diagnostics Check	OK	00:19	
Air Blank	0.000	00:19	
Control Test	0.079	00:19	
Air Blank	0.000	00:20	
Subject Sample	# 1	00:20	Subject Test Refused
Breath Volume	0.000	00:20	
Air Blank	0.000	00:21	
Control Test	0.079	00:21	
Air Blank	0.000	00:21	
Diagnostics Check	OK	00:21	

Data Download Date/Time: 05/31/2020 08:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: FUGATE,V
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 05/23/2020 18:58:31

Diagnostics Check	OK	18:55	
Air Blank	0.000	18:56	
Control Test	0.080	18:56	
Air Blank	0.000	18:56	
Subject Sample	# 1	18:57	Subject Test Refused
Breath Volume	0.000	18:57	
Air Blank	0.000	18:57	
Control Test	0.080	18:57	
Air Blank	0.000	18:58	
Diagnostics Check	OK	18:58	

Data Download Date/Time: 05/31/2020 08:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: THOMPSON,S
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/23/2020 22:28:59

Diagnostics Check	OK	22:22
Air Blank	0.000	22:23
Control Test	0.080	22:23
Air Blank	0.000	22:23
Subject Sample # 1	0.181	22:24
Breath Volume	1.355	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.182	22:27
Breath Volume	1.644	22:27
Air Blank	0.000	22:28
Control Test	0.079	22:28
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 05/31/2020 08:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: CONTRERAS,ANIBAL
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: GOMEZBORGES

DATE/TIME OF TEST: 05/30/2020 02:13:24

Diagnostics Check	OK	02:08	
Air Blank	0.000	02:09	
Control Test	0.080	02:09	
Air Blank	0.000	02:09	
Subject Sample	# 1	02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:12	
Control Test	0.080	02:12	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 05/31/2020 08:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,K,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: HOLMES,K
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/30/2020 17:16:13

Diagnostics Check	OK	17:10	
Air Blank	0.000	17:10	
Control Test	0.080	17:11	
Air Blank	0.000	17:11	
Subject Sample # 1	0.000	17:12	
Breath Volume	1.546	17:12	
Air Blank	0.000	17:12	
Air Blank	0.000	17:14	
Subject Sample # 2		17:14	Subject Test Refused
Breath Volume	0.000	17:14	
Air Blank	0.000	17:15	
Control Test	0.081	17:15	
Air Blank	0.000	17:16	
Diagnostics Check	OK	17:16	

Data Download Date/Time: 05/31/2020 08:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ACEVEDO
PERALTA,EMMANUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male

OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: MORAN

DATE/TIME OF TEST: 05/31/2020 00:40:15

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:37	
Control Test	0.079	00:37	
Air Blank	0.000	00:38	
Subject Sample	# 1	00:38	Subject Test Refused
Breath Volume	0.000	00:38	
Air Blank	0.000	00:39	
Control Test	0.079	00:39	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:40	

Data Download Date/Time: 05/31/2020 08:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GARRISON,JOHN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: STAFFIERI

DATE/TIME OF TEST: 05/31/2020 02:56:20

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:53	
Control Test	0.081	02:53	
Air Blank	0.000	02:54	
Subject Sample	# 1	02:54	Subject Test Refused
Breath Volume	0.000	02:54	
Air Blank	0.000	02:55	
Control Test	0.080	02:55	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

Data Download Date/Time: 05/31/2020 08:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENYON,KYLE,A
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: KENYON,KYLE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 04/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 05/11/2020 20:17:49

Diagnostics Check	OK	20:11
Air Blank	0.000	20:12
Control Test	0.078	20:12
Air Blank	0.000	20:12
Subject Sample # 1	0.000	20:13
Breath Volume	4.179	20:13
Air Blank	0.000	20:14
Air Blank	0.000	20:16
Subject Sample # 2	0.000	20:16
Breath Volume	4.187	20:16
Air Blank	0.000	20:16
Control Test	0.079	20:17
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 05/24/2020 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENYON,KYLE,A
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: KENYON,KYLE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 04/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: SACKKEY

DATE/TIME OF TEST: 05/21/2020 23:18:30

Diagnostics Check	OK	23:12
Air Blank	0.000	23:12
Control Test	0.078	23:13
Air Blank	0.000	23:13
Subject Sample # 1	0.000	23:14
Breath Volume	1.328	23:14
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.000	23:17
Breath Volume	1.367	23:17
Air Blank	0.000	23:17
Control Test	0.078	23:17
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 05/24/2020 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: ROBBINS,MARK
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: BERGERON

DATE/TIME OF TEST: 05/29/2020 03:53:32

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.077	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.000	03:49
Breath Volume	2.710	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.000	03:52
Breath Volume	2.503	03:52
Air Blank	0.000	03:52
Control Test	0.077	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 06/15/2020 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: ZIMMERMAN,JASON
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:08

SUBJECT'S LAST NAME: PADEN

DATE/TIME OF TEST: 05/03/2020 08:34:43

Diagnostics Check	OK	08:28
Air Blank	0.000	08:28
Control Test	0.077	08:29
Air Blank	0.000	08:29
Subject Sample # 1	0.000	08:30
Breath Volume	1.363	08:30
Air Blank	0.000	08:30
Air Blank	0.000	08:32
Subject Sample # 2	0.000	08:33
Breath Volume	1.714	08:33
Air Blank	0.000	08:33
Control Test	0.077	08:34
Air Blank	0.000	08:34
Diagnostics Check	OK	08:34

Data Download Date/Time: 05/28/2020 02:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONDOS,ROBERT,E
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: REYES,MARIO
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: ENNS

DATE/TIME OF TEST: 05/13/2020 19:21:39

Diagnostics Check	OK	19:15
Air Blank	0.000	19:15
Control Test	0.077	19:15
Air Blank	0.000	19:16
Subject Sample # 1	0.258	19:16
Breath Volume	1.781	19:16
Air Blank	0.000	19:17
Air Blank	0.000	19:19
Subject Sample # 2	0.251	19:20
Breath Volume	1.839	19:20
Air Blank	0.000	19:20
Control Test	0.077	19:21
Air Blank	0.000	19:21
Diagnostics Check	OK	19:21

Data Download Date/Time: 06/02/2020 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,JEFFREY,D
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: PALMER,LEEANN
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: ALBARRAN

DATE/TIME OF TEST: 05/18/2020 02:21:10

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:12	
Control Test	0.077	02:13	
Air Blank	0.000	02:13	
Subject Sample	# 1	02:16	No Sample Provided
Breath Volume	0.000	02:16	
Air Blank	0.000	02:17	
Air Blank	0.000	02:19	
Subject Sample	# 2	02:19	Subject Test Refused
Breath Volume	0.000	02:19	
Air Blank	0.000	02:20	
Control Test	0.077	02:20	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time: 06/02/2020 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONDOS,ROBERT,E
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: RONDOS,ROBERT
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 05/27/2020 19:19:00

Diagnostics Check	OK	19:16	
Air Blank	0.000	19:16	
Control Test	0.078	19:16	
Air Blank	0.000	19:17	
Subject Sample	# 1	19:17	Subject Test Refused
Breath Volume	0.000	19:17	
Air Blank	0.000	19:18	
Control Test	0.077	19:18	
Air Blank	0.000	19:18	
Diagnostics Check	OK	19:19	

Data Download Date/Time: 06/02/2020 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPATZ,JOHN,M
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: HAZIME,SABRINA
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: HOULLIS

DATE/TIME OF TEST: 05/31/2020 01:09:37

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:07	
Control Test	0.079	01:07	
Air Blank	0.000	01:08	
Subject Sample	# 1	01:08	Subject Test Refused
Breath Volume	0.000	01:08	
Air Blank	0.000	01:08	
Control Test	0.079	01:09	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 06/02/2020 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HUFF,ZACH
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 15:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:41

SUBJECT'S LAST NAME: CHALLIS

DATE/TIME OF TEST: 05/02/2020 16:15:10

Diagnostics Check	OK	16:01	
Air Blank	0.000	16:02	
Control Test	0.082	16:02	
Air Blank	0.000	16:02	
Subject Sample	# 1 0.000	16:05	
Breath Volume	1.394	16:05	
Air Blank	0.000	16:05	
Air Blank	0.000	16:07	
Subject Sample	# 2 0.000	16:10	Volume Not Met
Breath Volume	0.136	16:10	
Air Blank	0.000	16:11	
Air Blank	0.000	16:13	
Subject Sample	# 3 0.000	16:13	
Breath Volume	2.589	16:13	
Air Blank	0.000	16:14	
Control Test	0.083	16:14	
Air Blank	0.000	16:15	
Diagnostics Check	OK	16:15	

Data Download Date/Time: 05/14/2020 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: COFFEY

DATE/TIME OF TEST: 05/06/2020 21:17:54

Diagnostics Check	OK	21:11
Air Blank	0.000	21:11
Control Test	0.082	21:11
Air Blank	0.000	21:12
Subject Sample # 1	0.117	21:13
Breath Volume	2.277	21:13
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.114	21:16
Breath Volume	1.777	21:16
Air Blank	0.000	21:17
Control Test	0.081	21:17
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 05/14/2020 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,JOHN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: DANG

DATE/TIME OF TEST: 05/07/2020 04:29:52

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:27	
Control Test	0.082	04:27	
Air Blank	0.000	04:28	
Subject Sample	# 1	04:28	Subject Test Refused
Breath Volume	0.000	04:28	
Air Blank	0.000	04:28	
Control Test	0.081	04:29	
Air Blank	0.000	04:29	
Diagnostics Check	OK	04:29	

Data Download Date/Time: 05/14/2020 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VICKERS,JAMES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: SOLOMON

DATE/TIME OF TEST: 05/07/2020 21:02:44

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.082	20:57
Air Blank	0.000	20:57
Subject Sample # 1	0.152	20:58
Breath Volume	1.621	20:58
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.151	21:01
Breath Volume	1.753	21:01
Air Blank	0.000	21:01
Control Test	0.081	21:02
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 05/14/2020 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: GONZALEZ REYES

DATE/TIME OF TEST: 05/09/2020 00:05:36

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.082	23:58
Air Blank	0.000	23:59
Subject Sample # 1	0.090	00:00
Breath Volume	2.769	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.093	00:04
Breath Volume	2.703	00:04
Air Blank	0.000	00:04
Control Test	0.081	00:05
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 05/14/2020 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GARDNER,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: DANHARDT

DATE/TIME OF TEST: 05/09/2020 02:57:49

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.082	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.204	02:53
Breath Volume	2.417	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.207	02:56
Breath Volume	2.460	02:56
Air Blank	0.000	02:56
Control Test	0.081	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 05/14/2020 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BUSH,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 15:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 05/11/2020 17:23:43

Diagnostics Check	OK	17:20	
Air Blank	0.000	17:21	
Control Test	0.082	17:21	
Air Blank	0.000	17:22	
Subject Sample	# 1	17:22	Subject Test Refused
Breath Volume	0.000	17:22	
Air Blank	0.000	17:22	
Control Test	0.081	17:23	
Air Blank	0.000	17:23	
Diagnostics Check	OK	17:23	

Data Download Date/Time: 05/14/2020 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 05/12/2020 22:02:55

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.083	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.103	21:58
Breath Volume	1.253	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:00
Subject Sample # 2	0.110	22:01
Breath Volume	1.578	22:01
Air Blank	0.000	22:02
Control Test	0.081	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 05/14/2020 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GREENE,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 05/13/2020 21:47:37

Diagnostics Check	OK	21:44	
Air Blank	0.000	21:44	
Control Test	0.081	21:45	
Air Blank	0.000	21:45	
Subject Sample	# 1	21:46	Subject Test Refused
Breath Volume	0.000	21:46	
Air Blank	0.000	21:46	
Control Test	0.081	21:47	
Air Blank	0.000	21:47	
Diagnostics Check	OK	21:47	

Data Download Date/Time: 05/14/2020 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: MARRA

DATE/TIME OF TEST: 05/14/2020 23:03:08

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:00	
Control Test	0.082	23:00	
Air Blank	0.000	23:01	
Subject Sample	# 1	23:01	Subject Test Refused
Breath Volume	0.000	23:01	
Air Blank	0.000	23:02	
Control Test	0.081	23:02	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:03	

Data Download Date/Time: 06/03/2020 23:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CONEBY, MEGAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: LETTS

DATE/TIME OF TEST: 05/15/2020 03:50:10

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:47	
Control Test	0.083	03:47	
Air Blank	0.000	03:48	
Subject Sample	# 1	03:48	Subject Test Refused
Breath Volume	0.000	03:48	
Air Blank	0.000	03:49	
Control Test	0.081	03:49	
Air Blank	0.000	03:50	
Diagnostics Check	OK	03:50	

Data Download Date/Time: 06/03/2020 23:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FIGUEROA,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: COTRONE

DATE/TIME OF TEST: 05/15/2020 21:42:36

Diagnostics Check	OK	21:35
Air Blank	0.000	21:35
Control Test	0.082	21:36
Air Blank	0.000	21:36
Subject Sample # 1	0.023	21:37
Breath Volume	1.585	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.026	21:41
Breath Volume	1.707	21:41
Air Blank	0.000	21:41
Control Test	0.083	21:41
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time: 06/03/2020 23:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GROFF,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: FARERI

DATE/TIME OF TEST: 05/16/2020 01:47:24

Diagnostics Check	OK	01:44	
Air Blank	0.000	01:44	
Control Test	0.082	01:44	
Air Blank	0.000	01:45	
Subject Sample	# 1	01:46	Subject Test Refused
Breath Volume	0.000	01:46	
Air Blank	0.000	01:46	
Control Test	0.081	01:46	
Air Blank	0.000	01:47	
Diagnostics Check	OK	01:47	

Data Download Date/Time: 06/03/2020 23:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORO,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SCHIFFER

DATE/TIME OF TEST: 05/16/2020 02:23:01

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:20	
Control Test	0.081	02:20	
Air Blank	0.000	02:21	
Subject Sample	# 1	02:21	Subject Test Refused
Breath Volume	0.000	02:21	
Air Blank	0.000	02:22	
Control Test	0.080	02:22	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:23	

Data Download Date/Time: 06/03/2020 23:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORO,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: TATE

DATE/TIME OF TEST: 05/16/2020 04:17:16

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.082	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.178	04:12
Breath Volume	2.613	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:14
Subject Sample # 2	0.179	04:15
Breath Volume	1.355	04:15
Air Blank	0.000	04:16
Control Test	0.080	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 06/03/2020 23:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,K,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CAREY,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 05/16/2020 05:06:27

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.081	04:58
Air Blank	0.000	04:59
Subject Sample # 1	0.169	05:01
Breath Volume	1.355	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:03
Subject Sample # 2	0.172	05:04
Breath Volume	1.562	05:04
Air Blank	0.000	05:05
Control Test	0.081	05:05
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 06/03/2020 23:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 05/16/2020 21:47:37

Diagnostics Check	OK	21:40
Air Blank	0.000	21:40
Control Test	0.082	21:41
Air Blank	0.000	21:41
Subject Sample # 1	0.280	21:42
Breath Volume	3.121	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.272	21:46
Breath Volume	3.441	21:46
Air Blank	0.000	21:46
Control Test	0.081	21:47
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 06/03/2020 23:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PETESON,W
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: RUSH

DATE/TIME OF TEST: 05/17/2020 00:57:55

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.082	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.176	00:52
Breath Volume	1.906	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.168	00:56
Breath Volume	1.472	00:56
Air Blank	0.000	00:56
Control Test	0.082	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BELCHER,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 14:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 05/17/2020 15:27:48

Diagnostics Check	OK	15:20
Air Blank	0.000	15:21
Control Test	0.082	15:21
Air Blank	0.000	15:22
Subject Sample # 1	0.193	15:23
Breath Volume	1.187	15:23
Air Blank	0.000	15:23
Air Blank	0.000	15:25
Subject Sample # 2	0.183	15:26
Breath Volume	1.128	15:26
Air Blank	0.000	15:26
Control Test	0.081	15:27
Air Blank	0.000	15:27
Diagnostics Check	OK	15:27

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROPER, COLIN, G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROPER, COLIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: BUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: NADER

DATE/TIME OF TEST: 05/17/2020 23:31:01

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:26	
Control Test	0.081	23:26	
Air Blank	0.000	23:27	
Subject Sample	# 1	23:29	Subject Test Refused
Breath Volume	0.000	23:29	
Air Blank	0.000	23:30	
Control Test	0.081	23:30	
Air Blank	0.000	23:30	
Diagnostics Check	OK	23:31	

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,JOHN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 05/18/2020 00:20:11

Diagnostics Check	OK	00:10
Air Blank	0.000	00:11
Control Test	0.082	00:11
Air Blank	0.000	00:11
Subject Sample # 1	0.206	00:14
Breath Volume	1.511	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:16
Subject Sample # 2	0.210	00:18
Breath Volume	3.152	00:18
Air Blank	0.000	00:19
Control Test	0.082	00:19
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MARSH

DATE/TIME OF TEST: 05/21/2020 02:13:41

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:11	
Control Test	0.082	02:11	
Air Blank	0.000	02:12	
Subject Sample	# 1	02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:12	
Control Test	0.081	02:13	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: STYLES,L
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: DACOSTA

DATE/TIME OF TEST: 05/21/2020 02:19:29

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:17	
Control Test	0.082	02:17	
Air Blank	0.000	02:17	
Subject Sample	# 1	02:18	Subject Test Refused
Breath Volume	0.000	02:18	
Air Blank	0.000	02:18	
Control Test	0.080	02:18	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 07:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:54

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 05/21/2020 07:41:33

Diagnostics Check	OK	07:35
Air Blank	0.000	07:35
Control Test	0.083	07:36
Air Blank	0.000	07:36
Subject Sample # 1	0.000	07:37
Breath Volume	3.410	07:37
Air Blank	0.000	07:37
Air Blank	0.000	07:39
Subject Sample # 2	0.000	07:40
Breath Volume	3.308	07:40
Air Blank	0.000	07:40
Control Test	0.082	07:40
Air Blank	0.000	07:41
Diagnostics Check	OK	07:41

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: REFUSE,RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 16:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: YOUMAN

DATE/TIME OF TEST: 05/21/2020 17:37:12

Diagnostics Check	OK	17:20	
Air Blank	0.000	17:21	
Control Test	0.082	17:21	
Air Blank	0.000	17:22	
Subject Sample	# 1 0.071	17:24	
Breath Volume	1.847	17:24	
Air Blank	0.000	17:25	
Air Blank	0.000	17:26	
Subject Sample	# 2 0.054	17:30	Volume Not Met
Breath Volume	0.265	17:30	
Air Blank	0.000	17:30	
Air Blank	0.000	17:32	
Subject Sample	# 3 0.066	17:35	
Breath Volume	1.320	17:35	
Air Blank	0.000	17:36	
Control Test	0.082	17:36	
Air Blank	0.000	17:37	
Diagnostics Check	OK	17:37	

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,K,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CAREY,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 05/22/2020 04:08:25

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:57	
Control Test	0.082	03:57	
Air Blank	0.000	03:58	
Subject Sample	# 1	04:01	No Sample Provided
Breath Volume	0.000	04:01	
Air Blank	0.000	04:01	
Air Blank	0.000	04:03	
Subject Sample	# 2	04:07	No Sample Provided
Breath Volume	0.000	04:07	
Air Blank	0.000	04:07	
Control Test	0.082	04:07	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: STONE,K
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: LANGFORD

DATE/TIME OF TEST: 05/22/2020 23:07:46

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.083	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.069	23:03
Breath Volume	2.335	23:03
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.072	23:06
Breath Volume	2.800	23:06
Air Blank	0.000	23:06
Control Test	0.082	23:07
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: OSTROSKI,N
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: LOVISCEK

DATE/TIME OF TEST: 05/23/2020 01:09:36

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:07	
Control Test	0.082	01:07	
Air Blank	0.000	01:07	
Subject Sample	# 1	01:08	Subject Test Refused
Breath Volume	0.000	01:08	
Air Blank	0.000	01:08	
Control Test	0.081	01:09	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CONEBY,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: ORDONEZ

DATE/TIME OF TEST: 05/23/2020 02:30:13

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:14	
Control Test	0.082	02:14	
Air Blank	0.000	02:15	
Subject Sample	# 1 0.221	02:17	
Breath Volume	1.996	02:17	
Air Blank	0.000	02:17	
Air Blank	0.000	02:19	
Subject Sample	# 2 0.200	02:22	Volume Not Met
Breath Volume	0.031	02:22	
Air Blank	0.000	02:23	
Air Blank	0.000	02:25	
Subject Sample	# 3 0.204	02:28	Volume Not Met
Breath Volume	0.019	02:28	
Air Blank	0.000	02:29	
Control Test	0.080	02:29	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GARDNER,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: BUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 05/23/2020 22:58:40

Diagnostics Check	OK	22:46	
Air Blank	0.000	22:46	
Control Test	0.082	22:46	
Air Blank	0.000	22:47	
Subject Sample	# 1 0.148	22:50	Slope Not Level
Breath Volume	1.449	22:50	
Air Blank	0.000	22:51	
Air Blank	0.000	22:52	
Subject Sample	# 2 0.159	22:53	No .020 agreement
Breath Volume	1.378	22:53	
Air Blank	0.000	22:54	
Air Blank	0.000	22:56	
Subject Sample	# 3 0.183	22:57	No .020 agreement
Breath Volume	1.550	22:57	
Air Blank	0.000	22:57	
Control Test	0.081	22:58	
Air Blank	0.000	22:58	
Diagnostics Check	OK	22:58	

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROURKE,JAMES,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WILLIAMS,TIMMY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: ROURKE

DATE/TIME OF TEST: 05/23/2020 23:43:28

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.082	23:38
Air Blank	0.000	23:38
Subject Sample # 1	0.156	23:38
Breath Volume	2.968	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.157	23:41
Breath Volume	2.917	23:41
Air Blank	0.000	23:42
Control Test	0.081	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 05/24/2020 02:29:39

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.082	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.140	02:25
Breath Volume	2.000	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.137	02:28
Breath Volume	1.855	02:28
Air Blank	0.000	02:28
Control Test	0.081	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NIXON,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/24/2020 19:50:42

Diagnostics Check	OK	19:47	
Air Blank	0.000	19:47	
Control Test	0.082	19:48	
Air Blank	0.000	19:48	
Subject Sample	# 1	19:49	Subject Test Refused
Breath Volume	0.000	19:49	
Air Blank	0.000	19:49	
Control Test	0.081	19:50	
Air Blank	0.000	19:50	
Diagnostics Check	OK	19:50	

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CONEBY,MEGAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: MINSTER

DATE/TIME OF TEST: 05/24/2020 20:44:09

Diagnostics Check	OK	20:36
Air Blank	0.000	20:37
Control Test	0.081	20:37
Air Blank	0.000	20:38
Subject Sample # 1	0.266	20:39
Breath Volume	1.394	20:39
Air Blank	0.000	20:40
Air Blank	0.000	20:41
Subject Sample # 2	0.261	20:42
Breath Volume	1.859	20:42
Air Blank	0.000	20:43
Control Test	0.080	20:43
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MARTINI,ISABELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: GASKINS

DATE/TIME OF TEST: 05/24/2020 23:48:22

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:37	
Control Test	0.081	23:38	
Air Blank	0.000	23:38	
Subject Sample	# 1	23:41	No Sample Provided
Breath Volume	0.000	23:41	
Air Blank	0.000	23:42	
Air Blank	0.000	23:44	
Subject Sample	# 2	23:47	Subject Test Refused
Breath Volume	0.000	23:47	
Air Blank	0.000	23:47	
Control Test	0.082	23:47	
Air Blank	0.000	23:48	
Diagnostics Check	OK	23:48	

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GARDNER,DANIEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 05/25/2020 01:20:43

Diagnostics Check	OK	01:14
Air Blank	0.000	01:15
Control Test	0.082	01:15
Air Blank	0.000	01:16
Subject Sample # 1	0.000	01:16
Breath Volume	2.410	01:16
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.000	01:19
Breath Volume	2.339	01:19
Air Blank	0.000	01:19
Control Test	0.081	01:20
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBSEN,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 05:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:16

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 05/25/2020 06:42:38

Diagnostics Check	OK	06:36
Air Blank	0.000	06:36
Control Test	0.081	06:37
Air Blank	0.000	06:37
Subject Sample # 1	0.212	06:38
Breath Volume	2.242	06:38
Air Blank	0.000	06:38
Air Blank	0.000	06:40
Subject Sample # 2	0.197	06:41
Breath Volume	1.503	06:41
Air Blank	0.000	06:41
Control Test	0.080	06:42
Air Blank	0.000	06:42
Diagnostics Check	OK	06:42

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 14:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: CURTIS

DATE/TIME OF TEST: 05/26/2020 15:57:24

Diagnostics Check	OK	15:51
Air Blank	0.000	15:51
Control Test	0.082	15:51
Air Blank	0.000	15:52
Subject Sample # 1	0.267	15:52
Breath Volume	1.757	15:52
Air Blank	0.000	15:53
Air Blank	0.000	15:55
Subject Sample # 2	0.271	15:55
Breath Volume	2.335	15:55
Air Blank	0.000	15:56
Control Test	0.080	15:56
Air Blank	0.000	15:57
Diagnostics Check	OK	15:57

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,JOHN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: FENLON

DATE/TIME OF TEST: 05/27/2020 04:33:39

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:31	
Control Test	0.082	04:31	
Air Blank	0.000	04:32	
Subject Sample	# 1	04:32	Subject Test Refused
Breath Volume	0.000	04:32	
Air Blank	0.000	04:32	
Control Test	0.080	04:33	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: MACMILLAN

DATE/TIME OF TEST: 05/29/2020 01:50:48

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:47	
Control Test	0.082	01:48	
Air Blank	0.000	01:48	
Subject Sample	# 1	01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:49	
Control Test	0.080	01:50	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:50	

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CONEBY, MEGAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: PETRILLO

DATE/TIME OF TEST: 05/29/2020 04:20:44

Diagnostics Check	OK	04:13
Air Blank	0.000	04:14
Control Test	0.082	04:14
Air Blank	0.000	04:15
Subject Sample # 1	0.130	04:16
Breath Volume	1.683	04:16
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.129	04:19
Breath Volume	1.308	04:19
Air Blank	0.000	04:19
Control Test	0.081	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 05/29/2020 21:17:25

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.082	21:10
Air Blank	0.000	21:11
Subject Sample # 1	0.134	21:12
Breath Volume	2.175	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.125	21:15
Breath Volume	1.714	21:15
Air Blank	0.000	21:16
Control Test	0.081	21:16
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MALDONADO,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916501
DRY GAS EXPIRY DATE: 06/14/2021
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: BURCHETTE

DATE/TIME OF TEST: 05/30/2020 03:23:38

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.083	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.104	03:19
Breath Volume	1.488	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.100	03:22
Breath Volume	1.324	03:22
Air Blank	0.000	03:22
Control Test	0.081	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 06/03/2020 23:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: MCINTOSH

DATE/TIME OF TEST: 05/01/2020 21:49:33

Diagnostics Check	OK	21:42
Air Blank	0.000	21:42
Control Test	0.081	21:43
Air Blank	0.000	21:43
Subject Sample # 1	0.176	21:44
Breath Volume	2.371	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.170	21:48
Breath Volume	3.445	21:48
Air Blank	0.000	21:48
Control Test	0.081	21:48
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 05/11/2020 17:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,OFFICER,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRUNEAU,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: MOHR

DATE/TIME OF TEST: 05/07/2020 21:02:33

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.080	20:56
Air Blank	0.000	20:57
Subject Sample # 1	0.128	20:57
Breath Volume	5.230	20:57
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.130	21:00
Breath Volume	5.027	21:00
Air Blank	0.000	21:01
Control Test	0.080	21:01
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 05/11/2020 17:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: CLABORN

DATE/TIME OF TEST: 05/07/2020 22:46:14

Diagnostics Check	OK	22:39
Air Blank	0.000	22:39
Control Test	0.080	22:40
Air Blank	0.000	22:40
Subject Sample # 1	0.036	22:41
Breath Volume	4.699	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:43
Subject Sample # 2	0.035	22:44
Breath Volume	4.003	22:44
Air Blank	0.000	22:45
Control Test	0.080	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 05/11/2020 17:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,OFFICER,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRUNEAU,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: LABAZZO

DATE/TIME OF TEST: 05/09/2020 20:35:51

Diagnostics Check	OK	20:29
Air Blank	0.000	20:30
Control Test	0.080	20:30
Air Blank	0.000	20:30
Subject Sample # 1	0.203	20:31
Breath Volume	1.546	20:31
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample # 2	0.206	20:34
Breath Volume	2.378	20:34
Air Blank	0.000	20:34
Control Test	0.080	20:35
Air Blank	0.000	20:35
Diagnostics Check	OK	20:35

Data Download Date/Time: 05/11/2020 17:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENEZIA,STEVEN,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VENEZIA,STEVEN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: GVOZDKOVA

DATE/TIME OF TEST: 05/09/2020 23:35:47

Diagnostics Check	OK	23:22	
Air Blank	0.000	23:23	
Control Test	0.080	23:23	
Air Blank	0.000	23:24	
Subject Sample	# 1 0.270	23:27	Volume Not Met
Breath Volume	0.683	23:27	
Air Blank	0.000	23:28	
Air Blank	0.000	23:29	
Subject Sample	# 2 0.252	23:30	
Breath Volume	1.816	23:30	
Air Blank	0.000	23:31	
Air Blank	0.000	23:33	
Subject Sample	# 3 0.250	23:34	
Breath Volume	1.917	23:34	
Air Blank	0.000	23:34	
Control Test	0.079	23:35	
Air Blank	0.000	23:35	
Diagnostics Check	OK	23:35	

Data Download Date/Time: 05/11/2020 17:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: FLORES GUIFARRO

DATE/TIME OF TEST: 05/16/2020 01:19:05

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.080	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.237	01:14
Breath Volume	3.082	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:16
Subject Sample # 2	0.235	01:17
Breath Volume	2.050	01:17
Air Blank	0.000	01:18
Control Test	0.080	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 05/27/2020 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 05/16/2020 19:32:26

Diagnostics Check	OK	19:26
Air Blank	0.000	19:26
Control Test	0.081	19:26
Air Blank	0.000	19:27
Subject Sample # 1	0.244	19:27
Breath Volume	2.882	19:27
Air Blank	0.000	19:28
Air Blank	0.000	19:30
Subject Sample # 2	0.233	19:30
Breath Volume	2.550	19:30
Air Blank	0.000	19:31
Control Test	0.080	19:31
Air Blank	0.000	19:32
Diagnostics Check	OK	19:32

Data Download Date/Time: 05/27/2020 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/17/2020 00:30:24

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.080	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.000	00:25
Breath Volume	2.492	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.000	00:29
Breath Volume	1.796	00:29
Air Blank	0.000	00:29
Control Test	0.079	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 05/27/2020 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRIDWELL,C
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: BUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: FAIL

DATE/TIME OF TEST: 05/17/2020 22:29:07

Diagnostics Check	OK	22:26	
Air Blank	0.000	22:26	
Control Test	0.081	22:27	
Air Blank	0.000	22:27	
Subject Sample	# 1	22:27	Subject Test Refused
Breath Volume	0.000	22:27	
Air Blank	0.000	22:28	
Control Test	0.081	22:28	
Air Blank	0.000	22:29	
Diagnostics Check	OK	22:29	

Data Download Date/Time: 05/27/2020 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 05/18/2020 03:50:00

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.080	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.190	03:45
Breath Volume	2.605	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.185	03:48
Breath Volume	1.132	03:48
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:50

Data Download Date/Time: 05/27/2020 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: DEAL

DATE/TIME OF TEST: 05/18/2020 22:40:03

Diagnostics Check	OK	22:33
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.082	22:35
Breath Volume	2.769	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:37
Subject Sample # 2	0.079	22:38
Breath Volume	2.113	22:38
Air Blank	0.000	22:39
Control Test	0.080	22:39
Air Blank	0.000	22:39
Diagnostics Check	OK	22:40

Data Download Date/Time: 05/27/2020 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN, D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 05/22/2020 22:35:51

Diagnostics Check	OK	22:29
Air Blank	0.000	22:29
Control Test	0.080	22:30
Air Blank	0.000	22:30
Subject Sample # 1	0.061	22:31
Breath Volume	2.527	22:31
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.059	22:34
Breath Volume	2.496	22:34
Air Blank	0.000	22:34
Control Test	0.080	22:35
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 05/27/2020 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN, D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: AGUIRRE

DATE/TIME OF TEST: 05/23/2020 21:06:10

Diagnostics Check	OK	20:59
Air Blank	0.000	20:59
Control Test	0.080	21:00
Air Blank	0.000	21:00
Subject Sample # 1	0.161	21:01
Breath Volume	2.500	21:01
Air Blank	0.000	21:02
Air Blank	0.000	21:04
Subject Sample # 2	0.161	21:04
Breath Volume	1.472	21:04
Air Blank	0.000	21:05
Control Test	0.080	21:05
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

Data Download Date/Time: 05/27/2020 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS, B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 05/23/2020 22:40:28

Diagnostics Check	OK	22:34
Air Blank	0.000	22:34
Control Test	0.080	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.108	22:36
Breath Volume	1.882	22:36
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.113	22:38
Breath Volume	2.003	22:38
Air Blank	0.000	22:39
Control Test	0.080	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 05/27/2020 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: ENGLISH

DATE/TIME OF TEST: 05/24/2020 01:37:18

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:23	
Control Test	0.080	01:23	
Air Blank	0.000	01:23	
Subject Sample # 1	0.112	01:26	
Breath Volume	2.121	01:26	
Air Blank	0.000	01:27	
Air Blank	0.000	01:28	
Subject Sample # 2	0.108	01:32	Volume Not Met
Breath Volume	0.039	01:32	
Air Blank	0.000	01:32	
Air Blank	0.000	01:34	
Subject Sample # 3	0.114	01:35	
Breath Volume	1.246	01:35	
Air Blank	0.000	01:36	
Control Test	0.079	01:36	
Air Blank	0.000	01:37	
Diagnostics Check	OK	01:37	

Data Download Date/Time: 05/27/2020 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN, D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: WEEKS

DATE/TIME OF TEST: 05/24/2020 02:16:44

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:14	
Control Test	0.079	02:14	
Air Blank	0.000	02:14	
Subject Sample	# 1	02:15	Subject Test Refused
Breath Volume	0.000	02:15	
Air Blank	0.000	02:15	
Control Test	0.080	02:16	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

Data Download Date/Time: 05/27/2020 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE, S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: SHAD

DATE/TIME OF TEST: 05/24/2020 04:39:49

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.079	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.204	04:35
Breath Volume	4.859	04:35
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.203	04:38
Breath Volume	3.777	04:38
Air Blank	0.000	04:38
Control Test	0.080	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 05/27/2020 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KASSIS, F
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/25/2020 21:10:20

Diagnostics Check	OK	21:07	
Air Blank	0.000	21:07	
Control Test	0.082	21:08	
Air Blank	0.000	21:08	
Subject Sample	# 1	21:08	Subject Test Refused
Breath Volume	0.000	21:08	
Air Blank	0.000	21:09	
Control Test	0.081	21:09	
Air Blank	0.000	21:10	
Diagnostics Check	OK	21:10	

Data Download Date/Time: 05/27/2020 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: BARCARO

DATE/TIME OF TEST: 05/25/2020 23:48:55

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.081	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.037	23:44
Breath Volume	3.050	23:44
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.038	23:47
Breath Volume	2.789	23:47
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 05/27/2020 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 05/26/2020 01:40:07

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.081	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.152	01:35
Breath Volume	1.750	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:37
Subject Sample # 2	0.146	01:38
Breath Volume	1.597	01:38
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 05/27/2020 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: HINMAN

DATE/TIME OF TEST: 05/26/2020 03:14:52

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.081	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.185	03:07
Breath Volume	2.050	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.142	03:10
Breath Volume	1.253	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 3	0.160	03:13
Breath Volume	1.859	03:13
Air Blank	0.000	03:13
Control Test	0.080	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 05/27/2020 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: CRUCE

DATE/TIME OF TEST: 05/27/2020 00:40:57

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.081	00:35
Air Blank	0.000	00:35
Subject Sample # 1	0.007	00:36
Breath Volume	2.968	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.000	00:39
Breath Volume	2.937	00:39
Air Blank	0.000	00:40
Control Test	0.081	00:40
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 05/27/2020 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: CHERIF

DATE/TIME OF TEST: 05/27/2020 02:49:55

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:47	
Control Test	0.081	02:47	
Air Blank	0.000	02:48	
Subject Sample	# 1	02:48	Subject Test Refused
Breath Volume	0.000	02:48	
Air Blank	0.000	02:49	
Control Test	0.081	02:49	
Air Blank	0.000	02:49	
Diagnostics Check	OK	02:49	

Data Download Date/Time: 05/27/2020 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: FARRELL

DATE/TIME OF TEST: 05/27/2020 03:18:29

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.081	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.127	03:13
Breath Volume	1.816	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.123	03:16
Breath Volume	2.082	03:16
Air Blank	0.000	03:17
Control Test	0.081	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 05/27/2020 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS, B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: ALDRIDGE

DATE/TIME OF TEST: 05/28/2020 01:14:19

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.137	01:09
Breath Volume	1.917	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.142	01:12
Breath Volume	1.960	01:12
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 06/17/2020 23:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/30/2020 01:14:52

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:12	
Control Test	0.081	01:12	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:13	Subject Test Refused
Breath Volume	0.000	01:13	
Air Blank	0.000	01:13	
Control Test	0.080	01:14	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

Data Download Date/Time: 06/17/2020 23:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: FIFE

DATE/TIME OF TEST: 05/30/2020 04:39:01

Diagnostics Check	OK	04:34	
Air Blank	0.000	04:35	
Control Test	0.081	04:35	
Air Blank	0.000	04:36	
Subject Sample	# 1	04:37	Slope Not Met
Breath Volume	1.101	04:37	
Air Blank	0.000	04:38	
Control Test	0.080	04:38	
Air Blank	0.000	04:38	
Diagnostics Check	OK	04:39	

Data Download Date/Time: 06/17/2020 23:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON, N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: FIFE

DATE/TIME OF TEST: 05/30/2020 05:08:02

Diagnostics Check	OK	04:57	
Air Blank	0.000	04:58	
Control Test	0.080	04:58	
Air Blank	0.000	04:58	
Subject Sample	# 1 0.153	05:02	Volume Not Met
Breath Volume	0.390	05:02	
Air Blank	0.000	05:02	
Air Blank	0.000	05:04	
Subject Sample	# 2	05:06	Subject Test Refused
Breath Volume	1.101	05:06	
Air Blank	0.000	05:07	
Control Test	0.079	05:07	
Air Blank	0.000	05:07	
Diagnostics Check	OK	05:08	

Data Download Date/Time: 06/17/2020 23:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRIDWELL,C
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: BUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: GIRALDO

DATE/TIME OF TEST: 05/31/2020 22:59:24

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.080	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.075	22:54
Breath Volume	4.324	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.078	22:57
Breath Volume	3.929	22:57
Air Blank	0.000	22:58
Control Test	0.080	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 06/17/2020 23:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAHEY,M,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: BARBER

DATE/TIME OF TEST: 05/01/2020 02:45:08

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.138	02:40
Breath Volume	3.671	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.130	02:43
Breath Volume	2.207	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 05/27/2020 01:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAHEY, MARILYN,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BARBER, EDWARD
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 05/02/2020 20:33:48

Diagnostics Check	OK	20:25	
Air Blank	0.000	20:25	
Control Test	0.080	20:26	
Air Blank	0.000	20:26	
Subject Sample	# 1	20:29	No Sample Provided
Breath Volume	0.000	20:29	
Air Blank	0.000	20:30	
Air Blank	0.000	20:32	
Subject Sample	# 2	20:32	Subject Test Refused
Breath Volume	0.000	20:32	
Air Blank	0.000	20:32	
Control Test	0.080	20:33	
Air Blank	0.000	20:33	
Diagnostics Check	OK	20:33	

Data Download Date/Time: 05/27/2020 01:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FALL,JORDAN
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: NOFFA

DATE/TIME OF TEST: 05/03/2020 19:31:20

Diagnostics Check	OK	19:25
Air Blank	0.000	19:25
Control Test	0.079	19:25
Air Blank	0.000	19:26
Subject Sample # 1	0.284	19:26
Breath Volume	1.476	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample # 2	0.281	19:29
Breath Volume	1.566	19:29
Air Blank	0.000	19:30
Control Test	0.079	19:30
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

Data Download Date/Time: 05/27/2020 01:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,JESSE,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WIGGINS,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: CALLAHAN

DATE/TIME OF TEST: 05/05/2020 19:53:42

Diagnostics Check	OK	19:46
Air Blank	0.000	19:47
Control Test	0.080	19:47
Air Blank	0.000	19:48
Subject Sample # 1	0.114	19:49
Breath Volume	1.980	19:49
Air Blank	0.000	19:49
Air Blank	0.000	19:51
Subject Sample # 2	0.105	19:52
Breath Volume	1.570	19:52
Air Blank	0.000	19:52
Control Test	0.079	19:53
Air Blank	0.000	19:53
Diagnostics Check	OK	19:53

Data Download Date/Time: 05/27/2020 01:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANOE,PAMELA,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GROSIDIER,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: CHARLTON

DATE/TIME OF TEST: 05/07/2020 01:27:13

Diagnostics Check	OK	01:19
Air Blank	0.000	01:20
Control Test	0.081	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.137	01:21
Breath Volume	1.539	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.139	01:25
Breath Volume	1.265	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 05/27/2020 01:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JOHNSON,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: CATES

DATE/TIME OF TEST: 05/09/2020 05:08:48

Diagnostics Check	OK	05:02
Air Blank	0.000	05:02
Control Test	0.080	05:03
Air Blank	0.000	05:03
Subject Sample # 1	0.124	05:04
Breath Volume	3.078	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:06
Subject Sample # 2	0.123	05:07
Breath Volume	2.593	05:07
Air Blank	0.000	05:07
Control Test	0.080	05:08
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 05/27/2020 01:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LANIER,H
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: COURREGÉ

DATE/TIME OF TEST: 05/10/2020 00:56:10

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.080	00:47	
Air Blank	0.000	00:48	
Subject Sample	# 1	00:51	No Sample Provided
Breath Volume	0.000	00:51	
Air Blank	0.000	00:52	
Air Blank	0.000	00:54	
Subject Sample	# 2	00:54	Subject Test Refused
Breath Volume	0.000	00:54	
Air Blank	0.000	00:55	
Control Test	0.080	00:55	
Air Blank	0.000	00:56	
Diagnostics Check	OK	00:56	

Data Download Date/Time: 05/27/2020 01:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANOE,PAMELA,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JOHNSON,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: LEBO

DATE/TIME OF TEST: 05/10/2020 02:40:36

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.079	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.195	02:36
Breath Volume	1.375	02:36
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.208	02:39
Breath Volume	1.761	02:39
Air Blank	0.000	02:39
Control Test	0.080	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 05/27/2020 01:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANELLA,E,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAUMGARTNER,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:27

SUBJECT'S LAST NAME: DAVE

DATE/TIME OF TEST: 05/10/2020 16:00:27

Diagnostics Check	OK	15:57	
Air Blank	0.000	15:58	
Control Test	0.080	15:58	
Air Blank	0.000	15:58	
Subject Sample	# 1	15:59	Subject Test Refused
Breath Volume	0.000	15:59	
Air Blank	0.000	15:59	
Control Test	0.079	15:59	
Air Blank	0.000	16:00	
Diagnostics Check	OK	16:00	

Data Download Date/Time: 05/27/2020 01:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WIDERGREN,DAVID
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: MAHAN

DATE/TIME OF TEST: 05/12/2020 03:59:50

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.080	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.000	03:55
Breath Volume	3.000	03:55
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.000	03:58
Breath Volume	3.355	03:58
Air Blank	0.000	03:58
Control Test	0.080	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 05/27/2020 01:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HARRIS,A.M.
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:22

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/12/2020 16:03:12

Diagnostics Check	OK	15:55
Air Blank	0.000	15:56
Control Test	0.080	15:56
Air Blank	0.000	15:56
Subject Sample # 1	0.000	15:58
Breath Volume	1.605	15:58
Air Blank	0.000	15:59
Air Blank	0.000	16:01
Subject Sample # 2	0.000	16:01
Breath Volume	1.667	16:01
Air Blank	0.000	16:02
Control Test	0.080	16:02
Air Blank	0.000	16:03
Diagnostics Check	OK	16:03

Data Download Date/Time: 05/27/2020 01:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANOE,P,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JOHNSON,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: ESTES

DATE/TIME OF TEST: 05/14/2020 02:31:14

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.080	02:22	
Air Blank	0.000	02:22	
Subject Sample	# 1	02:25	No Sample Provided
Breath Volume	0.000	02:25	
Air Blank	0.000	02:26	
Air Blank	0.000	02:28	
Subject Sample	# 2	02:29	Subject Test Refused
Breath Volume	0.000	02:29	
Air Blank	0.000	02:30	
Control Test	0.079	02:30	
Air Blank	0.000	02:31	
Diagnostics Check	OK	02:31	

Data Download Date/Time: 05/27/2020 01:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAHEY, MARILYN,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS, S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 05/16/2020 04:35:37

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.080	04:30
Air Blank	0.000	04:30
Subject Sample # 1	0.000	04:31
Breath Volume	3.152	04:31
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.000	04:34
Breath Volume	3.097	04:34
Air Blank	0.000	04:34
Control Test	0.079	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 05/27/2020 01:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HANEY,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: BUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: MAHONEY

DATE/TIME OF TEST: 05/16/2020 19:01:22

Diagnostics Check	OK	18:55
Air Blank	0.000	18:55
Control Test	0.079	18:55
Air Blank	0.000	18:56
Subject Sample # 1	0.076	18:56
Breath Volume	1.597	18:56
Air Blank	0.000	18:57
Air Blank	0.000	18:59
Subject Sample # 2	0.074	18:59
Breath Volume	1.722	18:59
Air Blank	0.000	19:00
Control Test	0.080	19:00
Air Blank	0.000	19:01
Diagnostics Check	OK	19:01

Data Download Date/Time: 05/27/2020 01:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAHEY, MARILYN,
BREATH TEST OPERATOR AGENCY: Orange Park PD
ARRESTING OFFICER: CRUZ, C
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: CABBAGESTALK

DATE/TIME OF TEST: 05/18/2020 03:29:40

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:21	
Control Test	0.079	03:21	
Air Blank	0.000	03:22	
Subject Sample	# 1	03:25	No Sample Provided
Breath Volume	0.000	03:25	
Air Blank	0.000	03:25	
Air Blank	0.000	03:27	
Subject Sample	# 2	03:28	Subject Test Refused
Breath Volume	0.000	03:28	
Air Blank	0.000	03:28	
Control Test	0.079	03:29	
Air Blank	0.000	03:29	
Diagnostics Check	OK	03:29	

Data Download Date/Time: 05/27/2020 01:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANOE,PAM,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SPAULDING,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 05/19/2020 03:25:41

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.079	03:18
Air Blank	0.000	03:18
Subject Sample # 1	0.203	03:21
Breath Volume	1.988	03:21
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.214	03:24
Breath Volume	2.726	03:24
Air Blank	0.000	03:24
Control Test	0.080	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 05/27/2020 01:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,JESSE,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HOWELL,L
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: LARKIN

DATE/TIME OF TEST: 05/19/2020 22:42:14

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.080	22:37
Air Blank	0.000	22:37
Subject Sample # 1	0.178	22:37
Breath Volume	1.699	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.182	22:40
Breath Volume	2.812	22:40
Air Blank	0.000	22:41
Control Test	0.079	22:41
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 05/27/2020 01:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANOE,PAM,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SPAULDING,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 05/23/2020 00:55:57

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.079	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.166	00:50
Breath Volume	1.453	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:52
Subject Sample # 2	0.166	00:54
Breath Volume	2.359	00:54
Air Blank	0.000	00:55
Control Test	0.079	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 05/27/2020 01:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANOE,PAM,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TOLICK,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 05/23/2020 02:48:30

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.079	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.139	02:43
Breath Volume	2.144	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.149	02:46
Breath Volume	2.031	02:46
Air Blank	0.000	02:47
Control Test	0.079	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 05/27/2020 01:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANELLA,E,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HANEY,M
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: BUI
ARREST TIME: 16:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:16

SUBJECT'S LAST NAME: TUDER

DATE/TIME OF TEST: 05/23/2020 17:46:02

Diagnostics Check	OK	17:39
Air Blank	0.000	17:39
Control Test	0.079	17:40
Air Blank	0.000	17:40
Subject Sample # 1	0.111	17:41
Breath Volume	1.652	17:41
Air Blank	0.000	17:42
Air Blank	0.000	17:44
Subject Sample # 2	0.117	17:44
Breath Volume	1.332	17:44
Air Blank	0.000	17:45
Control Test	0.078	17:45
Air Blank	0.000	17:45
Diagnostics Check	OK	17:46

Data Download Date/Time: 05/27/2020 01:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANOE,PAM,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MACK,A.
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: RAWLINS

DATE/TIME OF TEST: 05/24/2020 02:49:47

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:37	
Control Test	0.079	02:37	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:40	No Sample Provided
Breath Volume	0.000	02:40	
Air Blank	0.000	02:41	
Air Blank	0.000	02:43	
Subject Sample	# 2	0.155	02:45
Breath Volume	1.210	02:45	
Air Blank	0.000	02:45	
Air Blank	0.000	02:47	
Subject Sample	# 3	0.156	02:48
Breath Volume	2.269	02:48	
Air Blank	0.000	02:48	
Control Test	0.078	02:49	
Air Blank	0.000	02:49	
Diagnostics Check	OK	02:49	

Data Download Date/Time: 05/27/2020 01:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANOE,PAM,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TAPAWAN,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: PRENDES

DATE/TIME OF TEST: 05/24/2020 04:21:32

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.263	04:15
Breath Volume	1.480	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.257	04:19
Breath Volume	1.421	04:19
Air Blank	0.000	04:20
Control Test	0.078	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 05/27/2020 01:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,J,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HANEY,M
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: TSCHUSCHKE

DATE/TIME OF TEST: 05/24/2020 20:16:20

Diagnostics Check	OK	20:04	
Air Blank	0.000	20:05	
Control Test	0.079	20:05	
Air Blank	0.000	20:05	
Subject Sample	# 1	20:09	No Sample Provided
Breath Volume	0.000	20:09	
Air Blank	0.000	20:09	
Air Blank	0.000	20:11	
Subject Sample	# 2	20:11	
Breath Volume	1.777	20:11	
Air Blank	0.000	20:12	
Air Blank	0.000	20:14	
Subject Sample	# 3	20:14	
Breath Volume	1.644	20:14	
Air Blank	0.000	20:15	
Control Test	0.079	20:15	
Air Blank	0.000	20:16	
Diagnostics Check	OK	20:16	

Data Download Date/Time: 05/27/2020 01:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANOE,P,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TOLICK,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: SORRELLS

DATE/TIME OF TEST: 05/24/2020 23:52:16

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:42	
Control Test	0.079	23:42	
Air Blank	0.000	23:42	
Subject Sample	# 1	23:46	No Sample Provided
Breath Volume	0.000	23:46	
Air Blank	0.000	23:46	
Air Blank	0.000	23:48	
Subject Sample	# 2	23:50	Subject Test Refused
Breath Volume	0.000	23:50	
Air Blank	0.000	23:51	
Control Test	0.080	23:51	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:52	

Data Download Date/Time: 05/27/2020 01:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,JESSE,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CAMP,B
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 05/28/2020 22:04:27

Diagnostics Check	OK	21:58
Air Blank	0.000	21:58
Control Test	0.080	21:58
Air Blank	0.000	21:59
Subject Sample # 1	0.000	21:59
Breath Volume	1.359	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.000	22:03
Breath Volume	1.203	22:03
Air Blank	0.000	22:03
Control Test	0.079	22:03
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 06/30/2020 05:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENBO,ANDREW,R
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: DIGIROLAMO,N
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: DOBSON

DATE/TIME OF TEST: 05/06/2020 19:16:55

Diagnostics Check	OK	19:10
Air Blank	0.000	19:10
Control Test	0.080	19:11
Air Blank	0.000	19:11
Subject Sample # 1	0.000	19:12
Breath Volume	2.792	19:12
Air Blank	0.000	19:12
Air Blank	0.000	19:14
Subject Sample # 2	0.000	19:15
Breath Volume	2.097	19:15
Air Blank	0.000	19:16
Control Test	0.075	19:16
Air Blank	0.000	19:16
Diagnostics Check	OK	19:16

Data Download Date/Time: 05/18/2020 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,DEBBIE,L
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HALL,LACEY
ARRESTING OFFICER AGENCY: Zephyrhills PD
LAST AGENCY INSPECTION: 04/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 05/11/2020 22:32:36

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:31	
Control Test	0.079	22:31	
Air Blank	0.000	22:32	
		22:32	RFI Detect
Air Blank		22:32	RFI Detect

Data Download Date/Time: 05/18/2020 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,DEBBIE,L
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HALL,LACEY
ARRESTING OFFICER AGENCY: Zephyrhills PD
LAST AGENCY INSPECTION: 04/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 05/11/2020 22:41:56

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.079	22:36
Air Blank	0.000	22:36
Subject Sample # 1	0.060	22:37
Breath Volume	1.824	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.062	22:40
Breath Volume	2.246	22:40
Air Blank	0.000	22:41
Control Test	0.079	22:41
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 05/18/2020 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,DEBBIE,L
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: WATSON,WILLIAM
ARRESTING OFFICER AGENCY: Dade City PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: ORDAZ

DATE/TIME OF TEST: 05/31/2020 04:06:14

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.078	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.110	04:01
Breath Volume	3.027	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.119	04:04
Breath Volume	4.023	04:04
Air Blank	0.000	04:05
Control Test	0.079	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 06/24/2020 10:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: FALTER

DATE/TIME OF TEST: 05/22/2020 00:33:05

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.080	00:27
Air Blank	0.000	00:27
Subject Sample # 1	0.175	00:28
Breath Volume	1.367	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:30
Subject Sample # 2	0.177	00:31
Breath Volume	1.890	00:31
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 06/03/2020 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 05/22/2020 23:42:10

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.080	23:37
Air Blank	0.000	23:37
Subject Sample # 1	0.180	23:37
Breath Volume	1.742	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.180	23:40
Breath Volume	1.738	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 06/03/2020 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: CHAMBERS

DATE/TIME OF TEST: 05/25/2020 23:17:33

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.081	23:12
Air Blank	0.000	23:12
Subject Sample # 1	0.195	23:13
Breath Volume	2.019	23:13
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.198	23:16
Breath Volume	1.714	23:16
Air Blank	0.000	23:16
Control Test	0.080	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 06/03/2020 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/05/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22419080A3

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 22:44

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: DAMSCHRODER

DATE/TIME OF TEST: 05/03/2020 23:39:07

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.080	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.209	23:29
Breath Volume	1.132	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.233	23:33
Breath Volume	1.449	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 3	0.212	23:37
Breath Volume	1.414	23:37
Air Blank	0.000	23:38
Control Test	0.079	23:38
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 05/10/2020 12:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/05/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22419080A3

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 02:51

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: ALI

DATE/TIME OF TEST: 05/06/2020 03:44:08

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:39	
Control Test	0.080	03:40	
Air Blank	0.000	03:40	
Subject Sample	# 1	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:43	
Control Test	0.080	03:43	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 05/10/2020 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/05/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22419080A3

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 22:50

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: ORFIELD

DATE/TIME OF TEST: 05/07/2020 23:31:51

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.081	23:26
Air Blank	0.000	23:26
Subject Sample # 1	0.231	23:27
Breath Volume	2.160	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:29
Subject Sample # 2	0.246	23:30
Breath Volume	2.054	23:30
Air Blank	0.000	23:30
Control Test	0.081	23:31
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 05/10/2020 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/05/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22419080A3

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 22:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: GILKEY

DATE/TIME OF TEST: 05/08/2020 22:57:06

Diagnostics Check	OK	22:50
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.276	22:52
Breath Volume	2.628	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.280	22:55
Breath Volume	1.953	22:55
Air Blank	0.000	22:56
Control Test	0.080	22:56
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 05/10/2020 12:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 04/05/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22419080A3

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 00:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 05/10/2020 01:35:45

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:27	
Control Test	0.081	01:27	
Air Blank	0.000	01:28	
Subject Sample	# 1	01:31	No Sample Provided
Breath Volume	0.000	01:31	
Air Blank	0.000	01:31	
Air Blank	0.000	01:33	
Subject Sample	# 2	01:34	Subject Test Refused
Breath Volume	0.000	01:34	
Air Blank	0.000	01:34	
Control Test	0.081	01:35	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 05/10/2020 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: EAMES

DATE/TIME OF TEST: 05/11/2020 00:06:24

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:55	
Control Test	0.080	23:55	
Air Blank	0.000	23:56	
Subject Sample	# 2	23:59	No Sample Provided
Breath Volume	0.000	23:59	
Air Blank	0.000	23:59	
Air Blank	0.000	00:01	
Subject Sample	# 1	00:05	No Sample Provided
Breath Volume	0.000	00:05	
Air Blank	0.000	00:05	
Control Test	0.080	00:05	
Air Blank	0.000	00:06	
Diagnostics Check	OK	00:06	

Data Download Date/Time: 06/03/2020 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22419080A3

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 23:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/15/2020 00:15:51

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.081	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.232	00:10
Breath Volume	2.687	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.239	00:14
Breath Volume	3.011	00:14
Air Blank	0.000	00:14
Control Test	0.080	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 06/03/2020 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILSON,CHRISTOPHER,R

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: WILSON,CHRISTOPHER

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22419080A3

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 14:54

GENDER: Female

OBSERVATION PERIOD BEGAN: 15:11

SUBJECT'S LAST NAME: HASTINGS

DATE/TIME OF TEST: 05/16/2020 15:40:18

Diagnostics Check	OK	15:32
Air Blank	0.000	15:32
Control Test	0.080	15:32
Air Blank	0.000	15:33
Subject Sample # 1	0.000	15:35
Breath Volume	1.417	15:35
Air Blank	0.000	15:36
Air Blank	0.000	15:38
Subject Sample # 2	0.000	15:38
Breath Volume	1.550	15:38
Air Blank	0.000	15:39
Control Test	0.081	15:39
Air Blank	0.000	15:40
Diagnostics Check	OK	15:40

Data Download Date/Time: 06/03/2020 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 05/21/2020 02:04:19

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:01	
Control Test	0.080	02:01	
Air Blank	0.000	02:02	
Subject Sample	# 1	02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:03	
Control Test	0.081	02:03	
Air Blank	0.000	02:04	
Diagnostics Check	OK	02:04	

Data Download Date/Time: 06/03/2020 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22419080A3

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 02:31

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: ROHRER

DATE/TIME OF TEST: 05/22/2020 03:25:19

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.249	03:20
Breath Volume	2.515	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.233	03:23
Breath Volume	1.476	03:23
Air Blank	0.000	03:24
Control Test	0.080	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 06/03/2020 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: BERTONIERE

DATE/TIME OF TEST: 05/24/2020 22:48:28

Diagnostics Check	OK	22:41
Air Blank	0.000	22:42
Control Test	0.081	22:42
Air Blank	0.000	22:43
Subject Sample # 1	0.220	22:44
Breath Volume	3.730	22:44
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.226	22:47
Breath Volume	2.171	22:47
Air Blank	0.000	22:47
Control Test	0.081	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 06/03/2020 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASBURY,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: ASBURY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: GROSS

DATE/TIME OF TEST: 05/24/2020 23:24:30

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.081	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.213	23:19
Breath Volume	1.351	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.208	23:23
Breath Volume	1.734	23:23
Air Blank	0.000	23:23
Control Test	0.081	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 06/03/2020 20:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER, MATTHEW, D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER, MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22419080A3

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 23:04

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: RUFF

DATE/TIME OF TEST: 05/24/2020 23:48:29

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:46	
Control Test	0.079	23:46	
Air Blank	0.000	23:46	
Subject Sample	# 1	23:47	Subject Test Refused
Breath Volume	0.000	23:47	
Air Blank	0.000	23:47	
Control Test	0.080	23:47	
Air Blank	0.000	23:48	
Diagnostics Check	OK	23:48	

Data Download Date/Time: 06/03/2020 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22419080A3

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 02:09

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 05/25/2020 03:01:26

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.081	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.150	02:56
Breath Volume	2.492	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.145	02:59
Breath Volume	2.152	02:59
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 06/03/2020 20:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22419080A3

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 02:28

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 05/25/2020 03:23:54

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:16	
Control Test	0.080	03:16	
Air Blank	0.000	03:17	
Subject Sample	# 1 0.238	03:19	
Breath Volume	1.218	03:19	
Air Blank	0.000	03:19	
Air Blank	0.000	03:21	
Subject Sample	# 2	03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:23	
Control Test	0.080	03:23	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 06/03/2020 20:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KATZ,JUSTIN,T
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: KATZ,JUSTIN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:51

SUBJECT'S LAST NAME: KINGERY

DATE/TIME OF TEST: 05/25/2020 14:15:13

Diagnostics Check	OK	14:11	
Air Blank	0.000	14:12	
Control Test	0.081	14:12	
Air Blank	0.000	14:13	
Subject Sample	# 1	14:13	Subject Test Refused
Breath Volume	0.000	14:13	
Air Blank	0.000	14:14	
Control Test	0.081	14:14	
Air Blank	0.000	14:15	
Diagnostics Check	OK	14:15	

Data Download Date/Time: 06/03/2020 20:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22419080A3

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 23:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: VEGA RAMIREZ

DATE/TIME OF TEST: 05/26/2020 00:04:25

Diagnostics Check	OK	23:49	
Air Blank	0.000	23:50	
Control Test	0.081	23:50	
Air Blank	0.000	23:51	
Subject Sample	# 2 0.081	23:53	
Breath Volume	1.953	23:53	
Air Blank	0.000	23:53	
Air Blank	0.000	23:55	
Subject Sample	# 3 0.073	23:58	Volume Not Met
Breath Volume	0.472	23:58	
Air Blank	0.000	23:59	
Air Blank	0.000	00:01	
Subject Sample	# 1 0.081	00:02	
Breath Volume	1.843	00:02	
Air Blank	0.000	00:03	
Control Test	0.080	00:03	
Air Blank	0.000	00:04	
Diagnostics Check	OK	00:04	

Data Download Date/Time: 06/03/2020 20:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22419080A3

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 21:57

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: BLASIOLI

DATE/TIME OF TEST: 05/29/2020 22:54:02

Diagnostics Check	OK	22:51	
Air Blank	0.000	22:51	
Control Test	0.080	22:51	
Air Blank	0.000	22:52	
Subject Sample	# 1	22:52	Subject Test Refused
Breath Volume	0.000	22:52	
Air Blank	0.000	22:53	
Control Test	0.080	22:53	
Air Blank	0.000	22:53	
Diagnostics Check	OK	22:54	

Data Download Date/Time: 06/03/2020 20:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22419080A3

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 00:11

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: HARTMAN

DATE/TIME OF TEST: 05/30/2020 00:48:53

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:46	
Control Test	0.080	00:46	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:47	Subject Test Refused
Breath Volume	0.000	00:47	
Air Blank	0.000	00:47	
Control Test	0.081	00:48	
Air Blank	0.000	00:48	
Diagnostics Check	OK	00:48	

Data Download Date/Time: 06/03/2020 20:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: KRUEGER

DATE/TIME OF TEST: 05/30/2020 20:04:48

Diagnostics Check	OK	19:51	
Air Blank	0.000	19:52	
Control Test	0.080	19:52	
Air Blank	0.000	19:53	
Subject Sample	# 1 0.196	19:56	Volume Not Met
Breath Volume	0.171	19:56	
Air Blank	0.000	19:56	
Air Blank	0.000	19:58	
Subject Sample	# 2 0.199	19:59	
Breath Volume	2.093	19:59	
Air Blank	0.000	20:00	
Air Blank	0.000	20:02	
Subject Sample	# 3 0.193	20:03	
Breath Volume	1.296	20:03	
Air Blank	0.000	20:03	
Control Test	0.079	20:04	
Air Blank	0.000	20:04	
Diagnostics Check	OK	20:04	

Data Download Date/Time: 06/03/2020 20:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22419080A3

DRY GAS EXPIRY DATE: 10/05/2021

VIOLATION CODE: DUI

ARREST TIME: 02:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SHEEHY

DATE/TIME OF TEST: 05/31/2020 03:32:05

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.064	03:27
Breath Volume	2.804	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.061	03:30
Breath Volume	2.085	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:32

Data Download Date/Time: 06/03/2020 20:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HAMMELMAN,NICHOLAS
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 05/07/2020 20:30:29

Diagnostics Check	OK	20:24
Air Blank	0.000	20:25
Control Test	0.082	20:25
Air Blank	0.000	20:25
Subject Sample # 1	0.000	20:26
Breath Volume	2.660	20:26
Air Blank	0.000	20:26
Air Blank	0.000	20:28
Subject Sample # 2	0.000	20:29
Breath Volume	1.878	20:29
Air Blank	0.000	20:29
Control Test	0.081	20:29
Air Blank	0.000	20:30
Diagnostics Check	OK	20:30

Data Download Date/Time: 05/16/2020 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER, BLAKE, E
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CARPENTER, BLAKE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 05/04/2020 22:18:28

Diagnostics Check	OK	22:12
Air Blank	0.000	22:12
Control Test	0.080	22:13
Air Blank	0.000	22:13
Subject Sample # 1	0.156	22:14
Breath Volume	2.011	22:14
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample # 2	0.149	22:16
Breath Volume	1.675	22:16
Air Blank	0.000	22:17
Control Test	0.079	22:17
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 05/27/2020 02:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUTIGLIANO,ROBERT,N
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: LOPEZ,HECTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: RUTIGLIANO

DATE/TIME OF TEST: 05/05/2020 20:24:58

Diagnostics Check	OK	20:18
Air Blank	0.000	20:18
Control Test	0.081	20:19
Air Blank	0.000	20:19
Subject Sample # 1	0.128	20:20
Breath Volume	2.464	20:20
Air Blank	0.000	20:21
Air Blank	0.000	20:22
Subject Sample # 2	0.128	20:23
Breath Volume	2.519	20:23
Air Blank	0.000	20:24
Control Test	0.080	20:24
Air Blank	0.000	20:24
Diagnostics Check	OK	20:24

Data Download Date/Time: 05/27/2020 02:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,TIMOTHY,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BLAIR,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: MERCANDETTI

DATE/TIME OF TEST: 05/23/2020 03:14:16

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.134	03:08
Breath Volume	2.117	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.134	03:12
Breath Volume	2.140	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 05/27/2020 02:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWISHER,EDWARD,T
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SWISHER,EDWARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: DODD

DATE/TIME OF TEST: 05/23/2020 03:24:32

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.201	03:20
Breath Volume	2.195	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.192	03:23
Breath Volume	2.277	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 05/27/2020 02:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,HECTOR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: LOPEZ,HECTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: RODES

DATE/TIME OF TEST: 05/30/2020 22:15:09

Diagnostics Check	OK	22:08
Air Blank	0.000	22:09
Control Test	0.081	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.260	22:10
Breath Volume	1.875	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.270	22:13
Breath Volume	2.582	22:13
Air Blank	0.000	22:14
Control Test	0.080	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 06/23/2020 01:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,LIN,
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: GRAY,LIN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 07:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: JONES JR

DATE/TIME OF TEST: 05/27/2020 08:47:12

Diagnostics Check	OK	08:40
Air Blank	0.000	08:41
Control Test	0.079	08:41
Air Blank	0.000	08:42
Subject Sample # 1	0.000	08:42
Breath Volume	3.109	08:42
Air Blank	0.000	08:43
Air Blank	0.000	08:45
Subject Sample # 2	0.000	08:45
Breath Volume	3.050	08:45
Air Blank	0.000	08:46
Control Test	0.079	08:46
Air Blank	0.000	08:47
Diagnostics Check	OK	08:47

Data Download Date/Time: 07/08/2020 07:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 05/02/2020 00:37:01

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:29	
Control Test	0.080	00:29	
Air Blank	0.000	00:29	
Subject Sample	# 1	00:33	No Sample Provided
Breath Volume	0.000	00:33	
Air Blank	0.000	00:33	
Air Blank	0.000	00:35	
Subject Sample	# 2	00:35	Subject Test Refused
Breath Volume	0.000	00:35	
Air Blank	0.000	00:36	
Control Test	0.079	00:36	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:37	

Data Download Date/Time: 05/16/2020 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETRU,MIROS,N
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PETRU,MIROS
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 04/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: BUI
ARREST TIME: 06:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 05/02/2020 21:00:07

Diagnostics Check	OK	20:54
Air Blank	0.000	20:54
Control Test	0.080	20:54
Air Blank	0.000	20:55
Subject Sample # 1	0.091	20:55
Breath Volume	2.445	20:55
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample # 2	0.085	20:58
Breath Volume	1.367	20:58
Air Blank	0.000	20:59
Control Test	0.080	20:59
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time: 05/16/2020 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,SUMMER,N
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,SUMMER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: MARKS

DATE/TIME OF TEST: 05/03/2020 02:26:28

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.079	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.211	02:21
Breath Volume	2.121	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.219	02:24
Breath Volume	2.550	02:24
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 05/16/2020 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: EDMOND

DATE/TIME OF TEST: 05/07/2020 22:03:21

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.079	21:57
Air Blank	0.000	21:57
Subject Sample # 1	0.160	21:59
Breath Volume	2.480	21:59
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.144	22:01
Breath Volume	1.507	22:01
Air Blank	0.000	22:02
Control Test	0.078	22:02
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 05/16/2020 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRINIDAD,RICKY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: TRINIDAD,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: SPRECHER

DATE/TIME OF TEST: 05/08/2020 20:18:29

Diagnostics Check	OK	20:15	
Air Blank	0.000	20:15	
Control Test	0.080	20:15	
Air Blank	0.000	20:16	
Subject Sample	# 1 0.121	20:16	
Breath Volume	2.210	20:16	
Air Blank		20:17	RFI Detect
Air Blank	0.000	20:17	
Control Test	0.081	20:17	
Air Blank	0.000	20:18	
Diagnostics Check	OK	20:18	

Data Download Date/Time: 05/16/2020 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRINIDAD,RICKY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: TRINIDAD,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: SPRECHER

DATE/TIME OF TEST: 05/08/2020 20:27:24

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.079	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.119	20:22
Breath Volume	1.964	20:22
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.114	20:25
Breath Volume	1.695	20:25
Air Blank	0.000	20:26
Control Test	0.080	20:26
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 05/16/2020 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSTOS,ALEJANDRO,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LEE,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: FRICKE

DATE/TIME OF TEST: 05/10/2020 02:26:42

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.079	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.130	02:22
Breath Volume	3.789	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.126	02:25
Breath Volume	3.550	02:25
Air Blank	0.000	02:25
Control Test	0.080	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 05/16/2020 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHANSEN,LLOYD,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HULL,DONALD
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: JENNINGS

DATE/TIME OF TEST: 05/10/2020 21:17:55

Diagnostics Check	OK	21:11
Air Blank	0.000	21:12
Control Test	0.080	21:12
Air Blank	0.000	21:12
Subject Sample # 1	0.000	21:13
Breath Volume	4.519	21:13
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.000	21:16
Breath Volume	3.968	21:16
Air Blank	0.000	21:17
Control Test	0.078	21:17
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 05/16/2020 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK, TYLER, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIRK, TYLER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:28

SUBJECT'S LAST NAME: SMALLWOOD

DATE/TIME OF TEST: 05/11/2020 15:51:16

Diagnostics Check	OK	15:48	
Air Blank	0.000	15:48	
Control Test	0.080	15:49	
Air Blank	0.000	15:49	
Subject Sample	# 1	15:49	Subject Test Refused
Breath Volume	0.000	15:49	
Air Blank	0.000	15:50	
Control Test	0.077	15:50	
Air Blank	0.000	15:51	
Diagnostics Check	OK	15:51	

Data Download Date/Time: 05/16/2020 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON, MATTHEW, F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: BENDER

DATE/TIME OF TEST: 05/13/2020 16:26:08

Diagnostics Check	OK	16:19
Air Blank	0.000	16:20
Control Test	0.080	16:20
Air Blank	0.000	16:20
Subject Sample # 1	0.008	16:21
Breath Volume	2.199	16:21
Air Blank	0.000	16:22
Air Blank	0.000	16:24
Subject Sample # 2	0.000	16:24
Breath Volume	2.566	16:24
Air Blank	0.000	16:25
Control Test	0.078	16:25
Air Blank	0.000	16:26
Diagnostics Check	OK	16:26

Data Download Date/Time: 05/16/2020 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETRU,MIROS,N
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PETRU,MIROS
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 04/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: BUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 05/14/2020 01:50:12

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.050	01:45
Breath Volume	1.628	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.051	01:48
Breath Volume	1.617	01:48
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 05/16/2020 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,SUMMER,N
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PATRICK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: HILLARD

DATE/TIME OF TEST: 05/16/2020 21:57:21

Diagnostics Check	OK	21:51
Air Blank	0.000	21:51
Control Test	0.080	21:51
Air Blank	0.000	21:52
Subject Sample # 1	0.296	21:52
Breath Volume	2.886	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.296	21:55
Breath Volume	2.644	21:55
Air Blank	0.000	21:56
Control Test	0.079	21:56
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

Data Download Date/Time: 07/01/2020 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,SUMMER,N
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,SUMMER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: FRENCH

DATE/TIME OF TEST: 05/16/2020 23:43:01

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.079	23:32
Air Blank	0.000	23:33
Subject Sample # 1	0.264	23:36
Breath Volume	1.769	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.258	23:41
Breath Volume	1.847	23:41
Air Blank	0.000	23:42
Control Test	0.078	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:43

Data Download Date/Time: 07/01/2020 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 05/17/2020 04:33:30

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.079	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.074	04:28
Breath Volume	1.816	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.073	04:31
Breath Volume	1.320	04:31
Air Blank	0.000	04:32
Control Test	0.079	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 07/01/2020 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEQUEIRA,NATHAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SEQUEIRA,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:26

SUBJECT'S LAST NAME: JELENOVA

DATE/TIME OF TEST: 05/17/2020 19:37:26

Diagnostics Check	OK	19:27
Air Blank	0.000	19:28
Control Test	0.081	19:28
Air Blank	0.000	19:28
Subject Sample # 1	0.311	19:31
Breath Volume	1.808	19:31
Air Blank	0.000	19:32
Air Blank	0.000	19:34
Subject Sample # 2	0.312	19:35
Breath Volume	1.175	19:35
Air Blank	0.000	19:36
Control Test	0.077	19:36
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time: 07/01/2020 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHANSEN,LLOYD,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: YANZER,NOAH
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 05/20/2020 04:31:01

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.079	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.249	04:26
Breath Volume	2.742	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:28
Subject Sample # 2	0.236	04:29
Breath Volume	1.792	04:29
Air Blank	0.000	04:30
Control Test	0.077	04:30
Air Blank	0.000	04:30
Diagnostics Check	OK	04:31

Data Download Date/Time: 07/01/2020 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON, MATTHEW, F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: O'DONNELL

DATE/TIME OF TEST: 05/23/2020 13:46:10

Diagnostics Check	OK	13:40
Air Blank	0.000	13:40
Control Test	0.081	13:40
Air Blank	0.000	13:41
Subject Sample # 1	0.000	13:41
Breath Volume	3.289	13:41
Air Blank	0.000	13:42
Air Blank	0.000	13:44
Subject Sample # 2	0.000	13:44
Breath Volume	3.933	13:44
Air Blank	0.000	13:45
Control Test	0.080	13:45
Air Blank	0.000	13:46
Diagnostics Check	OK	13:46

Data Download Date/Time: 07/01/2020 13:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONNELL,JACOB,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CONNELL,JACOB
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 05/23/2020 16:44:38

Diagnostics Check	OK	16:41	
Air Blank	0.000	16:42	
Control Test	0.081	16:42	
Air Blank	0.000	16:43	
Subject Sample	# 1	16:43	Subject Test Refused
Breath Volume	0.000	16:43	
Air Blank	0.000	16:43	
Control Test	0.081	16:44	
Air Blank	0.000	16:44	
Diagnostics Check	OK	16:44	

Data Download Date/Time: 07/01/2020 13:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCE,SHNAYDINE,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SPENCE,SHNAYDINE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:16

SUBJECT'S LAST NAME: SCHIPP

DATE/TIME OF TEST: 05/24/2020 13:20:42

Diagnostics Check	OK	13:14
Air Blank	0.000	13:15
Control Test	0.079	13:15
Air Blank	0.000	13:15
Subject Sample # 1	0.276	13:16
Breath Volume	2.207	13:16
Air Blank	0.000	13:16
Air Blank	0.000	13:18
Subject Sample # 2	0.268	13:19
Breath Volume	2.031	13:19
Air Blank	0.000	13:19
Control Test	0.078	13:20
Air Blank	0.000	13:20
Diagnostics Check	OK	13:20

Data Download Date/Time: 07/01/2020 13:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,JOSEPH,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FRANKLIN,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: AMMONS

DATE/TIME OF TEST: 05/24/2020 22:33:27

Diagnostics Check	OK	22:24	
Air Blank	0.000	22:25	
Control Test	0.079	22:25	
Air Blank	0.000	22:26	
Subject Sample	# 1	22:29	No Sample Provided
Breath Volume	0.000	22:29	
Air Blank	0.000	22:29	
Air Blank	0.000	22:31	
Subject Sample	# 2	22:32	Subject Test Refused
Breath Volume	0.000	22:32	
Air Blank	0.000	22:32	
Control Test	0.079	22:32	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:33	

Data Download Date/Time: 07/01/2020 13:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 05/29/2020 14:47:35

Diagnostics Check	OK	14:41
Air Blank	0.000	14:42
Control Test	0.080	14:42
Air Blank	0.000	14:43
Subject Sample # 1	0.000	14:43
Breath Volume	2.265	14:43
Air Blank	0.000	14:43
Air Blank	0.000	14:45
Subject Sample # 2	0.000	14:46
Breath Volume	2.429	14:46
Air Blank	0.000	14:46
Control Test	0.081	14:46
Air Blank	0.000	14:47
Diagnostics Check	OK	14:47

Data Download Date/Time: 07/01/2020 13:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEQUEIRA,NATHAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SEQUEIRA,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:13

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 05/29/2020 18:40:43

Diagnostics Check	OK	18:34
Air Blank	0.000	18:34
Control Test	0.080	18:35
Air Blank	0.000	18:35
Subject Sample # 1	0.350	18:36
Breath Volume	4.496	18:36
Air Blank	0.000	18:36
Air Blank	0.000	18:38
Subject Sample # 2	0.349	18:39
Breath Volume	3.765	18:39
Air Blank	0.000	18:39
Control Test	0.078	18:40
Air Blank	0.000	18:40
Diagnostics Check	OK	18:40

Data Download Date/Time: 07/01/2020 13:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH-LONG,JEFFERSON,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SMITH-LONG,JEFFERSON
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: COCHRAN

DATE/TIME OF TEST: 05/31/2020 00:42:17

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.080	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.095	00:37
Breath Volume	3.898	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.096	00:40
Breath Volume	3.761	00:40
Air Blank	0.000	00:41
Control Test	0.080	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 07/01/2020 13:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK, TYLER, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MEIER, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:03

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 05/31/2020 09:30:10

Diagnostics Check	OK	09:23
Air Blank	0.000	09:24
Control Test	0.080	09:24
Air Blank	0.000	09:25
Subject Sample # 1	0.124	09:25
Breath Volume	1.296	09:25
Air Blank	0.000	09:26
Air Blank	0.000	09:28
Subject Sample # 2	0.122	09:28
Breath Volume	2.425	09:28
Air Blank	0.000	09:29
Control Test	0.080	09:29
Air Blank	0.000	09:30
Diagnostics Check	OK	09:30

Data Download Date/Time: 07/01/2020 13:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: MCGILL

DATE/TIME OF TEST: 05/24/2020 16:32:59

Diagnostics Check	Fail	16:32	Temperature Regulation Test
Air Blank	0.000	16:32	

Data Download Date/Time: 06/01/2020 16:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: MCGILL

DATE/TIME OF TEST: 05/24/2020 17:17:38

Diagnostics Check	OK	17:11
Air Blank	0.000	17:11
Control Test	0.079	17:12
Air Blank	0.000	17:12
Subject Sample # 1	0.000	17:13
Breath Volume	2.085	17:13
Air Blank	0.000	17:13
Air Blank	0.000	17:15
Subject Sample # 2	0.000	17:16
Breath Volume	1.742	17:16
Air Blank	0.000	17:16
Control Test	0.080	17:17
Air Blank	0.000	17:17
Diagnostics Check	OK	17:17

Data Download Date/Time: 06/01/2020 16:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEYMANS,MORNE,
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: SCHOTT,HEATHER
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: NEEL

DATE/TIME OF TEST: 05/30/2020 21:16:25

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.079	21:11
Air Blank	0.000	21:11
Subject Sample # 1	0.171	21:12
Breath Volume	1.230	21:12
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.172	21:14
Breath Volume	1.859	21:14
Air Blank	0.000	21:15
Control Test	0.079	21:15
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 07/27/2020 18:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: NAYLOR,MESHELLE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: FRANCOEUR

DATE/TIME OF TEST: 05/08/2020 18:17:50

Diagnostics Check	OK	18:11
Air Blank	0.000	18:11
Control Test	0.081	18:12
Air Blank	0.000	18:12
Subject Sample # 1	0.000	18:13
Breath Volume	2.054	18:13
Air Blank	0.000	18:13
Air Blank	0.000	18:15
Subject Sample # 2	0.000	18:16
Breath Volume	1.339	18:16
Air Blank	0.000	18:16
Control Test	0.081	18:17
Air Blank	0.000	18:17
Diagnostics Check	OK	18:17

Data Download Date/Time: 05/28/2020 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HILERIO,JOSE
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 04/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 05/23/2020 20:46:47

Diagnostics Check	OK	20:37	
Air Blank	0.000	20:37	
Control Test	0.081	20:38	
Air Blank	0.000	20:38	
Subject Sample	# 1 0.238	20:39	
Breath Volume	1.628	20:39	
Air Blank	0.000	20:39	
Air Blank	0.000	20:41	
Subject Sample	# 2 0.217	20:42	
Breath Volume	1.437	20:42	
Air Blank	0.000	20:43	
Air Blank	0.000	20:44	
Subject Sample	# 3 0.223	20:45	
Breath Volume	2.613	20:45	
Air Blank		20:46	RFI Detect
Air Blank	0.000	20:46	

Data Download Date/Time: 05/28/2020 09:56