

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AUILERA,ANDREW,E  
BREATH TEST OPERATOR AGENCY: OTHER LOCAL  
ARRESTING OFFICER: AGUILERA,ANDREW  
ARRESTING OFFICER AGENCY: OTHER LOCAL  
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SAENZ

DATE/TIME OF TEST: 06/01/2020 00:28:54

---

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.080	00:21
Air Blank		0.000	00:21
Subject Sample	# 1	0.081	00:23
Breath Volume		1.367	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:25
Subject Sample	# 2	0.082	00:27
Breath Volume		1.277	00:27
Air Blank		0.000	00:27
Control Test		0.076	00:28
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

---

Data Download Date/Time: 06/25/2020 11:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUZ,DANIEL,A  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: CRUZ,DANIEL  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: MATHURIN

DATE/TIME OF TEST: 06/01/2020 07:44:04

---

Diagnostics Check	OK	07:41	
Air Blank	0.000	07:41	
Control Test	0.079	07:41	
Air Blank	0.000	07:42	
Subject Sample	# 1	07:42	Subject Test Refused
Breath Volume	0.000	07:42	
Air Blank	0.000	07:43	
Control Test	0.079	07:43	
Air Blank	0.000	07:43	
Diagnostics Check	OK	07:44	

---

Data Download Date/Time: 06/25/2020 11:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FERNANDEZ,O,  
BREATH TEST OPERATOR AGENCY: SWEETWATER PD  
ARRESTING OFFICER: O,FERNANDEZ  
ARRESTING OFFICER AGENCY: SWEETWATER PD  
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: RODRIGUEZACOSTA

DATE/TIME OF TEST: 06/05/2020 00:52:23

---

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:49	
Control Test	0.081	00:49	
Air Blank	0.000	00:50	
Subject Sample	# 1	00:50	Subject Test Refused
Breath Volume	0.000	00:50	
Air Blank	0.000	00:51	
Control Test	0.080	00:51	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

---

Data Download Date/Time: 06/25/2020 11:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FERNANDEZ,O,  
BREATH TEST OPERATOR AGENCY: SWEETWATER PD  
ARRESTING OFFICER: FERNANDEZ,O  
ARRESTING OFFICER AGENCY: SWEETWATER PD  
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: RODRIGUEZACOSTA

DATE/TIME OF TEST: 06/05/2020 01:18:27

---

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.081	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.182	01:13
Breath Volume	1.292	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.183	01:16
Breath Volume	1.773	01:16
Air Blank	0.000	01:17
Control Test	0.080	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

---

Data Download Date/Time: 06/25/2020 11:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: FLORIAN,ROLANDO  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 06/07/2020 02:49:22

---

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.081	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.116	02:44
Breath Volume	4.132	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.117	02:47
Breath Volume	3.910	02:47
Air Blank	0.000	02:48
Control Test	0.078	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

---

Data Download Date/Time: 06/25/2020 11:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: HINOJOSA,DANIEL  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: DIAZ CASTRO

DATE/TIME OF TEST: 06/07/2020 03:16:26

---

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.080	03:13	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:15	
Control Test	0.080	03:15	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

---

Data Download Date/Time: 06/25/2020 11:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: FLORIAN,ROLANDO  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/08/2020 03:37:42

---

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:35	
Control Test	0.081	03:35	
Air Blank	0.000	03:36	
Subject Sample	# 1	03:36	Subject Test Refused
Breath Volume	0.000	03:36	
Air Blank	0.000	03:36	
Control Test	0.081	03:37	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

---

Data Download Date/Time: 06/25/2020 11:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: FLORIAN,R  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: RESTREPO

DATE/TIME OF TEST: 06/10/2020 07:00:38

---

Diagnostics Check	OK	06:57	
Air Blank	0.000	06:58	
Control Test	0.081	06:58	
Air Blank	0.000	06:59	
Subject Sample	# 1	06:59	Subject Test Refused
Breath Volume	0.000	06:59	
Air Blank	0.000	06:59	
Control Test	0.080	07:00	
Air Blank	0.000	07:00	
Diagnostics Check	OK	07:00	

---

Data Download Date/Time: 06/25/2020 11:41



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CUADRA,M,J  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: MANNO,J  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: GUERRERO MONTERREY

DATE/TIME OF TEST: 06/10/2020 20:09:00

---

Diagnostics Check	OK	20:02
Air Blank	0.000	20:03
Control Test	0.080	20:03
Air Blank	0.000	20:03
Subject Sample # 1	0.269	20:04
Breath Volume	1.714	20:04
Air Blank	0.000	20:05
Air Blank	0.000	20:06
Subject Sample # 2	0.253	20:07
Breath Volume	1.570	20:07
Air Blank	0.000	20:08
Control Test	0.079	20:08
Air Blank	0.000	20:08
Diagnostics Check	OK	20:09

---

Data Download Date/Time: 06/25/2020 11:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: GUTIERREZ,KRISTINA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 06/11/2020 08:15:05

---

Diagnostics Check	OK	08:05
Air Blank	0.000	08:06
Control Test	0.080	08:06
Air Blank	0.000	08:07
Subject Sample # 1	0.209	08:09
Breath Volume	1.195	08:09
Air Blank	0.000	08:09
Air Blank	0.000	08:11
Subject Sample # 2	0.217	08:13
Breath Volume	2.113	08:13
Air Blank	0.000	08:14
Control Test	0.077	08:14
Air Blank	0.000	08:15
Diagnostics Check	OK	08:15

---

Data Download Date/Time: 06/25/2020 11:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PACHECO,GEORGE,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: PACHECO,GEORGE  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: LOPEZ BANEGAS

DATE/TIME OF TEST: 06/13/2020 01:49:17

---

Diagnostics Check	OK	01:44	
Air Blank	0.000	01:45	
Control Test	0.081	01:45	
Air Blank	0.000	01:46	
Subject Sample	# 1	01:47	Subject Test Refused
Breath Volume	0.054	01:47	
Air Blank	0.000	01:48	
Control Test	0.081	01:48	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

---

Data Download Date/Time: 06/25/2020 11:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: GUTIERREZ,KRISTINA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 08:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:28

SUBJECT'S LAST NAME: TRUJILLO

DATE/TIME OF TEST: 06/13/2020 09:57:21

---

Diagnostics Check	OK	09:53	
Air Blank	0.000	09:54	
Control Test	0.081	09:54	
Air Blank	0.000	09:55	
Subject Sample	# 1	09:55	Subject Test Refused
Breath Volume	0.000	09:55	
Air Blank	0.000	09:56	
Control Test	0.081	09:56	
Air Blank	0.000	09:57	
Diagnostics Check	OK	09:57	

---

Data Download Date/Time: 06/25/2020 11:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PACHECO,GEORGE,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: PACHECO,GEORGE  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 06/14/2020 00:18:02

---

Diagnostics Check		OK	00:09
Air Blank		0.000	00:09
Control Test		0.080	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.324	00:11
Breath Volume		1.335	00:11
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.328	00:16
Breath Volume		2.015	00:16
Air Blank		0.000	00:17
Control Test		0.079	00:17
Air Blank		0.000	00:17
Diagnostics Check		OK	00:18

---

Data Download Date/Time: 06/25/2020 11:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA,HENRY,E  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: GARCIA,HENRY  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: RAMLOGAN

DATE/TIME OF TEST: 06/14/2020 04:23:19

---

Diagnostics Check		OK	04:17
Air Blank		0.000	04:17
Control Test		0.081	04:18
Air Blank		0.000	04:18
Subject Sample	# 1	0.181	04:18
Breath Volume		2.570	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.183	04:21
Breath Volume		2.617	04:21
Air Blank		0.000	04:22
Control Test		0.079	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

---

Data Download Date/Time: 06/25/2020 11:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAVINON,CARLOS,M  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: SAVINON,CARLOS  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: BERNAL RUIZ

DATE/TIME OF TEST: 06/20/2020 02:13:15

---

Diagnostics Check		OK	02:06
Air Blank		0.000	02:07
Control Test		0.080	02:07
Air Blank		0.000	02:08
Subject Sample	# 1	0.201	02:08
Breath Volume		1.417	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.218	02:11
Breath Volume		2.363	02:11
Air Blank		0.000	02:12
Control Test		0.080	02:12
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

---

Data Download Date/Time: 06/25/2020 11:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FERNANDEZ,O,  
BREATH TEST OPERATOR AGENCY: SWEETWATER PD  
ARRESTING OFFICER: FERNANDEZ,O  
ARRESTING OFFICER AGENCY: SWEETWATER PD  
LAST AGENCY INSPECTION: 06/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: LEIVA CANTILLANO

DATE/TIME OF TEST: 06/28/2020 00:56:43

---

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:48	
Control Test	0.080	00:48	
Air Blank	0.000	00:49	
Subject Sample	# 1	00:52	No Sample Provided
Breath Volume	0.000	00:52	
Air Blank	0.000	00:53	
Air Blank	0.000	00:55	
Subject Sample	# 2	00:55	Subject Test Refused
Breath Volume	0.000	00:55	
Air Blank	0.000	00:55	
Control Test	0.081	00:56	
Air Blank	0.000	00:56	
Diagnostics Check	OK	00:56	

---

Data Download Date/Time: 07/29/2020 14:48



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: HINOJOSA,DANIEL  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 06/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 06/28/2020 03:57:34

---

Diagnostics Check	OK	03:45		
Air Blank	0.000	03:46		
Control Test	0.080	03:46		
Air Blank	0.000	03:46		
Subject Sample	# 1	03:50	No Sample Provided	
Breath Volume	0.000	03:50		
Air Blank	0.000	03:50		
Air Blank	0.000	03:52		
Subject Sample	# 2	0.156	03:55	Volume Not Met
Breath Volume	0.097	03:55		
Air Blank	0.000	03:56		
Control Test	0.081	03:56		
Air Blank	0.000	03:57		
Diagnostics Check	OK	03:57		

---

Data Download Date/Time: 07/29/2020 14:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: HINOJOSA,DANIEL  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 06/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A2  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 06/28/2020 04:09:31

---

Diagnostics Check		OK	04:00	
Air Blank		0.000	04:00	
Control Test		0.079	04:01	
Air Blank		0.000	04:01	
Subject Sample	# 1	0.164	04:04	Volume Not Met
Breath Volume		0.066	04:04	
Air Blank		0.000	04:05	
Air Blank		0.000	04:07	
Subject Sample	# 2		04:08	Subject Test Refused
Breath Volume		0.000	04:08	
Air Blank		0.000	04:08	
Control Test		0.078	04:08	
Air Blank		0.000	04:09	
Diagnostics Check		OK	04:09	

---

Data Download Date/Time: 07/29/2020 14:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARIDA,K,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNSON,G  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:18

SUBJECT'S LAST NAME: CASTILLO VARELA

DATE/TIME OF TEST: 06/06/2020 08:41:43

---

Diagnostics Check	OK	08:38	
Air Blank	0.000	08:39	
Control Test	0.080	08:39	
Air Blank	0.000	08:40	
Subject Sample	# 1	08:40	Subject Test Refused
Breath Volume	0.000	08:40	
Air Blank	0.000	08:40	
Control Test	0.080	08:41	
Air Blank	0.000	08:41	
Diagnostics Check	OK	08:41	

---

Data Download Date/Time: 06/17/2020 23:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARIDA,K,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: TAPLEY,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 06/06/2020 18:32:59

---

Diagnostics Check	OK	18:30	
Air Blank	0.000	18:30	
Control Test	0.080	18:30	
Air Blank	0.000	18:31	
Subject Sample	# 1	18:31	Subject Test Refused
Breath Volume	0.000	18:31	
Air Blank	0.000	18:32	
Control Test	0.080	18:32	
Air Blank	0.000	18:32	
Diagnostics Check	OK	18:32	

---

Data Download Date/Time: 06/17/2020 23:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,CHUCK,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNS,B  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: GOMEZ RUIZ

DATE/TIME OF TEST: 06/07/2020 03:14:26

---

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.081	03:09
Air Blank	0.000	03:09
Subject Sample # 1	0.199	03:10
Breath Volume	2.792	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.204	03:12
Breath Volume	3.074	03:12
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

---

Data Download Date/Time: 06/17/2020 23:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,CHUCK,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FARLEY,J  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: RICHTER

DATE/TIME OF TEST: 06/07/2020 22:13:41

---

Diagnostics Check	OK	22:07
Air Blank	0.000	22:08
Control Test	0.081	22:08
Air Blank	0.000	22:08
Subject Sample # 1	0.000	22:09
Breath Volume	2.695	22:09
Air Blank	0.000	22:09
Air Blank	0.000	22:11
Subject Sample # 2	0.000	22:12
Breath Volume	2.453	22:12
Air Blank	0.000	22:12
Control Test	0.080	22:13
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

---

Data Download Date/Time: 06/17/2020 23:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TUCKER,R,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNS,B  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: DYKES

DATE/TIME OF TEST: 06/08/2020 23:52:46

---

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.080	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.118	23:48
Breath Volume	2.554	23:48
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.107	23:51
Breath Volume	1.777	23:51
Air Blank	0.000	23:51
Control Test	0.080	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

---

Data Download Date/Time: 06/17/2020 23:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,CHUCK,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CHAVEZ,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 06/09/2020 03:56:39

---

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.080	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.058	03:51
Breath Volume	2.339	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.054	03:55
Breath Volume	1.296	03:55
Air Blank	0.000	03:55
Control Test	0.079	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

---

Data Download Date/Time: 06/17/2020 23:11



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,CHUCK,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CHAVEZ,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 06/09/2020 22:52:56

---

Diagnostics Check	OK	22:49	
Air Blank	0.000	22:50	
Control Test	0.080	22:50	
Air Blank	0.000	22:51	
Subject Sample	# 1	22:51	Subject Test Refused
Breath Volume	0.000	22:51	
Air Blank	0.000	22:52	
Control Test	0.080	22:52	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

---

Data Download Date/Time: 06/17/2020 23:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,CHUCK,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SIMMONS,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: CHLEWICKA

DATE/TIME OF TEST: 06/10/2020 05:30:34

---

Diagnostics Check	OK	05:24
Air Blank	0.000	05:24
Control Test	0.080	05:24
Air Blank	0.000	05:25
Subject Sample # 1	0.263	05:26
Breath Volume	2.582	05:26
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample # 2	0.255	05:29
Breath Volume	1.820	05:29
Air Blank	0.000	05:29
Control Test	0.080	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

---

Data Download Date/Time: 06/17/2020 23:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TUCKER,R,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SIMMONS,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: FEY

DATE/TIME OF TEST: 06/11/2020 01:14:31

---

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.080	01:09
Air Blank	0.000	01:09
Subject Sample # 1	0.080	01:10
Breath Volume	2.714	01:10
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.071	01:13
Breath Volume	2.878	01:13
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

---

Data Download Date/Time: 06/17/2020 23:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOVACS, TYLER, K  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: PARRISH, R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: HAMZA

DATE/TIME OF TEST: 06/13/2020 00:40:21

---

Diagnostics Check	OK	00:33
Air Blank	0.000	00:34
Control Test	0.079	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.148	00:35
Breath Volume	2.882	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.147	00:38
Breath Volume	2.546	00:38
Air Blank	0.000	00:39
Control Test	0.079	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

---

Data Download Date/Time: 06/17/2020 23:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT,J,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: TORRES,W  
ARRESTING OFFICER AGENCY: Neptune Beach PD  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: STERRANTINO

DATE/TIME OF TEST: 06/15/2020 03:46:42

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.080	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.072	03:42
Breath Volume	3.375	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.076	03:45
Breath Volume	3.355	03:45
Air Blank	0.000	03:45
Control Test	0.080	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

---

Data Download Date/Time: 06/17/2020 23:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAY,L,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VENTRONE,V  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 11:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: RICHMOND

DATE/TIME OF TEST: 06/15/2020 16:16:47

---

Diagnostics Check	OK	16:10
Air Blank	0.000	16:11
Control Test	0.079	16:11
Air Blank	0.000	16:12
Subject Sample # 1	0.000	16:12
Breath Volume	1.335	16:12
Air Blank	0.000	16:12
Air Blank	0.000	16:14
Subject Sample # 2	0.000	16:15
Breath Volume	1.449	16:15
Air Blank	0.000	16:15
Control Test	0.080	16:16
Air Blank	0.000	16:16
Diagnostics Check	OK	16:16

---

Data Download Date/Time: 06/17/2020 23:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAY,L,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FISETTE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: HENRICKS

DATE/TIME OF TEST: 06/15/2020 18:59:49

---

Diagnostics Check	OK	18:51
Air Blank	0.000	18:52
Control Test	0.080	18:52
Air Blank	0.000	18:52
Subject Sample # 1	0.104	18:54
Breath Volume	1.238	18:54
Air Blank	0.000	18:55
Air Blank	0.000	18:56
Subject Sample # 2	0.102	18:58
Breath Volume	1.847	18:58
Air Blank	0.000	18:58
Control Test	0.079	18:59
Air Blank	0.000	18:59
Diagnostics Check	OK	18:59

---

Data Download Date/Time: 06/17/2020 23:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,CHUCK,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HILLIER,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: LUSK

DATE/TIME OF TEST: 06/17/2020 01:14:26

---

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.079	01:09
Air Blank	0.000	01:09
Subject Sample # 1	0.176	01:09
Breath Volume	1.687	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.176	01:12
Breath Volume	1.843	01:12
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

---

Data Download Date/Time: 06/17/2020 23:11



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT,J,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: ROE,B  
ARRESTING OFFICER AGENCY: Neptune Beach PD  
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: PERMENTER

DATE/TIME OF TEST: 06/20/2020 03:32:54

---

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:30	
Control Test	0.079	03:30	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:31	Subject Test Refused
Breath Volume	0.000	03:31	
Air Blank	0.000	03:32	
Control Test	0.080	03:32	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:32	

---

Data Download Date/Time: 07/15/2020 04:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT,J,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: RODRIGUEZ,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 06/20/2020 20:38:41

---

Diagnostics Check	OK	20:35	
Air Blank	0.000	20:36	
Control Test	0.079	20:36	
Air Blank	0.000	20:37	
Subject Sample	# 1	20:37	Subject Test Refused
Breath Volume	0.000	20:37	
Air Blank	0.000	20:37	
Control Test	0.079	20:38	
Air Blank	0.000	20:38	
Diagnostics Check	OK	20:38	

---

Data Download Date/Time: 07/15/2020 04:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT,J,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: SMILEY

DATE/TIME OF TEST: 06/21/2020 00:19:42

---

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.079	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.000	00:15
Breath Volume	2.644	00:15
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.000	00:18
Breath Volume	1.851	00:18
Air Blank	0.000	00:18
Control Test	0.080	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

---

Data Download Date/Time: 07/15/2020 04:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT,J,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/21/2020 03:41:19

---

Diagnostics Check	OK	03:38	
Air Blank	0.000	03:38	
Control Test	0.078	03:39	
Air Blank	0.000	03:39	
Subject Sample	# 1	03:39	Subject Test Refused
Breath Volume	0.000	03:39	
Air Blank	0.000	03:40	
Control Test	0.079	03:40	
Air Blank	0.000	03:41	
Diagnostics Check	OK	03:41	

---

Data Download Date/Time: 07/15/2020 04:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT,J,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DOUGHFMAN,D  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: NUBER

DATE/TIME OF TEST: 06/24/2020 01:02:12

---

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.079	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.154	00:57
Breath Volume		2.632	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.148	01:00
Breath Volume		2.406	01:00
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

---

Data Download Date/Time: 07/15/2020 04:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: TINGLE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: BRUNSWICK

DATE/TIME OF TEST: 06/18/2020 04:45:49

---

Diagnostics Check	OK	04:45	
Air Blank		04:45	Sequence Aborted
Air Blank	0.000	04:45	

---

Data Download Date/Time: 07/15/2020 05:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKNS,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: TINGLE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: BRUNSWICK

DATE/TIME OF TEST: 06/18/2020 05:04:38

---

Diagnostics Check		OK	04:58
Air Blank		0.000	04:58
Control Test		0.078	04:59
Air Blank		0.000	04:59
Subject Sample	# 1	0.194	05:00
Breath Volume		2.250	05:00
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.182	05:03
Breath Volume		2.523	05:03
Air Blank		0.000	05:03
Control Test		0.078	05:04
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

---

Data Download Date/Time: 07/15/2020 05:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUTTON, BRADEN, J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: PARRISH, R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: DAVIS JR

DATE/TIME OF TEST: 06/23/2020 23:54:05

---

Diagnostics Check	OK	23:47
Air Blank	0.000	23:48
Control Test	0.078	23:48
Air Blank	0.000	23:49
Subject Sample # 1	0.000	23:49
Breath Volume	1.492	23:49
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.000	23:52
Breath Volume	1.535	23:52
Air Blank	0.000	23:53
Control Test	0.078	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

---

Data Download Date/Time: 07/15/2020 05:30



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUTTON, BRADEN, J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: RODRIGUEZ, M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: BROPHY

DATE/TIME OF TEST: 06/24/2020 03:16:25

---

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.078	03:13	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:15	
Control Test	0.079	03:15	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

---

Data Download Date/Time: 07/15/2020 05:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SIMMONS,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: RIVERA BELL

DATE/TIME OF TEST: 06/26/2020 01:49:58

---

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.078	01:44
Air Blank		0.000	01:44
Subject Sample	# 1	0.242	01:45
Breath Volume		2.445	01:45
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.235	01:48
Breath Volume		1.382	01:48
Air Blank		0.000	01:49
Control Test		0.078	01:49
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

---

Data Download Date/Time: 07/15/2020 05:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: PASCUAL REYES

DATE/TIME OF TEST: 06/28/2020 02:21:57

---

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.078	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.165	02:17
Breath Volume	2.312	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.170	02:20
Breath Volume	1.906	02:20
Air Blank	0.000	02:21
Control Test	0.078	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

---

Data Download Date/Time: 07/15/2020 05:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MOORE,W  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/28/2020 06:05:24

---

Diagnostics Check	OK	05:59
Air Blank	0.000	05:59
Control Test	0.078	05:59
Air Blank	0.000	06:00
Subject Sample # 1	0.112	06:01
Breath Volume	2.390	06:01
Air Blank	0.000	06:01
Air Blank	0.000	06:03
Subject Sample # 2	0.113	06:03
Breath Volume	3.359	06:03
Air Blank	0.000	06:04
Control Test	0.078	06:04
Air Blank	0.000	06:05
Diagnostics Check	OK	06:05

---

Data Download Date/Time: 07/15/2020 05:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MORGAN, J  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: STIENSTRA

DATE/TIME OF TEST: 06/02/2020 20:41:32

---

Diagnostics Check	OK	20:38	
Air Blank	0.000	20:39	
Control Test	0.080	20:39	
Air Blank	0.000	20:39	
Subject Sample	# 1	20:40	Subject Test Refused
Breath Volume	0.000	20:40	
Air Blank	0.000	20:40	
Control Test	0.080	20:40	
Air Blank	0.000	20:41	
Diagnostics Check	OK	20:41	

---

Data Download Date/Time: 06/17/2020 23:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: TORRES, W  
ARRESTING OFFICER AGENCY: Neptune Beach PD  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/04/2020 00:49:22

---

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:47	
Control Test	0.080	00:47	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:48	
Control Test	0.079	00:48	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

---

Data Download Date/Time: 06/17/2020 23:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT,J,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: TAVAREZ

DATE/TIME OF TEST: 06/04/2020 02:41:13

---

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:38	
Control Test	0.081	02:38	
Air Blank	0.000	02:39	
Subject Sample	# 1	02:39	Subject Test Refused
Breath Volume	0.000	02:39	
Air Blank	0.000	02:40	
Control Test	0.081	02:40	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

---

Data Download Date/Time: 06/17/2020 23:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT,J,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: RODRIGUEZ,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: OQUENDO

DATE/TIME OF TEST: 06/05/2020 02:12:44

---

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.079	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.000	02:08
Breath Volume	3.488	02:08
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.000	02:11
Breath Volume	3.808	02:11
Air Blank	0.000	02:11
Control Test	0.080	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

---

Data Download Date/Time: 06/17/2020 23:03



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DOUGHFMAN, D  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/12/2020 22:50:08

---

Diagnostics Check	OK	22:43
Air Blank	0.000	22:43
Control Test	0.080	22:44
Air Blank	0.000	22:44
Subject Sample # 1	0.126	22:45
Breath Volume	1.300	22:45
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.125	22:48
Breath Volume	2.035	22:48
Air Blank	0.000	22:49
Control Test	0.079	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

---

Data Download Date/Time: 06/17/2020 23:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: RODRIGUEZ, M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: ROYCE

DATE/TIME OF TEST: 06/13/2020 01:17:01

---

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:14	
Control Test	0.080	01:14	
Air Blank	0.000	01:15	
Subject Sample	# 1	01:15	Subject Test Refused
Breath Volume	0.000	01:15	
Air Blank	0.000	01:16	
Control Test	0.080	01:16	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:17	

---

Data Download Date/Time: 06/17/2020 23:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY,SHANON,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: ULYSSE,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: BLUE

DATE/TIME OF TEST: 06/15/2020 01:24:50

---

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:22	
Control Test	0.080	01:22	
Air Blank	0.000	01:23	
Subject Sample	# 1	01:23	Subject Test Refused
Breath Volume	0.000	01:23	
Air Blank	0.000	01:23	
Control Test	0.080	01:24	
Air Blank	0.000	01:24	
Diagnostics Check	OK	01:24	

---

Data Download Date/Time: 06/17/2020 23:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRY, SHANON, E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: RODRIGUEZ, M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: BELGRAVE SHORTER

DATE/TIME OF TEST: 06/15/2020 03:52:20

---

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.080	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.029	03:47
Breath Volume	1.925	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.032	03:50
Breath Volume	2.398	03:50
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

---

Data Download Date/Time: 06/17/2020 23:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MURPHY,JAVON,D  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DAVIDSON,EARL  
ARRESTING OFFICER AGENCY: Gilchrist County SO  
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1048227  
DRY GAS EXPIRY DATE: 11/15/2021  
VIOLATION CODE: ADMIN  
ARREST TIME: 23:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: EPPS

DATE/TIME OF TEST: 06/04/2020 02:00:26

---

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.079	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.108	01:55
Breath Volume	1.433	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:57
Subject Sample # 2	0.107	01:58
Breath Volume	1.261	01:58
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

---

Data Download Date/Time: 07/03/2020 18:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,RICK,D  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BROWN,RICK  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1048227  
DRY GAS EXPIRY DATE: 11/15/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: CZARNIAK

DATE/TIME OF TEST: 06/09/2020 19:41:20

---

Diagnostics Check	OK	19:34
Air Blank	0.000	19:35
Control Test	0.079	19:35
Air Blank	0.000	19:36
Subject Sample # 1	0.170	19:36
Breath Volume	2.179	19:36
Air Blank	0.000	19:37
Air Blank	0.000	19:39
Subject Sample # 2	0.176	19:39
Breath Volume	1.882	19:39
Air Blank	0.000	19:40
Control Test	0.078	19:40
Air Blank	0.000	19:41
Diagnostics Check	OK	19:41

---

Data Download Date/Time: 07/03/2020 18:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUKE,JUSTIN,B  
BREATH TEST OPERATOR AGENCY: FL Dept of AG  
ARRESTING OFFICER: DUKE,JUSTIN  
ARRESTING OFFICER AGENCY: FL Dept of AG  
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1048227  
DRY GAS EXPIRY DATE: 11/15/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/10/2020 02:53:37

---

Diagnostics Check	OK	02:46	
Air Blank	0.000	02:46	
Control Test	0.080	02:47	
Air Blank	0.000	02:47	
Subject Sample	# 1 0.148	02:48	
Breath Volume	1.496	02:48	
Air Blank	0.000	02:48	
Air Blank	0.000	02:50	
Subject Sample	# 2	02:51	Slope Not Met
Breath Volume	1.437	02:51	
Air Blank	0.000	02:52	
Control Test	0.081	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

---

Data Download Date/Time: 07/03/2020 18:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

INSTRUMENT REGISTERED TO: UNION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOFMANN, TRAVIS, A  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: HOFMANN, TRAVIS  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: SEAY

DATE/TIME OF TEST: 06/07/2020 05:55:33

---

Diagnostics Check	OK	05:52	
Air Blank	0.000	05:53	
Control Test	0.080	05:53	
Air Blank	0.000	05:54	
Subject Sample	# 1	05:54	Subject Test Refused
Breath Volume	0.000	05:54	
Air Blank	0.000	05:54	
Control Test	0.080	05:54	
Air Blank	0.000	05:55	
Diagnostics Check	OK	05:55	

---

Data Download Date/Time: 06/29/2020 10:42



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUMNER,DALTON,R  
BREATH TEST OPERATOR AGENCY: Bradford County SO  
ARRESTING OFFICER: ALSTON,BRANDON  
ARRESTING OFFICER AGENCY: Bradford County SO  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG000606  
DRY GAS EXPIRY DATE: 01/06/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: MILTON

DATE/TIME OF TEST: 06/11/2020 19:00:40

---

Diagnostics Check	OK	18:53
Air Blank	0.000	18:54
Control Test	0.078	18:54
Air Blank	0.000	18:54
Subject Sample # 1	0.195	18:56
Breath Volume	2.250	18:56
Air Blank	0.000	18:56
Air Blank	0.000	18:58
Subject Sample # 2	0.186	18:59
Breath Volume	2.394	18:59
Air Blank	0.000	18:59
Control Test	0.075	19:00
Air Blank	0.000	19:00
Diagnostics Check	OK	19:00

---

Data Download Date/Time: 08/01/2020 09:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUMNER,DALTON,R  
BREATH TEST OPERATOR AGENCY: Bradford County SO  
ARRESTING OFFICER: FLYNN,HALEY  
ARRESTING OFFICER AGENCY: Bradford County SO  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG000606  
DRY GAS EXPIRY DATE: 01/06/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: FULGHAM

DATE/TIME OF TEST: 06/25/2020 23:05:05

---

Diagnostics Check	OK	22:56	
Air Blank	0.000	22:56	
Control Test	0.079	22:57	
Air Blank	0.000	22:57	
Subject Sample	# 1	23:00	No Sample Provided
Breath Volume	0.000	23:00	
Air Blank	0.000	23:01	
Air Blank	0.000	23:03	
Subject Sample	# 2	23:03	Subject Test Refused
Breath Volume	0.000	23:03	
Air Blank	0.000	23:04	
Control Test	0.079	23:04	
Air Blank	0.000	23:05	
Diagnostics Check	OK	23:05	

---

Data Download Date/Time: 08/01/2020 09:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TOMPKINS,DANIEL,H  
BREATH TEST OPERATOR AGENCY: Lawtey PD  
ARRESTING OFFICER: TOMPKINS,DANIEL  
ARRESTING OFFICER AGENCY: Lawtey PD  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG000606  
DRY GAS EXPIRY DATE: 01/06/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: MERRIWHEATHER

DATE/TIME OF TEST: 06/27/2020 23:43:17

---

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:32	
Control Test	0.079	23:32	
Air Blank	0.000	23:33	
Subject Sample	# 1	23:36	No Sample Provided
Breath Volume	0.000	23:36	
Air Blank	0.000	23:36	
Air Blank	0.000	23:38	
Subject Sample	# 2	23:41	No Sample Provided
Breath Volume	0.000	23:41	
Air Blank	0.000	23:42	
Control Test	0.079	23:42	
Air Blank	0.000	23:43	
Diagnostics Check	OK	23:43	

---

Data Download Date/Time: 08/01/2020 09:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VISMAN,CLAYTON,A  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: VISMAN,CLAYTON  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG000606  
DRY GAS EXPIRY DATE: 01/06/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: NORMAN

DATE/TIME OF TEST: 06/30/2020 00:12:50

---

Diagnostics Check	OK	00:09	
Air Blank	0.000	00:09	
Control Test	0.079	00:10	
Air Blank	0.000	00:10	
Subject Sample	# 1	00:11	Subject Test Refused
Breath Volume	0.000	00:11	
Air Blank	0.000	00:11	
Control Test	0.079	00:12	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:12	

---

Data Download Date/Time: 08/01/2020 09:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VISMAN,CLAYTON,A  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: VISMAN,CLAYTON  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG000606  
DRY GAS EXPIRY DATE: 01/06/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: NORMAN

DATE/TIME OF TEST: 06/30/2020 00:37:29

---

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:35	
Control Test	0.078	00:35	
Air Blank	0.000	00:35	
Subject Sample	# 1	00:36	Subject Test Refused
Breath Volume	0.000	00:36	
Air Blank	0.000	00:36	
Control Test	0.079	00:36	
Air Blank	0.000	00:37	
Diagnostics Check	OK	00:37	

---

Data Download Date/Time: 08/01/2020 09:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VISMAN,CLAYTON,A  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: VISMAN,CLAYTON  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG000606  
DRY GAS EXPIRY DATE: 01/06/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: NORMAN

DATE/TIME OF TEST: 06/30/2020 01:02:19

---

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:55	
Control Test	0.078	00:55	
Air Blank	0.000	00:56	
Subject Sample # 1	0.146	00:57	
Breath Volume	1.875	00:57	
Air Blank	0.000	00:58	
Air Blank	0.000	01:00	
Subject Sample # 2	0.147	01:00	
Breath Volume	1.574	01:00	
Air Blank	0.000	01:01	
Control Test	0.078	01:01	
Air Blank		01:01	RFI Detect
Air Blank	0.000	01:02	

---

Data Download Date/Time: 08/01/2020 09:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VISMAN,CLAYTON,A  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: VISMAN,CLAYTON  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG000606  
DRY GAS EXPIRY DATE: 01/06/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: NORMAN

DATE/TIME OF TEST: 06/30/2020 01:32:53

---

Diagnostics Check		OK	01:26
Air Blank		0.000	01:27
Control Test		0.079	01:27
Air Blank		0.000	01:27
Subject Sample	# 1	0.137	01:28
Breath Volume		1.199	01:28
Air Blank		0.000	01:29
Air Blank		0.000	01:30
Subject Sample	# 2	0.133	01:31
Breath Volume		1.765	01:31
Air Blank		0.000	01:31
Control Test		0.078	01:32
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

---

Data Download Date/Time: 08/01/2020 09:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANDIS,BRYAN,S  
BREATH TEST OPERATOR AGENCY: Williston PD  
ARRESTING OFFICER: LANDIS,BRYAN  
ARRESTING OFFICER AGENCY: Williston PD  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 937240  
DRY GAS EXPIRY DATE: 04/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: TELL0-VELASQUEZ

DATE/TIME OF TEST: 06/06/2020 23:49:18

---

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.079	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.248	23:44
Breath Volume	1.738	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.243	23:47
Breath Volume	2.359	23:47
Air Blank	0.000	23:48
Control Test	0.076	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

---

Data Download Date/Time: 07/08/2020 19:44



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASOLO,DOREEN,M  
BREATH TEST OPERATOR AGENCY: Williston PD  
ARRESTING OFFICER: LONG,KATHY  
ARRESTING OFFICER AGENCY: Williston PD  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 937240  
DRY GAS EXPIRY DATE: 04/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: BARBER

DATE/TIME OF TEST: 06/13/2020 17:57:15

---

Diagnostics Check	OK	17:50
Air Blank	0.000	17:50
Control Test	0.080	17:50
Air Blank	0.000	17:51
Subject Sample # 1	0.000	17:52
Breath Volume	2.621	17:52
Air Blank	0.000	17:53
Air Blank	0.000	17:55
Subject Sample # 2	0.000	17:55
Breath Volume	3.914	17:55
Air Blank	0.000	17:56
Control Test	0.080	17:56
Air Blank	0.000	17:57
Diagnostics Check	OK	17:57

---

Data Download Date/Time: 07/08/2020 19:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: APPLING,JEFF,M  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: APPLING,JEFF  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: SURIANO

DATE/TIME OF TEST: 06/01/2020 20:41:21

---

Diagnostics Check	OK	20:29	
Air Blank	0.000	20:30	
Control Test	0.079	20:30	
Air Blank	0.000	20:30	
Subject Sample	# 1 0.187	20:34	Volume Not Met
Breath Volume	1.011	20:34	
Air Blank	0.000	20:34	
Air Blank	0.000	20:36	
Subject Sample	# 2 0.172	20:39	Volume Not Met
Breath Volume	0.292	20:39	
Air Blank	0.000	20:40	
Control Test	0.076	20:40	
Air Blank	0.000	20:41	
Diagnostics Check	OK	20:41	

---

Data Download Date/Time: 06/29/2020 18:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHAPMAN,MAVERICK,E  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: HALFHILL,TYLER  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 06/01/2020 22:11:57

---

Diagnostics Check	OK	22:06
Air Blank	0.000	22:06
Control Test	0.080	22:06
Air Blank	0.000	22:07
Subject Sample # 1	0.000	22:07
Breath Volume	2.554	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.000	22:10
Breath Volume	1.441	22:10
Air Blank	0.000	22:11
Control Test	0.078	22:11
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

---

Data Download Date/Time: 06/29/2020 18:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOGETTE,MASON,J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: ELENZ,TIMOTHY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:13

SUBJECT'S LAST NAME: BRANDIS

DATE/TIME OF TEST: 06/04/2020 18:37:47

---

Diagnostics Check	OK	18:33	
Air Blank	0.000	18:34	
Control Test	0.079	18:34	
Air Blank	0.000	18:35	
Subject Sample	# 1	18:36	Subject Test Refused
Breath Volume	0.000	18:36	
Air Blank	0.000	18:36	
Control Test	0.079	18:37	
Air Blank	0.000	18:37	
Diagnostics Check	OK	18:37	

---

Data Download Date/Time: 06/29/2020 18:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOYD,J,J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: STEINER,Z  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 06/05/2020 19:07:34

---

Diagnostics Check	OK	19:00
Air Blank	0.000	19:01
Control Test	0.080	19:01
Air Blank	0.000	19:02
Subject Sample # 1	0.000	19:02
Breath Volume	1.921	19:02
Air Blank	0.000	19:03
Air Blank	0.000	19:05
Subject Sample # 2	0.000	19:06
Breath Volume	1.589	19:06
Air Blank	0.000	19:06
Control Test	0.079	19:06
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

---

Data Download Date/Time: 06/29/2020 18:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHAPMAN,MAVERICK,E  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CHAPMAN,MAVERICK  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 06/05/2020 22:20:09

---

Diagnostics Check	OK	22:17	
Air Blank	0.000	22:17	
Control Test	0.080	22:17	
Air Blank	0.000	22:18	
Subject Sample	# 1	22:18	Subject Test Refused
Breath Volume	0.000	22:18	
Air Blank	0.000	22:19	
Control Test	0.080	22:19	
Air Blank	0.000	22:20	
Diagnostics Check	OK	22:20	

---

Data Download Date/Time: 06/29/2020 18:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPENCER, RAYMOND, L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: SPENCER, RAYMOND  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: GOLLEY

DATE/TIME OF TEST: 06/07/2020 04:52:07

---

Diagnostics Check	OK	04:43
Air Blank	0.000	04:44
Control Test	0.079	04:44
Air Blank	0.000	04:44
Subject Sample # 1	0.139	04:46
Breath Volume	1.648	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.139	04:50
Breath Volume	3.027	04:50
Air Blank	0.000	04:51
Control Test	0.079	04:51
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

---

Data Download Date/Time: 06/29/2020 18:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORENE,AIDAN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: NORENE,AIDAN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: FAIRCLOTH

DATE/TIME OF TEST: 06/11/2020 02:13:58

---

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.179	02:09
Breath Volume	2.406	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.170	02:12
Breath Volume	2.515	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

---

Data Download Date/Time: 06/29/2020 18:50



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,PAUL,M  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: RODRIGUEZ,PAUL  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 06/13/2020 02:12:00

---

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.187	02:07
Breath Volume	2.906	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.190	02:10
Breath Volume	3.437	02:10
Air Blank	0.000	02:11
Control Test	0.079	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:12

---

Data Download Date/Time: 06/29/2020 18:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPENCER, RAYMOND, L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: SPENCER, RAYMOND  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: ERVIN

DATE/TIME OF TEST: 06/13/2020 02:48:00

---

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.133	02:43
Breath Volume	2.949	02:43
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.134	02:46
Breath Volume	2.328	02:46
Air Blank	0.000	02:47
Control Test	0.080	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:48

---

Data Download Date/Time: 06/29/2020 18:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVERA,DANTE,E  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: MILLS,SIERRA  
ARRESTING OFFICER AGENCY: Milton PD  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: COSTILLA-ESPINOZA

DATE/TIME OF TEST: 06/13/2020 04:03:51

---

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.080	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.273	03:59
Breath Volume	3.070	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:01
Subject Sample # 2	0.273	04:02
Breath Volume	1.425	04:02
Air Blank	0.000	04:02
Control Test	0.078	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

---

Data Download Date/Time: 06/29/2020 18:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MERRILL,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: MERRILL,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 06/14/2020 02:39:00

---

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.129	02:34
Breath Volume	2.800	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.125	02:37
Breath Volume	2.722	02:37
Air Blank	0.000	02:38
Control Test	0.079	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:39

---

Data Download Date/Time: 06/29/2020 18:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOORE,JAMES,I  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: MOORE,JAMES  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: WILEY

DATE/TIME OF TEST: 06/15/2020 03:22:04

---

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.079	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.204	03:17
Breath Volume		1.113	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.186	03:20
Breath Volume		1.210	03:20
Air Blank		0.000	03:21
Control Test		0.078	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:22

---

Data Download Date/Time: 06/29/2020 18:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACKSON, DAKOTA, S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: JACKSON, DAKOTA  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: ERICKSEN

DATE/TIME OF TEST: 06/18/2020 22:21:17

---

Diagnostics Check	OK	22:18	
Air Blank	0.000	22:18	
Control Test	0.079	22:18	
Air Blank	0.000	22:19	
Subject Sample # 1	0.197	22:20	
Breath Volume	1.394	22:20	
Air Blank		22:20	Ambient Fail
Air Blank	0.000	22:21	

---

Data Download Date/Time: 06/29/2020 18:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACKSON, DAKOTA, S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: JACKSON, DAKOTA  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: ERICKSEN

DATE/TIME OF TEST: 06/19/2020 00:32:12

---

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.078	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.150	00:27
Breath Volume	1.316	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.151	00:30
Breath Volume	1.480	00:30
Air Blank	0.000	00:31
Control Test	0.078	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

---

Data Download Date/Time: 06/29/2020 18:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPENCER, RAYMOND, L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CRAWFORD, BILL  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 06/19/2020 00:55:27

---

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.079	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.112	00:50
Breath Volume	1.363	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.117	00:53
Breath Volume	1.972	00:53
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

---

Data Download Date/Time: 06/29/2020 18:50



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUSNAK, BRANDEN, M  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: RUSNAK, BRANDEN  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 13:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:46

SUBJECT'S LAST NAME: JOLLY

DATE/TIME OF TEST: 06/19/2020 14:17:32

---

Diagnostics Check	OK	14:06	
Air Blank	0.000	14:06	
Control Test	0.078	14:06	
Air Blank	0.000	14:07	
Subject Sample	# 1	14:10	No Sample Provided
Breath Volume	0.000	14:10	
Air Blank	0.000	14:11	
Air Blank	0.000	14:13	
Subject Sample	# 2	14:16	No Sample Provided
Breath Volume	0.000	14:16	
Air Blank	0.000	14:16	
Control Test	0.079	14:16	
Air Blank	0.000	14:17	
Diagnostics Check	OK	14:17	

---

Data Download Date/Time: 06/29/2020 18:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYNN,BAMBI,J  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: LYNN,BAMBI  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 13:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:14

SUBJECT'S LAST NAME: CHENEY

DATE/TIME OF TEST: 06/19/2020 14:51:00

---

Diagnostics Check	OK	14:38	
Air Blank	0.000	14:39	
Control Test	0.078	14:39	
Air Blank	0.000	14:39	
Subject Sample	# 1	14:43	No Sample Provided
Breath Volume	0.000	14:43	
Air Blank	0.000	14:43	
Air Blank	0.000	14:45	
Subject Sample	# 2 0.374	14:45	
Breath Volume	1.851	14:45	
Air Blank	0.000	14:46	
Air Blank	0.000	14:48	
Subject Sample	# 3 0.357	14:49	
Breath Volume	1.824	14:49	
Air Blank	0.000	14:50	
Control Test	0.078	14:50	
Air Blank	0.000	14:50	
Diagnostics Check	OK	14:51	

---

Data Download Date/Time: 06/29/2020 18:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOYE,JAMES,M  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: MOYE,JAMES  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: STARR

DATE/TIME OF TEST: 06/21/2020 03:09:22

---

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:01	
Control Test	0.080	03:01	
Air Blank	0.000	03:01	
Subject Sample	# 1	03:05	No Sample Provided
Breath Volume	0.000	03:05	
Air Blank	0.000	03:05	
Air Blank	0.000	03:07	
Subject Sample	# 2	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:08	
Control Test	0.079	03:08	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

---

Data Download Date/Time: 06/29/2020 18:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARTZOG,ROBERT,  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: HARTZOG,ROBERT  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 06/21/2020 22:19:36

---

Diagnostics Check	OK	22:16	
Air Blank	0.000	22:16	
Control Test	0.080	22:16	
Air Blank	0.000	22:17	
Subject Sample	# 1	22:18	Subject Test Refused
Breath Volume	0.000	22:18	
Air Blank	0.000	22:18	
Control Test	0.079	22:19	
Air Blank	0.000	22:19	
Diagnostics Check	OK	22:19	

---

Data Download Date/Time: 06/29/2020 18:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURKE,MATTHEW,S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: BURKE,MATTHEW  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: TYNER

DATE/TIME OF TEST: 06/22/2020 03:53:04

---

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:43	
Control Test	0.079	03:43	
Air Blank	0.000	03:44	
Subject Sample	# 1 0.100	03:47	Volume Not Met
Breath Volume	0.148	03:47	
Air Blank	0.000	03:48	
Air Blank	0.000	03:50	
Subject Sample	# 2	03:52	Sequence Aborted
Breath Volume	0.109	03:52	
Air Blank	0.000	03:53	

---

Data Download Date/Time: 06/29/2020 18:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAHR,DANIEL,E  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: HAHR,DANIEL  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: BUI  
ARREST TIME: 19:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/27/2020 20:55:06

---

Diagnostics Check	OK	20:48
Air Blank	0.000	20:48
Control Test	0.080	20:49
Air Blank	0.000	20:49
Subject Sample # 1	0.156	20:50
Breath Volume	3.875	20:50
Air Blank	0.000	20:51
Air Blank	0.000	20:52
Subject Sample # 2	0.157	20:53
Breath Volume	4.222	20:53
Air Blank	0.000	20:54
Control Test	0.078	20:54
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

---

Data Download Date/Time: 06/29/2020 18:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARLSON,JOSHUA,L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: ADAMS,BRIANA  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: STALL

DATE/TIME OF TEST: 06/28/2020 20:11:17

---

Diagnostics Check	OK	20:07	
Air Blank	0.000	20:07	
Control Test	0.080	20:08	
Air Blank	0.000	20:08	
Subject Sample	# 1	20:09	Subject Test Refused
Breath Volume	0.000	20:09	
Air Blank	0.000	20:10	
Control Test	0.080	20:10	
Air Blank	0.000	20:11	
Diagnostics Check	OK	20:11	

---

Data Download Date/Time: 06/29/2020 18:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BANKS,SEAN,C  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: BANKS,SEAN  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 06/02/2020 20:20:52

---

Diagnostics Check	OK	20:14
Air Blank	0.000	20:15
Control Test	0.078	20:15
Air Blank	0.000	20:16
Subject Sample # 1	0.161	20:16
Breath Volume	3.902	20:16
Air Blank	0.000	20:17
Air Blank	0.000	20:18
Subject Sample # 2	0.159	20:19
Breath Volume	2.218	20:19
Air Blank	0.000	20:19
Control Test	0.078	20:20
Air Blank	0.000	20:20
Diagnostics Check	OK	20:20

---

Data Download Date/Time: 06/04/2020 12:28



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BANKS,SEAN,C  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: BANKS,SEAN  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: HAGERMAN

DATE/TIME OF TEST: 06/17/2020 04:07:10

---

Diagnostics Check		OK	04:00
Air Blank		0.000	04:01
Control Test		0.079	04:01
Air Blank		0.000	04:02
Subject Sample	# 1	0.178	04:02
Breath Volume		2.011	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.179	04:05
Breath Volume		3.519	04:05
Air Blank		0.000	04:06
Control Test		0.079	04:06
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

---

Data Download Date/Time: 07/01/2020 14:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BANKS,SEAN,C  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: BANKS,SEAN  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 06/21/2020 01:50:01

---

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.079	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.166	01:45
Breath Volume	3.750	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:47
Subject Sample # 2	0.166	01:48
Breath Volume	4.222	01:48
Air Blank	0.000	01:49
Control Test	0.078	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:50

---

Data Download Date/Time: 07/01/2020 14:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BANKS,SEAN,C  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: BANKS,SEAN  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: PETROSSI

DATE/TIME OF TEST: 06/29/2020 22:10:48

---

Diagnostics Check	OK	22:04
Air Blank	0.000	22:04
Control Test	0.078	22:05
Air Blank	0.000	22:05
Subject Sample # 1	0.090	22:06
Breath Volume	3.156	22:06
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.089	22:09
Breath Volume	2.906	22:09
Air Blank	0.000	22:09
Control Test	0.079	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

---

Data Download Date/Time: 07/01/2020 14:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRELL,CHAD,A  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: DELANEY, TYLER  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG923908  
DRY GAS EXPIRY DATE: 08/27/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 06/08/2020 22:25:25

---

Diagnostics Check	OK	22:19
Air Blank	0.000	22:19
Control Test	0.081	22:19
Air Blank	0.000	22:20
Subject Sample # 1	0.249	22:20
Breath Volume	2.671	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.246	22:23
Breath Volume	1.160	22:23
Air Blank	0.000	22:24
Control Test	0.080	22:24
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

---

Data Download Date/Time: 06/26/2020 16:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ALEX,R  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: YOUNG,ALEX  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG923908  
DRY GAS EXPIRY DATE: 08/27/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: POGUE

DATE/TIME OF TEST: 06/09/2020 03:16:36

---

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:14	
Control Test	0.080	03:14	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:15	
Control Test	0.080	03:16	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

---

Data Download Date/Time: 06/26/2020 16:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ALEX,R  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: YOUNG,ALEX  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG923908  
DRY GAS EXPIRY DATE: 08/27/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: EL HANNAOUI

DATE/TIME OF TEST: 06/10/2020 01:05:02

---

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:56	
Control Test	0.081	00:56	
Air Blank	0.000	00:56	
Subject Sample	# 1	01:00	No Sample Provided
Breath Volume	0.000	01:00	
Air Blank	0.000	01:00	
Air Blank	0.000	01:02	
Subject Sample	# 2	01:03	Subject Test Refused
Breath Volume	0.000	01:03	
Air Blank	0.000	01:04	
Control Test	0.081	01:04	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:05	

---

Data Download Date/Time: 06/26/2020 16:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ALEX,R  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: YOUNG,ALEX  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG923908  
DRY GAS EXPIRY DATE: 08/27/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: RAGOONANAN

DATE/TIME OF TEST: 06/14/2020 02:15:24

---

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.081	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.108	02:11
Breath Volume	3.800	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.101	02:13
Breath Volume	3.976	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

---

Data Download Date/Time: 06/26/2020 16:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARNER,JOSHUA,C  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: MORRISON,ALLISON  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG923908  
DRY GAS EXPIRY DATE: 08/27/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: PROFFITT

DATE/TIME OF TEST: 06/22/2020 03:10:19

---

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.081	03:04
Air Blank	0.000	03:04
Subject Sample # 1	0.195	03:05
Breath Volume	3.652	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.187	03:08
Breath Volume	2.898	03:08
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

---

Data Download Date/Time: 06/26/2020 16:20



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ALEX,R  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: YOUNG,ALEX  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG923908  
DRY GAS EXPIRY DATE: 08/27/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: RIDER

DATE/TIME OF TEST: 06/23/2020 03:07:02

---

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.081	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.198	03:02
Breath Volume		2.468	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.205	03:05
Breath Volume		2.363	03:05
Air Blank		0.000	03:06
Control Test		0.081	03:06
Air Blank		0.000	03:06
Diagnostics Check		OK	03:07

---

Data Download Date/Time: 06/26/2020 16:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,CARLON,E  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: DELANEY,T  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG923908  
DRY GAS EXPIRY DATE: 08/27/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: WOODWARD

DATE/TIME OF TEST: 06/23/2020 23:19:05

---

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.080	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.151	23:14
Breath Volume	3.234	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.140	23:17
Breath Volume	2.394	23:17
Air Blank	0.000	23:18
Control Test	0.079	23:18
Air Blank	0.000	23:18
Diagnostics Check	OK	23:19

---

Data Download Date/Time: 06/26/2020 16:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POVEDA,CARLOS,M  
BREATH TEST OPERATOR AGENCY: Niceville PD  
ARRESTING OFFICER: CARRICO,KELLY  
ARRESTING OFFICER AGENCY: Niceville PD  
LAST AGENCY INSPECTION: 06/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 869968  
DRY GAS EXPIRY DATE: 12/06/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:42

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/15/2020 07:29:00

---

Diagnostics Check		OK	07:22
Air Blank		0.000	07:23
Control Test		0.077	07:23
Air Blank		0.000	07:24
Subject Sample	# 1	0.284	07:24
Breath Volume		2.660	07:24
Air Blank		0.000	07:25
Air Blank		0.000	07:26
Subject Sample	# 2	0.285	07:27
Breath Volume		2.781	07:27
Air Blank		0.000	07:28
Control Test		0.075	07:28
Air Blank		0.000	07:28
Diagnostics Check		OK	07:29

---

Data Download Date/Time: 07/20/2020 11:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OTILUS,LEBEDOPH,  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: OTILUS,LEBEDOPH  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 05/18/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CHIRUM

DATE/TIME OF TEST: 06/03/2020 23:41:48

---

Diagnostics Check	OK	23:32	
Air Blank	0.000	23:32	
Control Test	0.080	23:32	
Air Blank	0.000	23:33	
Subject Sample	# 1	23:36	No Sample Provided
Breath Volume	0.000	23:36	
Air Blank	0.000	23:37	
Air Blank	0.000	23:39	
Subject Sample	# 2	23:40	Subject Test Refused
Breath Volume	1.355	23:40	
Air Blank	0.000	23:40	
Control Test	0.080	23:41	
Air Blank	0.000	23:41	
Diagnostics Check	OK	23:41	

---

Data Download Date/Time: 06/28/2020 09:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAREY,ANDREW,S  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: PAN,JACKY  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 05/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: RITTER

DATE/TIME OF TEST: 06/04/2020 04:17:12

---

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.079	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 1	04:10	No Sample Provided
Breath Volume	0.000	04:10	
Air Blank	0.000	04:10	
Air Blank	0.000	04:12	
Subject Sample	# 2	04:15	No Sample Provided
Breath Volume	0.000	04:15	
Air Blank	0.000	04:16	
Control Test	0.080	04:16	
Air Blank	0.000	04:17	
Diagnostics Check	OK	04:17	

---

Data Download Date/Time: 06/28/2020 09:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NICHOLSON, HORACE, O  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: SATALLANTE, DEVON  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1076482  
DRY GAS EXPIRY DATE: 01/15/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 06/15/2020 21:12:04

---

Diagnostics Check	OK	21:04
Air Blank	0.000	21:05
Control Test	0.080	21:05
Air Blank	0.000	21:06
Subject Sample # 1	0.000	21:07
Breath Volume	1.437	21:07
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.000	21:10
Breath Volume	1.796	21:10
Air Blank	0.000	21:11
Control Test	0.079	21:11
Air Blank	0.000	21:11
Diagnostics Check	OK	21:12

---

Data Download Date/Time: 07/23/2020 12:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WRIGHT,MARK,A  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: VARNADORE,R  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1076482  
DRY GAS EXPIRY DATE: 01/15/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:13

SUBJECT'S LAST NAME: LANDRESS

DATE/TIME OF TEST: 06/17/2020 06:36:27

---

Diagnostics Check	OK	06:33	
Air Blank	0.000	06:33	
Control Test	0.079	06:34	
Air Blank	0.000	06:34	
Subject Sample	# 1	06:35	Subject Test Refused
Breath Volume	0.000	06:35	
Air Blank	0.000	06:35	
Control Test	0.079	06:35	
Air Blank	0.000	06:36	
Diagnostics Check	OK	06:36	

---

Data Download Date/Time: 07/23/2020 12:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROGERS,DAVID,E  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: GREIS,BRANDON  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1076482  
DRY GAS EXPIRY DATE: 01/15/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: GILLIS

DATE/TIME OF TEST: 06/24/2020 22:39:16

---

Diagnostics Check	OK	22:32
Air Blank	0.000	22:33
Control Test	0.079	22:33
Air Blank	0.000	22:34
Subject Sample # 1	0.000	22:34
Breath Volume	2.394	22:34
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.000	22:37
Breath Volume	3.597	22:37
Air Blank	0.000	22:38
Control Test	0.079	22:38
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

---

Data Download Date/Time: 07/23/2020 12:40



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NICHOLSON, HORACE, O  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: TULLIO, CHRISTOPHER  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1076482  
DRY GAS EXPIRY DATE: 01/15/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PONCE

DATE/TIME OF TEST: 06/26/2020 02:36:53

---

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.130	02:32
Breath Volume	2.304	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.132	02:35
Breath Volume	2.625	02:35
Air Blank	0.000	02:35
Control Test	0.080	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

---

Data Download Date/Time: 07/23/2020 12:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAJERE,ARIC,  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: CAULEY,MICHAEL  
ARRESTING OFFICER AGENCY: Okeechobee County SO  
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1076482  
DRY GAS EXPIRY DATE: 01/15/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 06/27/2020 18:29:19

---

Diagnostics Check	OK	18:22
Air Blank	0.000	18:23
Control Test	0.079	18:23
Air Blank	0.000	18:24
Subject Sample # 1	0.212	18:24
Breath Volume	1.957	18:24
Air Blank	0.000	18:25
Air Blank	0.000	18:27
Subject Sample # 2	0.208	18:27
Breath Volume	1.847	18:27
Air Blank	0.000	18:28
Control Test	0.079	18:28
Air Blank	0.000	18:29
Diagnostics Check	OK	18:29

---

Data Download Date/Time: 07/23/2020 12:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NICHOLSON, HORACE, O  
BREATH TEST OPERATOR AGENCY: Okeechobee County SO  
ARRESTING OFFICER: CROSBY, JASON  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1076482  
DRY GAS EXPIRY DATE: 01/15/2022  
VIOLATION CODE: BUI  
ARREST TIME: 20:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/07/2020 22:13:57

---

Diagnostics Check	OK	22:11	
Air Blank	0.000	22:11	
Control Test	0.080	22:11	
Air Blank	0.000	22:12	
Subject Sample	# 1	22:12	Subject Test Refused
Breath Volume	0.000	22:12	
Air Blank	0.000	22:13	
Control Test	0.080	22:13	
Air Blank	0.000	22:13	
Diagnostics Check	OK	22:13	

---

Data Download Date/Time: 06/09/2020 13:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICCARDI,A,  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: OCHOA,N  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 06/05/2020 23:38:46

---

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:25	
Control Test	0.080	23:26	
Air Blank	0.000	23:26	
Subject Sample	# 1	23:29	No Sample Provided
Breath Volume	0.000	23:29	
Air Blank	0.000	23:30	
Air Blank	0.000	23:32	
Subject Sample	# 2	0.257	23:34
Breath Volume	1.234	23:34	
Air Blank	0.000	23:35	
Air Blank	0.000	23:36	
Subject Sample	# 3	23:37	Slope Not Met
Breath Volume	1.328	23:37	
Air Blank	0.000	23:37	
Control Test	0.077	23:38	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

---

Data Download Date/Time: 06/30/2020 13:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICCARDI,A,  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: OCHOA,N  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 06/05/2020 23:50:50

---

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.078	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.252	23:45
Breath Volume	1.417	23:45
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.255	23:49
Breath Volume	1.503	23:49
Air Blank	0.000	23:49
Control Test	0.077	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

---

Data Download Date/Time: 06/30/2020 13:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,MICHAEL,C  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: WILLIAMS,MICHAEL  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 07:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: CALMO CASTILLO

DATE/TIME OF TEST: 06/07/2020 08:26:22

---

Diagnostics Check	OK	08:15	
Air Blank	0.000	08:15	
Control Test	0.080	08:15	
Air Blank	0.000	08:16	
Subject Sample	# 1	08:19	No Sample Provided
Breath Volume	0.000	08:19	
Air Blank	0.000	08:19	
Air Blank	0.000	08:21	
Subject Sample	# 2	08:25	No Sample Provided
Breath Volume	0.000	08:25	
Air Blank	0.000	08:25	
Control Test	0.080	08:25	
Air Blank	0.000	08:26	
Diagnostics Check	OK	08:26	

---

Data Download Date/Time: 06/30/2020 13:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,MICHAEL,C  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: THOMAS,BRYCE  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:38

SUBJECT'S LAST NAME: MYRTER

DATE/TIME OF TEST: 06/10/2020 16:06:53

---

Diagnostics Check	OK	15:59
Air Blank	0.000	16:00
Control Test	0.080	16:00
Air Blank	0.000	16:01
Subject Sample # 1	0.186	16:01
Breath Volume	1.351	16:01
Air Blank	0.000	16:02
Air Blank	0.000	16:04
Subject Sample # 2	0.184	16:05
Breath Volume	1.394	16:05
Air Blank	0.000	16:05
Control Test	0.078	16:06
Air Blank	0.000	16:06
Diagnostics Check	OK	16:06

---

Data Download Date/Time: 06/30/2020 13:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OJEDA,NELSON,J  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: GASPAR,BENJAMIN  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: AIUDI

DATE/TIME OF TEST: 06/14/2020 03:50:19

---

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:42	
Control Test	0.079	03:42	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:46	No Sample Provided
Breath Volume	0.000	03:46	
Air Blank	0.000	03:46	
Air Blank	0.000	03:48	
Subject Sample	# 2	03:48	Subject Test Refused
Breath Volume	0.000	03:48	
Air Blank	0.000	03:49	
Control Test	0.080	03:49	
Air Blank	0.000	03:50	
Diagnostics Check	OK	03:50	

---

Data Download Date/Time: 06/30/2020 13:06



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,MICHAEL,C  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: WILLIAMS,MICHAEL  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 10:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: ISHERWOOD

DATE/TIME OF TEST: 06/14/2020 11:14:19

---

Diagnostics Check	OK	11:11	
Air Blank	0.000	11:11	
Control Test	0.080	11:12	
Air Blank	0.000	11:12	
Subject Sample	# 1	11:12	RFI Detect
Breath Volume	0.000	11:12	
Air Blank	0.000	11:13	
Control Test	0.079	11:13	
Air Blank	0.000	11:14	
Diagnostics Check	OK	11:14	

---

Data Download Date/Time: 06/30/2020 13:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,MICHAEL,C  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: WILLIAMS,MICHAEL  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 10:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: ISHERWOOD

DATE/TIME OF TEST: 06/14/2020 11:16:22

---

Diagnostics Check	OK	11:16	
Air Blank		11:16	RFI Detect
Air Blank		11:16	RFI Detect

---

Data Download Date/Time: 06/30/2020 13:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,MICHAEL,C  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: WILLIAMS,MICHAEL  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 10:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: ISHERWOOD

DATE/TIME OF TEST: 06/14/2020 11:40:01

---

Diagnostics Check	OK	11:29
Air Blank	0.000	11:29
Control Test	0.079	11:30
Air Blank	0.000	11:30
Subject Sample # 1	0.229	11:31
Breath Volume	1.875	11:31
Air Blank	0.000	11:32
Air Blank	0.000	11:34
Subject Sample # 2	0.205	11:34
Breath Volume	1.859	11:34
Air Blank	0.000	11:35
Air Blank	0.000	11:37
Subject Sample # 3	0.208	11:38
Breath Volume	2.277	11:38
Air Blank	0.000	11:39
Control Test	0.077	11:39
Air Blank	0.000	11:39
Diagnostics Check	OK	11:40

---

Data Download Date/Time: 06/30/2020 13:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICCARDI,A,  
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD  
ARRESTING OFFICER: RICCARDI,A  
ARRESTING OFFICER AGENCY: Port St. Lucie PD  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A4  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 06/24/2020 03:09:16

---

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.079	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.112	03:04
Breath Volume		1.351	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.125	03:07
Breath Volume		2.472	03:07
Air Blank		0.000	03:08
Control Test		0.078	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

---

Data Download Date/Time: 06/30/2020 13:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIGNORELLIA,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: NEGRON,S  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: SCHWARZ

DATE/TIME OF TEST: 06/04/2020 23:40:58

---

Diagnostics Check	OK	23:38	
Air Blank	0.000	23:38	
Control Test	0.079	23:38	
Air Blank	0.000	23:39	
Subject Sample	# 1	23:39	Subject Test Refused
Breath Volume	0.000	23:39	
Air Blank	0.000	23:40	
Control Test	0.079	23:40	
Air Blank	0.000	23:40	
Diagnostics Check	OK	23:40	

---

Data Download Date/Time: 06/05/2020 10:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CINTRON,J,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: CASTANO,C  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: HAZZARD

DATE/TIME OF TEST: 06/06/2020 04:01:05

---

Diagnostics Check	OK	03:58	
Air Blank	0.000	03:58	
Control Test	0.080	03:58	
Air Blank	0.000	03:59	
Subject Sample	# 1	03:59	Subject Test Refused
Breath Volume	0.000	03:59	
Air Blank	0.000	04:00	
Control Test	0.079	04:00	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:01	

---

Data Download Date/Time: 07/03/2020 13:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUDOLPH,B,D  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: YORK,J  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: PEACOCK

DATE/TIME OF TEST: 06/06/2020 05:35:57

---

Diagnostics Check	OK	05:29
Air Blank	0.000	05:29
Control Test	0.080	05:30
Air Blank	0.000	05:30
Subject Sample # 1	0.168	05:31
Breath Volume	1.976	05:31
Air Blank	0.000	05:31
Air Blank	0.000	05:33
Subject Sample # 2	0.165	05:34
Breath Volume	1.492	05:34
Air Blank	0.000	05:35
Control Test	0.080	05:35
Air Blank	0.000	05:35
Diagnostics Check	OK	05:35

---

Data Download Date/Time: 07/03/2020 13:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CINTRON,J,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: RUIZ,M  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: JAVIER

DATE/TIME OF TEST: 06/07/2020 01:36:19

---

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:29	
Control Test	0.079	01:29	
Air Blank	0.000	01:30	
Subject Sample # 1	0.204	01:32	
Breath Volume	2.082	01:32	
Air Blank	0.000	01:32	
Air Blank	0.000	01:34	
Subject Sample # 2		01:34	Subject Test Refused
Breath Volume	0.000	01:34	
Air Blank	0.000	01:35	
Control Test	0.079	01:35	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

---

Data Download Date/Time: 07/03/2020 13:26



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,T  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: OREROSTHY,DEPUTY  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: MACHADO

DATE/TIME OF TEST: 06/07/2020 19:52:02

---

Diagnostics Check	OK	19:51	
Air Blank		19:51	RFI Detect
Air Blank	0.000	19:52	

---

Data Download Date/Time: 07/03/2020 13:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: OREHOSTHY,DEPUTY  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: MACHADO

DATE/TIME OF TEST: 06/07/2020 20:00:05

---

Diagnostics Check	OK	19:53
Air Blank	0.000	19:54
Control Test	0.080	19:54
Air Blank	0.000	19:55
Subject Sample # 1	0.219	19:55
Breath Volume	3.128	19:55
Air Blank	0.000	19:56
Air Blank	0.000	19:58
Subject Sample # 2	0.219	19:58
Breath Volume	2.218	19:58
Air Blank	0.000	19:59
Control Test	0.080	19:59
Air Blank	0.000	20:00
Diagnostics Check	OK	20:00

---

Data Download Date/Time: 07/03/2020 13:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KANE,JAMES,G  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: DOTSON,NATALIE  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 07:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: ASDELL

DATE/TIME OF TEST: 06/08/2020 08:16:27

---

Diagnostics Check	OK	08:13	
Air Blank	0.000	08:13	
Control Test	0.080	08:14	
Air Blank	0.000	08:14	
Subject Sample	# 1	08:15	RFI Detect
Breath Volume	0.000	08:15	
Air Blank	0.000	08:15	
Control Test	0.079	08:15	
Air Blank	0.000	08:16	
Diagnostics Check	OK	08:16	

---

Data Download Date/Time: 07/03/2020 13:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KANE,JAMES,G  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: DOTSON,NATALIE  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 07:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: ASDELL

DATE/TIME OF TEST: 06/08/2020 08:23:52

---

Diagnostics Check	OK	08:17
Air Blank	0.000	08:18
Control Test	0.079	08:18
Air Blank	0.000	08:18
Subject Sample # 1	0.000	08:19
Breath Volume	2.062	08:19
Air Blank	0.000	08:20
Air Blank	0.000	08:22
Subject Sample # 2	0.000	08:22
Breath Volume	1.851	08:22
Air Blank	0.000	08:22
Control Test	0.079	08:23
Air Blank	0.000	08:23
Diagnostics Check	OK	08:23

---

Data Download Date/Time: 07/03/2020 13:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KANE,JAMES,G  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: LUTHER,CHRIS  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 11:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:47

SUBJECT'S LAST NAME: ELLISON

DATE/TIME OF TEST: 06/08/2020 13:18:39

---

Diagnostics Check	OK	13:09	
Air Blank	0.000	13:10	
Control Test	0.080	13:10	
Air Blank	0.000	13:10	
Subject Sample	# 1	13:14	No Sample Provided
Breath Volume	0.000	13:14	
Air Blank	0.000	13:14	
Air Blank	0.000	13:16	
Subject Sample	# 2	13:17	Subject Test Refused
Breath Volume	0.000	13:17	
Air Blank	0.000	13:17	
Control Test	0.080	13:18	
Air Blank	0.000	13:18	
Diagnostics Check	OK	13:18	

---

Data Download Date/Time: 07/03/2020 13:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KANTORSKI,MARK,A  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: MACKENZIE,DOUG  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 10:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: HOLDERFIELD

DATE/TIME OF TEST: 06/09/2020 11:30:26

---

Diagnostics Check	OK	11:22
Air Blank	0.000	11:22
Control Test	0.079	11:23
Air Blank	0.000	11:23
Subject Sample # 1	0.000	11:24
Breath Volume	2.132	11:24
Air Blank	0.000	11:25
Air Blank	0.000	11:27
Subject Sample # 2	0.000	11:29
Breath Volume	1.726	11:29
Air Blank	0.000	11:29
Control Test	0.079	11:29
Air Blank	0.000	11:30
Diagnostics Check	OK	11:30

---

Data Download Date/Time: 07/03/2020 13:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KANTORSKI,MARK,A  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: OREHOSTKY,G  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 10:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: BARBOSA

DATE/TIME OF TEST: 06/09/2020 11:52:17

---

Diagnostics Check	OK	11:50	
Air Blank	0.000	11:51	
Control Test	0.080	11:51	
Air Blank	0.000	11:52	
		11:52	RFI Detect
Air Blank		11:52	RFI Detect

---

Data Download Date/Time: 07/03/2020 13:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KANTORSKI,MARK,A  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: OREHOSKY,G  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 10:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: BARBOSA

DATE/TIME OF TEST: 06/09/2020 12:02:56

---

Diagnostics Check	OK	11:54
Air Blank	0.000	11:54
Control Test	0.080	11:55
Air Blank	0.000	11:55
Subject Sample # 1	0.000	11:56
Breath Volume	1.531	11:56
Air Blank	0.000	11:56
Air Blank	0.000	11:58
Subject Sample # 2	0.000	12:01
Breath Volume	1.472	12:01
Air Blank	0.000	12:02
Control Test	0.080	12:02
Air Blank	0.000	12:02
Diagnostics Check	OK	12:02

---

Data Download Date/Time: 07/03/2020 13:26



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PLASSE, RAYMOND, J  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: MEMILLEN, TIFFIANY  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: AZIZ

DATE/TIME OF TEST: 06/11/2020 01:03:19

---

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.079	00:57
Air Blank	0.000	00:57
Subject Sample # 1	0.069	00:58
Breath Volume	2.058	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.073	01:01
Breath Volume	2.097	01:01
Air Blank	0.000	01:02
Control Test	0.080	01:02
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

---

Data Download Date/Time: 07/03/2020 13:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRIQUEZ ZUNIGA,MARCIO,E  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: HENRIQUEZ ZUNIGA,MARCIO  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 13:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:49

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 06/11/2020 15:21:00

---

Diagnostics Check	OK	15:14
Air Blank	0.000	15:15
Control Test	0.080	15:15
Air Blank	0.000	15:15
Subject Sample # 1	0.000	15:16
Breath Volume	1.394	15:16
Air Blank	0.000	15:17
Air Blank	0.000	15:19
Subject Sample # 2	0.000	15:19
Breath Volume	1.964	15:19
Air Blank	0.000	15:20
Control Test	0.080	15:20
Air Blank	0.000	15:20
Diagnostics Check	OK	15:21

---

Data Download Date/Time: 07/03/2020 13:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KEARSE,ANGIE,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: RUIZ,M  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: STEFFENS

DATE/TIME OF TEST: 06/12/2020 03:57:26

---

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.151	03:52
Breath Volume	1.199	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.156	03:55
Breath Volume	1.175	03:55
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

---

Data Download Date/Time: 07/03/2020 13:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUDOLPH,B,D  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: MANGEL,M  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 06/13/2020 00:37:57

---

Diagnostics Check	OK	00:35	
Air Blank	0.000	00:35	
Control Test	0.080	00:35	
Air Blank	0.000	00:36	
Subject Sample	# 1	00:36	Subject Test Refused
Breath Volume	0.000	00:36	
Air Blank	0.000	00:37	
Control Test	0.079	00:37	
Air Blank	0.000	00:37	
Diagnostics Check	OK	00:37	

---

Data Download Date/Time: 07/03/2020 13:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: LEVERING, BRYAN  
ARRESTING OFFICER AGENCY: Indian River Shores  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: QUINONES GONZALEZ

DATE/TIME OF TEST: 06/13/2020 21:08:24

---

Diagnostics Check	OK	21:01
Air Blank	0.000	21:01
Control Test	0.080	21:01
Air Blank	0.000	21:02
Subject Sample # 1	0.121	21:03
Breath Volume	1.789	21:03
Air Blank	0.000	21:04
Air Blank	0.000	21:06
Subject Sample # 2	0.124	21:06
Breath Volume	1.179	21:06
Air Blank	0.000	21:07
Control Test	0.081	21:07
Air Blank	0.000	21:08
Diagnostics Check	OK	21:08

---

Data Download Date/Time: 07/03/2020 13:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: SEBRIS,DEPUTY  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MUNGO

DATE/TIME OF TEST: 06/14/2020 22:32:05

---

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:29	
Control Test	0.080	22:29	
Air Blank	0.000	22:30	
Subject Sample	# 1	22:30	Subject Test Refused
Breath Volume	0.000	22:30	
Air Blank	0.000	22:31	
Control Test	0.080	22:31	
Air Blank	0.000	22:32	
Diagnostics Check	OK	22:32	

---

Data Download Date/Time: 07/03/2020 13:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: SKOVGARD,K  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/14/2020 23:43:53

---

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:41	
Control Test	0.080	23:41	
Air Blank	0.000	23:42	
Subject Sample	# 1	23:42	Subject Test Refused
Breath Volume	0.000	23:42	
Air Blank	0.000	23:42	
Control Test	0.079	23:43	
Air Blank	0.000	23:43	
Diagnostics Check	OK	23:43	

---

Data Download Date/Time: 07/03/2020 13:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OOLEY,JAMES,W  
BREATH TEST OPERATOR AGENCY: Vero Beach PD  
ARRESTING OFFICER: OOLEY,JAMES  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: HERBIG

DATE/TIME OF TEST: 06/16/2020 18:16:46

---

Diagnostics Check	OK	18:10
Air Blank	0.000	18:11
Control Test	0.080	18:11
Air Blank	0.000	18:12
Subject Sample # 1	0.000	18:12
Breath Volume	1.308	18:12
Air Blank	0.000	18:12
Air Blank	0.000	18:14
Subject Sample # 2	0.000	18:15
Breath Volume	1.558	18:15
Air Blank	0.000	18:15
Control Test	0.079	18:16
Air Blank	0.000	18:16
Diagnostics Check	OK	18:16

---

Data Download Date/Time: 07/03/2020 13:26



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CINTRON,J,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: HATTON,N  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: PAGUADA CONTRERAS

DATE/TIME OF TEST: 06/17/2020 02:16:33

---

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:14	
Control Test	0.079	02:14	
Air Blank	0.000	02:14	
Subject Sample	# 1	02:15	Subject Test Refused
Breath Volume	0.000	02:15	
Air Blank	0.000	02:15	
Control Test	0.080	02:15	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

---

Data Download Date/Time: 07/03/2020 13:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHILE,C,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: NEGRON,S  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 06/19/2020 22:19:47

---

Diagnostics Check	OK	22:16	
Air Blank	0.000	22:17	
Control Test	0.080	22:17	
Air Blank	0.000	22:18	
Subject Sample	# 1	22:18	Subject Test Refused
Breath Volume	0.000	22:18	
Air Blank	0.000	22:18	
Control Test	0.079	22:19	
Air Blank	0.000	22:19	
Diagnostics Check	OK	22:19	

---

Data Download Date/Time: 07/03/2020 13:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUDOLPH,B,D  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: DAUGHERTY,P  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: TERRACE

DATE/TIME OF TEST: 06/20/2020 02:56:04

---

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.079	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.216	02:50
Breath Volume	1.640	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:52
Subject Sample # 2	0.214	02:54
Breath Volume	1.558	02:54
Air Blank	0.000	02:55
Control Test	0.079	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:56

---

Data Download Date/Time: 07/03/2020 13:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: WARD,DEPUTY  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 06/20/2020 19:11:55

---

Diagnostics Check	OK	19:08	
Air Blank	0.000	19:09	
Control Test	0.080	19:09	
Air Blank	0.000	19:10	
Subject Sample	# 1	19:10	Subject Test Refused
Breath Volume	0.000	19:10	
Air Blank	0.000	19:10	
Control Test	0.080	19:11	
Air Blank		19:11	RFI Detect
Air Blank	0.000	19:11	

---

Data Download Date/Time: 07/03/2020 13:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CINTRON,J,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: MANGEL,M  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 06/21/2020 00:33:42

---

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.349	00:28
Breath Volume	2.828	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:30
Subject Sample # 2	0.356	00:32
Breath Volume	3.343	00:32
Air Blank	0.000	00:32
Control Test	0.078	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

---

Data Download Date/Time: 07/03/2020 13:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIGMORELLIA,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: NIEVES,L  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: LARA

DATE/TIME OF TEST: 06/22/2020 18:07:04

---

Diagnostics Check	OK	17:55	
Air Blank	0.000	17:56	
Control Test	0.080	17:56	
Air Blank	0.000	17:56	
Subject Sample	# 1	18:00	No Sample Provided
Breath Volume	0.000	18:00	
Air Blank	0.000	18:00	
Air Blank	0.000	18:02	
Subject Sample	# 2	18:05	No Sample Provided
Breath Volume	0.000	18:05	
Air Blank	0.000	18:06	
Control Test	0.080	18:06	
Air Blank	0.000	18:06	
Diagnostics Check	OK	18:07	

---

Data Download Date/Time: 07/03/2020 13:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIGNOELLI,A,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: JAWORISKI,K  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: MARIA

DATE/TIME OF TEST: 06/22/2020 19:50:30

---

Diagnostics Check	OK	19:43
Air Blank	0.000	19:44
Control Test	0.080	19:44
Air Blank	0.000	19:45
Subject Sample # 1	0.000	19:46
Breath Volume	1.585	19:46
Air Blank	0.000	19:46
Air Blank	0.000	19:48
Subject Sample # 2	0.000	19:49
Breath Volume	1.554	19:49
Air Blank	0.000	19:49
Control Test	0.080	19:49
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

---

Data Download Date/Time: 07/03/2020 13:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIGNORELLIA,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: BRASHEARS,E  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 06/23/2020 19:35:27

---

Diagnostics Check	OK	19:32	
Air Blank	0.000	19:33	
Control Test	0.081	19:33	
Air Blank	0.000	19:33	
Subject Sample	# 1	19:34	Subject Test Refused
Breath Volume	0.000	19:34	
Air Blank	0.000	19:34	
Control Test	0.080	19:34	
Air Blank	0.000	19:35	
Diagnostics Check	OK	19:35	

---

Data Download Date/Time: 07/03/2020 13:26



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OOLEY, JAMES, W  
BREATH TEST OPERATOR AGENCY: Vero Beach PD  
ARRESTING OFFICER: OOLEY, JAMES  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 11:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:24

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 06/24/2020 12:59:14

---

Diagnostics Check		OK	12:46	
Air Blank		0.000	12:47	
Control Test		0.080	12:47	
Air Blank		0.000	12:48	
Subject Sample	# 1	0.150	12:51	Volume Not Met
Breath Volume		0.535	12:51	
Air Blank		0.000	12:52	
Air Blank		0.000	12:53	
Subject Sample	# 2	0.169	12:54	
Breath Volume		1.347	12:54	
Air Blank		0.000	12:55	
Air Blank		0.000	12:56	
Subject Sample	# 3	0.166	12:57	
Breath Volume		1.339	12:57	
Air Blank		0.000	12:58	
Control Test		0.077	12:58	
Air Blank		0.000	12:59	
Diagnostics Check		OK	12:59	

---

Data Download Date/Time: 07/03/2020 13:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CINTRON,J,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: MANGEL,M  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: RANDAZZO

DATE/TIME OF TEST: 06/26/2020 03:58:46

---

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.211	03:53
Breath Volume	3.710	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.198	03:57
Breath Volume	2.835	03:57
Air Blank	0.000	03:57
Control Test	0.079	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

---

Data Download Date/Time: 07/03/2020 13:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: SWEENEY, C  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: BRANDT

DATE/TIME OF TEST: 06/28/2020 21:46:12

---

Diagnostics Check	OK	21:43	
Air Blank	0.000	21:43	
Control Test	0.079	21:44	
Air Blank	0.000	21:44	
Subject Sample	# 1	21:44	Subject Test Refused
Breath Volume	0.000	21:44	
Air Blank	0.000	21:45	
Control Test	0.080	21:45	
Air Blank	0.000	21:46	
Diagnostics Check	OK	21:46	

---

Data Download Date/Time: 07/03/2020 13:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREEN,BRIAN,P  
BREATH TEST OPERATOR AGENCY: Vero Beach PD  
ARRESTING OFFICER: AGUILAR,RODRIGO  
ARRESTING OFFICER AGENCY: Vero Beach PD  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: DISMUKE

DATE/TIME OF TEST: 06/29/2020 04:16:49

---

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.079	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.237	04:11
Breath Volume	1.488	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.232	04:15
Breath Volume	1.296	04:15
Air Blank	0.000	04:15
Control Test	0.079	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

---

Data Download Date/Time: 07/03/2020 13:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIGNORELLIA,  
BREATH TEST OPERATOR AGENCY: Indian River Co SO  
ARRESTING OFFICER: DOTSON,N  
ARRESTING OFFICER AGENCY: Indian River Co SO  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: BAUMANN

DATE/TIME OF TEST: 06/30/2020 21:34:56

---

Diagnostics Check	OK	21:32	
Air Blank	0.000	21:32	
Control Test	0.080	21:32	
Air Blank	0.000	21:33	
Subject Sample	# 1	21:33	Subject Test Refused
Breath Volume	0.000	21:33	
Air Blank	0.000	21:34	
Control Test	0.080	21:34	
Air Blank	0.000	21:34	
Diagnostics Check	OK	21:34	

---

Data Download Date/Time: 07/03/2020 13:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRANDON,LONEIJA,P  
BREATH TEST OPERATOR AGENCY: DeSoto County SO  
ARRESTING OFFICER: NUNEZ,NIURVIS  
ARRESTING OFFICER AGENCY: DeSoto County SO  
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: KNAPP

DATE/TIME OF TEST: 06/05/2020 03:40:09

---

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:30	
Control Test	0.081	03:30	
Air Blank	0.000	03:31	
Subject Sample	# 1 0.152	03:34	Volume Not Met
Breath Volume	0.500	03:34	
Air Blank	0.000	03:35	
Air Blank	0.000	03:37	
Subject Sample	# 2	03:38	Subject Test Refused
Breath Volume	0.140	03:38	
Air Blank	0.000	03:39	
Control Test	0.079	03:39	
Air Blank	0.000	03:40	
Diagnostics Check	OK	03:40	

---

Data Download Date/Time: 06/30/2020 14:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COWELL,DECKITA,R  
BREATH TEST OPERATOR AGENCY: DeSoto County SO  
ARRESTING OFFICER: ULLRICH,LAURA  
ARRESTING OFFICER AGENCY: DeSoto County SO  
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:33

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 06/12/2020 16:06:52

---

Diagnostics Check	OK	15:58
Air Blank	0.000	15:58
Control Test	0.081	15:59
Air Blank	0.000	15:59
Subject Sample # 1	0.206	16:02
Breath Volume	3.917	16:02
Air Blank	0.000	16:02
Air Blank	0.000	16:04
Subject Sample # 2	0.208	16:05
Breath Volume	3.296	16:05
Air Blank	0.000	16:05
Control Test	0.080	16:06
Air Blank	0.000	16:06
Diagnostics Check	OK	16:06

---

Data Download Date/Time: 06/30/2020 14:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VEGA,JASON,  
BREATH TEST OPERATOR AGENCY: DeSoto County SO  
ARRESTING OFFICER: CARROLL,STEVEN  
ARRESTING OFFICER AGENCY: Arcadia PD  
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: GONZALEZ-MARTINEZ

DATE/TIME OF TEST: 06/14/2020 01:13:30

---

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:11	
Control Test	0.080	01:11	
Air Blank	0.000	01:11	
Subject Sample	# 1	01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:12	
Control Test	0.080	01:12	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

---

Data Download Date/Time: 06/30/2020 14:59



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,JOHN,P  
BREATH TEST OPERATOR AGENCY: DeSoto County SO  
ARRESTING OFFICER: PAELOW,KYLE  
ARRESTING OFFICER AGENCY: DeSoto County SO  
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 06/22/2020 00:53:17

---

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:50	
Control Test	0.081	00:51	
Air Blank	0.000	00:51	
Subject Sample	# 1	00:51	Subject Test Refused
Breath Volume	0.000	00:51	
Air Blank	0.000	00:52	
Control Test	0.081	00:52	
Air Blank	0.000	00:53	
Diagnostics Check	OK	00:53	

---

Data Download Date/Time: 06/30/2020 14:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORTIZ,ROBERTO,R  
BREATH TEST OPERATOR AGENCY: DeSoto County SO  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: DeSoto County SO  
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: FORMAN

DATE/TIME OF TEST: 06/27/2020 00:30:57

---

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:28	
Control Test	0.081	00:28	
Air Blank	0.000	00:29	
Subject Sample	# 1	00:29	Subject Test Refused
Breath Volume	0.000	00:29	
Air Blank	0.000	00:30	
Control Test	0.081	00:30	
Air Blank	0.000	00:30	
Diagnostics Check	OK	00:30	

---

Data Download Date/Time: 06/30/2020 14:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: LUNA

DATE/TIME OF TEST: 06/02/2020 22:10:43

---

Diagnostics Check	OK	22:07	
Air Blank	0.000	22:07	
Control Test	0.081	22:08	
Air Blank	0.000	22:08	
Subject Sample	# 1	22:09	Subject Test Refused
Breath Volume	0.000	22:09	
Air Blank	0.000	22:09	
Control Test	0.080	22:10	
Air Blank	0.000	22:10	
Diagnostics Check	OK	22:10	

---

Data Download Date/Time: 06/30/2020 06:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MITCHELL,NICKKI,N  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: ROWELL,J  
ARRESTING OFFICER AGENCY: Venice PD  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BAYGULOVA

DATE/TIME OF TEST: 06/03/2020 03:16:18

---

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:15	
Control Test	0.080	03:15	
Air Blank		03:15	RFI Detect
Air Blank	0.000	03:16	

---

Data Download Date/Time: 06/30/2020 06:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MITCHELL,N,N  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: ROWELL,J  
ARRESTING OFFICER AGENCY: Venice PD  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BAYGULOVA

DATE/TIME OF TEST: 06/03/2020 03:26:56

---

Diagnostics Check		OK	03:20
Air Blank		0.000	03:20
Control Test		0.080	03:20
Air Blank		0.000	03:21
Subject Sample	# 1	0.202	03:22
Breath Volume		2.652	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:24
Subject Sample	# 2	0.185	03:25
Breath Volume		2.781	03:25
Air Blank		0.000	03:26
Control Test		0.077	03:26
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

---

Data Download Date/Time: 06/30/2020 06:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANDERS,DEPUTY,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: SANDERS,DEPUTY J.  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: DURAKU

DATE/TIME OF TEST: 06/03/2020 23:07:46

---

Diagnostics Check	OK	23:01
Air Blank	0.000	23:02
Control Test	0.080	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.008	23:03
Breath Volume	2.445	23:03
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.009	23:06
Breath Volume	1.957	23:06
Air Blank	0.000	23:06
Control Test	0.080	23:07
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

---

Data Download Date/Time: 06/30/2020 06:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,D,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BELL,D  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: ARMONTROUT

DATE/TIME OF TEST: 06/04/2020 00:04:23

---

Diagnostics Check	OK	23:57
Air Blank	0.000	23:58
Control Test	0.079	23:58
Air Blank	0.000	23:59
Subject Sample # 2	0.150	23:59
Breath Volume	1.574	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 1	0.146	00:02
Breath Volume	2.488	00:02
Air Blank	0.000	00:03
Control Test	0.078	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

---

Data Download Date/Time: 06/30/2020 06:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENCKLE,DEP,T  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BRENCKLE,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: GARTH ZAMORA

DATE/TIME OF TEST: 06/08/2020 01:03:09

---

Diagnostics Check	OK	00:49	
Air Blank	0.000	00:49	
Control Test	0.079	00:49	
Air Blank	0.000	00:50	
Subject Sample	# 1	00:53	No Sample Provided
Breath Volume	0.000	00:53	
Air Blank	0.000	00:53	
Air Blank	0.000	00:56	
Subject Sample	# 2	0.238	00:57
Breath Volume	1.222	00:57	
Air Blank	0.000	00:58	
Air Blank	0.000	01:00	
Subject Sample	# 3	0.226	01:01
Breath Volume	1.285	01:01	
Air Blank	0.000	01:02	
Control Test	0.078	01:02	
Air Blank	0.000	01:03	
Diagnostics Check	OK	01:03	

---

Data Download Date/Time: 06/30/2020 06:40



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: HALEMAN

DATE/TIME OF TEST: 06/08/2020 01:17:34

---

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.079	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.171	01:12
Breath Volume	1.832	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.171	01:16
Breath Volume	2.187	01:16
Air Blank	0.000	01:16
Control Test	0.079	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

---

Data Download Date/Time: 06/30/2020 06:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANDERS,DEPUTY,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: SANDERS,DEPUTY J.  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: ESQUIVEL MARQUEZ

DATE/TIME OF TEST: 06/09/2020 02:57:40

---

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.080	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.000	02:50
Breath Volume	1.824	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.000	02:56
Breath Volume	1.453	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

---

Data Download Date/Time: 06/30/2020 06:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,D,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BELL,D  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: TURNBULL

DATE/TIME OF TEST: 06/09/2020 22:39:49

---

Diagnostics Check	OK	22:33
Air Blank	0.000	22:33
Control Test	0.080	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.105	22:35
Breath Volume	3.765	22:35
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.099	22:38
Breath Volume	3.300	22:38
Air Blank	0.000	22:38
Control Test	0.076	22:39
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

---

Data Download Date/Time: 06/30/2020 06:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANDERS,DEPUTY,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: SANDERS,DEPUTY J.  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: WARING

DATE/TIME OF TEST: 06/10/2020 00:01:10

---

Diagnostics Check	OK	23:54
Air Blank	0.000	23:55
Control Test	0.079	23:55
Air Blank	0.000	23:56
Subject Sample # 1	0.156	23:56
Breath Volume	4.550	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 2	0.157	23:59
Breath Volume	4.812	23:59
Air Blank	0.000	00:00
Control Test	0.078	00:00
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

---

Data Download Date/Time: 06/30/2020 06:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,KYLE,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MARTIN,KYLE  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: JUDSON

DATE/TIME OF TEST: 06/11/2020 00:04:10

---

Diagnostics Check	OK	23:57
Air Blank	0.000	23:58
Control Test	0.080	23:58
Air Blank	0.000	23:59
Subject Sample # 2	0.062	23:59
Breath Volume	1.593	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 1	0.063	00:02
Breath Volume	1.738	00:02
Air Blank	0.000	00:03
Control Test	0.078	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

---

Data Download Date/Time: 06/30/2020 06:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUEHLER,DEP,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BUEHLER,M  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: MOUHOT

DATE/TIME OF TEST: 06/11/2020 02:26:30

---

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:23	
Control Test	0.080	02:24	
Air Blank	0.000	02:24	
Subject Sample	# 1 0.007	02:25	
Breath Volume	1.812	02:25	
Air Blank		02:25	RFI Detect
Air Blank		02:26	RFI Detect
Air Blank	0.000	02:26	

---

Data Download Date/Time: 06/30/2020 06:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUEHLER,DEP,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BUEHLER,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: MOUHOT

DATE/TIME OF TEST: 06/11/2020 02:35:05

---

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.079	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.000	02:30
Breath Volume	1.335	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.000	02:33
Breath Volume	1.835	02:33
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

---

Data Download Date/Time: 06/30/2020 06:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,D,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BELL,D  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: GOYTIA

DATE/TIME OF TEST: 06/12/2020 20:43:12

---

Diagnostics Check	OK	20:36
Air Blank	0.000	20:37
Control Test	0.080	20:37
Air Blank	0.000	20:37
Subject Sample # 1	0.206	20:38
Breath Volume	4.121	20:38
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.199	20:41
Breath Volume	3.871	20:41
Air Blank	0.000	20:42
Control Test	0.077	20:42
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

---

Data Download Date/Time: 06/30/2020 06:40



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WATSON,D,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: WATSON,D  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: KOBZA

DATE/TIME OF TEST: 06/13/2020 00:26:36

---

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.079	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.217	00:22
Breath Volume	1.988	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.209	00:25
Breath Volume	1.378	00:25
Air Blank	0.000	00:25
Control Test	0.078	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

---

Data Download Date/Time: 06/30/2020 06:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,D,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BELL,D  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 06/13/2020 03:43:56

---

Diagnostics Check		OK	03:37
Air Blank		0.000	03:38
Control Test		0.080	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.163	03:39
Breath Volume		1.269	03:39
Air Blank		0.000	03:40
Air Blank		0.000	03:41
Subject Sample	# 2	0.159	03:42
Breath Volume		1.207	03:42
Air Blank		0.000	03:43
Control Test		0.078	03:43
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

---

Data Download Date/Time: 06/30/2020 06:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,D,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: APPLEBEE,A  
ARRESTING OFFICER AGENCY: Sarasota PD  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: KLETCHKA

DATE/TIME OF TEST: 06/13/2020 03:59:21

---

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.079	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.177	03:54
Breath Volume	1.714	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:56
Subject Sample # 2	0.169	03:57
Breath Volume	1.414	03:57
Air Blank	0.000	03:58
Control Test	0.078	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

---

Data Download Date/Time: 06/30/2020 06:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANDERS,DEPUTY,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: SANDERS,DEPUTY J.  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: HOSPODOR

DATE/TIME OF TEST: 06/13/2020 18:59:52

---

Diagnostics Check	OK	18:53
Air Blank	0.000	18:54
Control Test	0.079	18:54
Air Blank	0.000	18:54
Subject Sample # 1	0.172	18:55
Breath Volume	4.738	18:55
Air Blank	0.000	18:56
Air Blank	0.000	18:57
Subject Sample # 2	0.164	18:58
Breath Volume	5.093	18:58
Air Blank	0.000	18:58
Control Test	0.079	18:59
Air Blank	0.000	18:59
Diagnostics Check	OK	18:59

---

Data Download Date/Time: 06/30/2020 06:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MATTHEWS,DEPUTY,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: CHAPMAN,J  
ARRESTING OFFICER AGENCY: Sarasota PD  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: RAYE

DATE/TIME OF TEST: 06/13/2020 22:57:57

---

Diagnostics Check	OK	22:55	
Air Blank	0.000	22:55	
Control Test	0.080	22:55	
Air Blank	0.000	22:56	
Subject Sample	# 1	22:56	Subject Test Refused
Breath Volume	0.000	22:56	
Air Blank	0.000	22:57	
Control Test	0.079	22:57	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

---

Data Download Date/Time: 06/30/2020 06:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,D,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BELL,D  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: INKOVA

DATE/TIME OF TEST: 06/14/2020 03:36:55

---

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.091	03:32
Breath Volume	3.957	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:34
Subject Sample # 2	0.089	03:35
Breath Volume	3.664	03:35
Air Blank	0.000	03:35
Control Test	0.079	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

---

Data Download Date/Time: 06/30/2020 06:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENCKLE,DEP,T  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BRENCKLE,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: DUCKER

DATE/TIME OF TEST: 06/15/2020 18:28:41

---

Diagnostics Check	OK	18:22
Air Blank	0.000	18:23
Control Test	0.079	18:23
Air Blank	0.000	18:24
Subject Sample # 1	0.000	18:24
Breath Volume	2.957	18:24
Air Blank	0.000	18:24
Air Blank	0.000	18:26
Subject Sample # 2	0.000	18:27
Breath Volume	3.167	18:27
Air Blank	0.000	18:27
Control Test	0.079	18:28
Air Blank	0.000	18:28
Diagnostics Check	OK	18:28

---

Data Download Date/Time: 06/30/2020 06:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: HERNANDEZ RAMIREZ

DATE/TIME OF TEST: 06/17/2020 22:46:35

---

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.080	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.176	22:42
Breath Volume	2.445	22:42
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.173	22:45
Breath Volume	1.898	22:45
Air Blank	0.000	22:45
Control Test	0.080	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

---

Data Download Date/Time: 06/30/2020 06:40



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WATSON,D,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: WATSON,D  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: VONDRASEK

DATE/TIME OF TEST: 06/18/2020 03:21:25

---

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:19	
Control Test	0.080	03:19	
Air Blank	0.000	03:19	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:20	
Control Test	0.079	03:20	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

---

Data Download Date/Time: 06/30/2020 06:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANDERS,DEPUTY,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: SANDERS,DEPUTY J  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: STARK

DATE/TIME OF TEST: 06/18/2020 05:06:24

---

Diagnostics Check	OK	04:55	
Air Blank	0.000	04:55	
Control Test	0.079	04:55	
Air Blank	0.000	04:56	
Subject Sample	# 1	04:59	No Sample Provided
Breath Volume	0.000	04:59	
Air Blank	0.000	04:59	
Air Blank	0.000	05:01	
Subject Sample	# 2	05:05	No Sample Provided
Breath Volume	0.000	05:05	
Air Blank	0.000	05:05	
Control Test	0.080	05:05	
Air Blank	0.000	05:06	
Diagnostics Check	OK	05:06	

---

Data Download Date/Time: 06/30/2020 06:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMBRIA,STEVEN,J  
BREATH TEST OPERATOR AGENCY: North Port PD  
ARRESTING OFFICER: CAMBRIA,STEVEN  
ARRESTING OFFICER AGENCY: North Port PD  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: MUCCI

DATE/TIME OF TEST: 06/18/2020 16:23:06

---

Diagnostics Check	OK	16:16
Air Blank	0.000	16:17
Control Test	0.079	16:17
Air Blank	0.000	16:18
Subject Sample # 1	0.147	16:18
Breath Volume	3.804	16:18
Air Blank	0.000	16:19
Air Blank	0.000	16:21
Subject Sample # 2	0.152	16:21
Breath Volume	3.945	16:21
Air Blank	0.000	16:22
Control Test	0.079	16:22
Air Blank	0.000	16:23
Diagnostics Check	OK	16:23

---

Data Download Date/Time: 06/30/2020 06:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WATSON,D,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: WATSON,D  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: DICKINSON

DATE/TIME OF TEST: 06/18/2020 20:27:29

---

Diagnostics Check		OK	20:10	
Air Blank		0.000	20:10	
Control Test		0.076	20:11	
Air Blank		0.000	20:11	
Subject Sample	# 1	0.164	20:14	
Breath Volume		1.109	20:14	
Air Blank		0.000	20:15	
Air Blank		0.000	20:17	
Subject Sample	# 2	0.154	20:20	Volume Not Met
Breath Volume		0.144	20:20	
Air Blank		0.000	20:20	
Air Blank		0.000	20:22	
Subject Sample	# 3	0.143	20:25	Volume Not Met
Breath Volume		0.492	20:25	
Air Blank		0.000	20:26	
Control Test		0.078	20:26	
Air Blank		0.000	20:27	
Diagnostics Check		OK	20:27	

---

Data Download Date/Time: 06/30/2020 06:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WATSON,D,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: WATSON,D  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: DICKINSON

DATE/TIME OF TEST: 06/18/2020 20:43:54

---

Diagnostics Check		OK	20:29	
Air Blank		0.000	20:29	
Control Test		0.077	20:29	
Air Blank		0.000	20:30	
Subject Sample	# 1	0.151	20:33	Volume Not Met
Breath Volume		1.066	20:33	
Air Blank		0.000	20:34	
Air Blank		0.000	20:36	
Subject Sample	# 2	0.158	20:37	
Breath Volume		1.136	20:37	
Air Blank		0.000	20:37	
Air Blank		0.000	20:39	
Subject Sample	# 3	0.152	20:42	
Breath Volume		1.113	20:42	
Air Blank		0.000	20:42	
Control Test		0.078	20:43	
Air Blank		0.000	20:43	
Diagnostics Check		OK	20:43	

---

Data Download Date/Time: 06/30/2020 06:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,D,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BELL,D  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: FULTON

DATE/TIME OF TEST: 06/19/2020 03:00:00

---

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.080	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.155	02:55
Breath Volume	1.375	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.151	02:58
Breath Volume	2.164	02:58
Air Blank	0.000	02:59
Control Test	0.078	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	03:00

---

Data Download Date/Time: 06/30/2020 06:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,KYLE,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MARTIN,KYLE  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: FELVER

DATE/TIME OF TEST: 06/20/2020 03:59:55

---

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.080	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 1	03:58	RFI Detect
Breath Volume	0.000	03:58	
Air Blank	0.000	03:59	
Control Test	0.080	03:59	
Air Blank	0.000	03:59	
Diagnostics Check	OK	03:59	

---

Data Download Date/Time: 06/30/2020 06:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,KYLE,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MARTIN,KYLE  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: FELVER

DATE/TIME OF TEST: 06/20/2020 04:08:48

---

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.078	04:02
Air Blank	0.000	04:03
Subject Sample # 1	0.223	04:04
Breath Volume	4.046	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.213	04:07
Breath Volume	3.964	04:07
Air Blank	0.000	04:07
Control Test	0.079	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

---

Data Download Date/Time: 06/30/2020 06:41



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: TAYLOR, OFC  
ARRESTING OFFICER AGENCY: Sarasota PD  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: IGLESIASPEDROSO

DATE/TIME OF TEST: 06/20/2020 23:58:30

---

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.080	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.177	23:53
Breath Volume	1.582	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.172	23:56
Breath Volume	1.468	23:56
Air Blank	0.000	23:57
Control Test	0.078	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

---

Data Download Date/Time: 06/30/2020 06:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENCKLE,DEP,T  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BRENCKLE,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: BARTOLOME

DATE/TIME OF TEST: 06/21/2020 20:04:01

---

Diagnostics Check	OK	20:00	
Air Blank	0.000	20:00	
Control Test	0.079	20:01	
Air Blank	0.000	20:01	
Subject Sample # 1	0.000	20:02	
Breath Volume	1.847	20:02	
Air Blank		20:02	RFI Detect
Air Blank	0.000	20:03	
Control Test	0.079	20:03	
Air Blank	0.000	20:03	
Diagnostics Check	OK	20:04	

---

Data Download Date/Time: 06/30/2020 06:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENCKLE,DEP,T  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BRENCKLE,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: BARTOLOME

DATE/TIME OF TEST: 06/21/2020 20:12:20

---

Diagnostics Check	OK	20:06	
Air Blank	0.000	20:06	
Control Test	0.080	20:07	
Air Blank	0.000	20:07	
Subject Sample	# 1 0.000	20:08	
Breath Volume	1.781	20:08	
Air Blank	0.000	20:08	
Air Blank	0.000	20:10	
Subject Sample	# 2 0.000	20:10	
Breath Volume	1.898	20:10	
Air Blank	0.000	20:11	
Control Test	0.079	20:11	
Air Blank		20:11	RFI Detect
Air Blank	0.000	20:12	

---

Data Download Date/Time: 06/30/2020 06:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,KYLE,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MARTIN,KYLE  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: FARLEY

DATE/TIME OF TEST: 06/21/2020 23:20:56

---

Diagnostics Check	OK	23:14
Air Blank	0.000	23:15
Control Test	0.080	23:15
Air Blank	0.000	23:15
Subject Sample # 1	0.082	23:16
Breath Volume	6.210	23:16
Air Blank	0.000	23:17
Air Blank	0.000	23:18
Subject Sample # 2	0.086	23:19
Breath Volume	6.472	23:19
Air Blank	0.000	23:20
Control Test	0.078	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

---

Data Download Date/Time: 06/30/2020 06:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WATSON,D,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: WATSON,D  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/23/2020 23:17:25

---

Diagnostics Check	OK	23:11
Air Blank	0.000	23:11
Control Test	0.078	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.213	23:12
Breath Volume	3.078	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.214	23:15
Breath Volume	3.328	23:15
Air Blank	0.000	23:16
Control Test	0.077	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

---

Data Download Date/Time: 06/30/2020 06:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,DEP,R  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BELL,DEP D  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: TOTTEN

DATE/TIME OF TEST: 06/24/2020 04:06:29

---

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.079	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.236	04:01
Breath Volume		3.656	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.238	04:04
Breath Volume		3.781	04:04
Air Blank		0.000	04:05
Control Test		0.078	04:05
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

---

Data Download Date/Time: 06/30/2020 06:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,KYLE,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MARTIN,KYLE  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MANGAS

DATE/TIME OF TEST: 06/25/2020 00:31:28

---

Diagnostics Check		OK	00:25
Air Blank		0.000	00:25
Control Test		0.079	00:25
Air Blank		0.000	00:26
Subject Sample	# 1	0.215	00:26
Breath Volume		1.382	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.226	00:29
Breath Volume		2.132	00:29
Air Blank		0.000	00:30
Control Test		0.079	00:30
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

---

Data Download Date/Time: 06/30/2020 06:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,KYLE,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MARTIN,KYLE  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: WATTS

DATE/TIME OF TEST: 06/25/2020 21:04:18

---

Diagnostics Check	OK	20:57
Air Blank	0.000	20:58
Control Test	0.080	20:58
Air Blank	0.000	20:59
Subject Sample # 1	0.000	20:59
Breath Volume	2.703	20:59
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample # 2	0.000	21:02
Breath Volume	1.636	21:02
Air Blank	0.000	21:03
Control Test	0.079	21:03
Air Blank	0.000	21:04
Diagnostics Check	OK	21:04

---

Data Download Date/Time: 06/30/2020 06:41



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENCKLE,DEP,T  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BRENCKLE,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: LITKE

DATE/TIME OF TEST: 06/25/2020 21:39:57

---

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.080	21:33
Air Blank	0.000	21:34
Subject Sample # 1	0.139	21:35
Breath Volume	1.375	21:35
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.137	21:38
Breath Volume	1.125	21:38
Air Blank	0.000	21:39
Control Test	0.080	21:39
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

---

Data Download Date/Time: 06/30/2020 06:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAEUSSLER,BRETT,W  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: HAEUSSLER,BRETT  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: COLWELL

DATE/TIME OF TEST: 06/26/2020 03:23:43

---

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:18
Air Blank	0.000	03:18
Subject Sample # 1	0.000	03:19
Breath Volume	2.328	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.000	03:22
Breath Volume	1.277	03:22
Air Blank	0.000	03:22
Control Test	0.081	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

---

Data Download Date/Time: 06/30/2020 06:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,D,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BELL,D  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 06/26/2020 19:05:02

---

Diagnostics Check	OK	18:59
Air Blank	0.000	18:59
Control Test	0.079	18:59
Air Blank	0.000	19:00
Subject Sample # 1	0.141	19:00
Breath Volume	2.121	19:00
Air Blank	0.000	19:01
Air Blank	0.000	19:03
Subject Sample # 2	0.134	19:03
Breath Volume	2.226	19:03
Air Blank	0.000	19:04
Control Test	0.077	19:04
Air Blank	0.000	19:04
Diagnostics Check	OK	19:05

---

Data Download Date/Time: 06/30/2020 06:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,DEP,R  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MILLER,DEP R  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/27/2020 22:32:31

---

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.080	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.082	22:27
Breath Volume	2.683	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.079	22:30
Breath Volume	3.722	22:30
Air Blank	0.000	22:31
Control Test	0.079	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

---

Data Download Date/Time: 06/30/2020 06:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENCKLE,DEP,T  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BRENCKLE,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/30/2020 23:59:17

---

Diagnostics Check	OK	23:58	
Air Blank		23:58	RFI Detect
Air Blank	0.000	23:59	

---

Data Download Date/Time: 07/29/2020 08:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WADE,CODY,L  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: WADE,CODY  
ARRESTING OFFICER AGENCY: Eustis PD  
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1031121  
DRY GAS EXPIRY DATE: 10/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 06/06/2020 02:27:01

---

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.080	02:18
Air Blank		0.000	02:19
Subject Sample	# 1	0.224	02:22
Breath Volume		2.769	02:22
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.222	02:25
Breath Volume		1.773	02:25
Air Blank		0.000	02:26
Control Test		0.077	02:26
Air Blank		0.000	02:26
Diagnostics Check		OK	02:27

---

Data Download Date/Time: 06/22/2020 08:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STARLING, WILLIAM, A  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: STARLING, WILLIAM  
ARRESTING OFFICER AGENCY: Eustis PD  
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1031121  
DRY GAS EXPIRY DATE: 10/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: SKOG-RIOS

DATE/TIME OF TEST: 06/17/2020 23:18:04

---

Diagnostics Check	OK	23:12	
Air Blank	0.000	23:13	
Control Test	0.080	23:13	
Air Blank	0.000	23:14	
Subject Sample	# 1	23:16	RFI Detect
Breath Volume	0.000	23:16	
Air Blank	0.000	23:17	
Control Test	0.080	23:17	
Air Blank	0.000	23:17	
Diagnostics Check	OK	23:18	

---

Data Download Date/Time: 06/22/2020 08:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STARLING, WILLIAM, A  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: STARLING, WILLIAM  
ARRESTING OFFICER AGENCY: Eustis PD  
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1031121  
DRY GAS EXPIRY DATE: 10/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: SKOG-RIOS

DATE/TIME OF TEST: 06/17/2020 23:24:07

---

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:21	
Control Test	0.080	23:21	
Air Blank	0.000	23:22	
Subject Sample	# 1	23:22	Subject Test Refused
Breath Volume	0.000	23:22	
Air Blank	0.000	23:23	
Control Test	0.080	23:23	
Air Blank	0.000	23:24	
Diagnostics Check	OK	23:24	

---

Data Download Date/Time: 06/22/2020 08:13



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,DEP,R  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MILLER,DEP R  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: LEWELLYN

DATE/TIME OF TEST: 06/10/2020 04:10:14

---

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.079	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.111	04:05
Breath Volume	3.632	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:07
Subject Sample # 2	0.103	04:08
Breath Volume	1.812	04:08
Air Blank	0.000	04:09
Control Test	0.079	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

---

Data Download Date/Time: 06/30/2020 06:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,DEP,R  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MILLER,DEP R  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:59

SUBJECT'S LAST NAME: ESCOBAR NATATEN

DATE/TIME OF TEST: 06/14/2020 18:26:03

---

Diagnostics Check	OK	18:20
Air Blank	0.000	18:20
Control Test	0.079	18:20
Air Blank	0.000	18:21
Subject Sample # 1	0.270	18:21
Breath Volume	4.191	18:21
Air Blank	0.000	18:22
Air Blank	0.000	18:24
Subject Sample # 2	0.272	18:24
Breath Volume	3.882	18:24
Air Blank	0.000	18:25
Control Test	0.079	18:25
Air Blank	0.000	18:25
Diagnostics Check	OK	18:26

---

Data Download Date/Time: 06/30/2020 06:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR,J,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BARR,J  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A4  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: BOURQUE

DATE/TIME OF TEST: 06/05/2020 20:46:05

---

Diagnostics Check	OK	20:40
Air Blank	0.000	20:40
Control Test	0.079	20:40
Air Blank	0.000	20:41
Subject Sample # 1	0.000	20:41
Breath Volume	1.773	20:41
Air Blank	0.000	20:42
Air Blank	0.000	20:44
Subject Sample # 2	0.000	20:44
Breath Volume	1.292	20:44
Air Blank	0.000	20:45
Control Test	0.079	20:45
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

---

Data Download Date/Time: 06/30/2020 06:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR,J,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: PERRY,D  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A4  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: EARLY

DATE/TIME OF TEST: 06/05/2020 23:29:59

---

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.078	23:23
Air Blank	0.000	23:24
Subject Sample # 1	0.118	23:25
Breath Volume	2.105	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:27
Subject Sample # 2	0.117	23:28
Breath Volume	2.050	23:28
Air Blank	0.000	23:29
Control Test	0.076	23:29
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

---

Data Download Date/Time: 06/30/2020 06:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR,J,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MARTIN,K  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A4  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 06/05/2020 23:40:42

---

Diagnostics Check		OK	23:34
Air Blank		0.000	23:34
Control Test		0.076	23:34
Air Blank		0.000	23:35
Subject Sample	# 1	0.251	23:36
Breath Volume		1.292	23:36
Air Blank		0.000	23:36
Air Blank		0.000	23:38
Subject Sample	# 2	0.248	23:39
Breath Volume		1.875	23:39
Air Blank		0.000	23:39
Control Test		0.076	23:40
Air Blank		0.000	23:40
Diagnostics Check		OK	23:40

---

Data Download Date/Time: 06/30/2020 06:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENCKLE,DEP,T  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BRENCKLE,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/06/2020 01:58:50

---

Diagnostics Check		OK	01:52
Air Blank		0.000	01:53
Control Test		0.078	01:53
Air Blank		0.000	01:54
Subject Sample	# 1	0.112	01:54
Breath Volume		3.171	01:54
Air Blank		0.000	01:55
Air Blank		0.000	01:57
Subject Sample	# 2	0.119	01:57
Breath Volume		3.476	01:57
Air Blank		0.000	01:57
Control Test		0.077	01:58
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

---

Data Download Date/Time: 06/30/2020 06:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR,J,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BARR,J  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: ANZELONE

DATE/TIME OF TEST: 06/06/2020 17:33:49

---

Diagnostics Check	OK	17:27
Air Blank	0.000	17:28
Control Test	0.078	17:28
Air Blank	0.000	17:28
Subject Sample # 1	0.171	17:29
Breath Volume	2.097	17:29
Air Blank	0.000	17:30
Air Blank	0.000	17:31
Subject Sample # 2	0.158	17:32
Breath Volume	1.218	17:32
Air Blank	0.000	17:32
Control Test	0.078	17:33
Air Blank	0.000	17:33
Diagnostics Check	OK	17:33

---

Data Download Date/Time: 06/30/2020 06:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARR,J,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BARR,J  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A2  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:14

SUBJECT'S LAST NAME: HELMLINGER

DATE/TIME OF TEST: 06/21/2020 16:41:15

---

Diagnostics Check	OK	16:34
Air Blank	0.000	16:35
Control Test	0.078	16:35
Air Blank	0.000	16:35
Subject Sample # 1	0.197	16:36
Breath Volume	1.164	16:36
Air Blank	0.000	16:37
Air Blank	0.000	16:39
Subject Sample # 2	0.196	16:39
Breath Volume	1.132	16:39
Air Blank	0.000	16:40
Control Test	0.076	16:40
Air Blank	0.000	16:41
Diagnostics Check	OK	16:41

---

Data Download Date/Time: 06/30/2020 06:33



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHWOCHO,ALAN,N  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: TURNER,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: BOSWELL

DATE/TIME OF TEST: 06/01/2020 19:00:38

---

Diagnostics Check	OK	18:57	
Air Blank	0.000	18:57	
Control Test	0.079	18:58	
Air Blank	0.000	18:58	
Subject Sample # 1	0.018	18:59	
Breath Volume	2.683	18:59	
Air Blank		18:59	Improper Sample
Air Blank	0.000	18:59	
Control Test	0.079	19:00	
Air Blank	0.000	19:00	
Diagnostics Check	OK	19:00	

---

Data Download Date/Time: 06/30/2020 07:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHWOCHO,ALAN,N  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: TURNER,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: BOSWELL

DATE/TIME OF TEST: 06/01/2020 19:10:13

---

Diagnostics Check	OK	19:04
Air Blank	0.000	19:04
Control Test	0.079	19:04
Air Blank	0.000	19:05
Subject Sample # 1	0.017	19:05
Breath Volume	2.949	19:05
Air Blank	0.000	19:06
Air Blank	0.000	19:08
Subject Sample # 2	0.016	19:08
Breath Volume	2.910	19:08
Air Blank	0.000	19:09
Control Test	0.079	19:09
Air Blank	0.000	19:10
Diagnostics Check	OK	19:10

---

Data Download Date/Time: 06/30/2020 07:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LARKINS,CLINTON,T  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: LUETHNER,N  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: VOS

DATE/TIME OF TEST: 06/03/2020 23:52:37

---

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.078	23:46
Air Blank	0.000	23:47
Subject Sample # 1	0.184	23:47
Breath Volume	2.621	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.185	23:51
Breath Volume	3.640	23:51
Air Blank	0.000	23:51
Control Test	0.078	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

---

Data Download Date/Time: 06/30/2020 07:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EGAN,MELISSA,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: JANKOWSKI,J  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: HEADRICK

DATE/TIME OF TEST: 06/06/2020 02:14:38

---

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.079	02:08
Air Blank	0.000	02:08
Subject Sample # 1	0.185	02:09
Breath Volume	2.078	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.181	02:13
Breath Volume	1.300	02:13
Air Blank	0.000	02:13
Control Test	0.078	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

---

Data Download Date/Time: 06/30/2020 07:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: ANDERSON,J  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 06/06/2020 02:43:00

---

Diagnostics Check		OK	02:29	
Air Blank		0.000	02:30	
Control Test		0.079	02:30	
Air Blank		0.000	02:31	
Subject Sample	# 1	0.120	02:34	Volume Not Met
Breath Volume		0.031	02:34	
Air Blank		0.000	02:34	
Air Blank		0.000	02:36	
Subject Sample	# 2	0.175	02:37	
Breath Volume		2.089	02:37	
Air Blank		0.000	02:38	
Air Blank		0.000	02:40	
Subject Sample	# 3	0.171	02:41	
Breath Volume		2.289	02:41	
Air Blank		0.000	02:42	
Control Test		0.077	02:42	
Air Blank		0.000	02:42	
Diagnostics Check		OK	02:43	

---

Data Download Date/Time: 06/30/2020 07:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COURTNEY,PATRICK,R  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: ROBLES,N  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: PARTRIDGE

DATE/TIME OF TEST: 06/06/2020 03:32:08

---

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.078	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.162	03:25
Breath Volume	1.332	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.147	03:30
Breath Volume	1.464	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

---

Data Download Date/Time: 06/30/2020 07:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEONARD,ALEYA,D  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: LEUTHNER,NICHOLAS  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/09/2020 06:11:18

---

Diagnostics Check	OK	06:03
Air Blank	0.000	06:04
Control Test	0.079	06:04
Air Blank	0.000	06:05
Subject Sample # 1	0.148	06:06
Breath Volume	1.101	06:06
Air Blank	0.000	06:07
Air Blank	0.000	06:09
Subject Sample # 2	0.142	06:09
Breath Volume	1.328	06:09
Air Blank	0.000	06:10
Control Test	0.079	06:10
Air Blank	0.000	06:11
Diagnostics Check	OK	06:11

---

Data Download Date/Time: 06/30/2020 07:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONN,C,R  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: KOVAL,K  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 06/12/2020 03:00:19

---

Diagnostics Check	OK	02:53
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:55
Subject Sample # 1	0.151	02:55
Breath Volume	2.902	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.147	02:58
Breath Volume	3.382	02:58
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

---

Data Download Date/Time: 06/30/2020 07:58



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ACEBO,ANDRES,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: KOVAL,K  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: KINKIN

DATE/TIME OF TEST: 06/16/2020 02:49:44

---

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.081	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.157	02:45
Breath Volume	3.148	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.146	02:48
Breath Volume	1.746	02:48
Air Blank	0.000	02:48
Control Test	0.080	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

---

Data Download Date/Time: 06/30/2020 07:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOSZE,DEPUTY,J  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: VAUGHN,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 12:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:25

SUBJECT'S LAST NAME: DION

DATE/TIME OF TEST: 06/16/2020 14:01:15

---

Diagnostics Check	OK	13:52
Air Blank	0.000	13:53
Control Test	0.080	13:53
Air Blank	0.000	13:53
Subject Sample # 1	0.307	13:56
Breath Volume	1.554	13:56
Air Blank	0.000	13:56
Air Blank	0.000	13:58
Subject Sample # 2	0.293	13:59
Breath Volume	1.113	13:59
Air Blank	0.000	14:00
Control Test	0.079	14:00
Air Blank	0.000	14:01
Diagnostics Check	OK	14:01

---

Data Download Date/Time: 06/30/2020 07:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COURTNEY,PATRICK,R  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: WILLIAMS,E  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: EAST

DATE/TIME OF TEST: 06/17/2020 01:13:51

---

Diagnostics Check	OK	01:04
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:05
Subject Sample # 1	0.167	01:08
Breath Volume	2.816	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.152	01:12
Breath Volume	1.773	01:12
Air Blank	0.000	01:12
Control Test	0.079	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

---

Data Download Date/Time: 06/30/2020 07:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUCKER,DEPUTY,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: OKEEFE,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: KANE

DATE/TIME OF TEST: 06/17/2020 15:20:03

---

Diagnostics Check	OK	15:13
Air Blank	0.000	15:13
Control Test	0.080	15:14
Air Blank	0.000	15:14
Subject Sample # 1	0.282	15:15
Breath Volume	1.585	15:15
Air Blank	0.000	15:15
Air Blank	0.000	15:17
Subject Sample # 2	0.296	15:18
Breath Volume	1.882	15:18
Air Blank	0.000	15:19
Control Test	0.078	15:19
Air Blank	0.000	15:19
Diagnostics Check	OK	15:20

---

Data Download Date/Time: 06/30/2020 07:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHWOCHO,ALAN,N  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: EPEARDS,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: FLECK

DATE/TIME OF TEST: 06/21/2020 19:22:13

---

Diagnostics Check	OK	19:15
Air Blank	0.000	19:15
Control Test	0.080	19:16
Air Blank	0.000	19:16
Subject Sample # 1	0.170	19:17
Breath Volume	3.792	19:17
Air Blank	0.000	19:18
Air Blank	0.000	19:20
Subject Sample # 2	0.169	19:20
Breath Volume	3.468	19:20
Air Blank	0.000	19:21
Control Test	0.080	19:21
Air Blank	0.000	19:22
Diagnostics Check	OK	19:22

---

Data Download Date/Time: 06/30/2020 07:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHWOCHO,ALAN,N  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: STEALEY,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: BIBBS

DATE/TIME OF TEST: 06/21/2020 21:11:16

---

Diagnostics Check	OK	21:05
Air Blank	0.000	21:05
Control Test	0.080	21:05
Air Blank	0.000	21:06
Subject Sample # 1	0.317	21:06
Breath Volume	3.371	21:06
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.325	21:09
Breath Volume	3.253	21:09
Air Blank	0.000	21:10
Control Test	0.079	21:10
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

---

Data Download Date/Time: 06/30/2020 07:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHWOCHO,ALAN,N  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: GANK,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 06/21/2020 23:56:28

---

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:53	
Control Test	0.081	23:53	
Air Blank	0.000	23:54	
Subject Sample	# 1	23:55	Sequence Aborted
Breath Volume	0.566	23:55	
Air Blank	0.000	23:56	

---

Data Download Date/Time: 06/30/2020 07:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHWOCHO,ALAN,N  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: GANK,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 06/22/2020 00:04:24

---

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.080	23:59
Air Blank	0.000	23:59
Subject Sample # 2	0.209	23:59
Breath Volume	2.656	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 1	0.190	00:02
Breath Volume	1.109	00:02
Air Blank	0.000	00:03
Control Test	0.079	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

---

Data Download Date/Time: 06/30/2020 07:58



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RONGA,DEPUTY,R  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: ANDERSON,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:11

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 06/23/2020 15:38:33

---

Diagnostics Check	OK	15:31
Air Blank	0.000	15:32
Control Test	0.080	15:32
Air Blank	0.000	15:33
Subject Sample # 1	0.124	15:33
Breath Volume	1.320	15:33
Air Blank	0.000	15:34
Air Blank	0.000	15:36
Subject Sample # 2	0.125	15:37
Breath Volume	1.695	15:37
Air Blank	0.000	15:37
Control Test	0.080	15:37
Air Blank	0.000	15:38
Diagnostics Check	OK	15:38

---

Data Download Date/Time: 06/30/2020 07:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDROP,P,W  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: THABIT,DFC  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: USTOJIN

DATE/TIME OF TEST: 06/24/2020 03:03:26

---

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.120	02:58
Breath Volume	2.066	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.119	03:01
Breath Volume	2.613	03:01
Air Blank	0.000	03:02
Control Test	0.080	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

---

Data Download Date/Time: 06/30/2020 07:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZAMORA,DEP,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: WEISSPFENNIG,DEP  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:56

SUBJECT'S LAST NAME: LOEWE

DATE/TIME OF TEST: 06/24/2020 16:24:19

---

Diagnostics Check	OK	16:17
Air Blank	0.000	16:17
Control Test	0.080	16:17
Air Blank	0.000	16:18
Subject Sample # 1	0.288	16:18
Breath Volume	1.820	16:18
Air Blank	0.000	16:19
Air Blank	0.000	16:21
Subject Sample # 2	0.279	16:22
Breath Volume	1.492	16:22
Air Blank	0.000	16:23
Control Test	0.078	16:23
Air Blank	0.000	16:24
Diagnostics Check	OK	16:24

---

Data Download Date/Time: 06/30/2020 07:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDROP,P,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: DFC THABIT,DFC  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ROTH

DATE/TIME OF TEST: 06/27/2020 01:36:36

---

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:32	
Control Test	0.080	01:32	
Air Blank	0.000	01:33	
Subject Sample	# 1	01:35	Subject Test Refused
Breath Volume	0.933	01:35	
Air Blank	0.000	01:35	
Control Test	0.080	01:36	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

---

Data Download Date/Time: 06/30/2020 07:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDROP,P,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: THABIT,DFC  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ROTH

DATE/TIME OF TEST: 06/27/2020 01:48:26

---

Diagnostics Check	OK	01:39
Air Blank	0.000	01:40
Control Test	0.080	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.076	01:42
Breath Volume	1.863	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.074	01:46
Breath Volume	2.195	01:46
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

---

Data Download Date/Time: 06/30/2020 07:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDROP,P,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: SCHWARTZ,DFC  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: CONNELL

DATE/TIME OF TEST: 06/27/2020 23:34:53

---

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:32	
Control Test	0.080	23:32	
Air Blank	0.000	23:33	
Subject Sample	# 1	23:33	Subject Test Refused
Breath Volume	0.000	23:33	
Air Blank	0.000	23:33	
Control Test	0.081	23:34	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

---

Data Download Date/Time: 06/30/2020 07:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDROP,P,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: SMITH,CPL  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 06/28/2020 03:27:34

---

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.080	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.117	03:20
Breath Volume	2.468	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.112	03:26
Breath Volume	1.652	03:26
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

---

Data Download Date/Time: 06/30/2020 07:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ACEBO,ANDRES,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: STEELEY,L  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 32518080A2  
DRY GAS EXPIRY DATE: 02/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: NUZUM

DATE/TIME OF TEST: 06/29/2020 22:48:03

---

Diagnostics Check		OK	22:37
Air Blank		0.000	22:37
Control Test		0.080	22:38
Air Blank		0.000	22:38
Subject Sample	# 1	0.258	22:39
Breath Volume		1.964	22:39
Air Blank		0.000	22:40
Air Blank		0.000	22:42
Subject Sample	# 2	0.233	22:43
Breath Volume		1.828	22:43
Air Blank		0.000	22:43
Air Blank		0.000	22:45
Subject Sample	# 3	0.232	22:46
Breath Volume		1.871	22:46
Air Blank		0.000	22:47
Control Test		0.079	22:47
Air Blank		0.000	22:47
Diagnostics Check		OK	22:48

---

Data Download Date/Time: 06/30/2020 07:59



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUYTON,JOHN,A  
BREATH TEST OPERATOR AGENCY: Gainesville PD  
ARRESTING OFFICER: SHRIGLEY,ROBERT  
ARRESTING OFFICER AGENCY: Gainesville PD  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 06/09/2020 19:11:08

---

Diagnostics Check	OK	19:02	
Air Blank	0.000	19:02	
Control Test	0.080	19:03	
Air Blank	0.000	19:03	
Subject Sample	# 1	19:06	No Sample Provided
Breath Volume	0.000	19:06	
Air Blank	0.000	19:07	
Air Blank	0.000	19:09	
Subject Sample	# 2	19:09	Subject Test Refused
Breath Volume	0.000	19:09	
Air Blank	0.000	19:10	
Control Test	0.080	19:10	
Air Blank	0.000	19:11	
Diagnostics Check	OK	19:11	

---

Data Download Date/Time: 06/30/2020 00:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUYTON,JOHN,A  
BREATH TEST OPERATOR AGENCY: Gainesville PD  
ARRESTING OFFICER: GUYTON,JOHN  
ARRESTING OFFICER AGENCY: Gainesville PD  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: HANKS

DATE/TIME OF TEST: 06/18/2020 02:02:33

---

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.169	01:57
Breath Volume	2.257	01:57
Air Blank	0.000	01:58
Air Blank	0.000	01:59
Subject Sample # 2	0.162	02:00
Breath Volume	1.121	02:00
Air Blank	0.000	02:01
Control Test	0.079	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

---

Data Download Date/Time: 06/30/2020 00:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER,SCOTT,C  
BREATH TEST OPERATOR AGENCY: Levy County SO  
ARRESTING OFFICER: WILLIAMSON,STEPHEN  
ARRESTING OFFICER AGENCY: Levy County SO  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG000606  
DRY GAS EXPIRY DATE: 01/06/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: STRIPP

DATE/TIME OF TEST: 06/05/2020 01:02:17

---

Diagnostics Check	OK	00:59	
Air Blank	0.000	00:59	
Control Test	0.080	01:00	
Air Blank	0.000	01:00	
Subject Sample	# 1	01:00	Subject Test Refused
Breath Volume	0.000	01:00	
Air Blank	0.000	01:01	
Control Test	0.080	01:01	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

---

Data Download Date/Time: 08/01/2020 08:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,MICHAEL,T  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: THOMAS,MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG000606  
DRY GAS EXPIRY DATE: 01/06/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HAVARD

DATE/TIME OF TEST: 06/06/2020 04:32:22

---

Diagnostics Check	OK	04:31	
Air Blank		04:32	RFI Detect
Air Blank		04:32	RFI Detect

---

Data Download Date/Time: 08/01/2020 08:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,MICHAEL,T  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: THOMAS,MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG000606  
DRY GAS EXPIRY DATE: 01/06/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HAVARD

DATE/TIME OF TEST: 06/06/2020 04:35:33

---

Diagnostics Check	OK	04:34	
Air Blank		04:35	RFI Detect
Air Blank	0.000	04:35	

---

Data Download Date/Time: 08/01/2020 08:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS, MICHAEL, T  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: THOMAS, MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG000606  
DRY GAS EXPIRY DATE: 01/06/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HAVARD

DATE/TIME OF TEST: 06/06/2020 04:44:28

---

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.081	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.166	04:39
Breath Volume	1.835	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.163	04:42
Breath Volume	1.585	04:42
Air Blank	0.000	04:43
Control Test	0.078	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

---

Data Download Date/Time: 08/01/2020 08:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLK,AUSTIN,B  
BREATH TEST OPERATOR AGENCY: Levy County SO  
ARRESTING OFFICER: LEE,JARED  
ARRESTING OFFICER AGENCY: Levy County SO  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG000606  
DRY GAS EXPIRY DATE: 01/06/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 06/07/2020 20:12:00

---

Diagnostics Check	OK	20:05
Air Blank	0.000	20:06
Control Test	0.079	20:06
Air Blank	0.000	20:06
Subject Sample # 1	0.161	20:07
Breath Volume	1.949	20:07
Air Blank	0.000	20:07
Air Blank	0.000	20:09
Subject Sample # 2	0.159	20:10
Breath Volume	2.617	20:10
Air Blank	0.000	20:11
Control Test	0.079	20:11
Air Blank	0.000	20:11
Diagnostics Check	OK	20:12

---

Data Download Date/Time: 08/01/2020 08:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLK,AUSTIN,B  
BREATH TEST OPERATOR AGENCY: Levy County SO  
ARRESTING OFFICER: FINKELL,JOHN  
ARRESTING OFFICER AGENCY: Levy County SO  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG000606  
DRY GAS EXPIRY DATE: 01/06/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 06/21/2020 04:09:58

---

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.199	04:05
Breath Volume	2.433	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.192	04:08
Breath Volume	2.804	04:08
Air Blank	0.000	04:09
Control Test	0.079	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

---

Data Download Date/Time: 08/01/2020 08:56



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER,SCOTT,C  
BREATH TEST OPERATOR AGENCY: Levy County SO  
ARRESTING OFFICER: WILLIAMSON,STEPHEN  
ARRESTING OFFICER AGENCY: Levy County SO  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG000606  
DRY GAS EXPIRY DATE: 01/06/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: GERALDO VILLAVICENI

DATE/TIME OF TEST: 06/23/2020 02:16:15

---

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:13	
Control Test	0.080	02:14	
Air Blank	0.000	02:14	
Subject Sample	# 1	02:14	Subject Test Refused
Breath Volume	0.000	02:14	
Air Blank	0.000	02:15	
Control Test	0.080	02:15	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

---

Data Download Date/Time: 08/01/2020 08:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER,SCOTT,C  
BREATH TEST OPERATOR AGENCY: Levy County SO  
ARRESTING OFFICER: DAVIS,KEVIN  
ARRESTING OFFICER AGENCY: Levy County SO  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG000606  
DRY GAS EXPIRY DATE: 01/06/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: HOLDER

DATE/TIME OF TEST: 06/27/2020 02:56:23

---

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.119	02:51
Breath Volume	1.375	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.124	02:54
Breath Volume	2.132	02:54
Air Blank	0.000	02:55
Control Test	0.079	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

---

Data Download Date/Time: 08/01/2020 08:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: EPPS,TIMOTHY  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: CARABALLO

DATE/TIME OF TEST: 06/04/2020 23:08:06

---

Diagnostics Check	OK	23:01
Air Blank	0.000	23:02
Control Test	0.082	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.202	23:03
Breath Volume	2.312	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.209	23:06
Breath Volume	2.585	23:06
Air Blank	0.000	23:07
Control Test	0.082	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

---

Data Download Date/Time: 06/11/2020 10:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,COURTNEY,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KAYFUS,DRU  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 06/05/2020 23:50:37

---

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.082	23:45
Air Blank	0.000	23:45
Subject Sample # 1	0.235	23:46
Breath Volume	1.628	23:46
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.231	23:49
Breath Volume	2.320	23:49
Air Blank	0.000	23:49
Control Test	0.081	23:49
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

---

Data Download Date/Time: 06/11/2020 10:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CHAPMAN,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: FADIGA

DATE/TIME OF TEST: 06/06/2020 01:40:52

---

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.082	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.084	01:36
Breath Volume	2.007	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.084	01:39
Breath Volume	2.429	01:39
Air Blank	0.000	01:39
Control Test	0.082	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

---

Data Download Date/Time: 06/11/2020 10:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: VENTURA,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: LOPEZ CASANOVA

DATE/TIME OF TEST: 06/06/2020 02:58:54

---

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.081	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.131	02:54
Breath Volume	1.726	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:56
Subject Sample # 2	0.139	02:57
Breath Volume	1.894	02:57
Air Blank	0.000	02:57
Control Test	0.081	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

---

Data Download Date/Time: 06/11/2020 10:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ZABALA,JOSE  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: GONZALEZ LOPEZ

DATE/TIME OF TEST: 06/06/2020 20:50:52

---

Diagnostics Check	OK	20:36	
Air Blank	0.000	20:36	
Control Test	0.082	20:37	
Air Blank	0.000	20:37	
Subject Sample	# 1 0.275	20:40	
Breath Volume	1.109	20:40	
Air Blank	0.000	20:41	
Air Blank	0.000	20:42	
Subject Sample	# 2 0.225	20:46	
Breath Volume	1.550	20:46	
Air Blank	0.000	20:47	
Air Blank	0.000	20:48	
Subject Sample	# 3	20:49	RFI Detect
Breath Volume	0.000	20:49	
Air Blank	0.000	20:49	
Control Test	0.074	20:50	Control Outside Tolerance
Air Blank	0.000	20:50	

---

Data Download Date/Time: 06/11/2020 10:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ZABALA,JOSE  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: GONZALEZLOPEZ

DATE/TIME OF TEST: 06/06/2020 20:55:46

---

Diagnostics Check	OK	20:52	
Air Blank	0.000	20:53	
Control Test	0.080	20:53	
Air Blank	0.000	20:54	
Subject Sample	# 1	20:54	Subject Test Refused
Breath Volume	0.000	20:54	
Air Blank	0.000	20:54	
Control Test	0.081	20:55	
Air Blank	0.000	20:55	
Diagnostics Check	OK	20:55	

---

Data Download Date/Time: 06/11/2020 10:23



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELVILLE,KELLY,M  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DRACH,STEFAN  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 10:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: RODRIGUEZSERRANO

DATE/TIME OF TEST: 06/08/2020 11:04:02

---

Diagnostics Check	OK	10:57
Air Blank	0.000	10:57
Control Test	0.082	10:57
Air Blank	0.000	10:58
Subject Sample # 1	0.127	10:59
Breath Volume	2.160	10:59
Air Blank	0.000	10:59
Air Blank	0.000	11:01
Subject Sample # 2	0.119	11:02
Breath Volume	2.222	11:02
Air Blank	0.000	11:03
Control Test	0.082	11:03
Air Blank	0.000	11:03
Diagnostics Check	OK	11:04

---

Data Download Date/Time: 06/11/2020 10:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MOREL, ABIGAIL  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 18:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: RUDDY

DATE/TIME OF TEST: 06/09/2020 19:33:19

---

Diagnostics Check	OK	19:26
Air Blank	0.000	19:27
Control Test	0.081	19:27
Air Blank	0.000	19:27
Subject Sample # 1	0.335	19:28
Breath Volume	1.683	19:28
Air Blank	0.000	19:29
Air Blank	0.000	19:30
Subject Sample # 2	0.327	19:31
Breath Volume	1.128	19:31
Air Blank	0.000	19:32
Control Test	0.080	19:32
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

---

Data Download Date/Time: 06/11/2020 10:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HYATT, MARCUS  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: CHABRIER

DATE/TIME OF TEST: 06/10/2020 02:27:05

---

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.082	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.057	02:22
Breath Volume		1.964	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.056	02:25
Breath Volume		2.097	02:25
Air Blank		0.000	02:26
Control Test		0.081	02:26
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

---

Data Download Date/Time: 06/11/2020 10:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,COURTNEY,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: VENTURA,RONNIE  
ARRESTING OFFICER AGENCY: Edgewood PD  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: GOYZUETA

DATE/TIME OF TEST: 06/10/2020 04:23:25

---

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.082	04:16
Air Blank	0.000	04:17
Subject Sample # 1	0.134	04:18
Breath Volume	1.167	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.136	04:21
Breath Volume	1.238	04:21
Air Blank	0.000	04:22
Control Test	0.081	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

---

Data Download Date/Time: 06/11/2020 10:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,COURTNEY,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RIVERA,VICTOR  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: POLLOCK

DATE/TIME OF TEST: 06/10/2020 22:50:09

---

Diagnostics Check	OK	22:42
Air Blank	0.000	22:43
Control Test	0.082	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.218	22:45
Breath Volume	1.562	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:47
Subject Sample # 2	0.219	22:48
Breath Volume	2.531	22:48
Air Blank	0.000	22:49
Control Test	0.079	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

---

Data Download Date/Time: 06/11/2020 10:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BRIGGS,DERIENNE  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: DEJESUSSOLER

DATE/TIME OF TEST: 06/12/2020 04:50:14

---

Diagnostics Check	OK	04:42
Air Blank	0.000	04:43
Control Test	0.081	04:43
Air Blank	0.000	04:44
Subject Sample # 1	0.163	04:45
Breath Volume	1.253	04:45
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.165	04:48
Breath Volume	1.800	04:48
Air Blank	0.000	04:49
Control Test	0.081	04:49
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

---

Data Download Date/Time: 06/17/2020 13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CASTELLANOS  
PADRON,BRYAN  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MCFARLANE

DATE/TIME OF TEST: 06/12/2020 23:54:04

---

Diagnostics Check		OK	23:47
Air Blank		0.000	23:48
Control Test		0.082	23:48
Air Blank		0.000	23:49
Subject Sample	# 1	0.080	23:49
Breath Volume		3.027	23:49
Air Blank		0.000	23:50
Air Blank		0.000	23:52
Subject Sample	# 2	0.074	23:52
Breath Volume		1.132	23:52
Air Blank		0.000	23:53
Control Test		0.082	23:53
Air Blank		0.000	23:53
Diagnostics Check		OK	23:54

---

Data Download Date/Time: 06/17/2020 13:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MCLEON, NICHOLAS  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 08:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: TELLEZCASTRO

DATE/TIME OF TEST: 06/13/2020 09:55:03

---

Diagnostics Check		OK	09:48
Air Blank		0.000	09:49
Control Test		0.082	09:49
Air Blank		0.000	09:49
Subject Sample	# 1	0.214	09:50
Breath Volume		1.871	09:50
Air Blank		0.000	09:50
Air Blank		0.000	09:52
Subject Sample	# 2	0.199	09:53
Breath Volume		1.378	09:53
Air Blank		0.000	09:54
Control Test		0.081	09:54
Air Blank		0.000	09:54
Diagnostics Check		OK	09:55

---

Data Download Date/Time: 06/17/2020 13:29



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DOMINGUEZ,JAYMES  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: HIDALGO

DATE/TIME OF TEST: 06/14/2020 04:16:59

---

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.082	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.235	04:12
Breath Volume	2.734	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.244	04:15
Breath Volume	1.250	04:15
Air Blank	0.000	04:16
Control Test	0.081	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

---

Data Download Date/Time: 06/17/2020 13:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RUSSO, ANDREW  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 04:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: PEREZ DANZA

DATE/TIME OF TEST: 06/14/2020 05:39:43

---

Diagnostics Check	OK	05:28	
Air Blank	0.000	05:28	
Control Test	0.081	05:29	
Air Blank	0.000	05:29	
Subject Sample	# 1	05:32	No Sample Provided
Breath Volume	0.000	05:32	
Air Blank	0.000	05:33	
Air Blank	0.000	05:35	
Subject Sample	# 2	05:38	No Sample Provided
Breath Volume	0.000	05:38	
Air Blank	0.000	05:38	
Control Test	0.082	05:39	
Air Blank	0.000	05:39	
Diagnostics Check	OK	05:39	

---

Data Download Date/Time: 06/17/2020 13:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HAMILTON, MATTHEW  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 20:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: BAL

DATE/TIME OF TEST: 06/14/2020 21:07:49

---

Diagnostics Check		OK	21:01
Air Blank		0.000	21:02
Control Test		0.082	21:02
Air Blank		0.000	21:02
Subject Sample	# 1	0.215	21:03
Breath Volume		1.906	21:03
Air Blank		0.000	21:04
Air Blank		0.000	21:05
Subject Sample	# 2	0.221	21:06
Breath Volume		2.023	21:06
Air Blank		0.000	21:06
Control Test		0.082	21:07
Air Blank		0.000	21:07
Diagnostics Check		OK	21:07

---

Data Download Date/Time: 06/17/2020 13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SWIFT,JALISA  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: RAUDALESCRUZ

DATE/TIME OF TEST: 06/15/2020 01:08:24

---

Diagnostics Check	OK	01:01
Air Blank	0.000	01:02
Control Test	0.082	01:02
Air Blank	0.000	01:03
Subject Sample # 1	0.156	01:03
Breath Volume	2.156	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:05
Subject Sample # 2	0.145	01:06
Breath Volume	1.210	01:06
Air Blank	0.000	01:07
Control Test	0.082	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

---

Data Download Date/Time: 06/17/2020 13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ROBBINS,NATHAN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 06:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:17

SUBJECT'S LAST NAME: KIAMA

DATE/TIME OF TEST: 06/15/2020 07:49:24

---

Diagnostics Check	OK	07:42
Air Blank	0.000	07:43
Control Test	0.081	07:43
Air Blank	0.000	07:44
Subject Sample # 1	0.158	07:44
Breath Volume	2.078	07:44
Air Blank	0.000	07:45
Air Blank	0.000	07:47
Subject Sample # 2	0.153	07:47
Breath Volume	1.628	07:47
Air Blank	0.000	07:48
Control Test	0.081	07:48
Air Blank	0.000	07:49
Diagnostics Check	OK	07:49

---

Data Download Date/Time: 06/17/2020 13:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELVILLE,KELLY,M  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KERN,ETHAN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:34

SUBJECT'S LAST NAME: TROMBINO

DATE/TIME OF TEST: 06/16/2020 16:21:27

---

Diagnostics Check	OK	16:05	
Air Blank	0.000	16:05	
Control Test	0.082	16:06	
Air Blank	0.000	16:06	
Subject Sample	# 1 0.229	16:09	Volume Not Met
Breath Volume	0.316	16:09	
Air Blank	0.000	16:10	
Air Blank	0.000	16:12	
Subject Sample	# 2 0.274	16:13	
Breath Volume	1.359	16:13	
Air Blank	0.000	16:14	
Air Blank	0.000	16:16	
Subject Sample	# 3 0.283	16:19	Volume Not Met
Breath Volume	0.867	16:19	
Air Blank		16:20	Purge Fail
Air Blank	0.000	16:21	

---

Data Download Date/Time: 06/17/2020 13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELVILLE,KELLY,M  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KERN,ETHAN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:34

SUBJECT'S LAST NAME: TROMBINO

DATE/TIME OF TEST: 06/16/2020 16:25:34

---

Diagnostics Check	OK	16:24	
Air Blank	0.000	16:24	
Control Test	0.072	16:24	Control Outside Tolerance
Air Blank	0.000	16:25	

---

Data Download Date/Time: 06/17/2020 13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELVILLE,KELLY,M  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KERN,ETHAN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:34

SUBJECT'S LAST NAME: TROMBINO

DATE/TIME OF TEST: 06/16/2020 16:41:32

---

Diagnostics Check		OK	16:27	
Air Blank		0.000	16:28	
Control Test		0.077	16:28	
Air Blank		0.000	16:28	
Subject Sample	# 1	0.229	16:32	Volume Not Met
Breath Volume		0.324	16:32	
Air Blank		0.000	16:32	
Air Blank		0.000	16:34	
Subject Sample	# 2	0.277	16:34	
Breath Volume		1.886	16:34	
Air Blank		0.000	16:35	
Air Blank		0.000	16:37	
Subject Sample	# 3	0.258	16:39	
Breath Volume		1.343	16:39	
Air Blank		0.000	16:40	
Control Test		0.081	16:40	
Air Blank		0.000	16:41	
Diagnostics Check		OK	16:41	

---

Data Download Date/Time: 06/17/2020 13:29



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: PRICE,ZACHARY  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BALLARD

DATE/TIME OF TEST: 06/17/2020 00:48:11

---

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.081	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.183	00:43
Breath Volume	3.601	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.187	00:46
Breath Volume	1.148	00:46
Air Blank	0.000	00:47
Control Test	0.079	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

---

Data Download Date/Time: 06/17/2020 13:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANOE,PAMELA,A  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08819080A1  
DRY GAS EXPIRY DATE: 06/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: HUFF

DATE/TIME OF TEST: 06/08/2020 03:33:14

---

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:26	
Control Test	0.080	03:26	
Air Blank	0.000	03:27	
Subject Sample	# 1 0.242	03:28	
Breath Volume	1.632	03:28	
Air Blank	0.000	03:29	
Air Blank	0.000	03:31	
Subject Sample	# 2	03:31	Subject Test Refused
Breath Volume	0.000	03:31	
Air Blank	0.000	03:32	
Control Test	0.080	03:32	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

---

Data Download Date/Time: 06/30/2020 05:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FAHEY,M,  
BREATH TEST OPERATOR AGENCY: Green Cove Springs  
ARRESTING OFFICER: CAMP,B  
ARRESTING OFFICER AGENCY: Green Cove Springs  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08819080A1  
DRY GAS EXPIRY DATE: 06/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: STRANG

DATE/TIME OF TEST: 06/13/2020 03:49:12

---

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:38	
Control Test	0.081	03:38	
Air Blank	0.000	03:39	
Subject Sample	# 1	03:42	No Sample Provided
Breath Volume	0.000	03:42	
Air Blank	0.000	03:42	
Air Blank	0.000	03:44	
Subject Sample	# 2	03:47	Subject Test Refused
Breath Volume	0.863	03:47	
Air Blank	0.000	03:48	
Control Test	0.081	03:48	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

---

Data Download Date/Time: 06/30/2020 05:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FAHEY,M,  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: CAMP,B  
ARRESTING OFFICER AGENCY: Green Cove Springs  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08819080A1  
DRY GAS EXPIRY DATE: 06/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: STRANG

DATE/TIME OF TEST: 06/13/2020 04:20:32

---

Diagnostics Check		OK	04:04	
Air Blank		0.000	04:05	
Control Test		0.080	04:05	
Air Blank		0.000	04:05	
Subject Sample	# 1	0.142	04:08	Volume Not Met
Breath Volume		0.054	04:08	
Air Blank		0.000	04:09	
Air Blank		0.000	04:11	
Subject Sample	# 2	0.148	04:14	
Breath Volume		1.605	04:14	
Air Blank		0.000	04:15	
Air Blank		0.000	04:17	
Subject Sample	# 3	0.153	04:19	
Breath Volume		1.554	04:19	
Air Blank		0.000	04:19	
Control Test		0.080	04:19	
Air Blank		0.000	04:20	
Diagnostics Check		OK	04:20	

---

Data Download Date/Time: 06/30/2020 05:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FAHEY, MARILYN,  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: BAUMGARTNER, N  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08819080A1  
DRY GAS EXPIRY DATE: 06/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: LANEY

DATE/TIME OF TEST: 06/23/2020 04:09:49

---

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:01	
Control Test	0.080	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1	04:05	No Sample Provided
Breath Volume	0.000	04:05	
Air Blank	0.000	04:05	
Air Blank	0.000	04:07	
Subject Sample	# 2	04:08	Subject Test Refused
Breath Volume	0.000	04:08	
Air Blank	0.000	04:08	
Control Test	0.080	04:09	
Air Blank	0.000	04:09	
Diagnostics Check	OK	04:09	

---

Data Download Date/Time: 06/30/2020 05:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FAHEY, MARILYN,  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: PITTMAN, B  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08819080A1  
DRY GAS EXPIRY DATE: 06/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: TARPLEY

DATE/TIME OF TEST: 06/26/2020 23:16:34

---

Diagnostics Check	OK	23:09
Air Blank	0.000	23:09
Control Test	0.080	23:10
Air Blank	0.000	23:10
Subject Sample # 1	0.127	23:12
Breath Volume	1.726	23:12
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.124	23:15
Breath Volume	1.687	23:15
Air Blank	0.000	23:15
Control Test	0.080	23:16
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

---

Data Download Date/Time: 06/30/2020 05:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FAHEY, MARILYN,  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: BAUMGARTNER, N  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08819080A1  
DRY GAS EXPIRY DATE: 06/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: DALE

DATE/TIME OF TEST: 06/28/2020 03:26:23

---

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:18	
Control Test	0.080	03:18	
Air Blank	0.000	03:18	
Subject Sample	# 1	03:22	No Sample Provided
Breath Volume	0.000	03:22	
Air Blank	0.000	03:22	
Air Blank	0.000	03:24	
Subject Sample	# 2	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:25	
Control Test	0.080	03:25	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

---

Data Download Date/Time: 06/30/2020 05:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GELSEY,RICHARD,D  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: GELSEY,RICHARD  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/14/2020 03:24:35

---

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.078	03:18	
Air Blank	0.000	03:19	
Subject Sample	# 1 0.138	03:19	
Breath Volume	2.023	03:19	
Air Blank	0.000	03:20	
Air Blank	0.000	03:22	
Subject Sample	# 2	03:23	Subject Test Refused
Breath Volume	0.000	03:23	
Air Blank	0.000	03:23	
Control Test	0.078	03:23	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

---

Data Download Date/Time: 07/27/2020 18:04



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARRETT,DAVID,S  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: MCKENZIE,RONALD  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: LASSITER

DATE/TIME OF TEST: 06/01/2020 07:30:07

---

Diagnostics Check	OK	07:25	
Air Blank	0.000	07:26	
Control Test	0.081	07:26	
Air Blank	0.000	07:26	
Subject Sample	# 1	07:28	Subject Test Refused
Breath Volume	0.000	07:28	
Air Blank	0.000	07:29	
Control Test	0.081	07:29	
Air Blank	0.000	07:30	
Diagnostics Check	OK	07:30	

---

Data Download Date/Time: 06/03/2020 08:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWRY,JAMIE,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: SMITH,EMILY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: MCLEHAN

DATE/TIME OF TEST: 06/01/2020 21:50:10

---

Diagnostics Check	OK	21:44
Air Blank	0.000	21:44
Control Test	0.081	21:44
Air Blank	0.000	21:45
Subject Sample # 1	0.192	21:45
Breath Volume	2.058	21:45
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample # 2	0.190	21:48
Breath Volume	2.289	21:48
Air Blank	0.000	21:49
Control Test	0.080	21:49
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

---

Data Download Date/Time: 06/03/2020 08:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCKENZIE, RONALD,  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: MCKENZIE, RONALD  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: RADCZENKO

DATE/TIME OF TEST: 06/03/2020 04:37:58

---

Diagnostics Check	OK	04:31
Air Blank	0.000	04:32
Control Test	0.081	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.000	04:33
Breath Volume	1.941	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.000	04:36
Breath Volume	2.417	04:36
Air Blank	0.000	04:37
Control Test	0.081	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

---

Data Download Date/Time: 06/03/2020 08:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUNN,COLLIN,J  
BREATH TEST OPERATOR AGENCY: St. Cloud PD  
ARRESTING OFFICER: HOSTETLER,KEVIN  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: NOWAK

DATE/TIME OF TEST: 06/13/2020 02:19:56

---

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.079	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.184	02:15
Breath Volume	3.132	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:17
Subject Sample # 2	0.188	02:18
Breath Volume	3.281	02:18
Air Blank	0.000	02:19
Control Test	0.078	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

---

Data Download Date/Time: 06/22/2020 14:46