

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 06/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 06/29/2020 01:28:30

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.079	01:21	
Air Blank	0.000	01:21	
Subject Sample	# 1 0.016	01:23	
Breath Volume	2.761	01:23	
Air Blank	0.000	01:23	
Air Blank	0.000	01:25	
Subject Sample	# 2 0.014	01:26	
Breath Volume	1.660	01:26	
Air Blank	0.000	01:27	
Control Test	0.079	01:27	
Air Blank	0.000	01:27	
Diagnostics Check	Fail	01:28	Voltage/Current Test
Air Blank	0.000	01:28	

Data Download Date/Time: 07/28/2020 17:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHALES PEREZ,OCTAVIO,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WARGO,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: CHALES PEREZ

DATE/TIME OF TEST: 06/06/2020 04:40:48

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:40	
Control Test	0.000	04:40	RFI Detect
Air Blank	0.000	04:40	

Data Download Date/Time: 07/14/2020 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ, JONATHAN, W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WARGO, TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: CHALES PEREZ

DATE/TIME OF TEST: 06/06/2020 04:49:27

Diagnostics Check		OK	04:42
Air Blank		0.000	04:43
Control Test		0.079	04:43
Air Blank		0.000	04:44
Subject Sample	# 1	0.249	04:44
Breath Volume		2.332	04:44
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.245	04:47
Breath Volume		3.246	04:47
Air Blank		0.000	04:48
Control Test		0.078	04:48
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 07/14/2020 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YETTER,ERIC,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUBOIS,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 06/07/2020 03:53:05

Diagnostics Check		OK	03:46
Air Blank		0.000	03:47
Control Test		0.079	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.185	03:48
Breath Volume		2.917	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:50
Subject Sample	# 2	0.188	03:51
Breath Volume		2.429	03:51
Air Blank		0.000	03:52
Control Test		0.079	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 07/14/2020 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAK,KYLE,F
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DZUGAN,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/07/2020 06:00:12

Diagnostics Check	OK	05:51
Air Blank	0.000	05:52
Control Test	0.080	05:52
Air Blank	0.000	05:52
Subject Sample # 1	0.060	05:55
Breath Volume	1.289	05:55
Air Blank	0.000	05:56
Air Blank	0.000	05:58
Subject Sample # 2	0.057	05:58
Breath Volume	1.425	05:58
Air Blank	0.000	05:59
Control Test	0.079	05:59
Air Blank	0.000	06:00
Diagnostics Check	OK	06:00

Data Download Date/Time: 07/14/2020 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAK,KYLE,F
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ROUSKEY,IAN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: GREENWAY

DATE/TIME OF TEST: 06/07/2020 16:36:45

Diagnostics Check	OK	16:30
Air Blank	0.000	16:30
Control Test	0.080	16:31
Air Blank	0.000	16:31
Subject Sample # 1	0.000	16:32
Breath Volume	2.617	16:32
Air Blank	0.000	16:32
Air Blank	0.000	16:34
Subject Sample # 2	0.000	16:35
Breath Volume	2.230	16:35
Air Blank	0.000	16:35
Control Test	0.079	16:36
Air Blank	0.000	16:36
Diagnostics Check	OK	16:36

Data Download Date/Time: 07/14/2020 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLINGAME,NATHALIE,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BLEDSOE,LUTHER
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: CIULA

DATE/TIME OF TEST: 06/07/2020 20:46:54

Diagnostics Check	OK	20:39
Air Blank	0.000	20:39
Control Test	0.079	20:40
Air Blank	0.000	20:40
Subject Sample # 1	0.160	20:42
Breath Volume	2.566	20:42
Air Blank	0.000	20:42
Air Blank	0.000	20:44
Subject Sample # 2	0.151	20:45
Breath Volume	1.675	20:45
Air Blank	0.000	20:45
Control Test	0.078	20:46
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

Data Download Date/Time: 07/14/2020 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REED,CHAD
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: CARDOZO ESCALERA

DATE/TIME OF TEST: 06/03/2020 00:13:06

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.078	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.189	00:08
Breath Volume	1.542	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.200	00:11
Breath Volume	1.171	00:11
Air Blank	0.000	00:12
Control Test	0.077	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 06/08/2020 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/03/2020 00:38:05

Diagnostics Check		OK	00:30
Air Blank		0.000	00:30
Control Test		0.077	00:31
Air Blank		0.000	00:31
Subject Sample	# 1	0.215	00:32
Breath Volume		1.820	00:32
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.208	00:36
Breath Volume		1.703	00:36
Air Blank		0.000	00:37
Control Test		0.077	00:37
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 06/08/2020 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SENTER,ZACHARY,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: SENTER,ZACHARY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 06/04/2020 08:19:31

Diagnostics Check	OK	08:12
Air Blank	0.000	08:13
Control Test	0.077	08:13
Air Blank	0.000	08:13
Subject Sample # 1	0.127	08:14
Breath Volume	4.058	08:14
Air Blank	0.000	08:15
Air Blank	0.000	08:17
Subject Sample # 2	0.128	08:17
Breath Volume	4.253	08:17
Air Blank	0.000	08:18
Control Test	0.076	08:18
Air Blank	0.000	08:19
Diagnostics Check	OK	08:19

Data Download Date/Time: 06/08/2020 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SENTER,ZACHARY,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: SENTER,ZACHARY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: DORNER

DATE/TIME OF TEST: 06/05/2020 14:50:53

Diagnostics Check	OK	14:43
Air Blank	0.000	14:43
Control Test	0.077	14:43
Air Blank	0.000	14:44
Subject Sample # 1	0.157	14:46
Breath Volume	2.019	14:46
Air Blank	0.000	14:46
Air Blank	0.000	14:48
Subject Sample # 2	0.159	14:49
Breath Volume	2.242	14:49
Air Blank	0.000	14:49
Control Test	0.076	14:50
Air Blank	0.000	14:50
Diagnostics Check	OK	14:50

Data Download Date/Time: 06/08/2020 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: MILLS/GARCIA,TONY/MARVIN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: NOBLE

DATE/TIME OF TEST: 06/05/2020 23:39:44

Diagnostics Check	OK	23:32
Air Blank	0.000	23:33
Control Test	0.078	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.000	23:34
Breath Volume	3.054	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.000	23:38
Breath Volume	1.664	23:38
Air Blank	0.000	23:38
Control Test	0.077	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 06/08/2020 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REED,CHAD
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: IGLESIAS JR

DATE/TIME OF TEST: 06/07/2020 00:10:41

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.077	00:03
Air Blank	0.000	00:03
Subject Sample # 1	0.182	00:05
Breath Volume	3.621	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.192	00:08
Breath Volume	2.421	00:08
Air Blank	0.000	00:09
Control Test	0.077	00:10
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 06/08/2020 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SENTER,ZACHARY,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: SENTER,ZACHARY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: PIPER

DATE/TIME OF TEST: 06/13/2020 17:32:54

Diagnostics Check	OK	17:25
Air Blank	0.000	17:26
Control Test	0.077	17:26
Air Blank	0.000	17:27
Subject Sample # 1	0.000	17:28
Breath Volume	2.007	17:28
Air Blank	0.000	17:28
Air Blank	0.000	17:30
Subject Sample # 2	0.000	17:31
Breath Volume	1.703	17:31
Air Blank	0.000	17:31
Control Test	0.077	17:32
Air Blank	0.000	17:32
Diagnostics Check	OK	17:32

Data Download Date/Time: 07/29/2020 23:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,OFFICER,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRUNEAU,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: BREWER

DATE/TIME OF TEST: 06/14/2020 00:12:03

Diagnostics Check		OK	00:05
Air Blank		0.000	00:06
Control Test		0.076	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.209	00:07
Breath Volume		1.917	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:09
Subject Sample	# 2	0.202	00:10
Breath Volume		3.378	00:10
Air Blank		0.000	00:11
Control Test		0.076	00:11
Air Blank		0.000	00:11
Diagnostics Check		OK	00:12

Data Download Date/Time: 07/29/2020 23:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SENTER,ZACHARY,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: SENTER,ZACHARY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 06/23/2020 14:40:11

Diagnostics Check		OK	14:32
Air Blank		0.000	14:32
Control Test		0.076	14:33
Air Blank		0.000	14:33
Subject Sample	# 1	0.274	14:34
Breath Volume		1.972	14:34
Air Blank		0.000	14:35
Air Blank		0.000	14:37
Subject Sample	# 2	0.280	14:38
Breath Volume		2.839	14:38
Air Blank		0.000	14:39
Control Test		0.076	14:39
Air Blank		0.000	14:40
Diagnostics Check		OK	14:40

Data Download Date/Time: 07/29/2020 23:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 06/24/2020 23:35:35

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.078	23:29
Air Blank	0.000	23:29
Subject Sample # 1	0.011	23:30
Breath Volume	1.929	23:30
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.010	23:34
Breath Volume	1.523	23:34
Air Blank	0.000	23:34
Control Test	0.078	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 07/29/2020 23:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,OFFICER,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRUNEAU,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: CHAVIANO

DATE/TIME OF TEST: 06/25/2020 01:08:57

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.076	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.133	01:04
Breath Volume	2.531	01:04
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.131	01:07
Breath Volume	3.246	01:07
Air Blank	0.000	01:08
Control Test	0.077	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 07/29/2020 23:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SENTER,ZACHARY,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: SENTER,ZACHARY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: GASSIOTT

DATE/TIME OF TEST: 06/25/2020 15:17:54

Diagnostics Check	OK	15:11
Air Blank	0.000	15:11
Control Test	0.078	15:11
Air Blank	0.000	15:12
Subject Sample # 1	0.273	15:12
Breath Volume	1.554	15:12
Air Blank	0.000	15:13
Air Blank	0.000	15:15
Subject Sample # 2	0.265	15:16
Breath Volume	1.335	15:16
Air Blank	0.000	15:16
Control Test	0.076	15:17
Air Blank	0.000	15:17
Diagnostics Check	OK	15:17

Data Download Date/Time: 07/29/2020 23:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: DEVINE

DATE/TIME OF TEST: 06/26/2020 02:01:48

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.078	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.042	01:56
Breath Volume	2.003	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.042	02:00
Breath Volume	2.226	02:00
Air Blank	0.000	02:00
Control Test	0.079	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 07/29/2020 23:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: RIVERA SANCHEZ

DATE/TIME OF TEST: 06/27/2020 22:24:52

Diagnostics Check	OK	22:17
Air Blank	0.000	22:18
Control Test	0.078	22:18
Air Blank	0.000	22:18
Subject Sample # 1	0.000	22:19
Breath Volume	3.039	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.000	22:23
Breath Volume	2.855	22:23
Air Blank	0.000	22:23
Control Test	0.077	22:24
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 07/29/2020 23:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

INSTRUMENT REGISTERED TO: NAS JACKSONVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,WILLIAM,A
BREATH TEST OPERATOR AGENCY: NAS Jacksonville PD
ARRESTING OFFICER: JOHNSON,JOSHUA
ARRESTING OFFICER AGENCY: NAS Jacksonville PD
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: PAIGE

DATE/TIME OF TEST: 06/03/2020 03:38:20

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.080	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.232	03:33
Breath Volume	1.914	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.224	03:36
Breath Volume	2.050	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 06/16/2020 08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNOZ,A,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: ALVARADO,B
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: GALVEZ

DATE/TIME OF TEST: 06/05/2020 22:21:49

Diagnostics Check	OK	22:11
Air Blank	0.000	22:12
Control Test	0.080	22:12
Air Blank	0.000	22:13
Subject Sample # 1	0.144	22:14
Breath Volume	2.113	22:14
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.139	22:19
Breath Volume	1.941	22:19
Air Blank	0.000	22:20
Control Test	0.078	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 06/24/2020 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DAWKINS,JUNIOR,A

BREATH TEST OPERATOR AGENCY: Hialeah PD

ARRESTING OFFICER: DAWKINS,JUNIOR

ARRESTING OFFICER AGENCY: Hialeah PD

LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1224982

DRY GAS EXPIRY DATE: 10/31/2022

VIOLATION CODE: DUI

ARREST TIME: 23:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: CASTANO

DATE/TIME OF TEST: 06/06/2020 00:47:57

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:42	
Control Test	0.080	00:43	
Air Blank	0.000	00:43	
Subject Sample	# 1	00:46	Subject Test Refused
Breath Volume	0.000	00:46	
Air Blank	0.000	00:47	
Control Test	0.081	00:47	
Air Blank	0.000	00:47	
Diagnostics Check	OK	00:47	

Data Download Date/Time: 06/24/2020 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNOZ,A,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: ECHEVARRIA,L
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: BLANCO

DATE/TIME OF TEST: 06/14/2020 18:25:43

Diagnostics Check	OK	18:22	
Air Blank	0.000	18:23	
Control Test	0.081	18:23	
Air Blank	0.000	18:24	
Subject Sample	# 1	18:24	Subject Test Refused
Breath Volume	0.000	18:24	
Air Blank	0.000	18:24	
Control Test	0.081	18:25	
Air Blank	0.000	18:25	
Diagnostics Check	OK	18:25	

Data Download Date/Time: 06/24/2020 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: WITRAK

DATE/TIME OF TEST: 06/13/2020 23:37:00

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.080	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.179	23:32
Breath Volume	4.343	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:34
Subject Sample # 2	0.177	23:35
Breath Volume	4.691	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:37

Data Download Date/Time: 06/28/2020 15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KROUSE,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 06/04/2020 16:17:24

Diagnostics Check	OK	16:10
Air Blank	0.000	16:11
Control Test	0.080	16:11
Air Blank	0.000	16:12
Subject Sample # 1	0.000	16:12
Breath Volume	1.519	16:12
Air Blank	0.000	16:13
Air Blank	0.000	16:15
Subject Sample # 2	0.000	16:16
Breath Volume	1.437	16:16
Air Blank	0.000	16:16
Control Test	0.080	16:16
Air Blank	0.000	16:17
Diagnostics Check	OK	16:17

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 06/06/2020 04:14:22

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.221	04:09
Breath Volume	3.664	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.225	04:12
Breath Volume	2.945	04:12
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: MOULD

DATE/TIME OF TEST: 06/07/2020 02:07:55

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.081	02:01
Air Blank	0.000	02:02
Subject Sample # 1	0.070	02:03
Breath Volume	3.027	02:03
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.065	02:06
Breath Volume	2.722	02:06
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: GARCIA SOLAR

DATE/TIME OF TEST: 06/07/2020 03:56:47

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.081	03:50
Air Blank		0.000	03:50
Subject Sample	# 1	0.151	03:52
Breath Volume		2.273	03:52
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.165	03:55
Breath Volume		1.312	03:55
Air Blank		0.000	03:55
Control Test		0.079	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PALFY, R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: ROSALES PACHECO

DATE/TIME OF TEST: 06/07/2020 21:33:28

Diagnostics Check		OK	21:19	
Air Blank		0.000	21:19	
Control Test		0.081	21:20	
Air Blank		0.000	21:20	
Subject Sample	# 1	0.192	21:23	Volume Not Met
Breath Volume		0.972	21:23	
Air Blank		0.000	21:24	
Air Blank		0.000	21:26	
Subject Sample	# 2	0.245	21:29	
Breath Volume		2.761	21:29	
Air Blank		0.000	21:29	
Air Blank		0.000	21:31	
Subject Sample	# 3	0.249	21:31	
Breath Volume		1.351	21:31	
Air Blank		0.000	21:32	
Control Test		0.079	21:32	
Air Blank		0.000	21:33	
Diagnostics Check		OK	21:33	

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 06/07/2020 23:24:11

Diagnostics Check	OK	23:19	
Air Blank	0.000	23:19	
Control Test	0.080	23:19	
Air Blank	0.000	23:20	
Subject Sample	# 1	23:22	Subject Test Refused
Breath Volume	0.000	23:22	
Air Blank	0.000	23:23	
Control Test	0.080	23:23	
Air Blank	0.000	23:24	
Diagnostics Check	OK	23:24	

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: CHANLOPEZ

DATE/TIME OF TEST: 06/10/2020 02:44:34

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.081	02:39
Air Blank		0.000	02:39
Subject Sample	# 1	0.085	02:40
Breath Volume		2.484	02:40
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.086	02:43
Breath Volume		2.164	02:43
Air Blank		0.000	02:43
Control Test		0.080	02:43
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: HERNANDEZ TORRES

DATE/TIME OF TEST: 06/11/2020 02:53:33

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.081	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.145	02:48
Breath Volume	3.265	02:48
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.141	02:52
Breath Volume	1.152	02:52
Air Blank	0.000	02:52
Control Test	0.079	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: GUILDER

DATE/TIME OF TEST: 06/12/2020 03:32:33

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.080	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.180	03:28
Breath Volume	1.449	03:28
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.188	03:31
Breath Volume	2.144	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 06/12/2020 03:49:01

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:41	
Control Test	0.080	03:42	
Air Blank	0.000	03:42	
Subject Sample	# 1 0.209	03:43	
Breath Volume	1.597	03:43	
Air Blank	0.000	03:44	
Air Blank	0.000	03:46	
Subject Sample	# 2	03:47	Subject Test Refused
Breath Volume	0.000	03:47	
Air Blank	0.000	03:48	
Control Test	0.079	03:48	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: WEIR

DATE/TIME OF TEST: 06/13/2020 01:19:04

Diagnostics Check		OK	01:13
Air Blank		0.000	01:13
Control Test		0.081	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.000	01:14
Breath Volume		2.472	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.000	01:17
Breath Volume		2.472	01:17
Air Blank		0.000	01:18
Control Test		0.081	01:18
Air Blank		0.000	01:18
Diagnostics Check		OK	01:19

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: VILORIO

DATE/TIME OF TEST: 06/13/2020 03:37:30

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.141	03:32
Breath Volume	1.917	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.133	03:36
Breath Volume	1.101	03:36
Air Blank	0.000	03:36
Control Test	0.081	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: IRVING

DATE/TIME OF TEST: 06/14/2020 00:27:10

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.080	00:21
Air Blank	0.000	00:22
Subject Sample # 1	0.139	00:22
Breath Volume	2.488	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.138	00:25
Breath Volume	1.960	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: BAUTISTA ZEFERINO

DATE/TIME OF TEST: 06/14/2020 04:00:28

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.165	03:55
Breath Volume	1.285	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:57
Subject Sample # 2	0.163	03:58
Breath Volume	2.042	03:58
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VINKEMULDER,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:48

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/14/2020 11:17:14

Diagnostics Check		OK	11:10
Air Blank		0.000	11:10
Control Test		0.080	11:10
Air Blank		0.000	11:11
Subject Sample	# 1	0.255	11:12
Breath Volume		1.257	11:12
Air Blank		0.000	11:13
Air Blank		0.000	11:15
Subject Sample	# 2	0.263	11:15
Breath Volume		1.910	11:15
Air Blank		0.000	11:16
Control Test		0.080	11:16
Air Blank		0.000	11:17
Diagnostics Check		OK	11:17

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 06/16/2020 04:23:16

Diagnostics Check		OK	04:12
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:13
Subject Sample	# 1	0.202	04:15
Breath Volume		1.128	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:17
Subject Sample	# 2	0.234	04:18
Breath Volume		1.425	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 3	0.231	04:21
Breath Volume		2.792	04:21
Air Blank		0.000	04:22
Control Test		0.080	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: REID, W
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/17/2020 00:54:27

Diagnostics Check		OK	00:48
Air Blank		0.000	00:49
Control Test		0.080	00:49
Air Blank		0.000	00:49
Subject Sample	# 1	0.182	00:50
Breath Volume		1.296	00:50
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.177	00:52
Breath Volume		1.648	00:52
Air Blank		0.000	00:53
Control Test		0.080	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: MARTINEZ GONZALEZ

DATE/TIME OF TEST: 06/18/2020 00:00:44

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.080	23:55
Air Blank	0.000	23:55
Subject Sample # 1	0.123	23:56
Breath Volume	1.398	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.124	23:59
Breath Volume	1.671	23:59
Air Blank	0.000	23:59
Control Test	0.079	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: AMBROISE

DATE/TIME OF TEST: 06/19/2020 04:26:24

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.081	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.177	04:21
Breath Volume	1.292	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.174	04:24
Breath Volume	1.660	04:24
Air Blank	0.000	04:25
Control Test	0.081	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOLSTENHOLM,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: EVELAND JR

DATE/TIME OF TEST: 06/19/2020 18:58:12

Diagnostics Check	OK	18:51
Air Blank	0.000	18:52
Control Test	0.081	18:52
Air Blank	0.000	18:52
Subject Sample # 1	0.000	18:53
Breath Volume	1.554	18:53
Air Blank	0.000	18:54
Air Blank	0.000	18:56
Subject Sample # 2	0.000	18:56
Breath Volume	1.898	18:56
Air Blank	0.000	18:57
Control Test	0.081	18:57
Air Blank	0.000	18:58
Diagnostics Check	OK	18:58

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 06/20/2020 03:08:28

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.082	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.297	03:03
Breath Volume	2.363	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.299	03:06
Breath Volume	2.074	03:06
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 06/21/2020 01:34:45

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.081	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.147	01:30
Breath Volume	2.843	01:30
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.144	01:33
Breath Volume	2.875	01:33
Air Blank	0.000	01:33
Control Test	0.080	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 06/21/2020 03:29:52

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.081	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.100	03:25
Breath Volume	3.011	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:27
Subject Sample # 2	0.086	03:28
Breath Volume	1.601	03:28
Air Blank	0.000	03:28
Control Test	0.081	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 07/08/2020 18:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 06/21/2020 04:20:10

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.080	04:14
Air Blank		0.000	04:14
Subject Sample	# 1	0.073	04:15
Breath Volume		2.968	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.071	04:18
Breath Volume		2.785	04:18
Air Blank		0.000	04:19
Control Test		0.081	04:19
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 07/08/2020 18:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: ADKINS

DATE/TIME OF TEST: 06/24/2020 02:00:10

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:49	
Control Test	0.082	01:49	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:53	No Sample Provided
Breath Volume	0.000	01:53	
Air Blank	0.000	01:53	
Air Blank	0.000	01:55	
Subject Sample	# 2	01:58	No Sample Provided
Breath Volume	0.000	01:58	
Air Blank	0.000	01:59	
Control Test	0.081	01:59	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

Data Download Date/Time: 07/08/2020 18:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/26/2020 21:48:35

Diagnostics Check		OK	21:34	
Air Blank		0.000	21:35	
Control Test		0.080	21:35	
Air Blank		0.000	21:36	
Subject Sample	# 1	0.202	21:39	Volume Not Met
Breath Volume		0.652	21:39	
Air Blank		0.000	21:40	
Air Blank		0.000	21:41	
Subject Sample	# 2	0.240	21:43	
Breath Volume		2.199	21:43	
Air Blank		0.000	21:44	
Air Blank		0.000	21:46	
Subject Sample	# 3	0.233	21:47	
Breath Volume		1.882	21:47	
Air Blank		0.000	21:47	
Control Test		0.079	21:47	
Air Blank		0.000	21:48	
Diagnostics Check		OK	21:48	

Data Download Date/Time: 07/08/2020 18:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: PEREZ JR

DATE/TIME OF TEST: 06/28/2020 02:53:10

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:50	
Control Test	0.080	02:50	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:52	
Control Test	0.080	02:52	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 07/08/2020 18:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 06/28/2020 04:27:38

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.081	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.062	04:22
Breath Volume	2.722	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.062	04:26
Breath Volume	2.867	04:26
Air Blank	0.000	04:26
Control Test	0.081	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 07/08/2020 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: HILLEY

DATE/TIME OF TEST: 06/28/2020 22:18:13

Diagnostics Check		OK	22:11
Air Blank		0.000	22:12
Control Test		0.081	22:12
Air Blank		0.000	22:13
Subject Sample	# 1	0.109	22:13
Breath Volume		2.257	22:13
Air Blank		0.000	22:14
Air Blank		0.000	22:16
Subject Sample	# 2	0.109	22:16
Breath Volume		2.253	22:16
Air Blank		0.000	22:17
Control Test		0.080	22:17
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 07/08/2020 18:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: DEVRIES

DATE/TIME OF TEST: 06/29/2020 00:02:12

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.080	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.168	23:57
Breath Volume	1.667	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.154	00:00
Breath Volume	1.550	00:00
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 07/08/2020 18:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHOBREY,SPENCER,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WHOBREY,SPENCER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 06/07/2020 23:41:35

Diagnostics Check	OK	23:40	
Air Blank		23:41	RFI Detect
Air Blank	0.000	23:41	

Data Download Date/Time: 06/22/2020 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHOBREY,SPENCER,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WHOBREY,SPENCER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: TORRES-PEREZ

DATE/TIME OF TEST: 06/11/2020 01:06:25

Diagnostics Check	OK	00:58
Air Blank	0.000	00:59
Control Test	0.079	00:59
Air Blank	0.000	00:59
Subject Sample # 1	0.183	01:01
Breath Volume	2.769	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.177	01:04
Breath Volume	2.843	01:04
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 06/22/2020 15:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHOBREY,SPENCER,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WHOBREY,SPENCER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: MENOR

DATE/TIME OF TEST: 06/20/2020 03:54:55

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.196	03:50
Breath Volume	2.105	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.200	03:53
Breath Volume	2.101	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 06/22/2020 15:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: CRUZ-PAGAN

DATE/TIME OF TEST: 06/19/2020 01:03:13

Diagnostics Check		OK	00:56
Air Blank		0.000	00:57
Control Test		0.079	00:57
Air Blank		0.000	00:57
Subject Sample	# 1	0.203	00:58
Breath Volume		1.656	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.192	01:01
Breath Volume		1.632	01:01
Air Blank		0.000	01:02
Control Test		0.080	01:02
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 06/22/2020 15:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNHAM,J,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BURNHAM,J
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: CUMMINS

DATE/TIME OF TEST: 06/01/2020 04:08:34

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.080	04:06	
Air Blank	0.000	04:06	
Subject Sample	# 1	04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:07	
Control Test	0.079	04:07	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 06/03/2020 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 06/03/2020 05:33:28

Diagnostics Check	OK	05:23
Air Blank	0.000	05:24
Control Test	0.080	05:24
Air Blank	0.000	05:25
Subject Sample # 1	0.062	05:27
Breath Volume	2.460	05:27
Air Blank	0.000	05:27
Air Blank	0.000	05:29
Subject Sample # 2	0.059	05:31
Breath Volume	1.656	05:31
Air Blank	0.000	05:32
Control Test	0.079	05:32
Air Blank	0.000	05:33
Diagnostics Check	OK	05:33

Data Download Date/Time: 06/03/2020 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,MASON,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BURGESS,MASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: COPLIN

DATE/TIME OF TEST: 06/05/2020 03:48:06

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:37	
Control Test	0.080	03:37	
Air Blank	0.000	03:38	
Subject Sample	# 1	03:41	No Sample Provided
Breath Volume	0.000	03:41	
Air Blank	0.000	03:41	
Air Blank	0.000	03:43	
Subject Sample	# 2	03:46	No Sample Provided
Breath Volume	0.000	03:46	
Air Blank	0.000	03:47	
Control Test	0.080	03:47	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 07/02/2020 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQUEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 06/05/2020 22:18:37

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:16	
Control Test	0.079	22:16	
Air Blank	0.000	22:16	
Subject Sample	# 1	22:17	Subject Test Refused
Breath Volume	0.000	22:17	
Air Blank	0.000	22:17	
Control Test	0.079	22:18	
Air Blank	0.000	22:18	
Diagnostics Check	OK	22:18	

Data Download Date/Time: 07/02/2020 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: PAINTER

DATE/TIME OF TEST: 06/06/2020 01:52:25

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:49	
Control Test	0.079	01:49	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:51	
Control Test	0.080	01:51	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 07/02/2020 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQUEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: SALAZAR DIAZ

DATE/TIME OF TEST: 06/06/2020 02:24:16

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.079	02:22	
Air Blank	0.000	02:22	
Subject Sample	# 1	02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:23	
Control Test	0.080	02:23	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

Data Download Date/Time: 07/02/2020 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDE, MATTHEW, B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILDE, MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: LETOURNEAU

DATE/TIME OF TEST: 06/06/2020 03:31:15

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:28	
Control Test	0.080	03:29	
Air Blank	0.000	03:29	
Subject Sample	# 1	03:29	Subject Test Refused
Breath Volume	0.000	03:29	
Air Blank	0.000	03:30	
Control Test	0.079	03:30	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 07/02/2020 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK, TAVISH, M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK, TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 06/06/2020 03:47:42

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.098	03:43
Breath Volume	4.527	03:43
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.097	03:46
Breath Volume	5.113	03:46
Air Blank	0.000	03:46
Control Test	0.078	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 07/02/2020 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ESNEUL,MICHAEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/06/2020 04:23:10

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:16
Subject Sample # 1	0.192	04:18
Breath Volume	3.574	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:20
Subject Sample # 2	0.188	04:21
Breath Volume	2.757	04:21
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 07/02/2020 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RHODES,PHILLIP
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: SEGOVIANO GUZMAN

DATE/TIME OF TEST: 06/06/2020 04:54:06

Diagnostics Check	OK	04:47
Air Blank	0.000	04:47
Control Test	0.079	04:47
Air Blank	0.000	04:48
Subject Sample # 1	0.137	04:49
Breath Volume	1.964	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.138	04:52
Breath Volume	2.417	04:52
Air Blank	0.000	04:53
Control Test	0.079	04:53
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 07/02/2020 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK, TAVISH, M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK, TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MATOS

DATE/TIME OF TEST: 06/08/2020 03:43:28

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:41	
Control Test	0.080	03:41	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:42	
Control Test	0.080	03:42	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 07/02/2020 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GONZALES,ANTHONY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: BRUCE

DATE/TIME OF TEST: 06/08/2020 05:27:38

Diagnostics Check	OK	05:19	
Air Blank	0.000	05:19	
Control Test	0.079	05:19	
Air Blank	0.000	05:20	
Subject Sample	# 1	05:23	No Sample Provided
Breath Volume	0.000	05:23	
Air Blank	0.000	05:23	
Air Blank	0.000	05:25	
Subject Sample	# 2	05:26	Subject Test Refused
Breath Volume	0.000	05:26	
Air Blank	0.000	05:26	
Control Test	0.080	05:27	
Air Blank	0.000	05:27	
Diagnostics Check	OK	05:27	

Data Download Date/Time: 07/02/2020 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 16:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:52

SUBJECT'S LAST NAME: HERMANSEN

DATE/TIME OF TEST: 06/10/2020 17:19:07

Diagnostics Check	OK	17:12
Air Blank	0.000	17:13
Control Test	0.080	17:13
Air Blank	0.000	17:14
Subject Sample # 1	0.220	17:14
Breath Volume	1.617	17:14
Air Blank	0.000	17:15
Air Blank	0.000	17:17
Subject Sample # 2	0.232	17:17
Breath Volume	2.027	17:17
Air Blank	0.000	17:18
Control Test	0.079	17:18
Air Blank	0.000	17:19
Diagnostics Check	OK	17:19

Data Download Date/Time: 07/02/2020 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 06/10/2020 23:59:22

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.056	23:54
Breath Volume	1.789	23:54
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.055	23:57
Breath Volume	1.824	23:57
Air Blank	0.000	23:58
Control Test	0.079	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 07/02/2020 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK,TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: RALSTON

DATE/TIME OF TEST: 06/11/2020 01:38:08

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:35	
Control Test	0.079	01:35	
Air Blank	0.000	01:36	
Subject Sample	# 1	01:36	Subject Test Refused
Breath Volume	0.000	01:36	
Air Blank	0.000	01:37	
Control Test	0.080	01:37	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK, TAVISH, M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK, TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 06/12/2020 02:56:54

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.081	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.000	02:52
Breath Volume	1.730	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.000	02:55
Breath Volume	1.425	02:55
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,LANDON,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WALKER,LANDON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: BREEDLOVE

DATE/TIME OF TEST: 06/12/2020 19:36:38

Diagnostics Check	OK	19:30
Air Blank	0.000	19:30
Control Test	0.080	19:30
Air Blank	0.000	19:31
Subject Sample # 1	0.000	19:32
Breath Volume	2.609	19:32
Air Blank	0.000	19:32
Air Blank	0.000	19:34
Subject Sample # 2	0.000	19:35
Breath Volume	2.078	19:35
Air Blank	0.000	19:35
Control Test	0.079	19:36
Air Blank	0.000	19:36
Diagnostics Check	OK	19:36

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNHAM,J,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BURNHAM,J
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: NATION

DATE/TIME OF TEST: 06/12/2020 21:47:37

Diagnostics Check	OK	21:41
Air Blank	0.000	21:42
Control Test	0.080	21:42
Air Blank	0.000	21:42
Subject Sample # 1	0.000	21:43
Breath Volume	1.792	21:43
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.000	21:46
Breath Volume	1.785	21:46
Air Blank	0.000	21:46
Control Test	0.080	21:47
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNHAM,J,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GLENECK,JORDAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 06/13/2020 05:23:31

Diagnostics Check		OK	05:14
Air Blank		0.000	05:15
Control Test		0.079	05:15
Air Blank		0.000	05:16
Subject Sample	# 1	0.169	05:17
Breath Volume		1.105	05:17
Air Blank		0.000	05:17
Air Blank		0.000	05:19
Subject Sample	# 2	0.170	05:21
Breath Volume		1.195	05:21
Air Blank		0.000	05:22
Control Test		0.077	05:22
Air Blank		0.000	05:23
Diagnostics Check		OK	05:23

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MILLER,RAYMOND
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/14/2020 04:39:36

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:31	
Control Test	0.079	04:31	
Air Blank	0.000	04:32	
Subject Sample	# 1	04:35	No Sample Provided
Breath Volume	0.000	04:35	
Air Blank	0.000	04:36	
Air Blank	0.000	04:38	
Subject Sample	# 2	04:38	Subject Test Refused
Breath Volume	0.000	04:38	
Air Blank	0.000	04:38	
Control Test	0.080	04:39	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:39	

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEENEY,ERIC,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FEENEY,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: NICHOLAS

DATE/TIME OF TEST: 06/14/2020 05:17:36

Diagnostics Check	OK	05:14	
Air Blank	0.000	05:15	
Control Test	0.079	05:15	
Air Blank	0.000	05:16	
Subject Sample	# 1	05:16	Subject Test Refused
Breath Volume	0.000	05:16	
Air Blank	0.000	05:16	
Control Test	0.080	05:17	
Air Blank	0.000	05:17	
Diagnostics Check	OK	05:17	

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,ROBERT,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIS,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 05:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:21

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 06/14/2020 06:50:02

Diagnostics Check	OK	06:47	
Air Blank	0.000	06:47	
Control Test	0.080	06:47	
Air Blank	0.000	06:48	
Subject Sample	# 1	06:48	Subject Test Refused
Breath Volume	0.000	06:48	
Air Blank	0.000	06:49	
Control Test	0.080	06:49	
Air Blank	0.000	06:49	
Diagnostics Check	OK	06:50	

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGUSA,NICHOLE,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RAGUSA,NICHOLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 06/15/2020 03:37:40

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.079	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.296	03:32
Breath Volume	1.179	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:34
Subject Sample # 2	0.293	03:35
Breath Volume	3.019	03:35
Air Blank	0.000	03:36
Control Test	0.078	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNHAM,J,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: JACKSON,TAYLOR
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 06/15/2020 06:07:45

Diagnostics Check	OK	06:00
Air Blank	0.000	06:01
Control Test	0.079	06:01
Air Blank	0.000	06:01
Subject Sample # 1	0.157	06:02
Breath Volume	1.406	06:02
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample # 2	0.155	06:06
Breath Volume	1.199	06:06
Air Blank	0.000	06:06
Control Test	0.077	06:07
Air Blank	0.000	06:07
Diagnostics Check	OK	06:07

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,SHELBY,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CARROLL,SHELBY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: OLIVA

DATE/TIME OF TEST: 06/16/2020 01:09:51

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.080	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.061	01:05
Breath Volume	1.894	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:07
Subject Sample # 2	0.062	01:08
Breath Volume	1.707	01:08
Air Blank	0.000	01:08
Control Test	0.080	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQUEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: HUNTINGTON

DATE/TIME OF TEST: 06/16/2020 03:09:20

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.080	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.151	03:04
Breath Volume	1.546	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.152	03:07
Breath Volume	1.402	03:07
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,MAREIKE,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK,TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 06/16/2020 05:46:34

Diagnostics Check		OK	05:40
Air Blank		0.000	05:40
Control Test		0.080	05:40
Air Blank		0.000	05:41
Subject Sample	# 1	0.134	05:42
Breath Volume		4.023	05:42
Air Blank		0.000	05:42
Air Blank		0.000	05:44
Subject Sample	# 2	0.124	05:45
Breath Volume		4.601	05:45
Air Blank		0.000	05:45
Control Test		0.080	05:45
Air Blank		0.000	05:46
Diagnostics Check		OK	05:46

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON, RAY, C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LAPEE, WYATT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: GILLUM

DATE/TIME OF TEST: 06/17/2020 21:39:39

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.080	21:33
Air Blank	0.000	21:34
Subject Sample # 1	0.271	21:35
Breath Volume	1.816	21:35
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.263	21:38
Breath Volume	2.574	21:38
Air Blank	0.000	21:38
Control Test	0.078	21:39
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: BRONX SKELTON FENZ

DATE/TIME OF TEST: 06/19/2020 21:22:28

Diagnostics Check	OK	21:15
Air Blank	0.000	21:16
Control Test	0.080	21:16
Air Blank	0.000	21:17
Subject Sample # 1	0.384	21:17
Breath Volume	1.554	21:17
Air Blank	0.000	21:18
Air Blank	0.000	21:20
Subject Sample # 2	0.371	21:20
Breath Volume	1.976	21:20
Air Blank	0.000	21:21
Control Test	0.079	21:21
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,SHELBY,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CARROLL,SHELBY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: PETRICCA

DATE/TIME OF TEST: 06/20/2020 02:11:53

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.080	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.141	02:07
Breath Volume		2.066	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:09
Subject Sample	# 2	0.136	02:10
Breath Volume		1.218	02:10
Air Blank		0.000	02:10
Control Test		0.078	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: MCCALL

DATE/TIME OF TEST: 06/20/2020 02:47:21

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.000	02:42
Breath Volume	2.460	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.000	02:46
Breath Volume	1.972	02:46
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQUEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: LAW

DATE/TIME OF TEST: 06/20/2020 04:37:24

Diagnostics Check		OK	04:30
Air Blank		0.000	04:31
Control Test		0.080	04:31
Air Blank		0.000	04:31
Subject Sample	# 1	0.170	04:32
Breath Volume		1.816	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:34
Subject Sample	# 2	0.164	04:35
Breath Volume		1.578	04:35
Air Blank		0.000	04:36
Control Test		0.079	04:36
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: NOBLE

DATE/TIME OF TEST: 06/21/2020 19:37:23

Diagnostics Check		OK	19:21	
Air Blank		0.000	19:22	
Control Test		0.080	19:22	
Air Blank		0.000	19:23	
Subject Sample	# 1	0.110	19:26	Volume Not Met
Breath Volume		0.402	19:26	
Air Blank		0.000	19:26	
Air Blank		0.000	19:28	
Subject Sample	# 2	0.123	19:32	
Breath Volume		1.570	19:32	
Air Blank		0.000	19:32	
Air Blank		0.000	19:34	
Subject Sample	# 3	0.123	19:35	
Breath Volume		1.722	19:35	
Air Blank		0.000	19:36	
Control Test		0.078	19:36	
Air Blank		0.000	19:37	
Diagnostics Check		OK	19:37	

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ABEL,ANTHONY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: ABDI

DATE/TIME OF TEST: 06/22/2020 01:18:16

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:07	
Control Test	0.080	01:07	
Air Blank	0.000	01:08	
Subject Sample	# 1	01:11	No Sample Provided
Breath Volume	0.000	01:11	
Air Blank	0.000	01:11	
Air Blank	0.000	01:13	
Subject Sample	# 2	01:16	No Sample Provided
Breath Volume	0.000	01:16	
Air Blank	0.000	01:17	
Control Test	0.080	01:17	
Air Blank	0.000	01:18	
Diagnostics Check	OK	01:18	

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ABEL,ANTHONY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: ABDI

DATE/TIME OF TEST: 06/22/2020 01:34:29

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.130	01:27
Breath Volume	3.468	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.134	01:32
Breath Volume	2.527	01:32
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGUSA,NICHOLE,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RAGUSA,NICHOLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: AUCOIN

DATE/TIME OF TEST: 06/23/2020 23:18:02

Diagnostics Check	OK	23:11
Air Blank	0.000	23:11
Control Test	0.080	23:12
Air Blank	0.000	23:12
Subject Sample # 1	0.136	23:13
Breath Volume	2.109	23:13
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.141	23:16
Breath Volume	1.429	23:16
Air Blank	0.000	23:17
Control Test	0.078	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:18

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RACER,JACOB,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RACER,JACOB
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 06:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:38

SUBJECT'S LAST NAME: YOUNGBLOOD

DATE/TIME OF TEST: 06/24/2020 08:03:48

Diagnostics Check	OK	07:59	
Air Blank	0.000	08:00	
Control Test	0.081	08:00	
Air Blank	0.000	08:01	
Subject Sample	# 1	08:02	Subject Test Refused
Breath Volume	0.000	08:02	
Air Blank	0.000	08:02	
Control Test	0.080	08:03	
Air Blank	0.000	08:03	
Diagnostics Check	OK	08:03	

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NABULSI,T,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NABULSI,T
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 14:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: DRY

DATE/TIME OF TEST: 06/24/2020 15:41:25

Diagnostics Check	OK	15:37	
Air Blank	0.000	15:37	
Control Test	0.080	15:37	
Air Blank	0.000	15:38	
Subject Sample	# 1	15:40	Subject Test Refused
Breath Volume	0.000	15:40	
Air Blank	0.000	15:40	
Control Test	0.080	15:40	
Air Blank	0.000	15:41	
Diagnostics Check	OK	15:41	

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,GABRIEL,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ROBERSON,GABRIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/26/2020 01:43:45

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.080	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.000	01:39
Breath Volume	1.457	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.000	01:42
Breath Volume	1.957	01:42
Air Blank	0.000	01:42
Control Test	0.079	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: GOODRUM

DATE/TIME OF TEST: 06/26/2020 02:39:18

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.079	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.120	02:34
Breath Volume	2.488	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.114	02:37
Breath Volume	1.382	02:37
Air Blank	0.000	02:38
Control Test	0.078	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER, LANDON, L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SIMPKINS, BRANDON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 14:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:08

SUBJECT'S LAST NAME: BOENNING

DATE/TIME OF TEST: 06/26/2020 15:39:40

Diagnostics Check	OK	15:33
Air Blank	0.000	15:33
Control Test	0.079	15:34
Air Blank	0.000	15:34
Subject Sample # 1	0.080	15:35
Breath Volume	3.421	15:35
Air Blank	0.000	15:35
Air Blank	0.000	15:37
Subject Sample # 2	0.080	15:38
Breath Volume	3.101	15:38
Air Blank	0.000	15:38
Control Test	0.079	15:39
Air Blank	0.000	15:39
Diagnostics Check	OK	15:39

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,AARON,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ZELLARS,TRENT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 06/27/2020 21:04:01

Diagnostics Check	OK	20:50	
Air Blank	0.000	20:50	
Control Test	0.079	20:50	
Air Blank	0.000	20:51	
Subject Sample # 1	0.211	20:53	
Breath Volume	1.203	20:53	
Air Blank	0.000	20:54	
Air Blank	0.000	20:55	
Subject Sample # 2	0.181	20:58	Volume Not Met
Breath Volume	0.015	20:58	
Air Blank	0.000	20:59	
Air Blank	0.000	21:01	
Subject Sample # 3	0.222	21:02	
Breath Volume	1.464	21:02	
Air Blank	0.000	21:03	
Control Test	0.074	21:03	Control Outside Tolerance
Air Blank	0.000	21:04	

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNHAM,J,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PLANCK,A
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: BUSBY

DATE/TIME OF TEST: 06/28/2020 20:51:57

Diagnostics Check	OK	20:45
Air Blank	0.000	20:45
Control Test	0.079	20:46
Air Blank	0.000	20:46
Subject Sample # 1	0.153	20:47
Breath Volume	1.562	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:49
Subject Sample # 2	0.159	20:50
Breath Volume	1.414	20:50
Air Blank	0.000	20:51
Control Test	0.079	20:51
Air Blank	0.000	20:51
Diagnostics Check	OK	20:51

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORTON,JAMES,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LAPEE,WYATT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: KAEWRASAMEE

DATE/TIME OF TEST: 06/28/2020 23:19:46

Diagnostics Check		OK	23:08	
Air Blank		0.000	23:08	
Control Test		0.079	23:08	
Air Blank		0.000	23:09	
Subject Sample	# 1	0.200	23:12	Volume Not Met
Breath Volume		0.523	23:12	
Air Blank		0.000	23:13	
Air Blank		0.000	23:14	
Subject Sample	# 2	0.215	23:18	Volume Not Met
Breath Volume		0.085	23:18	
Air Blank		0.000	23:18	
Control Test		0.078	23:19	
Air Blank		0.000	23:19	
Diagnostics Check		OK	23:19	

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 06:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: BENNISON

DATE/TIME OF TEST: 06/30/2020 07:52:14

Diagnostics Check	OK	07:46
Air Blank	0.000	07:46
Control Test	0.081	07:46
Air Blank	0.000	07:47
Subject Sample # 1	0.000	07:47
Breath Volume	4.003	07:47
Air Blank	0.000	07:48
Air Blank	0.000	07:50
Subject Sample # 2	0.000	07:50
Breath Volume	2.074	07:50
Air Blank	0.000	07:51
Control Test	0.080	07:51
Air Blank	0.000	07:52
Diagnostics Check	OK	07:52

Data Download Date/Time: 07/02/2020 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: BEHM

DATE/TIME OF TEST: 06/07/2020 00:17:37

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:01	
Control Test	0.078	00:01	
Air Blank	0.000	00:01	
Subject Sample # 1	0.115	00:04	
Breath Volume	1.488	00:04	
Air Blank	0.000	00:05	
Air Blank	0.000	00:07	
Subject Sample # 2	0.103	00:10	Volume Not Met
Breath Volume	0.308	00:10	
Air Blank	0.000	00:11	
Air Blank	0.000	00:13	
Subject Sample # 3		00:16	No Sample Provided
Breath Volume	0.000	00:16	
Air Blank	0.000	00:16	
Control Test	0.077	00:17	
Air Blank	0.000	00:17	
Diagnostics Check	OK	00:17	

Data Download Date/Time: 06/30/2020 07:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: FUEHRER

DATE/TIME OF TEST: 06/01/2020 05:22:36

Diagnostics Check		OK	05:16
Air Blank		0.000	05:16
Control Test		0.082	05:17
Air Blank		0.000	05:17
Subject Sample	# 1	0.207	05:17
Breath Volume		1.671	05:17
Air Blank		0.000	05:18
Air Blank		0.000	05:20
Subject Sample	# 2	0.199	05:21
Breath Volume		1.558	05:21
Air Blank		0.000	05:21
Control Test		0.081	05:21
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 06/04/2020 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: LEE II

DATE/TIME OF TEST: 06/05/2020 00:35:51

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.080	00:30
Air Blank	0.000	00:30
Subject Sample # 1	0.000	00:31
Breath Volume	2.828	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.000	00:34
Breath Volume	1.500	00:34
Air Blank	0.000	00:34
Control Test	0.081	00:35
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 07/21/2020 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: DRACOS

DATE/TIME OF TEST: 06/07/2020 05:21:47

Diagnostics Check	OK	05:14
Air Blank	0.000	05:15
Control Test	0.080	05:15
Air Blank	0.000	05:16
Subject Sample # 1	0.122	05:16
Breath Volume	1.296	05:16
Air Blank	0.000	05:17
Air Blank	0.000	05:19
Subject Sample # 2	0.138	05:20
Breath Volume	1.359	05:20
Air Blank	0.000	05:20
Control Test	0.080	05:21
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 07/21/2020 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: SCHNURMAN

DATE/TIME OF TEST: 06/10/2020 20:58:13

Diagnostics Check	OK	20:50
Air Blank	0.000	20:50
Control Test	0.080	20:51
Air Blank	0.000	20:51
Subject Sample # 1	0.217	20:52
Breath Volume	1.171	20:52
Air Blank	0.000	20:52
Air Blank	0.000	20:54
Subject Sample # 2	0.210	20:56
Breath Volume	1.492	20:56
Air Blank	0.000	20:57
Control Test	0.081	20:57
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 07/21/2020 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 06/11/2020 03:32:48

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.081	03:27
Air Blank		0.000	03:27
Subject Sample	# 1	0.218	03:28
Breath Volume		2.621	03:28
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.229	03:31
Breath Volume		3.328	03:31
Air Blank		0.000	03:31
Control Test		0.081	03:32
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 07/21/2020 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: HERNANDEZ GONZALEZ

DATE/TIME OF TEST: 06/12/2020 02:33:59

Diagnostics Check		OK	02:22	
Air Blank		0.000	02:22	
Control Test		0.081	02:23	
Air Blank		0.000	02:23	
Subject Sample	# 1	0.217	02:26	Volume Not Met
Breath Volume		0.039	02:26	
Air Blank		0.000	02:27	
Air Blank		0.000	02:29	
Subject Sample	# 2	0.201	02:32	Volume Not Met
Breath Volume		0.238	02:32	
Air Blank		0.000	02:33	
Control Test		0.080	02:33	
Air Blank		0.000	02:33	
Diagnostics Check		OK	02:33	

Data Download Date/Time: 07/21/2020 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: HERNANDEZ GONZALEZ

DATE/TIME OF TEST: 06/12/2020 03:04:51

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:02	
Control Test	0.081	03:02	
Air Blank	0.000	03:03	
Subject Sample	# 1	03:03	Subject Test Refused
Breath Volume	0.000	03:03	
Air Blank	0.000	03:03	
Control Test	0.080	03:04	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:04	

Data Download Date/Time: 07/21/2020 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/13/2020 23:34:43

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.081	23:29
Air Blank	0.000	23:29
Subject Sample # 1	0.155	23:30
Breath Volume	1.929	23:30
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.140	23:33
Breath Volume	1.253	23:33
Air Blank	0.000	23:33
Control Test	0.081	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 07/21/2020 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ARCHAMBEAU,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: TATEM III

DATE/TIME OF TEST: 06/14/2020 19:23:54

Diagnostics Check		OK	19:14
Air Blank		0.000	19:14
Control Test		0.081	19:15
Air Blank		0.000	19:15
Subject Sample	# 1	0.328	19:17
Breath Volume		1.695	19:17
Air Blank		0.000	19:18
Air Blank		0.000	19:19
Subject Sample	# 2	0.322	19:22
Breath Volume		1.105	19:22
Air Blank		0.000	19:22
Control Test		0.080	19:23
Air Blank		0.000	19:23
Diagnostics Check		OK	19:23

Data Download Date/Time: 07/21/2020 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: EVERETT

DATE/TIME OF TEST: 06/20/2020 03:55:53

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.081	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.138	03:51
Breath Volume		2.152	03:51
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.136	03:54
Breath Volume		2.695	03:54
Air Blank		0.000	03:54
Control Test		0.080	03:55
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 07/21/2020 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 06/20/2020 20:25:01

Diagnostics Check	OK	20:21	
Air Blank	0.000	20:22	
Control Test	0.081	20:22	
Air Blank	0.000	20:23	
Subject Sample	# 1	20:23	Subject Test Refused
Breath Volume	0.000	20:23	
Air Blank	0.000	20:24	
Control Test	0.081	20:24	
Air Blank	0.000	20:24	
Diagnostics Check	OK	20:25	

Data Download Date/Time: 07/21/2020 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: TWIGGS SR

DATE/TIME OF TEST: 06/28/2020 23:52:05

Diagnostics Check		OK	23:39	
Air Blank		0.000	23:39	
Control Test		0.080	23:40	
Air Blank		0.000	23:40	
Subject Sample	# 1	0.059	23:43	Volume Not Met
Breath Volume		0.214	23:43	
Air Blank		0.000	23:44	
Air Blank		0.000	23:46	
Subject Sample	# 2	0.070	23:46	
Breath Volume		1.960	23:46	
Air Blank		0.000	23:47	
Air Blank		0.000	23:49	
Subject Sample	# 3	0.067	23:50	
Breath Volume		1.875	23:50	
Air Blank		0.000	23:51	
Control Test		0.081	23:51	
Air Blank		0.000	23:52	
Diagnostics Check		OK	23:52	

Data Download Date/Time: 07/21/2020 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 06/29/2020 22:53:26

Diagnostics Check	OK	22:45
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.142	22:47
Breath Volume	1.613	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:49
Subject Sample # 2	0.142	22:51
Breath Volume	1.390	22:51
Air Blank	0.000	22:52
Control Test	0.082	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 07/21/2020 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005146

INSTRUMENT REGISTERED TO: PATRICK AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRUMM,MATTHEW,P
BREATH TEST OPERATOR AGENCY: Patrick AFB PD
ARRESTING OFFICER: ROBINSON,JAMES
ARRESTING OFFICER AGENCY: Patrick AFB PD
LAST AGENCY INSPECTION: 06/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 89547
DRY GAS EXPIRY DATE: 06/17/2021
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:56

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 06/25/2020 16:23:43

Diagnostics Check	OK	16:17
Air Blank	0.000	16:18
Control Test	0.081	16:18
Air Blank	0.000	16:18
Subject Sample # 1	0.000	16:19
Breath Volume	2.593	16:19
Air Blank	0.000	16:19
Air Blank	0.000	16:21
Subject Sample # 2	0.000	16:22
Breath Volume	3.218	16:22
Air Blank	0.000	16:22
Control Test	0.079	16:23
Air Blank	0.000	16:23
Diagnostics Check	OK	16:23

Data Download Date/Time: 07/10/2020 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: MUSSER

DATE/TIME OF TEST: 06/05/2020 23:44:16

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:41	
Control Test	0.078	23:42	
Air Blank	0.000	23:42	
Subject Sample	# 1	23:42	Subject Test Refused
Breath Volume	0.000	23:42	
Air Blank	0.000	23:43	
Control Test	0.077	23:43	
Air Blank	0.000	23:44	
Diagnostics Check	OK	23:44	

Data Download Date/Time: 06/16/2020 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,BILLY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ORTIZ,BILLY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: LILES

DATE/TIME OF TEST: 06/06/2020 02:27:22

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.077	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.170	02:22
Breath Volume	3.867	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.153	02:25
Breath Volume	2.031	02:25
Air Blank	0.000	02:26
Control Test	0.076	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 06/16/2020 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATIF, ANTHONY, R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: HOAGLAN, BRANDON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:43

SUBJECT'S LAST NAME: BETHUNE

DATE/TIME OF TEST: 06/07/2020 09:14:10

Diagnostics Check	OK	09:03
Air Blank	0.000	09:04
Control Test	0.079	09:04
Air Blank	0.000	09:05
Subject Sample # 1	0.000	09:08
Breath Volume	1.226	09:08
Air Blank	0.000	09:08
Air Blank	0.000	09:10
Subject Sample # 2	0.000	09:12
Breath Volume	1.394	09:12
Air Blank	0.000	09:13
Control Test	0.079	09:13
Air Blank	0.000	09:14
Diagnostics Check	OK	09:14

Data Download Date/Time: 06/16/2020 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARAMIELLO,CATEY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WITHERELL,B
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: DUSZAK

DATE/TIME OF TEST: 06/11/2020 17:47:47

Diagnostics Check	OK	17:41
Air Blank	0.000	17:41
Control Test	0.078	17:41
Air Blank	0.000	17:42
Subject Sample # 1	0.264	17:43
Breath Volume	3.921	17:43
Air Blank	0.000	17:43
Air Blank	0.000	17:45
Subject Sample # 2	0.273	17:46
Breath Volume	3.667	17:46
Air Blank	0.000	17:46
Control Test	0.078	17:47
Air Blank	0.000	17:47
Diagnostics Check	OK	17:47

Data Download Date/Time: 06/16/2020 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: FOUNTAIN

DATE/TIME OF TEST: 06/11/2020 22:31:40

Diagnostics Check	OK	22:25
Air Blank	0.000	22:25
Control Test	0.079	22:25
Air Blank	0.000	22:26
Subject Sample # 1	0.000	22:27
Breath Volume	1.136	22:27
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.000	22:30
Breath Volume	1.128	22:30
Air Blank	0.000	22:30
Control Test	0.079	22:31
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 06/16/2020 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 06/13/2020 03:23:56

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.080	03:21	
Air Blank	0.000	03:22	
Subject Sample	# 1	03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:23	
Control Test	0.079	03:23	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 06/16/2020 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEIRCE,HAROLD,F
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RAWALD,PAUL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: WYERS

DATE/TIME OF TEST: 06/13/2020 17:17:34

Diagnostics Check	OK	17:10
Air Blank	0.000	17:11
Control Test	0.079	17:11
Air Blank	0.000	17:11
Subject Sample # 1	0.227	17:12
Breath Volume	1.441	17:12
Air Blank	0.000	17:13
Air Blank	0.000	17:14
Subject Sample # 2	0.217	17:15
Breath Volume	1.910	17:15
Air Blank	0.000	17:16
Control Test	0.078	17:16
Air Blank	0.000	17:17
Diagnostics Check	OK	17:17

Data Download Date/Time: 06/16/2020 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIBBEY',BRADLEY,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIBBEY,BRADLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: WIESER

DATE/TIME OF TEST: 06/14/2020 00:57:16

Diagnostics Check		OK	00:45	
Air Blank		0.000	00:46	
Control Test		0.078	00:46	
Air Blank		0.000	00:46	
Subject Sample	# 1	0.000	00:50	Volume Not Met
Breath Volume		0.015	00:50	
Air Blank		0.000	00:50	
Air Blank		0.000	00:52	
Subject Sample	# 2	0.172	00:55	Volume Not Met
Breath Volume		0.000	00:55	
Air Blank		0.000	00:56	
Control Test		0.078	00:56	
Air Blank		0.000	00:57	
Diagnostics Check		OK	00:57	

Data Download Date/Time: 06/16/2020 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTIGLIONE,SALVATORE,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CASTIGLIONE,SALVATORE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/24/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 06/14/2020 03:25:53

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.078	03:22	
Air Blank	0.000	03:22	
Subject Sample	# 1	03:24	Subject Test Refused
Breath Volume	0.000	03:24	
Air Blank	0.000	03:24	
Control Test	0.079	03:25	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 06/16/2020 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HECKMAN,DANIEL,O
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WHITAKER,G
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: VARELA

DATE/TIME OF TEST: 06/14/2020 04:32:30

Diagnostics Check		OK	04:25
Air Blank		0.000	04:26
Control Test		0.079	04:26
Air Blank		0.000	04:27
Subject Sample	# 1	0.174	04:27
Breath Volume		1.976	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.177	04:30
Breath Volume		2.707	04:30
Air Blank		0.000	04:31
Control Test		0.079	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 06/16/2020 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEIRCE,HAROLD,F
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SMITH,BRIAN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: MCMANUS

DATE/TIME OF TEST: 06/14/2020 08:33:22

Diagnostics Check	OK	08:30	
Air Blank	0.000	08:30	
Control Test	0.077	08:31	
Air Blank	0.000	08:31	
Subject Sample	# 1	08:32	Subject Test Refused
Breath Volume	0.000	08:32	
Air Blank	0.000	08:32	
Control Test	0.078	08:32	
Air Blank	0.000	08:33	
Diagnostics Check	OK	08:33	

Data Download Date/Time: 06/16/2020 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 06/15/2020 00:58:16

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.079	00:52
Air Blank		0.000	00:53
Subject Sample	# 1	0.177	00:53
Breath Volume		2.289	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.183	00:56
Breath Volume		2.527	00:56
Air Blank		0.000	00:57
Control Test		0.078	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 06/16/2020 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: LOWTHER

DATE/TIME OF TEST: 06/16/2020 00:10:21

Diagnostics Check		OK	00:02
Air Blank		0.000	00:03
Control Test		0.078	00:03
Air Blank		0.000	00:03
Subject Sample	# 1	0.236	00:05
Breath Volume		2.160	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.256	00:08
Breath Volume		2.125	00:08
Air Blank		0.000	00:09
Control Test		0.079	00:09
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 06/16/2020 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAVALLE,D,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LAVALLE,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: MOYNIHAN

DATE/TIME OF TEST: 06/18/2020 01:08:56

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.079	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.121	01:04
Breath Volume	3.500	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:06
Subject Sample # 2	0.118	01:07
Breath Volume	2.980	01:07
Air Blank	0.000	01:08
Control Test	0.077	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 07/27/2020 07:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLITO,BRAD,R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SMITH,BRIAN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:56

SUBJECT'S LAST NAME: BETHUNE

DATE/TIME OF TEST: 06/18/2020 13:28:26

Diagnostics Check	OK	13:17	
Air Blank	0.000	13:17	
Control Test	0.079	13:17	
Air Blank	0.000	13:18	
Subject Sample	# 1 0.000	13:21	Volume Not Met
Breath Volume	0.468	13:21	
Air Blank	0.000	13:21	
Air Blank	0.000	13:23	
Subject Sample	# 2 0.000	13:27	Volume Not Met
Breath Volume	0.453	13:27	
Air Blank	0.000	13:27	
Control Test	0.078	13:27	
Air Blank	0.000	13:28	
Diagnostics Check	OK	13:28	

Data Download Date/Time: 07/27/2020 07:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATIF, ANTHONY, R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MITRO, JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:42

SUBJECT'S LAST NAME: LIND

DATE/TIME OF TEST: 06/21/2020 15:09:23

Diagnostics Check	OK	15:02
Air Blank	0.000	15:03
Control Test	0.079	15:03
Air Blank	0.000	15:04
Subject Sample # 1	0.000	15:04
Breath Volume	1.355	15:04
Air Blank	0.000	15:05
Air Blank	0.000	15:07
Subject Sample # 2	0.000	15:08
Breath Volume	1.152	15:08
Air Blank	0.000	15:08
Control Test	0.078	15:08
Air Blank	0.000	15:09
Diagnostics Check	OK	15:09

Data Download Date/Time: 07/27/2020 07:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAVALLE,DANIEL,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LAVALLE,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: LEITAO

DATE/TIME OF TEST: 06/25/2020 04:52:38

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.079	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.000	04:47
Breath Volume	3.320	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.000	04:51
Breath Volume	1.796	04:51
Air Blank	0.000	04:51
Control Test	0.079	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 07/27/2020 07:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: KROLL

DATE/TIME OF TEST: 06/26/2020 18:35:13

Diagnostics Check	OK	18:28
Air Blank	0.000	18:29
Control Test	0.079	18:29
Air Blank	0.000	18:30
Subject Sample # 1	0.181	18:30
Breath Volume	1.968	18:30
Air Blank	0.000	18:31
Air Blank	0.000	18:33
Subject Sample # 2	0.184	18:33
Breath Volume	2.281	18:33
Air Blank	0.000	18:34
Control Test	0.078	18:34
Air Blank	0.000	18:35
Diagnostics Check	OK	18:35

Data Download Date/Time: 07/27/2020 07:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELM,DEREK,W
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HELM,DEREK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: GREINER

DATE/TIME OF TEST: 06/28/2020 04:41:21

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.079	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.196	04:36
Breath Volume	1.808	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.216	04:39
Breath Volume	1.394	04:39
Air Blank	0.000	04:40
Control Test	0.078	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 07/27/2020 07:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIBBEY,BRADLEY,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIBBEY,BRADLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: CAIN

DATE/TIME OF TEST: 06/28/2020 23:44:32

Diagnostics Check	OK	23:37
Air Blank	0.000	23:38
Control Test	0.079	23:38
Air Blank	0.000	23:39
Subject Sample # 1	0.000	23:40
Breath Volume	2.308	23:40
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.000	23:43
Breath Volume	2.003	23:43
Air Blank	0.000	23:43
Control Test	0.079	23:43
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 07/27/2020 07:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAILLY,JOEL,N
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 06/10/2020 05:46:54

Diagnostics Check	OK	05:40
Air Blank	0.000	05:41
Control Test	0.080	05:41
Air Blank	0.000	05:42
Subject Sample # 1	0.000	05:42
Breath Volume	3.226	05:42
Air Blank	0.000	05:43
Air Blank	0.000	05:45
Subject Sample # 2	0.000	05:45
Breath Volume	3.382	05:45
Air Blank	0.000	05:46
Control Test	0.080	05:46
Air Blank	0.000	05:46
Diagnostics Check	OK	05:46

Data Download Date/Time: 07/15/2020 10:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAFFIN,TRACE,D
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: SWINDLE,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 06/28/2020 14:19:00

Diagnostics Check	OK	14:16	
Air Blank	0.000	14:17	
Control Test	0.080	14:17	
Air Blank	0.000	14:18	
Subject Sample	# 1	14:18	Sequence Aborted
Breath Volume	0.000	14:18	
Air Blank	0.000	14:19	

Data Download Date/Time: 07/15/2020 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAFFIN, TRACE, D
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: SWINDLE, CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 06/28/2020 14:41:29

Diagnostics Check	OK	14:35
Air Blank	0.000	14:35
Control Test	0.080	14:36
Air Blank	0.000	14:36
Subject Sample # 1	0.000	14:37
Breath Volume	1.238	14:37
Air Blank	0.000	14:37
Air Blank	0.000	14:39
Subject Sample # 2	0.000	14:40
Breath Volume	2.164	14:40
Air Blank	0.000	14:40
Control Test	0.079	14:40
Air Blank	0.000	14:41
Diagnostics Check	OK	14:41

Data Download Date/Time: 07/15/2020 10:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF,MICHAEL,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: OTHER
ARREST TIME: 10:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: WEISKOPF

DATE/TIME OF TEST: 06/03/2020 10:52:40

Diagnostics Check	OK	10:49
Air Blank	0.000	10:50
Control Test	0.078	10:50
Air Blank	0.000	10:51
Subject Sample # 1	0.000	10:51
Breath Volume	1.171	10:51
Air Blank	0.000	10:51
Air Blank	0.000	10:52

Data Download Date/Time: 07/14/2020 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAFFERTY,NAOMI
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: DIETRICH

DATE/TIME OF TEST: 06/05/2020 01:30:39

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:19	
Control Test	0.079	01:20	
Air Blank	0.000	01:20	
Subject Sample	# 1	01:23	No Sample Provided
Breath Volume	0.000	01:23	
Air Blank	0.000	01:24	
Air Blank	0.000	01:26	
Subject Sample	# 2	01:29	No Sample Provided
Breath Volume	0.000	01:29	
Air Blank	0.000	01:29	
Control Test	0.080	01:30	
Air Blank	0.000	01:30	
Diagnostics Check	OK	01:30	

Data Download Date/Time: 07/14/2020 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: TEST,J
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: OTHER
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 06/05/2020 16:56:06

Diagnostics Check	OK	16:55	
Air Blank		16:55	RFI Detect
Air Blank	0.000	16:56	

Data Download Date/Time: 07/14/2020 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: TEST,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: OTHER
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 06/05/2020 17:05:07

Diagnostics Check	OK	16:58
Air Blank	0.000	16:59
Control Test	0.080	16:59
Air Blank	0.000	17:00
Subject Sample # 1	0.000	17:00
Breath Volume	2.828	17:00
Air Blank	0.000	17:01
Air Blank	0.000	17:03
Subject Sample # 2	0.000	17:03
Breath Volume	2.464	17:03
Air Blank	0.000	17:04
Control Test	0.080	17:04
Air Blank	0.000	17:05
Diagnostics Check	OK	17:05

Data Download Date/Time: 07/14/2020 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAFFERTY,NAOMI
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: ALDOSSARI

DATE/TIME OF TEST: 06/05/2020 23:07:25

Diagnostics Check		OK	23:01
Air Blank		0.000	23:01
Control Test		0.079	23:01
Air Blank		0.000	23:02
Subject Sample	# 1	0.249	23:03
Breath Volume		1.601	23:03
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.252	23:05
Breath Volume		2.625	23:05
Air Blank		0.000	23:06
Control Test		0.077	23:06
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 07/14/2020 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 06/06/2020 02:05:32

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:54	
Control Test	0.079	01:54	
Air Blank	0.000	01:55	
Subject Sample	# 1	01:58	No Sample Provided
Breath Volume	0.000	01:58	
Air Blank	0.000	01:59	
Air Blank	0.000	02:01	
Subject Sample	# 2	02:04	No Sample Provided
Breath Volume	0.000	02:04	
Air Blank	0.000	02:04	
Control Test	0.079	02:04	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 07/14/2020 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCULLOUGH,SEAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: NALE

DATE/TIME OF TEST: 06/06/2020 03:32:19

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.203	03:27
Breath Volume	2.414	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.193	03:30
Breath Volume	2.058	03:30
Air Blank	0.000	03:31
Control Test	0.075	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 07/14/2020 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE, RONALD,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE, RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 06/09/2020 02:45:58

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:36	
Control Test	0.079	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 1 0.211	02:38	
Breath Volume	1.828	02:38	
Air Blank	0.000	02:39	
Air Blank	0.000	02:41	
Subject Sample	# 2 0.190	02:41	
Breath Volume	2.093	02:41	
Air Blank	0.000	02:42	
Air Blank	0.000	02:44	
Subject Sample	# 3 0.186	02:44	
Breath Volume	2.496	02:44	
Air Blank	0.000	02:45	
Control Test	0.074	02:45	Control Outside Tolerance
Air Blank	0.000	02:45	

Data Download Date/Time: 07/14/2020 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAFFERTY,NAOMI
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: BRADFORD

DATE/TIME OF TEST: 06/11/2020 17:25:19

Diagnostics Check	OK	17:13	
Air Blank	0.000	17:14	
Control Test	0.078	17:14	
Air Blank	0.000	17:15	
Subject Sample	# 1	17:18	No Sample Provided
Breath Volume	0.000	17:18	
Air Blank	0.000	17:18	
Air Blank	0.000	17:20	
Subject Sample	# 2	17:23	No Sample Provided
Breath Volume	0.000	17:23	
Air Blank	0.000	17:24	
Control Test	0.078	17:24	
Air Blank	0.000	17:25	
Diagnostics Check	OK	17:25	

Data Download Date/Time: 07/14/2020 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCULLOUGH,SEAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: TRAVIS

DATE/TIME OF TEST: 06/11/2020 23:02:20

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.079	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.216	22:57
Breath Volume	2.828	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.212	23:00
Breath Volume	3.066	23:00
Air Blank	0.000	23:01
Control Test	0.076	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 07/14/2020 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAFFERTY,NAOMI
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: RAMSAY

DATE/TIME OF TEST: 06/12/2020 21:31:51

Diagnostics Check	OK	21:25
Air Blank	0.000	21:26
Control Test	0.079	21:26
Air Blank	0.000	21:26
Subject Sample # 1	0.179	21:27
Breath Volume	3.367	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:29
Subject Sample # 2	0.183	21:30
Breath Volume	3.605	21:30
Air Blank	0.000	21:30
Control Test	0.079	21:31
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

Data Download Date/Time: 07/14/2020 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAIA,MICHAEL,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SAIA,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 06/14/2020 02:00:53

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.000	01:56
Breath Volume	2.480	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.000	01:59
Breath Volume	3.218	01:59
Air Blank	0.000	01:59
Control Test	0.078	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 07/14/2020 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HOBART,CLAYTON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: BEANE

DATE/TIME OF TEST: 06/16/2020 19:18:50

Diagnostics Check	OK	19:12
Air Blank	0.000	19:12
Control Test	0.080	19:13
Air Blank	0.000	19:13
Subject Sample # 1	0.152	19:14
Breath Volume	2.535	19:14
Air Blank	0.000	19:15
Air Blank	0.000	19:16
Subject Sample # 2	0.157	19:17
Breath Volume	3.042	19:17
Air Blank	0.000	19:17
Control Test	0.077	19:18
Air Blank	0.000	19:18
Diagnostics Check	OK	19:18

Data Download Date/Time: 07/14/2020 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRETT,DAVID,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLATTA,VLADAMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:12

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 06/17/2020 12:39:42

Diagnostics Check	OK	12:33
Air Blank	0.000	12:33
Control Test	0.080	12:34
Air Blank	0.000	12:34
Subject Sample # 1	0.000	12:35
Breath Volume	1.574	12:35
Air Blank	0.000	12:35
Air Blank	0.000	12:37
Subject Sample # 2	0.000	12:38
Breath Volume	1.578	12:38
Air Blank	0.000	12:38
Control Test	0.078	12:39
Air Blank	0.000	12:39
Diagnostics Check	OK	12:39

Data Download Date/Time: 07/14/2020 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAFFERTY,NAOMI,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: KEYS

DATE/TIME OF TEST: 06/20/2020 00:39:22

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.079	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.180	00:30
Breath Volume	1.554	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.148	00:34
Breath Volume	1.664	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 3	0.175	00:37
Breath Volume	1.175	00:37
Air Blank	0.000	00:38
Control Test	0.077	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 07/14/2020 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: VOLPE

DATE/TIME OF TEST: 06/20/2020 23:23:02

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.080	23:17
Air Blank	0.000	23:18
Subject Sample # 1	0.000	23:18
Breath Volume	3.929	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.000	23:21
Breath Volume	3.605	23:21
Air Blank	0.000	23:22
Control Test	0.080	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:23

Data Download Date/Time: 07/14/2020 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: GALLO

DATE/TIME OF TEST: 06/21/2020 01:29:04

Diagnostics Check	OK	01:26	
Air Blank	0.000	01:26	
Control Test	0.080	01:27	
Air Blank	0.000	01:27	
Subject Sample	# 1	01:27	Subject Test Refused
Breath Volume	0.000	01:27	
Air Blank	0.000	01:28	
Control Test	0.080	01:28	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:29	

Data Download Date/Time: 07/14/2020 09:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:11

SUBJECT'S LAST NAME: SAMPSON

DATE/TIME OF TEST: 06/24/2020 17:38:33

Diagnostics Check	OK	17:31
Air Blank	0.000	17:31
Control Test	0.079	17:32
Air Blank	0.000	17:32
Subject Sample # 1	0.291	17:34
Breath Volume	1.261	17:34
Air Blank	0.000	17:34
Air Blank	0.000	17:36
Subject Sample # 2	0.295	17:36
Breath Volume	1.425	17:36
Air Blank	0.000	17:37
Control Test	0.078	17:37
Air Blank	0.000	17:38
Diagnostics Check	OK	17:38

Data Download Date/Time: 07/14/2020 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: CAPOLINO

DATE/TIME OF TEST: 06/27/2020 21:50:18

Diagnostics Check	OK	21:47	
Air Blank	0.000	21:47	
Control Test	0.079	21:48	
Air Blank	0.000	21:48	
Subject Sample	# 1	21:48	Subject Test Refused
Breath Volume	0.000	21:48	
Air Blank	0.000	21:49	
Control Test	0.079	21:49	
Air Blank	0.000	21:50	
Diagnostics Check	OK	21:50	

Data Download Date/Time: 07/14/2020 09:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVILA,NICOLE,M
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: RETTE,CODY
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2021
VIOLATION CODE: DUI
ARREST TIME: 10:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: POWERS

DATE/TIME OF TEST: 06/03/2020 11:26:46

Diagnostics Check	OK	11:14	
Air Blank	0.000	11:15	
Control Test	0.081	11:15	
Air Blank	0.000	11:16	
Subject Sample # 1	0.000	11:16	
Breath Volume	1.542	11:16	
Air Blank	0.000	11:17	
Air Blank	0.000	11:19	
Subject Sample # 2	0.000	11:22	Volume Not Met
Breath Volume	0.152	11:22	
Air Blank	0.000	11:22	
Air Blank	0.000	11:24	
Subject Sample # 3		11:25	Subject Test Refused
Breath Volume	0.000	11:25	
Air Blank	0.000	11:25	
Control Test	0.078	11:26	
Air Blank	0.000	11:26	
Diagnostics Check	OK	11:26	

Data Download Date/Time: 07/10/2020 19:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KURTZ,THEODORE,F
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: GUILLERM,MICHAEL
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: SHIPKOWITZ

DATE/TIME OF TEST: 06/08/2020 04:44:02

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:38
Subject Sample # 1	0.136	04:39
Breath Volume	1.875	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:41
Subject Sample # 2	0.135	04:42
Breath Volume	1.746	04:42
Air Blank	0.000	04:43
Control Test	0.079	04:43
Air Blank	0.000	04:43
Diagnostics Check	OK	04:44

Data Download Date/Time: 07/10/2020 19:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEAN,WAYNE,G
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: SLUDER,JAMES
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: RANDOLPH

DATE/TIME OF TEST: 06/14/2020 23:42:51

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.081	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.159	23:38
Breath Volume	2.148	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:40
Subject Sample # 2	0.163	23:41
Breath Volume	2.054	23:41
Air Blank	0.000	23:41
Control Test	0.081	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 07/10/2020 19:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMBSH,CARLYTON,R
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: HAMBSH,CARLYTON
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: SALIM

DATE/TIME OF TEST: 06/16/2020 03:33:09

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.081	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.000	03:28
Breath Volume	2.093	03:28
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.000	03:31
Breath Volume	3.449	03:31
Air Blank	0.000	03:32
Control Test	0.080	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 07/10/2020 19:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMBSH,CARLYTON,R
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: SANDERSON,TAYLOR
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 05/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: CONNER

DATE/TIME OF TEST: 06/21/2020 03:08:58

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.080	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.189	03:04
Breath Volume		1.804	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:06
Subject Sample	# 2	0.187	03:07
Breath Volume		2.218	03:07
Air Blank		0.000	03:08
Control Test		0.079	03:08
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 07/10/2020 19:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMBSH,CARLYTON,R
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: CALEND,NICHOLAS
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 06/21/2020 04:24:29

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.081	04:18
Air Blank	0.000	04:18
Subject Sample # 1	0.163	04:19
Breath Volume	5.472	04:19
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.152	04:22
Breath Volume	4.179	04:22
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 07/10/2020 19:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMBSH,CARLYTON,R
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: HAMBSH,CARLYTON
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 06/25/2020 03:42:42

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.081	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.101	03:38
Breath Volume	1.835	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:40
Subject Sample # 2	0.096	03:41
Breath Volume	1.316	03:41
Air Blank	0.000	03:41
Control Test	0.081	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 07/10/2020 19:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BODKIN,MICHEAL,W
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: OLIVER,ALEXIS
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/08/2022
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 06/01/2020 03:08:08

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.077	03:01
Air Blank		0.000	03:01
Subject Sample	# 1	0.248	03:02
Breath Volume		1.855	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.259	03:06
Breath Volume		2.644	03:06
Air Blank		0.000	03:07
Control Test		0.076	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 06/24/2020 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNNE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: DUNNE,JUSTIN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/08/2022
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: LILLEROVDE

DATE/TIME OF TEST: 06/02/2020 01:26:04

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.226	01:18
Breath Volume		1.421	01:18
Air Blank		0.000	01:19
Air Blank		0.000	01:20
Subject Sample	# 2	0.248	01:21
Breath Volume		1.933	01:21
Air Blank		0.000	01:22
Air Blank		0.000	01:23
Subject Sample	# 3	0.219	01:24
Breath Volume		1.542	01:24
Air Blank		0.000	01:25
Control Test		0.079	01:25
Air Blank		0.000	01:25
Diagnostics Check		OK	01:26

Data Download Date/Time: 06/24/2020 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BODKIN,MICHEAL,W
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: WEKONY,ELIJAH
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/08/2022
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: DANIEL

DATE/TIME OF TEST: 06/09/2020 04:21:14

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:16
Subject Sample # 1	0.068	04:16
Breath Volume	3.664	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.065	04:19
Breath Volume	3.316	04:19
Air Blank	0.000	04:20
Control Test	0.079	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 06/24/2020 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BODKIN,MICHEAL,W
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: NELSON,MARC
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/08/2022
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: WETHERELL

DATE/TIME OF TEST: 06/13/2020 02:47:16

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.151	02:42
Breath Volume	2.910	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.145	02:45
Breath Volume	1.148	02:45
Air Blank	0.000	02:46
Control Test	0.079	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 06/24/2020 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,STERLING,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: BARRETT,TRAVIS
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/08/2022
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/13/2020 21:52:03

Diagnostics Check	OK	21:42	
Air Blank	0.000	21:43	
Control Test	0.080	21:43	
Air Blank	0.000	21:44	
Subject Sample	# 1 0.257	21:45	
Breath Volume	2.863	21:45	
Air Blank	0.000	21:45	
Air Blank	0.000	21:47	
Subject Sample	# 2 0.233	21:48	
Breath Volume	2.082	21:48	
Air Blank	0.000	21:48	
Air Blank	0.000	21:50	
Subject Sample	# 3	21:51	RFI Detect
Breath Volume	0.000	21:51	
Air Blank		21:51	RFI Detect
Air Blank	0.000	21:52	

Data Download Date/Time: 06/24/2020 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: PAUTH,JORGE
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/08/2022
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/16/2020 03:12:02

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.080	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.264	03:07
Breath Volume		1.625	03:07
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.266	03:10
Breath Volume		1.761	03:10
Air Blank		0.000	03:11
Control Test		0.080	03:11
Air Blank		0.000	03:11
Diagnostics Check		OK	03:12

Data Download Date/Time: 06/24/2020 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: PAUTH,JORGE
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/08/2022
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: LEBARON

DATE/TIME OF TEST: 06/20/2020 00:18:34

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:14	
Control Test	0.080	00:14	
Air Blank	0.000	00:14	
Subject Sample	# 1	00:16	Subject Test Refused
Breath Volume	0.644	00:16	
Air Blank	0.000	00:17	
Control Test	0.080	00:17	
Air Blank	0.000	00:18	
Diagnostics Check	OK	00:18	

Data Download Date/Time: 06/24/2020 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORREST,RYAN,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: FORREST,RYAN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/08/2022
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: LAYE

DATE/TIME OF TEST: 06/20/2020 21:13:25

Diagnostics Check	OK	21:01
Air Blank	0.000	21:01
Control Test	0.081	21:02
Air Blank	0.000	21:02
Subject Sample # 1	0.194	21:04
Breath Volume	1.277	21:04
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.171	21:08
Breath Volume	1.847	21:08
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample # 3	0.181	21:11
Breath Volume	1.632	21:11
Air Blank	0.000	21:12
Control Test	0.079	21:12
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

Data Download Date/Time: 06/24/2020 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BODKIN,MICHEAL,W
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: BODKIN,MICHEAL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/08/2022
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: BALLOU

DATE/TIME OF TEST: 06/23/2020 19:59:31

Diagnostics Check	OK	19:55	
Air Blank	0.000	19:55	
Control Test	0.079	19:55	
Air Blank	0.000	19:56	
Subject Sample	# 1	19:59	Sequence Aborted
Breath Volume	0.367	19:59	
Air Blank	0.000	19:59	

Data Download Date/Time: 06/24/2020 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VALDEZ,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 06/14/2020 02:44:07

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.078	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.000	02:39
Breath Volume	2.523	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.000	02:42
Breath Volume	1.667	02:42
Air Blank	0.000	02:43
Control Test	0.078	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 06/25/2020 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: HERRERA OVALLE

DATE/TIME OF TEST: 06/14/2020 06:16:27

Diagnostics Check	OK	06:09
Air Blank	0.000	06:10
Control Test	0.076	06:10
Air Blank	0.000	06:11
Subject Sample # 1	0.033	06:11
Breath Volume	1.550	06:11
Air Blank	0.000	06:12
Air Blank	0.000	06:14
Subject Sample # 2	0.037	06:14
Breath Volume	2.230	06:14
Air Blank	0.000	06:15
Control Test	0.077	06:15
Air Blank	0.000	06:16
Diagnostics Check	OK	06:16

Data Download Date/Time: 06/25/2020 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: PATARROYO

DATE/TIME OF TEST: 06/07/2020 02:31:14

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.235	02:26
Breath Volume	2.414	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.222	02:29
Breath Volume	1.792	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 06/25/2020 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 06/13/2020 00:51:54

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.079	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.202	00:47
Breath Volume	3.210	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:49
Subject Sample # 2	0.213	00:50
Breath Volume	3.769	00:50
Air Blank	0.000	00:50
Control Test	0.080	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 06/25/2020 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: SANCHEZ GONZALEZ

DATE/TIME OF TEST: 06/13/2020 05:37:17

Diagnostics Check	OK	05:33	
Air Blank	0.000	05:34	
Control Test	0.081	05:34	
Air Blank	0.000	05:35	
Subject Sample	# 1	05:35	Subject Test Refused
Breath Volume	0.000	05:35	
Air Blank	0.000	05:36	
Control Test	0.081	05:36	
Air Blank		05:36	RFI Detect
Air Blank	0.000	05:37	

Data Download Date/Time: 06/25/2020 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: SANCHEZ GONZALEZ

DATE/TIME OF TEST: 06/13/2020 05:44:26

Diagnostics Check	OK	05:41	
Air Blank	0.000	05:41	
Control Test	0.080	05:42	
Air Blank	0.000	05:42	
Subject Sample	# 1	05:43	Subject Test Refused
Breath Volume	0.000	05:43	
Air Blank	0.000	05:43	
Control Test	0.081	05:43	
Air Blank	0.000	05:44	
Diagnostics Check	OK	05:44	

Data Download Date/Time: 06/25/2020 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CHAVEZ,ANA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 06/13/2020 23:22:19

Diagnostics Check		OK	23:09	
Air Blank		0.000	23:10	
Control Test		0.079	23:10	
Air Blank		0.000	23:10	
Subject Sample	# 1	0.142	23:14	Volume Not Met
Breath Volume		0.875	23:14	
Air Blank		0.000	23:14	
Air Blank		0.000	23:16	
Subject Sample	# 2	0.149	23:17	
Breath Volume		2.472	23:17	
Air Blank		0.000	23:17	
Air Blank		0.000	23:19	
Subject Sample	# 3	0.149	23:20	
Breath Volume		2.183	23:20	
Air Blank		0.000	23:21	
Control Test		0.079	23:21	
Air Blank		0.000	23:22	
Diagnostics Check		OK	23:22	

Data Download Date/Time: 06/25/2020 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HUGGINS,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: PEDRIANES

DATE/TIME OF TEST: 06/17/2020 00:41:37

Diagnostics Check	OK	00:38	
Air Blank	0.000	00:39	
Control Test	0.080	00:39	
Air Blank	0.000	00:40	
Subject Sample	# 1	00:40	Subject Test Refused
Breath Volume	0.000	00:40	
Air Blank	0.000	00:40	
Control Test	0.080	00:40	
Air Blank	0.000	00:41	
Diagnostics Check	OK	00:41	

Data Download Date/Time: 06/25/2020 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HUGGIS,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: PEDRIANES

DATE/TIME OF TEST: 06/17/2020 01:15:50

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.161	01:10
Breath Volume	2.746	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.155	01:14
Breath Volume	2.605	01:14
Air Blank	0.000	01:14
Control Test	0.079	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 06/25/2020 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: NIEVES

DATE/TIME OF TEST: 06/13/2020 21:47:29

Diagnostics Check	OK	21:39
Air Blank	0.000	21:40
Control Test	0.078	21:40
Air Blank	0.000	21:41
Subject Sample # 1	0.225	21:41
Breath Volume	1.242	21:41
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.222	21:45
Breath Volume	2.261	21:45
Air Blank	0.000	21:46
Control Test	0.078	21:46
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 06/25/2020 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: BERMUDEZ

DATE/TIME OF TEST: 06/17/2020 00:42:36

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:34	
Control Test	0.078	00:34	
Air Blank	0.000	00:35	
Subject Sample	# 1	00:38	No Sample Provided
Breath Volume	0.000	00:38	
Air Blank	0.000	00:38	
Air Blank	0.000	00:40	
Subject Sample	# 2	00:41	Subject Test Refused
Breath Volume	0.000	00:41	
Air Blank	0.000	00:41	
Control Test	0.078	00:41	
Air Blank	0.000	00:42	
Diagnostics Check	OK	00:42	

Data Download Date/Time: 06/25/2020 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: POTTS

DATE/TIME OF TEST: 06/01/2020 20:54:19

Diagnostics Check	OK	20:48	
Air Blank	0.000	20:49	
Control Test	0.079	20:49	
Air Blank	0.000	20:50	
Subject Sample	# 1	20:52	RFI Detect
Breath Volume	0.570	20:52	
Air Blank	0.000	20:53	
Control Test	0.080	20:53	
Air Blank	0.000	20:54	
Diagnostics Check	OK	20:54	

Data Download Date/Time: 06/05/2020 19:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: POTTS

DATE/TIME OF TEST: 06/01/2020 21:00:36

Diagnostics Check	OK	20:56	
Air Blank	0.000	20:56	
Control Test	0.079	20:56	
Air Blank	0.000	20:57	
Subject Sample	# 1	20:59	RFI Detect
Breath Volume	0.351	20:59	
Air Blank	0.000	20:59	
Control Test	0.080	21:00	
Air Blank	0.000	21:00	
Diagnostics Check	OK	21:00	

Data Download Date/Time: 06/05/2020 19:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DANID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: OTHER
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: POTTS

DATE/TIME OF TEST: 06/01/2020 23:19:32

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:08	
Control Test	0.080	23:08	
Air Blank	0.000	23:09	
Subject Sample	# 1	23:12	No Sample Provided
Breath Volume	0.000	23:12	
Air Blank	0.000	23:13	
Air Blank	0.000	23:15	
Subject Sample	# 2	23:18	No Sample Provided
Breath Volume	0.000	23:18	
Air Blank	0.000	23:18	
Control Test	0.077	23:18	
Air Blank	0.000	23:19	
Diagnostics Check	OK	23:19	

Data Download Date/Time: 06/05/2020 19:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: SYS CHECK
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: JETSON

DATE/TIME OF TEST: 06/05/2020 19:41:53

Diagnostics Check	OK	19:35
Air Blank	0.000	19:36
Control Test	0.079	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.000	19:37
Breath Volume	2.011	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.000	19:40
Breath Volume	1.964	19:40
Air Blank	0.000	19:40
Control Test	0.078	19:41
Air Blank	0.000	19:41
Diagnostics Check	OK	19:41

Data Download Date/Time: 06/22/2020 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HELLER,MICHAEL
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: HERASME

DATE/TIME OF TEST: 06/08/2020 03:30:26

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.079	03:24
Air Blank		0.000	03:24
Subject Sample	# 1	0.214	03:25
Breath Volume		1.500	03:25
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.194	03:28
Breath Volume		1.671	03:28
Air Blank		0.000	03:29
Control Test		0.079	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 06/22/2020 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: ROJAS REYES

DATE/TIME OF TEST: 06/11/2020 00:58:25

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:55	
Control Test	0.078	00:55	
Air Blank	0.000	00:56	
Subject Sample	# 1	00:56	RFI Detect
Breath Volume	0.867	00:56	
Air Blank	0.000	00:57	
Control Test	0.079	00:57	
Air Blank	0.000	00:58	
Diagnostics Check	OK	00:58	

Data Download Date/Time: 06/22/2020 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: ROJAS REYES

DATE/TIME OF TEST: 06/11/2020 01:03:23

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:00	
Control Test	0.078	01:00	
Air Blank	0.000	01:01	
Subject Sample	# 1	01:01	RFI Detect
Breath Volume	0.031	01:01	
Air Blank	0.000	01:02	
Control Test	0.079	01:02	
Air Blank	0.000	01:03	
Diagnostics Check	OK	01:03	

Data Download Date/Time: 06/22/2020 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 06/19/2020 22:01:12

Diagnostics Check	OK	21:55	
Air Blank	0.000	21:55	
Control Test	0.079	21:56	
Air Blank	0.000	21:56	
Subject Sample # 1	0.177	21:57	
Breath Volume	3.679	21:57	
Air Blank	0.000	21:57	
Air Blank	0.000	21:59	
		21:59	RFI Detect
Air Blank	0.000	22:00	
Control Test	0.079	22:00	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

Data Download Date/Time: 06/22/2020 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/13/2020 22:45:39

Diagnostics Check	OK	22:38
Air Blank	0.000	22:39
Control Test	0.079	22:39
Air Blank	0.000	22:40
Subject Sample # 1	0.193	22:40
Breath Volume	1.820	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.189	22:44
Breath Volume	2.546	22:44
Air Blank	0.000	22:44
Control Test	0.078	22:45
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 06/30/2020 05:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULAIMAN,RAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: DINKINS

DATE/TIME OF TEST: 06/18/2020 03:13:00

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.078	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.152	03:08
Breath Volume	1.218	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.143	03:11
Breath Volume	1.226	03:11
Air Blank	0.000	03:12
Control Test	0.079	03:12
Air Blank	0.000	03:12
Diagnostics Check	OK	03:13

Data Download Date/Time: 06/30/2020 05:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: CUMMINGS

DATE/TIME OF TEST: 06/22/2020 20:30:27

Diagnostics Check	OK	20:22
Air Blank	0.000	20:23
Control Test	0.077	20:23
Air Blank	0.000	20:24
Subject Sample # 1	0.007	20:24
Breath Volume	1.660	20:24
Air Blank	0.000	20:25
Air Blank	0.000	20:27
Subject Sample # 2	0.007	20:29
Breath Volume	1.976	20:29
Air Blank	0.000	20:29
Control Test	0.079	20:29
Air Blank	0.000	20:30
Diagnostics Check	OK	20:30

Data Download Date/Time: 06/30/2020 05:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTAYA, GUILLERMO,
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: CARTAYA, GUILLERMO
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: BUI
ARREST TIME: 14:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:21

SUBJECT'S LAST NAME: ALAYON RAMOS

DATE/TIME OF TEST: 06/14/2020 15:25:41

Diagnostics Check	OK	15:19
Air Blank	0.000	15:20
Control Test	0.079	15:20
Air Blank	0.000	15:20
Subject Sample # 1	0.015	15:21
Breath Volume	3.230	15:21
Air Blank	0.000	15:22
Air Blank	0.000	15:23
Subject Sample # 2	0.011	15:24
Breath Volume	2.789	15:24
Air Blank	0.000	15:24
Control Test	0.077	15:25
Air Blank	0.000	15:25
Diagnostics Check	OK	15:25

Data Download Date/Time: 06/25/2020 14:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 06/17/2020 00:24:32

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:21	
Control Test	0.080	00:21	
Air Blank	0.000	00:22	
Subject Sample	# 1 0.188	00:22	
Breath Volume	4.386	00:22	
Air Blank		00:23	Purge Fail
Air Blank		00:24	Purge Fail

Data Download Date/Time: 07/02/2020 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 06/17/2020 00:33:11

Diagnostics Check		OK	00:26
Air Blank		0.000	00:26
Control Test		0.079	00:27
Air Blank		0.000	00:27
Subject Sample	# 1	0.184	00:28
Breath Volume		4.304	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:30
Subject Sample	# 2	0.175	00:31
Breath Volume		4.507	00:31
Air Blank		0.000	00:32
Control Test		0.078	00:32
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 07/02/2020 16:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: WUNSCH

DATE/TIME OF TEST: 06/19/2020 22:20:52

Diagnostics Check	OK	22:17	
Air Blank	0.000	22:17	
Control Test	0.079	22:18	
Air Blank	0.000	22:18	
Subject Sample	# 1 0.172	22:19	
Breath Volume	3.250	22:19	
Air Blank		22:20	Purge Fail
Air Blank		22:20	Purge Fail

Data Download Date/Time: 07/02/2020 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: WUNSCH

DATE/TIME OF TEST: 06/19/2020 22:31:37

Diagnostics Check	OK	22:23	
Air Blank	0.000	22:23	
Control Test	0.079	22:24	
Air Blank	0.000	22:24	
Subject Sample # 1	0.154	22:25	
Breath Volume	2.777	22:25	
Air Blank	0.000	22:26	
Air Blank	0.000	22:28	
Subject Sample # 2	0.155	22:30	
Breath Volume	3.531	22:30	
Air Blank	0.000	22:30	
Control Test	0.080	22:31	Interferent Detect
Air Blank	0.000	22:31	

Data Download Date/Time: 07/02/2020 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: LENTINE

DATE/TIME OF TEST: 06/20/2020 20:32:12

Diagnostics Check	OK	20:24	
Air Blank	0.000	20:25	
Control Test	0.079	20:25	
Air Blank	0.000	20:26	
Subject Sample	# 1 0.209	20:26	
Breath Volume	1.496	20:26	
Air Blank	0.000	20:27	
Air Blank	0.000	20:29	
Subject Sample	# 2 0.211	20:30	
Breath Volume	1.968	20:30	
Air Blank	0.000	20:30	
Control Test	0.078	20:31	
Air Blank	0.000	20:31	
Diagnostics Check	Fail	20:31	Voltage/Current Test
Air Blank	0.000	20:32	

Data Download Date/Time: 07/02/2020 16:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: LENTINE

DATE/TIME OF TEST: 06/20/2020 20:40:34

Diagnostics Check	OK	20:33
Air Blank	0.000	20:34
Control Test	0.078	20:34
Air Blank	0.000	20:35
Subject Sample # 1	0.213	20:35
Breath Volume	2.257	20:35
Air Blank	0.000	20:36
Air Blank	0.000	20:38
Subject Sample # 2	0.213	20:38
Breath Volume	1.820	20:38
Air Blank	0.000	20:39
Control Test	0.076	20:39
Air Blank	0.000	20:40
Diagnostics Check	OK	20:40

Data Download Date/Time: 07/02/2020 16:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: MILTON

DATE/TIME OF TEST: 06/21/2020 21:45:40

Diagnostics Check	OK	21:44	
Air Blank		21:44	Ambient Fail
Air Blank		21:45	Purge Fail

Data Download Date/Time: 07/02/2020 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: MILTON

DATE/TIME OF TEST: 06/21/2020 21:59:18

Diagnostics Check	OK	21:47	
Air Blank	0.000	21:48	
Control Test	0.081	21:48	
Air Blank	0.000	21:49	
Subject Sample	# 1	21:52	No Sample Provided
Breath Volume	0.000	21:52	
Air Blank	0.000	21:52	
Air Blank	0.000	21:54	
Subject Sample	# 2	21:57	No Sample Provided
Breath Volume	0.000	21:57	
Air Blank	0.000	21:58	
Control Test	0.079	21:58	
Air Blank	0.000	21:59	
Diagnostics Check	OK	21:59	

Data Download Date/Time: 07/02/2020 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 06/25/2020 23:45:56

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.079	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.099	23:41
Breath Volume	1.234	23:41
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.103	23:44
Breath Volume	2.625	23:44
Air Blank	0.000	23:44
Control Test	0.080	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 07/02/2020 16:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:29

SUBJECT'S LAST NAME: TEBOR

DATE/TIME OF TEST: 06/26/2020 19:12:51

Diagnostics Check	OK	19:04	
Air Blank	0.000	19:04	
Control Test	0.078	19:05	
Air Blank	0.000	19:05	
Subject Sample	# 1	19:08	No Sample Provided
Breath Volume	0.000	19:08	
Air Blank	0.000	19:09	
Air Blank	0.000	19:11	
Subject Sample	# 2	19:11	Subject Test Refused
Breath Volume	0.000	19:11	
Air Blank	0.000	19:11	
Control Test	0.078	19:12	
Air Blank	0.000	19:12	
Diagnostics Check	OK	19:12	

Data Download Date/Time: 07/02/2020 16:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: TODD

DATE/TIME OF TEST: 06/27/2020 01:25:16

Diagnostics Check	OK	01:18
Air Blank	0.000	01:19
Control Test	0.079	01:19
Air Blank	0.000	01:20
Subject Sample # 1	0.221	01:20
Breath Volume	3.335	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:22
Subject Sample # 2	0.225	01:23
Breath Volume	3.738	01:23
Air Blank	0.000	01:24
Control Test	0.078	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 07/02/2020 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: GASPAR FRANCISCO

DATE/TIME OF TEST: 06/30/2020 00:06:11

Diagnostics Check	OK	23:59	
Air Blank	0.000	00:00	
Control Test	0.079	00:00	
Air Blank	0.000	00:01	
Subject Sample # 1	0.194	00:01	
Breath Volume	1.570	00:01	
Air Blank	0.000	00:02	
Air Blank	0.000	00:04	
Subject Sample # 2	0.188	00:04	
Breath Volume	2.023	00:04	
Air Blank	0.000	00:05	
Control Test	0.077	00:05	RFI Detect
Air Blank	0.000	00:06	

Data Download Date/Time: 07/02/2020 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: GASPAR FRANCISCO

DATE/TIME OF TEST: 06/30/2020 00:15:55

Diagnostics Check	OK	00:09	
Air Blank	0.000	00:09	
Control Test	0.077	00:09	
Air Blank	0.000	00:10	
Subject Sample # 1	0.191	00:11	
Breath Volume	2.035	00:11	
Air Blank	0.000	00:11	
Air Blank	0.000	00:13	
Subject Sample # 2	0.178	00:13	
Breath Volume	1.316	00:13	
Air Blank	0.000	00:15	
Control Test	0.080	00:15	Interferent Detect
Air Blank	0.000	00:15	

Data Download Date/Time: 07/02/2020 16:54