

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CASON,CHRISTOPHER,N

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: AMAR,G

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13518080A5

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 23:03

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: HASKINS

DATE/TIME OF TEST: 06/01/2020 00:22:10

---

Diagnostics Check		OK	00:14
Air Blank		0.000	00:14
Control Test		0.076	00:14
Air Blank		0.000	00:15
Subject Sample	# 1	0.107	00:16
Breath Volume		1.355	00:16
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.101	00:20
Breath Volume		1.144	00:20
Air Blank		0.000	00:21
Control Test		0.077	00:21
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

---

Data Download Date/Time: 06/03/2020 23:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FLOYD,DANIEL,Z  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: BROWNING,JON  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: BUI  
ARREST TIME: 20:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: FLOM

DATE/TIME OF TEST: 06/01/2020 22:13:48

---

Diagnostics Check	OK	22:10	
Air Blank	0.000	22:11	
Control Test	0.076	22:11	
Air Blank	0.000	22:12	
Subject Sample	# 1	22:12	Subject Test Refused
Breath Volume	0.000	22:12	
Air Blank	0.000	22:12	
Control Test	0.077	22:13	
Air Blank	0.000	22:13	
Diagnostics Check	OK	22:13	

---

Data Download Date/Time: 06/03/2020 23:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH,DEVIN,A

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: WATTS,R

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13518080A5

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 23:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: DOSS

DATE/TIME OF TEST: 06/06/2020 00:08:57

---

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.075	00:03
Air Blank	0.000	00:03
Subject Sample # 1	0.000	00:04
Breath Volume	2.171	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.000	00:07
Breath Volume	1.832	00:07
Air Blank	0.000	00:08
Control Test	0.076	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

---

Data Download Date/Time: 08/04/2020 14:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEVORAK, MATTHEW, J  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: HOLTZHOUSE, J  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 15:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/06/2020 16:34:16

---

Diagnostics Check	OK	16:26	
Air Blank	0.000	16:27	
Control Test	0.075	16:27	
Air Blank	0.000	16:27	
Subject Sample # 1	0.128	16:29	
Breath Volume	1.742	16:29	
Air Blank	0.000	16:30	
Air Blank	0.000	16:32	
Subject Sample # 2	0.117	16:32	
Breath Volume	1.921	16:32	
Air Blank	0.000	16:33	
Control Test	0.074	16:33	Control Outside Tolerance
Air Blank	0.000	16:34	

---

Data Download Date/Time: 08/04/2020 14:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ORTIZ,BILLY,

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: ORTIZ,BILLY

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13518080A5

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 22:23

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: SCHENCK

DATE/TIME OF TEST: 06/06/2020 23:47:06

---

Diagnostics Check		OK	23:40
Air Blank		0.000	23:41
Control Test		0.076	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.214	23:42
Breath Volume		2.527	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:44
Subject Sample	# 2	0.219	23:45
Breath Volume		3.371	23:45
Air Blank		0.000	23:46
Control Test		0.075	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

---

Data Download Date/Time: 08/04/2020 14:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HELM,DEREK,W  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HELM,DEREK  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: WOOD  
DATE/TIME OF TEST: 06/13/2020 01:02:02

---

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.076	00:55
Air Blank	0.000	00:56
Subject Sample # 1	0.179	00:57
Breath Volume	1.144	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.181	01:00
Breath Volume	1.613	01:00
Air Blank	0.000	01:01
Control Test	0.076	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:02

---

Data Download Date/Time: 08/04/2020 14:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: POLK, BRANDON, A

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: BOSWELL, DEPUTY

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13518080A5

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: OTHER

ARREST TIME: 20:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 06/19/2020 20:45:48

---

Diagnostics Check	OK	20:39
Air Blank	0.000	20:39
Control Test	0.076	20:40
Air Blank	0.000	20:40
Subject Sample # 1	0.067	20:41
Breath Volume	2.820	20:41
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.066	20:44
Breath Volume	2.421	20:44
Air Blank	0.000	20:44
Control Test	0.076	20:45
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

---

Data Download Date/Time: 08/04/2020 14:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVENPORT,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: DAVENPORT,JOSHUA  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/20/2020 02:39:46

---

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:35	
Control Test	0.076	02:35	
Air Blank	0.000	02:35	
Subject Sample	# 1 0.140	02:37	
Breath Volume	1.902	02:37	
Air Blank		02:38	RFI Detect
Air Blank	0.000	02:38	
Control Test	0.078	02:39	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:39	

---

Data Download Date/Time: 08/04/2020 14:26



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVENPORT, JOSHUA, J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: DAVENPORT, JOSHUA  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/20/2020 02:48:59

---

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.075	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.133	02:44
Breath Volume	2.128	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:46
Subject Sample # 2	0.133	02:47
Breath Volume	2.113	02:47
Air Blank	0.000	02:48
Control Test	0.076	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

---

Data Download Date/Time: 08/04/2020 14:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH,DEVIN,A

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: HOWARD,R

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13518080A5

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 18:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: BALLOU

DATE/TIME OF TEST: 06/22/2020 19:34:35

---

Diagnostics Check	OK	19:27
Air Blank	0.000	19:27
Control Test	0.075	19:27
Air Blank	0.000	19:28
Subject Sample # 1	0.206	19:29
Breath Volume	2.476	19:29
Air Blank	0.000	19:30
Air Blank	0.000	19:31
Subject Sample # 2	0.209	19:33
Breath Volume	2.109	19:33
Air Blank	0.000	19:33
Control Test	0.075	19:34
Air Blank	0.000	19:34
Diagnostics Check	OK	19:34

---

Data Download Date/Time: 08/04/2020 14:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH,DEVIN,A

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: HORTON,T

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13518080A5

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 19:08

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: SILCOTT

DATE/TIME OF TEST: 06/27/2020 19:51:39

---

Diagnostics Check	OK	19:48	
Air Blank	0.000	19:49	
Control Test	0.076	19:49	
Air Blank	0.000	19:50	
Subject Sample	# 1	19:50	Subject Test Refused
Breath Volume	0.000	19:50	
Air Blank	0.000	19:50	
Control Test	0.076	19:51	
Air Blank	0.000	19:51	
Diagnostics Check	OK	19:51	

---

Data Download Date/Time: 08/04/2020 14:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOWERMASTER,MARY,E  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: JOHNSON,DALE  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13518080A5  
DRY GAS EXPIRY DATE: 08/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 17:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:47

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 06/29/2020 18:20:33

---

Diagnostics Check	OK	18:14
Air Blank	0.000	18:14
Control Test	0.075	18:15
Air Blank	0.000	18:15
Subject Sample # 1	0.132	18:16
Breath Volume	1.316	18:16
Air Blank	0.000	18:16
Air Blank	0.000	18:18
Subject Sample # 2	0.134	18:19
Breath Volume	1.949	18:19
Air Blank	0.000	18:19
Control Test	0.076	18:19
Air Blank	0.000	18:20
Diagnostics Check	OK	18:20

---

Data Download Date/Time: 08/04/2020 14:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BARRETT,CHASE,A

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: CARPENTER,T

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13518080A5

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 20:42

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: KOWALIK

DATE/TIME OF TEST: 06/29/2020 21:30:20

---

Diagnostics Check	OK	21:27	
Air Blank	0.000	21:28	
Control Test	0.075	21:28	
Air Blank	0.000	21:28	
Subject Sample	# 1	21:28	Subject Test Refused
Breath Volume	0.000	21:28	
Air Blank	0.000	21:29	
Control Test	0.076	21:29	
Air Blank	0.000	21:30	
Diagnostics Check	OK	21:30	

---

Data Download Date/Time: 08/04/2020 14:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BARRETT,CHASE,A

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: ROBINSON,RANDY

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13518080A5

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 21:17

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 06/29/2020 22:21:21

---

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:15	
Control Test	0.075	22:15	
Air Blank	0.000	22:16	
Subject Sample	# 1 0.124	22:16	
Breath Volume	2.820	22:16	
Air Blank	0.000	22:17	
Air Blank	0.000	22:19	
Subject Sample	# 2 0.130	22:19	
Breath Volume	2.007	22:19	
Air Blank	0.000	22:20	
Control Test	0.074	22:20	Control Outside Tolerance
Air Blank	0.000	22:21	

---

Data Download Date/Time: 08/04/2020 14:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURINGTON,PATRICK,A  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: PURINGTON,PATRICK  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: DEES

DATE/TIME OF TEST: 06/02/2020 01:13:18

---

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:10	
Control Test	0.079	01:11	
Air Blank	0.000	01:11	
Subject Sample	# 1	01:11	Subject Test Refused
Breath Volume	0.000	01:11	
Air Blank	0.000	01:12	
Control Test	0.079	01:12	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

---

Data Download Date/Time: 06/29/2020 10:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TELLE,MADDISON,J  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: TELLE,MADDISON  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 06/02/2020 01:39:56

---

Diagnostics Check	OK	01:39	
Air Blank		01:39	RFI Detect
Air Blank		01:39	RFI Detect

---

Data Download Date/Time: 06/29/2020 10:56



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TELLE,MADDISON,J  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: TELLE,MADDISON  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 06/02/2020 01:44:49

---

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:42	
Control Test	0.080	01:42	
Air Blank	0.000	01:43	
Subject Sample	# 1	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:43	
Control Test	0.079	01:44	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

---

Data Download Date/Time: 06/29/2020 10:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TELLE,MADDISON,J  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: MCVEIGH,TREVOR  
ARRESTING OFFICER AGENCY: Alachua PD  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BESCHEN

DATE/TIME OF TEST: 06/03/2020 00:54:02

---

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.079	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.115	00:48
Breath Volume	1.281	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.114	00:52
Breath Volume	1.390	00:52
Air Blank	0.000	00:53
Control Test	0.077	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:54

---

Data Download Date/Time: 06/29/2020 10:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CABRAL,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: CABRAL,JONATHAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: COURTNEY

DATE/TIME OF TEST: 06/06/2020 02:34:16

---

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.080	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.000	02:29
Breath Volume	3.652	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.000	02:32
Breath Volume	2.480	02:32
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

---

Data Download Date/Time: 06/29/2020 10:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CABRAL,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: CABRAL,JONATHAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 06/11/2020 00:29:41

---

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.079	00:21
Air Blank	0.000	00:22
Subject Sample # 1	0.411	00:24
Breath Volume	3.027	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.429	00:28
Breath Volume	3.343	00:28
Air Blank	0.000	00:28
Control Test	0.079	00:29
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

---

Data Download Date/Time: 06/29/2020 10:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRICKER,JEFFREY,S  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: SIMS,CHRIS  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: MCKAY

DATE/TIME OF TEST: 06/13/2020 01:17:56

---

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.079	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.189	01:10
Breath Volume	2.652	01:10
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.164	01:13
Breath Volume	1.273	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 3	0.181	01:16
Breath Volume	3.484	01:16
Air Blank	0.000	01:17
Control Test	0.077	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

---

Data Download Date/Time: 06/29/2020 10:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURINGTON,PATRICK,A  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: MYERS,BENJAMIN  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 06/16/2020 23:21:33

---

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.079	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.124	23:17
Breath Volume	1.117	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.118	23:19
Breath Volume	1.324	23:19
Air Blank	0.000	23:20
Control Test	0.077	23:20
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

---

Data Download Date/Time: 06/29/2020 10:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LITZKOW,CLAYTON,P  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: LITZKOW,CLAYTON  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 23:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 06/18/2020 00:24:05

---

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:13	
Control Test	0.080	00:13	
Air Blank	0.000	00:13	
Subject Sample	# 1	00:17	No Sample Provided
Breath Volume	0.000	00:17	
Air Blank	0.000	00:17	
Air Blank	0.000	00:19	
Subject Sample	# 2	00:22	No Sample Provided
Breath Volume	0.000	00:22	
Air Blank	0.000	00:23	
Control Test	0.081	00:23	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

---

Data Download Date/Time: 06/29/2020 10:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURINGTON,PATRICK,A  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: PURINGTON,PATRICK  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 06/20/2020 00:03:23

---

Diagnostics Check	OK	23:54
Air Blank	0.000	23:55
Control Test	0.080	23:55
Air Blank	0.000	23:55
Subject Sample # 2	0.161	23:58
Breath Volume	2.035	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.147	00:01
Breath Volume	2.089	00:01
Air Blank	0.000	00:02
Control Test	0.076	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

---

Data Download Date/Time: 06/29/2020 10:56



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURINGTON,PATRICK,A  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: PURINGTON,PATRICK  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 06/21/2020 00:01:50

---

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:59	
Control Test	0.080	23:59	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:00	Subject Test Refused
Breath Volume	0.000	00:00	
Air Blank	0.000	00:00	
Control Test	0.080	00:01	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

---

Data Download Date/Time: 06/29/2020 10:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,JACK,W  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: MILLER,JACK  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 12:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:38

SUBJECT'S LAST NAME: RISNER

DATE/TIME OF TEST: 06/23/2020 13:13:02

---

Diagnostics Check	OK	13:06
Air Blank	0.000	13:07
Control Test	0.080	13:07
Air Blank	0.000	13:08
Subject Sample # 1	0.000	13:08
Breath Volume	2.816	13:08
Air Blank	0.000	13:09
Air Blank	0.000	13:11
Subject Sample # 2	0.000	13:11
Breath Volume	1.265	13:11
Air Blank	0.000	13:12
Control Test	0.077	13:12
Air Blank	0.000	13:12
Diagnostics Check	OK	13:13

---

Data Download Date/Time: 06/29/2020 10:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: HOWARD,ANTHONY  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 14:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:49

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 06/23/2020 15:15:36

---

Diagnostics Check	OK	15:09
Air Blank	0.000	15:09
Control Test	0.080	15:10
Air Blank	0.000	15:10
Subject Sample # 1	0.245	15:11
Breath Volume	2.363	15:11
Air Blank	0.000	15:11
Air Blank	0.000	15:13
Subject Sample # 2	0.243	15:13
Breath Volume	2.652	15:13
Air Blank	0.000	15:14
Control Test	0.078	15:14
Air Blank	0.000	15:15
Diagnostics Check	OK	15:15

---

Data Download Date/Time: 06/29/2020 10:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DILLON,JUSTIN,P  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: DILLON,JUSTIN  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: CELESTIN

DATE/TIME OF TEST: 06/26/2020 05:24:19

---

Diagnostics Check		OK	05:12	
Air Blank		0.000	05:13	
Control Test		0.079	05:13	
Air Blank		0.000	05:13	
Subject Sample	# 1	0.142	05:17	Volume Not Met
Breath Volume		0.214	05:17	
Air Blank		0.000	05:17	
Air Blank		0.000	05:19	
Subject Sample	# 2	0.137	05:22	Volume Not Met
Breath Volume		0.257	05:22	
Air Blank		0.000	05:23	
Control Test		0.078	05:23	
Air Blank		0.000	05:24	
Diagnostics Check		OK	05:24	

---

Data Download Date/Time: 06/29/2020 10:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CABRAL,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: CABRAL,JONATHAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: FORTNER

DATE/TIME OF TEST: 06/30/2020 00:01:20

---

Diagnostics Check	OK	23:49	
Air Blank	0.000	23:50	
Control Test	0.079	23:50	
Air Blank	0.000	23:51	
Subject Sample	# 1	23:54	No Sample Provided
Breath Volume	0.000	23:54	
Air Blank	0.000	23:54	
Air Blank	0.000	23:56	
Subject Sample	# 2	23:59	No Sample Provided
Breath Volume	0.000	23:59	
Air Blank	0.000	00:00	
Control Test	0.081	00:00	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

---

Data Download Date/Time: 07/20/2020 13:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DILLON,JUSTIN,P  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: ADKINS,MICHAEL  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 12:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 06/30/2020 02:29:52

---

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.141	02:24
Breath Volume	1.875	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.148	02:28
Breath Volume	2.160	02:28
Air Blank	0.000	02:28
Control Test	0.081	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

---

Data Download Date/Time: 07/20/2020 14:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: USECHE,IVAN,D  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: MYERS,BENJAMIN  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: PELLETIER

DATE/TIME OF TEST: 06/30/2020 22:44:40

---

Diagnostics Check	Fail	22:44	Voltage/Current Test
Air Blank	0.000	22:44	

---

Data Download Date/Time: 07/20/2020 14:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: USECHE,IVAN,D  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: MYERS,BENJAMIN  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24818080A2  
DRY GAS EXPIRY DATE: 10/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 21:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: PELLETIER

DATE/TIME OF TEST: 06/30/2020 22:55:01

---

Diagnostics Check	OK	22:47
Air Blank	0.000	22:48
Control Test	0.079	22:48
Air Blank	0.000	22:48
Subject Sample # 1	0.172	22:49
Breath Volume	1.890	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:51
Subject Sample # 2	0.181	22:53
Breath Volume	1.964	22:53
Air Blank	0.000	22:54
Control Test	0.078	22:54
Air Blank	0.000	22:54
Diagnostics Check	OK	22:55

---

Data Download Date/Time: 07/20/2020 14:07



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SEARS,TODD,A  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: CRIPE,ANNALISA  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 06/06/2020 02:26:55

---

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.081	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.147	02:22
Breath Volume	1.269	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:24
Subject Sample # 2	0.146	02:25
Breath Volume	2.648	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

---

Data Download Date/Time: 06/29/2020 14:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANDREW, MATTHEW, A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: ANDREW, MATTHEW  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: LUNDY

DATE/TIME OF TEST: 06/07/2020 04:10:36

---

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.081	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.000	04:06
Breath Volume	2.808	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.000	04:09
Breath Volume	1.460	04:09
Air Blank	0.000	04:09
Control Test	0.080	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

---

Data Download Date/Time: 06/29/2020 14:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PETERS,MICHAEL,E  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: PETERS,MICHAEL  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: BRADEN

DATE/TIME OF TEST: 06/10/2020 21:09:55

---

Diagnostics Check	OK	20:56	
Air Blank	0.000	20:56	
Control Test	0.081	20:56	
Air Blank	0.000	20:57	
Subject Sample	# 1 0.162	21:00	
Breath Volume	3.125	21:00	
Air Blank	0.000	21:00	
Air Blank	0.000	21:02	
Subject Sample	# 2 0.152	21:05	Volume Not Met
Breath Volume	0.996	21:05	
Air Blank	0.000	21:06	
Air Blank	0.000	21:08	
Subject Sample	# 3	21:08	Subject Test Refused
Breath Volume	0.000	21:08	
Air Blank	0.000	21:09	
Control Test	0.079	21:09	
Air Blank	0.000	21:09	
Diagnostics Check	OK	21:09	

---

Data Download Date/Time: 06/29/2020 14:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIMBRIL,JOSHUA,B  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: KIMBRIL,JOSHUA  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: LABOY

DATE/TIME OF TEST: 06/12/2020 00:13:10

---

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.080	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.000	00:07
Breath Volume	3.691	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.000	00:11
Breath Volume	3.773	00:11
Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

---

Data Download Date/Time: 06/29/2020 14:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIMBRIL,JOSHUA,B  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: KIMBRIL,JOSHUA  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: BUCKLEY

DATE/TIME OF TEST: 06/12/2020 02:55:12

---

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.081	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.000	02:50
Breath Volume	1.917	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.000	02:53
Breath Volume	2.769	02:53
Air Blank	0.000	02:54
Control Test	0.081	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

---

Data Download Date/Time: 06/29/2020 14:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIMBRIL,JOSHUA,B  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: KIMBRIL,JOSHUA  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 06/13/2020 01:11:02

---

Diagnostics Check	OK	00:59	
Air Blank	0.000	01:00	
Control Test	0.080	01:00	
Air Blank	0.000	01:00	
Subject Sample	# 1	01:04	No Sample Provided
Breath Volume	0.000	01:04	
Air Blank	0.000	01:04	
Air Blank	0.000	01:06	
Subject Sample	# 2	01:09	No Sample Provided
Breath Volume	0.000	01:09	
Air Blank	0.000	01:10	
Control Test	0.080	01:10	
Air Blank	0.000	01:10	
Diagnostics Check	OK	01:11	

---

Data Download Date/Time: 06/29/2020 14:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAHR,DANIEL,E  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: HAHR,DANIEL  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: HOFFMANN

DATE/TIME OF TEST: 06/15/2020 00:34:45

---

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.080	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.000	00:30
Breath Volume	2.207	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.000	00:33
Breath Volume	1.867	00:33
Air Blank	0.000	00:33
Control Test	0.080	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

---

Data Download Date/Time: 06/29/2020 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURGESS,MASON,W  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: BURGESS,MASON  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: KENYON

DATE/TIME OF TEST: 06/18/2020 23:15:15

---

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:11	
Control Test	0.081	23:12	
Air Blank	0.000	23:12	
Subject Sample	# 1	23:13	RFI Detect
Breath Volume	0.000	23:13	
Air Blank	0.000	23:14	
Control Test	0.081	23:14	
Air Blank	0.000	23:15	
Diagnostics Check	OK	23:15	

---

Data Download Date/Time: 06/29/2020 14:10



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURGESS,MASON,W  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: BURGESS,MASON  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: KENYON

DATE/TIME OF TEST: 06/18/2020 23:28:49

---

Diagnostics Check	OK	23:17	
Air Blank	0.000	23:17	
Control Test	0.081	23:18	
Air Blank	0.000	23:18	
Subject Sample	# 1	23:21	No Sample Provided
Breath Volume	0.000	23:21	
Air Blank	0.000	23:22	
Air Blank	0.000	23:24	
Subject Sample	# 2	23:27	No Sample Provided
Breath Volume	0.000	23:27	
Air Blank	0.000	23:27	
Control Test	0.080	23:28	
Air Blank	0.000	23:28	
Diagnostics Check	OK	23:28	

---

Data Download Date/Time: 06/29/2020 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUS,CHRISTINA,J  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: DUS,CHRISTINA  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: WISEMAN

DATE/TIME OF TEST: 06/21/2020 06:27:06

---

Diagnostics Check	OK	06:20
Air Blank	0.000	06:21
Control Test	0.079	06:21
Air Blank	0.000	06:22
Subject Sample # 1	0.094	06:22
Breath Volume	3.824	06:22
Air Blank	0.000	06:23
Air Blank	0.000	06:25
Subject Sample # 2	0.098	06:25
Breath Volume	2.136	06:25
Air Blank	0.000	06:26
Control Test	0.080	06:26
Air Blank	0.000	06:27
Diagnostics Check	OK	06:27

---

Data Download Date/Time: 06/29/2020 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIMBRIL,JOSHUA,B  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: STEINER,ZACHARY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 08:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:21

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/24/2020 09:50:03

---

Diagnostics Check	OK	09:41	
Air Blank	0.000	09:41	
Control Test	0.080	09:42	
Air Blank	0.000	09:42	
Subject Sample	# 1	09:45	No Sample Provided
Breath Volume	0.000	09:45	
Air Blank	0.000	09:46	
Air Blank	0.000	09:48	
Subject Sample	# 2	09:48	Subject Test Refused
Breath Volume	0.000	09:48	
Air Blank	0.000	09:49	
Control Test	0.080	09:49	
Air Blank	0.000	09:49	
Diagnostics Check	OK	09:50	

---

Data Download Date/Time: 06/29/2020 14:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLONNA, TYLER, S  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: SPRAGUE, JONATHON  
ARRESTING OFFICER AGENCY: Crestview PD  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 06/27/2020 00:58:34

---

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.181	00:53
Breath Volume	4.253	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.172	00:56
Breath Volume	2.343	00:56
Air Blank	0.000	00:57
Control Test	0.078	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

---

Data Download Date/Time: 06/29/2020 14:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COOK,RUSSELL,L  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: COOK,RUSSELL  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: JASPER

DATE/TIME OF TEST: 06/28/2020 01:03:57

---

Diagnostics Check	OK	00:56
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.156	00:59
Breath Volume	1.628	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:01
Subject Sample # 2	0.151	01:02
Breath Volume	3.027	01:02
Air Blank	0.000	01:03
Control Test	0.076	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

---

Data Download Date/Time: 06/29/2020 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNETT,JAMES,C  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: BARNETT,JAMES  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 05/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: RASH

DATE/TIME OF TEST: 06/06/2020 21:41:41

---

Diagnostics Check	OK	21:34
Air Blank	0.000	21:35
Control Test	0.082	21:35
Air Blank	0.000	21:36
Subject Sample # 1	0.144	21:36
Breath Volume	1.878	21:36
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.140	21:40
Breath Volume	1.687	21:40
Air Blank	0.000	21:40
Control Test	0.081	21:41
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

---

Data Download Date/Time: 06/19/2020 16:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNETT,JAMES,C  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: BARNETT,JAMES  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 05/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: MYRICK

DATE/TIME OF TEST: 06/08/2020 00:59:40

---

Diagnostics Check	OK	00:52
Air Blank	0.000	00:53
Control Test	0.082	00:53
Air Blank	0.000	00:54
Subject Sample # 1	0.219	00:54
Breath Volume	1.835	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.209	00:58
Breath Volume	1.113	00:58
Air Blank	0.000	00:58
Control Test	0.082	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

---

Data Download Date/Time: 06/19/2020 16:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYNCH,CHAD,J  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: LYNCH,CHAD  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 06/08/2020 21:52:30

---

Diagnostics Check		OK	21:35	
Air Blank		0.000	21:36	
Control Test		0.082	21:36	
Air Blank		0.000	21:37	
Subject Sample	# 1	0.260	21:40	Volume Not Met
Breath Volume		0.664	21:40	
Air Blank		0.000	21:41	
Air Blank		0.000	21:42	
Subject Sample	# 2	0.282	21:44	
Breath Volume		1.296	21:44	
Air Blank		0.000	21:45	
Air Blank		0.000	21:47	
Subject Sample	# 3	0.266	21:50	Volume Not Met
Breath Volume		0.363	21:50	
Air Blank		0.000	21:51	
Control Test		0.079	21:51	
Air Blank		0.000	21:52	
Diagnostics Check		OK	21:52	

---

Data Download Date/Time: 06/19/2020 16:58



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEEKS, JEREMIAH, S  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: MEEKS, JEREMIAH  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 05/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MEJIA CORDOVA

DATE/TIME OF TEST: 06/10/2020 02:38:45

---

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.081	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.190	02:34
Breath Volume	3.289	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.180	02:36
Breath Volume	3.207	02:36
Air Blank	0.000	02:37
Control Test	0.079	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

---

Data Download Date/Time: 06/19/2020 16:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINH,TAM,T  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: GUSTER,CORY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 06/13/2020 17:47:05

---

Diagnostics Check	OK	17:40
Air Blank	0.000	17:41
Control Test	0.082	17:41
Air Blank	0.000	17:42
Subject Sample # 1	0.190	17:42
Breath Volume	2.761	17:42
Air Blank	0.000	17:43
Air Blank	0.000	17:45
Subject Sample # 2	0.183	17:45
Breath Volume	2.820	17:45
Air Blank	0.000	17:46
Control Test	0.081	17:46
Air Blank	0.000	17:47
Diagnostics Check	OK	17:47

---

Data Download Date/Time: 06/19/2020 16:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANNING,CARL,T  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: MANNING,CARL  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: FINNEY

DATE/TIME OF TEST: 06/14/2020 02:49:01

---

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.082	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.126	02:44
Breath Volume	3.214	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:46
Subject Sample # 2	0.132	02:47
Breath Volume	3.746	02:47
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:49

---

Data Download Date/Time: 06/19/2020 16:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYNCH,CHAD,J  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: LYNCH,CHAD  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 05/18/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 14:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 06/14/2020 15:07:30

---

Diagnostics Check	OK	15:00
Air Blank	0.000	15:00
Control Test	0.082	15:01
Air Blank	0.000	15:01
Subject Sample # 1	0.008	15:02
Breath Volume	3.312	15:02
Air Blank	0.000	15:03
Air Blank	0.000	15:05
Subject Sample # 2	0.008	15:06
Breath Volume	3.246	15:06
Air Blank	0.000	15:06
Control Test	0.081	15:06
Air Blank	0.000	15:07
Diagnostics Check	OK	15:07

---

Data Download Date/Time: 06/19/2020 16:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNETT,JAMES,C  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: BARNETT,JAMES  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 06/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1045940  
DRY GAS EXPIRY DATE: 11/13/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: BEECROFT

DATE/TIME OF TEST: 06/15/2020 22:27:06

---

Diagnostics Check	OK	22:19
Air Blank	0.000	22:20
Control Test	0.082	22:20
Air Blank	0.000	22:21
Subject Sample # 1	0.260	22:22
Breath Volume	1.414	22:22
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.245	22:25
Breath Volume	1.488	22:25
Air Blank	0.000	22:26
Control Test	0.081	22:26
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

---

Data Download Date/Time: 07/13/2020 10:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: SHELLEY,SAMUEL  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 06/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1045940  
DRY GAS EXPIRY DATE: 11/13/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/19/2020 21:58:08

---

Diagnostics Check	OK	21:51
Air Blank	0.000	21:51
Control Test	0.082	21:52
Air Blank	0.000	21:52
Subject Sample # 1	0.337	21:53
Breath Volume	2.726	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:55
Subject Sample # 2	0.331	21:56
Breath Volume	2.796	21:56
Air Blank	0.000	21:57
Control Test	0.081	21:57
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

---

Data Download Date/Time: 07/13/2020 10:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEITE,STEVEN,G  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: LEITE,STEVEN  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 06/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1045940  
DRY GAS EXPIRY DATE: 11/13/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: HASTINGS

DATE/TIME OF TEST: 06/22/2020 02:10:05

---

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.083	02:04
Air Blank	0.000	02:04
Subject Sample # 1	0.000	02:05
Breath Volume	1.417	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.000	02:08
Breath Volume	1.390	02:08
Air Blank	0.000	02:09
Control Test	0.081	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:10

---

Data Download Date/Time: 07/13/2020 10:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCLELLAN,DETRICK,S  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: MCCLELLAN,DETRICK  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1076482  
DRY GAS EXPIRY DATE: 01/15/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: HOPPER

DATE/TIME OF TEST: 06/18/2020 19:23:10

---

Diagnostics Check	OK	19:19	
Air Blank	0.000	19:19	
Control Test	0.082	19:20	
Air Blank	0.000	19:20	
Subject Sample	# 1	19:21	Subject Test Refused
Breath Volume	0.000	19:21	
Air Blank	0.000	19:22	
Control Test	0.082	19:22	
Air Blank	0.000	19:23	
Diagnostics Check	OK	19:23	

---

Data Download Date/Time: 07/08/2020 14:41



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCLELLAN,DETRICK,S  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: MCCLELLAN,DETRICK  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1076482  
DRY GAS EXPIRY DATE: 01/15/2022  
VIOLATION CODE: DUI  
ARREST TIME: 12:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/28/2020 13:52:16

---

Diagnostics Check	OK	13:42
Air Blank	0.000	13:42
Control Test	0.081	13:42
Air Blank	0.000	13:43
Subject Sample # 1	0.170	13:46
Breath Volume	2.500	13:46
Air Blank	0.000	13:47
Air Blank	0.000	13:48
Subject Sample # 2	0.159	13:50
Breath Volume	2.191	13:50
Air Blank	0.000	13:51
Control Test	0.079	13:51
Air Blank	0.000	13:52
Diagnostics Check	OK	13:52

---

Data Download Date/Time: 07/08/2020 14:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYNN,JAMISON,C  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: LYNN,JAMISON  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: POTTER

DATE/TIME OF TEST: 06/30/2020 20:11:26

---

Diagnostics Check	OK	20:09	
Air Blank	0.000	20:09	
Control Test	0.083	20:09	
Air Blank	0.000	20:10	
Subject Sample	# 1	20:10	Sequence Aborted
Breath Volume	0.000	20:10	
Air Blank	0.000	20:11	

---

Data Download Date/Time: 07/30/2020 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYNN,JAMISON,C  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: LYNN,JAMISON  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: POTTER

DATE/TIME OF TEST: 06/30/2020 20:31:49

---

Diagnostics Check	OK	20:25
Air Blank	0.000	20:25
Control Test	0.083	20:25
Air Blank	0.000	20:26
Subject Sample # 1	0.123	20:27
Breath Volume	2.800	20:27
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample # 2	0.124	20:30
Breath Volume	2.183	20:30
Air Blank	0.000	20:30
Control Test	0.082	20:31
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

---

Data Download Date/Time: 07/30/2020 17:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SISTRUNK,TRENT,D  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: SISTRUNK,TRENT  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: LEATHEM

DATE/TIME OF TEST: 06/01/2020 02:27:05

---

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:24	
Control Test	0.079	02:25	
Air Blank	0.000	02:25	
Subject Sample	# 1	02:25	Subject Test Refused
Breath Volume	0.000	02:25	
Air Blank	0.000	02:26	
Control Test	0.079	02:26	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:27	

---

Data Download Date/Time: 06/30/2020 10:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: HARVEY,DAVID  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: WALTHALL

DATE/TIME OF TEST: 06/06/2020 02:13:44

---

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:11	
Control Test	0.079	02:11	
Air Blank	0.000	02:11	
Subject Sample	# 1	02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:12	
Control Test	0.079	02:13	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

---

Data Download Date/Time: 06/30/2020 10:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: HARVEY,DAVID  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 06/07/2020 00:33:30

---

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:30	
Control Test	0.079	00:31	
Air Blank	0.000	00:31	
Subject Sample	# 1	00:32	Subject Test Refused
Breath Volume	0.000	00:32	
Air Blank	0.000	00:32	
Control Test	0.079	00:32	
Air Blank	0.000	00:33	
Diagnostics Check	OK	00:33	

---

Data Download Date/Time: 06/30/2020 10:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARVEY,DAVID,A  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: HARVEY,DAVID  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: LAVELLE

DATE/TIME OF TEST: 06/07/2020 03:39:09

---

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:30	
Control Test	0.079	03:31	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:34	No Sample Provided
Breath Volume	0.000	03:34	
Air Blank	0.000	03:35	
Air Blank	0.000	03:37	
Subject Sample	# 2	03:37	Subject Test Refused
Breath Volume	0.000	03:37	
Air Blank	0.000	03:38	
Control Test	0.078	03:38	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

---

Data Download Date/Time: 06/30/2020 10:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLIGAN,CONNIE,A  
BREATH TEST OPERATOR AGENCY: Lake City PD  
ARRESTING OFFICER: CHEWNING,J  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: STAILEY

DATE/TIME OF TEST: 06/13/2020 03:06:25

---

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.077	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.168	03:01
Breath Volume	2.199	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.167	03:04
Breath Volume	2.222	03:04
Air Blank	0.000	03:05
Control Test	0.077	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

---

Data Download Date/Time: 06/30/2020 10:01



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BATES,BRIAN,C  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BATES,BRIAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: LYLE

DATE/TIME OF TEST: 06/13/2020 21:41:54

---

Diagnostics Check	OK	21:35
Air Blank	0.000	21:36
Control Test	0.078	21:36
Air Blank	0.000	21:36
Subject Sample # 1	0.102	21:37
Breath Volume	2.804	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:39
Subject Sample # 2	0.098	21:40
Breath Volume	3.410	21:40
Air Blank	0.000	21:40
Control Test	0.077	21:41
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

---

Data Download Date/Time: 06/30/2020 10:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLE,DANDRA,M  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: HARVEY,DAVID  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: STURDIVANT

DATE/TIME OF TEST: 06/14/2020 00:52:11

---

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:41	
Control Test	0.080	00:41	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:45	No Sample Provided
Breath Volume	0.000	00:45	
Air Blank	0.000	00:45	
Air Blank	0.000	00:47	
Subject Sample	# 2	00:50	Subject Test Refused
Breath Volume	0.000	00:50	
Air Blank	0.000	00:51	
Control Test	0.079	00:51	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

---

Data Download Date/Time: 06/30/2020 10:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,JOSEPH,D  
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)  
ARRESTING OFFICER: JOHNSTON,JOSEPH  
ARRESTING OFFICER AGENCY: FFWCC (Lake City)  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: BUI  
ARREST TIME: 01:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: GLICK

DATE/TIME OF TEST: 06/14/2020 01:48:40

---

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:46	
Control Test	0.079	01:46	
Air Blank	0.000	01:47	
Subject Sample	# 1	01:47	Subject Test Refused
Breath Volume	0.000	01:47	
Air Blank	0.000	01:47	
Control Test	0.078	01:48	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

---

Data Download Date/Time: 06/30/2020 10:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLE,DANDRA,M  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: HARVEY,DAVID  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: SUH

DATE/TIME OF TEST: 06/14/2020 02:52:53

---

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.079	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.099	02:48
Breath Volume	1.550	02:48
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.099	02:51
Breath Volume	1.851	02:51
Air Blank	0.000	02:51
Control Test	0.076	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

---

Data Download Date/Time: 06/30/2020 10:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUSGROVE,DUSTIN,K  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: PLEMMONS,LAQUINTA  
ARRESTING OFFICER AGENCY: Lake City PD  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 06/17/2020 01:17:46

---

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:04	
Control Test	0.080	01:04	
Air Blank	0.000	01:05	
Subject Sample	# 1 0.145	01:07	
Breath Volume	1.339	01:07	
Air Blank	0.000	01:08	
Air Blank	0.000	01:09	
Subject Sample	# 2 0.136	01:13	Volume Not Met
Breath Volume	0.464	01:13	
Air Blank	0.000	01:13	
Air Blank	0.000	01:15	
Subject Sample	# 3 0.145	01:16	
Breath Volume	1.851	01:16	
Air Blank	0.000	01:16	
Control Test	0.078	01:17	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

---

Data Download Date/Time: 06/30/2020 10:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERRE-LOUIS,LYNX,  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: HARVEY,DAVID  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 11:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:35

SUBJECT'S LAST NAME: LYNX

DATE/TIME OF TEST: 06/17/2020 12:07:04

---

Diagnostics Check	OK	11:56
Air Blank	0.000	11:56
Control Test	0.078	11:56
Air Blank	0.000	11:57
Subject Sample # 1	0.239	11:59
Breath Volume	2.453	11:59
Air Blank	0.000	11:59
Air Blank	0.000	12:01
Subject Sample # 2	0.218	12:02
Breath Volume	1.757	12:02
Air Blank	0.000	12:03
Air Blank	0.000	12:04
Subject Sample # 3	0.259	12:05
Breath Volume	3.027	12:05
Air Blank	0.000	12:06
Control Test	0.077	12:06
Air Blank	0.000	12:06
Diagnostics Check	OK	12:07

---

Data Download Date/Time: 06/30/2020 10:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,DANIEL,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MARTIN,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 06/18/2020 01:29:34

---

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.079	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.232	01:21
Breath Volume	2.699	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.209	01:24
Breath Volume	2.539	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 3	0.194	01:27
Breath Volume	2.808	01:27
Air Blank	0.000	01:28
Control Test	0.076	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

---

Data Download Date/Time: 06/30/2020 10:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BATES,BRIAN,C  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BATES,BRIAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: YOTT

DATE/TIME OF TEST: 06/19/2020 04:58:17

---

Diagnostics Check	OK	04:55	
Air Blank	0.000	04:55	
Control Test	0.079	04:56	
Air Blank	0.000	04:56	
Subject Sample	# 1	04:56	Subject Test Refused
Breath Volume	0.000	04:56	
Air Blank	0.000	04:57	
Control Test	0.079	04:57	
Air Blank	0.000	04:58	
Diagnostics Check	OK	04:58	

---

Data Download Date/Time: 06/30/2020 10:01



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUGHES,ALFRED,L  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: PETERSON,ARIELLE  
ARRESTING OFFICER AGENCY: Lake City PD  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 06/25/2020 21:03:00

---

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.079	20:57
Air Blank	0.000	20:57
Subject Sample # 1	0.157	20:58
Breath Volume	2.273	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:00
Subject Sample # 2	0.157	21:01
Breath Volume	2.808	21:01
Air Blank	0.000	21:02
Control Test	0.077	21:02
Air Blank	0.000	21:02
Diagnostics Check	OK	21:03

---

Data Download Date/Time: 06/30/2020 10:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,DANIEL,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MARTIN,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: DEMOSS

DATE/TIME OF TEST: 06/27/2020 23:49:09

---

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.079	23:42
Air Blank	0.000	23:43
Subject Sample # 1	0.136	23:43
Breath Volume	1.542	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.144	23:47
Breath Volume	1.937	23:47
Air Blank	0.000	23:48
Control Test	0.077	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

---

Data Download Date/Time: 06/30/2020 10:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SINGLETARY, MATHEW, C  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: SINGLETARY, MATHEW  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: NEVAREZ

DATE/TIME OF TEST: 06/28/2020 23:58:57

---

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:56	
Control Test	0.079	23:56	
Air Blank	0.000	23:57	
Subject Sample	# 1	23:57	Subject Test Refused
Breath Volume	0.000	23:57	
Air Blank	0.000	23:58	
Control Test	0.079	23:58	
Air Blank	0.000	23:58	
Diagnostics Check	OK	23:58	

---

Data Download Date/Time: 06/30/2020 10:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LLANES,GABRIEL,M  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: LLANES,GABRIEL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: GRAF

DATE/TIME OF TEST: 06/28/2020 01:16:30

---

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.078	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.142	01:12
Breath Volume	2.230	01:12
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.148	01:14
Breath Volume	1.781	01:14
Air Blank	0.000	01:15
Control Test	0.078	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

---

Data Download Date/Time: 08/05/2020 07:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AKINS,JOSHUA,D  
BREATH TEST OPERATOR AGENCY: Madison County SO  
ARRESTING OFFICER: PIKE,BEN  
ARRESTING OFFICER AGENCY: Madison County SO  
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A2  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 08:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: SWAILS

DATE/TIME OF TEST: 06/29/2020 08:39:10

---

Diagnostics Check		OK	08:32
Air Blank		0.000	08:33
Control Test		0.078	08:33
Air Blank		0.000	08:34
Subject Sample	# 1	0.260	08:34
Breath Volume		1.761	08:34
Air Blank		0.000	08:35
Air Blank		0.000	08:37
Subject Sample	# 2	0.254	08:37
Breath Volume		2.214	08:37
Air Blank		0.000	08:38
Control Test		0.076	08:38
Air Blank		0.000	08:39
Diagnostics Check		OK	08:39

---

Data Download Date/Time: 08/05/2020 07:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KASTOR,JOSEPH,B  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: KASTOR,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: DANIEL

DATE/TIME OF TEST: 06/03/2020 20:09:31

---

Diagnostics Check	OK	20:02
Air Blank	0.000	20:03
Control Test	0.083	20:03
Air Blank	0.000	20:03
Subject Sample # 1	0.069	20:04
Breath Volume	1.238	20:04
Air Blank	0.000	20:05
Air Blank	0.000	20:07
Subject Sample # 2	0.070	20:07
Breath Volume	2.214	20:07
Air Blank	0.000	20:08
Control Test	0.082	20:08
Air Blank	0.000	20:09
Diagnostics Check	OK	20:09

---

Data Download Date/Time: 07/01/2020 23:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KASTOR,JOSEPH,B  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: KASTOR,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: ROUNDTREE

DATE/TIME OF TEST: 06/06/2020 03:59:58

---

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.083	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 1 0.159	03:58	
Breath Volume	2.367	03:58	
Air Blank		03:59	Purge Fail
Air Blank	0.000	03:59	

---

Data Download Date/Time: 07/01/2020 23:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KASTOR,JOSEPH,B  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: KASTOR,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 05/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: ROUNDTREE

DATE/TIME OF TEST: 06/06/2020 04:10:25

---

Diagnostics Check	OK	04:06	
Air Blank	0.000	04:07	
Control Test	0.082	04:07	
Air Blank	0.000	04:08	
Subject Sample # 1	0.156	04:09	
Breath Volume	2.886	04:09	
Air Blank		04:09	Purge Fail
Air Blank	0.000	04:10	

---

Data Download Date/Time: 07/01/2020 23:52



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN, BRANDIE, L  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: LANDIS, MICHAEL  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 05/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ALTMAN

DATE/TIME OF TEST: 06/08/2020 02:33:34

---

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:31	
Control Test	0.083	02:31	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:32	Subject Test Refused
Breath Volume	0.000	02:32	
Air Blank	0.000	02:32	
Control Test	0.084	02:32	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:33	

---

Data Download Date/Time: 07/01/2020 23:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN, BRANDIE, L  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: SLAUGHTER, DEREK  
ARRESTING OFFICER AGENCY: Live Oak PD  
LAST AGENCY INSPECTION: 05/13/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 06/19/2020 22:46:37

---

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.082	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.152	22:41
Breath Volume	2.519	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.155	22:44
Breath Volume	2.769	22:44
Air Blank	0.000	22:45
Control Test	0.081	22:46
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

---

Data Download Date/Time: 07/01/2020 23:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YASUDA,SHOHEI,  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: LEWIS,TIMOTHY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/07/2020  
VIOLATION CODE: DUI  
ARREST TIME: 05:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: HENAGEN

DATE/TIME OF TEST: 06/07/2020 06:29:59

---

Diagnostics Check	OK	06:25	
Air Blank	0.000	06:26	
Control Test	0.081	06:26	
Air Blank	0.000	06:27	
Subject Sample # 1	0.124	06:28	
Breath Volume	1.988	06:28	
Air Blank		06:28	RFI Detect
Air Blank	0.000	06:29	
Control Test	0.081	06:29	
Air Blank	0.000	06:29	
Diagnostics Check	OK	06:29	

---

Data Download Date/Time: 07/02/2020 13:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YASUDA,SHOHEI,  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: LEWIS,TIMOTHY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 770784  
DRY GAS EXPIRY DATE: 04/07/2020  
VIOLATION CODE: DUI  
ARREST TIME: 05:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: HENAGEN

DATE/TIME OF TEST: 06/07/2020 06:39:49

---

Diagnostics Check	OK	06:33
Air Blank	0.000	06:33
Control Test	0.079	06:33
Air Blank	0.000	06:34
Subject Sample # 1	0.123	06:35
Breath Volume	1.898	06:35
Air Blank	0.000	06:36
Air Blank	0.000	06:37
Subject Sample # 2	0.109	06:38
Breath Volume	1.824	06:38
Air Blank	0.000	06:38
Control Test	0.078	06:39
Air Blank	0.000	06:39
Diagnostics Check	OK	06:39

---

Data Download Date/Time: 07/02/2020 13:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: LEWIS,TIMOTHY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 06/07/2020 06:53:55

---

Diagnostics Check	OK	06:50	
Air Blank	0.000	06:51	
Control Test	0.080	06:51	
Air Blank	0.000	06:52	
Subject Sample	# 1	06:52	RFI Detect
Breath Volume	0.000	06:52	
Air Blank	0.000	06:52	
Control Test	0.081	06:53	
Air Blank	0.000	06:53	
Diagnostics Check	OK	06:53	

---

Data Download Date/Time: 07/02/2020 13:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: LEWIS,TIMOTHY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 06/07/2020 07:03:22

---

Diagnostics Check	OK	06:56
Air Blank	0.000	06:57
Control Test	0.080	06:57
Air Blank	0.000	06:57
Subject Sample # 1	0.097	06:58
Breath Volume	2.000	06:58
Air Blank	0.000	06:59
Air Blank	0.000	07:00
Subject Sample # 2	0.098	07:01
Breath Volume	3.070	07:01
Air Blank	0.000	07:02
Control Test	0.079	07:02
Air Blank	0.000	07:03
Diagnostics Check	OK	07:03

---

Data Download Date/Time: 07/02/2020 13:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: LEWIS,TIMOTHY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 06/13/2020 20:14:01

---

Diagnostics Check	OK	20:07
Air Blank	0.000	20:07
Control Test	0.082	20:08
Air Blank	0.000	20:08
Subject Sample # 1	0.044	20:09
Breath Volume	2.078	20:09
Air Blank	0.000	20:09
Air Blank	0.000	20:11
Subject Sample # 2	0.044	20:12
Breath Volume	1.785	20:12
Air Blank	0.000	20:13
Control Test	0.080	20:13
Air Blank	0.000	20:13
Diagnostics Check	OK	20:14

---

Data Download Date/Time: 07/02/2020 13:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SISK,CARLOS,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: SISK,CARLOS  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 06/19/2020 04:52:22

---

Diagnostics Check	OK	04:47	
Air Blank	0.000	04:47	
Control Test	0.082	04:47	
Air Blank	0.000	04:48	
Subject Sample # 1	0.189	04:51	
Breath Volume	1.109	04:51	
Air Blank		04:51	Purge Fail
Air Blank	0.000	04:52	

---

Data Download Date/Time: 07/02/2020 13:08



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SISK,CARLOS,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: SISK,CARLOS  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 06/19/2020 04:56:01

---

Diagnostics Check	OK	04:54	
Air Blank	0.000	04:55	
Control Test	0.074	04:55	Control Outside Tolerance
Air Blank	0.000	04:56	

---

Data Download Date/Time: 07/02/2020 13:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SISK,CARLOS,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: SISK,CARLOS  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 06/19/2020 05:06:12

---

Diagnostics Check	OK	04:57
Air Blank	0.000	04:58
Control Test	0.080	04:58
Air Blank	0.000	04:59
Subject Sample # 1	0.196	04:59
Breath Volume	1.828	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.196	05:04
Breath Volume	1.523	05:04
Air Blank	0.000	05:05
Control Test	0.078	05:05
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

---

Data Download Date/Time: 07/02/2020 13:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT,ALLEN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: SCOTT,ALLEN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 06/21/2020 00:40:50

---

Diagnostics Check	OK	00:32
Air Blank	0.000	00:33
Control Test	0.082	00:33
Air Blank	0.000	00:34
Subject Sample # 1	0.142	00:35
Breath Volume	2.339	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:37
Subject Sample # 2	0.133	00:39
Breath Volume	1.476	00:39
Air Blank	0.000	00:39
Control Test	0.079	00:40
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

---

Data Download Date/Time: 07/02/2020 13:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SISK,CARLOS,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: SISK,CARLOS  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: MCCLELLAN

DATE/TIME OF TEST: 06/23/2020 23:41:15

---

Diagnostics Check	OK	23:39	
Air Blank	0.000	23:40	
Control Test	0.000	23:40	Control Outside Tolerance
Air Blank	0.000	23:41	

---

Data Download Date/Time: 07/02/2020 13:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SISK,CARLOS,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: SISK,CARLOS  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: MCCLELLAN

DATE/TIME OF TEST: 06/23/2020 23:44:25

---

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:43	
Control Test	0.000	23:43	Control Outside Tolerance
Air Blank	0.000	23:44	

---

Data Download Date/Time: 07/02/2020 13:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SISK,CARLOS,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: SISK,CARLOS  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: MCCLELLAN

DATE/TIME OF TEST: 06/23/2020 23:48:00

---

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:47	
Control Test	0.000	23:47	Control Outside Tolerance
Air Blank	0.000	23:48	

---

Data Download Date/Time: 07/02/2020 13:08



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SELLERS,JOEY,C  
BREATH TEST OPERATOR AGENCY: Holmes County SO  
ARRESTING OFFICER: GORDON,GREG  
ARRESTING OFFICER AGENCY: Holmes County SO  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 06/06/2020 22:31:33

---

Diagnostics Check	OK	22:28	
Air Blank	0.000	22:29	
Control Test	0.082	22:29	
Air Blank	0.000	22:29	
Subject Sample	# 1	22:30	Subject Test Refused
Breath Volume	0.000	22:30	
Air Blank	0.000	22:30	
Control Test	0.082	22:30	
Air Blank	0.000	22:31	
Diagnostics Check	OK	22:31	

---

Data Download Date/Time: 07/08/2020 14:51





**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BANKS,SEAN,C  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: BYRD,BRANDON  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 06/07/2020 04:46:50

---

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.077	04:41
Air Blank	0.000	04:41
Subject Sample # 1	0.211	04:42
Breath Volume	2.542	04:42
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.196	04:45
Breath Volume	2.128	04:45
Air Blank	0.000	04:45
Control Test	0.078	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

---

Data Download Date/Time: 07/01/2020 13:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMS,KYLE,N  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: SIMMS,KYLE  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: BACHER

DATE/TIME OF TEST: 06/14/2020 20:54:15

---

Diagnostics Check	OK	20:47
Air Blank	0.000	20:48
Control Test	0.078	20:48
Air Blank	0.000	20:48
Subject Sample # 1	0.224	20:49
Breath Volume	2.207	20:49
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.237	20:52
Breath Volume	2.121	20:52
Air Blank	0.000	20:53
Control Test	0.078	20:53
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

---

Data Download Date/Time: 07/01/2020 13:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBINSON,ANDREW,J  
BREATH TEST OPERATOR AGENCY: OTHER Fed  
ARRESTING OFFICER: ROBINSON,ANDREW  
ARRESTING OFFICER AGENCY: OTHER Fed  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1094494  
DRY GAS EXPIRY DATE: 02/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 06/19/2020 19:07:30

---

Diagnostics Check	OK	19:01
Air Blank	0.000	19:01
Control Test	0.077	19:01
Air Blank	0.000	19:02
Subject Sample # 1	0.105	19:03
Breath Volume	3.476	19:03
Air Blank	0.000	19:03
Air Blank	0.000	19:05
Subject Sample # 2	0.098	19:05
Breath Volume	2.246	19:05
Air Blank	0.000	19:06
Control Test	0.078	19:06
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

---

Data Download Date/Time: 07/01/2020 13:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: PEREZ Y PEREZ

DATE/TIME OF TEST: 06/01/2020 02:16:07

---

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.080	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.169	02:11
Breath Volume	1.648	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.165	02:14
Breath Volume	1.347	02:14
Air Blank	0.000	02:15
Control Test	0.079	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

---

Data Download Date/Time: 06/24/2020 15:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRAWDY,LAINE,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: DRAWDY,LAINE  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: GOODLAND

DATE/TIME OF TEST: 06/01/2020 23:41:42

---

Diagnostics Check	OK	23:35
Air Blank	0.000	23:35
Control Test	0.080	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.133	23:37
Breath Volume	1.660	23:37
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.131	23:40
Breath Volume	2.242	23:40
Air Blank	0.000	23:40
Control Test	0.080	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

---

Data Download Date/Time: 06/24/2020 15:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,CODY,L  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: RODRIGUEZ,ANDRES  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 11:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/02/2020 12:53:08

---

Diagnostics Check	OK	12:47
Air Blank	0.000	12:47
Control Test	0.080	12:47
Air Blank	0.000	12:48
Subject Sample # 1	0.000	12:48
Breath Volume	4.417	12:48
Air Blank	0.000	12:49
Air Blank	0.000	12:51
Subject Sample # 2	0.000	12:51
Breath Volume	3.777	12:51
Air Blank	0.000	12:52
Control Test	0.079	12:52
Air Blank	0.000	12:53
Diagnostics Check	OK	12:53

---

Data Download Date/Time: 06/24/2020 15:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: SHAFFER

DATE/TIME OF TEST: 06/03/2020 19:31:52

---

Diagnostics Check	OK	19:25
Air Blank	0.000	19:26
Control Test	0.080	19:26
Air Blank	0.000	19:27
Subject Sample # 1	0.190	19:27
Breath Volume	2.425	19:27
Air Blank	0.000	19:28
Air Blank	0.000	19:29
Subject Sample # 2	0.187	19:30
Breath Volume	1.996	19:30
Air Blank	0.000	19:30
Control Test	0.079	19:31
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

---

Data Download Date/Time: 06/24/2020 15:16



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: CUMMINGS

DATE/TIME OF TEST: 06/04/2020 21:54:51

---

Diagnostics Check	OK	21:46
Air Blank	0.000	21:46
Control Test	0.080	21:46
Air Blank	0.000	21:47
Subject Sample # 1	0.000	21:49
Breath Volume	1.957	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.000	21:53
Breath Volume	2.363	21:53
Air Blank	0.000	21:53
Control Test	0.080	21:54
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

---

Data Download Date/Time: 06/24/2020 15:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMB,HUNTER,  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LIMB,HUNTER  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: SALDANA VELAZQUEZ

DATE/TIME OF TEST: 06/06/2020 00:57:01

---

Diagnostics Check		OK	00:40	
Air Blank		0.000	00:41	
Control Test		0.080	00:41	
Air Blank		0.000	00:42	
Subject Sample	# 1	0.167	00:45	Volume Not Met
Breath Volume		0.066	00:45	
Air Blank		0.000	00:45	
Air Blank		0.000	00:47	
Subject Sample	# 2	0.162	00:49	
Breath Volume		1.109	00:49	
Air Blank		0.000	00:50	
Air Blank		0.000	00:52	
Subject Sample	# 3	0.152	00:55	Volume Not Met
Breath Volume		0.164	00:55	
Air Blank		0.000	00:56	
Control Test		0.078	00:56	
Air Blank		0.000	00:56	
Diagnostics Check		OK	00:57	

---

Data Download Date/Time: 06/24/2020 15:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAMPIASI,JAKE,L  
BREATH TEST OPERATOR AGENCY: FHP Troop K  
ARRESTING OFFICER: LAMPIASI,JAKE  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/06/2020 02:34:23

---

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:32	
Control Test	0.080	02:32	
Air Blank	0.000	02:33	
Subject Sample	# 1	02:33	RFI Detect
Breath Volume	0.000	02:33	
Air Blank		02:34	RFI Detect
Air Blank		02:34	RFI Detect

---

Data Download Date/Time: 06/24/2020 15:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMB,HUNTER,  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LIMB,HUNTER  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 06/06/2020 20:11:14

---

Diagnostics Check	OK	20:05
Air Blank	0.000	20:05
Control Test	0.080	20:06
Air Blank	0.000	20:06
Subject Sample # 1	0.000	20:07
Breath Volume	3.214	20:07
Air Blank	0.000	20:07
Air Blank	0.000	20:09
Subject Sample # 2	0.000	20:09
Breath Volume	3.210	20:09
Air Blank	0.000	20:10
Control Test	0.079	20:10
Air Blank	0.000	20:11
Diagnostics Check	OK	20:11

---

Data Download Date/Time: 06/24/2020 15:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRAWDY,LAINE,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: DRAWDY,LAINE  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 06/08/2020 03:03:11

---

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.113	02:58
Breath Volume	3.253	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.108	03:01
Breath Volume	3.003	03:01
Air Blank	0.000	03:02
Control Test	0.079	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

---

Data Download Date/Time: 06/24/2020 15:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRAWDY,LAINE,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: DRAWDY,LAINE  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: GOOCH

DATE/TIME OF TEST: 06/12/2020 04:45:25

---

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.080	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.000	04:41
Breath Volume	3.000	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.000	04:44
Breath Volume	2.492	04:44
Air Blank	0.000	04:44
Control Test	0.080	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

---

Data Download Date/Time: 06/24/2020 15:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FREDERIC,JEAN,  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: FREDERIC,JEAN  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: SHELDON

DATE/TIME OF TEST: 06/12/2020 19:44:18

---

Diagnostics Check	OK	19:38
Air Blank	0.000	19:38
Control Test	0.080	19:39
Air Blank	0.000	19:39
Subject Sample # 1	0.000	19:40
Breath Volume	2.332	19:40
Air Blank	0.000	19:40
Air Blank	0.000	19:42
Subject Sample # 2	0.000	19:42
Breath Volume	1.785	19:42
Air Blank	0.000	19:43
Control Test	0.080	19:43
Air Blank	0.000	19:44
Diagnostics Check	OK	19:44

---

Data Download Date/Time: 06/24/2020 15:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: KIFFER

DATE/TIME OF TEST: 06/12/2020 20:47:54

---

Diagnostics Check	OK	20:40
Air Blank	0.000	20:40
Control Test	0.079	20:41
Air Blank	0.000	20:41
Subject Sample # 1	0.260	20:42
Breath Volume	2.175	20:42
Air Blank	0.000	20:42
Air Blank	0.000	20:44
Subject Sample # 2	0.249	20:46
Breath Volume	1.765	20:46
Air Blank	0.000	20:47
Control Test	0.079	20:47
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

---

Data Download Date/Time: 06/24/2020 15:16



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMB,HUNTER,  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LIMB,HUNTER  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: RIZZI

DATE/TIME OF TEST: 06/12/2020 22:40:46

---

Diagnostics Check	OK	22:34
Air Blank	0.000	22:35
Control Test	0.080	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.186	22:36
Breath Volume	3.277	22:36
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.185	22:39
Breath Volume	3.066	22:39
Air Blank	0.000	22:39
Control Test	0.079	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

---

Data Download Date/Time: 06/24/2020 15:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRCHNER, MATTHEW, J  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: COLANGELO, CODY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MCGEE

DATE/TIME OF TEST: 06/13/2020 04:08:39

---

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.080	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.147	04:02
Breath Volume	1.703	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.140	04:07
Breath Volume	1.820	04:07
Air Blank	0.000	04:07
Control Test	0.078	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

---

Data Download Date/Time: 06/24/2020 15:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SCUOPPO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BRODOCK

DATE/TIME OF TEST: 06/13/2020 23:07:28

---

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.080	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.058	23:02
Breath Volume	1.710	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.062	23:05
Breath Volume	2.167	23:05
Air Blank	0.000	23:06
Control Test	0.080	23:06
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

---

Data Download Date/Time: 06/24/2020 15:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,RICKY,  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: PEREZ,RICKY  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 06/16/2020 22:33:21

---

Diagnostics Check	OK	22:32	
Air Blank	0.000	22:32	
Control Test	0.080	22:32	RFI Detect
Air Blank	0.000	22:33	

---

Data Download Date/Time: 06/24/2020 15:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,RICKY,  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: PEREZ,RICKY  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 06/16/2020 22:41:48

---

Diagnostics Check	OK	22:35	
Air Blank	0.000	22:35	
Control Test	0.080	22:36	
Air Blank	0.000	22:36	
Subject Sample	# 1 0.248	22:37	
Breath Volume	1.699	22:37	
Air Blank	0.000	22:38	
Air Blank	0.000	22:40	
Subject Sample	# 2 0.248	22:41	
Breath Volume	1.558	22:41	
Air Blank		22:41	RFI Detect
Air Blank		22:41	RFI Detect

---

Data Download Date/Time: 06/24/2020 15:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,RICKY,  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: PEREZ,RICKY  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: CRAWNFORD

DATE/TIME OF TEST: 06/16/2020 22:51:49

---

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.079	22:46
Air Blank	0.000	22:46
Subject Sample # 1	0.246	22:47
Breath Volume	1.164	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:49
Subject Sample # 2	0.247	22:50
Breath Volume	1.683	22:50
Air Blank	0.000	22:50
Control Test	0.080	22:51
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

---

Data Download Date/Time: 06/24/2020 15:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAGLIONI,DAVID,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: GILKEY,JOSHUA  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: DELCASTILLO

DATE/TIME OF TEST: 06/17/2020 14:51:24

---

Diagnostics Check		OK	14:45
Air Blank		0.000	14:45
Control Test		0.080	14:46
Air Blank		0.000	14:46
Subject Sample	# 1	0.278	14:47
Breath Volume		1.707	14:47
Air Blank		0.000	14:47
Air Blank		0.000	14:49
Subject Sample	# 2	0.265	14:49
Breath Volume		1.511	14:49
Air Blank		0.000	14:50
Control Test		0.079	14:50
Air Blank		0.000	14:51
Diagnostics Check		OK	14:51

---

Data Download Date/Time: 06/24/2020 15:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMB,HUNTER,  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LIMB,HUNTER  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: GALLO

DATE/TIME OF TEST: 06/18/2020 01:12:25

---

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.080	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.000	01:08
Breath Volume	3.398	01:08
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.000	01:11
Breath Volume	3.539	01:11
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

---

Data Download Date/Time: 06/24/2020 15:16



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAMPIASI,JAKE,L  
BREATH TEST OPERATOR AGENCY: FHP Troop K  
ARRESTING OFFICER: LAMPIASI,JAKE  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 05/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: JULIAN

DATE/TIME OF TEST: 06/18/2020 01:39:38

---

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:36	
Control Test	0.079	01:36	
Air Blank	0.000	01:37	
Subject Sample	# 1	01:38	RFI Detect
Breath Volume	0.000	01:38	
Air Blank	0.000	01:38	
Control Test	0.079	01:39	
Air Blank	0.000	01:39	
Diagnostics Check	OK	01:39	

---

Data Download Date/Time: 06/24/2020 15:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GILKEY,JOSHUA,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: GILKEY,JOSHUA  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:48

SUBJECT'S LAST NAME: DAVIE

DATE/TIME OF TEST: 06/24/2020 16:20:57

---

Diagnostics Check	OK	16:14
Air Blank	0.000	16:14
Control Test	0.080	16:15
Air Blank	0.000	16:15
Subject Sample # 1	0.000	16:16
Breath Volume	3.570	16:16
Air Blank	0.000	16:17
Air Blank	0.000	16:19
Subject Sample # 2	0.000	16:19
Breath Volume	3.808	16:19
Air Blank	0.000	16:20
Control Test	0.079	16:20
Air Blank	0.000	16:20
Diagnostics Check	OK	16:20

---

Data Download Date/Time: 07/29/2020 10:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HALL,RICHARD,D  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: HALL,RICHARD  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/26/2020 04:57:01

---

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.080	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.120	04:52
Breath Volume	2.304	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.116	04:55
Breath Volume	1.972	04:55
Air Blank	0.000	04:56
Control Test	0.078	04:56
Air Blank	0.000	04:56
Diagnostics Check	OK	04:57

---

Data Download Date/Time: 07/29/2020 10:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PATTERSON,KYLE,L  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: PATTERSON,KYLE  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 06/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: BUI  
ARREST TIME: 19:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: MEEHAN

DATE/TIME OF TEST: 06/27/2020 20:03:39

---

Diagnostics Check	OK	19:57
Air Blank	0.000	19:57
Control Test	0.080	19:58
Air Blank	0.000	19:58
Subject Sample # 1	0.156	19:59
Breath Volume	3.074	19:59
Air Blank	0.000	19:59
Air Blank	0.000	20:01
Subject Sample # 2	0.158	20:02
Breath Volume	3.457	20:02
Air Blank	0.000	20:02
Control Test	0.080	20:03
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

---

Data Download Date/Time: 07/29/2020 10:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMB,HUNTER,  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LIMB,HUNTER  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 06/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/27/2020 22:05:13

---

Diagnostics Check	OK	21:58
Air Blank	0.000	21:59
Control Test	0.080	21:59
Air Blank	0.000	22:00
Subject Sample # 1	0.203	22:00
Breath Volume	1.886	22:00
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.203	22:03
Breath Volume	1.335	22:03
Air Blank	0.000	22:04
Control Test	0.079	22:04
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

---

Data Download Date/Time: 07/29/2020 10:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEEKS,BRADLEY,A  
BREATH TEST OPERATOR AGENCY: Hamilton County SO  
ARRESTING OFFICER: DOWDY,JOHNATHAN  
ARRESTING OFFICER AGENCY: White Springs PD  
LAST AGENCY INSPECTION: 05/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00919080A3  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 06/28/2020 04:32:48

---

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.077	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.222	04:28
Breath Volume	3.246	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:30
Subject Sample # 2	0.225	04:31
Breath Volume	2.996	04:31
Air Blank	0.000	04:31
Control Test	0.076	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

---

Data Download Date/Time: 07/02/2020 20:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADKINS, JAKE, M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: SINGLETON, NATHAN  
ARRESTING OFFICER AGENCY: Dixie County SO  
LAST AGENCY INSPECTION: 06/10/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 2241908A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:27

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 06/25/2020 15:06:05

---

Diagnostics Check	OK	15:00
Air Blank	0.000	15:00
Control Test	0.078	15:00
Air Blank	0.000	15:01
Subject Sample # 1	0.000	15:01
Breath Volume	1.898	15:01
Air Blank	0.000	15:02
Air Blank	0.000	15:04
Subject Sample # 2	0.000	15:04
Breath Volume	2.207	15:04
Air Blank	0.000	15:05
Control Test	0.078	15:05
Air Blank	0.000	15:06
Diagnostics Check	OK	15:06

---

Data Download Date/Time: 07/08/2020 10:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASCITELLI,V,D  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: ORIGON,L  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 06/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: SARTOR

DATE/TIME OF TEST: 06/27/2020 21:17:49

---

Diagnostics Check		OK	21:11
Air Blank		0.000	21:11
Control Test		0.081	21:12
Air Blank		0.000	21:12
Subject Sample	# 1	0.126	21:13
Breath Volume		2.210	21:13
Air Blank		0.000	21:13
Air Blank		0.000	21:15
Subject Sample	# 2	0.118	21:16
Breath Volume		1.746	21:16
Air Blank		0.000	21:16
Control Test		0.080	21:17
Air Blank		0.000	21:17
Diagnostics Check		OK	21:17

---

Data Download Date/Time: 07/27/2020 06:50



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: BURGESS,JAMES  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: HAVENS

DATE/TIME OF TEST: 06/30/2020 18:38:11

---

Diagnostics Check	OK	18:37	
Air Blank		18:37	RFI Detect
Air Blank	0.000	18:38	

---

Data Download Date/Time: 07/23/2020 09:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: BURGESS,JAMES  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: HAVENS

DATE/TIME OF TEST: 06/30/2020 18:41:46

---

Diagnostics Check	OK	18:41	
Air Blank		18:41	RFI Detect
Air Blank		18:41	RFI Detect

---

Data Download Date/Time: 07/23/2020 09:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: BURGESS,JAMES  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 06/24/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: HAVENS

DATE/TIME OF TEST: 06/30/2020 18:51:57

---

Diagnostics Check	OK	18:47	
Air Blank	0.000	18:48	
Control Test	0.080	18:48	
Air Blank	0.000	18:49	
Subject Sample	# 1	18:51	RFI Detect
Breath Volume	0.113	18:51	
Air Blank		18:51	RFI Detect
Air Blank	0.000	18:51	

---

Data Download Date/Time: 07/23/2020 09:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LETSON,REBECCA,G  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: LETSON,REBECCA  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 06/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: DEMAREST

DATE/TIME OF TEST: 06/28/2020 22:36:07

---

Diagnostics Check	OK	22:29
Air Blank	0.000	22:30
Control Test	0.079	22:30
Air Blank	0.000	22:30
Subject Sample # 1	0.230	22:31
Breath Volume	3.140	22:31
Air Blank	0.000	22:32
Air Blank	0.000	22:33
Subject Sample # 2	0.225	22:34
Breath Volume	3.175	22:34
Air Blank	0.000	22:35
Control Test	0.078	22:35
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

---

Data Download Date/Time: 07/19/2020 19:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ISHALL,JALEN,L  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: WILSON,DAKOTA  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 06/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 06/29/2020 22:12:05

---

Diagnostics Check	OK	22:05
Air Blank	0.000	22:06
Control Test	0.079	22:06
Air Blank	0.000	22:07
Subject Sample # 1	0.000	22:07
Breath Volume	3.351	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.000	22:10
Breath Volume	1.628	22:10
Air Blank	0.000	22:11
Control Test	0.078	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

---

Data Download Date/Time: 07/19/2020 19:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LETSON,REBECCA,G  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: GUINANE,JOHN  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 05/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: DUCAT

DATE/TIME OF TEST: 06/09/2020 18:15:33

---

Diagnostics Check		OK	18:09
Air Blank		0.000	18:09
Control Test		0.078	18:10
Air Blank		0.000	18:10
Subject Sample	# 1	0.214	18:10
Breath Volume		2.109	18:10
Air Blank		0.000	18:11
Air Blank		0.000	18:13
Subject Sample	# 2	0.215	18:13
Breath Volume		2.476	18:13
Air Blank		0.000	18:14
Control Test		0.077	18:14
Air Blank		0.000	18:15
Diagnostics Check		OK	18:15

---

Data Download Date/Time: 06/21/2020 00:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONG,ERIC,W  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: TEMPLIN,RICHARD  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 05/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 10:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:03

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 06/15/2020 11:30:53

---

Diagnostics Check	OK	11:24
Air Blank	0.000	11:24
Control Test	0.077	11:24
Air Blank	0.000	11:25
Subject Sample # 1	0.000	11:26
Breath Volume	1.484	11:26
Air Blank	0.000	11:26
Air Blank	0.000	11:28
Subject Sample # 2	0.000	11:29
Breath Volume	1.175	11:29
Air Blank	0.000	11:29
Control Test	0.078	11:30
Air Blank	0.000	11:30
Diagnostics Check	OK	11:30

---

Data Download Date/Time: 06/21/2020 00:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: WILSON,DAKOTA  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 05/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: MASSEY

DATE/TIME OF TEST: 06/15/2020 22:39:47

---

Diagnostics Check	OK	22:32
Air Blank	0.000	22:32
Control Test	0.078	22:33
Air Blank	0.000	22:33
Subject Sample # 1	0.000	22:35
Breath Volume	1.246	22:35
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.000	22:38
Breath Volume	1.464	22:38
Air Blank	0.000	22:38
Control Test	0.078	22:39
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

---

Data Download Date/Time: 06/21/2020 00:46



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITCHER,DARRYL,P  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: CHESSHER,SANDI  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MAHAN

DATE/TIME OF TEST: 06/05/2020 01:02:50

---

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:01	
Control Test	0.074	01:02	Control Outside Tolerance
Air Blank	0.000	01:02	

---

Data Download Date/Time: 06/24/2020 17:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITCHER,DARRYL,P  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: CHESSHER,SANDI  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MAHAN

DATE/TIME OF TEST: 06/05/2020 01:05:45

---

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:04	
Control Test	0.074	01:05	Control Outside Tolerance
Air Blank	0.000	01:05	

---

Data Download Date/Time: 06/24/2020 17:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALSH,SHANNON,L  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: CHESSHER,S  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 06/20/2020 01:22:33

---

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:21	
Control Test	0.074	01:21	Control Outside Tolerance
Air Blank	0.000	01:22	

---

Data Download Date/Time: 06/24/2020 17:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHESSHER,SANDRA,W  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: CHESSHER,SANDRA  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: SYS CHECK  
ARREST TIME: 16:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 06/20/2020 16:27:52

---

Diagnostics Check	OK	16:21
Air Blank	0.000	16:21
Control Test	0.076	16:22
Air Blank	0.000	16:22
Subject Sample # 1	0.000	16:23
Breath Volume	2.027	16:23
Air Blank	0.000	16:24
Air Blank	0.000	16:26
Subject Sample # 2	0.000	16:26
Breath Volume	1.855	16:26
Air Blank	0.000	16:26
Control Test	0.075	16:27
Air Blank	0.000	16:27
Diagnostics Check	OK	16:27

---

Data Download Date/Time: 06/24/2020 17:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHESSHER,SANDRA,W  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: BINDER,BRONSON  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: SERPA

DATE/TIME OF TEST: 06/21/2020 00:00:40

---

Diagnostics Check	OK	23:54
Air Blank	0.000	23:55
Control Test	0.077	23:55
Air Blank	0.000	23:55
Subject Sample # 1	0.105	23:56
Breath Volume	1.457	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.110	23:59
Breath Volume	1.238	23:59
Air Blank	0.000	23:59
Control Test	0.077	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

---

Data Download Date/Time: 06/24/2020 17:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHESSHER,SANDRA,W  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: BINDER,B  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 06/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1099238  
DRY GAS EXPIRY DATE: 03/01/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/25/2020 21:09:09

---

Diagnostics Check	OK	21:03	
Air Blank	0.000	21:04	
Control Test	0.080	21:04	
Air Blank	0.000	21:04	
Subject Sample	# 1	21:07	Subject Test Refused
Breath Volume	0.000	21:07	
Air Blank	0.000	21:08	
Control Test	0.081	21:08	
Air Blank	0.000	21:09	
Diagnostics Check	OK	21:09	

---

Data Download Date/Time: 07/11/2020 00:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALSH,SHANNON,L  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: BRONSON,B  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 06/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1099238  
DRY GAS EXPIRY DATE: 03/01/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: TOMASSIAN

DATE/TIME OF TEST: 06/26/2020 01:20:25

---

Diagnostics Check	OK	01:13
Air Blank	0.000	01:14
Control Test	0.081	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.112	01:15
Breath Volume	3.675	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.108	01:18
Breath Volume	2.402	01:18
Air Blank	0.000	01:19
Control Test	0.081	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

---

Data Download Date/Time: 07/11/2020 00:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALSH,SHANNON,L  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: BRONSON,B  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 06/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1099238  
DRY GAS EXPIRY DATE: 03/01/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: ELSZASZ

DATE/TIME OF TEST: 06/27/2020 00:50:23

---

Diagnostics Check	OK	00:43
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:44
Subject Sample # 1	0.272	00:45
Breath Volume	1.914	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.265	00:48
Breath Volume	1.968	00:48
Air Blank	0.000	00:49
Control Test	0.080	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

---

Data Download Date/Time: 07/11/2020 00:57



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PETERSON,KIP,J  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: PETERSON,KIP  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181692  
DRY GAS EXPIRY DATE: 03/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 06/14/2020 03:20:18

---

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.078	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.045	03:15
Breath Volume	2.011	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.045	03:18
Breath Volume	1.800	03:18
Air Blank	0.000	03:19
Control Test	0.078	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

---

Data Download Date/Time: 06/24/2020 11:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: HOWARD,D  
ARRESTING OFFICER AGENCY: FHP TROOP C  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: HANZLIK JR

DATE/TIME OF TEST: 06/01/2020 22:38:31

---

Diagnostics Check	OK	22:32
Air Blank	0.000	22:32
Control Test	0.080	22:33
Air Blank	0.000	22:33
Subject Sample # 1	0.260	22:34
Breath Volume	3.691	22:34
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.257	22:37
Breath Volume	2.945	22:37
Air Blank	0.000	22:37
Control Test	0.078	22:37
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

---

Data Download Date/Time: 06/04/2020 14:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: ANDREWS,S  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: MEEK

DATE/TIME OF TEST: 06/03/2020 05:32:31

---

Diagnostics Check	OK	05:25
Air Blank	0.000	05:25
Control Test	0.079	05:26
Air Blank	0.000	05:26
Subject Sample # 1	0.244	05:27
Breath Volume	1.191	05:27
Air Blank	0.000	05:28
Air Blank	0.000	05:29
Subject Sample # 2	0.240	05:31
Breath Volume	1.187	05:31
Air Blank	0.000	05:31
Control Test	0.079	05:31
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

---

Data Download Date/Time: 06/04/2020 14:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: HALL,K  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 06/06/2020 01:15:47

---

Diagnostics Check		OK	01:04	
Air Blank		0.000	01:04	
Control Test		0.080	01:05	
Air Blank		0.000	01:05	
Subject Sample	# 1	0.146	01:08	Volume Not Met
Breath Volume		0.089	01:08	
Air Blank		0.000	01:09	
Air Blank		0.000	01:11	
Subject Sample	# 2	0.157	01:14	Volume Not Met
Breath Volume		0.785	01:14	
Air Blank		0.000	01:14	
Control Test		0.078	01:15	
Air Blank		0.000	01:15	
Diagnostics Check		OK	01:15	

---

Data Download Date/Time: 07/21/2020 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: HALL,K  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 06/06/2020 01:34:14

---

Diagnostics Check		OK	01:17	
Air Blank		0.000	01:17	
Control Test		0.078	01:18	
Air Blank		0.000	01:18	
Subject Sample	# 1	0.139	01:21	Volume Not Met
Breath Volume		0.109	01:21	
Air Blank		0.000	01:22	
Air Blank		0.000	01:24	
Subject Sample	# 2	0.158	01:27	
Breath Volume		1.578	01:27	
Air Blank		0.000	01:27	
Air Blank		0.000	01:29	
Subject Sample	# 3	0.158	01:32	
Breath Volume		1.500	01:32	
Air Blank		0.000	01:33	
Control Test		0.078	01:33	
Air Blank		0.000	01:34	
Diagnostics Check		OK	01:34	

---

Data Download Date/Time: 07/21/2020 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: WASHINGTON,D  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 06/09/2020 02:42:03

---

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:26	
Control Test	0.080	02:26	
Air Blank	0.000	02:27	
Subject Sample	# 1 0.174	02:29	
Breath Volume	1.410	02:29	
Air Blank	0.000	02:29	
Air Blank	0.000	02:31	
Subject Sample	# 2 0.163	02:34	Volume Not Met
Breath Volume	0.660	02:34	
Air Blank	0.000	02:35	
Air Blank	0.000	02:37	
Subject Sample	# 3 0.161	02:40	Volume Not Met
Breath Volume	0.121	02:40	
Air Blank	0.000	02:41	
Control Test	0.079	02:41	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:42	

---

Data Download Date/Time: 07/21/2020 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: WASHINGTON,D  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 06/09/2020 02:58:23

---

Diagnostics Check		OK	02:44	
Air Blank		0.000	02:44	
Control Test		0.076	02:45	
Air Blank		0.000	02:45	
Subject Sample	# 1	0.187	02:48	Volume Not Met
Breath Volume		0.000	02:48	
Air Blank		0.000	02:49	
Air Blank		0.000	02:51	
Subject Sample	# 2	0.175	02:53	
Breath Volume		1.941	02:53	
Air Blank		0.000	02:54	
Air Blank		0.000	02:56	
Subject Sample	# 3	0.193	02:56	
Breath Volume		1.593	02:56	
Air Blank		0.000	02:57	
Control Test		0.075	02:57	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:58	

---

Data Download Date/Time: 07/21/2020 14:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: RODRIGUEZ,D  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 06/09/2020 21:16:27

---

Diagnostics Check	OK	21:09
Air Blank	0.000	21:10
Control Test	0.079	21:10
Air Blank	0.000	21:11
Subject Sample # 1	0.122	21:11
Breath Volume	1.230	21:11
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.120	21:14
Breath Volume	1.125	21:14
Air Blank	0.000	21:15
Control Test	0.078	21:15
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

---

Data Download Date/Time: 07/21/2020 14:10



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: ARCHEMBEAU,A  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: HERNANDEZ TAMAYO

DATE/TIME OF TEST: 06/10/2020 00:03:41

---

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.080	23:58
Air Blank	0.000	23:58
Subject Sample # 2	0.178	23:59
Breath Volume	1.824	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.182	00:02
Breath Volume	2.105	00:02
Air Blank	0.000	00:02
Control Test	0.079	00:03
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

---

Data Download Date/Time: 07/21/2020 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: TAMPA PD  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: MESA

DATE/TIME OF TEST: 06/11/2020 02:50:18

---

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.135	02:45
Breath Volume	1.960	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.138	02:48
Breath Volume	2.347	02:48
Air Blank	0.000	02:49
Control Test	0.079	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

---

Data Download Date/Time: 07/21/2020 14:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: FERGUSON,D  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 06/11/2020 03:21:44

---

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.079	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.000	03:17
Breath Volume	1.742	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.000	03:20
Breath Volume	1.117	03:20
Air Blank	0.000	03:20
Control Test	0.078	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

---

Data Download Date/Time: 07/21/2020 14:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: WOLSTENHOLM, A  
ARRESTING OFFICER AGENCY: FHP TROOP C  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 06/12/2020 02:24:09

---

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.202	02:19
Breath Volume	2.023	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.201	02:22
Breath Volume	1.710	02:22
Air Blank	0.000	02:23
Control Test	0.079	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

---

Data Download Date/Time: 07/21/2020 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: ARCHAMBEAU,M  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: BAUTISTA-PONCE

DATE/TIME OF TEST: 06/12/2020 21:40:24

---

Diagnostics Check	OK	21:33
Air Blank	0.000	21:34
Control Test	0.080	21:34
Air Blank	0.000	21:35
Subject Sample # 1	0.199	21:35
Breath Volume	1.765	21:35
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.205	21:38
Breath Volume	2.007	21:38
Air Blank	0.000	21:39
Control Test	0.079	21:39
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

---

Data Download Date/Time: 07/21/2020 14:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: TAMPA PD  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/14/2020 00:52:49

---

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.080	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.141	00:45
Breath Volume	1.175	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.134	00:51
Breath Volume	1.453	00:51
Air Blank	0.000	00:51
Control Test	0.078	00:52
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

---

Data Download Date/Time: 07/21/2020 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: TAMPA PD  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: LUGO ALVAREZ

DATE/TIME OF TEST: 06/14/2020 04:23:37

---

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.049	04:19
Breath Volume	1.367	04:19
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.049	04:22
Breath Volume	1.621	04:22
Air Blank	0.000	04:22
Control Test	0.079	04:23
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

---

Data Download Date/Time: 07/21/2020 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARNES AVERY,D,C  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: PATRON,A  
ARRESTING OFFICER AGENCY: TAMPA PD  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: OROZCO LARA

DATE/TIME OF TEST: 06/14/2020 04:42:40

---

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.079	04:37
Air Blank	0.000	04:37
Subject Sample # 1	0.127	04:38
Breath Volume	1.109	04:38
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.126	04:41
Breath Volume	1.578	04:41
Air Blank	0.000	04:41
Control Test	0.080	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

---

Data Download Date/Time: 07/21/2020 14:10



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: RODRIGUEZ,D  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ZELINSKAS

DATE/TIME OF TEST: 06/19/2020 02:27:21

---

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.079	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.191	02:22
Breath Volume	3.796	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.197	02:25
Breath Volume	4.347	02:25
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

---

Data Download Date/Time: 07/21/2020 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: BAAR,J  
ARRESTING OFFICER AGENCY: TAMPA PD  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: PENLEY

DATE/TIME OF TEST: 06/20/2020 02:55:43

---

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.079	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.172	02:50
Breath Volume	3.042	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.169	02:54
Breath Volume	3.335	02:54
Air Blank	0.000	02:54
Control Test	0.079	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

---

Data Download Date/Time: 07/21/2020 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DISMORE,NANCY,J  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: ANDREWS,S  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/21/2020 02:13:33

---

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.064	02:09
Breath Volume	2.957	02:09
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.058	02:12
Breath Volume	1.210	02:12
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

---

Data Download Date/Time: 07/21/2020 14:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: VISSER,A  
ARRESTING OFFICER AGENCY: TAMPA PD  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 06/26/2020 22:36:34

---

Diagnostics Check		OK	22:25	
Air Blank		0.000	22:25	
Control Test		0.080	22:25	
Air Blank		0.000	22:26	
Subject Sample	# 1	0.205	22:29	Volume Not Met
Breath Volume		0.222	22:29	
Air Blank		0.000	22:30	
Air Blank		0.000	22:31	
Subject Sample	# 2	0.201	22:35	Volume Not Met
Breath Volume		0.175	22:35	
Air Blank		0.000	22:35	
Control Test		0.078	22:35	
Air Blank		0.000	22:36	
Diagnostics Check		OK	22:36	

---

Data Download Date/Time: 07/21/2020 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,D,C  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: RENDON,D  
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: GONZALEZ QUINTERO

DATE/TIME OF TEST: 06/28/2020 23:18:19

---

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.080	23:12
Air Blank	0.000	23:12
Subject Sample # 1	0.114	23:14
Breath Volume	1.804	23:14
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.113	23:16
Breath Volume	1.843	23:16
Air Blank	0.000	23:17
Control Test	0.079	23:17
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

---

Data Download Date/Time: 07/21/2020 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: TAMPA PD  
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: GAFFNEY

DATE/TIME OF TEST: 06/29/2020 23:49:38

---

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.079	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.127	23:45
Breath Volume	3.742	23:45
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.129	23:48
Breath Volume	3.488	23:48
Air Blank	0.000	23:48
Control Test	0.079	23:49
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

---

Data Download Date/Time: 07/21/2020 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: CHESSHER,SANDRA  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 06/19/2020 02:20:11

---

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:16	
Control Test	0.078	02:16	
Air Blank	0.000	02:17	
Subject Sample	# 1	02:18	Interferent Detect
Breath Volume	2.324	02:18	
Air Blank	0.000	02:19	
Control Test	0.081	02:19	
Air Blank	0.000	02:20	
Diagnostics Check	OK	02:20	

---

Data Download Date/Time: 06/24/2020 18:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: CHESSHER,SANDRA  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 06/19/2020 02:29:07

---

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.078	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.149	02:23
Breath Volume	2.507	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.150	02:27
Breath Volume	2.359	02:27
Air Blank	0.000	02:28
Control Test	0.080	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

---

Data Download Date/Time: 06/24/2020 18:17



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALSH,SHANNON,L  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: CHESSHER,SANDI  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 06/20/2020 01:37:58

---

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.078	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.132	01:33
Breath Volume	1.125	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.140	01:36
Breath Volume	1.421	01:36
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

---

Data Download Date/Time: 06/24/2020 18:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: WALSH,SHANNON  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: O'KEEFE

DATE/TIME OF TEST: 06/20/2020 02:51:06

---

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:48	
Control Test	0.078	02:49	
Air Blank	0.000	02:49	
Subject Sample	# 1	02:50	Subject Test Refused
Breath Volume	0.000	02:50	
Air Blank	0.000	02:50	
		02:50	RFI Detect
Air Blank	0.000	02:51	

---

Data Download Date/Time: 06/24/2020 18:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: WALSH,SHANNON  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: OKEEFE

DATE/TIME OF TEST: 06/20/2020 02:56:49

---

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:54	
Control Test	0.076	02:54	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:55	Subject Test Refused
Breath Volume	0.000	02:55	
Air Blank	0.000	02:55	
Control Test	0.079	02:56	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

---

Data Download Date/Time: 06/24/2020 18:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINDER,BRONSON,M  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: CHESSHER,SANDRA  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 06/24/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 970670  
DRY GAS EXPIRY DATE: 06/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: HASSEY

DATE/TIME OF TEST: 06/26/2020 00:16:49

---

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.078	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.147	00:12
Breath Volume	2.218	00:12
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.156	00:15
Breath Volume	1.894	00:15
Air Blank	0.000	00:15
Control Test	0.078	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

---

Data Download Date/Time: 07/11/2020 00:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: WEBER,JEFFREY  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 06/05/2020 06:11:42

---

Diagnostics Check	OK	06:05
Air Blank	0.000	06:06
Control Test	0.080	06:06
Air Blank	0.000	06:06
Subject Sample # 1	0.191	06:07
Breath Volume	1.136	06:07
Air Blank	0.000	06:07
Air Blank	0.000	06:09
Subject Sample # 2	0.188	06:10
Breath Volume	1.363	06:10
Air Blank	0.000	06:10
Control Test	0.080	06:11
Air Blank	0.000	06:11
Diagnostics Check	OK	06:11

---

Data Download Date/Time: 07/06/2020 09:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSS,ROBERT,R  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MOSS,ROBERT  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: DELOACH

DATE/TIME OF TEST: 06/06/2020 22:17:22

---

Diagnostics Check	OK	22:14	
Air Blank	0.000	22:14	
Control Test	0.080	22:15	
Air Blank	0.000	22:15	
Subject Sample	# 1	22:16	Subject Test Refused
Breath Volume	0.000	22:16	
Air Blank	0.000	22:16	
Control Test	0.080	22:16	
Air Blank	0.000	22:17	
Diagnostics Check	OK	22:17	

---

Data Download Date/Time: 07/06/2020 09:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: WEBER,JEFFREY  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: HUGGINS

DATE/TIME OF TEST: 06/07/2020 00:22:13

---

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.080	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.149	00:17
Breath Volume	1.960	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.151	00:20
Breath Volume	2.800	00:20
Air Blank	0.000	00:21
Control Test	0.080	00:21
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

---

Data Download Date/Time: 07/06/2020 09:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: WEBER,JEFFREY  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: INGLE

DATE/TIME OF TEST: 06/13/2020 04:41:07

---

Diagnostics Check	OK	04:38	
Air Blank	0.000	04:38	
Control Test	0.081	04:39	
Air Blank	0.000	04:39	
Subject Sample	# 1	04:39	Subject Test Refused
Breath Volume	0.000	04:39	
Air Blank	0.000	04:40	
Control Test	0.080	04:40	
Air Blank	0.000	04:41	
Diagnostics Check	OK	04:41	

---

Data Download Date/Time: 07/06/2020 09:36



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: WEBER,JEFFREY  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: PETSKEI

DATE/TIME OF TEST: 06/15/2020 05:22:54

---

Diagnostics Check	OK	05:20	
Air Blank	0.000	05:20	
Control Test	0.080	05:20	
Air Blank	0.000	05:21	
Subject Sample	# 1	05:21	Subject Test Refused
Breath Volume	0.000	05:21	
Air Blank	0.000	05:22	
Control Test	0.080	05:22	
Air Blank	0.000	05:22	
Diagnostics Check	OK	05:22	

---

Data Download Date/Time: 07/06/2020 09:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,STEPHANIE,N  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: THOMAS,STEPHANIE  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 06/18/2020 19:12:40

---

Diagnostics Check	OK	19:02
Air Blank	0.000	19:03
Control Test	0.080	19:03
Air Blank	0.000	19:04
Subject Sample # 1	0.210	19:06
Breath Volume	1.679	19:06
Air Blank	0.000	19:07
Air Blank	0.000	19:09
Subject Sample # 2	0.205	19:11
Breath Volume	1.214	19:11
Air Blank	0.000	19:11
Control Test	0.079	19:12
Air Blank	0.000	19:12
Diagnostics Check	OK	19:12

---

Data Download Date/Time: 07/06/2020 09:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PELFREY,PAUL,D  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MCKEEBY,WADE  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ROCHELLE

DATE/TIME OF TEST: 06/20/2020 04:17:58

---

Diagnostics Check		OK	04:11
Air Blank		0.000	04:11
Control Test		0.079	04:11
Air Blank		0.000	04:12
Subject Sample	# 1	0.135	04:13
Breath Volume		2.000	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:15
Subject Sample	# 2	0.136	04:16
Breath Volume		1.648	04:16
Air Blank		0.000	04:17
Control Test		0.079	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

---

Data Download Date/Time: 07/06/2020 09:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSS,ROBERT,R  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MOSS,ROBERT  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/20/2020 20:31:46

---

Diagnostics Check	OK	20:24	
Air Blank	0.000	20:25	
Control Test	0.080	20:25	
Air Blank	0.000	20:26	
Subject Sample # 1	0.134	20:27	
Breath Volume	1.187	20:27	
Air Blank	0.000	20:27	
Air Blank	0.000	20:29	
Subject Sample # 2	0.128	20:30	
Breath Volume	1.257	20:30	
Air Blank	0.000	20:30	
Control Test	0.080	20:31	
Air Blank		20:31	RFI Detect
Air Blank	0.000	20:31	

---

Data Download Date/Time: 07/06/2020 09:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSS,R,R  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: NEWSOME,DONNIE  
ARRESTING OFFICER AGENCY: Naval Lab Pana City  
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: ALONSO

DATE/TIME OF TEST: 06/27/2020 22:01:45

---

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.186	21:57
Breath Volume	1.742	21:57
Air Blank	0.000	21:58
Air Blank	0.000	21:59
Subject Sample # 2	0.182	22:00
Breath Volume	1.500	22:00
Air Blank	0.000	22:00
Control Test	0.079	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

---

Data Download Date/Time: 07/06/2020 09:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: MARTIN VILLA

DATE/TIME OF TEST: 06/03/2020 00:11:20

---

Diagnostics Check		OK	00:03
Air Blank		0.000	00:03
Control Test		0.079	00:03
Air Blank		0.000	00:04
Subject Sample	# 1	0.103	00:06
Breath Volume		2.621	00:06
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.123	00:09
Breath Volume		2.148	00:09
Air Blank		0.000	00:10
Control Test		0.080	00:10
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

---

Data Download Date/Time: 06/03/2020 11:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: DOTSON,WILLIAM  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 12:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:57

SUBJECT'S LAST NAME: LATSHAW

DATE/TIME OF TEST: 06/03/2020 13:23:59

---

Diagnostics Check	OK	13:17
Air Blank	0.000	13:18
Control Test	0.080	13:18
Air Blank	0.000	13:19
Subject Sample # 1	0.000	13:19
Breath Volume	1.664	13:19
Air Blank	0.000	13:20
Air Blank	0.000	13:22
Subject Sample # 2	0.000	13:22
Breath Volume	1.542	13:22
Air Blank	0.000	13:23
Control Test	0.080	13:23
Air Blank	0.000	13:23
Diagnostics Check	OK	13:23

---

Data Download Date/Time: 07/09/2020 12:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FILLAK, MICHAEL, O  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: PRIETO PEREZ, CARLOS  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: DA EGLING SOCARRAS

DATE/TIME OF TEST: 06/05/2020 00:47:17

---

Diagnostics Check	OK	00:40
Air Blank	0.000	00:40
Control Test	0.080	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.186	00:42
Breath Volume	1.859	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.185	00:45
Breath Volume	3.597	00:45
Air Blank	0.000	00:46
Control Test	0.079	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

---

Data Download Date/Time: 07/09/2020 12:33



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: SHIPPY,DYLAN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: PRESCOTT

DATE/TIME OF TEST: 06/06/2020 02:57:01

---

Diagnostics Check		OK	02:49
Air Blank		0.000	02:49
Control Test		0.080	02:50
Air Blank		0.000	02:50
Subject Sample	# 1	0.187	02:52
Breath Volume		2.281	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.190	02:55
Breath Volume		1.468	02:55
Air Blank		0.000	02:56
Control Test		0.079	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:57

---

Data Download Date/Time: 07/09/2020 12:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,DANNY,W  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: PEREZ,LUIS  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: PERRY-MAZZETTI

DATE/TIME OF TEST: 06/06/2020 21:26:10

---

Diagnostics Check	OK	21:19
Air Blank	0.000	21:20
Control Test	0.081	21:20
Air Blank	0.000	21:21
Subject Sample # 1	0.000	21:21
Breath Volume	1.214	21:21
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample # 2	0.000	21:24
Breath Volume	1.507	21:24
Air Blank	0.000	21:25
Control Test	0.080	21:25
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

---

Data Download Date/Time: 07/09/2020 12:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,DANNY,W  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: BUSCH,CAROLINE  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: HILLEBOLD

DATE/TIME OF TEST: 06/06/2020 23:32:32

---

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.077	23:27
Air Blank	0.000	23:27
Subject Sample # 1	0.178	23:28
Breath Volume	1.285	23:28
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.174	23:31
Breath Volume	1.410	23:31
Air Blank	0.000	23:31
Control Test	0.077	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

---

Data Download Date/Time: 07/09/2020 12:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: LOWE,ALEXANDER  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: GUSTAVSEN

DATE/TIME OF TEST: 06/07/2020 01:46:53

---

Diagnostics Check	OK	01:39
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.173	01:41
Breath Volume	1.585	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.179	01:45
Breath Volume	1.527	01:45
Air Blank	0.000	01:45
Control Test	0.080	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

---

Data Download Date/Time: 07/09/2020 12:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,ALFONSO,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: SHIPPY,DYLAN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/08/2020 01:51:30

---

Diagnostics Check		OK	01:39	
Air Blank		0.000	01:40	
Control Test		0.080	01:40	
Air Blank		0.000	01:41	
Subject Sample	# 1	0.163	01:44	Volume Not Met
Breath Volume		0.191	01:44	
Air Blank		0.000	01:45	
Air Blank		0.000	01:46	
Subject Sample	# 2	0.154	01:50	Volume Not Met
Breath Volume		0.101	01:50	
Air Blank		0.000	01:50	
Control Test		0.079	01:50	
Air Blank		0.000	01:51	
Diagnostics Check		OK	01:51	

---

Data Download Date/Time: 07/09/2020 12:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: BARKLEY,JEFFREY  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: TURPIN JR

DATE/TIME OF TEST: 06/08/2020 18:38:16

---

Diagnostics Check	OK	18:31
Air Blank	0.000	18:32
Control Test	0.079	18:32
Air Blank	0.000	18:33
Subject Sample # 1	0.133	18:33
Breath Volume	1.167	18:33
Air Blank	0.000	18:34
Air Blank	0.000	18:36
Subject Sample # 2	0.144	18:36
Breath Volume	2.195	18:36
Air Blank	0.000	18:37
Control Test	0.079	18:37
Air Blank	0.000	18:38
Diagnostics Check	OK	18:38

---

Data Download Date/Time: 07/09/2020 12:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,SAMANTHA,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: BROWN,SAMANTHA  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: GAVIN

DATE/TIME OF TEST: 06/10/2020 03:19:37

---

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.184	03:15
Breath Volume	2.101	03:15
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.182	03:18
Breath Volume	1.753	03:18
Air Blank	0.000	03:18
Control Test	0.080	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

---

Data Download Date/Time: 07/09/2020 12:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: FRITZ,TYLER  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: BEVELLE

DATE/TIME OF TEST: 06/17/2020 19:13:03

---

Diagnostics Check	OK	19:06
Air Blank	0.000	19:06
Control Test	0.079	19:07
Air Blank	0.000	19:07
Subject Sample # 1	0.188	19:08
Breath Volume	1.191	19:08
Air Blank	0.000	19:08
Air Blank	0.000	19:10
Subject Sample # 2	0.205	19:11
Breath Volume	1.453	19:11
Air Blank	0.000	19:12
Control Test	0.079	19:12
Air Blank	0.000	19:12
Diagnostics Check	OK	19:13

---

Data Download Date/Time: 07/09/2020 12:33



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 06/20/2020 02:08:37

---

Diagnostics Check		OK	01:57	
Air Blank		0.000	01:57	
Control Test		0.080	01:57	
Air Blank		0.000	01:58	
Subject Sample	# 1	0.000	02:01	Volume Not Met
Breath Volume		0.000	02:01	
Air Blank		0.000	02:01	
Air Blank		0.000	02:03	
Subject Sample	# 2	0.153	02:07	Volume Not Met
Breath Volume		0.953	02:07	
Air Blank		0.000	02:07	
Control Test		0.080	02:08	
Air Blank		0.000	02:08	
Diagnostics Check		OK	02:08	

---

Data Download Date/Time: 07/09/2020 12:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 06/20/2020 02:19:42

---

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.079	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.141	02:12
Breath Volume	2.082	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.143	02:18
Breath Volume	1.316	02:18
Air Blank	0.000	02:18
Control Test	0.079	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

---

Data Download Date/Time: 07/09/2020 12:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,ALFONSO,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: SCARLATO,MICHAEL  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: FERMIN

DATE/TIME OF TEST: 06/20/2020 03:17:51

---

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:11	
Control Test	0.080	03:11	
Air Blank	0.000	03:11	
Subject Sample	# 1 0.127	03:13	
Breath Volume	1.179	03:13	
Air Blank	0.000	03:13	
Air Blank	0.000	03:15	
Subject Sample	# 2	03:16	Subject Test Refused
Breath Volume	0.000	03:16	
Air Blank	0.000	03:16	
Control Test	0.079	03:17	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

---

Data Download Date/Time: 07/09/2020 12:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARSON,JAMES,A  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: CARSON,JAMES  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 13:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 06/23/2020 14:55:33

---

Diagnostics Check	OK	14:47
Air Blank	0.000	14:47
Control Test	0.081	14:48
Air Blank	0.000	14:48
Subject Sample # 1	0.000	14:50
Breath Volume	2.082	14:50
Air Blank	0.000	14:51
Air Blank	0.000	14:53
Subject Sample # 2	0.000	14:54
Breath Volume	2.296	14:54
Air Blank	0.000	14:54
Control Test	0.080	14:54
Air Blank	0.000	14:55
Diagnostics Check	OK	14:55

---

Data Download Date/Time: 07/09/2020 12:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANTRELL,ROBERT,V  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: DOTSON,WILLIAM  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 12:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:08

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 06/26/2020 13:35:10

---

Diagnostics Check	OK	13:29
Air Blank	0.000	13:29
Control Test	0.080	13:29
Air Blank	0.000	13:30
Subject Sample # 1	0.000	13:30
Breath Volume	2.648	13:30
Air Blank	0.000	13:31
Air Blank	0.000	13:33
Subject Sample # 2	0.000	13:33
Breath Volume	1.929	13:33
Air Blank	0.000	13:34
Control Test	0.080	13:34
Air Blank	0.000	13:35
Diagnostics Check	OK	13:35

---

Data Download Date/Time: 07/09/2020 12:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MADURO,VICTOR,A  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MADURO,VICTOR  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A2  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BRODZIK

DATE/TIME OF TEST: 06/27/2020 02:58:09

---

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:55	
Control Test	0.080	02:56	
Air Blank	0.000	02:56	
Subject Sample	# 1	02:56	Subject Test Refused
Breath Volume	0.000	02:56	
Air Blank	0.000	02:57	
Control Test	0.080	02:57	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

---

Data Download Date/Time: 07/09/2020 12:33