

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: COTUGNO

DATE/TIME OF TEST: 06/05/2020 19:59:57

Diagnostics Check	OK	19:56	
Air Blank	0.000	19:56	
Control Test	0.079	19:57	
Air Blank	0.000	19:57	
Subject Sample	# 1	19:58	Interferent Detect
Breath Volume	2.621	19:58	
Air Blank	0.000	19:59	
Control Test	0.081	19:59	
Air Blank	0.000	19:59	
Diagnostics Check	OK	19:59	

Data Download Date/Time: 06/24/2020 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: COTUGNO

DATE/TIME OF TEST: 06/05/2020 20:05:30

Diagnostics Check	OK	20:02	
Air Blank	0.000	20:02	
Control Test	0.080	20:02	
Air Blank	0.000	20:03	
Subject Sample	# 1	20:03	Interferent Detect
Breath Volume	1.375	20:03	
Air Blank	0.000	20:04	
Control Test	0.080	20:04	
Air Blank	0.000	20:05	
Diagnostics Check	OK	20:05	

Data Download Date/Time: 06/24/2020 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: PEREZ BARTOLON

DATE/TIME OF TEST: 06/06/2020 22:26:02

Diagnostics Check	OK	22:18
Air Blank	0.000	22:19
Control Test	0.081	22:19
Air Blank	0.000	22:20
Subject Sample # 1	0.317	22:21
Breath Volume	2.148	22:21
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.322	22:24
Breath Volume	2.335	22:24
Air Blank	0.000	22:25
Control Test	0.079	22:25
Air Blank	0.000	22:25
Diagnostics Check	OK	22:26

Data Download Date/Time: 06/24/2020 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,W,L
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: RAMIREZ,W
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:01

SUBJECT'S LAST NAME: MIGUEL

DATE/TIME OF TEST: 06/13/2020 18:59:04

Diagnostics Check	OK	18:47		
Air Blank	0.000	18:48		
Control Test	0.082	18:48		
Air Blank	0.000	18:49		
Subject Sample	# 1	18:52	No Sample Provided	
Breath Volume	0.000	18:52		
Air Blank	0.000	18:52		
Air Blank	0.000	18:54		
Subject Sample	# 2	0.188	18:57	Volume Not Met
Breath Volume	0.136	18:57		
Air Blank		18:58		Purge Fail
Air Blank	0.000	18:59		

Data Download Date/Time: 06/24/2020 09:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,W,L
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: RAMIREZ,W
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:01

SUBJECT'S LAST NAME: MIGUEL

DATE/TIME OF TEST: 06/13/2020 19:07:56

Diagnostics Check	OK	19:05	
Air Blank	0.000	19:05	
Control Test	0.077	19:05	
Air Blank	0.000	19:06	
Subject Sample	# 1	19:06	Subject Test Refused
Breath Volume	0.000	19:06	
Air Blank	0.000	19:07	
Control Test	0.079	19:07	
Air Blank	0.000	19:07	
Diagnostics Check	OK	19:07	

Data Download Date/Time: 06/24/2020 09:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,W,L
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: ROGERS,W
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: SIERRA SERRATO

DATE/TIME OF TEST: 06/13/2020 22:05:41

Diagnostics Check	OK	21:59	
Air Blank	0.000	22:00	
Control Test	0.081	22:00	
Air Blank	0.000	22:01	
Subject Sample # 1	0.226	22:04	Volume Not Met
Breath Volume	0.050	22:04	
Air Blank		22:05	Purge Fail
Air Blank	0.000	22:05	

Data Download Date/Time: 06/24/2020 09:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,W,L
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: ROGERS,W
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: SIERRA SERRATO

DATE/TIME OF TEST: 06/13/2020 22:19:34

Diagnostics Check		OK	22:10	
Air Blank		0.000	22:11	
Control Test		0.076	22:11	
Air Blank		0.000	22:12	
Subject Sample	# 1	0.226	22:15	Volume Not Met
Breath Volume		0.156	22:15	
Air Blank		0.000	22:16	
Air Blank		0.000	22:17	
Subject Sample	# 2		22:18	Subject Test Refused
Breath Volume		0.000	22:18	
Air Blank		0.000	22:18	
Control Test		0.076	22:18	
Air Blank		0.000	22:19	
Diagnostics Check		OK	22:19	

Data Download Date/Time: 06/24/2020 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,W,L
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: ROGERS,W
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 06/21/2020 23:35:36

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.080	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.187	23:30
Breath Volume	1.468	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:32
Subject Sample # 2	0.192	23:33
Breath Volume	2.074	23:33
Air Blank	0.000	23:34
Control Test	0.078	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 06/24/2020 09:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,W,L
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: ROGERS,W
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 06/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 06/25/2020 22:21:22

Diagnostics Check		OK	22:11	
Air Blank		0.000	22:11	
Control Test		0.080	22:12	
Air Blank		0.000	22:12	
Subject Sample	# 1	0.196	22:15	Volume Not Met
Breath Volume		0.000	22:15	
Air Blank		0.000	22:16	
Air Blank		0.000	22:18	
Subject Sample	# 2		22:19	Subject Test Refused
Breath Volume		0.175	22:19	
Air Blank		0.000	22:20	
Control Test		0.077	22:20	
Air Blank		0.000	22:21	
Diagnostics Check		OK	22:21	

Data Download Date/Time: 07/27/2020 18:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: COMBS

DATE/TIME OF TEST: 06/05/2020 00:00:31

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:53	
Control Test	0.077	23:53	
Air Blank	0.000	23:54	
Subject Sample	# 1 0.220	23:55	
Breath Volume	1.488	23:55	
Air Blank	0.000	23:56	
Air Blank	0.000	23:57	
Subject Sample	# 2 0.211	23:58	
Breath Volume	1.523	23:58	
Air Blank	0.000	23:59	
Control Test	0.074	23:59	Control Outside Tolerance
Air Blank	0.000	00:00	

Data Download Date/Time: 06/30/2020 22:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: COMBS

DATE/TIME OF TEST: 06/05/2020 00:04:53

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:03	
Control Test	0.074	00:04	Control Outside Tolerance
Air Blank	0.000	00:04	

Data Download Date/Time: 06/30/2020 22:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KIP,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PETERSON,KIP
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: PAGAN

DATE/TIME OF TEST: 06/06/2020 01:03:22

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.075	00:57
Air Blank		0.000	00:58
Subject Sample	# 1	0.262	00:58
Breath Volume		1.769	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.267	01:01
Breath Volume		2.109	01:01
Air Blank		0.000	01:02
Control Test		0.075	01:02
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 06/30/2020 22:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: SNEED

DATE/TIME OF TEST: 06/14/2020 00:36:47

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.077	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.117	00:32
Breath Volume	3.289	00:32
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.112	00:35
Breath Volume	3.386	00:35
Air Blank	0.000	00:35
Control Test	0.076	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 06/30/2020 22:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 06/27/2020 01:50:57

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.077	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.215	01:46
Breath Volume		1.292	01:46
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.220	01:49
Breath Volume		1.464	01:49
Air Blank		0.000	01:49
Control Test		0.076	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 06/30/2020 22:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDI
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 06/13/2020 01:32:24

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.077	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.178	01:27
Breath Volume	1.890	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.160	01:30
Breath Volume	1.464	01:30
Air Blank	0.000	01:31
Control Test	0.076	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 06/24/2020 17:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIALOBRZESKI,M,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIALOBRZESKI,M
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 06/25/2020 01:39:51

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.000	01:35
Breath Volume	1.789	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.000	01:38
Breath Volume	1.226	01:38
Air Blank	0.000	01:38
Control Test	0.081	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 07/31/2020 23:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,EVEN,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: MORALES,RAFAEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: OSCEOLA

DATE/TIME OF TEST: 06/22/2020 23:24:53

Diagnostics Check	OK	23:19	
Air Blank	0.000	23:20	
Control Test	0.080	23:20	
Air Blank	0.000	23:21	
Subject Sample	# 1	23:23	Subject Test Refused
Breath Volume	0.000	23:23	
Air Blank	0.000	23:23	
Control Test	0.081	23:24	
Air Blank	0.000	23:24	
Diagnostics Check	OK	23:24	

Data Download Date/Time: 06/23/2020 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2020
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: HERNANDEZ LOPEZ

DATE/TIME OF TEST: 06/19/2020 22:40:58

Diagnostics Check	OK	22:33
Air Blank	0.000	22:34
Control Test	0.078	22:34
Air Blank	0.000	22:35
Subject Sample # 1	0.064	22:36
Breath Volume	2.363	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:38
Subject Sample # 2	0.063	22:39
Breath Volume	1.796	22:39
Air Blank	0.000	22:40
Control Test	0.078	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 07/16/2020 13:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: PANDYA

DATE/TIME OF TEST: 06/01/2020 23:53:53

Diagnostics Check	OK	23:48
Air Blank	0.000	23:48
Control Test	0.081	23:48
Air Blank	0.000	23:49
Subject Sample # 1	0.000	23:49
Breath Volume	1.136	23:49
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.000	23:52
Breath Volume	1.113	23:52
Air Blank	0.000	23:52
Control Test	0.081	23:53
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 06/26/2020 16:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CASTRO VAIL

DATE/TIME OF TEST: 06/03/2020 03:23:57

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.226	03:19
Breath Volume	2.136	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:21
Subject Sample # 2	0.228	03:22
Breath Volume	2.652	03:22
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCGILLICUDDY,STEVEN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: WARNER

DATE/TIME OF TEST: 06/03/2020 03:52:25

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:49	
Control Test	0.080	03:49	
Air Blank	0.000	03:50	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:51	
Control Test	0.080	03:51	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:52	

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH, GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: YOO

DATE/TIME OF TEST: 06/03/2020 06:24:01

Diagnostics Check		OK	06:12	
Air Blank		0.000	06:12	
Control Test		0.081	06:12	
Air Blank		0.000	06:13	
Subject Sample	# 1	0.119	06:16	Volume Not Met
Breath Volume		0.253	06:16	
Air Blank		0.000	06:17	
Air Blank		0.000	06:19	
Subject Sample	# 2	0.141	06:19	
Breath Volume		2.085	06:19	
Air Blank		0.000	06:20	
Air Blank		0.000	06:22	
Subject Sample	# 3	0.148	06:22	
Breath Volume		1.695	06:22	
Air Blank		0.000	06:23	
Control Test		0.080	06:23	
Air Blank		0.000	06:23	
Diagnostics Check		OK	06:24	

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BUDISH, HARISON
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 06/03/2020 22:57:59

Diagnostics Check	OK	22:51
Air Blank	0.000	22:52
Control Test	0.081	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.000	22:53
Breath Volume	1.957	22:53
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.000	22:56
Breath Volume	1.402	22:56
Air Blank	0.000	22:57
Control Test	0.081	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CASTRO,DENNIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: WILDERIII

DATE/TIME OF TEST: 06/04/2020 17:03:19

Diagnostics Check	OK	16:57
Air Blank	0.000	16:57
Control Test	0.081	16:57
Air Blank	0.000	16:58
Subject Sample # 1	0.000	16:58
Breath Volume	1.394	16:58
Air Blank	0.000	16:59
Air Blank	0.000	17:01
Subject Sample # 2	0.000	17:01
Breath Volume	2.585	17:01
Air Blank	0.000	17:02
Control Test	0.081	17:02
Air Blank	0.000	17:03
Diagnostics Check	OK	17:03

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SEMPLE,WILLIAM
ARRESTING OFFICER AGENCY: Highland Beach PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:32

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 06/04/2020 18:16:09

Diagnostics Check	OK	18:09
Air Blank	0.000	18:10
Control Test	0.081	18:10
Air Blank	0.000	18:10
Subject Sample # 1	0.156	18:11
Breath Volume	1.179	18:11
Air Blank	0.000	18:11
Air Blank	0.000	18:13
Subject Sample # 2	0.152	18:14
Breath Volume	1.816	18:14
Air Blank	0.000	18:15
Control Test	0.080	18:15
Air Blank	0.000	18:16
Diagnostics Check	OK	18:16

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CASAS,JAVIER
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: SOPOCHANA

DATE/TIME OF TEST: 06/05/2020 01:27:43

Diagnostics Check		OK	01:18	
Air Blank		0.000	01:19	
Control Test		0.081	01:19	
Air Blank		0.000	01:19	
Subject Sample	# 1	0.144	01:23	Volume Not Met
Breath Volume		0.632	01:23	
Air Blank		0.000	01:23	
Air Blank		0.000	01:25	
Subject Sample	# 2		01:26	Subject Test Refused
Breath Volume		0.000	01:26	
Air Blank		0.000	01:26	
Control Test		0.080	01:27	
Air Blank		0.000	01:27	
Diagnostics Check		OK	01:27	

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH, GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GIRON BARRIOS

DATE/TIME OF TEST: 06/05/2020 03:27:43

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.081	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.141	03:22
Breath Volume	2.886	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.143	03:26
Breath Volume	2.554	03:26
Air Blank	0.000	03:26
Control Test	0.081	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAUNG,NGIN
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/05/2020 19:46:42

Diagnostics Check	OK	19:38
Air Blank	0.000	19:39
Control Test	0.081	19:39
Air Blank	0.000	19:40
Subject Sample # 1	0.259	19:41
Breath Volume	1.648	19:41
Air Blank	0.000	19:42
Air Blank	0.000	19:43
Subject Sample # 2	0.247	19:45
Breath Volume	1.945	19:45
Air Blank	0.000	19:45
Control Test	0.079	19:46
Air Blank	0.000	19:46
Diagnostics Check	OK	19:46

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: HERNANDEZ LOPEZ

DATE/TIME OF TEST: 06/06/2020 03:34:56

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.081	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.231	03:30
Breath Volume	2.140	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:32
Subject Sample # 2	0.226	03:33
Breath Volume	2.609	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAVARES,LUCAS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 06/06/2020 23:16:03

Diagnostics Check	OK	23:09
Air Blank	0.000	23:10
Control Test	0.081	23:10
Air Blank	0.000	23:11
Subject Sample # 1	0.130	23:11
Breath Volume	1.359	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.131	23:14
Breath Volume	1.179	23:14
Air Blank	0.000	23:15
Control Test	0.081	23:15
Air Blank	0.000	23:15
Diagnostics Check	OK	23:16

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: VELASQUEZ ARREAGA

DATE/TIME OF TEST: 06/07/2020 02:25:55

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.081	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.149	02:20
Breath Volume	2.082	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.136	02:24
Breath Volume	1.171	02:24
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: GRANADA GARCIA

DATE/TIME OF TEST: 06/10/2020 04:02:34

Diagnostics Check	OK	03:54
Air Blank	0.000	03:55
Control Test	0.081	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.217	03:56
Breath Volume	1.269	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.221	04:00
Breath Volume	2.832	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: THOMAS,MICAH
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: VELASQUEZ MAZARIEGOS

DATE/TIME OF TEST: 06/11/2020 00:32:15

Diagnostics Check		OK	00:22	
Air Blank		0.000	00:22	
Control Test		0.081	00:23	
Air Blank		0.000	00:23	
Subject Sample	# 1	0.223	00:26	Volume Not Met
Breath Volume		0.019	00:26	
Air Blank		0.000	00:27	
Air Blank		0.000	00:29	
Subject Sample	# 2		00:30	Subject Test Refused
Breath Volume		0.000	00:30	
Air Blank		0.000	00:31	
Control Test		0.079	00:31	
Air Blank		0.000	00:32	
Diagnostics Check		OK	00:32	

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TUANG,NGIN
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 06/11/2020 04:17:02

Diagnostics Check		OK	04:01	
Air Blank		0.000	04:01	
Control Test		0.080	04:01	
Air Blank		0.000	04:02	
Subject Sample	# 1	0.132	04:05	Volume Not Met
Breath Volume		0.113	04:05	
Air Blank		0.000	04:06	
Air Blank		0.000	04:07	
Subject Sample	# 2	0.170	04:09	
Breath Volume		1.175	04:09	
Air Blank		0.000	04:10	
Air Blank		0.000	04:12	
Subject Sample	# 3	0.160	04:15	
Breath Volume		2.070	04:15	
Air Blank		0.000	04:16	
Control Test		0.080	04:16	
Air Blank		0.000	04:16	
Diagnostics Check		OK	04:17	

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GRUEBEL,K
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:25

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 06/11/2020 09:54:42

Diagnostics Check	OK	09:48
Air Blank	0.000	09:48
Control Test	0.081	09:49
Air Blank	0.000	09:49
Subject Sample # 1	0.000	09:50
Breath Volume	2.542	09:50
Air Blank	0.000	09:50
Air Blank	0.000	09:52
Subject Sample # 2	0.000	09:53
Breath Volume	1.992	09:53
Air Blank	0.000	09:53
Control Test	0.080	09:54
Air Blank	0.000	09:54
Diagnostics Check	OK	09:54

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: HASLETT

DATE/TIME OF TEST: 06/11/2020 23:00:43

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.080	22:54
Air Blank	0.000	22:54
Subject Sample # 1	0.000	22:56
Breath Volume	2.164	22:56
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.000	22:59
Breath Volume	1.625	22:59
Air Blank	0.000	22:59
Control Test	0.078	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTER, ERIC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 06/12/2020 03:22:31

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.081	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.133	03:16
Breath Volume	2.167	03:16
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.128	03:20
Breath Volume	2.546	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MACK,C
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: WHYTE

DATE/TIME OF TEST: 06/12/2020 17:32:37

Diagnostics Check	OK	17:26
Air Blank	0.000	17:26
Control Test	0.081	17:27
Air Blank	0.000	17:27
Subject Sample # 1	0.000	17:28
Breath Volume	1.187	17:28
Air Blank	0.000	17:28
Air Blank	0.000	17:30
Subject Sample # 2	0.000	17:31
Breath Volume	1.792	17:31
Air Blank	0.000	17:31
Control Test	0.081	17:32
Air Blank	0.000	17:32
Diagnostics Check	OK	17:32

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CASTRO,DENNIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/12/2020 21:17:21

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.080	21:11
Air Blank	0.000	21:11
Subject Sample # 1	0.161	21:12
Breath Volume	1.410	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.159	21:15
Breath Volume	1.378	21:15
Air Blank	0.000	21:16
Control Test	0.080	21:16
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MENDEZ,JOSE
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: KOPINS

DATE/TIME OF TEST: 06/12/2020 23:38:46

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.081	23:32
Air Blank	0.000	23:33
Subject Sample # 1	0.150	23:33
Breath Volume	3.011	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.151	23:37
Breath Volume	2.957	23:37
Air Blank	0.000	23:37
Control Test	0.081	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROMERO,JUAN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: TANNER

DATE/TIME OF TEST: 06/13/2020 01:03:10

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:00	
Control Test	0.081	01:00	
Air Blank	0.000	01:01	
Subject Sample	# 1	01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:02	
Control Test	0.080	01:02	
Air Blank	0.000	01:03	
Diagnostics Check	OK	01:03	

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: BATISTA

DATE/TIME OF TEST: 06/13/2020 03:15:57

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.081	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.162	03:11
Breath Volume		2.000	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:13
Subject Sample	# 2	0.178	03:14
Breath Volume		1.175	03:14
Air Blank		0.000	03:15
Control Test		0.080	03:15
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CRUZ, GARDENIA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:35

SUBJECT'S LAST NAME: ZEITLER

DATE/TIME OF TEST: 06/13/2020 14:14:06

Diagnostics Check		OK	14:01	
Air Blank		0.000	14:02	
Control Test		0.081	14:02	
Air Blank		0.000	14:02	
Subject Sample	# 1	0.050	14:06	Volume Not Met
Breath Volume		0.890	14:06	
Air Blank		0.000	14:06	
Air Blank		0.000	14:08	
Subject Sample	# 2	0.053	14:09	
Breath Volume		2.394	14:09	
Air Blank		0.000	14:09	
Air Blank		0.000	14:11	
Subject Sample	# 3	0.051	14:12	
Breath Volume		2.125	14:12	
Air Blank		0.000	14:13	
Control Test		0.080	14:13	
Air Blank		0.000	14:14	
Diagnostics Check		OK	14:14	

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: CARRION

DATE/TIME OF TEST: 06/13/2020 22:35:35

Diagnostics Check		OK	22:28
Air Blank		0.000	22:29
Control Test		0.081	22:29
Air Blank		0.000	22:30
Subject Sample	# 1	0.043	22:30
Breath Volume		2.031	22:30
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.043	22:34
Breath Volume		2.441	22:34
Air Blank		0.000	22:34
Control Test		0.080	22:34
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER,CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: HATTON

DATE/TIME OF TEST: 06/14/2020 01:21:04

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.081	01:15
Air Blank	0.000	01:15
Subject Sample # 1	0.147	01:16
Breath Volume	3.535	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:18
Subject Sample # 2	0.151	01:19
Breath Volume	3.519	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:20
Diagnostics Check	OK	01:21

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCGILLICUDDY,STEVEN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 06/14/2020 04:46:12

Diagnostics Check	OK	04:39
Air Blank	0.000	04:40
Control Test	0.081	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.192	04:41
Breath Volume	2.109	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:43
Subject Sample # 2	0.183	04:44
Breath Volume	2.238	04:44
Air Blank	0.000	04:45
Control Test	0.080	04:45
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: D'ANDREA,ADAM
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: LENZ

DATE/TIME OF TEST: 06/14/2020 08:37:32

Diagnostics Check		OK	08:30
Air Blank		0.000	08:30
Control Test		0.081	08:31
Air Blank		0.000	08:31
Subject Sample	# 1	0.205	08:32
Breath Volume		2.273	08:32
Air Blank		0.000	08:33
Air Blank		0.000	08:35
Subject Sample	# 2	0.197	08:35
Breath Volume		3.320	08:35
Air Blank		0.000	08:36
Control Test		0.080	08:36
Air Blank		0.000	08:37
Diagnostics Check		OK	08:37

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOCHUM,CRAIG
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 06/15/2020 00:13:22

Diagnostics Check	OK	00:07
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:08
Subject Sample # 1	0.000	00:09
Breath Volume	1.437	00:09
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.000	00:11
Breath Volume	1.128	00:11
Air Blank	0.000	00:12
Control Test	0.081	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SOLOWAY,ALAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: DEAL

DATE/TIME OF TEST: 06/15/2020 01:43:00

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.081	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.000	01:38
Breath Volume	2.011	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.000	01:41
Breath Volume	1.945	01:41
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:43

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: EBEL,VICTORIA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: LOPEZ RAMIREZ

DATE/TIME OF TEST: 06/18/2020 02:20:01

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:11	
Control Test	0.081	02:12	
Air Blank	0.000	02:12	
Subject Sample	# 1 0.208	02:15	Volume Not Met
Breath Volume	0.238	02:15	
Air Blank	0.000	02:16	
Air Blank	0.000	02:18	
Subject Sample	# 2	02:18	Subject Test Refused
Breath Volume	0.000	02:18	
Air Blank	0.000	02:19	
Control Test	0.078	02:19	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:20	

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MENDEZ,JOSE
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 06/18/2020 05:13:34

Diagnostics Check	OK	05:04
Air Blank	0.000	05:05
Control Test	0.081	05:05
Air Blank	0.000	05:06
Subject Sample # 1	0.164	05:07
Breath Volume	1.589	05:07
Air Blank	0.000	05:08
Air Blank	0.000	05:09
Subject Sample # 2	0.155	05:11
Breath Volume	1.386	05:11
Air Blank	0.000	05:12
Control Test	0.079	05:12
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MACH,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: CHOPIN

DATE/TIME OF TEST: 06/19/2020 17:44:44

Diagnostics Check	OK	17:34
Air Blank	0.000	17:34
Control Test	0.081	17:34
Air Blank	0.000	17:35
Subject Sample # 1	0.146	17:36
Breath Volume	1.273	17:36
Air Blank	0.000	17:36
Air Blank	0.000	17:38
Subject Sample # 2	0.125	17:39
Breath Volume	1.664	17:39
Air Blank	0.000	17:40
Air Blank	0.000	17:42
Subject Sample # 3	0.115	17:43
Breath Volume	1.156	17:43
Air Blank	0.000	17:43
Control Test	0.080	17:44
Air Blank	0.000	17:44
Diagnostics Check	OK	17:44

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CASTRO,DENNIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: LALLI

DATE/TIME OF TEST: 06/19/2020 22:44:06

Diagnostics Check	OK	22:37
Air Blank	0.000	22:38
Control Test	0.081	22:38
Air Blank	0.000	22:38
Subject Sample # 1	0.000	22:39
Breath Volume	4.117	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.000	22:42
Breath Volume	2.464	22:42
Air Blank	0.000	22:43
Control Test	0.079	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PRADO,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: RODRIGUEZ RODRIGUEZ

DATE/TIME OF TEST: 06/20/2020 00:00:16

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:52	
Control Test	0.082	23:52	
Air Blank	0.000	23:53	
Subject Sample	# 1 0.178	23:56	Volume Not Met
Breath Volume	0.183	23:56	
Air Blank	0.000	23:57	
Air Blank	0.000	23:58	
Subject Sample	# 2	23:58	Subject Test Refused
Breath Volume	0.000	23:58	
Air Blank	0.000	23:59	
Control Test	0.078	23:59	
Air Blank	0.000	00:00	
Diagnostics Check	OK	00:00	

Data Download Date/Time: 06/26/2020 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR,N
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: HURTADO

DATE/TIME OF TEST: 06/20/2020 03:45:24

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.178	03:40
Breath Volume	2.839	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.175	03:43
Breath Volume	1.890	03:43
Air Blank	0.000	03:44
Control Test	0.081	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 06/26/2020 16:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: LAHTI

DATE/TIME OF TEST: 06/21/2020 01:44:13

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:41	
Control Test	0.080	01:41	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:42	Subject Test Refused
Breath Volume	0.000	01:42	
Air Blank	0.000	01:43	
Control Test	0.080	01:43	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

Data Download Date/Time: 06/26/2020 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SMITH, MARK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: DURANTES SALINAS

DATE/TIME OF TEST: 06/22/2020 22:57:08

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.081	22:51
Air Blank	0.000	22:51
Subject Sample # 1	0.125	22:52
Breath Volume	2.027	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.125	22:55
Breath Volume	2.421	22:55
Air Blank	0.000	22:56
Control Test	0.081	22:56
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 06/26/2020 16:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH, GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: GONZALEZ ZAVALA

DATE/TIME OF TEST: 06/24/2020 01:14:58

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.082	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.179	01:10
Breath Volume	1.445	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:12
Subject Sample # 2	0.188	01:13
Breath Volume	2.230	01:13
Air Blank	0.000	01:14
Control Test	0.081	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 06/26/2020 16:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/24/2020 03:07:33

Diagnostics Check	OK	03:03	
Air Blank	0.000	03:03	
Control Test	0.081	03:03	
Air Blank	0.000	03:04	
Subject Sample	# 1	03:06	Subject Test Refused
Breath Volume	0.000	03:06	
Air Blank	0.000	03:06	
Control Test	0.081	03:06	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

Data Download Date/Time: 06/26/2020 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: QUEVEDO, MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/25/2020 01:04:02

Diagnostics Check		OK	00:57
Air Blank		0.000	00:58
Control Test		0.081	00:58
Air Blank		0.000	00:59
Subject Sample	# 1	0.236	00:59
Breath Volume		1.410	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:01
Subject Sample	# 2	0.235	01:02
Breath Volume		1.906	01:02
Air Blank		0.000	01:03
Control Test		0.080	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:04

Data Download Date/Time: 06/26/2020 16:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: TOUSSAINT

DATE/TIME OF TEST: 06/25/2020 01:47:21

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.081	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.189	01:41
Breath Volume	1.890	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.173	01:45
Breath Volume	1.429	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 06/26/2020 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BLACK,BRITTNEY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: BERGER

DATE/TIME OF TEST: 06/25/2020 13:20:20

Diagnostics Check	OK	13:13
Air Blank	0.000	13:14
Control Test	0.082	13:14
Air Blank	0.000	13:15
Subject Sample # 1	0.000	13:15
Breath Volume	3.007	13:15
Air Blank	0.000	13:16
Air Blank	0.000	13:18
Subject Sample # 2	0.000	13:18
Breath Volume	3.234	13:18
Air Blank	0.000	13:19
Control Test	0.081	13:19
Air Blank	0.000	13:20
Diagnostics Check	OK	13:20

Data Download Date/Time: 06/26/2020 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: REUS,DAIANI
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: HAMIDEH

DATE/TIME OF TEST: 06/25/2020 18:17:00

Diagnostics Check	OK	18:10
Air Blank	0.000	18:10
Control Test	0.081	18:11
Air Blank	0.000	18:11
Subject Sample # 1	0.000	18:12
Breath Volume	1.625	18:12
Air Blank	0.000	18:12
Air Blank	0.000	18:14
Subject Sample # 2	0.000	18:15
Breath Volume	2.191	18:15
Air Blank	0.000	18:16
Control Test	0.082	18:16
Air Blank	0.000	18:16
Diagnostics Check	OK	18:17

Data Download Date/Time: 06/26/2020 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VAN CAMP,JEFF
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: SWIFT

DATE/TIME OF TEST: 06/25/2020 23:01:37

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.080	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.000	22:57
Breath Volume	2.210	22:57
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.000	23:00
Breath Volume	2.515	23:00
Air Blank	0.000	23:00
Control Test	0.081	23:01
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 06/26/2020 16:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINSOR, NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: BELANGER

DATE/TIME OF TEST: 06/26/2020 01:15:01

Diagnostics Check		OK	01:07
Air Blank		0.000	01:08
Control Test		0.081	01:08
Air Blank		0.000	01:08
Subject Sample	# 1	0.238	01:10
Breath Volume		2.300	01:10
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.245	01:13
Breath Volume		1.875	01:13
Air Blank		0.000	01:14
Control Test		0.080	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:15

Data Download Date/Time: 06/26/2020 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SAAVEDRO,ALONSO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: SIRMULIS

DATE/TIME OF TEST: 06/26/2020 02:07:24

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.081	02:01
Air Blank	0.000	02:01
Subject Sample # 1	0.195	02:02
Breath Volume	3.472	02:02
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.196	02:05
Breath Volume	1.808	02:05
Air Blank	0.000	02:06
Control Test	0.079	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 06/26/2020 16:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARI,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILLIARD,KARI
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: COLLINS SHIVER

DATE/TIME OF TEST: 06/07/2020 20:32:15

Diagnostics Check	OK	20:25
Air Blank	0.000	20:26
Control Test	0.079	20:26
Air Blank	0.000	20:26
Subject Sample # 1	0.166	20:27
Breath Volume	2.281	20:27
Air Blank	0.000	20:28
Air Blank	0.000	20:29
Subject Sample # 2	0.158	20:30
Breath Volume	2.554	20:30
Air Blank	0.000	20:31
Control Test	0.078	20:31
Air Blank	0.000	20:32
Diagnostics Check	OK	20:32

Data Download Date/Time: 06/29/2020 14:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARLA
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: JOHNSON,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CISNEROS SANTIAGO

DATE/TIME OF TEST: 06/08/2020 03:19:50

Diagnostics Check		OK	03:13
Air Blank		0.000	03:13
Control Test		0.077	03:14
Air Blank		0.000	03:14
Subject Sample	# 1	0.132	03:15
Breath Volume		1.710	03:15
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.132	03:18
Breath Volume		2.785	03:18
Air Blank		0.000	03:18
Control Test		0.077	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 06/29/2020 14:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAYNE,ANDREW,S
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: CANO,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: LEONARDIS

DATE/TIME OF TEST: 06/06/2020 02:28:52

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.275	02:23
Breath Volume	3.105	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.281	02:27
Breath Volume	3.023	02:27
Air Blank	0.000	02:27
Control Test	0.079	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 06/16/2020 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BEECH

DATE/TIME OF TEST: 06/05/2020 01:43:12

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.080	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.215	01:38
Breath Volume		2.210	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:40
Subject Sample	# 2	0.215	01:41
Breath Volume		3.324	01:41
Air Blank		0.000	01:42
Control Test		0.079	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 07/08/2020 18:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: CAVRERA

DATE/TIME OF TEST: 06/06/2020 03:56:45

Diagnostics Check		OK	03:43	
Air Blank		0.000	03:44	
Control Test		0.080	03:44	
Air Blank		0.000	03:44	
Subject Sample	# 1	0.085	03:48	Volume Not Met
Breath Volume		0.710	03:48	
Air Blank		0.000	03:48	
Air Blank		0.000	03:50	
Subject Sample	# 2	0.111	03:51	
Breath Volume		1.550	03:51	
Air Blank		0.000	03:52	
Air Blank		0.000	03:54	
Subject Sample	# 3	0.114	03:55	
Breath Volume		1.265	03:55	
Air Blank		0.000	03:55	
Control Test		0.079	03:56	
Air Blank		0.000	03:56	
Diagnostics Check		OK	03:56	

Data Download Date/Time: 07/08/2020 18:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: SLUKHANOVA

DATE/TIME OF TEST: 06/06/2020 04:54:55

Diagnostics Check		OK	04:43	
Air Blank		0.000	04:43	
Control Test		0.079	04:43	
Air Blank		0.000	04:44	
Subject Sample	# 1	0.183	04:47	Volume Not Met
Breath Volume		0.132	04:47	
Air Blank		0.000	04:48	
Air Blank		0.000	04:50	
Subject Sample	# 2	0.162	04:53	Volume Not Met
Breath Volume		0.085	04:53	
Air Blank		0.000	04:53	
Control Test		0.076	04:54	
Air Blank		0.000	04:54	
Diagnostics Check		OK	04:54	

Data Download Date/Time: 07/08/2020 18:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: SLUKHANOVA

DATE/TIME OF TEST: 06/06/2020 05:08:43

Diagnostics Check		OK	04:56	
Air Blank		0.000	04:57	
Control Test		0.078	04:57	
Air Blank		0.000	04:58	
Subject Sample	# 1	0.170	05:01	Volume Not Met
Breath Volume		0.089	05:01	
Air Blank		0.000	05:02	
Air Blank		0.000	05:03	
Subject Sample	# 2	0.155	05:06	Volume Not Met
Breath Volume		0.160	05:06	
Air Blank		0.000	05:07	
Control Test		0.077	05:08	
Air Blank		0.000	05:08	
Diagnostics Check		OK	05:08	

Data Download Date/Time: 07/08/2020 18:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARNA,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/07/2020 01:13:05

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.081	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.149	01:08
Breath Volume	2.109	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.145	01:11
Breath Volume	1.472	01:11
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 07/08/2020 18:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: SUAREZ CHAVEZ

DATE/TIME OF TEST: 06/07/2020 05:50:40

Diagnostics Check		OK	05:42
Air Blank		0.000	05:43
Control Test		0.080	05:43
Air Blank		0.000	05:44
Subject Sample	# 1	0.258	05:46
Breath Volume		1.925	05:46
Air Blank		0.000	05:46
Air Blank		0.000	05:48
Subject Sample	# 2	0.249	05:48
Breath Volume		1.453	05:48
Air Blank		0.000	05:49
Control Test		0.075	05:50
Air Blank		0.000	05:50
Diagnostics Check		OK	05:50

Data Download Date/Time: 07/08/2020 18:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARCIA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:21

SUBJECT'S LAST NAME: SANTAMARIA CABANAS

DATE/TIME OF TEST: 06/07/2020 15:54:07

Diagnostics Check	OK	15:44
Air Blank	0.000	15:45
Control Test	0.081	15:45
Air Blank	0.000	15:45
Subject Sample # 1	0.247	15:46
Breath Volume	4.167	15:46
Air Blank	0.000	15:47
Air Blank	0.000	15:48
Subject Sample # 2	0.204	15:49
Breath Volume	1.355	15:49
Air Blank	0.000	15:50
Air Blank	0.000	15:51
Subject Sample # 3	0.235	15:52
Breath Volume	3.011	15:52
Air Blank	0.000	15:53
Control Test	0.079	15:53
Air Blank	0.000	15:54
Diagnostics Check	OK	15:54

Data Download Date/Time: 07/08/2020 18:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: REYES JR

DATE/TIME OF TEST: 06/08/2020 04:45:30

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.081	04:37
Air Blank	0.000	04:38
Subject Sample # 1	0.201	04:39
Breath Volume	1.652	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.215	04:43
Breath Volume	2.062	04:43
Air Blank	0.000	04:44
Control Test	0.079	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 07/08/2020 18:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/08/2020 23:52:12

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.146	23:47
Breath Volume	1.464	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.151	23:50
Breath Volume	1.183	23:50
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 07/08/2020 18:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: MCMURRAY

DATE/TIME OF TEST: 06/09/2020 20:45:11

Diagnostics Check	OK	20:38
Air Blank	0.000	20:38
Control Test	0.080	20:38
Air Blank	0.000	20:39
Subject Sample # 1	0.189	20:40
Breath Volume	2.617	20:40
Air Blank	0.000	20:41
Air Blank	0.000	20:42
Subject Sample # 2	0.193	20:43
Breath Volume	1.410	20:43
Air Blank	0.000	20:44
Control Test	0.079	20:44
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 07/08/2020 18:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D 249590
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:52

SUBJECT'S LAST NAME: SOWERS

DATE/TIME OF TEST: 06/10/2020 09:28:24

Diagnostics Check	OK	09:21
Air Blank	0.000	09:22
Control Test	0.081	09:22
Air Blank	0.000	09:23
Subject Sample # 1	0.000	09:23
Breath Volume	2.984	09:23
Air Blank	0.000	09:24
Air Blank	0.000	09:26
Subject Sample # 2	0.000	09:27
Breath Volume	2.062	09:27
Air Blank	0.000	09:27
Control Test	0.080	09:27
Air Blank	0.000	09:28
Diagnostics Check	OK	09:28

Data Download Date/Time: 07/08/2020 18:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KENNEDY, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: THERIAULT

DATE/TIME OF TEST: 06/10/2020 17:08:59

Diagnostics Check	OK	17:00	
Air Blank	0.000	17:00	
Control Test	0.080	17:01	
Air Blank	0.000	17:01	
Subject Sample	# 1	17:04	No Sample Provided
Breath Volume	0.000	17:04	
Air Blank	0.000	17:05	
Air Blank	0.000	17:07	
Subject Sample	# 2	17:07	Subject Test Refused
Breath Volume	0.000	17:07	
Air Blank	0.000	17:08	
Control Test	0.080	17:08	
Air Blank	0.000	17:08	
Diagnostics Check	OK	17:08	

Data Download Date/Time: 07/08/2020 18:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 06/11/2020 01:09:55

Diagnostics Check		OK	00:57	
Air Blank		0.000	00:58	
Control Test		0.080	00:58	
Air Blank		0.000	00:58	
Subject Sample	# 1	0.000	01:02	Volume Not Met
Breath Volume		0.160	01:02	
Air Blank		0.000	01:02	
Air Blank		0.000	01:04	
Subject Sample	# 2	0.000	01:04	
Breath Volume		2.003	01:04	
Air Blank		0.000	01:05	
Air Blank		0.000	01:07	
Subject Sample	# 3	0.000	01:08	
Breath Volume		1.625	01:08	
Air Blank		0.000	01:08	
Control Test		0.080	01:09	
Air Blank		0.000	01:09	
Diagnostics Check		OK	01:09	

Data Download Date/Time: 07/08/2020 18:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURROUGHS,H
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 06/11/2020 02:55:43

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.160	02:50
Breath Volume	1.855	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.164	02:54
Breath Volume	1.367	02:54
Air Blank	0.000	02:54
Control Test	0.079	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 07/08/2020 18:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GRAHAM,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: SONGER

DATE/TIME OF TEST: 06/11/2020 17:50:24

Diagnostics Check	OK	17:42
Air Blank	0.000	17:43
Control Test	0.080	17:43
Air Blank	0.000	17:44
Subject Sample # 1	0.242	17:45
Breath Volume	2.148	17:45
Air Blank	0.000	17:46
Air Blank	0.000	17:47
Subject Sample # 2	0.251	17:48
Breath Volume	3.000	17:48
Air Blank	0.000	17:49
Control Test	0.078	17:49
Air Blank	0.000	17:50
Diagnostics Check	OK	17:50

Data Download Date/Time: 07/08/2020 18:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: FLYNN

DATE/TIME OF TEST: 06/11/2020 23:37:17

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.081	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.126	23:32
Breath Volume	2.667	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:34
Subject Sample # 2	0.117	23:35
Breath Volume	1.179	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 07/08/2020 18:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 06/12/2020 03:32:14

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.211	03:27
Breath Volume	3.566	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:29
Subject Sample # 2	0.219	03:30
Breath Volume	3.679	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 07/08/2020 18:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: RODRIGUEZ AVILA

DATE/TIME OF TEST: 06/12/2020 04:01:32

Diagnostics Check		OK	03:54
Air Blank		0.000	03:54
Control Test		0.080	03:55
Air Blank		0.000	03:55
Subject Sample	# 1	0.148	03:56
Breath Volume		2.800	03:56
Air Blank		0.000	03:57
Air Blank		0.000	03:58
Subject Sample	# 2	0.135	03:59
Breath Volume		1.609	03:59
Air Blank		0.000	04:00
Control Test		0.079	04:00
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 07/08/2020 18:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 06/12/2020 23:57:27

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.081	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.100	23:52
Breath Volume	2.332	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.103	23:55
Breath Volume	4.660	23:55
Air Blank	0.000	23:56
Control Test	0.080	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 07/08/2020 18:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 06/13/2020 02:53:58

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.080	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.147	02:48
Breath Volume	1.605	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:50
Subject Sample # 2	0.149	02:52
Breath Volume	1.292	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 07/08/2020 18:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: BERNAL LOPEZ

DATE/TIME OF TEST: 06/13/2020 06:22:34

Diagnostics Check		OK	06:15
Air Blank		0.000	06:16
Control Test		0.080	06:16
Air Blank		0.000	06:17
Subject Sample	# 1	0.202	06:18
Breath Volume		1.554	06:18
Air Blank		0.000	06:18
Air Blank		0.000	06:20
Subject Sample	# 2	0.193	06:20
Breath Volume		1.273	06:20
Air Blank		0.000	06:21
Control Test		0.079	06:21
Air Blank		0.000	06:22
Diagnostics Check		OK	06:22

Data Download Date/Time: 07/08/2020 18:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: COSME NIEVES

DATE/TIME OF TEST: 06/13/2020 23:10:30

Diagnostics Check	OK	23:04
Air Blank	0.000	23:05
Control Test	0.080	23:05
Air Blank	0.000	23:05
Subject Sample # 1	0.000	23:06
Breath Volume	2.007	23:06
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.000	23:09
Breath Volume	2.042	23:09
Air Blank	0.000	23:09
Control Test	0.081	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 07/08/2020 18:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 06/14/2020 04:06:06

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.081	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.215	04:01
Breath Volume	1.753	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:03
Subject Sample # 2	0.218	04:04
Breath Volume	3.390	04:04
Air Blank	0.000	04:05
Control Test	0.079	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 07/08/2020 18:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: MORIN

DATE/TIME OF TEST: 06/14/2020 19:15:39

Diagnostics Check	OK	19:08
Air Blank	0.000	19:09
Control Test	0.081	19:09
Air Blank	0.000	19:09
Subject Sample # 1	0.166	19:10
Breath Volume	3.253	19:10
Air Blank	0.000	19:11
Air Blank	0.000	19:13
Subject Sample # 2	0.164	19:13
Breath Volume	2.234	19:13
Air Blank	0.000	19:14
Control Test	0.079	19:15
Air Blank	0.000	19:15
Diagnostics Check	OK	19:15

Data Download Date/Time: 07/08/2020 18:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: CIPOLLA

DATE/TIME OF TEST: 06/15/2020 23:51:49

Diagnostics Check	OK	23:44
Air Blank	0.000	23:45
Control Test	0.081	23:45
Air Blank	0.000	23:45
Subject Sample # 1	0.108	23:47
Breath Volume	3.296	23:47
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.095	23:50
Breath Volume	1.496	23:50
Air Blank	0.000	23:50
Control Test	0.079	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 07/08/2020 18:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: NEWTON

DATE/TIME OF TEST: 06/17/2020 00:58:18

Diagnostics Check		OK	00:49
Air Blank		0.000	00:50
Control Test		0.081	00:50
Air Blank		0.000	00:51
Subject Sample	# 1	0.163	00:53
Breath Volume		2.187	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:55
Subject Sample	# 2	0.173	00:56
Breath Volume		2.433	00:56
Air Blank		0.000	00:57
Control Test		0.078	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 07/08/2020 18:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLILE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 06/17/2020 01:24:00

Diagnostics Check	OK	01:17
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.000	01:19
Breath Volume	2.941	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.000	01:22
Breath Volume	3.265	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:24

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 06/18/2020 00:02:50

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:55
Subject Sample # 2	0.232	23:55
Breath Volume	2.593	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 3	0.208	23:58
Breath Volume	1.679	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.205	00:01
Breath Volume	1.855	00:01
Air Blank	0.000	00:01
Control Test	0.079	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/19/2020 04:21:25

Diagnostics Check		OK	04:09
Air Blank		0.000	04:10
Control Test		0.080	04:10
Air Blank		0.000	04:11
Subject Sample	# 1	0.159	04:13
Breath Volume		1.664	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.183	04:16
Breath Volume		1.667	04:16
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 3	0.178	04:19
Breath Volume		1.722	04:19
Air Blank		0.000	04:20
Control Test		0.078	04:20
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 06/20/2020 02:33:54

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.081	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.184	02:28
Breath Volume	1.593	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.176	02:32
Breath Volume	1.582	02:32
Air Blank	0.000	02:32
Control Test	0.079	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: DE LOS SANTOS-AQUINO

DATE/TIME OF TEST: 06/20/2020 05:13:00

Diagnostics Check	OK	05:06
Air Blank	0.000	05:06
Control Test	0.080	05:06
Air Blank	0.000	05:07
Subject Sample # 1	0.222	05:08
Breath Volume	2.136	05:08
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 2	0.217	05:11
Breath Volume	2.511	05:11
Air Blank	0.000	05:12
Control Test	0.080	05:12
Air Blank	0.000	05:12
Diagnostics Check	OK	05:13

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 06/20/2020 05:40:49

Diagnostics Check	OK	05:34
Air Blank	0.000	05:35
Control Test	0.080	05:35
Air Blank	0.000	05:35
Subject Sample # 1	0.155	05:36
Breath Volume	1.683	05:36
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.151	05:39
Breath Volume	1.617	05:39
Air Blank	0.000	05:39
Control Test	0.079	05:40
Air Blank	0.000	05:40
Diagnostics Check	OK	05:40

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOLSTENHOLM, A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: PANZO GARCIA

DATE/TIME OF TEST: 06/21/2020 00:30:19

Diagnostics Check		OK	00:23
Air Blank		0.000	00:23
Control Test		0.080	00:24
Air Blank		0.000	00:24
Subject Sample	# 1	0.126	00:25
Breath Volume		1.878	00:25
Air Blank		0.000	00:26
Air Blank		0.000	00:28
Subject Sample	# 2	0.124	00:28
Breath Volume		1.218	00:28
Air Blank		0.000	00:29
Control Test		0.079	00:29
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER, T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: PERSHAD

DATE/TIME OF TEST: 06/21/2020 03:27:37

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:16	
Control Test	0.080	03:16	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:20	No Sample Provided
Breath Volume	0.000	03:20	
Air Blank	0.000	03:21	
Air Blank	0.000	03:23	
Subject Sample	# 2	03:26	No Sample Provided
Breath Volume	0.000	03:26	
Air Blank	0.000	03:26	
Control Test	0.081	03:26	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:27	

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: MOJICA PABON

DATE/TIME OF TEST: 06/21/2020 04:41:47

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.127	04:36
Breath Volume	1.406	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.123	04:40
Breath Volume	1.628	04:40
Air Blank	0.000	04:40
Control Test	0.079	04:41
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: DOOLEY

DATE/TIME OF TEST: 06/23/2020 05:20:50

Diagnostics Check	OK	05:13
Air Blank	0.000	05:14
Control Test	0.081	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.163	05:16
Breath Volume	3.273	05:16
Air Blank	0.000	05:17
Air Blank	0.000	05:18
Subject Sample # 2	0.164	05:19
Breath Volume	3.769	05:19
Air Blank	0.000	05:19
Control Test	0.079	05:20
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROSALES,F
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: VEGA NIEVES

DATE/TIME OF TEST: 06/24/2020 02:10:23

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.081	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.000	02:05
Breath Volume	1.484	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.000	02:08
Breath Volume	1.250	02:08
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: OVERMAN

DATE/TIME OF TEST: 06/24/2020 05:50:56

Diagnostics Check	OK	05:44
Air Blank	0.000	05:44
Control Test	0.080	05:45
Air Blank	0.000	05:45
Subject Sample # 1	0.000	05:46
Breath Volume	2.238	05:46
Air Blank	0.000	05:47
Air Blank	0.000	05:49
Subject Sample # 2	0.000	05:49
Breath Volume	2.046	05:49
Air Blank	0.000	05:50
Control Test	0.080	05:50
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 06/24/2020 23:24:02

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.080	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.090	23:18
Breath Volume	1.179	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.093	23:22
Breath Volume	1.695	23:22
Air Blank	0.000	23:23
Control Test	0.080	23:23
Air Blank	0.000	23:23
Diagnostics Check	OK	23:24

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISCO,S
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:44

SUBJECT'S LAST NAME: CALDER

DATE/TIME OF TEST: 06/26/2020 11:13:36

Diagnostics Check	OK	11:06
Air Blank	0.000	11:06
Control Test	0.080	11:07
Air Blank	0.000	11:07
Subject Sample # 1	0.000	11:09
Breath Volume	3.082	11:09
Air Blank	0.000	11:09
Air Blank	0.000	11:11
Subject Sample # 2	0.000	11:12
Breath Volume	3.105	11:12
Air Blank	0.000	11:12
Control Test	0.081	11:12
Air Blank	0.000	11:13
Diagnostics Check	OK	11:13

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: SANABRIA

DATE/TIME OF TEST: 06/27/2020 00:05:26

Diagnostics Check	OK	23:58
Air Blank	0.000	23:59
Control Test	0.081	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.000	00:00
Breath Volume	1.468	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.000	00:04
Breath Volume	1.636	00:04
Air Blank	0.000	00:04
Control Test	0.081	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 06/27/2020 06:36:45

Diagnostics Check	OK	06:30
Air Blank	0.000	06:30
Control Test	0.081	06:30
Air Blank	0.000	06:31
Subject Sample # 1	0.013	06:32
Breath Volume	1.347	06:32
Air Blank	0.000	06:32
Air Blank	0.000	06:34
Subject Sample # 2	0.014	06:35
Breath Volume	2.062	06:35
Air Blank	0.000	06:35
Control Test	0.081	06:36
Air Blank	0.000	06:36
Diagnostics Check	OK	06:36

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: SYKES

DATE/TIME OF TEST: 06/27/2020 15:51:36

Diagnostics Check	OK	15:44
Air Blank	0.000	15:45
Control Test	0.080	15:45
Air Blank	0.000	15:46
Subject Sample # 1	0.354	15:46
Breath Volume	2.414	15:46
Air Blank	0.000	15:47
Air Blank	0.000	15:49
Subject Sample # 2	0.345	15:49
Breath Volume	2.476	15:49
Air Blank	0.000	15:50
Control Test	0.078	15:50
Air Blank	0.000	15:51
Diagnostics Check	OK	15:51

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 06/27/2020 19:46:14

Diagnostics Check	OK	19:39
Air Blank	0.000	19:39
Control Test	0.080	19:39
Air Blank	0.000	19:40
Subject Sample # 1	0.000	19:41
Breath Volume	3.449	19:41
Air Blank	0.000	19:42
Air Blank	0.000	19:44
Subject Sample # 2	0.000	19:44
Breath Volume	3.355	19:44
Air Blank	0.000	19:45
Control Test	0.080	19:45
Air Blank	0.000	19:46
Diagnostics Check	OK	19:46

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: RODRIGUEZ ALONSO

DATE/TIME OF TEST: 06/28/2020 05:37:56

Diagnostics Check		OK	05:25	
Air Blank		0.000	05:25	
Control Test		0.081	05:26	
Air Blank		0.000	05:26	
Subject Sample	# 1	0.073	05:29	Volume Not Met
Breath Volume		1.078	05:29	
Air Blank		0.000	05:30	
Air Blank		0.000	05:32	
Subject Sample	# 2	0.087	05:33	
Breath Volume		1.125	05:33	
Air Blank		0.000	05:33	
Air Blank		0.000	05:35	
Subject Sample	# 3	0.081	05:36	
Breath Volume		1.390	05:36	
Air Blank		0.000	05:37	
Control Test		0.078	05:37	
Air Blank		0.000	05:37	
Diagnostics Check		OK	05:37	

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/28/2020 20:29:56

Diagnostics Check	OK	20:23	
Air Blank	0.000	20:24	
Control Test	0.080	20:24	
Air Blank	0.000	20:25	
Subject Sample # 1	0.244	20:25	
Breath Volume	1.687	20:25	
Air Blank	0.000	20:26	
Air Blank	0.000	20:28	
		20:28	Improper Sample
Air Blank	0.000	20:28	
Control Test	0.080	20:29	
Air Blank	0.000	20:29	
Diagnostics Check	OK	20:29	

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/28/2020 20:38:31

Diagnostics Check	OK	20:31	
Air Blank	0.000	20:31	
Control Test	0.078	20:31	
Air Blank	0.000	20:32	
Subject Sample	# 1 0.228	20:33	
Breath Volume	1.367	20:33	
Air Blank	0.000	20:33	
Air Blank	0.000	20:35	
Subject Sample	# 2	20:36	Slope Not Met
Breath Volume	0.503	20:36	
Air Blank	0.000	20:37	
Control Test	0.078	20:37	
Air Blank	0.000	20:38	
Diagnostics Check	OK	20:38	

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/28/2020 21:09:30

Diagnostics Check		OK	21:00
Air Blank		0.000	21:00
Control Test		0.079	21:00
Air Blank		0.000	21:01
Subject Sample	# 1	0.217	21:02
Breath Volume		2.066	21:02
Air Blank		0.000	21:03
Air Blank		0.000	21:05
Subject Sample	# 2	0.201	21:07
Breath Volume		1.562	21:07
Air Blank		0.000	21:08
Control Test		0.078	21:08
Air Blank		0.000	21:09
Diagnostics Check		OK	21:09

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 06/29/2020 03:27:03

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.081	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.165	03:22
Breath Volume	2.015	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.160	03:25
Breath Volume	2.058	03:25
Air Blank	0.000	03:26
Control Test	0.078	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:27

Data Download Date/Time: 07/08/2020 18:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/29/2020 21:57:05

Diagnostics Check	OK	21:48	
Air Blank	0.000	21:49	
Control Test	0.080	21:49	
Air Blank	0.000	21:50	
Subject Sample	# 1 0.204	21:51	
Breath Volume	2.929	21:51	
Air Blank	0.000	21:52	
Air Blank	0.000	21:53	
Subject Sample	# 2	21:55	Subject Test Refused
Breath Volume	0.000	21:55	
Air Blank	0.000	21:56	
Control Test	0.079	21:56	
Air Blank	0.000	21:57	
Diagnostics Check	OK	21:57	

Data Download Date/Time: 07/08/2020 18:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/29/2020 22:10:13

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.080	22:02
Air Blank	0.000	22:02
Subject Sample # 1	0.196	22:03
Breath Volume	1.160	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.178	22:08
Breath Volume	3.277	22:08
Air Blank	0.000	22:09
Control Test	0.079	22:09
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 07/08/2020 18:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,ROBERT,
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: RAMOS,ROBERT
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: BUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:39

SUBJECT'S LAST NAME: O'CONNOR

DATE/TIME OF TEST: 06/14/2020 02:00:20

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:57	
Control Test	0.083	01:58	
Air Blank	0.000	01:58	
Subject Sample	# 1	01:58	Subject Test Refused
Breath Volume	0.000	01:58	
Air Blank	0.000	01:59	
Control Test	0.083	01:59	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

Data Download Date/Time: 06/15/2020 19:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEITE,STEVEN,G
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LEITE,STEVEN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 05/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: KOLB

DATE/TIME OF TEST: 06/14/2020 03:23:51

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.083	03:18
Air Blank	0.000	03:18
Subject Sample # 1	0.163	03:19
Breath Volume	1.765	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.156	03:22
Breath Volume	1.617	03:22
Air Blank	0.000	03:22
Control Test	0.081	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 06/15/2020 19:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS, JEREMIAH, S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: BYRD, JERRY
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: KULICKA

DATE/TIME OF TEST: 06/17/2020 00:41:23

Diagnostics Check	OK	00:38	
Air Blank	0.000	00:38	
Control Test	0.082	00:39	
Air Blank	0.000	00:39	
Subject Sample	# 1	00:40	Subject Test Refused
Breath Volume	0.000	00:40	
Air Blank	0.000	00:40	
Control Test	0.082	00:40	
Air Blank	0.000	00:41	
Diagnostics Check	OK	00:41	

Data Download Date/Time: 07/13/2020 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,RUSSELL,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: COOK,RUSSELL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: MONREAL ROSTRO

DATE/TIME OF TEST: 06/24/2020 03:32:29

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.083	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.107	03:25
Breath Volume	1.257	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:27
Subject Sample # 2	0.104	03:30
Breath Volume	1.910	03:30
Air Blank	0.000	03:31
Control Test	0.081	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 07/13/2020 11:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIDDELL,ANGELA,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HARRISON,BOBBY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: LIDDELL

DATE/TIME OF TEST: 06/25/2020 00:34:59

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:31	
Control Test	0.082	00:32	
Air Blank	0.000	00:32	
Subject Sample	# 1	00:33	Subject Test Refused
Breath Volume	0.000	00:33	
Air Blank	0.000	00:34	
Control Test	0.082	00:34	
Air Blank	0.000	00:34	
Diagnostics Check	OK	00:34	

Data Download Date/Time: 07/13/2020 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMBERT,BRIAN,K
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: ALLGOOD,JEFF
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:34

SUBJECT'S LAST NAME: OWEN

DATE/TIME OF TEST: 06/27/2020 18:24:58

Diagnostics Check	OK	18:17
Air Blank	0.000	18:18
Control Test	0.081	18:18
Air Blank	0.000	18:19
Subject Sample # 1	0.107	18:19
Breath Volume	1.289	18:19
Air Blank	0.000	18:20
Air Blank	0.000	18:22
Subject Sample # 2	0.112	18:23
Breath Volume	1.726	18:23
Air Blank	0.000	18:24
Control Test	0.082	18:24
Air Blank	0.000	18:24
Diagnostics Check	OK	18:24

Data Download Date/Time: 07/13/2020 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANNING,CARL,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MANNING,CARL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: GANTAR

DATE/TIME OF TEST: 06/27/2020 23:52:56

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.082	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.174	23:48
Breath Volume	2.441	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:50
Subject Sample # 2	0.167	23:51
Breath Volume	1.714	23:51
Air Blank	0.000	23:52
Control Test	0.080	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 07/13/2020 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: HEATON

DATE/TIME OF TEST: 06/28/2020 21:02:18

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.082	20:56
Air Blank	0.000	20:57
Subject Sample # 1	0.000	20:57
Breath Volume	3.851	20:57
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.000	21:00
Breath Volume	3.484	21:00
Air Blank	0.000	21:01
Control Test	0.082	21:01
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 07/13/2020 11:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RODRIGUEZ,ALBERTO,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: RODRIGUEZ,ALBERTO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 11:56

GENDER: Male

OBSERVATION PERIOD BEGAN: 11:56

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 06/07/2020 12:41:50

Diagnostics Check	OK	12:38	
Air Blank	0.000	12:38	
Control Test	0.081	12:39	
Air Blank	0.000	12:39	
Subject Sample	# 1	12:40	Subject Test Refused
Breath Volume	0.000	12:40	
Air Blank	0.000	12:40	
Control Test	0.081	12:41	
Air Blank	0.000	12:41	
Diagnostics Check	OK	12:41	

Data Download Date/Time: 06/29/2020 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LASTER,WILLIE,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HALL,MATTHEW
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 07:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:23

SUBJECT'S LAST NAME: CANO

DATE/TIME OF TEST: 06/09/2020 09:49:49

Diagnostics Check	OK	09:43
Air Blank	0.000	09:44
Control Test	0.082	09:44
Air Blank	0.000	09:44
Subject Sample # 1	0.000	09:45
Breath Volume	1.429	09:45
Air Blank	0.000	09:46
Air Blank	0.000	09:48
Subject Sample # 2	0.000	09:48
Breath Volume	2.117	09:48
Air Blank	0.000	09:48
Control Test	0.080	09:49
Air Blank	0.000	09:49
Diagnostics Check	OK	09:49

Data Download Date/Time: 06/29/2020 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,RUBEN,G
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: CATALANO,BRANDON
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: BUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: TYSON

DATE/TIME OF TEST: 06/28/2020 22:57:43

Diagnostics Check	OK	22:54	
Air Blank	0.000	22:54	
Control Test	0.082	22:55	
Air Blank	0.000	22:55	
Subject Sample	# 1	22:56	Subject Test Refused
Breath Volume	0.000	22:56	
Air Blank	0.000	22:56	
Control Test	0.082	22:57	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

Data Download Date/Time: 06/29/2020 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: SANCHEZ, AQUILES

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 00:36

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 06/03/2020 01:52:46

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:50	
Control Test	0.079	01:50	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:51	
Control Test	0.079	01:52	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 06/29/2020 16:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ACUNA JR,CARLOS,A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: HALL,MATHEW

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 07:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: CANO

DATE/TIME OF TEST: 06/09/2020 09:06:05

Diagnostics Check	OK	09:05	
Air Blank		09:05	Ambient Fail
Air Blank	0.000	09:06	

Data Download Date/Time: 06/29/2020 16:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ACUNA JR,CARLOS,A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: HALL,MATHEW

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 07:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: CANO

DATE/TIME OF TEST: 06/09/2020 09:11:05

Diagnostics Check	OK	09:09	
Air Blank	0.000	09:09	
Control Test	0.075	09:10	
Air Blank		09:10	Ambient Fail
Air Blank		09:11	Ambient Fail

Data Download Date/Time: 06/29/2020 16:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CAMILA,P
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,CAMILA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: QUIROZ

DATE/TIME OF TEST: 06/13/2020 00:00:36

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:58	
Control Test	0.079	23:58	
Air Blank	0.000	23:58	
Subject Sample	# 1	23:59	Subject Test Refused
Breath Volume	0.000	23:59	
Air Blank	0.000	23:59	
Control Test	0.079	00:00	
Air Blank	0.000	00:00	
Diagnostics Check	OK	00:00	

Data Download Date/Time: 06/29/2020 16:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,WILLIAMS,E
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RODRIGUEZ,LUIS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: HORMAZA

DATE/TIME OF TEST: 06/13/2020 03:36:06

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:27	
Control Test	0.080	03:27	
Air Blank	0.000	03:28	
Subject Sample	# 1	03:31	No Sample Provided
Breath Volume	0.000	03:31	
Air Blank	0.000	03:31	
Air Blank	0.000	03:33	
Subject Sample	# 2	03:34	Subject Test Refused
Breath Volume	0.000	03:34	
Air Blank	0.000	03:35	
Control Test	0.079	03:35	
Air Blank	0.000	03:36	
Diagnostics Check	OK	03:36	

Data Download Date/Time: 06/29/2020 16:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINLEY,STEVEN,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BRUDER,ALEX
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: BUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: C220650993670

DATE/TIME OF TEST: 06/16/2020 23:07:25

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:05	
Control Test	0.079	23:05	
Air Blank	0.000	23:05	
Subject Sample	# 1	23:06	Subject Test Refused
Breath Volume	0.000	23:06	
Air Blank	0.000	23:06	
Control Test	0.083	23:06	
Air Blank	0.000	23:07	
Diagnostics Check	OK	23:07	

Data Download Date/Time: 06/29/2020 16:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,JAVIER,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MESA,JAVIER
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: KERSHAW

DATE/TIME OF TEST: 06/28/2020 23:12:18

Diagnostics Check	OK	23:09	
Air Blank	0.000	23:09	
Control Test	0.080	23:10	
Air Blank	0.000	23:10	
Subject Sample	# 1	23:10	Subject Test Refused
Breath Volume	0.000	23:10	
Air Blank	0.000	23:11	
Control Test	0.080	23:11	
Air Blank	0.000	23:12	
Diagnostics Check	OK	23:12	

Data Download Date/Time: 06/29/2020 16:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FERNANDEZ,ANGEL,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: FERNANDEZ,ANGEL

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 21:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: CASTINEIRAS

DATE/TIME OF TEST: 06/29/2020 22:16:31

Diagnostics Check	OK	22:08	
Air Blank	0.000	22:08	
Control Test	0.080	22:08	
Air Blank	0.000	22:09	
Subject Sample	# 1	22:12	No Sample Provided
Breath Volume	0.000	22:12	
Air Blank	0.000	22:12	
Air Blank	0.000	22:14	
Subject Sample	# 2	22:15	Subject Test Refused
Breath Volume	0.000	22:15	
Air Blank	0.000	22:15	
Control Test	0.079	22:15	
Air Blank	0.000	22:16	
Diagnostics Check	OK	22:16	

Data Download Date/Time: 07/30/2020 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/01/2020 03:35:00

Diagnostics Check		OK	03:25
Air Blank		0.000	03:26
Control Test		0.077	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.243	03:27
Breath Volume		2.722	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:29
Subject Sample	# 2	0.265	03:30
Breath Volume		3.140	03:30
Air Blank		0.000	03:31
Air Blank		0.000	03:32
Subject Sample	# 3	0.262	03:33
Breath Volume		2.824	03:33
Air Blank		0.000	03:34
Control Test		0.077	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:35

Data Download Date/Time: 06/08/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEJIA,ELVIS,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: ARRIAGA LOPEZ

DATE/TIME OF TEST: 06/02/2020 23:08:06

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.078	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.136	23:03
Breath Volume	2.554	23:03
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.135	23:06
Breath Volume	1.976	23:06
Air Blank	0.000	23:07
Control Test	0.076	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 06/08/2020 15:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,L
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MORALES,VICTOR
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: CENTENO

DATE/TIME OF TEST: 06/03/2020 17:47:32

Diagnostics Check	OK	17:39
Air Blank	0.000	17:39
Control Test	0.077	17:39
Air Blank	0.000	17:40
Subject Sample # 1	0.000	17:43
Breath Volume	1.433	17:43
Air Blank	0.000	17:43
Air Blank	0.000	17:45
Subject Sample # 2	0.000	17:46
Breath Volume	2.171	17:46
Air Blank	0.000	17:46
Control Test	0.076	17:46
Air Blank	0.000	17:47
Diagnostics Check	OK	17:47

Data Download Date/Time: 06/08/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 06/04/2020 01:15:48

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.078	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.154	01:11
Breath Volume	1.554	01:11
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.155	01:14
Breath Volume	1.859	01:14
Air Blank	0.000	01:14
Control Test	0.078	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 06/08/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER, EDWARD, A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: SWOGGER, EDWARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 06/04/2020 01:45:49

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:43	
Control Test	0.077	01:43	
Air Blank	0.000	01:44	
Subject Sample	# 1	01:44	Subject Test Refused
Breath Volume	0.000	01:44	
Air Blank	0.000	01:44	
Control Test	0.078	01:45	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 06/08/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: BRITTON

DATE/TIME OF TEST: 06/05/2020 00:47:35

Diagnostics Check	OK	00:44	
Air Blank	0.000	00:45	
Control Test	0.078	00:45	
Air Blank	0.000	00:45	
Subject Sample	# 1	00:46	Subject Test Refused
Breath Volume	0.000	00:46	
Air Blank	0.000	00:46	
Control Test	0.078	00:46	
Air Blank	0.000	00:47	
Diagnostics Check	OK	00:47	

Data Download Date/Time: 06/08/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER, EDWARD, A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: SWOGGER, EDWARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: VILLENAS

DATE/TIME OF TEST: 06/05/2020 20:21:54

Diagnostics Check	OK	20:19	
Air Blank	0.000	20:19	
Control Test	0.077	20:19	
Air Blank	0.000	20:20	
Subject Sample	# 1	20:20	Subject Test Refused
Breath Volume	0.000	20:20	
Air Blank	0.000	20:20	
Control Test	0.077	20:21	
Air Blank	0.000	20:21	
Diagnostics Check	OK	20:21	

Data Download Date/Time: 06/08/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER, EDWARD, A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HARRILL, ANDERSON
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/06/2020 03:39:14

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:36	
Control Test	0.077	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 1	03:37	Subject Test Refused
Breath Volume	0.000	03:37	
Air Blank	0.000	03:38	
Control Test	0.079	03:38	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 06/08/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,R
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: ROCHELA

DATE/TIME OF TEST: 06/06/2020 22:23:59

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:21	
Control Test	0.077	22:21	
Air Blank	0.000	22:22	
Subject Sample	# 1	22:22	Subject Test Refused
Breath Volume	0.000	22:22	
Air Blank	0.000	22:23	
Control Test	0.077	22:23	
Air Blank	0.000	22:23	
Diagnostics Check	OK	22:23	

Data Download Date/Time: 06/08/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER, EDWARD, A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: COSTA, SCOTT
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: POLLOCK

DATE/TIME OF TEST: 06/06/2020 23:38:01

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.078	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.175	23:33
Breath Volume	2.214	23:33
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.165	23:36
Breath Volume	2.042	23:36
Air Blank	0.000	23:37
Control Test	0.077	23:37
Air Blank	0.000	23:37
Diagnostics Check	OK	23:38

Data Download Date/Time: 06/08/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER, EDWARD, A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: SWOGGER, EDWARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MAXWELL

DATE/TIME OF TEST: 06/07/2020 03:13:41

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:11	
Control Test	0.078	03:11	
Air Blank	0.000	03:12	
Subject Sample	# 1	03:12	Subject Test Refused
Breath Volume	0.000	03:12	
Air Blank	0.000	03:12	
Control Test	0.078	03:13	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 06/08/2020 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: SANTOS,CRISTIAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 06/07/2020 21:07:45

Diagnostics Check	OK	21:04	
Air Blank	0.000	21:05	
Control Test	0.076	21:05	
Air Blank	0.000	21:06	
Subject Sample	# 1	21:06	Subject Test Refused
Breath Volume	0.000	21:06	
Air Blank	0.000	21:06	
Control Test	0.076	21:07	
Air Blank	0.000	21:07	
Diagnostics Check	OK	21:07	

Data Download Date/Time: 06/08/2020 15:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,JOSHUA,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HRADECKY,ALEXANDRIA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: DEMBLANS

DATE/TIME OF TEST: 06/07/2020 23:46:52

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.077	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.000	23:42
Breath Volume	1.757	23:42
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.000	23:45
Breath Volume	1.343	23:45
Air Blank	0.000	23:45
Control Test	0.078	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 06/08/2020 15:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,R
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/09/2020 19:28:23

Diagnostics Check	OK	19:20
Air Blank	0.000	19:21
Control Test	0.076	19:21
Air Blank	0.000	19:22
Subject Sample # 1	0.313	19:23
Breath Volume	3.097	19:23
Air Blank	0.000	19:24
Air Blank	0.000	19:26
Subject Sample # 2	0.311	19:26
Breath Volume	2.988	19:26
Air Blank	0.000	19:27
Control Test	0.075	19:27
Air Blank	0.000	19:28
Diagnostics Check	OK	19:28

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,JOEL,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MILLER,JAMIE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: MANNIX

DATE/TIME OF TEST: 06/09/2020 20:55:21

Diagnostics Check	OK	20:48
Air Blank	0.000	20:48
Control Test	0.077	20:48
Air Blank	0.000	20:49
Subject Sample # 1	0.087	20:49
Breath Volume	2.324	20:49
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.090	20:53
Breath Volume	2.003	20:53
Air Blank	0.000	20:54
Control Test	0.077	20:54
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER, EDWARD, A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HRADECKY, ALEXANDRIA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: UPPER

DATE/TIME OF TEST: 06/11/2020 01:06:08

Diagnostics Check		OK	01:00
Air Blank		0.000	01:00
Control Test		0.078	01:00
Air Blank		0.000	01:01
Subject Sample	# 1	0.060	01:01
Breath Volume		2.218	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.057	01:04
Breath Volume		2.539	01:04
Air Blank		0.000	01:05
Control Test		0.078	01:05
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER, EDWARD, A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: SWOGGER, EDWARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: REINHARDT

DATE/TIME OF TEST: 06/12/2020 00:14:29

Diagnostics Check		OK	00:08
Air Blank		0.000	00:08
Control Test		0.077	00:08
Air Blank		0.000	00:09
Subject Sample	# 1	0.203	00:09
Breath Volume		2.945	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:12
Subject Sample	# 2	0.200	00:12
Breath Volume		2.238	00:12
Air Blank		0.000	00:13
Control Test		0.077	00:13
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 06/12/2020 00:36:27

Diagnostics Check	OK	00:33	
Air Blank	0.000	00:33	
Control Test	0.078	00:34	
Air Blank	0.000	00:34	
Subject Sample	# 1	00:35	Subject Test Refused
Breath Volume	0.000	00:35	
Air Blank	0.000	00:35	
Control Test	0.077	00:35	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,E
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:39

SUBJECT'S LAST NAME: VINCENT

DATE/TIME OF TEST: 06/12/2020 16:06:45

Diagnostics Check	OK	16:00
Air Blank	0.000	16:01
Control Test	0.078	16:01
Air Blank	0.000	16:01
Subject Sample # 1	0.000	16:02
Breath Volume	1.855	16:02
Air Blank	0.000	16:02
Air Blank	0.000	16:04
Subject Sample # 2	0.000	16:05
Breath Volume	2.078	16:05
Air Blank	0.000	16:05
Control Test	0.078	16:06
Air Blank	0.000	16:06
Diagnostics Check	OK	16:06

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER, EDWARD, A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: SWOGGER, EDWARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 06/12/2020 22:00:49

Diagnostics Check	OK	21:57	
Air Blank	0.000	21:58	
Control Test	0.078	21:58	
Air Blank	0.000	21:59	
Subject Sample	# 1	21:59	Subject Test Refused
Breath Volume	0.000	21:59	
Air Blank	0.000	21:59	
Control Test	0.078	22:00	
Air Blank	0.000	22:00	
Diagnostics Check	OK	22:00	

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER, EDWARD, A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: SWOGGER, EDWARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 06/13/2020 02:57:46

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:55	
Control Test	0.079	02:55	
Air Blank	0.000	02:56	
Subject Sample	# 1	02:56	Subject Test Refused
Breath Volume	0.000	02:56	
Air Blank	0.000	02:56	
Control Test	0.078	02:57	
Air Blank	0.000	02:57	
Diagnostics Check	OK	02:57	

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,L
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MORALES,VICTOR
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:47

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/13/2020 07:11:28

Diagnostics Check	OK	07:09	
Air Blank	0.000	07:09	
Control Test	0.077	07:10	
Air Blank	0.000	07:10	
Subject Sample	# 1	07:10	RFI Detect
Breath Volume	0.000	07:10	
Air Blank		07:11	RFI Detect
Air Blank	0.000	07:11	

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,L
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MORALES,VICTOR
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:47

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/13/2020 07:16:51

Diagnostics Check	OK	07:13	
Air Blank	0.000	07:14	
Control Test	0.078	07:14	
Air Blank	0.000	07:15	
Subject Sample	# 1	07:15	Subject Test Refused
Breath Volume	0.000	07:15	
Air Blank	0.000	07:15	
Control Test	0.078	07:16	
Air Blank	0.000	07:16	
Diagnostics Check	OK	07:16	

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,JOEL,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,ANFERNEE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/14/2020 00:11:04

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.078	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.000	00:06
Breath Volume	2.898	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.000	00:09
Breath Volume	2.886	00:09
Air Blank	0.000	00:10
Control Test	0.082	00:10
Air Blank	0.000	00:10
Diagnostics Check	OK	00:11

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ DEL TORO,RUBEN,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: GUTIERREZ DEL TORO,RUBEN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: ONWUZO

DATE/TIME OF TEST: 06/15/2020 20:34:52

Diagnostics Check	OK	20:23	
Air Blank	0.000	20:23	
Control Test	0.076	20:24	
Air Blank	0.000	20:24	
Subject Sample	# 1 0.000	20:27	Volume Not Met
Breath Volume	0.535	20:27	
Air Blank	0.000	20:28	
Air Blank	0.000	20:30	
Subject Sample	# 2 0.000	20:33	Volume Not Met
Breath Volume	0.175	20:33	
Air Blank	0.000	20:33	
Control Test	0.076	20:34	
Air Blank	0.000	20:34	
Diagnostics Check	OK	20:34	

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ DEL TORO,RUBEN,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: GUTIERREZ DEL TORO,RUBEN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: ONWUZO

DATE/TIME OF TEST: 06/15/2020 20:49:26

Diagnostics Check	OK	20:43
Air Blank	0.000	20:43
Control Test	0.076	20:44
Air Blank	0.000	20:44
Subject Sample # 1	0.000	20:45
Breath Volume	1.742	20:45
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.000	20:48
Breath Volume	1.757	20:48
Air Blank	0.000	20:48
Control Test	0.077	20:48
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: VIDAL,NERELYS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 06/18/2020 03:39:37

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.078	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.124	03:35
Breath Volume	2.656	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.130	03:38
Breath Volume	2.027	03:38
Air Blank	0.000	03:38
Control Test	0.078	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: SPAULDING

DATE/TIME OF TEST: 06/18/2020 23:54:48

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:52	
Control Test	0.078	23:52	
Air Blank	0.000	23:53	
Subject Sample	# 1	23:53	Subject Test Refused
Breath Volume	0.000	23:53	
Air Blank	0.000	23:53	
Control Test	0.079	23:54	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:54	

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER, EDWARD, A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: SWOGGER, EDWARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: CAMBO

DATE/TIME OF TEST: 06/20/2020 21:24:47

Diagnostics Check	OK	21:21	
Air Blank	0.000	21:22	
Control Test	0.078	21:22	
Air Blank	0.000	21:23	
Subject Sample	# 1	21:23	Subject Test Refused
Breath Volume	0.000	21:23	
Air Blank	0.000	21:23	
Control Test	0.078	21:24	
Air Blank	0.000	21:24	
Diagnostics Check	OK	21:24	

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,JEAN,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: JACKSON,CODY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:59

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 06/21/2020 14:22:28

Diagnostics Check	OK	14:19	
Air Blank	0.000	14:20	
Control Test	0.078	14:20	
Air Blank	0.000	14:20	
Subject Sample	# 1	14:21	Subject Test Refused
Breath Volume	0.000	14:21	
Air Blank	0.000	14:21	
Control Test	0.077	14:21	
Air Blank	0.000	14:22	
Diagnostics Check	OK	14:22	

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 06/21/2020 21:52:21

Diagnostics Check		OK	21:45
Air Blank		0.000	21:45
Control Test		0.077	21:46
Air Blank		0.000	21:46
Subject Sample	# 1	0.164	21:47
Breath Volume		2.117	21:47
Air Blank		0.000	21:48
Air Blank		0.000	21:50
Subject Sample	# 2	0.163	21:50
Breath Volume		2.746	21:50
Air Blank		0.000	21:51
Control Test		0.077	21:51
Air Blank		0.000	21:52
Diagnostics Check		OK	21:52

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: AGUIRRE

DATE/TIME OF TEST: 06/22/2020 02:06:51

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:04	
Control Test	0.079	02:04	
Air Blank	0.000	02:05	
Subject Sample	# 1	02:05	Subject Test Refused
Breath Volume	0.000	02:05	
Air Blank	0.000	02:05	
Control Test	0.078	02:06	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:06	

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HARRILL,ANDERSEN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: DISLA

DATE/TIME OF TEST: 06/22/2020 02:45:16

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:40	
Control Test	0.078	02:41	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:43	Subject Test Refused
Breath Volume	0.000	02:43	
Air Blank	0.000	02:44	
Control Test	0.079	02:44	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY, JOSHUA, J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HARRILL, ANDERSON
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: ERCOLE

DATE/TIME OF TEST: 06/25/2020 01:37:52

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.078	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.253	01:33
Breath Volume	1.656	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:35
Subject Sample # 2	0.250	01:36
Breath Volume	2.058	01:36
Air Blank	0.000	01:36
Control Test	0.077	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ZANDER

DATE/TIME OF TEST: 06/27/2020 01:41:58

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.078	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.145	01:37
Breath Volume	2.484	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.143	01:40
Breath Volume	2.660	01:40
Air Blank	0.000	01:41
Control Test	0.078	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JARRETT,RASHAHN,C
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: JARRETT,RASHAHN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: IANNACCO

DATE/TIME OF TEST: 06/28/2020 01:42:13

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:38	
Control Test	0.077	01:38	
Air Blank	0.000	01:38	
Subject Sample	# 1	01:40	Subject Test Refused
Breath Volume	0.000	01:40	
Air Blank	0.000	01:41	
Control Test	0.078	01:41	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:42	

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,JOEL,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: BAUMGARDNER

DATE/TIME OF TEST: 06/28/2020 02:22:09

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.078	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.000	02:17
Breath Volume	1.914	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.000	02:20
Breath Volume	1.652	02:20
Air Blank	0.000	02:21
Control Test	0.078	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,JOSHUA,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HARRILL,ANDERSEN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: SIMOGA

DATE/TIME OF TEST: 06/29/2020 22:10:13

Diagnostics Check	OK	22:03
Air Blank	0.000	22:04
Control Test	0.077	22:04
Air Blank	0.000	22:05
Subject Sample # 1	0.238	22:05
Breath Volume	1.660	22:05
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.220	22:08
Breath Volume	1.863	22:08
Air Blank	0.000	22:09
Control Test	0.075	22:09
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JARRETT,RASHAHN,C
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HRADECKY,ALEXANDRIA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 06/30/2020 05:55:51

Diagnostics Check	OK	05:54	
Air Blank	0.000	05:54	
Control Test	0.078	05:55	
Air Blank		05:55	RFI Detect
Air Blank	0.000	05:55	

Data Download Date/Time: 07/21/2020 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JARRTT,RASHAHN,C
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HRADECKY,ALEXANDRIA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181702
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 06/30/2020 06:08:31

Diagnostics Check		OK	06:00
Air Blank		0.000	06:01
Control Test		0.078	06:01
Air Blank		0.000	06:02
Subject Sample	# 1	0.214	06:03
Breath Volume		1.386	06:03
Air Blank		0.000	06:04
Air Blank		0.000	06:05
Subject Sample	# 2	0.213	06:06
Breath Volume		1.457	06:06
Air Blank		0.000	06:07
Control Test		0.075	06:07
Air Blank		0.000	06:08
Diagnostics Check		OK	06:08

Data Download Date/Time: 07/21/2020 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COBO,JEANETTE,
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: COBO,JEANETTE
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 06/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/29/2020 00:12:59

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.078	00:07
Air Blank	0.000	00:08
Subject Sample # 1	0.000	00:08
Breath Volume	1.859	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.000	00:11
Breath Volume	2.230	00:11
Air Blank	0.000	00:12
Control Test	0.076	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 07/29/2020 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ODOR,ROBERT,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ODOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: DIGUGLIELMO

DATE/TIME OF TEST: 06/16/2020 01:37:42

Diagnostics Check		OK	01:31
Air Blank		0.000	01:32
Control Test		0.078	01:32
Air Blank		0.000	01:32
Subject Sample	# 1	0.000	01:33
Breath Volume		2.023	01:33
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.000	01:36
Breath Volume		2.332	01:36
Air Blank		0.000	01:36
Control Test		0.078	01:37
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 07/19/2020 20:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ODOR,ROBERT,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ODOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: SHAH

DATE/TIME OF TEST: 06/20/2020 22:56:08

Diagnostics Check	OK	22:46	
Air Blank	0.000	22:46	
Control Test	0.076	22:47	
Air Blank	0.000	22:47	
Subject Sample	# 1 0.235	22:50	
Breath Volume	1.187	22:50	
Air Blank	0.000	22:51	
Air Blank	0.000	22:53	
Subject Sample	# 2 0.250	22:54	
Breath Volume	1.585	22:54	
Air Blank	0.000	22:55	
Control Test	0.074	22:55	Control Outside Tolerance
Air Blank	0.000	22:56	

Data Download Date/Time: 07/19/2020 20:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YEPEZ,GHERMAINE
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: MOYA

DATE/TIME OF TEST: 06/02/2020 21:31:20

Diagnostics Check	OK	21:23
Air Blank	0.000	21:24
Control Test	0.080	21:24
Air Blank	0.000	21:24
Subject Sample # 1	0.144	21:26
Breath Volume	1.398	21:26
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.152	21:29
Breath Volume	2.820	21:29
Air Blank	0.000	21:30
Control Test	0.079	21:30
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

Data Download Date/Time: 06/26/2020 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALEZ,ROBERTO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: HALSEY

DATE/TIME OF TEST: 06/03/2020 19:30:47

Diagnostics Check	OK	19:25	
Air Blank	0.000	19:26	
Control Test	0.079	19:26	
Air Blank	0.000	19:27	
Subject Sample	# 1	19:28	Subject Test Refused
Breath Volume	1.066	19:28	
Air Blank	0.000	19:29	
Control Test	0.079	19:30	
Air Blank	0.000	19:30	
Diagnostics Check	OK	19:30	

Data Download Date/Time: 06/26/2020 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MARTINI,BRYAN
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 06/03/2020 20:10:49

Diagnostics Check	OK	20:04
Air Blank	0.000	20:04
Control Test	0.079	20:05
Air Blank	0.000	20:05
Subject Sample # 1	0.000	20:06
Breath Volume	2.417	20:06
Air Blank	0.000	20:06
Air Blank	0.000	20:08
Subject Sample # 2	0.000	20:09
Breath Volume	3.281	20:09
Air Blank	0.000	20:09
Control Test	0.079	20:10
Air Blank	0.000	20:10
Diagnostics Check	OK	20:10

Data Download Date/Time: 06/26/2020 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MITCHELL,COLE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: WEIL

DATE/TIME OF TEST: 06/07/2020 01:40:19

Diagnostics Check		OK	01:32
Air Blank		0.000	01:32
Control Test		0.080	01:32
Air Blank		0.000	01:33
Subject Sample	# 1	0.055	01:34
Breath Volume		2.097	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:36
Subject Sample	# 2	0.056	01:38
Breath Volume		1.289	01:38
Air Blank		0.000	01:39
Control Test		0.079	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 06/26/2020 13:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: LOPEZ HERNANDZ

DATE/TIME OF TEST: 06/07/2020 03:42:37

Diagnostics Check		OK	03:35
Air Blank		0.000	03:36
Control Test		0.081	03:36
Air Blank		0.000	03:37
Subject Sample	# 1	0.195	03:37
Breath Volume		2.066	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.196	03:40
Breath Volume		1.949	03:40
Air Blank		0.000	03:41
Control Test		0.079	03:41
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 06/26/2020 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SITZ,IAN
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: TUOTO

DATE/TIME OF TEST: 06/07/2020 18:54:02

Diagnostics Check	OK	18:41
Air Blank	0.000	18:42
Control Test	0.081	18:42
Air Blank	0.000	18:43
Subject Sample # 1	0.221	18:44
Breath Volume	3.031	18:44
Air Blank	0.000	18:45
Air Blank	0.000	18:46
Subject Sample # 2	0.196	18:47
Breath Volume	1.558	18:47
Air Blank	0.000	18:48
Air Blank	0.000	18:49
Subject Sample # 3	0.192	18:52
Breath Volume	1.406	18:52
Air Blank	0.000	18:53
Control Test	0.080	18:53
Air Blank	0.000	18:53
Diagnostics Check	OK	18:54

Data Download Date/Time: 06/26/2020 13:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PRADO,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 06/07/2020 20:47:50

Diagnostics Check	OK	20:40
Air Blank	0.000	20:40
Control Test	0.080	20:41
Air Blank	0.000	20:41
Subject Sample # 1	0.327	20:42
Breath Volume	2.878	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.321	20:45
Breath Volume	2.902	20:45
Air Blank	0.000	20:46
Control Test	0.079	20:47
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 06/26/2020 13:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FORTEZA,LUIS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: BRENGLE

DATE/TIME OF TEST: 06/07/2020 21:38:29

Diagnostics Check	OK	21:31
Air Blank	0.000	21:31
Control Test	0.079	21:32
Air Blank	0.000	21:32
Subject Sample # 1	0.000	21:33
Breath Volume	2.644	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.000	21:37
Breath Volume	2.656	21:37
Air Blank	0.000	21:37
Control Test	0.079	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 06/26/2020 13:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MACK,CHEYENNE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: OROZCO VALENZUELA

DATE/TIME OF TEST: 06/10/2020 20:17:47

Diagnostics Check	OK	20:10
Air Blank	0.000	20:10
Control Test	0.080	20:11
Air Blank	0.000	20:11
Subject Sample # 1	0.309	20:12
Breath Volume	2.988	20:12
Air Blank	0.000	20:13
Air Blank	0.000	20:15
Subject Sample # 2	0.310	20:15
Breath Volume	3.093	20:15
Air Blank	0.000	20:16
Control Test	0.078	20:17
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 06/26/2020 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: THOMAS,MICAH
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: VELASQUEZ MAZARIEGOS

DATE/TIME OF TEST: 06/11/2020 00:06:07

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:01	
Control Test	0.080	00:01	
Air Blank	0.000	00:02	
Subject Sample # 1	0.238	00:04	
Breath Volume	2.492	00:04	
Air Blank		00:05	Purge Fail
Air Blank	0.000	00:06	

Data Download Date/Time: 06/26/2020 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR, NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: BARNARD

DATE/TIME OF TEST: 06/11/2020 00:59:13

Diagnostics Check		OK	00:51
Air Blank		0.000	00:51
Control Test		0.078	00:51
Air Blank		0.000	00:52
Subject Sample	# 1	0.235	00:53
Breath Volume		2.398	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:55
Subject Sample	# 2	0.246	00:57
Breath Volume		1.839	00:57
Air Blank		0.000	00:58
Control Test		0.077	00:58
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

Data Download Date/Time: 06/26/2020 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR,NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: BURTON

DATE/TIME OF TEST: 06/12/2020 02:50:45

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.195	02:45
Breath Volume	1.425	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:47
Subject Sample # 2	0.208	02:48
Breath Volume	1.515	02:48
Air Blank	0.000	02:49
Control Test	0.078	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 06/26/2020 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU,YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: CONNORS

DATE/TIME OF TEST: 06/13/2020 04:03:13

Diagnostics Check		OK	03:53
Air Blank		0.000	03:54
Control Test		0.081	03:54
Air Blank		0.000	03:55
Subject Sample	# 1	0.128	03:57
Breath Volume		1.312	03:57
Air Blank		0.000	03:58
Air Blank		0.000	03:59
Subject Sample	# 2	0.125	04:01
Breath Volume		1.171	04:01
Air Blank		0.000	04:02
Control Test		0.076	04:02
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 06/26/2020 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCGILLICUDDY,STEVEN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: MUCHUCHCHONAY

DATE/TIME OF TEST: 06/13/2020 22:40:11

Diagnostics Check	OK	22:33
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.377	22:35
Breath Volume	2.449	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:37
Subject Sample # 2	0.374	22:38
Breath Volume	2.332	22:38
Air Blank	0.000	22:39
Control Test	0.077	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 06/26/2020 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CRUZ,GARDENIA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:51

SUBJECT'S LAST NAME: TIRADO

DATE/TIME OF TEST: 06/14/2020 18:25:42

Diagnostics Check	OK	18:18
Air Blank	0.000	18:19
Control Test	0.079	18:19
Air Blank	0.000	18:20
Subject Sample # 1	0.000	18:21
Breath Volume	2.152	18:21
Air Blank	0.000	18:21
Air Blank	0.000	18:23
Subject Sample # 2	0.000	18:24
Breath Volume	2.144	18:24
Air Blank	0.000	18:24
Control Test	0.079	18:25
Air Blank	0.000	18:25
Diagnostics Check	OK	18:25

Data Download Date/Time: 06/26/2020 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CLAYTON,BRANDI
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: WISCH

DATE/TIME OF TEST: 06/16/2020 13:51:18

Diagnostics Check	OK	13:42
Air Blank	0.000	13:42
Control Test	0.080	13:43
Air Blank	0.000	13:43
Subject Sample # 1	0.202	13:46
Breath Volume	1.640	13:46
Air Blank	0.000	13:47
Air Blank	0.000	13:48
Subject Sample # 2	0.203	13:49
Breath Volume	1.894	13:49
Air Blank	0.000	13:50
Control Test	0.075	13:50
Air Blank	0.000	13:51
Diagnostics Check	OK	13:51

Data Download Date/Time: 06/26/2020 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MENDEZ,JOSE
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: SZAVA

DATE/TIME OF TEST: 06/17/2020 20:19:10

Diagnostics Check	OK	20:15	
Air Blank	0.000	20:15	
Control Test	0.081	20:15	
Air Blank	0.000	20:16	
Subject Sample	# 1	20:17	Subject Test Refused
Breath Volume	0.000	20:17	
Air Blank	0.000	20:18	
Control Test	0.079	20:18	
Air Blank	0.000	20:19	
Diagnostics Check	OK	20:19	

Data Download Date/Time: 06/26/2020 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CASAS,JAVIER
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: BRONARD

DATE/TIME OF TEST: 06/17/2020 21:50:12

Diagnostics Check		OK	21:41
Air Blank		0.000	21:41
Control Test		0.081	21:42
Air Blank		0.000	21:42
Subject Sample	# 1	0.248	21:44
Breath Volume		2.453	21:44
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.238	21:48
Breath Volume		2.136	21:48
Air Blank		0.000	21:49
Control Test		0.079	21:49
Air Blank		0.000	21:50
Diagnostics Check		OK	21:50

Data Download Date/Time: 06/26/2020 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR, NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: SHANK III

DATE/TIME OF TEST: 06/18/2020 02:55:59

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.000	02:51
Breath Volume	4.246	02:51
Air Blank	0.000	02:51
Air Blank	0.000	02:54
Subject Sample # 2	0.000	02:54
Breath Volume	3.207	02:54
Air Blank	0.000	02:55
Control Test	0.078	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 06/26/2020 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: DANIEL

DATE/TIME OF TEST: 06/19/2020 22:49:24

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.080	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.101	22:43
Breath Volume	2.300	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.111	22:47
Breath Volume	1.328	22:47
Air Blank	0.000	22:48
Control Test	0.076	22:48
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 06/26/2020 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCGILLICUDDY,STEVEN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: CONTE

DATE/TIME OF TEST: 06/20/2020 23:40:49

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.080	23:34
Air Blank	0.000	23:34
Subject Sample # 1	0.201	23:35
Breath Volume	1.246	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.206	23:38
Breath Volume	1.421	23:38
Air Blank	0.000	23:39
Control Test	0.077	23:40
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 06/26/2020 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR,NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: KURTZO

DATE/TIME OF TEST: 06/21/2020 04:50:53

Diagnostics Check	OK	04:44
Air Blank	0.000	04:44
Control Test	0.080	04:44
Air Blank	0.000	04:45
Subject Sample # 1	0.141	04:45
Breath Volume	2.406	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.138	04:49
Breath Volume	2.335	04:49
Air Blank	0.000	04:49
Control Test	0.078	04:50
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 06/26/2020 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH, GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: QUIJAS JIMENEZ

DATE/TIME OF TEST: 06/25/2020 02:26:46

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.164	02:21
Breath Volume	3.410	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.162	02:24
Breath Volume	3.078	02:24
Air Blank	0.000	02:25
Control Test	0.078	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 06/26/2020 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH, GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: PAWELSKI

DATE/TIME OF TEST: 06/25/2020 22:20:27

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.081	22:14
Air Blank	0.000	22:14
Subject Sample # 1	0.175	22:15
Breath Volume	2.539	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.177	22:18
Breath Volume	1.625	22:18
Air Blank	0.000	22:19
Control Test	0.079	22:19
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 06/26/2020 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TUANG, NGIN
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 05/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 06/26/2020 02:18:09

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.000	02:13
Breath Volume	2.128	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.000	02:16
Breath Volume	1.296	02:16
Air Blank	0.000	02:17
Control Test	0.079	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 06/26/2020 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: THOMAS,MICAH
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:37

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 06/26/2020 18:11:37

Diagnostics Check	OK	18:02
Air Blank	0.000	18:03
Control Test	0.080	18:03
Air Blank	0.000	18:04
Subject Sample # 1	0.000	18:07
Breath Volume	4.804	18:07
Air Blank	0.000	18:07
Air Blank	0.000	18:09
Subject Sample # 2	0.000	18:10
Breath Volume	4.855	18:10
Air Blank	0.000	18:10
Control Test	0.079	18:10
Air Blank	0.000	18:11
Diagnostics Check	OK	18:11

Data Download Date/Time: 07/17/2020 17:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ACOSTA,NERELYS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: CHATMAN

DATE/TIME OF TEST: 06/27/2020 02:37:21

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.000	02:33
Breath Volume	2.777	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.000	02:35
Breath Volume	2.734	02:35
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 07/17/2020 17:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 06/27/2020 21:49:25

Diagnostics Check		OK	21:42
Air Blank		0.000	21:43
Control Test		0.081	21:43
Air Blank		0.000	21:43
Subject Sample	# 1	0.158	21:44
Breath Volume		2.039	21:44
Air Blank		0.000	21:45
Air Blank		0.000	21:46
Subject Sample	# 2	0.165	21:47
Breath Volume		2.679	21:47
Air Blank		0.000	21:48
Control Test		0.079	21:48
Air Blank		0.000	21:49
Diagnostics Check		OK	21:49

Data Download Date/Time: 07/17/2020 17:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MENDEZ,JOSE
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/27/2020 23:14:33

Diagnostics Check	OK	23:07
Air Blank	0.000	23:08
Control Test	0.080	23:08
Air Blank	0.000	23:09
Subject Sample # 1	0.297	23:09
Breath Volume	2.234	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.295	23:12
Breath Volume	1.890	23:12
Air Blank	0.000	23:13
Control Test	0.079	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 07/17/2020 17:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SORIA,KEITH
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: HANNIGAN

DATE/TIME OF TEST: 06/28/2020 00:26:43

Diagnostics Check		OK	00:19
Air Blank		0.000	00:20
Control Test		0.079	00:20
Air Blank		0.000	00:21
Subject Sample	# 1	0.165	00:22
Breath Volume		2.046	00:22
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.164	00:24
Breath Volume		1.820	00:24
Air Blank		0.000	00:25
Control Test		0.078	00:26
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 07/17/2020 17:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WITT,TYLER
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: JOHNS

DATE/TIME OF TEST: 06/28/2020 02:44:57

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.080	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.081	02:40
Breath Volume	1.757	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:42
Subject Sample # 2	0.085	02:43
Breath Volume	2.207	02:43
Air Blank	0.000	02:43
Control Test	0.079	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 07/17/2020 17:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROMERO,J
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: LYON JR

DATE/TIME OF TEST: 06/29/2020 01:50:43

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.047	01:45
Breath Volume	1.550	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.047	01:49
Breath Volume	2.394	01:49
Air Blank	0.000	01:49
Control Test	0.079	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 07/17/2020 17:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:27

SUBJECT'S LAST NAME: YANCHURA

DATE/TIME OF TEST: 06/29/2020 06:59:34

Diagnostics Check		OK	06:52
Air Blank		0.000	06:53
Control Test		0.080	06:53
Air Blank		0.000	06:54
Subject Sample	# 1	0.197	06:54
Breath Volume		4.300	06:54
Air Blank		0.000	06:55
Air Blank		0.000	06:57
Subject Sample	# 2	0.198	06:57
Breath Volume		4.660	06:57
Air Blank		0.000	06:58
Control Test		0.078	06:58
Air Blank		0.000	06:59
Diagnostics Check		OK	06:59

Data Download Date/Time: 07/17/2020 17:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH, GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: MONGIOVI

DATE/TIME OF TEST: 06/29/2020 23:11:45

Diagnostics Check		OK	22:59	
Air Blank		0.000	23:00	
Control Test		0.080	23:00	
Air Blank		0.000	23:01	
Subject Sample	# 1	0.052	23:04	Volume Not Met
Breath Volume		0.070	23:04	
Air Blank		0.000	23:05	
Air Blank		0.000	23:06	
Subject Sample	# 2	0.046	23:09	Volume Not Met
Breath Volume		0.871	23:09	
Air Blank		0.000	23:10	
Control Test		0.077	23:11	
Air Blank		0.000	23:11	
Diagnostics Check		OK	23:11	

Data Download Date/Time: 07/17/2020 17:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH, GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: MONGIOVI

DATE/TIME OF TEST: 06/29/2020 23:17:49

Diagnostics Check	OK	23:13	
Air Blank	0.000	23:13	
Control Test	0.077	23:14	
Air Blank	0.000	23:14	
Subject Sample	# 1	23:16	Subject Test Refused
Breath Volume	0.828	23:16	
Air Blank	0.000	23:16	
Control Test	0.078	23:17	
Air Blank	0.000	23:17	
Diagnostics Check	OK	23:17	

Data Download Date/Time: 07/17/2020 17:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: LORENZO

DATE/TIME OF TEST: 06/26/2020 23:08:12

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.080	23:02
Air Blank	0.000	23:02
Subject Sample # 1	0.123	23:03
Breath Volume	1.304	23:03
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.123	23:06
Breath Volume	2.406	23:06
Air Blank	0.000	23:07
Control Test	0.080	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 07/17/2020 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 06/27/2020 02:25:06

Diagnostics Check		OK	02:18
Air Blank		0.000	02:19
Control Test		0.080	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.175	02:20
Breath Volume		3.394	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:22
Subject Sample	# 2	0.177	02:23
Breath Volume		3.777	02:23
Air Blank		0.000	02:24
Control Test		0.079	02:24
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 07/17/2020 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: AGUILAR OLIVA

DATE/TIME OF TEST: 06/27/2020 23:53:46

Diagnostics Check	OK	23:46
Air Blank	0.000	23:47
Control Test	0.080	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.249	23:48
Breath Volume	2.535	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.246	23:52
Breath Volume	2.257	23:52
Air Blank	0.000	23:52
Control Test	0.079	23:53
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 07/17/2020 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: ZELLA

DATE/TIME OF TEST: 06/28/2020 01:56:05

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.115	01:51
Breath Volume	3.023	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:53
Subject Sample # 2	0.115	01:54
Breath Volume	2.664	01:54
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 07/17/2020 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: IVORY,BRELAN
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: SMITH DODGIN

DATE/TIME OF TEST: 06/28/2020 03:06:13

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.080	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.168	03:01
Breath Volume	2.785	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:03
Subject Sample # 2	0.152	03:04
Breath Volume	1.503	03:04
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 07/17/2020 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MASTERSON,MONIQUE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:07

SUBJECT'S LAST NAME: PICKENS

DATE/TIME OF TEST: 06/29/2020 16:36:15

Diagnostics Check	OK	16:31	
Air Blank	0.000	16:31	
Control Test	0.080	16:32	
Air Blank	0.000	16:32	
Subject Sample	# 1	16:34	Subject Test Refused
Breath Volume	0.000	16:34	
Air Blank	0.000	16:35	
Control Test	0.081	16:35	
Air Blank	0.000	16:36	
Diagnostics Check	OK	16:36	

Data Download Date/Time: 07/17/2020 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: REED,CHARLES
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: SLOANE

DATE/TIME OF TEST: 06/29/2020 19:33:56

Diagnostics Check	OK	19:27
Air Blank	0.000	19:27
Control Test	0.081	19:27
Air Blank	0.000	19:28
Subject Sample # 1	0.000	19:29
Breath Volume	1.898	19:29
Air Blank	0.000	19:29
Air Blank	0.000	19:31
Subject Sample # 2	0.000	19:32
Breath Volume	2.000	19:32
Air Blank	0.000	19:32
Control Test	0.080	19:33
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

Data Download Date/Time: 07/17/2020 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCGILLICUDDY,STEVEN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: TAMBRIS

DATE/TIME OF TEST: 06/29/2020 20:22:26

Diagnostics Check	OK	20:15
Air Blank	0.000	20:16
Control Test	0.080	20:16
Air Blank	0.000	20:16
Subject Sample # 1	0.277	20:17
Breath Volume	1.246	20:17
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample # 2	0.264	20:20
Breath Volume	1.226	20:20
Air Blank	0.000	20:21
Control Test	0.080	20:21
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

Data Download Date/Time: 07/17/2020 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: BRENNECKE

DATE/TIME OF TEST: 06/29/2020 21:55:51

Diagnostics Check	OK	21:49
Air Blank	0.000	21:49
Control Test	0.080	21:49
Air Blank	0.000	21:50
Subject Sample # 1	0.201	21:51
Breath Volume	1.496	21:51
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.201	21:54
Breath Volume	1.605	21:54
Air Blank	0.000	21:54
Control Test	0.080	21:55
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 07/17/2020 14:01