

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: BOLOSH

DATE/TIME OF TEST: 06/02/2020 01:18:37

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.080	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.016	01:14
Breath Volume	3.285	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.013	01:17
Breath Volume	2.945	01:17
Air Blank	0.000	01:17
Control Test	0.079	01:18
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 06/08/2020 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: VAN OSTENBRIDGE

DATE/TIME OF TEST: 06/15/2020 00:27:50

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:25	
Control Test	0.080	00:25	
Air Blank	0.000	00:26	
Subject Sample	# 1	00:26	Subject Test Refused
Breath Volume	0.000	00:26	
Air Blank	0.000	00:26	
Control Test	0.080	00:27	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

Data Download Date/Time: 07/13/2020 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/15/2020 00:41:23

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:30	
Control Test	0.080	00:30	
Air Blank	0.000	00:30	
Subject Sample	# 1 0.237	00:33	Volume Not Met
Breath Volume	0.761	00:33	
Air Blank	0.000	00:34	
Air Blank	0.000	00:36	
Subject Sample	# 2 0.265	00:39	Volume Not Met
Breath Volume	0.156	00:39	
Air Blank	0.000	00:40	
Control Test	0.077	00:40	
Air Blank	0.000	00:41	
Diagnostics Check	OK	00:41	

Data Download Date/Time: 07/13/2020 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,LUCCIANO,R
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: WILSON,MATTHEW
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: HANSON

DATE/TIME OF TEST: 06/19/2020 19:46:32

Diagnostics Check	OK	19:39
Air Blank	0.000	19:40
Control Test	0.080	19:40
Air Blank	0.000	19:41
Subject Sample # 1	0.000	19:42
Breath Volume	1.847	19:42
Air Blank	0.000	19:42
Air Blank	0.000	19:44
Subject Sample # 2	0.000	19:45
Breath Volume	1.687	19:45
Air Blank	0.000	19:45
Control Test	0.079	19:45
Air Blank	0.000	19:46
Diagnostics Check	OK	19:46

Data Download Date/Time: 07/13/2020 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,ROBERT,J
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: DAVIS,D
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: MANLEY

DATE/TIME OF TEST: 06/20/2020 04:03:38

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.184	03:58
Breath Volume	2.058	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.183	04:01
Breath Volume	1.308	04:01
Air Blank	0.000	04:02
Control Test	0.078	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 07/13/2020 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: SANDS,OFC
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: EBERHART

DATE/TIME OF TEST: 06/20/2020 23:30:02

Diagnostics Check	OK	23:23
Air Blank	0.000	23:24
Control Test	0.080	23:24
Air Blank	0.000	23:25
Subject Sample # 1	0.205	23:25
Breath Volume	2.464	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.199	23:28
Breath Volume	2.449	23:28
Air Blank	0.000	23:29
Control Test	0.078	23:29
Air Blank	0.000	23:29
Diagnostics Check	OK	23:30

Data Download Date/Time: 07/13/2020 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: LACEY

DATE/TIME OF TEST: 06/27/2020 23:08:31

Diagnostics Check	OK	23:01
Air Blank	0.000	23:02
Control Test	0.080	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.164	23:03
Breath Volume	2.066	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.160	23:06
Breath Volume	1.875	23:06
Air Blank	0.000	23:07
Control Test	0.080	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 07/13/2020 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMBERLIN,KEVIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TOMBERLIN,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: VENTURA

DATE/TIME OF TEST: 06/28/2020 02:54:19

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.116	02:49
Breath Volume	1.656	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.124	02:52
Breath Volume	3.039	02:52
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 07/13/2020 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMBERLIN,KEVIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TOMBERLIN,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: HICKEY

DATE/TIME OF TEST: 06/28/2020 23:47:41

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.260	23:42
Breath Volume	1.488	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.241	23:46
Breath Volume	2.093	23:46
Air Blank	0.000	23:46
Control Test	0.077	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 07/13/2020 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: MENDEZ MUNGUIA

DATE/TIME OF TEST: 06/29/2020 04:22:37

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.079	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.187	04:17
Breath Volume	2.214	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.189	04:21
Breath Volume	1.515	04:21
Air Blank	0.000	04:21
Control Test	0.080	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 07/13/2020 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/01/2020 02:21:31

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.197	02:15
Breath Volume	1.550	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.206	02:19
Breath Volume	2.617	02:19
Air Blank	0.000	02:20
Control Test	0.079	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 06/11/2020 10:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PODHYSKI,PHILIP,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PODHYSKI,PHILIP
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: RAMIREZ VELASQUEZ

DATE/TIME OF TEST: 06/01/2020 04:49:38

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:46	
Control Test	0.080	04:47	
Air Blank	0.000	04:47	
Subject Sample	# 1	04:48	Subject Test Refused
Breath Volume	0.000	04:48	
Air Blank	0.000	04:48	
Control Test	0.080	04:49	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:49	

Data Download Date/Time: 06/11/2020 10:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/01/2020 20:38:24

Diagnostics Check	OK	20:30
Air Blank	0.000	20:31
Control Test	0.080	20:31
Air Blank	0.000	20:31
Subject Sample # 1	0.000	20:33
Breath Volume	1.593	20:33
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample # 2	0.000	20:37
Breath Volume	1.523	20:37
Air Blank	0.000	20:37
Control Test	0.080	20:37
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

Data Download Date/Time: 06/04/2020 14:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ARCHAMBEAU,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: BATCHIK

DATE/TIME OF TEST: 06/03/2020 23:38:11

Diagnostics Check	OK	23:31
Air Blank	0.000	23:32
Control Test	0.080	23:32
Air Blank	0.000	23:33
Subject Sample # 1	0.196	23:33
Breath Volume	1.226	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.204	23:36
Breath Volume	1.648	23:36
Air Blank	0.000	23:37
Control Test	0.079	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 06/04/2020 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: VINA

DATE/TIME OF TEST: 06/04/2020 01:36:06

Diagnostics Check		OK	01:29
Air Blank		0.000	01:30
Control Test		0.079	01:30
Air Blank		0.000	01:30
Subject Sample	# 1	0.265	01:31
Breath Volume		2.585	01:31
Air Blank		0.000	01:32
Air Blank		0.000	01:33
Subject Sample	# 2	0.259	01:34
Breath Volume		2.433	01:34
Air Blank		0.000	01:35
Control Test		0.079	01:35
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 06/04/2020 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: RUDD

DATE/TIME OF TEST: 06/02/2020 11:07:24

Diagnostics Check	OK	11:00
Air Blank	0.000	11:01
Control Test	0.081	11:01
Air Blank	0.000	11:01
Subject Sample # 1	0.000	11:02
Breath Volume	1.515	11:02
Air Blank	0.000	11:03
Air Blank	0.000	11:05
Subject Sample # 2	0.000	11:06
Breath Volume	1.468	11:06
Air Blank	0.000	11:06
Control Test	0.081	11:06
Air Blank	0.000	11:07
Diagnostics Check	OK	11:07

Data Download Date/Time: 06/27/2020 16:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOGDON,BRAD,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: NGUYEN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/23/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:29

SUBJECT'S LAST NAME: RUDD

DATE/TIME OF TEST: 06/07/2020 13:55:28

Diagnostics Check	OK	13:49
Air Blank	0.000	13:49
Control Test	0.079	13:50
Air Blank	0.000	13:50
Subject Sample # 1	0.000	13:51
Breath Volume	1.710	13:51
Air Blank	0.000	13:51
Air Blank	0.000	13:53
Subject Sample # 2	0.000	13:54
Breath Volume	1.585	13:54
Air Blank	0.000	13:54
Control Test	0.080	13:54
Air Blank	0.000	13:55
Diagnostics Check	OK	13:55

Data Download Date/Time: 06/27/2020 16:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRCLOTH, EDWARD, W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: ANDERSON, MICHAEL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: LIEVANO

DATE/TIME OF TEST: 06/08/2020 01:48:49

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:34	
Control Test	0.080	01:34	
Air Blank	0.000	01:35	
Subject Sample # 1	0.066	01:35	
Breath Volume	1.652	01:35	
Air Blank	0.000	01:36	
Air Blank	0.000	01:38	
Subject Sample # 2	0.050	01:41	Volume Not Met
Breath Volume	0.820	01:41	
Air Blank	0.000	01:42	
Air Blank	0.000	01:44	
Subject Sample # 3	0.047	01:47	Volume Not Met
Breath Volume	0.386	01:47	
Air Blank	0.000	01:47	
Control Test	0.078	01:48	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

Data Download Date/Time: 06/27/2020 16:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRCLOTH, EDWARD, W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: ANDERSON, MICHAEL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: LIEVANO

DATE/TIME OF TEST: 06/08/2020 02:29:05

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:17	
Control Test	0.080	02:17	
Air Blank	0.000	02:18	
Subject Sample # 1	0.040	02:18	
Breath Volume	1.121	02:18	
Air Blank	0.000	02:19	
Air Blank	0.000	02:21	
Subject Sample # 2	0.040	02:24	Volume Not Met
Breath Volume	0.816	02:24	
Air Blank	0.000	02:25	
Air Blank	0.000	02:26	
Subject Sample # 3	0.044	02:27	
Breath Volume	2.718	02:27	
Air Blank	0.000	02:28	
Control Test	0.080	02:28	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

Data Download Date/Time: 06/27/2020 16:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: LICHTY,DAVID
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/13/2020 04:26:48

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:24	
Control Test	0.081	04:24	
Air Blank	0.000	04:25	
Subject Sample	# 1	04:25	Subject Test Refused
Breath Volume	0.000	04:25	
Air Blank	0.000	04:25	
Control Test	0.080	04:26	
Air Blank	0.000	04:26	
Diagnostics Check	OK	04:26	

Data Download Date/Time: 06/27/2020 16:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ST.JOHN,J,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: ST.JOHN,J
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 06/18/2020 02:56:47

Diagnostics Check		OK	02:49
Air Blank		0.000	02:50
Control Test		0.080	02:50
Air Blank		0.000	02:51
Subject Sample	# 1	0.250	02:51
Breath Volume		2.917	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.231	02:55
Breath Volume		2.203	02:55
Air Blank		0.000	02:55
Control Test		0.075	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 06/27/2020 16:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRCLOTH,EDWARD,W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: YEOMAN,T
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/20/2020 22:02:27

Diagnostics Check		OK	21:56
Air Blank		0.000	21:56
Control Test		0.079	21:56
Air Blank		0.000	21:57
Subject Sample	# 1	0.248	21:57
Breath Volume		1.691	21:57
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.262	22:00
Breath Volume		1.667	22:00
Air Blank		0.000	22:01
Control Test		0.079	22:01
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 06/27/2020 16:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRCLOTH,EDWARD,W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: LEWIS,R
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: HARMAN

DATE/TIME OF TEST: 06/21/2020 00:26:08

Diagnostics Check		OK	00:19
Air Blank		0.000	00:19
Control Test		0.080	00:19
Air Blank		0.000	00:20
Subject Sample	# 1	0.138	00:21
Breath Volume		1.320	00:21
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.130	00:24
Breath Volume		1.726	00:24
Air Blank		0.000	00:25
Control Test		0.080	00:25
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 06/27/2020 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRCLOTH,EDWARD,W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: LEWIS,R
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 06/21/2020 18:58:57

Diagnostics Check	OK	18:52
Air Blank	0.000	18:52
Control Test	0.080	18:53
Air Blank	0.000	18:53
Subject Sample # 1	0.189	18:54
Breath Volume	1.937	18:54
Air Blank	0.000	18:55
Air Blank	0.000	18:56
Subject Sample # 2	0.191	18:57
Breath Volume	1.785	18:57
Air Blank	0.000	18:58
Control Test	0.079	18:58
Air Blank	0.000	18:58
Diagnostics Check	OK	18:58

Data Download Date/Time: 06/27/2020 16:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRCLOTH, EDWARD, W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: KOTOWSKI, P
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/21/2020 22:33:05

Diagnostics Check		OK	22:20	
Air Blank		0.000	22:20	
Control Test		0.080	22:21	
Air Blank		0.000	22:21	
Subject Sample	# 1	0.185	22:24	Slope Not Level
Breath Volume		1.105	22:24	
Air Blank		0.000	22:25	
Air Blank		0.000	22:27	
Subject Sample	# 2	0.225	22:27	
Breath Volume		2.511	22:27	
Air Blank		0.000	22:28	
Air Blank		0.000	22:30	
Subject Sample	# 3	0.226	22:31	
Breath Volume		1.605	22:31	
Air Blank		0.000	22:32	
Control Test		0.078	22:32	
Air Blank		0.000	22:33	
Diagnostics Check		OK	22:33	

Data Download Date/Time: 06/27/2020 16:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRCLOTH,EDWARD,W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GORDON,K
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: BAYOUS

DATE/TIME OF TEST: 06/26/2020 00:51:29

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.118	00:45
Breath Volume	1.718	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.106	00:49
Breath Volume	1.968	00:49
Air Blank	0.000	00:50
Control Test	0.081	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 06/27/2020 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: DENKER,BRYAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 06/27/2020 01:16:16

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.081	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.142	01:10
Breath Volume	1.375	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.135	01:14
Breath Volume	1.441	01:14
Air Blank	0.000	01:15
Control Test	0.077	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 06/27/2020 16:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: DOMINGUEZ,JOSE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: BERRIOS MORALES

DATE/TIME OF TEST: 06/22/2020 01:45:08

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.081	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.198	01:40
Breath Volume	3.074	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.195	01:43
Breath Volume	3.769	01:43
Air Blank	0.000	01:44
Control Test	0.081	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 08/01/2020 04:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,NELSON,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PEREZ,NELSON
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 06/22/2020 21:42:09

Diagnostics Check	OK	21:36
Air Blank	0.000	21:36
Control Test	0.082	21:37
Air Blank	0.000	21:37
Subject Sample # 1	0.000	21:37
Breath Volume	1.937	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.000	21:40
Breath Volume	2.230	21:40
Air Blank	0.000	21:41
Control Test	0.082	21:41
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time: 08/01/2020 04:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SAVINON,CARLOS
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: LEMES

DATE/TIME OF TEST: 06/26/2020 02:02:19

Diagnostics Check	OK	01:59	
Air Blank	0.000	01:59	
Control Test	0.082	02:00	
Air Blank	0.000	02:00	
Subject Sample	# 1	02:00	Subject Test Refused
Breath Volume	0.000	02:00	
Air Blank	0.000	02:01	
Control Test	0.082	02:01	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 08/01/2020 04:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACHADO,ERICK,J
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: MACHADO,ERICK
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/18/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: LAFAURIE

DATE/TIME OF TEST: 06/26/2020 05:43:18

Diagnostics Check		OK	05:36
Air Blank		0.000	05:36
Control Test		0.082	05:37
Air Blank		0.000	05:37
Subject Sample	# 1	0.199	05:38
Breath Volume		1.535	05:38
Air Blank		0.000	05:39
Air Blank		0.000	05:41
Subject Sample	# 2	0.204	05:41
Breath Volume		1.339	05:41
Air Blank		0.000	05:42
Control Test		0.079	05:42
Air Blank		0.000	05:43
Diagnostics Check		OK	05:43

Data Download Date/Time: 08/01/2020 04:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: DOMINGUEZ,JOSE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: SANCHEZ CASTRO

DATE/TIME OF TEST: 06/29/2020 02:03:21

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:48	
Control Test	0.082	01:49	
Air Blank	0.000	01:49	
Subject Sample	# 1 0.141	01:50	
Breath Volume	1.371	01:50	
Air Blank	0.000	01:51	
Air Blank	0.000	01:53	
Subject Sample	# 2 0.120	01:56	Volume Not Met
Breath Volume	0.753	01:56	
Air Blank	0.000	01:57	
Air Blank	0.000	01:58	
Subject Sample	# 3 0.138	02:01	
Breath Volume	1.351	02:01	
Air Blank	0.000	02:02	
Control Test	0.079	02:02	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 08/01/2020 04:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: VILLATORO VASQUEZ

DATE/TIME OF TEST: 06/01/2020 02:32:39

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.176	02:28
Breath Volume	2.261	02:28
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.170	02:31
Breath Volume	1.566	02:31
Air Blank	0.000	02:31
Control Test	0.079	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 06/11/2020 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: RAMOS MATEO

DATE/TIME OF TEST: 06/06/2020 02:11:21

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:57	
Control Test	0.080	01:58	
Air Blank	0.000	01:58	
Subject Sample	# 1 0.197	02:01	Volume Not Met
Breath Volume	0.664	02:01	
Air Blank	0.000	02:02	
Air Blank	0.000	02:04	
Subject Sample	# 2 0.224	02:06	
Breath Volume	2.140	02:06	
Air Blank	0.000	02:06	
Air Blank	0.000	02:08	
Subject Sample	# 3 0.225	02:09	
Breath Volume	2.648	02:09	
Air Blank	0.000	02:10	
Control Test	0.077	02:10	
Air Blank	0.000	02:11	
Diagnostics Check	OK	02:11	

Data Download Date/Time: 06/11/2020 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENYAN,M,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MOORE,A
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: GENTSCH

DATE/TIME OF TEST: 06/08/2020 21:13:30

Diagnostics Check	OK	21:05
Air Blank	0.000	21:05
Control Test	0.078	21:06
Air Blank	0.000	21:06
Subject Sample # 1	0.033	21:08
Breath Volume	1.507	21:08
Air Blank	0.000	21:09
Air Blank	0.000	21:10
Subject Sample # 2	0.030	21:11
Breath Volume	2.812	21:11
Air Blank	0.000	21:12
Control Test	0.076	21:12
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

Data Download Date/Time: 07/13/2020 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: FUENTES ARAUJO

DATE/TIME OF TEST: 06/09/2020 02:40:50

Diagnostics Check		OK	02:34
Air Blank		0.000	02:34
Control Test		0.077	02:34
Air Blank		0.000	02:35
Subject Sample	# 1	0.156	02:36
Breath Volume		2.289	02:36
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.158	02:39
Breath Volume		2.707	02:39
Air Blank		0.000	02:39
Control Test		0.077	02:40
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 07/13/2020 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENYAN,M,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: DAVIS JR,D
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: ZIPPRIAN

DATE/TIME OF TEST: 06/10/2020 05:37:13

Diagnostics Check	OK	05:30
Air Blank	0.000	05:31
Control Test	0.077	05:31
Air Blank	0.000	05:31
Subject Sample # 1	0.129	05:32
Breath Volume	3.281	05:32
Air Blank	0.000	05:33
Air Blank	0.000	05:34
Subject Sample # 2	0.130	05:35
Breath Volume	3.648	05:35
Air Blank	0.000	05:36
Control Test	0.077	05:36
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 07/13/2020 12:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,KATRYNA,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: JENNINGS

DATE/TIME OF TEST: 06/11/2020 03:06:49

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.077	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.110	03:02
Breath Volume	2.281	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:04
Subject Sample # 2	0.109	03:05
Breath Volume	2.980	03:05
Air Blank	0.000	03:05
Control Test	0.077	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 07/13/2020 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 06/14/2020 01:03:51

Diagnostics Check		OK	00:56
Air Blank		0.000	00:56
Control Test		0.077	00:57
Air Blank		0.000	00:57
Subject Sample	# 1	0.114	00:58
Breath Volume		2.417	00:58
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.106	01:02
Breath Volume		1.605	01:02
Air Blank		0.000	01:02
Control Test		0.077	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 07/13/2020 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 06/14/2020 02:34:59

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.077	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.179	02:30
Breath Volume	1.351	02:30
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.170	02:33
Breath Volume	1.628	02:33
Air Blank	0.000	02:34
Control Test	0.077	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 07/13/2020 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/14/2020 03:15:17

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:12	
Control Test	0.078	03:12	
Air Blank	0.000	03:13	
Subject Sample	# 1	03:13	Subject Test Refused
Breath Volume	0.000	03:13	
Air Blank	0.000	03:14	
Control Test	0.078	03:14	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:15	

Data Download Date/Time: 07/13/2020 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: KALISH

DATE/TIME OF TEST: 06/14/2020 04:24:33

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.077	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.138	04:19
Breath Volume	2.175	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.138	04:22
Breath Volume	1.933	04:22
Air Blank	0.000	04:23
Control Test	0.077	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 07/13/2020 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/15/2020 00:20:55

Diagnostics Check	OK	00:05	
Air Blank	0.000	00:05	
Control Test	0.077	00:06	
Air Blank	0.000	00:06	
Subject Sample	# 1 0.277	00:09	Volume Not Met
Breath Volume	0.472	00:09	
Air Blank	0.000	00:11	
Air Blank	0.000	00:12	
Subject Sample	# 2 0.320	00:13	
Breath Volume	1.187	00:13	
Air Blank	0.000	00:14	
Air Blank	0.000	00:15	
Subject Sample	# 3 0.221	00:18	Volume Not Met
Breath Volume	0.632	00:18	
Air Blank	0.000	00:19	
Control Test	0.073	00:20	Control Outside Tolerance
Air Blank	0.000	00:20	

Data Download Date/Time: 07/13/2020 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VONADOR, JERARD, E
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: DAVIS, DONALD
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: HEARNS

DATE/TIME OF TEST: 06/17/2020 01:06:22

Diagnostics Check	OK	01:02	
Air Blank	0.000	01:03	
Control Test	0.077	01:03	
Air Blank	0.000	01:03	
Subject Sample	# 1	01:04	Subject Test Refused
Breath Volume	0.000	01:04	
Air Blank	0.000	01:05	
Control Test	0.077	01:05	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 07/13/2020 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VONADOR,JERARD,E
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: VONADOR,JERARD
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 06/17/2020 03:17:09

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:14	
Control Test	0.078	03:14	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:16	
Control Test	0.078	03:16	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 07/13/2020 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: CONRAD

DATE/TIME OF TEST: 06/17/2020 22:13:28

Diagnostics Check	OK	22:06
Air Blank	0.000	22:07
Control Test	0.077	22:07
Air Blank	0.000	22:08
Subject Sample # 1	0.099	22:08
Breath Volume	3.906	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:11
Subject Sample # 2	0.101	22:11
Breath Volume	3.117	22:11
Air Blank	0.000	22:12
Control Test	0.077	22:12
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

Data Download Date/Time: 07/13/2020 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SHERRY

DATE/TIME OF TEST: 06/18/2020 00:35:38

Diagnostics Check		OK	00:28
Air Blank		0.000	00:29
Control Test		0.078	00:29
Air Blank		0.000	00:30
Subject Sample	# 1	0.220	00:31
Breath Volume		2.269	00:31
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.231	00:33
Breath Volume		2.214	00:33
Air Blank		0.000	00:34
Control Test		0.077	00:35
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 07/13/2020 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: MARTINEZ BAUTISTA

DATE/TIME OF TEST: 06/18/2020 22:29:46

Diagnostics Check	OK	22:23
Air Blank	0.000	22:23
Control Test	0.078	22:23
Air Blank	0.000	22:24
Subject Sample # 1	0.182	22:25
Breath Volume	1.425	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:27
Subject Sample # 2	0.191	22:28
Breath Volume	2.171	22:28
Air Blank	0.000	22:28
Control Test	0.076	22:29
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 07/13/2020 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: MORENO GARZA

DATE/TIME OF TEST: 06/19/2020 02:03:50

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.077	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.204	01:58
Breath Volume	1.570	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.224	02:02
Breath Volume	1.652	02:02
Air Blank	0.000	02:02
Control Test	0.076	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 07/13/2020 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 06/19/2020 04:19:50

Diagnostics Check	OK	04:14	
Air Blank	0.000	04:15	
Control Test	0.077	04:15	
Air Blank	0.000	04:15	
Subject Sample	# 1	04:18	Subject Test Refused
Breath Volume	0.234	04:18	
Air Blank	0.000	04:18	
Control Test	0.078	04:19	
Air Blank	0.000	04:19	
Diagnostics Check	OK	04:19	

Data Download Date/Time: 07/13/2020 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,KATRYNA,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TAYLOR,KATRYNA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: GADSON-DAVIS

DATE/TIME OF TEST: 06/22/2020 06:17:46

Diagnostics Check	OK	06:14	
Air Blank	0.000	06:14	
Control Test	0.078	06:15	
Air Blank	0.000	06:15	
Subject Sample	# 1	06:16	Subject Test Refused
Breath Volume	0.000	06:16	
Air Blank	0.000	06:16	
Control Test	0.078	06:17	
Air Blank	0.000	06:17	
Diagnostics Check	OK	06:17	

Data Download Date/Time: 07/13/2020 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: SEASON

DATE/TIME OF TEST: 06/22/2020 22:11:15

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.078	22:03
Air Blank	0.000	22:03
Subject Sample # 1	0.097	22:06
Breath Volume	1.992	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.102	22:09
Breath Volume	1.535	22:09
Air Blank	0.000	22:10
Control Test	0.076	22:10
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 07/13/2020 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:53

SUBJECT'S LAST NAME: STOLTIE

DATE/TIME OF TEST: 06/23/2020 16:30:38

Diagnostics Check	OK	16:24
Air Blank	0.000	16:24
Control Test	0.077	16:24
Air Blank	0.000	16:25
Subject Sample # 1	0.304	16:25
Breath Volume	2.070	16:25
Air Blank	0.000	16:26
Air Blank	0.000	16:28
Subject Sample # 2	0.292	16:28
Breath Volume	1.511	16:28
Air Blank	0.000	16:29
Control Test	0.077	16:30
Air Blank	0.000	16:30
Diagnostics Check	OK	16:30

Data Download Date/Time: 07/13/2020 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: ROCKMAN

DATE/TIME OF TEST: 06/24/2020 02:07:20

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.078	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.215	02:02
Breath Volume	1.371	02:02
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.207	02:05
Breath Volume	1.597	02:05
Air Blank	0.000	02:06
Control Test	0.077	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 07/13/2020 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: SEARLS

DATE/TIME OF TEST: 06/27/2020 03:21:07

Diagnostics Check		OK	03:05	
Air Blank		0.000	03:06	
Control Test		0.078	03:06	
Air Blank		0.000	03:07	
Subject Sample	# 1	0.165	03:08	
Breath Volume		1.703	03:08	
Air Blank		0.000	03:08	
Air Blank		0.000	03:10	
Subject Sample	# 2	0.120	03:13	Volume Not Met
Breath Volume		0.433	03:13	
Air Blank		0.000	03:14	
Air Blank		0.000	03:16	
Subject Sample	# 3	0.146	03:19	Volume Not Met
Breath Volume		0.785	03:19	
Air Blank		0.000	03:20	
Control Test		0.076	03:20	
Air Blank		0.000	03:21	
Diagnostics Check		OK	03:21	

Data Download Date/Time: 07/13/2020 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: SEARLS

DATE/TIME OF TEST: 06/27/2020 03:32:27

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:22	
Control Test	0.077	03:23	
Air Blank	0.000	03:23	
Subject Sample	# 1 0.119	03:26	Volume Not Met
Breath Volume	0.085	03:26	
Air Blank	0.000	03:27	
Air Blank	0.000	03:29	
Subject Sample	# 2	03:30	Subject Test Refused
Breath Volume	0.281	03:30	
Air Blank	0.000	03:31	
Control Test	0.074	03:31	Control Outside Tolerance
Air Blank	0.000	03:32	

Data Download Date/Time: 07/13/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: PARSLEY

DATE/TIME OF TEST: 06/28/2020 00:21:40

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.077	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.200	00:16
Breath Volume	4.957	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.185	00:19
Breath Volume	3.023	00:19
Air Blank	0.000	00:20
Control Test	0.075	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 07/13/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: SEREBRYANSKIY

DATE/TIME OF TEST: 06/28/2020 00:40:57

Diagnostics Check	OK	00:34
Air Blank	0.000	00:35
Control Test	0.077	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.048	00:36
Breath Volume	1.171	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:38
Subject Sample # 2	0.048	00:39
Breath Volume	1.937	00:39
Air Blank	0.000	00:40
Control Test	0.078	00:40
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 07/13/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 06/28/2020 20:34:39

Diagnostics Check	OK	20:28
Air Blank	0.000	20:28
Control Test	0.077	20:28
Air Blank	0.000	20:29
Subject Sample # 1	0.236	20:29
Breath Volume	1.808	20:29
Air Blank	0.000	20:30
Air Blank	0.000	20:32
Subject Sample # 2	0.236	20:32
Breath Volume	1.976	20:32
Air Blank	0.000	20:33
Control Test	0.076	20:34
Air Blank	0.000	20:34
Diagnostics Check	OK	20:34

Data Download Date/Time: 07/13/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BLANTON

DATE/TIME OF TEST: 06/18/2020 02:58:34

Diagnostics Check		OK	02:52
Air Blank		0.000	02:53
Control Test		0.079	02:53
Air Blank		0.000	02:53
Subject Sample	# 1	0.133	02:54
Breath Volume		1.171	02:54
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.133	02:56
Breath Volume		1.210	02:56
Air Blank		0.000	02:57
Control Test		0.079	02:57
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 06/23/2020 01:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRASER, THOMAS, J
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: DESANTIS, ADAM
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175970
DRY GAS EXPIRY DATE: 08/02/2022
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: JOSWIAK

DATE/TIME OF TEST: 06/12/2020 23:23:44

Diagnostics Check	OK	23:13	
Air Blank	0.000	23:14	
Control Test	0.077	23:14	
Air Blank	0.000	23:15	
Subject Sample	# 1 0.154	23:18	Volume Not Met
Breath Volume	0.671	23:18	
Air Blank	0.000	23:19	
Air Blank	0.000	23:20	
Subject Sample	# 2	23:23	Sequence Aborted
Breath Volume	0.566	23:23	
Air Blank	0.000	23:23	

Data Download Date/Time: 07/15/2020 14:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: HURT,ALEXANDER
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175970
DRY GAS EXPIRY DATE: 08/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: LAZZARA

DATE/TIME OF TEST: 06/16/2020 03:36:35

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:29	
Control Test	0.076	03:30	
Air Blank	0.000	03:30	
Subject Sample	# 1 0.258	03:32	
Breath Volume	4.582	03:32	
Air Blank	0.000	03:32	
Air Blank	0.000	03:34	
Subject Sample	# 2 0.247	03:35	
Breath Volume	4.968	03:35	
Air Blank	0.000	03:35	
Control Test	0.074	03:35	Control Outside Tolerance
Air Blank	0.000	03:36	

Data Download Date/Time: 07/15/2020 14:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: HURT,ALEXANDER
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175970
DRY GAS EXPIRY DATE: 08/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: LAZZARO

DATE/TIME OF TEST: 06/16/2020 04:06:47

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.077	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.234	04:02
Breath Volume		4.980	04:02
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.233	04:05
Breath Volume		5.093	04:05
Air Blank		0.000	04:05
Control Test		0.075	04:06
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 07/15/2020 14:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: VANHORN,MICHAEL
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175970
DRY GAS EXPIRY DATE: 08/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: JAMESON

DATE/TIME OF TEST: 06/22/2020 05:12:49

Diagnostics Check	OK	05:09	
Air Blank	0.000	05:10	
Control Test	0.078	05:10	
Air Blank	0.000	05:11	
Subject Sample	# 1	05:11	Subject Test Refused
Breath Volume	0.000	05:11	
Air Blank	0.000	05:11	
Control Test	0.078	05:12	
Air Blank	0.000	05:12	
Diagnostics Check	OK	05:12	

Data Download Date/Time: 07/15/2020 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,RYAN,A
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MILLER,RYAN
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: BUI
ARREST TIME: 16:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: MALECKI

DATE/TIME OF TEST: 06/20/2020 17:35:30

Diagnostics Check	OK	17:29
Air Blank	0.000	17:29
Control Test	0.078	17:30
Air Blank	0.000	17:30
Subject Sample # 1	0.100	17:31
Breath Volume	2.910	17:31
Air Blank	0.000	17:31
Air Blank	0.000	17:33
Subject Sample # 2	0.103	17:33
Breath Volume	2.968	17:33
Air Blank	0.000	17:34
Control Test	0.078	17:34
Air Blank	0.000	17:35
Diagnostics Check	OK	17:35

Data Download Date/Time: 07/15/2020 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: CASIANO RAMIREZ

DATE/TIME OF TEST: 06/26/2020 02:23:41

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.079	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.170	02:19
Breath Volume	2.507	02:19
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.167	02:22
Breath Volume	1.578	02:22
Air Blank	0.000	02:22
Control Test	0.078	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 07/15/2020 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KHUNE,RONALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: STOKES

DATE/TIME OF TEST: 06/05/2020 22:39:40

Diagnostics Check	OK	22:33
Air Blank	0.000	22:33
Control Test	0.079	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.051	22:35
Breath Volume	2.585	22:35
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.047	22:38
Breath Volume	2.082	22:38
Air Blank	0.000	22:38
Control Test	0.079	22:39
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 07/02/2020 12:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WAGNER,BRETT
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 06/26/2020 04:59:35

Diagnostics Check	OK	04:52
Air Blank	0.000	04:53
Control Test	0.079	04:53
Air Blank	0.000	04:53
Subject Sample # 1	0.127	04:54
Breath Volume	1.281	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:56
Subject Sample # 2	0.131	04:57
Breath Volume	1.859	04:57
Air Blank	0.000	04:58
Control Test	0.079	04:58
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 07/02/2020 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: KURTIS

DATE/TIME OF TEST: 06/10/2020 05:36:45

Diagnostics Check	OK	05:30
Air Blank	0.000	05:31
Control Test	0.078	05:31
Air Blank	0.000	05:32
Subject Sample # 1	0.000	05:32
Breath Volume	2.359	05:32
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.000	05:35
Breath Volume	1.707	05:35
Air Blank	0.000	05:35
Control Test	0.076	05:36
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 06/26/2020 18:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 06/28/2020 00:27:43

Diagnostics Check		OK	00:21
Air Blank		0.000	00:21
Control Test		0.076	00:22
Air Blank		0.000	00:22
Subject Sample	# 1	0.201	00:23
Breath Volume		2.484	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:25
Subject Sample	# 2	0.185	00:26
Breath Volume		1.269	00:26
Air Blank		0.000	00:26
Control Test		0.076	00:27
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 07/07/2020 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WARD,GLENN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: FALK

DATE/TIME OF TEST: 06/28/2020 22:17:38

Diagnostics Check		OK	22:11
Air Blank		0.000	22:11
Control Test		0.077	22:12
Air Blank		0.000	22:12
Subject Sample	# 1	0.294	22:13
Breath Volume		1.738	22:13
Air Blank		0.000	22:14
Air Blank		0.000	22:15
Subject Sample	# 2	0.285	22:16
Breath Volume		1.714	22:16
Air Blank		0.000	22:16
Control Test		0.077	22:17
Air Blank		0.000	22:17
Diagnostics Check		OK	22:17

Data Download Date/Time: 07/07/2020 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: KELLY,CHARLTON
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: BELINKY

DATE/TIME OF TEST: 06/01/2020 20:23:15

Diagnostics Check	OK	20:16
Air Blank	0.000	20:16
Control Test	0.081	20:17
Air Blank	0.000	20:17
Subject Sample # 1	0.147	20:18
Breath Volume	2.281	20:18
Air Blank	0.000	20:19
Air Blank	0.000	20:21
Subject Sample # 2	0.135	20:21
Breath Volume	1.371	20:21
Air Blank	0.000	20:22
Control Test	0.079	20:22
Air Blank	0.000	20:23
Diagnostics Check	OK	20:23

Data Download Date/Time: 06/26/2020 17:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: ALDERMAN

DATE/TIME OF TEST: 06/02/2020 22:53:37

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.081	22:48
Air Blank	0.000	22:48
Subject Sample # 1	0.000	22:49
Breath Volume	2.234	22:49
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.000	22:52
Breath Volume	2.355	22:52
Air Blank	0.000	22:52
Control Test	0.080	22:53
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 06/26/2020 17:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: JOHNSON,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: HOO-HING

DATE/TIME OF TEST: 06/04/2020 23:02:14

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.079	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.133	22:57
Breath Volume	1.175	22:57
Air Blank	0.000	22:58
Air Blank	0.000	22:59
Subject Sample # 2	0.139	23:00
Breath Volume	2.132	23:00
Air Blank	0.000	23:01
Control Test	0.078	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 06/26/2020 17:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MARROQUIN,RICHARD
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: BUI
ARREST TIME: 14:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: EDMONSON

DATE/TIME OF TEST: 06/05/2020 15:57:57

Diagnostics Check	OK	15:51
Air Blank	0.000	15:51
Control Test	0.080	15:51
Air Blank	0.000	15:52
Subject Sample # 1	0.145	15:53
Breath Volume	3.296	15:53
Air Blank	0.000	15:53
Air Blank	0.000	15:55
Subject Sample # 2	0.145	15:56
Breath Volume	2.765	15:56
Air Blank	0.000	15:57
Control Test	0.079	15:57
Air Blank	0.000	15:57
Diagnostics Check	OK	15:57

Data Download Date/Time: 06/26/2020 17:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: MCCRAY 3

DATE/TIME OF TEST: 06/06/2020 02:04:07

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.081	01:58
Air Blank		0.000	01:58
Subject Sample	# 1	0.158	01:59
Breath Volume		2.449	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.148	02:02
Breath Volume		1.410	02:02
Air Blank		0.000	02:03
Control Test		0.078	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 06/26/2020 17:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: VALENCIA

DATE/TIME OF TEST: 06/07/2020 03:14:00

Diagnostics Check		OK	03:06
Air Blank		0.000	03:07
Control Test		0.079	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.104	03:09
Breath Volume		2.109	03:09
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.106	03:12
Breath Volume		2.359	03:12
Air Blank		0.000	03:13
Control Test		0.078	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:14

Data Download Date/Time: 06/26/2020 17:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: HARRELL-LEYSATH

DATE/TIME OF TEST: 06/06/2020 03:14:08

Diagnostics Check		OK	03:07
Air Blank		0.000	03:08
Control Test		0.079	03:08
Air Blank		0.000	03:08
Subject Sample	# 1	0.180	03:09
Breath Volume		3.433	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.181	03:12
Breath Volume		3.320	03:12
Air Blank		0.000	03:13
Control Test		0.079	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 06/07/2020 21:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUGHEY, MICHAEL, D
BREATH TEST OPERATOR AGENCY: OTHER STATE
ARRESTING OFFICER: NA, NA
ARRESTING OFFICER AGENCY: OTHER STATE
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: SYS CHECK
ARREST TIME: 08:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: LICENSE

DATE/TIME OF TEST: 06/18/2020 09:33:59

Diagnostics Check	OK	09:28
Air Blank	0.000	09:28
Control Test	0.079	09:29
Air Blank	0.000	09:29
Subject Sample # 1	0.000	09:29
Breath Volume	1.246	09:29
Air Blank	0.000	09:30
Air Blank	0.000	09:32
Subject Sample # 2	0.000	09:32
Breath Volume	1.722	09:32
Air Blank	0.000	09:33
Control Test	0.080	09:33
Air Blank	0.000	09:33
Diagnostics Check	OK	09:33

Data Download Date/Time: 06/18/2020 12:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: BODEN JR

DATE/TIME OF TEST: 06/02/2020 02:21:05

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.201	02:15
Breath Volume	2.378	02:15
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.183	02:19
Breath Volume	1.734	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:21

Data Download Date/Time: 06/12/2020 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: CIANCI

DATE/TIME OF TEST: 06/03/2020 02:21:49

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.079	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.200	02:17
Breath Volume		1.203	02:17
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.203	02:20
Breath Volume		1.558	02:20
Air Blank		0.000	02:20
Control Test		0.079	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 06/12/2020 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: FARRINGTON

DATE/TIME OF TEST: 06/05/2020 00:20:58

Diagnostics Check	OK	00:13
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.000	00:16
Breath Volume	2.175	00:16
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.000	00:19
Breath Volume	1.917	00:19
Air Blank	0.000	00:20
Control Test	0.080	00:20
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 06/12/2020 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: REID,FRANCINE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 15:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 06/05/2020 18:31:24

Diagnostics Check	OK	18:22
Air Blank	0.000	18:23
Control Test	0.080	18:23
Air Blank	0.000	18:23
Subject Sample # 1	0.159	18:24
Breath Volume	2.179	18:24
Air Blank	0.000	18:25
Air Blank	0.000	18:27
Subject Sample # 2	0.141	18:29
Breath Volume	1.398	18:29
Air Blank	0.000	18:30
Control Test	0.081	18:30
Air Blank	0.000	18:31
Diagnostics Check	OK	18:31

Data Download Date/Time: 06/12/2020 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: CHACON

DATE/TIME OF TEST: 06/06/2020 04:26:35

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:15	
Control Test	0.080	04:15	
Air Blank	0.000	04:16	
Subject Sample	# 1	04:19	No Sample Provided
Breath Volume	0.000	04:19	
Air Blank	0.000	04:20	
Air Blank	0.000	04:22	
Subject Sample	# 2	04:25	No Sample Provided
Breath Volume	0.000	04:25	
Air Blank	0.000	04:25	
Control Test	0.080	04:25	
Air Blank	0.000	04:26	
Diagnostics Check	OK	04:26	

Data Download Date/Time: 06/12/2020 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PRESUME,JEHUD
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/10/2020 23:25:24

Diagnostics Check	OK	23:13	
Air Blank	0.000	23:14	
Control Test	0.081	23:14	
Air Blank	0.000	23:15	
Subject Sample	# 1	23:18	No Sample Provided
Breath Volume	0.000	23:18	
Air Blank	0.000	23:18	
Air Blank	0.000	23:20	
Subject Sample	# 2	23:24	No Sample Provided
Breath Volume	0.000	23:24	
Air Blank	0.000	23:24	
Control Test	0.081	23:24	
Air Blank	0.000	23:25	
Diagnostics Check	OK	23:25	

Data Download Date/Time: 06/12/2020 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIRMANS,ABNER,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: SIRMANS,ABNER
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: HULSEY

DATE/TIME OF TEST: 06/15/2020 04:51:59

Diagnostics Check		OK	04:44
Air Blank		0.000	04:45
Control Test		0.081	04:45
Air Blank		0.000	04:46
Subject Sample	# 1	0.308	04:47
Breath Volume		2.300	04:47
Air Blank		0.000	04:48
Air Blank		0.000	04:49
Subject Sample	# 2	0.318	04:50
Breath Volume		1.429	04:50
Air Blank		0.000	04:51
Control Test		0.077	04:51
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 07/12/2020 03:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CHAPIN,NATHAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ANICETO DA SILVA

DATE/TIME OF TEST: 06/20/2020 03:58:53

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.080	03:52
Air Blank		0.000	03:53
Subject Sample	# 1	0.186	03:54
Breath Volume		2.093	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:56
Subject Sample	# 2	0.186	03:57
Breath Volume		1.281	03:57
Air Blank		0.000	03:57
Control Test		0.079	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 07/12/2020 03:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: SWENSON,BRIAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: KOLLER

DATE/TIME OF TEST: 06/27/2020 06:08:12

Diagnostics Check	OK	06:01
Air Blank	0.000	06:02
Control Test	0.080	06:02
Air Blank	0.000	06:02
Subject Sample # 1	0.179	06:03
Breath Volume	1.199	06:03
Air Blank	0.000	06:04
Air Blank	0.000	06:05
Subject Sample # 2	0.177	06:06
Breath Volume	1.843	06:06
Air Blank	0.000	06:07
Control Test	0.080	06:07
Air Blank	0.000	06:08
Diagnostics Check	OK	06:08

Data Download Date/Time: 07/12/2020 03:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOHLIN,MARK,T
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CHAPIN,NATHAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: STARLING

DATE/TIME OF TEST: 06/28/2020 01:43:42

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.080	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.192	01:39
Breath Volume	2.339	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.187	01:41
Breath Volume	2.132	01:41
Air Blank	0.000	01:42
Control Test	0.079	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 07/14/2020 21:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEATING,KAYLA,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/09/2020 23:48:38

Diagnostics Check		OK	23:39
Air Blank		0.000	23:40
Control Test		0.077	23:40
Air Blank		0.000	23:41
Subject Sample	# 1	0.219	23:42
Breath Volume		1.996	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.238	23:46
Breath Volume		2.652	23:46
Air Blank		0.000	23:47
Control Test		0.077	23:47
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 06/12/2020 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: LLESHI

DATE/TIME OF TEST: 06/07/2020 00:41:54

Diagnostics Check		OK	00:30	
Air Blank		0.000	00:30	
Control Test		0.079	00:30	
Air Blank		0.000	00:31	
Subject Sample	# 1	0.210	00:34	Volume Not Met
Breath Volume		0.054	00:34	
Air Blank		0.000	00:35	
Air Blank		0.000	00:37	
Subject Sample	# 2	0.209	00:40	Volume Not Met
Breath Volume		0.000	00:40	
Air Blank		0.000	00:40	
Control Test		0.079	00:41	
Air Blank		0.000	00:41	
Diagnostics Check		OK	00:41	

Data Download Date/Time: 06/12/2020 14:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 06/07/2020 00:55:22

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.079	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.138	00:50
Breath Volume	2.332	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:52
Subject Sample # 2	0.137	00:53
Breath Volume	1.925	00:53
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 06/12/2020 14:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOONE,MIKE
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 06/07/2020 03:58:41

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.079	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.201	03:52
Breath Volume	2.515	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.206	03:57
Breath Volume	1.484	03:57
Air Blank	0.000	03:57
Control Test	0.077	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 06/12/2020 14:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: AUGUSTIN JR

DATE/TIME OF TEST: 06/07/2020 04:36:24

Diagnostics Check		OK	04:29
Air Blank		0.000	04:30
Control Test		0.077	04:30
Air Blank		0.000	04:31
Subject Sample	# 1	0.308	04:31
Breath Volume		2.179	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.303	04:34
Breath Volume		1.164	04:34
Air Blank		0.000	04:35
Control Test		0.078	04:35
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 06/12/2020 14:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SGT. BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: MELKONIAN

DATE/TIME OF TEST: 06/13/2020 22:19:25

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.080	22:14
Air Blank	0.000	22:14
Subject Sample # 1	0.140	22:14
Breath Volume	1.640	22:14
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.155	22:17
Breath Volume	2.261	22:17
Air Blank	0.000	22:18
Control Test	0.080	22:18
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 07/02/2020 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,DANNY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: MALCOLM

DATE/TIME OF TEST: 06/14/2020 04:56:00

Diagnostics Check	OK	04:49
Air Blank	0.000	04:49
Control Test	0.080	04:50
Air Blank	0.000	04:50
Subject Sample # 1	0.077	04:51
Breath Volume	1.457	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:53
Subject Sample # 2	0.078	04:54
Breath Volume	1.628	04:54
Air Blank	0.000	04:55
Control Test	0.079	04:55
Air Blank	0.000	04:55
Diagnostics Check	OK	04:56

Data Download Date/Time: 07/02/2020 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SABAN,SHAHEED
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/14/2020 05:38:23

Diagnostics Check		OK	05:32
Air Blank		0.000	05:32
Control Test		0.080	05:32
Air Blank		0.000	05:33
Subject Sample	# 1	0.203	05:33
Breath Volume		1.484	05:33
Air Blank		0.000	05:34
Air Blank		0.000	05:36
Subject Sample	# 2	0.202	05:36
Breath Volume		1.558	05:36
Air Blank		0.000	05:37
Control Test		0.079	05:37
Air Blank		0.000	05:38
Diagnostics Check		OK	05:38

Data Download Date/Time: 07/02/2020 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ULLOA,ADRIAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: FLANAGAN

DATE/TIME OF TEST: 06/14/2020 06:18:53

Diagnostics Check		OK	06:12
Air Blank		0.000	06:12
Control Test		0.079	06:13
Air Blank		0.000	06:13
Subject Sample	# 1	0.152	06:14
Breath Volume		2.078	06:14
Air Blank		0.000	06:15
Air Blank		0.000	06:17
Subject Sample	# 2	0.147	06:17
Breath Volume		1.101	06:17
Air Blank		0.000	06:17
Control Test		0.079	06:18
Air Blank		0.000	06:18
Diagnostics Check		OK	06:18

Data Download Date/Time: 07/02/2020 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MCGUIRE,THOMAS
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 15:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: PROKHORV

DATE/TIME OF TEST: 06/21/2020 17:56:19

Diagnostics Check	OK	17:48
Air Blank	0.000	17:49
Control Test	0.079	17:49
Air Blank	0.000	17:49
Subject Sample # 1	0.282	17:50
Breath Volume	1.894	17:50
Air Blank	0.000	17:51
Air Blank	0.000	17:53
Subject Sample # 2	0.288	17:54
Breath Volume	1.640	17:54
Air Blank	0.000	17:55
Control Test	0.079	17:55
Air Blank	0.000	17:56
Diagnostics Check	OK	17:56

Data Download Date/Time: 07/02/2020 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: CHI

DATE/TIME OF TEST: 06/22/2020 04:03:18

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.000	03:59
Breath Volume	1.816	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.000	04:01
Breath Volume	1.343	04:01
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 07/02/2020 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CHINCHILLA,SERGIO
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: RANDALL

DATE/TIME OF TEST: 06/24/2020 00:39:17

Diagnostics Check		OK	00:32
Air Blank		0.000	00:33
Control Test		0.080	00:33
Air Blank		0.000	00:33
Subject Sample	# 1	0.251	00:34
Breath Volume		2.363	00:34
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.245	00:37
Breath Volume		1.847	00:37
Air Blank		0.000	00:38
Control Test		0.078	00:38
Air Blank		0.000	00:39
Diagnostics Check		OK	00:39

Data Download Date/Time: 07/02/2020 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLIGAN,DANIEL
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: SHEROOD

DATE/TIME OF TEST: 06/26/2020 00:32:36

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.000	00:28
Breath Volume	1.480	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.000	00:31
Breath Volume	1.429	00:31
Air Blank	0.000	00:31
Control Test	0.078	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 07/02/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,DANNY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: SPERBER

DATE/TIME OF TEST: 06/26/2020 01:19:18

Diagnostics Check		OK	01:12
Air Blank		0.000	01:13
Control Test		0.080	01:13
Air Blank		0.000	01:13
Subject Sample	# 1	0.102	01:14
Breath Volume		2.695	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.102	01:17
Breath Volume		2.617	01:17
Air Blank		0.000	01:18
Control Test		0.080	01:18
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 07/02/2020 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTNEY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/27/2020 05:32:56

Diagnostics Check		OK	05:17	
Air Blank		0.000	05:18	
Control Test		0.080	05:18	
Air Blank		0.000	05:18	
Subject Sample	# 1	0.138	05:22	Volume Not Met
Breath Volume		0.406	05:22	
Air Blank		0.000	05:22	
Air Blank		0.000	05:24	
Subject Sample	# 2	0.166	05:25	
Breath Volume		1.929	05:25	
Air Blank		0.000	05:26	
Air Blank		0.000	05:28	
Subject Sample	# 3		05:31	No Sample Provided
Breath Volume		0.000	05:31	
Air Blank		0.000	05:32	
Control Test		0.079	05:32	
Air Blank		0.000	05:32	
Diagnostics Check		OK	05:32	

Data Download Date/Time: 07/02/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTNEY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/27/2020 05:45:52

Diagnostics Check	OK	05:36	
Air Blank	0.000	05:37	
Control Test	0.079	05:37	
Air Blank	0.000	05:38	
Subject Sample	# 1	05:41	No Sample Provided
Breath Volume	0.000	05:41	
Air Blank	0.000	05:41	
Air Blank	0.000	05:43	
Subject Sample	# 2	05:44	Subject Test Refused
Breath Volume	0.000	05:44	
Air Blank	0.000	05:44	
Control Test	0.080	05:45	
Air Blank	0.000	05:45	
Diagnostics Check	OK	05:45	

Data Download Date/Time: 07/02/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HERNANDEZ,ERNESTO
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: PEREZ VEGA

DATE/TIME OF TEST: 06/27/2020 23:05:20

Diagnostics Check	OK	22:57
Air Blank	0.000	22:58
Control Test	0.079	22:58
Air Blank	0.000	22:59
Subject Sample # 1	0.152	23:00
Breath Volume	3.566	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.144	23:03
Breath Volume	3.707	23:03
Air Blank	0.000	23:04
Control Test	0.080	23:04
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 07/02/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: SAMSON

DATE/TIME OF TEST: 06/27/2020 23:54:28

Diagnostics Check	OK	23:48
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:49
Subject Sample # 1	0.158	23:50
Breath Volume	3.886	23:50
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.150	23:52
Breath Volume	3.660	23:52
Air Blank	0.000	23:53
Control Test	0.079	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 07/02/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: IACONO

DATE/TIME OF TEST: 06/28/2020 03:55:54

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:42	
Control Test		0.080	03:42	
Air Blank		0.000	03:43	
Subject Sample	# 1		03:46	No Sample Provided
Breath Volume		0.000	03:46	
Air Blank		0.000	03:47	
Air Blank		0.000	03:49	
Subject Sample	# 2	0.185	03:51	No .020 agreement
Breath Volume		1.312	03:51	
Air Blank		0.000	03:51	
Air Blank		0.000	03:53	
Subject Sample	# 3	0.212	03:54	No .020 agreement
Breath Volume		1.750	03:54	
Air Blank		0.000	03:55	
Control Test		0.079	03:55	
Air Blank		0.000	03:55	
Diagnostics Check		OK	03:55	

Data Download Date/Time: 07/02/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: IACONO

DATE/TIME OF TEST: 06/28/2020 04:07:01

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.079	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.206	04:02
Breath Volume		2.000	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:04
Subject Sample	# 2	0.216	04:05
Breath Volume		2.121	04:05
Air Blank		0.000	04:06
Control Test		0.079	04:06
Air Blank		0.000	04:06
Diagnostics Check		OK	04:07

Data Download Date/Time: 07/02/2020 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 06/11/2020
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: GROENEN

DATE/TIME OF TEST: 06/18/2020 23:55:09

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.077	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.157	23:50
Breath Volume	1.847	23:50
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.153	23:53
Breath Volume	2.097	23:53
Air Blank	0.000	23:54
Control Test	0.077	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 07/02/2020 12:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG, BRITTANY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG, B
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: PRATT

DATE/TIME OF TEST: 06/06/2020 04:39:08

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.077	04:29
Air Blank	0.000	04:30
Subject Sample # 1	0.221	04:30
Breath Volume	1.117	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.194	04:34
Breath Volume	1.707	04:34
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 3	0.188	04:37
Breath Volume	2.203	04:37
Air Blank	0.000	04:38
Control Test	0.076	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 06/12/2020 20:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,BRITTANY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0091908A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: BARRUETA

DATE/TIME OF TEST: 06/11/2020 23:48:01

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:34	
Control Test	0.077	23:34	
Air Blank	0.000	23:34	
Subject Sample	# 1	23:38	No Sample Provided
Breath Volume	0.000	23:38	
Air Blank	0.000	23:38	
Air Blank	0.000	23:40	
Subject Sample	# 2	0.349	23:40
Breath Volume	1.175	23:40	
Air Blank	0.000	23:41	
Air Blank	0.000	23:43	
Subject Sample	# 3	23:46	No Sample Provided
Breath Volume	0.000	23:46	
Air Blank	0.000	23:47	
Control Test	0.077	23:47	
Air Blank	0.000	23:47	
Diagnostics Check	OK	23:48	

Data Download Date/Time: 06/12/2020 20:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,BRITTANY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0091908A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: DE OLIVEIRA

DATE/TIME OF TEST: 06/13/2020 19:48:58

Diagnostics Check	OK	19:41
Air Blank	0.000	19:42
Control Test	0.078	19:42
Air Blank	0.000	19:43
Subject Sample # 1	0.203	19:43
Breath Volume	1.378	19:43
Air Blank	0.000	19:44
Air Blank	0.000	19:46
Subject Sample # 2	0.206	19:47
Breath Volume	1.125	19:47
Air Blank	0.000	19:48
Control Test	0.078	19:48
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 07/02/2020 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 06/09/2020 22:25:49

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:21	
Control Test	0.078	22:22	
Air Blank	0.000	22:22	
Subject Sample	# 1	22:24	Interferent Detect
Breath Volume	1.453	22:24	
Air Blank	0.000	22:24	
Control Test	0.078	22:25	
Air Blank	0.000	22:25	
Diagnostics Check	OK	22:25	

Data Download Date/Time: 06/12/2020 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 06/09/2020 22:54:11

Diagnostics Check	OK	22:47
Air Blank	0.000	22:48
Control Test	0.078	22:48
Air Blank	0.000	22:49
Subject Sample # 1	0.000	22:49
Breath Volume	1.609	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.000	22:52
Breath Volume	1.734	22:52
Air Blank	0.000	22:53
Control Test	0.079	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 06/12/2020 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:21

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 06/14/2020 17:48:32

Diagnostics Check	OK	17:42
Air Blank	0.000	17:42
Control Test	0.079	17:42
Air Blank	0.000	17:43
Subject Sample # 1	0.164	17:43
Breath Volume	4.812	17:43
Air Blank	0.000	17:44
Air Blank	0.000	17:46
Subject Sample # 2	0.153	17:46
Breath Volume	2.703	17:46
Air Blank	0.000	17:47
Control Test	0.079	17:47
Air Blank	0.000	17:48
Diagnostics Check	OK	17:48

Data Download Date/Time: 07/02/2020 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/21/2020 22:57:42

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:46	
Control Test	0.077	22:46	
Air Blank	0.000	22:47	
Subject Sample	# 1	22:50	No Sample Provided
Breath Volume	0.000	22:50	
Air Blank	0.000	22:50	
Air Blank	0.000	22:52	
Subject Sample	# 2 0.218	22:55	Volume Not Met
Breath Volume	0.121	22:55	
Air Blank	0.000	22:56	
Control Test	0.079	22:57	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

Data Download Date/Time: 07/02/2020 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 06/23/2020 20:25:29

Diagnostics Check	OK	20:18
Air Blank	0.000	20:19
Control Test	0.078	20:19
Air Blank	0.000	20:20
Subject Sample # 1	0.152	20:20
Breath Volume	2.460	20:20
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample # 2	0.153	20:23
Breath Volume	2.785	20:23
Air Blank	0.000	20:24
Control Test	0.078	20:24
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 07/02/2020 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,P,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PEREZ,M
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: STIPEK

DATE/TIME OF TEST: 06/04/2020 01:08:21

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.078	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.130	01:03
Breath Volume	3.351	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.131	01:06
Breath Volume	4.371	01:06
Air Blank	0.000	01:07
Control Test	0.078	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 07/08/2020 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAFFEE,C,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CORWIN,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: KOSINSKI

DATE/TIME OF TEST: 06/08/2020 23:01:57

Diagnostics Check	OK	22:47	
Air Blank	0.000	22:47	
Control Test	0.079	22:48	
Air Blank	0.000	22:48	
Subject Sample	# 1 0.137	22:51	
Breath Volume	1.500	22:51	
Air Blank	0.000	22:52	
Air Blank	0.000	22:53	
Subject Sample	# 2 0.126	22:56	Volume Not Met
Breath Volume	0.097	22:56	
Air Blank	0.000	22:57	
Air Blank	0.000	22:59	
Subject Sample	# 3 0.140	23:00	
Breath Volume	2.320	23:00	
Air Blank	0.000	23:01	
Control Test	0.074	23:01	Control Outside Tolerance
Air Blank	0.000	23:01	

Data Download Date/Time: 07/08/2020 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRATT,N,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PRATT,N
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BLAKEMAN

DATE/TIME OF TEST: 06/11/2020 02:44:20

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.078	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.280	02:39
Breath Volume	2.917	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.283	02:42
Breath Volume	2.089	02:42
Air Blank	0.000	02:43
Control Test	0.078	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 07/08/2020 15:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRATT,N,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PRATT,N
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: DOWNEY

DATE/TIME OF TEST: 06/12/2020 04:05:33

Diagnostics Check		OK	03:58
Air Blank		0.000	03:59
Control Test		0.078	03:59
Air Blank		0.000	04:00
Subject Sample	# 1	0.214	04:00
Breath Volume		2.343	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:03
Subject Sample	# 2	0.204	04:03
Breath Volume		2.175	04:03
Air Blank		0.000	04:04
Control Test		0.079	04:04
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 07/08/2020 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,SHANNON,N
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CORWIN,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: ORDONEZ

DATE/TIME OF TEST: 06/12/2020 06:17:55

Diagnostics Check	OK	06:10
Air Blank	0.000	06:11
Control Test	0.078	06:11
Air Blank	0.000	06:12
Subject Sample # 1	0.171	06:13
Breath Volume	1.531	06:13
Air Blank	0.000	06:13
Air Blank	0.000	06:15
Subject Sample # 2	0.184	06:16
Breath Volume	1.121	06:16
Air Blank	0.000	06:16
Control Test	0.078	06:17
Air Blank	0.000	06:17
Diagnostics Check	OK	06:17

Data Download Date/Time: 07/08/2020 15:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 06/13/2020 03:37:06

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.078	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.123	03:32
Breath Volume	2.781	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.121	03:35
Breath Volume	2.406	03:35
Air Blank	0.000	03:36
Control Test	0.078	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 07/08/2020 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBBONS,S,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: REILLY,S
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: ORTEGA

DATE/TIME OF TEST: 06/13/2020 23:30:11

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:25	
Control Test	0.078	23:25	
Air Blank	0.000	23:26	
Subject Sample	# 1	23:29	Sequence Aborted
Breath Volume	0.097	23:29	
Air Blank	0.000	23:30	

Data Download Date/Time: 07/08/2020 15:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARR,L,J
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: TERUZ,G
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:41

SUBJECT'S LAST NAME: RIESEN

DATE/TIME OF TEST: 06/14/2020 08:12:19

Diagnostics Check	OK	08:05
Air Blank	0.000	08:05
Control Test	0.078	08:06
Air Blank	0.000	08:06
Subject Sample # 1	0.000	08:07
Breath Volume	3.253	08:07
Air Blank	0.000	08:08
Air Blank	0.000	08:10
Subject Sample # 2	0.000	08:10
Breath Volume	2.656	08:10
Air Blank	0.000	08:11
Control Test	0.078	08:11
Air Blank	0.000	08:12
Diagnostics Check	OK	08:12

Data Download Date/Time: 07/08/2020 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,S,C
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GROENDYK,R
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 06/17/2020 01:28:17

Diagnostics Check		OK	01:13	
Air Blank		0.000	01:13	
Control Test		0.079	01:13	
Air Blank		0.000	01:14	
Subject Sample	# 1	0.121	01:17	Slope Not Level
Breath Volume		1.171	01:17	
Air Blank		0.000	01:18	
Air Blank		0.000	01:20	
Subject Sample	# 2	0.133	01:20	
Breath Volume		1.226	01:20	
Air Blank		0.000	01:21	
Air Blank		0.000	01:23	
Subject Sample	# 3	0.120	01:26	Slope Not Level
Breath Volume		1.234	01:26	
Air Blank		0.000	01:27	
Control Test		0.078	01:27	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 07/08/2020 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARVIN,ANGELINA,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ZYNDA,T
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: PEREZ MORALES

DATE/TIME OF TEST: 06/18/2020 01:41:52

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.079	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.219	01:37
Breath Volume		2.691	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:39
Subject Sample	# 2	0.222	01:40
Breath Volume		2.769	01:40
Air Blank		0.000	01:40
Control Test		0.078	01:41
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 07/08/2020 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS, MATTHEW, L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PEREZ, M
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: OQUENDO CEDENO

DATE/TIME OF TEST: 06/19/2020 02:02:36

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.078	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.185	01:57
Breath Volume	2.878	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.185	02:01
Breath Volume	3.250	02:01
Air Blank	0.000	02:01
Control Test	0.078	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 07/08/2020 15:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUONPANE,V,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: VAUGHAN,B
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: ALLISON

DATE/TIME OF TEST: 06/19/2020 02:14:54

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:09	
Control Test	0.078	02:10	
Air Blank	0.000	02:10	
Subject Sample	# 1	02:13	Subject Test Refused
Breath Volume	0.160	02:13	
Air Blank	0.000	02:13	
Control Test	0.079	02:14	
Air Blank	0.000	02:14	
Diagnostics Check	OK	02:14	

Data Download Date/Time: 07/08/2020 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRICE,MARCUS,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: TERUZ,G
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: GERONIMO REINOSO

DATE/TIME OF TEST: 06/19/2020 02:51:06

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.078	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.207	02:46
Breath Volume	2.660	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:48
Subject Sample # 2	0.214	02:49
Breath Volume	1.937	02:49
Air Blank	0.000	02:50
Control Test	0.078	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 07/08/2020 15:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,E,W
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KORPOLINSKI,M
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 06/23/2020 02:10:29

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.078	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.139	02:05
Breath Volume	1.308	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.137	02:08
Breath Volume	2.019	02:08
Air Blank	0.000	02:09
Control Test	0.077	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 07/08/2020 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,SEAN,R
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ELLIS,SEAN
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:43

SUBJECT'S LAST NAME: MITA

DATE/TIME OF TEST: 06/23/2020 09:17:52

Diagnostics Check	OK	09:09
Air Blank	0.000	09:10
Control Test	0.078	09:10
Air Blank	0.000	09:10
Subject Sample # 1	0.000	09:13
Breath Volume	1.515	09:13
Air Blank	0.000	09:13
Air Blank	0.000	09:15
Subject Sample # 2	0.000	09:16
Breath Volume	1.226	09:16
Air Blank	0.000	09:16
Control Test	0.077	09:17
Air Blank	0.000	09:17
Diagnostics Check	OK	09:17

Data Download Date/Time: 07/08/2020 15:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLMSTEAD,J,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GROSSKLAS,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: ZICHT

DATE/TIME OF TEST: 06/24/2020 00:32:20

Diagnostics Check		OK	00:19	
Air Blank		0.000	00:19	
Control Test		0.078	00:19	
Air Blank		0.000	00:20	
Subject Sample	# 1	0.210	00:23	Volume Not Met
Breath Volume		0.378	00:23	
Air Blank		0.000	00:24	
Air Blank		0.000	00:26	
Subject Sample	# 2	0.190	00:26	No .020 agreement
Breath Volume		1.585	00:26	
Air Blank		0.000	00:27	
Air Blank		0.000	00:29	
Subject Sample	# 3	0.214	00:30	No .020 agreement
Breath Volume		2.171	00:30	
Air Blank		0.000	00:31	
Control Test		0.077	00:31	
Air Blank		0.000	00:32	
Diagnostics Check		OK	00:32	

Data Download Date/Time: 07/08/2020 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLMSTEAD,J,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GROSSKLAS,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: ZICHT

DATE/TIME OF TEST: 06/24/2020 00:46:30

Diagnostics Check	OK	00:38
Air Blank	0.000	00:39
Control Test	0.077	00:39
Air Blank	0.000	00:40
Subject Sample # 1	0.191	00:41
Breath Volume	1.382	00:41
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.208	00:44
Breath Volume	2.699	00:44
Air Blank	0.000	00:45
Control Test	0.078	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 07/08/2020 15:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBBONS,S,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: LOPEZ,J
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: DALBY

DATE/TIME OF TEST: 06/24/2020 04:39:57

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.078	04:34
Air Blank	0.000	04:34
Subject Sample # 1	0.092	04:35
Breath Volume	2.500	04:35
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.093	04:38
Breath Volume	2.085	04:38
Air Blank	0.000	04:39
Control Test	0.078	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 07/08/2020 15:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,YVETTE,R
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: NEGRON,.
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:31

SUBJECT'S LAST NAME: FRAZEE

DATE/TIME OF TEST: 06/25/2020 13:59:48

Diagnostics Check	OK	13:53
Air Blank	0.000	13:53
Control Test	0.079	13:54
Air Blank	0.000	13:54
Subject Sample # 1	0.227	13:55
Breath Volume	1.507	13:55
Air Blank	0.000	13:55
Air Blank	0.000	13:57
Subject Sample # 2	0.217	13:58
Breath Volume	1.554	13:58
Air Blank	0.000	13:58
Control Test	0.078	13:59
Air Blank	0.000	13:59
Diagnostics Check	OK	13:59

Data Download Date/Time: 07/08/2020 15:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATSON,ANDRAE,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MCLAIN,K
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: MCCLELLAN

DATE/TIME OF TEST: 06/25/2020 20:28:04

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.079	20:22
Air Blank	0.000	20:22
Subject Sample # 1	0.000	20:23
Breath Volume	1.503	20:23
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.000	20:26
Breath Volume	1.355	20:26
Air Blank	0.000	20:27
Control Test	0.078	20:27
Air Blank	0.000	20:27
Diagnostics Check	OK	20:28

Data Download Date/Time: 07/08/2020 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERMAN,BRENT,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: POUNDS,R
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: MAZZIOTTI

DATE/TIME OF TEST: 06/28/2020 03:39:17

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.078	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.149	03:34
Breath Volume	1.699	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.143	03:37
Breath Volume	1.335	03:37
Air Blank	0.000	03:38
Control Test	0.078	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 07/08/2020 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: OSUNA

DATE/TIME OF TEST: 06/11/2020 20:41:00

Diagnostics Check	OK	20:34
Air Blank	0.000	20:35
Control Test	0.080	20:35
Air Blank	0.000	20:35
Subject Sample # 1	0.307	20:36
Breath Volume	2.167	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:38
Subject Sample # 2	0.295	20:39
Breath Volume	2.347	20:39
Air Blank	0.000	20:40
Control Test	0.080	20:40
Air Blank	0.000	20:40
Diagnostics Check	OK	20:41

Data Download Date/Time: 06/12/2020 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUNA,KRISTINA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 06/12/2020 02:47:41

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.157	02:43
Breath Volume	1.820	02:43
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.144	02:46
Breath Volume	1.527	02:46
Air Blank	0.000	02:46
Control Test	0.079	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 06/12/2020 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALTAMIRANDA,ANTONIO
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: AMADOR FLORES

DATE/TIME OF TEST: 06/13/2020 01:08:31

Diagnostics Check	OK	01:01
Air Blank	0.000	01:02
Control Test	0.079	01:02
Air Blank	0.000	01:03
Subject Sample # 1	0.135	01:03
Breath Volume	2.835	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.140	01:06
Breath Volume	1.714	01:06
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 07/02/2020 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: AKERS,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 10:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:10

SUBJECT'S LAST NAME: SALDANA BONILLA

DATE/TIME OF TEST: 06/14/2020 12:46:40

Diagnostics Check	OK	12:40
Air Blank	0.000	12:40
Control Test	0.079	12:41
Air Blank	0.000	12:41
Subject Sample # 1	0.037	12:42
Breath Volume	2.644	12:42
Air Blank	0.000	12:42
Air Blank	0.000	12:44
Subject Sample # 2	0.036	12:45
Breath Volume	4.968	12:45
Air Blank	0.000	12:45
Control Test	0.079	12:46
Air Blank	0.000	12:46
Diagnostics Check	OK	12:46

Data Download Date/Time: 07/02/2020 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HILSON,BERNARD
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 11:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: SAWYER

DATE/TIME OF TEST: 06/14/2020 13:27:14

Diagnostics Check	OK	13:20
Air Blank	0.000	13:21
Control Test	0.079	13:21
Air Blank	0.000	13:22
Subject Sample # 1	0.103	13:22
Breath Volume	2.425	13:22
Air Blank	0.000	13:23
Air Blank	0.000	13:25
Subject Sample # 2	0.096	13:25
Breath Volume	1.214	13:25
Air Blank	0.000	13:26
Control Test	0.080	13:26
Air Blank	0.000	13:27
Diagnostics Check	OK	13:27

Data Download Date/Time: 07/02/2020 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: EDWARDS,DERRICK
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: BURGAMY

DATE/TIME OF TEST: 06/14/2020 18:20:22

Diagnostics Check	OK	18:14
Air Blank	0.000	18:14
Control Test	0.080	18:14
Air Blank	0.000	18:15
Subject Sample # 1	0.195	18:15
Breath Volume	3.089	18:15
Air Blank	0.000	18:16
Air Blank	0.000	18:18
Subject Sample # 2	0.184	18:18
Breath Volume	2.425	18:18
Air Blank	0.000	18:19
Control Test	0.078	18:19
Air Blank	0.000	18:20
Diagnostics Check	OK	18:20

Data Download Date/Time: 07/02/2020 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LALANE,LYNN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/14/2020 20:03:29

Diagnostics Check	OK	19:56
Air Blank	0.000	19:57
Control Test	0.079	19:57
Air Blank	0.000	19:58
Subject Sample # 1	0.180	19:58
Breath Volume	2.222	19:58
Air Blank	0.000	19:59
Air Blank	0.000	20:01
Subject Sample # 2	0.175	20:01
Breath Volume	2.121	20:01
Air Blank	0.000	20:02
Control Test	0.079	20:02
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

Data Download Date/Time: 07/02/2020 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: MEDINA VALDEZ

DATE/TIME OF TEST: 06/15/2020 02:47:53

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.080	02:42
Air Blank		0.000	02:42
Subject Sample	# 1	0.242	02:43
Breath Volume		1.230	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:45
Subject Sample	# 2	0.236	02:46
Breath Volume		1.566	02:46
Air Blank		0.000	02:46
Control Test		0.078	02:47
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 07/02/2020 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALSHAIER,E
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: DOSANJH SEPIELLI

DATE/TIME OF TEST: 06/15/2020 03:26:20

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.079	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.148	03:21
Breath Volume	1.914	03:21
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.135	03:24
Breath Volume	1.847	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 07/02/2020 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BEAUREGARD,JEFF
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: CASIMIR

DATE/TIME OF TEST: 06/20/2020 01:13:07

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.246	01:08
Breath Volume		2.609	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:10
Subject Sample	# 2	0.250	01:11
Breath Volume		2.136	01:11
Air Blank		0.000	01:12
Control Test		0.078	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 07/02/2020 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 06/20/2020 22:57:36

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.079	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.148	22:52
Breath Volume	2.175	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.142	22:56
Breath Volume	1.773	22:56
Air Blank	0.000	22:56
Control Test	0.079	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 07/02/2020 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SHAH

DATE/TIME OF TEST: 06/21/2020 00:33:34

Diagnostics Check		OK	00:26
Air Blank		0.000	00:26
Control Test		0.076	00:26
Air Blank		0.000	00:27
Subject Sample	# 1	0.252	00:28
Breath Volume		1.472	00:28
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.241	00:31
Breath Volume		1.230	00:31
Air Blank		0.000	00:32
Control Test		0.076	00:32
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 07/02/2020 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: ROSALES HERNANDEZ

DATE/TIME OF TEST: 06/21/2020 04:57:39

Diagnostics Check	OK	04:51
Air Blank	0.000	04:51
Control Test	0.079	04:52
Air Blank	0.000	04:52
Subject Sample # 1	0.113	04:53
Breath Volume	2.804	04:53
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.113	04:56
Breath Volume	3.203	04:56
Air Blank	0.000	04:56
Control Test	0.079	04:57
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 07/02/2020 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: CARDONA LOZANO

DATE/TIME OF TEST: 06/27/2020 00:02:02

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.080	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.139	23:57
Breath Volume	1.261	23:57
Air Blank	0.000	23:58
Air Blank	0.000	23:59
Subject Sample # 1	0.137	00:00
Breath Volume	2.589	00:00
Air Blank	0.000	00:01
Control Test	0.079	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:02

Data Download Date/Time: 07/02/2020 13:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BARRUTIA,JESSENIA
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: FOISY

DATE/TIME OF TEST: 06/29/2020 19:49:15

Diagnostics Check	OK	19:43
Air Blank	0.000	19:43
Control Test	0.079	19:43
Air Blank	0.000	19:44
Subject Sample # 1	0.000	19:44
Breath Volume	2.371	19:44
Air Blank	0.000	19:45
Air Blank	0.000	19:47
Subject Sample # 2	0.000	19:47
Breath Volume	2.265	19:47
Air Blank	0.000	19:48
Control Test	0.079	19:48
Air Blank	0.000	19:49
Diagnostics Check	OK	19:49

Data Download Date/Time: 07/02/2020 13:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBSON, BEAU, T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: MEKKAZHA NARAYANAN

DATE/TIME OF TEST: 06/13/2020 23:41:48

Diagnostics Check		OK	23:28	
Air Blank		0.000	23:29	
Control Test		0.079	23:29	
Air Blank		0.000	23:29	
Subject Sample	# 1	0.184	23:33	Volume Not Met
Breath Volume		1.070	23:33	
Air Blank		0.000	23:33	
Air Blank		0.000	23:35	
Subject Sample	# 2	0.192	23:35	
Breath Volume		1.367	23:35	
Air Blank		0.000	23:36	
Air Blank		0.000	23:38	
Subject Sample	# 3	0.184	23:40	
Breath Volume		1.328	23:40	
Air Blank		0.000	23:40	
Control Test		0.075	23:41	
Air Blank		0.000	23:41	
Diagnostics Check		OK	23:41	

Data Download Date/Time: 07/08/2020 17:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBSON, BEAU, T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: RIVERA RIVERA

DATE/TIME OF TEST: 06/14/2020 02:02:14

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:56	
Control Test	0.077	01:56	
Air Blank	0.000	01:57	
Subject Sample	# 1 0.170	01:57	
Breath Volume	4.363	01:57	
Air Blank	0.000	01:58	
Air Blank	0.000	02:00	
Subject Sample	# 2 0.160	02:00	
Breath Volume	4.320	02:00	
Air Blank	0.000	02:01	
Control Test	0.066	02:01	Control Outside Tolerance
Air Blank	0.000	02:02	

Data Download Date/Time: 07/08/2020 17:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBSON, BEAU, T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: RIVERA RIVERA

DATE/TIME OF TEST: 06/14/2020 02:06:55

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:06	
Control Test	0.082	02:06	Interferent Detect
Air Blank	0.000	02:06	

Data Download Date/Time: 07/08/2020 17:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBSON, BEAU, T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: RIVERA RIVERA

DATE/TIME OF TEST: 06/14/2020 02:15:10

Diagnostics Check		OK	02:08
Air Blank		0.000	02:08
Control Test		0.076	02:09
Air Blank		0.000	02:09
Subject Sample	# 1	0.165	02:10
Breath Volume		4.398	02:10
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.156	02:13
Breath Volume		4.242	02:13
Air Blank		0.000	02:14
Control Test		0.077	02:14
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 07/08/2020 17:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBSON, BEAU, T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: CARABALLO GARCIA

DATE/TIME OF TEST: 06/14/2020 02:28:28

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.079	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.200	02:23
Breath Volume	3.019	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:25
Subject Sample # 2	0.195	02:26
Breath Volume	2.437	02:26
Air Blank	0.000	02:27
Control Test	0.083	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 07/08/2020 17:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUSE, RONNIE, L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: JOHNSON, DARIUS
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 05/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: MAYBELL

DATE/TIME OF TEST: 06/07/2020 06:29:03

Diagnostics Check	OK	06:25	
Air Blank	0.000	06:26	
Control Test	0.079	06:26	
Air Blank	0.000	06:27	
Subject Sample	# 1	06:27	Subject Test Refused
Breath Volume	0.000	06:27	
Air Blank	0.000	06:28	
Control Test	0.080	06:28	
Air Blank	0.000	06:28	
Diagnostics Check	OK	06:29	

Data Download Date/Time: 07/06/2020 09:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANDY,JAMIE,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: RAGLIN,JEREMY
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 05/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: ARDON MENDEZ

DATE/TIME OF TEST: 06/27/2020 02:28:41

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:26	
Control Test	0.080	02:26	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:27	Subject Test Refused
Breath Volume	0.000	02:27	
Air Blank	0.000	02:27	
Control Test	0.080	02:28	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:28	

Data Download Date/Time: 07/06/2020 09:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FIERRO,ALEX
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: KARCHER

DATE/TIME OF TEST: 06/01/2020 00:16:04

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:13	
Control Test	0.079	00:14	
Air Blank	0.000	00:14	
Subject Sample	# 1	00:14	Subject Test Refused
Breath Volume	0.000	00:14	
Air Blank	0.000	00:15	
Control Test	0.079	00:15	
Air Blank	0.000	00:15	
Diagnostics Check	OK	00:16	

Data Download Date/Time: 06/09/2020 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LANE,JONATHAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 06/01/2020 18:12:41

Diagnostics Check	OK	18:09	
Air Blank	0.000	18:10	
Control Test	0.080	18:10	
Air Blank	0.000	18:11	
Subject Sample	# 1	18:11	Subject Test Refused
Breath Volume	0.000	18:11	
Air Blank	0.000	18:11	
Control Test	0.079	18:12	
Air Blank	0.000	18:12	
Diagnostics Check	OK	18:12	

Data Download Date/Time: 06/09/2020 10:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: WAITE, K
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: AUGUSTINE

DATE/TIME OF TEST: 06/04/2020 02:42:33

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:40	
Control Test	0.080	02:40	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:41	Subject Test Refused
Breath Volume	0.000	02:41	
Air Blank	0.000	02:41	
Control Test	0.080	02:41	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

Data Download Date/Time: 06/09/2020 10:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE, JONATHAN, M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BUSCEMI, ANTHONY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: RICKETTS

DATE/TIME OF TEST: 06/06/2020 00:41:20

Diagnostics Check	OK	00:34
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.016	00:36
Breath Volume	3.277	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.015	00:39
Breath Volume	1.707	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 06/09/2020 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLIER, JEREMY,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: KEOHANE, KUNIKO
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: JOHNSTONE

DATE/TIME OF TEST: 06/06/2020 05:45:30

Diagnostics Check	OK	05:41	
Air Blank	0.000	05:42	
Control Test	0.080	05:42	
Air Blank	0.000	05:43	
Subject Sample # 1	0.163	05:43	
Breath Volume	1.761	05:43	
Air Blank		05:43	RFI Detect
Air Blank	0.000	05:44	
Control Test	0.080	05:44	
Air Blank	0.000	05:45	
Diagnostics Check	OK	05:45	

Data Download Date/Time: 06/09/2020 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLIER, JEREMEY, J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: KEOHANE, NIKO
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: JOHNSTONE

DATE/TIME OF TEST: 06/06/2020 05:54:28

Diagnostics Check		OK	05:48
Air Blank		0.000	05:48
Control Test		0.078	05:49
Air Blank		0.000	05:49
Subject Sample	# 1	0.149	05:49
Breath Volume		1.250	05:49
Air Blank		0.000	05:50
Air Blank		0.000	05:52
Subject Sample	# 2	0.163	05:52
Breath Volume		1.332	05:52
Air Blank		0.000	05:53
Control Test		0.078	05:53
Air Blank		0.000	05:54
Diagnostics Check		OK	05:54

Data Download Date/Time: 06/09/2020 10:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: MATTINGLY

DATE/TIME OF TEST: 06/07/2020 02:34:34

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:32	
Control Test	0.079	02:32	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:33	Subject Test Refused
Breath Volume	0.000	02:33	
Air Blank	0.000	02:33	
Control Test	0.079	02:33	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:34	

Data Download Date/Time: 06/09/2020 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: WILKINSON

DATE/TIME OF TEST: 06/07/2020 02:51:41

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.079	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.236	02:46
Breath Volume	1.324	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.236	02:49
Breath Volume	1.375	02:49
Air Blank	0.000	02:50
Control Test	0.078	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 06/09/2020 10:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLIER, JEREMEY, J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HUDSON, A
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: TORRES TRUJILLO

DATE/TIME OF TEST: 06/07/2020 04:47:58

Diagnostics Check	OK	04:38	
Air Blank	0.000	04:38	
Control Test	0.079	04:39	
Air Blank	0.000	04:39	
Subject Sample # 1	0.139	04:41	
Breath Volume	1.406	04:41	
Air Blank	0.000	04:42	
Air Blank	0.000	04:43	
Subject Sample # 2	0.133	04:46	
Breath Volume	1.472	04:46	
Air Blank	0.000	04:47	
		04:47	RFI Detect
Air Blank	0.000	04:47	

Data Download Date/Time: 06/09/2020 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: WAITE,KENNETH
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: ALTERIO

DATE/TIME OF TEST: 06/13/2020 23:39:03

Diagnostics Check	OK	23:32
Air Blank	0.000	23:33
Control Test	0.080	23:33
Air Blank	0.000	23:34
Subject Sample # 1	0.034	23:34
Breath Volume	2.191	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.030	23:37
Breath Volume	2.546	23:37
Air Blank	0.000	23:38
Control Test	0.079	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:39

Data Download Date/Time: 07/13/2020 12:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMSON, BRANDEE, E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: WILLIAMSON, BRANDEE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: FRICKE

DATE/TIME OF TEST: 06/16/2020 00:41:34

Diagnostics Check	OK	00:38	
Air Blank	0.000	00:38	
Control Test	0.079	00:39	
Air Blank	0.000	00:39	
Subject Sample	# 1	00:40	Subject Test Refused
Breath Volume	0.000	00:40	
Air Blank	0.000	00:40	
Control Test	0.079	00:40	
Air Blank	0.000	00:41	
Diagnostics Check	OK	00:41	

Data Download Date/Time: 07/13/2020 12:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLIER, JEREMEY, J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HUDSON, ANDREW
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: MCCABE

DATE/TIME OF TEST: 06/16/2020 20:52:41

Diagnostics Check	OK	20:49	
Air Blank	0.000	20:50	
Control Test	0.079	20:50	
Air Blank	0.000	20:51	
Subject Sample	# 1	20:51	Subject Test Refused
Breath Volume	0.000	20:51	
Air Blank	0.000	20:51	
Control Test	0.079	20:52	
Air Blank	0.000	20:52	
Diagnostics Check	OK	20:52	

Data Download Date/Time: 07/13/2020 12:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: WHITE,BRANDON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: SPRINGER

DATE/TIME OF TEST: 06/20/2020 00:34:02

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:31	
Control Test	0.079	00:31	
Air Blank	0.000	00:32	
Subject Sample	# 1	00:32	Subject Test Refused
Breath Volume	0.000	00:32	
Air Blank	0.000	00:33	
Control Test	0.079	00:33	
Air Blank	0.000	00:33	
Diagnostics Check	OK	00:34	

Data Download Date/Time: 07/13/2020 12:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOGERLEY,NEIL,F
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MOGERLEY,NEIL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: NARANJO

DATE/TIME OF TEST: 06/20/2020 02:01:38

Diagnostics Check	OK	02:01	
Air Blank		02:01	RFI Detect
Air Blank	0.000	02:01	

Data Download Date/Time: 07/13/2020 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOGERLEY,NEIL,F
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MOGERLEY,NEIL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: NARANJO

DATE/TIME OF TEST: 06/20/2020 02:11:43

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.079	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.125	02:06
Breath Volume	1.363	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.124	02:10
Breath Volume	1.164	02:10
Air Blank	0.000	02:10
Control Test	0.079	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 07/13/2020 12:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLIER, JEREMEY, J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: TELLIER, JEREMEY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: FENSTERMAKER

DATE/TIME OF TEST: 06/21/2020 02:28:54

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:26	
Control Test	0.079	02:26	
Air Blank	0.000	02:27	
Subject Sample	# 1	02:27	Subject Test Refused
Breath Volume	0.000	02:27	
Air Blank	0.000	02:27	
Control Test	0.079	02:28	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:28	

Data Download Date/Time: 07/13/2020 12:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 06/21/2020 03:04:18

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:01	
Control Test	0.079	03:02	
Air Blank	0.000	03:02	
Subject Sample	# 1	03:02	Subject Test Refused
Breath Volume	0.000	03:02	
Air Blank	0.000	03:03	
Control Test	0.079	03:03	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:04	

Data Download Date/Time: 07/13/2020 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: MASSENGILL

DATE/TIME OF TEST: 06/23/2020 17:51:42

Diagnostics Check	OK	17:45
Air Blank	0.000	17:45
Control Test	0.078	17:46
Air Blank	0.000	17:46
Subject Sample # 1	0.277	17:47
Breath Volume	3.753	17:47
Air Blank	0.000	17:47
Air Blank	0.000	17:49
Subject Sample # 2	0.267	17:50
Breath Volume	3.757	17:50
Air Blank	0.000	17:50
Control Test	0.079	17:51
Air Blank	0.000	17:51
Diagnostics Check	OK	17:51

Data Download Date/Time: 07/13/2020 12:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 06/27/2020 22:47:15

Diagnostics Check	OK	22:38
Air Blank	0.000	22:39
Control Test	0.079	22:39
Air Blank	0.000	22:40
Subject Sample # 1	0.170	22:40
Breath Volume	2.113	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.153	22:45
Breath Volume	2.453	22:45
Air Blank	0.000	22:46
Control Test	0.079	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 07/13/2020 12:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,ANDREW,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HUDSON,ANDREW
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: SELLERS

DATE/TIME OF TEST: 06/30/2020 00:33:59

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:31	
Control Test	0.080	00:31	
Air Blank	0.000	00:32	
Subject Sample	# 1	00:32	Subject Test Refused
Breath Volume	0.000	00:32	
Air Blank	0.000	00:33	
Control Test	0.079	00:33	
Air Blank	0.000	00:33	
Diagnostics Check	OK	00:33	

Data Download Date/Time: 07/13/2020 12:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MORENO,JORGE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: MCLEROY

DATE/TIME OF TEST: 06/05/2020 23:38:05

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:35	
Control Test	0.079	23:36	
Air Blank	0.000	23:36	
Subject Sample	# 1	23:37	RFI Detect
Breath Volume	1.664	23:37	
Air Blank		23:37	RFI Detect
Air Blank	0.000	23:38	

Data Download Date/Time: 06/08/2020 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MORENO,JORGE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: MCLEROY

DATE/TIME OF TEST: 06/05/2020 23:46:18

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.079	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.202	23:41
Breath Volume	2.511	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.197	23:44
Breath Volume	2.648	23:44
Air Blank	0.000	23:45
Control Test	0.079	23:45
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 06/08/2020 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: DIEGUEZ

DATE/TIME OF TEST: 06/06/2020 01:15:32

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.079	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.115	01:11
Breath Volume	2.640	01:11
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.109	01:13
Breath Volume	1.320	01:13
Air Blank	0.000	01:14
Control Test	0.080	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 06/08/2020 11:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MORENO,JORGE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: GELPI PEREZ

DATE/TIME OF TEST: 06/07/2020 03:26:05

Diagnostics Check	OK	03:25	
Air Blank		03:25	RFI Detect
Air Blank	0.000	03:26	

Data Download Date/Time: 06/08/2020 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MORENO,JORGE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: GELPI PEREZ

DATE/TIME OF TEST: 06/07/2020 03:34:01

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.080	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.100	03:29
Breath Volume	2.144	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.100	03:32
Breath Volume	1.410	03:32
Air Blank	0.000	03:33
Control Test	0.079	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:34

Data Download Date/Time: 06/08/2020 11:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONES,SHAUN,G
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LONES,SHAUN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: FOLEY

DATE/TIME OF TEST: 06/19/2020 19:02:35

Diagnostics Check	OK	18:53
Air Blank	0.000	18:53
Control Test	0.080	18:54
Air Blank	0.000	18:54
Subject Sample # 1	0.230	18:56
Breath Volume	1.777	18:56
Air Blank	0.000	18:57
Air Blank	0.000	18:59
Subject Sample # 2	0.217	19:00
Breath Volume	2.019	19:00
Air Blank	0.000	19:01
Control Test	0.078	19:01
Air Blank	0.000	19:02
Diagnostics Check	OK	19:02

Data Download Date/Time: 07/21/2020 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: CABA

DATE/TIME OF TEST: 06/21/2020 04:54:33

Diagnostics Check		OK	04:41	
Air Blank		0.000	04:41	
Control Test		0.079	04:41	
Air Blank		0.000	04:42	
Subject Sample	# 1	0.305	04:45	Volume Not Met
Breath Volume		0.082	04:45	
Air Blank		0.000	04:46	
Air Blank		0.000	04:48	
Subject Sample	# 2	0.312	04:48	
Breath Volume		1.437	04:48	
Air Blank		0.000	04:49	
Air Blank		0.000	04:50	
Subject Sample	# 3	0.293	04:52	
Breath Volume		1.398	04:52	
Air Blank		0.000	04:53	
Control Test		0.078	04:53	
Air Blank		0.000	04:54	
Diagnostics Check		OK	04:54	

Data Download Date/Time: 07/21/2020 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTERO,LISSETTE,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHANDLER,KYLE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: MEDLEN

DATE/TIME OF TEST: 06/23/2020 01:49:11

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.038	01:44
Breath Volume	2.656	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.035	01:47
Breath Volume	3.445	01:47
Air Blank	0.000	01:48
Control Test	0.079	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 07/21/2020 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTERO,LISSETTE,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHANDLER,KYLE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 06/28/2020 01:17:53

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:15	
Control Test	0.079	01:15	
Air Blank	0.000	01:16	
Subject Sample	# 1	01:16	Subject Test Refused
Breath Volume	0.000	01:16	
Air Blank	0.000	01:16	
Control Test	0.079	01:17	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 07/21/2020 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTERO,LISSETTE,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FAISON,MEGAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: KIMBRO

DATE/TIME OF TEST: 06/29/2020 01:18:22

Diagnostics Check	OK	01:16	
Air Blank	0.000	01:16	
Control Test	0.079	01:16	
Air Blank	0.000	01:17	
Subject Sample	# 1	01:17	Subject Test Refused
Breath Volume	0.000	01:17	
Air Blank		01:17	RFI Detect
Air Blank	0.000	01:18	

Data Download Date/Time: 07/21/2020 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GURRIERI,JOHN,V
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GURRIERI,JOHN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 12:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: LAROSE

DATE/TIME OF TEST: 06/02/2020 14:30:23

Diagnostics Check	OK	14:23
Air Blank	0.000	14:23
Control Test	0.081	14:23
Air Blank	0.000	14:24
Subject Sample # 1	0.179	14:25
Breath Volume	2.613	14:25
Air Blank	0.000	14:26
Air Blank	0.000	14:28
Subject Sample # 2	0.174	14:28
Breath Volume	2.121	14:28
Air Blank	0.000	14:29
Control Test	0.080	14:29
Air Blank	0.000	14:30
Diagnostics Check	OK	14:30

Data Download Date/Time: 06/24/2020 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: COMBS

DATE/TIME OF TEST: 06/05/2020 00:56:22

Diagnostics Check		OK	00:48
Air Blank		0.000	00:49
Control Test		0.080	00:49
Air Blank		0.000	00:50
Subject Sample	# 1	0.210	00:51
Breath Volume		1.433	00:51
Air Blank		0.000	00:52
Air Blank		0.000	00:53
Subject Sample	# 2	0.209	00:54
Breath Volume		1.859	00:54
Air Blank		0.000	00:55
Control Test		0.079	00:55
Air Blank		0.000	00:56
Diagnostics Check		OK	00:56

Data Download Date/Time: 06/24/2020 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELINSKI,H,T
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: HELINSKI,H
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: KARKOSKI

DATE/TIME OF TEST: 06/05/2020 05:55:04

Diagnostics Check	OK	05:47
Air Blank	0.000	05:48
Control Test	0.081	05:48
Air Blank	0.000	05:49
Subject Sample # 1	0.183	05:49
Breath Volume	2.156	05:49
Air Blank	0.000	05:50
Air Blank	0.000	05:52
Subject Sample # 2	0.173	05:53
Breath Volume	1.871	05:53
Air Blank	0.000	05:54
Control Test	0.081	05:54
Air Blank	0.000	05:54
Diagnostics Check	OK	05:55

Data Download Date/Time: 06/24/2020 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORTORA,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: TORTORA,JUSTIN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/05/2020 23:54:54

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.081	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.178	23:49
Breath Volume	1.660	23:49
Air Blank	0.000	23:50
Air Blank	0.000	23:51
Subject Sample # 2	0.164	23:52
Breath Volume	1.402	23:52
Air Blank	0.000	23:53
Control Test	0.076	23:54
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 06/24/2020 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALL,CODY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALL,CODY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: STICKLEY

DATE/TIME OF TEST: 06/06/2020 01:04:48

Diagnostics Check	OK	00:57
Air Blank	0.000	00:58
Control Test	0.081	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.179	00:59
Breath Volume	2.839	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:01
Subject Sample # 2	0.179	01:03
Breath Volume	3.054	01:03
Air Blank	0.000	01:03
Control Test	0.080	01:04
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 06/24/2020 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN, JONATHAN, M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: COLEMAN, JONATHAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: DUNTON

DATE/TIME OF TEST: 06/07/2020 01:40:09

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.081	01:32	
Air Blank	0.000	01:32	
Subject Sample # 1	0.232	01:33	
Breath Volume	4.164	01:33	
Air Blank	0.000	01:34	
Air Blank	0.000	01:35	
Subject Sample # 2	0.209	01:35	
Breath Volume	1.820	01:35	
Air Blank	0.000	01:37	
Air Blank	0.000	01:38	
Subject Sample # 3	0.213	01:38	
Breath Volume	1.855	01:38	
Air Blank		01:39	Purge Fail
Air Blank	0.000	01:40	

Data Download Date/Time: 06/24/2020 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALL,CODY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALL,CODY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 06/07/2020 03:37:38

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.081	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.060	03:32
Breath Volume	1.710	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.060	03:36
Breath Volume	1.972	03:36
Air Blank	0.000	03:36
Control Test	0.081	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 06/24/2020 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALL,CODY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALL,CODY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: JAMESON

DATE/TIME OF TEST: 06/08/2020 00:22:09

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.082	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.109	00:17
Breath Volume	3.792	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:19
Subject Sample # 2	0.105	00:20
Breath Volume	2.035	00:20
Air Blank	0.000	00:21
Control Test	0.082	00:21
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 06/24/2020 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGANO,JORDAN,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: PAGANO,JORDAN
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 06/10/2020 01:12:09

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.081	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.245	01:07
Breath Volume	2.894	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:09
Subject Sample # 2	0.233	01:10
Breath Volume	3.144	01:10
Air Blank	0.000	01:11
Control Test	0.079	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 06/24/2020 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 06/10/2020 20:07:11

Diagnostics Check	OK	19:57
Air Blank	0.000	19:57
Control Test	0.082	19:58
Air Blank	0.000	19:58
Subject Sample # 1	0.147	20:01
Breath Volume	2.125	20:01
Air Blank	0.000	20:01
Air Blank	0.000	20:03
Subject Sample # 2	0.141	20:05
Breath Volume	1.367	20:05
Air Blank	0.000	20:06
Control Test	0.080	20:06
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 06/24/2020 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORCHARD,DUSTIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ORCHARD,DUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: SADLER

DATE/TIME OF TEST: 06/13/2020 19:12:26

Diagnostics Check	OK	19:05
Air Blank	0.000	19:06
Control Test	0.081	19:06
Air Blank	0.000	19:06
Subject Sample # 1	0.146	19:07
Breath Volume	1.789	19:07
Air Blank	0.000	19:08
Air Blank	0.000	19:09
Subject Sample # 2	0.139	19:10
Breath Volume	1.140	19:10
Air Blank	0.000	19:11
Control Test	0.081	19:11
Air Blank	0.000	19:12
Diagnostics Check	OK	19:12

Data Download Date/Time: 06/24/2020 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRINGTON,A,N
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: HARRINGTON,A
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 06/14/2020 20:58:58

Diagnostics Check	OK	20:51
Air Blank	0.000	20:52
Control Test	0.082	20:52
Air Blank	0.000	20:53
Subject Sample # 1	0.000	20:53
Breath Volume	1.335	20:53
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.000	20:57
Breath Volume	1.195	20:57
Air Blank	0.000	20:58
Control Test	0.082	20:58
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 06/24/2020 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,AUSTIN',D
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: COON,AUSTIN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: ELYAMAN

DATE/TIME OF TEST: 06/17/2020 02:19:23

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.081	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.000	02:14
Breath Volume	1.476	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.000	02:17
Breath Volume	3.019	02:17
Air Blank	0.000	02:18
Control Test	0.081	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 06/24/2020 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSACI,NEIL,R
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ROSACI,NEIL
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: CARSEY

DATE/TIME OF TEST: 06/18/2020 21:47:12

Diagnostics Check	OK	21:40
Air Blank	0.000	21:41
Control Test	0.081	21:41
Air Blank	0.000	21:41
Subject Sample # 1	0.157	21:42
Breath Volume	2.460	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:44
Subject Sample # 2	0.142	21:45
Breath Volume	1.691	21:45
Air Blank	0.000	21:46
Control Test	0.079	21:46
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 06/24/2020 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITMORE,JAMES,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WHITMORE,JAMES
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 10:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/19/2020 00:26:55

Diagnostics Check		OK	00:11	
Air Blank		0.000	00:11	
Control Test		0.081	00:12	
Air Blank		0.000	00:12	
Subject Sample	# 1	0.185	00:13	
Breath Volume		1.890	00:13	
Air Blank		0.000	00:14	
Air Blank		0.000	00:16	
Subject Sample	# 2	0.175	00:19	Volume Not Met
Breath Volume		0.359	00:19	
Air Blank		0.000	00:20	
Air Blank		0.000	00:21	
Subject Sample	# 3	0.160	00:25	Volume Not Met
Breath Volume		0.753	00:25	
Air Blank		0.000	00:25	
Control Test		0.078	00:26	
Air Blank		0.000	00:26	
Diagnostics Check		OK	00:26	

Data Download Date/Time: 06/24/2020 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,R,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WEST,R
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/19/2020 02:08:21

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:05	
Control Test	0.080	02:06	
Air Blank	0.000	02:06	
Subject Sample	# 1	02:06	Subject Test Refused
Breath Volume	0.000	02:06	
Air Blank	0.000	02:07	
Control Test	0.080	02:07	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

Data Download Date/Time: 06/24/2020 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 06/20/2020 01:52:20

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:42	
Control Test	0.081	01:43	
Air Blank	0.000	01:43	
Subject Sample	# 1	01:46	No Sample Provided
Breath Volume	0.000	01:46	
Air Blank	0.000	01:47	
Air Blank	0.000	01:49	
Subject Sample	# 2	01:50	Subject Test Refused
Breath Volume	0.000	01:50	
Air Blank	0.000	01:51	
Control Test	0.081	01:51	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 06/24/2020 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 06/20/2020 22:47:29

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.081	22:42
Air Blank	0.000	22:42
Subject Sample # 1	0.000	22:43
Breath Volume	2.644	22:43
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.000	22:46
Breath Volume	3.078	22:46
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 06/24/2020 10:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STITH,RYAN,C
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: STITH,RYAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: MOATZ

DATE/TIME OF TEST: 06/21/2020 20:07:33

Diagnostics Check	OK	20:01
Air Blank	0.000	20:01
Control Test	0.081	20:01
Air Blank	0.000	20:02
Subject Sample # 1	0.247	20:02
Breath Volume	2.558	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.239	20:05
Breath Volume	3.085	20:05
Air Blank	0.000	20:06
Control Test	0.080	20:06
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 06/24/2020 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER\,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: LLOYD

DATE/TIME OF TEST: 06/24/2020 22:45:02

Diagnostics Check	OK	22:37
Air Blank	0.000	22:38
Control Test	0.080	22:38
Air Blank	0.000	22:38
Subject Sample # 1	0.174	22:40
Breath Volume	1.992	22:40
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.163	22:43
Breath Volume	1.761	22:43
Air Blank	0.000	22:44
Control Test	0.079	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:45

Data Download Date/Time: 07/23/2020 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALL,CODY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALL,CODY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: CROOK

DATE/TIME OF TEST: 06/29/2020 21:44:56

Diagnostics Check	OK	21:37
Air Blank	0.000	21:38
Control Test	0.080	21:38
Air Blank	0.000	21:39
Subject Sample # 1	0.140	21:39
Breath Volume	2.027	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.129	21:43
Breath Volume	1.148	21:43
Air Blank	0.000	21:43
Control Test	0.080	21:44
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 07/23/2020 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,JAMES,W
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BURGESS,JAMES
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: HAVENS

DATE/TIME OF TEST: 06/30/2020 20:09:41

Diagnostics Check		OK	19:55	
Air Blank		0.000	19:55	
Control Test		0.080	19:55	
Air Blank		0.000	19:56	
Subject Sample	# 1	0.134	19:59	Volume Not Met
Breath Volume		0.121	19:59	
Air Blank		0.000	20:00	
Air Blank		0.000	20:02	
Subject Sample	# 2	0.132	20:04	
Breath Volume		1.488	20:04	
Air Blank		0.000	20:04	
Air Blank		0.000	20:06	
Subject Sample	# 3	0.115	20:08	
Breath Volume		1.554	20:08	
Air Blank		0.000	20:08	
Control Test		0.079	20:09	
Air Blank		0.000	20:09	
Diagnostics Check		OK	20:09	

Data Download Date/Time: 07/23/2020 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CATHERINE,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HALL,CATHERINE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: FEAGIN

DATE/TIME OF TEST: 06/30/2020 22:48:01

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:45	
Control Test	0.080	22:45	
Air Blank	0.000	22:46	
Subject Sample	# 1	22:46	Subject Test Refused
Breath Volume	0.000	22:46	
Air Blank	0.000	22:47	
Control Test	0.080	22:47	
Air Blank	0.000	22:47	
Diagnostics Check	OK	22:48	

Data Download Date/Time: 07/23/2020 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PAN,JACKY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 06/16/2020 03:03:59

Diagnostics Check		OK	02:54
Air Blank		0.000	02:55
Control Test		0.079	02:55
Air Blank		0.000	02:56
Subject Sample	# 1	0.208	02:58
Breath Volume		2.210	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.207	03:02
Breath Volume		1.343	03:02
Air Blank		0.000	03:03
Control Test		0.078	03:03
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 06/17/2020 19:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PAN,JACKY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: HESS JR

DATE/TIME OF TEST: 06/20/2020 19:41:44

Diagnostics Check	OK	19:35
Air Blank	0.000	19:35
Control Test	0.078	19:35
Air Blank	0.000	19:36
Subject Sample # 1	0.205	19:37
Breath Volume	4.656	19:37
Air Blank	0.000	19:37
Air Blank	0.000	19:39
Subject Sample # 2	0.191	19:40
Breath Volume	4.605	19:40
Air Blank	0.000	19:40
Control Test	0.080	19:41
Air Blank	0.000	19:41
Diagnostics Check	OK	19:41

Data Download Date/Time: 07/15/2020 20:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PAN,JACKY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: FUNEZ GONZALES

DATE/TIME OF TEST: 06/21/2020 23:39:22

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.078	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.167	23:34
Breath Volume	3.378	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.169	23:37
Breath Volume	2.667	23:37
Air Blank	0.000	23:38
Control Test	0.080	23:38
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 07/15/2020 20:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYDOS,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GAYDOS,JONATHAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: HENSHAW

DATE/TIME OF TEST: 06/22/2020 23:57:58

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.078	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.149	23:53
Breath Volume	2.359	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:55
Subject Sample # 2	0.153	23:56
Breath Volume	2.273	23:56
Air Blank	0.000	23:57
Control Test	0.079	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 07/15/2020 20:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: VENTRY,WILLIS
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MULLER

DATE/TIME OF TEST: 06/25/2020 00:49:25

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:38	
Control Test	0.079	00:38	
Air Blank	0.000	00:39	
Subject Sample	# 1	00:42	No Sample Provided
Breath Volume	0.000	00:42	
Air Blank	0.000	00:42	
Air Blank	0.000	00:44	
Subject Sample	# 2	00:47	No Sample Provided
Breath Volume	0.000	00:47	
Air Blank	0.000	00:48	
Control Test	0.080	00:48	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 07/15/2020 20:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYDOS,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MOSE,JOSHUA
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: ESPINOSA BALCAZAR

DATE/TIME OF TEST: 06/28/2020 22:22:14

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.078	22:15
Air Blank	0.000	22:16
Subject Sample # 1	0.215	22:17
Breath Volume	3.765	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.209	22:20
Breath Volume	3.015	22:20
Air Blank	0.000	22:21
Control Test	0.079	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 07/15/2020 20:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABDELHADI,AUSTIN,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: ABDELHADI,AUSTIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 13:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:34

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 06/01/2020 14:06:13

Diagnostics Check	OK	13:59
Air Blank	0.000	14:00
Control Test	0.080	14:00
Air Blank	0.000	14:01
Subject Sample # 1	0.000	14:01
Breath Volume	2.457	14:01
Air Blank	0.000	14:02
Air Blank	0.000	14:04
Subject Sample # 2	0.000	14:04
Breath Volume	2.441	14:04
Air Blank	0.000	14:05
Control Test	0.079	14:05
Air Blank	0.000	14:06
Diagnostics Check	OK	14:06

Data Download Date/Time: 06/29/2020 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: VICKERY

DATE/TIME OF TEST: 06/13/2020 02:04:42

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.126	02:00
Breath Volume	2.808	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.126	02:03
Breath Volume	2.835	02:03
Air Blank	0.000	02:03
Control Test	0.078	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 06/29/2020 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JAMIE,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: CHECHAK,LANA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 06/13/2020 04:37:22

Diagnostics Check		OK	04:31
Air Blank		0.000	04:31
Control Test		0.079	04:31
Air Blank		0.000	04:32
Subject Sample	# 1	0.167	04:32
Breath Volume		1.464	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.180	04:35
Breath Volume		2.414	04:35
Air Blank		0.000	04:36
Control Test		0.078	04:36
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 06/29/2020 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: GALLUP

DATE/TIME OF TEST: 06/14/2020 01:49:57

Diagnostics Check	OK	01:42
Air Blank	0.000	01:43
Control Test	0.079	01:43
Air Blank	0.000	01:44
Subject Sample # 1	0.135	01:45
Breath Volume	3.539	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.135	01:48
Breath Volume	3.343	01:48
Air Blank	0.000	01:49
Control Test	0.078	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 06/29/2020 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEARN,SHELDON,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PARENTE,JOSEPH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 05:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: ROBERSON

DATE/TIME OF TEST: 06/18/2020 06:21:28

Diagnostics Check		OK	06:14
Air Blank		0.000	06:15
Control Test		0.079	06:15
Air Blank		0.000	06:16
Subject Sample	# 1	0.212	06:16
Breath Volume		2.453	06:16
Air Blank		0.000	06:17
Air Blank		0.000	06:19
Subject Sample	# 2	0.204	06:19
Breath Volume		1.519	06:19
Air Blank		0.000	06:20
Control Test		0.078	06:20
Air Blank		0.000	06:21
Diagnostics Check		OK	06:21

Data Download Date/Time: 06/29/2020 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,CHASE,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COLLINS,CHASE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HAZLEY

DATE/TIME OF TEST: 06/19/2020 02:32:35

Diagnostics Check		OK	02:20	
Air Blank		0.000	02:21	
Control Test		0.079	02:21	
Air Blank		0.000	02:22	
Subject Sample	# 1	0.166	02:25	Volume Not Met
Breath Volume		0.214	02:25	
Air Blank		0.000	02:26	
Air Blank		0.000	02:27	
Subject Sample	# 2	0.138	02:30	Volume Not Met
Breath Volume		0.941	02:30	
Air Blank		0.000	02:31	
Control Test		0.078	02:31	
Air Blank		0.000	02:32	
Diagnostics Check		OK	02:32	

Data Download Date/Time: 06/29/2020 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,CHASE,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COLLINS,CHASE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HAZLEY

DATE/TIME OF TEST: 06/19/2020 02:46:02

Diagnostics Check		OK	02:34	
Air Blank		0.000	02:34	
Control Test		0.078	02:35	
Air Blank		0.000	02:35	
Subject Sample	# 1	0.119	02:38	Volume Not Met
Breath Volume		0.375	02:38	
Air Blank		0.000	02:39	
Air Blank		0.000	02:41	
Subject Sample	# 2	0.160	02:44	Volume Not Met
Breath Volume		0.542	02:44	
Air Blank		0.000	02:45	
Control Test		0.078	02:45	
Air Blank		0.000	02:45	
Diagnostics Check		OK	02:46	

Data Download Date/Time: 06/29/2020 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LEACH,ZACHARY,M

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: TUESBURG,TROY

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1114319

DRY GAS EXPIRY DATE: 03/27/2022

VIOLATION CODE: DUI

ARREST TIME: 19:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: RIALS

DATE/TIME OF TEST: 06/19/2020 19:54:20

Diagnostics Check	OK	19:53	
Air Blank	0.000	19:53	
Control Test	0.079	19:53	
Air Blank		19:54	RFI Detect
Air Blank		19:54	RFI Detect

Data Download Date/Time: 06/29/2020 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEACH,ZACHARY,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: TUESBURG,TROY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: RIALS

DATE/TIME OF TEST: 06/19/2020 20:02:18

Diagnostics Check	OK	19:56
Air Blank	0.000	19:56
Control Test	0.079	19:56
Air Blank	0.000	19:57
Subject Sample # 1	0.162	19:57
Breath Volume	3.066	19:57
Air Blank	0.000	19:58
Air Blank	0.000	20:00
Subject Sample # 2	0.162	20:00
Breath Volume	2.769	20:00
Air Blank	0.000	20:01
Control Test	0.078	20:01
Air Blank	0.000	20:02
Diagnostics Check	OK	20:02

Data Download Date/Time: 06/29/2020 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GARCIA,MICHAEL,A

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: ABED,ABYGAEI

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1114319

DRY GAS EXPIRY DATE: 03/27/2022

VIOLATION CODE: DUI

ARREST TIME: 05:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: KUHNHAUSEN

DATE/TIME OF TEST: 06/21/2020 05:38:58

Diagnostics Check	OK	05:35	
Air Blank	0.000	05:35	
Control Test	0.079	05:36	
Air Blank	0.000	05:36	
Subject Sample	# 1	05:37	RFI Detect
Breath Volume	2.968	05:37	
Air Blank	0.000	05:38	
Control Test	0.078	05:38	
Air Blank	0.000	05:38	
Diagnostics Check	OK	05:38	

Data Download Date/Time: 06/29/2020 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GARCIA,MICHAEL,A

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: ABED,ABYGAE

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1114319

DRY GAS EXPIRY DATE: 03/27/2022

VIOLATION CODE: DUI

ARREST TIME: 05:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: KUHNHAUSEN

DATE/TIME OF TEST: 06/21/2020 05:45:04

Diagnostics Check	OK	05:40	
Air Blank	0.000	05:41	
Control Test	0.077	05:41	
Air Blank	0.000	05:42	
Subject Sample	# 1	05:43	RFI Detect
Breath Volume	1.046	05:43	
Air Blank	0.000	05:44	
Control Test	0.078	05:44	
Air Blank	0.000	05:44	
Diagnostics Check	OK	05:45	

Data Download Date/Time: 06/29/2020 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GARCIA, MICHAEL, A

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: ABED, ABYGAEEL

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1114319

DRY GAS EXPIRY DATE: 03/27/2022

VIOLATION CODE: DUI

ARREST TIME: 05:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: KUHNHAUSEN

DATE/TIME OF TEST: 06/21/2020 05:50:43

Diagnostics Check	OK	05:46	
Air Blank	0.000	05:47	
Control Test	0.078	05:47	
Air Blank	0.000	05:47	
Subject Sample	# 1 0.142	05:48	
Breath Volume	1.296	05:48	
Air Blank		05:49	RFI Detect
Air Blank	0.000	05:49	
Control Test	0.078	05:50	
Air Blank	0.000	05:50	
Diagnostics Check	OK	05:50	

Data Download Date/Time: 06/29/2020 18:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GARCIA,MICHAEL,A

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: ABED,ABYGAEI

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1114319

DRY GAS EXPIRY DATE: 03/27/2022

VIOLATION CODE: DUI

ARREST TIME: 05:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: KUHNHAUSEN

DATE/TIME OF TEST: 06/21/2020 05:55:21

Diagnostics Check	OK	05:52	
Air Blank	0.000	05:52	
Control Test	0.076	05:53	
Air Blank	0.000	05:53	
Subject Sample	# 1	05:54	Sequence Aborted
Breath Volume	0.289	05:54	
Air Blank		05:55	RFI Detect

Data Download Date/Time: 06/29/2020 18:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: PEARCE,BRANDON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 02/23/2020
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: ALLTOP

DATE/TIME OF TEST: 06/02/2020 00:08:21

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:58	
Control Test	0.079	23:59	
Air Blank	0.000	23:59	
Subject Sample	# 1 0.319	00:02	Volume Not Met
Breath Volume	0.265	00:02	
Air Blank	0.000	00:03	
Air Blank	0.000	00:05	
Subject Sample	# 2	00:06	Subject Test Refused
Breath Volume	0.308	00:06	
Air Blank	0.000	00:07	
Control Test	0.075	00:07	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:08	

Data Download Date/Time: 06/24/2020 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: DORON

DATE/TIME OF TEST: 06/07/2020 04:16:32

Diagnostics Check	OK	04:13	
Air Blank	0.000	04:14	
Control Test	0.079	04:14	
Air Blank	0.000	04:15	
Subject Sample	# 1	04:15	Subject Test Refused
Breath Volume	0.000	04:15	
Air Blank	0.000	04:15	
Control Test	0.078	04:15	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 06/24/2020 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: SINNO

DATE/TIME OF TEST: 06/16/2020 04:17:29

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.079	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.170	04:13
Breath Volume	2.218	04:13
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.173	04:15
Breath Volume	2.000	04:15
Air Blank	0.000	04:16
Control Test	0.078	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 06/24/2020 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLLIFF, ABIGAIL, L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: HARVEY, CALEB
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 02/23/2020
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: ALKHAJA

DATE/TIME OF TEST: 06/24/2020 02:48:44

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.000	02:44
Breath Volume	1.578	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.000	02:47
Breath Volume	1.378	02:47
Air Blank	0.000	02:47
Control Test	0.077	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 06/24/2020 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YARBOROUGH,ERIC,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: YARBOROUGH,ERIC
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 06/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HEISLITZ

DATE/TIME OF TEST: 06/28/2020 03:56:24

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:51	
Control Test	0.079	03:51	
Air Blank	0.000	03:52	
Subject Sample	# 1	03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:55	
Control Test	0.079	03:55	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 07/22/2020 11:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYWARD,KYLE,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCNETT,ELI
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 06/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 16:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:14

SUBJECT'S LAST NAME: DEBITETTO

DATE/TIME OF TEST: 06/29/2020 17:59:57

Diagnostics Check	OK	17:51	
Air Blank	0.000	17:52	
Control Test	0.077	17:52	
Air Blank	0.000	17:52	
Subject Sample # 1	0.206	17:54	
Breath Volume	1.207	17:54	
Air Blank	0.000	17:55	
Air Blank	0.000	17:57	
Subject Sample # 2	0.194	17:58	
Breath Volume	1.222	17:58	
Air Blank	0.000	17:59	
Control Test	0.068	17:59	Control Outside Tolerance
Air Blank	0.000	17:59	

Data Download Date/Time: 07/22/2020 11:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENNING,LUKE,
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SYVERSON,PAUL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: RUSHING

DATE/TIME OF TEST: 06/02/2020 23:36:53

Diagnostics Check	OK	23:36	
Air Blank		23:36	RFI Detect
Air Blank	0.000	23:36	

Data Download Date/Time: 06/12/2020 21:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENNING,LUKE,
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SYVERSON,PAUL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: RUSHING

DATE/TIME OF TEST: 06/02/2020 23:49:49

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:43	
Control Test	0.077	23:43	
Air Blank	0.000	23:44	
Subject Sample	# 1 0.152	23:45	
Breath Volume	2.414	23:45	
Air Blank	0.000	23:45	
Air Blank	0.000	23:47	
Subject Sample	# 2 0.149	23:48	
Breath Volume	2.183	23:48	
Air Blank	0.000	23:48	
Control Test	0.074	23:49	Control Outside Tolerance
Air Blank	0.000	23:49	

Data Download Date/Time: 06/12/2020 21:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,ALEXIS,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: PINCUS

DATE/TIME OF TEST: 06/03/2020 01:46:26

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.076	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.102	01:41
Breath Volume	3.476	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.100	01:44
Breath Volume	2.714	01:44
Air Blank	0.000	01:45
Control Test	0.076	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 06/12/2020 21:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITEHEAD,JEFFREY,A
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: BOBLITT,RICK
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: ATCHISON

DATE/TIME OF TEST: 06/06/2020 06:04:26

Diagnostics Check	OK	05:56	
Air Blank	0.000	05:56	
Control Test	0.077	05:56	
Air Blank	0.000	05:57	
Subject Sample	# 1 0.139	05:58	
Breath Volume	1.414	05:58	
Air Blank	0.000	05:59	
Air Blank	0.000	06:01	
Subject Sample	# 2 0.137	06:02	
Breath Volume	1.253	06:02	
Air Blank	0.000	06:03	
Control Test	0.071	06:03	Control Outside Tolerance
Air Blank	0.000	06:04	

Data Download Date/Time: 06/12/2020 21:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLEOD,JOEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCLEOD,JOEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: BUCKLE

DATE/TIME OF TEST: 06/06/2020 08:04:10

Diagnostics Check	OK	07:57
Air Blank	0.000	07:58
Control Test	0.076	07:58
Air Blank	0.000	07:59
Subject Sample # 1	0.147	07:59
Breath Volume	2.429	07:59
Air Blank	0.000	08:00
Air Blank	0.000	08:02
Subject Sample # 2	0.135	08:02
Breath Volume	1.636	08:02
Air Blank	0.000	08:03
Control Test	0.075	08:03
Air Blank	0.000	08:04
Diagnostics Check	OK	08:04

Data Download Date/Time: 06/12/2020 21:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: KENNEL

DATE/TIME OF TEST: 06/06/2020 21:10:21

Diagnostics Check	OK	21:07	
Air Blank	0.000	21:07	
Control Test	0.075	21:08	
Air Blank	0.000	21:08	
Subject Sample	# 1	21:08	Subject Test Refused
Breath Volume	0.000	21:08	
Air Blank	0.000	21:09	
Control Test	0.075	21:09	
Air Blank	0.000	21:10	
Diagnostics Check	OK	21:10	

Data Download Date/Time: 06/12/2020 21:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNETT,DAVID,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BENNETT,DAVID
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 06/13/2020 01:49:26

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:47	
Control Test	0.075	01:47	
Air Blank	0.000	01:47	
Subject Sample	# 1	01:48	Subject Test Refused
Breath Volume	0.000	01:48	
Air Blank	0.000	01:48	
Control Test	0.076	01:48	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

Data Download Date/Time: 07/15/2020 21:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERY,JOHN,W
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: BOROS,CHUWAN
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 06/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: KEY

DATE/TIME OF TEST: 06/13/2020 17:25:03

Diagnostics Check		OK	17:10	
Air Blank		0.000	17:10	
Control Test		0.077	17:11	
Air Blank		0.000	17:11	
Subject Sample	# 1	0.306	17:14	Volume Not Met
Breath Volume		0.320	17:14	
Air Blank		0.000	17:15	
Air Blank		0.000	17:17	
Subject Sample	# 2	0.338	17:17	
Breath Volume		1.691	17:17	
Air Blank		0.000	17:18	
Air Blank		0.000	17:20	
Subject Sample	# 3	0.291	17:23	Volume Not Met
Breath Volume		0.000	17:23	
Air Blank		0.000	17:24	
Control Test		0.076	17:24	
Air Blank		0.000	17:24	
Diagnostics Check		OK	17:25	

Data Download Date/Time: 07/15/2020 21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUGATE,MITCHELL,W
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TALLEY,DAVID
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: BACA

DATE/TIME OF TEST: 06/14/2020 23:29:04

Diagnostics Check	OK	23:23	
Air Blank	0.000	23:23	
Control Test	0.075	23:23	
Air Blank	0.000	23:24	
Subject Sample	# 1 0.047	23:24	
Breath Volume	1.964	23:24	
Air Blank	0.000	23:25	
Air Blank	0.000	23:27	
Subject Sample	# 2 0.043	23:27	
Breath Volume	1.292	23:27	
Air Blank	0.000	23:28	
Control Test	0.074	23:28	Control Outside Tolerance
Air Blank	0.000	23:29	

Data Download Date/Time: 07/15/2020 21:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLEOD,JOEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCLEOD,JOEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: HERNANDEZ-KING

DATE/TIME OF TEST: 06/17/2020 22:33:24

Diagnostics Check	OK	22:26
Air Blank	0.000	22:27
Control Test	0.077	22:27
Air Blank	0.000	22:28
Subject Sample # 1	0.200	22:28
Breath Volume	2.031	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.204	22:31
Breath Volume	2.523	22:31
Air Blank	0.000	22:32
Control Test	0.076	22:32
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 07/15/2020 21:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,ALEXIS,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCALIA,ADAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: LOYDA-CRUZ

DATE/TIME OF TEST: 06/20/2020 00:44:59

Diagnostics Check		OK	00:36
Air Blank		0.000	00:36
Control Test		0.077	00:36
Air Blank		0.000	00:37
Subject Sample	# 1	0.163	00:40
Breath Volume		1.765	00:40
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.155	00:43
Breath Volume		1.656	00:43
Air Blank		0.000	00:44
Control Test		0.075	00:44
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 07/15/2020 21:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: BIEGLER

DATE/TIME OF TEST: 06/20/2020 05:05:25

Diagnostics Check	OK	05:02	
Air Blank	0.000	05:03	
Control Test	0.077	05:03	
Air Blank	0.000	05:03	
Subject Sample	# 1	05:04	Subject Test Refused
Breath Volume	0.000	05:04	
Air Blank	0.000	05:04	
Control Test	0.077	05:04	
Air Blank	0.000	05:05	
Diagnostics Check	OK	05:05	

Data Download Date/Time: 07/15/2020 21:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,ALEXIS,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SPAULDING

DATE/TIME OF TEST: 06/21/2020 00:16:50

Diagnostics Check	OK	00:10
Air Blank	0.000	00:11
Control Test	0.077	00:11
Air Blank	0.000	00:11
Subject Sample # 1	0.061	00:12
Breath Volume	2.058	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:14
Subject Sample # 2	0.060	00:15
Breath Volume	2.085	00:15
Air Blank	0.000	00:15
Control Test	0.077	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 07/15/2020 21:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAY,JODY,L
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DAY,JODY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 06/26/2020 20:33:26

Diagnostics Check	OK	20:30	
Air Blank	0.000	20:31	
Control Test	0.076	20:31	
Air Blank	0.000	20:31	
Subject Sample	# 1	20:32	Subject Test Refused
Breath Volume	0.000	20:32	
Air Blank	0.000	20:32	
Control Test	0.075	20:32	
Air Blank	0.000	20:33	
Diagnostics Check	OK	20:33	

Data Download Date/Time: 07/15/2020 21:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONTENOT,DOMINIC,M
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FONTENOT,DOMINIC
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 06/06/2020 01:35:00

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.000	01:30
Breath Volume	1.843	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.000	01:33
Breath Volume	2.062	01:33
Air Blank	0.000	01:34
Control Test	0.078	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:35

Data Download Date/Time: 06/24/2020 09:26