

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,GONZALO,
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: GARCIA,GONZALO
ARRESTING OFFICER AGENCY: MEDLEY PD
LAST AGENCY INSPECTION: 06/25/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: CHAVARRIA

DATE/TIME OF TEST: 07/08/2020 20:46:07

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 20:43 | |
| Air Blank | 0.000 | 20:43 | |
| Control Test | 0.079 | 20:43 | |
| Air Blank | 0.000 | 20:44 | |
| Subject Sample | # 1 | 20:44 | Subject Test Refused |
| Breath Volume | 0.000 | 20:44 | |
| Air Blank | 0.000 | 20:45 | |
| Control Test | 0.079 | 20:45 | |
| Air Blank | 0.000 | 20:46 | |
| Diagnostics Check | OK | 20:46 | |

Data Download Date/Time: 07/29/2020 15:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,O,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: FERNANDEZ,O
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 06/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: QUINTANILLA

DATE/TIME OF TEST: 07/01/2020 23:10:11

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 23:07 | |
| Air Blank | 0.000 | 23:07 | |
| Control Test | 0.080 | 23:07 | |
| Air Blank | 0.000 | 23:08 | |
| Subject Sample | # 1 | 23:08 | Subject Test Refused |
| Breath Volume | 0.000 | 23:08 | |
| Air Blank | 0.000 | 23:09 | |
| Control Test | 0.081 | 23:09 | |
| Air Blank | 0.000 | 23:10 | |
| Diagnostics Check | OK | 23:10 | |

Data Download Date/Time: 07/29/2020 14:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGUILERA,ANDREW,E
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: AGUILERA,ANDREW
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 06/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PINERO

DATE/TIME OF TEST: 07/09/2020 01:50:56

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 01:46 | |
| Air Blank | 0.000 | 01:47 | |
| Control Test | 0.081 | 01:47 | |
| Air Blank | 0.000 | 01:48 | |
| Subject Sample | # 1 | 01:49 | Subject Test Refused |
| Breath Volume | 0.000 | 01:49 | |
| Air Blank | 0.000 | 01:50 | |
| Control Test | 0.081 | 01:50 | |
| Air Blank | 0.000 | 01:50 | |
| Diagnostics Check | OK | 01:50 | |

Data Download Date/Time: 07/29/2020 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NANAN,KEIL,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: FERNANDEZ,MICHAEL
ARRESTING OFFICER AGENCY: MICCOSUKEE PD
LAST AGENCY INSPECTION: 06/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: CACERES FONTANES

DATE/TIME OF TEST: 07/13/2020 04:44:47

| | | | |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK | 04:37 | |
| Air Blank | 0.000 | 04:37 | |
| Control Test | 0.081 | 04:37 | |
| Air Blank | 0.000 | 04:38 | |
| Subject Sample | # 1 0.167 | 04:40 | |
| Breath Volume | 1.179 | 04:40 | |
| Air Blank | 0.000 | 04:41 | |
| Air Blank | 0.000 | 04:42 | |
| Subject Sample | # 2 | 04:43 | Subject Test Refused |
| Breath Volume | 0.000 | 04:43 | |
| Air Blank | 0.000 | 04:43 | |
| Control Test | 0.077 | 04:44 | |
| Air Blank | 0.000 | 04:44 | |
| Diagnostics Check | OK | 04:44 | |

Data Download Date/Time: 07/29/2020 14:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 07/04/2020 02:57:44

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 02:50 |
| Air Blank | | 0.000 | 02:51 |
| Control Test | | 0.080 | 02:51 |
| Air Blank | | 0.000 | 02:52 |
| Subject Sample | # 1 | 0.014 | 02:53 |
| Breath Volume | | 4.414 | 02:53 |
| Air Blank | | 0.000 | 02:53 |
| Air Blank | | 0.000 | 02:55 |
| Subject Sample | # 2 | 0.015 | 02:56 |
| Breath Volume | | 4.371 | 02:56 |
| Air Blank | | 0.000 | 02:56 |
| Control Test | | 0.080 | 02:57 |
| Air Blank | | 0.000 | 02:57 |
| Diagnostics Check | | OK | 02:57 |

Data Download Date/Time: 07/15/2020 04:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/09/2020 21:39:35

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:33 |
| Air Blank | 0.000 | 21:33 |
| Control Test | 0.080 | 21:34 |
| Air Blank | 0.000 | 21:34 |
| Subject Sample # 1 | 0.132 | 21:35 |
| Breath Volume | 2.054 | 21:35 |
| Air Blank | 0.000 | 21:35 |
| Air Blank | 0.000 | 21:37 |
| Subject Sample # 2 | 0.127 | 21:38 |
| Breath Volume | 2.460 | 21:38 |
| Air Blank | 0.000 | 21:38 |
| Control Test | 0.080 | 21:38 |
| Air Blank | 0.000 | 21:39 |
| Diagnostics Check | OK | 21:39 |

Data Download Date/Time: 07/15/2020 04:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: HOXHA

DATE/TIME OF TEST: 07/10/2020 00:22:02

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:15 |
| Air Blank | 0.000 | 00:15 |
| Control Test | 0.081 | 00:15 |
| Air Blank | 0.000 | 00:16 |
| Subject Sample # 1 | 0.000 | 00:17 |
| Breath Volume | 2.781 | 00:17 |
| Air Blank | 0.000 | 00:17 |
| Air Blank | 0.000 | 00:19 |
| Subject Sample # 2 | 0.000 | 00:20 |
| Breath Volume | 2.429 | 00:20 |
| Air Blank | 0.000 | 00:21 |
| Control Test | 0.080 | 00:21 |
| Air Blank | 0.000 | 00:21 |
| Diagnostics Check | OK | 00:22 |

Data Download Date/Time: 07/15/2020 04:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 07/29/2020 20:55:38

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:49 |
| Air Blank | 0.000 | 20:50 |
| Control Test | 0.078 | 20:50 |
| Air Blank | 0.000 | 20:50 |
| Subject Sample # 1 | 0.000 | 20:51 |
| Breath Volume | 2.257 | 20:51 |
| Air Blank | 0.000 | 20:51 |
| Air Blank | 0.000 | 20:53 |
| Subject Sample # 2 | 0.000 | 20:54 |
| Breath Volume | 1.851 | 20:54 |
| Air Blank | 0.000 | 20:54 |
| Control Test | 0.079 | 20:55 |
| Air Blank | 0.000 | 20:55 |
| Diagnostics Check | OK | 20:55 |

Data Download Date/Time: 08/17/2020 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: GARCIA GARCIA

DATE/TIME OF TEST: 07/31/2020 20:09:24

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:02 |
| Air Blank | 0.000 | 20:03 |
| Control Test | 0.080 | 20:03 |
| Air Blank | 0.000 | 20:04 |
| Subject Sample # 1 | 0.112 | 20:04 |
| Breath Volume | 1.800 | 20:04 |
| Air Blank | 0.000 | 20:05 |
| Air Blank | 0.000 | 20:07 |
| Subject Sample # 2 | 0.114 | 20:07 |
| Breath Volume | 1.367 | 20:07 |
| Air Blank | 0.000 | 20:08 |
| Control Test | 0.079 | 20:08 |
| Air Blank | 0.000 | 20:09 |
| Diagnostics Check | OK | 20:09 |

Data Download Date/Time: 08/17/2020 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ, M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 07/01/2020 22:44:28

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:38 |
| Air Blank | 0.000 | 22:38 |
| Control Test | 0.078 | 22:39 |
| Air Blank | 0.000 | 22:39 |
| Subject Sample # 1 | 0.000 | 22:40 |
| Breath Volume | 2.441 | 22:40 |
| Air Blank | 0.000 | 22:40 |
| Air Blank | 0.000 | 22:42 |
| Subject Sample # 2 | 0.000 | 22:43 |
| Breath Volume | 2.121 | 22:43 |
| Air Blank | 0.000 | 22:43 |
| Control Test | 0.078 | 22:43 |
| Air Blank | 0.000 | 22:44 |
| Diagnostics Check | OK | 22:44 |

Data Download Date/Time: 07/15/2020 05:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARRISH, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: PHILLIPPY

DATE/TIME OF TEST: 07/02/2020 01:03:52

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:57 |
| Air Blank | 0.000 | 00:58 |
| Control Test | 0.078 | 00:58 |
| Air Blank | 0.000 | 00:58 |
| Subject Sample # 1 | 0.083 | 00:59 |
| Breath Volume | 2.242 | 00:59 |
| Air Blank | 0.000 | 01:00 |
| Air Blank | 0.000 | 01:02 |
| Subject Sample # 2 | 0.079 | 01:02 |
| Breath Volume | 1.816 | 01:02 |
| Air Blank | 0.000 | 01:02 |
| Control Test | 0.077 | 01:03 |
| Air Blank | 0.000 | 01:03 |
| Diagnostics Check | OK | 01:03 |

Data Download Date/Time: 07/15/2020 05:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARRISH, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: BOSWELL

DATE/TIME OF TEST: 07/03/2020 23:25:01

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:18 |
| Air Blank | 0.000 | 23:19 |
| Control Test | 0.078 | 23:19 |
| Air Blank | 0.000 | 23:20 |
| Subject Sample # 1 | 0.096 | 23:20 |
| Breath Volume | 3.507 | 23:20 |
| Air Blank | 0.000 | 23:21 |
| Air Blank | 0.000 | 23:23 |
| Subject Sample # 2 | 0.095 | 23:23 |
| Breath Volume | 3.558 | 23:23 |
| Air Blank | 0.000 | 23:24 |
| Control Test | 0.078 | 23:24 |
| Air Blank | 0.000 | 23:24 |
| Diagnostics Check | OK | 23:25 |

Data Download Date/Time: 07/15/2020 05:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ, M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: BAUGH

DATE/TIME OF TEST: 07/04/2020 00:47:48

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:41 |
| Air Blank | 0.000 | 00:41 |
| Control Test | 0.078 | 00:42 |
| Air Blank | 0.000 | 00:42 |
| Subject Sample # 1 | 0.117 | 00:43 |
| Breath Volume | 3.296 | 00:43 |
| Air Blank | 0.000 | 00:43 |
| Air Blank | 0.000 | 00:45 |
| Subject Sample # 2 | 0.118 | 00:46 |
| Breath Volume | 2.804 | 00:46 |
| Air Blank | 0.000 | 00:46 |
| Control Test | 0.079 | 00:47 |
| Air Blank | 0.000 | 00:47 |
| Diagnostics Check | OK | 00:47 |

Data Download Date/Time: 07/15/2020 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 07/04/2020 01:59:13

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 01:52 |
| Air Blank | | 0.000 | 01:53 |
| Control Test | | 0.077 | 01:53 |
| Air Blank | | 0.000 | 01:54 |
| Subject Sample | # 1 | 0.236 | 01:54 |
| Breath Volume | | 2.718 | 01:54 |
| Air Blank | | 0.000 | 01:55 |
| Air Blank | | 0.000 | 01:57 |
| Subject Sample | # 2 | 0.244 | 01:57 |
| Breath Volume | | 3.183 | 01:57 |
| Air Blank | | 0.000 | 01:58 |
| Control Test | | 0.078 | 01:58 |
| Air Blank | | 0.000 | 01:59 |
| Diagnostics Check | | OK | 01:59 |

Data Download Date/Time: 07/15/2020 05:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SWEAT,T
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: BUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 07/04/2020 20:03:57

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 20:01 | |
| Air Blank | 0.000 | 20:01 | |
| Control Test | 0.077 | 20:01 | |
| Air Blank | 0.000 | 20:02 | |
| Subject Sample | # 1 | 20:02 | Subject Test Refused |
| Breath Volume | 0.000 | 20:02 | |
| Air Blank | 0.000 | 20:03 | |
| Control Test | 0.078 | 20:03 | |
| Air Blank | 0.000 | 20:03 | |
| Diagnostics Check | OK | 20:03 | |

Data Download Date/Time: 07/15/2020 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 07/05/2020 03:35:21

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:28 |
| Air Blank | 0.000 | 03:29 |
| Control Test | 0.079 | 03:29 |
| Air Blank | 0.000 | 03:29 |
| Subject Sample # 1 | 0.078 | 03:31 |
| Breath Volume | 2.406 | 03:31 |
| Air Blank | 0.000 | 03:31 |
| Air Blank | 0.000 | 03:33 |
| Subject Sample # 2 | 0.073 | 03:33 |
| Breath Volume | 1.199 | 03:33 |
| Air Blank | 0.000 | 03:34 |
| Control Test | 0.078 | 03:34 |
| Air Blank | 0.000 | 03:35 |
| Diagnostics Check | OK | 03:35 |

Data Download Date/Time: 07/15/2020 05:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VANHOF,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 07/09/2020 00:21:49

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 00:18 | |
| Air Blank | 0.000 | 00:19 | |
| Control Test | 0.078 | 00:19 | |
| Air Blank | 0.000 | 00:20 | |
| Subject Sample | # 1 | 00:20 | Subject Test Refused |
| Breath Volume | 0.000 | 00:20 | |
| Air Blank | 0.000 | 00:20 | |
| Control Test | 0.078 | 00:21 | |
| Air Blank | 0.000 | 00:21 | |
| Diagnostics Check | OK | 00:21 | |

Data Download Date/Time: 07/15/2020 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: SANDERSON

DATE/TIME OF TEST: 07/10/2020 19:39:12

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 19:32 |
| Air Blank | 0.000 | 19:32 |
| Control Test | 0.078 | 19:32 |
| Air Blank | 0.000 | 19:33 |
| Subject Sample # 1 | 0.194 | 19:34 |
| Breath Volume | 1.574 | 19:34 |
| Air Blank | 0.000 | 19:34 |
| Air Blank | 0.000 | 19:36 |
| Subject Sample # 2 | 0.194 | 19:37 |
| Breath Volume | 1.976 | 19:37 |
| Air Blank | 0.000 | 19:38 |
| Control Test | 0.078 | 19:38 |
| Air Blank | 0.000 | 19:39 |
| Diagnostics Check | OK | 19:39 |

Data Download Date/Time: 07/15/2020 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/10/2020 23:49:30

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:42 |
| Air Blank | 0.000 | 23:42 |
| Control Test | 0.077 | 23:42 |
| Air Blank | 0.000 | 23:43 |
| Subject Sample # 1 | 0.113 | 23:44 |
| Breath Volume | 1.203 | 23:44 |
| Air Blank | 0.000 | 23:45 |
| Air Blank | 0.000 | 23:47 |
| Subject Sample # 2 | 0.110 | 23:48 |
| Breath Volume | 1.894 | 23:48 |
| Air Blank | 0.000 | 23:48 |
| Control Test | 0.077 | 23:48 |
| Air Blank | 0.000 | 23:49 |
| Diagnostics Check | OK | 23:49 |

Data Download Date/Time: 07/15/2020 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HALBRITTER, S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: DA SILVA

DATE/TIME OF TEST: 07/11/2020 00:57:02

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 00:54 | |
| Air Blank | 0.000 | 00:54 | |
| Control Test | 0.077 | 00:55 | |
| Air Blank | 0.000 | 00:55 | |
| Subject Sample | # 1 | 00:55 | Subject Test Refused |
| Breath Volume | 0.000 | 00:55 | |
| Air Blank | 0.000 | 00:56 | |
| Control Test | 0.078 | 00:56 | |
| Air Blank | 0.000 | 00:56 | |
| Diagnostics Check | OK | 00:57 | |

Data Download Date/Time: 07/15/2020 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN, C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: POWERS

DATE/TIME OF TEST: 07/11/2020 01:35:20

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:29 |
| Air Blank | 0.000 | 01:29 |
| Control Test | 0.078 | 01:29 |
| Air Blank | 0.000 | 01:30 |
| Subject Sample # 1 | 0.144 | 01:30 |
| Breath Volume | 2.023 | 01:30 |
| Air Blank | 0.000 | 01:31 |
| Air Blank | 0.000 | 01:33 |
| Subject Sample # 2 | 0.140 | 01:33 |
| Breath Volume | 3.710 | 01:33 |
| Air Blank | 0.000 | 01:34 |
| Control Test | 0.078 | 01:34 |
| Air Blank | 0.000 | 01:35 |
| Diagnostics Check | OK | 01:35 |

Data Download Date/Time: 07/15/2020 05:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WEST, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: FLORESTAL

DATE/TIME OF TEST: 07/11/2020 20:16:17

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 20:13 | |
| Air Blank | 0.000 | 20:13 | |
| Control Test | 0.078 | 20:14 | |
| Air Blank | 0.000 | 20:14 | |
| Subject Sample | # 1 | 20:14 | Subject Test Refused |
| Breath Volume | 0.000 | 20:14 | |
| Air Blank | 0.000 | 20:15 | |
| Control Test | 0.079 | 20:15 | |
| Air Blank | 0.000 | 20:16 | |
| Diagnostics Check | OK | 20:16 | |

Data Download Date/Time: 07/15/2020 05:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARRISH, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/12/2020 21:31:17

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 21:28 | |
| Air Blank | 0.000 | 21:28 | |
| Control Test | 0.078 | 21:29 | |
| Air Blank | 0.000 | 21:29 | |
| Subject Sample | # 1 | 21:29 | Subject Test Refused |
| Breath Volume | 0.000 | 21:29 | |
| Air Blank | 0.000 | 21:30 | |
| Control Test | 0.078 | 21:30 | |
| Air Blank | 0.000 | 21:31 | |
| Diagnostics Check | OK | 21:31 | |

Data Download Date/Time: 07/15/2020 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: LADSON

DATE/TIME OF TEST: 07/12/2020 23:55:04

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:47 |
| Air Blank | 0.000 | 23:48 |
| Control Test | 0.078 | 23:48 |
| Air Blank | 0.000 | 23:49 |
| Subject Sample # 1 | 0.000 | 23:50 |
| Breath Volume | 1.558 | 23:50 |
| Air Blank | 0.000 | 23:50 |
| Air Blank | 0.000 | 23:52 |
| Subject Sample # 2 | 0.000 | 23:53 |
| Breath Volume | 1.355 | 23:53 |
| Air Blank | 0.000 | 23:54 |
| Control Test | 0.078 | 23:54 |
| Air Blank | 0.000 | 23:54 |
| Diagnostics Check | OK | 23:55 |

Data Download Date/Time: 07/15/2020 05:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PROPPER, N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: BUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: DESARO

DATE/TIME OF TEST: 07/13/2020 01:30:57

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 01:28 | |
| Air Blank | 0.000 | 01:28 | |
| Control Test | 0.078 | 01:28 | |
| Air Blank | 0.000 | 01:29 | |
| Subject Sample | # 1 | 01:29 | Subject Test Refused |
| Breath Volume | 0.000 | 01:29 | |
| Air Blank | 0.000 | 01:30 | |
| Control Test | 0.079 | 01:30 | |
| Air Blank | 0.000 | 01:30 | |
| Diagnostics Check | OK | 01:30 | |

Data Download Date/Time: 07/15/2020 05:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ, M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 07/13/2020 02:31:36

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 02:28 | |
| Air Blank | 0.000 | 02:29 | |
| Control Test | 0.079 | 02:29 | |
| Air Blank | 0.000 | 02:30 | |
| Subject Sample | # 1 | 02:30 | Subject Test Refused |
| Breath Volume | 0.000 | 02:30 | |
| Air Blank | 0.000 | 02:30 | |
| Control Test | 0.079 | 02:31 | |
| Air Blank | 0.000 | 02:31 | |
| Diagnostics Check | OK | 02:31 | |

Data Download Date/Time: 07/15/2020 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: OCONNELL,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:31

SUBJECT'S LAST NAME: LANGEL

DATE/TIME OF TEST: 07/18/2020 07:54:34

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 07:51 | |
| Air Blank | 0.000 | 07:52 | |
| Control Test | 0.079 | 07:52 | |
| Air Blank | 0.000 | 07:53 | |
| Subject Sample | # 1 | 07:53 | Subject Test Refused |
| Breath Volume | 0.000 | 07:53 | |
| Air Blank | 0.000 | 07:53 | |
| Control Test | 0.078 | 07:53 | |
| Air Blank | 0.000 | 07:54 | |
| Diagnostics Check | OK | 07:54 | |

Data Download Date/Time: 07/30/2020 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 07/22/2020 00:09:14

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:03 |
| Air Blank | 0.000 | 00:03 |
| Control Test | 0.078 | 00:03 |
| Air Blank | 0.000 | 00:04 |
| Subject Sample # 1 | 0.119 | 00:04 |
| Breath Volume | 1.656 | 00:04 |
| Air Blank | 0.000 | 00:05 |
| Air Blank | 0.000 | 00:07 |
| Subject Sample # 2 | 0.116 | 00:07 |
| Breath Volume | 3.187 | 00:07 |
| Air Blank | 0.000 | 00:08 |
| Control Test | 0.078 | 00:08 |
| Air Blank | 0.000 | 00:09 |
| Diagnostics Check | OK | 00:09 |

Data Download Date/Time: 07/30/2020 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS, TYLER, K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: ZAWACKI

DATE/TIME OF TEST: 07/22/2020 01:50:46

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 01:47 | |
| Air Blank | 0.000 | 01:48 | |
| Control Test | 0.079 | 01:48 | |
| Air Blank | 0.000 | 01:48 | |
| Subject Sample | # 1 | 01:49 | Subject Test Refused |
| Breath Volume | 0.000 | 01:49 | |
| Air Blank | 0.000 | 01:49 | |
| Control Test | 0.078 | 01:50 | |
| Air Blank | 0.000 | 01:50 | |
| Diagnostics Check | OK | 01:50 | |

Data Download Date/Time: 07/30/2020 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS, TYLER, K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARRISH, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: VICCHRILLI

DATE/TIME OF TEST: 07/22/2020 20:27:38

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:21 |
| Air Blank | 0.000 | 20:21 |
| Control Test | 0.078 | 20:22 |
| Air Blank | 0.000 | 20:22 |
| Subject Sample # 1 | 0.000 | 20:23 |
| Breath Volume | 3.480 | 20:23 |
| Air Blank | 0.000 | 20:23 |
| Air Blank | 0.000 | 20:25 |
| Subject Sample # 2 | 0.000 | 20:26 |
| Breath Volume | 3.656 | 20:26 |
| Air Blank | 0.000 | 20:26 |
| Control Test | 0.078 | 20:27 |
| Air Blank | 0.000 | 20:27 |
| Diagnostics Check | OK | 20:27 |

Data Download Date/Time: 07/30/2020 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS, TYLER, K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROE, B
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 07/24/2020 06:02:03

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 05:54 |
| Air Blank | | 0.000 | 05:55 |
| Control Test | | 0.079 | 05:55 |
| Air Blank | | 0.000 | 05:55 |
| Subject Sample | # 1 | 0.162 | 05:57 |
| Breath Volume | | 2.058 | 05:57 |
| Air Blank | | 0.000 | 05:58 |
| Air Blank | | 0.000 | 06:00 |
| Subject Sample | # 2 | 0.157 | 06:00 |
| Breath Volume | | 2.316 | 06:00 |
| Air Blank | | 0.000 | 06:01 |
| Control Test | | 0.078 | 06:01 |
| Air Blank | | 0.000 | 06:01 |
| Diagnostics Check | | OK | 06:02 |

Data Download Date/Time: 07/30/2020 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: DARR

DATE/TIME OF TEST: 07/27/2020 23:26:41

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 23:20 |
| Air Blank | | 0.000 | 23:20 |
| Control Test | | 0.078 | 23:21 |
| Air Blank | | 0.000 | 23:21 |
| Subject Sample | # 1 | 0.277 | 23:22 |
| Breath Volume | | 1.375 | 23:22 |
| Air Blank | | 0.000 | 23:22 |
| Air Blank | | 0.000 | 23:24 |
| Subject Sample | # 2 | 0.282 | 23:25 |
| Breath Volume | | 1.574 | 23:25 |
| Air Blank | | 0.000 | 23:25 |
| Control Test | | 0.078 | 23:26 |
| Air Blank | | 0.000 | 23:26 |
| Diagnostics Check | | OK | 23:26 |

Data Download Date/Time: 07/30/2020 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: TIMMRECK

DATE/TIME OF TEST: 07/28/2020 04:18:34

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 04:12 |
| Air Blank | | 0.000 | 04:12 |
| Control Test | | 0.078 | 04:13 |
| Air Blank | | 0.000 | 04:13 |
| Subject Sample | # 1 | 0.198 | 04:14 |
| Breath Volume | | 2.695 | 04:14 |
| Air Blank | | 0.000 | 04:14 |
| Air Blank | | 0.000 | 04:16 |
| Subject Sample | # 2 | 0.201 | 04:17 |
| Breath Volume | | 2.621 | 04:17 |
| Air Blank | | 0.000 | 04:17 |
| Control Test | | 0.078 | 04:17 |
| Air Blank | | 0.000 | 04:18 |
| Diagnostics Check | | OK | 04:18 |

Data Download Date/Time: 07/30/2020 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: KULBACKI

DATE/TIME OF TEST: 07/29/2020 01:00:38

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:54 |
| Air Blank | 0.000 | 00:54 |
| Control Test | 0.078 | 00:55 |
| Air Blank | 0.000 | 00:55 |
| Subject Sample # 1 | 0.114 | 00:56 |
| Breath Volume | 1.832 | 00:56 |
| Air Blank | 0.000 | 00:56 |
| Air Blank | 0.000 | 00:58 |
| Subject Sample # 2 | 0.106 | 00:59 |
| Breath Volume | 1.972 | 00:59 |
| Air Blank | 0.000 | 00:59 |
| Control Test | 0.078 | 01:00 |
| Air Blank | 0.000 | 01:00 |
| Diagnostics Check | OK | 01:00 |

Data Download Date/Time: 07/30/2020 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 07/30/2020 00:33:54

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:27 |
| Air Blank | 0.000 | 00:28 |
| Control Test | 0.078 | 00:28 |
| Air Blank | 0.000 | 00:29 |
| Subject Sample # 1 | 0.133 | 00:29 |
| Breath Volume | 2.183 | 00:29 |
| Air Blank | 0.000 | 00:30 |
| Air Blank | 0.000 | 00:32 |
| Subject Sample # 2 | 0.131 | 00:32 |
| Breath Volume | 2.238 | 00:32 |
| Air Blank | 0.000 | 00:33 |
| Control Test | 0.078 | 00:33 |
| Air Blank | 0.000 | 00:33 |
| Diagnostics Check | OK | 00:33 |

Data Download Date/Time: 07/30/2020 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/30/2020 03:05:17

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:59 |
| Air Blank | 0.000 | 02:59 |
| Control Test | 0.078 | 02:59 |
| Air Blank | 0.000 | 03:00 |
| Subject Sample # 1 | 0.113 | 03:00 |
| Breath Volume | 2.246 | 03:00 |
| Air Blank | 0.000 | 03:01 |
| Air Blank | 0.000 | 03:03 |
| Subject Sample # 2 | 0.112 | 03:03 |
| Breath Volume | 1.949 | 03:03 |
| Air Blank | 0.000 | 03:04 |
| Control Test | 0.078 | 03:04 |
| Air Blank | 0.000 | 03:05 |
| Diagnostics Check | OK | 03:05 |

Data Download Date/Time: 07/30/2020 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 07/30/2020 23:35:20

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:28 |
| Air Blank | 0.000 | 23:29 |
| Control Test | 0.078 | 23:29 |
| Air Blank | 0.000 | 23:30 |
| Subject Sample # 1 | 0.221 | 23:30 |
| Breath Volume | 1.488 | 23:30 |
| Air Blank | 0.000 | 23:31 |
| Air Blank | 0.000 | 23:33 |
| Subject Sample # 2 | 0.219 | 23:33 |
| Breath Volume | 1.593 | 23:33 |
| Air Blank | 0.000 | 23:34 |
| Control Test | 0.078 | 23:34 |
| Air Blank | 0.000 | 23:35 |
| Diagnostics Check | OK | 23:35 |

Data Download Date/Time: 08/14/2020 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHOWDHURY,P
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: AHSAN BASHIR

DATE/TIME OF TEST: 07/31/2020 19:27:51

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 19:20 |
| Air Blank | 0.000 | 19:20 |
| Control Test | 0.077 | 19:21 |
| Air Blank | 0.000 | 19:21 |
| Subject Sample # 1 | 0.180 | 19:22 |
| Breath Volume | 1.242 | 19:22 |
| Air Blank | 0.000 | 19:23 |
| Air Blank | 0.000 | 19:25 |
| Subject Sample # 2 | 0.186 | 19:26 |
| Breath Volume | 1.718 | 19:26 |
| Air Blank | 0.000 | 19:26 |
| Control Test | 0.077 | 19:27 |
| Air Blank | 0.000 | 19:27 |
| Diagnostics Check | OK | 19:27 |

Data Download Date/Time: 08/14/2020 08:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 07/31/2020 22:41:13

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 22:38 | |
| Air Blank | 0.000 | 22:38 | |
| Control Test | 0.077 | 22:39 | |
| Air Blank | 0.000 | 22:39 | |
| Subject Sample | # 1 | 22:39 | Subject Test Refused |
| Breath Volume | 0.000 | 22:39 | |
| Air Blank | 0.000 | 22:40 | |
| Control Test | 0.078 | 22:40 | |
| Air Blank | 0.000 | 22:41 | |
| Diagnostics Check | OK | 22:41 | |

Data Download Date/Time: 08/14/2020 08:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JONES,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: BRIGNER

DATE/TIME OF TEST: 07/19/2020 18:19:12

| | | | |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK | 18:17 | |
| Air Blank | 0.000 | 18:18 | |
| Control Test | 0.077 | 18:18 | |
| Air Blank | | 18:18 | RFI Detect |
| Air Blank | 0.000 | 18:19 | |

Data Download Date/Time: 07/29/2020 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JONES,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: BRIGNER

DATE/TIME OF TEST: 07/19/2020 18:29:48

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 18:23 |
| Air Blank | 0.000 | 18:23 |
| Control Test | 0.077 | 18:24 |
| Air Blank | 0.000 | 18:24 |
| Subject Sample # 1 | 0.138 | 18:25 |
| Breath Volume | 2.593 | 18:25 |
| Air Blank | 0.000 | 18:25 |
| Air Blank | 0.000 | 18:27 |
| Subject Sample # 2 | 0.141 | 18:28 |
| Breath Volume | 3.546 | 18:28 |
| Air Blank | 0.000 | 18:28 |
| Control Test | 0.077 | 18:29 |
| Air Blank | 0.000 | 18:29 |
| Diagnostics Check | OK | 18:29 |

Data Download Date/Time: 07/29/2020 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: AMBS

DATE/TIME OF TEST: 07/24/2020 22:52:20

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:45 |
| Air Blank | 0.000 | 22:46 |
| Control Test | 0.077 | 22:46 |
| Air Blank | 0.000 | 22:46 |
| Subject Sample # 1 | 0.121 | 22:47 |
| Breath Volume | 3.066 | 22:47 |
| Air Blank | 0.000 | 22:48 |
| Air Blank | 0.000 | 22:50 |
| Subject Sample # 2 | 0.117 | 22:50 |
| Breath Volume | 3.265 | 22:50 |
| Air Blank | 0.000 | 22:51 |
| Control Test | 0.078 | 22:51 |
| Air Blank | 0.000 | 22:52 |
| Diagnostics Check | OK | 22:52 |

Data Download Date/Time: 07/29/2020 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: OCONNELL,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: CRAIG

DATE/TIME OF TEST: 07/25/2020 05:54:34

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 05:47 |
| Air Blank | 0.000 | 05:48 |
| Control Test | 0.077 | 05:48 |
| Air Blank | 0.000 | 05:48 |
| Subject Sample # 1 | 0.170 | 05:49 |
| Breath Volume | 1.386 | 05:49 |
| Air Blank | 0.000 | 05:50 |
| Air Blank | 0.000 | 05:52 |
| Subject Sample # 2 | 0.175 | 05:53 |
| Breath Volume | 2.554 | 05:53 |
| Air Blank | 0.000 | 05:53 |
| Control Test | 0.078 | 05:53 |
| Air Blank | 0.000 | 05:54 |
| Diagnostics Check | OK | 05:54 |

Data Download Date/Time: 07/29/2020 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VINTI,STEPHEN,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: VINTI

DATE/TIME OF TEST: 07/25/2020 20:29:18

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:20 |
| Air Blank | 0.000 | 20:20 |
| Control Test | 0.077 | 20:21 |
| Air Blank | 0.000 | 20:21 |
| Subject Sample # 1 | 0.234 | 20:23 |
| Breath Volume | 1.281 | 20:23 |
| Air Blank | 0.000 | 20:24 |
| Air Blank | 0.000 | 20:26 |
| Subject Sample # 2 | 0.233 | 20:27 |
| Breath Volume | 1.671 | 20:27 |
| Air Blank | 0.000 | 20:28 |
| Control Test | 0.075 | 20:28 |
| Air Blank | 0.000 | 20:29 |
| Diagnostics Check | OK | 20:29 |

Data Download Date/Time: 07/29/2020 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SNELLGROVE

DATE/TIME OF TEST: 07/26/2020 04:27:26

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 04:20 |
| Air Blank | 0.000 | 04:21 |
| Control Test | 0.078 | 04:21 |
| Air Blank | 0.000 | 04:22 |
| Subject Sample # 1 | 0.057 | 04:22 |
| Breath Volume | 2.769 | 04:22 |
| Air Blank | 0.000 | 04:23 |
| Air Blank | 0.000 | 04:25 |
| Subject Sample # 2 | 0.055 | 04:25 |
| Breath Volume | 2.066 | 04:25 |
| Air Blank | 0.000 | 04:26 |
| Control Test | 0.077 | 04:26 |
| Air Blank | 0.000 | 04:27 |
| Diagnostics Check | OK | 04:27 |

Data Download Date/Time: 07/29/2020 14:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MOTEN

DATE/TIME OF TEST: 07/27/2020 01:24:03

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 01:15 |
| Air Blank | | 0.000 | 01:16 |
| Control Test | | 0.077 | 01:16 |
| Air Blank | | 0.000 | 01:17 |
| Subject Sample | # 1 | 0.163 | 01:18 |
| Breath Volume | | 1.437 | 01:18 |
| Air Blank | | 0.000 | 01:18 |
| Air Blank | | 0.000 | 01:20 |
| Subject Sample | # 2 | 0.163 | 01:22 |
| Breath Volume | | 2.511 | 01:22 |
| Air Blank | | 0.000 | 01:23 |
| Control Test | | 0.076 | 01:23 |
| Air Blank | | 0.000 | 01:23 |
| Diagnostics Check | | OK | 01:24 |

Data Download Date/Time: 07/29/2020 14:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS, JAKE, M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ADKINS, JAKE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: CIFUENTES MORALES

DATE/TIME OF TEST: 07/17/2020 21:20:34

| | | | |
|--------------------|-------|-------|----------------|
| Diagnostics Check | OK | 21:04 | |
| Air Blank | 0.000 | 21:05 | |
| Control Test | 0.079 | 21:05 | |
| Air Blank | 0.000 | 21:06 | |
| Subject Sample # 1 | 0.087 | 21:09 | |
| Breath Volume | 1.449 | 21:09 | |
| Air Blank | 0.000 | 21:09 | |
| Air Blank | 0.000 | 21:11 | |
| Subject Sample # 2 | 0.063 | 21:14 | Volume Not Met |
| Breath Volume | 0.035 | 21:14 | |
| Air Blank | 0.000 | 21:15 | |
| Air Blank | 0.000 | 21:17 | |
| Subject Sample # 3 | 0.075 | 21:19 | |
| Breath Volume | 1.976 | 21:19 | |
| Air Blank | 0.000 | 21:19 | |
| Control Test | 0.078 | 21:19 | |
| Air Blank | 0.000 | 21:20 | |
| Diagnostics Check | OK | 21:20 | |

Data Download Date/Time: 09/04/2020 00:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAEYAERT,JEFF,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: NAEYAERT,JEFF
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 07/15/2020 22:42:43

| | | | |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK | 22:41 | |
| Air Blank | 0.000 | 22:41 | |
| Control Test | 0.079 | 22:41 | |
| Air Blank | | 22:42 | RFI Detect |
| Air Blank | 0.000 | 22:42 | |

Data Download Date/Time: 08/04/2020 15:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAEYAERT,JEFF,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: NAEYAERT,JEFF
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 07/15/2020 22:57:23

| | | | |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK | 22:44 | |
| Air Blank | 0.000 | 22:44 | |
| Control Test | 0.079 | 22:45 | |
| Air Blank | 0.000 | 22:45 | |
| Subject Sample | # 1 0.150 | 22:46 | |
| Breath Volume | 1.527 | 22:46 | |
| Air Blank | 0.000 | 22:46 | |
| Air Blank | 0.000 | 22:48 | |
| Subject Sample | # 2 0.159 | 22:51 | Volume Not Met |
| Breath Volume | 0.226 | 22:51 | |
| Air Blank | 0.000 | 22:52 | |
| Air Blank | 0.000 | 22:54 | |
| Subject Sample | # 3 0.162 | 22:55 | |
| Breath Volume | 3.093 | 22:55 | |
| Air Blank | 0.000 | 22:56 | |
| Control Test | 0.078 | 22:56 | |
| Air Blank | 0.000 | 22:57 | |
| Diagnostics Check | OK | 22:57 | |

Data Download Date/Time: 08/04/2020 15:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,KEVIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FROST,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 07/28/2020 18:55:45

| | | | |
|-------------------|----|-------|------------|
| Diagnostics Check | OK | 18:55 | |
| Air Blank | | 18:55 | RFI Detect |
| Air Blank | | 18:55 | RFI Detect |

Data Download Date/Time: 08/04/2020 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,KEVIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FROST,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 07/28/2020 19:01:52

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 18:55 |
| Air Blank | 0.000 | 18:56 |
| Control Test | 0.081 | 18:56 |
| Air Blank | 0.000 | 18:57 |
| Subject Sample # 1 | 0.099 | 18:57 |
| Breath Volume | 2.199 | 18:57 |
| Air Blank | 0.000 | 18:58 |
| Air Blank | 0.000 | 19:00 |
| Subject Sample # 2 | 0.090 | 19:00 |
| Breath Volume | 2.023 | 19:00 |
| Air Blank | 0.000 | 19:00 |
| Control Test | 0.081 | 19:01 |
| Air Blank | 0.000 | 19:01 |
| Diagnostics Check | OK | 19:01 |

Data Download Date/Time: 08/04/2020 14:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMPKINS,DANIEL,H
BREATH TEST OPERATOR AGENCY: Lawtey PD
ARRESTING OFFICER: WARD,G
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/07/2020 23:18:28

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 23:09 | |
| Air Blank | 0.000 | 23:10 | |
| Control Test | 0.078 | 23:10 | |
| Air Blank | 0.000 | 23:11 | |
| Subject Sample | # 1 | 23:14 | No Sample Provided |
| Breath Volume | 0.000 | 23:14 | |
| Air Blank | 0.000 | 23:14 | |
| Air Blank | 0.000 | 23:16 | |
| Subject Sample | # 2 | 23:17 | Subject Test Refused |
| Breath Volume | 0.000 | 23:17 | |
| Air Blank | 0.000 | 23:17 | |
| Control Test | 0.077 | 23:17 | |
| Air Blank | 0.000 | 23:18 | |
| Diagnostics Check | OK | 23:18 | |

Data Download Date/Time: 08/01/2020 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONKEL,SCOTT,M
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: HULLENDER,ROBERT
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 13:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:42

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 07/13/2020 14:32:02

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 14:26 |
| Air Blank | 0.000 | 14:26 |
| Control Test | 0.079 | 14:26 |
| Air Blank | 0.000 | 14:27 |
| Subject Sample # 1 | 0.000 | 14:27 |
| Breath Volume | 3.183 | 14:27 |
| Air Blank | 0.000 | 14:28 |
| Air Blank | 0.000 | 14:30 |
| Subject Sample # 2 | 0.000 | 14:30 |
| Breath Volume | 3.320 | 14:30 |
| Air Blank | 0.000 | 14:31 |
| Control Test | 0.078 | 14:31 |
| Air Blank | 0.000 | 14:31 |
| Diagnostics Check | OK | 14:32 |

Data Download Date/Time: 08/01/2020 09:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUMNER,DALTON,R
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: RAULERSON,JAY
ARRESTING OFFICER AGENCY: Lawtey PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ABIYOW

DATE/TIME OF TEST: 07/13/2020 22:10:28

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:03 |
| Air Blank | 0.000 | 22:04 |
| Control Test | 0.079 | 22:04 |
| Air Blank | 0.000 | 22:04 |
| Subject Sample # 1 | 0.206 | 22:05 |
| Breath Volume | 1.144 | 22:05 |
| Air Blank | 0.000 | 22:06 |
| Air Blank | 0.000 | 22:08 |
| Subject Sample # 2 | 0.212 | 22:08 |
| Breath Volume | 1.402 | 22:08 |
| Air Blank | 0.000 | 22:09 |
| Control Test | 0.077 | 22:09 |
| Air Blank | 0.000 | 22:10 |
| Diagnostics Check | OK | 22:10 |

Data Download Date/Time: 08/01/2020 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMS,KYLE,N
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SIMMS,KYLE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: ALAMRI

DATE/TIME OF TEST: 07/30/2020 22:12:44

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:05 |
| Air Blank | 0.000 | 22:06 |
| Control Test | 0.080 | 22:06 |
| Air Blank | 0.000 | 22:07 |
| Subject Sample # 1 | 0.000 | 22:08 |
| Breath Volume | 2.441 | 22:08 |
| Air Blank | 0.000 | 22:08 |
| Air Blank | 0.000 | 22:10 |
| Subject Sample # 2 | 0.000 | 22:11 |
| Breath Volume | 2.625 | 22:11 |
| Air Blank | 0.000 | 22:11 |
| Control Test | 0.079 | 22:12 |
| Air Blank | 0.000 | 22:12 |
| Diagnostics Check | OK | 22:12 |

Data Download Date/Time: 08/31/2020 20:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERRILL,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MERRILL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: RUTLEDGE

DATE/TIME OF TEST: 07/31/2020 00:06:29

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 00:02 | |
| Air Blank | 0.000 | 00:02 | |
| Control Test | 0.079 | 00:02 | |
| Air Blank | 0.000 | 00:03 | |
| Subject Sample | # 1 | 00:05 | Subject Test Refused |
| Breath Volume | 0.000 | 00:05 | |
| Air Blank | 0.000 | 00:05 | |
| Control Test | 0.080 | 00:05 | |
| Air Blank | 0.000 | 00:06 | |
| Diagnostics Check | OK | 00:06 | |

Data Download Date/Time: 08/31/2020 20:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,DANIEL,K
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DAVIS,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: EVERETT

DATE/TIME OF TEST: 07/31/2020 01:23:13

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 01:11 | |
| Air Blank | | 0.000 | 01:12 | |
| Control Test | | 0.079 | 01:12 | |
| Air Blank | | 0.000 | 01:12 | |
| Subject Sample | # 1 | 0.133 | 01:16 | Volume Not Met |
| Breath Volume | | 0.449 | 01:16 | |
| Air Blank | | 0.000 | 01:16 | |
| Air Blank | | 0.000 | 01:18 | |
| Subject Sample | # 2 | 0.118 | 01:21 | Volume Not Met |
| Breath Volume | | 0.734 | 01:21 | |
| Air Blank | | 0.000 | 01:22 | |
| Control Test | | 0.076 | 01:22 | |
| Air Blank | | 0.000 | 01:23 | |
| Diagnostics Check | | OK | 01:23 | |

Data Download Date/Time: 08/31/2020 20:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JAMES,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DAVIS,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: HODGES

DATE/TIME OF TEST: 07/31/2020 23:47:29

| | | | |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK | 23:36 | |
| Air Blank | 0.000 | 23:36 | |
| Control Test | 0.079 | 23:36 | |
| Air Blank | 0.000 | 23:37 | |
| Subject Sample | # 1 | 23:40 | No Sample Provided |
| Breath Volume | 0.000 | 23:40 | |
| Air Blank | 0.000 | 23:40 | |
| Air Blank | 0.000 | 23:42 | |
| Subject Sample | # 2 | 23:46 | No Sample Provided |
| Breath Volume | 0.000 | 23:46 | |
| Air Blank | 0.000 | 23:46 | |
| Control Test | 0.079 | 23:46 | |
| Air Blank | 0.000 | 23:47 | |
| Diagnostics Check | OK | 23:47 | |

Data Download Date/Time: 08/31/2020 20:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWELL,KEENAN,B
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: HOWELL,DAVID
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 07/10/2020 21:38:19

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 21:34 | |
| Air Blank | 0.000 | 21:34 | |
| Control Test | 0.080 | 21:34 | |
| Air Blank | 0.000 | 21:35 | |
| Subject Sample | # 1 | 21:36 | Subject Test Refused |
| Breath Volume | 0.000 | 21:36 | |
| Air Blank | 0.000 | 21:37 | |
| Control Test | 0.080 | 21:37 | |
| Air Blank | 0.000 | 21:38 | |
| Diagnostics Check | OK | 21:38 | |

Data Download Date/Time: 08/03/2020 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASSALLO,STEVEN,A
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: SCARBOROUGH,TYLER
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 13:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:14

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/29/2020 14:14:46

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 14:06 |
| Air Blank | 0.000 | 14:06 |
| Control Test | 0.080 | 14:07 |
| Air Blank | 0.000 | 14:07 |
| Subject Sample # 1 | 0.084 | 14:09 |
| Breath Volume | 1.156 | 14:09 |
| Air Blank | 0.000 | 14:10 |
| Air Blank | 0.000 | 14:11 |
| Subject Sample # 2 | 0.077 | 14:13 |
| Breath Volume | 1.152 | 14:13 |
| Air Blank | 0.000 | 14:13 |
| Control Test | 0.078 | 14:14 |
| Air Blank | 0.000 | 14:14 |
| Diagnostics Check | OK | 14:14 |

Data Download Date/Time: 08/03/2020 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 07/08/2020 19:12:50

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 19:08 | |
| Air Blank | 0.000 | 19:08 | |
| Control Test | 0.078 | 19:09 | |
| Air Blank | 0.000 | 19:09 | |
| Subject Sample | # 1 | 19:11 | Subject Test Refused |
| Breath Volume | 0.000 | 19:11 | |
| Air Blank | 0.000 | 19:11 | |
| Control Test | 0.079 | 19:12 | |
| Air Blank | 0.000 | 19:12 | |
| Diagnostics Check | OK | 19:12 | |

Data Download Date/Time: 08/03/2020 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 07/08/2020 19:28:46

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 19:24 | |
| Air Blank | 0.000 | 19:24 | |
| Control Test | 0.080 | 19:25 | |
| Air Blank | 0.000 | 19:25 | |
| Subject Sample | # 1 | 19:27 | Subject Test Refused |
| Breath Volume | 0.757 | 19:27 | |
| Air Blank | 0.000 | 19:27 | |
| Control Test | 0.080 | 19:28 | |
| Air Blank | 0.000 | 19:28 | |
| Diagnostics Check | OK | 19:28 | |

Data Download Date/Time: 08/03/2020 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 07/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: LEKARCZYK

DATE/TIME OF TEST: 07/26/2020 20:58:52

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:52 |
| Air Blank | 0.000 | 20:52 |
| Control Test | 0.080 | 20:53 |
| Air Blank | 0.000 | 20:53 |
| Subject Sample # 1 | 0.177 | 20:54 |
| Breath Volume | 2.023 | 20:54 |
| Air Blank | 0.000 | 20:54 |
| Air Blank | 0.000 | 20:56 |
| Subject Sample # 2 | 0.175 | 20:57 |
| Breath Volume | 2.898 | 20:57 |
| Air Blank | 0.000 | 20:57 |
| Control Test | 0.079 | 20:58 |
| Air Blank | 0.000 | 20:58 |
| Diagnostics Check | OK | 20:58 |

Data Download Date/Time: 08/03/2020 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LLANES,GABRIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MYERS,DAVID
ARRESTING OFFICER AGENCY: Madison County SO
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: FERRELL

DATE/TIME OF TEST: 07/11/2020 20:36:43

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 20:31 | |
| Air Blank | 0.000 | 20:31 | |
| Control Test | 0.078 | 20:32 | |
| Air Blank | 0.000 | 20:32 | |
| Subject Sample | # 1 | 20:35 | Subject Test Refused |
| Breath Volume | 0.000 | 20:35 | |
| Air Blank | 0.000 | 20:35 | |
| Control Test | 0.078 | 20:36 | |
| Air Blank | 0.000 | 20:36 | |
| Diagnostics Check | OK | 20:36 | |

Data Download Date/Time: 08/05/2020 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,JOSHUA,N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WRIGHT,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:47

SUBJECT'S LAST NAME: CONRAD

DATE/TIME OF TEST: 07/12/2020 09:15:05

| | | | |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK | 09:13 | |
| Air Blank | 0.000 | 09:13 | |
| Control Test | 0.078 | 09:14 | |
| Air Blank | | 09:14 | RFI Detect |
| Air Blank | 0.000 | 09:15 | |

Data Download Date/Time: 08/05/2020 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT, JOSHUA, N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WRIGHT, JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:47

SUBJECT'S LAST NAME: CONRAD

DATE/TIME OF TEST: 07/12/2020 09:19:47

| | | | |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK | 09:16 | |
| Air Blank | 0.000 | 09:17 | |
| Control Test | 0.078 | 09:17 | |
| Air Blank | 0.000 | 09:18 | |
| Subject Sample | # 1 | 09:18 | RFI Detect |
| Breath Volume | 0.000 | 09:18 | |
| Air Blank | 0.000 | 09:18 | |
| Control Test | 0.078 | 09:19 | |
| Air Blank | | 09:19 | RFI Detect |
| Air Blank | 0.000 | 09:19 | |

Data Download Date/Time: 08/05/2020 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,JOSHUA,N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WRIGHT,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:47

SUBJECT'S LAST NAME: CONRAD

DATE/TIME OF TEST: 07/12/2020 09:21:48

| | | | |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK | 09:21 | |
| Air Blank | | 09:21 | RFI Detect |
| Air Blank | 0.000 | 09:21 | |

Data Download Date/Time: 08/05/2020 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,JOSHUA,N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WRIGHT,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:47

SUBJECT'S LAST NAME: CONRAD

DATE/TIME OF TEST: 07/12/2020 09:29:56

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 09:23 |
| Air Blank | 0.000 | 09:23 |
| Control Test | 0.078 | 09:24 |
| Air Blank | 0.000 | 09:24 |
| Subject Sample # 1 | 0.068 | 09:25 |
| Breath Volume | 5.128 | 09:25 |
| Air Blank | 0.000 | 09:25 |
| Air Blank | 0.000 | 09:27 |
| Subject Sample # 2 | 0.067 | 09:28 |
| Breath Volume | 5.222 | 09:28 |
| Air Blank | 0.000 | 09:29 |
| Control Test | 0.077 | 09:29 |
| Air Blank | 0.000 | 09:29 |
| Diagnostics Check | OK | 09:29 |

Data Download Date/Time: 08/05/2020 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AKINS,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Madison County SO
ARRESTING OFFICER: SANDERS,ROBERT
ARRESTING OFFICER AGENCY: Madison PD
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 07/19/2020 00:54:39

| | | | |
|-------------------|-------|-------|---------------|
| Diagnostics Check | OK | 00:51 | |
| Air Blank | 0.000 | 00:51 | |
| Control Test | 0.078 | 00:51 | |
| Air Blank | 0.000 | 00:52 | |
| Subject Sample | # 1 | 00:53 | Slope Not Met |
| Breath Volume | 1.574 | 00:53 | |
| Air Blank | 0.000 | 00:53 | |
| Control Test | 0.079 | 00:54 | |
| Air Blank | 0.000 | 00:54 | |
| Diagnostics Check | OK | 00:54 | |

Data Download Date/Time: 08/05/2020 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LLANES,GABRIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: LLANES,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: WILCOX THOMPSON

DATE/TIME OF TEST: 07/24/2020 23:34:00

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:27 |
| Air Blank | 0.000 | 23:28 |
| Control Test | 0.078 | 23:28 |
| Air Blank | 0.000 | 23:28 |
| Subject Sample # 1 | 0.171 | 23:29 |
| Breath Volume | 2.457 | 23:29 |
| Air Blank | 0.000 | 23:30 |
| Air Blank | 0.000 | 23:32 |
| Subject Sample # 2 | 0.170 | 23:32 |
| Breath Volume | 2.125 | 23:32 |
| Air Blank | 0.000 | 23:33 |
| Control Test | 0.077 | 23:33 |
| Air Blank | 0.000 | 23:33 |
| Diagnostics Check | OK | 23:34 |

Data Download Date/Time: 08/05/2020 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AKINS,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Madison County SO
ARRESTING OFFICER: BRANNON,EDDIE
ARRESTING OFFICER AGENCY: Madison County SO
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/31/2020 23:59:58

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:53 |
| Air Blank | 0.000 | 23:53 |
| Control Test | 0.079 | 23:54 |
| Air Blank | 0.000 | 23:54 |
| Subject Sample # 1 | 0.153 | 23:55 |
| Breath Volume | 2.417 | 23:55 |
| Air Blank | 0.000 | 23:55 |
| Air Blank | 0.000 | 23:57 |
| Subject Sample # 2 | 0.149 | 23:58 |
| Breath Volume | 2.480 | 23:58 |
| Air Blank | 0.000 | 23:59 |
| Control Test | 0.077 | 23:59 |
| Air Blank | 0.000 | 23:59 |
| Diagnostics Check | OK | 23:59 |

Data Download Date/Time: 08/05/2020 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,BRETT,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WAGNER,BRETT
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 07/02/2020 05:59:22

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 05:52 |
| Air Blank | 0.000 | 05:52 |
| Control Test | 0.081 | 05:53 |
| Air Blank | 0.000 | 05:53 |
| Subject Sample # 1 | 0.239 | 05:54 |
| Breath Volume | 1.488 | 05:54 |
| Air Blank | 0.000 | 05:55 |
| Air Blank | 0.000 | 05:56 |
| Subject Sample # 2 | 0.251 | 05:57 |
| Breath Volume | 1.882 | 05:57 |
| Air Blank | 0.000 | 05:58 |
| Control Test | 0.082 | 05:58 |
| Air Blank | 0.000 | 05:59 |
| Diagnostics Check | OK | 05:59 |

Data Download Date/Time: 07/30/2020 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOODWIN,NANETTE,M
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: MORRISON,ALLISON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BOLE

DATE/TIME OF TEST: 07/15/2020 00:11:47

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:05 |
| Air Blank | 0.000 | 00:06 |
| Control Test | 0.081 | 00:06 |
| Air Blank | 0.000 | 00:06 |
| Subject Sample # 1 | 0.125 | 00:07 |
| Breath Volume | 3.519 | 00:07 |
| Air Blank | 0.000 | 00:08 |
| Air Blank | 0.000 | 00:09 |
| Subject Sample # 2 | 0.119 | 00:10 |
| Breath Volume | 2.656 | 00:10 |
| Air Blank | 0.000 | 00:10 |
| Control Test | 0.080 | 00:11 |
| Air Blank | 0.000 | 00:11 |
| Diagnostics Check | OK | 00:11 |

Data Download Date/Time: 07/30/2020 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRELL, CHAD, A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: DELANEY, TYLER
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: BROTHERS

DATE/TIME OF TEST: 07/27/2020 01:30:11

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 01:27 | |
| Air Blank | 0.000 | 01:27 | |
| Control Test | 0.080 | 01:28 | |
| Air Blank | 0.000 | 01:28 | |
| Subject Sample | # 1 | 01:28 | Subject Test Refused |
| Breath Volume | 0.000 | 01:28 | |
| Air Blank | 0.000 | 01:29 | |
| Control Test | 0.080 | 01:29 | |
| Air Blank | 0.000 | 01:30 | |
| Diagnostics Check | OK | 01:30 | |

Data Download Date/Time: 07/30/2020 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRELL,CHAD,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: ALLEN,FORREST
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 06/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 07/29/2020 21:45:21

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:38 |
| Air Blank | 0.000 | 21:39 |
| Control Test | 0.081 | 21:39 |
| Air Blank | 0.000 | 21:40 |
| Subject Sample # 1 | 0.142 | 21:40 |
| Breath Volume | 3.593 | 21:40 |
| Air Blank | 0.000 | 21:41 |
| Air Blank | 0.000 | 21:43 |
| Subject Sample # 2 | 0.144 | 21:43 |
| Breath Volume | 3.203 | 21:43 |
| Air Blank | 0.000 | 21:44 |
| Control Test | 0.080 | 21:44 |
| Air Blank | 0.000 | 21:45 |
| Diagnostics Check | OK | 21:45 |

Data Download Date/Time: 07/30/2020 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT, MARK, A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: DOREUS, W
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:17

SUBJECT'S LAST NAME: TRUMPP

DATE/TIME OF TEST: 07/13/2020 06:41:34

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 06:37 | |
| Air Blank | 0.000 | 06:37 | |
| Control Test | 0.080 | 06:37 | |
| Air Blank | 0.000 | 06:38 | |
| Subject Sample | # 1 | 06:40 | Subject Test Refused |
| Breath Volume | 0.019 | 06:40 | |
| Air Blank | 0.000 | 06:40 | |
| Control Test | 0.080 | 06:40 | |
| Air Blank | 0.000 | 06:41 | |
| Diagnostics Check | OK | 06:41 | |

Data Download Date/Time: 07/23/2020 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARK,A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: WENRICK,N
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 06/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 07/14/2020 20:14:59

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 20:01 |
| Air Blank | | 0.000 | 20:01 |
| Control Test | | 0.079 | 20:01 |
| Air Blank | | 0.000 | 20:02 |
| Subject Sample | # 1 | 0.285 | 20:05 |
| Breath Volume | | 1.445 | 20:05 |
| Air Blank | | 0.000 | 20:05 |
| Air Blank | | 0.000 | 20:07 |
| Subject Sample | # 2 | 0.314 | 20:09 |
| Breath Volume | | 1.390 | 20:09 |
| Air Blank | | 0.000 | 20:10 |
| Air Blank | | 0.000 | 20:11 |
| Subject Sample | # 3 | 0.295 | 20:13 |
| Breath Volume | | 1.625 | 20:13 |
| Air Blank | | 0.000 | 20:14 |
| Control Test | | 0.078 | 20:14 |
| Air Blank | | 0.000 | 20:14 |
| Diagnostics Check | | OK | 20:14 |

Data Download Date/Time: 07/23/2020 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGGARD,RICKY,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: HAGGARD,RICKY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 16:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: LEGGIO

DATE/TIME OF TEST: 07/29/2020 17:38:57

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 17:33 |
| Air Blank | 0.000 | 17:33 |
| Control Test | 0.080 | 17:33 |
| Air Blank | 0.000 | 17:34 |
| Subject Sample # 1 | 0.000 | 17:34 |
| Breath Volume | 2.792 | 17:34 |
| Air Blank | 0.000 | 17:35 |
| Air Blank | 0.000 | 17:37 |
| Subject Sample # 2 | 0.000 | 17:37 |
| Breath Volume | 2.703 | 17:37 |
| Air Blank | 0.000 | 17:38 |
| Control Test | 0.079 | 17:38 |
| Air Blank | 0.000 | 17:38 |
| Diagnostics Check | OK | 17:38 |

Data Download Date/Time: 08/19/2020 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKUS,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: KIRKUS,JOSHUA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/02/2020 22:19:38

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:13 |
| Air Blank | 0.000 | 22:13 |
| Control Test | 0.078 | 22:14 |
| Air Blank | 0.000 | 22:14 |
| Subject Sample # 1 | 0.199 | 22:15 |
| Breath Volume | 1.894 | 22:15 |
| Air Blank | 0.000 | 22:15 |
| Air Blank | 0.000 | 22:17 |
| Subject Sample # 2 | 0.208 | 22:18 |
| Breath Volume | 1.214 | 22:18 |
| Air Blank | 0.000 | 22:18 |
| Control Test | 0.079 | 22:19 |
| Air Blank | 0.000 | 22:19 |
| Diagnostics Check | OK | 22:19 |

Data Download Date/Time: 07/29/2020 16:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINERNEY,JESSE,L
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: COOK,RYAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: DE MAIO

DATE/TIME OF TEST: 07/26/2020 03:56:02

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 03:49 |
| Air Blank | | 0.000 | 03:49 |
| Control Test | | 0.079 | 03:50 |
| Air Blank | | 0.000 | 03:50 |
| Subject Sample | # 1 | 0.084 | 03:51 |
| Breath Volume | | 2.503 | 03:51 |
| Air Blank | | 0.000 | 03:52 |
| Air Blank | | 0.000 | 03:54 |
| Subject Sample | # 2 | 0.076 | 03:54 |
| Breath Volume | | 1.429 | 03:54 |
| Air Blank | | 0.000 | 03:55 |
| Control Test | | 0.078 | 03:55 |
| Air Blank | | 0.000 | 03:55 |
| Diagnostics Check | | OK | 03:56 |

Data Download Date/Time: 07/29/2020 16:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCARDI,A,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: PETRY,D
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 07/29/2020 00:11:22

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 00:07 | |
| Air Blank | 0.000 | 00:08 | |
| Control Test | 0.080 | 00:08 | |
| Air Blank | 0.000 | 00:08 | |
| Subject Sample | # 1 | 00:10 | Subject Test Refused |
| Breath Volume | 0.000 | 00:10 | |
| Air Blank | | 00:10 | RFI Detect |
| Air Blank | 0.000 | 00:11 | |

Data Download Date/Time: 07/29/2020 16:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,B,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: DRUMMER

DATE/TIME OF TEST: 07/03/2020 00:38:31

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:31 |
| Air Blank | 0.000 | 00:32 |
| Control Test | 0.080 | 00:32 |
| Air Blank | 0.000 | 00:33 |
| Subject Sample # 1 | 0.000 | 00:34 |
| Breath Volume | 1.257 | 00:34 |
| Air Blank | 0.000 | 00:34 |
| Air Blank | 0.000 | 00:36 |
| Subject Sample # 2 | 0.000 | 00:37 |
| Breath Volume | 1.316 | 00:37 |
| Air Blank | 0.000 | 00:37 |
| Control Test | 0.080 | 00:37 |
| Air Blank | 0.000 | 00:38 |
| Diagnostics Check | OK | 00:38 |

Data Download Date/Time: 07/03/2020 13:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 07/09/2020 00:29:34

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:21 |
| Air Blank | 0.000 | 00:22 |
| Control Test | 0.080 | 00:22 |
| Air Blank | 0.000 | 00:23 |
| Subject Sample # 1 | 0.134 | 00:24 |
| Breath Volume | 1.132 | 00:24 |
| Air Blank | 0.000 | 00:25 |
| Air Blank | 0.000 | 00:26 |
| Subject Sample # 2 | 0.146 | 00:28 |
| Breath Volume | 1.625 | 00:28 |
| Air Blank | 0.000 | 00:28 |
| Control Test | 0.080 | 00:28 |
| Air Blank | 0.000 | 00:29 |
| Diagnostics Check | OK | 00:29 |

Data Download Date/Time: 08/07/2020 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MASLAK

DATE/TIME OF TEST: 07/10/2020 23:47:12

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 23:44 | |
| Air Blank | 0.000 | 23:44 | |
| Control Test | 0.080 | 23:45 | |
| Air Blank | 0.000 | 23:45 | |
| Subject Sample | # 1 | 23:45 | Subject Test Refused |
| Breath Volume | 0.000 | 23:45 | |
| Air Blank | 0.000 | 23:46 | |
| Control Test | 0.080 | 23:46 | |
| Air Blank | 0.000 | 23:47 | |
| Diagnostics Check | OK | 23:47 | |

Data Download Date/Time: 08/07/2020 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FARINACCI,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 07/11/2020 00:35:44

| | | | |
|--------------------|-------|-------|------------|
| Diagnostics Check | OK | 00:30 | |
| Air Blank | 0.000 | 00:30 | |
| Control Test | 0.080 | 00:31 | |
| Air Blank | 0.000 | 00:31 | |
| Subject Sample # 1 | 0.060 | 00:32 | |
| Breath Volume | 2.300 | 00:32 | |
| Air Blank | 0.000 | 00:32 | |
| Air Blank | 0.000 | 00:34 | |
| Subject Sample # 2 | 0.059 | 00:35 | |
| Breath Volume | 2.093 | 00:35 | |
| Air Blank | | 00:35 | RFI Detect |
| Air Blank | 0.000 | 00:35 | |

Data Download Date/Time: 08/07/2020 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLIA,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DOTSON,N
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: ROMER

DATE/TIME OF TEST: 07/15/2020 00:03:19

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 23:57 | |
| Air Blank | 0.000 | 23:58 | |
| Control Test | 0.079 | 23:58 | |
| Air Blank | 0.000 | 23:58 | |
| Subject Sample | # 1 | 00:01 | Subject Test Refused |
| Breath Volume | 0.175 | 00:01 | |
| Air Blank | 0.000 | 00:02 | |
| Control Test | 0.079 | 00:02 | |
| Air Blank | 0.000 | 00:03 | |
| Diagnostics Check | OK | 00:03 | |

Data Download Date/Time: 08/07/2020 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGMORELLIA,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GRUBER,E
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: GRICE

DATE/TIME OF TEST: 07/15/2020 22:56:31

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:48 |
| Air Blank | 0.000 | 22:48 |
| Control Test | 0.080 | 22:48 |
| Air Blank | 0.000 | 22:49 |
| Subject Sample # 1 | 0.195 | 22:51 |
| Breath Volume | 1.425 | 22:51 |
| Air Blank | 0.000 | 22:51 |
| Air Blank | 0.000 | 22:53 |
| Subject Sample # 2 | 0.190 | 22:54 |
| Breath Volume | 1.601 | 22:54 |
| Air Blank | 0.000 | 22:55 |
| Control Test | 0.078 | 22:55 |
| Air Blank | 0.000 | 22:56 |
| Diagnostics Check | OK | 22:56 |

Data Download Date/Time: 08/07/2020 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLASSE, RAYMOND, J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EVANS, Z
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:52

SUBJECT'S LAST NAME: BARTHOLOMEW

DATE/TIME OF TEST: 07/18/2020 15:19:24

| | | | |
|--------------------|-------|-------|------------|
| Diagnostics Check | OK | 15:12 | |
| Air Blank | 0.000 | 15:12 | |
| Control Test | 0.080 | 15:13 | |
| Air Blank | 0.000 | 15:13 | |
| Subject Sample # 1 | 0.220 | 15:15 | |
| Breath Volume | 1.234 | 15:15 | |
| Air Blank | 0.000 | 15:16 | |
| Air Blank | 0.000 | 15:17 | |
| Subject Sample # 2 | 0.211 | 15:19 | |
| Breath Volume | 1.937 | 15:19 | |
| Air Blank | | 15:19 | RFI Detect |
| Air Blank | | 15:19 | RFI Detect |

Data Download Date/Time: 08/07/2020 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: WHITTINGTON, D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 14:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 07/19/2020 16:26:36

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 16:20 |
| Air Blank | 0.000 | 16:21 |
| Control Test | 0.080 | 16:21 |
| Air Blank | 0.000 | 16:21 |
| Subject Sample # 1 | 0.000 | 16:22 |
| Breath Volume | 1.875 | 16:22 |
| Air Blank | 0.000 | 16:22 |
| Air Blank | 0.000 | 16:24 |
| Subject Sample # 2 | 0.000 | 16:25 |
| Breath Volume | 1.671 | 16:25 |
| Air Blank | 0.000 | 16:25 |
| Control Test | 0.080 | 16:26 |
| Air Blank | 0.000 | 16:26 |
| Diagnostics Check | OK | 16:26 |

Data Download Date/Time: 08/07/2020 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,SHAKAYLA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PRIMAVERA,SEAN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BEHEN

DATE/TIME OF TEST: 07/20/2020 01:40:04

| | | | |
|--------------------|-------|-------|------------|
| Diagnostics Check | OK | 01:36 | |
| Air Blank | 0.000 | 01:37 | |
| Control Test | 0.080 | 01:37 | |
| Air Blank | 0.000 | 01:37 | |
| Subject Sample # 1 | 0.000 | 01:38 | |
| Breath Volume | 3.492 | 01:38 | |
| Air Blank | | 01:38 | RFI Detect |
| Air Blank | 0.000 | 01:39 | |
| Control Test | 0.080 | 01:39 | |
| Air Blank | 0.000 | 01:39 | |
| Diagnostics Check | OK | 01:40 | |

Data Download Date/Time: 08/07/2020 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,SHAKAYLA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PRIMAVERA,SEAN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BEHEN

DATE/TIME OF TEST: 07/20/2020 02:08:28

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:02 |
| Air Blank | 0.000 | 02:03 |
| Control Test | 0.080 | 02:03 |
| Air Blank | 0.000 | 02:03 |
| Subject Sample # 1 | 0.000 | 02:04 |
| Breath Volume | 2.648 | 02:04 |
| Air Blank | 0.000 | 02:04 |
| Air Blank | 0.000 | 02:06 |
| Subject Sample # 2 | 0.000 | 02:07 |
| Breath Volume | 2.285 | 02:07 |
| Air Blank | 0.000 | 02:07 |
| Control Test | 0.079 | 02:07 |
| Air Blank | 0.000 | 02:08 |
| Diagnostics Check | OK | 02:08 |

Data Download Date/Time: 08/07/2020 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINTRON,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GRUBER,E
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 06:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:14

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 07/20/2020 06:38:32

| | | | |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK | 06:37 | |
| Air Blank | | 06:38 | RFI Detect |
| Air Blank | 0.000 | 06:38 | |

Data Download Date/Time: 08/07/2020 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINTRON,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GRUBER,E
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 06:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:14

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 07/20/2020 06:49:23

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 06:40 |
| Air Blank | 0.000 | 06:41 |
| Control Test | 0.079 | 06:41 |
| Air Blank | 0.000 | 06:41 |
| Subject Sample # 1 | 0.207 | 06:44 |
| Breath Volume | 2.671 | 06:44 |
| Air Blank | 0.000 | 06:45 |
| Air Blank | 0.000 | 06:46 |
| Subject Sample # 2 | 0.200 | 06:47 |
| Breath Volume | 1.269 | 06:47 |
| Air Blank | 0.000 | 06:48 |
| Control Test | 0.077 | 06:48 |
| Air Blank | 0.000 | 06:49 |
| Diagnostics Check | OK | 06:49 |

Data Download Date/Time: 08/07/2020 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLIA,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DELEON,T
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: BRAGG

DATE/TIME OF TEST: 07/23/2020 19:03:27

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 19:00 | |
| Air Blank | 0.000 | 19:01 | |
| Control Test | 0.080 | 19:01 | |
| Air Blank | 0.000 | 19:01 | |
| Subject Sample | # 1 | 19:02 | Subject Test Refused |
| Breath Volume | 0.000 | 19:02 | |
| Air Blank | 0.000 | 19:02 | |
| Control Test | 0.080 | 19:02 | |
| Air Blank | 0.000 | 19:03 | |
| Diagnostics Check | OK | 19:03 | |

Data Download Date/Time: 08/07/2020 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGMORELLIA,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HOILETT,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: MAIGE

DATE/TIME OF TEST: 07/23/2020 20:49:34

| | | | |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK | 20:40 | |
| Air Blank | 0.000 | 20:40 | |
| Control Test | 0.080 | 20:41 | |
| Air Blank | 0.000 | 20:41 | |
| Subject Sample | # 1 0.116 | 20:44 | Volume Not Met |
| Breath Volume | 0.160 | 20:44 | |
| Air Blank | 0.000 | 20:45 | |
| Air Blank | 0.000 | 20:47 | |
| Subject Sample | # 2 | 20:48 | Subject Test Refused |
| Breath Volume | 0.707 | 20:48 | |
| Air Blank | 0.000 | 20:48 | |
| Control Test | 0.079 | 20:48 | |
| Air Blank | 0.000 | 20:49 | |
| Diagnostics Check | OK | 20:49 | |

Data Download Date/Time: 08/07/2020 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGMORELLIA,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: LINDENBERGER

DATE/TIME OF TEST: 07/23/2020 22:42:53

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:35 |
| Air Blank | 0.000 | 22:35 |
| Control Test | 0.080 | 22:36 |
| Air Blank | 0.000 | 22:36 |
| Subject Sample # 1 | 0.000 | 22:37 |
| Breath Volume | 1.460 | 22:37 |
| Air Blank | 0.000 | 22:38 |
| Air Blank | 0.000 | 22:40 |
| Subject Sample # 2 | 0.000 | 22:41 |
| Breath Volume | 1.429 | 22:41 |
| Air Blank | 0.000 | 22:41 |
| Control Test | 0.080 | 22:42 |
| Air Blank | 0.000 | 22:42 |
| Diagnostics Check | OK | 22:42 |

Data Download Date/Time: 08/07/2020 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGMORELLIA,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: WHITTINGTON,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: VOLPE

DATE/TIME OF TEST: 07/24/2020 22:34:22

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:26 |
| Air Blank | 0.000 | 22:27 |
| Control Test | 0.080 | 22:27 |
| Air Blank | 0.000 | 22:28 |
| Subject Sample # 1 | 0.083 | 22:29 |
| Breath Volume | 1.691 | 22:29 |
| Air Blank | 0.000 | 22:30 |
| Air Blank | 0.000 | 22:32 |
| Subject Sample # 2 | 0.081 | 22:32 |
| Breath Volume | 1.730 | 22:32 |
| Air Blank | 0.000 | 22:33 |
| Control Test | 0.079 | 22:33 |
| Air Blank | 0.000 | 22:34 |
| Diagnostics Check | OK | 22:34 |

Data Download Date/Time: 08/07/2020 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGMORELLIA,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SANDERS,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: ALMEIDA

DATE/TIME OF TEST: 07/24/2020 23:37:40

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 23:23 | |
| Air Blank | | 0.000 | 23:23 | |
| Control Test | | 0.080 | 23:24 | |
| Air Blank | | 0.000 | 23:24 | |
| Subject Sample | # 1 | 0.085 | 23:27 | Volume Not Met |
| Breath Volume | | 0.726 | 23:27 | |
| Air Blank | | 0.000 | 23:28 | |
| Air Blank | | 0.000 | 23:30 | |
| Subject Sample | # 2 | 0.106 | 23:31 | |
| Breath Volume | | 1.710 | 23:31 | |
| Air Blank | | 0.000 | 23:32 | |
| Air Blank | | 0.000 | 23:33 | |
| Subject Sample | # 3 | 0.104 | 23:36 | |
| Breath Volume | | 1.519 | 23:36 | |
| Air Blank | | 0.000 | 23:36 | |
| Control Test | | 0.079 | 23:37 | |
| Air Blank | | 0.000 | 23:37 | |
| Diagnostics Check | | OK | 23:37 | |

Data Download Date/Time: 08/07/2020 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLIA,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GONZALEZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: FELICIANO

DATE/TIME OF TEST: 07/27/2020 19:06:37

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 19:03 | |
| Air Blank | 0.000 | 19:04 | |
| Control Test | 0.080 | 19:04 | |
| Air Blank | 0.000 | 19:05 | |
| Subject Sample | # 1 | 19:05 | Subject Test Refused |
| Breath Volume | 0.000 | 19:05 | |
| Air Blank | 0.000 | 19:05 | |
| Control Test | 0.080 | 19:06 | |
| Air Blank | 0.000 | 19:06 | |
| Diagnostics Check | OK | 19:06 | |

Data Download Date/Time: 08/07/2020 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,S,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SWEENEY,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BARRON

DATE/TIME OF TEST: 07/28/2020 03:35:54

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 03:33 | |
| Air Blank | 0.000 | 03:33 | |
| Control Test | 0.080 | 03:33 | |
| Air Blank | 0.000 | 03:34 | |
| Subject Sample | # 1 | 03:34 | Subject Test Refused |
| Breath Volume | 0.000 | 03:34 | |
| Air Blank | 0.000 | 03:35 | |
| Control Test | 0.080 | 03:35 | |
| Air Blank | 0.000 | 03:35 | |
| Diagnostics Check | OK | 03:35 | |

Data Download Date/Time: 08/07/2020 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANTORSKI,MARK,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 12:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: BECKER

DATE/TIME OF TEST: 07/28/2020 13:26:28

| | | | |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK | 13:25 | |
| Air Blank | | 13:25 | RFI Detect |
| Air Blank | 0.000 | 13:26 | |

Data Download Date/Time: 08/07/2020 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANTORSKI,MARK,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 12:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: BRYAN

DATE/TIME OF TEST: 07/28/2020 13:37:20

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 13:28 |
| Air Blank | 0.000 | 13:28 |
| Control Test | 0.080 | 13:28 |
| Air Blank | 0.000 | 13:29 |
| Subject Sample # 1 | 0.212 | 13:29 |
| Breath Volume | 2.867 | 13:29 |
| Air Blank | 0.000 | 13:30 |
| Air Blank | 0.000 | 13:32 |
| Subject Sample # 2 | 0.191 | 13:32 |
| Breath Volume | 1.976 | 13:32 |
| Air Blank | 0.000 | 13:33 |
| Air Blank | 0.000 | 13:35 |
| Subject Sample # 3 | 0.204 | 13:35 |
| Breath Volume | 3.445 | 13:35 |
| Air Blank | 0.000 | 13:36 |
| Control Test | 0.080 | 13:36 |
| Air Blank | 0.000 | 13:37 |
| Diagnostics Check | OK | 13:37 |

Data Download Date/Time: 08/07/2020 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINTRON,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: NEWBERRY,L
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 07/31/2020 03:35:20

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 03:32 | |
| Air Blank | 0.000 | 03:32 | |
| Control Test | 0.080 | 03:33 | |
| Air Blank | 0.000 | 03:33 | |
| Subject Sample | # 1 | 03:33 | Subject Test Refused |
| Breath Volume | 0.000 | 03:33 | |
| Air Blank | 0.000 | 03:34 | |
| Control Test | 0.080 | 03:34 | |
| Air Blank | 0.000 | 03:35 | |
| Diagnostics Check | OK | 03:35 | |

Data Download Date/Time: 08/07/2020 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILIAMS,DEPUTY.,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: ROSCHACH

DATE/TIME OF TEST: 07/04/2020 21:04:41

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:57 |
| Air Blank | 0.000 | 20:57 |
| Control Test | 0.083 | 20:57 |
| Air Blank | 0.000 | 20:58 |
| Subject Sample # 1 | 0.066 | 20:59 |
| Breath Volume | 1.281 | 20:59 |
| Air Blank | 0.000 | 20:59 |
| Air Blank | 0.000 | 21:01 |
| Subject Sample # 2 | 0.060 | 21:03 |
| Breath Volume | 1.121 | 21:03 |
| Air Blank | 0.000 | 21:03 |
| Control Test | 0.083 | 21:04 |
| Air Blank | 0.000 | 21:04 |
| Diagnostics Check | OK | 21:04 |

Data Download Date/Time: 08/07/2020 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCCONNELL, CHAD
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: MILES

DATE/TIME OF TEST: 07/07/2020 22:08:27

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 22:05 | |
| Air Blank | 0.000 | 22:06 | |
| Control Test | 0.084 | 22:06 | |
| Air Blank | 0.000 | 22:06 | |
| Subject Sample | # 1 | 22:07 | Subject Test Refused |
| Breath Volume | 0.000 | 22:07 | |
| Air Blank | 0.000 | 22:07 | |
| Control Test | 0.083 | 22:07 | |
| Air Blank | 0.000 | 22:08 | |
| Diagnostics Check | OK | 22:08 | |

Data Download Date/Time: 08/07/2020 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BRASHEARS, E
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 07/07/2020 23:47:34

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:40 |
| Air Blank | 0.000 | 23:41 |
| Control Test | 0.083 | 23:41 |
| Air Blank | 0.000 | 23:42 |
| Subject Sample # 1 | 0.169 | 23:42 |
| Breath Volume | 1.699 | 23:42 |
| Air Blank | 0.000 | 23:43 |
| Air Blank | 0.000 | 23:45 |
| Subject Sample # 2 | 0.172 | 23:45 |
| Breath Volume | 1.496 | 23:45 |
| Air Blank | 0.000 | 23:46 |
| Control Test | 0.082 | 23:46 |
| Air Blank | 0.000 | 23:47 |
| Diagnostics Check | OK | 23:47 |

Data Download Date/Time: 08/07/2020 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,ROBERTO,R
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: GONZALEZ,JUAN
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 07/01/2020 22:58:47

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 22:55 | |
| Air Blank | 0.000 | 22:56 | |
| Control Test | 0.079 | 22:56 | |
| Air Blank | 0.000 | 22:57 | |
| Subject Sample | # 1 | 22:57 | Subject Test Refused |
| Breath Volume | 0.000 | 22:57 | |
| Air Blank | 0.000 | 22:57 | |
| Control Test | 0.079 | 22:58 | |
| Air Blank | 0.000 | 22:58 | |
| Diagnostics Check | OK | 22:58 | |

Data Download Date/Time: 07/23/2020 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REICHEL, RICH, J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: VASQUEZ, BRYANT
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CHANCEY

DATE/TIME OF TEST: 07/10/2020 02:03:41

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:56 |
| Air Blank | 0.000 | 01:56 |
| Control Test | 0.081 | 01:57 |
| Air Blank | 0.000 | 01:57 |
| Subject Sample # 1 | 0.102 | 01:58 |
| Breath Volume | 1.472 | 01:58 |
| Air Blank | 0.000 | 01:59 |
| Air Blank | 0.000 | 02:01 |
| Subject Sample # 2 | 0.102 | 02:02 |
| Breath Volume | 1.804 | 02:02 |
| Air Blank | 0.000 | 02:02 |
| Control Test | 0.080 | 02:03 |
| Air Blank | 0.000 | 02:03 |
| Diagnostics Check | OK | 02:03 |

Data Download Date/Time: 07/23/2020 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COWELL,DECKITA,R
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: ULLRICH,LAURA
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 07/10/2020 22:40:39

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:34 |
| Air Blank | 0.000 | 22:34 |
| Control Test | 0.080 | 22:35 |
| Air Blank | 0.000 | 22:35 |
| Subject Sample # 1 | 0.221 | 22:36 |
| Breath Volume | 3.242 | 22:36 |
| Air Blank | 0.000 | 22:36 |
| Air Blank | 0.000 | 22:38 |
| Subject Sample # 2 | 0.219 | 22:39 |
| Breath Volume | 2.640 | 22:39 |
| Air Blank | 0.000 | 22:39 |
| Control Test | 0.080 | 22:40 |
| Air Blank | 0.000 | 22:40 |
| Diagnostics Check | OK | 22:40 |

Data Download Date/Time: 07/23/2020 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COWELL,DECKITA,R
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: STRUBE,SAMUEL
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/15/2020 22:58:29

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:50 |
| Air Blank | 0.000 | 22:51 |
| Control Test | 0.081 | 22:51 |
| Air Blank | 0.000 | 22:52 |
| Subject Sample # 1 | 0.184 | 22:53 |
| Breath Volume | 2.722 | 22:53 |
| Air Blank | 0.000 | 22:54 |
| Air Blank | 0.000 | 22:56 |
| Subject Sample # 2 | 0.178 | 22:56 |
| Breath Volume | 1.996 | 22:56 |
| Air Blank | 0.000 | 22:57 |
| Control Test | 0.079 | 22:57 |
| Air Blank | 0.000 | 22:58 |
| Diagnostics Check | OK | 22:58 |

Data Download Date/Time: 07/23/2020 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COWELL,DECKITA,R
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: MORRIS,JACK
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 07/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/25/2020 21:00:51

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:53 |
| Air Blank | 0.000 | 20:54 |
| Control Test | 0.080 | 20:54 |
| Air Blank | 0.000 | 20:55 |
| Subject Sample # 1 | 0.192 | 20:56 |
| Breath Volume | 3.179 | 20:56 |
| Air Blank | 0.000 | 20:56 |
| Air Blank | 0.000 | 20:58 |
| Subject Sample # 2 | 0.189 | 20:59 |
| Breath Volume | 2.894 | 20:59 |
| Air Blank | 0.000 | 20:59 |
| Control Test | 0.080 | 21:00 |
| Air Blank | 0.000 | 21:00 |
| Diagnostics Check | OK | 21:00 |

Data Download Date/Time: 08/27/2020 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/01/2020 00:06:29

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:00 |
| Air Blank | 0.000 | 00:01 |
| Control Test | 0.081 | 00:01 |
| Air Blank | 0.000 | 00:01 |
| Subject Sample # 1 | 0.000 | 00:02 |
| Breath Volume | 2.484 | 00:02 |
| Air Blank | 0.000 | 00:02 |
| Air Blank | 0.000 | 00:04 |
| Subject Sample # 2 | 0.000 | 00:05 |
| Breath Volume | 2.648 | 00:05 |
| Air Blank | 0.000 | 00:05 |
| Control Test | 0.079 | 00:05 |
| Air Blank | 0.000 | 00:06 |
| Diagnostics Check | OK | 00:06 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 07/02/2020 20:11:10

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:04 |
| Air Blank | 0.000 | 20:04 |
| Control Test | 0.081 | 20:05 |
| Air Blank | 0.000 | 20:05 |
| Subject Sample # 1 | 0.000 | 20:06 |
| Breath Volume | 1.152 | 20:06 |
| Air Blank | 0.000 | 20:07 |
| Air Blank | 0.000 | 20:09 |
| Subject Sample # 2 | 0.000 | 20:09 |
| Breath Volume | 1.343 | 20:09 |
| Air Blank | 0.000 | 20:10 |
| Control Test | 0.079 | 20:10 |
| Air Blank | 0.000 | 20:11 |
| Diagnostics Check | OK | 20:11 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEPUTY M.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: KENNEDY-LEE

DATE/TIME OF TEST: 07/02/2020 22:40:13

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:33 |
| Air Blank | 0.000 | 22:33 |
| Control Test | 0.080 | 22:34 |
| Air Blank | 0.000 | 22:34 |
| Subject Sample # 1 | 0.201 | 22:35 |
| Breath Volume | 2.929 | 22:35 |
| Air Blank | 0.000 | 22:35 |
| Air Blank | 0.000 | 22:37 |
| Subject Sample # 2 | 0.196 | 22:38 |
| Breath Volume | 3.195 | 22:38 |
| Air Blank | 0.000 | 22:39 |
| Control Test | 0.080 | 22:39 |
| Air Blank | 0.000 | 22:40 |
| Diagnostics Check | OK | 22:40 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: WATSON,DEPUTY D.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 07/03/2020 03:29:53

| | | | |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK | 03:18 | |
| Air Blank | 0.000 | 03:18 | |
| Control Test | 0.080 | 03:19 | |
| Air Blank | 0.000 | 03:19 | |
| Subject Sample | # 1 | 03:22 | No Sample Provided |
| Breath Volume | 0.000 | 03:22 | |
| Air Blank | 0.000 | 03:23 | |
| Air Blank | 0.000 | 03:25 | |
| Subject Sample | # 2 | 03:28 | No Sample Provided |
| Breath Volume | 0.000 | 03:28 | |
| Air Blank | 0.000 | 03:28 | |
| Control Test | 0.080 | 03:29 | |
| Air Blank | 0.000 | 03:29 | |
| Diagnostics Check | OK | 03:29 | |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,KYLE,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,KYLE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: STREITLER

DATE/TIME OF TEST: 07/04/2020 00:12:50

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 00:06 |
| Air Blank | | 0.000 | 00:07 |
| Control Test | | 0.080 | 00:07 |
| Air Blank | | 0.000 | 00:07 |
| Subject Sample | # 1 | 0.000 | 00:08 |
| Breath Volume | | 4.093 | 00:08 |
| Air Blank | | 0.000 | 00:08 |
| Air Blank | | 0.000 | 00:10 |
| Subject Sample | # 2 | 0.000 | 00:11 |
| Breath Volume | | 3.546 | 00:11 |
| Air Blank | | 0.000 | 00:11 |
| Control Test | | 0.080 | 00:12 |
| Air Blank | | 0.000 | 00:12 |
| Diagnostics Check | | OK | 00:12 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKERT,GABRIEL,B
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ECKERT,GABRIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 07/04/2020 03:10:42

| | | | |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK | 03:05 | |
| Air Blank | 0.000 | 03:06 | |
| Control Test | 0.080 | 03:06 | |
| Air Blank | 0.000 | 03:06 | |
| Subject Sample | # 1 0.215 | 03:08 | |
| Breath Volume | 1.449 | 03:08 | |
| Air Blank | | 03:09 | RFI Detect |
| Air Blank | 0.000 | 03:09 | |
| Control Test | 0.080 | 03:10 | |
| Air Blank | 0.000 | 03:10 | |
| Diagnostics Check | OK | 03:10 | |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKERT,GABRIEL,B
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ECKERT,GABRIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 07/04/2020 03:22:32

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 03:12 |
| Air Blank | | 0.000 | 03:13 |
| Control Test | | 0.078 | 03:13 |
| Air Blank | | 0.000 | 03:14 |
| Subject Sample | # 1 | 0.203 | 03:17 |
| Breath Volume | | 1.968 | 03:17 |
| Air Blank | | 0.000 | 03:17 |
| Air Blank | | 0.000 | 03:19 |
| Subject Sample | # 2 | 0.190 | 03:20 |
| Breath Volume | | 2.359 | 03:20 |
| Air Blank | | 0.000 | 03:21 |
| Control Test | | 0.077 | 03:21 |
| Air Blank | | 0.000 | 03:22 |
| Diagnostics Check | | OK | 03:22 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKERT,GABRIEL,B
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: WHITE,L
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: TILSON

DATE/TIME OF TEST: 07/04/2020 04:06:20

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 04:00 |
| Air Blank | 0.000 | 04:00 |
| Control Test | 0.079 | 04:01 |
| Air Blank | 0.000 | 04:01 |
| Subject Sample # 1 | 0.111 | 04:01 |
| Breath Volume | 1.699 | 04:01 |
| Air Blank | 0.000 | 04:02 |
| Air Blank | 0.000 | 04:04 |
| Subject Sample # 2 | 0.110 | 04:04 |
| Breath Volume | 2.746 | 04:04 |
| Air Blank | 0.000 | 04:05 |
| Control Test | 0.080 | 04:05 |
| Air Blank | 0.000 | 04:06 |
| Diagnostics Check | OK | 04:06 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICK\A.,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: NICK,A.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: MESSAI

DATE/TIME OF TEST: 07/04/2020 23:13:28

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:06 |
| Air Blank | 0.000 | 23:07 |
| Control Test | 0.080 | 23:07 |
| Air Blank | 0.000 | 23:08 |
| Subject Sample # 1 | 0.100 | 23:08 |
| Breath Volume | 4.355 | 23:08 |
| Air Blank | 0.000 | 23:09 |
| Air Blank | 0.000 | 23:11 |
| Subject Sample # 2 | 0.084 | 23:11 |
| Breath Volume | 2.312 | 23:11 |
| Air Blank | 0.000 | 23:12 |
| Control Test | 0.079 | 23:12 |
| Air Blank | 0.000 | 23:13 |
| Diagnostics Check | OK | 23:13 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: GOULET

DATE/TIME OF TEST: 07/07/2020 01:13:58

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 01:11 | |
| Air Blank | 0.000 | 01:11 | |
| Control Test | 0.080 | 01:11 | |
| Air Blank | 0.000 | 01:12 | |
| Subject Sample | # 1 | 01:12 | Subject Test Refused |
| Breath Volume | 0.000 | 01:12 | |
| Air Blank | 0.000 | 01:13 | |
| Control Test | 0.080 | 01:13 | |
| Air Blank | 0.000 | 01:13 | |
| Diagnostics Check | OK | 01:13 | |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,N,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NELSON,A
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 07/09/2020 04:44:56

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 04:37 |
| Air Blank | 0.000 | 04:38 |
| Control Test | 0.080 | 04:38 |
| Air Blank | 0.000 | 04:39 |
| Subject Sample # 1 | 0.222 | 04:39 |
| Breath Volume | 2.574 | 04:39 |
| Air Blank | 0.000 | 04:40 |
| Air Blank | 0.000 | 04:42 |
| Subject Sample # 2 | 0.217 | 04:43 |
| Breath Volume | 2.660 | 04:43 |
| Air Blank | 0.000 | 04:44 |
| Control Test | 0.079 | 04:44 |
| Air Blank | 0.000 | 04:44 |
| Diagnostics Check | OK | 04:44 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CORONADO,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:38

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 07/09/2020 15:32:19

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 15:25 |
| Air Blank | 0.000 | 15:26 |
| Control Test | 0.081 | 15:26 |
| Air Blank | 0.000 | 15:27 |
| Subject Sample # 1 | 0.000 | 15:27 |
| Breath Volume | 2.101 | 15:27 |
| Air Blank | 0.000 | 15:28 |
| Air Blank | 0.000 | 15:30 |
| Subject Sample # 2 | 0.000 | 15:30 |
| Breath Volume | 2.210 | 15:30 |
| Air Blank | 0.000 | 15:31 |
| Control Test | 0.081 | 15:31 |
| Air Blank | 0.000 | 15:32 |
| Diagnostics Check | OK | 15:32 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BOWERS,JOSEPH
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: COLANTUONO

DATE/TIME OF TEST: 07/10/2020 18:31:38

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 18:24 |
| Air Blank | 0.000 | 18:24 |
| Control Test | 0.079 | 18:24 |
| Air Blank | 0.000 | 18:25 |
| Subject Sample # 1 | 0.162 | 18:26 |
| Breath Volume | 2.339 | 18:26 |
| Air Blank | 0.000 | 18:27 |
| Air Blank | 0.000 | 18:29 |
| Subject Sample # 2 | 0.154 | 18:30 |
| Breath Volume | 1.558 | 18:30 |
| Air Blank | 0.000 | 18:30 |
| Control Test | 0.077 | 18:31 |
| Air Blank | 0.000 | 18:31 |
| Diagnostics Check | OK | 18:31 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICK,A,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: NICK,A.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: PRITTS

DATE/TIME OF TEST: 07/10/2020 20:01:17

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 19:55 |
| Air Blank | 0.000 | 19:55 |
| Control Test | 0.080 | 19:56 |
| Air Blank | 0.000 | 19:56 |
| Subject Sample # 1 | 0.000 | 19:57 |
| Breath Volume | 2.199 | 19:57 |
| Air Blank | 0.000 | 19:57 |
| Air Blank | 0.000 | 19:59 |
| Subject Sample # 2 | 0.000 | 19:59 |
| Breath Volume | 2.144 | 19:59 |
| Air Blank | 0.000 | 20:00 |
| Control Test | 0.079 | 20:00 |
| Air Blank | 0.000 | 20:01 |
| Diagnostics Check | OK | 20:01 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: BENKHATAR

DATE/TIME OF TEST: 07/11/2020 04:06:48

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 04:00 |
| Air Blank | 0.000 | 04:00 |
| Control Test | 0.081 | 04:00 |
| Air Blank | 0.000 | 04:01 |
| Subject Sample # 1 | 0.080 | 04:02 |
| Breath Volume | 3.992 | 04:02 |
| Air Blank | 0.000 | 04:02 |
| Air Blank | 0.000 | 04:04 |
| Subject Sample # 2 | 0.079 | 04:05 |
| Breath Volume | 3.664 | 04:05 |
| Air Blank | 0.000 | 04:05 |
| Control Test | 0.079 | 04:06 |
| Air Blank | 0.000 | 04:06 |
| Diagnostics Check | OK | 04:06 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: AYARS

DATE/TIME OF TEST: 07/11/2020 21:49:15

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 21:43 |
| Air Blank | | 0.000 | 21:43 |
| Control Test | | 0.080 | 21:43 |
| Air Blank | | 0.000 | 21:44 |
| Subject Sample | # 1 | 0.200 | 21:44 |
| Breath Volume | | 2.066 | 21:44 |
| Air Blank | | 0.000 | 21:45 |
| Air Blank | | 0.000 | 21:47 |
| Subject Sample | # 2 | 0.188 | 21:47 |
| Breath Volume | | 2.816 | 21:47 |
| Air Blank | | 0.000 | 21:48 |
| Control Test | | 0.079 | 21:48 |
| Air Blank | | 0.000 | 21:49 |
| Diagnostics Check | | OK | 21:49 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DEP D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: OSWALD

DATE/TIME OF TEST: 07/12/2020 20:13:38

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:07 |
| Air Blank | 0.000 | 20:07 |
| Control Test | 0.078 | 20:08 |
| Air Blank | 0.000 | 20:08 |
| Subject Sample # 1 | 0.118 | 20:09 |
| Breath Volume | 1.941 | 20:09 |
| Air Blank | 0.000 | 20:09 |
| Air Blank | 0.000 | 20:11 |
| Subject Sample # 2 | 0.122 | 20:12 |
| Breath Volume | 2.757 | 20:12 |
| Air Blank | 0.000 | 20:12 |
| Control Test | 0.079 | 20:13 |
| Air Blank | 0.000 | 20:13 |
| Diagnostics Check | OK | 20:13 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MINGS,DEP G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: KAUFFMAN

DATE/TIME OF TEST: 07/12/2020 23:38:17

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:32 |
| Air Blank | 0.000 | 23:32 |
| Control Test | 0.077 | 23:32 |
| Air Blank | 0.000 | 23:33 |
| Subject Sample # 1 | 0.122 | 23:33 |
| Breath Volume | 3.078 | 23:33 |
| Air Blank | 0.000 | 23:34 |
| Air Blank | 0.000 | 23:36 |
| Subject Sample # 2 | 0.122 | 23:36 |
| Breath Volume | 2.722 | 23:36 |
| Air Blank | 0.000 | 23:37 |
| Control Test | 0.078 | 23:37 |
| Air Blank | 0.000 | 23:38 |
| Diagnostics Check | OK | 23:38 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTWELL,J,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CANTWELL,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: BRAY

DATE/TIME OF TEST: 07/14/2020 14:21:37

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 14:15 |
| Air Blank | 0.000 | 14:15 |
| Control Test | 0.079 | 14:16 |
| Air Blank | 0.000 | 14:16 |
| Subject Sample # 1 | 0.000 | 14:17 |
| Breath Volume | 2.152 | 14:17 |
| Air Blank | 0.000 | 14:17 |
| Air Blank | 0.000 | 14:19 |
| Subject Sample # 2 | 0.000 | 14:20 |
| Breath Volume | 2.199 | 14:20 |
| Air Blank | 0.000 | 14:20 |
| Control Test | 0.079 | 14:21 |
| Air Blank | 0.000 | 14:21 |
| Diagnostics Check | OK | 14:21 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOW,JOSHUA,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FLOW,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: GRIMES

DATE/TIME OF TEST: 07/15/2020 11:07:09

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 11:01 |
| Air Blank | 0.000 | 11:01 |
| Control Test | 0.079 | 11:01 |
| Air Blank | 0.000 | 11:02 |
| Subject Sample # 1 | 0.000 | 11:02 |
| Breath Volume | 2.894 | 11:02 |
| Air Blank | 0.000 | 11:03 |
| Air Blank | 0.000 | 11:05 |
| Subject Sample # 2 | 0.000 | 11:05 |
| Breath Volume | 2.136 | 11:05 |
| Air Blank | 0.000 | 11:06 |
| Control Test | 0.079 | 11:06 |
| Air Blank | 0.000 | 11:07 |
| Diagnostics Check | OK | 11:07 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 07/17/2020 02:53:44

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:46 |
| Air Blank | 0.000 | 02:46 |
| Control Test | 0.079 | 02:47 |
| Air Blank | 0.000 | 02:47 |
| Subject Sample # 1 | 0.242 | 02:48 |
| Breath Volume | 2.761 | 02:48 |
| Air Blank | 0.000 | 02:49 |
| Air Blank | 0.000 | 02:51 |
| Subject Sample # 2 | 0.237 | 02:52 |
| Breath Volume | 2.320 | 02:52 |
| Air Blank | 0.000 | 02:52 |
| Control Test | 0.077 | 02:53 |
| Air Blank | 0.000 | 02:53 |
| Diagnostics Check | OK | 02:53 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: STOLL,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: POSEY

DATE/TIME OF TEST: 07/18/2020 21:10:30

| | | | |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK | 21:09 | |
| Air Blank | | 21:10 | RFI Detect |
| Air Blank | 0.000 | 21:10 | |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: STOLL,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: POSEY

DATE/TIME OF TEST: 07/18/2020 21:23:25

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:17 |
| Air Blank | 0.000 | 21:17 |
| Control Test | 0.080 | 21:18 |
| Air Blank | 0.000 | 21:18 |
| Subject Sample # 1 | 0.000 | 21:19 |
| Breath Volume | 2.796 | 21:19 |
| Air Blank | 0.000 | 21:19 |
| Air Blank | 0.000 | 21:21 |
| Subject Sample # 2 | 0.000 | 21:22 |
| Breath Volume | 2.878 | 21:22 |
| Air Blank | 0.000 | 21:22 |
| Control Test | 0.079 | 21:22 |
| Air Blank | 0.000 | 21:23 |
| Diagnostics Check | OK | 21:23 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKERT,GABRIEL,B
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ECKERT,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: TERRY

DATE/TIME OF TEST: 07/18/2020 23:48:22

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:42 |
| Air Blank | 0.000 | 23:42 |
| Control Test | 0.080 | 23:42 |
| Air Blank | 0.000 | 23:43 |
| Subject Sample # 1 | 0.195 | 23:43 |
| Breath Volume | 1.878 | 23:43 |
| Air Blank | 0.000 | 23:44 |
| Air Blank | 0.000 | 23:46 |
| Subject Sample # 2 | 0.179 | 23:46 |
| Breath Volume | 1.550 | 23:46 |
| Air Blank | 0.000 | 23:47 |
| Control Test | 0.080 | 23:47 |
| Air Blank | 0.000 | 23:48 |
| Diagnostics Check | OK | 23:48 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: DUCKER

DATE/TIME OF TEST: 07/19/2020 03:48:21

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:42 |
| Air Blank | 0.000 | 03:42 |
| Control Test | 0.080 | 03:43 |
| Air Blank | 0.000 | 03:43 |
| Subject Sample # 1 | 0.000 | 03:44 |
| Breath Volume | 2.949 | 03:44 |
| Air Blank | 0.000 | 03:44 |
| Air Blank | 0.000 | 03:46 |
| Subject Sample # 2 | 0.000 | 03:46 |
| Breath Volume | 1.101 | 03:46 |
| Air Blank | 0.000 | 03:47 |
| Control Test | 0.079 | 03:47 |
| Air Blank | 0.000 | 03:48 |
| Diagnostics Check | OK | 03:48 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,N,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: COKER,J
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: BURNETT

DATE/TIME OF TEST: 07/19/2020 20:45:44

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:39 |
| Air Blank | 0.000 | 20:39 |
| Control Test | 0.079 | 20:40 |
| Air Blank | 0.000 | 20:40 |
| Subject Sample # 1 | 0.137 | 20:41 |
| Breath Volume | 2.566 | 20:41 |
| Air Blank | 0.000 | 20:41 |
| Air Blank | 0.000 | 20:43 |
| Subject Sample # 2 | 0.146 | 20:44 |
| Breath Volume | 4.160 | 20:44 |
| Air Blank | 0.000 | 20:44 |
| Control Test | 0.078 | 20:45 |
| Air Blank | 0.000 | 20:45 |
| Diagnostics Check | OK | 20:45 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHALE,SGT C
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: MAGLIANO

DATE/TIME OF TEST: 07/20/2020 22:43:44

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:37 |
| Air Blank | 0.000 | 22:38 |
| Control Test | 0.080 | 22:38 |
| Air Blank | 0.000 | 22:38 |
| Subject Sample # 1 | 0.182 | 22:39 |
| Breath Volume | 3.320 | 22:39 |
| Air Blank | 0.000 | 22:40 |
| Air Blank | 0.000 | 22:41 |
| Subject Sample # 2 | 0.179 | 22:42 |
| Breath Volume | 2.812 | 22:42 |
| Air Blank | 0.000 | 22:42 |
| Control Test | 0.078 | 22:43 |
| Air Blank | 0.000 | 22:43 |
| Diagnostics Check | OK | 22:43 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BOWERS,JOSEPH
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: HAIGHT

DATE/TIME OF TEST: 07/20/2020 22:53:36

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 22:49 | |
| Air Blank | 0.000 | 22:50 | |
| Control Test | 0.079 | 22:50 | |
| Air Blank | 0.000 | 22:51 | |
| Subject Sample | # 1 | 22:52 | Subject Test Refused |
| Breath Volume | 0.000 | 22:52 | |
| Air Blank | 0.000 | 22:52 | |
| Control Test | 0.079 | 22:53 | |
| Air Blank | 0.000 | 22:53 | |
| Diagnostics Check | OK | 22:53 | |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHALE,SGT C
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 07/21/2020 21:19:06

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 21:12 |
| Air Blank | | 0.000 | 21:13 |
| Control Test | | 0.079 | 21:13 |
| Air Blank | | 0.000 | 21:14 |
| Subject Sample | # 1 | 0.264 | 21:14 |
| Breath Volume | | 2.820 | 21:14 |
| Air Blank | | 0.000 | 21:15 |
| Air Blank | | 0.000 | 21:17 |
| Subject Sample | # 2 | 0.247 | 21:17 |
| Breath Volume | | 2.835 | 21:17 |
| Air Blank | | 0.000 | 21:18 |
| Control Test | | 0.077 | 21:18 |
| Air Blank | | 0.000 | 21:19 |
| Diagnostics Check | | OK | 21:19 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/23/2020 21:13:27

| | | | |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK | 21:01 | |
| Air Blank | 0.000 | 21:01 | |
| Control Test | 0.080 | 21:02 | |
| Air Blank | 0.000 | 21:02 | |
| Subject Sample | # 1 | 21:05 | No Sample Provided |
| Breath Volume | 0.000 | 21:05 | |
| Air Blank | 0.000 | 21:06 | |
| Air Blank | 0.000 | 21:08 | |
| Subject Sample | # 2 | 0.181 | 21:09 |
| Breath Volume | 2.910 | 21:09 | |
| Air Blank | 0.000 | 21:09 | |
| Air Blank | 0.000 | 21:11 | |
| Subject Sample | # 3 | 0.182 | 21:11 |
| Breath Volume | 2.878 | 21:11 | |
| Air Blank | 0.000 | 21:12 | |
| Control Test | 0.079 | 21:12 | |
| Air Blank | 0.000 | 21:13 | |
| Diagnostics Check | OK | 21:13 | |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 07/24/2020 18:15:40

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 18:06 |
| Air Blank | 0.000 | 18:06 |
| Control Test | 0.079 | 18:07 |
| Air Blank | 0.000 | 18:07 |
| Subject Sample # 1 | 0.048 | 18:09 |
| Breath Volume | 2.726 | 18:09 |
| Air Blank | 0.000 | 18:09 |
| Air Blank | 0.000 | 18:11 |
| Subject Sample # 2 | 0.043 | 18:14 |
| Breath Volume | 1.480 | 18:14 |
| Air Blank | 0.000 | 18:14 |
| Control Test | 0.078 | 18:15 |
| Air Blank | 0.000 | 18:15 |
| Diagnostics Check | OK | 18:15 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: CRAMER

DATE/TIME OF TEST: 07/24/2020 21:52:02

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:45 |
| Air Blank | 0.000 | 21:46 |
| Control Test | 0.079 | 21:46 |
| Air Blank | 0.000 | 21:46 |
| Subject Sample # 1 | 0.179 | 21:47 |
| Breath Volume | 2.371 | 21:47 |
| Air Blank | 0.000 | 21:48 |
| Air Blank | 0.000 | 21:49 |
| Subject Sample # 2 | 0.180 | 21:50 |
| Breath Volume | 2.507 | 21:50 |
| Air Blank | 0.000 | 21:51 |
| Control Test | 0.078 | 21:51 |
| Air Blank | 0.000 | 21:51 |
| Diagnostics Check | OK | 21:52 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUERO,JOYRICK,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SUERO,JOYRICK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:48

SUBJECT'S LAST NAME: FRYE

DATE/TIME OF TEST: 07/25/2020 08:19:42

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 08:13 |
| Air Blank | 0.000 | 08:14 |
| Control Test | 0.079 | 08:14 |
| Air Blank | 0.000 | 08:14 |
| Subject Sample # 1 | 0.000 | 08:15 |
| Breath Volume | 3.355 | 08:15 |
| Air Blank | 0.000 | 08:15 |
| Air Blank | 0.000 | 08:17 |
| Subject Sample # 2 | 0.000 | 08:18 |
| Breath Volume | 2.171 | 08:18 |
| Air Blank | 0.000 | 08:18 |
| Control Test | 0.080 | 08:19 |
| Air Blank | 0.000 | 08:19 |
| Diagnostics Check | OK | 08:19 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: ESCLANGON

DATE/TIME OF TEST: 07/27/2020 23:52:01

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:46 |
| Air Blank | 0.000 | 23:46 |
| Control Test | 0.080 | 23:46 |
| Air Blank | 0.000 | 23:47 |
| Subject Sample # 1 | 0.007 | 23:47 |
| Breath Volume | 1.218 | 23:47 |
| Air Blank | 0.000 | 23:48 |
| Air Blank | 0.000 | 23:50 |
| Subject Sample # 2 | 0.007 | 23:50 |
| Breath Volume | 1.210 | 23:50 |
| Air Blank | 0.000 | 23:51 |
| Control Test | 0.080 | 23:51 |
| Air Blank | 0.000 | 23:51 |
| Diagnostics Check | OK | 23:52 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,N,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MONTES,S
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: GUIDO

DATE/TIME OF TEST: 07/28/2020 04:47:39

| | | | |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK | 04:40 | |
| Air Blank | 0.000 | 04:40 | |
| Control Test | 0.080 | 04:41 | |
| Air Blank | 0.000 | 04:41 | |
| Subject Sample | # 1 0.064 | 04:42 | |
| Breath Volume | 1.929 | 04:42 | |
| Air Blank | 0.000 | 04:43 | |
| Air Blank | 0.000 | 04:45 | |
| Subject Sample | # 2 0.059 | 04:47 | |
| Breath Volume | 1.257 | 04:47 | |
| Air Blank | | 04:47 | RFI Detect |
| Air Blank | | 04:47 | RFI Detect |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,N,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MONTES,S
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: GUIDO

DATE/TIME OF TEST: 07/28/2020 05:05:43

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 04:58 |
| Air Blank | 0.000 | 04:59 |
| Control Test | 0.079 | 04:59 |
| Air Blank | 0.000 | 05:00 |
| Subject Sample # 1 | 0.059 | 05:01 |
| Breath Volume | 1.437 | 05:01 |
| Air Blank | 0.000 | 05:01 |
| Air Blank | 0.000 | 05:03 |
| Subject Sample # 2 | 0.054 | 05:04 |
| Breath Volume | 1.417 | 05:04 |
| Air Blank | 0.000 | 05:04 |
| Control Test | 0.079 | 05:05 |
| Air Blank | 0.000 | 05:05 |
| Diagnostics Check | OK | 05:05 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: WALTERS-D'ASCENZO

DATE/TIME OF TEST: 07/28/2020 23:46:50

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:40 |
| Air Blank | 0.000 | 23:41 |
| Control Test | 0.080 | 23:41 |
| Air Blank | 0.000 | 23:41 |
| Subject Sample # 1 | 0.000 | 23:42 |
| Breath Volume | 1.554 | 23:42 |
| Air Blank | 0.000 | 23:42 |
| Air Blank | 0.000 | 23:44 |
| Subject Sample # 2 | 0.000 | 23:45 |
| Breath Volume | 2.851 | 23:45 |
| Air Blank | 0.000 | 23:45 |
| Control Test | 0.080 | 23:46 |
| Air Blank | 0.000 | 23:46 |
| Diagnostics Check | OK | 23:46 |

Data Download Date/Time: 07/29/2020 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 07/30/2020 17:16:27

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 17:10 |
| Air Blank | 0.000 | 17:10 |
| Control Test | 0.080 | 17:11 |
| Air Blank | 0.000 | 17:11 |
| Subject Sample # 1 | 0.000 | 17:12 |
| Breath Volume | 2.417 | 17:12 |
| Air Blank | 0.000 | 17:12 |
| Air Blank | 0.000 | 17:14 |
| Subject Sample # 2 | 0.000 | 17:15 |
| Breath Volume | 2.289 | 17:15 |
| Air Blank | 0.000 | 17:15 |
| Control Test | 0.079 | 17:15 |
| Air Blank | 0.000 | 17:16 |
| Diagnostics Check | OK | 17:16 |

Data Download Date/Time: 08/28/2020 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: RISSER

DATE/TIME OF TEST: 07/07/2020 22:32:48

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:26 |
| Air Blank | 0.000 | 22:27 |
| Control Test | 0.079 | 22:27 |
| Air Blank | 0.000 | 22:28 |
| Subject Sample # 1 | 0.000 | 22:28 |
| Breath Volume | 3.550 | 22:28 |
| Air Blank | 0.000 | 22:28 |
| Air Blank | 0.000 | 22:30 |
| Subject Sample # 2 | 0.000 | 22:31 |
| Breath Volume | 4.421 | 22:31 |
| Air Blank | 0.000 | 22:31 |
| Control Test | 0.079 | 22:32 |
| Air Blank | 0.000 | 22:32 |
| Diagnostics Check | OK | 22:32 |

Data Download Date/Time: 07/29/2020 08:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:58

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 07/10/2020 03:27:59

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 03:24 | |
| Air Blank | 0.000 | 03:25 | |
| Control Test | 0.078 | 03:25 | |
| Air Blank | 0.000 | 03:26 | |
| Subject Sample | # 1 | 03:26 | Subject Test Refused |
| Breath Volume | 0.000 | 03:26 | |
| Air Blank | 0.000 | 03:27 | |
| Control Test | 0.078 | 03:27 | |
| Air Blank | 0.000 | 03:27 | |
| Diagnostics Check | OK | 03:27 | |

Data Download Date/Time: 07/29/2020 08:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:58

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 07/10/2020 17:24:27

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 17:18 |
| Air Blank | 0.000 | 17:18 |
| Control Test | 0.078 | 17:19 |
| Air Blank | 0.000 | 17:19 |
| Subject Sample # 1 | 0.201 | 17:20 |
| Breath Volume | 1.289 | 17:20 |
| Air Blank | 0.000 | 17:20 |
| Air Blank | 0.000 | 17:22 |
| Subject Sample # 2 | 0.192 | 17:22 |
| Breath Volume | 1.285 | 17:22 |
| Air Blank | 0.000 | 17:23 |
| Control Test | 0.078 | 17:23 |
| Air Blank | 0.000 | 17:24 |
| Diagnostics Check | OK | 17:24 |

Data Download Date/Time: 07/29/2020 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEP J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 07/13/2020 00:33:57

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:27 |
| Air Blank | 0.000 | 00:28 |
| Control Test | 0.079 | 00:28 |
| Air Blank | 0.000 | 00:29 |
| Subject Sample # 1 | 0.186 | 00:29 |
| Breath Volume | 1.542 | 00:29 |
| Air Blank | 0.000 | 00:30 |
| Air Blank | 0.000 | 00:32 |
| Subject Sample # 2 | 0.173 | 00:32 |
| Breath Volume | 1.367 | 00:32 |
| Air Blank | 0.000 | 00:33 |
| Control Test | 0.078 | 00:33 |
| Air Blank | 0.000 | 00:33 |
| Diagnostics Check | OK | 00:33 |

Data Download Date/Time: 07/29/2020 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: KLIMANTOVICH

DATE/TIME OF TEST: 07/17/2020 02:37:55

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:32 |
| Air Blank | 0.000 | 02:32 |
| Control Test | 0.079 | 02:32 |
| Air Blank | 0.000 | 02:33 |
| Subject Sample # 1 | 0.000 | 02:33 |
| Breath Volume | 3.667 | 02:33 |
| Air Blank | 0.000 | 02:34 |
| Air Blank | 0.000 | 02:36 |
| Subject Sample # 2 | 0.000 | 02:36 |
| Breath Volume | 1.902 | 02:36 |
| Air Blank | 0.000 | 02:37 |
| Control Test | 0.078 | 02:37 |
| Air Blank | 0.000 | 02:37 |
| Diagnostics Check | OK | 02:37 |

Data Download Date/Time: 07/29/2020 08:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 07/17/2020 19:36:52

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 19:30 |
| Air Blank | 0.000 | 19:31 |
| Control Test | 0.079 | 19:31 |
| Air Blank | 0.000 | 19:31 |
| Subject Sample # 1 | 0.202 | 19:32 |
| Breath Volume | 4.539 | 19:32 |
| Air Blank | 0.000 | 19:33 |
| Air Blank | 0.000 | 19:34 |
| Subject Sample # 2 | 0.204 | 19:35 |
| Breath Volume | 4.792 | 19:35 |
| Air Blank | 0.000 | 19:35 |
| Control Test | 0.079 | 19:36 |
| Air Blank | 0.000 | 19:36 |
| Diagnostics Check | OK | 19:36 |

Data Download Date/Time: 07/29/2020 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: AVROS

DATE/TIME OF TEST: 07/21/2020 17:42:21

| | | | |
|--------------------|-------|-------|--------------|
| Diagnostics Check | OK | 17:36 | |
| Air Blank | 0.000 | 17:36 | |
| Control Test | 0.078 | 17:36 | |
| Air Blank | 0.000 | 17:37 | |
| Subject Sample # 1 | 0.000 | 17:37 | |
| Breath Volume | 2.265 | 17:37 | |
| Air Blank | 0.000 | 17:38 | |
| Air Blank | 0.000 | 17:40 | |
| Subject Sample # 2 | 0.000 | 17:40 | |
| Breath Volume | 1.464 | 17:40 | |
| Air Blank | 0.000 | 17:40 | |
| Control Test | 0.079 | 17:41 | |
| Air Blank | | 17:41 | Ambient Fail |
| Air Blank | 0.000 | 17:42 | |

Data Download Date/Time: 07/29/2020 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: AVROS

DATE/TIME OF TEST: 07/21/2020 18:12:11

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 18:05 |
| Air Blank | 0.000 | 18:06 |
| Control Test | 0.079 | 18:06 |
| Air Blank | 0.000 | 18:07 |
| Subject Sample # 1 | 0.000 | 18:07 |
| Breath Volume | 1.175 | 18:07 |
| Air Blank | 0.000 | 18:07 |
| Air Blank | 0.000 | 18:09 |
| Subject Sample # 2 | 0.000 | 18:10 |
| Breath Volume | 2.496 | 18:10 |
| Air Blank | 0.000 | 18:11 |
| Control Test | 0.079 | 18:11 |
| Air Blank | 0.000 | 18:12 |
| Diagnostics Check | OK | 18:12 |

Data Download Date/Time: 07/29/2020 08:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: ABELARDO

DATE/TIME OF TEST: 07/25/2020 01:45:01

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:38 |
| Air Blank | 0.000 | 01:39 |
| Control Test | 0.079 | 01:39 |
| Air Blank | 0.000 | 01:40 |
| Subject Sample # 1 | 0.174 | 01:40 |
| Breath Volume | 2.539 | 01:40 |
| Air Blank | 0.000 | 01:41 |
| Air Blank | 0.000 | 01:43 |
| Subject Sample # 2 | 0.175 | 01:43 |
| Breath Volume | 2.218 | 01:43 |
| Air Blank | 0.000 | 01:44 |
| Control Test | 0.078 | 01:44 |
| Air Blank | 0.000 | 01:44 |
| Diagnostics Check | OK | 01:45 |

Data Download Date/Time: 07/29/2020 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/26/2020 05:17:50

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 05:11 |
| Air Blank | 0.000 | 05:11 |
| Control Test | 0.079 | 05:12 |
| Air Blank | 0.000 | 05:12 |
| Subject Sample # 1 | 0.180 | 05:13 |
| Breath Volume | 2.656 | 05:13 |
| Air Blank | 0.000 | 05:13 |
| Air Blank | 0.000 | 05:15 |
| Subject Sample # 2 | 0.171 | 05:16 |
| Breath Volume | 2.300 | 05:16 |
| Air Blank | 0.000 | 05:16 |
| Control Test | 0.079 | 05:17 |
| Air Blank | 0.000 | 05:17 |
| Diagnostics Check | OK | 05:17 |

Data Download Date/Time: 07/29/2020 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: BRANKIN

DATE/TIME OF TEST: 07/01/2020 00:14:19

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:08 |
| Air Blank | 0.000 | 00:08 |
| Control Test | 0.077 | 00:08 |
| Air Blank | 0.000 | 00:09 |
| Subject Sample # 1 | 0.272 | 00:09 |
| Breath Volume | 2.738 | 00:09 |
| Air Blank | 0.000 | 00:10 |
| Air Blank | 0.000 | 00:12 |
| Subject Sample # 2 | 0.272 | 00:12 |
| Breath Volume | 3.027 | 00:12 |
| Air Blank | 0.000 | 00:13 |
| Control Test | 0.077 | 00:13 |
| Air Blank | 0.000 | 00:14 |
| Diagnostics Check | OK | 00:14 |

Data Download Date/Time: 07/29/2020 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 07/09/2020 17:07:37

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 17:00 |
| Air Blank | | 0.000 | 17:00 |
| Control Test | | 0.078 | 17:01 |
| Air Blank | | 0.000 | 17:01 |
| Subject Sample | # 1 | 0.248 | 17:02 |
| Breath Volume | | 1.808 | 17:02 |
| Air Blank | | 0.000 | 17:03 |
| Air Blank | | 0.000 | 17:04 |
| Subject Sample | # 2 | 0.254 | 17:06 |
| Breath Volume | | 1.757 | 17:06 |
| Air Blank | | 0.000 | 17:06 |
| Control Test | | 0.079 | 17:07 |
| Air Blank | | 0.000 | 17:07 |
| Diagnostics Check | | OK | 17:07 |

Data Download Date/Time: 07/29/2020 08:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: CAPASSO

DATE/TIME OF TEST: 07/14/2020 00:52:53

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 00:46 |
| Air Blank | | 0.000 | 00:46 |
| Control Test | | 0.078 | 00:46 |
| Air Blank | | 0.000 | 00:47 |
| Subject Sample | # 1 | 0.249 | 00:48 |
| Breath Volume | | 1.902 | 00:48 |
| Air Blank | | 0.000 | 00:48 |
| Air Blank | | 0.000 | 00:50 |
| Subject Sample | # 2 | 0.253 | 00:51 |
| Breath Volume | | 2.468 | 00:51 |
| Air Blank | | 0.000 | 00:51 |
| Control Test | | 0.077 | 00:52 |
| Air Blank | | 0.000 | 00:52 |
| Diagnostics Check | | OK | 00:52 |

Data Download Date/Time: 07/29/2020 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NELSON,OFC
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: KAPONIS

DATE/TIME OF TEST: 07/17/2020 21:46:14

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:39 |
| Air Blank | 0.000 | 21:40 |
| Control Test | 0.079 | 21:40 |
| Air Blank | 0.000 | 21:41 |
| Subject Sample # 1 | 0.000 | 21:41 |
| Breath Volume | 1.507 | 21:41 |
| Air Blank | 0.000 | 21:42 |
| Air Blank | 0.000 | 21:44 |
| Subject Sample # 2 | 0.000 | 21:44 |
| Breath Volume | 1.820 | 21:44 |
| Air Blank | 0.000 | 21:45 |
| Control Test | 0.077 | 21:45 |
| Air Blank | 0.000 | 21:46 |
| Diagnostics Check | OK | 21:46 |

Data Download Date/Time: 07/29/2020 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: COLAIZZI

DATE/TIME OF TEST: 07/18/2020 02:22:09

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 02:15 |
| Air Blank | | 0.000 | 02:16 |
| Control Test | | 0.078 | 02:16 |
| Air Blank | | 0.000 | 02:16 |
| Subject Sample | # 1 | 0.227 | 02:17 |
| Breath Volume | | 2.125 | 02:17 |
| Air Blank | | 0.000 | 02:18 |
| Air Blank | | 0.000 | 02:19 |
| Subject Sample | # 2 | 0.222 | 02:20 |
| Breath Volume | | 2.250 | 02:20 |
| Air Blank | | 0.000 | 02:21 |
| Control Test | | 0.079 | 02:21 |
| Air Blank | | 0.000 | 02:22 |
| Diagnostics Check | | OK | 02:22 |

Data Download Date/Time: 07/29/2020 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ROWELL,OFC
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 07/19/2020 03:36:57

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 03:30 |
| Air Blank | | 0.000 | 03:30 |
| Control Test | | 0.078 | 03:31 |
| Air Blank | | 0.000 | 03:31 |
| Subject Sample | # 1 | 0.133 | 03:32 |
| Breath Volume | | 2.093 | 03:32 |
| Air Blank | | 0.000 | 03:33 |
| Air Blank | | 0.000 | 03:35 |
| Subject Sample | # 2 | 0.137 | 03:35 |
| Breath Volume | | 2.082 | 03:35 |
| Air Blank | | 0.000 | 03:36 |
| Control Test | | 0.079 | 03:36 |
| Air Blank | | 0.000 | 03:36 |
| Diagnostics Check | | OK | 03:36 |

Data Download Date/Time: 07/29/2020 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: CHILDRESS

DATE/TIME OF TEST: 07/23/2020 20:27:06

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:20 |
| Air Blank | 0.000 | 20:20 |
| Control Test | 0.078 | 20:21 |
| Air Blank | 0.000 | 20:21 |
| Subject Sample # 1 | 0.306 | 20:22 |
| Breath Volume | 3.000 | 20:22 |
| Air Blank | 0.000 | 20:23 |
| Air Blank | 0.000 | 20:25 |
| Subject Sample # 2 | 0.292 | 20:25 |
| Breath Volume | 1.308 | 20:25 |
| Air Blank | 0.000 | 20:26 |
| Control Test | 0.077 | 20:26 |
| Air Blank | 0.000 | 20:27 |
| Diagnostics Check | OK | 20:27 |

Data Download Date/Time: 07/29/2020 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: HASTON

DATE/TIME OF TEST: 07/27/2020 20:53:55

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:47 |
| Air Blank | 0.000 | 20:48 |
| Control Test | 0.079 | 20:48 |
| Air Blank | 0.000 | 20:48 |
| Subject Sample # 1 | 0.000 | 20:49 |
| Breath Volume | 1.597 | 20:49 |
| Air Blank | 0.000 | 20:49 |
| Air Blank | 0.000 | 20:51 |
| Subject Sample # 2 | 0.000 | 20:52 |
| Breath Volume | 2.050 | 20:52 |
| Air Blank | 0.000 | 20:53 |
| Control Test | 0.078 | 20:53 |
| Air Blank | 0.000 | 20:53 |
| Diagnostics Check | OK | 20:53 |

Data Download Date/Time: 07/29/2020 08:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 07/28/2020 19:48:20

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 19:42 |
| Air Blank | 0.000 | 19:43 |
| Control Test | 0.078 | 19:43 |
| Air Blank | 0.000 | 19:43 |
| Subject Sample # 1 | 0.000 | 19:44 |
| Breath Volume | 1.953 | 19:44 |
| Air Blank | 0.000 | 19:44 |
| Air Blank | 0.000 | 19:46 |
| Subject Sample # 2 | 0.000 | 19:46 |
| Breath Volume | 1.824 | 19:46 |
| Air Blank | 0.000 | 19:47 |
| Control Test | 0.078 | 19:47 |
| Air Blank | 0.000 | 19:48 |
| Diagnostics Check | OK | 19:48 |

Data Download Date/Time: 07/29/2020 08:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACEBO,ANDRES,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CIOLINO,L
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/02/2020 22:15:45

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:08 |
| Air Blank | 0.000 | 22:09 |
| Control Test | 0.080 | 22:09 |
| Air Blank | 0.000 | 22:09 |
| Subject Sample # 1 | 0.098 | 22:11 |
| Breath Volume | 2.226 | 22:11 |
| Air Blank | 0.000 | 22:11 |
| Air Blank | 0.000 | 22:13 |
| Subject Sample # 2 | 0.100 | 22:14 |
| Breath Volume | 3.492 | 22:14 |
| Air Blank | 0.000 | 22:14 |
| Control Test | 0.077 | 22:15 |
| Air Blank | 0.000 | 22:15 |
| Diagnostics Check | OK | 22:15 |

Data Download Date/Time: 07/23/2020 23:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ANDERSON,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: PERISCO

DATE/TIME OF TEST: 07/04/2020 23:01:23

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:53 |
| Air Blank | 0.000 | 22:53 |
| Control Test | 0.080 | 22:54 |
| Air Blank | 0.000 | 22:54 |
| Subject Sample # 1 | 0.236 | 22:56 |
| Breath Volume | 2.187 | 22:56 |
| Air Blank | 0.000 | 22:57 |
| Air Blank | 0.000 | 22:59 |
| Subject Sample # 2 | 0.255 | 22:59 |
| Breath Volume | 1.925 | 22:59 |
| Air Blank | 0.000 | 23:00 |
| Control Test | 0.080 | 23:00 |
| Air Blank | 0.000 | 23:01 |
| Diagnostics Check | OK | 23:01 |

Data Download Date/Time: 07/23/2020 23:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DAVIDSON,MICHAEL
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: WOJTASZEK

DATE/TIME OF TEST: 07/07/2020 00:08:32

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 00:05 | |
| Air Blank | 0.000 | 00:06 | |
| Control Test | 0.080 | 00:06 | |
| Air Blank | 0.000 | 00:07 | |
| Subject Sample | # 1 | 00:07 | Subject Test Refused |
| Breath Volume | 0.000 | 00:07 | |
| Air Blank | 0.000 | 00:07 | |
| Control Test | 0.080 | 00:07 | |
| Air Blank | 0.000 | 00:08 | |
| Diagnostics Check | OK | 00:08 | |

Data Download Date/Time: 07/23/2020 23:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARKINS,CLINTON,T
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CURTIS,R
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: DEVERE

DATE/TIME OF TEST: 07/07/2020 03:21:04

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:14 |
| Air Blank | 0.000 | 03:14 |
| Control Test | 0.080 | 03:15 |
| Air Blank | 0.000 | 03:15 |
| Subject Sample # 1 | 0.007 | 03:16 |
| Breath Volume | 2.117 | 03:16 |
| Air Blank | 0.000 | 03:16 |
| Air Blank | 0.000 | 03:18 |
| Subject Sample # 2 | 0.000 | 03:19 |
| Breath Volume | 1.511 | 03:19 |
| Air Blank | 0.000 | 03:20 |
| Control Test | 0.080 | 03:20 |
| Air Blank | 0.000 | 03:20 |
| Diagnostics Check | OK | 03:21 |

Data Download Date/Time: 07/23/2020 23:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ANDERSON,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: SILVER

DATE/TIME OF TEST: 07/07/2020 19:20:16

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 19:13 |
| Air Blank | 0.000 | 19:13 |
| Control Test | 0.080 | 19:14 |
| Air Blank | 0.000 | 19:14 |
| Subject Sample # 1 | 0.126 | 19:15 |
| Breath Volume | 2.082 | 19:15 |
| Air Blank | 0.000 | 19:16 |
| Air Blank | 0.000 | 19:17 |
| Subject Sample # 2 | 0.119 | 19:18 |
| Breath Volume | 1.679 | 19:18 |
| Air Blank | 0.000 | 19:19 |
| Control Test | 0.080 | 19:19 |
| Air Blank | 0.000 | 19:20 |
| Diagnostics Check | OK | 19:20 |

Data Download Date/Time: 07/23/2020 23:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: OWENS,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: LEGGE

DATE/TIME OF TEST: 07/11/2020 01:25:37

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 01:18 |
| Air Blank | | 0.000 | 01:19 |
| Control Test | | 0.080 | 01:19 |
| Air Blank | | 0.000 | 01:19 |
| Subject Sample | # 1 | 0.194 | 01:20 |
| Breath Volume | | 4.160 | 01:20 |
| Air Blank | | 0.000 | 01:21 |
| Air Blank | | 0.000 | 01:23 |
| Subject Sample | # 2 | 0.193 | 01:24 |
| Breath Volume | | 4.164 | 01:24 |
| Air Blank | | 0.000 | 01:24 |
| Control Test | | 0.079 | 01:25 |
| Air Blank | | 0.000 | 01:25 |
| Diagnostics Check | | OK | 01:25 |

Data Download Date/Time: 07/23/2020 23:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEUTHNER,N,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LEUTHNER,N
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: EASLEY

DATE/TIME OF TEST: 07/11/2020 18:55:04

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 18:48 |
| Air Blank | 0.000 | 18:49 |
| Control Test | 0.081 | 18:49 |
| Air Blank | 0.000 | 18:49 |
| Subject Sample # 1 | 0.161 | 18:50 |
| Breath Volume | 1.472 | 18:50 |
| Air Blank | 0.000 | 18:51 |
| Air Blank | 0.000 | 18:53 |
| Subject Sample # 2 | 0.152 | 18:53 |
| Breath Volume | 2.031 | 18:53 |
| Air Blank | 0.000 | 18:54 |
| Control Test | 0.079 | 18:54 |
| Air Blank | 0.000 | 18:54 |
| Diagnostics Check | OK | 18:55 |

Data Download Date/Time: 07/23/2020 23:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MADDAUGH,OFF
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: LANSDALE

DATE/TIME OF TEST: 07/12/2020 00:25:07

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:18 |
| Air Blank | 0.000 | 00:19 |
| Control Test | 0.080 | 00:19 |
| Air Blank | 0.000 | 00:19 |
| Subject Sample # 1 | 0.000 | 00:20 |
| Breath Volume | 1.851 | 00:20 |
| Air Blank | 0.000 | 00:21 |
| Air Blank | 0.000 | 00:23 |
| Subject Sample # 2 | 0.000 | 00:23 |
| Breath Volume | 2.128 | 00:23 |
| Air Blank | 0.000 | 00:24 |
| Control Test | 0.081 | 00:24 |
| Air Blank | 0.000 | 00:25 |
| Diagnostics Check | OK | 00:25 |

Data Download Date/Time: 07/23/2020 23:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORITZ,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LEUTHNER,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 07/12/2020 03:58:40

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:51 |
| Air Blank | 0.000 | 03:51 |
| Control Test | 0.080 | 03:52 |
| Air Blank | 0.000 | 03:52 |
| Subject Sample # 1 | 0.266 | 03:53 |
| Breath Volume | 3.722 | 03:53 |
| Air Blank | 0.000 | 03:54 |
| Air Blank | 0.000 | 03:56 |
| Subject Sample # 2 | 0.255 | 03:57 |
| Breath Volume | 2.820 | 03:57 |
| Air Blank | 0.000 | 03:57 |
| Control Test | 0.080 | 03:58 |
| Air Blank | 0.000 | 03:58 |
| Diagnostics Check | OK | 03:58 |

Data Download Date/Time: 07/23/2020 23:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUCKER,DEPUTY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HEGWOOD,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: MENINGA

DATE/TIME OF TEST: 07/12/2020 17:12:49

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 17:04 |
| Air Blank | 0.000 | 17:05 |
| Control Test | 0.080 | 17:05 |
| Air Blank | 0.000 | 17:06 |
| Subject Sample # 1 | 0.120 | 17:07 |
| Breath Volume | 1.539 | 17:07 |
| Air Blank | 0.000 | 17:08 |
| Air Blank | 0.000 | 17:09 |
| Subject Sample # 2 | 0.113 | 17:11 |
| Breath Volume | 1.496 | 17:11 |
| Air Blank | 0.000 | 17:11 |
| Control Test | 0.079 | 17:12 |
| Air Blank | 0.000 | 17:12 |
| Diagnostics Check | OK | 17:12 |

Data Download Date/Time: 07/23/2020 23:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORITZ,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: THABIT,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/13/2020 05:23:14

| | | | |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK | 05:09 | |
| Air Blank | 0.000 | 05:09 | |
| Control Test | 0.080 | 05:10 | |
| Air Blank | 0.000 | 05:10 | |
| Subject Sample | # 1 | 05:13 | No Sample Provided |
| Breath Volume | 0.000 | 05:13 | |
| Air Blank | 0.000 | 05:14 | |
| Air Blank | 0.000 | 05:16 | |
| Subject Sample | # 2 | 0.094 | 05:17 |
| Breath Volume | 1.417 | 05:17 | |
| Air Blank | 0.000 | 05:18 | |
| Air Blank | 0.000 | 05:20 | |
| Subject Sample | # 3 | 0.088 | 05:21 |
| Breath Volume | 1.914 | 05:21 | |
| Air Blank | 0.000 | 05:22 | |
| Control Test | 0.080 | 05:22 | |
| Air Blank | 0.000 | 05:23 | |
| Diagnostics Check | OK | 05:23 | |

Data Download Date/Time: 07/23/2020 23:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCILLA,RICHARD,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TORRES,JUAN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: SEARS III

DATE/TIME OF TEST: 07/17/2020 22:09:01

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:02 |
| Air Blank | 0.000 | 22:03 |
| Control Test | 0.080 | 22:03 |
| Air Blank | 0.000 | 22:04 |
| Subject Sample # 1 | 0.164 | 22:04 |
| Breath Volume | 2.214 | 22:04 |
| Air Blank | 0.000 | 22:05 |
| Air Blank | 0.000 | 22:06 |
| Subject Sample # 2 | 0.166 | 22:07 |
| Breath Volume | 2.230 | 22:07 |
| Air Blank | 0.000 | 22:08 |
| Control Test | 0.080 | 22:08 |
| Air Blank | 0.000 | 22:08 |
| Diagnostics Check | OK | 22:09 |

Data Download Date/Time: 07/23/2020 23:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,HENRY,E
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DELHOMME,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: BUNNELL

DATE/TIME OF TEST: 07/22/2020 17:50:14

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 17:44 |
| Air Blank | 0.000 | 17:44 |
| Control Test | 0.079 | 17:44 |
| Air Blank | 0.000 | 17:45 |
| Subject Sample # 1 | 0.287 | 17:45 |
| Breath Volume | 2.636 | 17:45 |
| Air Blank | 0.000 | 17:46 |
| Air Blank | 0.000 | 17:48 |
| Subject Sample # 2 | 0.279 | 17:48 |
| Breath Volume | 2.519 | 17:48 |
| Air Blank | 0.000 | 17:49 |
| Control Test | 0.078 | 17:49 |
| Air Blank | 0.000 | 17:50 |
| Diagnostics Check | OK | 17:50 |

Data Download Date/Time: 07/23/2020 23:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAFFARAS,CELINA,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MATZICOFF,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: WAID

DATE/TIME OF TEST: 07/23/2020 20:32:25

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:25 |
| Air Blank | 0.000 | 20:26 |
| Control Test | 0.080 | 20:26 |
| Air Blank | 0.000 | 20:26 |
| Subject Sample # 1 | 0.240 | 20:27 |
| Breath Volume | 1.925 | 20:27 |
| Air Blank | 0.000 | 20:28 |
| Air Blank | 0.000 | 20:30 |
| Subject Sample # 2 | 0.250 | 20:30 |
| Breath Volume | 1.324 | 20:30 |
| Air Blank | 0.000 | 20:31 |
| Control Test | 0.080 | 20:31 |
| Air Blank | 0.000 | 20:32 |
| Diagnostics Check | OK | 20:32 |

Data Download Date/Time: 07/23/2020 23:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORDSTROM,DEPUTY,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JANKOWSKI,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 07/24/2020 06:36:23

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 06:30 |
| Air Blank | 0.000 | 06:30 |
| Control Test | 0.081 | 06:31 |
| Air Blank | 0.000 | 06:31 |
| Subject Sample # 1 | 0.133 | 06:32 |
| Breath Volume | 2.562 | 06:32 |
| Air Blank | 0.000 | 06:32 |
| Air Blank | 0.000 | 06:34 |
| Subject Sample # 2 | 0.136 | 06:34 |
| Breath Volume | 2.148 | 06:34 |
| Air Blank | 0.000 | 06:35 |
| Control Test | 0.080 | 06:35 |
| Air Blank | 0.000 | 06:36 |
| Diagnostics Check | OK | 06:36 |

Data Download Date/Time: 08/17/2020 02:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONGA,DEPUTY,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CUADRA,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: GORMAN

DATE/TIME OF TEST: 07/26/2020 17:47:48

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 17:41 |
| Air Blank | 0.000 | 17:41 |
| Control Test | 0.080 | 17:41 |
| Air Blank | 0.000 | 17:42 |
| Subject Sample # 1 | 0.270 | 17:43 |
| Breath Volume | 1.890 | 17:43 |
| Air Blank | 0.000 | 17:43 |
| Air Blank | 0.000 | 17:45 |
| Subject Sample # 2 | 0.266 | 17:46 |
| Breath Volume | 1.367 | 17:46 |
| Air Blank | 0.000 | 17:46 |
| Control Test | 0.080 | 17:47 |
| Air Blank | 0.000 | 17:47 |
| Diagnostics Check | OK | 17:47 |

Data Download Date/Time: 08/17/2020 02:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARKINS,CLINTON,T
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: STEALEY,L
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: KOUKIN

DATE/TIME OF TEST: 07/26/2020 19:44:24

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 19:37 |
| Air Blank | 0.000 | 19:37 |
| Control Test | 0.080 | 19:38 |
| Air Blank | 0.000 | 19:38 |
| Subject Sample # 1 | 0.150 | 19:39 |
| Breath Volume | 1.398 | 19:39 |
| Air Blank | 0.000 | 19:40 |
| Air Blank | 0.000 | 19:42 |
| Subject Sample # 2 | 0.147 | 19:42 |
| Breath Volume | 1.953 | 19:42 |
| Air Blank | 0.000 | 19:43 |
| Control Test | 0.080 | 19:43 |
| Air Blank | 0.000 | 19:44 |
| Diagnostics Check | OK | 19:44 |

Data Download Date/Time: 08/17/2020 02:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MEDDAUGH,B
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 07/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: KINGTON

DATE/TIME OF TEST: 07/26/2020 21:19:28

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 21:06 | |
| Air Blank | | 0.000 | 21:07 | |
| Control Test | | 0.080 | 21:07 | |
| Air Blank | | 0.000 | 21:08 | |
| Subject Sample | # 1 | 0.000 | 21:11 | Volume Not Met |
| Breath Volume | | 0.023 | 21:11 | |
| Air Blank | | 0.000 | 21:11 | |
| Air Blank | | 0.000 | 21:13 | |
| Subject Sample | # 2 | 0.143 | 21:14 | |
| Breath Volume | | 1.285 | 21:14 | |
| Air Blank | | 0.000 | 21:15 | |
| Air Blank | | 0.000 | 21:17 | |
| Subject Sample | # 3 | 0.150 | 21:17 | |
| Breath Volume | | 1.199 | 21:17 | |
| Air Blank | | 0.000 | 21:18 | |
| Control Test | | 0.080 | 21:18 | |
| Air Blank | | 0.000 | 21:19 | |
| Diagnostics Check | | OK | 21:19 | |

Data Download Date/Time: 08/17/2020 02:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACEBO,ANDRES,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LEUTHNER,N
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: JACINTO CORTES

DATE/TIME OF TEST: 07/27/2020 04:35:54

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 04:25 |
| Air Blank | | 0.000 | 04:25 |
| Control Test | | 0.080 | 04:26 |
| Air Blank | | 0.000 | 04:26 |
| Subject Sample | # 1 | 0.271 | 04:27 |
| Breath Volume | | 2.773 | 04:27 |
| Air Blank | | 0.000 | 04:28 |
| Air Blank | | 0.000 | 04:30 |
| Subject Sample | # 2 | 0.240 | 04:31 |
| Breath Volume | | 1.476 | 04:31 |
| Air Blank | | 0.000 | 04:31 |
| Air Blank | | 0.000 | 04:33 |
| Subject Sample | # 3 | 0.258 | 04:34 |
| Breath Volume | | 2.679 | 04:34 |
| Air Blank | | 0.000 | 04:35 |
| Control Test | | 0.080 | 04:35 |
| Air Blank | | 0.000 | 04:35 |
| Diagnostics Check | | OK | 04:35 |

Data Download Date/Time: 08/17/2020 02:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEPARD,C,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTORFER,C
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: QUEEN

DATE/TIME OF TEST: 07/27/2020 15:32:11

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 15:26 |
| Air Blank | 0.000 | 15:26 |
| Control Test | 0.080 | 15:26 |
| Air Blank | 0.000 | 15:27 |
| Subject Sample # 1 | 0.000 | 15:27 |
| Breath Volume | 2.550 | 15:27 |
| Air Blank | 0.000 | 15:28 |
| Air Blank | 0.000 | 15:30 |
| Subject Sample # 2 | 0.000 | 15:30 |
| Breath Volume | 2.175 | 15:30 |
| Air Blank | 0.000 | 15:31 |
| Control Test | 0.079 | 15:31 |
| Air Blank | 0.000 | 15:32 |
| Diagnostics Check | OK | 15:32 |

Data Download Date/Time: 08/17/2020 02:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: FISH

DATE/TIME OF TEST: 07/04/2020 01:42:07

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:35 |
| Air Blank | 0.000 | 01:36 |
| Control Test | 0.079 | 01:36 |
| Air Blank | 0.000 | 01:37 |
| Subject Sample # 1 | 0.167 | 01:37 |
| Breath Volume | 2.375 | 01:37 |
| Air Blank | 0.000 | 01:38 |
| Air Blank | 0.000 | 01:40 |
| Subject Sample # 2 | 0.159 | 01:40 |
| Breath Volume | 1.937 | 01:40 |
| Air Blank | 0.000 | 01:41 |
| Control Test | 0.079 | 01:41 |
| Air Blank | 0.000 | 01:42 |
| Diagnostics Check | OK | 01:42 |

Data Download Date/Time: 07/29/2020 01:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLON,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: DILLON,JUSTIN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 07/19/2020 21:14:47

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 21:07 |
| Air Blank | | 0.000 | 21:08 |
| Control Test | | 0.080 | 21:08 |
| Air Blank | | 0.000 | 21:08 |
| Subject Sample | # 1 | 0.261 | 21:10 |
| Breath Volume | | 1.539 | 21:10 |
| Air Blank | | 0.000 | 21:10 |
| Air Blank | | 0.000 | 21:12 |
| Subject Sample | # 2 | 0.244 | 21:13 |
| Breath Volume | | 1.257 | 21:13 |
| Air Blank | | 0.000 | 21:13 |
| Control Test | | 0.079 | 21:14 |
| Air Blank | | 0.000 | 21:14 |
| Diagnostics Check | | OK | 21:14 |

Data Download Date/Time: 07/29/2020 01:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: CHRISTIE,LUCAS
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: LEMMONS

DATE/TIME OF TEST: 07/24/2020 00:39:31

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 00:33 |
| Air Blank | | 0.000 | 00:34 |
| Control Test | | 0.079 | 00:34 |
| Air Blank | | 0.000 | 00:34 |
| Subject Sample | # 1 | 0.134 | 00:35 |
| Breath Volume | | 2.070 | 00:35 |
| Air Blank | | 0.000 | 00:35 |
| Air Blank | | 0.000 | 00:37 |
| Subject Sample | # 2 | 0.135 | 00:38 |
| Breath Volume | | 1.800 | 00:38 |
| Air Blank | | 0.000 | 00:38 |
| Control Test | | 0.079 | 00:38 |
| Air Blank | | 0.000 | 00:39 |
| Diagnostics Check | | OK | 00:39 |

Data Download Date/Time: 07/29/2020 01:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SHUTTERLY,BROOKE
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 07/27/2020 03:22:15

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:12 |
| Air Blank | 0.000 | 03:13 |
| Control Test | 0.080 | 03:13 |
| Air Blank | 0.000 | 03:14 |
| Subject Sample # 1 | 0.211 | 03:15 |
| Breath Volume | 1.808 | 03:15 |
| Air Blank | 0.000 | 03:16 |
| Air Blank | 0.000 | 03:18 |
| Subject Sample # 2 | 0.210 | 03:20 |
| Breath Volume | 1.265 | 03:20 |
| Air Blank | 0.000 | 03:21 |
| Control Test | 0.080 | 03:21 |
| Air Blank | 0.000 | 03:22 |
| Diagnostics Check | OK | 03:22 |

Data Download Date/Time: 07/29/2020 01:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,SCOTT,C
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: MIMS

DATE/TIME OF TEST: 07/01/2020 21:55:40

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 21:52 | |
| Air Blank | 0.000 | 21:53 | |
| Control Test | 0.080 | 21:53 | |
| Air Blank | 0.000 | 21:54 | |
| Subject Sample | # 1 | 21:54 | Subject Test Refused |
| Breath Volume | 0.000 | 21:54 | |
| Air Blank | 0.000 | 21:54 | |
| Control Test | 0.080 | 21:55 | |
| Air Blank | 0.000 | 21:55 | |
| Diagnostics Check | OK | 21:55 | |

Data Download Date/Time: 08/01/2020 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NETTLES, JONATHAN, K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: NETTLES, JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 10:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: DYER

DATE/TIME OF TEST: 07/03/2020 01:08:39

| | | | |
|--------------------|-------|-------|--------------------|
| Diagnostics Check | OK | 00:54 | |
| Air Blank | 0.000 | 00:54 | |
| Control Test | 0.080 | 00:55 | |
| Air Blank | 0.000 | 00:55 | |
| Subject Sample # 1 | 0.144 | 00:56 | |
| Breath Volume | 2.582 | 00:56 | |
| Air Blank | 0.000 | 00:56 | |
| Air Blank | 0.000 | 00:58 | |
| Subject Sample # 2 | | 01:01 | No Sample Provided |
| Breath Volume | 0.000 | 01:01 | |
| Air Blank | 0.000 | 01:02 | |
| Air Blank | 0.000 | 01:04 | |
| Subject Sample # 3 | | 01:07 | No Sample Provided |
| Breath Volume | 0.000 | 01:07 | |
| Air Blank | 0.000 | 01:07 | |
| Control Test | 0.080 | 01:08 | |
| Air Blank | 0.000 | 01:08 | |
| Diagnostics Check | OK | 01:08 | |

Data Download Date/Time: 08/01/2020 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANESH,TOTORAM,G
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: THOMAS,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:32

SUBJECT'S LAST NAME: GANESH

DATE/TIME OF TEST: 07/03/2020 16:00:37

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 15:52 |
| Air Blank | 0.000 | 15:53 |
| Control Test | 0.080 | 15:53 |
| Air Blank | 0.000 | 15:54 |
| Subject Sample # 1 | 0.118 | 15:55 |
| Breath Volume | 2.476 | 15:55 |
| Air Blank | 0.000 | 15:56 |
| Air Blank | 0.000 | 15:58 |
| Subject Sample # 2 | 0.111 | 15:59 |
| Breath Volume | 1.457 | 15:59 |
| Air Blank | 0.000 | 15:59 |
| Control Test | 0.078 | 16:00 |
| Air Blank | 0.000 | 16:00 |
| Diagnostics Check | OK | 16:00 |

Data Download Date/Time: 08/01/2020 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWLING, DAKOTA, A
BREATH TEST OPERATOR AGENCY: Cedar Keys PD
ARRESTING OFFICER: DOWLING, DAKOTA
ARRESTING OFFICER AGENCY: Cedar Keys PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/18/2020 01:14:36

| | | | |
|--------------------|-------|-------|-----------------|
| Diagnostics Check | OK | 01:08 | |
| Air Blank | 0.000 | 01:08 | |
| Control Test | 0.079 | 01:09 | |
| Air Blank | 0.000 | 01:09 | |
| Subject Sample # 1 | 0.172 | 01:10 | |
| Breath Volume | 1.832 | 01:10 | |
| Air Blank | 0.000 | 01:11 | |
| Air Blank | 0.000 | 01:13 | |
| Subject Sample # 2 | 0.185 | 01:13 | |
| Breath Volume | 2.031 | 01:13 | |
| Air Blank | | 01:14 | Improper Sample |
| Air Blank | 0.000 | 01:14 | |

Data Download Date/Time: 08/01/2020 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANUS,GLEN,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GANUS,GLEN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: FLOYD

DATE/TIME OF TEST: 07/19/2020 22:17:33

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:11 |
| Air Blank | 0.000 | 22:11 |
| Control Test | 0.079 | 22:11 |
| Air Blank | 0.000 | 22:12 |
| Subject Sample # 1 | 0.228 | 22:12 |
| Breath Volume | 1.839 | 22:12 |
| Air Blank | 0.000 | 22:13 |
| Air Blank | 0.000 | 22:15 |
| Subject Sample # 2 | 0.239 | 22:15 |
| Breath Volume | 1.984 | 22:15 |
| Air Blank | 0.000 | 22:16 |
| Control Test | 0.078 | 22:16 |
| Air Blank | 0.000 | 22:17 |
| Diagnostics Check | OK | 22:17 |

Data Download Date/Time: 08/01/2020 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MICHAEL,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: THOMAS,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 07/27/2020 19:06:08

| | | | |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK | 19:05 | |
| Air Blank | | 19:05 | RFI Detect |
| Air Blank | 0.000 | 19:06 | |

Data Download Date/Time: 08/01/2020 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MICHAEL,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: THOMAS,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 07/27/2020 19:11:51

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 19:08 | |
| Air Blank | 0.000 | 19:09 | |
| Control Test | 0.080 | 19:09 | |
| Air Blank | 0.000 | 19:10 | |
| Subject Sample | # 1 | 19:10 | Subject Test Refused |
| Breath Volume | 0.000 | 19:10 | |
| Air Blank | 0.000 | 19:10 | |
| Control Test | 0.080 | 19:11 | |
| Air Blank | 0.000 | 19:11 | |
| Diagnostics Check | OK | 19:11 | |

Data Download Date/Time: 08/01/2020 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLOSKA,EDWARD,J
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: POLLOCK,DEREK
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 07/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 07/23/2020 23:30:59

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:25 |
| Air Blank | 0.000 | 23:25 |
| Control Test | 0.079 | 23:25 |
| Air Blank | 0.000 | 23:26 |
| Subject Sample # 1 | 0.244 | 23:26 |
| Breath Volume | 3.480 | 23:26 |
| Air Blank | 0.000 | 23:27 |
| Air Blank | 0.000 | 23:29 |
| Subject Sample # 2 | 0.250 | 23:29 |
| Breath Volume | 3.742 | 23:29 |
| Air Blank | 0.000 | 23:30 |
| Control Test | 0.079 | 23:30 |
| Air Blank | 0.000 | 23:30 |
| Diagnostics Check | OK | 23:30 |

Data Download Date/Time: 08/13/2020 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTINEZ,YAISEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: SIQUE

DATE/TIME OF TEST: 07/11/2020 03:23:56

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 03:17 |
| Air Blank | | 0.000 | 03:18 |
| Control Test | | 0.080 | 03:18 |
| Air Blank | | 0.000 | 03:19 |
| Subject Sample | # 1 | 0.156 | 03:19 |
| Breath Volume | | 2.281 | 03:19 |
| Air Blank | | 0.000 | 03:20 |
| Air Blank | | 0.000 | 03:22 |
| Subject Sample | # 2 | 0.159 | 03:22 |
| Breath Volume | | 2.839 | 03:22 |
| Air Blank | | 0.000 | 03:23 |
| Control Test | | 0.080 | 03:23 |
| Air Blank | | 0.000 | 03:23 |
| Diagnostics Check | | OK | 03:23 |

Data Download Date/Time: 07/16/2020 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ, JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: ZAMORA

DATE/TIME OF TEST: 07/12/2020 05:25:13

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 05:18 |
| Air Blank | 0.000 | 05:19 |
| Control Test | 0.081 | 05:19 |
| Air Blank | 0.000 | 05:20 |
| Subject Sample # 1 | 0.007 | 05:20 |
| Breath Volume | 1.390 | 05:20 |
| Air Blank | 0.000 | 05:21 |
| Air Blank | 0.000 | 05:23 |
| Subject Sample # 2 | 0.000 | 05:23 |
| Breath Volume | 1.222 | 05:23 |
| Air Blank | 0.000 | 05:24 |
| Control Test | 0.080 | 05:24 |
| Air Blank | 0.000 | 05:25 |
| Diagnostics Check | OK | 05:25 |

Data Download Date/Time: 07/16/2020 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CARPENTER,BLAKE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/15/2020 02:57:10

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:51 |
| Air Blank | 0.000 | 02:51 |
| Control Test | 0.080 | 02:51 |
| Air Blank | 0.000 | 02:52 |
| Subject Sample # 1 | 0.169 | 02:52 |
| Breath Volume | 1.976 | 02:52 |
| Air Blank | 0.000 | 02:53 |
| Air Blank | 0.000 | 02:55 |
| Subject Sample # 2 | 0.176 | 02:55 |
| Breath Volume | 1.683 | 02:55 |
| Air Blank | 0.000 | 02:56 |
| Control Test | 0.079 | 02:56 |
| Air Blank | 0.000 | 02:57 |
| Diagnostics Check | OK | 02:57 |

Data Download Date/Time: 07/16/2020 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JEANBAPTISTE,ALBENS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 07/17/2020 00:35:28

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 00:32 | |
| Air Blank | 0.000 | 00:33 | |
| Control Test | 0.079 | 00:33 | |
| Air Blank | 0.000 | 00:33 | |
| Subject Sample | # 1 | 00:34 | Subject Test Refused |
| Breath Volume | 0.000 | 00:34 | |
| Air Blank | 0.000 | 00:34 | |
| Control Test | 0.078 | 00:34 | |
| Air Blank | 0.000 | 00:35 | |
| Diagnostics Check | OK | 00:35 | |

Data Download Date/Time: 07/23/2020 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERN,ETHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: SIMONS

DATE/TIME OF TEST: 07/18/2020 04:28:30

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 04:22 |
| Air Blank | | 0.000 | 04:22 |
| Control Test | | 0.080 | 04:22 |
| Air Blank | | 0.000 | 04:23 |
| Subject Sample | # 1 | 0.183 | 04:23 |
| Breath Volume | | 2.464 | 04:23 |
| Air Blank | | 0.000 | 04:24 |
| Air Blank | | 0.000 | 04:26 |
| Subject Sample | # 2 | 0.190 | 04:27 |
| Breath Volume | | 2.597 | 04:27 |
| Air Blank | | 0.000 | 04:27 |
| Control Test | | 0.079 | 04:27 |
| Air Blank | | 0.000 | 04:28 |
| Diagnostics Check | | OK | 04:28 |

Data Download Date/Time: 07/23/2020 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BANOS,FRANK
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: STEGALL

DATE/TIME OF TEST: 07/19/2020 04:25:05

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 04:18 |
| Air Blank | | 0.000 | 04:18 |
| Control Test | | 0.080 | 04:19 |
| Air Blank | | 0.000 | 04:19 |
| Subject Sample | # 1 | 0.126 | 04:20 |
| Breath Volume | | 1.871 | 04:20 |
| Air Blank | | 0.000 | 04:21 |
| Air Blank | | 0.000 | 04:22 |
| Subject Sample | # 2 | 0.121 | 04:23 |
| Breath Volume | | 1.269 | 04:23 |
| Air Blank | | 0.000 | 04:24 |
| Control Test | | 0.079 | 04:24 |
| Air Blank | | 0.000 | 04:25 |
| Diagnostics Check | | OK | 04:25 |

Data Download Date/Time: 07/23/2020 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CARBONE,ANDRES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: ACOSTA RAMOS

DATE/TIME OF TEST: 07/19/2020 05:32:31

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 05:25 |
| Air Blank | | 0.000 | 05:26 |
| Control Test | | 0.079 | 05:26 |
| Air Blank | | 0.000 | 05:27 |
| Subject Sample | # 1 | 0.154 | 05:27 |
| Breath Volume | | 1.210 | 05:27 |
| Air Blank | | 0.000 | 05:28 |
| Air Blank | | 0.000 | 05:30 |
| Subject Sample | # 2 | 0.148 | 05:31 |
| Breath Volume | | 1.394 | 05:31 |
| Air Blank | | 0.000 | 05:31 |
| Control Test | | 0.079 | 05:31 |
| Air Blank | | 0.000 | 05:32 |
| Diagnostics Check | | OK | 05:32 |

Data Download Date/Time: 07/23/2020 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BINET,MARLON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: RODRIGUEZNUNEZ

DATE/TIME OF TEST: 07/19/2020 19:53:07

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 19:46 |
| Air Blank | 0.000 | 19:47 |
| Control Test | 0.080 | 19:47 |
| Air Blank | 0.000 | 19:48 |
| Subject Sample # 1 | 0.210 | 19:48 |
| Breath Volume | 1.507 | 19:48 |
| Air Blank | 0.000 | 19:49 |
| Air Blank | 0.000 | 19:51 |
| Subject Sample # 2 | 0.222 | 19:51 |
| Breath Volume | 2.582 | 19:51 |
| Air Blank | 0.000 | 19:52 |
| Control Test | 0.079 | 19:52 |
| Air Blank | 0.000 | 19:53 |
| Diagnostics Check | OK | 19:53 |

Data Download Date/Time: 07/23/2020 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BANOS,FRANK
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: RHINE

DATE/TIME OF TEST: 07/20/2020 01:05:47

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 00:59 |
| Air Blank | | 0.000 | 00:59 |
| Control Test | | 0.079 | 00:59 |
| Air Blank | | 0.000 | 01:00 |
| Subject Sample | # 1 | 0.289 | 01:00 |
| Breath Volume | | 2.406 | 01:00 |
| Air Blank | | 0.000 | 01:01 |
| Air Blank | | 0.000 | 01:03 |
| Subject Sample | # 2 | 0.297 | 01:04 |
| Breath Volume | | 1.722 | 01:04 |
| Air Blank | | 0.000 | 01:04 |
| Control Test | | 0.079 | 01:05 |
| Air Blank | | 0.000 | 01:05 |
| Diagnostics Check | | OK | 01:05 |

Data Download Date/Time: 07/23/2020 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DENIJS,SAMIRA
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 07/20/2020 20:49:31

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:43 |
| Air Blank | 0.000 | 20:43 |
| Control Test | 0.081 | 20:44 |
| Air Blank | 0.000 | 20:44 |
| Subject Sample # 1 | 0.000 | 20:45 |
| Breath Volume | 1.636 | 20:45 |
| Air Blank | 0.000 | 20:45 |
| Air Blank | 0.000 | 20:47 |
| Subject Sample # 2 | 0.000 | 20:48 |
| Breath Volume | 1.746 | 20:48 |
| Air Blank | 0.000 | 20:48 |
| Control Test | 0.080 | 20:48 |
| Air Blank | 0.000 | 20:49 |
| Diagnostics Check | OK | 20:49 |

Data Download Date/Time: 07/23/2020 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOMINGUEZ,JAYMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: CAPP

DATE/TIME OF TEST: 07/21/2020 03:05:58

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:59 |
| Air Blank | 0.000 | 03:00 |
| Control Test | 0.081 | 03:00 |
| Air Blank | 0.000 | 03:00 |
| Subject Sample # 1 | 0.133 | 03:01 |
| Breath Volume | 3.347 | 03:01 |
| Air Blank | 0.000 | 03:02 |
| Air Blank | 0.000 | 03:04 |
| Subject Sample # 2 | 0.133 | 03:04 |
| Breath Volume | 3.937 | 03:04 |
| Air Blank | 0.000 | 03:05 |
| Control Test | 0.080 | 03:05 |
| Air Blank | 0.000 | 03:05 |
| Diagnostics Check | OK | 03:05 |

Data Download Date/Time: 07/23/2020 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: RIOSLOPEZ

DATE/TIME OF TEST: 07/21/2020 23:34:21

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:28 |
| Air Blank | 0.000 | 23:28 |
| Control Test | 0.080 | 23:29 |
| Air Blank | 0.000 | 23:29 |
| Subject Sample # 1 | 0.000 | 23:30 |
| Breath Volume | 1.566 | 23:30 |
| Air Blank | 0.000 | 23:30 |
| Air Blank | 0.000 | 23:32 |
| Subject Sample # 2 | 0.000 | 23:33 |
| Breath Volume | 1.722 | 23:33 |
| Air Blank | 0.000 | 23:33 |
| Control Test | 0.081 | 23:33 |
| Air Blank | 0.000 | 23:34 |
| Diagnostics Check | OK | 23:34 |

Data Download Date/Time: 07/23/2020 14:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: CASSON

DATE/TIME OF TEST: 07/09/2020 17:49:25

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 17:37 |
| Air Blank | 0.000 | 17:37 |
| Control Test | 0.075 | 17:38 |
| Air Blank | 0.000 | 17:38 |
| Subject Sample # 1 | 0.266 | 17:39 |
| Breath Volume | 2.300 | 17:39 |
| Air Blank | 0.000 | 17:40 |
| Air Blank | 0.000 | 17:42 |
| Subject Sample # 2 | 0.232 | 17:44 |
| Breath Volume | 1.261 | 17:44 |
| Air Blank | 0.000 | 17:45 |
| Air Blank | 0.000 | 17:47 |
| Subject Sample # 3 | 0.241 | 17:47 |
| Breath Volume | 2.472 | 17:47 |
| Air Blank | 0.000 | 17:48 |
| Control Test | 0.077 | 17:48 |
| Air Blank | 0.000 | 17:49 |
| Diagnostics Check | OK | 17:49 |

Data Download Date/Time: 08/09/2020 22:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: BRUNHOLZL

DATE/TIME OF TEST: 07/13/2020 21:28:38

| | | | |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK | 21:17 | |
| Air Blank | 0.000 | 21:17 | |
| Control Test | 0.077 | 21:18 | |
| Air Blank | 0.000 | 21:18 | |
| Subject Sample | # 1 | 21:21 | No Sample Provided |
| Breath Volume | 0.000 | 21:21 | |
| Air Blank | 0.000 | 21:22 | |
| Air Blank | 0.000 | 21:24 | |
| Subject Sample | # 2 | 21:27 | No Sample Provided |
| Breath Volume | 0.000 | 21:27 | |
| Air Blank | 0.000 | 21:27 | |
| Control Test | 0.073 | 21:28 | Interferent Detect |
| Air Blank | 0.000 | 21:28 | |

Data Download Date/Time: 08/09/2020 22:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: BRUNHOLZL

DATE/TIME OF TEST: 07/13/2020 21:46:52

| | | | |
|--------------------|-------|-------|------------|
| Diagnostics Check | OK | 21:39 | |
| Air Blank | 0.000 | 21:40 | |
| Control Test | 0.077 | 21:40 | |
| Air Blank | 0.000 | 21:41 | |
| Subject Sample # 1 | 0.172 | 21:41 | |
| Breath Volume | 2.953 | 21:41 | |
| Air Blank | 0.000 | 21:42 | |
| Air Blank | 0.000 | 21:44 | |
| Subject Sample # 2 | 0.179 | 21:45 | |
| Breath Volume | 2.609 | 21:45 | |
| Air Blank | | 21:46 | Purge Fail |
| Air Blank | 0.000 | 21:46 | |

Data Download Date/Time: 08/09/2020 22:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: BRUNHOLZL

DATE/TIME OF TEST: 07/13/2020 21:59:42

| | | | |
|--------------------|-------|-------|------------|
| Diagnostics Check | OK | 21:52 | |
| Air Blank | 0.000 | 21:53 | |
| Control Test | 0.077 | 21:53 | |
| Air Blank | 0.000 | 21:54 | |
| Subject Sample # 1 | 0.170 | 21:54 | |
| Breath Volume | 3.058 | 21:54 | |
| Air Blank | 0.000 | 21:55 | |
| Air Blank | 0.000 | 21:57 | |
| Subject Sample # 2 | 0.174 | 21:58 | |
| Breath Volume | 3.621 | 21:58 | |
| Air Blank | | 21:59 | Purge Fail |
| Air Blank | 0.000 | 21:59 | |

Data Download Date/Time: 08/09/2020 22:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: GORMAN

DATE/TIME OF TEST: 07/17/2020 21:06:42

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:00 |
| Air Blank | 0.000 | 21:00 |
| Control Test | 0.079 | 21:00 |
| Air Blank | 0.000 | 21:01 |
| Subject Sample # 1 | 0.017 | 21:02 |
| Breath Volume | 1.484 | 21:02 |
| Air Blank | 0.000 | 21:02 |
| Air Blank | 0.000 | 21:04 |
| Subject Sample # 2 | 0.016 | 21:05 |
| Breath Volume | 1.222 | 21:05 |
| Air Blank | 0.000 | 21:05 |
| Control Test | 0.077 | 21:06 |
| Air Blank | 0.000 | 21:06 |
| Diagnostics Check | OK | 21:06 |

Data Download Date/Time: 08/09/2020 22:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: CHUN SICA

DATE/TIME OF TEST: 07/19/2020 02:49:16

| | | | |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK | 02:45 | |
| Air Blank | 0.000 | 02:46 | |
| Control Test | 0.077 | 02:46 | |
| Air Blank | 0.000 | 02:46 | |
| Subject Sample | # 1 0.219 | 02:47 | |
| Breath Volume | 2.632 | 02:47 | |
| Air Blank | | 02:48 | Purge Fail |
| Air Blank | | 02:49 | Purge Fail |

Data Download Date/Time: 08/09/2020 22:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: MENGES

DATE/TIME OF TEST: 07/23/2020 20:13:35

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 20:05 |
| Air Blank | | 0.000 | 20:06 |
| Control Test | | 0.076 | 20:06 |
| Air Blank | | 0.000 | 20:06 |
| Subject Sample | # 1 | 0.299 | 20:07 |
| Breath Volume | | 2.246 | 20:07 |
| Air Blank | | 0.000 | 20:08 |
| Air Blank | | 0.000 | 20:10 |
| Subject Sample | # 2 | 0.296 | 20:11 |
| Breath Volume | | 1.886 | 20:11 |
| Air Blank | | 0.000 | 20:12 |
| Control Test | | 0.076 | 20:12 |
| Air Blank | | 0.000 | 20:13 |
| Diagnostics Check | | OK | 20:13 |

Data Download Date/Time: 08/09/2020 22:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON, MATTHEW, P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CAMP, B
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 06/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: MOSER

DATE/TIME OF TEST: 07/20/2020 21:27:26

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:20 |
| Air Blank | 0.000 | 21:20 |
| Control Test | 0.081 | 21:21 |
| Air Blank | 0.000 | 21:21 |
| Subject Sample # 1 | 0.058 | 21:22 |
| Breath Volume | 1.222 | 21:22 |
| Air Blank | 0.000 | 21:23 |
| Air Blank | 0.000 | 21:25 |
| Subject Sample # 2 | 0.058 | 21:26 |
| Breath Volume | 1.121 | 21:26 |
| Air Blank | 0.000 | 21:26 |
| Control Test | 0.079 | 21:26 |
| Air Blank | 0.000 | 21:27 |
| Diagnostics Check | OK | 21:27 |

Data Download Date/Time: 07/27/2020 01:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,MATTHEW
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 07/21/2020 00:25:27

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:19 |
| Air Blank | 0.000 | 00:19 |
| Control Test | 0.081 | 00:19 |
| Air Blank | 0.000 | 00:20 |
| Subject Sample # 1 | 0.192 | 00:20 |
| Breath Volume | 2.816 | 00:20 |
| Air Blank | 0.000 | 00:21 |
| Air Blank | 0.000 | 00:23 |
| Subject Sample # 2 | 0.200 | 00:24 |
| Breath Volume | 2.265 | 00:24 |
| Air Blank | 0.000 | 00:24 |
| Control Test | 0.079 | 00:24 |
| Air Blank | 0.000 | 00:25 |
| Diagnostics Check | OK | 00:25 |

Data Download Date/Time: 07/27/2020 01:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,J,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 07/22/2020 18:38:59

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 18:31 |
| Air Blank | 0.000 | 18:31 |
| Control Test | 0.080 | 18:31 |
| Air Blank | 0.000 | 18:32 |
| Subject Sample # 1 | 0.159 | 18:34 |
| Breath Volume | 2.992 | 18:34 |
| Air Blank | 0.000 | 18:35 |
| Air Blank | 0.000 | 18:37 |
| Subject Sample # 2 | 0.159 | 18:37 |
| Breath Volume | 2.082 | 18:37 |
| Air Blank | 0.000 | 18:38 |
| Control Test | 0.079 | 18:38 |
| Air Blank | 0.000 | 18:38 |
| Diagnostics Check | OK | 18:38 |

Data Download Date/Time: 07/27/2020 01:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAHEY, MARILYN,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAUMGARTNER, N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: MOSLEY

DATE/TIME OF TEST: 07/25/2020 18:07:01

| | | | |
|-------------------|-------|-------|-----------------|
| Diagnostics Check | OK | 18:06 | |
| Air Blank | | 18:06 | Improper Sample |
| Air Blank | 0.000 | 18:07 | |

Data Download Date/Time: 07/27/2020 01:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAHEY, MARILYN,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAUMGARTNER, N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: MOSLEY

DATE/TIME OF TEST: 07/25/2020 18:22:11

| | | | |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK | 18:09 | |
| Air Blank | 0.000 | 18:10 | |
| Control Test | 0.080 | 18:10 | |
| Air Blank | 0.000 | 18:11 | |
| Subject Sample | # 1 0.000 | 18:14 | Volume Not Met |
| Breath Volume | 0.628 | 18:14 | |
| Air Blank | 0.000 | 18:14 | |
| Air Blank | 0.000 | 18:16 | |
| Subject Sample | # 2 0.000 | 18:17 | |
| Breath Volume | 1.390 | 18:17 | |
| Air Blank | 0.000 | 18:18 | |
| Air Blank | 0.000 | 18:20 | |
| Subject Sample | # 3 0.000 | 18:20 | |
| Breath Volume | 2.125 | 18:20 | |
| Air Blank | 0.000 | 18:21 | |
| Control Test | 0.079 | 18:21 | |
| Air Blank | 0.000 | 18:22 | |
| Diagnostics Check | OK | 18:22 | |

Data Download Date/Time: 07/27/2020 01:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON, MATTHEW, P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: NEWTON, N
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 06/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: MAXWELL

DATE/TIME OF TEST: 07/25/2020 20:42:16

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:35 |
| Air Blank | 0.000 | 20:36 |
| Control Test | 0.080 | 20:36 |
| Air Blank | 0.000 | 20:36 |
| Subject Sample # 1 | 0.170 | 20:37 |
| Breath Volume | 2.144 | 20:37 |
| Air Blank | 0.000 | 20:38 |
| Air Blank | 0.000 | 20:40 |
| Subject Sample # 2 | 0.162 | 20:40 |
| Breath Volume | 1.843 | 20:40 |
| Air Blank | 0.000 | 20:41 |
| Control Test | 0.079 | 20:41 |
| Air Blank | 0.000 | 20:42 |
| Diagnostics Check | OK | 20:42 |

Data Download Date/Time: 07/27/2020 01:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAUMGARTNER,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: ARNETT

DATE/TIME OF TEST: 07/26/2020 13:33:52

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 13:27 |
| Air Blank | 0.000 | 13:27 |
| Control Test | 0.080 | 13:27 |
| Air Blank | 0.000 | 13:28 |
| Subject Sample # 1 | 0.000 | 13:29 |
| Breath Volume | 1.867 | 13:29 |
| Air Blank | 0.000 | 13:29 |
| Air Blank | 0.000 | 13:31 |
| Subject Sample # 2 | 0.000 | 13:32 |
| Breath Volume | 1.468 | 13:32 |
| Air Blank | 0.000 | 13:32 |
| Control Test | 0.081 | 13:33 |
| Air Blank | 0.000 | 13:33 |
| Diagnostics Check | OK | 13:33 |

Data Download Date/Time: 07/27/2020 01:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON, MATTHEW, P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PATRICK, K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 07/26/2020 21:53:33

| | | | |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK | 21:43 | |
| Air Blank | 0.000 | 21:43 | |
| Control Test | 0.080 | 21:44 | |
| Air Blank | 0.000 | 21:44 | |
| Subject Sample | # 1 0.089 | 21:47 | Volume Not Met |
| Breath Volume | 0.000 | 21:47 | |
| Air Blank | 0.000 | 21:48 | |
| Air Blank | 0.000 | 21:50 | |
| Subject Sample | # 2 | 21:52 | Subject Test Refused |
| Breath Volume | 0.523 | 21:52 | |
| Air Blank | 0.000 | 21:52 | |
| Control Test | 0.080 | 21:52 | |
| Air Blank | 0.000 | 21:53 | |
| Diagnostics Check | OK | 21:53 | |

Data Download Date/Time: 07/27/2020 01:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,E,W
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: HENDERSON,K
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:49

SUBJECT'S LAST NAME: WEDGE

DATE/TIME OF TEST: 07/07/2020 18:46:55

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 18:40 |
| Air Blank | 0.000 | 18:41 |
| Control Test | 0.078 | 18:41 |
| Air Blank | 0.000 | 18:41 |
| Subject Sample # 1 | 0.176 | 18:42 |
| Breath Volume | 2.074 | 18:42 |
| Air Blank | 0.000 | 18:43 |
| Air Blank | 0.000 | 18:45 |
| Subject Sample # 2 | 0.170 | 18:45 |
| Breath Volume | 1.542 | 18:45 |
| Air Blank | 0.000 | 18:45 |
| Control Test | 0.077 | 18:46 |
| Air Blank | 0.000 | 18:46 |
| Diagnostics Check | OK | 18:46 |

Data Download Date/Time: 08/04/2020 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIS,KEVIN,
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: KLEIS,KEVIN
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 07/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: BURGOS

DATE/TIME OF TEST: 07/09/2020 23:50:35

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:44 |
| Air Blank | 0.000 | 23:45 |
| Control Test | 0.078 | 23:45 |
| Air Blank | 0.000 | 23:45 |
| Subject Sample # 1 | 0.000 | 23:46 |
| Breath Volume | 1.402 | 23:46 |
| Air Blank | 0.000 | 23:46 |
| Air Blank | 0.000 | 23:48 |
| Subject Sample # 2 | 0.000 | 23:49 |
| Breath Volume | 1.851 | 23:49 |
| Air Blank | 0.000 | 23:49 |
| Control Test | 0.079 | 23:49 |
| Air Blank | 0.000 | 23:50 |
| Diagnostics Check | OK | 23:50 |

Data Download Date/Time: 08/04/2020 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 07/17/2020 01:30:34

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:24 |
| Air Blank | 0.000 | 01:24 |
| Control Test | 0.079 | 01:25 |
| Air Blank | 0.000 | 01:25 |
| Subject Sample # 1 | 0.133 | 01:26 |
| Breath Volume | 1.949 | 01:26 |
| Air Blank | 0.000 | 01:26 |
| Air Blank | 0.000 | 01:28 |
| Subject Sample # 2 | 0.129 | 01:29 |
| Breath Volume | 2.125 | 01:29 |
| Air Blank | 0.000 | 01:29 |
| Control Test | 0.078 | 01:29 |
| Air Blank | 0.000 | 01:30 |
| Diagnostics Check | OK | 01:30 |

Data Download Date/Time: 08/04/2020 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOEL,Y,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CONROY,J
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 07/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 07/18/2020 01:11:01

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:04 |
| Air Blank | 0.000 | 01:05 |
| Control Test | 0.079 | 01:05 |
| Air Blank | 0.000 | 01:06 |
| Subject Sample # 1 | 0.000 | 01:06 |
| Breath Volume | 2.593 | 01:06 |
| Air Blank | 0.000 | 01:07 |
| Air Blank | 0.000 | 01:09 |
| Subject Sample # 2 | 0.000 | 01:09 |
| Breath Volume | 1.679 | 01:09 |
| Air Blank | 0.000 | 01:10 |
| Control Test | 0.079 | 01:10 |
| Air Blank | 0.000 | 01:10 |
| Diagnostics Check | OK | 01:11 |

Data Download Date/Time: 08/04/2020 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PEREZ,M
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 07/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: LANGWEIL

DATE/TIME OF TEST: 07/18/2020 03:41:13

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:35 |
| Air Blank | 0.000 | 03:35 |
| Control Test | 0.078 | 03:35 |
| Air Blank | 0.000 | 03:36 |
| Subject Sample # 1 | 0.114 | 03:36 |
| Breath Volume | 2.269 | 03:36 |
| Air Blank | 0.000 | 03:37 |
| Air Blank | 0.000 | 03:39 |
| Subject Sample # 2 | 0.109 | 03:39 |
| Breath Volume | 2.171 | 03:39 |
| Air Blank | 0.000 | 03:40 |
| Control Test | 0.078 | 03:40 |
| Air Blank | 0.000 | 03:41 |
| Diagnostics Check | OK | 03:41 |

Data Download Date/Time: 08/04/2020 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANEDA ROMO,KARLA,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ROWLEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: STEPHEN

DATE/TIME OF TEST: 07/19/2020 03:29:55

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 03:18 | |
| Air Blank | | 0.000 | 03:18 | |
| Control Test | | 0.079 | 03:19 | |
| Air Blank | | 0.000 | 03:19 | |
| Subject Sample | # 1 | 0.201 | 03:22 | Volume Not Met |
| Breath Volume | | 1.023 | 03:22 | |
| Air Blank | | 0.000 | 03:23 | |
| Air Blank | | 0.000 | 03:25 | |
| Subject Sample | # 2 | 0.192 | 03:28 | Volume Not Met |
| Breath Volume | | 0.898 | 03:28 | |
| Air Blank | | 0.000 | 03:28 | |
| Control Test | | 0.078 | 03:29 | |
| Air Blank | | 0.000 | 03:29 | |
| Diagnostics Check | | OK | 03:29 | |

Data Download Date/Time: 08/04/2020 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANEDA ROMO,KARLA,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ROWLEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: STEPHEN

DATE/TIME OF TEST: 07/19/2020 03:43:35

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 03:31 | |
| Air Blank | | 0.000 | 03:32 | |
| Control Test | | 0.079 | 03:32 | |
| Air Blank | | 0.000 | 03:33 | |
| Subject Sample | # 1 | 0.204 | 03:36 | Volume Not Met |
| Breath Volume | | 0.039 | 03:36 | |
| Air Blank | | 0.000 | 03:37 | |
| Air Blank | | 0.000 | 03:38 | |
| Subject Sample | # 2 | 0.191 | 03:41 | Volume Not Met |
| Breath Volume | | 1.015 | 03:41 | |
| Air Blank | | 0.000 | 03:42 | |
| Control Test | | 0.077 | 03:42 | |
| Air Blank | | 0.000 | 03:43 | |
| Diagnostics Check | | OK | 03:43 | |

Data Download Date/Time: 08/04/2020 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANEDA ROMO,KARLA,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ROWLEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: STEPHEN

DATE/TIME OF TEST: 07/19/2020 04:31:17

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 04:19 | |
| Air Blank | | 0.000 | 04:20 | |
| Control Test | | 0.079 | 04:20 | |
| Air Blank | | 0.000 | 04:20 | |
| Subject Sample | # 1 | 0.188 | 04:24 | Volume Not Met |
| Breath Volume | | 0.820 | 04:24 | |
| Air Blank | | 0.000 | 04:24 | |
| Air Blank | | 0.000 | 04:26 | |
| Subject Sample | # 2 | 0.190 | 04:29 | Volume Not Met |
| Breath Volume | | 0.800 | 04:29 | |
| Air Blank | | 0.000 | 04:30 | |
| Control Test | | 0.078 | 04:30 | |
| Air Blank | | 0.000 | 04:31 | |
| Diagnostics Check | | OK | 04:31 | |

Data Download Date/Time: 08/04/2020 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,P,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: DIAZ,R
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: DENMARK

DATE/TIME OF TEST: 07/21/2020 04:07:05

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:59 |
| Air Blank | 0.000 | 04:00 |
| Control Test | 0.079 | 04:00 |
| Air Blank | 0.000 | 04:01 |
| Subject Sample # 1 | 0.193 | 04:02 |
| Breath Volume | 1.765 | 04:02 |
| Air Blank | 0.000 | 04:02 |
| Air Blank | 0.000 | 04:04 |
| Subject Sample # 2 | 0.180 | 04:05 |
| Breath Volume | 1.457 | 04:05 |
| Air Blank | 0.000 | 04:06 |
| Control Test | 0.078 | 04:06 |
| Air Blank | 0.000 | 04:07 |
| Diagnostics Check | OK | 04:07 |

Data Download Date/Time: 08/04/2020 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATSON,ANDRAE,M
BREATH TEST OPERATOR AGENCY: NAPLES PD
ARRESTING OFFICER: WALBERT,A
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 07/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: MARIANI

DATE/TIME OF TEST: 07/22/2020 21:54:10

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:47 |
| Air Blank | 0.000 | 21:48 |
| Control Test | 0.079 | 21:48 |
| Air Blank | 0.000 | 21:49 |
| Subject Sample # 1 | 0.134 | 21:49 |
| Breath Volume | 1.847 | 21:49 |
| Air Blank | 0.000 | 21:50 |
| Air Blank | 0.000 | 21:52 |
| Subject Sample # 2 | 0.132 | 21:52 |
| Breath Volume | 1.906 | 21:52 |
| Air Blank | 0.000 | 21:53 |
| Control Test | 0.079 | 21:53 |
| Air Blank | 0.000 | 21:54 |
| Diagnostics Check | OK | 21:54 |

Data Download Date/Time: 08/04/2020 16:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: TERCERO,K
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: SHALLIES

DATE/TIME OF TEST: 07/24/2020 00:28:05

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:21 |
| Air Blank | 0.000 | 00:22 |
| Control Test | 0.079 | 00:22 |
| Air Blank | 0.000 | 00:22 |
| Subject Sample # 1 | 0.091 | 00:23 |
| Breath Volume | 1.648 | 00:23 |
| Air Blank | 0.000 | 00:24 |
| Air Blank | 0.000 | 00:26 |
| Subject Sample # 2 | 0.087 | 00:26 |
| Breath Volume | 1.433 | 00:26 |
| Air Blank | 0.000 | 00:27 |
| Control Test | 0.079 | 00:27 |
| Air Blank | 0.000 | 00:28 |
| Diagnostics Check | OK | 00:28 |

Data Download Date/Time: 08/04/2020 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: THIBODEAU

DATE/TIME OF TEST: 07/24/2020 21:49:08

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:42 |
| Air Blank | 0.000 | 21:43 |
| Control Test | 0.078 | 21:43 |
| Air Blank | 0.000 | 21:44 |
| Subject Sample # 1 | 0.116 | 21:44 |
| Breath Volume | 2.117 | 21:44 |
| Air Blank | 0.000 | 21:45 |
| Air Blank | 0.000 | 21:47 |
| Subject Sample # 2 | 0.115 | 21:47 |
| Breath Volume | 2.105 | 21:47 |
| Air Blank | 0.000 | 21:48 |
| Control Test | 0.078 | 21:48 |
| Air Blank | 0.000 | 21:49 |
| Diagnostics Check | OK | 21:49 |

Data Download Date/Time: 08/04/2020 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRICE,MARCUS,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: VAUGHAN,B
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: GONZALO

DATE/TIME OF TEST: 07/26/2020 01:01:12

| | | | |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK | 00:58 | |
| Air Blank | 0.000 | 00:58 | |
| Control Test | 0.079 | 00:58 | |
| Air Blank | 0.000 | 00:59 | |
| Subject Sample | # 1 | 00:59 | RFI Detect |
| Breath Volume | 0.000 | 00:59 | |
| Air Blank | 0.000 | 01:00 | |
| Control Test | 0.079 | 01:00 | |
| Air Blank | 0.000 | 01:01 | |
| Diagnostics Check | OK | 01:01 | |

Data Download Date/Time: 08/04/2020 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRICE,MARCUS,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: VAUGHAN,B
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: GONZALO

DATE/TIME OF TEST: 07/26/2020 01:11:54

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:04 |
| Air Blank | 0.000 | 01:04 |
| Control Test | 0.080 | 01:05 |
| Air Blank | 0.000 | 01:05 |
| Subject Sample # 1 | 0.124 | 01:07 |
| Breath Volume | 1.371 | 01:07 |
| Air Blank | 0.000 | 01:07 |
| Air Blank | 0.000 | 01:09 |
| Subject Sample # 2 | 0.130 | 01:10 |
| Breath Volume | 1.816 | 01:10 |
| Air Blank | 0.000 | 01:10 |
| Control Test | 0.078 | 01:11 |
| Air Blank | 0.000 | 01:11 |
| Diagnostics Check | OK | 01:11 |

Data Download Date/Time: 08/04/2020 16:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBBONS,S,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PULIZZOTTO,S
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: SAJOJO AGUILAR

DATE/TIME OF TEST: 07/26/2020 20:45:31

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:38 |
| Air Blank | 0.000 | 20:39 |
| Control Test | 0.079 | 20:39 |
| Air Blank | 0.000 | 20:40 |
| Subject Sample # 1 | 0.175 | 20:40 |
| Breath Volume | 2.167 | 20:40 |
| Air Blank | 0.000 | 20:41 |
| Air Blank | 0.000 | 20:43 |
| Subject Sample # 2 | 0.171 | 20:43 |
| Breath Volume | 1.964 | 20:43 |
| Air Blank | 0.000 | 20:44 |
| Control Test | 0.078 | 20:44 |
| Air Blank | 0.000 | 20:45 |
| Diagnostics Check | OK | 20:45 |

Data Download Date/Time: 08/04/2020 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,E,W
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MCCOY,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: QUINLAN

DATE/TIME OF TEST: 07/27/2020 03:55:59

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:49 |
| Air Blank | 0.000 | 03:49 |
| Control Test | 0.079 | 03:49 |
| Air Blank | 0.000 | 03:50 |
| Subject Sample # 1 | 0.239 | 03:51 |
| Breath Volume | 3.183 | 03:51 |
| Air Blank | 0.000 | 03:52 |
| Air Blank | 0.000 | 03:54 |
| Subject Sample # 2 | 0.244 | 03:54 |
| Breath Volume | 2.246 | 03:54 |
| Air Blank | 0.000 | 03:55 |
| Control Test | 0.077 | 03:55 |
| Air Blank | 0.000 | 03:55 |
| Diagnostics Check | OK | 03:55 |

Data Download Date/Time: 08/04/2020 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,S,C
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: APONTE,C
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: LEBUFF

DATE/TIME OF TEST: 07/29/2020 01:28:41

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:22 |
| Air Blank | 0.000 | 01:22 |
| Control Test | 0.079 | 01:23 |
| Air Blank | 0.000 | 01:23 |
| Subject Sample # 1 | 0.000 | 01:24 |
| Breath Volume | 2.332 | 01:24 |
| Air Blank | 0.000 | 01:24 |
| Air Blank | 0.000 | 01:26 |
| Subject Sample # 2 | 0.000 | 01:27 |
| Breath Volume | 1.820 | 01:27 |
| Air Blank | 0.000 | 01:27 |
| Control Test | 0.079 | 01:28 |
| Air Blank | 0.000 | 01:28 |
| Diagnostics Check | OK | 01:28 |

Data Download Date/Time: 08/04/2020 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAGE,LYNN,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: COSTELLO,EDWARD
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 07/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:54

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 07/30/2020 10:21:19

| | | | |
|--------------------|-------|-------|---------------------------|
| Diagnostics Check | OK | 10:15 | |
| Air Blank | 0.000 | 10:15 | |
| Control Test | 0.078 | 10:15 | |
| Air Blank | 0.000 | 10:16 | |
| Subject Sample # 1 | 0.000 | 10:16 | |
| Breath Volume | 1.894 | 10:16 | |
| Air Blank | 0.000 | 10:17 | |
| Air Blank | 0.000 | 10:19 | |
| Subject Sample # 2 | 0.000 | 10:19 | |
| Breath Volume | 1.820 | 10:19 | |
| Air Blank | 0.000 | 10:20 | |
| Control Test | 0.074 | 10:20 | Control Outside Tolerance |
| Air Blank | 0.000 | 10:21 | |

Data Download Date/Time: 08/04/2020 16:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: NORTH MIAMI PD
ARRESTING OFFICER: DOUGALL,D
ARRESTING OFFICER AGENCY: NORTH MIAMI PD
LAST AGENCY INSPECTION: 06/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34418080A3
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: CELESTIN

DATE/TIME OF TEST: 07/05/2020 00:10:03

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 00:06 | |
| Air Blank | 0.000 | 00:07 | |
| Control Test | 0.082 | 00:07 | |
| Air Blank | 0.000 | 00:08 | |
| Subject Sample | # 1 | 00:08 | Subject Test Refused |
| Breath Volume | 0.000 | 00:08 | |
| Air Blank | 0.000 | 00:09 | |
| Control Test | 0.082 | 00:09 | |
| Air Blank | 0.000 | 00:09 | |
| Diagnostics Check | OK | 00:10 | |

Data Download Date/Time: 07/09/2020 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,SHAWN,N
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SWEENEY,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 08:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: ALVAREZ ABREU

DATE/TIME OF TEST: 07/11/2020 08:29:33

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 08:13 | |
| Air Blank | | 0.000 | 08:14 | |
| Control Test | | 0.080 | 08:14 | |
| Air Blank | | 0.000 | 08:14 | |
| Subject Sample | # 1 | 0.115 | 08:16 | |
| Breath Volume | | 1.125 | 08:16 | |
| Air Blank | | 0.000 | 08:17 | |
| Air Blank | | 0.000 | 08:19 | |
| Subject Sample | # 2 | 0.096 | 08:22 | Volume Not Met |
| Breath Volume | | 0.058 | 08:22 | |
| Air Blank | | 0.000 | 08:23 | |
| Air Blank | | 0.000 | 08:24 | |
| Subject Sample | # 3 | 0.097 | 08:27 | Volume Not Met |
| Breath Volume | | 0.082 | 08:27 | |
| Air Blank | | 0.000 | 08:28 | |
| Control Test | | 0.077 | 08:28 | |
| Air Blank | | 0.000 | 08:29 | |
| Diagnostics Check | | OK | 08:29 | |

Data Download Date/Time: 07/14/2020 17:27