

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,CRISTINO,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: PEREZ,CRISTINO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 06/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: BRAVO RODRIGUEZ

DATE/TIME OF TEST: 07/22/2020 20:58:30

Diagnostics Check	OK	20:54	
Air Blank	0.000	20:54	
Control Test	0.081	20:55	
Air Blank	0.000	20:55	
Subject Sample	# 1	20:57	Subject Test Refused
Breath Volume	0.000	20:57	
Air Blank	0.000	20:57	
Control Test	0.082	20:57	
Air Blank	0.000	20:58	
Diagnostics Check	OK	20:58	

Data Download Date/Time: 07/24/2020 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKER,ALEXA,C
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: SMITH,DANIEL
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 06/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:58

SUBJECT'S LAST NAME: CANALES

DATE/TIME OF TEST: 07/24/2020 07:21:27

Diagnostics Check	OK	07:18	
Air Blank	0.000	07:19	
Control Test	0.082	07:19	
Air Blank	0.000	07:19	
Subject Sample	# 1	07:20	Subject Test Refused
Breath Volume	0.000	07:20	
Air Blank	0.000	07:20	
Control Test	0.081	07:20	
Air Blank	0.000	07:21	
Diagnostics Check	OK	07:21	

Data Download Date/Time: 07/24/2020 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONEY,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ILER,ALVIN
ARRESTING OFFICER AGENCY: Calhoun County SO
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 15:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: BURRIS

DATE/TIME OF TEST: 07/04/2020 15:53:58

Diagnostics Check	OK	15:50	
Air Blank	0.000	15:50	
Control Test	0.081	15:51	
Air Blank	0.000	15:51	
Subject Sample	# 1	15:52	Subject Test Refused
Breath Volume	0.000	15:52	
Air Blank	0.000	15:53	
Control Test	0.081	15:53	
Air Blank	0.000	15:53	
Diagnostics Check	OK	15:53	

Data Download Date/Time: 08/03/2020 16:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOCARRAS,PEDRO,A
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: RIESGO,LAZARO
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 06/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 08:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 07/12/2020 10:03:14

Diagnostics Check	OK	10:00	
Air Blank	0.000	10:00	
Control Test	0.082	10:01	
Air Blank	0.000	10:01	
Subject Sample	# 1	10:01	Subject Test Refused
Breath Volume	0.000	10:01	
Air Blank	0.000	10:02	
Control Test	0.082	10:02	
Air Blank	0.000	10:03	
Diagnostics Check	OK	10:03	

Data Download Date/Time: 07/29/2020 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POMARES, MAKIN,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: POMARES, MAKIN
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 06/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 16:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:58

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 07/20/2020 18:34:51

Diagnostics Check	OK	18:31	
Air Blank	0.000	18:32	
Control Test	0.082	18:32	
Air Blank	0.000	18:32	
Subject Sample	# 1	18:33	Subject Test Refused
Breath Volume	0.000	18:33	
Air Blank	0.000	18:33	
Control Test	0.082	18:34	
Air Blank	0.000	18:34	
Diagnostics Check	OK	18:34	

Data Download Date/Time: 07/29/2020 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOCARRAS,PEDRO,A
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: DOMINGUEZ,ERIC
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 06/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 07:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:35

SUBJECT'S LAST NAME: LEGUISAMON

DATE/TIME OF TEST: 07/24/2020 08:10:54

Diagnostics Check		OK	08:04
Air Blank		0.000	08:04
Control Test		0.080	08:05
Air Blank		0.000	08:05
Subject Sample	# 1	0.238	08:06
Breath Volume		2.929	08:06
Air Blank		0.000	08:07
Air Blank		0.000	08:08
Subject Sample	# 2	0.235	08:09
Breath Volume		1.507	08:09
Air Blank		0.000	08:10
Control Test		0.080	08:10
Air Blank		0.000	08:10
Diagnostics Check		OK	08:10

Data Download Date/Time: 07/29/2020 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: UZCATEGUI FERNANDEZ

DATE/TIME OF TEST: 07/03/2020 00:56:34

Diagnostics Check	OK	00:49
Air Blank	0.000	00:50
Control Test	0.079	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.176	00:51
Breath Volume	1.890	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.177	00:55
Breath Volume	2.007	00:55
Air Blank	0.000	00:55
Control Test	0.078	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 07/29/2020 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PACHECO,G
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/11/2020 23:55:31

Diagnostics Check		OK	23:41	
Air Blank		0.000	23:42	
Control Test		0.078	23:42	
Air Blank		0.000	23:43	
Subject Sample	# 1	0.168	23:46	Volume Not Met
Breath Volume		0.652	23:46	
Air Blank		0.000	23:46	
Air Blank		0.000	23:48	
Subject Sample	# 2	0.171	23:50	
Breath Volume		1.234	23:50	
Air Blank		0.000	23:51	
Air Blank		0.000	23:53	
Subject Sample	# 3	0.178	23:53	
Breath Volume		1.593	23:53	
Air Blank		0.000	23:54	
Control Test		0.076	23:54	
Air Blank		0.000	23:55	
Diagnostics Check		OK	23:55	

Data Download Date/Time: 07/29/2020 14:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:09

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 07/25/2020 16:36:48

Diagnostics Check	OK	16:29
Air Blank	0.000	16:30
Control Test	0.079	16:30
Air Blank	0.000	16:31
Subject Sample # 1	0.118	16:32
Breath Volume	1.519	16:32
Air Blank	0.000	16:32
Air Blank	0.000	16:34
Subject Sample # 2	0.121	16:35
Breath Volume	1.457	16:35
Air Blank	0.000	16:35
Control Test	0.077	16:36
Air Blank	0.000	16:36
Diagnostics Check	OK	16:36

Data Download Date/Time: 07/29/2020 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORBITT,KYLE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DOWNING,CADEN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: BUI
ARREST TIME: 14:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:58

SUBJECT'S LAST NAME: BRENNAN

DATE/TIME OF TEST: 07/03/2020 15:26:51

Diagnostics Check	OK	15:24	
Air Blank	0.000	15:24	
Control Test	0.080	15:24	
Air Blank	0.000	15:25	
Subject Sample	# 1	15:25	Subject Test Refused
Breath Volume	0.000	15:25	
Air Blank	0.000	15:25	
Control Test	0.080	15:26	
Air Blank	0.000	15:26	
Diagnostics Check	OK	15:26	

Data Download Date/Time: 07/27/2020 15:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORBITT,KYLE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CORBITT,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: BUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:51

SUBJECT'S LAST NAME: MONCLA
DATE/TIME OF TEST: 07/03/2020 18:22:12

Diagnostics Check	OK	18:16
Air Blank	0.000	18:16
Control Test	0.080	18:16
Air Blank	0.000	18:17
Subject Sample # 1	0.052	18:17
Breath Volume	5.273	18:17
Air Blank	0.000	18:18
Air Blank	0.000	18:20
Subject Sample # 2	0.050	18:20
Breath Volume	2.792	18:20
Air Blank	0.000	18:21
Control Test	0.080	18:21
Air Blank	0.000	18:22
Diagnostics Check	OK	18:22

Data Download Date/Time: 07/27/2020 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONOVAN,CHARLES,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LECOCQ,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: BUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: DURON

DATE/TIME OF TEST: 07/03/2020 18:42:21

Diagnostics Check	OK	18:36
Air Blank	0.000	18:36
Control Test	0.080	18:36
Air Blank	0.000	18:37
Subject Sample # 1	0.039	18:37
Breath Volume	2.457	18:37
Air Blank	0.000	18:38
Air Blank	0.000	18:40
Subject Sample # 2	0.040	18:40
Breath Volume	2.238	18:40
Air Blank	0.000	18:41
Control Test	0.080	18:41
Air Blank	0.000	18:42
Diagnostics Check	OK	18:42

Data Download Date/Time: 07/27/2020 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCVANEY,JEFFREY,M
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MCVANEY,JEFFREY
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: BUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: CHUCK

DATE/TIME OF TEST: 07/03/2020 19:13:56

Diagnostics Check	OK	19:02		
Air Blank	0.000	19:02		
Control Test	0.080	19:03		
Air Blank	0.000	19:03		
Subject Sample	# 1	19:06	No Sample Provided	
Breath Volume	0.000	19:06		
Air Blank	0.000	19:07		
Air Blank	0.000	19:09		
Subject Sample	# 2	0.099	19:12	Volume Not Met
Breath Volume	0.125	19:12		
Air Blank	0.000	19:13		
Control Test	0.080	19:13		
Air Blank	0.000	19:13		
Diagnostics Check	OK	19:13		

Data Download Date/Time: 07/27/2020 15:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCVANEY,JEFFREY,B
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MCVANEY,JEFFREY
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: BUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: CHUCK

DATE/TIME OF TEST: 07/03/2020 19:26:48

Diagnostics Check	OK	19:18
Air Blank	0.000	19:19
Control Test	0.079	19:19
Air Blank	0.000	19:20
Subject Sample # 1	0.087	19:21
Breath Volume	1.359	19:21
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.085	19:25
Breath Volume	1.617	19:25
Air Blank	0.000	19:25
Control Test	0.077	19:26
Air Blank	0.000	19:26
Diagnostics Check	OK	19:26

Data Download Date/Time: 07/27/2020 15:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONOVAN,CHARLES,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DOWNING,CADEN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: BUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: KNIGHT
DATE/TIME OF TEST: 07/03/2020 19:58:59

Diagnostics Check	OK	19:58	
Air Blank		19:58	RFI Detect
Air Blank	0.000	19:58	

Data Download Date/Time: 07/27/2020 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONOVAN,CHARLES,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DOWNING,CADEN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: BUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 07/03/2020 20:07:31

Diagnostics Check	OK	20:00
Air Blank	0.000	20:01
Control Test	0.079	20:01
Air Blank	0.000	20:02
Subject Sample # 1	0.150	20:02
Breath Volume	2.453	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.143	20:05
Breath Volume	2.417	20:05
Air Blank	0.000	20:06
Control Test	0.079	20:06
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 07/27/2020 15:52

**Florida Department of Law Enforcement
Alcohol Testing Program**

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORBITT,KYLE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DONOVAN,CHARLES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: BUI
ARREST TIME: 17:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: PRIEST

DATE/TIME OF TEST: 07/04/2020 17:36:41

Diagnostics Check	OK	17:30
Air Blank	0.000	17:30
Control Test	0.079	17:31
Air Blank	0.000	17:31
Subject Sample # 1	0.087	17:32
Breath Volume	2.832	17:32
Air Blank	0.000	17:32
Air Blank	0.000	17:34
Subject Sample # 2	0.084	17:35
Breath Volume	3.539	17:35
Air Blank	0.000	17:35
Control Test	0.078	17:36
Air Blank	0.000	17:36
Diagnostics Check	OK	17:36

Data Download Date/Time: 07/27/2020 15:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCVANEY,JEFFREY,B
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MATECHIK,GREGORY
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: BUI
ARREST TIME: 17:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:13

SUBJECT'S LAST NAME: CHALHOUB

DATE/TIME OF TEST: 07/04/2020 17:57:28

Diagnostics Check	OK	17:52	
Air Blank	0.000	17:53	
Control Test	0.079	17:53	
Air Blank	0.000	17:53	
Subject Sample	# 1	17:56	Subject Test Refused
Breath Volume	0.000	17:56	
Air Blank	0.000	17:56	
Control Test	0.079	17:56	
Air Blank	0.000	17:57	
Diagnostics Check	OK	17:57	

Data Download Date/Time: 07/27/2020 15:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORBITT,KYLE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CORBITT,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: BUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: AKIN

DATE/TIME OF TEST: 07/04/2020 19:46:55

Diagnostics Check	OK	19:44	
Air Blank	0.000	19:44	
Control Test	0.079	19:44	
Air Blank	0.000	19:45	
Subject Sample	# 1	19:45	Subject Test Refused
Breath Volume	0.000	19:45	
Air Blank	0.000	19:46	
Control Test	0.079	19:46	
Air Blank	0.000	19:46	
Diagnostics Check	OK	19:46	

Data Download Date/Time: 07/27/2020 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAHR,DANIEL,E
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: HAHR,DANIEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: BUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: GLORIOSO

DATE/TIME OF TEST: 07/04/2020 22:57:42

Diagnostics Check	OK	22:48
Air Blank	0.000	22:48
Control Test	0.080	22:49
Air Blank	0.000	22:49
Subject Sample # 1	0.167	22:52
Breath Volume	2.449	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:54
Subject Sample # 2	0.174	22:55
Breath Volume	1.359	22:55
Air Blank	0.000	22:56
Control Test	0.077	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 07/27/2020 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCVANEY,JEFFREY,B
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MCVANEY,JEFFREY
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: BUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: COBLE

DATE/TIME OF TEST: 07/04/2020 23:22:54

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.078	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.111	23:18
Breath Volume	3.914	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:20
Subject Sample # 2	0.116	23:21
Breath Volume	2.531	23:21
Air Blank	0.000	23:22
Control Test	0.079	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 07/27/2020 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORBITT,KYLE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LECOCQ,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: BUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: HANNA

DATE/TIME OF TEST: 07/05/2020 19:16:50

Diagnostics Check	OK	19:10
Air Blank	0.000	19:11
Control Test	0.080	19:11
Air Blank	0.000	19:12
Subject Sample # 1	0.092	19:12
Breath Volume	3.425	19:12
Air Blank	0.000	19:13
Air Blank	0.000	19:14
Subject Sample # 2	0.093	19:15
Breath Volume	2.589	19:15
Air Blank	0.000	19:15
Control Test	0.080	19:16
Air Blank	0.000	19:16
Diagnostics Check	OK	19:16

Data Download Date/Time: 07/27/2020 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1060911
BREATH TEST OPERATOR: DONOVAN,CHARLES4,L	DRY GAS EXPIRY DATE: 12/11/2021
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: BUI
ARRESTING OFFICER: DONOVAN,CHARLES	ARREST TIME: 18:04
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 06/22/2020	OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: HERRERA DATE/TIME OF TEST: 07/11/2020 18:35:57

Diagnostics Check	OK	18:29
Air Blank	0.000	18:29
Control Test	0.080	18:30
Air Blank	0.000	18:30
Subject Sample # 1	0.130	18:31
Breath Volume	3.738	18:31
Air Blank	0.000	18:31
Air Blank	0.000	18:33
Subject Sample # 2	0.136	18:34
Breath Volume	3.750	18:34
Air Blank	0.000	18:35
Control Test	0.080	18:35
Air Blank	0.000	18:35
Diagnostics Check	OK	18:35

Data Download Date/Time: 07/27/2020 15:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1060911
BREATH TEST OPERATOR: DONOVAN,CHARLES,L	DRY GAS EXPIRY DATE: 12/11/2021
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: BUI
ARRESTING OFFICER: DONOVAN,CHARLES	ARREST TIME: 16:47
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 06/22/2020	OBSERVATION PERIOD BEGAN: 16:47

SUBJECT'S LAST NAME: MCCORMACK DATE/TIME OF TEST: 07/16/2020 17:26:45

Diagnostics Check	OK	17:23	
Air Blank	0.000	17:24	
Control Test	0.079	17:24	
Air Blank	0.000	17:25	
Subject Sample	# 1	17:25	Subject Test Refused
Breath Volume	0.000	17:25	
Air Blank	0.000	17:25	
Control Test	0.080	17:26	
Air Blank	0.000	17:26	
Diagnostics Check	OK	17:26	

Data Download Date/Time: 07/27/2020 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAHR,DANIEL,E
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: CORBIN,BRITTON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: BUI
ARREST TIME: 16:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: WALLS

DATE/TIME OF TEST: 07/17/2020 17:40:22

Diagnostics Check		OK	17:32
Air Blank		0.000	17:32
Control Test		0.080	17:32
Air Blank		0.000	17:33
Subject Sample	# 1	0.060	17:35
Breath Volume		2.687	17:35
Air Blank		0.000	17:36
Air Blank		0.000	17:38
Subject Sample	# 2	0.058	17:38
Breath Volume		2.593	17:38
Air Blank		0.000	17:39
Control Test		0.078	17:39
Air Blank		0.000	17:40
Diagnostics Check		OK	17:40

Data Download Date/Time: 07/27/2020 15:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LECOCQ,MATTHEW,E
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LECOCQ,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: BUI
ARREST TIME: 17:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: DARES

DATE/TIME OF TEST: 07/17/2020 17:54:27

Diagnostics Check	OK	17:51	
Air Blank	0.000	17:51	
Control Test	0.079	17:51	
Air Blank	0.000	17:52	
Subject Sample	# 1	17:53	Subject Test Refused
Breath Volume	0.000	17:53	
Air Blank	0.000	17:53	
Control Test	0.079	17:53	
Air Blank	0.000	17:54	
Diagnostics Check	OK	17:54	

Data Download Date/Time: 07/27/2020 15:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORBITT,KYLE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CORBITT,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: BUI
ARREST TIME: 16:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 07/18/2020 17:06:18

Diagnostics Check	OK	17:03	
Air Blank	0.000	17:03	
Control Test	0.080	17:04	
Air Blank	0.000	17:04	
Subject Sample	# 1	17:04	Subject Test Refused
Breath Volume	0.000	17:04	
Air Blank	0.000	17:05	
Control Test	0.079	17:05	
Air Blank	0.000	17:06	
Diagnostics Check	OK	17:06	

Data Download Date/Time: 07/27/2020 15:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LECOCQ,MATTHEW,E
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LECOCQ,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: BUI
ARREST TIME: 17:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:49

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 07/18/2020 17:26:21

Diagnostics Check	OK	17:21	
Air Blank	0.000	17:22	
Control Test	0.079	17:22	
Air Blank	0.000	17:22	
Subject Sample	# 1	17:24	Subject Test Refused
Breath Volume	0.000	17:24	
Air Blank	0.000	17:25	
Control Test	0.079	17:25	
Air Blank	0.000	17:26	
Diagnostics Check	OK	17:26	

Data Download Date/Time: 07/27/2020 15:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORBIN,BRITTON,L
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: CORBIN,BRITTON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: BUI
ARREST TIME: 18:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: LUCERO

DATE/TIME OF TEST: 07/19/2020 18:33:51

Diagnostics Check	OK	18:25	
Air Blank	0.000	18:25	
Control Test	0.080	18:26	
Air Blank	0.000	18:26	
Subject Sample	# 1	18:29	No Sample Provided
Breath Volume	0.000	18:29	
Air Blank	0.000	18:30	
Air Blank	0.000	18:32	
Subject Sample	# 2	18:32	Subject Test Refused
Breath Volume	0.000	18:32	
Air Blank	0.000	18:32	
Control Test	0.080	18:33	
Air Blank	0.000	18:33	
Diagnostics Check	OK	18:33	

Data Download Date/Time: 07/27/2020 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORBIN,BRITTON,L
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: HAHR,DANIEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: BUI
ARREST TIME: 16:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: WEIGART

DATE/TIME OF TEST: 07/24/2020 16:39:59

Diagnostics Check	OK	16:36	
Air Blank	0.000	16:37	
Control Test	0.080	16:37	
Air Blank	0.000	16:38	
Subject Sample	# 1	16:38	Subject Test Refused
Breath Volume	0.000	16:38	
Air Blank	0.000	16:39	
Control Test	0.080	16:39	
Air Blank	0.000	16:39	
Diagnostics Check	OK	16:39	

Data Download Date/Time: 07/27/2020 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALSOBROOKS,KEELIN,G
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: POSAR,ADAM
ARRESTING OFFICER AGENCY: Naval Lab Pana City
LAST AGENCY INSPECTION: 06/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/21/2021
VIOLATION CODE: DUI
ARREST TIME: 07:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:55

SUBJECT'S LAST NAME: MATHIS

DATE/TIME OF TEST: 07/04/2020 08:28:15

Diagnostics Check	OK	08:21
Air Blank	0.000	08:21
Control Test	0.080	08:22
Air Blank	0.000	08:22
Subject Sample # 1	0.222	08:23
Breath Volume	3.027	08:23
Air Blank	0.000	08:24
Air Blank	0.000	08:25
Subject Sample # 2	0.207	08:26
Breath Volume	2.281	08:26
Air Blank	0.000	08:27
Control Test	0.080	08:27
Air Blank	0.000	08:28
Diagnostics Check	OK	08:28

Data Download Date/Time: 07/27/2020 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROTHCHILD,FOREST,W
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: ROTHCHILD,FOREST
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A1
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: STEVENSON

DATE/TIME OF TEST: 07/02/2020 23:17:00

Diagnostics Check	OK	23:11
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.000	23:12
Breath Volume	3.988	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.000	23:15
Breath Volume	2.191	23:15
Air Blank	0.000	23:16
Control Test	0.080	23:16
Air Blank	0.000	23:16
Diagnostics Check	OK	23:17

Data Download Date/Time: 07/13/2020 21:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULASKI,DAMON,J
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PULASKI,DAMON
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A2
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: BUI
ARREST TIME: 18:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: KINNEY

DATE/TIME OF TEST: 07/26/2020 19:50:03

Diagnostics Check	OK	19:43
Air Blank	0.000	19:43
Control Test	0.078	19:43
Air Blank	0.000	19:44
Subject Sample # 1	0.211	19:45
Breath Volume	1.281	19:45
Air Blank	0.000	19:45
Air Blank	0.000	19:47
Subject Sample # 2	0.212	19:48
Breath Volume	1.972	19:48
Air Blank	0.000	19:49
Control Test	0.078	19:49
Air Blank	0.000	19:49
Diagnostics Check	OK	19:50

Data Download Date/Time: 09/17/2020 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,MALACHI,C
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: ALVIS,REECE
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: BUI
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/04/2020 17:50:06

Diagnostics Check	OK	17:44
Air Blank	0.000	17:44
Control Test	0.079	17:44
Air Blank	0.000	17:45
Subject Sample # 1	0.000	17:45
Breath Volume	2.210	17:45
Air Blank	0.000	17:46
Air Blank	0.000	17:48
Subject Sample # 2	0.000	17:48
Breath Volume	2.878	17:48
Air Blank	0.000	17:49
Control Test	0.079	17:49
Air Blank	0.000	17:50
Diagnostics Check	OK	17:50

Data Download Date/Time: 07/13/2020 20:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,MALACHI,C
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PASS,JUSTIN
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: BUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: VILLEDADA

DATE/TIME OF TEST: 07/04/2020 21:46:43

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:35	
Control Test	0.080	21:36	
Air Blank	0.000	21:36	
Subject Sample	# 1	21:39	No Sample Provided
Breath Volume	0.000	21:39	
Air Blank	0.000	21:40	
Air Blank	0.000	21:42	
Subject Sample	# 2	21:45	No Sample Provided
Breath Volume	0.000	21:45	
Air Blank	0.000	21:45	
Control Test	0.080	21:46	
Air Blank	0.000	21:46	
Diagnostics Check	OK	21:46	

Data Download Date/Time: 07/13/2020 20:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: ALVAREZ,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: BUI
ARREST TIME: 16:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:09

SUBJECT'S LAST NAME: ALONSO

DATE/TIME OF TEST: 07/05/2020 17:17:43

Diagnostics Check	OK	17:09	
Air Blank	0.000	17:09	
Control Test	0.082	17:10	
Air Blank	0.000	17:10	
Subject Sample	# 1	17:13	No Sample Provided
Breath Volume	0.000	17:13	
Air Blank	0.000	17:14	
Air Blank	0.000	17:16	
Subject Sample	# 2	17:16	Subject Test Refused
Breath Volume	0.000	17:16	
Air Blank	0.000	17:16	
Control Test	0.082	17:17	
Air Blank	0.000	17:17	
Diagnostics Check	OK	17:17	

Data Download Date/Time: 07/27/2020 19:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAFOYA,THOMAS,R
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: TAFOYA,THOMAS
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: BUI
ARREST TIME: 17:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 07/25/2020 18:43:27

Diagnostics Check	OK	18:40	
Air Blank	0.000	18:41	
Control Test	0.081	18:41	
Air Blank	0.000	18:41	
Subject Sample	# 1	18:42	Subject Test Refused
Breath Volume	0.000	18:42	
Air Blank	0.000	18:42	
Control Test	0.080	18:42	
Air Blank	0.000	18:43	
Diagnostics Check	OK	18:43	

Data Download Date/Time: 07/27/2020 19:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,WILLIAM,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: THOMPSON,WILLIAM
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: BUI
ARREST TIME: 13:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: MULLA

DATE/TIME OF TEST: 07/29/2020 14:25:06

Diagnostics Check	OK	14:22	
Air Blank	0.000	14:22	
Control Test	0.081	14:23	
Air Blank	0.000	14:23	
Subject Sample	# 1	14:23	Subject Test Refused
Breath Volume	0.000	14:23	
Air Blank	0.000	14:24	
Control Test	0.081	14:24	
Air Blank	0.000	14:25	
Diagnostics Check	OK	14:25	

Data Download Date/Time: 09/04/2020 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIS,KEVIN,W
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: KLEIS,KEVIN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: BUI
ARREST TIME: 17:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: CURLS

DATE/TIME OF TEST: 07/30/2020 18:13:47

Diagnostics Check	OK	18:10	
Air Blank	0.000	18:11	
Control Test	0.080	18:11	
Air Blank	0.000	18:12	
Subject Sample	# 1	18:12	Subject Test Refused
Breath Volume	0.000	18:12	
Air Blank	0.000	18:12	
Control Test	0.080	18:13	
Air Blank	0.000	18:13	
Diagnostics Check	OK	18:13	

Data Download Date/Time: 09/04/2020 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000906

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUSA,KYLE,A
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: PLUSA,KYLE
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 07/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 121704
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: BUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: HOLZKNECHT

DATE/TIME OF TEST: 07/18/2020 19:27:57

Diagnostics Check	OK	19:20
Air Blank	0.000	19:20
Control Test	0.079	19:21
Air Blank	0.000	19:21
Subject Sample # 1	0.181	19:23
Breath Volume	3.722	19:23
Air Blank	0.000	19:24
Air Blank	0.000	19:26
Subject Sample # 2	0.184	19:26
Breath Volume	2.773	19:26
Air Blank	0.000	19:27
Control Test	0.077	19:27
Air Blank	0.000	19:27
Diagnostics Check	OK	19:27

Data Download Date/Time: 08/03/2020 17:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DZUGAN 4694,M
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/10/2020 01:42:12

Diagnostics Check	OK	01:41	
Air Blank		01:42	RFI Detect
Air Blank		01:42	RFI Detect

Data Download Date/Time: 08/12/2020 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,KATIE,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GONZALES,DAVID
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 07/11/2020 21:09:32

Diagnostics Check	OK	21:02
Air Blank	0.000	21:03
Control Test	0.079	21:03
Air Blank	0.000	21:04
Subject Sample # 1	0.000	21:04
Breath Volume	2.550	21:04
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.000	21:08
Breath Volume	2.335	21:08
Air Blank	0.000	21:08
Control Test	0.079	21:08
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

Data Download Date/Time: 08/12/2020 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SUZUKI

DATE/TIME OF TEST: 07/12/2020 04:07:23

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.000	04:03
Breath Volume	1.878	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.000	04:06
Breath Volume	2.136	04:06
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 08/12/2020 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAWLINS,DUSTIN,P
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CAFFRONI,IVAN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 07/14/2020 05:17:10

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.079	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.153	05:12
Breath Volume	1.656	05:12
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.162	05:15
Breath Volume	1.406	05:15
Air Blank	0.000	05:16
Control Test	0.077	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 08/12/2020 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,DAVID,J
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: CAMERON,DAVID
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: VALEZ II

DATE/TIME OF TEST: 07/15/2020 01:15:16

Diagnostics Check		OK	01:09
Air Blank		0.000	01:09
Control Test		0.078	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.186	01:10
Breath Volume		2.570	01:10
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.185	01:13
Breath Volume		2.699	01:13
Air Blank		0.000	01:14
Control Test		0.076	01:14
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 08/12/2020 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEINER,RYAN,T
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: BEINER,RYAN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: PETRILLI

DATE/TIME OF TEST: 07/16/2020 03:35:05

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.078	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.177	03:30
Breath Volume	1.800	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:32
Subject Sample # 2	0.167	03:33
Breath Volume	1.707	03:33
Air Blank	0.000	03:34
Control Test	0.078	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 08/12/2020 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,KATIE,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: RIDGEWAY,CAITLIN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: TIQUIRAM TOJIN

DATE/TIME OF TEST: 07/16/2020 20:25:23

Diagnostics Check	OK	20:18
Air Blank	0.000	20:18
Control Test	0.078	20:18
Air Blank	0.000	20:19
Subject Sample # 1	0.163	20:20
Breath Volume	2.117	20:20
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample # 2	0.161	20:23
Breath Volume	2.167	20:23
Air Blank	0.000	20:24
Control Test	0.077	20:24
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 08/12/2020 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOTO,MICHAEL,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SOTO,MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:25

SUBJECT'S LAST NAME: SCIMIO

DATE/TIME OF TEST: 07/18/2020 08:58:02

Diagnostics Check	OK	08:51
Air Blank	0.000	08:51
Control Test	0.079	08:52
Air Blank	0.000	08:52
Subject Sample # 1	0.093	08:53
Breath Volume	2.179	08:53
Air Blank	0.000	08:54
Air Blank	0.000	08:55
Subject Sample # 2	0.095	08:56
Breath Volume	1.988	08:56
Air Blank	0.000	08:57
Control Test	0.078	08:57
Air Blank	0.000	08:57
Diagnostics Check	OK	08:58

Data Download Date/Time: 08/12/2020 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINTER,CHRISTIAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WINTER,CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:57

SUBJECT'S LAST NAME: FRANCISCO MIGUEL

DATE/TIME OF TEST: 07/18/2020 14:36:29

Diagnostics Check	OK	14:30
Air Blank	0.000	14:30
Control Test	0.079	14:31
Air Blank	0.000	14:31
Subject Sample # 1	0.027	14:32
Breath Volume	1.917	14:32
Air Blank	0.000	14:32
Air Blank	0.000	14:34
Subject Sample # 2	0.025	14:35
Breath Volume	2.003	14:35
Air Blank	0.000	14:35
Control Test	0.077	14:35
Air Blank	0.000	14:36
Diagnostics Check	OK	14:36

Data Download Date/Time: 08/12/2020 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,LISA,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN,R
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 07/18/2020 16:25:12

Diagnostics Check	OK	16:17
Air Blank	0.000	16:18
Control Test	0.078	16:18
Air Blank	0.000	16:19
Subject Sample # 1	0.236	16:20
Breath Volume	2.125	16:20
Air Blank	0.000	16:21
Air Blank	0.000	16:23
Subject Sample # 2	0.238	16:23
Breath Volume	1.429	16:23
Air Blank	0.000	16:24
Control Test	0.077	16:24
Air Blank	0.000	16:25
Diagnostics Check	OK	16:25

Data Download Date/Time: 08/12/2020 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN 07102,R
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 07/18/2020 19:58:01

Diagnostics Check	OK	19:51
Air Blank	0.000	19:52
Control Test	0.079	19:52
Air Blank	0.000	19:52
Subject Sample # 1	0.181	19:53
Breath Volume	1.804	19:53
Air Blank	0.000	19:54
Air Blank	0.000	19:55
Subject Sample # 2	0.185	19:56
Breath Volume	2.300	19:56
Air Blank	0.000	19:57
Control Test	0.078	19:57
Air Blank	0.000	19:57
Diagnostics Check	OK	19:58

Data Download Date/Time: 08/12/2020 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: PAHLKE 17045,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: RAMOS ROSALES

DATE/TIME OF TEST: 07/18/2020 22:24:36

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.079	22:19
Air Blank	0.000	22:19
Subject Sample # 1	0.161	22:20
Breath Volume	3.066	22:20
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.162	22:23
Breath Volume	2.703	22:23
Air Blank	0.000	22:23
Control Test	0.079	22:24
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 08/12/2020 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DZUGAN 4694,M
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: FRANCISCO PEDRO

DATE/TIME OF TEST: 07/19/2020 04:38:32

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.079	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.198	04:33
Breath Volume	2.937	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.200	04:36
Breath Volume	2.589	04:36
Air Blank	0.000	04:37
Control Test	0.079	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 08/12/2020 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINTER,CHRISTIAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WINTER,CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:25

SUBJECT'S LAST NAME: PASCUAL VICENTE

DATE/TIME OF TEST: 07/19/2020 09:59:32

Diagnostics Check		OK	09:53
Air Blank		0.000	09:53
Control Test		0.079	09:53
Air Blank		0.000	09:54
Subject Sample	# 1	0.111	09:54
Breath Volume		1.667	09:54
Air Blank		0.000	09:55
Air Blank		0.000	09:57
Subject Sample	# 2	0.106	09:57
Breath Volume		1.582	09:57
Air Blank		0.000	09:58
Control Test		0.078	09:58
Air Blank		0.000	09:59
Diagnostics Check		OK	09:59

Data Download Date/Time: 08/12/2020 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,KATIE,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: VALCIN,JEFFREY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:29

SUBJECT'S LAST NAME: KRAGH

DATE/TIME OF TEST: 07/25/2020 19:49:40

Diagnostics Check	OK	19:43
Air Blank	0.000	19:43
Control Test	0.079	19:44
Air Blank	0.000	19:44
Subject Sample # 1	0.153	19:45
Breath Volume	3.511	19:45
Air Blank	0.000	19:45
Air Blank	0.000	19:47
Subject Sample # 2	0.150	19:48
Breath Volume	2.515	19:48
Air Blank	0.000	19:48
Control Test	0.079	19:49
Air Blank	0.000	19:49
Diagnostics Check	OK	19:49

Data Download Date/Time: 08/12/2020 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEMEN,LOGAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEMEN,LOGAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: FRANSISCO-JUAN

DATE/TIME OF TEST: 07/25/2020 21:35:26

Diagnostics Check	OK	21:28
Air Blank	0.000	21:29
Control Test	0.078	21:29
Air Blank	0.000	21:30
Subject Sample # 1	0.298	21:30
Breath Volume	3.378	21:30
Air Blank	0.000	21:31
Air Blank	0.000	21:33
Subject Sample # 2	0.279	21:33
Breath Volume	2.808	21:33
Air Blank	0.000	21:34
Control Test	0.078	21:34
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

Data Download Date/Time: 08/12/2020 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINTON, ADAM, H
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINTON, ADAM
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: RISNER

DATE/TIME OF TEST: 07/27/2020 01:05:07

Diagnostics Check	OK	00:58
Air Blank	0.000	00:59
Control Test	0.079	00:59
Air Blank	0.000	01:00
Subject Sample # 1	0.054	01:00
Breath Volume	2.343	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.056	01:03
Breath Volume	2.664	01:03
Air Blank	0.000	01:04
Control Test	0.079	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 08/12/2020 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/02/2020 19:50:03

Diagnostics Check		OK	19:44
Air Blank		0.000	19:44
Control Test		0.080	19:44
Air Blank		0.000	19:45
Subject Sample	# 1	0.286	19:45
Breath Volume		2.046	19:45
Air Blank		0.000	19:46
Air Blank		0.000	19:48
Subject Sample	# 2	0.270	19:48
Breath Volume		2.207	19:48
Air Blank		0.000	19:49
Control Test		0.079	19:49
Air Blank		0.000	19:49
Diagnostics Check		OK	19:50

Data Download Date/Time: 07/14/2020 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HOYOS 19022,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: CORLEY

DATE/TIME OF TEST: 07/04/2020 02:55:19

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.196	02:50
Breath Volume	1.351	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.209	02:53
Breath Volume	2.191	02:53
Air Blank	0.000	02:54
Control Test	0.081	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 07/14/2020 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ECKERT 1573,K
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 07/04/2020 19:46:24

Diagnostics Check		OK	19:36	
Air Blank		0.000	19:36	
Control Test		0.080	19:37	
Air Blank		0.000	19:37	
Subject Sample	# 1	0.185	19:40	Volume Not Met
Breath Volume		0.253	19:40	
Air Blank		0.000	19:41	
Air Blank		0.000	19:43	
Subject Sample	# 2		19:44	Subject Test Refused
Breath Volume		0.222	19:44	
Air Blank		0.000	19:45	
Control Test		0.078	19:45	
Air Blank		0.000	19:46	
Diagnostics Check		OK	19:46	

Data Download Date/Time: 07/14/2020 09:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRAGO,JAMES,A
BREATH TEST OPERATOR AGENCY: Sanibel PD
ARRESTING OFFICER: SERRAGO,JAMES
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: NARGIS

DATE/TIME OF TEST: 07/05/2020 01:13:15

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:01	
Control Test	0.080	01:01	
Air Blank	0.000	01:01	
Subject Sample # 1	0.204	01:02	
Breath Volume	1.578	01:02	
Air Blank	0.000	01:03	
Air Blank	0.000	01:05	
Subject Sample # 2	0.162	01:08	Volume Not Met
Breath Volume	0.589	01:08	
Air Blank	0.000	01:09	
Air Blank	0.000	01:10	
Subject Sample # 3		01:11	Subject Test Refused
Breath Volume	0.867	01:11	
Air Blank	0.000	01:12	
Control Test	0.077	01:12	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 07/14/2020 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEMEN,LOGAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEMEN,LOGAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/07/2020 01:48:06

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.126	01:43
Breath Volume	4.546	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.125	01:46
Breath Volume	3.429	01:46
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 07/14/2020 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EPPS,QUINTELLA,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CHARLES,CLEEFORD
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 07/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: DESAI

DATE/TIME OF TEST: 07/25/2020 02:44:48

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.076	02:39
Air Blank		0.000	02:39
Subject Sample	# 1	0.250	02:40
Breath Volume		1.210	02:40
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.260	02:43
Breath Volume		1.355	02:43
Air Blank		0.000	02:43
Control Test		0.076	02:44
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 08/27/2020 05:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER, DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER, DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 07/20/2020 23:33:15

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.078	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.257	23:28
Breath Volume	2.042	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.244	23:31
Breath Volume	1.714	23:31
Air Blank	0.000	23:32
Control Test	0.079	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 07/31/2020 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEWITT,JARED,S
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: SHIVER,ROBERT
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 06/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: NIU

DATE/TIME OF TEST: 07/20/2020 00:15:35

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:12	
Control Test	0.079	00:12	
Air Blank	0.000	00:13	
Subject Sample # 1	0.226	00:14	
Breath Volume	2.160	00:14	
Air Blank		00:15	Purge Fail
Air Blank	0.000	00:15	

Data Download Date/Time: 07/29/2020 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEWITT,JARED,S
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: SHIVER,ROBERT
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 06/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: NIU

DATE/TIME OF TEST: 07/20/2020 00:19:43

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:18	
Control Test	0.078	00:19	
Air Blank		00:19	RFI Detect
Air Blank	0.000	00:19	

Data Download Date/Time: 07/29/2020 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEWITT,JARED,S
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: SHIVER,ROBERT
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 06/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: NIU

DATE/TIME OF TEST: 07/20/2020 00:28:46

Diagnostics Check		OK	00:22
Air Blank		0.000	00:22
Control Test		0.078	00:22
Air Blank		0.000	00:23
Subject Sample	# 1	0.232	00:24
Breath Volume		2.667	00:24
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.216	00:27
Breath Volume		2.113	00:27
Air Blank		0.000	00:27
Control Test		0.079	00:28
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 07/29/2020 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TALLMAN,JOHN,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BETHEA,COLBY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: STEPHAN

DATE/TIME OF TEST: 07/15/2020 20:20:06

Diagnostics Check	OK	20:13
Air Blank	0.000	20:13
Control Test	0.079	20:14
Air Blank	0.000	20:14
Subject Sample # 1	0.225	20:15
Breath Volume	2.316	20:15
Air Blank	0.000	20:16
Air Blank	0.000	20:17
Subject Sample # 2	0.231	20:18
Breath Volume	1.953	20:18
Air Blank	0.000	20:19
Control Test	0.078	20:19
Air Blank	0.000	20:20
Diagnostics Check	OK	20:20

Data Download Date/Time: 08/18/2020 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: BROADBENT

DATE/TIME OF TEST: 07/02/2020 00:53:31

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.082	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.151	00:49
Breath Volume	3.000	00:49
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.147	00:52
Breath Volume	2.050	00:52
Air Blank	0.000	00:52
Control Test	0.082	00:52
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 07/06/2020 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUCHOCKI, RAYMOND, J
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SUCHOCKI, RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: AMIN

DATE/TIME OF TEST: 07/02/2020 23:15:26

Diagnostics Check	OK	23:09
Air Blank	0.000	23:09
Control Test	0.082	23:10
Air Blank	0.000	23:10
Subject Sample # 1	0.194	23:10
Breath Volume	2.265	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.179	23:13
Breath Volume	1.773	23:13
Air Blank	0.000	23:14
Control Test	0.082	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 07/06/2020 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,JAMES,D
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WEST,JAMES
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: ANASTACIO

DATE/TIME OF TEST: 07/04/2020 02:11:50

Diagnostics Check	OK	02:11	
Air Blank		02:11	Ambient Fail
Air Blank		02:11	Improper Sample

Data Download Date/Time: 07/06/2020 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,JAMES,D
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WEST,JAMES
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: ANASTACIO

DATE/TIME OF TEST: 07/04/2020 02:14:43

Diagnostics Check	OK	02:14	
Air Blank		02:14	Ambient Fail
Air Blank		02:14	Improper Sample

Data Download Date/Time: 07/06/2020 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,JAMES,D
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WEST,JAMES
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: ANASTACIO

DATE/TIME OF TEST: 07/04/2020 02:18:49

Diagnostics Check	OK	02:18	
Air Blank		02:18	Improper Sample
Air Blank		02:18	Improper Sample

Data Download Date/Time: 07/06/2020 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,JAMES,D
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WESTMAN,PATRICK
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 06/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: ANASTACIO

DATE/TIME OF TEST: 07/04/2020 02:22:22

Diagnostics Check	OK	02:21	
Air Blank		02:21	Ambient Fail
Air Blank	0.000	02:22	

Data Download Date/Time: 07/06/2020 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BARROW, JESHUA
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 07/11/2020 21:26:57

Diagnostics Check		OK	21:18	
Air Blank		0.000	21:19	
Control Test		0.081	21:19	
Air Blank		0.000	21:19	
Subject Sample	# 1	0.197	21:22	Volume Not Met
Breath Volume		0.019	21:22	
Air Blank		0.000	21:23	
Air Blank		0.000	21:25	
Subject Sample	# 2		21:25	Subject Test Refused
Breath Volume		0.000	21:25	
Air Blank		0.000	21:26	
Control Test		0.079	21:26	
Air Blank		0.000	21:26	
Diagnostics Check		OK	21:26	

Data Download Date/Time: 08/06/2020 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/12/2020 21:02:42

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.082	20:57
Air Blank	0.000	20:57
Subject Sample # 1	0.000	20:58
Breath Volume	3.019	20:58
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.000	21:01
Breath Volume	3.070	21:01
Air Blank	0.000	21:01
Control Test	0.082	21:02
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 08/06/2020 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: GEFFRARD

DATE/TIME OF TEST: 07/14/2020 20:16:27

Diagnostics Check	OK	20:10
Air Blank	0.000	20:10
Control Test	0.081	20:10
Air Blank	0.000	20:11
Subject Sample # 1	0.000	20:12
Breath Volume	3.917	20:12
Air Blank	0.000	20:12
Air Blank	0.000	20:14
Subject Sample # 2	0.000	20:15
Breath Volume	3.613	20:15
Air Blank	0.000	20:15
Control Test	0.081	20:15
Air Blank	0.000	20:16
Diagnostics Check	OK	20:16

Data Download Date/Time: 08/06/2020 08:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD,LEE,M
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: HOLT,TIERRA
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 07/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: CARNAHAN

DATE/TIME OF TEST: 07/30/2020 14:30:24

Diagnostics Check	OK	14:23
Air Blank	0.000	14:24
Control Test	0.080	14:24
Air Blank	0.000	14:24
Subject Sample # 1	0.290	14:25
Breath Volume	1.160	14:25
Air Blank	0.000	14:26
Air Blank	0.000	14:27
Subject Sample # 2	0.290	14:28
Breath Volume	2.082	14:28
Air Blank	0.000	14:29
Control Test	0.082	14:29
Air Blank	0.000	14:30
Diagnostics Check	OK	14:30

Data Download Date/Time: 08/06/2020 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARR,JARED,R
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MIDDLEBROOKS,ALAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: BRANCH

DATE/TIME OF TEST: 07/02/2020 23:42:45

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.080	23:37
Air Blank	0.000	23:37
Subject Sample # 1	0.180	23:38
Breath Volume	2.019	23:38
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.183	23:41
Breath Volume	2.175	23:41
Air Blank	0.000	23:41
Control Test	0.080	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 07/30/2020 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,GEORGE,H
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: EDWARDS,GEORGE
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 07/04/2020 19:09:20

Diagnostics Check	OK	19:00
Air Blank	0.000	19:01
Control Test	0.080	19:01
Air Blank	0.000	19:02
Subject Sample # 1	0.099	19:04
Breath Volume	1.875	19:04
Air Blank	0.000	19:05
Air Blank	0.000	19:07
Subject Sample # 2	0.094	19:07
Breath Volume	1.675	19:07
Air Blank	0.000	19:08
Control Test	0.078	19:08
Air Blank	0.000	19:09
Diagnostics Check	OK	19:09

Data Download Date/Time: 07/30/2020 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/04/2020 23:02:18

Diagnostics Check	OK	22:57	
Air Blank	0.000	22:57	
Control Test	0.079	22:57	
Air Blank	0.000	22:58	
Subject Sample	# 1	23:00	Subject Test Refused
Breath Volume	0.000	23:00	
Air Blank	0.000	23:01	
Control Test	0.080	23:01	
Air Blank	0.000	23:02	
Diagnostics Check	OK	23:02	

Data Download Date/Time: 07/30/2020 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: GUSTAFSON

DATE/TIME OF TEST: 07/05/2020 04:02:33

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.000	03:58
Breath Volume	3.296	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.000	04:01
Breath Volume	2.347	04:01
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 07/30/2020 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REVELL,DENNIS,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: REVELL,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 07/13/2020 20:56:36

Diagnostics Check	OK	20:44	
Air Blank	0.000	20:45	
Control Test	0.080	20:45	
Air Blank	0.000	20:46	
Subject Sample	# 1	20:49	No Sample Provided
Breath Volume	1.179	20:49	
Air Blank	0.000	20:49	
Air Blank	0.000	20:51	
Subject Sample	# 2	20:52	
Breath Volume	3.687	20:52	
Air Blank	0.000	20:52	
Air Blank	0.000	20:54	
Subject Sample	# 3	20:55	
Breath Volume	3.855	20:55	
Air Blank	0.000	20:55	
Control Test	0.080	20:56	
Air Blank	0.000	20:56	
Diagnostics Check	OK	20:56	

Data Download Date/Time: 07/30/2020 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,VICKI,N
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: OWENS,JONATHAN
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: NUNAMAKER

DATE/TIME OF TEST: 07/18/2020 05:38:28

Diagnostics Check	OK	05:36	
Air Blank	0.000	05:37	
Control Test	0.160	05:37	Control Outside Tolerance
Air Blank	0.000	05:38	

Data Download Date/Time: 07/30/2020 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,VICKI,N
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: OWENS,JOHN
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 06/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: NUNAMAKER

DATE/TIME OF TEST: 07/18/2020 05:46:57

Diagnostics Check	OK	05:40	
Air Blank	0.000	05:41	
Control Test	0.079	05:41	
Air Blank	0.000	05:42	
Subject Sample	# 1 0.190	05:42	
Breath Volume	3.699	05:42	
Air Blank	0.000	05:43	
Air Blank	0.000	05:44	
Subject Sample	# 2 0.193	05:45	
Breath Volume	3.683	05:45	
Air Blank	0.000	05:46	
Control Test	0.079	05:46	RFI Detect
Air Blank	0.000	05:46	

Data Download Date/Time: 07/30/2020 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BROWNING,STEVEN
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: POTOCKI

DATE/TIME OF TEST: 07/25/2020 04:13:02

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:08	
Control Test	0.079	04:08	
Air Blank	0.000	04:09	
Subject Sample	# 1	04:11	Subject Test Refused
Breath Volume	0.000	04:11	
Air Blank	0.000	04:12	
Control Test	0.079	04:12	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 08/03/2020 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: COBB

DATE/TIME OF TEST: 07/28/2020 23:05:43

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.079	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.000	23:01
Breath Volume	3.218	23:01
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.000	23:04
Breath Volume	3.457	23:04
Air Blank	0.000	23:04
Control Test	0.080	23:05
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 08/06/2020 10:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRODTMAN,ERIC,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BLAS,KRISTIAN
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: KHAN

DATE/TIME OF TEST: 07/01/2020 22:33:56

Diagnostics Check	OK	22:31	
Air Blank	0.000	22:31	
Control Test	0.078	22:31	
Air Blank	0.000	22:32	
Subject Sample	# 1	22:32	Subject Test Refused
Breath Volume	0.000	22:32	
Air Blank	0.000	22:33	
Control Test	0.077	22:33	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:33	

Data Download Date/Time: 07/06/2020 20:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHOBREY,SPENCER,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WHOBREY,SPENCER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: RODRIGUEZ RIVERA

DATE/TIME OF TEST: 07/06/2020 04:47:00

Diagnostics Check	OK	04:38
Air Blank	0.000	04:39
Control Test	0.077	04:39
Air Blank	0.000	04:39
Subject Sample # 1	0.148	04:41
Breath Volume	2.578	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:43
Subject Sample # 2	0.147	04:45
Breath Volume	2.367	04:45
Air Blank	0.000	04:46
Control Test	0.078	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:47

Data Download Date/Time: 07/06/2020 20:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,SHAWN,N
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SWEENEY,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 08:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: ALVAREZ ABREU

DATE/TIME OF TEST: 07/11/2020 08:39:53

Diagnostics Check	OK	08:31
Air Blank	0.000	08:32
Control Test	0.079	08:32
Air Blank	0.000	08:33
Subject Sample # 1	0.092	08:35
Breath Volume	3.500	08:35
Air Blank	0.000	08:35
Air Blank	0.000	08:37
Subject Sample # 2	0.091	08:38
Breath Volume	2.132	08:38
Air Blank	0.000	08:38
Control Test	0.079	08:39
Air Blank	0.000	08:39
Diagnostics Check	OK	08:39

Data Download Date/Time: 07/14/2020 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 07/15/2020 06:38:23

Diagnostics Check	OK	06:32
Air Blank	0.000	06:32
Control Test	0.081	06:33
Air Blank	0.000	06:33
Subject Sample # 1	0.000	06:34
Breath Volume	3.429	06:34
Air Blank	0.000	06:34
Air Blank	0.000	06:36
Subject Sample # 2	0.000	06:37
Breath Volume	3.261	06:37
Air Blank	0.000	06:37
Control Test	0.080	06:37
Air Blank	0.000	06:38
Diagnostics Check	OK	06:38

Data Download Date/Time: 08/03/2020 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ,BAYRON,L
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: PHILLIPS,ADAM
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 07/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 10:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:47

SUBJECT'S LAST NAME: APONTE

DATE/TIME OF TEST: 07/16/2020 11:14:19

Diagnostics Check	OK	11:08
Air Blank	0.000	11:08
Control Test	0.081	11:09
Air Blank	0.000	11:09
Subject Sample # 1	0.000	11:10
Breath Volume	1.980	11:10
Air Blank	0.000	11:10
Air Blank	0.000	11:12
Subject Sample # 2	0.000	11:12
Breath Volume	2.421	11:12
Air Blank	0.000	11:13
Control Test	0.080	11:13
Air Blank	0.000	11:14
Diagnostics Check	OK	11:14

Data Download Date/Time: 08/03/2020 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSAK,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RUSSAK,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: ABEL

DATE/TIME OF TEST: 07/16/2020 21:27:28

Diagnostics Check	OK	21:20
Air Blank	0.000	21:21
Control Test	0.078	21:21
Air Blank	0.000	21:22
Subject Sample # 1	0.165	21:22
Breath Volume	1.992	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.170	21:25
Breath Volume	2.367	21:25
Air Blank	0.000	21:26
Control Test	0.078	21:26
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 08/03/2020 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JOSHUA
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: SALINAS-CALDERON

DATE/TIME OF TEST: 07/18/2020 01:54:28

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.080	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.124	01:49
Breath Volume	2.832	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.112	01:52
Breath Volume	1.699	01:52
Air Blank	0.000	01:53
Control Test	0.078	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 08/03/2020 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOUVRES,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: DOUVRES,MICHAEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 07/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: TRIGO DE SERRANO VAS

DATE/TIME OF TEST: 07/26/2020 04:37:17

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:22	
Control Test	0.080	04:23	
Air Blank	0.000	04:23	
Subject Sample	# 1 0.170	04:24	
Breath Volume	1.453	04:24	
Air Blank	0.000	04:25	
Air Blank	0.000	04:26	
Subject Sample	# 2 0.142	04:30	Volume Not Met
Breath Volume	0.019	04:30	
Air Blank	0.000	04:30	
Air Blank	0.000	04:32	
Subject Sample	# 3 0.165	04:35	Volume Not Met
Breath Volume	0.128	04:35	
Air Blank	0.000	04:36	
Control Test	0.079	04:36	
Air Blank	0.000	04:37	
Diagnostics Check	OK	04:37	

Data Download Date/Time: 08/03/2020 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARONEY,GEORGE,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: RAMIREZ,JOSE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ESCOLASTICO

DATE/TIME OF TEST: 07/27/2020 03:13:24

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.078	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.085	03:08
Breath Volume	1.402	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.092	03:11
Breath Volume	2.605	03:11
Air Blank	0.000	03:12
Control Test	0.078	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 08/03/2020 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,DAVID,
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: ORTIZ,DAVID
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/22/2020 21:19:30

Diagnostics Check	OK	21:14	
Air Blank	0.000	21:15	
Control Test	0.079	21:15	
Air Blank	0.000	21:16	
Subject Sample # 1	0.200	21:16	
Breath Volume	5.316	21:16	
Air Blank	0.000	21:16	
Air Blank		21:18	Ambient Fail
Air Blank	0.000	21:19	

Data Download Date/Time: 08/28/2020 17:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,DAVID,
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: ORTIZ,DAVID
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/22/2020 21:30:02

Diagnostics Check	OK	21:22	
Air Blank	0.000	21:22	
Control Test	0.078	21:23	
Air Blank	0.000	21:23	
Subject Sample # 1	0.222	21:24	
Breath Volume	4.355	21:24	
Air Blank	0.000	21:25	
Air Blank	0.000	21:26	
Subject Sample # 2	0.215	21:28	
Breath Volume	4.710	21:28	
Air Blank	0.000	21:29	
Control Test	0.080	21:29	
Air Blank		21:29	RFI Detect
Air Blank	0.000	21:30	

Data Download Date/Time: 08/28/2020 17:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSEPH,A
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: SMITH,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: BAUM

DATE/TIME OF TEST: 07/26/2020 02:31:17

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:25	
Control Test	0.080	02:25	
Air Blank	0.000	02:25	
Subject Sample # 1	0.000	02:26	
Breath Volume	2.375	02:26	
Air Blank	0.000	02:27	
Air Blank	0.000	02:29	
Subject Sample # 2	0.000	02:29	
Breath Volume	1.582	02:29	
Air Blank	0.000	02:30	
Control Test	0.079	02:30	
Air Blank		02:30	RFI Detect
Air Blank	0.000	02:31	

Data Download Date/Time: 08/28/2020 17:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,JACOB,T
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: BOND,BRIAN
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: HALLOCK

DATE/TIME OF TEST: 07/20/2020 23:49:53

Diagnostics Check		OK	23:42
Air Blank		0.000	23:43
Control Test		0.081	23:43
Air Blank		0.000	23:43
Subject Sample	# 1	0.283	23:44
Breath Volume		2.078	23:44
Air Blank		0.000	23:45
Air Blank		0.000	23:47
Subject Sample	# 2	0.290	23:48
Breath Volume		2.207	23:48
Air Blank		0.000	23:48
Control Test		0.080	23:49
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 08/05/2020 08:48