

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PATSCH,JOSHUA,K  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: PATSCH,JOSHUA  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: MANDEVILLE

DATE/TIME OF TEST: 09/05/2020 01:32:17

---

|                   |     |       |       |                      |
|-------------------|-----|-------|-------|----------------------|
| Diagnostics Check |     | OK    | 01:22 |                      |
| Air Blank         |     | 0.000 | 01:23 |                      |
| Control Test      |     | 0.079 | 01:23 |                      |
| Air Blank         |     | 0.000 | 01:24 |                      |
| Subject Sample    | # 1 | 0.163 | 01:27 | Volume Not Met       |
| Breath Volume     |     | 0.425 | 01:27 |                      |
| Air Blank         |     | 0.000 | 01:27 |                      |
| Air Blank         |     | 0.000 | 01:29 |                      |
| Subject Sample    | # 2 |       | 01:30 | Subject Test Refused |
| Breath Volume     |     | 0.000 | 01:30 |                      |
| Air Blank         |     | 0.000 | 01:31 |                      |
| Control Test      |     | 0.078 | 01:31 |                      |
| Air Blank         |     | 0.000 | 01:32 |                      |
| Diagnostics Check |     | OK    | 01:32 |                      |

---

Data Download Date/Time: 09/14/2020 15:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,THOMAS,H  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: EVANS,THOMAS  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: SHUGART

DATE/TIME OF TEST: 09/02/2020 00:18:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:10 |
| Air Blank          | 0.000 | 00:10 |
| Control Test       | 0.080 | 00:11 |
| Air Blank          | 0.000 | 00:11 |
| Subject Sample # 1 | 0.312 | 00:13 |
| Breath Volume      | 2.113 | 00:13 |
| Air Blank          | 0.000 | 00:14 |
| Air Blank          | 0.000 | 00:15 |
| Subject Sample # 2 | 0.306 | 00:16 |
| Breath Volume      | 1.402 | 00:16 |
| Air Blank          | 0.000 | 00:17 |
| Control Test       | 0.078 | 00:17 |
| Air Blank          | 0.000 | 00:18 |
| Diagnostics Check  | OK    | 00:18 |

---

Data Download Date/Time: 09/21/2020 11:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,MASON,E  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: CAMDEN,DENNIS  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/02/2020 03:16:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:10 |
| Air Blank          | 0.000 | 03:10 |
| Control Test       | 0.079 | 03:11 |
| Air Blank          | 0.000 | 03:11 |
| Subject Sample # 1 | 0.120 | 03:12 |
| Breath Volume      | 2.949 | 03:12 |
| Air Blank          | 0.000 | 03:12 |
| Air Blank          | 0.000 | 03:14 |
| Subject Sample # 2 | 0.118 | 03:15 |
| Breath Volume      | 3.484 | 03:15 |
| Air Blank          | 0.000 | 03:15 |
| Control Test       | 0.080 | 03:15 |
| Air Blank          | 0.000 | 03:16 |
| Diagnostics Check  | OK    | 03:16 |

---

Data Download Date/Time: 09/21/2020 11:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,MASON,E  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: HIGHLANDER,JACOB  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: MUSSOLINE

DATE/TIME OF TEST: 09/04/2020 18:39:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:32 |
| Air Blank          | 0.000 | 18:32 |
| Control Test       | 0.080 | 18:33 |
| Air Blank          | 0.000 | 18:33 |
| Subject Sample # 1 | 0.031 | 18:35 |
| Breath Volume      | 2.875 | 18:35 |
| Air Blank          | 0.000 | 18:35 |
| Air Blank          | 0.000 | 18:37 |
| Subject Sample # 2 | 0.028 | 18:38 |
| Breath Volume      | 3.375 | 18:38 |
| Air Blank          | 0.000 | 18:38 |
| Control Test       | 0.080 | 18:39 |
| Air Blank          | 0.000 | 18:39 |
| Diagnostics Check  | OK    | 18:39 |

---

Data Download Date/Time: 09/21/2020 11:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRUSCIO,JESSE,J  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: EVANS,IAN  
ARRESTING OFFICER AGENCY: St. Augustine Beach  
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 11:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: KATZ

DATE/TIME OF TEST: 09/06/2020 00:29:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:20 |
| Air Blank          | 0.000 | 00:21 |
| Control Test       | 0.080 | 00:21 |
| Air Blank          | 0.000 | 00:22 |
| Subject Sample # 1 | 0.157 | 00:24 |
| Breath Volume      | 1.375 | 00:24 |
| Air Blank          | 0.000 | 00:25 |
| Air Blank          | 0.000 | 00:27 |
| Subject Sample # 2 | 0.157 | 00:27 |
| Breath Volume      | 1.179 | 00:27 |
| Air Blank          | 0.000 | 00:28 |
| Control Test       | 0.078 | 00:28 |
| Air Blank          | 0.000 | 00:29 |
| Diagnostics Check  | OK    | 00:29 |

---

Data Download Date/Time: 09/21/2020 11:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRUSCIO,JESSE,J  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: BRIGGS,JASON  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: ORTEGA

DATE/TIME OF TEST: 09/06/2020 01:12:37

---

|                   |     |       |       |                    |
|-------------------|-----|-------|-------|--------------------|
| Diagnostics Check |     | OK    | 00:56 |                    |
| Air Blank         |     | 0.000 | 00:56 |                    |
| Control Test      |     | 0.080 | 00:56 |                    |
| Air Blank         |     | 0.000 | 00:57 |                    |
| Subject Sample    | # 1 |       | 01:00 | No Sample Provided |
| Breath Volume     |     | 0.000 | 01:00 |                    |
| Air Blank         |     | 0.000 | 01:00 |                    |
| Air Blank         |     | 0.000 | 01:02 |                    |
| Subject Sample    | # 2 | 0.181 | 01:05 |                    |
| Breath Volume     |     | 2.011 | 01:05 |                    |
| Air Blank         |     | 0.000 | 01:06 |                    |
| Air Blank         |     | 0.000 | 01:08 |                    |
| Subject Sample    | # 3 | 0.161 | 01:10 |                    |
| Breath Volume     |     | 1.222 | 01:10 |                    |
| Air Blank         |     | 0.000 | 01:11 |                    |
| Control Test      |     | 0.078 | 01:11 |                    |
| Air Blank         |     | 0.000 | 01:12 |                    |
| Diagnostics Check |     | OK    | 01:12 |                    |

---

Data Download Date/Time: 09/21/2020 11:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRUSCIO,JESSE,J  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: CASH,JUSTIN  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: GRENNAN

DATE/TIME OF TEST: 09/06/2020 03:28:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:19 |
| Air Blank          | 0.000 | 03:20 |
| Control Test       | 0.080 | 03:20 |
| Air Blank          | 0.000 | 03:21 |
| Subject Sample # 1 | 0.150 | 03:23 |
| Breath Volume      | 1.359 | 03:23 |
| Air Blank          | 0.000 | 03:24 |
| Air Blank          | 0.000 | 03:26 |
| Subject Sample # 2 | 0.148 | 03:26 |
| Breath Volume      | 1.296 | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Control Test       | 0.079 | 03:27 |
| Air Blank          | 0.000 | 03:28 |
| Diagnostics Check  | OK    | 03:28 |

---

Data Download Date/Time: 09/21/2020 11:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMS,WALKER,L  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: BRIGGS,JASON  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 12:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 09/17/2020 01:17:42

---

|                    |       |       |                           |
|--------------------|-------|-------|---------------------------|
| Diagnostics Check  | OK    | 01:07 |                           |
| Air Blank          | 0.000 | 01:07 |                           |
| Control Test       | 0.080 | 01:07 |                           |
| Air Blank          | 0.000 | 01:08 |                           |
| Subject Sample # 1 | 0.240 | 01:10 |                           |
| Breath Volume      | 2.347 | 01:10 |                           |
| Air Blank          | 0.000 | 01:11 |                           |
| Air Blank          | 0.000 | 01:13 |                           |
| Subject Sample # 2 | 0.241 | 01:16 |                           |
| Breath Volume      | 1.750 | 01:16 |                           |
| Air Blank          | 0.000 | 01:16 |                           |
| Control Test       | 0.073 | 01:17 | Control Outside Tolerance |
| Air Blank          | 0.000 | 01:17 |                           |

---

Data Download Date/Time: 09/21/2020 11:20



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMS,WALKER,L  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: POWELL,AARON  
ARRESTING OFFICER AGENCY: St. Augustine Beach  
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 12:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: URBANIAK

DATE/TIME OF TEST: 09/18/2020 01:04:31

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 00:58 |
| Air Blank         |     | 0.000 | 00:58 |
| Control Test      |     | 0.080 | 00:59 |
| Air Blank         |     | 0.000 | 00:59 |
| Subject Sample    | # 1 | 0.213 | 01:00 |
| Breath Volume     |     | 1.218 | 01:00 |
| Air Blank         |     | 0.000 | 01:00 |
| Air Blank         |     | 0.000 | 01:02 |
| Subject Sample    | # 2 | 0.213 | 01:03 |
| Breath Volume     |     | 1.109 | 01:03 |
| Air Blank         |     | 0.000 | 01:03 |
| Control Test      |     | 0.080 | 01:03 |
| Air Blank         |     | 0.000 | 01:04 |
| Diagnostics Check |     | OK    | 01:04 |

---

Data Download Date/Time: 09/21/2020 11:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMS,WALKER,L  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: TROTZKE,JENNIFER  
ARRESTING OFFICER AGENCY: St. Augustine PD  
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: GANTT

DATE/TIME OF TEST: 09/18/2020 01:32:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:26 |
| Air Blank          | 0.000 | 01:26 |
| Control Test       | 0.081 | 01:27 |
| Air Blank          | 0.000 | 01:27 |
| Subject Sample # 1 | 0.250 | 01:28 |
| Breath Volume      | 2.078 | 01:28 |
| Air Blank          | 0.000 | 01:28 |
| Air Blank          | 0.000 | 01:30 |
| Subject Sample # 2 | 0.249 | 01:31 |
| Breath Volume      | 1.976 | 01:31 |
| Air Blank          | 0.000 | 01:31 |
| Control Test       | 0.080 | 01:32 |
| Air Blank          | 0.000 | 01:32 |
| Diagnostics Check  | OK    | 01:32 |

---

Data Download Date/Time: 09/21/2020 11:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TRUSCIO,JESSE,J  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: BRIGGS,J  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: ROMAINE

DATE/TIME OF TEST: 09/20/2020 02:53:24

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:48 |                      |
| Air Blank         | 0.000 | 02:48 |                      |
| Control Test      | 0.080 | 02:49 |                      |
| Air Blank         | 0.000 | 02:49 |                      |
| Subject Sample    | # 1   | 02:52 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:52 |                      |
| Air Blank         | 0.000 | 02:52 |                      |
| Control Test      | 0.080 | 02:52 |                      |
| Air Blank         | 0.000 | 02:53 |                      |
| Diagnostics Check | OK    | 02:53 |                      |

---

Data Download Date/Time: 09/21/2020 11:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRESTON,MIKAYLA,D  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: MANGOLD,KENNY  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 09/21/2020 01:25:39

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 01:14 |                    |
| Air Blank         | 0.000 | 01:14 |                    |
| Control Test      | 0.081 | 01:15 |                    |
| Air Blank         | 0.000 | 01:15 |                    |
| Subject Sample    | # 1   | 01:18 | No Sample Provided |
| Breath Volume     | 0.000 | 01:18 |                    |
| Air Blank         | 0.000 | 01:19 |                    |
| Air Blank         | 0.000 | 01:21 |                    |
| Subject Sample    | # 2   | 01:24 | No Sample Provided |
| Breath Volume     | 0.000 | 01:24 |                    |
| Air Blank         | 0.000 | 01:24 |                    |
| Control Test      | 0.080 | 01:25 |                    |
| Air Blank         | 0.000 | 01:25 |                    |
| Diagnostics Check | OK    | 01:25 |                    |

---

Data Download Date/Time: 09/21/2020 11:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,SCOTT,B  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: NARVAEZ-PEREZ,ANDRES  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: FRENCH

DATE/TIME OF TEST: 09/22/2020 02:56:10

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 02:49 |
| Air Blank         |     | 0.000 | 02:50 |
| Control Test      |     | 0.082 | 02:50 |
| Air Blank         |     | 0.000 | 02:51 |
| Subject Sample    | # 1 | 0.216 | 02:51 |
| Breath Volume     |     | 1.226 | 02:51 |
| Air Blank         |     | 0.000 | 02:52 |
| Air Blank         |     | 0.000 | 02:54 |
| Subject Sample    | # 2 | 0.214 | 02:54 |
| Breath Volume     |     | 1.535 | 02:54 |
| Air Blank         |     | 0.000 | 02:55 |
| Control Test      |     | 0.080 | 02:55 |
| Air Blank         |     | 0.000 | 02:56 |
| Diagnostics Check |     | OK    | 02:56 |

---

Data Download Date/Time: 10/19/2020 11:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRESTON,MIKAYLA,D  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: PRESTON,MIKAYLA  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: MORAN

DATE/TIME OF TEST: 09/23/2020 23:36:02

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:32 |                      |
| Air Blank         | 0.000 | 23:32 |                      |
| Control Test      | 0.082 | 23:33 |                      |
| Air Blank         | 0.000 | 23:33 |                      |
| Subject Sample    | # 1   | 23:34 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:34 |                      |
| Air Blank         | 0.000 | 23:35 |                      |
| Control Test      | 0.082 | 23:35 |                      |
| Air Blank         | 0.000 | 23:35 |                      |
| Diagnostics Check | OK    | 23:36 |                      |

---

Data Download Date/Time: 10/19/2020 11:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRESTON,MIKAYLA,D  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: PRESTON,MIKAYLA  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 09/25/2020 20:32:59

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:30 |                      |
| Air Blank         | 0.000 | 20:30 |                      |
| Control Test      | 0.082 | 20:30 |                      |
| Air Blank         | 0.000 | 20:31 |                      |
| Subject Sample    | # 1   | 20:31 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:31 |                      |
| Air Blank         | 0.000 | 20:32 |                      |
| Control Test      | 0.082 | 20:32 |                      |
| Air Blank         | 0.000 | 20:32 |                      |
| Diagnostics Check | OK    | 20:32 |                      |

---

Data Download Date/Time: 10/19/2020 11:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,STERLIN,R  
BREATH TEST OPERATOR AGENCY: Daytona Beach PD  
ARRESTING OFFICER: DEMNER,AUTUMN  
ARRESTING OFFICER AGENCY: Daytona Beach PD  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/08/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: ROIG

DATE/TIME OF TEST: 09/05/2020 03:48:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:41 |
| Air Blank          | 0.000 | 03:42 |
| Control Test       | 0.082 | 03:42 |
| Air Blank          | 0.000 | 03:42 |
| Subject Sample # 1 | 0.158 | 03:43 |
| Breath Volume      | 1.773 | 03:43 |
| Air Blank          | 0.000 | 03:44 |
| Air Blank          | 0.000 | 03:45 |
| Subject Sample # 2 | 0.157 | 03:46 |
| Breath Volume      | 2.023 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Control Test       | 0.081 | 03:47 |
| Air Blank          | 0.000 | 03:48 |
| Diagnostics Check  | OK    | 03:48 |

---

Data Download Date/Time: 09/21/2020 13:24



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BODKIN,MICHEAL,W  
BREATH TEST OPERATOR AGENCY: Daytona Beach PD  
ARRESTING OFFICER: BODKIN,MICHEAL  
ARRESTING OFFICER AGENCY: Daytona Beach PD  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/08/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DOMINICK

DATE/TIME OF TEST: 09/09/2020 02:16:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:10 |
| Air Blank          | 0.000 | 02:10 |
| Control Test       | 0.081 | 02:10 |
| Air Blank          | 0.000 | 02:11 |
| Subject Sample # 1 | 0.211 | 02:11 |
| Breath Volume      | 1.714 | 02:11 |
| Air Blank          | 0.000 | 02:12 |
| Air Blank          | 0.000 | 02:14 |
| Subject Sample # 2 | 0.202 | 02:14 |
| Breath Volume      | 1.539 | 02:14 |
| Air Blank          | 0.000 | 02:15 |
| Control Test       | 0.081 | 02:16 |
| Air Blank          | 0.000 | 02:16 |
| Diagnostics Check  | OK    | 02:16 |

---

Data Download Date/Time: 09/21/2020 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BODKIN,MICHEAL,W  
BREATH TEST OPERATOR AGENCY: Daytona Beach PD  
ARRESTING OFFICER: BODKIN,MICHEAL  
ARRESTING OFFICER AGENCY: Daytona Beach PD  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/08/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: PRATT

DATE/TIME OF TEST: 09/12/2020 02:04:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:54 |
| Air Blank          | 0.000 | 01:54 |
| Control Test       | 0.082 | 01:55 |
| Air Blank          | 0.000 | 01:55 |
| Subject Sample # 1 | 0.211 | 01:55 |
| Breath Volume      | 1.289 | 01:55 |
| Air Blank          | 0.000 | 01:56 |
| Air Blank          | 0.000 | 01:58 |
| Subject Sample # 2 | 0.186 | 01:59 |
| Breath Volume      | 2.082 | 01:59 |
| Air Blank          | 0.000 | 02:00 |
| Air Blank          | 0.000 | 02:01 |
| Subject Sample # 3 | 0.180 | 02:02 |
| Breath Volume      | 2.011 | 02:02 |
| Air Blank          | 0.000 | 02:03 |
| Control Test       | 0.079 | 02:03 |
| Air Blank          | 0.000 | 02:04 |
| Diagnostics Check  | OK    | 02:04 |

---

Data Download Date/Time: 09/21/2020 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BODKIN,MICHEAL,W  
BREATH TEST OPERATOR AGENCY: Daytona Beach PD  
ARRESTING OFFICER: BODKIN,MICHEAL  
ARRESTING OFFICER AGENCY: Daytona Beach PD  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/08/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: WARNOCK

DATE/TIME OF TEST: 09/13/2020 01:35:39

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 01:29 |
| Air Blank         |     | 0.000 | 01:29 |
| Control Test      |     | 0.081 | 01:29 |
| Air Blank         |     | 0.000 | 01:30 |
| Subject Sample    | # 1 | 0.239 | 01:30 |
| Breath Volume     |     | 2.671 | 01:30 |
| Air Blank         |     | 0.000 | 01:31 |
| Air Blank         |     | 0.000 | 01:33 |
| Subject Sample    | # 2 | 0.238 | 01:33 |
| Breath Volume     |     | 3.289 | 01:33 |
| Air Blank         |     | 0.000 | 01:34 |
| Control Test      |     | 0.080 | 01:35 |
| Air Blank         |     | 0.000 | 01:35 |
| Diagnostics Check |     | OK    | 01:35 |

---

Data Download Date/Time: 09/21/2020 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BISSA,JAMES,D  
BREATH TEST OPERATOR AGENCY: Daytona Beach PD  
ARRESTING OFFICER: WEBB,ADAM  
ARRESTING OFFICER AGENCY: Daytona Beach PD  
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/08/2022  
VIOLATION CODE: DUI  
ARREST TIME: 07:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:57

SUBJECT'S LAST NAME: BIRCHETT

DATE/TIME OF TEST: 09/26/2020 08:26:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:20 |
| Air Blank          | 0.000 | 08:20 |
| Control Test       | 0.082 | 08:20 |
| Air Blank          | 0.000 | 08:21 |
| Subject Sample # 1 | 0.155 | 08:22 |
| Breath Volume      | 3.460 | 08:22 |
| Air Blank          | 0.000 | 08:22 |
| Air Blank          | 0.000 | 08:24 |
| Subject Sample # 2 | 0.155 | 08:25 |
| Breath Volume      | 2.445 | 08:25 |
| Air Blank          | 0.000 | 08:25 |
| Control Test       | 0.081 | 08:26 |
| Air Blank          | 0.000 | 08:26 |
| Diagnostics Check  | OK    | 08:26 |

---

Data Download Date/Time: 10/29/2020 13:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO  
LAST AGENCY INSPECTION: 08/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: RIVERA HERNANDEZ

DATE/TIME OF TEST: 09/11/2020 22:09:52

---

|                   |       |       |              |
|-------------------|-------|-------|--------------|
| Diagnostics Check | OK    | 22:08 |              |
| Air Blank         |       | 22:09 | Ambient Fail |
| Air Blank         | 0.000 | 22:09 |              |

---

Data Download Date/Time: 09/29/2020 11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO  
LAST AGENCY INSPECTION: 08/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: RIVERA HERNANDEZ

DATE/TIME OF TEST: 09/11/2020 22:13:54

---

|                   |       |       |              |
|-------------------|-------|-------|--------------|
| Diagnostics Check | OK    | 22:12 |              |
| Air Blank         | 0.000 | 22:12 |              |
| Control Test      | 0.080 | 22:12 |              |
| Air Blank         |       | 22:13 | Ambient Fail |
| Air Blank         | 0.000 | 22:13 |              |

---

Data Download Date/Time: 09/29/2020 11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO  
LAST AGENCY INSPECTION: 08/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: RIVERA HERNANDEZ

DATE/TIME OF TEST: 09/11/2020 22:18:04

---

|                   |       |       |              |
|-------------------|-------|-------|--------------|
| Diagnostics Check | OK    | 22:16 |              |
| Air Blank         |       | 22:17 | Ambient Fail |
| Air Blank         | 0.000 | 22:18 |              |

---

Data Download Date/Time: 09/29/2020 11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO  
LAST AGENCY INSPECTION: 08/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/13/2020 02:49:51

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 02:45 |                    |
| Air Blank         | 0.000 | 02:46 |                    |
| Control Test      | 0.081 | 02:46 |                    |
| Air Blank         | 0.000 | 02:46 |                    |
| Subject Sample    | # 1   | 02:47 | Interferent Detect |
| Breath Volume     | 1.585 | 02:47 |                    |
| Air Blank         | 0.000 | 02:48 |                    |
| Control Test      | 0.080 | 02:49 |                    |
| Air Blank         | 0.000 | 02:49 |                    |
| Diagnostics Check | OK    | 02:49 |                    |

---

Data Download Date/Time: 09/29/2020 11:04



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO  
LAST AGENCY INSPECTION: 08/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/13/2020 02:57:43

---

|                    |       |       |            |
|--------------------|-------|-------|------------|
| Diagnostics Check  | OK    | 02:51 |            |
| Air Blank          | 0.000 | 02:51 |            |
| Control Test       | 0.082 | 02:52 |            |
| Air Blank          | 0.000 | 02:52 |            |
| Subject Sample # 1 | 0.180 | 02:53 |            |
| Breath Volume      | 1.441 | 02:53 |            |
| Air Blank          | 0.000 | 02:54 |            |
| Air Blank          | 0.000 | 02:55 |            |
| Subject Sample # 2 | 0.174 | 02:56 |            |
| Breath Volume      | 1.128 | 02:56 |            |
| Air Blank          |       | 02:57 | Purge Fail |
| Air Blank          | 0.000 | 02:57 |            |

---

Data Download Date/Time: 09/29/2020 11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO  
LAST AGENCY INSPECTION: 08/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/13/2020 02:59:56

---

|                   |       |       |              |
|-------------------|-------|-------|--------------|
| Diagnostics Check | OK    | 02:59 |              |
| Air Blank         |       | 02:59 | Ambient Fail |
| Air Blank         | 0.000 | 02:59 |              |

---

Data Download Date/Time: 09/29/2020 11:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO  
LAST AGENCY INSPECTION: 08/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: SYS CHECK  
ARREST TIME: 16:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 09/13/2020 16:56:51

---

|                    |       |       |                    |
|--------------------|-------|-------|--------------------|
| Diagnostics Check  | OK    | 16:50 |                    |
| Air Blank          | 0.000 | 16:50 |                    |
| Control Test       | 0.080 | 16:51 |                    |
| Air Blank          | 0.000 | 16:51 |                    |
| Subject Sample # 1 | 0.000 | 16:52 |                    |
| Breath Volume      | 2.375 | 16:52 |                    |
| Air Blank          | 0.000 | 16:52 |                    |
| Air Blank          | 0.000 | 16:54 |                    |
| Subject Sample # 2 |       | 16:54 | Interferent Detect |
| Breath Volume      | 2.335 | 16:54 |                    |
| Air Blank          | 0.000 | 16:55 |                    |
| Control Test       | 0.084 | 16:56 |                    |
| Air Blank          | 0.000 | 16:56 |                    |
| Diagnostics Check  | OK    | 16:56 |                    |

---

Data Download Date/Time: 09/29/2020 11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO  
LAST AGENCY INSPECTION: 08/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: SYS CHECK  
ARREST TIME: 16:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:01

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 09/13/2020 17:01:12

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 16:59 |                    |
| Air Blank         | 0.000 | 17:00 |                    |
| Control Test      | 0.076 | 17:00 | Interferent Detect |
| Air Blank         | 0.000 | 17:01 |                    |

---

Data Download Date/Time: 09/29/2020 11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C  
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO  
ARRESTING OFFICER: MONTEMAGNO,CARLO  
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO  
LAST AGENCY INSPECTION: 08/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: SYS CHECK  
ARREST TIME: 17:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 09/13/2020 17:29:10

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 17:28 |            |
| Air Blank         |       | 17:28 | RFI Detect |
| Air Blank         | 0.000 | 17:29 |            |

---

Data Download Date/Time: 09/29/2020 11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C  
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO  
ARRESTING OFFICER: MONTEMAGNO,CARLO  
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO  
LAST AGENCY INSPECTION: 08/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: SYS CHECK  
ARREST TIME: 17:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 09/13/2020 17:32:37

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 17:30 |            |
| Air Blank         | 0.000 | 17:30 |            |
| Control Test      | 0.080 | 17:31 |            |
| Air Blank         | 0.000 | 17:31 |            |
| Subject Sample    | # 1   | 17:31 | RFI Detect |
| Breath Volume     | 0.000 | 17:31 |            |
| Air Blank         |       | 17:32 | RFI Detect |
| Air Blank         | 0.000 | 17:32 |            |

---

Data Download Date/Time: 09/29/2020 11:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C  
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO  
ARRESTING OFFICER: MONTEMAGNO,CARLO  
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO  
LAST AGENCY INSPECTION: 08/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: SYS CHECK  
ARREST TIME: 17:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 09/13/2020 17:37:34

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 17:34 |            |
| Air Blank         | 0.000 | 17:35 |            |
| Control Test      | 0.080 | 17:35 |            |
| Air Blank         | 0.000 | 17:35 |            |
| Subject Sample    | # 1   | 17:36 | RFI Detect |
| Breath Volume     | 0.000 | 17:36 |            |
| Air Blank         | 0.000 | 17:36 |            |
| Control Test      | 0.080 | 17:36 |            |
| Air Blank         | 0.000 | 17:37 |            |
| Diagnostics Check | OK    | 17:37 |            |

---

Data Download Date/Time: 09/29/2020 11:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C  
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO  
ARRESTING OFFICER: MONTEMAGNO,CARLO  
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO  
LAST AGENCY INSPECTION: 08/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: SYS CHECK  
ARREST TIME: 17:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 09/13/2020 17:45:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:39 |
| Air Blank          | 0.000 | 17:39 |
| Control Test       | 0.079 | 17:39 |
| Air Blank          | 0.000 | 17:40 |
| Subject Sample # 1 | 0.000 | 17:40 |
| Breath Volume      | 1.839 | 17:40 |
| Air Blank          | 0.000 | 17:41 |
| Air Blank          | 0.000 | 17:43 |
| Subject Sample # 2 | 0.000 | 17:43 |
| Breath Volume      | 1.375 | 17:43 |
| Air Blank          | 0.000 | 17:44 |
| Control Test       | 0.079 | 17:44 |
| Air Blank          | 0.000 | 17:45 |
| Diagnostics Check  | OK    | 17:45 |

---

Data Download Date/Time: 09/29/2020 11:04



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO  
LAST AGENCY INSPECTION: 08/17/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: SYS CHECK  
ARREST TIME: 17:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 09/13/2020 18:00:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:54 |
| Air Blank          | 0.000 | 17:54 |
| Control Test       | 0.080 | 17:55 |
| Air Blank          | 0.000 | 17:55 |
| Subject Sample # 1 | 0.000 | 17:56 |
| Breath Volume      | 2.117 | 17:56 |
| Air Blank          | 0.000 | 17:56 |
| Air Blank          | 0.000 | 17:58 |
| Subject Sample # 2 | 0.000 | 17:58 |
| Breath Volume      | 2.265 | 17:58 |
| Air Blank          | 0.000 | 17:59 |
| Control Test       | 0.079 | 17:59 |
| Air Blank          | 0.000 | 18:00 |
| Diagnostics Check  | OK    | 18:00 |

---

Data Download Date/Time: 09/29/2020 11:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KALLMEYER,DANIEL,K  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: N/A,N/A  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: PROB/PAROL  
ARREST TIME: 00:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: HOKE

DATE/TIME OF TEST: 09/01/2020 14:37:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:30 |
| Air Blank          | 0.000 | 14:31 |
| Control Test       | 0.082 | 14:31 |
| Air Blank          | 0.000 | 14:32 |
| Subject Sample # 1 | 0.000 | 14:32 |
| Breath Volume      | 1.910 | 14:32 |
| Air Blank          | 0.000 | 14:33 |
| Air Blank          | 0.000 | 14:35 |
| Subject Sample # 2 | 0.000 | 14:35 |
| Breath Volume      | 1.703 | 14:35 |
| Air Blank          | 0.000 | 14:36 |
| Control Test       | 0.082 | 14:36 |
| Air Blank          | 0.000 | 14:36 |
| Diagnostics Check  | OK    | 14:37 |

---

Data Download Date/Time: 09/28/2020 07:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNER,SEAN,R  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: BRUNER,SEAN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: MOISE

DATE/TIME OF TEST: 09/06/2020 21:46:53

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 21:35 |                |
| Air Blank         |     | 0.000 | 21:35 |                |
| Control Test      |     | 0.083 | 21:36 |                |
| Air Blank         |     | 0.000 | 21:36 |                |
| Subject Sample    | # 1 | 0.033 | 21:39 | Volume Not Met |
| Breath Volume     |     | 0.000 | 21:39 |                |
| Air Blank         |     | 0.000 | 21:40 |                |
| Air Blank         |     | 0.000 | 21:42 |                |
| Subject Sample    | # 2 | 0.035 | 21:45 | Volume Not Met |
| Breath Volume     |     | 0.113 | 21:45 |                |
| Air Blank         |     | 0.000 | 21:45 |                |
| Control Test      |     | 0.083 | 21:46 |                |
| Air Blank         |     | 0.000 | 21:46 |                |
| Diagnostics Check |     | OK    | 21:46 |                |

---

Data Download Date/Time: 09/28/2020 07:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER, RAY, M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: CRUZ, EDGARDO  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 12:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:43

SUBJECT'S LAST NAME: ELLINGTON

DATE/TIME OF TEST: 09/08/2020 14:11:38

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 14:08 |                      |
| Air Blank         | 0.000 | 14:09 |                      |
| Control Test      | 0.082 | 14:09 |                      |
| Air Blank         | 0.000 | 14:09 |                      |
| Subject Sample    | # 1   | 14:10 | Subject Test Refused |
| Breath Volume     | 0.000 | 14:10 |                      |
| Air Blank         | 0.000 | 14:10 |                      |
| Control Test      | 0.081 | 14:11 |                      |
| Air Blank         | 0.000 | 14:11 |                      |
| Diagnostics Check | OK    | 14:11 |                      |

---

Data Download Date/Time: 09/28/2020 07:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER, RAY, M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: WEBER, RAY  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 12:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: HOKE

DATE/TIME OF TEST: 09/15/2020 13:18:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:12 |
| Air Blank          | 0.000 | 13:12 |
| Control Test       | 0.083 | 13:13 |
| Air Blank          | 0.000 | 13:13 |
| Subject Sample # 1 | 0.000 | 13:14 |
| Breath Volume      | 1.792 | 13:14 |
| Air Blank          | 0.000 | 13:14 |
| Air Blank          | 0.000 | 13:16 |
| Subject Sample # 2 | 0.000 | 13:16 |
| Breath Volume      | 2.175 | 13:16 |
| Air Blank          | 0.000 | 13:17 |
| Control Test       | 0.083 | 13:17 |
| Air Blank          | 0.000 | 13:18 |
| Diagnostics Check  | OK    | 13:18 |

---

Data Download Date/Time: 09/28/2020 07:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNER,SEAN,R  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: BRUNER,SEAN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: FRANCO

DATE/TIME OF TEST: 09/20/2020 03:40:43

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:37 |                      |
| Air Blank         | 0.000 | 03:38 |                      |
| Control Test      | 0.083 | 03:38 |                      |
| Air Blank         | 0.000 | 03:39 |                      |
| Subject Sample    | # 1   | 03:39 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:39 |                      |
| Air Blank         | 0.000 | 03:39 |                      |
| Control Test      | 0.081 | 03:40 |                      |
| Air Blank         | 0.000 | 03:40 |                      |
| Diagnostics Check | OK    | 03:40 |                      |

---

Data Download Date/Time: 09/28/2020 07:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,CHAD,B  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: CRUZ,EDGARDO  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 11:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:33

SUBJECT'S LAST NAME: EMMITT

DATE/TIME OF TEST: 09/27/2020 14:12:56

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 14:09 |                    |
| Air Blank         | 0.000 | 14:10 |                    |
| Control Test      | 0.082 | 14:10 |                    |
| Air Blank         | 0.000 | 14:10 |                    |
| Subject Sample    | # 1   | 14:11 | Interferent Detect |
| Breath Volume     | 1.847 | 14:11 |                    |
| Air Blank         | 0.000 | 14:12 |                    |
| Control Test      | 0.082 | 14:12 |                    |
| Air Blank         | 0.000 | 14:12 |                    |
| Diagnostics Check | OK    | 14:12 |                    |

---

Data Download Date/Time: 09/28/2020 07:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER, RAY, M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: WEBER, RAY  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 12:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/29/2020 13:03:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 12:58 |
| Air Blank          | 0.000 | 12:58 |
| Control Test       | 0.082 | 12:58 |
| Air Blank          | 0.000 | 12:59 |
| Subject Sample # 1 | 0.000 | 12:59 |
| Breath Volume      | 1.664 | 12:59 |
| Air Blank          | 0.000 | 13:00 |
| Air Blank          | 0.000 | 13:02 |
| Subject Sample # 2 | 0.000 | 13:02 |
| Breath Volume      | 1.746 | 13:02 |
| Air Blank          | 0.000 | 13:03 |
| Control Test       | 0.080 | 13:03 |
| Air Blank          | 0.000 | 13:03 |
| Diagnostics Check  | OK    | 13:03 |

---

Data Download Date/Time: 10/24/2020 06:53



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER, RAY, M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: WEBER, RAY  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 13:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: JACINTHE

DATE/TIME OF TEST: 09/29/2020 14:13:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:07 |
| Air Blank          | 0.000 | 14:08 |
| Control Test       | 0.081 | 14:08 |
| Air Blank          | 0.000 | 14:08 |
| Subject Sample # 1 | 0.000 | 14:09 |
| Breath Volume      | 1.757 | 14:09 |
| Air Blank          | 0.000 | 14:09 |
| Air Blank          | 0.000 | 14:11 |
| Subject Sample # 2 | 0.000 | 14:12 |
| Breath Volume      | 1.742 | 14:12 |
| Air Blank          | 0.000 | 14:12 |
| Control Test       | 0.080 | 14:12 |
| Air Blank          | 0.000 | 14:13 |
| Diagnostics Check  | OK    | 14:13 |

---

Data Download Date/Time: 10/24/2020 06:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REVE,JASON,R  
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD  
ARRESTING OFFICER: HUNT,RYAN  
ARRESTING OFFICER AGENCY: New Smyrna Beach PD  
LAST AGENCY INSPECTION: 08/20/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10818080A5  
DRY GAS EXPIRY DATE: 06/05/2020  
VIOLATION CODE: DUI  
ARREST TIME: 10:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 09/05/2020 11:37:07

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 11:30 |
| Air Blank         |     | 0.000 | 11:30 |
| Control Test      |     | 0.081 | 11:31 |
| Air Blank         |     | 0.000 | 11:31 |
| Subject Sample    | # 1 | 0.244 | 11:32 |
| Breath Volume     |     | 1.914 | 11:32 |
| Air Blank         |     | 0.000 | 11:32 |
| Air Blank         |     | 0.000 | 11:34 |
| Subject Sample    | # 2 | 0.259 | 11:35 |
| Breath Volume     |     | 2.363 | 11:35 |
| Air Blank         |     | 0.000 | 11:36 |
| Control Test      |     | 0.081 | 11:36 |
| Air Blank         |     | 0.000 | 11:37 |
| Diagnostics Check |     | OK    | 11:37 |

---

Data Download Date/Time: 09/20/2020 09:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HELMS,WILLIAM,A  
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD  
ARRESTING OFFICER: EASON,BEN  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 08/20/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: HELLER

DATE/TIME OF TEST: 09/05/2020 22:59:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:53 |
| Air Blank          | 0.000 | 22:53 |
| Control Test       | 0.081 | 22:54 |
| Air Blank          | 0.000 | 22:54 |
| Subject Sample # 1 | 0.071 | 22:55 |
| Breath Volume      | 2.292 | 22:55 |
| Air Blank          | 0.000 | 22:56 |
| Air Blank          | 0.000 | 22:57 |
| Subject Sample # 2 | 0.067 | 22:58 |
| Breath Volume      | 1.847 | 22:58 |
| Air Blank          | 0.000 | 22:58 |
| Control Test       | 0.080 | 22:59 |
| Air Blank          | 0.000 | 22:59 |
| Diagnostics Check  | OK    | 22:59 |

---

Data Download Date/Time: 09/20/2020 09:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUNNEFELD,RALPH,A  
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD  
ARRESTING OFFICER: TRZASKA,JOSEPH  
ARRESTING OFFICER AGENCY: New Smyrna Beach PD  
LAST AGENCY INSPECTION: 08/20/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: FARRELLY

DATE/TIME OF TEST: 09/07/2020 23:07:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:01 |
| Air Blank          | 0.000 | 23:01 |
| Control Test       | 0.081 | 23:02 |
| Air Blank          | 0.000 | 23:02 |
| Subject Sample # 1 | 0.171 | 23:03 |
| Breath Volume      | 1.855 | 23:03 |
| Air Blank          | 0.000 | 23:03 |
| Air Blank          | 0.000 | 23:05 |
| Subject Sample # 2 | 0.171 | 23:05 |
| Breath Volume      | 2.285 | 23:05 |
| Air Blank          | 0.000 | 23:06 |
| Control Test       | 0.080 | 23:06 |
| Air Blank          | 0.000 | 23:07 |
| Diagnostics Check  | OK    | 23:07 |

---

Data Download Date/Time: 09/20/2020 09:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: INGRAM,TRISTAN,N  
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD  
ARRESTING OFFICER: NORMAN,BENJAMIN  
ARRESTING OFFICER AGENCY: New Smyrna Beach PD  
LAST AGENCY INSPECTION: 08/20/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30119080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 09/14/2020 00:52:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:46 |
| Air Blank          | 0.000 | 00:46 |
| Control Test       | 0.081 | 00:47 |
| Air Blank          | 0.000 | 00:47 |
| Subject Sample # 1 | 0.122 | 00:47 |
| Breath Volume      | 1.742 | 00:47 |
| Air Blank          | 0.000 | 00:48 |
| Air Blank          | 0.000 | 00:50 |
| Subject Sample # 2 | 0.125 | 00:51 |
| Breath Volume      | 1.882 | 00:51 |
| Air Blank          | 0.000 | 00:51 |
| Control Test       | 0.081 | 00:52 |
| Air Blank          | 0.000 | 00:52 |
| Diagnostics Check  | OK    | 00:52 |

---

Data Download Date/Time: 09/20/2020 09:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: INGRAM,TRISTAN,N  
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD  
ARRESTING OFFICER: INGRAM,TRISTAN  
ARRESTING OFFICER AGENCY: New Smyrna Beach PD  
LAST AGENCY INSPECTION: 08/20/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: LABARRE

DATE/TIME OF TEST: 09/14/2020 22:23:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:16 |
| Air Blank          | 0.000 | 22:16 |
| Control Test       | 0.081 | 22:16 |
| Air Blank          | 0.000 | 22:17 |
| Subject Sample # 1 | 0.000 | 22:17 |
| Breath Volume      | 1.449 | 22:17 |
| Air Blank          | 0.000 | 22:18 |
| Air Blank          | 0.000 | 22:20 |
| Subject Sample # 2 | 0.000 | 22:21 |
| Breath Volume      | 1.300 | 22:21 |
| Air Blank          | 0.000 | 22:22 |
| Control Test       | 0.082 | 22:22 |
| Air Blank          | 0.000 | 22:23 |
| Diagnostics Check  | OK    | 22:23 |

---

Data Download Date/Time: 09/20/2020 09:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRYAN,STEVEN,A  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: PALMIER,PEDRO  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 09/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: TOOLE

DATE/TIME OF TEST: 09/14/2020 23:31:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:25 |
| Air Blank          | 0.000 | 23:25 |
| Control Test       | 0.080 | 23:25 |
| Air Blank          | 0.000 | 23:26 |
| Subject Sample # 1 | 0.127 | 23:27 |
| Breath Volume      | 2.464 | 23:27 |
| Air Blank          | 0.000 | 23:27 |
| Air Blank          | 0.000 | 23:29 |
| Subject Sample # 2 | 0.126 | 23:30 |
| Breath Volume      | 1.945 | 23:30 |
| Air Blank          | 0.000 | 23:31 |
| Control Test       | 0.079 | 23:31 |
| Air Blank          | 0.000 | 23:31 |
| Diagnostics Check  | OK    | 23:31 |

---

Data Download Date/Time: 10/27/2020 16:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRYAN,STEVEN,A  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: PALMIER,PEDRO  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 09/14/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: MEINCKE

DATE/TIME OF TEST: 09/24/2020 23:08:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:02 |
| Air Blank          | 0.000 | 23:02 |
| Control Test       | 0.080 | 23:03 |
| Air Blank          | 0.000 | 23:03 |
| Subject Sample # 1 | 0.144 | 23:04 |
| Breath Volume      | 1.472 | 23:04 |
| Air Blank          | 0.000 | 23:05 |
| Air Blank          | 0.000 | 23:06 |
| Subject Sample # 2 | 0.140 | 23:07 |
| Breath Volume      | 2.414 | 23:07 |
| Air Blank          | 0.000 | 23:08 |
| Control Test       | 0.080 | 23:08 |
| Air Blank          | 0.000 | 23:08 |
| Diagnostics Check  | OK    | 23:08 |

---

Data Download Date/Time: 10/27/2020 16:41



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARNER, MICHAEL, D  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: OGLE, SHAWN  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 09/29/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/29/2020 21:22:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:16 |
| Air Blank          | 0.000 | 21:16 |
| Control Test       | 0.082 | 21:17 |
| Air Blank          | 0.000 | 21:17 |
| Subject Sample # 1 | 0.007 | 21:18 |
| Breath Volume      | 2.773 | 21:18 |
| Air Blank          | 0.000 | 21:18 |
| Air Blank          | 0.000 | 21:20 |
| Subject Sample # 2 | 0.000 | 21:21 |
| Breath Volume      | 2.351 | 21:21 |
| Air Blank          | 0.000 | 21:21 |
| Control Test       | 0.082 | 21:22 |
| Air Blank          | 0.000 | 21:22 |
| Diagnostics Check  | OK    | 21:22 |

---

Data Download Date/Time: 10/26/2020 15:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIBBE,BRIAN,J  
BREATH TEST OPERATOR AGENCY: Orange City PD  
ARRESTING OFFICER: PAUL,JAYSON  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A3  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 08:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:28

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 09/13/2020 09:02:03

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 08:55 |
| Air Blank         |     | 0.000 | 08:56 |
| Control Test      |     | 0.080 | 08:56 |
| Air Blank         |     | 0.000 | 08:57 |
| Subject Sample    | # 1 | 0.165 | 08:57 |
| Breath Volume     |     | 2.792 | 08:57 |
| Air Blank         |     | 0.000 | 08:58 |
| Air Blank         |     | 0.000 | 09:00 |
| Subject Sample    | # 2 | 0.169 | 09:00 |
| Breath Volume     |     | 1.960 | 09:00 |
| Air Blank         |     | 0.000 | 09:01 |
| Control Test      |     | 0.080 | 09:01 |
| Air Blank         |     | 0.000 | 09:01 |
| Diagnostics Check |     | OK    | 09:02 |

---

Data Download Date/Time: 10/05/2020 14:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BELL,ETHAN,C

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: KRUSE,KRIS

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 08/22/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1149031

DRY GAS EXPIRY DATE: 05/30/2022

VIOLATION CODE: DUI

ARREST TIME: 05:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: BOWERS

DATE/TIME OF TEST: 09/01/2020 06:13:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:07 |
| Air Blank          | 0.000 | 06:07 |
| Control Test       | 0.079 | 06:07 |
| Air Blank          | 0.000 | 06:08 |
| Subject Sample # 1 | 0.000 | 06:08 |
| Breath Volume      | 1.164 | 06:08 |
| Air Blank          | 0.000 | 06:09 |
| Air Blank          | 0.000 | 06:11 |
| Subject Sample # 2 | 0.000 | 06:11 |
| Breath Volume      | 1.761 | 06:11 |
| Air Blank          | 0.000 | 06:12 |
| Control Test       | 0.078 | 06:12 |
| Air Blank          | 0.000 | 06:13 |
| Diagnostics Check  | OK    | 06:13 |

---

Data Download Date/Time: 09/19/2020 15:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DUBI,UZIEL,Y

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: GUY,KEVIN

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 08/22/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1149031

DRY GAS EXPIRY DATE: 05/30/2022

VIOLATION CODE: DUI

ARREST TIME: 21:53

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: CHISHOLM

DATE/TIME OF TEST: 09/16/2020 22:48:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:38 |
| Air Blank          | 0.000 | 22:39 |
| Control Test       | 0.081 | 22:39 |
| Air Blank          | 0.000 | 22:39 |
| Subject Sample # 1 | 0.386 | 22:40 |
| Breath Volume      | 3.515 | 22:40 |
| Air Blank          | 0.000 | 22:41 |
| Air Blank          | 0.000 | 22:42 |
| Subject Sample # 2 | 0.359 | 22:43 |
| Breath Volume      | 3.339 | 22:43 |
| Air Blank          | 0.000 | 22:44 |
| Air Blank          | 0.000 | 22:45 |
| Subject Sample # 3 | 0.367 | 22:46 |
| Breath Volume      | 1.964 | 22:46 |
| Air Blank          | 0.000 | 22:47 |
| Control Test       | 0.079 | 22:47 |
| Air Blank          | 0.000 | 22:47 |
| Diagnostics Check  | OK    | 22:48 |

---

Data Download Date/Time: 09/19/2020 15:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KRUSE, KRISTOPHER, D  
BREATH TEST OPERATOR AGENCY: Clermont PD  
ARRESTING OFFICER: COONS, SEAN  
ARRESTING OFFICER AGENCY: Clermont PD  
LAST AGENCY INSPECTION: 08/22/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1149031  
DRY GAS EXPIRY DATE: 05/30/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: CLAY

DATE/TIME OF TEST: 09/19/2020 03:46:30

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 03:37 |
| Air Blank         |     | 0.000 | 03:38 |
| Control Test      |     | 0.081 | 03:38 |
| Air Blank         |     | 0.000 | 03:38 |
| Subject Sample    | # 1 | 0.229 | 03:41 |
| Breath Volume     |     | 2.406 | 03:41 |
| Air Blank         |     | 0.000 | 03:42 |
| Air Blank         |     | 0.000 | 03:44 |
| Subject Sample    | # 2 | 0.233 | 03:44 |
| Breath Volume     |     | 2.371 | 03:44 |
| Air Blank         |     | 0.000 | 03:45 |
| Control Test      |     | 0.080 | 03:45 |
| Air Blank         |     | 0.000 | 03:46 |
| Diagnostics Check |     | OK    | 03:46 |

---

Data Download Date/Time: 09/19/2020 15:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARPER,J,

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: REBANDO,B

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 09/19/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1149031

DRY GAS EXPIRY DATE: 05/30/2022

VIOLATION CODE: DUI

ARREST TIME: 00:44

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: HIGH

DATE/TIME OF TEST: 09/21/2020 01:13:52

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:11 |                      |
| Air Blank         | 0.000 | 01:11 |                      |
| Control Test      | 0.082 | 01:11 |                      |
| Air Blank         | 0.000 | 01:12 |                      |
| Subject Sample    | # 1   | 01:12 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:12 |                      |
| Air Blank         | 0.000 | 01:12 |                      |
| Control Test      | 0.081 | 01:13 |                      |
| Air Blank         | 0.000 | 01:13 |                      |
| Diagnostics Check | OK    | 01:13 |                      |

---

Data Download Date/Time: 10/21/2020 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KRUSE, KRISTOPHER, D  
BREATH TEST OPERATOR AGENCY: Clermont PD  
ARRESTING OFFICER: PEREZ, CALEB  
ARRESTING OFFICER AGENCY: Clermont PD  
LAST AGENCY INSPECTION: 09/19/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1149031  
DRY GAS EXPIRY DATE: 05/30/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: KAISER

DATE/TIME OF TEST: 09/25/2020 01:26:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:17 |
| Air Blank          | 0.000 | 01:17 |
| Control Test       | 0.082 | 01:18 |
| Air Blank          | 0.000 | 01:18 |
| Subject Sample # 1 | 0.026 | 01:21 |
| Breath Volume      | 1.558 | 01:21 |
| Air Blank          | 0.000 | 01:21 |
| Air Blank          | 0.000 | 01:23 |
| Subject Sample # 2 | 0.024 | 01:25 |
| Breath Volume      | 1.402 | 01:25 |
| Air Blank          | 0.000 | 01:25 |
| Control Test       | 0.081 | 01:26 |
| Air Blank          | 0.000 | 01:26 |
| Diagnostics Check  | OK    | 01:26 |

---

Data Download Date/Time: 10/21/2020 12:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: LADY LAKE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRONK,A,  
BREATH TEST OPERATOR AGENCY: Lady Lake PD  
ARRESTING OFFICER: RICKETTS,J  
ARRESTING OFFICER AGENCY: Fruitland Park PD  
LAST AGENCY INSPECTION: 08/25/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17919080A1  
DRY GAS EXPIRY DATE: 08/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: LOLLAR

DATE/TIME OF TEST: 09/03/2020 22:51:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:45 |
| Air Blank          | 0.000 | 22:45 |
| Control Test       | 0.080 | 22:45 |
| Air Blank          | 0.000 | 22:46 |
| Subject Sample # 1 | 0.000 | 22:46 |
| Breath Volume      | 1.500 | 22:46 |
| Air Blank          | 0.000 | 22:47 |
| Air Blank          | 0.000 | 22:49 |
| Subject Sample # 2 | 0.000 | 22:49 |
| Breath Volume      | 1.585 | 22:49 |
| Air Blank          | 0.000 | 22:50 |
| Control Test       | 0.078 | 22:50 |
| Air Blank          | 0.000 | 22:50 |
| Diagnostics Check  | OK    | 22:51 |

---

Data Download Date/Time: 12/08/2020 12:18



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: POLIZZI,CHRISTINE  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: FORRESTER

DATE/TIME OF TEST: 09/03/2020 01:17:36

---

|                    |       |       |                      |
|--------------------|-------|-------|----------------------|
| Diagnostics Check  | OK    | 01:10 |                      |
| Air Blank          | 0.000 | 01:10 |                      |
| Control Test       | 0.082 | 01:11 |                      |
| Air Blank          | 0.000 | 01:11 |                      |
| Subject Sample # 1 | 0.069 | 01:12 |                      |
| Breath Volume      | 1.652 | 01:12 |                      |
| Air Blank          | 0.000 | 01:13 |                      |
| Air Blank          | 0.000 | 01:14 |                      |
| Subject Sample # 2 |       | 01:16 | Subject Test Refused |
| Breath Volume      | 0.000 | 01:16 |                      |
| Air Blank          | 0.000 | 01:16 |                      |
| Control Test       | 0.082 | 01:16 |                      |
| Air Blank          | 0.000 | 01:17 |                      |
| Diagnostics Check  | OK    | 01:17 |                      |

---

Data Download Date/Time: 09/16/2020 22:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: POLIZZI,CHRISTINE  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: GRANATO

DATE/TIME OF TEST: 09/03/2020 22:29:35

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:24 |                      |
| Air Blank         | 0.000 | 22:25 |                      |
| Control Test      | 0.082 | 22:25 |                      |
| Air Blank         | 0.000 | 22:26 |                      |
| Subject Sample    | # 1   | 22:28 | Subject Test Refused |
| Breath Volume     | 0.058 | 22:28 |                      |
| Air Blank         | 0.000 | 22:28 |                      |
| Control Test      | 0.082 | 22:28 |                      |
| Air Blank         | 0.000 | 22:29 |                      |
| Diagnostics Check | OK    | 22:29 |                      |

---

Data Download Date/Time: 09/16/2020 22:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRADY,REBECCA,A  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: BRADY,REBECCA  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CHAVEZ LOPEZ

DATE/TIME OF TEST: 09/06/2020 01:33:08

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 01:25 |
| Air Blank         |     | 0.000 | 01:25 |
| Control Test      |     | 0.083 | 01:26 |
| Air Blank         |     | 0.000 | 01:26 |
| Subject Sample    | # 1 | 0.271 | 01:28 |
| Breath Volume     |     | 2.960 | 01:28 |
| Air Blank         |     | 0.000 | 01:28 |
| Air Blank         |     | 0.000 | 01:30 |
| Subject Sample    | # 2 | 0.268 | 01:31 |
| Breath Volume     |     | 2.734 | 01:31 |
| Air Blank         |     | 0.000 | 01:32 |
| Control Test      |     | 0.080 | 01:32 |
| Air Blank         |     | 0.000 | 01:33 |
| Diagnostics Check |     | OK    | 01:33 |

---

Data Download Date/Time: 09/16/2020 22:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRADY,REBECCA,A  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: BRADY,REBECCA  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:33

SUBJECT'S LAST NAME: VELASQUEZ-JUAN

DATE/TIME OF TEST: 09/07/2020 18:07:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:01 |
| Air Blank          | 0.000 | 18:01 |
| Control Test       | 0.078 | 18:02 |
| Air Blank          | 0.000 | 18:02 |
| Subject Sample # 1 | 0.072 | 18:03 |
| Breath Volume      | 1.929 | 18:03 |
| Air Blank          | 0.000 | 18:03 |
| Air Blank          | 0.000 | 18:05 |
| Subject Sample # 2 | 0.068 | 18:06 |
| Breath Volume      | 2.179 | 18:06 |
| Air Blank          | 0.000 | 18:06 |
| Control Test       | 0.078 | 18:06 |
| Air Blank          | 0.000 | 18:07 |
| Diagnostics Check  | OK    | 18:07 |

---

Data Download Date/Time: 09/16/2020 22:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: POLIZZI,CHRISTINE  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: HROBAK

DATE/TIME OF TEST: 09/08/2020 19:10:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:03 |
| Air Blank          | 0.000 | 19:04 |
| Control Test       | 0.079 | 19:04 |
| Air Blank          | 0.000 | 19:05 |
| Subject Sample # 1 | 0.000 | 19:05 |
| Breath Volume      | 2.875 | 19:05 |
| Air Blank          | 0.000 | 19:06 |
| Air Blank          | 0.000 | 19:08 |
| Subject Sample # 2 | 0.000 | 19:08 |
| Breath Volume      | 2.992 | 19:08 |
| Air Blank          | 0.000 | 19:09 |
| Control Test       | 0.082 | 19:09 |
| Air Blank          | 0.000 | 19:10 |
| Diagnostics Check  | OK    | 19:10 |

---

Data Download Date/Time: 09/16/2020 22:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P  
BREATH TEST OPERATOR AGENCY: Stuart PD  
ARRESTING OFFICER: ANGERVIL,CLAUDY  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/11/2020 00:38:02

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 00:26 |                    |
| Air Blank         | 0.000 | 00:27 |                    |
| Control Test      | 0.082 | 00:27 |                    |
| Air Blank         | 0.000 | 00:27 |                    |
| Subject Sample    | # 1   | 00:31 | No Sample Provided |
| Breath Volume     | 0.000 | 00:31 |                    |
| Air Blank         | 0.000 | 00:31 |                    |
| Air Blank         | 0.000 | 00:33 |                    |
| Subject Sample    | # 2   | 00:36 | No Sample Provided |
| Breath Volume     | 0.000 | 00:36 |                    |
| Air Blank         | 0.000 | 00:37 |                    |
| Control Test      | 0.082 | 00:37 |                    |
| Air Blank         | 0.000 | 00:37 |                    |
| Diagnostics Check | OK    | 00:38 |                    |

---

Data Download Date/Time: 09/16/2020 22:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: POLIZZI,CHRISTINE  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:29

SUBJECT'S LAST NAME: RIKER

DATE/TIME OF TEST: 09/11/2020 19:18:18

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 19:06 |                |
| Air Blank         | 0.000     | 19:06 |                |
| Control Test      | 0.082     | 19:07 |                |
| Air Blank         | 0.000     | 19:07 |                |
| Subject Sample    | # 1 0.219 | 19:08 |                |
| Breath Volume     | 1.378     | 19:08 |                |
| Air Blank         | 0.000     | 19:08 |                |
| Air Blank         | 0.000     | 19:10 |                |
| Subject Sample    | # 2 0.174 | 19:13 | Volume Not Met |
| Breath Volume     | 0.476     | 19:13 |                |
| Air Blank         | 0.000     | 19:14 |                |
| Air Blank         | 0.000     | 19:16 |                |
| Subject Sample    | # 3 0.208 | 19:16 |                |
| Breath Volume     | 1.738     | 19:16 |                |
| Air Blank         | 0.000     | 19:17 |                |
| Control Test      | 0.080     | 19:17 |                |
| Air Blank         | 0.000     | 19:18 |                |
| Diagnostics Check | OK        | 19:18 |                |

---

Data Download Date/Time: 09/16/2020 22:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: POLIZZI,CHRISTINE  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: BURNSED III

DATE/TIME OF TEST: 09/12/2020 03:20:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:14 |
| Air Blank          | 0.000 | 03:15 |
| Control Test       | 0.082 | 03:15 |
| Air Blank          | 0.000 | 03:15 |
| Subject Sample # 1 | 0.113 | 03:16 |
| Breath Volume      | 2.792 | 03:16 |
| Air Blank          | 0.000 | 03:17 |
| Air Blank          | 0.000 | 03:18 |
| Subject Sample # 2 | 0.115 | 03:19 |
| Breath Volume      | 2.515 | 03:19 |
| Air Blank          | 0.000 | 03:19 |
| Control Test       | 0.081 | 03:20 |
| Air Blank          | 0.000 | 03:20 |
| Diagnostics Check  | OK    | 03:20 |

---

Data Download Date/Time: 09/16/2020 22:49



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: POLIZZI,CHRISTINE  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: GRISSOM

DATE/TIME OF TEST: 09/13/2020 00:19:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:12 |
| Air Blank          | 0.000 | 00:12 |
| Control Test       | 0.082 | 00:12 |
| Air Blank          | 0.000 | 00:13 |
| Subject Sample # 1 | 0.193 | 00:14 |
| Breath Volume      | 1.730 | 00:14 |
| Air Blank          | 0.000 | 00:14 |
| Air Blank          | 0.000 | 00:16 |
| Subject Sample # 2 | 0.188 | 00:17 |
| Breath Volume      | 2.425 | 00:17 |
| Air Blank          | 0.000 | 00:18 |
| Control Test       | 0.081 | 00:18 |
| Air Blank          | 0.000 | 00:18 |
| Diagnostics Check  | OK    | 00:19 |

---

Data Download Date/Time: 09/16/2020 22:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAY,ERICK,J  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: DAY,ERICK  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:32

SUBJECT'S LAST NAME: OCASIO ROLON

DATE/TIME OF TEST: 09/13/2020 17:32:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:26 |
| Air Blank          | 0.000 | 17:27 |
| Control Test       | 0.082 | 17:27 |
| Air Blank          | 0.000 | 17:28 |
| Subject Sample # 1 | 0.215 | 17:28 |
| Breath Volume      | 2.218 | 17:28 |
| Air Blank          | 0.000 | 17:29 |
| Air Blank          | 0.000 | 17:30 |
| Subject Sample # 2 | 0.208 | 17:31 |
| Breath Volume      | 2.250 | 17:31 |
| Air Blank          | 0.000 | 17:31 |
| Control Test       | 0.080 | 17:32 |
| Air Blank          | 0.000 | 17:32 |
| Diagnostics Check  | OK    | 17:32 |

---

Data Download Date/Time: 09/16/2020 22:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKES,JOHN,D  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: WILKES,JOHN  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: ROSATO

DATE/TIME OF TEST: 09/13/2020 20:34:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:27 |
| Air Blank          | 0.000 | 20:28 |
| Control Test       | 0.081 | 20:28 |
| Air Blank          | 0.000 | 20:29 |
| Subject Sample # 1 | 0.265 | 20:29 |
| Breath Volume      | 1.957 | 20:29 |
| Air Blank          | 0.000 | 20:30 |
| Air Blank          | 0.000 | 20:32 |
| Subject Sample # 2 | 0.261 | 20:33 |
| Breath Volume      | 2.402 | 20:33 |
| Air Blank          | 0.000 | 20:33 |
| Control Test       | 0.081 | 20:33 |
| Air Blank          | 0.000 | 20:34 |
| Diagnostics Check  | OK    | 20:34 |

---

Data Download Date/Time: 09/16/2020 22:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FREDERIC,JEAN,E  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: FREDERIC,JEAN  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: LEONE

DATE/TIME OF TEST: 09/14/2020 23:11:46

---

|                   |       |       |                  |
|-------------------|-------|-------|------------------|
| Diagnostics Check | OK    | 23:08 |                  |
| Air Blank         | 0.000 | 23:09 |                  |
| Control Test      | 0.082 | 23:09 |                  |
| Air Blank         | 0.000 | 23:09 |                  |
| Subject Sample    | # 1   | 23:11 | Sequence Aborted |
| Breath Volume     | 0.000 | 23:11 |                  |
| Air Blank         | 0.000 | 23:11 |                  |

---

Data Download Date/Time: 09/16/2020 22:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FREDERIC,JEAN,  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: FREDERIC,JEAN  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: LEONE

DATE/TIME OF TEST: 09/14/2020 23:19:44

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:15 |                      |
| Air Blank         | 0.000 | 23:16 |                      |
| Control Test      | 0.082 | 23:16 |                      |
| Air Blank         | 0.000 | 23:17 |                      |
| Subject Sample    | # 1   | 23:18 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:18 |                      |
| Air Blank         | 0.000 | 23:18 |                      |
| Control Test      | 0.082 | 23:19 |                      |
| Air Blank         | 0.000 | 23:19 |                      |
| Diagnostics Check | OK    | 23:19 |                      |

---

Data Download Date/Time: 09/16/2020 22:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P  
BREATH TEST OPERATOR AGENCY: Stuart PD  
ARRESTING OFFICER: ANGERVIL,CLAUDY  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 09/15/2020 03:16:00

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:13 |                      |
| Air Blank         | 0.000 | 03:13 |                      |
| Control Test      | 0.081 | 03:13 |                      |
| Air Blank         | 0.000 | 03:14 |                      |
| Subject Sample    | # 1   | 03:14 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:14 |                      |
| Air Blank         | 0.000 | 03:15 |                      |
| Control Test      | 0.082 | 03:15 |                      |
| Air Blank         | 0.000 | 03:15 |                      |
| Diagnostics Check | OK    | 03:16 |                      |

---

Data Download Date/Time: 09/16/2020 22:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKES,JOHN,D  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: WHITE,CAMERON  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 09/16/2020 04:18:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:09 |
| Air Blank          | 0.000 | 04:10 |
| Control Test       | 0.082 | 04:10 |
| Air Blank          | 0.000 | 04:11 |
| Subject Sample # 1 | 0.179 | 04:13 |
| Breath Volume      | 1.367 | 04:13 |
| Air Blank          | 0.000 | 04:13 |
| Air Blank          | 0.000 | 04:15 |
| Subject Sample # 2 | 0.170 | 04:16 |
| Breath Volume      | 1.308 | 04:16 |
| Air Blank          | 0.000 | 04:17 |
| Control Test       | 0.080 | 04:17 |
| Air Blank          | 0.000 | 04:18 |
| Diagnostics Check  | OK    | 04:18 |

---

Data Download Date/Time: 09/16/2020 22:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: POLIZZI,CHRISTINE  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 09/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: BERNARDINI

DATE/TIME OF TEST: 09/18/2020 00:07:43

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 00:01 |
| Air Blank         |     | 0.000 | 00:01 |
| Control Test      |     | 0.082 | 00:01 |
| Air Blank         |     | 0.000 | 00:02 |
| Subject Sample    | # 1 | 0.000 | 00:03 |
| Breath Volume     |     | 2.687 | 00:03 |
| Air Blank         |     | 0.000 | 00:03 |
| Air Blank         |     | 0.000 | 00:05 |
| Subject Sample    | # 2 | 0.000 | 00:06 |
| Breath Volume     |     | 2.226 | 00:06 |
| Air Blank         |     | 0.000 | 00:06 |
| Control Test      |     | 0.082 | 00:07 |
| Air Blank         |     | 0.000 | 00:07 |
| Diagnostics Check |     | OK    | 00:07 |

---

Data Download Date/Time: 10/01/2020 02:05



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CERNUTO, MATTHEW, D  
BREATH TEST OPERATOR AGENCY: Stuart PD  
ARRESTING OFFICER: CERNUTO, MATTHEW  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 09/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 11:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:43

SUBJECT'S LAST NAME: HOWE

DATE/TIME OF TEST: 09/21/2020 12:14:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 12:08 |
| Air Blank          | 0.000 | 12:08 |
| Control Test       | 0.082 | 12:09 |
| Air Blank          | 0.000 | 12:09 |
| Subject Sample # 1 | 0.000 | 12:10 |
| Breath Volume      | 1.558 | 12:10 |
| Air Blank          | 0.000 | 12:10 |
| Air Blank          | 0.000 | 12:12 |
| Subject Sample # 2 | 0.000 | 12:12 |
| Breath Volume      | 2.328 | 12:12 |
| Air Blank          | 0.000 | 12:13 |
| Control Test       | 0.081 | 12:13 |
| Air Blank          | 0.000 | 12:14 |
| Diagnostics Check  | OK    | 12:14 |

---

Data Download Date/Time: 10/01/2020 02:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: THOMAS,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 09/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: PABLO

DATE/TIME OF TEST: 09/26/2020 02:23:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:16 |
| Air Blank          | 0.000 | 02:17 |
| Control Test       | 0.081 | 02:17 |
| Air Blank          | 0.000 | 02:17 |
| Subject Sample # 1 | 0.194 | 02:19 |
| Breath Volume      | 1.480 | 02:19 |
| Air Blank          | 0.000 | 02:19 |
| Air Blank          | 0.000 | 02:21 |
| Subject Sample # 2 | 0.193 | 02:22 |
| Breath Volume      | 1.402 | 02:22 |
| Air Blank          | 0.000 | 02:23 |
| Control Test       | 0.080 | 02:23 |
| Air Blank          | 0.000 | 02:23 |
| Diagnostics Check  | OK    | 02:23 |

---

Data Download Date/Time: 10/01/2020 02:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: PALMIERI,A  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 09/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: JUARBE

DATE/TIME OF TEST: 09/26/2020 03:52:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:45 |
| Air Blank          | 0.000 | 03:46 |
| Control Test       | 0.081 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Subject Sample # 1 | 0.089 | 03:47 |
| Breath Volume      | 4.136 | 03:47 |
| Air Blank          | 0.000 | 03:48 |
| Air Blank          | 0.000 | 03:50 |
| Subject Sample # 2 | 0.088 | 03:51 |
| Breath Volume      | 3.824 | 03:51 |
| Air Blank          | 0.000 | 03:51 |
| Control Test       | 0.081 | 03:51 |
| Air Blank          | 0.000 | 03:52 |
| Diagnostics Check  | OK    | 03:52 |

---

Data Download Date/Time: 10/01/2020 02:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKES,JOHN,D  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: GITLIN,CORY  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 09/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MEJIA MORALES

DATE/TIME OF TEST: 09/28/2020 00:12:03

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 00:04 |
| Air Blank         |     | 0.000 | 00:04 |
| Control Test      |     | 0.082 | 00:05 |
| Air Blank         |     | 0.000 | 00:05 |
| Subject Sample    | # 1 | 0.212 | 00:07 |
| Breath Volume     |     | 1.917 | 00:07 |
| Air Blank         |     | 0.000 | 00:07 |
| Air Blank         |     | 0.000 | 00:09 |
| Subject Sample    | # 2 | 0.209 | 00:10 |
| Breath Volume     |     | 1.972 | 00:10 |
| Air Blank         |     | 0.000 | 00:11 |
| Control Test      |     | 0.080 | 00:11 |
| Air Blank         |     | 0.000 | 00:11 |
| Diagnostics Check |     | OK    | 00:12 |

---

Data Download Date/Time: 10/01/2020 02:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUBOIS,BROOKS,E  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: DUBOIS,BROOKS  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 09/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:29

SUBJECT'S LAST NAME: FARGO

DATE/TIME OF TEST: 09/28/2020 15:58:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:50 |
| Air Blank          | 0.000 | 15:51 |
| Control Test       | 0.081 | 15:51 |
| Air Blank          | 0.000 | 15:52 |
| Subject Sample # 1 | 0.182 | 15:53 |
| Breath Volume      | 2.445 | 15:53 |
| Air Blank          | 0.000 | 15:54 |
| Air Blank          | 0.000 | 15:55 |
| Subject Sample # 2 | 0.179 | 15:56 |
| Breath Volume      | 3.050 | 15:56 |
| Air Blank          | 0.000 | 15:57 |
| Control Test       | 0.081 | 15:57 |
| Air Blank          | 0.000 | 15:58 |
| Diagnostics Check  | OK    | 15:58 |

---

Data Download Date/Time: 10/01/2020 02:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P  
BREATH TEST OPERATOR AGENCY: Stuart PD  
ARRESTING OFFICER: ANGERVIL,CLAUDY  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 09/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: FORTNER

DATE/TIME OF TEST: 09/29/2020 00:16:59

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 00:10 |
| Air Blank         |     | 0.000 | 00:11 |
| Control Test      |     | 0.082 | 00:11 |
| Air Blank         |     | 0.000 | 00:12 |
| Subject Sample    | # 1 | 0.311 | 00:12 |
| Breath Volume     |     | 2.480 | 00:12 |
| Air Blank         |     | 0.000 | 00:13 |
| Air Blank         |     | 0.000 | 00:14 |
| Subject Sample    | # 2 | 0.322 | 00:15 |
| Breath Volume     |     | 2.703 | 00:15 |
| Air Blank         |     | 0.000 | 00:16 |
| Control Test      |     | 0.080 | 00:16 |
| Air Blank         |     | 0.000 | 00:16 |
| Diagnostics Check |     | OK    | 00:16 |

---

Data Download Date/Time: 10/01/2020 02:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CERNUTO, MATTHEW, D  
BREATH TEST OPERATOR AGENCY: Stuart PD  
ARRESTING OFFICER: CERNUTO, MATTHEW  
ARRESTING OFFICER AGENCY: Stuart PD  
LAST AGENCY INSPECTION: 09/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:02

SUBJECT'S LAST NAME: SIKORA

DATE/TIME OF TEST: 09/30/2020 16:37:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:25 |
| Air Blank          | 0.000 | 16:26 |
| Control Test       | 0.081 | 16:26 |
| Air Blank          | 0.000 | 16:26 |
| Subject Sample # 1 | 0.294 | 16:29 |
| Breath Volume      | 1.589 | 16:29 |
| Air Blank          | 0.000 | 16:30 |
| Air Blank          | 0.000 | 16:31 |
| Subject Sample # 2 | 0.270 | 16:32 |
| Breath Volume      | 1.238 | 16:32 |
| Air Blank          | 0.000 | 16:33 |
| Air Blank          | 0.000 | 16:35 |
| Subject Sample # 3 | 0.279 | 16:36 |
| Breath Volume      | 1.269 | 16:36 |
| Air Blank          | 0.000 | 16:36 |
| Control Test       | 0.080 | 16:36 |
| Air Blank          | 0.000 | 16:37 |
| Diagnostics Check  | OK    | 16:37 |

---

Data Download Date/Time: 10/01/2020 02:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: POLIZZI,CHRISTINE  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 09/16/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03519080A5  
DRY GAS EXPIRY DATE: 04/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 09/30/2020 23:24:15

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:18 |                      |
| Air Blank         | 0.000 | 23:18 |                      |
| Control Test      | 0.082 | 23:19 |                      |
| Air Blank         | 0.000 | 23:19 |                      |
| Subject Sample    | # 1   | 23:22 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:22 |                      |
| Air Blank         | 0.000 | 23:23 |                      |
| Control Test      | 0.082 | 23:23 |                      |
| Air Blank         | 0.000 | 23:24 |                      |
| Diagnostics Check | OK    | 23:24 |                      |

---

Data Download Date/Time: 10/01/2020 02:06



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: WEBER,JEFFREY  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: BASKIN

DATE/TIME OF TEST: 09/05/2020 03:32:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:23 |
| Air Blank          | 0.000 | 03:24 |
| Control Test       | 0.079 | 03:24 |
| Air Blank          | 0.000 | 03:25 |
| Subject Sample # 1 | 0.116 | 03:26 |
| Breath Volume      | 1.148 | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Air Blank          | 0.000 | 03:29 |
| Subject Sample # 2 | 0.115 | 03:30 |
| Breath Volume      | 2.265 | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Control Test       | 0.079 | 03:31 |
| Air Blank          | 0.000 | 03:31 |
| Diagnostics Check  | OK    | 03:32 |

---

Data Download Date/Time: 10/01/2020 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: WEBER,JEFFREY  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 09/06/2020 04:07:53

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:05 |                      |
| Air Blank         | 0.000 | 04:05 |                      |
| Control Test      | 0.080 | 04:05 |                      |
| Air Blank         | 0.000 | 04:06 |                      |
| Subject Sample    | # 1   | 04:06 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:06 |                      |
| Air Blank         | 0.000 | 04:06 |                      |
| Control Test      | 0.080 | 04:07 |                      |
| Air Blank         | 0.000 | 04:07 |                      |
| Diagnostics Check | OK    | 04:07 |                      |

---

Data Download Date/Time: 10/01/2020 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: WEBER,JEFFREY  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: TERRY

DATE/TIME OF TEST: 09/11/2020 02:35:45

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:32 |                      |
| Air Blank         | 0.000 | 02:33 |                      |
| Control Test      | 0.080 | 02:33 |                      |
| Air Blank         | 0.000 | 02:34 |                      |
| Subject Sample    | # 1   | 02:34 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:34 |                      |
| Air Blank         | 0.000 | 02:34 |                      |
| Control Test      | 0.081 | 02:35 |                      |
| Air Blank         | 0.000 | 02:35 |                      |
| Diagnostics Check | OK    | 02:35 |                      |

---

Data Download Date/Time: 10/01/2020 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: WEBER,JEFFREY  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 09/14/2020 22:30:26

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:27 |                      |
| Air Blank         | 0.000 | 22:27 |                      |
| Control Test      | 0.080 | 22:27 |                      |
| Air Blank         | 0.000 | 22:28 |                      |
| Subject Sample    | # 1   | 22:29 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:29 |                      |
| Air Blank         | 0.000 | 22:29 |                      |
| Control Test      | 0.081 | 22:29 |                      |
| Air Blank         | 0.000 | 22:30 |                      |
| Diagnostics Check | OK    | 22:30 |                      |

---

Data Download Date/Time: 10/01/2020 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: WEBER,JEFFREY  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 09/20/2020 01:46:10

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:43 |                      |
| Air Blank         | 0.000 | 01:43 |                      |
| Control Test      | 0.080 | 01:44 |                      |
| Air Blank         | 0.000 | 01:44 |                      |
| Subject Sample    | # 1   | 01:44 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:44 |                      |
| Air Blank         | 0.000 | 01:45 |                      |
| Control Test      | 0.080 | 01:45 |                      |
| Air Blank         | 0.000 | 01:46 |                      |
| Diagnostics Check | OK    | 01:46 |                      |

---

Data Download Date/Time: 10/01/2020 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEBER,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: WEBER,JEFFREY  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: CONNELL

DATE/TIME OF TEST: 09/25/2020 01:23:43

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 01:16 |
| Air Blank         |     | 0.000 | 01:16 |
| Control Test      |     | 0.081 | 01:17 |
| Air Blank         |     | 0.000 | 01:17 |
| Subject Sample    | # 1 | 0.283 | 01:18 |
| Breath Volume     |     | 3.640 | 01:18 |
| Air Blank         |     | 0.000 | 01:19 |
| Air Blank         |     | 0.000 | 01:21 |
| Subject Sample    | # 2 | 0.274 | 01:22 |
| Breath Volume     |     | 2.046 | 01:22 |
| Air Blank         |     | 0.000 | 01:22 |
| Control Test      |     | 0.081 | 01:23 |
| Air Blank         |     | 0.000 | 01:23 |
| Diagnostics Check |     | OK    | 01:23 |

---

Data Download Date/Time: 10/01/2020 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHERI,DIANE,M  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: JOHNSON,ROBERT  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 906364  
DRY GAS EXPIRY DATE: 02/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: CANTILLA

DATE/TIME OF TEST: 09/27/2020 04:11:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:04 |
| Air Blank          | 0.000 | 04:05 |
| Control Test       | 0.079 | 04:05 |
| Air Blank          | 0.000 | 04:06 |
| Subject Sample # 1 | 0.235 | 04:06 |
| Breath Volume      | 1.562 | 04:06 |
| Air Blank          | 0.000 | 04:07 |
| Air Blank          | 0.000 | 04:09 |
| Subject Sample # 2 | 0.249 | 04:09 |
| Breath Volume      | 2.597 | 04:09 |
| Air Blank          | 0.000 | 04:10 |
| Control Test       | 0.077 | 04:10 |
| Air Blank          | 0.000 | 04:11 |
| Diagnostics Check  | OK    | 04:11 |

---

Data Download Date/Time: 10/01/2020 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOUSE,STEVEN,D  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: HOUSE,STEVEN  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1349789  
DRY GAS EXPIRY DATE: 07/15/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: VELASQUEZ-EUCEDA

DATE/TIME OF TEST: 09/11/2020 21:43:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:36 |
| Air Blank          | 0.000 | 21:36 |
| Control Test       | 0.081 | 21:37 |
| Air Blank          | 0.000 | 21:37 |
| Subject Sample # 1 | 0.179 | 21:38 |
| Breath Volume      | 1.972 | 21:38 |
| Air Blank          | 0.000 | 21:38 |
| Air Blank          | 0.000 | 21:40 |
| Subject Sample # 2 | 0.186 | 21:41 |
| Breath Volume      | 1.300 | 21:41 |
| Air Blank          | 0.000 | 21:42 |
| Control Test       | 0.080 | 21:42 |
| Air Blank          | 0.000 | 21:42 |
| Diagnostics Check  | OK    | 21:43 |

---

Data Download Date/Time: 09/30/2020 13:24



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAGUADOG,SHAUN,J  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: SMOLKIN,DMITRY  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1349789  
DRY GAS EXPIRY DATE: 07/15/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: SOWELL

DATE/TIME OF TEST: 09/16/2020 02:29:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:23 |
| Air Blank          | 0.000 | 02:23 |
| Control Test       | 0.081 | 02:24 |
| Air Blank          | 0.000 | 02:24 |
| Subject Sample # 1 | 0.000 | 02:25 |
| Breath Volume      | 2.765 | 02:25 |
| Air Blank          | 0.000 | 02:25 |
| Air Blank          | 0.000 | 02:27 |
| Subject Sample # 2 | 0.000 | 02:28 |
| Breath Volume      | 2.695 | 02:28 |
| Air Blank          | 0.000 | 02:28 |
| Control Test       | 0.080 | 02:29 |
| Air Blank          | 0.000 | 02:29 |
| Diagnostics Check  | OK    | 02:29 |

---

Data Download Date/Time: 09/30/2020 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOUSE,STEVEN,D  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: HUTCHINSON,JOSEPH  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1349789  
DRY GAS EXPIRY DATE: 07/15/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: HERNANDEZ-BAUTISTA

DATE/TIME OF TEST: 09/19/2020 21:25:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:18 |
| Air Blank          | 0.000 | 21:19 |
| Control Test       | 0.081 | 21:19 |
| Air Blank          | 0.000 | 21:20 |
| Subject Sample # 1 | 0.166 | 21:21 |
| Breath Volume      | 2.804 | 21:21 |
| Air Blank          | 0.000 | 21:21 |
| Air Blank          | 0.000 | 21:23 |
| Subject Sample # 2 | 0.168 | 21:24 |
| Breath Volume      | 2.976 | 21:24 |
| Air Blank          | 0.000 | 21:24 |
| Control Test       | 0.079 | 21:25 |
| Air Blank          | 0.000 | 21:25 |
| Diagnostics Check  | OK    | 21:25 |

---

Data Download Date/Time: 09/30/2020 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOUSE,STEVEN,D  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: HUTCHINSON,JOSEPH  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1349789  
DRY GAS EXPIRY DATE: 07/15/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: COLLIER

DATE/TIME OF TEST: 09/21/2020 00:07:22

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 00:01 |
| Air Blank         |     | 0.000 | 00:01 |
| Control Test      |     | 0.081 | 00:01 |
| Air Blank         |     | 0.000 | 00:02 |
| Subject Sample    | # 1 | 0.311 | 00:02 |
| Breath Volume     |     | 3.207 | 00:02 |
| Air Blank         |     | 0.000 | 00:03 |
| Air Blank         |     | 0.000 | 00:05 |
| Subject Sample    | # 2 | 0.292 | 00:05 |
| Breath Volume     |     | 2.269 | 00:05 |
| Air Blank         |     | 0.000 | 00:06 |
| Control Test      |     | 0.080 | 00:06 |
| Air Blank         |     | 0.000 | 00:07 |
| Diagnostics Check |     | OK    | 00:07 |

---

Data Download Date/Time: 09/30/2020 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOUSE,STEVEN,D  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: ZHU,HENRY  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1349789  
DRY GAS EXPIRY DATE: 07/15/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: TENNYSON

DATE/TIME OF TEST: 09/27/2020 01:54:40

---

|                   |     |       |       |                 |
|-------------------|-----|-------|-------|-----------------|
| Diagnostics Check |     | OK    | 01:42 |                 |
| Air Blank         |     | 0.000 | 01:43 |                 |
| Control Test      |     | 0.081 | 01:43 |                 |
| Air Blank         |     | 0.000 | 01:43 |                 |
| Subject Sample    | # 1 | 0.171 | 01:47 | Slope Not Level |
| Breath Volume     |     | 1.367 | 01:47 |                 |
| Air Blank         |     | 0.000 | 01:47 |                 |
| Air Blank         |     | 0.000 | 01:49 |                 |
| Subject Sample    | # 2 | 0.211 | 01:50 |                 |
| Breath Volume     |     | 1.714 | 01:50 |                 |
| Air Blank         |     | 0.000 | 01:50 |                 |
| Air Blank         |     | 0.000 | 01:52 |                 |
| Subject Sample    | # 3 | 0.195 | 01:53 |                 |
| Breath Volume     |     | 2.378 | 01:53 |                 |
| Air Blank         |     | 0.000 | 01:53 |                 |
| Control Test      |     | 0.080 | 01:54 |                 |
| Air Blank         |     | 0.000 | 01:54 |                 |
| Diagnostics Check |     | OK    | 01:54 |                 |

---

Data Download Date/Time: 09/30/2020 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOUSE,STEVEN,D  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: ZHU,HENRY  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1349789  
DRY GAS EXPIRY DATE: 07/15/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: TENNYSON

DATE/TIME OF TEST: 09/27/2020 02:11:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:05 |
| Air Blank          | 0.000 | 02:06 |
| Control Test       | 0.081 | 02:06 |
| Air Blank          | 0.000 | 02:06 |
| Subject Sample # 1 | 0.188 | 02:07 |
| Breath Volume      | 1.414 | 02:07 |
| Air Blank          | 0.000 | 02:07 |
| Air Blank          | 0.000 | 02:09 |
| Subject Sample # 2 | 0.192 | 02:10 |
| Breath Volume      | 1.429 | 02:10 |
| Air Blank          | 0.000 | 02:10 |
| Control Test       | 0.080 | 02:11 |
| Air Blank          | 0.000 | 02:11 |
| Diagnostics Check  | OK    | 02:11 |

---

Data Download Date/Time: 09/30/2020 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOUSE,STEVEN,D  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: HUTCHINSON,JOSEPH  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1349789  
DRY GAS EXPIRY DATE: 07/15/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: ROULHAC

DATE/TIME OF TEST: 09/27/2020 04:56:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:50 |
| Air Blank          | 0.000 | 04:51 |
| Control Test       | 0.081 | 04:51 |
| Air Blank          | 0.000 | 04:52 |
| Subject Sample # 1 | 0.086 | 04:52 |
| Breath Volume      | 2.843 | 04:52 |
| Air Blank          | 0.000 | 04:53 |
| Air Blank          | 0.000 | 04:55 |
| Subject Sample # 2 | 0.085 | 04:55 |
| Breath Volume      | 2.703 | 04:55 |
| Air Blank          | 0.000 | 04:56 |
| Control Test       | 0.081 | 04:56 |
| Air Blank          | 0.000 | 04:56 |
| Diagnostics Check  | OK    | 04:56 |

---

Data Download Date/Time: 09/30/2020 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOUSE,STEVEN,D  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: HUTCHINSON,JOSEPH  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1349789  
DRY GAS EXPIRY DATE: 07/15/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: CARBAJAL

DATE/TIME OF TEST: 09/28/2020 23:06:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:00 |
| Air Blank          | 0.000 | 23:00 |
| Control Test       | 0.081 | 23:00 |
| Air Blank          | 0.000 | 23:01 |
| Subject Sample # 1 | 0.146 | 23:01 |
| Breath Volume      | 2.386 | 23:01 |
| Air Blank          | 0.000 | 23:02 |
| Air Blank          | 0.000 | 23:04 |
| Subject Sample # 2 | 0.148 | 23:04 |
| Breath Volume      | 2.066 | 23:04 |
| Air Blank          | 0.000 | 23:05 |
| Control Test       | 0.080 | 23:05 |
| Air Blank          | 0.000 | 23:06 |
| Diagnostics Check  | OK    | 23:06 |

---

Data Download Date/Time: 09/30/2020 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOUSE,STEVEN,D  
BREATH TEST OPERATOR AGENCY: Panama City PD  
ARRESTING OFFICER: HUTCHINSON,JOSEPH  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1349789  
DRY GAS EXPIRY DATE: 07/15/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/30/2020 03:38:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:32 |
| Air Blank          | 0.000 | 03:32 |
| Control Test       | 0.081 | 03:33 |
| Air Blank          | 0.000 | 03:33 |
| Subject Sample # 1 | 0.000 | 03:34 |
| Breath Volume      | 2.867 | 03:34 |
| Air Blank          | 0.000 | 03:34 |
| Air Blank          | 0.000 | 03:36 |
| Subject Sample # 2 | 0.000 | 03:37 |
| Breath Volume      | 2.683 | 03:37 |
| Air Blank          | 0.000 | 03:37 |
| Control Test       | 0.079 | 03:37 |
| Air Blank          | 0.000 | 03:38 |
| Diagnostics Check  | OK    | 03:38 |

---

Data Download Date/Time: 09/30/2020 13:24



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARDO, ADALBERTO,  
BREATH TEST OPERATOR AGENCY: Clewiston PD  
ARRESTING OFFICER: SCELFO, ZACHARY  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 09/15/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 09/17/2020 02:21:36

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:18 |                      |
| Air Blank         | 0.000 | 02:19 |                      |
| Control Test      | 0.080 | 02:19 |                      |
| Air Blank         | 0.000 | 02:20 |                      |
| Subject Sample    | # 1   | 02:20 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:20 |                      |
| Air Blank         | 0.000 | 02:20 |                      |
| Control Test      | 0.081 | 02:21 |                      |
| Air Blank         | 0.000 | 02:21 |                      |
| Diagnostics Check | OK    | 02:21 |                      |

---

Data Download Date/Time: 10/15/2020 17:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVENDANO,JUAN,M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: VALCIN,JEFFREY  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: HERNANDEZ-ALBA

DATE/TIME OF TEST: 09/01/2020 02:05:13

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 01:58 |
| Air Blank         |     | 0.000 | 01:58 |
| Control Test      |     | 0.079 | 01:59 |
| Air Blank         |     | 0.000 | 01:59 |
| Subject Sample    | # 1 | 0.218 | 02:00 |
| Breath Volume     |     | 2.457 | 02:00 |
| Air Blank         |     | 0.000 | 02:01 |
| Air Blank         |     | 0.000 | 02:02 |
| Subject Sample    | # 2 | 0.227 | 02:03 |
| Breath Volume     |     | 2.324 | 02:03 |
| Air Blank         |     | 0.000 | 02:04 |
| Control Test      |     | 0.079 | 02:04 |
| Air Blank         |     | 0.000 | 02:05 |
| Diagnostics Check |     | OK    | 02:05 |

---

Data Download Date/Time: 09/23/2020 10:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVENDANO,JUAN,M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: RIDGEWAY,CAITLIN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: CHAJAL BENITO

DATE/TIME OF TEST: 09/02/2020 01:31:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:25 |
| Air Blank          | 0.000 | 01:26 |
| Control Test       | 0.080 | 01:26 |
| Air Blank          | 0.000 | 01:26 |
| Subject Sample # 1 | 0.201 | 01:27 |
| Breath Volume      | 3.269 | 01:27 |
| Air Blank          | 0.000 | 01:28 |
| Air Blank          | 0.000 | 01:29 |
| Subject Sample # 2 | 0.196 | 01:30 |
| Breath Volume      | 2.835 | 01:30 |
| Air Blank          | 0.000 | 01:30 |
| Control Test       | 0.079 | 01:31 |
| Air Blank          | 0.000 | 01:31 |
| Diagnostics Check  | OK    | 01:31 |

---

Data Download Date/Time: 09/23/2020 10:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINDHOLM,JOHN,T  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SOMMERS,E  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: GUILLEN

DATE/TIME OF TEST: 09/04/2020 02:50:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:44 |
| Air Blank          | 0.000 | 02:44 |
| Control Test       | 0.080 | 02:44 |
| Air Blank          | 0.000 | 02:45 |
| Subject Sample # 1 | 0.129 | 02:45 |
| Breath Volume      | 3.890 | 02:45 |
| Air Blank          | 0.000 | 02:46 |
| Air Blank          | 0.000 | 02:48 |
| Subject Sample # 2 | 0.142 | 02:48 |
| Breath Volume      | 3.242 | 02:48 |
| Air Blank          | 0.000 | 02:49 |
| Control Test       | 0.079 | 02:49 |
| Air Blank          | 0.000 | 02:49 |
| Diagnostics Check  | OK    | 02:50 |

---

Data Download Date/Time: 09/23/2020 10:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVENDANO,JUAN,M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: OTERO,TYLER  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: VELAZCO ALEJO

DATE/TIME OF TEST: 09/09/2020 20:36:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:30 |
| Air Blank          | 0.000 | 20:30 |
| Control Test       | 0.079 | 20:30 |
| Air Blank          | 0.000 | 20:31 |
| Subject Sample # 1 | 0.163 | 20:31 |
| Breath Volume      | 2.457 | 20:31 |
| Air Blank          | 0.000 | 20:32 |
| Air Blank          | 0.000 | 20:34 |
| Subject Sample # 2 | 0.162 | 20:34 |
| Breath Volume      | 2.648 | 20:34 |
| Air Blank          | 0.000 | 20:35 |
| Control Test       | 0.080 | 20:35 |
| Air Blank          | 0.000 | 20:36 |
| Diagnostics Check  | OK    | 20:36 |

---

Data Download Date/Time: 09/23/2020 10:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMYRNIOS,MANUEL,J  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: SMYRNIOS,MANUEL  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 09:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: KELCHNER

DATE/TIME OF TEST: 09/12/2020 11:47:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 11:40 |
| Air Blank          | 0.000 | 11:41 |
| Control Test       | 0.080 | 11:41 |
| Air Blank          | 0.000 | 11:41 |
| Subject Sample # 1 | 0.131 | 11:42 |
| Breath Volume      | 3.929 | 11:42 |
| Air Blank          | 0.000 | 11:43 |
| Air Blank          | 0.000 | 11:45 |
| Subject Sample # 2 | 0.128 | 11:45 |
| Breath Volume      | 4.035 | 11:45 |
| Air Blank          | 0.000 | 11:46 |
| Control Test       | 0.078 | 11:46 |
| Air Blank          | 0.000 | 11:47 |
| Diagnostics Check  | OK    | 11:47 |

---

Data Download Date/Time: 09/23/2020 10:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TOTU,ROBERT,M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SAMMUT,ALEXANDER  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 13:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:24

SUBJECT'S LAST NAME: MARTINEZ PEREZ

DATE/TIME OF TEST: 09/12/2020 14:55:46

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 14:49 |
| Air Blank         |     | 0.000 | 14:50 |
| Control Test      |     | 0.078 | 14:50 |
| Air Blank         |     | 0.000 | 14:51 |
| Subject Sample    | # 1 | 0.123 | 14:51 |
| Breath Volume     |     | 4.253 | 14:51 |
| Air Blank         |     | 0.000 | 14:52 |
| Air Blank         |     | 0.000 | 14:53 |
| Subject Sample    | # 2 | 0.119 | 14:54 |
| Breath Volume     |     | 2.175 | 14:54 |
| Air Blank         |     | 0.000 | 14:54 |
| Control Test      |     | 0.078 | 14:55 |
| Air Blank         |     | 0.000 | 14:55 |
| Diagnostics Check |     | OK    | 14:55 |

---

Data Download Date/Time: 09/23/2020 10:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMERON,DAVID,J  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: LUTZ,DARIAN  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: GARLAND

DATE/TIME OF TEST: 09/13/2020 05:31:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:25 |
| Air Blank          | 0.000 | 05:25 |
| Control Test       | 0.080 | 05:26 |
| Air Blank          | 0.000 | 05:26 |
| Subject Sample # 1 | 0.160 | 05:27 |
| Breath Volume      | 2.222 | 05:27 |
| Air Blank          | 0.000 | 05:27 |
| Air Blank          | 0.000 | 05:29 |
| Subject Sample # 2 | 0.153 | 05:30 |
| Breath Volume      | 3.074 | 05:30 |
| Air Blank          | 0.000 | 05:30 |
| Control Test       | 0.079 | 05:30 |
| Air Blank          | 0.000 | 05:31 |
| Diagnostics Check  | OK    | 05:31 |

---

Data Download Date/Time: 09/23/2020 10:05



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WITT,KELLY,  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: WITT,KELLY  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/17/2020 23:40:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:32 |
| Air Blank          | 0.000 | 23:33 |
| Control Test       | 0.080 | 23:33 |
| Air Blank          | 0.000 | 23:33 |
| Subject Sample # 1 | 0.150 | 23:35 |
| Breath Volume      | 1.691 | 23:35 |
| Air Blank          | 0.000 | 23:36 |
| Air Blank          | 0.000 | 23:38 |
| Subject Sample # 2 | 0.145 | 23:38 |
| Breath Volume      | 2.000 | 23:38 |
| Air Blank          | 0.000 | 23:39 |
| Control Test       | 0.079 | 23:39 |
| Air Blank          | 0.000 | 23:40 |
| Diagnostics Check  | OK    | 23:40 |

---

Data Download Date/Time: 09/23/2020 10:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVENDANO,JUAN,M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: OTERO,TYLER  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: QUINTERO SANCHEZ

DATE/TIME OF TEST: 09/20/2020 01:24:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:18 |
| Air Blank          | 0.000 | 01:18 |
| Control Test       | 0.081 | 01:19 |
| Air Blank          | 0.000 | 01:19 |
| Subject Sample # 1 | 0.092 | 01:20 |
| Breath Volume      | 4.527 | 01:20 |
| Air Blank          | 0.000 | 01:20 |
| Air Blank          | 0.000 | 01:22 |
| Subject Sample # 2 | 0.091 | 01:23 |
| Breath Volume      | 5.074 | 01:23 |
| Air Blank          | 0.000 | 01:23 |
| Control Test       | 0.080 | 01:24 |
| Air Blank          | 0.000 | 01:24 |
| Diagnostics Check  | OK    | 01:24 |

---

Data Download Date/Time: 09/23/2020 10:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEAK,KYLE,F  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: TAVERNA,DANIEL  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BOWERMAN

DATE/TIME OF TEST: 09/20/2020 01:47:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:41 |
| Air Blank          | 0.000 | 01:41 |
| Control Test       | 0.080 | 01:41 |
| Air Blank          | 0.000 | 01:42 |
| Subject Sample # 1 | 0.161 | 01:42 |
| Breath Volume      | 3.058 | 01:42 |
| Air Blank          | 0.000 | 01:43 |
| Air Blank          | 0.000 | 01:45 |
| Subject Sample # 2 | 0.155 | 01:45 |
| Breath Volume      | 2.003 | 01:45 |
| Air Blank          | 0.000 | 01:46 |
| Control Test       | 0.079 | 01:46 |
| Air Blank          | 0.000 | 01:47 |
| Diagnostics Check  | OK    | 01:47 |

---

Data Download Date/Time: 09/23/2020 10:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVENDANO,JUAN,M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LOETHEN,KEITH  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: PHUNG

DATE/TIME OF TEST: 09/20/2020 04:13:54

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 04:03 |                      |
| Air Blank         | 0.000     | 04:03 |                      |
| Control Test      | 0.081     | 04:04 |                      |
| Air Blank         | 0.000     | 04:04 |                      |
| Subject Sample    | # 1 0.000 | 04:07 | Volume Not Met       |
| Breath Volume     | 0.000     | 04:07 |                      |
| Air Blank         | 0.000     | 04:08 |                      |
| Air Blank         | 0.000     | 04:10 |                      |
| Subject Sample    | # 2       | 04:12 | Subject Test Refused |
| Breath Volume     | 0.105     | 04:12 |                      |
| Air Blank         | 0.000     | 04:13 |                      |
| Control Test      | 0.079     | 04:13 |                      |
| Air Blank         | 0.000     | 04:13 |                      |
| Diagnostics Check | OK        | 04:13 |                      |

---

Data Download Date/Time: 09/23/2020 10:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: WINCELOWICZ,LARRY  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 09/23/2020 22:46:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:40 |
| Air Blank          | 0.000 | 22:40 |
| Control Test       | 0.079 | 22:40 |
| Air Blank          | 0.000 | 22:41 |
| Subject Sample # 1 | 0.160 | 22:41 |
| Breath Volume      | 3.273 | 22:41 |
| Air Blank          | 0.000 | 22:42 |
| Air Blank          | 0.000 | 22:44 |
| Subject Sample # 2 | 0.157 | 22:45 |
| Breath Volume      | 2.054 | 22:45 |
| Air Blank          | 0.000 | 22:45 |
| Control Test       | 0.079 | 22:45 |
| Air Blank          | 0.000 | 22:46 |
| Diagnostics Check  | OK    | 22:46 |

---

Data Download Date/Time: 10/13/2020 15:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TISDALE,KATIE,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: ALAGO SONERA,J  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: OLSEN

DATE/TIME OF TEST: 09/24/2020 04:35:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:28 |
| Air Blank          | 0.000 | 04:29 |
| Control Test       | 0.080 | 04:29 |
| Air Blank          | 0.000 | 04:29 |
| Subject Sample # 1 | 0.000 | 04:30 |
| Breath Volume      | 2.785 | 04:30 |
| Air Blank          | 0.000 | 04:31 |
| Air Blank          | 0.000 | 04:33 |
| Subject Sample # 2 | 0.000 | 04:33 |
| Breath Volume      | 2.824 | 04:33 |
| Air Blank          | 0.000 | 04:34 |
| Control Test       | 0.081 | 04:34 |
| Air Blank          | 0.000 | 04:34 |
| Diagnostics Check  | OK    | 04:35 |

---

Data Download Date/Time: 10/13/2020 15:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TISDALE,KATIE,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: HARGIS,KYLE  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: PATTON

DATE/TIME OF TEST: 09/24/2020 19:46:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:40 |
| Air Blank          | 0.000 | 19:40 |
| Control Test       | 0.081 | 19:41 |
| Air Blank          | 0.000 | 19:41 |
| Subject Sample # 1 | 0.158 | 19:42 |
| Breath Volume      | 2.144 | 19:42 |
| Air Blank          | 0.000 | 19:42 |
| Air Blank          | 0.000 | 19:44 |
| Subject Sample # 2 | 0.166 | 19:44 |
| Breath Volume      | 1.804 | 19:44 |
| Air Blank          | 0.000 | 19:45 |
| Control Test       | 0.080 | 19:45 |
| Air Blank          | 0.000 | 19:46 |
| Diagnostics Check  | OK    | 19:46 |

---

Data Download Date/Time: 10/13/2020 15:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TISDALE,KATIE,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LA SPIA,GIORGIO  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: SEBASTIAN JOSE

DATE/TIME OF TEST: 09/24/2020 23:13:33

---

|                   |     |       |       |                      |
|-------------------|-----|-------|-------|----------------------|
| Diagnostics Check |     | OK    | 22:59 |                      |
| Air Blank         |     | 0.000 | 23:00 |                      |
| Control Test      |     | 0.080 | 23:00 |                      |
| Air Blank         |     | 0.000 | 23:01 |                      |
| Subject Sample    | # 1 | 0.000 | 23:04 | Volume Not Met       |
| Breath Volume     |     | 0.042 | 23:04 |                      |
| Air Blank         |     | 0.000 | 23:04 |                      |
| Air Blank         |     | 0.000 | 23:06 |                      |
| Subject Sample    | # 2 | 0.189 | 23:08 |                      |
| Breath Volume     |     | 2.460 | 23:08 |                      |
| Air Blank         |     | 0.000 | 23:09 |                      |
| Air Blank         |     | 0.000 | 23:11 |                      |
| Subject Sample    | # 3 |       | 23:12 | Subject Test Refused |
| Breath Volume     |     | 0.308 | 23:12 |                      |
| Air Blank         |     | 0.000 | 23:12 |                      |
| Control Test      |     | 0.078 | 23:12 |                      |
| Air Blank         |     | 0.000 | 23:13 |                      |
| Diagnostics Check |     | OK    | 23:13 |                      |

---

Data Download Date/Time: 10/13/2020 15:27



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD 18096,A,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BYRD 18120,B  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: LADE

DATE/TIME OF TEST: 09/27/2020 03:38:09

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 03:30 |                      |
| Air Blank         | 0.000     | 03:31 |                      |
| Control Test      | 0.080     | 03:31 |                      |
| Air Blank         | 0.000     | 03:31 |                      |
| Subject Sample    | # 1 0.200 | 03:32 |                      |
| Breath Volume     | 1.546     | 03:32 |                      |
| Air Blank         | 0.000     | 03:33 |                      |
| Air Blank         | 0.000     | 03:35 |                      |
| Subject Sample    | # 2       | 03:36 | Subject Test Refused |
| Breath Volume     | 0.000     | 03:36 |                      |
| Air Blank         | 0.000     | 03:37 |                      |
| Control Test      | 0.080     | 03:37 |                      |
| Air Blank         | 0.000     | 03:38 |                      |
| Diagnostics Check | OK        | 03:38 |                      |

---

Data Download Date/Time: 10/13/2020 15:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD 18096,A,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BYRD 18120,B  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: LADE

DATE/TIME OF TEST: 09/27/2020 04:09:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:03 |
| Air Blank          | 0.000 | 04:03 |
| Control Test       | 0.080 | 04:04 |
| Air Blank          | 0.000 | 04:04 |
| Subject Sample # 1 | 0.181 | 04:05 |
| Breath Volume      | 2.457 | 04:05 |
| Air Blank          | 0.000 | 04:05 |
| Air Blank          | 0.000 | 04:07 |
| Subject Sample # 2 | 0.170 | 04:08 |
| Breath Volume      | 1.437 | 04:08 |
| Air Blank          | 0.000 | 04:08 |
| Control Test       | 0.079 | 04:09 |
| Air Blank          | 0.000 | 04:09 |
| Diagnostics Check  | OK    | 04:09 |

---

Data Download Date/Time: 10/13/2020 15:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD 18096,A,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SCARLATO 1613,M  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 09/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 09/28/2020 04:08:44

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:02 |                      |
| Air Blank         | 0.000 | 04:03 |                      |
| Control Test      | 0.080 | 04:03 |                      |
| Air Blank         | 0.000 | 04:04 |                      |
| Subject Sample    | # 1   | 04:07 | Subject Test Refused |
| Breath Volume     | 0.609 | 04:07 |                      |
| Air Blank         | 0.000 | 04:07 |                      |
| Control Test      | 0.081 | 04:08 |                      |
| Air Blank         | 0.000 | 04:08 |                      |
| Diagnostics Check | OK    | 04:08 |                      |

---

Data Download Date/Time: 10/13/2020 15:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANFIELD 18096,A,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SCARLATO 1613,M  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 09/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 09/28/2020 05:04:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:57 |
| Air Blank          | 0.000 | 04:58 |
| Control Test       | 0.080 | 04:58 |
| Air Blank          | 0.000 | 04:58 |
| Subject Sample # 1 | 0.229 | 04:59 |
| Breath Volume      | 3.535 | 04:59 |
| Air Blank          | 0.000 | 05:00 |
| Air Blank          | 0.000 | 05:01 |
| Subject Sample # 2 | 0.236 | 05:02 |
| Breath Volume      | 2.371 | 05:02 |
| Air Blank          | 0.000 | 05:03 |
| Control Test       | 0.080 | 05:03 |
| Air Blank          | 0.000 | 05:04 |
| Diagnostics Check  | OK    | 05:04 |

---

Data Download Date/Time: 10/13/2020 15:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PINGRY,BRIAN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: VICIOSO,YEURIS  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 09/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: BROUGHTON

DATE/TIME OF TEST: 09/28/2020 17:40:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:32 |
| Air Blank          | 0.000 | 17:33 |
| Control Test       | 0.081 | 17:33 |
| Air Blank          | 0.000 | 17:34 |
| Subject Sample # 1 | 0.000 | 17:35 |
| Breath Volume      | 1.898 | 17:35 |
| Air Blank          | 0.000 | 17:36 |
| Air Blank          | 0.000 | 17:38 |
| Subject Sample # 2 | 0.000 | 17:38 |
| Breath Volume      | 1.871 | 17:38 |
| Air Blank          | 0.000 | 17:39 |
| Control Test       | 0.080 | 17:39 |
| Air Blank          | 0.000 | 17:40 |
| Diagnostics Check  | OK    | 17:40 |

---

Data Download Date/Time: 10/13/2020 15:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINTON, ADAM, H  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: WINTON, ADAM  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: IWANSKI

DATE/TIME OF TEST: 09/28/2020 18:20:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:13 |
| Air Blank          | 0.000 | 18:14 |
| Control Test       | 0.079 | 18:14 |
| Air Blank          | 0.000 | 18:14 |
| Subject Sample # 1 | 0.117 | 18:15 |
| Breath Volume      | 1.925 | 18:15 |
| Air Blank          | 0.000 | 18:16 |
| Air Blank          | 0.000 | 18:17 |
| Subject Sample # 2 | 0.114 | 18:18 |
| Breath Volume      | 1.828 | 18:18 |
| Air Blank          | 0.000 | 18:19 |
| Control Test       | 0.079 | 18:19 |
| Air Blank          | 0.000 | 18:20 |
| Diagnostics Check  | OK    | 18:20 |

---

Data Download Date/Time: 10/13/2020 15:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: URENA,ROSEMARY,A  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: LUNA,MICHAEL  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 09/23/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A3  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/30/2020 01:16:00

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 01:09 |
| Air Blank         |     | 0.000 | 01:09 |
| Control Test      |     | 0.080 | 01:10 |
| Air Blank         |     | 0.000 | 01:10 |
| Subject Sample    | # 1 | 0.192 | 01:11 |
| Breath Volume     |     | 2.304 | 01:11 |
| Air Blank         |     | 0.000 | 01:12 |
| Air Blank         |     | 0.000 | 01:14 |
| Subject Sample    | # 2 | 0.178 | 01:14 |
| Breath Volume     |     | 1.761 | 01:14 |
| Air Blank         |     | 0.000 | 01:15 |
| Control Test      |     | 0.079 | 01:15 |
| Air Blank         |     | 0.000 | 01:15 |
| Diagnostics Check |     | OK    | 01:16 |

---

Data Download Date/Time: 10/13/2020 15:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,DAVID,B  
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores  
ARRESTING OFFICER: MEDDERS,NORMAN  
ARRESTING OFFICER AGENCY: Daytona Beach Shores  
LAST AGENCY INSPECTION: 09/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1305520  
DRY GAS EXPIRY DATE: 04/01/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MONDELL

DATE/TIME OF TEST: 09/19/2020 04:01:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:52 |
| Air Blank          | 0.000 | 03:52 |
| Control Test       | 0.080 | 03:53 |
| Air Blank          | 0.000 | 03:53 |
| Subject Sample # 1 | 0.060 | 03:56 |
| Breath Volume      | 1.781 | 03:56 |
| Air Blank          | 0.000 | 03:57 |
| Air Blank          | 0.000 | 03:59 |
| Subject Sample # 2 | 0.063 | 04:00 |
| Breath Volume      | 2.453 | 04:00 |
| Air Blank          | 0.000 | 04:00 |
| Control Test       | 0.078 | 04:01 |
| Air Blank          | 0.000 | 04:01 |
| Diagnostics Check  | OK    | 04:01 |

---

Data Download Date/Time: 10/11/2020 22:48



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTELLANO,VINCENT,J  
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores  
ARRESTING OFFICER: CASTELLANO,VINCENT  
ARRESTING OFFICER AGENCY: Daytona Beach Shores  
LAST AGENCY INSPECTION: 09/07/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1305520  
DRY GAS EXPIRY DATE: 04/01/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: FORT

DATE/TIME OF TEST: 09/29/2020 03:51:49

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 03:37 |                |
| Air Blank         |     | 0.000 | 03:38 |                |
| Control Test      |     | 0.080 | 03:38 |                |
| Air Blank         |     | 0.000 | 03:39 |                |
| Subject Sample    | # 1 | 0.125 | 03:42 | Volume Not Met |
| Breath Volume     |     | 0.960 | 03:42 |                |
| Air Blank         |     | 0.000 | 03:42 |                |
| Air Blank         |     | 0.000 | 03:44 |                |
| Subject Sample    | # 2 | 0.124 | 03:45 |                |
| Breath Volume     |     | 1.417 | 03:45 |                |
| Air Blank         |     | 0.000 | 03:46 |                |
| Air Blank         |     | 0.000 | 03:48 |                |
| Subject Sample    | # 3 | 0.129 | 03:50 |                |
| Breath Volume     |     | 1.601 | 03:50 |                |
| Air Blank         |     | 0.000 | 03:50 |                |
| Control Test      |     | 0.079 | 03:51 |                |
| Air Blank         |     | 0.000 | 03:51 |                |
| Diagnostics Check |     | OK    | 03:51 |                |

---

Data Download Date/Time: 10/11/2020 22:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SLOAN,JULIAN,E  
BREATH TEST OPERATOR AGENCY: Edgewater PD  
ARRESTING OFFICER: HARRIS,LODRICK  
ARRESTING OFFICER AGENCY: New Smyrna Beach PD  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1217042  
DRY GAS EXPIRY DATE: 10/10/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 09/19/2020 22:05:33

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 21:54 |                    |
| Air Blank         | 0.000 | 21:54 |                    |
| Control Test      | 0.078 | 21:54 |                    |
| Air Blank         | 0.000 | 21:55 |                    |
| Subject Sample    | # 1   | 21:58 | No Sample Provided |
| Breath Volume     | 0.000 | 21:58 |                    |
| Air Blank         | 0.000 | 21:59 |                    |
| Air Blank         | 0.000 | 22:01 |                    |
| Subject Sample    | # 2   | 22:04 | No Sample Provided |
| Breath Volume     | 0.000 | 22:04 |                    |
| Air Blank         | 0.000 | 22:04 |                    |
| Control Test      | 0.078 | 22:04 |                    |
| Air Blank         | 0.000 | 22:05 |                    |
| Diagnostics Check | OK    | 22:05 |                    |

---

Data Download Date/Time: 10/30/2020 11:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRESENZI,G,  
BREATH TEST OPERATOR AGENCY: Edgewater PD  
ARRESTING OFFICER: CRESENZI,G  
ARRESTING OFFICER AGENCY: Edgewater PD  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1217042  
DRY GAS EXPIRY DATE: 10/10/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 09/26/2020 03:03:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:57 |
| Air Blank          | 0.000 | 02:57 |
| Control Test       | 0.079 | 02:58 |
| Air Blank          | 0.000 | 02:58 |
| Subject Sample # 1 | 0.150 | 02:59 |
| Breath Volume      | 3.851 | 02:59 |
| Air Blank          | 0.000 | 02:59 |
| Air Blank          | 0.000 | 03:01 |
| Subject Sample # 2 | 0.143 | 03:01 |
| Breath Volume      | 3.667 | 03:01 |
| Air Blank          | 0.000 | 03:02 |
| Control Test       | 0.078 | 03:02 |
| Air Blank          | 0.000 | 03:03 |
| Diagnostics Check  | OK    | 03:03 |

---

Data Download Date/Time: 10/30/2020 11:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FEGER,JOSHUA,D  
BREATH TEST OPERATOR AGENCY: Edgewater PD  
ARRESTING OFFICER: FEGER,JOSHUA  
ARRESTING OFFICER AGENCY: Edgewater PD  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1217042  
DRY GAS EXPIRY DATE: 10/10/2022  
VIOLATION CODE: DUI  
ARREST TIME: 09:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: WHALEY

DATE/TIME OF TEST: 09/28/2020 11:04:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:57 |
| Air Blank          | 0.000 | 10:58 |
| Control Test       | 0.079 | 10:58 |
| Air Blank          | 0.000 | 10:59 |
| Subject Sample # 1 | 0.000 | 10:59 |
| Breath Volume      | 3.347 | 10:59 |
| Air Blank          | 0.000 | 11:00 |
| Air Blank          | 0.000 | 11:02 |
| Subject Sample # 2 | 0.000 | 11:02 |
| Breath Volume      | 1.902 | 11:02 |
| Air Blank          | 0.000 | 11:03 |
| Control Test       | 0.078 | 11:03 |
| Air Blank          | 0.000 | 11:03 |
| Diagnostics Check  | OK    | 11:04 |

---

Data Download Date/Time: 10/30/2020 11:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SCHOEN,ELLIS  
ARRESTING OFFICER AGENCY: Winter Garden PD  
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: ROBY

DATE/TIME OF TEST: 09/01/2020 22:38:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:32 |
| Air Blank          | 0.000 | 22:32 |
| Control Test       | 0.082 | 22:32 |
| Air Blank          | 0.000 | 22:33 |
| Subject Sample # 1 | 0.000 | 22:33 |
| Breath Volume      | 2.710 | 22:33 |
| Air Blank          | 0.000 | 22:34 |
| Air Blank          | 0.000 | 22:36 |
| Subject Sample # 2 | 0.000 | 22:36 |
| Breath Volume      | 2.746 | 22:36 |
| Air Blank          | 0.000 | 22:37 |
| Control Test       | 0.081 | 22:37 |
| Air Blank          | 0.000 | 22:38 |
| Diagnostics Check  | OK    | 22:38 |

---

Data Download Date/Time: 09/08/2020 14:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CINTRON,KRISTIN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: SHARPE

DATE/TIME OF TEST: 09/05/2020 04:36:58

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 04:30 |
| Air Blank         |     | 0.000 | 04:30 |
| Control Test      |     | 0.081 | 04:31 |
| Air Blank         |     | 0.000 | 04:31 |
| Subject Sample    | # 1 | 0.132 | 04:32 |
| Breath Volume     |     | 2.859 | 04:32 |
| Air Blank         |     | 0.000 | 04:33 |
| Air Blank         |     | 0.000 | 04:35 |
| Subject Sample    | # 2 | 0.123 | 04:35 |
| Breath Volume     |     | 1.230 | 04:35 |
| Air Blank         |     | 0.000 | 04:36 |
| Control Test      |     | 0.081 | 04:36 |
| Air Blank         |     | 0.000 | 04:36 |
| Diagnostics Check |     | OK    | 04:36 |

---

Data Download Date/Time: 09/08/2020 14:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CAMPBELL,ROBERT  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BOSTIC

DATE/TIME OF TEST: 09/05/2020 23:09:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:03 |
| Air Blank          | 0.000 | 23:04 |
| Control Test       | 0.083 | 23:04 |
| Air Blank          | 0.000 | 23:05 |
| Subject Sample # 1 | 0.058 | 23:05 |
| Breath Volume      | 2.484 | 23:05 |
| Air Blank          | 0.000 | 23:06 |
| Air Blank          | 0.000 | 23:08 |
| Subject Sample # 2 | 0.054 | 23:08 |
| Breath Volume      | 1.902 | 23:08 |
| Air Blank          | 0.000 | 23:09 |
| Control Test       | 0.081 | 23:09 |
| Air Blank          | 0.000 | 23:09 |
| Diagnostics Check  | OK    | 23:09 |

---

Data Download Date/Time: 09/08/2020 14:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RODRIGUEZ,STEPHINE  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 09/06/2020 20:59:59

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 20:44 |                |
| Air Blank         |     | 0.000 | 20:45 |                |
| Control Test      |     | 0.080 | 20:45 |                |
| Air Blank         |     | 0.000 | 20:45 |                |
| Subject Sample    | # 1 | 0.229 | 20:46 |                |
| Breath Volume     |     | 1.511 | 20:46 |                |
| Air Blank         |     | 0.000 | 20:47 |                |
| Air Blank         |     | 0.000 | 20:49 |                |
| Subject Sample    | # 2 | 0.207 | 20:52 | Volume Not Met |
| Breath Volume     |     | 0.035 | 20:52 |                |
| Air Blank         |     | 0.000 | 20:53 |                |
| Air Blank         |     | 0.000 | 20:54 |                |
| Subject Sample    | # 3 | 0.200 | 20:58 | Volume Not Met |
| Breath Volume     |     | 0.242 | 20:58 |                |
| Air Blank         |     | 0.000 | 20:59 |                |
| Control Test      |     | 0.076 | 20:59 |                |
| Air Blank         |     | 0.000 | 20:59 |                |
| Diagnostics Check |     | OK    | 20:59 |                |

---

Data Download Date/Time: 09/08/2020 14:52



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RODRIGUEZ,STEPHINE  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 09/06/2020 21:03:46

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 21:02 |                           |
| Air Blank         | 0.000 | 21:02 |                           |
| Control Test      | 0.074 | 21:03 | Control Outside Tolerance |
| Air Blank         | 0.000 | 21:03 |                           |

---

Data Download Date/Time: 09/08/2020 14:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RODRIGUEZ,STEPHAINE  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 09/06/2020 21:12:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:05 |
| Air Blank          | 0.000 | 21:06 |
| Control Test       | 0.075 | 21:06 |
| Air Blank          | 0.000 | 21:07 |
| Subject Sample # 1 | 0.186 | 21:07 |
| Breath Volume      | 1.136 | 21:07 |
| Air Blank          | 0.000 | 21:08 |
| Air Blank          | 0.000 | 21:10 |
| Subject Sample # 2 | 0.198 | 21:11 |
| Breath Volume      | 1.578 | 21:11 |
| Air Blank          | 0.000 | 21:11 |
| Control Test       | 0.075 | 21:12 |
| Air Blank          | 0.000 | 21:12 |
| Diagnostics Check  | OK    | 21:12 |

---

Data Download Date/Time: 09/08/2020 14:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WILK,JULIE  
ARRESTING OFFICER AGENCY: Central FL Univ PD  
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: DEMAI0

DATE/TIME OF TEST: 09/02/2020 06:15:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:06 |
| Air Blank          | 0.000 | 06:07 |
| Control Test       | 0.080 | 06:07 |
| Air Blank          | 0.000 | 06:08 |
| Subject Sample # 1 | 0.216 | 06:08 |
| Breath Volume      | 1.253 | 06:08 |
| Air Blank          | 0.000 | 06:09 |
| Air Blank          | 0.000 | 06:11 |
| Subject Sample # 2 | 0.198 | 06:14 |
| Breath Volume      | 3.507 | 06:14 |
| Air Blank          | 0.000 | 06:14 |
| Control Test       | 0.080 | 06:14 |
| Air Blank          | 0.000 | 06:15 |
| Diagnostics Check  | OK    | 06:15 |

---

Data Download Date/Time: 09/08/2020 15:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DEL CASTILLO,BRET  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: BARRE

DATE/TIME OF TEST: 09/03/2020 03:59:44

---

|                   |     |       |       |                   |
|-------------------|-----|-------|-------|-------------------|
| Diagnostics Check |     | OK    | 03:46 |                   |
| Air Blank         |     | 0.000 | 03:47 |                   |
| Control Test      |     | 0.078 | 03:47 |                   |
| Air Blank         |     | 0.000 | 03:48 |                   |
| Subject Sample    | # 1 | 0.222 | 03:51 | Slope Not Level   |
| Breath Volume     |     | 1.250 | 03:51 |                   |
| Air Blank         |     | 0.000 | 03:51 |                   |
| Air Blank         |     | 0.000 | 03:53 |                   |
| Subject Sample    | # 2 | 0.248 | 03:54 | No .020 agreement |
| Breath Volume     |     | 2.230 | 03:54 |                   |
| Air Blank         |     | 0.000 | 03:55 |                   |
| Air Blank         |     | 0.000 | 03:57 |                   |
| Subject Sample    | # 3 | 0.227 | 03:58 | No .020 agreement |
| Breath Volume     |     | 1.105 | 03:58 |                   |
| Air Blank         |     | 0.000 | 03:58 |                   |
| Control Test      |     | 0.078 | 03:59 |                   |
| Air Blank         |     | 0.000 | 03:59 |                   |
| Diagnostics Check |     | OK    | 03:59 |                   |

---

Data Download Date/Time: 09/08/2020 15:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DEL CASTILLO,BRET  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: BARRE

DATE/TIME OF TEST: 09/03/2020 04:10:45

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 04:04 |
| Air Blank         |     | 0.000 | 04:04 |
| Control Test      |     | 0.078 | 04:04 |
| Air Blank         |     | 0.000 | 04:05 |
| Subject Sample    | # 1 | 0.241 | 04:06 |
| Breath Volume     |     | 1.722 | 04:06 |
| Air Blank         |     | 0.000 | 04:06 |
| Air Blank         |     | 0.000 | 04:08 |
| Subject Sample    | # 2 | 0.228 | 04:09 |
| Breath Volume     |     | 1.425 | 04:09 |
| Air Blank         |     | 0.000 | 04:09 |
| Control Test      |     | 0.079 | 04:10 |
| Air Blank         |     | 0.000 | 04:10 |
| Diagnostics Check |     | OK    | 04:10 |

---

Data Download Date/Time: 09/08/2020 15:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELVILLE,KELLY,M  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: TUERO,ADRIAN  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 09:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:51

SUBJECT'S LAST NAME: TOOLE

DATE/TIME OF TEST: 09/03/2020 10:22:16

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 10:16 |
| Air Blank         |     | 0.000 | 10:16 |
| Control Test      |     | 0.077 | 10:17 |
| Air Blank         |     | 0.000 | 10:17 |
| Subject Sample    | # 1 | 0.208 | 10:18 |
| Breath Volume     |     | 2.273 | 10:18 |
| Air Blank         |     | 0.000 | 10:18 |
| Air Blank         |     | 0.000 | 10:20 |
| Subject Sample    | # 2 | 0.201 | 10:20 |
| Breath Volume     |     | 1.789 | 10:20 |
| Air Blank         |     | 0.000 | 10:21 |
| Control Test      |     | 0.077 | 10:21 |
| Air Blank         |     | 0.000 | 10:22 |
| Diagnostics Check |     | OK    | 10:22 |

---

Data Download Date/Time: 09/08/2020 15:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MCLEAN,NICHOLAS  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/05/2020 00:54:41

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 00:48 |
| Air Blank         |     | 0.000 | 00:49 |
| Control Test      |     | 0.079 | 00:49 |
| Air Blank         |     | 0.000 | 00:49 |
| Subject Sample    | # 1 | 0.158 | 00:50 |
| Breath Volume     |     | 2.984 | 00:50 |
| Air Blank         |     | 0.000 | 00:50 |
| Air Blank         |     | 0.000 | 00:52 |
| Subject Sample    | # 2 | 0.152 | 00:53 |
| Breath Volume     |     | 1.695 | 00:53 |
| Air Blank         |     | 0.000 | 00:53 |
| Control Test      |     | 0.079 | 00:54 |
| Air Blank         |     | 0.000 | 00:54 |
| Diagnostics Check |     | OK    | 00:54 |

---

Data Download Date/Time: 09/08/2020 15:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON,TRAVIS,W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SANCHEZ,KEVIN  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: URDANETA

DATE/TIME OF TEST: 09/05/2020 03:37:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Control Test       | 0.079 | 03:27 |
| Air Blank          | 0.000 | 03:28 |
| Subject Sample # 1 | 0.204 | 03:30 |
| Breath Volume      | 1.652 | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Air Blank          | 0.000 | 03:32 |
| Subject Sample # 2 | 0.200 | 03:35 |
| Breath Volume      | 1.332 | 03:35 |
| Air Blank          | 0.000 | 03:36 |
| Control Test       | 0.077 | 03:36 |
| Air Blank          | 0.000 | 03:37 |
| Diagnostics Check  | OK    | 03:37 |

---

Data Download Date/Time: 09/08/2020 15:03



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,COURTNEY,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CAMARGO,WILLINGTON  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: CONDOR LEONARDO

DATE/TIME OF TEST: 09/05/2020 04:05:23

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 03:54 |                    |
| Air Blank         | 0.000 | 03:54 |                    |
| Control Test      | 0.078 | 03:54 |                    |
| Air Blank         | 0.000 | 03:55 |                    |
| Subject Sample    | # 1   | 03:58 | No Sample Provided |
| Breath Volume     | 0.000 | 03:58 |                    |
| Air Blank         | 0.000 | 03:58 |                    |
| Air Blank         | 0.000 | 04:00 |                    |
| Subject Sample    | # 2   | 04:04 | No Sample Provided |
| Breath Volume     | 0.000 | 04:04 |                    |
| Air Blank         | 0.000 | 04:04 |                    |
| Control Test      | 0.079 | 04:04 |                    |
| Air Blank         | 0.000 | 04:05 |                    |
| Diagnostics Check | OK    | 04:05 |                    |

---

Data Download Date/Time: 09/08/2020 15:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SANCHEZ, KEVIN  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: VELEZ MONTERO

DATE/TIME OF TEST: 09/05/2020 21:50:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:43 |
| Air Blank          | 0.000 | 21:43 |
| Control Test       | 0.078 | 21:44 |
| Air Blank          | 0.000 | 21:44 |
| Subject Sample # 1 | 0.197 | 21:46 |
| Breath Volume      | 1.855 | 21:46 |
| Air Blank          | 0.000 | 21:46 |
| Air Blank          | 0.000 | 21:48 |
| Subject Sample # 2 | 0.199 | 21:49 |
| Breath Volume      | 1.535 | 21:49 |
| Air Blank          | 0.000 | 21:49 |
| Control Test       | 0.078 | 21:50 |
| Air Blank          | 0.000 | 21:50 |
| Diagnostics Check  | OK    | 21:50 |

---

Data Download Date/Time: 09/08/2020 15:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: GONZALEZ,JOEL  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: VILLARES BOUDET

DATE/TIME OF TEST: 09/06/2020 01:50:35

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 01:39 |                    |
| Air Blank         | 0.000 | 01:39 |                    |
| Control Test      | 0.079 | 01:40 |                    |
| Air Blank         | 0.000 | 01:40 |                    |
| Subject Sample    | # 1   | 01:43 | No Sample Provided |
| Breath Volume     | 0.000 | 01:43 |                    |
| Air Blank         | 0.000 | 01:44 |                    |
| Air Blank         | 0.000 | 01:46 |                    |
| Subject Sample    | # 2   | 01:49 | No Sample Provided |
| Breath Volume     | 0.000 | 01:49 |                    |
| Air Blank         | 0.000 | 01:49 |                    |
| Control Test      | 0.079 | 01:50 |                    |
| Air Blank         | 0.000 | 01:50 |                    |
| Diagnostics Check | OK    | 01:50 |                    |

---

Data Download Date/Time: 09/08/2020 15:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CASTELLANOS  
PADRON,BRYAN  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: EMMANUEL

DATE/TIME OF TEST: 09/06/2020 04:20:27

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 04:14 |
| Air Blank         |     | 0.000 | 04:14 |
| Control Test      |     | 0.079 | 04:15 |
| Air Blank         |     | 0.000 | 04:15 |
| Subject Sample    | # 1 | 0.151 | 04:16 |
| Breath Volume     |     | 2.226 | 04:16 |
| Air Blank         |     | 0.000 | 04:16 |
| Air Blank         |     | 0.000 | 04:18 |
| Subject Sample    | # 2 | 0.151 | 04:19 |
| Breath Volume     |     | 2.527 | 04:19 |
| Air Blank         |     | 0.000 | 04:19 |
| Control Test      |     | 0.079 | 04:19 |
| Air Blank         |     | 0.000 | 04:20 |
| Diagnostics Check |     | OK    | 04:20 |

---

Data Download Date/Time: 09/08/2020 15:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WATKINS,ROMAN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: SOTO JR

DATE/TIME OF TEST: 09/06/2020 05:28:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:22 |
| Air Blank          | 0.000 | 05:23 |
| Control Test       | 0.080 | 05:23 |
| Air Blank          | 0.000 | 05:23 |
| Subject Sample # 1 | 0.279 | 05:24 |
| Breath Volume      | 2.304 | 05:24 |
| Air Blank          | 0.000 | 05:24 |
| Air Blank          | 0.000 | 05:26 |
| Subject Sample # 2 | 0.278 | 05:27 |
| Breath Volume      | 3.066 | 05:27 |
| Air Blank          | 0.000 | 05:27 |
| Control Test       | 0.079 | 05:28 |
| Air Blank          | 0.000 | 05:28 |
| Diagnostics Check  | OK    | 05:28 |

---

Data Download Date/Time: 09/08/2020 15:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: PETERSON, CAMERON  
ARRESTING OFFICER AGENCY: Winter Garden PD  
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/06/2020 14:38:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:30 |
| Air Blank          | 0.000 | 14:30 |
| Control Test       | 0.080 | 14:31 |
| Air Blank          | 0.000 | 14:31 |
| Subject Sample # 1 | 0.000 | 14:33 |
| Breath Volume      | 1.167 | 14:33 |
| Air Blank          | 0.000 | 14:34 |
| Air Blank          | 0.000 | 14:36 |
| Subject Sample # 2 | 0.000 | 14:37 |
| Breath Volume      | 1.222 | 14:37 |
| Air Blank          | 0.000 | 14:37 |
| Control Test       | 0.080 | 14:38 |
| Air Blank          | 0.000 | 14:38 |
| Diagnostics Check  | OK    | 14:38 |

---

Data Download Date/Time: 09/08/2020 15:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ROGERS,MICHELLE  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: RUDZKI

DATE/TIME OF TEST: 09/07/2020 01:16:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:08 |
| Air Blank          | 0.000 | 01:08 |
| Control Test       | 0.076 | 01:09 |
| Air Blank          | 0.000 | 01:09 |
| Subject Sample # 1 | 0.176 | 01:10 |
| Breath Volume      | 1.121 | 01:10 |
| Air Blank          | 0.000 | 01:11 |
| Air Blank          | 0.000 | 01:13 |
| Subject Sample # 2 | 0.164 | 01:15 |
| Breath Volume      | 1.355 | 01:15 |
| Air Blank          | 0.000 | 01:15 |
| Control Test       | 0.077 | 01:16 |
| Air Blank          | 0.000 | 01:16 |
| Diagnostics Check  | OK    | 01:16 |

---

Data Download Date/Time: 09/08/2020 15:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BANOS,FRANK  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: SHANDIK

DATE/TIME OF TEST: 09/12/2020 02:47:03

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 02:40 |
| Air Blank         |     | 0.000 | 02:40 |
| Control Test      |     | 0.081 | 02:41 |
| Air Blank         |     | 0.000 | 02:41 |
| Subject Sample    | # 1 | 0.265 | 02:42 |
| Breath Volume     |     | 1.710 | 02:42 |
| Air Blank         |     | 0.000 | 02:43 |
| Air Blank         |     | 0.000 | 02:44 |
| Subject Sample    | # 2 | 0.272 | 02:45 |
| Breath Volume     |     | 2.332 | 02:45 |
| Air Blank         |     | 0.000 | 02:46 |
| Control Test      |     | 0.080 | 02:46 |
| Air Blank         |     | 0.000 | 02:46 |
| Diagnostics Check |     | OK    | 02:47 |

---

Data Download Date/Time: 10/07/2020 14:51



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HALL,ANDREW  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 09/13/2020 01:26:41

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 01:20 |
| Air Blank         |     | 0.000 | 01:20 |
| Control Test      |     | 0.082 | 01:21 |
| Air Blank         |     | 0.000 | 01:21 |
| Subject Sample    | # 1 | 0.204 | 01:22 |
| Breath Volume     |     | 1.316 | 01:22 |
| Air Blank         |     | 0.000 | 01:22 |
| Air Blank         |     | 0.000 | 01:24 |
| Subject Sample    | # 2 | 0.202 | 01:25 |
| Breath Volume     |     | 1.242 | 01:25 |
| Air Blank         |     | 0.000 | 01:25 |
| Control Test      |     | 0.082 | 01:26 |
| Air Blank         |     | 0.000 | 01:26 |
| Diagnostics Check |     | OK    | 01:26 |

---

Data Download Date/Time: 10/07/2020 14:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ROCCHILD, MIGUEL  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: BLACKMON

DATE/TIME OF TEST: 09/14/2020 03:29:17

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 03:17 |                |
| Air Blank         |     | 0.000 | 03:18 |                |
| Control Test      |     | 0.082 | 03:18 |                |
| Air Blank         |     | 0.000 | 03:18 |                |
| Subject Sample    | # 1 | 0.144 | 03:22 | Volume Not Met |
| Breath Volume     |     | 0.074 | 03:22 |                |
| Air Blank         |     | 0.000 | 03:22 |                |
| Air Blank         |     | 0.000 | 03:24 |                |
| Subject Sample    | # 2 | 0.200 | 03:27 | Volume Not Met |
| Breath Volume     |     | 0.593 | 03:27 |                |
| Air Blank         |     | 0.000 | 03:28 |                |
| Control Test      |     | 0.081 | 03:28 |                |
| Air Blank         |     | 0.000 | 03:29 |                |
| Diagnostics Check |     | OK    | 03:29 |                |

---

Data Download Date/Time: 10/07/2020 14:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MELLA, MELISSA  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: RIVERA MORALES

DATE/TIME OF TEST: 09/15/2020 01:34:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:28 |
| Air Blank          | 0.000 | 01:28 |
| Control Test       | 0.082 | 01:29 |
| Air Blank          | 0.000 | 01:29 |
| Subject Sample # 1 | 0.000 | 01:30 |
| Breath Volume      | 3.777 | 01:30 |
| Air Blank          | 0.000 | 01:30 |
| Air Blank          | 0.000 | 01:32 |
| Subject Sample # 2 | 0.000 | 01:33 |
| Breath Volume      | 3.792 | 01:33 |
| Air Blank          | 0.000 | 01:33 |
| Control Test       | 0.082 | 01:34 |
| Air Blank          | 0.000 | 01:34 |
| Diagnostics Check  | OK    | 01:34 |

---

Data Download Date/Time: 10/07/2020 14:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CHAPMAN,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 09/15/2020 05:32:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:26 |
| Air Blank          | 0.000 | 05:26 |
| Control Test       | 0.080 | 05:26 |
| Air Blank          | 0.000 | 05:27 |
| Subject Sample # 1 | 0.139 | 05:28 |
| Breath Volume      | 1.550 | 05:28 |
| Air Blank          | 0.000 | 05:28 |
| Air Blank          | 0.000 | 05:30 |
| Subject Sample # 2 | 0.137 | 05:31 |
| Breath Volume      | 1.160 | 05:31 |
| Air Blank          | 0.000 | 05:31 |
| Control Test       | 0.080 | 05:32 |
| Air Blank          | 0.000 | 05:32 |
| Diagnostics Check  | OK    | 05:32 |

---

Data Download Date/Time: 10/07/2020 14:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MOONDA,SHUAIB  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 09/15/2020 07:42:41

---

|                    |       |       |                |
|--------------------|-------|-------|----------------|
| Diagnostics Check  | OK    | 07:29 |                |
| Air Blank          | 0.000 | 07:30 |                |
| Control Test       | 0.082 | 07:30 |                |
| Air Blank          | 0.000 | 07:31 |                |
| Subject Sample # 1 | 0.000 | 07:32 |                |
| Breath Volume      | 1.175 | 07:32 |                |
| Air Blank          | 0.000 | 07:32 |                |
| Air Blank          | 0.000 | 07:34 |                |
| Subject Sample # 2 | 0.000 | 07:37 | Volume Not Met |
| Breath Volume      | 0.050 | 07:37 |                |
| Air Blank          | 0.000 | 07:38 |                |
| Air Blank          | 0.000 | 07:40 |                |
| Subject Sample # 3 | 0.000 | 07:41 |                |
| Breath Volume      | 1.292 | 07:41 |                |
| Air Blank          | 0.000 | 07:41 |                |
| Control Test       | 0.082 | 07:42 |                |
| Air Blank          | 0.000 | 07:42 |                |
| Diagnostics Check  | OK    | 07:42 |                |

---

Data Download Date/Time: 10/07/2020 14:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DELLOVA,,JOSEPH  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: MORALES GODINEZ

DATE/TIME OF TEST: 09/19/2020 01:10:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:04 |
| Air Blank          | 0.000 | 01:04 |
| Control Test       | 0.081 | 01:05 |
| Air Blank          | 0.000 | 01:05 |
| Subject Sample # 1 | 0.151 | 01:06 |
| Breath Volume      | 1.960 | 01:06 |
| Air Blank          | 0.000 | 01:06 |
| Air Blank          | 0.000 | 01:08 |
| Subject Sample # 2 | 0.148 | 01:09 |
| Breath Volume      | 1.941 | 01:09 |
| Air Blank          | 0.000 | 01:09 |
| Control Test       | 0.081 | 01:09 |
| Air Blank          | 0.000 | 01:10 |
| Diagnostics Check  | OK    | 01:10 |

---

Data Download Date/Time: 10/07/2020 14:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BODE,JOSHUA  
ARRESTING OFFICER AGENCY: Ocoee PD  
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: IXLAJVASQUEZ

DATE/TIME OF TEST: 09/19/2020 03:49:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:42 |
| Air Blank          | 0.000 | 03:42 |
| Control Test       | 0.081 | 03:42 |
| Air Blank          | 0.000 | 03:43 |
| Subject Sample # 1 | 0.262 | 03:45 |
| Breath Volume      | 2.835 | 03:45 |
| Air Blank          | 0.000 | 03:45 |
| Air Blank          | 0.000 | 03:47 |
| Subject Sample # 2 | 0.254 | 03:48 |
| Breath Volume      | 1.195 | 03:48 |
| Air Blank          | 0.000 | 03:49 |
| Control Test       | 0.080 | 03:49 |
| Air Blank          | 0.000 | 03:49 |
| Diagnostics Check  | OK    | 03:49 |

---

Data Download Date/Time: 10/07/2020 14:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SALAZAR,JUAN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: FERNANDESDESOUZA

DATE/TIME OF TEST: 09/19/2020 23:50:24

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 23:37 |                |
| Air Blank         |     | 0.000 | 23:38 |                |
| Control Test      |     | 0.081 | 23:38 |                |
| Air Blank         |     | 0.000 | 23:38 |                |
| Subject Sample    | # 1 | 0.111 | 23:42 | Volume Not Met |
| Breath Volume     |     | 0.539 | 23:42 |                |
| Air Blank         |     | 0.000 | 23:42 |                |
| Air Blank         |     | 0.000 | 23:44 |                |
| Subject Sample    | # 2 | 0.142 | 23:45 |                |
| Breath Volume     |     | 2.011 | 23:45 |                |
| Air Blank         |     | 0.000 | 23:46 |                |
| Air Blank         |     | 0.000 | 23:48 |                |
| Subject Sample    | # 3 | 0.140 | 23:48 |                |
| Breath Volume     |     | 1.777 | 23:48 |                |
| Air Blank         |     | 0.000 | 23:49 |                |
| Control Test      |     | 0.080 | 23:49 |                |
| Air Blank         |     | 0.000 | 23:50 |                |
| Diagnostics Check |     | OK    | 23:50 |                |

---

Data Download Date/Time: 10/07/2020 14:51



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ERICKSON,STEPHEN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: VEGAMARTINEZ

DATE/TIME OF TEST: 09/20/2020 00:49:41

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 00:43 |
| Air Blank         |     | 0.000 | 00:43 |
| Control Test      |     | 0.081 | 00:43 |
| Air Blank         |     | 0.000 | 00:44 |
| Subject Sample    | # 1 | 0.205 | 00:45 |
| Breath Volume     |     | 1.613 | 00:45 |
| Air Blank         |     | 0.000 | 00:45 |
| Air Blank         |     | 0.000 | 00:47 |
| Subject Sample    | # 2 | 0.203 | 00:48 |
| Breath Volume     |     | 1.726 | 00:48 |
| Air Blank         |     | 0.000 | 00:48 |
| Control Test      |     | 0.081 | 00:49 |
| Air Blank         |     | 0.000 | 00:49 |
| Diagnostics Check |     | OK    | 00:49 |

---

Data Download Date/Time: 10/07/2020 14:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: PARRISH,KAYLA  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: BATISTA

DATE/TIME OF TEST: 09/21/2020 19:07:12

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:03 |                      |
| Air Blank         | 0.000 | 19:04 |                      |
| Control Test      | 0.080 | 19:04 |                      |
| Air Blank         | 0.000 | 19:05 |                      |
| Subject Sample    | # 1   | 19:05 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:05 |                      |
| Air Blank         | 0.000 | 19:06 |                      |
| Control Test      | 0.080 | 19:06 |                      |
| Air Blank         | 0.000 | 19:07 |                      |
| Diagnostics Check | OK    | 19:07 |                      |

---

Data Download Date/Time: 10/07/2020 14:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SANCHEZ,KEVIN  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: LOCK

DATE/TIME OF TEST: 09/24/2020 00:17:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:10 |
| Air Blank          | 0.000 | 00:10 |
| Control Test       | 0.081 | 00:11 |
| Air Blank          | 0.000 | 00:11 |
| Subject Sample # 1 | 0.195 | 00:12 |
| Breath Volume      | 2.933 | 00:12 |
| Air Blank          | 0.000 | 00:13 |
| Air Blank          | 0.000 | 00:14 |
| Subject Sample # 2 | 0.192 | 00:15 |
| Breath Volume      | 1.910 | 00:15 |
| Air Blank          | 0.000 | 00:16 |
| Control Test       | 0.080 | 00:16 |
| Air Blank          | 0.000 | 00:17 |
| Diagnostics Check  | OK    | 00:17 |

---

Data Download Date/Time: 10/07/2020 14:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HALLOCK,STEPHEN  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: THIEME

DATE/TIME OF TEST: 09/25/2020 02:32:17

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 02:26 |
| Air Blank         |     | 0.000 | 02:26 |
| Control Test      |     | 0.081 | 02:26 |
| Air Blank         |     | 0.000 | 02:27 |
| Subject Sample    | # 1 | 0.216 | 02:27 |
| Breath Volume     |     | 1.769 | 02:27 |
| Air Blank         |     | 0.000 | 02:28 |
| Air Blank         |     | 0.000 | 02:30 |
| Subject Sample    | # 2 | 0.221 | 02:30 |
| Breath Volume     |     | 1.746 | 02:30 |
| Air Blank         |     | 0.000 | 02:31 |
| Control Test      |     | 0.081 | 02:31 |
| Air Blank         |     | 0.000 | 02:32 |
| Diagnostics Check |     | OK    | 02:32 |

---

Data Download Date/Time: 10/07/2020 14:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HARRIS,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: GUPPENBERGER

DATE/TIME OF TEST: 09/25/2020 18:30:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:23 |
| Air Blank          | 0.000 | 18:24 |
| Control Test       | 0.081 | 18:24 |
| Air Blank          | 0.000 | 18:25 |
| Subject Sample # 1 | 0.212 | 18:25 |
| Breath Volume      | 1.925 | 18:25 |
| Air Blank          | 0.000 | 18:26 |
| Air Blank          | 0.000 | 18:28 |
| Subject Sample # 2 | 0.211 | 18:28 |
| Breath Volume      | 1.785 | 18:28 |
| Air Blank          | 0.000 | 18:29 |
| Control Test       | 0.081 | 18:29 |
| Air Blank          | 0.000 | 18:30 |
| Diagnostics Check  | OK    | 18:30 |

---

Data Download Date/Time: 10/07/2020 14:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HARWOOD,ARMANDO  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: WYLAM

DATE/TIME OF TEST: 09/26/2020 02:40:33

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:37 |                      |
| Air Blank         | 0.000 | 02:38 |                      |
| Control Test      | 0.081 | 02:38 |                      |
| Air Blank         | 0.000 | 02:39 |                      |
| Subject Sample    | # 1   | 02:39 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:39 |                      |
| Air Blank         | 0.000 | 02:39 |                      |
| Control Test      | 0.081 | 02:39 |                      |
| Air Blank         | 0.000 | 02:40 |                      |
| Diagnostics Check | OK    | 02:40 |                      |

---

Data Download Date/Time: 10/07/2020 14:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HEBEL,GRIFFIN  
ARRESTING OFFICER AGENCY: Windermere PD  
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 09/26/2020 21:22:11

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 21:16 |
| Air Blank         |     | 0.000 | 21:16 |
| Control Test      |     | 0.081 | 21:16 |
| Air Blank         |     | 0.000 | 21:17 |
| Subject Sample    | # 1 | 0.205 | 21:17 |
| Breath Volume     |     | 1.289 | 21:17 |
| Air Blank         |     | 0.000 | 21:18 |
| Air Blank         |     | 0.000 | 21:20 |
| Subject Sample    | # 2 | 0.211 | 21:20 |
| Breath Volume     |     | 1.652 | 21:20 |
| Air Blank         |     | 0.000 | 21:21 |
| Control Test      |     | 0.081 | 21:21 |
| Air Blank         |     | 0.000 | 21:22 |
| Diagnostics Check |     | OK    | 21:22 |

---

Data Download Date/Time: 10/07/2020 14:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MONTGOMERY,JAMES  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: SALEH

DATE/TIME OF TEST: 09/27/2020 05:45:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:39 |
| Air Blank          | 0.000 | 05:39 |
| Control Test       | 0.081 | 05:40 |
| Air Blank          | 0.000 | 05:40 |
| Subject Sample # 1 | 0.175 | 05:41 |
| Breath Volume      | 1.191 | 05:41 |
| Air Blank          | 0.000 | 05:41 |
| Air Blank          | 0.000 | 05:43 |
| Subject Sample # 2 | 0.167 | 05:43 |
| Breath Volume      | 1.296 | 05:43 |
| Air Blank          | 0.000 | 05:44 |
| Control Test       | 0.081 | 05:44 |
| Air Blank          | 0.000 | 05:45 |
| Diagnostics Check  | OK    | 05:45 |

---

Data Download Date/Time: 10/07/2020 14:52



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SIMPSON,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:04

SUBJECT'S LAST NAME: OSORIORODRIGUEZ

DATE/TIME OF TEST: 09/27/2020 06:45:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:39 |
| Air Blank          | 0.000 | 06:40 |
| Control Test       | 0.081 | 06:40 |
| Air Blank          | 0.000 | 06:40 |
| Subject Sample # 1 | 0.119 | 06:41 |
| Breath Volume      | 1.582 | 06:41 |
| Air Blank          | 0.000 | 06:42 |
| Air Blank          | 0.000 | 06:43 |
| Subject Sample # 2 | 0.123 | 06:44 |
| Breath Volume      | 1.742 | 06:44 |
| Air Blank          | 0.000 | 06:44 |
| Control Test       | 0.082 | 06:45 |
| Air Blank          | 0.000 | 06:45 |
| Diagnostics Check  | OK    | 06:45 |

---

Data Download Date/Time: 10/07/2020 14:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ROBBINS,NATHAN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 05:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:41

SUBJECT'S LAST NAME: GONZALEZVARGAS

DATE/TIME OF TEST: 09/27/2020 07:45:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:38 |
| Air Blank          | 0.000 | 07:39 |
| Control Test       | 0.082 | 07:39 |
| Air Blank          | 0.000 | 07:40 |
| Subject Sample # 1 | 0.128 | 07:41 |
| Breath Volume      | 2.089 | 07:41 |
| Air Blank          | 0.000 | 07:41 |
| Air Blank          | 0.000 | 07:43 |
| Subject Sample # 2 | 0.128 | 07:44 |
| Breath Volume      | 2.007 | 07:44 |
| Air Blank          | 0.000 | 07:44 |
| Control Test       | 0.081 | 07:45 |
| Air Blank          | 0.000 | 07:45 |
| Diagnostics Check  | OK    | 07:45 |

---

Data Download Date/Time: 10/07/2020 14:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ERICKSON,STEPHEN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 09/30/2020 01:50:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:44 |
| Air Blank          | 0.000 | 01:45 |
| Control Test       | 0.081 | 01:45 |
| Air Blank          | 0.000 | 01:46 |
| Subject Sample # 1 | 0.000 | 01:46 |
| Breath Volume      | 2.382 | 01:46 |
| Air Blank          | 0.000 | 01:47 |
| Air Blank          | 0.000 | 01:49 |
| Subject Sample # 2 | 0.000 | 01:49 |
| Breath Volume      | 2.492 | 01:49 |
| Air Blank          | 0.000 | 01:49 |
| Control Test       | 0.081 | 01:50 |
| Air Blank          | 0.000 | 01:50 |
| Diagnostics Check  | OK    | 01:50 |

---

Data Download Date/Time: 10/07/2020 14:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: JACKSON,ALLEN  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06419080A1  
DRY GAS EXPIRY DATE: 05/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 09/30/2020 23:06:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:00 |
| Air Blank          | 0.000 | 23:01 |
| Control Test       | 0.082 | 23:01 |
| Air Blank          | 0.000 | 23:01 |
| Subject Sample # 1 | 0.245 | 23:02 |
| Breath Volume      | 2.015 | 23:02 |
| Air Blank          | 0.000 | 23:02 |
| Air Blank          | 0.000 | 23:04 |
| Subject Sample # 2 | 0.247 | 23:04 |
| Breath Volume      | 1.953 | 23:04 |
| Air Blank          | 0.000 | 23:05 |
| Control Test       | 0.081 | 23:05 |
| Air Blank          | 0.000 | 23:06 |
| Diagnostics Check  | OK    | 23:06 |

---

Data Download Date/Time: 10/07/2020 14:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PICHECO,JOHN,A

BREATH TEST OPERATOR AGENCY: Cocoa PD

ARRESTING OFFICER: DEAN,ISAAC

ARRESTING OFFICER AGENCY: Cocoa PD

LAST AGENCY INSPECTION: 08/13/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30219080A4

DRY GAS EXPIRY DATE: 01/05/2022

VIOLATION CODE: DUI

ARREST TIME: 00:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: OROZCO

DATE/TIME OF TEST: 09/08/2020 00:46:43

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 00:45 |                           |
| Air Blank         | 0.000 | 00:45 |                           |
| Control Test      | 0.072 | 00:46 | Control Outside Tolerance |
| Air Blank         | 0.000 | 00:46 |                           |

---

Data Download Date/Time: 09/19/2020 19:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PICHECO,JOHN,A

BREATH TEST OPERATOR AGENCY: Cocoa PD

ARRESTING OFFICER: DEAN,ISAAC

ARRESTING OFFICER AGENCY: Cocoa PD

LAST AGENCY INSPECTION: 08/13/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30219080A4

DRY GAS EXPIRY DATE: 01/05/2022

VIOLATION CODE: DUI

ARREST TIME: 00:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: OROZCO

DATE/TIME OF TEST: 09/08/2020 00:53:23

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 00:52 |                           |
| Air Blank         | 0.000 | 00:52 |                           |
| Control Test      | 0.072 | 00:52 | Control Outside Tolerance |
| Air Blank         | 0.000 | 00:53 |                           |

---

Data Download Date/Time: 09/19/2020 19:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PENDER,DEREK,A  
BREATH TEST OPERATOR AGENCY: Rockledge PD  
ARRESTING OFFICER: PENDER,DEREK  
ARRESTING OFFICER AGENCY: Rockledge PD  
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1175968  
DRY GAS EXPIRY DATE: 07/22/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: CRADDOCK

DATE/TIME OF TEST: 09/05/2020 02:56:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:49 |
| Air Blank          | 0.000 | 02:50 |
| Control Test       | 0.079 | 02:50 |
| Air Blank          | 0.000 | 02:50 |
| Subject Sample # 1 | 0.232 | 02:51 |
| Breath Volume      | 1.839 | 02:51 |
| Air Blank          | 0.000 | 02:52 |
| Air Blank          | 0.000 | 02:54 |
| Subject Sample # 2 | 0.235 | 02:54 |
| Breath Volume      | 1.671 | 02:54 |
| Air Blank          | 0.000 | 02:55 |
| Control Test       | 0.078 | 02:55 |
| Air Blank          | 0.000 | 02:56 |
| Diagnostics Check  | OK    | 02:56 |

---

Data Download Date/Time: 10/16/2020 15:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENNEDY,PATRICK,W  
BREATH TEST OPERATOR AGENCY: Rockledge PD  
ARRESTING OFFICER: LEITCH,RICKFORD  
ARRESTING OFFICER AGENCY: Cocoa PD  
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1175968  
DRY GAS EXPIRY DATE: 07/22/2022  
VIOLATION CODE: DUI  
ARREST TIME: 12:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:17

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 09/10/2020 13:47:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:40 |
| Air Blank          | 0.000 | 13:40 |
| Control Test       | 0.078 | 13:41 |
| Air Blank          | 0.000 | 13:41 |
| Subject Sample # 1 | 0.272 | 13:42 |
| Breath Volume      | 1.394 | 13:42 |
| Air Blank          | 0.000 | 13:43 |
| Air Blank          | 0.000 | 13:44 |
| Subject Sample # 2 | 0.256 | 13:46 |
| Breath Volume      | 1.765 | 13:46 |
| Air Blank          | 0.000 | 13:46 |
| Control Test       | 0.076 | 13:47 |
| Air Blank          | 0.000 | 13:47 |
| Diagnostics Check  | OK    | 13:47 |

---

Data Download Date/Time: 10/16/2020 15:34



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOMEZ,IVETTE,  
BREATH TEST OPERATOR AGENCY: Rockledge PD  
ARRESTING OFFICER: MOREHEAD,AVERY  
ARRESTING OFFICER AGENCY: Rockledge PD  
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1175968  
DRY GAS EXPIRY DATE: 07/22/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: HOOG

DATE/TIME OF TEST: 09/12/2020 21:46:31

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:43 |                      |
| Air Blank         | 0.000 | 21:44 |                      |
| Control Test      | 0.078 | 21:44 |                      |
| Air Blank         | 0.000 | 21:44 |                      |
| Subject Sample    | # 1   | 21:45 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:45 |                      |
| Air Blank         | 0.000 | 21:45 |                      |
| Control Test      | 0.079 | 21:45 |                      |
| Air Blank         | 0.000 | 21:46 |                      |
| Diagnostics Check | OK    | 21:46 |                      |

---

Data Download Date/Time: 10/16/2020 15:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARNER, TYRELL, J  
BREATH TEST OPERATOR AGENCY: Rockledge PD  
ARRESTING OFFICER: ROWELL, JENNA  
ARRESTING OFFICER AGENCY: Rockledge PD  
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1175968  
DRY GAS EXPIRY DATE: 07/22/2022  
VIOLATION CODE: DUI  
ARREST TIME: 07:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:07

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/21/2020 07:50:34

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 07:45 |                      |
| Air Blank         | 0.000 | 07:46 |                      |
| Control Test      | 0.077 | 07:46 |                      |
| Air Blank         | 0.000 | 07:47 |                      |
| Subject Sample    | # 1   | 07:49 | Subject Test Refused |
| Breath Volume     | 0.000 | 07:49 |                      |
| Air Blank         | 0.000 | 07:49 |                      |
| Control Test      | 0.077 | 07:49 |                      |
| Air Blank         | 0.000 | 07:50 |                      |
| Diagnostics Check | OK    | 07:50 |                      |

---

Data Download Date/Time: 10/16/2020 15:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOMEZ,IVETTE,  
BREATH TEST OPERATOR AGENCY: Rockledge PD  
ARRESTING OFFICER: CACERES,JACOB  
ARRESTING OFFICER AGENCY: Rockledge PD  
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1775968  
DRY GAS EXPIRY DATE: 07/22/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 09/26/2020 21:37:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:30 |
| Air Blank          | 0.000 | 21:30 |
| Control Test       | 0.078 | 21:30 |
| Air Blank          | 0.000 | 21:31 |
| Subject Sample # 1 | 0.229 | 21:32 |
| Breath Volume      | 1.187 | 21:32 |
| Air Blank          | 0.000 | 21:32 |
| Air Blank          | 0.000 | 21:34 |
| Subject Sample # 2 | 0.230 | 21:35 |
| Breath Volume      | 1.382 | 21:35 |
| Air Blank          | 0.000 | 21:36 |
| Control Test       | 0.077 | 21:36 |
| Air Blank          | 0.000 | 21:37 |
| Diagnostics Check  | OK    | 21:37 |

---

Data Download Date/Time: 10/16/2020 15:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,A,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: VAZQUEZ,A  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: RANEY

DATE/TIME OF TEST: 09/03/2020 00:48:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:40 |
| Air Blank          | 0.000 | 00:40 |
| Control Test       | 0.081 | 00:41 |
| Air Blank          | 0.000 | 00:41 |
| Subject Sample # 1 | 0.183 | 00:43 |
| Breath Volume      | 1.750 | 00:43 |
| Air Blank          | 0.000 | 00:44 |
| Air Blank          | 0.000 | 00:45 |
| Subject Sample # 2 | 0.189 | 00:47 |
| Breath Volume      | 1.304 | 00:47 |
| Air Blank          | 0.000 | 00:47 |
| Control Test       | 0.079 | 00:48 |
| Air Blank          | 0.000 | 00:48 |
| Diagnostics Check  | OK    | 00:48 |

---

Data Download Date/Time: 09/29/2020 13:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIVINGSTON,CHRISTOPHER,S  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: MITCHELL,B  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 14:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: GRUTTADAURIA

DATE/TIME OF TEST: 09/05/2020 15:49:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:43 |
| Air Blank          | 0.000 | 15:43 |
| Control Test       | 0.081 | 15:44 |
| Air Blank          | 0.000 | 15:44 |
| Subject Sample # 1 | 0.000 | 15:45 |
| Breath Volume      | 2.695 | 15:45 |
| Air Blank          | 0.000 | 15:45 |
| Air Blank          | 0.000 | 15:47 |
| Subject Sample # 2 | 0.000 | 15:47 |
| Breath Volume      | 3.082 | 15:47 |
| Air Blank          | 0.000 | 15:48 |
| Control Test       | 0.081 | 15:48 |
| Air Blank          | 0.000 | 15:49 |
| Diagnostics Check  | OK    | 15:49 |

---

Data Download Date/Time: 09/29/2020 13:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRISON,A,  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: PEREZ,L  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: CREWS

DATE/TIME OF TEST: 09/06/2020 02:55:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:49 |
| Air Blank          | 0.000 | 02:49 |
| Control Test       | 0.081 | 02:49 |
| Air Blank          | 0.000 | 02:50 |
| Subject Sample # 1 | 0.154 | 02:50 |
| Breath Volume      | 1.378 | 02:50 |
| Air Blank          | 0.000 | 02:51 |
| Air Blank          | 0.000 | 02:53 |
| Subject Sample # 2 | 0.140 | 02:53 |
| Breath Volume      | 1.347 | 02:53 |
| Air Blank          | 0.000 | 02:54 |
| Control Test       | 0.078 | 02:54 |
| Air Blank          | 0.000 | 02:55 |
| Diagnostics Check  | OK    | 02:55 |

---

Data Download Date/Time: 09/29/2020 13:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,A,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: BLAZEWICZ,A  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: SANTOS IGLESIAS

DATE/TIME OF TEST: 09/08/2020 21:51:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:44 |
| Air Blank          | 0.000 | 21:45 |
| Control Test       | 0.081 | 21:45 |
| Air Blank          | 0.000 | 21:46 |
| Subject Sample # 1 | 0.000 | 21:46 |
| Breath Volume      | 1.812 | 21:46 |
| Air Blank          | 0.000 | 21:47 |
| Air Blank          | 0.000 | 21:49 |
| Subject Sample # 2 | 0.000 | 21:49 |
| Breath Volume      | 1.562 | 21:49 |
| Air Blank          | 0.000 | 21:50 |
| Control Test       | 0.079 | 21:50 |
| Air Blank          | 0.000 | 21:50 |
| Diagnostics Check  | OK    | 21:51 |

---

Data Download Date/Time: 09/29/2020 13:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,A,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: BLAZEWICZ,A  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: VERA

DATE/TIME OF TEST: 09/11/2020 19:28:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:22 |
| Air Blank          | 0.000 | 19:22 |
| Control Test       | 0.082 | 19:22 |
| Air Blank          | 0.000 | 19:23 |
| Subject Sample # 1 | 0.000 | 19:23 |
| Breath Volume      | 1.515 | 19:23 |
| Air Blank          | 0.000 | 19:24 |
| Air Blank          | 0.000 | 19:26 |
| Subject Sample # 2 | 0.000 | 19:27 |
| Breath Volume      | 1.660 | 19:27 |
| Air Blank          | 0.000 | 19:27 |
| Control Test       | 0.081 | 19:27 |
| Air Blank          | 0.000 | 19:28 |
| Diagnostics Check  | OK    | 19:28 |

---

Data Download Date/Time: 09/29/2020 13:09



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,A,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: BLAZEWICZ,A  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 09/13/2020 04:05:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:58 |
| Air Blank          | 0.000 | 03:59 |
| Control Test       | 0.079 | 03:59 |
| Air Blank          | 0.000 | 04:00 |
| Subject Sample # 1 | 0.095 | 04:00 |
| Breath Volume      | 1.339 | 04:00 |
| Air Blank          | 0.000 | 04:01 |
| Air Blank          | 0.000 | 04:03 |
| Subject Sample # 2 | 0.100 | 04:03 |
| Breath Volume      | 2.011 | 04:03 |
| Air Blank          | 0.000 | 04:04 |
| Control Test       | 0.080 | 04:04 |
| Air Blank          | 0.000 | 04:05 |
| Diagnostics Check  | OK    | 04:05 |

---

Data Download Date/Time: 09/29/2020 13:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,A,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: VAZQUEZ,A  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: BLADEK

DATE/TIME OF TEST: 09/13/2020 04:24:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:18 |
| Air Blank          | 0.000 | 04:19 |
| Control Test       | 0.081 | 04:19 |
| Air Blank          | 0.000 | 04:20 |
| Subject Sample # 1 | 0.000 | 04:20 |
| Breath Volume      | 2.355 | 04:20 |
| Air Blank          | 0.000 | 04:21 |
| Air Blank          | 0.000 | 04:23 |
| Subject Sample # 2 | 0.000 | 04:23 |
| Breath Volume      | 2.437 | 04:23 |
| Air Blank          | 0.000 | 04:23 |
| Control Test       | 0.080 | 04:24 |
| Air Blank          | 0.000 | 04:24 |
| Diagnostics Check  | OK    | 04:24 |

---

Data Download Date/Time: 09/29/2020 13:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHILBIN,SKYLER,J  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: KEENER,BAYLEE  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: KAISER

DATE/TIME OF TEST: 09/14/2020 04:42:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:36 |
| Air Blank          | 0.000 | 04:37 |
| Control Test       | 0.080 | 04:37 |
| Air Blank          | 0.000 | 04:37 |
| Subject Sample # 1 | 0.000 | 04:38 |
| Breath Volume      | 1.781 | 04:38 |
| Air Blank          | 0.000 | 04:38 |
| Air Blank          | 0.000 | 04:40 |
| Subject Sample # 2 | 0.000 | 04:41 |
| Breath Volume      | 1.570 | 04:41 |
| Air Blank          | 0.000 | 04:41 |
| Control Test       | 0.080 | 04:42 |
| Air Blank          | 0.000 | 04:42 |
| Diagnostics Check  | OK    | 04:42 |

---

Data Download Date/Time: 09/29/2020 13:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,A,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: DUMMER,L  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: WHITCOMB

DATE/TIME OF TEST: 09/17/2020 21:24:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:18 |
| Air Blank          | 0.000 | 21:19 |
| Control Test       | 0.080 | 21:19 |
| Air Blank          | 0.000 | 21:19 |
| Subject Sample # 1 | 0.000 | 21:20 |
| Breath Volume      | 1.835 | 21:20 |
| Air Blank          | 0.000 | 21:20 |
| Air Blank          | 0.000 | 21:22 |
| Subject Sample # 2 | 0.000 | 21:23 |
| Breath Volume      | 1.808 | 21:23 |
| Air Blank          | 0.000 | 21:23 |
| Control Test       | 0.080 | 21:24 |
| Air Blank          | 0.000 | 21:24 |
| Diagnostics Check  | OK    | 21:24 |

---

Data Download Date/Time: 09/29/2020 13:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIVINGSTON,CHRISTOPHER,S  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: BLACKMON,TERRANCE  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 13:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:23

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 09/19/2020 14:43:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:37 |
| Air Blank          | 0.000 | 14:38 |
| Control Test       | 0.082 | 14:38 |
| Air Blank          | 0.000 | 14:39 |
| Subject Sample # 1 | 0.000 | 14:39 |
| Breath Volume      | 3.082 | 14:39 |
| Air Blank          | 0.000 | 14:40 |
| Air Blank          | 0.000 | 14:42 |
| Subject Sample # 2 | 0.000 | 14:42 |
| Breath Volume      | 2.722 | 14:42 |
| Air Blank          | 0.000 | 14:43 |
| Control Test       | 0.079 | 14:43 |
| Air Blank          | 0.000 | 14:43 |
| Diagnostics Check  | OK    | 14:43 |

---

Data Download Date/Time: 09/29/2020 13:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRISON,A,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: GOULD,W  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: OTOOLE

DATE/TIME OF TEST: 09/19/2020 20:51:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:44 |
| Air Blank          | 0.000 | 20:45 |
| Control Test       | 0.080 | 20:45 |
| Air Blank          | 0.000 | 20:46 |
| Subject Sample # 1 | 0.359 | 20:46 |
| Breath Volume      | 1.570 | 20:46 |
| Air Blank          | 0.000 | 20:47 |
| Air Blank          | 0.000 | 20:48 |
| Subject Sample # 2 | 0.360 | 20:49 |
| Breath Volume      | 1.246 | 20:49 |
| Air Blank          | 0.000 | 20:50 |
| Control Test       | 0.079 | 20:50 |
| Air Blank          | 0.000 | 20:50 |
| Diagnostics Check  | OK    | 20:51 |

---

Data Download Date/Time: 09/29/2020 13:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,A,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: VAZQUEZ,A  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 09/21/2020 06:37:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:31 |
| Air Blank          | 0.000 | 06:31 |
| Control Test       | 0.082 | 06:31 |
| Air Blank          | 0.000 | 06:32 |
| Subject Sample # 1 | 0.055 | 06:33 |
| Breath Volume      | 1.804 | 06:33 |
| Air Blank          | 0.000 | 06:33 |
| Air Blank          | 0.000 | 06:35 |
| Subject Sample # 2 | 0.051 | 06:36 |
| Breath Volume      | 1.746 | 06:36 |
| Air Blank          | 0.000 | 06:36 |
| Control Test       | 0.080 | 06:37 |
| Air Blank          | 0.000 | 06:37 |
| Diagnostics Check  | OK    | 06:37 |

---

Data Download Date/Time: 09/29/2020 13:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,A,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: TUBAJ,N  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: BONZAR

DATE/TIME OF TEST: 09/21/2020 11:59:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 11:53 |
| Air Blank          | 0.000 | 11:53 |
| Control Test       | 0.081 | 11:53 |
| Air Blank          | 0.000 | 11:54 |
| Subject Sample # 1 | 0.000 | 11:54 |
| Breath Volume      | 1.500 | 11:54 |
| Air Blank          | 0.000 | 11:55 |
| Air Blank          | 0.000 | 11:57 |
| Subject Sample # 2 | 0.000 | 11:57 |
| Breath Volume      | 1.773 | 11:57 |
| Air Blank          | 0.000 | 11:58 |
| Control Test       | 0.081 | 11:58 |
| Air Blank          | 0.000 | 11:59 |
| Diagnostics Check  | OK    | 11:59 |

---

Data Download Date/Time: 09/29/2020 13:09



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD, WILLIAM,  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: HOWARD, WILLIAM  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: SALANTY

DATE/TIME OF TEST: 09/23/2020 09:55:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:46 |
| Air Blank          | 0.000 | 09:46 |
| Control Test       | 0.082 | 09:46 |
| Air Blank          | 0.000 | 09:47 |
| Subject Sample # 1 | 0.112 | 09:48 |
| Breath Volume      | 1.695 | 09:48 |
| Air Blank          | 0.000 | 09:49 |
| Air Blank          | 0.000 | 09:50 |
| Subject Sample # 2 | 0.108 | 09:53 |
| Breath Volume      | 1.488 | 09:53 |
| Air Blank          | 0.000 | 09:54 |
| Control Test       | 0.082 | 09:54 |
| Air Blank          | 0.000 | 09:55 |
| Diagnostics Check  | OK    | 09:55 |

---

Data Download Date/Time: 09/29/2020 13:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,A,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: BLAZEWICZ,A  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: MOTOWYLAK

DATE/TIME OF TEST: 09/26/2020 08:00:26

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 07:53 |
| Air Blank         |     | 0.000 | 07:53 |
| Control Test      |     | 0.079 | 07:54 |
| Air Blank         |     | 0.000 | 07:54 |
| Subject Sample    | # 1 | 0.158 | 07:55 |
| Breath Volume     |     | 2.292 | 07:55 |
| Air Blank         |     | 0.000 | 07:56 |
| Air Blank         |     | 0.000 | 07:58 |
| Subject Sample    | # 2 | 0.151 | 07:58 |
| Breath Volume     |     | 1.687 | 07:58 |
| Air Blank         |     | 0.000 | 07:59 |
| Control Test      |     | 0.079 | 07:59 |
| Air Blank         |     | 0.000 | 08:00 |
| Diagnostics Check |     | OK    | 08:00 |

---

Data Download Date/Time: 09/29/2020 13:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,A,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: FREITAG,J  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 09/26/2020 09:39:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:32 |
| Air Blank          | 0.000 | 09:33 |
| Control Test       | 0.080 | 09:33 |
| Air Blank          | 0.000 | 09:34 |
| Subject Sample # 1 | 0.123 | 09:35 |
| Breath Volume      | 1.656 | 09:35 |
| Air Blank          | 0.000 | 09:35 |
| Air Blank          | 0.000 | 09:37 |
| Subject Sample # 2 | 0.118 | 09:38 |
| Breath Volume      | 1.230 | 09:38 |
| Air Blank          | 0.000 | 09:38 |
| Control Test       | 0.078 | 09:39 |
| Air Blank          | 0.000 | 09:39 |
| Diagnostics Check  | OK    | 09:39 |

---

Data Download Date/Time: 09/29/2020 13:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,A,M  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: VAZQUEZ,A  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1249794  
DRY GAS EXPIRY DATE: 12/13/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: SEXTON

DATE/TIME OF TEST: 09/27/2020 07:03:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:56 |
| Air Blank          | 0.000 | 06:57 |
| Control Test       | 0.081 | 06:57 |
| Air Blank          | 0.000 | 06:58 |
| Subject Sample # 1 | 0.122 | 06:58 |
| Breath Volume      | 1.617 | 06:58 |
| Air Blank          | 0.000 | 06:59 |
| Air Blank          | 0.000 | 07:01 |
| Subject Sample # 2 | 0.125 | 07:01 |
| Breath Volume      | 1.394 | 07:01 |
| Air Blank          | 0.000 | 07:02 |
| Control Test       | 0.079 | 07:02 |
| Air Blank          | 0.000 | 07:03 |
| Diagnostics Check  | OK    | 07:03 |

---

Data Download Date/Time: 09/29/2020 13:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLACRES,NICHOLAS,J  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: VILLACRES,NICHOLAS  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/03/2020 00:05:05

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 23:57 |
| Air Blank         |     | 0.000 | 23:58 |
| Control Test      |     | 0.077 | 23:58 |
| Air Blank         |     | 0.000 | 23:58 |
| Subject Sample    | # 1 | 0.189 | 00:00 |
| Breath Volume     |     | 2.250 | 00:00 |
| Air Blank         |     | 0.000 | 00:01 |
| Air Blank         |     | 0.000 | 00:02 |
| Subject Sample    | # 2 | 0.183 | 00:03 |
| Breath Volume     |     | 1.621 | 00:03 |
| Air Blank         |     | 0.000 | 00:04 |
| Control Test      |     | 0.077 | 00:04 |
| Air Blank         |     | 0.000 | 00:05 |
| Diagnostics Check |     | OK    | 00:05 |

---

Data Download Date/Time: 09/30/2020 14:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLACRES,NICHOLAS,J  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: ITRI,ANGELA  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 09/03/2020 03:13:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:04 |
| Air Blank          | 0.000 | 03:04 |
| Control Test       | 0.078 | 03:05 |
| Air Blank          | 0.000 | 03:05 |
| Subject Sample # 1 | 0.072 | 03:06 |
| Breath Volume      | 1.394 | 03:06 |
| Air Blank          | 0.000 | 03:07 |
| Air Blank          | 0.000 | 03:09 |
| Subject Sample # 2 | 0.073 | 03:11 |
| Breath Volume      | 1.210 | 03:11 |
| Air Blank          | 0.000 | 03:12 |
| Control Test       | 0.078 | 03:12 |
| Air Blank          | 0.000 | 03:12 |
| Diagnostics Check  | OK    | 03:13 |

---

Data Download Date/Time: 09/30/2020 14:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEWIS,JASON,E  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: MORIARTY,RORY  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: VILCHEZ

DATE/TIME OF TEST: 09/05/2020 06:55:54

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 06:48 |
| Air Blank         |     | 0.000 | 06:49 |
| Control Test      |     | 0.078 | 06:49 |
| Air Blank         |     | 0.000 | 06:50 |
| Subject Sample    | # 1 | 0.204 | 06:50 |
| Breath Volume     |     | 1.421 | 06:50 |
| Air Blank         |     | 0.000 | 06:51 |
| Air Blank         |     | 0.000 | 06:53 |
| Subject Sample    | # 2 | 0.197 | 06:54 |
| Breath Volume     |     | 1.140 | 06:54 |
| Air Blank         |     | 0.000 | 06:54 |
| Control Test      |     | 0.078 | 06:55 |
| Air Blank         |     | 0.000 | 06:55 |
| Diagnostics Check |     | OK    | 06:55 |

---

Data Download Date/Time: 09/30/2020 14:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBIN,GASTON,A  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: LEE,NEIL  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BIERI JR

DATE/TIME OF TEST: 09/06/2020 03:17:33

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 03:11 |
| Air Blank         |     | 0.000 | 03:11 |
| Control Test      |     | 0.079 | 03:12 |
| Air Blank         |     | 0.000 | 03:12 |
| Subject Sample    | # 1 | 0.000 | 03:13 |
| Breath Volume     |     | 1.695 | 03:13 |
| Air Blank         |     | 0.000 | 03:13 |
| Air Blank         |     | 0.000 | 03:15 |
| Subject Sample    | # 2 | 0.000 | 03:16 |
| Breath Volume     |     | 1.867 | 03:16 |
| Air Blank         |     | 0.000 | 03:16 |
| Control Test      |     | 0.078 | 03:16 |
| Air Blank         |     | 0.000 | 03:17 |
| Diagnostics Check |     | OK    | 03:17 |

---

Data Download Date/Time: 09/30/2020 14:22



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KEARNEY,BRANDON,G  
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)  
ARRESTING OFFICER: HALLSTEN,CALEB  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: BUI  
ARREST TIME: 14:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 09/06/2020 16:16:08

---

|                   |       |       |                  |
|-------------------|-------|-------|------------------|
| Diagnostics Check | OK    | 16:13 |                  |
| Air Blank         | 0.000 | 16:14 |                  |
| Control Test      | 0.077 | 16:14 |                  |
| Air Blank         | 0.000 | 16:14 |                  |
| Subject Sample    | # 1   | 16:15 | Sequence Aborted |
| Breath Volume     | 0.000 | 16:15 |                  |
| Air Blank         | 0.000 | 16:16 |                  |

---

Data Download Date/Time: 09/30/2020 14:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KEARNEY, BRANDON, G  
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)  
ARRESTING OFFICER: HALLSTEN, CALEB  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: BUI  
ARREST TIME: 14:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 09/06/2020 16:24:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:17 |
| Air Blank          | 0.000 | 16:18 |
| Control Test       | 0.077 | 16:18 |
| Air Blank          | 0.000 | 16:19 |
| Subject Sample # 1 | 0.083 | 16:20 |
| Breath Volume      | 2.042 | 16:20 |
| Air Blank          | 0.000 | 16:20 |
| Air Blank          | 0.000 | 16:22 |
| Subject Sample # 2 | 0.074 | 16:22 |
| Breath Volume      | 1.230 | 16:22 |
| Air Blank          | 0.000 | 16:23 |
| Control Test       | 0.075 | 16:23 |
| Air Blank          | 0.000 | 16:24 |
| Diagnostics Check  | OK    | 16:24 |

---

Data Download Date/Time: 09/30/2020 14:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLACRES,NICHOLAS,J  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: ITRI,ANGELA  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: ATWILL

DATE/TIME OF TEST: 09/08/2020 22:35:26

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:31 |                      |
| Air Blank         | 0.000 | 22:32 |                      |
| Control Test      | 0.079 | 22:32 |                      |
| Air Blank         | 0.000 | 22:32 |                      |
| Subject Sample    | # 1   | 22:34 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:34 |                      |
| Air Blank         | 0.000 | 22:34 |                      |
| Control Test      | 0.079 | 22:34 |                      |
| Air Blank         | 0.000 | 22:35 |                      |
| Diagnostics Check | OK    | 22:35 |                      |

---

Data Download Date/Time: 09/30/2020 14:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLACRES,NICHOLAS,J  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: SMITH,HERBERT  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: CHANG

DATE/TIME OF TEST: 09/12/2020 01:52:37

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 01:38 |                |
| Air Blank         | 0.000     | 01:38 |                |
| Control Test      | 0.079     | 01:38 |                |
| Air Blank         | 0.000     | 01:39 |                |
| Subject Sample    | # 1 0.207 | 01:41 |                |
| Breath Volume     | 1.187     | 01:41 |                |
| Air Blank         | 0.000     | 01:42 |                |
| Air Blank         | 0.000     | 01:44 |                |
| Subject Sample    | # 2 0.224 | 01:47 | Volume Not Met |
| Breath Volume     | 0.074     | 01:47 |                |
| Air Blank         | 0.000     | 01:47 |                |
| Air Blank         | 0.000     | 01:49 |                |
| Subject Sample    | # 3 0.210 | 01:50 |                |
| Breath Volume     | 2.046     | 01:50 |                |
| Air Blank         | 0.000     | 01:51 |                |
| Control Test      | 0.076     | 01:52 |                |
| Air Blank         | 0.000     | 01:52 |                |
| Diagnostics Check | OK        | 01:52 |                |

---

Data Download Date/Time: 09/30/2020 14:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,CONNOR,M  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: GRAY,CONNOR  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: FAUNCE

DATE/TIME OF TEST: 09/12/2020 02:02:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:56 |
| Air Blank          | 0.000 | 01:56 |
| Control Test       | 0.077 | 01:57 |
| Air Blank          | 0.000 | 01:57 |
| Subject Sample # 1 | 0.131 | 01:58 |
| Breath Volume      | 3.578 | 01:58 |
| Air Blank          | 0.000 | 01:59 |
| Air Blank          | 0.000 | 02:00 |
| Subject Sample # 2 | 0.129 | 02:01 |
| Breath Volume      | 3.433 | 02:01 |
| Air Blank          | 0.000 | 02:02 |
| Control Test       | 0.076 | 02:02 |
| Air Blank          | 0.000 | 02:02 |
| Diagnostics Check  | OK    | 02:02 |

---

Data Download Date/Time: 09/30/2020 14:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: WINSTEAD,JUSTIN  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: ROMACK

DATE/TIME OF TEST: 09/13/2020 00:44:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:37 |
| Air Blank          | 0.000 | 00:37 |
| Control Test       | 0.078 | 00:37 |
| Air Blank          | 0.000 | 00:38 |
| Subject Sample # 1 | 0.101 | 00:39 |
| Breath Volume      | 2.832 | 00:39 |
| Air Blank          | 0.000 | 00:40 |
| Air Blank          | 0.000 | 00:42 |
| Subject Sample # 2 | 0.100 | 00:42 |
| Breath Volume      | 2.890 | 00:42 |
| Air Blank          | 0.000 | 00:43 |
| Control Test       | 0.076 | 00:43 |
| Air Blank          | 0.000 | 00:44 |
| Diagnostics Check  | OK    | 00:44 |

---

Data Download Date/Time: 09/30/2020 14:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAN, TYLER, M  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: PATTERSON, TYLER  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 09/19/2020 23:46:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:40 |
| Air Blank          | 0.000 | 23:40 |
| Control Test       | 0.079 | 23:41 |
| Air Blank          | 0.000 | 23:41 |
| Subject Sample # 1 | 0.198 | 23:42 |
| Breath Volume      | 3.957 | 23:42 |
| Air Blank          | 0.000 | 23:42 |
| Air Blank          | 0.000 | 23:44 |
| Subject Sample # 2 | 0.187 | 23:45 |
| Breath Volume      | 3.664 | 23:45 |
| Air Blank          | 0.000 | 23:45 |
| Control Test       | 0.076 | 23:46 |
| Air Blank          | 0.000 | 23:46 |
| Diagnostics Check  | OK    | 23:46 |

---

Data Download Date/Time: 09/30/2020 14:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIALOBRZESKI,M,  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: HERMAN,K  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: KUYPERS

DATE/TIME OF TEST: 09/27/2020 00:01:47

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 23:50 |                |
| Air Blank         |     | 0.000 | 23:50 |                |
| Control Test      |     | 0.078 | 23:51 |                |
| Air Blank         |     | 0.000 | 23:51 |                |
| Subject Sample    | # 2 | 0.209 | 23:54 | Volume Not Met |
| Breath Volume     |     | 0.183 | 23:54 |                |
| Air Blank         |     | 0.000 | 23:55 |                |
| Air Blank         |     | 0.000 | 23:57 |                |
| Subject Sample    | # 1 | 0.194 | 00:00 | Volume Not Met |
| Breath Volume     |     | 0.761 | 00:00 |                |
| Air Blank         |     | 0.000 | 00:00 |                |
| Control Test      |     | 0.076 | 00:01 |                |
| Air Blank         |     | 0.000 | 00:01 |                |
| Diagnostics Check |     | OK    | 00:01 |                |

---

Data Download Date/Time: 09/30/2020 14:22



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER,MARCUS,H  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: BAKER,MARCUS  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: STARKS

DATE/TIME OF TEST: 09/27/2020 02:32:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:25 |
| Air Blank          | 0.000 | 02:25 |
| Control Test       | 0.078 | 02:25 |
| Air Blank          | 0.000 | 02:26 |
| Subject Sample # 1 | 0.054 | 02:27 |
| Breath Volume      | 1.671 | 02:27 |
| Air Blank          | 0.000 | 02:28 |
| Air Blank          | 0.000 | 02:29 |
| Subject Sample # 2 | 0.048 | 02:30 |
| Breath Volume      | 1.386 | 02:30 |
| Air Blank          | 0.000 | 02:31 |
| Control Test       | 0.078 | 02:31 |
| Air Blank          | 0.000 | 02:32 |
| Diagnostics Check  | OK    | 02:32 |

---

Data Download Date/Time: 09/30/2020 14:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: UNDERWOOD

DATE/TIME OF TEST: 09/01/2020 01:19:41

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 01:12 |
| Air Blank         |     | 0.000 | 01:13 |
| Control Test      |     | 0.080 | 01:13 |
| Air Blank         |     | 0.000 | 01:14 |
| Subject Sample    | # 1 | 0.185 | 01:15 |
| Breath Volume     |     | 1.824 | 01:15 |
| Air Blank         |     | 0.000 | 01:15 |
| Air Blank         |     | 0.000 | 01:17 |
| Subject Sample    | # 2 | 0.185 | 01:18 |
| Breath Volume     |     | 1.726 | 01:18 |
| Air Blank         |     | 0.000 | 01:18 |
| Control Test      |     | 0.080 | 01:19 |
| Air Blank         |     | 0.000 | 01:19 |
| Diagnostics Check |     | OK    | 01:19 |

---

Data Download Date/Time: 10/06/2020 10:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28719080A1

DRY GAS EXPIRY DATE: 12/05/2021

VIOLATION CODE: DUI

ARREST TIME: 22:50

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: PELLERIN

DATE/TIME OF TEST: 09/14/2020 23:00:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:53 |
| Air Blank          | 0.000 | 22:54 |
| Control Test       | 0.078 | 22:54 |
| Air Blank          | 0.000 | 22:55 |
| Subject Sample # 1 | 0.077 | 22:55 |
| Breath Volume      | 2.539 | 22:55 |
| Air Blank          | 0.000 | 22:56 |
| Air Blank          | 0.000 | 22:58 |
| Subject Sample # 2 | 0.081 | 22:59 |
| Breath Volume      | 2.460 | 22:59 |
| Air Blank          | 0.000 | 22:59 |
| Control Test       | 0.079 | 23:00 |
| Air Blank          | 0.000 | 23:00 |
| Diagnostics Check  | OK    | 23:00 |

---

Data Download Date/Time: 10/06/2020 10:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:01

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 09/15/2020 18:31:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:25 |
| Air Blank          | 0.000 | 18:25 |
| Control Test       | 0.080 | 18:26 |
| Air Blank          | 0.000 | 18:26 |
| Subject Sample # 1 | 0.012 | 18:27 |
| Breath Volume      | 2.742 | 18:27 |
| Air Blank          | 0.000 | 18:27 |
| Air Blank          | 0.000 | 18:29 |
| Subject Sample # 2 | 0.011 | 18:30 |
| Breath Volume      | 2.722 | 18:30 |
| Air Blank          | 0.000 | 18:30 |
| Control Test       | 0.079 | 18:30 |
| Air Blank          | 0.000 | 18:31 |
| Diagnostics Check  | OK    | 18:31 |

---

Data Download Date/Time: 10/06/2020 10:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 09/15/2020 23:28:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:22 |
| Air Blank          | 0.000 | 23:22 |
| Control Test       | 0.081 | 23:22 |
| Air Blank          | 0.000 | 23:23 |
| Subject Sample # 1 | 0.000 | 23:23 |
| Breath Volume      | 3.093 | 23:23 |
| Air Blank          | 0.000 | 23:24 |
| Air Blank          | 0.000 | 23:26 |
| Subject Sample # 2 | 0.000 | 23:26 |
| Breath Volume      | 3.082 | 23:26 |
| Air Blank          | 0.000 | 23:27 |
| Control Test       | 0.080 | 23:27 |
| Air Blank          | 0.000 | 23:28 |
| Diagnostics Check  | OK    | 23:28 |

---

Data Download Date/Time: 10/06/2020 10:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: COSTON

DATE/TIME OF TEST: 09/16/2020 19:59:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:52 |
| Air Blank          | 0.000 | 19:53 |
| Control Test       | 0.080 | 19:53 |
| Air Blank          | 0.000 | 19:54 |
| Subject Sample # 1 | 0.000 | 19:54 |
| Breath Volume      | 1.226 | 19:54 |
| Air Blank          | 0.000 | 19:55 |
| Air Blank          | 0.000 | 19:57 |
| Subject Sample # 2 | 0.000 | 19:57 |
| Breath Volume      | 1.320 | 19:57 |
| Air Blank          | 0.000 | 19:58 |
| Control Test       | 0.080 | 19:58 |
| Air Blank          | 0.000 | 19:59 |
| Diagnostics Check  | OK    | 19:59 |

---

Data Download Date/Time: 10/06/2020 10:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28719080A1

DRY GAS EXPIRY DATE: 12/05/2021

VIOLATION CODE: DUI

ARREST TIME: 01:55

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 09/18/2020 02:19:31

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 02:06 |                |
| Air Blank         |     | 0.000 | 02:06 |                |
| Control Test      |     | 0.079 | 02:06 |                |
| Air Blank         |     | 0.000 | 02:07 |                |
| Subject Sample    | # 1 | 0.041 | 02:10 | Volume Not Met |
| Breath Volume     |     | 0.605 | 02:10 |                |
| Air Blank         |     | 0.000 | 02:11 |                |
| Air Blank         |     | 0.000 | 02:13 |                |
| Subject Sample    | # 2 | 0.055 | 02:14 |                |
| Breath Volume     |     | 1.578 | 02:14 |                |
| Air Blank         |     | 0.000 | 02:15 |                |
| Air Blank         |     | 0.000 | 02:17 |                |
| Subject Sample    | # 3 | 0.056 | 02:18 |                |
| Breath Volume     |     | 1.964 | 02:18 |                |
| Air Blank         |     | 0.000 | 02:18 |                |
| Control Test      |     | 0.080 | 02:18 |                |
| Air Blank         |     | 0.000 | 02:19 |                |
| Diagnostics Check |     | OK    | 02:19 |                |

---

Data Download Date/Time: 10/06/2020 10:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: CAPLIN,M  
ARRESTING OFFICER AGENCY: West Melbourne PD  
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: MULLINS

DATE/TIME OF TEST: 09/22/2020 00:15:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:07 |
| Air Blank          | 0.000 | 00:08 |
| Control Test       | 0.080 | 00:08 |
| Air Blank          | 0.000 | 00:08 |
| Subject Sample # 1 | 0.114 | 00:10 |
| Breath Volume      | 1.261 | 00:10 |
| Air Blank          | 0.000 | 00:11 |
| Air Blank          | 0.000 | 00:13 |
| Subject Sample # 2 | 0.117 | 00:14 |
| Breath Volume      | 1.304 | 00:14 |
| Air Blank          | 0.000 | 00:14 |
| Control Test       | 0.080 | 00:15 |
| Air Blank          | 0.000 | 00:15 |
| Diagnostics Check  | OK    | 00:15 |

---

Data Download Date/Time: 10/06/2020 10:11



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28719080A1

DRY GAS EXPIRY DATE: 12/05/2021

VIOLATION CODE: DUI

ARREST TIME: 18:29

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: BISSONNETTE

DATE/TIME OF TEST: 09/22/2020 18:42:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:36 |
| Air Blank          | 0.000 | 18:36 |
| Control Test       | 0.080 | 18:37 |
| Air Blank          | 0.000 | 18:37 |
| Subject Sample # 1 | 0.319 | 18:38 |
| Breath Volume      | 2.074 | 18:38 |
| Air Blank          | 0.000 | 18:38 |
| Air Blank          | 0.000 | 18:40 |
| Subject Sample # 2 | 0.304 | 18:41 |
| Breath Volume      | 1.445 | 18:41 |
| Air Blank          | 0.000 | 18:41 |
| Control Test       | 0.080 | 18:41 |
| Air Blank          | 0.000 | 18:42 |
| Diagnostics Check  | OK    | 18:42 |

---

Data Download Date/Time: 10/06/2020 10:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28719080A1

DRY GAS EXPIRY DATE: 12/05/2021

VIOLATION CODE: DUI

ARREST TIME: 04:43

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: SMITH-JOLLIFFE

DATE/TIME OF TEST: 09/23/2020 05:02:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:56 |
| Air Blank          | 0.000 | 04:56 |
| Control Test       | 0.081 | 04:57 |
| Air Blank          | 0.000 | 04:57 |
| Subject Sample # 1 | 0.049 | 04:57 |
| Breath Volume      | 2.640 | 04:57 |
| Air Blank          | 0.000 | 04:58 |
| Air Blank          | 0.000 | 05:00 |
| Subject Sample # 2 | 0.049 | 05:01 |
| Breath Volume      | 2.183 | 05:01 |
| Air Blank          | 0.000 | 05:01 |
| Control Test       | 0.080 | 05:01 |
| Air Blank          | 0.000 | 05:02 |
| Diagnostics Check  | OK    | 05:02 |

---

Data Download Date/Time: 10/06/2020 10:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28719080A1  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: COSTELLO

DATE/TIME OF TEST: 09/24/2020 02:23:00

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 02:16 |
| Air Blank         |     | 0.000 | 02:17 |
| Control Test      |     | 0.080 | 02:17 |
| Air Blank         |     | 0.000 | 02:18 |
| Subject Sample    | # 1 | 0.140 | 02:18 |
| Breath Volume     |     | 1.664 | 02:18 |
| Air Blank         |     | 0.000 | 02:19 |
| Air Blank         |     | 0.000 | 02:21 |
| Subject Sample    | # 2 | 0.141 | 02:21 |
| Breath Volume     |     | 1.839 | 02:21 |
| Air Blank         |     | 0.000 | 02:22 |
| Control Test      |     | 0.080 | 02:22 |
| Air Blank         |     | 0.000 | 02:22 |
| Diagnostics Check |     | OK    | 02:23 |

---

Data Download Date/Time: 10/06/2020 10:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28719080A1

DRY GAS EXPIRY DATE: 12/05/2021

VIOLATION CODE: DUI

ARREST TIME: 21:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MERCER

DATE/TIME OF TEST: 09/26/2020 22:25:07

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:21 |                      |
| Air Blank         | 0.000 | 22:21 |                      |
| Control Test      | 0.079 | 22:21 |                      |
| Air Blank         | 0.000 | 22:22 |                      |
| Subject Sample    | # 1   | 22:23 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:23 |                      |
| Air Blank         | 0.000 | 22:24 |                      |
| Control Test      | 0.080 | 22:24 |                      |
| Air Blank         | 0.000 | 22:25 |                      |
| Diagnostics Check | OK    | 22:25 |                      |

---

Data Download Date/Time: 10/06/2020 10:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28719080A1

DRY GAS EXPIRY DATE: 12/05/2021

VIOLATION CODE: ZERO TOL

ARREST TIME: 00:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: PLUMMER

DATE/TIME OF TEST: 09/27/2020 19:26:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:20 |
| Air Blank          | 0.000 | 19:20 |
| Control Test       | 0.081 | 19:20 |
| Air Blank          | 0.000 | 19:21 |
| Subject Sample # 1 | 0.035 | 19:22 |
| Breath Volume      | 1.179 | 19:22 |
| Air Blank          | 0.000 | 19:22 |
| Air Blank          | 0.000 | 19:24 |
| Subject Sample # 2 | 0.039 | 19:25 |
| Breath Volume      | 2.398 | 19:25 |
| Air Blank          | 0.000 | 19:26 |
| Control Test       | 0.080 | 19:26 |
| Air Blank          | 0.000 | 19:26 |
| Diagnostics Check  | OK    | 19:26 |

---

Data Download Date/Time: 10/06/2020 10:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28719080A1

DRY GAS EXPIRY DATE: 12/05/2021

VIOLATION CODE: DUI

ARREST TIME: 02:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: BOLDT

DATE/TIME OF TEST: 09/28/2020 02:32:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:22 |
| Air Blank          | 0.000 | 02:23 |
| Control Test       | 0.081 | 02:23 |
| Air Blank          | 0.000 | 02:24 |
| Subject Sample # 1 | 0.250 | 02:26 |
| Breath Volume      | 1.210 | 02:26 |
| Air Blank          | 0.000 | 02:26 |
| Air Blank          | 0.000 | 02:28 |
| Subject Sample # 2 | 0.234 | 02:31 |
| Breath Volume      | 1.296 | 02:31 |
| Air Blank          | 0.000 | 02:31 |
| Control Test       | 0.079 | 02:32 |
| Air Blank          | 0.000 | 02:32 |
| Diagnostics Check  | OK    | 02:32 |

---

Data Download Date/Time: 10/06/2020 10:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOUROUMOUSIS,GEORGE,T  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: MCDONNELL,BRANDON  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: KOHLHOF

DATE/TIME OF TEST: 09/08/2020 20:37:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:31 |
| Air Blank          | 0.000 | 20:31 |
| Control Test       | 0.081 | 20:32 |
| Air Blank          | 0.000 | 20:32 |
| Subject Sample # 1 | 0.098 | 20:32 |
| Breath Volume      | 2.199 | 20:32 |
| Air Blank          | 0.000 | 20:33 |
| Air Blank          | 0.000 | 20:35 |
| Subject Sample # 2 | 0.093 | 20:35 |
| Breath Volume      | 2.398 | 20:35 |
| Air Blank          | 0.000 | 20:36 |
| Control Test       | 0.079 | 20:36 |
| Air Blank          | 0.000 | 20:37 |
| Diagnostics Check  | OK    | 20:37 |

---

Data Download Date/Time: 09/28/2020 21:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PONDER,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: SIMPSON,CARL  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:38

SUBJECT'S LAST NAME: FORTUNATO

DATE/TIME OF TEST: 09/26/2020 18:05:08

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 18:02 |                      |
| Air Blank         | 0.000 | 18:02 |                      |
| Control Test      | 0.080 | 18:03 |                      |
| Air Blank         | 0.000 | 18:03 |                      |
| Subject Sample    | # 1   | 18:03 | Subject Test Refused |
| Breath Volume     | 0.000 | 18:03 |                      |
| Air Blank         | 0.000 | 18:04 |                      |
| Control Test      | 0.080 | 18:04 |                      |
| Air Blank         | 0.000 | 18:05 |                      |
| Diagnostics Check | OK    | 18:05 |                      |

---

Data Download Date/Time: 09/28/2020 21:11



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANCION,CALIX,N  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: JEUDI,M  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 09/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 09:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:33

SUBJECT'S LAST NAME: NAPOLI

DATE/TIME OF TEST: 09/28/2020 12:07:43

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 12:02 |            |
| Air Blank         | 0.000     | 12:02 |            |
| Control Test      | 0.081     | 12:02 |            |
| Air Blank         | 0.000     | 12:03 |            |
| Subject Sample    | # 1 0.000 | 12:03 |            |
| Breath Volume     | 2.781     | 12:03 |            |
| Air Blank         | 0.000     | 12:04 |            |
| Air Blank         | 0.000     | 12:06 |            |
| Subject Sample    | # 2       | 12:06 | RFI Detect |
| Breath Volume     | 0.000     | 12:06 |            |
| Air Blank         | 0.000     | 12:06 |            |
| Control Test      | 0.081     | 12:07 |            |
| Air Blank         | 0.000     | 12:07 |            |
| Diagnostics Check | OK        | 12:07 |            |

---

Data Download Date/Time: 09/28/2020 21:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANCION,CALIX,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: JEUDI,MAC  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 09/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 09:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:33

SUBJECT'S LAST NAME: NAPOLI

DATE/TIME OF TEST: 09/28/2020 12:17:01

---

|                    |       |       |            |
|--------------------|-------|-------|------------|
| Diagnostics Check  | OK    | 12:11 |            |
| Air Blank          | 0.000 | 12:11 |            |
| Control Test       | 0.081 | 12:11 |            |
| Air Blank          | 0.000 | 12:12 |            |
| Subject Sample # 1 | 0.000 | 12:12 |            |
| Breath Volume      | 2.496 | 12:12 |            |
| Air Blank          | 0.000 | 12:13 |            |
| Air Blank          | 0.000 | 12:15 |            |
| Subject Sample # 2 |       | 12:15 | RFI Detect |
| Breath Volume      | 0.000 | 12:15 |            |
| Air Blank          | 0.000 | 12:16 |            |
| Control Test       | 0.081 | 12:16 |            |
| Air Blank          | 0.000 | 12:16 |            |
| Diagnostics Check  | OK    | 12:17 |            |

---

Data Download Date/Time: 09/28/2020 21:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANCION,CALIX,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: JEUDI,MAC  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 09/27/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 915310  
DRY GAS EXPIRY DATE: 03/07/2021  
VIOLATION CODE: DUI  
ARREST TIME: 09:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:33

SUBJECT'S LAST NAME: NAPOLI

DATE/TIME OF TEST: 09/28/2020 12:28:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 12:22 |
| Air Blank          | 0.000 | 12:23 |
| Control Test       | 0.080 | 12:23 |
| Air Blank          | 0.000 | 12:23 |
| Subject Sample # 1 | 0.000 | 12:24 |
| Breath Volume      | 2.207 | 12:24 |
| Air Blank          | 0.000 | 12:24 |
| Air Blank          | 0.000 | 12:26 |
| Subject Sample # 2 | 0.000 | 12:27 |
| Breath Volume      | 2.296 | 12:27 |
| Air Blank          | 0.000 | 12:27 |
| Control Test       | 0.080 | 12:27 |
| Air Blank          | 0.000 | 12:28 |
| Diagnostics Check  | OK    | 12:28 |

---

Data Download Date/Time: 09/28/2020 21:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERGSTROM,STEPHEN,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MILLER,HOWARD  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916501  
DRY GAS EXPIRY DATE: 06/14/2021  
VIOLATION CODE: DUI  
ARREST TIME: 08:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: LAMB

DATE/TIME OF TEST: 09/07/2020 09:29:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:22 |
| Air Blank          | 0.000 | 09:23 |
| Control Test       | 0.079 | 09:23 |
| Air Blank          | 0.000 | 09:24 |
| Subject Sample # 1 | 0.169 | 09:24 |
| Breath Volume      | 2.714 | 09:24 |
| Air Blank          | 0.000 | 09:25 |
| Air Blank          | 0.000 | 09:27 |
| Subject Sample # 2 | 0.176 | 09:27 |
| Breath Volume      | 2.996 | 09:27 |
| Air Blank          | 0.000 | 09:28 |
| Control Test       | 0.078 | 09:29 |
| Air Blank          | 0.000 | 09:29 |
| Diagnostics Check  | OK    | 09:29 |

---

Data Download Date/Time: 10/16/2020 11:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CROSS,M,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MITCHELL,A  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916501  
DRY GAS EXPIRY DATE: 06/14/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: DOUDS

DATE/TIME OF TEST: 09/19/2020 05:39:52

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 05:36 |                      |
| Air Blank         | 0.000 | 05:37 |                      |
| Control Test      | 0.079 | 05:37 |                      |
| Air Blank         | 0.000 | 05:38 |                      |
| Subject Sample    | # 1   | 05:38 | Subject Test Refused |
| Breath Volume     | 0.000 | 05:38 |                      |
| Air Blank         | 0.000 | 05:38 |                      |
| Control Test      | 0.079 | 05:39 |                      |
| Air Blank         | 0.000 | 05:39 |                      |
| Diagnostics Check | OK    | 05:39 |                      |

---

Data Download Date/Time: 10/16/2020 11:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERGSTROM,STEPHEN,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MORGAN,WILLIAM  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916501  
DRY GAS EXPIRY DATE: 06/14/2021  
VIOLATION CODE: DUI  
ARREST TIME: 09:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: HEMPFLING

DATE/TIME OF TEST: 09/21/2020 10:29:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:21 |
| Air Blank          | 0.000 | 10:21 |
| Control Test       | 0.079 | 10:21 |
| Air Blank          | 0.000 | 10:22 |
| Subject Sample # 1 | 0.199 | 10:24 |
| Breath Volume      | 2.281 | 10:24 |
| Air Blank          | 0.000 | 10:25 |
| Air Blank          | 0.000 | 10:27 |
| Subject Sample # 2 | 0.203 | 10:27 |
| Breath Volume      | 1.609 | 10:27 |
| Air Blank          | 0.000 | 10:28 |
| Control Test       | 0.078 | 10:29 |
| Air Blank          | 0.000 | 10:29 |
| Diagnostics Check  | OK    | 10:29 |

---

Data Download Date/Time: 10/16/2020 11:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: OREILLY,T  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916501  
DRY GAS EXPIRY DATE: 06/14/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: DIX

DATE/TIME OF TEST: 09/10/2020 04:03:57

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 03:57 |
| Air Blank         |     | 0.000 | 03:57 |
| Control Test      |     | 0.080 | 03:57 |
| Air Blank         |     | 0.000 | 03:58 |
| Subject Sample    | # 1 | 0.195 | 03:59 |
| Breath Volume     |     | 1.750 | 03:59 |
| Air Blank         |     | 0.000 | 03:59 |
| Air Blank         |     | 0.000 | 04:01 |
| Subject Sample    | # 2 | 0.201 | 04:02 |
| Breath Volume     |     | 1.546 | 04:02 |
| Air Blank         |     | 0.000 | 04:03 |
| Control Test      |     | 0.079 | 04:03 |
| Air Blank         |     | 0.000 | 04:03 |
| Diagnostics Check |     | OK    | 04:03 |

---

Data Download Date/Time: 10/16/2020 11:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: TORO,J  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916501  
DRY GAS EXPIRY DATE: 06/14/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: SCHILLER

DATE/TIME OF TEST: 09/20/2020 04:10:20

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:07 |                      |
| Air Blank         | 0.000 | 04:08 |                      |
| Control Test      | 0.080 | 04:08 |                      |
| Air Blank         | 0.000 | 04:08 |                      |
| Subject Sample    | # 1   | 04:08 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:08 |                      |
| Air Blank         | 0.000 | 04:09 |                      |
| Control Test      | 0.080 | 04:09 |                      |
| Air Blank         | 0.000 | 04:10 |                      |
| Diagnostics Check | OK    | 04:10 |                      |

---

Data Download Date/Time: 10/16/2020 11:27



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MULLINS,J  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916501  
DRY GAS EXPIRY DATE: 06/14/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: GREEFIELD

DATE/TIME OF TEST: 09/20/2020 04:39:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:32 |
| Air Blank          | 0.000 | 04:33 |
| Control Test       | 0.080 | 04:33 |
| Air Blank          | 0.000 | 04:34 |
| Subject Sample # 1 | 0.169 | 04:34 |
| Breath Volume      | 2.695 | 04:34 |
| Air Blank          | 0.000 | 04:35 |
| Air Blank          | 0.000 | 04:37 |
| Subject Sample # 2 | 0.170 | 04:37 |
| Breath Volume      | 2.121 | 04:37 |
| Air Blank          | 0.000 | 04:38 |
| Control Test       | 0.079 | 04:38 |
| Air Blank          | 0.000 | 04:39 |
| Diagnostics Check  | OK    | 04:39 |

---

Data Download Date/Time: 10/16/2020 11:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANEY,DAMON,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BEELER,BRIAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG916501  
DRY GAS EXPIRY DATE: 06/14/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 09/25/2020 19:30:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:24 |
| Air Blank          | 0.000 | 19:24 |
| Control Test       | 0.079 | 19:25 |
| Air Blank          | 0.000 | 19:25 |
| Subject Sample # 1 | 0.023 | 19:26 |
| Breath Volume      | 2.121 | 19:26 |
| Air Blank          | 0.000 | 19:26 |
| Air Blank          | 0.000 | 19:28 |
| Subject Sample # 2 | 0.022 | 19:29 |
| Breath Volume      | 1.972 | 19:29 |
| Air Blank          | 0.000 | 19:29 |
| Control Test       | 0.077 | 19:29 |
| Air Blank          | 0.000 | 19:30 |
| Diagnostics Check  | OK    | 19:30 |

---

Data Download Date/Time: 10/16/2020 11:27