

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEXANDER,E
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: VAZQUEZ,WILLIAM
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 01/02/2021 00:51:56

Diagnostics Check	OK	00:49	
Air Blank	0.000	00:49	
Control Test	0.080	00:49	
Air Blank	0.000	00:50	
Subject Sample	# 1	00:50	Subject Test Refused
Breath Volume	0.000	00:50	
Air Blank	0.000	00:51	
Control Test	0.080	00:51	
Air Blank	0.000	00:51	
Diagnostics Check	OK	00:51	

Data Download Date/Time: 01/27/2021 16:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYA,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: HERNANDEZ,PETER
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: OCASIO

DATE/TIME OF TEST: 01/21/2021 02:13:36

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:11	
Control Test	0.080	02:11	
Air Blank	0.000	02:12	
Subject Sample	# 1	02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:12	
Control Test	0.080	02:13	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 01/27/2021 16:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEXANDER,E
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: PINO,ALEJANDRO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080AI
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: KILCREASE

DATE/TIME OF TEST: 01/23/2021 01:46:12

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.079	01:44	
Air Blank	0.000	01:44	
Subject Sample	# 1	01:44	Subject Test Refused
Breath Volume	0.000	01:44	
Air Blank	0.000	01:45	
Control Test	0.079	01:45	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 01/27/2021 16:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHEETZ,JOHN,W
BREATH TEST OPERATOR AGENCY: Calhoun County SO
ARRESTING OFFICER: SCHEETZ,JOHN
ARRESTING OFFICER AGENCY: Calhoun County SO
LAST AGENCY INSPECTION: 12/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: PROB/PAROL
ARREST TIME: 07:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: FERRELL

DATE/TIME OF TEST: 01/11/2021 08:22:06

Diagnostics Check	OK	08:16
Air Blank	0.000	08:16
Control Test	0.081	08:17
Air Blank	0.000	08:17
Subject Sample # 1	0.000	08:17
Breath Volume	2.046	08:17
Air Blank	0.000	08:18
Air Blank	0.000	08:20
Subject Sample # 2	0.000	08:20
Breath Volume	2.195	08:20
Air Blank	0.000	08:21
Control Test	0.081	08:21
Air Blank	0.000	08:22
Diagnostics Check	OK	08:22

Data Download Date/Time: 01/25/2021 09:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANO,ADRIEL,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: DELGADO,ARMANDO
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: TABARES

DATE/TIME OF TEST: 01/20/2021 20:33:20

Diagnostics Check	OK	20:30	
Air Blank	0.000	20:30	
Control Test	0.079	20:31	
Air Blank	0.000	20:31	
Subject Sample	# 1	20:31	Subject Test Refused
Breath Volume	0.000	20:31	
Air Blank	0.000	20:32	
Control Test	0.078	20:32	
Air Blank	0.000	20:33	
Diagnostics Check	OK	20:33	

Data Download Date/Time: 01/28/2021 09:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,PHILIP,R
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: SANCHEZ,ANAHI
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 12/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 10:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:18

SUBJECT'S LAST NAME: FLOREZ BUESAQUILLO

DATE/TIME OF TEST: 01/01/2021 11:17:12

Diagnostics Check	OK	11:14	
Air Blank	0.000	11:14	
Control Test	0.080	11:15	
Air Blank	0.000	11:15	
Subject Sample	# 1	11:15	Subject Test Refused
Breath Volume	0.000	11:15	
Air Blank	0.000	11:16	
Control Test	0.080	11:16	
Air Blank	0.000	11:17	
Diagnostics Check	OK	11:17	

Data Download Date/Time: 01/25/2021 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,FREDERICK,A
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: WOODSTOCK,KENDRA
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 12/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 08:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:28

SUBJECT'S LAST NAME: RAMIREZ HERNANDEZ

DATE/TIME OF TEST: 01/02/2021 09:07:49

Diagnostics Check	OK	09:04	
Air Blank	0.000	09:05	
Control Test	0.080	09:05	
Air Blank	0.000	09:06	
Subject Sample	# 1	09:06	Subject Test Refused
Breath Volume	0.000	09:06	
Air Blank	0.000	09:06	
Control Test	0.079	09:07	
Air Blank	0.000	09:07	
Diagnostics Check	OK	09:07	

Data Download Date/Time: 01/25/2021 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,MARGARITA,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: RODRIGUEZ,MONICA
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 12/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: DI SARNO

DATE/TIME OF TEST: 01/03/2021 22:58:16

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.210	22:53
Breath Volume	2.648	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.199	22:56
Breath Volume	1.460	22:56
Air Blank	0.000	22:57
Control Test	0.077	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 01/25/2021 14:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN-
 ROSQUETE,STEPHANNE,J
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: ROMERO,HANSEL
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 12/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022

VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: HUERTAS

DATE/TIME OF TEST: 01/11/2021 00:59:32

Diagnostics Check	OK	00:59	
Air Blank		00:59	RFI Detect
Air Blank		00:59	RFI Detect

Data Download Date/Time: 01/25/2021 14:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN-
ROSQUETE,STEPHANNE,J
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: ROMERO,HANSEL
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 12/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022

VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: HUERTAS

DATE/TIME OF TEST: 01/11/2021 01:10:33

Diagnostics Check		OK	01:04
Air Blank		0.000	01:04
Control Test		0.080	01:05
Air Blank		0.000	01:05
Subject Sample	# 1	0.220	01:06
Breath Volume		1.433	01:06
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.224	01:09
Breath Volume		2.472	01:09
Air Blank		0.000	01:09
Control Test		0.080	01:09
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 01/25/2021 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,MARGARITA,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: SUAREZ,ANDY
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 12/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 01/13/2021 02:17:32

Diagnostics Check		OK	02:09
Air Blank		0.000	02:09
Control Test		0.080	02:09
Air Blank		0.000	02:10
Subject Sample	# 1	0.250	02:13
Breath Volume		1.238	02:13
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.249	02:16
Breath Volume		1.402	02:16
Air Blank		0.000	02:16
Control Test		0.078	02:16
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 01/25/2021 14:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIND,ABEL,D
BREATH TEST OPERATOR AGENCY: NORTH BAY VILLAGE PD
ARRESTING OFFICER: CHERASIA,ETHAN
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD
LAST AGENCY INSPECTION: 12/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: VARGAS
DATE/TIME OF TEST: 01/16/2021 02:36:55

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:34	
Control Test	0.080	02:34	
Air Blank	0.000	02:35	
Subject Sample	# 1	02:35	Subject Test Refused
Breath Volume	0.000	02:35	
Air Blank	0.000	02:36	
Control Test	0.080	02:36	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

Data Download Date/Time: 01/25/2021 14:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMO,JORGE,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ROMO,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:01

SUBJECT'S LAST NAME: LAGUERRE

DATE/TIME OF TEST: 01/01/2021 13:37:28

Diagnostics Check	OK	13:34	
Air Blank	0.000	13:35	
Control Test	0.080	13:35	
Air Blank	0.000	13:35	
Subject Sample	# 1	13:36	Subject Test Refused
Breath Volume	0.000	13:36	
Air Blank	0.000	13:36	
Control Test	0.080	13:36	
Air Blank	0.000	13:37	
Diagnostics Check	OK	13:37	

Data Download Date/Time: 01/28/2021 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 12/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: CHANCANASTUJ

DATE/TIME OF TEST: 01/10/2021 06:27:25

Diagnostics Check	OK	06:24	
Air Blank	0.000	06:25	
Control Test	0.080	06:25	
Air Blank	0.000	06:25	
Subject Sample	# 1	06:26	Subject Test Refused
Breath Volume	0.000	06:26	
Air Blank	0.000	06:26	
Control Test	0.080	06:26	
Air Blank	0.000	06:27	
Diagnostics Check	OK	06:27	

Data Download Date/Time: 01/26/2021 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUIROGA,DIEGO,O
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: QUIROGA,DIEGO
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 12/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: DE LEON SALES

DATE/TIME OF TEST: 01/16/2021 04:04:35

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:01	
Control Test	0.079	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1 0.185	04:02	
Breath Volume	2.417	04:02	
Air Blank		04:03	Purge Fail
Air Blank	0.000	04:04	

Data Download Date/Time: 01/26/2021 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUIROGA,DIEGO,O
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: QUIROGA,DIEGO
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 12/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: DE LEON SALES

DATE/TIME OF TEST: 01/16/2021 04:12:57

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.079	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.186	04:08
Breath Volume	1.769	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:10
Subject Sample # 2	0.194	04:11
Breath Volume	1.921	04:11
Air Blank	0.000	04:12
Control Test	0.078	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 01/26/2021 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CASTILLO,CHRISTIAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: MENDEZ ORTEGA

DATE/TIME OF TEST: 01/17/2021 22:01:24

Diagnostics Check	OK	21:53
Air Blank	0.000	21:53
Control Test	0.080	21:54
Air Blank	0.000	21:54
Subject Sample # 1	0.160	21:56
Breath Volume	1.917	21:56
Air Blank	0.000	21:57
Air Blank	0.000	21:58
Subject Sample # 2	0.163	21:59
Breath Volume	1.812	21:59
Air Blank	0.000	22:00
Control Test	0.079	22:00
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 01/26/2021 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ALVAREZ,ISAIAS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: MENENDEZ

DATE/TIME OF TEST: 01/17/2021 22:25:12

Diagnostics Check	OK	22:13	
Air Blank	0.000	22:14	
Control Test	0.079	22:14	
Air Blank	0.000	22:15	
Subject Sample	# 1	22:18	No Sample Provided
Breath Volume	0.000	22:18	
Air Blank	0.000	22:18	
Air Blank	0.000	22:20	
Subject Sample	# 2	22:23	No Sample Provided
Breath Volume	0.000	22:23	
Air Blank	0.000	22:24	
Control Test	0.080	22:24	
Air Blank	0.000	22:25	
Diagnostics Check	OK	22:25	

Data Download Date/Time: 01/26/2021 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NADAL,JOSE,V
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MUNOZ,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: NICHLOSON

DATE/TIME OF TEST: 01/21/2021 12:49:44

Diagnostics Check	OK	12:46	
Air Blank	0.000	12:47	
Control Test	0.079	12:47	
Air Blank	0.000	12:48	
Subject Sample	# 1	12:48	Subject Test Refused
Breath Volume	0.000	12:48	
Air Blank	0.000	12:48	
Control Test	0.080	12:49	
Air Blank	0.000	12:49	
Diagnostics Check	OK	12:49	

Data Download Date/Time: 01/26/2021 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,DARREN,B
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DIAZ,DARREN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34917080A4
DRY GAS EXPIRY DATE: 02/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 01/30/2021 19:51:01

Diagnostics Check	OK	19:43
Air Blank	0.000	19:43
Control Test	0.080	19:44
Air Blank	0.000	19:44
Subject Sample # 1	0.000	19:46
Breath Volume	3.164	19:46
Air Blank	0.000	19:46
Air Blank	0.000	19:48
Subject Sample # 2	0.000	19:49
Breath Volume	3.425	19:49
Air Blank	0.000	19:50
Control Test	0.081	19:50
Air Blank	0.000	19:50
Diagnostics Check	OK	19:51

Data Download Date/Time: 02/20/2021 01:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ALEX,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: MELCHER,ANDREW
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/21/2021
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: ORTEGA-RESENDIZ

DATE/TIME OF TEST: 01/09/2021 19:43:15

Diagnostics Check	OK	19:35
Air Blank	0.000	19:35
Control Test	0.082	19:36
Air Blank	0.000	19:36
Subject Sample # 1	0.153	19:38
Breath Volume	2.710	19:38
Air Blank	0.000	19:39
Air Blank	0.000	19:41
Subject Sample # 2	0.147	19:41
Breath Volume	2.585	19:41
Air Blank	0.000	19:42
Control Test	0.081	19:42
Air Blank	0.000	19:43
Diagnostics Check	OK	19:43

Data Download Date/Time: 01/25/2021 09:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICKEY, WILLIAM, Q
BREATH TEST OPERATOR AGENCY: PORT ST. JOE PD
ARRESTING OFFICER: YOUNG, P
ARRESTING OFFICER AGENCY: GULF COUNTY SO
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: AMISON

DATE/TIME OF TEST: 01/18/2021 18:32:02

Diagnostics Check	OK	18:29	
Air Blank	0.000	18:29	
Control Test	0.079	18:29	
Air Blank	0.000	18:30	
Subject Sample	# 1	18:30	Subject Test Refused
Breath Volume	0.000	18:30	
Air Blank	0.000	18:31	
Control Test	0.080	18:31	
Air Blank	0.000	18:31	
Diagnostics Check	OK	18:32	

Data Download Date/Time: 02/01/2021 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,PAUL,G
BREATH TEST OPERATOR AGENCY: GULF COUNTY SO
ARRESTING OFFICER: NICHOLS,RICHARD
ARRESTING OFFICER AGENCY: PORT ST. JOE PD
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 01/20/2021 05:26:58

Diagnostics Check	OK	05:20
Air Blank	0.000	05:21
Control Test	0.080	05:21
Air Blank	0.000	05:22
Subject Sample # 1	0.009	05:22
Breath Volume	2.371	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.008	05:25
Breath Volume	2.277	05:25
Air Blank	0.000	05:26
Control Test	0.080	05:26
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 02/01/2021 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DARIEN,
BREATH TEST OPERATOR AGENCY: South Miami PD
ARRESTING OFFICER: RODRIGUEZ,DARIEN
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 01/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:57

SUBJECT'S LAST NAME: BULL

DATE/TIME OF TEST: 01/27/2021 08:25:16

Diagnostics Check	OK	08:19
Air Blank	0.000	08:19
Control Test	0.080	08:19
Air Blank	0.000	08:20
Subject Sample # 1	0.000	08:20
Breath Volume	2.394	08:20
Air Blank	0.000	08:21
Air Blank	0.000	08:23
Subject Sample # 2	0.000	08:23
Breath Volume	2.453	08:23
Air Blank	0.000	08:24
Control Test	0.080	08:24
Air Blank	0.000	08:25
Diagnostics Check	OK	08:25

Data Download Date/Time: 02/16/2021 19:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEMEN,LOGAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEMEN,LOGAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MARQUEZ GONZALEZ

DATE/TIME OF TEST: 01/13/2021 00:23:33

Diagnostics Check	OK	00:13
Air Blank	0.000	00:14
Control Test	0.079	00:14
Air Blank	0.000	00:14
Subject Sample # 1	0.171	00:17
Breath Volume	1.476	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:19
Subject Sample # 2	0.162	00:21
Breath Volume	1.707	00:21
Air Blank	0.000	00:22
Control Test	0.077	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 02/09/2021 13:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD	Date or Time Changed: Yes
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 07220080A3
BREATH TEST OPERATOR: LEMEN,LOGAN,C	DRY GAS EXPIRY DATE: 05/05/2022
BREATH TEST OPERATOR AGENCY: FHP Troop F	VIOLATION CODE: DUI
ARRESTING OFFICER: LEMEN,LOGAN	ARREST TIME: 23:01
ARRESTING OFFICER AGENCY: FHP Troop F	GENDER: Male
LAST AGENCY INSPECTION: 01/11/2021	OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MARQUEZ GONZALEZ DATE/TIME OF TEST: 01/15/2021 00:32:20

Diagnostics Check	OK	00:22	
Air Blank	0.000	00:22	
Control Test	0.078	00:22	
Air Blank	0.000	00:23	
Subject Sample	# 1 0.135	00:26	Volume Not Met
Breath Volume	0.285	00:26	
Air Blank	0.000	00:27	
Air Blank	0.000	00:28	
Subject Sample	# 2	00:30	Subject Test Refused
Breath Volume	0.214	00:30	
Air Blank	0.000	00:31	
Control Test	0.078	00:31	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 02/09/2021 13:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERO ORTIZ,JOHAN,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BARRERO ORTIZ,JOHAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: UECKER

DATE/TIME OF TEST: 01/15/2021 06:05:10

Diagnostics Check	OK	05:58
Air Blank	0.000	05:58
Control Test	0.078	05:58
Air Blank	0.000	05:59
Subject Sample # 1	0.144	06:00
Breath Volume	1.753	06:00
Air Blank	0.000	06:01
Air Blank	0.000	06:02
Subject Sample # 2	0.145	06:03
Breath Volume	1.789	06:03
Air Blank	0.000	06:04
Control Test	0.077	06:04
Air Blank	0.000	06:05
Diagnostics Check	OK	06:05

Data Download Date/Time: 02/09/2021 13:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT,KELLY,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: RODRIGUEZ,LAYNOR
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 01/17/2021 19:52:25

Diagnostics Check	OK	19:45
Air Blank	0.000	19:46
Control Test	0.079	19:46
Air Blank	0.000	19:47
Subject Sample # 1	0.310	19:47
Breath Volume	2.292	19:47
Air Blank	0.000	19:48
Air Blank	0.000	19:50
Subject Sample # 2	0.325	19:50
Breath Volume	2.503	19:50
Air Blank	0.000	19:51
Control Test	0.079	19:51
Air Blank	0.000	19:52
Diagnostics Check	OK	19:52

Data Download Date/Time: 02/09/2021 13:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERONI, KRISTIA
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: BRADFORD, HARLEE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: ULLRICH

DATE/TIME OF TEST: 01/20/2021 15:56:34

Diagnostics Check	OK	15:49
Air Blank	0.000	15:49
Control Test	0.079	15:49
Air Blank	0.000	15:50
Subject Sample # 1	0.264	15:51
Breath Volume	4.867	15:51
Air Blank	0.000	15:52
Air Blank	0.000	15:53
Subject Sample # 2	0.271	15:55
Breath Volume	5.203	15:55
Air Blank	0.000	15:55
Control Test	0.078	15:55
Air Blank	0.000	15:56
Diagnostics Check	OK	15:56

Data Download Date/Time: 02/09/2021 13:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT,KELLY,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: RODRIGUEZ,LAYNOR
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: ALVAREZ GONZALEZ

DATE/TIME OF TEST: 01/20/2021 20:24:06

Diagnostics Check	OK	20:17
Air Blank	0.000	20:18
Control Test	0.080	20:18
Air Blank	0.000	20:18
Subject Sample # 1	0.193	20:19
Breath Volume	1.847	20:19
Air Blank	0.000	20:20
Air Blank	0.000	20:22
Subject Sample # 2	0.199	20:22
Breath Volume	2.324	20:22
Air Blank	0.000	20:23
Control Test	0.079	20:23
Air Blank	0.000	20:24
Diagnostics Check	OK	20:24

Data Download Date/Time: 02/09/2021 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,DANIEL,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: JOHNSTON,TORI
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 01/21/2021 20:15:44

Diagnostics Check	OK	20:09
Air Blank	0.000	20:09
Control Test	0.079	20:09
Air Blank	0.000	20:10
Subject Sample # 1	0.116	20:11
Breath Volume	3.656	20:11
Air Blank	0.000	20:11
Air Blank	0.000	20:13
Subject Sample # 2	0.114	20:14
Breath Volume	3.328	20:14
Air Blank	0.000	20:14
Control Test	0.079	20:15
Air Blank	0.000	20:15
Diagnostics Check	OK	20:15

Data Download Date/Time: 02/09/2021 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: KREBS

DATE/TIME OF TEST: 01/26/2021 00:03:25

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.079	23:58
Air Blank	0.000	23:58
Subject Sample # 2	0.205	23:58
Breath Volume	2.160	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.193	00:01
Breath Volume	1.429	00:01
Air Blank	0.000	00:02
Control Test	0.079	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 02/09/2021 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIED 18096,A,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CANFIELD 18096,A
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:49

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 01/28/2021 11:22:14

Diagnostics Check	OK	11:16
Air Blank	0.000	11:16
Control Test	0.080	11:16
Air Blank	0.000	11:17
Subject Sample # 1	0.036	11:17
Breath Volume	1.535	11:17
Air Blank	0.000	11:18
Air Blank	0.000	11:20
Subject Sample # 2	0.034	11:20
Breath Volume	1.851	11:20
Air Blank	0.000	11:21
Control Test	0.079	11:21
Air Blank	0.000	11:22
Diagnostics Check	OK	11:22

Data Download Date/Time: 02/09/2021 13:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ADAMS,JAMES
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: VALASQUEZ TROCHEZ

DATE/TIME OF TEST: 01/30/2021 03:39:14

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:24	
Control Test	0.080	03:24	
Air Blank	0.000	03:25	
Subject Sample	# 1 0.197	03:26	
Breath Volume	1.183	03:26	
Air Blank	0.000	03:27	
Air Blank	0.000	03:28	
Subject Sample	# 2 0.153	03:32	Volume Not Met
Breath Volume	0.566	03:32	
Air Blank	0.000	03:32	
Air Blank	0.000	03:34	
Subject Sample	# 3 0.175	03:37	Volume Not Met
Breath Volume	0.109	03:37	
Air Blank	0.000	03:38	
Control Test	0.078	03:38	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 02/09/2021 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTERIO,CHRIS,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: KIRK,JUSTIN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: BIRCH

DATE/TIME OF TEST: 01/08/2021 21:42:45

Diagnostics Check		OK	21:31	
Air Blank		0.000	21:31	
Control Test		0.080	21:32	
Air Blank		0.000	21:32	
Subject Sample	# 1	0.132	21:35	Volume Not Met
Breath Volume		0.148	21:35	
Air Blank		0.000	21:36	
Air Blank		0.000	21:38	
Subject Sample	# 2	0.129	21:41	Volume Not Met
Breath Volume		0.742	21:41	
Air Blank		0.000	21:41	
Control Test		0.081	21:42	
Air Blank		0.000	21:42	
Diagnostics Check		OK	21:42	

Data Download Date/Time: 02/24/2021 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTERIO,CHRIS,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: KIRK,JUSTIN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: BIRCH

DATE/TIME OF TEST: 01/08/2021 21:55:58

Diagnostics Check	OK	21:47	
Air Blank	0.000	21:47	
Control Test	0.079	21:48	
Air Blank	0.000	21:48	
Subject Sample	# 1 0.128	21:51	Volume Not Met
Breath Volume	0.371	21:51	
Air Blank	0.000	21:52	
Air Blank	0.000	21:54	
Subject Sample	# 2	21:54	Subject Test Refused
Breath Volume	0.000	21:54	
Air Blank	0.000	21:55	
Control Test	0.078	21:55	
Air Blank	0.000	21:55	
Diagnostics Check	OK	21:55	

Data Download Date/Time: 02/24/2021 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GASH,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: CARRILLO LEIVA

DATE/TIME OF TEST: 01/01/2021 01:06:41

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.082	00:58
Air Blank		0.000	00:58
Subject Sample	# 1	0.216	01:00
Breath Volume		2.703	01:00
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.210	01:05
Breath Volume		2.292	01:05
Air Blank		0.000	01:05
Control Test		0.080	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SACKMANN,A
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 01/01/2021 01:41:25

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.087	01:36
Breath Volume	2.433	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.084	01:39
Breath Volume	1.953	01:39
Air Blank	0.000	01:40
Control Test	0.081	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,S,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PEREZ,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: ABENE

DATE/TIME OF TEST: 01/03/2021 05:09:48

Diagnostics Check		OK	05:00
Air Blank		0.000	05:01
Control Test		0.082	05:01
Air Blank		0.000	05:01
Subject Sample	# 1	0.226	05:04
Breath Volume		1.273	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:06
Subject Sample	# 2	0.221	05:08
Breath Volume		1.570	05:08
Air Blank		0.000	05:08
Control Test		0.080	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,S,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ACCIME,L
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/06/2021 02:28:16

Diagnostics Check		OK	02:21
Air Blank		0.000	02:22
Control Test		0.082	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.118	02:23
Breath Volume		1.609	02:23
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.119	02:26
Breath Volume		1.621	02:26
Air Blank		0.000	02:27
Control Test		0.082	02:27
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELIAS,D,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ELIAS,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: MCSHEFFERY

DATE/TIME OF TEST: 01/08/2021 01:37:17

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.082	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.208	01:32
Breath Volume	1.914	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.198	01:35
Breath Volume	2.199	01:35
Air Blank	0.000	01:36
Control Test	0.082	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,SHANNON,N
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SELLER,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: RAMTAHAL

DATE/TIME OF TEST: 01/08/2021 01:50:10

Diagnostics Check	OK	01:42
Air Blank	0.000	01:43
Control Test	0.082	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.179	01:45
Breath Volume	1.183	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.179	01:48
Breath Volume	1.171	01:48
Air Blank	0.000	01:49
Control Test	0.081	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILLIARD,MARTINA,F
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: JEAN,P
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: BUI
ARREST TIME: 07:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:08

SUBJECT'S LAST NAME: BATISTA

DATE/TIME OF TEST: 01/08/2021 09:45:12

Diagnostics Check	OK	09:38	
Air Blank	0.000	09:39	
Control Test	0.081	09:39	
Air Blank	0.000	09:40	
Subject Sample	# 1 0.094	09:40	
Breath Volume	2.199	09:40	
Air Blank	0.000	09:41	
Air Blank	0.000	09:43	
Subject Sample	# 2 0.079	09:43	
Breath Volume	2.074	09:43	
Air Blank	0.000	09:44	
Control Test	0.069	09:44	Control Outside Tolerance
Air Blank	0.000	09:45	

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBBONS,S,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CAREY,E
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: KESLER

DATE/TIME OF TEST: 01/10/2021 03:43:52

Diagnostics Check		OK	03:31	
Air Blank		0.000	03:31	
Control Test		0.082	03:32	
Air Blank		0.000	03:32	
Subject Sample	# 1	0.129	03:35	Volume Not Met
Breath Volume		0.707	03:35	
Air Blank		0.000	03:36	
Air Blank		0.000	03:38	
Subject Sample	# 2	0.146	03:38	
Breath Volume		1.878	03:38	
Air Blank		0.000	03:39	
Air Blank		0.000	03:41	
Subject Sample	# 3	0.150	03:42	
Breath Volume		2.082	03:42	
Air Blank		0.000	03:42	
Control Test		0.082	03:43	
Air Blank		0.000	03:43	
Diagnostics Check		OK	03:43	

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: CHRISTOPHER

DATE/TIME OF TEST: 01/11/2021 20:07:51

Diagnostics Check	OK	20:01
Air Blank	0.000	20:01
Control Test	0.081	20:02
Air Blank	0.000	20:02
Subject Sample # 1	0.249	20:03
Breath Volume	2.804	20:03
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.241	20:06
Breath Volume	2.265	20:06
Air Blank	0.000	20:06
Control Test	0.081	20:07
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GASH,C,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GASH,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 01/11/2021 21:42:25

Diagnostics Check	OK	21:33
Air Blank	0.000	21:34
Control Test	0.082	21:34
Air Blank	0.000	21:35
Subject Sample # 1	0.233	21:36
Breath Volume	1.484	21:36
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.229	21:40
Breath Volume	1.671	21:40
Air Blank	0.000	21:41
Control Test	0.081	21:41
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER JR,SCOTT,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MOLCAN,M
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: BURKE IV

DATE/TIME OF TEST: 01/12/2021 20:38:28

Diagnostics Check	OK	20:31
Air Blank	0.000	20:31
Control Test	0.081	20:32
Air Blank	0.000	20:32
Subject Sample # 1	0.228	20:33
Breath Volume	3.648	20:33
Air Blank	0.000	20:34
Air Blank	0.000	20:36
Subject Sample # 2	0.235	20:36
Breath Volume	3.371	20:36
Air Blank	0.000	20:37
Control Test	0.080	20:37
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GASH,C,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GASH,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: ARGUETA-YANES

DATE/TIME OF TEST: 01/12/2021 23:51:10

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.082	23:45
Air Blank	0.000	23:46
Subject Sample # 1	0.057	23:46
Breath Volume	4.964	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.054	23:49
Breath Volume	4.324	23:49
Air Blank	0.000	23:50
Control Test	0.081	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GASH,C,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ROWLEY,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: BURPEE

DATE/TIME OF TEST: 01/13/2021 00:03:24

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.082	23:57
Air Blank	0.000	23:58
Subject Sample # 2	0.163	23:58
Breath Volume	3.824	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.164	00:01
Breath Volume	3.960	00:01
Air Blank	0.000	00:02
Control Test	0.081	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ZACHARY,H
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WILLIAMS,ZACHARY
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: BARANCYK

DATE/TIME OF TEST: 01/14/2021 02:06:44

Diagnostics Check	OK	01:59
Air Blank	0.000	02:00
Control Test	0.081	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.135	02:02
Breath Volume	1.847	02:02
Air Blank	0.000	02:03
Air Blank	0.000	02:04
Subject Sample # 2	0.153	02:05
Breath Volume	1.695	02:05
Air Blank	0.000	02:05
Control Test	0.079	02:06
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THROESCH,PATRICK,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CAREY,E
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: DESAMOUR

DATE/TIME OF TEST: 01/14/2021 04:41:20

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.081	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.247	04:36
Breath Volume	2.671	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.250	04:39
Breath Volume	3.179	04:39
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,E,W
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LEONI,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: ROMEROHERNANDEZ

DATE/TIME OF TEST: 01/15/2021 02:13:14

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:11	
Control Test	0.081	02:12	
Air Blank		02:12	RFI Detect
Air Blank	0.000	02:13	

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,E,W
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LEONI,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: ROMEROHERNANDES

DATE/TIME OF TEST: 01/15/2021 02:26:33

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.081	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.138	02:21
Breath Volume	1.578	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.139	02:25
Breath Volume	2.328	02:25
Air Blank	0.000	02:25
Control Test	0.082	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLMSTEAD,J,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PEREZ,M
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: KARVINEN

DATE/TIME OF TEST: 01/15/2021 03:19:56

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.082	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.112	03:15
Breath Volume	1.355	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:17
Subject Sample # 2	0.111	03:18
Breath Volume	1.460	03:18
Air Blank	0.000	03:19
Control Test	0.081	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: VAUGHAN,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: DRUCKMAN

DATE/TIME OF TEST: 01/16/2021 01:54:30

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.081	01:49
Air Blank	0.000	01:49
Subject Sample # 1	0.119	01:50
Breath Volume	2.980	01:50
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.116	01:52
Breath Volume	2.085	01:52
Air Blank	0.000	01:53
Control Test	0.082	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOEL,Y,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ALUSMA,J
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/16/2021 04:50:42

Diagnostics Check		OK	04:38
Air Blank		0.000	04:39
Control Test		0.081	04:39
Air Blank		0.000	04:39
Subject Sample	# 1	0.191	04:41
Breath Volume		1.394	04:41
Air Blank		0.000	04:42
Air Blank		0.000	04:43
Subject Sample	# 2	0.224	04:44
Breath Volume		2.281	04:44
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 3	0.207	04:49
Breath Volume		2.250	04:49
Air Blank		0.000	04:49
Control Test		0.083	04:50
Air Blank		0.000	04:50
Diagnostics Check		OK	04:50

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATSON,ANDRAE,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ENDELMANN,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: DRABAVICIUS

DATE/TIME OF TEST: 01/17/2021 02:02:29

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.082	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.180	01:58
Breath Volume	1.304	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.170	02:00
Breath Volume	1.710	02:00
Air Blank	0.000	02:01
Control Test	0.081	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DEAR,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ABDELAZIZ

DATE/TIME OF TEST: 01/17/2021 03:13:44

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.081	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.179	03:08
Breath Volume		2.015	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.183	03:12
Breath Volume		2.199	03:12
Air Blank		0.000	03:12
Control Test		0.081	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,S,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DEAR,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/18/2021 01:45:28

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.081	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.135	01:39
Breath Volume	2.585	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.127	01:43
Breath Volume	1.656	01:43
Air Blank	0.000	01:44
Control Test	0.081	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GASH,C,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GASH,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: TOWNSEND

DATE/TIME OF TEST: 01/18/2021 20:01:27

Diagnostics Check	OK	19:55
Air Blank	0.000	19:55
Control Test	0.082	19:55
Air Blank	0.000	19:56
Subject Sample # 1	0.103	19:56
Breath Volume	2.421	19:56
Air Blank	0.000	19:57
Air Blank	0.000	19:59
Subject Sample # 2	0.099	19:59
Breath Volume	2.800	19:59
Air Blank	0.000	20:00
Control Test	0.082	20:00
Air Blank	0.000	20:01
Diagnostics Check	OK	20:01

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNORAT,ADELY,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KELLY,M
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: BOWMAN

DATE/TIME OF TEST: 01/18/2021 20:48:11

Diagnostics Check	OK	20:45	
Air Blank	0.000	20:45	
Control Test	0.081	20:45	
Air Blank	0.000	20:46	
Subject Sample	# 1	20:46	RFI Detect
Breath Volume	0.082	20:46	
Air Blank	0.000	20:47	
Control Test	0.081	20:47	
Air Blank	0.000	20:48	
Diagnostics Check	OK	20:48	

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNORAT,ADELY,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KELLY,M
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: BOWMAN

DATE/TIME OF TEST: 01/18/2021 21:02:33

Diagnostics Check		OK	20:51	
Air Blank		0.000	20:51	
Control Test		0.082	20:51	
Air Blank		0.000	20:52	
Subject Sample	# 1	0.155	20:55	Volume Not Met
Breath Volume		0.128	20:55	
Air Blank		0.000	20:56	
Air Blank		0.000	20:57	
Subject Sample	# 2	0.134	21:01	Volume Not Met
Breath Volume		0.128	21:01	
Air Blank		0.000	21:01	
Control Test		0.080	21:01	
Air Blank		0.000	21:02	
Diagnostics Check		OK	21:02	

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,S,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GROSSKLAS,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: EMICK

DATE/TIME OF TEST: 01/19/2021 23:00:27

Diagnostics Check	OK	22:54
Air Blank	0.000	22:54
Control Test	0.082	22:54
Air Blank	0.000	22:55
Subject Sample # 1	0.202	22:55
Breath Volume	1.554	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.214	22:58
Breath Volume	1.210	22:58
Air Blank	0.000	22:59
Control Test	0.081	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THROESCH,PATRICK,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: TERUZ,G
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: QUILES

DATE/TIME OF TEST: 01/20/2021 03:50:45

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.082	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.160	03:46
Breath Volume	3.812	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:48
Subject Sample # 2	0.157	03:49
Breath Volume	3.855	03:49
Air Blank	0.000	03:49
Control Test	0.080	03:50
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 01/27/2021 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER JR,SCOTT,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CHAPMAN,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: MINER

DATE/TIME OF TEST: 01/20/2021 19:38:05

Diagnostics Check	OK	19:32	
Air Blank	0.000	19:33	
Control Test	0.081	19:33	
Air Blank	0.000	19:34	
Subject Sample	# 1	19:36	Subject Test Refused
Breath Volume	0.000	19:36	
Air Blank	0.000	19:37	
Control Test	0.081	19:37	
Air Blank	0.000	19:37	
Diagnostics Check	OK	19:38	

Data Download Date/Time: 02/04/2021 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER JR,SCOTT,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CHAPMAN,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: MINER

DATE/TIME OF TEST: 01/20/2021 19:49:32

Diagnostics Check	OK	19:41
Air Blank	0.000	19:41
Control Test	0.081	19:42
Air Blank	0.000	19:42
Subject Sample # 1	0.196	19:44
Breath Volume	2.875	19:44
Air Blank	0.000	19:45
Air Blank	0.000	19:47
Subject Sample # 2	0.187	19:47
Breath Volume	2.257	19:47
Air Blank	0.000	19:48
Control Test	0.079	19:48
Air Blank	0.000	19:49
Diagnostics Check	OK	19:49

Data Download Date/Time: 02/04/2021 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LECLERC,A,S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: VAUGHAN,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: KREAMER

DATE/TIME OF TEST: 01/22/2021 04:37:04

Diagnostics Check	OK	04:30
Air Blank	0.000	04:30
Control Test	0.081	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.157	04:32
Breath Volume	2.164	04:32
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.157	04:35
Breath Volume	2.375	04:35
Air Blank	0.000	04:36
Control Test	0.081	04:36
Air Blank	0.000	04:36
Diagnostics Check	OK	04:37

Data Download Date/Time: 02/04/2021 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PEREZ, M
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: HARDEN

DATE/TIME OF TEST: 01/23/2021 02:52:42

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.081	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.199	02:48
Breath Volume	3.480	02:48
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.210	02:51
Breath Volume	3.593	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 02/04/2021 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,M,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GROSSKLAS,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: HERNANDEZ HERNANDEZ

DATE/TIME OF TEST: 01/23/2021 04:18:13

Diagnostics Check	OK	04:11
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:13
Subject Sample # 1	0.173	04:13
Breath Volume	1.320	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.171	04:16
Breath Volume	1.117	04:16
Air Blank	0.000	04:17
Control Test	0.081	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 02/04/2021 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,THOMAS,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DEAR,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 01/26/2021 02:11:41

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:00	
Control Test	0.081	02:01	
Air Blank	0.000	02:01	
Subject Sample	# 1	02:04	No Sample Provided
Breath Volume	0.000	02:04	
Air Blank	0.000	02:05	
Air Blank	0.000	02:07	
Subject Sample	# 2	02:10	No Sample Provided
Breath Volume	0.000	02:10	
Air Blank	0.000	02:10	
Control Test	0.080	02:11	
Air Blank	0.000	02:11	
Diagnostics Check	OK	02:11	

Data Download Date/Time: 02/04/2021 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: TERCERO,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: MARIANI

DATE/TIME OF TEST: 01/26/2021 03:03:33

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.081	02:58
Air Blank	0.000	02:58
Subject Sample # 1	0.041	02:59
Breath Volume	2.878	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.038	03:02
Breath Volume	2.324	03:02
Air Blank	0.000	03:02
Control Test	0.081	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 02/04/2021 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CASELLI,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: WOOSTER

DATE/TIME OF TEST: 01/30/2021 02:21:26

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:18	
Control Test	0.081	02:19	
Air Blank	0.000	02:19	
Subject Sample	# 1	02:20	Subject Test Refused
Breath Volume	0.000	02:20	
Air Blank	0.000	02:20	
Control Test	0.081	02:20	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time: 02/04/2021 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,S,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ENDELMANN,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: BERGT

DATE/TIME OF TEST: 01/31/2021 01:53:48

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:50	
Control Test	0.081	01:51	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:53	Sequence Aborted
Breath Volume	0.000	01:53	
Air Blank	0.000	01:53	

Data Download Date/Time: 02/04/2021 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKS,D,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PARKS,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: MIRABALRUIZ

DATE/TIME OF TEST: 01/31/2021 22:04:35

Diagnostics Check	OK	21:58	
Air Blank	0.000	21:59	
Control Test	0.081	21:59	
Air Blank	0.000	21:59	
Subject Sample # 1	0.162	22:00	
Breath Volume	2.472	22:00	
Air Blank	0.000	22:01	
Air Blank	0.000	22:03	
Subject Sample # 2	0.156	22:03	
Breath Volume	2.070	22:03	
Air Blank		22:04	RFI Detect
Air Blank	0.000	22:04	

Data Download Date/Time: 02/04/2021 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACEBO,ANDRES,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JANKOWSKI,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 01/17/2021 00:19:14

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.080	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.123	00:14
Breath Volume	1.828	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.120	00:17
Breath Volume	2.230	00:17
Air Blank	0.000	00:18
Control Test	0.080	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 01/22/2021 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EPPS,QUINTELLA,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: WILLIAMS,CHARLES
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 12/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: CANALES

DATE/TIME OF TEST: 01/01/2021 00:24:38

Diagnostics Check	OK	00:17
Air Blank	0.000	00:17
Control Test	0.082	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.073	00:20
Breath Volume	1.105	00:20
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.076	00:23
Breath Volume	1.804	00:23
Air Blank	0.000	00:23
Control Test	0.079	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 02/02/2021 15:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EPPS,QUINTELLA,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: LOIS,MARCUS
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 01/07/2021 20:35:56

Diagnostics Check	OK	20:30	
Air Blank	0.000	20:30	
Control Test	0.081	20:31	
Air Blank	0.000	20:31	
Subject Sample	# 1	20:34	No Sample Provided
Breath Volume	0.000	20:34	
Air Blank		20:35	Ambient Fail
Air Blank	0.000	20:35	

Data Download Date/Time: 02/02/2021 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EPPS,QUINTELLA,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: LOIS,MARCUS
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 01/07/2021 20:47:21

Diagnostics Check	OK	20:41
Air Blank	0.000	20:41
Control Test	0.081	20:41
Air Blank	0.000	20:42
Subject Sample # 1	0.213	20:42
Breath Volume	2.945	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.194	20:45
Breath Volume	2.066	20:45
Air Blank	0.000	20:46
Control Test	0.079	20:46
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 02/02/2021 16:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EPPS,QUINTELLA,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: LOIS,MARCUS
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BETHEL

DATE/TIME OF TEST: 01/07/2021 23:38:44

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.081	23:30
Air Blank	0.000	23:31
Subject Sample # 1	0.203	23:33
Breath Volume	1.679	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.213	23:36
Breath Volume	1.757	23:36
Air Blank	0.000	23:37
Control Test	0.076	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 02/02/2021 16:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA, JOSUE, D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: LOIS, MARCUS
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 01/08/2021 08:16:34

Diagnostics Check	OK	08:10
Air Blank	0.000	08:11
Control Test	0.080	08:11
Air Blank	0.000	08:12
Subject Sample # 1	0.000	08:12
Breath Volume	2.570	08:12
Air Blank	0.000	08:12
Air Blank	0.000	08:14
Subject Sample # 2	0.000	08:15
Breath Volume	1.867	08:15
Air Blank	0.000	08:15
Control Test	0.081	08:15
Air Blank	0.000	08:16
Diagnostics Check	OK	08:16

Data Download Date/Time: 02/02/2021 16:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA, JOSUE, D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: SEMPLE, W
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: BETHEL

DATE/TIME OF TEST: 01/08/2021 08:45:25

Diagnostics Check		OK	08:38
Air Blank		0.000	08:39
Control Test		0.080	08:39
Air Blank		0.000	08:39
Subject Sample	# 1	0.038	08:40
Breath Volume		1.410	08:40
Air Blank		0.000	08:41
Air Blank		0.000	08:42
Subject Sample	# 2	0.039	08:43
Breath Volume		1.832	08:43
Air Blank		0.000	08:44
Control Test		0.080	08:44
Air Blank		0.000	08:45
Diagnostics Check		OK	08:45

Data Download Date/Time: 02/02/2021 16:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA, JOSUE, D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: SAMPLE, W
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: BETHEL

DATE/TIME OF TEST: 01/08/2021 10:45:51

Diagnostics Check	OK	10:39
Air Blank	0.000	10:40
Control Test	0.081	10:40
Air Blank	0.000	10:41
Subject Sample # 1	0.008	10:41
Breath Volume	1.269	10:41
Air Blank	0.000	10:42
Air Blank	0.000	10:44
Subject Sample # 2	0.007	10:44
Breath Volume	1.683	10:44
Air Blank	0.000	10:44
Control Test	0.080	10:45
Air Blank	0.000	10:45
Diagnostics Check	OK	10:45

Data Download Date/Time: 02/02/2021 16:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVINGSTON,JESSICA,L
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: LOIS,M
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ARIAS

DATE/TIME OF TEST: 01/17/2021 03:34:20

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:30	
Control Test	0.082	03:30	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:33	
Control Test	0.081	03:33	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 02/02/2021 16:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,GENE,A
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: SMITH,LARRY
ARRESTING OFFICER AGENCY: Apalachicola PD
LAST AGENCY INSPECTION: 12/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: GREENFIELD

DATE/TIME OF TEST: 01/02/2021 22:28:21

Diagnostics Check	OK	22:18	
Air Blank	0.000	22:19	
Control Test	0.078	22:19	
Air Blank	0.000	22:19	
Subject Sample	# 1 0.332	22:20	
Breath Volume	1.906	22:20	
Air Blank	0.000	22:21	
Air Blank	0.000	22:22	
Subject Sample	# 2 0.379	22:23	
Breath Volume	3.140	22:23	
Air Blank	0.000	22:24	
Air Blank	0.000	22:25	
Subject Sample	# 3 0.374	22:26	
Breath Volume	3.488	22:26	
Air Blank	0.000	22:27	
Control Test	0.073	22:27	Control Outside Tolerance
Air Blank	0.000	22:28	

Data Download Date/Time: 01/20/2021 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,GENE,A
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: SMITH,LARRY
ARRESTING OFFICER AGENCY: Apalachicola PD
LAST AGENCY INSPECTION: 12/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: GREENFIELD

DATE/TIME OF TEST: 01/02/2021 22:42:35

Diagnostics Check	OK	22:41	
Air Blank	0.000	22:41	
Control Test	0.073	22:41	Control Outside Tolerance
Air Blank	0.000	22:42	

Data Download Date/Time: 01/20/2021 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,GENE,A
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: DAVIS,TIMMY
ARRESTING OFFICER AGENCY: Apalachicola PD
LAST AGENCY INSPECTION: 12/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: GREENFIELD

DATE/TIME OF TEST: 01/02/2021 22:59:54

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:53	
Control Test	0.076	22:53	
Air Blank	0.000	22:54	
Subject Sample	# 1 0.359	22:54	
Breath Volume	4.179	22:54	
Air Blank	0.000	22:55	
Air Blank	0.000	22:57	
Subject Sample	# 2 0.357	22:58	
Breath Volume	3.343	22:58	
Air Blank	0.000	22:58	
Control Test	0.074	22:59	Control Outside Tolerance
Air Blank	0.000	22:59	

Data Download Date/Time: 01/20/2021 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,GENE,A
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: DAVIS,TIMMY
ARRESTING OFFICER AGENCY: Apalachicola PD
LAST AGENCY INSPECTION: 12/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: GREENFIELD

DATE/TIME OF TEST: 01/02/2021 23:29:06

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:26	
Control Test	0.075	23:26	
Air Blank	0.000	23:27	
Subject Sample # 1	0.341	23:27	
Breath Volume	3.562	23:27	
Air Blank		23:28	Purge Fail
Air Blank	0.000	23:29	

Data Download Date/Time: 01/20/2021 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,GENE,A
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: DAVIS,TIMMY
ARRESTING OFFICER AGENCY: Apalachicola PD
LAST AGENCY INSPECTION: 12/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: GREENFIELD

DATE/TIME OF TEST: 01/03/2021 00:09:54

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:00	
Control Test	0.075	00:00	
Air Blank	0.000	00:01	
Subject Sample # 1	0.289	00:01	
Breath Volume	1.828	00:01	
Air Blank	0.000	00:02	
Air Blank	0.000	00:04	
Subject Sample # 2	0.316	00:04	
Breath Volume	2.777	00:04	
Air Blank	0.000	00:05	
Air Blank	0.000	00:07	
Subject Sample # 3	0.318	00:08	
Breath Volume	2.003	00:08	
Air Blank	0.000	00:08	
Control Test	0.073	00:09	Control Outside Tolerance
Air Blank	0.000	00:09	

Data Download Date/Time: 01/20/2021 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: FULLARD

DATE/TIME OF TEST: 01/01/2021 01:05:56

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.080	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.158	01:01
Breath Volume	2.554	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.156	01:04
Breath Volume	2.574	01:04
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 01/05/2021 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: JENKINS II

DATE/TIME OF TEST: 01/01/2021 02:34:33

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.081	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.083	02:30
Breath Volume	2.570	02:30
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.078	02:33
Breath Volume	1.851	02:33
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 01/05/2021 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: PENDLETON

DATE/TIME OF TEST: 01/01/2021 03:42:25

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:40	
Control Test	0.081	03:40	
Air Blank	0.000	03:40	
Subject Sample	# 1	03:41	Subject Test Refused
Breath Volume	0.000	03:41	
Air Blank	0.000	03:41	
Control Test	0.081	03:41	
Air Blank	0.000	03:42	
Diagnostics Check	OK	03:42	

Data Download Date/Time: 01/05/2021 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: MARIN

DATE/TIME OF TEST: 01/02/2021 19:57:16

Diagnostics Check	OK	19:51
Air Blank	0.000	19:51
Control Test	0.081	19:52
Air Blank	0.000	19:52
Subject Sample # 1	0.000	19:53
Breath Volume	2.929	19:53
Air Blank	0.000	19:53
Air Blank	0.000	19:55
Subject Sample # 2	0.000	19:55
Breath Volume	1.945	19:55
Air Blank	0.000	19:56
Control Test	0.080	19:56
Air Blank	0.000	19:57
Diagnostics Check	OK	19:57

Data Download Date/Time: 01/05/2021 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: CHRISTOPHER

DATE/TIME OF TEST: 01/03/2021 01:01:45

Diagnostics Check	OK	00:55
Air Blank	0.000	00:55
Control Test	0.082	00:55
Air Blank	0.000	00:56
Subject Sample # 1	0.134	00:57
Breath Volume	1.531	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.134	01:00
Breath Volume	1.562	01:00
Air Blank	0.000	01:00
Control Test	0.080	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 01/05/2021 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,A.,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: GONZALEZ,A.
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: BAIN

DATE/TIME OF TEST: 01/20/2021 00:24:33

Diagnostics Check		OK	00:18
Air Blank		0.000	00:18
Control Test		0.080	00:19
Air Blank		0.000	00:19
Subject Sample	# 1	0.227	00:19
Breath Volume		2.312	00:19
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.217	00:22
Breath Volume		3.488	00:22
Air Blank		0.000	00:23
Control Test		0.079	00:23
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 02/11/2021 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTON, JEREMY, R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MARTINA, RYAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 01/24/2021 03:30:40

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.080	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.149	03:25
Breath Volume	1.562	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.144	03:29
Breath Volume	2.292	03:29
Air Blank	0.000	03:29
Control Test	0.080	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 02/11/2021 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTON, JEREMY, R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: RABEN, MATHEW
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 01/24/2021 04:39:58

Diagnostics Check	OK	04:32
Air Blank	0.000	04:33
Control Test	0.080	04:33
Air Blank	0.000	04:33
Subject Sample # 1	0.196	04:35
Breath Volume	2.179	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:37
Subject Sample # 2	0.189	04:38
Breath Volume	1.628	04:38
Air Blank	0.000	04:39
Control Test	0.079	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 02/11/2021 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: HAFHEY

DATE/TIME OF TEST: 01/24/2021 04:49:41

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.079	04:44
Air Blank	0.000	04:44
Subject Sample # 1	0.111	04:45
Breath Volume	2.574	04:45
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.111	04:48
Breath Volume	2.914	04:48
Air Blank	0.000	04:48
Control Test	0.079	04:49
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 02/11/2021 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: CARLOS

DATE/TIME OF TEST: 01/29/2021 01:24:49

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.081	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.162	01:20
Breath Volume	3.082	01:20
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.164	01:23
Breath Volume	3.570	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:24
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 02/11/2021 12:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000960

INSTRUMENT REGISTERED TO: LIBERTY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHULER, NEIL, A
BREATH TEST OPERATOR AGENCY: Liberty County SO
ARRESTING OFFICER: FINUFF, DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: SELLERS

DATE/TIME OF TEST: 01/10/2021 15:12:05

Diagnostics Check	OK	15:09	
Air Blank	0.000	15:09	
Control Test	0.082	15:09	
Air Blank	0.000	15:10	
Subject Sample	# 1	15:10	Subject Test Refused
Breath Volume	0.000	15:10	
Air Blank	0.000	15:11	
Control Test	0.082	15:11	
Air Blank	0.000	15:12	
Diagnostics Check	OK	15:12	

Data Download Date/Time: 02/16/2021 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,GEORGE,H
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: EDWARDS,GEORGE
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 01/10/2021 23:01:40

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:59	
Control Test	0.081	22:59	
Air Blank	0.000	23:00	
Subject Sample	# 1	23:01	Sequence Aborted
Breath Volume	0.000	23:01	
Air Blank	0.000	23:01	

Data Download Date/Time: 01/21/2021 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,GEORGE,H
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: EDWARDS,GEORGE
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 01/10/2021 23:59:43

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:57	
Control Test	0.081	23:57	
Air Blank	0.000	23:58	
Subject Sample	# 1	23:58	Subject Test Refused
Breath Volume	0.000	23:58	
Air Blank	0.000	23:58	
Control Test	0.081	23:59	
Air Blank	0.000	23:59	
Diagnostics Check	OK	23:59	

Data Download Date/Time: 01/21/2021 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARY,BRIAN,G
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MILLER,KENYON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: TOLLIVER

DATE/TIME OF TEST: 01/17/2021 21:52:16

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:46	
Control Test	0.081	21:46	
Air Blank	0.000	21:47	
Subject Sample	# 1 0.143	21:50	Volume Not Met
Breath Volume	0.183	21:50	
Air Blank		21:50	RFI Detect
Air Blank		21:51	Purge Fail
Air Blank		21:52	Purge Fail

Data Download Date/Time: 01/21/2021 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARY,BRIAN,G
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MILLER,KENYON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: TOLLIVER

DATE/TIME OF TEST: 01/17/2021 21:56:32

Diagnostics Check	OK	21:55	
Air Blank		21:56	RFI Detect
Air Blank	0.000	21:56	

Data Download Date/Time: 01/21/2021 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARY,BRIAN,G
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MILLER,KENYON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: TOLLIVER

DATE/TIME OF TEST: 01/17/2021 22:04:50

Diagnostics Check	OK	21:59	
Air Blank	0.000	21:59	
Control Test	0.080	21:59	
Air Blank	0.000	22:00	
Subject Sample	# 1 0.150	22:01	
Breath Volume	1.570	22:01	
Air Blank	0.000	22:02	
Air Blank	0.000	22:03	
Subject Sample	# 2 0.147	22:04	
Breath Volume	1.488	22:04	
Air Blank		22:04	RFI Detect
Air Blank	0.000	22:04	

Data Download Date/Time: 01/21/2021 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENING,FRED,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MILLER,KENYON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: TOLLIVER

DATE/TIME OF TEST: 01/17/2021 22:30:56

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:26	
Control Test	0.081	22:26	
Air Blank	0.000	22:26	
Subject Sample	# 1 0.143	22:29	
Breath Volume	1.910	22:29	
Air Blank		22:30	Purge Fail
Air Blank	0.000	22:30	

Data Download Date/Time: 01/21/2021 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENING,FRED,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MILLER,KENYON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: TOLLIVER

DATE/TIME OF TEST: 01/17/2021 22:53:13

Diagnostics Check		OK	22:38	
Air Blank		0.000	22:38	
Control Test		0.080	22:39	
Air Blank		0.000	22:39	
Subject Sample	# 1	0.143	22:40	
Breath Volume		1.351	22:40	
Air Blank		0.000	22:41	
Air Blank		0.000	22:42	
Subject Sample	# 2	0.123	22:45	Volume Not Met
Breath Volume		0.253	22:45	
Air Blank		0.000	22:46	
Air Blank		0.000	22:48	
Subject Sample	# 3	0.118	22:51	Volume Not Met
Breath Volume		0.320	22:51	
Air Blank		0.000	22:52	
Control Test		0.076	22:52	
Air Blank		0.000	22:53	
Diagnostics Check		OK	22:53	

Data Download Date/Time: 01/21/2021 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENING,FRED,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MILLER,KENYON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/17/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: TOLLIVER

DATE/TIME OF TEST: 01/17/2021 23:09:15

Diagnostics Check	OK	23:06	
Air Blank	0.000	23:06	
Control Test	0.080	23:07	
Air Blank	0.000	23:07	
Subject Sample	# 1	23:07	Subject Test Refused
Breath Volume	0.000	23:07	
Air Blank	0.000	23:08	
Control Test	0.080	23:08	
Air Blank	0.000	23:09	
Diagnostics Check	OK	23:09	

Data Download Date/Time: 01/21/2021 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REVELL,DENNIS,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: REVELL,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: WHITWORTH

DATE/TIME OF TEST: 01/29/2021 21:23:34

Diagnostics Check	OK	21:22	
Air Blank	0.000	21:22	
Control Test	0.000	21:23	RFI Detect
Air Blank	0.000	21:23	

Data Download Date/Time: 03/03/2021 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REVELL,DENNIS,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: REVELL,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: WHITWORTH

DATE/TIME OF TEST: 01/29/2021 21:34:29

Diagnostics Check	OK	21:28
Air Blank	0.000	21:28
Control Test	0.080	21:28
Air Blank	0.000	21:29
Subject Sample # 1	0.124	21:30
Breath Volume	2.441	21:30
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample # 2	0.119	21:32
Breath Volume	2.902	21:32
Air Blank	0.000	21:33
Control Test	0.081	21:33
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

Data Download Date/Time: 03/03/2021 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VENTURA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: ANKNEY

DATE/TIME OF TEST: 01/07/2021 01:04:34

Diagnostics Check	OK	00:58
Air Blank	0.000	00:58
Control Test	0.080	00:59
Air Blank	0.000	00:59
Subject Sample # 1	0.000	01:00
Breath Volume	3.058	01:00
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.000	01:03
Breath Volume	3.492	01:03
Air Blank	0.000	01:03
Control Test	0.079	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 01/12/2021 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES, JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: HEWICK

DATE/TIME OF TEST: 01/07/2021 23:00:41

Diagnostics Check	OK	22:54
Air Blank	0.000	22:54
Control Test	0.079	22:54
Air Blank	0.000	22:55
Subject Sample # 1	0.190	22:56
Breath Volume	2.699	22:56
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.189	22:59
Breath Volume	2.628	22:59
Air Blank	0.000	22:59
Control Test	0.080	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 01/12/2021 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TATTOLI,RYAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: STOLOFF

DATE/TIME OF TEST: 01/08/2021 02:28:15

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.079	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.137	02:23
Breath Volume	3.250	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.136	02:26
Breath Volume	2.750	02:26
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 01/12/2021 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK,STEPHEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: HIRALDO HERNANDEZ

DATE/TIME OF TEST: 01/09/2021 03:43:30

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.170	03:39
Breath Volume	1.839	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.172	03:42
Breath Volume	1.476	03:42
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 01/12/2021 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTINEZ,ERNESTO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: CARABALLO

DATE/TIME OF TEST: 01/16/2021 03:51:27

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.079	03:40	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:44	No Sample Provided
Breath Volume	0.000	03:44	
Air Blank	0.000	03:44	
Air Blank	0.000	03:46	
Subject Sample	# 2	03:50	No Sample Provided
Breath Volume	0.000	03:50	
Air Blank	0.000	03:50	
Control Test	0.080	03:50	
Air Blank	0.000	03:51	
Diagnostics Check	OK	03:51	

Data Download Date/Time: 02/03/2021 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEWAN,JANICE,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SLAFER,GAGE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: GONZALESEVANS

DATE/TIME OF TEST: 01/17/2021 03:21:09

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.080	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.121	03:16
Breath Volume	3.023	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.129	03:19
Breath Volume	3.375	03:19
Air Blank	0.000	03:20
Control Test	0.080	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 02/03/2021 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 01/17/2021 05:43:02

Diagnostics Check	OK	05:39	
Air Blank	0.000	05:39	
Control Test	0.080	05:39	
Air Blank	0.000	05:40	
Subject Sample	# 1	05:41	Subject Test Refused
Breath Volume	0.000	05:41	
Air Blank	0.000	05:42	
Control Test	0.080	05:42	
Air Blank	0.000	05:42	
Diagnostics Check	OK	05:43	

Data Download Date/Time: 02/03/2021 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CERVELLERA,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: GERMAN

DATE/TIME OF TEST: 01/17/2021 21:47:04

Diagnostics Check	OK	21:40
Air Blank	0.000	21:40
Control Test	0.080	21:40
Air Blank	0.000	21:41
Subject Sample # 1	0.129	21:42
Breath Volume	1.414	21:42
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.122	21:45
Breath Volume	1.242	21:45
Air Blank	0.000	21:46
Control Test	0.079	21:46
Air Blank	0.000	21:46
Diagnostics Check	OK	21:47

Data Download Date/Time: 02/03/2021 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SINGSON,ROLAND
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: BEAUMILUS

DATE/TIME OF TEST: 01/18/2021 22:48:14

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.080	22:40
Air Blank	0.000	22:40
Subject Sample # 1	0.109	22:42
Breath Volume	1.359	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.101	22:46
Breath Volume	1.121	22:46
Air Blank	0.000	22:47
Control Test	0.080	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 02/03/2021 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: OGLETREE,JERRELL
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 01/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: ALVESPAIVA

DATE/TIME OF TEST: 01/23/2021 03:47:08

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.139	03:42
Breath Volume	1.347	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.141	03:45
Breath Volume	2.402	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 02/03/2021 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MELLA,MELISSA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MEJIASALGUERO

DATE/TIME OF TEST: 01/24/2021 00:45:37

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.080	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.206	00:41
Breath Volume		2.406	00:41
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.208	00:44
Breath Volume		2.261	00:44
Air Blank		0.000	00:44
Control Test		0.079	00:45
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 02/03/2021 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CARBONE,ANDRES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: TESENIAR

DATE/TIME OF TEST: 01/24/2021 04:15:30

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.080	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.218	04:11
Breath Volume	1.371	04:11
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.229	04:14
Breath Volume	2.941	04:14
Air Blank	0.000	04:14
Control Test	0.079	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 02/03/2021 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BORDEAUX,LAUREN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 01/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: BURGER

DATE/TIME OF TEST: 01/26/2021 00:51:47

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.081	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.000	00:47
Breath Volume	2.222	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.000	00:50
Breath Volume	1.910	00:50
Air Blank	0.000	00:50
Control Test	0.080	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 02/03/2021 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CISOWSKI,ANDREW
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 01/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/26/2021 19:40:25

Diagnostics Check	OK	19:33
Air Blank	0.000	19:34
Control Test	0.080	19:34
Air Blank	0.000	19:35
Subject Sample # 1	0.188	19:35
Breath Volume	1.175	19:35
Air Blank	0.000	19:36
Air Blank	0.000	19:38
Subject Sample # 2	0.181	19:38
Breath Volume	1.167	19:38
Air Blank	0.000	19:39
Control Test	0.079	19:39
Air Blank	0.000	19:40
Diagnostics Check	OK	19:40

Data Download Date/Time: 02/03/2021 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ELLISON,LINDSEY
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 01/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: CANO

DATE/TIME OF TEST: 01/27/2021 02:29:09

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.079	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.161	02:24
Breath Volume	2.105	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.164	02:27
Breath Volume	2.019	02:27
Air Blank	0.000	02:28
Control Test	0.080	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 02/03/2021 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ERICKSON,JORDAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: GRIGGS

DATE/TIME OF TEST: 01/28/2021 02:02:26

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.080	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.139	01:58
Breath Volume	1.804	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.150	02:00
Breath Volume	2.675	02:00
Air Blank	0.000	02:01
Control Test	0.079	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 02/03/2021 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VESCO,ALEX
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 01/29/2021 00:10:15

Diagnostics Check	OK	23:54	
Air Blank	0.000	23:54	
Control Test	0.080	23:55	
Air Blank	0.000	23:55	
Subject Sample	# 3 0.137	23:57	
Breath Volume	1.320	23:57	
Air Blank	0.000	23:58	
Air Blank	0.000	23:59	
Subject Sample	# 1 0.105	00:03	Volume Not Met
Breath Volume	0.261	00:03	
Air Blank	0.000	00:03	
Air Blank	0.000	00:05	
Subject Sample	# 2 0.124	00:08	Volume Not Met
Breath Volume	0.960	00:08	
Air Blank	0.000	00:09	
Control Test	0.079	00:09	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 02/03/2021 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VESCO,ALEX
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 01/29/2021 00:18:18

Diagnostics Check	OK	00:12
Air Blank	0.000	00:12
Control Test	0.079	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.127	00:14
Breath Volume	1.496	00:14
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.134	00:16
Breath Volume	1.652	00:16
Air Blank	0.000	00:17
Control Test	0.079	00:17
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 02/03/2021 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KNIGHT,RYAN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 01/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 01/30/2021 04:04:45

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.191	03:59
Breath Volume	1.230	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.187	04:03
Breath Volume	1.992	04:03
Air Blank	0.000	04:03
Control Test	0.079	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 02/03/2021 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BANOS,FRANK
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 01/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: WHITTAKER

DATE/TIME OF TEST: 01/31/2021 04:01:45

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.079	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.145	03:57
Breath Volume	1.175	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.142	04:00
Breath Volume	1.625	04:00
Air Blank	0.000	04:00
Control Test	0.079	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 02/03/2021 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,JOSEPH,B
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FRANKLIN,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: MENJIVAR

DATE/TIME OF TEST: 01/01/2021 03:10:37

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.079	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.152	03:06
Breath Volume		2.605	03:06
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.143	03:09
Breath Volume		2.914	03:09
Air Blank		0.000	03:09
Control Test		0.076	03:10
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 01/12/2021 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: SAMPSON

DATE/TIME OF TEST: 01/02/2021 05:01:23

Diagnostics Check	OK	05:00	
Air Blank	0.000	05:00	
		05:00	RFI Detect
Air Blank	0.000	05:01	

Data Download Date/Time: 01/12/2021 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: SAMPSON

DATE/TIME OF TEST: 01/02/2021 05:08:13

Diagnostics Check	OK	05:03	
Air Blank	0.000	05:03	
Control Test	0.078	05:03	
Air Blank	0.000	05:04	
Subject Sample	# 1	05:06	Subject Test Refused
Breath Volume	0.000	05:06	
Air Blank	0.000	05:07	
Control Test	0.079	05:07	
Air Blank	0.000	05:08	
Diagnostics Check	OK	05:08	

Data Download Date/Time: 01/12/2021 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,JOSEPH,B
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 01/17/2021 20:07:40

Diagnostics Check	OK	20:00
Air Blank	0.000	20:01
Control Test	0.078	20:01
Air Blank	0.000	20:02
Subject Sample # 1	0.248	20:03
Breath Volume	1.164	20:03
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.249	20:06
Breath Volume	1.667	20:06
Air Blank	0.000	20:06
Control Test	0.080	20:07
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 02/15/2021 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROHECKER,KYNDAL,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MILLER,KENYON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: MCMILLON

DATE/TIME OF TEST: 01/22/2021 05:07:13

Diagnostics Check		OK	04:58
Air Blank		0.000	04:59
Control Test		0.078	04:59
Air Blank		0.000	04:59
Subject Sample	# 1	0.131	05:02
Breath Volume		1.640	05:02
Air Blank		0.000	05:03
Air Blank		0.000	05:04
Subject Sample	# 2	0.127	05:05
Breath Volume		1.410	05:05
Air Blank		0.000	05:06
Control Test		0.078	05:06
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 02/15/2021 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTON,JEREMY,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BARTON,JEREMY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: SURA

DATE/TIME OF TEST: 01/24/2021 01:46:13

Diagnostics Check		OK	01:39
Air Blank		0.000	01:40
Control Test		0.079	01:40
Air Blank		0.000	01:41
Subject Sample	# 1	0.213	01:41
Breath Volume		2.136	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.220	01:44
Breath Volume		1.105	01:44
Air Blank		0.000	01:45
Control Test		0.079	01:45
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 02/15/2021 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,RONDRICKA,N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: LOGAN

DATE/TIME OF TEST: 01/27/2021 13:41:47

Diagnostics Check	OK	13:35
Air Blank	0.000	13:36
Control Test	0.078	13:36
Air Blank	0.000	13:36
Subject Sample # 1	0.047	13:37
Breath Volume	2.203	13:37
Air Blank	0.000	13:38
Air Blank	0.000	13:39
Subject Sample # 2	0.046	13:40
Breath Volume	2.640	13:40
Air Blank	0.000	13:40
Control Test	0.079	13:41
Air Blank	0.000	13:41
Diagnostics Check	OK	13:41

Data Download Date/Time: 02/15/2021 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTONE,LACEY,M
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: SANTONE,LACEY
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: BANTON

DATE/TIME OF TEST: 01/01/2021 01:16:41

Diagnostics Check		OK	01:05	
Air Blank		0.000	01:05	
Control Test		0.078	01:05	
Air Blank		0.000	01:06	
Subject Sample	# 1	0.131	01:09	Volume Not Met
Breath Volume		0.289	01:09	
Air Blank		0.000	01:10	
Air Blank		0.000	01:12	
Subject Sample	# 2	0.000	01:15	Volume Not Met
Breath Volume		0.242	01:15	
Air Blank		0.000	01:15	
Control Test		0.079	01:16	
Air Blank		0.000	01:16	
Diagnostics Check		OK	01:16	

Data Download Date/Time: 01/19/2021 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOUVRES,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: DOUVRES,MICHAEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: BURGOS

DATE/TIME OF TEST: 01/03/2021 04:50:22

Diagnostics Check	OK	04:43
Air Blank	0.000	04:44
Control Test	0.079	04:44
Air Blank	0.000	04:45
Subject Sample # 1	0.188	04:45
Breath Volume	3.062	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.186	04:48
Breath Volume	2.738	04:48
Air Blank	0.000	04:49
Control Test	0.079	04:49
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 01/19/2021 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: CALDERON JR

DATE/TIME OF TEST: 01/04/2021 03:00:49

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.079	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.000	02:56
Breath Volume	3.824	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.000	02:59
Breath Volume	3.734	02:59
Air Blank	0.000	02:59
Control Test	0.078	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 01/19/2021 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SOLIS,GUSTAVO
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 08:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:09

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 01/06/2021 09:41:59

Diagnostics Check	OK	09:30	
Air Blank	0.000	09:30	
Control Test	0.080	09:31	
Air Blank	0.000	09:31	
Subject Sample	# 1	09:34	No Sample Provided
Breath Volume	0.000	09:34	
Air Blank	0.000	09:35	
Air Blank	0.000	09:37	
Subject Sample	# 2	0.163	09:37
Breath Volume	2.070	09:37	
Air Blank	0.000	09:38	
Air Blank	0.000	09:40	
Subject Sample	# 3	0.143	09:40
Breath Volume	1.210	09:40	
Air Blank	0.000	09:41	
Control Test	0.078	09:41	
Air Blank	0.000	09:41	
Diagnostics Check	OK	09:41	

Data Download Date/Time: 01/19/2021 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WYBIRAL,JUSTIN
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: AROCHO PEREZ

DATE/TIME OF TEST: 01/16/2021 02:07:29

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.080	02:00
Air Blank	0.000	02:01
Subject Sample # 1	0.082	02:02
Breath Volume	1.957	02:02
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.079	02:05
Breath Volume	2.453	02:05
Air Blank	0.000	02:06
Control Test	0.078	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 01/19/2021 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOUVRES,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: DOUVRES,MICHAEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 01/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: CARABDALLO

DATE/TIME OF TEST: 01/27/2021 03:37:21

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.079	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.227	03:31
Breath Volume	1.714	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:33
Subject Sample # 2	0.231	03:35
Breath Volume	1.281	03:35
Air Blank	0.000	03:36
Control Test	0.078	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 02/16/2021 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 01/30/2021 04:26:28

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.079	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.112	04:21
Breath Volume	2.671	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.110	04:24
Breath Volume	2.679	04:24
Air Blank	0.000	04:25
Control Test	0.078	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 02/16/2021 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONTRERAS,BRYANT,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CONTRERAS,BRYANT
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 13:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:12

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 01/30/2021 14:43:50

Diagnostics Check	OK	14:36
Air Blank	0.000	14:36
Control Test	0.079	14:36
Air Blank	0.000	14:37
Subject Sample # 1	0.087	14:38
Breath Volume	3.152	14:38
Air Blank	0.000	14:38
Air Blank	0.000	14:40
Subject Sample # 2	0.090	14:42
Breath Volume	2.394	14:42
Air Blank	0.000	14:42
Control Test	0.080	14:43
Air Blank	0.000	14:43
Diagnostics Check	OK	14:43

Data Download Date/Time: 02/16/2021 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INGERSOLL,NICHOLAS,F
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: INGERSOLL,NICHOLAS
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 12/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: GRIMES

DATE/TIME OF TEST: 01/02/2021 00:26:10

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.081	00:20
Air Blank	0.000	00:20
Subject Sample # 1	0.150	00:21
Breath Volume	1.753	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.151	00:24
Breath Volume	2.761	00:24
Air Blank	0.000	00:25
Control Test	0.080	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 02/03/2021 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN, BRADLEY,
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: MCLAUGHLIN, BRADLEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 01/03/2021 21:49:17

Diagnostics Check	OK	21:37	
Air Blank	0.000	21:38	
Control Test	0.076	21:38	
Air Blank	0.000	21:39	
Subject Sample	# 1	21:42	No Sample Provided
Breath Volume	0.000	21:42	
Air Blank	0.000	21:42	
Air Blank	0.000	21:44	
Subject Sample	# 2	21:47	No Sample Provided
Breath Volume	0.000	21:47	
Air Blank	0.000	21:48	
Control Test	0.078	21:48	
Air Blank	0.000	21:49	
Diagnostics Check	OK	21:49	

Data Download Date/Time: 02/03/2021 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JUSTIN,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ADAMS,JUSTIN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: EDDY

DATE/TIME OF TEST: 01/04/2021 20:47:19

Diagnostics Check	OK	20:35		
Air Blank	0.000	20:36		
Control Test	0.080	20:36		
Air Blank	0.000	20:36		
Subject Sample	# 1	20:40	No Sample Provided	
Breath Volume	0.000	20:40		
Air Blank	0.000	20:40		
Air Blank	0.000	20:42		
Subject Sample	# 2	0.139	20:45	Volume Not Met
Breath Volume	0.226	20:45		
Air Blank	0.000	20:46		
Control Test	0.081	20:46		
Air Blank	0.000	20:47		
Diagnostics Check	OK	20:47		

Data Download Date/Time: 02/03/2021 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAROCHE, CHARLES, W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LAROCHE, CHARLES
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 01/04/2021 22:07:58

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.081	22:02
Air Blank	0.000	22:02
Subject Sample # 1	0.245	22:03
Breath Volume	2.378	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:05
Subject Sample # 2	0.260	22:06
Breath Volume	1.371	22:06
Air Blank	0.000	22:07
Control Test	0.080	22:07
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 02/03/2021 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTMAN, MATTHEW, R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: ROBERTS, MICHAEL
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 12/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: GASTUNAS

DATE/TIME OF TEST: 01/16/2021 22:49:07

Diagnostics Check	OK	22:41
Air Blank	0.000	22:42
Control Test	0.081	22:42
Air Blank	0.000	22:43
Subject Sample # 1	0.220	22:44
Breath Volume	2.527	22:44
Air Blank	0.000	22:45
Air Blank	0.000	22:46
Subject Sample # 2	0.230	22:47
Breath Volume	3.359	22:47
Air Blank	0.000	22:48
Control Test	0.080	22:48
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 02/03/2021 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN,BRADLEY,F
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: MCLAUGHLIN,BRADLEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: CROFT

DATE/TIME OF TEST: 01/22/2021 01:02:13

Diagnostics Check		OK	00:56
Air Blank		0.000	00:56
Control Test		0.079	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.172	00:57
Breath Volume		1.601	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.167	01:00
Breath Volume		1.718	01:00
Air Blank		0.000	01:01
Control Test		0.080	01:01
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 02/03/2021 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN, BRADLEY, F
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HATZEL, BRANDON
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 12/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: BETSINGER

DATE/TIME OF TEST: 01/22/2021 03:37:39

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.000	03:33
Breath Volume	1.394	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.000	03:36
Breath Volume	1.250	03:36
Air Blank	0.000	03:36
Control Test	0.079	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 02/03/2021 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INGERSOLL,NICHOLAS,F
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: OLSEN,IAN
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 12/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: CRESSON

DATE/TIME OF TEST: 01/26/2021 21:43:22

Diagnostics Check	OK	21:34
Air Blank	0.000	21:35
Control Test	0.079	21:35
Air Blank	0.000	21:35
Subject Sample # 1	0.232	21:38
Breath Volume	1.628	21:38
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.223	21:41
Breath Volume	1.101	21:41
Air Blank	0.000	21:42
Control Test	0.078	21:42
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 02/03/2021 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MCDONOUGH,ANTHONY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: HOOD

DATE/TIME OF TEST: 01/03/2021 23:12:24

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.082	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.166	23:07
Breath Volume	1.429	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.149	23:10
Breath Volume	1.187	23:10
Air Blank	0.000	23:11
Control Test	0.081	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 01/08/2021 13:52