

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUERTA, EDWIN,  
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD  
ARRESTING OFFICER: LORENZO, FRANCISCO  
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD  
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: GARCIA PEREZ

DATE/TIME OF TEST: 10/06/2021 21:55:26

---

Diagnostics Check	OK	21:48
Air Blank	0.000	21:48
Control Test	0.080	21:48
Air Blank	0.000	21:49
Subject Sample # 1	0.218	21:51
Breath Volume	1.218	21:51
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.217	21:53
Breath Volume	1.281	21:53
Air Blank	0.000	21:54
Control Test	0.077	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

---

Data Download Date/Time: 10/30/2021 01:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHAVARRIA,DEGLYS,E  
BREATH TEST OPERATOR AGENCY: MEDLEY PD  
ARRESTING OFFICER: CHAVARRIA,DEGLYS  
ARRESTING OFFICER AGENCY: MEDLEY PD  
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:22

SUBJECT'S LAST NAME: MARIN

DATE/TIME OF TEST: 10/09/2021 07:46:59

---

Diagnostics Check	OK	07:44	
Air Blank	0.000	07:44	
Control Test	0.080	07:44	
Air Blank	0.000	07:45	
Subject Sample	# 1	07:45	Subject Test Refused
Breath Volume	0.000	07:45	
Air Blank	0.000	07:46	
Control Test	0.081	07:46	
Air Blank	0.000	07:46	
Diagnostics Check	OK	07:46	

---

Data Download Date/Time: 10/30/2021 01:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ, RICARDO, A  
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD  
ARRESTING OFFICER: PEREZ, RICARDO  
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD  
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: VERGARA

DATE/TIME OF TEST: 10/10/2021 02:57:33

---

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.000	02:53
Breath Volume	1.937	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.000	02:56
Breath Volume	2.429	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

---

Data Download Date/Time: 10/30/2021 01:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUERTA, EDWIN,  
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD  
ARRESTING OFFICER: LORENZO, FRANCISCO  
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD  
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: ACOSTA GONZALEZ

DATE/TIME OF TEST: 10/15/2021 22:55:18

---

Diagnostics Check	OK	22:48
Air Blank	0.000	22:48
Control Test	0.080	22:49
Air Blank	0.000	22:49
Subject Sample # 1	0.095	22:50
Breath Volume	1.359	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.097	22:53
Breath Volume	2.320	22:53
Air Blank	0.000	22:54
Control Test	0.077	22:54
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

---

Data Download Date/Time: 10/30/2021 01:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUERTA, EDWIN,  
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD  
ARRESTING OFFICER: LORENZO, FRANCISCO  
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD  
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CUBA BARRETO

DATE/TIME OF TEST: 10/16/2021 03:24:41

---

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:22	
Control Test	0.080	03:22	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:23	Subject Test Refused
Breath Volume	0.000	03:23	
Air Blank	0.000	03:23	
Control Test	0.081	03:24	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

---

Data Download Date/Time: 10/30/2021 01:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUERTA, EDWIN,  
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD  
ARRESTING OFFICER: LORENZO, FRANCISCO  
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD  
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: TORREZ GUERRERO

DATE/TIME OF TEST: 10/17/2021 04:55:10

---

Diagnostics Check	OK	04:48
Air Blank	0.000	04:48
Control Test	0.081	04:48
Air Blank	0.000	04:49
Subject Sample # 1	0.190	04:50
Breath Volume	1.281	04:50
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.194	04:53
Breath Volume	1.437	04:53
Air Blank	0.000	04:54
Control Test	0.080	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

---

Data Download Date/Time: 10/30/2021 01:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALDANA,ERICK,N  
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD  
ARRESTING OFFICER: LORENZO,FRANCISCO  
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD  
LAST AGENCY INSPECTION: 10/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: GAGO

DATE/TIME OF TEST: 10/30/2021 03:26:17

---

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:23	
Control Test	0.080	03:24	
Air Blank	0.000	03:24	
Subject Sample	# 1	03:24	Subject Test Refused
Breath Volume	0.000	03:24	
Air Blank	0.000	03:25	
Control Test	0.079	03:25	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

---

Data Download Date/Time: 11/30/2021 21:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALDANA,ERICK,  
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD  
ARRESTING OFFICER: REYES,ISRAEL  
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD  
LAST AGENCY INSPECTION: 10/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: FERNANDEZ MONTERREY

DATE/TIME OF TEST: 10/31/2021 01:54:22

---

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.081	01:52	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:53	
Control Test	0.081	01:53	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

---

Data Download Date/Time: 11/30/2021 21:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OCEGUERA,YASMANI,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: OCEGUERA,YASMANI  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 09/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: SALGUEDO

DATE/TIME OF TEST: 10/01/2021 23:44:40

---

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:38	
Control Test	0.078	23:38	
Air Blank	0.000	23:39	
Subject Sample	# 1 0.241	23:39	
Breath Volume	1.109	23:39	
Air Blank	0.000	23:40	
Air Blank	0.000	23:42	
Subject Sample	# 2	23:43	Subject Test Refused
Breath Volume	0.031	23:43	
Air Blank	0.000	23:43	
Control Test	0.077	23:44	
Air Blank	0.000	23:44	
Diagnostics Check	OK	23:44	

---

Data Download Date/Time: 10/07/2021 07:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FERNANDEZ,O,  
BREATH TEST OPERATOR AGENCY: SWEETWATER PD  
ARRESTING OFFICER: LAPLACELIERE,L  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: BRACAMONTEMARTINEZ

DATE/TIME OF TEST: 10/02/2021 22:09:08

---

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.081	22:02
Air Blank	0.000	22:03
Subject Sample # 1	0.226	22:04
Breath Volume	1.722	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.224	22:07
Breath Volume	1.742	22:07
Air Blank	0.000	22:08
Control Test	0.077	22:08
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

---

Data Download Date/Time: 10/14/2021 03:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: CLOSIUS,RICHARD  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: TONSMEIRE

DATE/TIME OF TEST: 10/03/2021 01:12:22

---

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.266	01:07
Breath Volume		2.953	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:09
Subject Sample	# 2	0.281	01:10
Breath Volume		3.320	01:10
Air Blank		0.000	01:11
Control Test		0.080	01:11
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

---

Data Download Date/Time: 10/14/2021 03:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERO,NATALIE,  
BREATH TEST OPERATOR AGENCY: OTHER LOCAL  
ARRESTING OFFICER: CASTILLO,JOSUET  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: JURE

DATE/TIME OF TEST: 10/04/2021 04:12:17

---

Diagnostics Check		OK	03:57	
Air Blank		0.000	03:58	
Control Test		0.081	03:58	
Air Blank		0.000	03:58	
Subject Sample	# 1	0.235	04:01	Volume Not Met
Breath Volume		0.378	04:01	
Air Blank		0.000	04:02	
Air Blank		0.000	04:04	
Subject Sample	# 2	0.256	04:07	
Breath Volume		1.222	04:07	
Air Blank		0.000	04:08	
Air Blank		0.000	04:09	
Subject Sample	# 3	0.245	04:10	
Breath Volume		2.230	04:10	
Air Blank		0.000	04:11	
Control Test		0.078	04:11	
Air Blank		0.000	04:12	
Diagnostics Check		OK	04:12	

---

Data Download Date/Time: 10/14/2021 03:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEY, LAZARO,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: LEY, LAZARO  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 10:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/04/2021 11:23:29

---

Diagnostics Check	OK	11:17
Air Blank	0.000	11:17
Control Test	0.081	11:17
Air Blank	0.000	11:18
Subject Sample # 1	0.000	11:19
Breath Volume	1.648	11:19
Air Blank	0.000	11:19
Air Blank	0.000	11:21
Subject Sample # 2	0.000	11:22
Breath Volume	1.808	11:22
Air Blank	0.000	11:22
Control Test	0.081	11:22
Air Blank	0.000	11:23
Diagnostics Check	OK	11:23

---

Data Download Date/Time: 10/14/2021 03:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEY, LAZARO,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: LEY, LAZARO  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 10:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/05/2021 11:03:37

---

Diagnostics Check	OK	10:57
Air Blank	0.000	10:57
Control Test	0.081	10:58
Air Blank	0.000	10:58
Subject Sample # 1	0.000	10:59
Breath Volume	1.632	10:59
Air Blank	0.000	10:59
Air Blank	0.000	11:01
Subject Sample # 2	0.000	11:02
Breath Volume	1.679	11:02
Air Blank	0.000	11:02
Control Test	0.081	11:02
Air Blank	0.000	11:03
Diagnostics Check	OK	11:03

---

Data Download Date/Time: 10/14/2021 03:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: FLORIAN,ROLANDO  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: SERNA

DATE/TIME OF TEST: 10/07/2021 05:48:46

---

Diagnostics Check	OK	05:41
Air Blank	0.000	05:42
Control Test	0.080	05:42
Air Blank	0.000	05:43
Subject Sample # 1	0.086	05:44
Breath Volume	1.964	05:44
Air Blank	0.000	05:44
Air Blank	0.000	05:46
Subject Sample # 2	0.074	05:47
Breath Volume	1.164	05:47
Air Blank	0.000	05:47
Control Test	0.079	05:48
Air Blank	0.000	05:48
Diagnostics Check	OK	05:48

---

Data Download Date/Time: 10/14/2021 03:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOUIS, WALEX,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: LOUIS, WALEX  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 10/10/2021 21:45:08

---

Diagnostics Check	OK	21:36
Air Blank	0.000	21:36
Control Test	0.079	21:37
Air Blank	0.000	21:37
Subject Sample # 1	0.191	21:38
Breath Volume	2.363	21:38
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.188	21:43
Breath Volume	2.460	21:43
Air Blank	0.000	21:44
Control Test	0.079	21:44
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

---

Data Download Date/Time: 10/14/2021 03:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEY, LAZARO,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: LEY, LAZARO  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 09:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/11/2021 10:02:00

---

Diagnostics Check	OK	09:55
Air Blank	0.000	09:56
Control Test	0.081	09:56
Air Blank	0.000	09:57
Subject Sample # 1	0.000	09:57
Breath Volume	1.773	09:57
Air Blank	0.000	09:58
Air Blank	0.000	10:00
Subject Sample # 2	0.000	10:00
Breath Volume	2.031	10:00
Air Blank	0.000	10:01
Control Test	0.081	10:01
Air Blank	0.000	10:01
Diagnostics Check	OK	10:02

---

Data Download Date/Time: 10/14/2021 03:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVALOS,DANIEL,A  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: LAPLACELIERE,LEONARDO  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 10/14/2021 22:13:44

---

Diagnostics Check	OK	22:10	
Air Blank	0.000	22:11	
Control Test	0.081	22:11	
Air Blank	0.000	22:12	
Subject Sample	# 1	22:12	Subject Test Refused
Breath Volume	0.000	22:12	
Air Blank	0.000	22:12	
Control Test	0.080	22:13	
Air Blank	0.000	22:13	
Diagnostics Check	OK	22:13	

---

Data Download Date/Time: 11/11/2021 12:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JUDAH,DAVID,B  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: JUDAH,DAVID  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 10/15/2021 20:53:50

---

Diagnostics Check	OK	20:46
Air Blank	0.000	20:46
Control Test	0.080	20:46
Air Blank	0.000	20:47
Subject Sample # 1	0.229	20:48
Breath Volume	1.226	20:48
Air Blank	0.000	20:49
Air Blank	0.000	20:50
Subject Sample # 2	0.242	20:52
Breath Volume	2.191	20:52
Air Blank	0.000	20:52
Control Test	0.080	20:53
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

---

Data Download Date/Time: 11/11/2021 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANCHEZ,CELESTE,E  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: SANCHEZ,CELESTE  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: SOCA

DATE/TIME OF TEST: 10/17/2021 20:15:02

---

Diagnostics Check	OK	20:06
Air Blank	0.000	20:06
Control Test	0.080	20:07
Air Blank	0.000	20:07
Subject Sample # 1	0.160	20:10
Breath Volume	1.351	20:10
Air Blank	0.000	20:11
Air Blank	0.000	20:12
Subject Sample # 2	0.163	20:13
Breath Volume	1.933	20:13
Air Blank	0.000	20:14
Control Test	0.078	20:14
Air Blank	0.000	20:14
Diagnostics Check	OK	20:15

---

Data Download Date/Time: 11/11/2021 12:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEY, LAZARO,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: LEY, LAZARO  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: COURT ORD  
ARREST TIME: 10:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/19/2021 11:18:47

---

Diagnostics Check	OK	11:12
Air Blank	0.000	11:12
Control Test	0.081	11:12
Air Blank	0.000	11:13
Subject Sample # 1	0.000	11:14
Breath Volume	1.703	11:14
Air Blank	0.000	11:14
Air Blank	0.000	11:16
Subject Sample # 2	0.000	11:17
Breath Volume	2.062	11:17
Air Blank	0.000	11:17
Control Test	0.081	11:18
Air Blank	0.000	11:18
Diagnostics Check	OK	11:18

---

Data Download Date/Time: 11/11/2021 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,MIGUEL,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: DELGADO,MIGUEL  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 10/23/2021 01:11:47

---

Diagnostics Check	OK	01:08	
Air Blank	0.000	01:09	
Control Test	0.081	01:09	
Air Blank	0.000	01:10	
Subject Sample	# 1	01:10	Subject Test Refused
Breath Volume	0.000	01:10	
Air Blank	0.000	01:10	
Control Test	0.081	01:11	
Air Blank	0.000	01:11	
Diagnostics Check	OK	01:11	

---

Data Download Date/Time: 11/11/2021 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,MIGUEL,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: DELGADO,MIGUEL  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 10/23/2021 01:40:12

---

Diagnostics Check		OK	01:26	
Air Blank		0.000	01:26	
Control Test		0.081	01:27	
Air Blank		0.000	01:27	
Subject Sample	# 1	0.236	01:28	
Breath Volume		1.617	01:28	
Air Blank		0.000	01:29	
Air Blank		0.000	01:31	
Subject Sample	# 2	0.195	01:34	Volume Not Met
Breath Volume		0.765	01:34	
Air Blank		0.000	01:35	
Air Blank		0.000	01:37	
Subject Sample	# 3		01:38	Subject Test Refused
Breath Volume		0.000	01:38	
Air Blank		0.000	01:39	
Control Test		0.078	01:39	
Air Blank		0.000	01:40	
Diagnostics Check		OK	01:40	

---

Data Download Date/Time: 11/11/2021 12:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAYARD,CLIFFORD,P  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: RILEY,LASHAUNTEY  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 08:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:35

SUBJECT'S LAST NAME: AGUILERA

DATE/TIME OF TEST: 10/23/2021 10:05:22

---

Diagnostics Check	OK	09:56	
Air Blank	0.000	09:57	
Control Test	0.080	09:57	
Air Blank	0.000	09:58	
Subject Sample	# 1	10:01	No Sample Provided
Breath Volume	0.000	10:01	
Air Blank	0.000	10:01	
Air Blank	0.000	10:03	
Subject Sample	# 2	10:03	Subject Test Refused
Breath Volume	0.000	10:03	
Air Blank	0.000	10:04	
Control Test	0.081	10:04	
Air Blank	0.000	10:05	
Diagnostics Check	OK	10:05	

---

Data Download Date/Time: 11/11/2021 12:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,MIGUEL,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: DELGADO,MIGUEL  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 10/24/2021 01:37:54

---

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:26	
Control Test		0.081	01:26	
Air Blank		0.000	01:27	
Subject Sample	# 1	0.329	01:30	Volume Not Met
Breath Volume		0.207	01:30	
Air Blank		0.000	01:31	
Air Blank		0.000	01:32	
Subject Sample	# 2	0.337	01:35	Volume Not Met
Breath Volume		0.183	01:35	
Air Blank		0.000	01:36	
Control Test		0.077	01:37	
Air Blank		0.000	01:37	
Diagnostics Check		OK	01:37	

---

Data Download Date/Time: 11/11/2021 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,MIGUEL,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: DELGADO,MIGUEL  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 10/24/2021 01:47:38

---

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.077	01:44	
Air Blank	0.000	01:44	
Subject Sample	# 1	01:45	Subject Test Refused
Breath Volume	0.429	01:45	
Air Blank	0.000	01:46	
Control Test	0.079	01:46	
Air Blank	0.000	01:47	
Diagnostics Check	OK	01:47	

---

Data Download Date/Time: 11/11/2021 12:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUINTANA,JORGE,F  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: QUINTANA,JORGE  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: MORA

DATE/TIME OF TEST: 10/24/2021 04:23:33

---

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:19	
Control Test	0.080	04:20	
Air Blank	0.000	04:20	
Subject Sample	# 1	04:21	Subject Test Refused
Breath Volume	0.425	04:21	
Air Blank	0.000	04:22	
Control Test	0.081	04:22	
Air Blank	0.000	04:23	
Diagnostics Check	OK	04:23	

---

Data Download Date/Time: 11/11/2021 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUTTON,B,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HARRELL,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: LANHAM 2

DATE/TIME OF TEST: 10/02/2021 03:27:54

---

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.081	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.074	03:23
Breath Volume	4.019	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.074	03:26
Breath Volume	3.691	03:26
Air Blank	0.000	03:26
Control Test	0.081	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

---

Data Download Date/Time: 10/26/2021 10:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEURER,D,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LAQUIDARA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 10/02/2021 04:54:12

---

Diagnostics Check	OK	04:47
Air Blank	0.000	04:47
Control Test	0.081	04:48
Air Blank	0.000	04:48
Subject Sample # 1	0.076	04:49
Breath Volume	1.886	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.075	04:52
Breath Volume	1.625	04:52
Air Blank	0.000	04:53
Control Test	0.081	04:53
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

---

Data Download Date/Time: 10/26/2021 10:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,JUSTIN,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: SANTA

DATE/TIME OF TEST: 10/05/2021 21:05:02

---

Diagnostics Check	OK	20:57
Air Blank	0.000	20:57
Control Test	0.081	20:58
Air Blank	0.000	20:58
Subject Sample # 1	0.120	20:59
Breath Volume	1.203	20:59
Air Blank	0.000	21:00
Air Blank	0.000	21:01
Subject Sample # 2	0.111	21:03
Breath Volume	1.808	21:03
Air Blank	0.000	21:04
Control Test	0.081	21:04
Air Blank	0.000	21:04
Diagnostics Check	OK	21:05

---

Data Download Date/Time: 10/26/2021 10:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,JUSTIN,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: YOC PEREZ

DATE/TIME OF TEST: 10/06/2021 01:58:48

---

Diagnostics Check	OK	01:51
Air Blank	0.000	01:52
Control Test	0.081	01:52
Air Blank	0.000	01:53
Subject Sample # 1	0.169	01:53
Breath Volume	2.242	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.167	01:57
Breath Volume	1.238	01:57
Air Blank	0.000	01:57
Control Test	0.080	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

---

Data Download Date/Time: 10/26/2021 10:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOSTON,D,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: PARRISH,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: YORK

DATE/TIME OF TEST: 10/07/2021 18:46:18

---

Diagnostics Check	OK	18:39
Air Blank	0.000	18:40
Control Test	0.081	18:40
Air Blank	0.000	18:41
Subject Sample # 1	0.286	18:41
Breath Volume	2.277	18:41
Air Blank	0.000	18:42
Air Blank	0.000	18:44
Subject Sample # 2	0.295	18:44
Breath Volume	2.652	18:44
Air Blank	0.000	18:45
Control Test	0.080	18:45
Air Blank	0.000	18:46
Diagnostics Check	OK	18:46

---

Data Download Date/Time: 10/26/2021 10:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,JUSTIN,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HAWKINS,A  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: JULIAN

DATE/TIME OF TEST: 10/08/2021 03:58:55

---

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.080	03:56	
Air Blank	0.000	03:56	
Subject Sample	# 1 0.161	03:57	
Breath Volume	1.769	03:57	
Air Blank		03:58	RFI Detect
Air Blank		03:58	RFI Detect
Air Blank	0.000	03:58	

---

Data Download Date/Time: 10/26/2021 10:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,JUSTIN,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HAWKINS,A  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: JULIAN

DATE/TIME OF TEST: 10/08/2021 04:09:48

---

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.080	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.157	04:05
Breath Volume	2.640	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.158	04:08
Breath Volume	3.121	04:08
Air Blank	0.000	04:08
Control Test	0.080	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

---

Data Download Date/Time: 10/26/2021 10:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEURER,D,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 10/10/2021 02:56:17

---

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.081	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.194	02:51
Breath Volume	1.511	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.195	02:54
Breath Volume	1.300	02:54
Air Blank	0.000	02:55
Control Test	0.081	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

---

Data Download Date/Time: 10/26/2021 10:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DESOUZA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 10/10/2021 05:10:54

---

Diagnostics Check	OK	05:03
Air Blank	0.000	05:04
Control Test	0.081	05:04
Air Blank	0.000	05:04
Subject Sample # 1	0.194	05:06
Breath Volume	1.625	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:08
Subject Sample # 2	0.187	05:09
Breath Volume	1.683	05:09
Air Blank	0.000	05:09
Control Test	0.080	05:10
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

---

Data Download Date/Time: 10/26/2021 10:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEURER,D,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LUGO,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: HARALSON

DATE/TIME OF TEST: 10/11/2021 04:07:20

---

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.081	04:00
Air Blank		0.000	04:01
Subject Sample	# 1	0.179	04:01
Breath Volume		1.570	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.185	04:05
Breath Volume		1.398	04:05
Air Blank		0.000	04:06
Control Test		0.080	04:06
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

---

Data Download Date/Time: 10/26/2021 10:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEURER,D,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNS,B  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:07

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/13/2021 06:31:46

---

Diagnostics Check	OK	06:28	
Air Blank	0.000	06:29	
Control Test	0.081	06:29	
Air Blank	0.000	06:30	
Subject Sample	# 1	06:30	Subject Test Refused
Breath Volume	0.000	06:30	
Air Blank	0.000	06:30	
Control Test	0.081	06:31	
Air Blank	0.000	06:31	
Diagnostics Check	OK	06:31	

---

Data Download Date/Time: 10/26/2021 10:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KEEN,C  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: CHAMBERLAIN

DATE/TIME OF TEST: 10/15/2021 04:21:52

---

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.082	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.096	04:17
Breath Volume	1.136	04:17
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.088	04:20
Breath Volume	1.296	04:20
Air Blank	0.000	04:20
Control Test	0.081	04:21
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

---

Data Download Date/Time: 10/26/2021 10:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HIXSON,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: BLOT

DATE/TIME OF TEST: 10/15/2021 23:26:26

---

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.081	23:20
Air Blank	0.000	23:21
Subject Sample # 1	0.144	23:21
Breath Volume	2.503	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.142	23:24
Breath Volume	2.250	23:24
Air Blank	0.000	23:25
Control Test	0.082	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

---

Data Download Date/Time: 10/26/2021 10:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SIMMONS,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 10/16/2021 02:09:08

---

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.082	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.036	02:04
Breath Volume	1.179	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.034	02:07
Breath Volume	1.320	02:07
Air Blank	0.000	02:08
Control Test	0.082	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

---

Data Download Date/Time: 10/26/2021 10:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KEEN,C  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/16/2021 02:58:15

---

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:55	
Control Test	0.081	02:56	
Air Blank	0.000	02:56	
Subject Sample	# 1	02:56	Subject Test Refused
Breath Volume	0.000	02:56	
Air Blank	0.000	02:57	
Control Test	0.081	02:57	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

---

Data Download Date/Time: 10/26/2021 10:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HILLIER,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: EVERY

DATE/TIME OF TEST: 10/16/2021 04:23:35

---

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.081	04:18
Air Blank	0.000	04:18
Subject Sample # 1	0.054	04:19
Breath Volume	1.933	04:19
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.053	04:22
Breath Volume	1.625	04:22
Air Blank	0.000	04:22
Control Test	0.081	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

---

Data Download Date/Time: 10/26/2021 10:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LOKMIC,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: NEWNAM

DATE/TIME OF TEST: 10/16/2021 19:52:10

---

Diagnostics Check	OK	19:49	
Air Blank	0.000	19:49	
Control Test	0.081	19:50	
Air Blank	0.000	19:50	
Subject Sample	# 1	19:50	Subject Test Refused
Breath Volume	0.000	19:50	
Air Blank	0.000	19:51	
Control Test	0.081	19:51	
Air Blank	0.000	19:52	
Diagnostics Check	OK	19:52	

---

Data Download Date/Time: 10/26/2021 10:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: KOZYARSKY

DATE/TIME OF TEST: 10/16/2021 22:51:10

---

Diagnostics Check	OK	22:41
Air Blank	0.000	22:42
Control Test	0.080	22:42
Air Blank	0.000	22:43
Subject Sample # 1	0.172	22:45
Breath Volume	1.480	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:47
Subject Sample # 2	0.165	22:49
Breath Volume	1.398	22:49
Air Blank	0.000	22:50
Control Test	0.080	22:50
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

---

Data Download Date/Time: 10/26/2021 10:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SIMMONS,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: MONDRAGON LIZARDI

DATE/TIME OF TEST: 10/17/2021 04:15:24

---

Diagnostics Check	OK	04:08
Air Blank	0.000	04:09
Control Test	0.081	04:09
Air Blank	0.000	04:09
Subject Sample # 1	0.205	04:10
Breath Volume	1.640	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:12
Subject Sample # 2	0.199	04:13
Breath Volume	2.296	04:13
Air Blank	0.000	04:14
Control Test	0.081	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

---

Data Download Date/Time: 10/26/2021 10:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LOKMIC,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: VAN SICKLE

DATE/TIME OF TEST: 10/18/2021 02:42:48

---

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:40	
Control Test	0.081	02:40	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:41	Subject Test Refused
Breath Volume	0.000	02:41	
Air Blank	0.000	02:41	
Control Test	0.082	02:42	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

---

Data Download Date/Time: 10/26/2021 10:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: PEREZ,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 10/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: BURNETTE

DATE/TIME OF TEST: 10/18/2021 05:56:41

---

Diagnostics Check	OK	05:53	
Air Blank	0.000	05:54	
Control Test	0.082	05:54	
Air Blank	0.000	05:55	
Subject Sample	# 1	05:55	Subject Test Refused
Breath Volume	0.000	05:55	
Air Blank	0.000	05:55	
Control Test	0.081	05:56	
Air Blank	0.000	05:56	
Diagnostics Check	OK	05:56	

---

Data Download Date/Time: 10/26/2021 10:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINSON,J,M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: JACKSON,R  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 10/01/2021 00:58:43

---

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:54	
Control Test	0.077	00:54	
Air Blank	0.000	00:54	
Subject Sample	# 1	00:57	Subject Test Refused
Breath Volume	0.000	00:57	
Air Blank	0.000	00:57	
Control Test	0.076	00:58	
Air Blank	0.000	00:58	
Diagnostics Check	OK	00:58	

---

Data Download Date/Time: 10/30/2021 02:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLOUNT,JUSTIN,M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: EDWARDS,I  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/03/2021 02:31:07

---

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.132	02:26
Breath Volume	2.832	02:26
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.132	02:29
Breath Volume	2.707	02:29
Air Blank	0.000	02:30
Control Test	0.076	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

---

Data Download Date/Time: 10/30/2021 02:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PICKETT,J,D  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: JACKSON,R  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 10/10/2021 02:33:52

---

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.079	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.111	02:29
Breath Volume	2.085	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.102	02:32
Breath Volume	1.781	02:32
Air Blank	0.000	02:32
Control Test	0.078	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

---

Data Download Date/Time: 10/30/2021 02:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLOUNT,J,M  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: PADGETT,M  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 10/17/2021 02:58:35

---

Diagnostics Check		OK	02:51
Air Blank		0.000	02:51
Control Test		0.078	02:52
Air Blank		0.000	02:52
Subject Sample	# 1	0.136	02:53
Breath Volume		2.246	02:53
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.134	02:57
Breath Volume		2.015	02:57
Air Blank		0.000	02:57
Control Test		0.076	02:57
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

---

Data Download Date/Time: 10/30/2021 02:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRISCOLL,SEAN,P  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HUFFMAN,J  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: HERON

DATE/TIME OF TEST: 10/17/2021 03:45:29

---

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.079	03:39
Air Blank		0.000	03:39
Subject Sample	# 1	0.157	03:40
Breath Volume		1.445	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.173	03:43
Breath Volume		2.140	03:43
Air Blank		0.000	03:44
Control Test		0.077	03:44
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

---

Data Download Date/Time: 10/30/2021 02:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KUTZ,T,E  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: JACKSON,R  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: VANDERHOOF

DATE/TIME OF TEST: 10/18/2021 20:26:08

---

Diagnostics Check	OK	20:23	
Air Blank	0.000	20:23	
Control Test	0.078	20:24	
Air Blank	0.000	20:24	
Subject Sample	# 1	20:24	Subject Test Refused
Breath Volume	0.000	20:24	
Air Blank	0.000	20:25	
Control Test	0.079	20:25	
Air Blank	0.000	20:26	
Diagnostics Check	OK	20:26	

---

Data Download Date/Time: 10/30/2021 02:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,J,  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: THOMASON,P  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BOONE

DATE/TIME OF TEST: 10/20/2021 03:11:25

---

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.077	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.180	03:07
Breath Volume		2.367	03:07
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.178	03:09
Breath Volume		2.453	03:09
Air Blank		0.000	03:10
Control Test		0.078	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

---

Data Download Date/Time: 10/30/2021 02:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KUTZ,T,E  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: JACKSON,R  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 10/20/2021 05:54:32

---

Diagnostics Check	OK	05:48
Air Blank	0.000	05:48
Control Test	0.078	05:49
Air Blank	0.000	05:49
Subject Sample # 1	0.132	05:50
Breath Volume	2.066	05:50
Air Blank	0.000	05:50
Air Blank	0.000	05:52
Subject Sample # 2	0.133	05:53
Breath Volume	1.312	05:53
Air Blank	0.000	05:53
Control Test	0.078	05:53
Air Blank	0.000	05:54
Diagnostics Check	OK	05:54

---

Data Download Date/Time: 10/30/2021 02:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARRIS, MICHAEL, T  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: FARRIS, MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: BOZARTH

DATE/TIME OF TEST: 10/20/2021 20:27:22

---

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.075	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.199	20:22
Breath Volume	2.531	20:22
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.198	20:25
Breath Volume	2.089	20:25
Air Blank	0.000	20:26
Control Test	0.077	20:26
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

---

Data Download Date/Time: 10/30/2021 02:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERON, WINSTON, G  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: HALL, J  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 10/21/2021 00:17:21

---

Diagnostics Check		OK	00:10
Air Blank		0.000	00:11
Control Test		0.077	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.045	00:12
Breath Volume		2.601	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.044	00:15
Breath Volume		2.921	00:15
Air Blank		0.000	00:16
Control Test		0.078	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

---

Data Download Date/Time: 10/30/2021 02:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOELZER, KRISTIE, A  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: WOODS, W  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 10/21/2021 02:31:27

---

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:30	
Control Test	0.079	02:30	
Air Blank		02:31	Improper Sample
Air Blank		02:31	Improper Sample

---

Data Download Date/Time: 10/30/2021 02:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOELZER, KRISTIE, A  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: WOODS, W  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 10/21/2021 02:41:59

---

Diagnostics Check		OK	02:35
Air Blank		0.000	02:35
Control Test		0.079	02:35
Air Blank		0.000	02:36
Subject Sample	# 1	0.000	02:37
Breath Volume		1.718	02:37
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.000	02:40
Breath Volume		1.710	02:40
Air Blank		0.000	02:41
Control Test		0.079	02:41
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

---

Data Download Date/Time: 10/30/2021 02:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOELZER, KRISTIE, A  
BREATH TEST OPERATOR AGENCY: Nassau County SO  
ARRESTING OFFICER: WOODS, W  
ARRESTING OFFICER AGENCY: Nassau County SO  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: REGAN

DATE/TIME OF TEST: 10/27/2021 02:15:20

---

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.078	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.162	02:07
Breath Volume	1.121	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.186	02:10
Breath Volume	3.296	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:12
Subject Sample # 3	0.179	02:13
Breath Volume	2.492	02:13
Air Blank	0.000	02:14
Control Test	0.078	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

---

Data Download Date/Time: 10/30/2021 02:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANGFORD,BRIAN,A  
BREATH TEST OPERATOR AGENCY: Gilchrist County SO  
ARRESTING OFFICER: VALENCIA,J  
ARRESTING OFFICER AGENCY: Gilchrist County SO  
LAST AGENCY INSPECTION: 09/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1224982  
DRY GAS EXPIRY DATE: 10/31/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 10/13/2021 17:31:45

---

Diagnostics Check	OK	17:25
Air Blank	0.000	17:25
Control Test	0.080	17:26
Air Blank	0.000	17:26
Subject Sample # 1	0.174	17:27
Breath Volume	1.554	17:27
Air Blank	0.000	17:27
Air Blank	0.000	17:29
Subject Sample # 2	0.176	17:30
Breath Volume	1.664	17:30
Air Blank	0.000	17:30
Control Test	0.078	17:31
Air Blank	0.000	17:31
Diagnostics Check	OK	17:31

---

Data Download Date/Time: 11/11/2021 16:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANGFORD,BRIAN,A  
BREATH TEST OPERATOR AGENCY: Gilchrist County SO  
ARRESTING OFFICER: REEVES,DAVID  
ARRESTING OFFICER AGENCY: Trenton PD  
LAST AGENCY INSPECTION: 09/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1224982  
DRY GAS EXPIRY DATE: 10/31/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/24/2021 03:10:19

---

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.079	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.205	03:03
Breath Volume		1.175	03:03
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.228	03:05
Breath Volume		1.734	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 3	0.227	03:08
Breath Volume		1.511	03:08
Air Blank		0.000	03:09
Control Test		0.078	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

---

Data Download Date/Time: 11/11/2021 16:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLEN, JOSHUA, B  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: ALLEN, JOSHUA  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 09/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: MCCAM

DATE/TIME OF TEST: 10/16/2021 23:55:13

---

Diagnostics Check		OK	23:44	
Air Blank		0.000	23:44	
Control Test		0.078	23:45	
Air Blank		0.000	23:45	
Subject Sample	# 1	0.114	23:48	Volume Not Met
Breath Volume		0.593	23:48	
Air Blank		0.000	23:49	
Air Blank		0.000	23:51	
Subject Sample	# 2		23:53	Subject Test Refused
Breath Volume		0.000	23:53	
Air Blank		0.000	23:54	
Control Test		0.076	23:54	
Air Blank		0.000	23:55	
Diagnostics Check		OK	23:55	

---

Data Download Date/Time: 11/02/2021 14:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,KEVIN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DAVIS,KEVIN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 09/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 10/25/2021 19:32:29

---

Diagnostics Check	OK	19:30	
Air Blank	0.000	19:31	
Control Test	0.078	19:31	
Air Blank	0.000	19:31	
		19:32	RFI Detect
Air Blank	0.000	19:32	

---

Data Download Date/Time: 11/02/2021 14:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,KEVIN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DAVIS,KEVIN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 09/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 10/25/2021 19:40:30

---

Diagnostics Check	OK	19:34
Air Blank	0.000	19:34
Control Test	0.079	19:34
Air Blank	0.000	19:35
Subject Sample # 1	0.262	19:35
Breath Volume	3.304	19:35
Air Blank	0.000	19:36
Air Blank	0.000	19:38
Subject Sample # 2	0.260	19:38
Breath Volume	3.160	19:38
Air Blank	0.000	19:39
Control Test	0.077	19:39
Air Blank	0.000	19:40
Diagnostics Check	OK	19:40

---

Data Download Date/Time: 11/02/2021 14:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCONNELL,JAIRUS,S  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MCCONNELL,JAIRUS  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 09/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 10/30/2021 02:09:28

---

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.078	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.072	02:04
Breath Volume	2.343	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.073	02:08
Breath Volume	2.285	02:08
Air Blank	0.000	02:08
Control Test	0.078	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

---

Data Download Date/Time: 11/02/2021 14:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CUPPLES,ZACHARY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 10/11/2021 00:51:34

---

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.080	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.000	00:47
Breath Volume	2.101	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.000	00:50
Breath Volume	2.636	00:50
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

---

Data Download Date/Time: 11/03/2021 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBINS II,JERRY,M  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: BLACK,KURT  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: GILSTRAP

DATE/TIME OF TEST: 10/12/2021 22:45:06

---

Diagnostics Check	OK	22:38
Air Blank	0.000	22:39
Control Test	0.080	22:39
Air Blank	0.000	22:40
Subject Sample # 1	0.187	22:40
Breath Volume	3.878	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.197	22:43
Breath Volume	4.175	22:43
Air Blank	0.000	22:44
Control Test	0.078	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:45

---

Data Download Date/Time: 11/03/2021 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYNN,JAMISON,C  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: LYNN,JAMISON  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 18:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: BATSON

DATE/TIME OF TEST: 10/15/2021 19:36:58

---

Diagnostics Check	OK	19:30
Air Blank	0.000	19:30
Control Test	0.080	19:30
Air Blank	0.000	19:31
Subject Sample # 1	0.127	19:32
Breath Volume	2.101	19:32
Air Blank	0.000	19:32
Air Blank	0.000	19:34
Subject Sample # 2	0.123	19:35
Breath Volume	1.746	19:35
Air Blank	0.000	19:36
Control Test	0.081	19:36
Air Blank	0.000	19:36
Diagnostics Check	OK	19:36

---

Data Download Date/Time: 11/03/2021 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUNTER,REGEN,B  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: GUNTER,REGEN  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 10/15/2021 20:54:03

---

Diagnostics Check	OK	20:45
Air Blank	0.000	20:45
Control Test	0.081	20:45
Air Blank	0.000	20:46
Subject Sample # 1	0.133	20:49
Breath Volume	1.351	20:49
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.132	20:52
Breath Volume	1.277	20:52
Air Blank	0.000	20:53
Control Test	0.078	20:53
Air Blank	0.000	20:53
Diagnostics Check	OK	20:54

---

Data Download Date/Time: 11/03/2021 12:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAREY,ANDREW,S  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CAREY,ANDREW  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 13:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: COMER

DATE/TIME OF TEST: 10/16/2021 15:23:06

---

Diagnostics Check		OK	15:16
Air Blank		0.000	15:17
Control Test		0.081	15:17
Air Blank		0.000	15:17
Subject Sample	# 1	0.245	15:18
Breath Volume		2.156	15:18
Air Blank		0.000	15:19
Air Blank		0.000	15:21
Subject Sample	# 2	0.250	15:21
Breath Volume		2.210	15:21
Air Blank		0.000	15:22
Control Test		0.080	15:22
Air Blank		0.000	15:23
Diagnostics Check		OK	15:23

---

Data Download Date/Time: 11/03/2021 12:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CUPPLES,ZACHARY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 10/16/2021 22:23:46

---

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:21	
Control Test	0.080	22:21	
Air Blank	0.000	22:22	
Subject Sample	# 1	22:22	Subject Test Refused
Breath Volume	0.000	22:22	
Air Blank	0.000	22:22	
Control Test	0.080	22:23	
Air Blank	0.000	22:23	
Diagnostics Check	OK	22:23	

---

Data Download Date/Time: 11/03/2021 12:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CUPPLES,ZACHARY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: NUTWELL

DATE/TIME OF TEST: 10/17/2021 02:10:25

---

Diagnostics Check	OK	02:07	
Air Blank	0.000	02:08	
Control Test	0.080	02:08	
Air Blank	0.000	02:08	
Subject Sample	# 1	02:09	Subject Test Refused
Breath Volume	0.000	02:09	
Air Blank	0.000	02:09	
Control Test	0.081	02:09	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

---

Data Download Date/Time: 11/03/2021 12:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPENCER, RAYMOND, L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: SPENCER, RAYMOND  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: BUCKLEY

DATE/TIME OF TEST: 10/17/2021 04:12:02

---

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:08	
Control Test	0.080	04:08	
Air Blank	0.000	04:09	
Subject Sample	# 1	04:10	Subject Test Refused
Breath Volume	0.000	04:10	
Air Blank	0.000	04:11	
Control Test	0.081	04:11	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:12	

---

Data Download Date/Time: 11/03/2021 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CUPPLES,ZACHARY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: CORKRAN

DATE/TIME OF TEST: 10/19/2021 02:49:15

---

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.124	02:44
Breath Volume	2.429	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.128	02:47
Breath Volume	3.398	02:47
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

---

Data Download Date/Time: 11/03/2021 12:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PINTO,FELIPE,G  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: PINTO,FELIPE  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/20/2021 23:01:28

---

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:53	
Control Test	0.081	22:53	
Air Blank	0.000	22:54	
Subject Sample	# 1 0.196	22:57	Volume Not Met
Breath Volume	0.484	22:57	
Air Blank	0.000	22:57	
Air Blank	0.000	22:59	
Subject Sample	# 2	23:00	Subject Test Refused
Breath Volume	0.000	23:00	
Air Blank	0.000	23:00	
Control Test	0.079	23:00	
Air Blank	0.000	23:01	
Diagnostics Check	OK	23:01	

---

Data Download Date/Time: 11/03/2021 12:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZOLLER,J,G  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CHAUVET,N  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: COMPTON

DATE/TIME OF TEST: 10/21/2021 05:27:23

---

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.080	05:17
Air Blank	0.000	05:17
Subject Sample # 1	0.134	05:19
Breath Volume	2.570	05:19
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample # 2	0.165	05:22
Breath Volume	1.464	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 3	0.162	05:25
Breath Volume	2.316	05:25
Air Blank	0.000	05:26
Control Test	0.078	05:26
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

---

Data Download Date/Time: 11/03/2021 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COPELAND,JUSTIN,M  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: STACHEWICZ,RILEY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: LAFFIN

DATE/TIME OF TEST: 10/22/2021 01:52:50

---

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.162	01:48
Breath Volume	2.632	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.174	01:51
Breath Volume	3.492	01:51
Air Blank	0.000	01:51
Control Test	0.080	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

---

Data Download Date/Time: 11/03/2021 12:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,KAYLAN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: DAVIS,KAYLAN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 08:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:43

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 10/22/2021 09:11:37

---

Diagnostics Check	OK	09:05
Air Blank	0.000	09:05
Control Test	0.080	09:06
Air Blank	0.000	09:06
Subject Sample # 1	0.000	09:07
Breath Volume	3.125	09:07
Air Blank	0.000	09:07
Air Blank	0.000	09:09
Subject Sample # 2	0.000	09:10
Breath Volume	2.902	09:10
Air Blank	0.000	09:10
Control Test	0.080	09:11
Air Blank	0.000	09:11
Diagnostics Check	OK	09:11

---

Data Download Date/Time: 11/03/2021 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORENE,AIDAN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: NORENE,AIDAN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: ASKEGARD

DATE/TIME OF TEST: 10/23/2021 01:02:06

---

Diagnostics Check		OK	00:54
Air Blank		0.000	00:55
Control Test		0.081	00:55
Air Blank		0.000	00:55
Subject Sample	# 1	0.134	00:57
Breath Volume		2.816	00:57
Air Blank		0.000	00:58
Air Blank		0.000	00:59
Subject Sample	# 2	0.134	01:00
Breath Volume		1.425	01:00
Air Blank		0.000	01:01
Control Test		0.081	01:01
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

---

Data Download Date/Time: 11/03/2021 12:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALEXANDER,LLOYD,A  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: ALEXANDER,LLOYD  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: BAUM

DATE/TIME OF TEST: 10/24/2021 21:09:45

---

Diagnostics Check	OK	21:05	
Air Blank	0.000	21:05	
Control Test	0.079	21:05	
Air Blank	0.000	21:06	
Subject Sample	# 1	21:08	Subject Test Refused
Breath Volume	0.320	21:08	
Air Blank	0.000	21:08	
Control Test	0.080	21:09	
Air Blank	0.000	21:09	
Diagnostics Check	OK	21:09	

---

Data Download Date/Time: 11/03/2021 12:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRANTHAM,WILLIAM,H  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: GRANTHAM,WILLIAM  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 10/26/2021 01:13:44

---

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.080	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.118	01:09
Breath Volume	1.894	01:09
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.120	01:12
Breath Volume	2.226	01:12
Air Blank	0.000	01:12
Control Test	0.080	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

---

Data Download Date/Time: 11/03/2021 12:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYNN,JAMISON,C  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: LYNN,JAMISON  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: KILLOUGH

DATE/TIME OF TEST: 10/26/2021 21:23:05

---

Diagnostics Check	OK	21:14
Air Blank	0.000	21:15
Control Test	0.080	21:15
Air Blank	0.000	21:16
Subject Sample # 1	0.172	21:17
Breath Volume	4.718	21:17
Air Blank	0.000	21:18
Air Blank	0.000	21:20
Subject Sample # 2	0.162	21:21
Breath Volume	3.363	21:21
Air Blank	0.000	21:22
Control Test	0.079	21:22
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

---

Data Download Date/Time: 11/03/2021 12:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARCUM,JOSEPH,K  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: ALEXANDER,C  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 10/29/2021 00:44:34

---

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.080	00:38
Air Blank	0.000	00:39
Subject Sample # 1	0.140	00:39
Breath Volume	2.539	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.137	00:43
Breath Volume	2.570	00:43
Air Blank	0.000	00:43
Control Test	0.080	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

---

Data Download Date/Time: 11/03/2021 12:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVERA,DANTE,E  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CHAUVET,NIKOLAS  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: WEBER

DATE/TIME OF TEST: 10/29/2021 04:53:36

---

Diagnostics Check		OK	04:41	
Air Blank		0.000	04:41	
Control Test		0.080	04:42	
Air Blank		0.000	04:42	
Subject Sample	# 1	0.266	04:45	Volume Not Met
Breath Volume		0.160	04:45	
Air Blank		0.000	04:46	
Air Blank		0.000	04:48	
Subject Sample	# 2	0.283	04:49	No .020 agreement
Breath Volume		3.023	04:49	
Air Blank		0.000	04:49	
Air Blank		0.000	04:51	
Subject Sample	# 3	0.254	04:52	No .020 agreement
Breath Volume		2.070	04:52	
Air Blank		0.000	04:52	
Control Test		0.077	04:53	
Air Blank		0.000	04:53	
Diagnostics Check		OK	04:53	

---

Data Download Date/Time: 11/03/2021 12:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CUPPLES,ZACHARY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 10/03/2021 04:46:49

---

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.084	04:41
Air Blank	0.000	04:41
Subject Sample # 1	0.000	04:42
Breath Volume	2.011	04:42
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.000	04:45
Breath Volume	1.269	04:45
Air Blank	0.000	04:45
Control Test	0.085	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

---

Data Download Date/Time: 11/08/2021 14:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CUPPLES,ZACHARY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 10/04/2021 05:49:16

---

Diagnostics Check	OK	05:42
Air Blank	0.000	05:42
Control Test	0.084	05:43
Air Blank	0.000	05:43
Subject Sample # 1	0.000	05:44
Breath Volume	2.882	05:44
Air Blank	0.000	05:45
Air Blank	0.000	05:47
Subject Sample # 2	0.000	05:47
Breath Volume	1.230	05:47
Air Blank	0.000	05:48
Control Test	0.082	05:48
Air Blank	0.000	05:49
Diagnostics Check	OK	05:49

---

Data Download Date/Time: 11/08/2021 14:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRANTHAM,WILLIAM,H  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: GRANTHAM,WILLIAM  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/07/2021 01:21:19

---

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:14	
Control Test	0.085	01:14	
Air Blank	0.000	01:15	
Subject Sample # 1	0.093	01:16	
Breath Volume	1.921	01:16	
Air Blank	0.000	01:17	
Air Blank	0.000	01:19	
Subject Sample # 2	0.089	01:19	
Breath Volume	2.398	01:19	
Air Blank	0.000	01:20	
Control Test	0.078	01:20	Interferent Detect
Air Blank	0.000	01:21	

---

Data Download Date/Time: 11/08/2021 14:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPENCER, RAYMOND, L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: SPENCER, RAYMOND  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 10/10/2021 01:28:58

---

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:28	
Control Test	0.078	01:28	Interferent Detect
Air Blank		01:28	RFI Detect

---

Data Download Date/Time: 11/08/2021 14:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPENCER, RAYMOND, L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: SPENCER, RAYMOND  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 10/10/2021 01:33:14

---

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.085	01:32	
Air Blank		01:32	Ambient Fail
Air Blank	0.000	01:33	

---

Data Download Date/Time: 11/08/2021 14:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPENCER, RAYMOND, L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: SPENCER, RAYMOND  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 10/10/2021 01:52:27

---

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:35	
Control Test	0.082	01:35	
Air Blank	0.000	01:36	
Subject Sample # 1	0.000	01:39	
Breath Volume	2.074	01:39	
Air Blank	0.000	01:39	
Air Blank	0.000	01:41	
Subject Sample # 2		01:44	No Sample Provided
Breath Volume	0.000	01:44	
Air Blank	0.000	01:45	
Air Blank	0.000	01:47	
Subject Sample # 3		01:50	No Sample Provided
Breath Volume	0.000	01:50	
Air Blank	0.000	01:51	
Control Test	0.084	01:51	
Air Blank	0.000	01:51	
Diagnostics Check	Fail	01:51	Analytical Stability Test
Air Blank	0.000	01:52	

---

Data Download Date/Time: 11/08/2021 14:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CUPPLES,ZACHARY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 10/11/2021 00:37:13

---

Diagnostics Check	OK	00:36	
Air Blank		00:36	RFI Detect
Air Blank	0.000	00:37	

---

Data Download Date/Time: 11/08/2021 14:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CUPPLES,ZACHARY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 10/11/2021 00:41:05

---

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:39	
Control Test	0.082	00:40	
Air Blank		00:40	Ambient Fail
Air Blank	0.000	00:41	

---

Data Download Date/Time: 11/08/2021 14:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBINS II,JERRY,M  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: BLACK,KURT  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: GILSTRAP

DATE/TIME OF TEST: 10/12/2021 22:25:01

---

Diagnostics Check	OK	22:23	
Air Blank	0.000	22:24	
Control Test	0.078	22:24	Interferent Detect
Air Blank	0.000	22:25	

---

Data Download Date/Time: 11/08/2021 14:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBINS II,JERRY,M  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: BLACK,KURT  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: GILSTRAP

DATE/TIME OF TEST: 10/12/2021 22:33:44

---

Diagnostics Check	OK	22:27	
Air Blank	0.000	22:27	
Control Test	0.084	22:28	
Air Blank	0.000	22:28	
Subject Sample	# 1 0.198	22:29	
Breath Volume	4.535	22:29	
Air Blank	0.000	22:29	
Air Blank	0.000	22:31	
Subject Sample	# 2 0.198	22:32	
Breath Volume	4.457	22:32	
Air Blank	0.000	22:32	
Control Test	0.084	22:33	Interferent Detect
Air Blank	0.000	22:33	

---

Data Download Date/Time: 11/08/2021 14:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWELL,KEENAN,B  
BREATH TEST OPERATOR AGENCY: Marianna PD  
ARRESTING OFFICER: SULLIVAN,ANDER  
ARRESTING OFFICER AGENCY: Sneads PD  
LAST AGENCY INSPECTION: 10/11/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1346146  
DRY GAS EXPIRY DATE: 07/07/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 10/24/2021 05:14:01

---

Diagnostics Check	OK	05:06
Air Blank	0.000	05:07
Control Test	0.083	05:07
Air Blank	0.000	05:08
Subject Sample # 1	0.015	05:09
Breath Volume	2.105	05:09
Air Blank	0.000	05:09
Air Blank	0.000	05:11
Subject Sample # 2	0.013	05:12
Breath Volume	1.906	05:12
Air Blank	0.000	05:13
Control Test	0.082	05:13
Air Blank	0.000	05:13
Diagnostics Check	OK	05:14

---

Data Download Date/Time: 12/13/2021 10:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMYRNIOS,GUS,T  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: SMYRNIOS,GUS  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 09/22/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34520080A1  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: SEAY

DATE/TIME OF TEST: 10/09/2021 20:27:27

---

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.080	20:22
Air Blank	0.000	20:22
Subject Sample # 1	0.196	20:23
Breath Volume	2.589	20:23
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.207	20:25
Breath Volume	2.652	20:25
Air Blank	0.000	20:26
Control Test	0.078	20:26
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

---

Data Download Date/Time: 10/20/2021 15:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOODALL,CORY,D  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: MANSON,TIMOTHY  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 09/17/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1305520  
DRY GAS EXPIRY DATE: 04/01/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: RUFFIN

DATE/TIME OF TEST: 10/02/2021 03:01:22

---

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:54	
Control Test	0.082	02:54	
Air Blank	0.000	02:55	
Subject Sample	# 1 0.092	02:55	
Breath Volume	1.882	02:55	
Air Blank	0.000	02:56	
Air Blank	0.000	02:58	
Subject Sample	# 2	03:00	Sequence Aborted
Breath Volume	0.535	03:00	
Air Blank	0.000	03:01	

---

Data Download Date/Time: 10/13/2021 12:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROSIUS,STANLEY,W  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: SMITH,C  
ARRESTING OFFICER AGENCY: Panama City PD  
LAST AGENCY INSPECTION: 09/17/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1305520  
DRY GAS EXPIRY DATE: 04/01/2023  
VIOLATION CODE: DUI  
ARREST TIME: 14:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:54

SUBJECT'S LAST NAME: DEASON

DATE/TIME OF TEST: 10/02/2021 15:26:43

---

Diagnostics Check		OK	15:14
Air Blank		0.000	15:14
Control Test		0.079	15:15
Air Blank		0.000	15:15
Subject Sample	# 1	0.295	15:16
Breath Volume		2.410	15:16
Air Blank		0.000	15:17
Air Blank		0.000	15:19
Subject Sample	# 2	0.262	15:20
Breath Volume		1.687	15:20
Air Blank		0.000	15:21
Air Blank		0.000	15:23
Subject Sample	# 3	0.245	15:25
Breath Volume		1.414	15:25
Air Blank		0.000	15:25
Control Test		0.078	15:26
Air Blank		0.000	15:26
Diagnostics Check		OK	15:26

---

Data Download Date/Time: 10/13/2021 12:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRICE,CARLON,E  
BREATH TEST OPERATOR AGENCY: Bay County SO  
ARRESTING OFFICER: PRICE,CARLON  
ARRESTING OFFICER AGENCY: Bay County SO  
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1305520  
DRY GAS EXPIRY DATE: 04/01/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/14/2021 02:13:42

---

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.081	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.118	02:07
Breath Volume	1.398	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.108	02:12
Breath Volume	2.238	02:12
Air Blank	0.000	02:12
Control Test	0.082	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

---

Data Download Date/Time: 11/17/2021 13:55

































































































































































































# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANDERS,DEPUTY,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: SANDERS,DEPUTY J  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: BRILL

DATE/TIME OF TEST: 10/11/2021 23:52:40

---

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.079	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.269	23:45
Breath Volume	3.886	23:45
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.235	23:48
Breath Volume	1.773	23:48
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 3	0.261	23:51
Breath Volume	3.558	23:51
Air Blank	0.000	23:51
Control Test	0.078	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

---

Data Download Date/Time: 10/27/2021 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NAMORS,JASON,R  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BERNIER,N  
ARRESTING OFFICER AGENCY: Sarasota PD  
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: GARCIA-TOVAR

DATE/TIME OF TEST: 10/12/2021 01:04:37

---

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:02	
Control Test	0.078	01:02	
Air Blank	0.000	01:03	
Subject Sample	# 1	01:03	Subject Test Refused
Breath Volume	0.000	01:03	
Air Blank	0.000	01:03	
Control Test	0.078	01:04	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

---

Data Download Date/Time: 10/27/2021 09:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERSON,CORY,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: PIERSON,CORY  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: HOAGLUND

DATE/TIME OF TEST: 10/13/2021 02:35:51

---

Diagnostics Check		OK	02:29
Air Blank		0.000	02:29
Control Test		0.080	02:30
Air Blank		0.000	02:30
Subject Sample	# 1	0.257	02:31
Breath Volume		4.496	02:31
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.249	02:34
Breath Volume		3.574	02:34
Air Blank		0.000	02:34
Control Test		0.079	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

---

Data Download Date/Time: 10/27/2021 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENCKLE,DEP,T  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BRINTON,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:03

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 10/14/2021 16:31:04

---

Diagnostics Check	OK	16:24
Air Blank	0.000	16:25
Control Test	0.079	16:25
Air Blank	0.000	16:25
Subject Sample # 1	0.270	16:26
Breath Volume	2.183	16:26
Air Blank	0.000	16:27
Air Blank	0.000	16:29
Subject Sample # 2	0.267	16:29
Breath Volume	2.394	16:29
Air Blank	0.000	16:30
Control Test	0.080	16:30
Air Blank	0.000	16:30
Diagnostics Check	OK	16:31

---

Data Download Date/Time: 10/27/2021 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: DAVENPORT,DANIEL  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: RADLOFF

DATE/TIME OF TEST: 10/15/2021 19:47:14

---

Diagnostics Check	OK	19:40
Air Blank	0.000	19:41
Control Test	0.078	19:41
Air Blank	0.000	19:41
Subject Sample # 1	0.184	19:42
Breath Volume	3.468	19:42
Air Blank	0.000	19:42
Air Blank	0.000	19:44
Subject Sample # 2	0.187	19:45
Breath Volume	2.132	19:45
Air Blank	0.000	19:46
Control Test	0.078	19:46
Air Blank	0.000	19:47
Diagnostics Check	OK	19:47

---

Data Download Date/Time: 10/27/2021 09:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIMITROV,OFFICER,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: NELSON,OFFICER  
ARRESTING OFFICER AGENCY: Venice PD  
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/20/2021 02:30:52

---

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.009	02:26
Breath Volume	1.242	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.009	02:29
Breath Volume	1.503	02:29
Air Blank	0.000	02:29
Control Test	0.079	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

---

Data Download Date/Time: 10/27/2021 09:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANDRADE,DEPUTY,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: LUTZ,SERGEANT  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 14:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:17

SUBJECT'S LAST NAME: BLOTNER

DATE/TIME OF TEST: 10/21/2021 15:57:10

---

Diagnostics Check		OK	15:45	
Air Blank		0.000	15:46	
Control Test		0.079	15:46	
Air Blank		0.000	15:47	
Subject Sample	# 1	0.000	15:50	Volume Not Met
Breath Volume		0.437	15:50	
Air Blank		0.000	15:50	
Air Blank		0.000	15:52	
Subject Sample	# 2	0.000	15:55	Volume Not Met
Breath Volume		0.437	15:55	
Air Blank		0.000	15:56	
Control Test		0.079	15:56	
Air Blank		0.000	15:57	
Diagnostics Check		OK	15:57	

---

Data Download Date/Time: 10/27/2021 09:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANDERS,DEPUTY,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: SANDERS,DEPUTY J  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 10/22/2021 03:38:20

---

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.078	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.089	03:33
Breath Volume	2.039	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.086	03:36
Breath Volume	2.136	03:36
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

---

Data Download Date/Time: 10/27/2021 09:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERSON,CORY,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: PIERSON,CORY  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: BUDIER

DATE/TIME OF TEST: 10/22/2021 05:27:31

---

Diagnostics Check	OK	05:21
Air Blank	0.000	05:21
Control Test	0.079	05:22
Air Blank	0.000	05:22
Subject Sample # 1	0.194	05:23
Breath Volume	1.699	05:23
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.190	05:25
Breath Volume	1.531	05:25
Air Blank	0.000	05:26
Control Test	0.079	05:26
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

---

Data Download Date/Time: 10/27/2021 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ACKLEY,DEP,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: ACKLEY,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 10/23/2021 00:35:55

---

Diagnostics Check	OK	00:29
Air Blank	0.000	00:29
Control Test	0.080	00:29
Air Blank	0.000	00:30
Subject Sample # 1	0.197	00:31
Breath Volume	1.542	00:31
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.194	00:34
Breath Volume	2.507	00:34
Air Blank	0.000	00:35
Control Test	0.077	00:35
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

---

Data Download Date/Time: 10/27/2021 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENCKLE,DEP,T  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BRENCKLE,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/23/2021 01:55:26

---

Diagnostics Check	OK	01:51
Air Blank	0.000	01:52
Control Test	0.080	01:52
Air Blank	0.000	01:52
Subject Sample # 1	0.071	01:53
Breath Volume	1.875	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:55

---

Data Download Date/Time: 10/27/2021 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,DAVID,A  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: RAINSFORD,BRITTANY  
ARRESTING OFFICER AGENCY: Eustis PD  
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1031121  
DRY GAS EXPIRY DATE: 10/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: DE LACRUZ

DATE/TIME OF TEST: 10/03/2021 02:48:06

---

Diagnostics Check	OK	02:47	
Air Blank		02:47	RFI Detect
Air Blank		02:48	RFI Detect

---

Data Download Date/Time: 10/11/2021 13:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,DAVID,A  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: RAINSFORD,BRITTANY  
ARRESTING OFFICER AGENCY: Eustis PD  
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1031121  
DRY GAS EXPIRY DATE: 10/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: DE LA CRUZ

DATE/TIME OF TEST: 10/03/2021 02:52:31

---

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:51	
Control Test	0.081	02:51	
Air Blank		02:52	RFI Detect
Air Blank		02:52	RFI Detect

---

Data Download Date/Time: 10/11/2021 13:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,DAVID,A  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: RAINSFORD,BRITTANY  
ARRESTING OFFICER AGENCY: Eustis PD  
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1031121  
DRY GAS EXPIRY DATE: 10/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: DE LA CRUZ

DATE/TIME OF TEST: 10/03/2021 03:01:47

---

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.081	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.223	02:57
Breath Volume		2.207	02:57
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.205	03:00
Breath Volume		2.054	03:00
Air Blank		0.000	03:00
Control Test		0.081	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

---

Data Download Date/Time: 10/11/2021 13:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STROUP,STEVE,A  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: BURTZLAFF,SYDNEE  
ARRESTING OFFICER AGENCY: Eustis PD  
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1031121  
DRY GAS EXPIRY DATE: 10/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: PALACIO

DATE/TIME OF TEST: 10/11/2021 02:52:24

---

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:49	
Control Test	0.079	02:50	
Air Blank	0.000	02:50	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:51	
Control Test	0.079	02:51	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

---

Data Download Date/Time: 10/11/2021 13:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,DAVID,A  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: PEREZ,DAVID  
ARRESTING OFFICER AGENCY: Eustis PD  
LAST AGENCY INSPECTION: 10/11/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1031121  
DRY GAS EXPIRY DATE: 10/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: FOUNTAIN

DATE/TIME OF TEST: 10/15/2021 20:51:34

---

Diagnostics Check	OK	20:48	
Air Blank	0.000	20:49	
Control Test	0.081	20:49	
Air Blank	0.000	20:50	
Subject Sample	# 1	20:50	Subject Test Refused
Breath Volume	0.000	20:50	
Air Blank	0.000	20:50	
Control Test	0.081	20:50	
Air Blank	0.000	20:51	
Diagnostics Check	OK	20:51	

---

Data Download Date/Time: 12/01/2021 12:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,DAVID,A  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: FREITAS,MARCIA  
ARRESTING OFFICER AGENCY: Eustis PD  
LAST AGENCY INSPECTION: 10/11/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1031121  
DRY GAS EXPIRY DATE: 10/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: DICK

DATE/TIME OF TEST: 10/22/2021 03:07:53

---

Diagnostics Check		OK	03:01
Air Blank		0.000	03:01
Control Test		0.082	03:02
Air Blank		0.000	03:02
Subject Sample	# 1	0.204	03:03
Breath Volume		3.605	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:05
Subject Sample	# 2	0.198	03:06
Breath Volume		3.066	03:06
Air Blank		0.000	03:06
Control Test		0.081	03:07
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

---

Data Download Date/Time: 12/01/2021 12:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STROUP,STEVE,A  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: COE,BARRY  
ARRESTING OFFICER AGENCY: Eustis PD  
LAST AGENCY INSPECTION: 10/11/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1031121  
DRY GAS EXPIRY DATE: 10/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: PARENT

DATE/TIME OF TEST: 10/31/2021 02:21:07

---

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.081	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.117	02:15
Breath Volume	2.148	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.099	02:19
Breath Volume	1.218	02:19
Air Blank	0.000	02:20
Control Test	0.081	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

---

Data Download Date/Time: 12/01/2021 12:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STROUP,STEVE,A  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: BURTZLAFF,SYDNEE  
ARRESTING OFFICER AGENCY: Eustis PD  
LAST AGENCY INSPECTION: 10/11/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1031121  
DRY GAS EXPIRY DATE: 10/12/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HASKINS

DATE/TIME OF TEST: 10/31/2021 03:58:14

---

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:54	
Control Test	0.080	03:55	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank	0.000	03:57	
Control Test	0.081	03:57	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

---

Data Download Date/Time: 12/01/2021 12:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENCKLE,DEP,T  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BRENCKLE,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: MOCCO

DATE/TIME OF TEST: 10/10/2021 01:24:17

---

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.000	01:20
Breath Volume	2.328	01:20
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.000	01:22
Breath Volume	2.121	01:22
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

---

Data Download Date/Time: 10/27/2021 09:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENCKLE,DEP,T  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BRENCKLE,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 10/15/2021 00:40:20

---

Diagnostics Check		OK	00:34
Air Blank		0.000	00:34
Control Test		0.079	00:35
Air Blank		0.000	00:35
Subject Sample	# 1	0.154	00:36
Breath Volume		2.074	00:36
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.154	00:38
Breath Volume		2.152	00:38
Air Blank		0.000	00:39
Control Test		0.078	00:39
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

---

Data Download Date/Time: 10/27/2021 09:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,DEPUTY,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MARTIN,KYLE  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: TROTTA

DATE/TIME OF TEST: 10/20/2021 02:26:30

---

Diagnostics Check		OK	02:18
Air Blank		0.000	02:19
Control Test		0.079	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.179	02:22
Breath Volume		1.730	02:22
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.185	02:25
Breath Volume		1.812	02:25
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

---

Data Download Date/Time: 10/27/2021 09:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENCKLE,DEP,T  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BRENCKLE,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: BILL

DATE/TIME OF TEST: 10/22/2021 21:11:11

---

Diagnostics Check	OK	21:04
Air Blank	0.000	21:05
Control Test	0.079	21:05
Air Blank	0.000	21:05
Subject Sample # 1	0.233	21:06
Breath Volume	1.937	21:06
Air Blank	0.000	21:07
Air Blank	0.000	21:08
Subject Sample # 2	0.238	21:09
Breath Volume	1.480	21:09
Air Blank	0.000	21:10
Control Test	0.079	21:10
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

---

Data Download Date/Time: 10/27/2021 09:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENCKLE,DEP,T  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BRENCKLE,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/23/2021 02:26:24

---

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.065	02:22
Breath Volume	2.109	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.066	02:24
Breath Volume	2.578	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

---

Data Download Date/Time: 10/27/2021 09:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,DEPUTY,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MARTIN,KYLE  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: SCHAFFER

DATE/TIME OF TEST: 10/24/2021 19:51:42

---

Diagnostics Check	OK	19:36	
Air Blank	0.000	19:36	
Control Test	0.079	19:37	
Air Blank	0.000	19:37	
Subject Sample	# 1 0.165	19:40	
Breath Volume	1.496	19:40	
Air Blank	0.000	19:41	
Air Blank	0.000	19:43	
Subject Sample	# 2 0.169	19:46	Volume Not Met
Breath Volume	0.535	19:46	
Air Blank	0.000	19:46	
Air Blank	0.000	19:48	
Subject Sample	# 3 0.163	19:50	
Breath Volume	1.199	19:50	
Air Blank	0.000	19:50	
Control Test	0.077	19:51	
Air Blank	0.000	19:51	
Diagnostics Check	OK	19:51	

---

Data Download Date/Time: 10/27/2021 09:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,SHAUN,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MARTIN,DEPUTY  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: COPPOCK

DATE/TIME OF TEST: 10/25/2021 02:49:10

---

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.151	02:44
Breath Volume	4.007	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.144	02:47
Breath Volume	1.382	02:47
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

---

Data Download Date/Time: 10/27/2021 09:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MATTHEWS,DEPUTY,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: KLEIN,OFFICER  
ARRESTING OFFICER AGENCY: North Port PD  
LAST AGENCY INSPECTION: 10/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: GOSSETT

DATE/TIME OF TEST: 10/25/2021 23:09:26

---

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.080	23:03
Air Blank		0.000	23:03
Subject Sample	# 1	0.173	23:04
Breath Volume		1.785	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.165	23:07
Breath Volume		1.812	23:07
Air Blank		0.000	23:08
Control Test		0.077	23:08
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

---

Data Download Date/Time: 10/27/2021 09:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANDERS,DEPUTY,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: SANDERS,DEPUTY J  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: GREENWOOD

DATE/TIME OF TEST: 10/26/2021 00:47:35

---

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.079	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.144	00:43
Breath Volume	1.312	00:43
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.144	00:46
Breath Volume	2.238	00:46
Air Blank	0.000	00:46
Control Test	0.079	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

---

Data Download Date/Time: 10/27/2021 09:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENCKLE,DEP,T  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: BRENCKLE,DEP  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: BONTRAGER

DATE/TIME OF TEST: 10/27/2021 21:16:34

---

Diagnostics Check		OK	21:09
Air Blank		0.000	21:10
Control Test		0.079	21:10
Air Blank		0.000	21:11
Subject Sample	# 1	0.115	21:12
Breath Volume		2.082	21:12
Air Blank		0.000	21:12
Air Blank		0.000	21:14
Subject Sample	# 2	0.122	21:15
Breath Volume		2.253	21:15
Air Blank		0.000	21:15
Control Test		0.079	21:15
Air Blank		0.000	21:16
Diagnostics Check		OK	21:16

---

Data Download Date/Time: 11/30/2021 18:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,DEPUTY,  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MARTIN,KYLE  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: MANLEY

DATE/TIME OF TEST: 10/28/2021 02:00:36

---

Diagnostics Check	OK	01:55	
Air Blank	0.000	01:55	
Control Test	0.079	01:55	
Air Blank	0.000	01:56	
Subject Sample	# 1	01:59	Subject Test Refused
Breath Volume	0.000	01:59	
Air Blank	0.000	01:59	
Control Test	0.080	02:00	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

---

Data Download Date/Time: 11/30/2021 18:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,DEP,R  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MILLER,DEP R  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: HUCULAK

DATE/TIME OF TEST: 10/31/2021 02:00:53

---

Diagnostics Check		OK	01:53
Air Blank		0.000	01:54
Control Test		0.080	01:54
Air Blank		0.000	01:55
Subject Sample	# 1	0.165	01:55
Breath Volume		1.875	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.174	01:59
Breath Volume		2.300	01:59
Air Blank		0.000	01:59
Control Test		0.078	02:00
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

---

Data Download Date/Time: 11/30/2021 18:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURLEY,S,M  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: ZWIEFELHOFER,A  
ARRESTING OFFICER AGENCY: North Port PD  
LAST AGENCY INSPECTION: 10/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 10/31/2021 03:18:07

---

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.131	03:13
Breath Volume	1.773	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.124	03:16
Breath Volume	2.554	03:16
Air Blank	0.000	03:17
Control Test	0.079	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

---

Data Download Date/Time: 11/30/2021 18:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANDRADE,DEPUTY,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: MONTANEZ,V  
ARRESTING OFFICER AGENCY: Venice PD  
LAST AGENCY INSPECTION: 10/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 12:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: CANNADY

DATE/TIME OF TEST: 10/31/2021 14:34:26

---

Diagnostics Check	OK	14:28
Air Blank	0.000	14:28
Control Test	0.080	14:29
Air Blank	0.000	14:29
Subject Sample # 1	0.179	14:30
Breath Volume	1.671	14:30
Air Blank	0.000	14:30
Air Blank	0.000	14:32
Subject Sample # 2	0.177	14:32
Breath Volume	2.140	14:32
Air Blank	0.000	14:33
Control Test	0.079	14:33
Air Blank	0.000	14:34
Diagnostics Check	OK	14:34

---

Data Download Date/Time: 11/30/2021 18:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANDERS,DEPUTY,J  
BREATH TEST OPERATOR AGENCY: Sarasota County SO  
ARRESTING OFFICER: SANDERS,DEPUTY J.  
ARRESTING OFFICER AGENCY: Sarasota County SO  
LAST AGENCY INSPECTION: 10/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: EARLE

DATE/TIME OF TEST: 10/31/2021 23:51:04

---

Diagnostics Check		OK	23:38	
Air Blank		0.000	23:39	
Control Test		0.079	23:39	
Air Blank		0.000	23:39	
Subject Sample	# 1	0.197	23:43	Volume Not Met
Breath Volume		1.097	23:43	
Air Blank		0.000	23:43	
Air Blank		0.000	23:45	
Subject Sample	# 2	0.219	23:45	
Breath Volume		1.617	23:45	
Air Blank		0.000	23:46	
Air Blank		0.000	23:48	
Subject Sample	# 3	0.237	23:49	
Breath Volume		1.593	23:49	
Air Blank		0.000	23:50	
Control Test		0.078	23:50	
Air Blank		0.000	23:50	
Diagnostics Check		OK	23:51	

---

Data Download Date/Time: 11/30/2021 18:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KILLON,JONATHAN  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ALVAREZ-AYALA

DATE/TIME OF TEST: 10/17/2021 03:45:37

---

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.081	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.098	03:41
Breath Volume		2.601	03:41
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.098	03:44
Breath Volume		3.707	03:44
Air Blank		0.000	03:44
Control Test		0.080	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

---

Data Download Date/Time: 11/03/2021 11:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RINCON,RENZO  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: BRELAND

DATE/TIME OF TEST: 10/17/2021 04:27:51

---

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.081	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.165	04:23
Breath Volume	1.253	04:23
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.152	04:26
Breath Volume	1.324	04:26
Air Blank	0.000	04:26
Control Test	0.080	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

---

Data Download Date/Time: 11/03/2021 11:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HENRIQUEZ,ENRIQUE  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 10/24/2021 05:41:38

---

Diagnostics Check	OK	05:35
Air Blank	0.000	05:35
Control Test	0.081	05:35
Air Blank	0.000	05:36
Subject Sample # 1	0.315	05:37
Breath Volume	3.546	05:37
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample # 2	0.326	05:40
Breath Volume	3.343	05:40
Air Blank	0.000	05:40
Control Test	0.080	05:41
Air Blank	0.000	05:41
Diagnostics Check	OK	05:41

---

Data Download Date/Time: 11/03/2021 11:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HARWOOD,ARMANDO  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: GONGORA SADA

DATE/TIME OF TEST: 10/30/2021 03:03:01

---

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.081	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.163	02:56
Breath Volume	1.464	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.169	03:01
Breath Volume	1.246	03:01
Air Blank	0.000	03:02
Control Test	0.081	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:03

---

Data Download Date/Time: 11/03/2021 11:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARDEN,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: VALLE,OMAR  
ARRESTING OFFICER AGENCY: Ocoee PD  
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: OROPEZA-MORENO

DATE/TIME OF TEST: 10/31/2021 03:26:11

---

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.081	03:18
Air Blank	0.000	03:18
Subject Sample # 1	0.110	03:21
Breath Volume	1.597	03:21
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.115	03:24
Breath Volume	2.714	03:24
Air Blank	0.000	03:25
Control Test	0.081	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

---

Data Download Date/Time: 11/03/2021 11:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKES,JOHN,D  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: WILKES,JOHN  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BRADFORD

DATE/TIME OF TEST: 10/02/2021 03:01:54

---

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.078	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.149	02:57
Breath Volume	3.058	02:57
Air Blank	0.000	02:58
Air Blank	0.000	02:59
Subject Sample # 2	0.148	03:00
Breath Volume	2.937	03:00
Air Blank	0.000	03:01
Control Test	0.077	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

---

Data Download Date/Time: 10/07/2021 06:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,JOSE,G  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: DELGADO,JOSE  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: BAMACA DOMINGO

DATE/TIME OF TEST: 10/16/2021 00:40:18

---

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.078	00:32
Air Blank		0.000	00:32
Subject Sample	# 1	0.327	00:33
Breath Volume		1.839	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.321	00:38
Breath Volume		1.648	00:38
Air Blank		0.000	00:39
Control Test		0.078	00:39
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

---

Data Download Date/Time: 11/03/2021 19:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKES,JOHN,D  
BREATH TEST OPERATOR AGENCY: Martin County SO  
ARRESTING OFFICER: WILKES,JOHN  
ARRESTING OFFICER AGENCY: Martin County SO  
LAST AGENCY INSPECTION: 10/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A3  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MOZZILLO

DATE/TIME OF TEST: 10/31/2021 23:20:41

---

Diagnostics Check		OK	23:14
Air Blank		0.000	23:14
Control Test		0.079	23:15
Air Blank		0.000	23:15
Subject Sample	# 1	0.191	23:16
Breath Volume		2.218	23:16
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.187	23:19
Breath Volume		1.800	23:19
Air Blank		0.000	23:19
Control Test		0.079	23:20
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

---

Data Download Date/Time: 11/03/2021 19:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARSH,MATTHEW,E  
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD  
ARRESTING OFFICER: ROMAN,RYAN  
ARRESTING OFFICER AGENCY: Tyndall AFB PD  
LAST AGENCY INSPECTION: 10/18/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 20520080A1  
DRY GAS EXPIRY DATE: 09/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 10/30/2021 02:17:37

---

Diagnostics Check		OK	02:10
Air Blank		0.000	02:11
Control Test		0.081	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.036	02:13
Breath Volume		3.781	02:13
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.032	02:16
Breath Volume		3.601	02:16
Air Blank		0.000	02:16
Control Test		0.079	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

---

Data Download Date/Time: 11/15/2021 13:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MYERS,JUSTIN  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: BONTRAGER

DATE/TIME OF TEST: 10/01/2021 01:00:09

---

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:55
Subject Sample # 1	0.041	00:55
Breath Volume	1.886	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.040	00:58
Breath Volume	2.007	00:58
Air Blank	0.000	00:59
Control Test	0.079	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

---

Data Download Date/Time: 10/07/2021 14:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'GRADY,PATRICK,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: GOULD,PATRICIA  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 16:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:52

SUBJECT'S LAST NAME: RUTLEDGE

DATE/TIME OF TEST: 10/04/2021 17:32:19

---

Diagnostics Check	OK	17:28	
Air Blank	0.000	17:29	
Control Test	0.080	17:29	
Air Blank	0.000	17:29	
Subject Sample	# 1	17:30	Subject Test Refused
Breath Volume	0.316	17:30	
Air Blank	0.000	17:31	
Control Test	0.081	17:31	
Air Blank	0.000	17:32	
Diagnostics Check	OK	17:32	

---

Data Download Date/Time: 10/07/2021 14:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANDLEY,ASHLEY,N  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: RODRIGUEZ,KELVIS  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 08:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: DISARNO

DATE/TIME OF TEST: 10/06/2021 09:57:08

---

Diagnostics Check	OK	09:51	
Air Blank	0.000	09:52	
Control Test	0.080	09:52	
Air Blank	0.000	09:53	
Subject Sample # 1	0.000	09:53	
Breath Volume	2.199	09:53	
Air Blank	0.000	09:54	
Air Blank		09:55	Improper Sample
Air Blank	0.000	09:56	
Control Test	0.080	09:56	
Air Blank	0.000	09:57	
Diagnostics Check	OK	09:57	

---

Data Download Date/Time: 10/07/2021 14:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANDLEY,ASHLEY,N  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: RODRIGUEZ,KELVIS  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 08:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: DISARNO

DATE/TIME OF TEST: 10/06/2021 10:05:41

---

Diagnostics Check	OK	09:59
Air Blank	0.000	10:00
Control Test	0.081	10:00
Air Blank	0.000	10:00
Subject Sample # 1	0.000	10:01
Breath Volume	1.914	10:01
Air Blank	0.000	10:01
Air Blank	0.000	10:03
Subject Sample # 2	0.000	10:04
Breath Volume	1.558	10:04
Air Blank	0.000	10:04
Control Test	0.080	10:05
Air Blank	0.000	10:05
Diagnostics Check	OK	10:05

---

Data Download Date/Time: 10/07/2021 14:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANDLEY,ASHLEY,N  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: JACKSON,MICHAEL  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 11:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 10/06/2021 12:42:18

---

Diagnostics Check	OK	12:35
Air Blank	0.000	12:36
Control Test	0.080	12:36
Air Blank	0.000	12:37
Subject Sample # 1	0.000	12:38
Breath Volume	2.250	12:38
Air Blank	0.000	12:38
Air Blank	0.000	12:40
Subject Sample # 2	0.000	12:40
Breath Volume	2.093	12:40
Air Blank	0.000	12:41
Control Test	0.080	12:41
Air Blank	0.000	12:42
Diagnostics Check	OK	12:42

---

Data Download Date/Time: 10/07/2021 14:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,KENNETH,L  
BREATH TEST OPERATOR AGENCY: Putnam County SO  
ARRESTING OFFICER: VICCARI,AMY  
ARRESTING OFFICER AGENCY: Putnam County SO  
LAST AGENCY INSPECTION: 10/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 10/03/2021 01:36:47

---

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:30	
Control Test	0.075	01:30	
Air Blank	0.000	01:31	
Subject Sample # 1	0.317	01:32	
Breath Volume	1.589	01:32	
Air Blank	0.000	01:32	
Air Blank	0.000	01:34	
Subject Sample # 2	0.302	01:35	
Breath Volume	2.605	01:35	
Air Blank	0.000	01:35	
Control Test	0.074	01:36	Control Outside Tolerance
Air Blank	0.000	01:36	

---

Data Download Date/Time: 11/09/2021 15:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRYANT,B,J  
BREATH TEST OPERATOR AGENCY: Putnam County SO  
ARRESTING OFFICER: NUNEZ,M  
ARRESTING OFFICER AGENCY: Putnam County SO  
LAST AGENCY INSPECTION: 10/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: WEATHERWAX

DATE/TIME OF TEST: 10/09/2021 01:42:50

---

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:36	
Control Test	0.075	01:36	
Air Blank	0.000	01:37	
Subject Sample	# 1 0.274	01:37	
Breath Volume	1.175	01:37	
Air Blank	0.000	01:38	
Air Blank	0.000	01:40	
Subject Sample	# 2 0.290	01:41	
Breath Volume	1.691	01:41	
Air Blank	0.000	01:41	
Control Test	0.074	01:42	Control Outside Tolerance
Air Blank	0.000	01:42	

---

Data Download Date/Time: 11/09/2021 15:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRYANT, BRANDEN, J  
BREATH TEST OPERATOR AGENCY: Putnam County SO  
ARRESTING OFFICER: PLEIMAN, E  
ARRESTING OFFICER AGENCY: Palatka PD  
LAST AGENCY INSPECTION: 10/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 12:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ALT

DATE/TIME OF TEST: 10/10/2021 01:27:41

---

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.076	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.189	01:22
Breath Volume	1.625	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.186	01:26
Breath Volume	1.359	01:26
Air Blank	0.000	01:26
Control Test	0.075	01:27
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

---

Data Download Date/Time: 11/09/2021 15:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRYANT, BRANDEN, J  
BREATH TEST OPERATOR AGENCY: Putnam County SO  
ARRESTING OFFICER: PLIEMAN, ERIC  
ARRESTING OFFICER AGENCY: Palatka PD  
LAST AGENCY INSPECTION: 10/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: CREWS

DATE/TIME OF TEST: 10/13/2021 23:41:50

---

Diagnostics Check	OK	23:34	
Air Blank	0.000	23:34	
Control Test	0.075	23:35	
Air Blank	0.000	23:35	
Subject Sample	# 1 0.203	23:36	
Breath Volume	1.539	23:36	
Air Blank	0.000	23:37	
Air Blank	0.000	23:39	
Subject Sample	# 2 0.208	23:40	
Breath Volume	1.132	23:40	
Air Blank	0.000	23:40	
Control Test	0.074	23:41	Control Outside Tolerance
Air Blank	0.000	23:41	

---

Data Download Date/Time: 11/09/2021 15:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,KENNETH,L  
BREATH TEST OPERATOR AGENCY: Putnam County SO  
ARRESTING OFFICER: GRYKA,SABASTIAN  
ARRESTING OFFICER AGENCY: Putnam County SO  
LAST AGENCY INSPECTION: 10/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 10/22/2021 00:29:50

---

Diagnostics Check		OK	00:23
Air Blank		0.000	00:23
Control Test		0.076	00:23
Air Blank		0.000	00:24
Subject Sample	# 1	0.085	00:25
Breath Volume		1.664	00:25
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.076	00:28
Breath Volume		1.164	00:28
Air Blank		0.000	00:28
Control Test		0.076	00:29
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

---

Data Download Date/Time: 11/09/2021 15:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAYLOR,STEPHEN,M  
BREATH TEST OPERATOR AGENCY: Putnam County SO  
ARRESTING OFFICER: PLEIMAN,ERIC  
ARRESTING OFFICER AGENCY: Palatka PD  
LAST AGENCY INSPECTION: 10/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A2  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/23/2021 05:07:27

---

Diagnostics Check		OK	05:01
Air Blank		0.000	05:01
Control Test		0.075	05:02
Air Blank		0.000	05:02
Subject Sample	# 1	0.081	05:03
Breath Volume		1.359	05:03
Air Blank		0.000	05:03
Air Blank		0.000	05:05
Subject Sample	# 2	0.083	05:05
Breath Volume		1.921	05:05
Air Blank		0.000	05:06
Control Test		0.076	05:06
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

---

Data Download Date/Time: 11/09/2021 15:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMA, MICHAEL, A  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: LIMA, 2S02  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 10/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: SEMMES

DATE/TIME OF TEST: 10/28/2021 00:04:19

---

Diagnostics Check	OK	23:57
Air Blank	0.000	23:58
Control Test	0.076	23:58
Air Blank	0.000	23:59
Subject Sample # 2	0.206	23:59
Breath Volume	1.683	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 1	0.211	00:02
Breath Volume	2.136	00:02
Air Blank	0.000	00:03
Control Test	0.076	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

---

Data Download Date/Time: 11/09/2021 15:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,DUSTIN,A  
BREATH TEST OPERATOR AGENCY: Putnam County SO  
ARRESTING OFFICER: WADE,COLT  
ARRESTING OFFICER AGENCY: Putnam County SO  
LAST AGENCY INSPECTION: 10/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29017080A3  
DRY GAS EXPIRY DATE: 12/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: BOLANOS

DATE/TIME OF TEST: 10/31/2021 05:03:02

---

Diagnostics Check		OK	04:55
Air Blank		0.000	04:55
Control Test		0.077	04:56
Air Blank		0.000	04:56
Subject Sample	# 1	0.213	04:58
Breath Volume		3.402	04:58
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.221	05:01
Breath Volume		3.648	05:01
Air Blank		0.000	05:02
Control Test		0.075	05:02
Air Blank		0.000	05:02
Diagnostics Check		OK	05:03

---

Data Download Date/Time: 11/09/2021 15:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBISON,CASEY,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: PEREZ,SAMANTHA  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 10/01/2021 00:31:30

---

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.081	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.017	00:27
Breath Volume	2.855	00:27
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.016	00:30
Breath Volume	2.304	00:30
Air Blank	0.000	00:30
Control Test	0.081	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

---

Data Download Date/Time: 10/28/2021 09:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBISON,CASEY,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: MCCULLOUGH,SEAN  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/01/2021 01:38:49

---

Diagnostics Check		OK	01:24	
Air Blank		0.000	01:25	
Control Test		0.081	01:25	
Air Blank		0.000	01:25	
Subject Sample	# 1	0.206	01:28	Volume Not Met
Breath Volume		1.046	01:28	
Air Blank		0.000	01:29	
Air Blank		0.000	01:31	
Subject Sample	# 2	0.212	01:33	
Breath Volume		2.375	01:33	
Air Blank		0.000	01:34	
Air Blank		0.000	01:36	
Subject Sample	# 3	0.214	01:37	
Breath Volume		2.003	01:37	
Air Blank		0.000	01:37	
Control Test		0.079	01:38	
Air Blank		0.000	01:38	
Diagnostics Check		OK	01:38	

---

Data Download Date/Time: 10/28/2021 09:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBISON,CASEY,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: PEREZ,SAMANTHA  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: GUIZZOTTI

DATE/TIME OF TEST: 10/01/2021 19:12:58

---

Diagnostics Check	OK	19:04	
Air Blank	0.000	19:05	
Control Test	0.081	19:05	
Air Blank	0.000	19:05	
Subject Sample	# 1	19:08	No Sample Provided
Breath Volume	0.000	19:08	
Air Blank	0.000	19:09	
Air Blank	0.000	19:11	
Subject Sample	# 2	19:11	Subject Test Refused
Breath Volume	0.000	19:11	
Air Blank	0.000	19:12	
Control Test	0.081	19:12	
Air Blank	0.000	19:12	
Diagnostics Check	OK	19:12	

---

Data Download Date/Time: 10/28/2021 09:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHRISTENSEN,STEPHANIE,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: HINES,PAUL  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: BATTLE

DATE/TIME OF TEST: 10/02/2021 04:31:27

---

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.081	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.079	04:27
Breath Volume	1.441	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.078	04:30
Breath Volume	1.371	04:30
Air Blank	0.000	04:30
Control Test	0.081	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

---

Data Download Date/Time: 10/28/2021 09:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACKASAL, TYVON, T  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: LENAHAN, RYAN  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 10/02/2021 19:50:01

---

Diagnostics Check		OK	19:42
Air Blank		0.000	19:42
Control Test		0.081	19:42
Air Blank		0.000	19:43
Subject Sample	# 1	0.243	19:43
Breath Volume		1.226	19:43
Air Blank		0.000	19:44
Air Blank		0.000	19:46
Subject Sample	# 2	0.230	19:48
Breath Volume		2.183	19:48
Air Blank		0.000	19:49
Control Test		0.081	19:49
Air Blank		0.000	19:49
Diagnostics Check		OK	19:50

---

Data Download Date/Time: 10/28/2021 09:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARK,BRENT,D  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: CLARK,BRENT  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 11:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:37

SUBJECT'S LAST NAME: MORROW

DATE/TIME OF TEST: 10/05/2021 12:06:36

---

Diagnostics Check	OK	11:58
Air Blank	0.000	11:58
Control Test	0.082	11:58
Air Blank	0.000	11:59
Subject Sample # 1	0.000	12:01
Breath Volume	1.828	12:01
Air Blank	0.000	12:01
Air Blank	0.000	12:03
Subject Sample # 2	0.000	12:05
Breath Volume	2.085	12:05
Air Blank	0.000	12:05
Control Test	0.082	12:06
Air Blank	0.000	12:06
Diagnostics Check	OK	12:06

---

Data Download Date/Time: 10/28/2021 09:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBISON,CASEY,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ROBISON,CASEY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:56

SUBJECT'S LAST NAME: BARTASIS

DATE/TIME OF TEST: 10/06/2021 17:36:08

---

Diagnostics Check		OK	17:23	
Air Blank		0.000	17:24	
Control Test		0.081	17:24	
Air Blank		0.000	17:24	
Subject Sample	# 1	0.148	17:27	Volume Not Met
Breath Volume		0.125	17:27	
Air Blank		0.000	17:28	
Air Blank		0.000	17:30	
Subject Sample	# 2	0.129	17:31	
Breath Volume		2.031	17:31	
Air Blank		0.000	17:32	
Air Blank		0.000	17:34	
Subject Sample	# 3	0.140	17:34	
Breath Volume		1.699	17:34	
Air Blank		0.000	17:35	
Control Test		0.080	17:35	
Air Blank		0.000	17:36	
Diagnostics Check		OK	17:36	

---

Data Download Date/Time: 10/28/2021 09:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBISON,CASEY,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: PEREZ,SAMANTHA  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: STILLMAN

DATE/TIME OF TEST: 10/07/2021 03:08:27

---

Diagnostics Check		OK	03:01
Air Blank		0.000	03:01
Control Test		0.081	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.142	03:03
Breath Volume		1.199	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:05
Subject Sample	# 2	0.136	03:06
Breath Volume		1.277	03:06
Air Blank		0.000	03:07
Control Test		0.081	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

---

Data Download Date/Time: 10/28/2021 09:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GERRY,TABATHA,B  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: PEREZ,SAMANTHA  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: GUERRA PINERO

DATE/TIME OF TEST: 10/07/2021 23:33:43

---

Diagnostics Check	OK	23:27
Air Blank	0.000	23:28
Control Test	0.081	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.140	23:29
Breath Volume	1.687	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.140	23:32
Breath Volume	1.562	23:32
Air Blank	0.000	23:32
Control Test	0.081	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

---

Data Download Date/Time: 10/28/2021 09:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ROBISON,CASEY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: MCLAY

DATE/TIME OF TEST: 10/08/2021 02:03:32

---

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.081	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.194	01:58
Breath Volume	4.437	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.186	02:02
Breath Volume	3.921	02:02
Air Blank	0.000	02:02
Control Test	0.081	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

---

Data Download Date/Time: 10/28/2021 09:25

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ROBISON,CASEY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: DELOACH

DATE/TIME OF TEST: 10/09/2021 02:49:10

---

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.081	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.289	02:44
Breath Volume	2.371	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.283	02:47
Breath Volume	1.843	02:47
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

---

Data Download Date/Time: 10/28/2021 09:25

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ROBISON,CASEY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: CORTES

DATE/TIME OF TEST: 10/09/2021 04:18:47

---

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.082	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.050	04:14
Breath Volume		1.136	04:14
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.048	04:17
Breath Volume		3.050	04:17
Air Blank		0.000	04:17
Control Test		0.082	04:18
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

---

Data Download Date/Time: 10/28/2021 09:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBISON,CASEY,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: PEREZ,SAMANTHA  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 10/09/2021 21:53:26

---

Diagnostics Check	OK	21:46
Air Blank	0.000	21:47
Control Test	0.081	21:47
Air Blank	0.000	21:48
Subject Sample # 1	0.205	21:49
Breath Volume	3.324	21:49
Air Blank	0.000	21:49
Air Blank	0.000	21:51
Subject Sample # 2	0.195	21:51
Breath Volume	4.292	21:51
Air Blank	0.000	21:52
Control Test	0.081	21:52
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

---

Data Download Date/Time: 10/28/2021 09:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ROBISON,CASEY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: HASSAN

DATE/TIME OF TEST: 10/13/2021 01:59:46

---

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:48	
Control Test	0.082	01:49	
Air Blank	0.000	01:49	
Subject Sample	# 1	01:52	No Sample Provided
Breath Volume	0.000	01:52	
Air Blank	0.000	01:53	
Air Blank	0.000	01:55	
Subject Sample	# 2	01:58	No Sample Provided
Breath Volume	0.000	01:58	
Air Blank	0.000	01:58	
Control Test	0.082	01:59	
Air Blank	0.000	01:59	
Diagnostics Check	OK	01:59	

---

Data Download Date/Time: 10/28/2021 09:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: PEREZ,SAMANTHA  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 10/13/2021 21:26:50

---

Diagnostics Check	OK	21:21
Air Blank	0.000	21:21
Control Test	0.082	21:21
Air Blank	0.000	21:22
Subject Sample # 1	0.000	21:22
Breath Volume	2.656	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.000	21:25
Breath Volume	2.113	21:25
Air Blank	0.000	21:25
Control Test	0.082	21:26
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

---

Data Download Date/Time: 10/28/2021 09:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ROBISON,CASEY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: PURCELL

DATE/TIME OF TEST: 10/14/2021 01:45:08

---

Diagnostics Check		OK	01:38
Air Blank		0.000	01:39
Control Test		0.082	01:39
Air Blank		0.000	01:39
Subject Sample	# 1	0.076	01:40
Breath Volume		1.355	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.078	01:43
Breath Volume		1.484	01:43
Air Blank		0.000	01:44
Control Test		0.081	01:44
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

---

Data Download Date/Time: 10/28/2021 09:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBISON,CASEY,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: PEREZ,SAMANTHA  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: CLAY

DATE/TIME OF TEST: 10/14/2021 23:14:20

---

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.081	23:09
Air Blank	0.000	23:09
Subject Sample # 1	0.000	23:10
Breath Volume	2.003	23:10
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.000	23:12
Breath Volume	2.117	23:12
Air Blank	0.000	23:13
Control Test	0.082	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

---

Data Download Date/Time: 10/28/2021 09:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ROBISON,CASEY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: BROXTON

DATE/TIME OF TEST: 10/15/2021 01:45:01

---

Diagnostics Check		OK	01:38
Air Blank		0.000	01:39
Control Test		0.081	01:39
Air Blank		0.000	01:40
Subject Sample	# 1	0.010	01:40
Breath Volume		2.000	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.010	01:43
Breath Volume		2.460	01:43
Air Blank		0.000	01:44
Control Test		0.082	01:44
Air Blank		0.000	01:44
Diagnostics Check		OK	01:45

---

Data Download Date/Time: 10/28/2021 09:25

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: BISHOP,CHRISTOPHER  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: CHEEK

DATE/TIME OF TEST: 10/15/2021 22:44:03

---

Diagnostics Check	OK	22:27	
Air Blank	0.000	22:28	
Control Test	0.082	22:28	
Air Blank	0.000	22:29	
Subject Sample	# 1	22:32	No Sample Provided
Breath Volume	0.000	22:32	
Air Blank	0.000	22:32	
Air Blank	0.000	22:34	
Subject Sample	# 2	22:37	
Breath Volume	3.085	22:37	
Air Blank	0.000	22:37	
Air Blank	0.000	22:39	
Subject Sample	# 3	22:42	No Sample Provided
Breath Volume	0.000	22:42	
Air Blank	0.000	22:43	
Control Test	0.082	22:43	
Air Blank	0.000	22:43	
Diagnostics Check	OK	22:44	

---

Data Download Date/Time: 10/28/2021 09:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GERRY,TABATHA,B  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: LENAHAN,RYAN  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: RENTZ

DATE/TIME OF TEST: 10/16/2021 01:52:53

---

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.081	01:49	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:51	
Control Test	0.082	01:52	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

---

Data Download Date/Time: 10/28/2021 09:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GERRY,TABATHA,B  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: BISHOP,CHRISTOPHER  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: KELEMAN

DATE/TIME OF TEST: 10/16/2021 02:23:23

---

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:21	
Control Test	0.082	02:21	
Air Blank	0.000	02:21	
Subject Sample	# 1	02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:22	
Control Test	0.081	02:22	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

---

Data Download Date/Time: 10/28/2021 09:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SEAY, III, KENNETH, M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ROBISON, CASEY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: MAGGIO

DATE/TIME OF TEST: 10/16/2021 03:31:11

---

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:19	
Control Test		0.082	03:19	
Air Blank		0.000	03:20	
Subject Sample	# 1	0.135	03:23	Volume Not Met
Breath Volume		0.296	03:23	
Air Blank		0.000	03:23	
Air Blank		0.000	03:25	
Subject Sample	# 2	0.123	03:26	
Breath Volume		1.296	03:26	
Air Blank		0.000	03:26	
Air Blank		0.000	03:28	
Subject Sample	# 3	0.124	03:29	
Breath Volume		1.246	03:29	
Air Blank		0.000	03:30	
Control Test		0.080	03:30	
Air Blank		0.000	03:31	
Diagnostics Check		OK	03:31	

---

Data Download Date/Time: 10/28/2021 09:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBISON,CASEY,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: PEREZ,SAMANTHA  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: BIRCH

DATE/TIME OF TEST: 10/16/2021 04:48:22

---

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.081	04:42
Air Blank	0.000	04:42
Subject Sample # 1	0.136	04:43
Breath Volume	1.667	04:43
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.134	04:46
Breath Volume	1.812	04:46
Air Blank	0.000	04:47
Control Test	0.081	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

---

Data Download Date/Time: 10/28/2021 09:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: BUTTERFIELD,KELLY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: BARRAGAN

DATE/TIME OF TEST: 10/17/2021 00:25:55

---

Diagnostics Check	OK	00:18
Air Blank	0.000	00:18
Control Test	0.081	00:19
Air Blank	0.000	00:19
Subject Sample # 1	0.242	00:21
Breath Volume	3.203	00:21
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.254	00:24
Breath Volume	4.457	00:24
Air Blank	0.000	00:25
Control Test	0.079	00:25
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

---

Data Download Date/Time: 10/28/2021 09:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,EMILY,K  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: LENAHAN,RYAN  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: OCONNOR

DATE/TIME OF TEST: 10/17/2021 01:20:45

---

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:18	
Control Test	0.082	01:18	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:19	
Control Test	0.081	01:20	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

---

Data Download Date/Time: 10/28/2021 09:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,EMILY,K  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: LENAHAN,RYAN  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: SCHANEMAN

DATE/TIME OF TEST: 10/18/2021 22:13:40

---

Diagnostics Check	OK	22:07
Air Blank	0.000	22:07
Control Test	0.081	22:08
Air Blank	0.000	22:08
Subject Sample # 1	0.160	22:09
Breath Volume	2.730	22:09
Air Blank	0.000	22:09
Air Blank	0.000	22:11
Subject Sample # 2	0.157	22:12
Breath Volume	2.468	22:12
Air Blank	0.000	22:12
Control Test	0.081	22:13
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

---

Data Download Date/Time: 10/28/2021 09:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BISHOP,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: THOMAS,KRISTEN  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: GIFFORD

DATE/TIME OF TEST: 10/20/2021 22:06:50

---

Diagnostics Check	OK	22:00
Air Blank	0.000	22:00
Control Test	0.082	22:01
Air Blank	0.000	22:01
Subject Sample # 1	0.134	22:02
Breath Volume	1.546	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:04
Subject Sample # 2	0.133	22:05
Breath Volume	1.257	22:05
Air Blank	0.000	22:05
Control Test	0.081	22:06
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

---

Data Download Date/Time: 10/28/2021 09:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BISHOP,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ROBISON,CASEY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: WEATHERALL

DATE/TIME OF TEST: 10/21/2021 02:29:30

---

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.081	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.228	02:24
Breath Volume		1.339	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.232	02:28
Breath Volume		2.000	02:28
Air Blank		0.000	02:28
Control Test		0.081	02:28
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

---

Data Download Date/Time: 10/28/2021 09:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BISHOP,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: PEREZ,SAMANTHA  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: DAVIDSMEIER

DATE/TIME OF TEST: 10/21/2021 02:44:47

---

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:33	
Control Test	0.081	02:34	
Air Blank	0.000	02:34	
Subject Sample	# 1	02:37	No Sample Provided
Breath Volume	0.000	02:37	
Air Blank	0.000	02:38	
Air Blank	0.000	02:40	
Subject Sample	# 2	02:43	No Sample Provided
Breath Volume	0.000	02:43	
Air Blank	0.000	02:43	
Control Test	0.081	02:44	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:44	

---

Data Download Date/Time: 10/28/2021 09:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCKENZIE, RONALD,  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: MCKENZIE, RONALD  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 11:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: LANG

DATE/TIME OF TEST: 10/21/2021 12:14:15

---

Diagnostics Check	OK	12:11	
Air Blank	0.000	12:11	
Control Test	0.082	12:12	
Air Blank	0.000	12:12	
Subject Sample	# 1	12:12	Subject Test Refused
Breath Volume	0.000	12:12	
Air Blank	0.000	12:13	
Control Test	0.082	12:13	
Air Blank	0.000	12:14	
Diagnostics Check	OK	12:14	

---

Data Download Date/Time: 10/28/2021 09:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ROBISON,CASEY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 10/23/2021 03:17:05

---

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.082	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.121	03:11
Breath Volume	1.410	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.119	03:15
Breath Volume	1.218	03:15
Air Blank	0.000	03:16
Control Test	0.081	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

---

Data Download Date/Time: 10/28/2021 09:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESTOCH,JEFFREY,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: WRIGHT,NAOMI  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 12:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 10/23/2021 13:21:12

---

Diagnostics Check	OK	13:20	
Air Blank		13:20	RFI Detect
Air Blank	0.000	13:21	

---

Data Download Date/Time: 10/28/2021 09:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESTOCH,JEFFREY,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: WRIGHT,NAOMI  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 12:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 10/23/2021 13:31:04

---

Diagnostics Check	OK	13:22
Air Blank	0.000	13:23
Control Test	0.082	13:23
Air Blank	0.000	13:23
Subject Sample # 1	0.000	13:26
Breath Volume	1.910	13:26
Air Blank	0.000	13:26
Air Blank	0.000	13:28
Subject Sample # 2	0.000	13:29
Breath Volume	1.132	13:29
Air Blank	0.000	13:30
Control Test	0.081	13:30
Air Blank	0.000	13:30
Diagnostics Check	OK	13:31

---

Data Download Date/Time: 10/28/2021 09:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ROBISON,CASEY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: KOCH

DATE/TIME OF TEST: 10/24/2021 03:50:33

---

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.082	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.144	03:46
Breath Volume	2.570	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.137	03:49
Breath Volume	2.292	03:49
Air Blank	0.000	03:49
Control Test	0.081	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

---

Data Download Date/Time: 10/28/2021 09:26

**Florida Department of Law Enforcement**

**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOBART,CLAYTON,T  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: LENAHAN,RYAN  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: WHITCOMB

DATE/TIME OF TEST: 10/24/2021 20:06:27

---

Diagnostics Check		OK	19:59
Air Blank		0.000	19:59
Control Test		0.082	20:00
Air Blank		0.000	20:00
Subject Sample	# 1	0.199	20:01
Breath Volume		1.839	20:01
Air Blank		0.000	20:02
Air Blank		0.000	20:04
Subject Sample	# 2	0.201	20:04
Breath Volume		1.550	20:04
Air Blank		0.000	20:05
Control Test		0.081	20:05
Air Blank		0.000	20:06
Diagnostics Check		OK	20:06

---

Data Download Date/Time: 10/28/2021 09:26

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: KARAYIANES,MICHAEL  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: HUTSON

DATE/TIME OF TEST: 10/26/2021 01:58:10

---

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.082	01:52
Air Blank	0.000	01:53
Subject Sample # 1	0.000	01:53
Breath Volume	2.542	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.000	01:56
Breath Volume	2.589	01:56
Air Blank	0.000	01:57
Control Test	0.081	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

---

Data Download Date/Time: 10/28/2021 09:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BISHOP,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: THOMAS,KRISTEN  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: RIBEIRO

DATE/TIME OF TEST: 10/27/2021 21:28:24

---

Diagnostics Check	OK	21:20
Air Blank	0.000	21:21
Control Test	0.081	21:21
Air Blank	0.000	21:22
Subject Sample # 1	0.000	21:24
Breath Volume	1.457	21:24
Air Blank	0.000	21:24
Air Blank	0.000	21:26
Subject Sample # 2	0.000	21:27
Breath Volume	1.253	21:27
Air Blank	0.000	21:27
Control Test	0.081	21:27
Air Blank	0.000	21:28
Diagnostics Check	OK	21:28

---

Data Download Date/Time: 10/28/2021 09:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BISHOP,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ROBISON,CASEY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: KILLENS

DATE/TIME OF TEST: 10/27/2021 22:09:25

---

Diagnostics Check	OK	22:03
Air Blank	0.000	22:03
Control Test	0.082	22:03
Air Blank	0.000	22:04
Subject Sample # 1	0.000	22:05
Breath Volume	1.574	22:05
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.000	22:08
Breath Volume	1.457	22:08
Air Blank	0.000	22:08
Control Test	0.082	22:08
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

---

Data Download Date/Time: 10/28/2021 09:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: ROBISON,CASEY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 10/28/2021 01:14:47

---

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:00	
Control Test	0.081	01:00	
Air Blank	0.000	01:01	
Subject Sample # 1	0.155	01:04	
Breath Volume	1.289	01:04	
Air Blank	0.000	01:05	
Air Blank	0.000	01:06	
Subject Sample # 2	0.149	01:10	Volume Not Met
Breath Volume	0.742	01:10	
Air Blank	0.000	01:10	
Air Blank	0.000	01:12	
Subject Sample # 3	0.156	01:13	
Breath Volume	1.390	01:13	
Air Blank	0.000	01:13	
Control Test	0.080	01:14	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

---

Data Download Date/Time: 10/28/2021 09:26

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBISON,CASEY,M  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: BISHOP,CHRISTOPHER  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: AVELARES

DATE/TIME OF TEST: 10/28/2021 02:33:21

---

Diagnostics Check		OK	02:25
Air Blank		0.000	02:25
Control Test		0.081	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.102	02:27
Breath Volume		1.691	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.095	02:31
Breath Volume		1.531	02:31
Air Blank		0.000	02:32
Control Test		0.081	02:32
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

---

Data Download Date/Time: 10/28/2021 09:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUAREZ,JOEL,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: LASTRA,SUNEY  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 08:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:35

SUBJECT'S LAST NAME: BOSCH

DATE/TIME OF TEST: 10/01/2021 10:04:40

---

Diagnostics Check		OK	09:57
Air Blank		0.000	09:58
Control Test		0.082	09:58
Air Blank		0.000	09:58
Subject Sample	# 1	0.212	09:59
Breath Volume		1.804	09:59
Air Blank		0.000	10:00
Air Blank		0.000	10:02
Subject Sample	# 2	0.198	10:02
Breath Volume		2.097	10:02
Air Blank		0.000	10:03
Control Test		0.081	10:04
Air Blank		0.000	10:04
Diagnostics Check		OK	10:04

---

Data Download Date/Time: 11/30/2021 12:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA,J,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: ROMERO,B  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 10/02/2021 01:53:08

---

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:50	
Control Test	0.081	01:51	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:52	
Control Test	0.080	01:52	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

---

Data Download Date/Time: 11/30/2021 12:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GONZALEZ,CAMILA,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: MEDEROS,BRYAN

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1497672

DRY GAS EXPIRY DATE: 05/20/2024

VIOLATION CODE: DUI

ARREST TIME: 21:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: ESTRADA

DATE/TIME OF TEST: 10/03/2021 21:58:47

---

Diagnostics Check	OK	21:57	
Air Blank		21:58	RFI Detect
Air Blank	0.000	21:58	

---

Data Download Date/Time: 11/30/2021 12:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GONZALEZ,CAMILA,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: MEDEROS,BRYAN

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1497672

DRY GAS EXPIRY DATE: 05/20/2024

VIOLATION CODE: DUI

ARREST TIME: 21:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: ESTRADA

DATE/TIME OF TEST: 10/03/2021 22:00:47

---

Diagnostics Check	OK	22:00	
Air Blank		22:00	RFI Detect
Air Blank		22:00	RFI Detect

---

Data Download Date/Time: 11/30/2021 12:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,DAVID,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: GONZALEZ,CAMILA  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/04/2021 02:35:04

---

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:32	
Control Test	0.082	02:32	
Air Blank	0.000	02:33	
Subject Sample	# 1	02:33	Subject Test Refused
Breath Volume	0.000	02:33	
Air Blank	0.000	02:34	
Control Test	0.082	02:34	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:35	

---

Data Download Date/Time: 11/30/2021 12:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ECHEVARRIA,D,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: VASQUEZ,JASON  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 07:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:26

SUBJECT'S LAST NAME: GOMEZ FUNEZ

DATE/TIME OF TEST: 10/04/2021 09:56:29

---

Diagnostics Check	OK	09:50
Air Blank	0.000	09:50
Control Test	0.082	09:50
Air Blank	0.000	09:51
Subject Sample # 1	0.113	09:51
Breath Volume	1.167	09:51
Air Blank	0.000	09:52
Air Blank	0.000	09:54
Subject Sample # 2	0.106	09:54
Breath Volume	1.890	09:54
Air Blank	0.000	09:55
Control Test	0.079	09:55
Air Blank	0.000	09:56
Diagnostics Check	OK	09:56

---

Data Download Date/Time: 11/30/2021 12:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ACUNA JR,C,A  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: BYARS,W  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 10:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/06/2021 11:04:17

---

Diagnostics Check	OK	11:01	
Air Blank	0.000	11:01	
Control Test	0.081	11:02	
Air Blank	0.000	11:02	
Subject Sample	# 1	11:02	Subject Test Refused
Breath Volume	0.000	11:02	
Air Blank	0.000	11:03	
Control Test	0.081	11:03	
Air Blank	0.000	11:04	
Diagnostics Check	OK	11:04	

---

Data Download Date/Time: 11/30/2021 12:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEL COLLADO,ADRIAN,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: BYARS,WALTER  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 09:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:12

SUBJECT'S LAST NAME: SOUZA GOMES  
DATE/TIME OF TEST: 10/08/2021 09:56:24

---

Diagnostics Check	OK	09:53	
Air Blank	0.000	09:54	
Control Test	0.082	09:54	
Air Blank	0.000	09:55	
Subject Sample	# 1	09:55	RFI Detect
Breath Volume	0.000	09:55	
Air Blank	0.000	09:55	
		09:55	RFI Detect
Air Blank	0.000	09:56	

---

Data Download Date/Time: 11/30/2021 12:43



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORA,MANUEL,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: MORA,MANUEL  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: COURT ORD  
ARREST TIME: 15:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/08/2021 15:56:47

---

Diagnostics Check	OK	15:51
Air Blank	0.000	15:51
Control Test	0.082	15:51
Air Blank	0.000	15:52
Subject Sample # 1	0.000	15:52
Breath Volume	1.996	15:52
Air Blank	0.000	15:53
Air Blank	0.000	15:55
Subject Sample # 2	0.000	15:55
Breath Volume	1.800	15:55
Air Blank	0.000	15:55
Control Test	0.082	15:56
Air Blank	0.000	15:56
Diagnostics Check	OK	15:56

---

Data Download Date/Time: 11/30/2021 12:43

**Florida Department of Law Enforcement**

**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SINGLETARY,CHIQUITA,M  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: ARIAS,LUIS  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 07:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: GRIJALVA-CASTILLO

DATE/TIME OF TEST: 10/09/2021 08:59:09

---

Diagnostics Check	OK	08:53	
Air Blank	0.000	08:54	
Control Test	0.081	08:54	
Air Blank	0.000	08:55	
Subject Sample	# 1	08:57	Subject Test Refused
Breath Volume	0.000	08:57	
Air Blank	0.000	08:58	
Control Test	0.082	08:58	
Air Blank	0.000	08:59	
Diagnostics Check	OK	08:59	

---

Data Download Date/Time: 11/30/2021 12:43



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,JOHN,J  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: IVY,D  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 12:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:07

SUBJECT'S LAST NAME: GARCIA-TINCOPIA

DATE/TIME OF TEST: 10/10/2021 13:34:52

---

Diagnostics Check	OK	13:28
Air Blank	0.000	13:28
Control Test	0.082	13:29
Air Blank	0.000	13:29
Subject Sample # 1	0.216	13:30
Breath Volume	1.718	13:30
Air Blank	0.000	13:31
Air Blank	0.000	13:32
Subject Sample # 2	0.211	13:33
Breath Volume	1.824	13:33
Air Blank	0.000	13:33
Control Test	0.081	13:34
Air Blank	0.000	13:34
Diagnostics Check	OK	13:34

---

Data Download Date/Time: 11/30/2021 12:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HICKS,TAVARIUS,C  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: HICKS,TAVARIUS  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 10/16/2021 03:11:57

---

Diagnostics Check	OK	03:11	
Air Blank		03:11	RFI Detect
Air Blank	0.000	03:11	

---

Data Download Date/Time: 11/30/2021 12:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HICKS,TAVARIUS,C  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: HICKS,TAVARIUS  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 10/16/2021 03:21:08

---

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.082	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.147	03:16
Breath Volume	1.902	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.141	03:19
Breath Volume	1.136	03:19
Air Blank	0.000	03:20
Control Test	0.079	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

---

Data Download Date/Time: 11/30/2021 12:44



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESTAPE,Z,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: ESTAPE,Z  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: ROSES

DATE/TIME OF TEST: 10/16/2021 22:47:24

---

Diagnostics Check	OK	22:44	
Air Blank	0.000	22:44	
Control Test	0.082	22:45	
Air Blank	0.000	22:45	
Subject Sample	# 1	22:46	Subject Test Refused
Breath Volume	0.000	22:46	
Air Blank	0.000	22:46	
Control Test	0.082	22:46	
Air Blank	0.000	22:47	
Diagnostics Check	OK	22:47	

---

Data Download Date/Time: 11/30/2021 12:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCLEOD,KYLE,P  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: MCLEOD,KYLE  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 07:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:08

SUBJECT'S LAST NAME: AYERS

DATE/TIME OF TEST: 10/17/2021 12:41:43

---

Diagnostics Check		OK	12:30	
Air Blank		0.000	12:30	
Control Test		0.082	12:30	
Air Blank		0.000	12:31	
Subject Sample	# 1	0.023	12:34	Volume Not Met
Breath Volume		0.996	12:34	
Air Blank		0.000	12:35	
Air Blank		0.000	12:36	
Subject Sample	# 2	0.028	12:40	Volume Not Met
Breath Volume		1.058	12:40	
Air Blank		0.000	12:40	
Control Test		0.081	12:41	
Air Blank		0.000	12:41	
Diagnostics Check		OK	12:41	

---

Data Download Date/Time: 11/30/2021 12:44

**Florida Department of Law Enforcement**

**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCLEOD,KYLE,P  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: MCLEOD,KYLE  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 07:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:47

SUBJECT'S LAST NAME: AYERS

DATE/TIME OF TEST: 10/17/2021 13:20:08

---

Diagnostics Check		OK	13:08	
Air Blank		0.000	13:09	
Control Test		0.081	13:09	
Air Blank		0.000	13:10	
Subject Sample	# 1	0.013	13:13	Volume Not Met
Breath Volume		0.324	13:13	
Air Blank		0.000	13:13	
Air Blank		0.000	13:15	
Subject Sample	# 2	0.011	13:18	Volume Not Met
Breath Volume		0.644	13:18	
Air Blank		0.000	13:19	
Control Test		0.081	13:19	
Air Blank		0.000	13:20	
Diagnostics Check		OK	13:20	

---

Data Download Date/Time: 11/30/2021 12:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,DAVID,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: GONZALEZ,CAMILA  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: BALLESTEROS

DATE/TIME OF TEST: 10/18/2021 01:10:46

---

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:08	
Control Test	0.082	01:08	
Air Blank	0.000	01:08	
Subject Sample	# 1	01:09	Subject Test Refused
Breath Volume	0.000	01:09	
Air Blank	0.000	01:09	
Control Test	0.081	01:10	
Air Blank	0.000	01:10	
Diagnostics Check	OK	01:10	

---

Data Download Date/Time: 11/30/2021 12:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITT,JESSE,T  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: HERNANDEZ,KEITH  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: BUI  
ARREST TIME: 23:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: LEON VELASCO

DATE/TIME OF TEST: 10/20/2021 00:12:15

---

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.081	00:04
Air Blank	0.000	00:05
Subject Sample # 1	0.088	00:07
Breath Volume	1.871	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.083	00:10
Breath Volume	1.566	00:10
Air Blank	0.000	00:11
Control Test	0.079	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

---

Data Download Date/Time: 11/30/2021 12:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LLANES,IOANNYS,  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: LLANES,IOANNYS  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: SOTO FRECH

DATE/TIME OF TEST: 10/20/2021 02:26:21

---

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:22	
Control Test	0.080	02:23	
Air Blank	0.000	02:23	
Subject Sample	# 1	02:24	Subject Test Refused
Breath Volume	0.000	02:24	
Air Blank	0.000	02:25	
Control Test	0.081	02:25	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

---

Data Download Date/Time: 11/30/2021 13:08



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GONZALEZ,CAMILA,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CARBONO,DOUGLAS

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1497672

DRY GAS EXPIRY DATE: 05/20/2024

VIOLATION CODE: DUI

ARREST TIME: 23:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: VIGIL

DATE/TIME OF TEST: 10/23/2021 04:22:48

---

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:20	
Control Test	0.082	04:20	
Air Blank	0.000	04:21	
Subject Sample	# 1	04:21	Subject Test Refused
Breath Volume	0.000	04:21	
Air Blank	0.000	04:21	
Control Test	0.080	04:22	
Air Blank	0.000	04:22	
Diagnostics Check	OK	04:22	

---

Data Download Date/Time: 11/30/2021 13:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROMERO,BRANDON,C  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: GUERRERO,RAFAEL  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PAEZ

DATE/TIME OF TEST: 10/24/2021 04:20:32

---

Diagnostics Check	OK	04:14	
Air Blank	0.000	04:15	
Control Test	0.082	04:15	
Air Blank	0.000	04:16	
Subject Sample	# 1	04:19	Subject Test Refused
Breath Volume	0.000	04:19	
Air Blank	0.000	04:19	
Control Test	0.082	04:19	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:20	

---

Data Download Date/Time: 11/30/2021 13:09





# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GUZMAN,GUILLERMO,J

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: GUZMAN,GUILLERMO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1497672

DRY GAS EXPIRY DATE: 05/20/2024

VIOLATION CODE: DUI

ARREST TIME: 05:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 07:02

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 10/24/2021 09:05:08

---

Diagnostics Check	OK	09:00	
Air Blank	0.000	09:01	
Control Test	0.079	09:01	
Air Blank	0.000	09:01	
Subject Sample	# 1	09:03	Subject Test Refused
Breath Volume	0.015	09:03	
Air Blank	0.000	09:04	
Control Test	0.080	09:04	
Air Blank	0.000	09:05	
Diagnostics Check	OK	09:05	

---

Data Download Date/Time: 11/30/2021 13:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARGITIS,NICHOLAS,J  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: MCGUIRE,AMBER  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: FOLEY

DATE/TIME OF TEST: 10/01/2021 21:11:12

---

Diagnostics Check		OK	21:04
Air Blank		0.000	21:05
Control Test		0.079	21:05
Air Blank		0.000	21:06
Subject Sample	# 1	0.203	21:06
Breath Volume		2.312	21:06
Air Blank		0.000	21:07
Air Blank		0.000	21:09
Subject Sample	# 2	0.194	21:09
Breath Volume		2.574	21:09
Air Blank		0.000	21:10
Control Test		0.079	21:10
Air Blank		0.000	21:11
Diagnostics Check		OK	21:11

---

Data Download Date/Time: 10/29/2021 11:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,J,G  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: HOWARD,J  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: MCGINLEY

DATE/TIME OF TEST: 10/06/2021 01:46:50

---

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.081	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.132	01:42
Breath Volume	1.667	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.137	01:45
Breath Volume	2.667	01:45
Air Blank	0.000	01:45
Control Test	0.080	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

---

Data Download Date/Time: 10/29/2021 11:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,J,G  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: HOWARD,J  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 10/07/2021 02:40:45

---

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.080	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.033	02:36
Breath Volume	2.003	02:36
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.031	02:39
Breath Volume	1.816	02:39
Air Blank	0.000	02:39
Control Test	0.079	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

---

Data Download Date/Time: 10/29/2021 11:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,MASON,E  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: THOMPSON,MASON  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 06:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: MARBRY

DATE/TIME OF TEST: 10/07/2021 06:57:57

---

Diagnostics Check	OK	06:51
Air Blank	0.000	06:52
Control Test	0.081	06:52
Air Blank	0.000	06:52
Subject Sample # 1	0.119	06:53
Breath Volume	1.300	06:53
Air Blank	0.000	06:54
Air Blank	0.000	06:55
Subject Sample # 2	0.120	06:56
Breath Volume	1.691	06:56
Air Blank	0.000	06:57
Control Test	0.079	06:57
Air Blank	0.000	06:57
Diagnostics Check	OK	06:57

---

Data Download Date/Time: 10/29/2021 11:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORREA,JONATHAN,E  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: CORREA,JONATHAN  
ARRESTING OFFICER AGENCY: St. Augustine PD  
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: WISSEL

DATE/TIME OF TEST: 10/08/2021 22:15:37

---

Diagnostics Check	OK	22:09
Air Blank	0.000	22:10
Control Test	0.082	22:10
Air Blank	0.000	22:10
Subject Sample # 1	0.147	22:11
Breath Volume	1.617	22:11
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.143	22:14
Breath Volume	1.804	22:14
Air Blank	0.000	22:14
Control Test	0.081	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

---

Data Download Date/Time: 10/29/2021 11:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,J,G  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: DANIELS,A  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 10/20/2021 02:31:51

---

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.186	02:24
Breath Volume	2.750	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:26
Subject Sample # 2	0.165	02:27
Breath Volume	1.863	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:29
Subject Sample # 3	0.190	02:30
Breath Volume	3.261	02:30
Air Blank	0.000	02:30
Control Test	0.079	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

---

Data Download Date/Time: 10/29/2021 11:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,J,G  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: HOWARD,J  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/11/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 10/30/2021 02:26:41

---

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.146	02:22
Breath Volume	2.648	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.147	02:25
Breath Volume	3.812	02:25
Air Blank	0.000	02:25
Control Test	0.079	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

---

Data Download Date/Time: 12/02/2021 20:48