

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CONTRERAS,ORLANDO,F  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: HICKEY,JOSHUA  
ARRESTING OFFICER AGENCY: South Miami PD  
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 11/12/2021 22:00:10

---

Diagnostics Check	OK	21:57	
Air Blank	0.000	21:57	
Control Test	0.080	21:58	
Air Blank	0.000	21:58	
Subject Sample	# 1	21:58	Subject Test Refused
Breath Volume	0.000	21:58	
Air Blank	0.000	21:59	
Control Test	0.080	21:59	
Air Blank	0.000	22:00	
Diagnostics Check	OK	22:00	

---

Data Download Date/Time: 11/29/2021 14:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMSON,ROBERT,H  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: PINO,A  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 11/14/2021 05:17:09

---

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:14	
Control Test	0.080	05:14	
Air Blank	0.000	05:15	
Subject Sample	# 1	05:15	Subject Test Refused
Breath Volume	0.000	05:15	
Air Blank	0.000	05:16	
Control Test	0.080	05:16	
Air Blank	0.000	05:17	
Diagnostics Check	OK	05:17	

---

Data Download Date/Time: 11/29/2021 14:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BECKER,ALEXA,C  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: ALZATE,MELENNI  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 15:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:04

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 11/15/2021 16:27:10

---

Diagnostics Check	OK	16:24	
Air Blank	0.000	16:24	
Control Test	0.079	16:25	
Air Blank	0.000	16:25	
Subject Sample	# 1	16:25	Subject Test Refused
Breath Volume	0.000	16:25	
Air Blank	0.000	16:26	
Control Test	0.080	16:26	
Air Blank	0.000	16:27	
Diagnostics Check	OK	16:27	

---

Data Download Date/Time: 11/29/2021 14:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,JUSTIN,J  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: JARAMILLO,DEVIN  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: LABANINO

DATE/TIME OF TEST: 11/17/2021 00:11:13

---

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.076	00:04
Air Blank	0.000	00:05
Subject Sample # 1	0.166	00:06
Breath Volume	2.046	00:06
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.164	00:09
Breath Volume	1.531	00:09
Air Blank	0.000	00:10
Control Test	0.078	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

---

Data Download Date/Time: 11/29/2021 14:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BECKER,ALEXA,C  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: MARTINEZ,JUAN  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 11/22/2021 23:07:34

---

Diagnostics Check		OK	22:55	
Air Blank		0.000	22:55	
Control Test		0.079	22:56	
Air Blank		0.000	22:56	
Subject Sample	# 1	0.162	22:59	Volume Not Met
Breath Volume		0.000	22:59	
Air Blank		0.000	23:00	
Air Blank		0.000	23:02	
Subject Sample	# 2	0.176	23:02	
Breath Volume		1.277	23:02	
Air Blank		0.000	23:03	
Air Blank		0.000	23:05	
Subject Sample	# 3	0.173	23:05	
Breath Volume		1.285	23:05	
Air Blank		0.000	23:06	
Control Test		0.078	23:06	
Air Blank		0.000	23:07	
Diagnostics Check		OK	23:07	

---

Data Download Date/Time: 11/29/2021 14:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CARBALLO,ANGEL,O

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: MORA,M

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 10/25/2021

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 22:05

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: BENITEZ

DATE/TIME OF TEST: 11/05/2021 22:52:20

---

Diagnostics Check	OK	22:51	
Air Blank		22:51	RFI Detect
Air Blank	0.000	22:52	

---

Data Download Date/Time: 11/30/2021 11:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CARBALLO,ANGEL,O

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: MORA,M

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 10/25/2021

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 22:05

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: BENITEZ

DATE/TIME OF TEST: 11/05/2021 22:57:06

---

Diagnostics Check	OK	22:54	
Air Blank	0.000	22:54	
Control Test	0.081	22:54	
Air Blank	0.000	22:55	
Subject Sample	# 1	22:55	Subject Test Refused
Breath Volume	0.000	22:55	
Air Blank	0.000	22:56	
Control Test	0.081	22:56	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

---

Data Download Date/Time: 11/30/2021 11:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PASTOR,JOSE,F  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: MARTINEZ,JOSANDRE  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 10/25/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1506449  
DRY GAS EXPIRY DATE: 06/10/2024  
VIOLATION CODE: DUI  
ARREST TIME: 09:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: HERRERA OVALLE

DATE/TIME OF TEST: 11/12/2021 10:40:50

---

Diagnostics Check	OK	10:33
Air Blank	0.000	10:33
Control Test	0.080	10:34
Air Blank	0.000	10:34
Subject Sample # 1	0.209	10:36
Breath Volume	1.128	10:36
Air Blank	0.000	10:36
Air Blank	0.000	10:38
Subject Sample # 2	0.206	10:39
Breath Volume	1.730	10:39
Air Blank	0.000	10:39
Control Test	0.079	10:40
Air Blank	0.000	10:40
Diagnostics Check	OK	10:40

---

Data Download Date/Time: 11/30/2021 11:48



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,MIGUEL,R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CORDOBA,A  
ARRESTING OFFICER AGENCY: North Miami PD  
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 11/07/2021 05:57:30

---

Diagnostics Check	OK	05:50
Air Blank	0.000	05:51
Control Test	0.077	05:51
Air Blank	0.000	05:51
Subject Sample # 1	0.096	05:52
Breath Volume	1.785	05:52
Air Blank	0.000	05:53
Air Blank	0.000	05:55
Subject Sample # 2	0.096	05:56
Breath Volume	1.871	05:56
Air Blank	0.000	05:56
Control Test	0.078	05:56
Air Blank	0.000	05:57
Diagnostics Check	OK	05:57

---

Data Download Date/Time: 12/09/2021 04:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRUZ,DANIEL,A  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CRUZ,DANIEL  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: MITHCELL

DATE/TIME OF TEST: 11/11/2021 21:20:18

---

Diagnostics Check	OK	21:16	
Air Blank	0.000	21:17	
Control Test	0.078	21:17	
Air Blank	0.000	21:18	
Subject Sample	# 1	21:18	Subject Test Refused
Breath Volume	0.000	21:18	
Air Blank	0.000	21:19	
Control Test	0.079	21:19	
Air Blank	0.000	21:20	
Diagnostics Check	OK	21:20	

---

Data Download Date/Time: 12/09/2021 04:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELGADO,MIGUEL,R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: DELGADO,M  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: BLOOD

DATE/TIME OF TEST: 11/12/2021 01:41:03

---

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.151	01:36
Breath Volume	3.066	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:38
Subject Sample # 2	0.147	01:39
Breath Volume	2.894	01:39
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:41

---

Data Download Date/Time: 12/09/2021 04:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OCEGUERA,YASMANI,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: OCEGUERA,YASMANI  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 11/05/2021 22:44:42

---

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.079	22:37
Air Blank	0.000	22:37
Subject Sample # 1	0.152	22:40
Breath Volume	3.292	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:42
Subject Sample # 2	0.136	22:43
Breath Volume	2.332	22:43
Air Blank	0.000	22:43
Control Test	0.077	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

---

Data Download Date/Time: 12/02/2021 06:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OCEGUERA,YASMANI,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: OCEGUERA,YASMANI  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MATOMDELEON

DATE/TIME OF TEST: 11/14/2021 23:48:36

---

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.080	23:42
Air Blank	0.000	23:43
Subject Sample # 1	0.296	23:43
Breath Volume	2.304	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.302	23:46
Breath Volume	1.359	23:46
Air Blank	0.000	23:47
Control Test	0.079	23:48
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

---

Data Download Date/Time: 12/02/2021 06:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAVINON,CARLOS,M  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: SAVINON,CARLOS  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 11/16/2021 08:43:35

---

Diagnostics Check	OK	08:40	
Air Blank	0.000	08:41	
Control Test	0.081	08:41	
Air Blank	0.000	08:42	
Subject Sample	# 1	08:42	Subject Test Refused
Breath Volume	0.000	08:42	
Air Blank	0.000	08:42	
Control Test	0.080	08:42	
Air Blank	0.000	08:43	
Diagnostics Check	OK	08:43	

---

Data Download Date/Time: 12/02/2021 06:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARAUJO,AARON,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: ARAUJO,AARON  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 07:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:33

SUBJECT'S LAST NAME: YANCE

DATE/TIME OF TEST: 11/18/2021 16:15:57

---

Diagnostics Check	OK	16:08
Air Blank	0.000	16:09
Control Test	0.081	16:09
Air Blank	0.000	16:10
Subject Sample # 1	0.124	16:11
Breath Volume	2.492	16:11
Air Blank	0.000	16:12
Air Blank	0.000	16:13
Subject Sample # 2	0.123	16:14
Breath Volume	3.203	16:14
Air Blank	0.000	16:15
Control Test	0.081	16:15
Air Blank	0.000	16:15
Diagnostics Check	OK	16:15

---

Data Download Date/Time: 12/02/2021 06:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OROL,MANUEL,J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: OROL,MANUEL  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 11/07/2021 23:48:33

---

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.080	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.220	23:41
Breath Volume	2.398	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.220	23:46
Breath Volume	1.125	23:46
Air Blank	0.000	23:47
Control Test	0.081	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

---

Data Download Date/Time: 12/02/2021 07:15



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TOME,VERA,L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: TOME,VERA  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 13:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: HERRERA PINEDA

DATE/TIME OF TEST: 11/13/2021 14:29:21

---

Diagnostics Check	OK	14:26	
Air Blank	0.000	14:26	
Control Test	0.080	14:26	
Air Blank	0.000	14:27	
Subject Sample	# 1	14:28	Subject Test Refused
Breath Volume	0.000	14:28	
Air Blank	0.000	14:28	
Control Test	0.081	14:28	
Air Blank	0.000	14:29	
Diagnostics Check	OK	14:29	

---

Data Download Date/Time: 12/02/2021 07:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORENO,FRANCISCO,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: QUINTANA,SILVIO  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:58

SUBJECT'S LAST NAME: SOTOLONGO

DATE/TIME OF TEST: 11/21/2021 07:25:31

---

Diagnostics Check	OK	07:18
Air Blank	0.000	07:19
Control Test	0.080	07:19
Air Blank	0.000	07:20
Subject Sample # 1	0.219	07:21
Breath Volume	1.695	07:21
Air Blank	0.000	07:21
Air Blank	0.000	07:23
Subject Sample # 2	0.213	07:23
Breath Volume	1.238	07:23
Air Blank	0.000	07:24
Control Test	0.079	07:24
Air Blank	0.000	07:25
Diagnostics Check	OK	07:25

---

Data Download Date/Time: 12/02/2021 07:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OROL,MANUEL,J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: OROL,MANUEL  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 13:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: BARRIOS VILLACINDE

DATE/TIME OF TEST: 11/21/2021 14:38:34

---

Diagnostics Check	OK	14:35	
Air Blank	0.000	14:36	
Control Test	0.080	14:36	
Air Blank	0.000	14:37	
Subject Sample	# 1	14:37	Subject Test Refused
Breath Volume	0.000	14:37	
Air Blank	0.000	14:37	
Control Test	0.080	14:37	
Air Blank	0.000	14:38	
Diagnostics Check	OK	14:38	

---

Data Download Date/Time: 12/02/2021 07:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FHP TROOP A  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MALONEY,CHRISTOPHER,R  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: ARROYO,WILFREDO  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 11/05/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1327124  
DRY GAS EXPIRY DATE: 05/14/2023  
VIOLATION CODE: DUI  
ARREST TIME: 15:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: HOLLEY

DATE/TIME OF TEST: 11/12/2021 16:12:03

---

Diagnostics Check	OK	16:02
Air Blank	0.000	16:02
Control Test	0.083	16:02
Air Blank	0.000	16:03
Subject Sample # 1	0.159	16:04
Breath Volume	1.218	16:04
Air Blank	0.000	16:04
Air Blank	0.000	16:06
Subject Sample # 2	0.181	16:07
Breath Volume	1.722	16:07
Air Blank	0.000	16:08
Air Blank	0.000	16:09
Subject Sample # 3	0.178	16:10
Breath Volume	1.878	16:10
Air Blank	0.000	16:11
Control Test	0.078	16:11
Air Blank	0.000	16:11
Diagnostics Check	OK	16:12

---

Data Download Date/Time: 12/20/2021 10:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FHP TROOP A  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MALONEY,CHRISTOPHER,R  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: MALONEY,CHRISTOPHER  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 11/05/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1327124  
DRY GAS EXPIRY DATE: 05/14/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: LEWIS  
DATE/TIME OF TEST: 11/26/2021 22:51:17

---

Diagnostics Check	OK	22:47	
Air Blank	0.000	22:47	
Control Test	0.080	22:48	
Air Blank	0.000	22:48	
Subject Sample	# 1	22:49	Interferent Detect
Breath Volume	2.750	22:49	
Air Blank	0.000	22:50	
Control Test	0.084	22:50	
Air Blank	0.000	22:51	
Diagnostics Check	OK	22:51	

---

Data Download Date/Time: 12/20/2021 10:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FHP TROOP A  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MALONEY,CHRISTOPHER,R  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: MALONEY,CHRISTOPHER  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 11/05/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1327124  
DRY GAS EXPIRY DATE: 05/14/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 11/26/2021 23:01:39

---

Diagnostics Check	OK	22:54
Air Blank	0.000	22:55
Control Test	0.079	22:55
Air Blank	0.000	22:55
Subject Sample # 1	0.129	22:57
Breath Volume	2.941	22:57
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.114	23:00
Breath Volume	1.253	23:00
Air Blank	0.000	23:00
Control Test	0.080	23:01
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

---

Data Download Date/Time: 12/20/2021 10:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000903

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRODBECK,JAMES,J  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: VILLARIN,JENNIFER  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1380119  
DRY GAS EXPIRY DATE: 09/22/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: TRAN

DATE/TIME OF TEST: 11/25/2021 00:04:58

---

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.080	23:58
Air Blank	0.000	23:58
Subject Sample # 1	0.141	00:00
Breath Volume	1.773	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.127	00:03
Breath Volume	1.687	00:03
Air Blank	0.000	00:04
Control Test	0.078	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

---

Data Download Date/Time: 12/29/2021 10:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000903

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRODBECK,JAMES,J  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: FLINK,ANDREW  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1380119  
DRY GAS EXPIRY DATE: 09/22/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: CHIARAVALLE

DATE/TIME OF TEST: 11/25/2021 01:10:20

---

Diagnostics Check	OK	01:03
Air Blank	0.000	01:04
Control Test	0.080	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.166	01:05
Breath Volume	2.042	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.162	01:08
Breath Volume	2.683	01:08
Air Blank	0.000	01:09
Control Test	0.079	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

---

Data Download Date/Time: 12/29/2021 10:18



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/18/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: CARROZZA

DATE/TIME OF TEST: 11/01/2021 01:12:30

---

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.078	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.198	01:07
Breath Volume	2.316	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.203	01:11
Breath Volume	2.847	01:11
Air Blank	0.000	01:11
Control Test	0.078	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

---

Data Download Date/Time: 11/16/2021 10:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENYAN,M,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: KENYAN,MA  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/18/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 11/04/2021 07:36:23

---

Diagnostics Check	OK	07:33	
Air Blank	0.000	07:34	
Control Test	0.077	07:34	
Air Blank	0.000	07:34	
Subject Sample	# 1	07:35	Subject Test Refused
Breath Volume	0.000	07:35	
Air Blank	0.000	07:35	
Control Test	0.078	07:35	
Air Blank	0.000	07:36	
Diagnostics Check	OK	07:36	

---

Data Download Date/Time: 11/16/2021 10:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/18/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: TOMLINSON III

DATE/TIME OF TEST: 11/06/2021 03:55:35

---

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.078	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.123	03:50
Breath Volume		1.781	03:50
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.129	03:54
Breath Volume		2.058	03:54
Air Blank		0.000	03:54
Control Test		0.078	03:55
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

---

Data Download Date/Time: 11/16/2021 10:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMANDL,REAGHN,S  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: SIMANDL,REAGHN  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/18/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 05:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:11

SUBJECT'S LAST NAME: SARMIENTO AMAYA

DATE/TIME OF TEST: 11/07/2021 06:40:22

---

Diagnostics Check	OK	06:32
Air Blank	0.000	06:32
Control Test	0.079	06:32
Air Blank	0.000	06:33
Subject Sample # 1	0.134	06:35
Breath Volume	2.457	06:35
Air Blank	0.000	06:36
Air Blank	0.000	06:37
Subject Sample # 2	0.135	06:38
Breath Volume	2.605	06:38
Air Blank	0.000	06:39
Control Test	0.076	06:39
Air Blank	0.000	06:40
Diagnostics Check	OK	06:40

---

Data Download Date/Time: 11/16/2021 10:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRERO,FRAICOR,  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: DAVIS,DONAIL  
ARRESTING OFFICER AGENCY: Bradenton PD  
LAST AGENCY INSPECTION: 10/18/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 11/09/2021 02:03:46

---

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.077	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.189	01:59
Breath Volume	1.371	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.186	02:02
Breath Volume	1.460	02:02
Air Blank	0.000	02:02
Control Test	0.078	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

---

Data Download Date/Time: 11/16/2021 10:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRERO,FRAICOR,  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: TERRERO,FRAICOR  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/18/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: WINNINGHAM

DATE/TIME OF TEST: 11/14/2021 02:01:21

---

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.078	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.157	01:57
Breath Volume	1.820	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.159	01:59
Breath Volume	2.250	01:59
Air Blank	0.000	02:00
Control Test	0.078	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

---

Data Download Date/Time: 11/16/2021 10:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/18/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: VASQUEZ GUZMAN

DATE/TIME OF TEST: 11/14/2021 03:30:31

---

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.077	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.097	03:26
Breath Volume	1.386	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.095	03:29
Breath Volume	2.074	03:29
Air Blank	0.000	03:29
Control Test	0.078	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

---

Data Download Date/Time: 11/16/2021 10:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 10/18/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: OCONNELL

DATE/TIME OF TEST: 11/14/2021 19:40:39

---

Diagnostics Check	OK	19:29	
Air Blank	0.000	19:29	
Control Test	0.078	19:30	
Air Blank	0.000	19:30	
Subject Sample	# 1 0.140	19:33	Volume Not Met
Breath Volume	0.015	19:33	
Air Blank	0.000	19:34	
Air Blank	0.000	19:36	
Subject Sample	# 2 0.016	19:39	Volume Not Met
Breath Volume	0.148	19:39	
Air Blank	0.000	19:39	
Control Test	0.000	19:40	Control Outside Tolerance
Air Blank	0.000	19:40	

---

Data Download Date/Time: 11/16/2021 10:53



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNOZ,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Village of Pinecrest  
ARRESTING OFFICER: CRUZ,EDISON  
ARRESTING OFFICER AGENCY: Village of Pinecrest  
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1380119  
DRY GAS EXPIRY DATE: 09/22/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: MONTANEZ RESTO

DATE/TIME OF TEST: 11/19/2021 23:38:52

---

Diagnostics Check	OK	23:36	
Air Blank	0.000	23:36	
Control Test	0.081	23:36	
Air Blank	0.000	23:37	
Subject Sample	# 1	23:37	Subject Test Refused
Breath Volume	0.000	23:37	
Air Blank	0.000	23:37	
Control Test	0.079	23:38	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

---

Data Download Date/Time: 11/25/2021 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNOZ,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Village of Pinecrest  
ARRESTING OFFICER: CARRASCO,JONATHAN  
ARRESTING OFFICER AGENCY: Village of Pinecrest  
LAST AGENCY INSPECTION: 10/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1380119  
DRY GAS EXPIRY DATE: 09/22/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: RODRIGUEZ GARCIA

DATE/TIME OF TEST: 11/20/2021 02:02:00

---

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:59	
Control Test	0.081	01:59	
Air Blank	0.000	01:59	
Subject Sample	# 1	02:00	Subject Test Refused
Breath Volume	0.000	02:00	
Air Blank	0.000	02:01	
Control Test	0.080	02:01	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:02	

---

Data Download Date/Time: 11/25/2021 15:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNOZ,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Village of Pinecrest  
ARRESTING OFFICER: SCHRY,HEATHER  
ARRESTING OFFICER AGENCY: Village of Pinecrest  
LAST AGENCY INSPECTION: 11/25/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1380119  
DRY GAS EXPIRY DATE: 09/22/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: DONES

DATE/TIME OF TEST: 11/25/2021 21:38:25

---

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:36	
Control Test	0.081	21:36	
Air Blank	0.000	21:36	
Subject Sample	# 1	21:37	Subject Test Refused
Breath Volume	0.000	21:37	
Air Blank	0.000	21:37	
Control Test	0.081	21:37	
Air Blank	0.000	21:38	
Diagnostics Check	OK	21:38	

---

Data Download Date/Time: 12/21/2021 16:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VEGA,OBEL,  
BREATH TEST OPERATOR AGENCY: Village of Pinecrest  
ARRESTING OFFICER: VEGA,OBEL  
ARRESTING OFFICER AGENCY: Village of Pinecrest  
LAST AGENCY INSPECTION: 11/25/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1380119  
DRY GAS EXPIRY DATE: 09/22/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: DE LA GRANA CHACON

DATE/TIME OF TEST: 11/28/2021 05:22:06

---

Diagnostics Check	OK	05:19	
Air Blank	0.000	05:19	
Control Test	0.081	05:19	
Air Blank	0.000	05:20	
Subject Sample	# 1	05:20	Subject Test Refused
Breath Volume	0.000	05:20	
Air Blank	0.000	05:21	
Control Test	0.081	05:21	
Air Blank	0.000	05:22	
Diagnostics Check	OK	05:22	

---

Data Download Date/Time: 12/21/2021 16:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KING, JONATHAN, L  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: HERNANDEZ BELLO, PABLO  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 10/20/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: PABLO PEREZ

DATE/TIME OF TEST: 11/01/2021 02:22:17

---

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.080	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.243	02:17
Breath Volume		1.218	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.239	02:20
Breath Volume		1.660	02:20
Air Blank		0.000	02:21
Control Test		0.077	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

---

Data Download Date/Time: 11/09/2021 09:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON, TORI, D  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: JOHNSTON, TORI  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 10/20/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: B255512707190

DATE/TIME OF TEST: 11/02/2021 20:22:30

---

Diagnostics Check	OK	20:16
Air Blank	0.000	20:16
Control Test	0.080	20:17
Air Blank	0.000	20:17
Subject Sample # 1	0.271	20:17
Breath Volume	2.859	20:17
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample # 2	0.269	20:20
Breath Volume	2.734	20:20
Air Blank	0.000	20:21
Control Test	0.077	20:21
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

---

Data Download Date/Time: 11/09/2021 09:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINK,CHRIS,D  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LINK,CHRIS  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/20/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 17:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 11/03/2021 18:04:58

---

Diagnostics Check	OK	17:58
Air Blank	0.000	17:59
Control Test	0.079	17:59
Air Blank	0.000	17:59
Subject Sample # 1	0.127	18:00
Breath Volume	1.570	18:00
Air Blank	0.000	18:01
Air Blank	0.000	18:02
Subject Sample # 2	0.113	18:03
Breath Volume	1.695	18:03
Air Blank	0.000	18:04
Control Test	0.078	18:04
Air Blank	0.000	18:04
Diagnostics Check	OK	18:04

---

Data Download Date/Time: 11/09/2021 09:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINK,CHRIS,D  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LINK,CHRIS  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/20/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 17:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:17

SUBJECT'S LAST NAME: ALLEVA

DATE/TIME OF TEST: 11/04/2021 18:45:41

---

Diagnostics Check	OK	18:39
Air Blank	0.000	18:39
Control Test	0.080	18:40
Air Blank	0.000	18:40
Subject Sample # 1	0.268	18:41
Breath Volume	2.921	18:41
Air Blank	0.000	18:41
Air Blank	0.000	18:43
Subject Sample # 2	0.262	18:44
Breath Volume	1.832	18:44
Air Blank	0.000	18:44
Control Test	0.078	18:45
Air Blank	0.000	18:45
Diagnostics Check	OK	18:45

---

Data Download Date/Time: 11/09/2021 09:21



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SELLERS,BRANDON,J  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: CARSON,TISHA  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/20/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 11/05/2021 20:42:15

---

Diagnostics Check	OK	20:36
Air Blank	0.000	20:36
Control Test	0.080	20:36
Air Blank	0.000	20:37
Subject Sample # 1	0.000	20:37
Breath Volume	3.226	20:37
Air Blank	0.000	20:38
Air Blank	0.000	20:40
Subject Sample # 2	0.000	20:40
Breath Volume	2.367	20:40
Air Blank	0.000	20:41
Control Test	0.079	20:41
Air Blank	0.000	20:42
Diagnostics Check	OK	20:42

---

Data Download Date/Time: 11/09/2021 09:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SELLERS,BRANDON,J  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: OTERD,TYLER  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/20/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: ARDILA BARLIS

DATE/TIME OF TEST: 11/05/2021 20:58:36

---

Diagnostics Check	OK	20:52	
Air Blank	0.000	20:53	
Control Test	0.080	20:53	
Air Blank	0.000	20:54	
Subject Sample	# 1 0.158	20:57	Volume Not Met
Breath Volume	0.796	20:57	
Air Blank	0.000	20:58	
Air Blank	0.000	20:58	

---

Data Download Date/Time: 11/09/2021 09:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON, TORI,  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: JOHNSTON, TORI  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 10/20/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 14:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: VIOLETTE

DATE/TIME OF TEST: 11/06/2021 15:37:41

---

Diagnostics Check	OK	15:31
Air Blank	0.000	15:31
Control Test	0.081	15:32
Air Blank	0.000	15:32
Subject Sample # 1	0.076	15:33
Breath Volume	3.007	15:33
Air Blank	0.000	15:33
Air Blank	0.000	15:35
Subject Sample # 2	0.072	15:36
Breath Volume	3.093	15:36
Air Blank	0.000	15:36
Control Test	0.078	15:37
Air Blank	0.000	15:37
Diagnostics Check	OK	15:37

---

Data Download Date/Time: 11/09/2021 09:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SELLERS,BRANDON,J  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: DUBOIS,MICHAEL  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/20/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: WYATT

DATE/TIME OF TEST: 11/07/2021 00:08:12

---

Diagnostics Check		OK	00:00
Air Blank		0.000	00:01
Control Test		0.080	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.212	00:03
Breath Volume		1.589	00:03
Air Blank		0.000	00:04
Air Blank		0.000	00:05
Subject Sample	# 2	0.215	00:06
Breath Volume		1.902	00:06
Air Blank		0.000	00:07
Control Test		0.078	00:07
Air Blank		0.000	00:08
Diagnostics Check		OK	00:08

---

Data Download Date/Time: 11/09/2021 09:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SELLERS,BRANDON,J  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: STRANGE,KATHERINE  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 10/20/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: MENA

DATE/TIME OF TEST: 11/07/2021 23:26:24

---

Diagnostics Check	OK	23:19
Air Blank	0.000	23:20
Control Test	0.080	23:20
Air Blank	0.000	23:21
Subject Sample # 1	0.202	23:21
Breath Volume	2.214	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.204	23:24
Breath Volume	2.476	23:24
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

---

Data Download Date/Time: 11/09/2021 09:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TOTU,ROBERT,M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: TOTU,ROBERT  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 12:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:23

SUBJECT'S LAST NAME: FUSSE

DATE/TIME OF TEST: 11/10/2021 13:52:51

---

Diagnostics Check	OK	13:46	
Air Blank	0.000	13:47	
Control Test	0.079	13:47	
Air Blank	0.000	13:47	
Subject Sample	# 1 0.316	13:48	
Breath Volume	2.777	13:48	
Air Blank	0.000	13:49	
Air Blank	0.000	13:50	
Subject Sample	# 2 0.303	13:51	
Breath Volume	2.386	13:51	
Air Blank		13:52	Purge Fail
Air Blank	0.000	13:52	

---

Data Download Date/Time: 12/08/2021 09:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, JOSIAH, R  
BREATH TEST OPERATOR AGENCY: Lee Port Authority  
ARRESTING OFFICER: SMITH, JOSIAH  
ARRESTING OFFICER AGENCY: Lee Port Authority  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: PICKETT

DATE/TIME OF TEST: 11/11/2021 01:26:41

---

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.079	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.000	01:22
Breath Volume	1.546	01:22
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.000	01:25
Breath Volume	1.710	01:25
Air Blank	0.000	01:25
Control Test	0.079	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

---

Data Download Date/Time: 12/08/2021 09:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: HARRISON,TRAVIS  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 11/11/2021 02:15:22

---

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.079	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.225	02:10
Breath Volume	3.414	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.227	02:13
Breath Volume	3.113	02:13
Air Blank	0.000	02:14
Control Test	0.079	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

---

Data Download Date/Time: 12/08/2021 09:49



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAHLKE,M,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: WEISS,M  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: DESROSIERS

DATE/TIME OF TEST: 11/12/2021 05:10:30

---

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.080	05:04
Air Blank	0.000	05:04
Subject Sample # 1	0.137	05:05
Breath Volume	3.417	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.132	05:08
Breath Volume	3.492	05:08
Air Blank	0.000	05:09
Control Test	0.079	05:09
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

---

Data Download Date/Time: 12/08/2021 09:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON, TORI, D  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: JOHNSTON, TORI  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: DE MAYO

DATE/TIME OF TEST: 11/13/2021 04:16:45

---

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.167	04:11
Breath Volume	2.511	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.158	04:15
Breath Volume	2.023	04:15
Air Blank	0.000	04:15
Control Test	0.078	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

---

Data Download Date/Time: 12/08/2021 09:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KING, JONATHAN, L  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: HERNANDEZ BELLO, PABLO  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HINDS

DATE/TIME OF TEST: 11/14/2021 03:02:36

---

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.153	02:57
Breath Volume	1.242	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.158	03:01
Breath Volume	1.312	03:01
Air Blank	0.000	03:01
Control Test	0.077	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

---

Data Download Date/Time: 12/08/2021 09:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINK,CHRIS,D  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LINK,CHRIS  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A1  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 11/14/2021 03:37:22

---

Diagnostics Check		OK	03:24	
Air Blank		0.000	03:25	
Control Test		0.079	03:25	
Air Blank		0.000	03:26	
Subject Sample	# 1	0.088	03:29	Volume Not Met
Breath Volume		0.121	03:29	
Air Blank		0.000	03:30	
Air Blank		0.000	03:31	
Subject Sample	# 2	0.106	03:32	
Breath Volume		1.335	03:32	
Air Blank		0.000	03:33	
Air Blank		0.000	03:35	
Subject Sample	# 3	0.101	03:35	
Breath Volume		1.656	03:35	
Air Blank		0.000	03:36	
Control Test		0.078	03:36	
Air Blank		0.000	03:37	
Diagnostics Check		OK	03:37	

---

Data Download Date/Time: 12/08/2021 09:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINK,CHRIS,D  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: CHIAPPY,C  
ARRESTING OFFICER AGENCY: Sanibel PD  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A1  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: CAVANAGH

DATE/TIME OF TEST: 11/14/2021 04:51:57

---

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.079	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.155	04:47
Breath Volume	2.058	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:49
Subject Sample # 2	0.153	04:50
Breath Volume	1.371	04:50
Air Blank	0.000	04:51
Control Test	0.078	04:51
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

---

Data Download Date/Time: 12/08/2021 09:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HALL,CHRISINE,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: HALL,CHRISTINE  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A1  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: MAHAN

DATE/TIME OF TEST: 11/14/2021 22:05:49

---

Diagnostics Check	OK	22:00	
Air Blank	0.000	22:01	
Control Test	0.080	22:01	
Air Blank	0.000	22:02	
Subject Sample	# 1	22:04	Subject Test Refused
Breath Volume	0.000	22:04	
Air Blank	0.000	22:04	
Control Test	0.079	22:05	
Air Blank	0.000	22:05	
Diagnostics Check	OK	22:05	

---

Data Download Date/Time: 12/08/2021 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINK,CHRIS,D  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LINK,CHRIS  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A1  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: SANTIAGOLAZARO

DATE/TIME OF TEST: 11/15/2021 02:29:46

---

Diagnostics Check		OK	02:17	
Air Blank		0.000	02:18	
Control Test		0.079	02:18	
Air Blank		0.000	02:18	
Subject Sample	# 1	0.117	02:22	Volume Not Met
Breath Volume		0.675	02:22	
Air Blank		0.000	02:23	
Air Blank		0.000	02:24	
Subject Sample	# 2	0.111	02:25	
Breath Volume		1.843	02:25	
Air Blank		0.000	02:25	
Air Blank		0.000	02:27	
Subject Sample	# 3	0.117	02:28	
Breath Volume		1.769	02:28	
Air Blank		0.000	02:28	
Control Test		0.078	02:29	
Air Blank		0.000	02:29	
Diagnostics Check		OK	02:29	

---

Data Download Date/Time: 12/08/2021 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWE,ALEXANDER,A  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MILLIGAN,MATTHEW  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: RYBICKI

DATE/TIME OF TEST: 11/16/2021 03:30:13

---

Diagnostics Check	OK	03:29	
Air Blank		03:29	RFI Detect
Air Blank	0.000	03:30	

---

Data Download Date/Time: 12/08/2021 09:50



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWE,ALEXANDER,A  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: MILLIGAN,MATTHEW  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: RYBICKI

DATE/TIME OF TEST: 11/16/2021 03:38:39

---

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.080	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.166	03:33
Breath Volume		1.472	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:35
Subject Sample	# 2	0.177	03:36
Breath Volume		1.351	03:36
Air Blank		0.000	03:37
Control Test		0.080	03:38
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

---

Data Download Date/Time: 12/08/2021 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOWE,ALEXANDER,A  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: BOCK,DAVID  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: DEIBERT

DATE/TIME OF TEST: 11/16/2021 03:48:52

---

Diagnostics Check		OK	03:41
Air Blank		0.000	03:42
Control Test		0.078	03:42
Air Blank		0.000	03:43
Subject Sample	# 1	0.232	03:44
Breath Volume		1.132	03:44
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.250	03:47
Breath Volume		1.675	03:47
Air Blank		0.000	03:47
Control Test		0.077	03:48
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

---

Data Download Date/Time: 12/08/2021 09:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BONSALL, MATTHEW, A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BYRD, B  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: AKERLEY

DATE/TIME OF TEST: 11/17/2021 02:16:25

---

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:11	
Control Test	0.080	02:11	
Air Blank	0.000	02:11	
Subject Sample # 1	0.107	02:12	
Breath Volume	1.855	02:12	
Air Blank	0.000	02:13	
Air Blank	0.000	02:15	
Subject Sample # 2	0.108	02:15	
Breath Volume	1.667	02:15	
Air Blank		02:15	RFI Detect
Air Blank	0.000	02:16	

---

Data Download Date/Time: 12/08/2021 09:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BONSALL, MATTHEW, A  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BYRD, B  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: AKERLEY

DATE/TIME OF TEST: 11/17/2021 02:26:48

---

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.079	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.107	02:21
Breath Volume	1.855	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.109	02:25
Breath Volume	1.722	02:25
Air Blank	0.000	02:25
Control Test	0.079	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

---

Data Download Date/Time: 12/08/2021 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANPELT,J,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: KAZIUK,M  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: FACELLA

DATE/TIME OF TEST: 11/17/2021 03:25:31

---

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.079	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.169	03:20
Breath Volume	3.000	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.154	03:23
Breath Volume	2.480	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

---

Data Download Date/Time: 12/08/2021 09:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KING, JONATHAN, L  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: LESSARD, JEREMY  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 11/18/2021 03:18:39

---

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.079	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.125	03:14
Breath Volume	3.082	03:14
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.121	03:17
Breath Volume	2.847	03:17
Air Blank	0.000	03:17
Control Test	0.078	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

---

Data Download Date/Time: 12/08/2021 09:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SELLERS,BRANDON,J  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BRUNICK,JACOB  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: KRUG

DATE/TIME OF TEST: 11/19/2021 20:07:58

---

Diagnostics Check	OK	20:05	
Air Blank	0.000	20:06	
Control Test	0.079	20:06	
Air Blank	0.000	20:07	
Subject Sample	# 1	20:07	Sequence Aborted
Breath Volume	0.000	20:07	
Air Blank	0.000	20:07	

---

Data Download Date/Time: 12/08/2021 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SELLERS,BRANDON,J  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BRUNICK,JACOB  
ARRESTING OFFICER AGENCY: Lee Port Authority  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: KRUG

DATE/TIME OF TEST: 11/19/2021 20:16:58

---

Diagnostics Check		OK	20:10
Air Blank		0.000	20:10
Control Test		0.079	20:10
Air Blank		0.000	20:11
Subject Sample	# 1	0.229	20:12
Breath Volume		2.933	20:12
Air Blank		0.000	20:13
Air Blank		0.000	20:14
Subject Sample	# 2	0.226	20:15
Breath Volume		2.539	20:15
Air Blank		0.000	20:16
Control Test		0.079	20:16
Air Blank		0.000	20:16
Diagnostics Check		OK	20:16

---

Data Download Date/Time: 12/08/2021 09:50



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SELLERS,BRANDON,J  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: KIRK,PATRICK  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 16:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 11/19/2021 21:49:44

---

Diagnostics Check	OK	21:42
Air Blank	0.000	21:43
Control Test	0.080	21:43
Air Blank	0.000	21:44
Subject Sample # 1	0.096	21:44
Breath Volume	1.757	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.096	21:48
Breath Volume	2.070	21:48
Air Blank	0.000	21:48
Control Test	0.080	21:49
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

---

Data Download Date/Time: 12/08/2021 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SELLERS,BRANDON,J  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: ECKERT,KYLE  
ARRESTING OFFICER AGENCY: Sanibel PD  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: WESSLING

DATE/TIME OF TEST: 11/20/2021 03:51:36

---

Diagnostics Check		OK	03:41
Air Blank		0.000	03:42
Control Test		0.079	03:42
Air Blank		0.000	03:43
Subject Sample	# 1	0.184	03:43
Breath Volume		1.988	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.209	03:46
Breath Volume		1.957	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 3	0.178	03:49
Breath Volume		1.691	03:49
Air Blank		0.000	03:50
Control Test		0.078	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

---

Data Download Date/Time: 12/08/2021 09:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,JUSTIN,L  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: LEHMANN,DAMIAN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 11/20/2021 23:51:09

---

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.080	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.215	23:45
Breath Volume	3.316	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.215	23:49
Breath Volume	4.726	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

---

Data Download Date/Time: 12/08/2021 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,JAMES,R  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: MILLIGAN,MATTHEW  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: PAETOW-FANNING

DATE/TIME OF TEST: 11/21/2021 02:21:36

---

Diagnostics Check		OK	02:14
Air Blank		0.000	02:15
Control Test		0.079	02:15
Air Blank		0.000	02:16
Subject Sample	# 1	0.155	02:17
Breath Volume		1.328	02:17
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.140	02:19
Breath Volume		1.386	02:19
Air Blank		0.000	02:20
Control Test		0.079	02:20
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

---

Data Download Date/Time: 12/08/2021 09:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,J,T  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: MILLER,J  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SCARNATO JR

DATE/TIME OF TEST: 11/21/2021 03:29:07

---

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.000	03:24
Breath Volume	2.386	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.000	03:27
Breath Volume	2.367	03:27
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

---

Data Download Date/Time: 12/08/2021 09:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVENDANO,JUAN,M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: CAIL,MATTHEW  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: PEREZ LOPEZ

DATE/TIME OF TEST: 11/23/2021 02:58:34

---

Diagnostics Check		OK	02:51
Air Blank		0.000	02:51
Control Test		0.080	02:52
Air Blank		0.000	02:52
Subject Sample	# 1	0.143	02:53
Breath Volume		2.285	02:53
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.139	02:56
Breath Volume		2.238	02:56
Air Blank		0.000	02:57
Control Test		0.079	02:57
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

---

Data Download Date/Time: 12/08/2021 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,J,T  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: MILLER,J  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SALGADO

DATE/TIME OF TEST: 11/25/2021 04:28:23

---

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:24	
Control Test	0.081	04:25	
Air Blank	0.000	04:25	
Subject Sample # 1	0.184	04:26	
Breath Volume	1.421	04:26	
Air Blank		04:27	Purge Fail
Air Blank	0.000	04:28	

---

Data Download Date/Time: 12/08/2021 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,J,T  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: MILLER,J  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SALGADO

DATE/TIME OF TEST: 11/25/2021 04:36:15

---

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.076	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.185	04:31
Breath Volume	2.070	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.173	04:34
Breath Volume	1.574	04:34
Air Blank	0.000	04:35
Control Test	0.076	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

---

Data Download Date/Time: 12/08/2021 09:50



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON, TORI, D  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: JOHNSTON, TORI  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: GERRITSEN

DATE/TIME OF TEST: 11/27/2021 02:00:49

---

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:57	
Control Test	0.079	01:58	
Air Blank	0.000	01:58	
Subject Sample	# 1	01:59	Subject Test Refused
Breath Volume	0.000	01:59	
Air Blank	0.000	01:59	
Control Test	0.079	02:00	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

---

Data Download Date/Time: 12/08/2021 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,TORI,D  
BREATH TEST OPERATOR AGENCY: Avon Park PD  
ARRESTING OFFICER: JOHNSTON,TORI  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 11/27/2021 02:56:28

---

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:51	
Control Test	0.080	02:52	
Air Blank	0.000	02:52	
Subject Sample	# 1 0.151	02:53	
Breath Volume	2.062	02:53	
Air Blank	0.000	02:53	
Air Blank	0.000	02:55	
Subject Sample	# 2 0.144	02:55	
Breath Volume	2.152	02:55	
Air Blank		02:56	RFI Detect
Air Blank		02:56	RFI Detect

---

Data Download Date/Time: 12/08/2021 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINK,CHRIS,D  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: LINK,CHRIS  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A1  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: MCKAY

DATE/TIME OF TEST: 11/27/2021 20:03:07

---

Diagnostics Check	OK	19:53
Air Blank	0.000	19:53
Control Test	0.080	19:54
Air Blank	0.000	19:54
Subject Sample # 1	0.128	19:55
Breath Volume	2.824	19:55
Air Blank	0.000	19:55
Air Blank	0.000	19:57
Subject Sample # 2	0.106	19:58
Breath Volume	1.535	19:58
Air Blank	0.000	19:59
Air Blank	0.000	20:00
Subject Sample # 3	0.115	20:01
Breath Volume	3.023	20:01
Air Blank	0.000	20:02
Control Test	0.078	20:02
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

---

Data Download Date/Time: 12/08/2021 09:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JAYDA,J  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: CHIAPPY,C  
ARRESTING OFFICER AGENCY: Sanibel PD  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A1  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 11/27/2021 20:40:01

---

Diagnostics Check	OK	20:33
Air Blank	0.000	20:33
Control Test	0.080	20:34
Air Blank	0.000	20:34
Subject Sample # 1	0.199	20:35
Breath Volume	1.898	20:35
Air Blank	0.000	20:36
Air Blank	0.000	20:37
Subject Sample # 2	0.214	20:38
Breath Volume	1.851	20:38
Air Blank	0.000	20:39
Control Test	0.077	20:39
Air Blank	0.000	20:39
Diagnostics Check	OK	20:40

---

Data Download Date/Time: 12/08/2021 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,J,T  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: MILLER,J  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GROTHENDICK

DATE/TIME OF TEST: 11/30/2021 02:44:29

---

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:42	
Control Test	0.080	02:42	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:43	Subject Test Refused
Breath Volume	0.000	02:43	
Air Blank	0.000	02:43	
Control Test	0.080	02:43	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:44	

---

Data Download Date/Time: 12/08/2021 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAHLKE,M,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: ALEJO,T  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22521080A  
DRY GAS EXPIRY DATE: 10/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 17:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: BAUGH

DATE/TIME OF TEST: 11/30/2021 19:37:51

---

Diagnostics Check	OK	19:30
Air Blank	0.000	19:31
Control Test	0.080	19:31
Air Blank	0.000	19:31
Subject Sample # 1	0.114	19:33
Breath Volume	1.195	19:33
Air Blank	0.000	19:34
Air Blank	0.000	19:35
Subject Sample # 2	0.109	19:36
Breath Volume	1.289	19:36
Air Blank	0.000	19:36
Control Test	0.079	19:37
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

---

Data Download Date/Time: 12/08/2021 09:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EPPS,QUINTELLA,  
BREATH TEST OPERATOR AGENCY: Glades County SO  
ARRESTING OFFICER: MURRAY,STACY  
ARRESTING OFFICER AGENCY: Glades County SO  
LAST AGENCY INSPECTION: 10/31/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 00:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 11/05/2021 01:13:53

---

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:11	
Control Test	0.078	01:11	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:12	
Control Test	0.079	01:13	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

---

Data Download Date/Time: 11/09/2021 09:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAXTER,AUDREY,D  
BREATH TEST OPERATOR AGENCY: Glades County SO  
ARRESTING OFFICER: MURRAY,STACY  
ARRESTING OFFICER AGENCY: Glades County SO  
LAST AGENCY INSPECTION: 10/31/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 12:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 11/05/2021 09:06:44

---

Diagnostics Check		OK	09:00
Air Blank		0.000	09:00
Control Test		0.078	09:01
Air Blank		0.000	09:01
Subject Sample	# 1	0.115	09:02
Breath Volume		1.699	09:02
Air Blank		0.000	09:02
Air Blank		0.000	09:04
Subject Sample	# 2	0.120	09:05
Breath Volume		2.011	09:05
Air Blank		0.000	09:05
Control Test		0.078	09:06
Air Blank		0.000	09:06
Diagnostics Check		OK	09:06

---

Data Download Date/Time: 11/09/2021 09:18



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAXTER,AUDREY,D  
BREATH TEST OPERATOR AGENCY: Glades County SO  
ARRESTING OFFICER: MURRAY,STACY  
ARRESTING OFFICER AGENCY: Glades County SO  
LAST AGENCY INSPECTION: 10/31/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 12:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 11/05/2021 10:57:21

---

Diagnostics Check	OK	10:51
Air Blank	0.000	10:51
Control Test	0.078	10:51
Air Blank	0.000	10:52
Subject Sample # 1	0.075	10:52
Breath Volume	1.734	10:52
Air Blank	0.000	10:53
Air Blank	0.000	10:55
Subject Sample # 2	0.070	10:55
Breath Volume	1.664	10:55
Air Blank	0.000	10:56
Control Test	0.076	10:56
Air Blank	0.000	10:57
Diagnostics Check	OK	10:57

---

Data Download Date/Time: 11/09/2021 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEWIS,XAVIER,T  
BREATH TEST OPERATOR AGENCY: Glades County SO  
ARRESTING OFFICER: DIAZ,ROBIN  
ARRESTING OFFICER AGENCY: Glades County SO  
LAST AGENCY INSPECTION: 10/31/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: COLUCCI

DATE/TIME OF TEST: 11/07/2021 20:25:46

---

Diagnostics Check	OK	20:14	
Air Blank	0.000	20:14	
Control Test	0.078	20:15	
Air Blank	0.000	20:15	
Subject Sample	# 1 0.073	20:18	Volume Not Met
Breath Volume	0.066	20:18	
Air Blank	0.000	20:19	
Air Blank	0.000	20:21	
Subject Sample	# 2	20:24	No Sample Provided
Breath Volume	0.000	20:24	
Air Blank	0.000	20:24	
Control Test	0.077	20:25	
Air Blank	0.000	20:25	
Diagnostics Check	OK	20:25	

---

Data Download Date/Time: 11/09/2021 09:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HEWITT,JARED,S  
BREATH TEST OPERATOR AGENCY: Franklin County SO  
ARRESTING OFFICER: COATNEY,MIRANDA  
ARRESTING OFFICER AGENCY: Franklin County SO  
LAST AGENCY INSPECTION: 10/11/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34520080A1  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: RONALD

DATE/TIME OF TEST: 11/06/2021 19:16:43

---

Diagnostics Check	OK	19:10
Air Blank	0.000	19:10
Control Test	0.080	19:11
Air Blank	0.000	19:11
Subject Sample # 1	0.128	19:12
Breath Volume	3.023	19:12
Air Blank	0.000	19:12
Air Blank	0.000	19:14
Subject Sample # 2	0.136	19:15
Breath Volume	1.832	19:15
Air Blank	0.000	19:15
Control Test	0.080	19:16
Air Blank	0.000	19:16
Diagnostics Check	OK	19:16

---

Data Download Date/Time: 11/11/2021 12:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HEWITT,JARED,S  
BREATH TEST OPERATOR AGENCY: Franklin County SO  
ARRESTING OFFICER: LENDERMAN,TYLER  
ARRESTING OFFICER AGENCY: Franklin County SO  
LAST AGENCY INSPECTION: 10/11/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34520080A1  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 11/06/2021 20:35:07

---

Diagnostics Check		OK	20:20	
Air Blank		0.000	20:20	
Control Test		0.080	20:21	
Air Blank		0.000	20:21	
Subject Sample	# 1	0.126	20:24	Slope Not Level
Breath Volume		1.492	20:24	
Air Blank		0.000	20:25	
Air Blank		0.000	20:27	
Subject Sample	# 2	0.124	20:27	
Breath Volume		1.441	20:27	
Air Blank		0.000	20:28	
Air Blank		0.000	20:30	
Subject Sample	# 3	0.120	20:33	
Breath Volume		1.652	20:33	
Air Blank		0.000	20:34	
Control Test		0.080	20:34	
Air Blank		0.000	20:35	
Diagnostics Check		OK	20:35	

---

Data Download Date/Time: 11/11/2021 12:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHILCOX,DAVID,B  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: KELLY,WILLIAM  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 10/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MULLIGAN

DATE/TIME OF TEST: 11/01/2021 02:10:40

---

Diagnostics Check	OK	02:07	
Air Blank	0.000	02:08	
Control Test	0.081	02:08	
Air Blank	0.000	02:09	
Subject Sample	# 1	02:09	Subject Test Refused
Breath Volume	0.000	02:09	
Air Blank	0.000	02:09	
Control Test	0.082	02:10	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

---

Data Download Date/Time: 11/02/2021 13:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHILCOX,DAVID,B  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: PHILCOX,DAVID  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 11/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: BARRINGTON

DATE/TIME OF TEST: 11/02/2021 23:54:48

---

Diagnostics Check		OK	23:48
Air Blank		0.000	23:48
Control Test		0.081	23:48
Air Blank		0.000	23:49
Subject Sample	# 1	0.158	23:49
Breath Volume		1.289	23:49
Air Blank		0.000	23:50
Air Blank		0.000	23:52
Subject Sample	# 2	0.160	23:53
Breath Volume		2.273	23:53
Air Blank		0.000	23:53
Control Test		0.082	23:54
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

---

Data Download Date/Time: 12/14/2021 11:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHELPS,WILLIAM,D  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: PHELPS,WILLIAM  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 11/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: JEFFERSON

DATE/TIME OF TEST: 11/04/2021 01:28:55

---

Diagnostics Check	OK	01:28	
Air Blank		01:28	RFI Detect
Air Blank	0.000	01:28	

---

Data Download Date/Time: 12/14/2021 11:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHELPS,WILLIAM,D  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: PHELPS,WILLIAM  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 11/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: JEFFERSON

DATE/TIME OF TEST: 11/04/2021 01:38:42

---

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.081	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.145	01:32
Breath Volume	2.316	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.142	01:36
Breath Volume	1.937	01:36
Air Blank	0.000	01:37
Control Test	0.084	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

---

Data Download Date/Time: 12/14/2021 11:52



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MATLOCK,LOGAN,R  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: MATLOCK,LOGAN  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 11/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 11/09/2021 04:36:29

---

Diagnostics Check	OK	04:30
Air Blank	0.000	04:30
Control Test	0.081	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.242	04:32
Breath Volume	3.054	04:32
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.224	04:34
Breath Volume	1.671	04:34
Air Blank	0.000	04:35
Control Test	0.082	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

---

Data Download Date/Time: 12/14/2021 11:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: VOLLERTSEN,CRAIG  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 11/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 11/10/2021 07:02:55

---

Diagnostics Check	OK	06:56
Air Blank	0.000	06:57
Control Test	0.081	06:57
Air Blank	0.000	06:58
Subject Sample # 1	0.100	06:58
Breath Volume	3.277	06:58
Air Blank	0.000	06:59
Air Blank	0.000	07:00
Subject Sample # 2	0.101	07:01
Breath Volume	3.050	07:01
Air Blank	0.000	07:02
Control Test	0.083	07:02
Air Blank	0.000	07:02
Diagnostics Check	OK	07:02

---

Data Download Date/Time: 12/14/2021 11:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHEFFIELD, BRANDON, S  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: HEDGES, MATTHEW  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 11/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:53

SUBJECT'S LAST NAME: EMANUEL

DATE/TIME OF TEST: 11/10/2021 16:16:35

---

Diagnostics Check	OK	16:13	
Air Blank	0.000	16:13	
Control Test	0.081	16:14	
Air Blank	0.000	16:14	
Subject Sample # 1	0.188	16:15	
Breath Volume	1.640	16:15	
Air Blank		16:16	Purge Fail
Air Blank	0.000	16:16	

---

Data Download Date/Time: 12/14/2021 11:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHEFFIELD, BRANDON, S  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: HEDGES, MATTHEW  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 11/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:53

SUBJECT'S LAST NAME: EMANUEL

DATE/TIME OF TEST: 11/10/2021 16:28:35

---

Diagnostics Check	OK	16:19
Air Blank	0.000	16:19
Control Test	0.080	16:19
Air Blank	0.000	16:20
Subject Sample # 1	0.176	16:20
Breath Volume	1.589	16:20
Air Blank	0.000	16:21
Air Blank	0.000	16:23
Subject Sample # 2	0.197	16:24
Breath Volume	2.269	16:24
Air Blank	0.000	16:24
Air Blank	0.000	16:26
Subject Sample # 3	0.194	16:27
Breath Volume	2.148	16:27
Air Blank	0.000	16:27
Control Test	0.082	16:27
Air Blank	0.000	16:28
Diagnostics Check	OK	16:28

---

Data Download Date/Time: 12/14/2021 11:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSFORD,C,B  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: HANSFORD,C  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 11/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: MILLION

DATE/TIME OF TEST: 11/14/2021 03:56:47

---

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.082	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.244	03:52
Breath Volume		2.035	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:54
Subject Sample	# 2	0.238	03:55
Breath Volume		1.554	03:55
Air Blank		0.000	03:55
Control Test		0.083	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

---

Data Download Date/Time: 12/14/2021 11:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSFORD,C,B  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: HINDS,L  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 11/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:29

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 11/19/2021 08:05:26

---

Diagnostics Check	OK	07:59
Air Blank	0.000	07:59
Control Test	0.082	07:59
Air Blank	0.000	08:00
Subject Sample # 1	0.153	08:00
Breath Volume	2.468	08:00
Air Blank	0.000	08:01
Air Blank	0.000	08:03
Subject Sample # 2	0.149	08:03
Breath Volume	1.847	08:03
Air Blank	0.000	08:04
Control Test	0.082	08:04
Air Blank	0.000	08:05
Diagnostics Check	OK	08:05

---

Data Download Date/Time: 12/14/2021 11:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REVELL,DENNIS,E  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: GRUEBEL,KARL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 11/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: TRAXLER

DATE/TIME OF TEST: 11/21/2021 03:32:18

---

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.082	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.320	03:27
Breath Volume	4.046	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.323	03:30
Breath Volume	4.136	03:30
Air Blank	0.000	03:31
Control Test	0.081	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

---

Data Download Date/Time: 12/14/2021 11:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHILLING,JOHN,R  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: CAUSSEAUX,DYLAN  
ARRESTING OFFICER AGENCY: Wakulla County SO  
LAST AGENCY INSPECTION: 10/11/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34520080A1  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 11:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:11

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 11/20/2021 12:43:54

---

Diagnostics Check	OK	12:40	
Air Blank	0.000	12:41	
Control Test	0.079	12:41	
Air Blank	0.000	12:42	
Subject Sample	# 1	12:42	Subject Test Refused
Breath Volume	0.000	12:42	
Air Blank	0.000	12:42	
Control Test	0.079	12:43	
Air Blank	0.000	12:43	
Diagnostics Check	OK	12:43	

---

Data Download Date/Time: 11/23/2021 10:11



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KAUFFMAN, MITCHELL, P  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: KAUFFMAN, MITCHELL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34520080A1  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 05:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: LOGSDON

DATE/TIME OF TEST: 11/26/2021 06:01:04

---

Diagnostics Check	OK	05:54
Air Blank	0.000	05:55
Control Test	0.080	05:55
Air Blank	0.000	05:56
Subject Sample # 1	0.172	05:56
Breath Volume	2.824	05:56
Air Blank	0.000	05:57
Air Blank	0.000	05:59
Subject Sample # 2	0.168	05:59
Breath Volume	2.589	05:59
Air Blank	0.000	06:00
Control Test	0.080	06:00
Air Blank	0.000	06:00
Diagnostics Check	OK	06:01

---

Data Download Date/Time: 12/22/2021 13:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: VOLLERTSEN,CRAIG  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34520080A1  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 07:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: EPPS

DATE/TIME OF TEST: 11/27/2021 08:37:20

---

Diagnostics Check	OK	08:30
Air Blank	0.000	08:31
Control Test	0.080	08:31
Air Blank	0.000	08:32
Subject Sample # 1	0.229	08:32
Breath Volume	2.808	08:32
Air Blank	0.000	08:33
Air Blank	0.000	08:35
Subject Sample # 2	0.238	08:35
Breath Volume	3.300	08:35
Air Blank	0.000	08:36
Control Test	0.082	08:36
Air Blank	0.000	08:37
Diagnostics Check	OK	08:37

---

Data Download Date/Time: 12/22/2021 13:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LECKINGER, MICHAEL, E  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: LECKINGER, MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34520080A1  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: RUDELL

DATE/TIME OF TEST: 11/27/2021 19:33:04

---

Diagnostics Check	OK	19:30	
Air Blank	0.000	19:30	
Control Test	0.081	19:31	
Air Blank	0.000	19:31	
Subject Sample	# 1	19:31	Subject Test Refused
Breath Volume	0.000	19:31	
Air Blank	0.000	19:32	
Control Test	0.080	19:32	
Air Blank	0.000	19:32	
Diagnostics Check	OK	19:33	

---

Data Download Date/Time: 12/22/2021 13:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LECKINGER, MICHAEL, E  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: LECKINGER, MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34520080A1  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: LAMBRIGHT

DATE/TIME OF TEST: 11/27/2021 22:01:45

---

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.081	21:55
Air Blank	0.000	21:56
Subject Sample # 1	0.192	21:57
Breath Volume	1.667	21:57
Air Blank	0.000	21:58
Air Blank	0.000	21:59
Subject Sample # 2	0.192	22:00
Breath Volume	1.210	22:00
Air Blank	0.000	22:00
Control Test	0.079	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

---

Data Download Date/Time: 12/22/2021 13:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RUSSO,ANDREW  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: COREANORIVERA

DATE/TIME OF TEST: 11/14/2021 02:38:08

---

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.080	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.208	02:33
Breath Volume		2.710	02:33
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.206	02:36
Breath Volume		1.839	02:36
Air Blank		0.000	02:37
Control Test		0.079	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

---

Data Download Date/Time: 12/07/2021 11:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON,TRAVIS,W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: SAUNDERS,HUNTER  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 11/14/2021 03:54:18

---

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.079	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.096	03:49
Breath Volume		2.460	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.095	03:52
Breath Volume		2.082	03:52
Air Blank		0.000	03:53
Control Test		0.078	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

---

Data Download Date/Time: 12/07/2021 11:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HUNT,FRANCIS  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: BAEZCAMACHO

DATE/TIME OF TEST: 11/14/2021 05:20:06

---

Diagnostics Check	OK	05:12
Air Blank	0.000	05:12
Control Test	0.080	05:13
Air Blank	0.000	05:13
Subject Sample # 1	0.171	05:15
Breath Volume	1.324	05:15
Air Blank	0.000	05:15
Air Blank	0.000	05:17
Subject Sample # 2	0.171	05:18
Breath Volume	1.445	05:18
Air Blank	0.000	05:19
Control Test	0.079	05:19
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

---

Data Download Date/Time: 12/07/2021 11:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HERNANDEZ,YASSEL  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: GIMENEZYUKUSHIMA

DATE/TIME OF TEST: 11/14/2021 06:45:01

---

Diagnostics Check		OK	06:29	
Air Blank		0.000	06:29	
Control Test		0.080	06:30	
Air Blank		0.000	06:30	
Subject Sample	# 1	0.078	06:33	Volume Not Met
Breath Volume		0.015	06:33	
Air Blank		0.000	06:34	
Air Blank		0.000	06:36	
Subject Sample	# 2	0.080	06:38	
Breath Volume		1.527	06:38	
Air Blank		0.000	06:39	
Air Blank		0.000	06:41	
Subject Sample	# 3	0.078	06:43	
Breath Volume		1.140	06:43	
Air Blank		0.000	06:44	
Control Test		0.078	06:44	
Air Blank		0.000	06:44	
Diagnostics Check		OK	06:45	

---

Data Download Date/Time: 12/07/2021 11:18



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KILLON,JONATHAN  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 11/16/2021 21:52:59

---

Diagnostics Check	OK	21:46
Air Blank	0.000	21:46
Control Test	0.079	21:47
Air Blank	0.000	21:47
Subject Sample # 1	0.102	21:48
Breath Volume	3.203	21:48
Air Blank	0.000	21:49
Air Blank	0.000	21:50
Subject Sample # 2	0.099	21:51
Breath Volume	3.699	21:51
Air Blank	0.000	21:52
Control Test	0.078	21:52
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

---

Data Download Date/Time: 12/07/2021 11:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: STAHL,BRANDON  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: SALGADOCASTILLO

DATE/TIME OF TEST: 11/17/2021 06:39:32

---

Diagnostics Check	OK	06:33
Air Blank	0.000	06:34
Control Test	0.080	06:34
Air Blank	0.000	06:34
Subject Sample # 1	0.102	06:35
Breath Volume	2.730	06:35
Air Blank	0.000	06:35
Air Blank	0.000	06:37
Subject Sample # 2	0.105	06:38
Breath Volume	1.203	06:38
Air Blank	0.000	06:38
Control Test	0.079	06:38
Air Blank	0.000	06:39
Diagnostics Check	OK	06:39

---

Data Download Date/Time: 12/07/2021 11:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MELLA,MELISSA  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 11/19/2021 23:46:57

---

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.079	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.232	23:42
Breath Volume	1.761	23:42
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.214	23:45
Breath Volume	1.152	23:45
Air Blank	0.000	23:46
Control Test	0.076	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

---

Data Download Date/Time: 12/07/2021 11:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: VINER,DENNIS  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: VELASQUEZCORONADO

DATE/TIME OF TEST: 11/20/2021 02:45:18

---

Diagnostics Check		OK	02:31	
Air Blank		0.000	02:32	
Control Test		0.079	02:32	
Air Blank		0.000	02:33	
Subject Sample	# 1	0.168	02:36	Volume Not Met
Breath Volume		0.597	02:36	
Air Blank		0.000	02:37	
Air Blank		0.000	02:38	
Subject Sample	# 2	0.164	02:40	
Breath Volume		1.757	02:40	
Air Blank		0.000	02:41	
Air Blank		0.000	02:43	
Subject Sample	# 3	0.161	02:43	
Breath Volume		1.253	02:43	
Air Blank		0.000	02:44	
Control Test		0.074	02:44	Control Outside Tolerance
Air Blank		0.000	02:45	

---

Data Download Date/Time: 12/07/2021 11:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: VINER,DENNIS  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: VELASQUEZCORONADO

DATE/TIME OF TEST: 11/20/2021 02:54:16

---

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.077	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.173	02:49
Breath Volume	1.351	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.174	02:52
Breath Volume	1.738	02:52
Air Blank	0.000	02:53
Control Test	0.077	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

---

Data Download Date/Time: 12/07/2021 11:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KISSANE, BRIAN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 11/20/2021 03:40:53

---

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.079	03:35
Air Blank	0.000	03:35
Subject Sample # 1	0.028	03:36
Breath Volume	1.585	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.029	03:39
Breath Volume	2.113	03:39
Air Blank	0.000	03:39
Control Test	0.078	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

---

Data Download Date/Time: 12/07/2021 11:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIS,COURTNEY,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HAMMOND,DANIEL  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 11/21/2021 03:55:52

---

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.080	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.169	03:51
Breath Volume	1.113	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.173	03:54
Breath Volume	2.019	03:54
Air Blank	0.000	03:54
Control Test	0.078	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

---

Data Download Date/Time: 12/07/2021 11:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BRULE,JASON  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 11/25/2021 00:39:16

---

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.170	00:29
Breath Volume	2.636	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.140	00:34
Breath Volume	1.273	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 3	0.149	00:37
Breath Volume	1.351	00:37
Air Blank	0.000	00:38
Control Test	0.077	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

---

Data Download Date/Time: 12/07/2021 11:19



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON,TRAVIS,W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HECTOR,JEFFREY  
ARRESTING OFFICER AGENCY: Central FL Univ PD  
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: RANDELL

DATE/TIME OF TEST: 11/25/2021 03:43:55

---

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.079	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.109	03:39
Breath Volume	1.757	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.113	03:42
Breath Volume	1.996	03:42
Air Blank	0.000	03:43
Control Test	0.079	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

---

Data Download Date/Time: 12/07/2021 11:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: TORRES,JOSUE  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: WHALEY

DATE/TIME OF TEST: 11/25/2021 04:42:45

---

Diagnostics Check		OK	04:36
Air Blank		0.000	04:37
Control Test		0.079	04:37
Air Blank		0.000	04:38
Subject Sample	# 1	0.066	04:38
Breath Volume		1.996	04:38
Air Blank		0.000	04:39
Air Blank		0.000	04:40
Subject Sample	# 2	0.064	04:41
Breath Volume		1.464	04:41
Air Blank		0.000	04:41
Control Test		0.079	04:42
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

---

Data Download Date/Time: 12/07/2021 11:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ZAMBALA,JOSE  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 11/25/2021 22:35:34

---

Diagnostics Check	OK	22:29
Air Blank	0.000	22:29
Control Test	0.078	22:30
Air Blank	0.000	22:30
Subject Sample # 1	0.125	22:31
Breath Volume	2.367	22:31
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.128	22:34
Breath Volume	2.453	22:34
Air Blank	0.000	22:34
Control Test	0.079	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

---

Data Download Date/Time: 12/07/2021 11:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KISSANE,BRIAN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 11/27/2021 03:40:35

---

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.080	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.209	03:36
Breath Volume		4.601	03:36
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.223	03:39
Breath Volume		3.726	03:39
Air Blank		0.000	03:39
Control Test		0.078	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

---

Data Download Date/Time: 12/07/2021 11:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: CERVELLERA,MATTHEW  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 11/27/2021 04:38:14

---

Diagnostics Check		OK	04:31
Air Blank		0.000	04:32
Control Test		0.080	04:32
Air Blank		0.000	04:32
Subject Sample	# 1	0.233	04:33
Breath Volume		1.429	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.228	04:36
Breath Volume		1.312	04:36
Air Blank		0.000	04:37
Control Test		0.079	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

---

Data Download Date/Time: 12/07/2021 11:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERVAIS,TABITHA,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: RIVERA,VICTOR  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: JIMENEZGUACARI

DATE/TIME OF TEST: 11/27/2021 06:05:33

---

Diagnostics Check		OK	05:59
Air Blank		0.000	05:59
Control Test		0.079	05:59
Air Blank		0.000	06:00
Subject Sample	# 1	0.144	06:01
Breath Volume		2.515	06:01
Air Blank		0.000	06:01
Air Blank		0.000	06:03
Subject Sample	# 2	0.142	06:04
Breath Volume		2.285	06:04
Air Blank		0.000	06:04
Control Test		0.079	06:04
Air Blank		0.000	06:05
Diagnostics Check		OK	06:05

---

Data Download Date/Time: 12/07/2021 11:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PADRON,MARCIE,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: LUCIANOMONTES,NICOLAS  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:37

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 11/29/2021 07:13:59

---

Diagnostics Check	OK	07:07
Air Blank	0.000	07:08
Control Test	0.080	07:08
Air Blank	0.000	07:08
Subject Sample # 1	0.226	07:09
Breath Volume	2.261	07:09
Air Blank	0.000	07:10
Air Blank	0.000	07:11
Subject Sample # 2	0.221	07:12
Breath Volume	1.558	07:12
Air Blank	0.000	07:13
Control Test	0.078	07:13
Air Blank	0.000	07:13
Diagnostics Check	OK	07:13

---

Data Download Date/Time: 12/07/2021 11:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PADRON,MARCIE,L  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MARRERO,EDWIN  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 07:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:54

SUBJECT'S LAST NAME: SALINASCEVEDA

DATE/TIME OF TEST: 11/01/2021 08:44:03

---

Diagnostics Check	OK	08:37
Air Blank	0.000	08:38
Control Test	0.079	08:38
Air Blank	0.000	08:38
Subject Sample # 1	0.170	08:39
Breath Volume	1.773	08:39
Air Blank	0.000	08:40
Air Blank	0.000	08:42
Subject Sample # 2	0.170	08:42
Breath Volume	1.136	08:42
Air Blank	0.000	08:43
Control Test	0.079	08:43
Air Blank	0.000	08:43
Diagnostics Check	OK	08:44

---

Data Download Date/Time: 11/03/2021 13:45



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FORAN,ETHAN,L  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: FORAN,ETHAN  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 11/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: RIOS VILLATORO

DATE/TIME OF TEST: 11/11/2021 22:34:20

---

Diagnostics Check	OK	22:28
Air Blank	0.000	22:28
Control Test	0.079	22:28
Air Blank	0.000	22:29
Subject Sample # 1	0.214	22:29
Breath Volume	1.335	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.213	22:32
Breath Volume	2.097	22:32
Air Blank	0.000	22:33
Control Test	0.079	22:33
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

---

Data Download Date/Time: 12/28/2021 14:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHOBREY,SPENCER,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: WHOBREY,SPENCER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: MANUEL

DATE/TIME OF TEST: 11/04/2021 00:35:21

---

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.080	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.000	00:30
Breath Volume	1.453	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.000	00:33
Breath Volume	1.125	00:33
Air Blank	0.000	00:34
Control Test	0.080	00:34
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

---

Data Download Date/Time: 11/24/2021 15:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILAN,VICTOR,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: MILAN,VICTOR  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 14:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: GUALBANCE

DATE/TIME OF TEST: 11/12/2021 18:24:38

---

Diagnostics Check	OK	18:16
Air Blank	0.000	18:17
Control Test	0.079	18:17
Air Blank	0.000	18:17
Subject Sample # 1	0.166	18:19
Breath Volume	1.480	18:19
Air Blank	0.000	18:19
Air Blank	0.000	18:21
Subject Sample # 2	0.155	18:23
Breath Volume	1.597	18:23
Air Blank	0.000	18:23
Control Test	0.077	18:24
Air Blank	0.000	18:24
Diagnostics Check	OK	18:24

---

Data Download Date/Time: 11/24/2021 15:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHOBREY,SPENCER,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: ROSARIO,ORLANDO  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/13/2021 01:56:05

---

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.093	01:50
Breath Volume	2.230	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.094	01:54
Breath Volume	1.457	01:54
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

---

Data Download Date/Time: 11/24/2021 15:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAWFORD,DAVID,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: CRAWFORD,DAVID  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 10/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 11/19/2021 21:21:53

---

Diagnostics Check	OK	21:15
Air Blank	0.000	21:15
Control Test	0.079	21:16
Air Blank	0.000	21:16
Subject Sample # 1	0.209	21:17
Breath Volume	1.628	21:17
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample # 2	0.202	21:20
Breath Volume	2.769	21:20
Air Blank	0.000	21:20
Control Test	0.079	21:21
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

---

Data Download Date/Time: 11/24/2021 15:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHOBREY,SPENCER,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: WHOBREY,SPENCER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 11/28/2021 01:08:30

---

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:06	
Control Test	0.080	01:06	
Air Blank	0.000	01:06	
Subject Sample	# 1	01:07	Subject Test Refused
Breath Volume	0.000	01:07	
Air Blank	0.000	01:07	
Control Test	0.080	01:07	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

---

Data Download Date/Time: 12/27/2021 16:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHOBREY,SPENCER,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: ROSARIO,ORLANDO  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 11/28/2021 03:49:42

---

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.080	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.191	03:45
Breath Volume	2.035	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.188	03:48
Breath Volume	1.453	03:48
Air Blank	0.000	03:48
Control Test	0.081	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

---

Data Download Date/Time: 12/27/2021 16:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: INGERSOLL,NICHOLAS,F  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: INGERSOLL,NICHOLAS  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A2  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 11/07/2021 21:09:45

---

Diagnostics Check	OK	20:58	
Air Blank	0.000	20:58	
Control Test	0.081	20:59	
Air Blank	0.000	20:59	
Subject Sample	# 1	21:02	No Sample Provided
Breath Volume	0.000	21:02	
Air Blank	0.000	21:03	
Air Blank	0.000	21:05	
Subject Sample	# 2	21:08	No Sample Provided
Breath Volume	0.000	21:08	
Air Blank	0.000	21:08	
Control Test	0.081	21:09	
Air Blank	0.000	21:09	
Diagnostics Check	OK	21:09	

---

Data Download Date/Time: 11/27/2021 16:03



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIALOBRZESKI,M,  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: BATCH,J  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A2  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: BAUTISTA

DATE/TIME OF TEST: 11/17/2021 04:20:17

---

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.080	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.097	04:15
Breath Volume	1.343	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.102	04:18
Breath Volume	2.832	04:18
Air Blank	0.000	04:19
Control Test	0.080	04:19
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

---

Data Download Date/Time: 11/27/2021 16:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: INGERSOLL,NICHOLAS,F  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: LEACH,BRANDON  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A2  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: CARRASQUILLO SUAREZ

DATE/TIME OF TEST: 11/20/2021 04:10:44

---

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.173	04:05
Breath Volume	1.429	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:07
Subject Sample # 2	0.164	04:09
Breath Volume	1.894	04:09
Air Blank	0.000	04:09
Control Test	0.079	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

---

Data Download Date/Time: 11/27/2021 16:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETTS, RONALD, E  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: TEAT, MATTHEW  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 11/26/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A2  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 14:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: MUNCH

DATE/TIME OF TEST: 11/27/2021 15:04:25

---

Diagnostics Check		OK	14:52	
Air Blank		0.000	14:52	
Control Test		0.081	14:52	
Air Blank		0.000	14:53	
Subject Sample	# 1	0.167	14:56	Volume Not Met
Breath Volume		0.464	14:56	
Air Blank		0.000	14:57	
Air Blank		0.000	14:58	
Subject Sample	# 2	0.203	14:59	
Breath Volume		1.234	14:59	
Air Blank		0.000	15:00	
Air Blank		0.000	15:01	
Subject Sample	# 3	0.195	15:02	
Breath Volume		1.464	15:02	
Air Blank		0.000	15:03	
Control Test		0.081	15:03	
Air Blank		0.000	15:04	
Diagnostics Check		OK	15:04	

---

Data Download Date/Time: 11/27/2021 16:03