

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AFONSO,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: AFONSO,MICHAEL  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 11/11/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 12/04/2021 23:11:47

---

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.082	23:05
Air Blank	0.000	23:05
Subject Sample # 1	0.067	23:06
Breath Volume	3.500	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.068	23:10
Breath Volume	3.738	23:10
Air Blank	0.000	23:10
Control Test	0.083	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

---

Data Download Date/Time: 12/15/2021 15:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLON, RAYMOND, M  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: EARL, CLIFFORD  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 11/11/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 14:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:43

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/06/2021 15:13:18

---

Diagnostics Check		OK	15:04
Air Blank		0.000	15:04
Control Test		0.082	15:04
Air Blank		0.000	15:05
Subject Sample	# 1	0.252	15:08
Breath Volume		2.179	15:08
Air Blank		0.000	15:08
Air Blank		0.000	15:10
Subject Sample	# 2	0.259	15:11
Breath Volume		2.179	15:11
Air Blank		0.000	15:12
Control Test		0.081	15:12
Air Blank		0.000	15:13
Diagnostics Check		OK	15:13

---

Data Download Date/Time: 12/15/2021 15:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AFONSO, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: AFONSO, MICHAEL  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 11/11/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 12/08/2021 19:47:30

---

Diagnostics Check	OK	19:44	
Air Blank	0.000	19:45	
Control Test	0.082	19:45	
Air Blank	0.000	19:45	
Subject Sample	# 1	19:46	Subject Test Refused
Breath Volume	0.000	19:46	
Air Blank	0.000	19:46	
Control Test	0.083	19:46	
Air Blank	0.000	19:47	
Diagnostics Check	OK	19:47	

---

Data Download Date/Time: 12/15/2021 15:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HULL,GARY,B  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: CORONILLA,DANIEL  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 11/11/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CONTRERAS RODRIGUEZ

DATE/TIME OF TEST: 12/12/2021 03:43:25

---

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.083	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.146	03:38
Breath Volume	3.347	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.141	03:41
Breath Volume	3.367	03:41
Air Blank	0.000	03:42
Control Test	0.078	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

---

Data Download Date/Time: 12/15/2021 15:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HULL,GARY,B  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: CORONILLA,DANIEL  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 11/11/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: ATKINSON

DATE/TIME OF TEST: 12/13/2021 02:06:17

---

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:03	
Control Test	0.083	02:04	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:04	Subject Test Refused
Breath Volume	0.000	02:04	
Air Blank	0.000	02:05	
Control Test	0.083	02:05	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:06	

---

Data Download Date/Time: 12/15/2021 15:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HULL,GARY,B  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: CORONILLA,DANIEL  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/16/2021 02:09:09

---

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:06	
Control Test	0.083	02:07	
Air Blank	0.000	02:07	
Subject Sample	# 1	02:07	Subject Test Refused
Breath Volume	0.000	02:07	
Air Blank	0.000	02:08	
Control Test	0.082	02:08	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

---

Data Download Date/Time: 02/02/2022 10:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HULL,GARY,B  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: CORONILLA,DANIEL  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 12/17/2021 02:00:07

---

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:57	
Control Test	0.082	01:58	
Air Blank	0.000	01:58	
Subject Sample	# 1	01:58	Subject Test Refused
Breath Volume	0.000	01:58	
Air Blank	0.000	01:59	
Control Test	0.083	01:59	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

---

Data Download Date/Time: 02/02/2022 10:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AFONSO, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: ELDRIDGE, ANDREW  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: IXMAY VICENTE

DATE/TIME OF TEST: 12/17/2021 22:37:40

---

Diagnostics Check	OK	22:33	
Air Blank	0.000	22:34	
Control Test	0.082	22:34	
Air Blank	0.000	22:35	
Subject Sample	# 1	22:36	Subject Test Refused
Breath Volume	0.000	22:36	
Air Blank	0.000	22:36	
Control Test	0.083	22:37	
Air Blank	0.000	22:37	
Diagnostics Check	OK	22:37	

---

Data Download Date/Time: 02/02/2022 10:15



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HULL,GARY,B  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: DREW,LIAM  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 11:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 12/21/2021 01:13:20

---

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.083	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.184	01:08
Breath Volume	1.945	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.182	01:11
Breath Volume	2.804	01:11
Air Blank	0.000	01:12
Control Test	0.080	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

---

Data Download Date/Time: 02/02/2022 10:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HULL,GARY,B  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: CHAMPAGNE,ALEXANDER  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: MORALES LOPEZ

DATE/TIME OF TEST: 12/21/2021 20:51:23

---

Diagnostics Check	OK	20:45
Air Blank	0.000	20:45
Control Test	0.082	20:45
Air Blank	0.000	20:46
Subject Sample # 1	0.167	20:46
Breath Volume	1.132	20:46
Air Blank	0.000	20:47
Air Blank	0.000	20:49
Subject Sample # 2	0.174	20:49
Breath Volume	1.101	20:49
Air Blank	0.000	20:50
Control Test	0.081	20:50
Air Blank	0.000	20:51
Diagnostics Check	OK	20:51

---

Data Download Date/Time: 02/02/2022 10:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HULL,GARY,B  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: RAMIREZ-REYNA,MAYRA  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 12/25/2021 00:04:46

---

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:02	
Control Test	0.083	00:02	
Air Blank	0.000	00:03	
Subject Sample	# 1	00:03	RFI Detect
Breath Volume	0.000	00:03	
Air Blank	0.000	00:03	
Control Test	0.083	00:04	
Air Blank	0.000	00:04	
Diagnostics Check	OK	00:04	

---

Data Download Date/Time: 02/02/2022 10:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HULL,GARY,B  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: RAMIREZ-REYNA,MAYRA  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 12/25/2021 00:14:21

---

Diagnostics Check		OK	00:06
Air Blank		0.000	00:06
Control Test		0.082	00:06
Air Blank		0.000	00:07
Subject Sample	# 1	0.248	00:09
Breath Volume		1.750	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:11
Subject Sample	# 2	0.248	00:12
Breath Volume		1.519	00:12
Air Blank		0.000	00:13
Control Test		0.078	00:13
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

---

Data Download Date/Time: 02/02/2022 10:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HULL,GARY,B  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: GEORGE,SHAQWAN  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 17:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: SALVADOR CRUZ PALMA

DATE/TIME OF TEST: 12/25/2021 19:06:31

---

Diagnostics Check		OK	19:00
Air Blank		0.000	19:00
Control Test		0.082	19:00
Air Blank		0.000	19:01
Subject Sample	# 1	0.223	19:02
Breath Volume		2.027	19:02
Air Blank		0.000	19:02
Air Blank		0.000	19:04
Subject Sample	# 2	0.213	19:04
Breath Volume		1.226	19:04
Air Blank		0.000	19:05
Control Test		0.079	19:05
Air Blank		0.000	19:06
Diagnostics Check		OK	19:06

---

Data Download Date/Time: 02/02/2022 10:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HULL,GARY,B  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: RAMIREZ-REYNA,MAYRA  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: SHEPARD

DATE/TIME OF TEST: 12/26/2021 01:25:48

---

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:23	
Control Test	0.083	01:23	
Air Blank	0.000	01:24	
Subject Sample	# 1	01:24	Subject Test Refused
Breath Volume	0.000	01:24	
Air Blank	0.000	01:24	
Control Test	0.082	01:25	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

---

Data Download Date/Time: 02/02/2022 10:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HULL,GARY,B  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: RAMIREZ-REYNA,MAYRA  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: SHEPARD

DATE/TIME OF TEST: 12/26/2021 01:31:50

---

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:29	
Control Test	0.082	01:29	
Air Blank	0.000	01:30	
Subject Sample	# 1	01:30	Subject Test Refused
Breath Volume	0.000	01:30	
Air Blank	0.000	01:30	
Control Test	0.083	01:31	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:31	

---

Data Download Date/Time: 02/02/2022 10:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HULL,GARY,B  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: CORONILLA,DANIEL  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1474747  
DRY GAS EXPIRY DATE: 04/12/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 12/26/2021 02:23:58

---

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.083	02:21	
Air Blank	0.000	02:22	
Subject Sample	# 1	02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:23	
Control Test	0.082	02:23	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

---

Data Download Date/Time: 02/02/2022 10:15



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLARIN,MISTY,D  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILRY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: TATE

DATE/TIME OF TEST: 12/18/2021 03:31:26

---

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:23	
Control Test	0.080	03:23	
Air Blank	0.000	03:24	
Subject Sample	# 1	03:27	No Sample Provided
Breath Volume	0.000	03:27	
Air Blank	0.000	03:27	
Air Blank	0.000	03:29	
Subject Sample	# 2	03:30	Subject Test Refused
Breath Volume	0.000	03:30	
Air Blank	0.000	03:30	
Control Test	0.080	03:30	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

---

Data Download Date/Time: 12/27/2021 10:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MCGILLICUDDY,STEVEN  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02021080A1  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: NUNN

DATE/TIME OF TEST: 12/02/2021 00:03:24

---

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.080	23:58
Air Blank	0.000	23:58
Subject Sample # 2	0.094	23:58
Breath Volume	1.199	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.093	00:01
Breath Volume	1.613	00:01
Air Blank	0.000	00:02
Control Test	0.080	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

---

Data Download Date/Time: 12/03/2021 14:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CISSON,JACKIE  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02021080A1  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 12/02/2021 23:30:41

---

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.080	23:23
Air Blank	0.000	23:24
Subject Sample # 1	0.245	23:26
Breath Volume	1.679	23:26
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.240	23:29
Breath Volume	1.105	23:29
Air Blank	0.000	23:29
Control Test	0.078	23:30
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

---

Data Download Date/Time: 12/03/2021 14:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: LOPEZ,BRIAN  
ARRESTING OFFICER AGENCY: Delray Beach PD  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02021080A1  
DRY GAS EXPIRY DATE: 03/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: VONDELL

DATE/TIME OF TEST: 12/03/2021 02:36:50

---

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.229	02:31
Breath Volume	1.214	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.213	02:35
Breath Volume	2.359	02:35
Air Blank	0.000	02:35
Control Test	0.078	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

---

Data Download Date/Time: 12/03/2021 14:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: HOWARD,EDWARD  
ARRESTING OFFICER AGENCY: West Palm Beach PD  
LAST AGENCY INSPECTION: 11/05/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GREY

DATE/TIME OF TEST: 12/01/2021 01:57:33

---

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.080	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.036	01:51
Breath Volume		1.449	01:51
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.036	01:56
Breath Volume		1.617	01:56
Air Blank		0.000	01:56
Control Test		0.080	01:56
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

---

Data Download Date/Time: 12/03/2021 15:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MCGILLICUDDY,STEVEN  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 11/05/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 12/01/2021 02:49:44

---

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.142	02:44
Breath Volume	1.218	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.138	02:48
Breath Volume	1.679	02:48
Air Blank	0.000	02:48
Control Test	0.079	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

---

Data Download Date/Time: 12/03/2021 15:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: ROSAS,DESIREE  
ARRESTING OFFICER AGENCY: Boynton Beach PD  
LAST AGENCY INSPECTION: 11/05/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 12/02/2021 00:04:15

---

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.080	23:57
Air Blank	0.000	23:58
Subject Sample # 2	0.106	23:59
Breath Volume	1.511	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.101	00:02
Breath Volume	1.140	00:02
Air Blank	0.000	00:03
Control Test	0.079	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

---

Data Download Date/Time: 12/03/2021 15:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: BENNETT,JONATHAN  
ARRESTING OFFICER AGENCY: Riviera Beach PD  
LAST AGENCY INSPECTION: 11/05/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: MILES

DATE/TIME OF TEST: 12/02/2021 05:21:33

---

Diagnostics Check	OK	05:15
Air Blank	0.000	05:15
Control Test	0.080	05:16
Air Blank	0.000	05:16
Subject Sample # 1	0.153	05:17
Breath Volume	1.230	05:17
Air Blank	0.000	05:17
Air Blank	0.000	05:19
Subject Sample # 2	0.144	05:19
Breath Volume	1.199	05:19
Air Blank	0.000	05:20
Control Test	0.080	05:20
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

---

Data Download Date/Time: 12/03/2021 15:20



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SOLOWAY,ALAN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 11/05/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: RAMIREZ LUGO

DATE/TIME OF TEST: 12/03/2021 01:13:43

---

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.150	01:09
Breath Volume	1.804	01:09
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.149	01:12
Breath Volume	2.000	01:12
Air Blank	0.000	01:12
Control Test	0.079	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

---

Data Download Date/Time: 12/03/2021 15:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: BENNETT, JONATHAN  
ARRESTING OFFICER AGENCY: Riviera Beach PD  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: SHEER

DATE/TIME OF TEST: 12/10/2021 02:39:44

---

Diagnostics Check		OK	02:33
Air Blank		0.000	02:33
Control Test		0.080	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.000	02:35
Breath Volume		1.261	02:35
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.000	02:38
Breath Volume		1.929	02:38
Air Blank		0.000	02:38
Control Test		0.080	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

---

Data Download Date/Time: 01/13/2022 13:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: JENNINGS, DAVON  
ARRESTING OFFICER AGENCY: Boynton Beach PD  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: BLACKBURN

DATE/TIME OF TEST: 12/11/2021 01:14:13

---

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.080	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.120	01:09
Breath Volume	3.046	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:11
Subject Sample # 2	0.105	01:12
Breath Volume	1.542	01:12
Air Blank	0.000	01:13
Control Test	0.078	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

---

Data Download Date/Time: 01/13/2022 13:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: POINTU, PATRICK  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: RAMIREZ PEREZ

DATE/TIME OF TEST: 12/11/2021 02:17:41

---

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.080	02:08
Air Blank		0.000	02:08
Subject Sample	# 1	0.219	02:10
Breath Volume		2.363	02:10
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.197	02:13
Breath Volume		1.300	02:13
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 3	0.211	02:16
Breath Volume		1.328	02:16
Air Blank		0.000	02:16
Control Test		0.079	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

---

Data Download Date/Time: 01/13/2022 13:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CISSON, JACKIE  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: TAHAY TZOC

DATE/TIME OF TEST: 12/11/2021 05:03:47

---

Diagnostics Check	OK	04:56
Air Blank	0.000	04:57
Control Test	0.080	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.187	04:59
Breath Volume	2.558	04:59
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.186	05:02
Breath Volume	2.253	05:02
Air Blank	0.000	05:02
Control Test	0.079	05:03
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

---

Data Download Date/Time: 01/13/2022 13:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: AMADON, WILL  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: AGUSTIN RAMIREZ

DATE/TIME OF TEST: 12/11/2021 23:42:28

---

Diagnostics Check	OK	23:34
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.226	23:36
Breath Volume	2.214	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.218	23:40
Breath Volume	1.789	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

---

Data Download Date/Time: 01/13/2022 13:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WINDSOR,NICHOLAS  
ARRESTING OFFICER AGENCY: Delray Beach PD  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 12/12/2021 01:09:48

---

Diagnostics Check		OK	01:03
Air Blank		0.000	01:04
Control Test		0.080	01:04
Air Blank		0.000	01:04
Subject Sample	# 1	0.154	01:05
Breath Volume		1.726	01:05
Air Blank		0.000	01:06
Air Blank		0.000	01:07
Subject Sample	# 2	0.145	01:08
Breath Volume		1.691	01:08
Air Blank		0.000	01:08
Control Test		0.079	01:09
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

---

Data Download Date/Time: 01/13/2022 13:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: KYSOR, MITCHEL  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: GONZALEZ PEREZ

DATE/TIME OF TEST: 12/13/2021 01:26:38

---

Diagnostics Check	OK	01:18
Air Blank	0.000	01:19
Control Test	0.080	01:19
Air Blank	0.000	01:19
Subject Sample # 1	0.143	01:21
Breath Volume	2.191	01:21
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.140	01:25
Breath Volume	2.054	01:25
Air Blank	0.000	01:25
Control Test	0.079	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

---

Data Download Date/Time: 01/13/2022 13:24



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: ALEXANDRE,DOMINIC  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 16:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: LISHINSKY

DATE/TIME OF TEST: 12/13/2021 17:56:20

---

Diagnostics Check	OK	17:49
Air Blank	0.000	17:49
Control Test	0.080	17:50
Air Blank	0.000	17:50
Subject Sample # 1	0.125	17:51
Breath Volume	1.476	17:51
Air Blank	0.000	17:52
Air Blank	0.000	17:53
Subject Sample # 2	0.124	17:54
Breath Volume	1.828	17:54
Air Blank	0.000	17:55
Control Test	0.079	17:55
Air Blank	0.000	17:56
Diagnostics Check	OK	17:56

---

Data Download Date/Time: 01/13/2022 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FISHER,CHRISTOPHER  
ARRESTING OFFICER AGENCY: West Palm Beach PD  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: SERGEYEV

DATE/TIME OF TEST: 12/13/2021 19:27:26

---

Diagnostics Check	OK	19:20
Air Blank	0.000	19:21
Control Test	0.079	19:21
Air Blank	0.000	19:21
Subject Sample # 1	0.000	19:22
Breath Volume	2.316	19:22
Air Blank	0.000	19:23
Air Blank	0.000	19:25
Subject Sample # 2	0.000	19:26
Breath Volume	2.570	19:26
Air Blank	0.000	19:26
Control Test	0.079	19:26
Air Blank	0.000	19:27
Diagnostics Check	OK	19:27

---

Data Download Date/Time: 01/13/2022 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FLINK,ANDREW  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 12/13/2021 20:55:38

---

Diagnostics Check	OK	20:49
Air Blank	0.000	20:49
Control Test	0.079	20:50
Air Blank	0.000	20:50
Subject Sample # 1	0.000	20:51
Breath Volume	1.882	20:51
Air Blank	0.000	20:51
Air Blank	0.000	20:53
Subject Sample # 2	0.000	20:54
Breath Volume	1.644	20:54
Air Blank	0.000	20:54
Control Test	0.079	20:55
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

---

Data Download Date/Time: 01/13/2022 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CISSON,JACKIE  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: VELLONE JR.

DATE/TIME OF TEST: 12/13/2021 22:35:40

---

Diagnostics Check	OK	22:29
Air Blank	0.000	22:29
Control Test	0.080	22:30
Air Blank	0.000	22:30
Subject Sample # 1	0.057	22:31
Breath Volume	2.140	22:31
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.055	22:34
Breath Volume	2.386	22:34
Air Blank	0.000	22:34
Control Test	0.080	22:35
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

---

Data Download Date/Time: 01/13/2022 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FLINK, ANDREW  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ARABIAT

DATE/TIME OF TEST: 12/16/2021 01:47:52

---

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.068	01:43
Breath Volume	1.332	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.067	01:46
Breath Volume	1.277	01:46
Air Blank	0.000	01:46
Control Test	0.080	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

---

Data Download Date/Time: 01/13/2022 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: TEJEDA,ANTHONY  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: JOYCE

DATE/TIME OF TEST: 12/17/2021 00:36:09

---

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.080	00:30
Air Blank	0.000	00:30
Subject Sample # 1	0.189	00:31
Breath Volume	1.539	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.190	00:34
Breath Volume	1.589	00:34
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

---

Data Download Date/Time: 01/13/2022 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: GONZALEZ,MARQUIS  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/17/2021 01:19:25

---

Diagnostics Check		OK	01:06	
Air Blank		0.000	01:07	
Control Test		0.080	01:07	
Air Blank		0.000	01:08	
Subject Sample	# 1	0.210	01:11	Slope Not Level
Breath Volume		1.968	01:11	
Air Blank		0.000	01:11	
Air Blank		0.000	01:13	
Subject Sample	# 2	0.213	01:14	
Breath Volume		2.515	01:14	
Air Blank		0.000	01:15	
Air Blank		0.000	01:17	
Subject Sample	# 3	0.209	01:17	
Breath Volume		1.921	01:17	
Air Blank		0.000	01:18	
Control Test		0.078	01:18	
Air Blank		0.000	01:19	
Diagnostics Check		OK	01:19	

---

Data Download Date/Time: 01/13/2022 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WHITE, EDWARD  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/17/2021 03:39:51

---

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.079	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.135	03:34
Breath Volume	1.402	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:36
Subject Sample # 2	0.135	03:38
Breath Volume	1.785	03:38
Air Blank	0.000	03:38
Control Test	0.079	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

---

Data Download Date/Time: 01/13/2022 13:24



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SOLOWAY, ALAN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: DESSALINES

DATE/TIME OF TEST: 12/17/2021 23:14:52

---

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.080	23:08
Air Blank	0.000	23:09
Subject Sample # 1	0.000	23:10
Breath Volume	2.921	23:10
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.000	23:13
Breath Volume	2.179	23:13
Air Blank	0.000	23:13
Control Test	0.079	23:14
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

---

Data Download Date/Time: 01/13/2022 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WHITE, EDWARD  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: ARGUETA RAMIREZ

DATE/TIME OF TEST: 12/18/2021 01:43:33

---

Diagnostics Check		OK	01:28	
Air Blank		0.000	01:29	
Control Test		0.080	01:29	
Air Blank		0.000	01:30	
Subject Sample	# 1	0.208	01:33	Slope Not Level
Breath Volume		1.527	01:33	
Air Blank		0.000	01:34	
Air Blank		0.000	01:35	
Subject Sample	# 2	0.193	01:37	
Breath Volume		1.453	01:37	
Air Blank		0.000	01:38	
Air Blank		0.000	01:40	
Subject Sample	# 3	0.200	01:41	
Breath Volume		2.191	01:41	
Air Blank		0.000	01:42	
Control Test		0.077	01:42	
Air Blank		0.000	01:43	
Diagnostics Check		OK	01:43	

---

Data Download Date/Time: 01/13/2022 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MULHAM, DONALD  
ARRESTING OFFICER AGENCY: Palm Beach PD  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 16:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: NIEVES

DATE/TIME OF TEST: 12/18/2021 18:22:24

---

Diagnostics Check	OK	18:13	
Air Blank	0.000	18:14	
Control Test	0.080	18:14	
Air Blank	0.000	18:15	
Subject Sample	# 1	18:18	No Sample Provided
Breath Volume	0.000	18:18	
Air Blank	0.000	18:18	
Air Blank	0.000	18:20	
Subject Sample	# 2	18:21	Subject Test Refused
Breath Volume	0.000	18:21	
Air Blank	0.000	18:21	
Control Test	0.080	18:21	
Air Blank	0.000	18:22	
Diagnostics Check	OK	18:22	

---

Data Download Date/Time: 01/13/2022 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MEDINA, CHRISTIAN  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: STENZEL

DATE/TIME OF TEST: 12/18/2021 22:27:53

---

Diagnostics Check	OK	22:21
Air Blank	0.000	22:22
Control Test	0.080	22:22
Air Blank	0.000	22:22
Subject Sample # 1	0.177	22:23
Breath Volume	1.457	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:25
Subject Sample # 2	0.181	22:26
Breath Volume	2.132	22:26
Air Blank	0.000	22:26
Control Test	0.080	22:27
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

---

Data Download Date/Time: 01/13/2022 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WINDSOR, NICHOLAS  
ARRESTING OFFICER AGENCY: Delray Beach PD  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 12/19/2021 00:01:43

---

Diagnostics Check	OK	23:54
Air Blank	0.000	23:55
Control Test	0.080	23:55
Air Blank	0.000	23:56
Subject Sample # 2	0.000	23:57
Breath Volume	1.238	23:57
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 1	0.000	00:00
Breath Volume	1.226	00:00
Air Blank	0.000	00:00
Control Test	0.080	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

---

Data Download Date/Time: 01/13/2022 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CLAUDIO,JOSE  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 07:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: CALLAHAN

DATE/TIME OF TEST: 12/19/2021 10:16:10

---

Diagnostics Check	OK	10:08
Air Blank	0.000	10:09
Control Test	0.079	10:09
Air Blank	0.000	10:10
Subject Sample # 1	0.145	10:11
Breath Volume	1.210	10:11
Air Blank	0.000	10:12
Air Blank	0.000	10:14
Subject Sample # 2	0.160	10:14
Breath Volume	2.007	10:14
Air Blank	0.000	10:15
Control Test	0.078	10:15
Air Blank	0.000	10:16
Diagnostics Check	OK	10:16

---

Data Download Date/Time: 01/13/2022 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: POINTU,PATRICK  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: FIGUEROA JR

DATE/TIME OF TEST: 12/20/2021 20:51:15

---

Diagnostics Check	OK	20:44
Air Blank	0.000	20:44
Control Test	0.079	20:45
Air Blank	0.000	20:45
Subject Sample # 1	0.284	20:46
Breath Volume	2.292	20:46
Air Blank	0.000	20:47
Air Blank	0.000	20:49
Subject Sample # 2	0.282	20:49
Breath Volume	1.523	20:49
Air Blank	0.000	20:50
Control Test	0.078	20:50
Air Blank	0.000	20:51
Diagnostics Check	OK	20:51

---

Data Download Date/Time: 01/13/2022 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CISSON,JACKIE  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MURILLO

DATE/TIME OF TEST: 12/21/2021 02:38:17

---

Diagnostics Check		OK	02:26	
Air Blank		0.000	02:27	
Control Test		0.079	02:27	
Air Blank		0.000	02:27	
Subject Sample	# 1	0.094	02:31	Volume Not Met
Breath Volume		0.390	02:31	
Air Blank		0.000	02:31	
Air Blank		0.000	02:33	
Subject Sample	# 2	0.095	02:36	Volume Not Met
Breath Volume		0.949	02:36	
Air Blank		0.000	02:37	
Control Test		0.078	02:37	
Air Blank		0.000	02:38	
Diagnostics Check		OK	02:38	

---

Data Download Date/Time: 01/13/2022 13:24



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: TEJEDA,ANTHONY  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: LENGYEL

DATE/TIME OF TEST: 12/21/2021 22:41:40

---

Diagnostics Check	OK	22:34
Air Blank	0.000	22:35
Control Test	0.080	22:35
Air Blank	0.000	22:36
Subject Sample # 1	0.202	22:37
Breath Volume	4.082	22:37
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.203	22:40
Breath Volume	3.902	22:40
Air Blank	0.000	22:40
Control Test	0.079	22:41
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

---

Data Download Date/Time: 01/13/2022 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: KYSOR, MICHELL  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: ALPIZAR FLORES

DATE/TIME OF TEST: 12/23/2021 01:03:38

---

Diagnostics Check	OK	00:56
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:57
Subject Sample # 1	0.132	00:58
Breath Volume	1.175	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:00
Subject Sample # 2	0.140	01:02
Breath Volume	1.695	01:02
Air Blank	0.000	01:02
Control Test	0.080	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

---

Data Download Date/Time: 01/13/2022 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MARCH, THOMAS  
ARRESTING OFFICER AGENCY: Palm Beach PD  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 12/23/2021 01:46:00

---

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.080	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.000	01:41
Breath Volume	2.488	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.000	01:44
Breath Volume	2.550	01:44
Air Blank	0.000	01:45
Control Test	0.080	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:46

---

Data Download Date/Time: 01/13/2022 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WHITE, EDWARD  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: GASPAR JACINTO

DATE/TIME OF TEST: 12/23/2021 23:17:22

---

Diagnostics Check	OK	23:10
Air Blank	0.000	23:11
Control Test	0.080	23:11
Air Blank	0.000	23:11
Subject Sample # 1	0.238	23:12
Breath Volume	3.632	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.238	23:15
Breath Volume	3.648	23:15
Air Blank	0.000	23:16
Control Test	0.080	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

---

Data Download Date/Time: 01/13/2022 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: POINTU, PATRICK  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: DAKOTA

DATE/TIME OF TEST: 12/24/2021 20:08:39

---

Diagnostics Check	OK	19:57	
Air Blank	0.000	19:58	
Control Test	0.081	19:58	
Air Blank	0.000	19:59	
Subject Sample	# 1 0.116	20:02	Volume Not Met
Breath Volume	0.015	20:02	
Air Blank	0.000	20:02	
Air Blank	0.000	20:04	
Subject Sample	# 2	20:07	Subject Test Refused
Breath Volume	0.035	20:07	
Air Blank	0.000	20:07	
Control Test	0.079	20:08	
Air Blank	0.000	20:08	
Diagnostics Check	OK	20:08	

---

Data Download Date/Time: 01/13/2022 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FLINK, ANDREW  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 12/24/2021 22:15:33

---

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.080	22:10
Air Blank	0.000	22:10
Subject Sample # 1	0.031	22:11
Breath Volume	3.937	22:11
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.031	22:14
Breath Volume	4.710	22:14
Air Blank	0.000	22:14
Control Test	0.080	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

---

Data Download Date/Time: 01/13/2022 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WHITE, EDWARD  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/25/2021 00:49:50

---

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.080	00:44
Air Blank	0.000	00:44
Subject Sample # 1	0.101	00:44
Breath Volume	2.644	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.102	00:48
Breath Volume	2.601	00:48
Air Blank	0.000	00:48
Control Test	0.080	00:49
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

---

Data Download Date/Time: 01/13/2022 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SOLOWAY, ALAN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: GOMEZ SANCHEZ

DATE/TIME OF TEST: 12/25/2021 03:52:14

---

Diagnostics Check	OK	03:42
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.151	03:44
Breath Volume	2.433	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.119	03:47
Breath Volume	1.535	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 3	0.141	03:50
Breath Volume	3.457	03:50
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

---

Data Download Date/Time: 01/13/2022 13:24



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: GONZALEZ, ROBERTO  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: CONVISER

DATE/TIME OF TEST: 12/25/2021 04:15:00

---

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.117	04:10
Breath Volume	1.937	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:12
Subject Sample # 2	0.113	04:13
Breath Volume	1.843	04:13
Air Blank	0.000	04:14
Control Test	0.079	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:15

---

Data Download Date/Time: 01/13/2022 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: POINTU, PATRICK  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: REMY

DATE/TIME OF TEST: 12/26/2021 00:13:08

---

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.081	00:04
Air Blank	0.000	00:05
Subject Sample # 1	0.116	00:07
Breath Volume	1.234	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.124	00:11
Breath Volume	1.644	00:11
Air Blank	0.000	00:12
Control Test	0.078	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

---

Data Download Date/Time: 01/13/2022 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: BARRERAS, ANTHONY  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: LAZARO RECENDIZ

DATE/TIME OF TEST: 12/26/2021 00:41:46

---

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.079	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.104	00:37
Breath Volume	2.082	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.106	00:40
Breath Volume	2.671	00:40
Air Blank	0.000	00:40
Control Test	0.079	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

---

Data Download Date/Time: 01/13/2022 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: POINTU,PATRICK  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: MORALES BERDUO

DATE/TIME OF TEST: 12/26/2021 21:10:43

---

Diagnostics Check	OK	21:03
Air Blank	0.000	21:03
Control Test	0.080	21:03
Air Blank	0.000	21:04
Subject Sample # 1	0.202	21:06
Breath Volume	2.523	21:06
Air Blank	0.000	21:06
Air Blank	0.000	21:08
Subject Sample # 2	0.203	21:09
Breath Volume	1.824	21:09
Air Blank	0.000	21:09
Control Test	0.080	21:10
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

---

Data Download Date/Time: 01/13/2022 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: ODEA,PATRICK  
ARRESTING OFFICER AGENCY: Palm Beach PD  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 12/27/2021 04:56:22

---

Diagnostics Check	OK	04:49
Air Blank	0.000	04:50
Control Test	0.080	04:50
Air Blank	0.000	04:50
Subject Sample # 1	0.148	04:52
Breath Volume	2.871	04:52
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.148	04:54
Breath Volume	1.636	04:54
Air Blank	0.000	04:55
Control Test	0.079	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

---

Data Download Date/Time: 01/13/2022 13:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WALKER,JODI  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: POL TOL

DATE/TIME OF TEST: 12/28/2021 23:51:13

---

Diagnostics Check	OK	23:44
Air Blank	0.000	23:45
Control Test	0.080	23:45
Air Blank	0.000	23:45
Subject Sample # 1	0.262	23:46
Breath Volume	2.054	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.269	23:49
Breath Volume	1.843	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

---

Data Download Date/Time: 01/13/2022 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: THOMAS, MICAH  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SANCHEZ SANCHEZ

DATE/TIME OF TEST: 12/30/2021 00:36:25

---

Diagnostics Check		OK	00:27
Air Blank		0.000	00:27
Control Test		0.080	00:28
Air Blank		0.000	00:28
Subject Sample	# 1	0.046	00:31
Breath Volume		1.277	00:31
Air Blank		0.000	00:32
Air Blank		0.000	00:33
Subject Sample	# 2	0.044	00:34
Breath Volume		1.351	00:34
Air Blank		0.000	00:35
Control Test		0.079	00:35
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

---

Data Download Date/Time: 01/13/2022 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: BENNETT, JONATHAN  
ARRESTING OFFICER AGENCY: Riviera Beach PD  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: HULCOLM

DATE/TIME OF TEST: 12/30/2021 03:21:11

---

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.080	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.000	03:16
Breath Volume	1.914	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.000	03:19
Breath Volume	1.875	03:19
Air Blank	0.000	03:20
Control Test	0.080	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

---

Data Download Date/Time: 01/13/2022 13:24



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FLINK, ANDREW  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: SKAAR

DATE/TIME OF TEST: 12/30/2021 23:06:46

---

Diagnostics Check	OK	23:00
Air Blank	0.000	23:00
Control Test	0.079	23:01
Air Blank	0.000	23:01
Subject Sample # 1	0.167	23:02
Breath Volume	3.187	23:02
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.163	23:05
Breath Volume	2.863	23:05
Air Blank	0.000	23:05
Control Test	0.080	23:06
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

---

Data Download Date/Time: 01/13/2022 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: KYSOR, MITCHELL  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: CALIHUA SANDOVAL

DATE/TIME OF TEST: 12/31/2021 00:51:01

---

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.080	00:42
Air Blank		0.000	00:43
Subject Sample	# 1	0.222	00:45
Breath Volume		2.300	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.221	00:49
Breath Volume		2.566	00:49
Air Blank		0.000	00:50
Control Test		0.079	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:51

---

Data Download Date/Time: 01/13/2022 13:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: POINTU, PATRICK  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 12/31/2021 02:16:00

---

Diagnostics Check		OK	02:09
Air Blank		0.000	02:09
Control Test		0.080	02:10
Air Blank		0.000	02:10
Subject Sample	# 1	0.190	02:11
Breath Volume		2.292	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:13
Subject Sample	# 2	0.203	02:14
Breath Volume		2.152	02:14
Air Blank		0.000	02:15
Control Test		0.079	02:15
Air Blank		0.000	02:15
Diagnostics Check		OK	02:16

---

Data Download Date/Time: 01/13/2022 13:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WHITE, EDWARD  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: CAPETILLO

DATE/TIME OF TEST: 12/31/2021 04:21:52

---

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:16
Subject Sample # 1	0.132	04:17
Breath Volume	1.742	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:19
Subject Sample # 2	0.130	04:20
Breath Volume	1.867	04:20
Air Blank	0.000	04:20
Control Test	0.080	04:21
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

---

Data Download Date/Time: 01/13/2022 13:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

INSTRUMENT REGISTERED TO: FHP TROOP F  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEMEN,LOGAN,C  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEMEN,LOGAN  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1167545  
DRY GAS EXPIRY DATE: 07/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: HUMPHREY  
DATE/TIME OF TEST: 12/30/2021 00:47:24

---

Diagnostics Check		OK	00:36
Air Blank		0.000	00:36
Control Test		0.077	00:37
Air Blank		0.000	00:37
Subject Sample	# 1	0.180	00:38
Breath Volume		1.503	00:38
Air Blank		0.000	00:39
Air Blank		0.000	00:40
Subject Sample	# 2	0.211	00:42
Breath Volume		1.246	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:44
Subject Sample	# 3	0.197	00:45
Breath Volume		2.027	00:45
Air Blank		0.000	00:46
Control Test		0.078	00:46
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

---

Data Download Date/Time: 01/10/2022 09:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLS,SCOTT,T  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MILLS,SCOTT  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35019080A3  
DRY GAS EXPIRY DATE: 02/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 11:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:09

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 12/19/2021 11:44:48

---

Diagnostics Check	OK	11:41	
Air Blank	0.000	11:42	
Control Test	0.078	11:42	
Air Blank	0.000	11:43	
Subject Sample	# 1	11:43	RFI Detect
Breath Volume	0.000	11:43	
Air Blank	0.000	11:43	
Control Test	0.080	11:44	
Air Blank	0.000	11:44	
Diagnostics Check	OK	11:44	

---

Data Download Date/Time: 01/18/2022 14:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLS,SCOTT,T  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MILLS,SCOTT  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35019080A3  
DRY GAS EXPIRY DATE: 02/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 11:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:09

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 12/19/2021 11:53:37

---

Diagnostics Check	OK	11:46
Air Blank	0.000	11:46
Control Test	0.078	11:46
Air Blank	0.000	11:47
Subject Sample # 1	0.128	11:48
Breath Volume	1.832	11:48
Air Blank	0.000	11:48
Air Blank	0.000	11:50
Subject Sample # 2	0.115	11:52
Breath Volume	1.449	11:52
Air Blank	0.000	11:52
Control Test	0.077	11:52
Air Blank	0.000	11:53
Diagnostics Check	OK	11:53

---

Data Download Date/Time: 01/18/2022 14:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LEISCH,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 12/04/2021 05:30:31

---

Diagnostics Check		OK	05:22
Air Blank		0.000	05:23
Control Test		0.080	05:23
Air Blank		0.000	05:24
Subject Sample	# 1	0.221	05:25
Breath Volume		2.468	05:25
Air Blank		0.000	05:26
Air Blank		0.000	05:28
Subject Sample	# 2	0.215	05:28
Breath Volume		1.859	05:28
Air Blank		0.000	05:29
Control Test		0.078	05:29
Air Blank		0.000	05:30
Diagnostics Check		OK	05:30

---

Data Download Date/Time: 12/09/2021 14:42



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: HARWELL

DATE/TIME OF TEST: 12/16/2021 03:45:50

---

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.155	03:41
Breath Volume	1.390	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:43
Subject Sample # 2	0.149	03:44
Breath Volume	1.277	03:44
Air Blank	0.000	03:44
Control Test	0.079	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

---

Data Download Date/Time: 01/06/2022 13:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: SCIAME

DATE/TIME OF TEST: 12/16/2021 21:21:17

---

Diagnostics Check	OK	21:13
Air Blank	0.000	21:14
Control Test	0.079	21:14
Air Blank	0.000	21:15
Subject Sample # 1	0.436	21:16
Breath Volume	3.562	21:16
Air Blank	0.000	21:17
Air Blank	0.000	21:18
Subject Sample # 2	0.441	21:19
Breath Volume	3.882	21:19
Air Blank	0.000	21:20
Control Test	0.077	21:20
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

---

Data Download Date/Time: 01/06/2022 13:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONEY, ANIA, L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RENDON, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: VASQUEZ-GONZALEZ

DATE/TIME OF TEST: 12/16/2021 22:54:34

---

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.079	22:48
Air Blank	0.000	22:48
Subject Sample # 1	0.235	22:49
Breath Volume	2.265	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.242	22:52
Breath Volume	2.347	22:52
Air Blank	0.000	22:53
Control Test	0.077	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

---

Data Download Date/Time: 01/06/2022 13:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LEISCH,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: BERDUO GOMEZ

DATE/TIME OF TEST: 12/18/2021 00:30:00

---

Diagnostics Check		OK	00:22
Air Blank		0.000	00:23
Control Test		0.080	00:23
Air Blank		0.000	00:24
Subject Sample	# 1	0.216	00:25
Breath Volume		1.257	00:25
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.215	00:28
Breath Volume		2.378	00:28
Air Blank		0.000	00:29
Control Test		0.080	00:29
Air Blank		0.000	00:29
Diagnostics Check		OK	00:30

---

Data Download Date/Time: 01/06/2022 13:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 12/18/2021 00:55:43

---

Diagnostics Check		OK	00:48
Air Blank		0.000	00:49
Control Test		0.079	00:49
Air Blank		0.000	00:50
Subject Sample	# 1	0.234	00:50
Breath Volume		2.769	00:50
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.220	00:54
Breath Volume		2.785	00:54
Air Blank		0.000	00:54
Control Test		0.078	00:55
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

---

Data Download Date/Time: 01/06/2022 13:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONEY, ANIA, L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LEISCH, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 12/18/2021 03:47:05

---

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.080	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.104	03:41
Breath Volume		1.375	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.102	03:45
Breath Volume		1.531	03:45
Air Blank		0.000	03:46
Control Test		0.079	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

---

Data Download Date/Time: 01/06/2022 13:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: MORRONE

DATE/TIME OF TEST: 12/18/2021 04:11:50

---

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.079	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 1 0.127	04:00	
Breath Volume	1.300	04:00	
Air Blank	0.000	04:01	
Air Blank	0.000	04:03	
Subject Sample	# 2 0.112	04:06	Volume Not Met
Breath Volume	0.390	04:06	
Air Blank	0.000	04:07	
Air Blank	0.000	04:08	
Subject Sample	# 3 0.137	04:10	
Breath Volume	2.035	04:10	
Air Blank	0.000	04:10	
Control Test	0.077	04:11	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:11	

---

Data Download Date/Time: 01/06/2022 13:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TRIEGLAFF,M  
ARRESTING OFFICER AGENCY: Macdill Air Force Ba  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: FRIDAY

DATE/TIME OF TEST: 12/18/2021 21:16:57

---

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.079	21:10
Air Blank	0.000	21:11
Subject Sample # 1	0.312	21:12
Breath Volume	3.460	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:14
Subject Sample # 2	0.317	21:15
Breath Volume	2.226	21:15
Air Blank	0.000	21:16
Control Test	0.078	21:16
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

---

Data Download Date/Time: 01/06/2022 13:57



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREENLEE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GUERRERO,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: ALBERICO

DATE/TIME OF TEST: 12/18/2021 22:08:00

---

Diagnostics Check	OK	22:01
Air Blank	0.000	22:02
Control Test	0.079	22:02
Air Blank	0.000	22:02
Subject Sample # 1	0.217	22:03
Breath Volume	3.132	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:05
Subject Sample # 2	0.223	22:06
Breath Volume	3.058	22:06
Air Blank	0.000	22:07
Control Test	0.078	22:07
Air Blank	0.000	22:07
Diagnostics Check	OK	22:08

---

Data Download Date/Time: 01/06/2022 13:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LEISCH,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: DELGADO JR

DATE/TIME OF TEST: 12/19/2021 00:08:44

---

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:04	
Control Test	0.079	00:05	
Air Blank	0.000	00:05	
Subject Sample	# 1	00:08	Sequence Aborted
Breath Volume	0.000	00:08	
Air Blank	0.000	00:08	

---

Data Download Date/Time: 01/06/2022 13:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LEISCH,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: DELGADO JR

DATE/TIME OF TEST: 12/19/2021 00:38:44

---

Diagnostics Check	OK	00:31
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:33
Subject Sample # 1	0.152	00:34
Breath Volume	2.117	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.147	00:37
Breath Volume	2.292	00:37
Air Blank	0.000	00:37
Control Test	0.079	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

---

Data Download Date/Time: 01/06/2022 13:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WINTER,C  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: JOSIPOVIC

DATE/TIME OF TEST: 12/19/2021 01:42:46

---

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.079	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.123	01:38
Breath Volume	1.828	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.121	01:41
Breath Volume	2.351	01:41
Air Blank	0.000	01:41
Control Test	0.078	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

---

Data Download Date/Time: 01/06/2022 13:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURKE,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: GILLEY

DATE/TIME OF TEST: 12/19/2021 04:14:12

---

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.142	04:09
Breath Volume	1.156	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.142	04:12
Breath Volume	1.234	04:12
Air Blank	0.000	04:13
Control Test	0.077	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

---

Data Download Date/Time: 01/06/2022 13:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GARNER,W  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/19/2021 05:37:41

---

Diagnostics Check	OK	05:31
Air Blank	0.000	05:31
Control Test	0.079	05:32
Air Blank	0.000	05:32
Subject Sample # 1	0.105	05:33
Breath Volume	2.070	05:33
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.108	05:36
Breath Volume	2.320	05:36
Air Blank	0.000	05:36
Control Test	0.078	05:37
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

---

Data Download Date/Time: 01/06/2022 13:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KNOX,STEPHANIE,R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNYDER,A  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 14:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:22

SUBJECT'S LAST NAME: GOLEMBIEWSKI

DATE/TIME OF TEST: 12/20/2021 15:49:18

---

Diagnostics Check	OK	15:42
Air Blank	0.000	15:43
Control Test	0.079	15:43
Air Blank	0.000	15:43
Subject Sample # 1	0.198	15:44
Breath Volume	2.515	15:44
Air Blank	0.000	15:45
Air Blank	0.000	15:47
Subject Sample # 2	0.184	15:47
Breath Volume	1.781	15:47
Air Blank	0.000	15:48
Control Test	0.078	15:48
Air Blank	0.000	15:49
Diagnostics Check	OK	15:49

---

Data Download Date/Time: 01/06/2022 13:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LANDRY,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 12/22/2021 02:29:42

---

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.080	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.000	02:25
Breath Volume	1.464	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.000	02:28
Breath Volume	1.804	02:28
Air Blank	0.000	02:28
Control Test	0.080	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

---

Data Download Date/Time: 01/06/2022 13:58



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WASHINGTON,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:48

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/22/2021 07:15:28

---

Diagnostics Check		OK	07:08
Air Blank		0.000	07:09
Control Test		0.080	07:09
Air Blank		0.000	07:10
Subject Sample	# 1	0.000	07:10
Breath Volume		1.394	07:10
Air Blank		0.000	07:11
Air Blank		0.000	07:13
Subject Sample	# 2	0.000	07:14
Breath Volume		1.367	07:14
Air Blank		0.000	07:14
Control Test		0.080	07:14
Air Blank		0.000	07:15
Diagnostics Check		OK	07:15

---

Data Download Date/Time: 01/06/2022 13:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONEY, ANIA, L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GARNER, W  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: SANDOVAL JR.

DATE/TIME OF TEST: 12/23/2021 00:56:50

---

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:49	
Control Test	0.081	00:49	
Air Blank	0.000	00:50	
Subject Sample	# 1 0.092	00:53	Slope Not Level
Breath Volume	1.808	00:53	
Air Blank	0.000	00:54	
Air Blank	0.000	00:55	
Subject Sample	# 2	00:56	Sequence Aborted
Breath Volume	0.000	00:56	
Air Blank	0.000	00:56	

---

Data Download Date/Time: 01/06/2022 13:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONEY, ANIA, L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GARNER, W  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: SANDOVAL JR.

DATE/TIME OF TEST: 12/23/2021 01:02:28

---

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:59	
Control Test	0.080	00:59	
Air Blank	0.000	00:59	
Subject Sample	# 1	01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:01	
Control Test	0.080	01:01	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

---

Data Download Date/Time: 01/06/2022 13:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: YAYLACI

DATE/TIME OF TEST: 12/24/2021 03:10:23

---

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.081	03:04
Air Blank		0.000	03:05
Subject Sample	# 1	0.222	03:05
Breath Volume		1.828	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.229	03:08
Breath Volume		1.707	03:08
Air Blank		0.000	03:09
Control Test		0.080	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

---

Data Download Date/Time: 01/06/2022 13:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTER,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 12/24/2021 03:43:29

---

Diagnostics Check		OK	03:36
Air Blank		0.000	03:37
Control Test		0.081	03:37
Air Blank		0.000	03:37
Subject Sample	# 1	0.235	03:38
Breath Volume		1.792	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.236	03:41
Breath Volume		1.367	03:41
Air Blank		0.000	03:42
Control Test		0.080	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

---

Data Download Date/Time: 01/06/2022 13:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 12/24/2021 23:32:01

---

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.081	23:26
Air Blank	0.000	23:26
Subject Sample # 1	0.192	23:27
Breath Volume	2.101	23:27
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.186	23:30
Breath Volume	1.394	23:30
Air Blank	0.000	23:31
Control Test	0.079	23:31
Air Blank	0.000	23:31
Diagnostics Check	OK	23:32

---

Data Download Date/Time: 01/06/2022 13:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HALL,L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/25/2021 05:14:45

---

Diagnostics Check	OK	05:07
Air Blank	0.000	05:08
Control Test	0.081	05:08
Air Blank	0.000	05:09
Subject Sample # 1	0.167	05:09
Breath Volume	2.371	05:09
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample # 2	0.161	05:13
Breath Volume	1.832	05:13
Air Blank	0.000	05:13
Control Test	0.079	05:14
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

---

Data Download Date/Time: 01/06/2022 13:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 12/26/2021 00:50:54

---

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.081	00:44
Air Blank	0.000	00:44
Subject Sample # 1	0.296	00:45
Breath Volume	2.304	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.285	00:49
Breath Volume	1.656	00:49
Air Blank	0.000	00:49
Control Test	0.079	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

---

Data Download Date/Time: 01/06/2022 13:59



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONEY, ANIA, L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN, T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 12/26/2021 03:13:01

---

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.080	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.178	03:08
Breath Volume		1.855	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:10
Subject Sample	# 2	0.167	03:11
Breath Volume		1.558	03:11
Air Blank		0.000	03:12
Control Test		0.079	03:12
Air Blank		0.000	03:12
Diagnostics Check		OK	03:13

---

Data Download Date/Time: 01/06/2022 13:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ROBERTS,J  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: NAGY

DATE/TIME OF TEST: 12/27/2021 05:01:33

---

Diagnostics Check		OK	04:54
Air Blank		0.000	04:54
Control Test		0.079	04:55
Air Blank		0.000	04:55
Subject Sample	# 1	0.283	04:56
Breath Volume		2.246	04:56
Air Blank		0.000	04:57
Air Blank		0.000	04:59
Subject Sample	# 2	0.274	04:59
Breath Volume		1.753	04:59
Air Blank		0.000	05:00
Control Test		0.079	05:00
Air Blank		0.000	05:01
Diagnostics Check		OK	05:01

---

Data Download Date/Time: 01/06/2022 13:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RIVERA,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: REVILLA

DATE/TIME OF TEST: 12/28/2021 02:57:20

---

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.080	02:51
Air Blank		0.000	02:51
Subject Sample	# 1	0.160	02:52
Breath Volume		3.835	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.163	02:55
Breath Volume		2.988	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

---

Data Download Date/Time: 01/06/2022 13:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VELAZQUEZ,B  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: HERNANDEZ-GARCIA

DATE/TIME OF TEST: 12/28/2021 03:43:41

---

Diagnostics Check		OK	03:31	
Air Blank		0.000	03:32	
Control Test		0.079	03:32	
Air Blank		0.000	03:33	
Subject Sample	# 1	0.148	03:36	Volume Not Met
Breath Volume		1.003	03:36	
Air Blank		0.000	03:37	
Air Blank		0.000	03:38	
Subject Sample	# 2	0.117	03:41	Volume Not Met
Breath Volume		0.109	03:41	
Air Blank		0.000	03:42	
Control Test		0.078	03:43	
Air Blank		0.000	03:43	
Diagnostics Check		OK	03:43	

---

Data Download Date/Time: 01/06/2022 13:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BRABAND,M  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/29/2021 01:50:29

---

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.170	01:45
Breath Volume	2.375	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.168	01:48
Breath Volume	1.636	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

---

Data Download Date/Time: 01/06/2022 13:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENDY,HALEY,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: CHU,A  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 11:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:13

SUBJECT'S LAST NAME: PORRAS MONTERO

DATE/TIME OF TEST: 12/29/2021 12:40:37

---

Diagnostics Check	OK	12:33
Air Blank	0.000	12:34
Control Test	0.079	12:34
Air Blank	0.000	12:34
Subject Sample # 1	0.134	12:35
Breath Volume	3.820	12:35
Air Blank	0.000	12:36
Air Blank	0.000	12:38
Subject Sample # 2	0.131	12:38
Breath Volume	3.800	12:38
Air Blank	0.000	12:39
Control Test	0.078	12:40
Air Blank	0.000	12:40
Diagnostics Check	OK	12:40

---

Data Download Date/Time: 01/06/2022 13:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ROBERTS,J  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: STARGARDT

DATE/TIME OF TEST: 12/29/2021 22:11:29

---

Diagnostics Check	OK	22:04
Air Blank	0.000	22:05
Control Test	0.080	22:05
Air Blank	0.000	22:06
Subject Sample # 1	0.147	22:06
Breath Volume	1.996	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.134	22:09
Breath Volume	1.398	22:09
Air Blank	0.000	22:10
Control Test	0.079	22:10
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

---

Data Download Date/Time: 01/06/2022 13:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BLANCHARD,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: KRUCKENBERG

DATE/TIME OF TEST: 12/30/2021 03:42:36

---

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.079	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.249	03:34
Breath Volume		1.808	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.228	03:37
Breath Volume		1.214	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 3	0.237	03:40
Breath Volume		2.117	03:40
Air Blank		0.000	03:41
Control Test		0.078	03:41
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

---

Data Download Date/Time: 01/06/2022 13:59



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENDY,HALEY,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 12/31/2021 01:51:06

---

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.171	01:46
Breath Volume	3.765	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.163	01:49
Breath Volume	2.496	01:49
Air Blank	0.000	01:50
Control Test	0.078	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

---

Data Download Date/Time: 01/06/2022 13:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WASHINGTON, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/31/2021 21:59:09

---

Diagnostics Check		OK	21:51
Air Blank		0.000	21:52
Control Test		0.080	21:52
Air Blank		0.000	21:53
Subject Sample	# 1	0.253	21:54
Breath Volume		2.117	21:54
Air Blank		0.000	21:55
Air Blank		0.000	21:56
Subject Sample	# 2	0.244	21:57
Breath Volume		2.601	21:57
Air Blank		0.000	21:58
Control Test		0.077	21:58
Air Blank		0.000	21:59
Diagnostics Check		OK	21:59

---

Data Download Date/Time: 01/06/2022 14:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GENTRY, BENJAMIN, W  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: LEDBETTER, CALEB  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 11/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 05:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:07

SUBJECT'S LAST NAME: SANCHEZE TELLIS

DATE/TIME OF TEST: 12/04/2021 06:40:39

---

Diagnostics Check		OK	06:29	
Air Blank		0.000	06:29	
Control Test		0.081	06:30	
Air Blank		0.000	06:30	
Subject Sample	# 1	0.136	06:33	Volume Not Met
Breath Volume		0.117	06:33	
Air Blank		0.000	06:34	
Air Blank		0.000	06:36	
Subject Sample	# 2		06:39	No Sample Provided
Breath Volume		0.000	06:39	
Air Blank		0.000	06:39	
Control Test		0.076	06:40	
Air Blank		0.000	06:40	
Diagnostics Check		OK	06:40	

---

Data Download Date/Time: 12/20/2021 13:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GENTRY, BENJAMIN, W  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: LEDBETTER, CALEB  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 11/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 05:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:39

SUBJECT'S LAST NAME: SANCHEZE TELLIS

DATE/TIME OF TEST: 12/04/2021 07:08:07

---

Diagnostics Check	OK	07:00
Air Blank	0.000	07:01
Control Test	0.081	07:01
Air Blank	0.000	07:02
Subject Sample # 1	0.123	07:02
Breath Volume	1.886	07:02
Air Blank	0.000	07:03
Air Blank	0.000	07:05
Subject Sample # 2	0.126	07:06
Breath Volume	1.867	07:06
Air Blank	0.000	07:07
Control Test	0.078	07:07
Air Blank	0.000	07:08
Diagnostics Check	OK	07:08

---

Data Download Date/Time: 12/20/2021 13:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GORDON,GREGORY,S  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: THOMPSON,GREGORY  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 11/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 12:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:08

SUBJECT'S LAST NAME: HINDS

DATE/TIME OF TEST: 12/05/2021 13:37:42

---

Diagnostics Check	OK	13:31
Air Blank	0.000	13:31
Control Test	0.082	13:31
Air Blank	0.000	13:32
Subject Sample # 1	0.000	13:33
Breath Volume	3.371	13:33
Air Blank	0.000	13:33
Air Blank	0.000	13:35
Subject Sample # 2	0.000	13:36
Breath Volume	2.144	13:36
Air Blank	0.000	13:36
Control Test	0.079	13:37
Air Blank	0.000	13:37
Diagnostics Check	OK	13:37

---

Data Download Date/Time: 12/20/2021 13:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BENNETT,ARETHA,  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: TENA,MICHAEL  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 11/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 12/07/2021 23:10:42

---

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.081	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.166	23:06
Breath Volume	1.929	23:06
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.163	23:09
Breath Volume	1.246	23:09
Air Blank	0.000	23:09
Control Test	0.079	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

---

Data Download Date/Time: 12/20/2021 13:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,BRENDAN,D  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: BERRY,JACOB  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 11/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CYR

DATE/TIME OF TEST: 12/09/2021 04:07:53

---

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:01	
Control Test	0.081	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1 0.170	04:02	
Breath Volume	4.113	04:02	
Air Blank	0.000	04:03	
Air Blank	0.000	04:05	
Subject Sample	# 2 0.159	04:06	
Breath Volume	3.164	04:06	
Air Blank	0.000	04:06	
Control Test	0.092	04:07	Control Outside Tolerance
Air Blank	0.000	04:07	

---

Data Download Date/Time: 12/20/2021 13:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ,HENDRICK,J  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: MAYO,TREVOR  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 11/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 06:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 12/12/2021 06:41:33

---

Diagnostics Check	OK	06:35
Air Blank	0.000	06:35
Control Test	0.082	06:36
Air Blank	0.000	06:36
Subject Sample # 1	0.000	06:37
Breath Volume	2.605	06:37
Air Blank	0.000	06:37
Air Blank	0.000	06:39
Subject Sample # 2	0.000	06:40
Breath Volume	2.550	06:40
Air Blank	0.000	06:40
Control Test	0.083	06:40
Air Blank	0.000	06:41
Diagnostics Check	OK	06:41

---

Data Download Date/Time: 12/20/2021 13:11



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GLASS,DREW,S  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: GLASS,DREW  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 12/17/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 12/18/2021 02:05:20

---

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.082	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.179	02:00
Breath Volume	1.242	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.171	02:03
Breath Volume	1.398	02:03
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

---

Data Download Date/Time: 12/20/2021 13:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,BRENDAN,D  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: COTNER,STEVEN  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 12/17/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: NETTLES

DATE/TIME OF TEST: 12/19/2021 03:04:19

---

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.082	02:58
Air Blank	0.000	02:58
Subject Sample # 1	0.097	02:59
Breath Volume	3.417	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.098	03:02
Breath Volume	2.785	03:02
Air Blank	0.000	03:03
Control Test	0.081	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

---

Data Download Date/Time: 12/20/2021 13:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GENTRY, BENJAMIN, W  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: GLASS, DREW  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 12/17/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 12/19/2021 05:17:17

---

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.082	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.222	05:11
Breath Volume	1.691	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:13
Subject Sample # 2	0.218	05:15
Breath Volume	2.132	05:15
Air Blank	0.000	05:16
Control Test	0.078	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

---

Data Download Date/Time: 12/20/2021 13:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TOMASEGOVICH,KODY,D  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: PINO,R  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 12/17/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: ROBLES

DATE/TIME OF TEST: 12/23/2021 04:36:08

---

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.082	04:29
Air Blank	0.000	04:29
Subject Sample # 1	0.166	04:31
Breath Volume	1.492	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.160	04:34
Breath Volume	1.214	04:34
Air Blank	0.000	04:35
Control Test	0.082	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

---

Data Download Date/Time: 01/28/2022 19:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORALES,KARISSA,C  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: MORALES,KARISSA  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 12/17/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/25/2021 01:38:58

---

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.082	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.159	01:34
Breath Volume	1.445	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:36
Subject Sample # 2	0.152	01:37
Breath Volume	1.250	01:37
Air Blank	0.000	01:38
Control Test	0.077	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

---

Data Download Date/Time: 01/28/2022 19:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCMAHON,KENNETH,J  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: BERRY,JACOB  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 12/17/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/28/2021 23:38:11

---

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.082	23:32
Air Blank	0.000	23:33
Subject Sample # 1	0.220	23:33
Breath Volume	2.250	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.236	23:36
Breath Volume	2.039	23:36
Air Blank	0.000	23:37
Control Test	0.080	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

---

Data Download Date/Time: 01/28/2022 19:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEACH,ZACHARY,M  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: LEACH,ZACHARY  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 12/17/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: CRIDGE

DATE/TIME OF TEST: 12/31/2021 01:34:58

---

Diagnostics Check	OK	01:27
Air Blank	0.000	01:28
Control Test	0.082	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.191	01:30
Breath Volume	1.519	01:30
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.199	01:33
Breath Volume	2.253	01:33
Air Blank	0.000	01:34
Control Test	0.078	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

---

Data Download Date/Time: 01/28/2022 19:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TOMASEGOVICH,KODY,D  
BREATH TEST OPERATOR AGENCY: Pensacola PD  
ARRESTING OFFICER: BAUTISTA,J  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 12/17/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: SHORT-KONOPKA

DATE/TIME OF TEST: 12/31/2021 21:38:34

---

Diagnostics Check	OK	21:31
Air Blank	0.000	21:31
Control Test	0.082	21:31
Air Blank	0.000	21:32
Subject Sample # 1	0.202	21:33
Breath Volume	1.984	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:35
Subject Sample # 2	0.187	21:36
Breath Volume	1.921	21:36
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

---

Data Download Date/Time: 01/28/2022 19:37



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PINO,BRYAN,C  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: PINO,BRYAN  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 11/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: COURT ORD  
ARREST TIME: 15:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/02/2021 16:42:04

---

Diagnostics Check	OK	16:35
Air Blank	0.000	16:36
Control Test	0.082	16:36
Air Blank	0.000	16:36
Subject Sample # 1	0.000	16:37
Breath Volume	1.996	16:37
Air Blank	0.000	16:38
Air Blank	0.000	16:40
Subject Sample # 2	0.000	16:40
Breath Volume	1.785	16:40
Air Blank	0.000	16:41
Control Test	0.080	16:41
Air Blank	0.000	16:41
Diagnostics Check	OK	16:42

---

Data Download Date/Time: 01/03/2022 16:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,DAVID,E  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: WILLIAMS,DAVID  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 14:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:34

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 12/01/2021 16:01:06

---

Diagnostics Check	OK	15:54
Air Blank	0.000	15:55
Control Test	0.079	15:55
Air Blank	0.000	15:55
Subject Sample # 1	0.000	15:56
Breath Volume	2.167	15:56
Air Blank	0.000	15:57
Air Blank	0.000	15:59
Subject Sample # 2	0.000	15:59
Breath Volume	2.390	15:59
Air Blank	0.000	16:00
Control Test	0.078	16:00
Air Blank	0.000	16:01
Diagnostics Check	OK	16:01

---

Data Download Date/Time: 12/13/2021 15:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,LANCE,P  
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO  
ARRESTING OFFICER: HERNANDEZ,LANCE  
ARRESTING OFFICER AGENCY: MONROE COUNTY SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 13:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: OHARA

DATE/TIME OF TEST: 12/02/2021 14:16:55

---

Diagnostics Check		OK	14:05	
Air Blank		0.000	14:05	
Control Test		0.078	14:06	
Air Blank		0.000	14:06	
Subject Sample	# 1	0.000	14:09	Volume Not Met
Breath Volume		0.175	14:09	
Air Blank		0.000	14:10	
Air Blank		0.000	14:12	
Subject Sample	# 2	0.000	14:15	Volume Not Met
Breath Volume		1.078	14:15	
Air Blank		0.000	14:15	
Control Test		0.078	14:16	
Air Blank		0.000	14:16	
Diagnostics Check		OK	14:16	

---

Data Download Date/Time: 12/13/2021 15:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,LANCE,P  
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO  
ARRESTING OFFICER: HERNANDEZ,LANCE  
ARRESTING OFFICER AGENCY: MONROE COUNTY SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 13:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: O'HARA

DATE/TIME OF TEST: 12/02/2021 14:44:01

---

Diagnostics Check	OK	14:35	
Air Blank	0.000	14:35	
Control Test	0.078	14:35	
Air Blank	0.000	14:36	
Subject Sample # 1	0.000	14:39	Volume Not Met
Breath Volume	0.082	14:39	
Air Blank	0.000	14:39	
Air Blank	0.000	14:41	
Subject Sample # 2	0.000	14:42	
Breath Volume	1.152	14:42	
Air Blank		14:42	RFI Detect
Air Blank	0.000	14:43	
Control Test	0.078	14:43	
Air Blank	0.000	14:43	
Diagnostics Check	OK	14:44	

---

Data Download Date/Time: 12/13/2021 15:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,LANCE,P  
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO  
ARRESTING OFFICER: HERNANDEZ,LANCE  
ARRESTING OFFICER AGENCY: MONROE COUNTY SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 13:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: O'HARA

DATE/TIME OF TEST: 12/02/2021 15:10:42

---

Diagnostics Check	OK	15:03	
Air Blank	0.000	15:03	
Control Test	0.078	15:03	
Air Blank	0.000	15:04	
Subject Sample	# 1 0.000	15:07	Volume Not Met
Breath Volume	0.058	15:07	
Air Blank	0.000	15:07	
Air Blank	0.000	15:09	
Subject Sample	# 2	15:10	RFI Detect
Breath Volume	0.000	15:10	
Air Blank		15:10	RFI Detect
Air Blank	0.000	15:10	

---

Data Download Date/Time: 12/13/2021 15:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORENO,JORGE,  
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO  
ARRESTING OFFICER: MORENO,JORGE  
ARRESTING OFFICER AGENCY: MONROE COUNTY SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: YAGERMAN

DATE/TIME OF TEST: 12/04/2021 01:36:24

---

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:33	
Control Test	0.079	01:34	
Air Blank	0.000	01:34	
Subject Sample	# 1	01:35	Subject Test Refused
Breath Volume	0.000	01:35	
Air Blank	0.000	01:35	
Control Test	0.078	01:35	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

---

Data Download Date/Time: 12/13/2021 15:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORENO,JORGE,  
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO  
ARRESTING OFFICER: HRADECKY,CORBIN  
ARRESTING OFFICER AGENCY: MONROE COUNTY SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/04/2021 20:59:49

---

Diagnostics Check	OK	20:56	
Air Blank	0.000	20:57	
Control Test	0.079	20:57	
Air Blank	0.000	20:58	
Subject Sample	# 1	20:58	Subject Test Refused
Breath Volume	0.000	20:58	
Air Blank	0.000	20:58	
Control Test	0.079	20:59	
Air Blank	0.000	20:59	
Diagnostics Check	OK	20:59	

---

Data Download Date/Time: 12/13/2021 15:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUNA,ANGEL,  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: LUNA,ANGEL  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: RADECKA

DATE/TIME OF TEST: 12/04/2021 21:51:29

---

Diagnostics Check	OK	21:45	
Air Blank	0.000	21:46	
Control Test	0.079	21:46	
Air Blank	0.000	21:47	
Subject Sample	# 1	21:49	Slope Not Met
Breath Volume	1.214	21:49	
Air Blank	0.000	21:50	
Control Test	0.079	21:50	
Air Blank	0.000	21:51	
Diagnostics Check	OK	21:51	

---

Data Download Date/Time: 12/13/2021 15:32



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUNA,ANGEL,  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: LUNA,ANGEL  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: RADECKA

DATE/TIME OF TEST: 12/04/2021 22:01:41

---

Diagnostics Check	OK	21:54
Air Blank	0.000	21:54
Control Test	0.078	21:55
Air Blank	0.000	21:55
Subject Sample # 1	0.300	21:56
Breath Volume	2.488	21:56
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.283	22:00
Breath Volume	1.265	22:00
Air Blank	0.000	22:00
Control Test	0.077	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

---

Data Download Date/Time: 12/13/2021 15:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HRADECKY,CORBIN,J  
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO  
ARRESTING OFFICER: VALDES,D  
ARRESTING OFFICER AGENCY: MONROE COUNTY SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: MCKEOWN

DATE/TIME OF TEST: 12/05/2021 02:45:43

---

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:43	
Control Test	0.079	02:43	
Air Blank	0.000	02:44	
Subject Sample	# 1	02:44	Subject Test Refused
Breath Volume	0.000	02:44	
Air Blank	0.000	02:44	
Control Test	0.078	02:45	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

---

Data Download Date/Time: 12/13/2021 15:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONES,SHAUN,G  
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO  
ARRESTING OFFICER: LONES,SHAUN  
ARRESTING OFFICER AGENCY: MONROE COUNTY SO  
LAST AGENCY INSPECTION: 11/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 13:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:31

SUBJECT'S LAST NAME: KEENEY

DATE/TIME OF TEST: 12/10/2021 13:58:18

---

Diagnostics Check	OK	13:51
Air Blank	0.000	13:52
Control Test	0.078	13:52
Air Blank	0.000	13:52
Subject Sample # 1	0.237	13:53
Breath Volume	1.519	13:53
Air Blank	0.000	13:54
Air Blank	0.000	13:56
Subject Sample # 2	0.239	13:56
Breath Volume	1.847	13:56
Air Blank	0.000	13:57
Control Test	0.077	13:57
Air Blank	0.000	13:58
Diagnostics Check	OK	13:58

---

Data Download Date/Time: 12/13/2021 15:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MATTHEWS,CASEY,A  
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO  
ARRESTING OFFICER: DOMINGUEZ,OSCAR  
ARRESTING OFFICER AGENCY: MONROE COUNTY SO  
LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 18:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: COSME

DATE/TIME OF TEST: 12/13/2021 19:09:58

---

Diagnostics Check	OK	19:06	
Air Blank	0.000	19:07	
Control Test	0.079	19:07	
Air Blank	0.000	19:08	
Subject Sample	# 1	19:08	Subject Test Refused
Breath Volume	0.000	19:08	
Air Blank	0.000	19:09	
Control Test	0.080	19:09	
Air Blank	0.000	19:09	
Diagnostics Check	OK	19:09	

---

Data Download Date/Time: 01/06/2022 14:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,  
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO  
ARRESTING OFFICER: RODRIGUEZ,RICHARD  
ARRESTING OFFICER AGENCY: MONROE COUNTY SO  
LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 12/17/2021 00:23:06

---

Diagnostics Check		OK	00:17
Air Blank		0.000	00:17
Control Test		0.079	00:17
Air Blank		0.000	00:18
Subject Sample	# 1	0.180	00:18
Breath Volume		1.207	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.194	00:21
Breath Volume		1.652	00:21
Air Blank		0.000	00:22
Control Test		0.078	00:22
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

---

Data Download Date/Time: 01/06/2022 14:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORENO,JORGE,  
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO  
ARRESTING OFFICER: DOMINGUEZ,OSCAR  
ARRESTING OFFICER AGENCY: MONROE COUNTY SO  
LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 12/17/2021 23:59:27

---

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:57	
Control Test	0.079	23:57	
Air Blank	0.000	23:57	
Subject Sample	# 1	23:58	Subject Test Refused
Breath Volume	0.000	23:58	
Air Blank	0.000	23:58	
Control Test	0.079	23:58	
Air Blank	0.000	23:59	
Diagnostics Check	OK	23:59	

---

Data Download Date/Time: 01/06/2022 14:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JOEL,M  
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO  
ARRESTING OFFICER: DOMINGUEZ,OSCAR  
ARRESTING OFFICER AGENCY: MONROE COUNTY SO  
LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: ARIAS

DATE/TIME OF TEST: 12/18/2021 22:44:09

---

Diagnostics Check	OK	22:37
Air Blank	0.000	22:38
Control Test	0.078	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.138	22:39
Breath Volume	1.976	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.140	22:42
Breath Volume	1.750	22:42
Air Blank	0.000	22:43
Control Test	0.079	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

---

Data Download Date/Time: 01/06/2022 14:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANNA,WARDELL,P  
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO  
ARRESTING OFFICER: HANNA,WARDELL  
ARRESTING OFFICER AGENCY: MONROE COUNTY SO  
LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 12/19/2021 22:35:21

---

Diagnostics Check	OK	22:28
Air Blank	0.000	22:29
Control Test	0.079	22:29
Air Blank	0.000	22:30
Subject Sample # 1	0.090	22:30
Breath Volume	1.230	22:30
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.090	22:33
Breath Volume	1.218	22:33
Air Blank	0.000	22:34
Control Test	0.079	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

---

Data Download Date/Time: 01/06/2022 14:37



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,  
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO  
ARRESTING OFFICER: RODRIGUEZ,RICHARD  
ARRESTING OFFICER AGENCY: MONROE COUNTY SO  
LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MANCUSO

DATE/TIME OF TEST: 12/21/2021 00:43:58

---

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:41	
Control Test	0.079	00:41	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:42	Subject Test Refused
Breath Volume	0.000	00:42	
Air Blank	0.000	00:43	
Control Test	0.079	00:43	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

---

Data Download Date/Time: 01/06/2022 14:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JOEL,M  
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO  
ARRESTING OFFICER: TORRES,JOEL  
ARRESTING OFFICER AGENCY: MONROE COUNTY SO  
LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: HODGES

DATE/TIME OF TEST: 12/23/2021 03:36:40

---

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:30	
Control Test	0.077	03:30	
Air Blank	0.000	03:31	
Subject Sample	# 1 0.099	03:32	
Breath Volume	1.187	03:32	
Air Blank	0.000	03:33	
Air Blank	0.000	03:35	
Subject Sample	# 2	03:35	Subject Test Refused
Breath Volume	0.000	03:35	
Air Blank	0.000	03:35	
Control Test	0.076	03:36	
Air Blank	0.000	03:36	
Diagnostics Check	OK	03:36	

---

Data Download Date/Time: 01/06/2022 14:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORALES,VICTOR,M  
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO  
ARRESTING OFFICER: MORALES,VICTOR  
ARRESTING OFFICER AGENCY: MONROE COUNTY SO  
LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 06:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:33

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 12/24/2021 07:00:44

---

Diagnostics Check	OK	06:53
Air Blank	0.000	06:54
Control Test	0.078	06:54
Air Blank	0.000	06:55
Subject Sample # 1	0.130	06:56
Breath Volume	1.605	06:56
Air Blank	0.000	06:56
Air Blank	0.000	06:58
Subject Sample # 2	0.133	06:59
Breath Volume	1.492	06:59
Air Blank	0.000	06:59
Control Test	0.078	07:00
Air Blank	0.000	07:00
Diagnostics Check	OK	07:00

---

Data Download Date/Time: 01/06/2022 14:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORENO,JORGE,  
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO  
ARRESTING OFFICER: MORENO,JORGE  
ARRESTING OFFICER AGENCY: MONROE COUNTY SO  
LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: ENGLISH

DATE/TIME OF TEST: 12/28/2021 23:24:18

---

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.079	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.027	23:19
Breath Volume	1.144	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:21
Subject Sample # 2	0.025	23:22
Breath Volume	2.007	23:22
Air Blank	0.000	23:23
Control Test	0.079	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

---

Data Download Date/Time: 01/06/2022 14:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JOEL,M  
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO  
ARRESTING OFFICER: TORRES,JOEL  
ARRESTING OFFICER AGENCY: MONROE COUNTY SO  
LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: STRANDQUIST

DATE/TIME OF TEST: 12/30/2021 00:41:05

---

Diagnostics Check		OK	00:34
Air Blank		0.000	00:34
Control Test		0.079	00:35
Air Blank		0.000	00:35
Subject Sample	# 1	0.201	00:36
Breath Volume		1.898	00:36
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.201	00:39
Breath Volume		1.769	00:39
Air Blank		0.000	00:40
Control Test		0.079	00:40
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

---

Data Download Date/Time: 01/06/2022 14:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,JOEL,M  
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO  
ARRESTING OFFICER: RAMIREZ,ANDRES  
ARRESTING OFFICER AGENCY: MONROE COUNTY SO  
LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: CARVAJAL CARVAJAL

DATE/TIME OF TEST: 12/30/2021 01:27:57

---

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.104	01:23
Breath Volume	1.582	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:25
Subject Sample # 2	0.103	01:26
Breath Volume	1.500	01:26
Air Blank	0.000	01:27
Control Test	0.078	01:27
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

---

Data Download Date/Time: 01/06/2022 14:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUIZ,PABLO,R  
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)  
ARRESTING OFFICER: RUIZ,PABLO  
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)  
LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: BUI  
ARREST TIME: 14:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:26

SUBJECT'S LAST NAME: FINCH

DATE/TIME OF TEST: 12/30/2021 16:51:55

---

Diagnostics Check	OK	16:48	
Air Blank	0.000	16:49	
Control Test	0.078	16:49	
Air Blank	0.000	16:50	
Subject Sample	# 1	16:50	Subject Test Refused
Breath Volume	0.000	16:50	
Air Blank	0.000	16:50	
Control Test	0.079	16:51	
Air Blank	0.000	16:51	
Diagnostics Check	OK	16:51	

---

Data Download Date/Time: 01/06/2022 14:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOGGAN,MARLON,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: BOGGAN,MARLON  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 12/10/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1355293  
DRY GAS EXPIRY DATE: 07/29/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: TEUTON

DATE/TIME OF TEST: 12/10/2021 20:30:26

---

Diagnostics Check	OK	20:19	
Air Blank	0.000	20:19	
Control Test	0.078	20:19	
Air Blank	0.000	20:20	
Subject Sample	# 1	20:23	No Sample Provided
Breath Volume	0.000	20:23	
Air Blank	0.000	20:23	
Air Blank	0.000	20:25	
Subject Sample	# 2	20:29	No Sample Provided
Breath Volume	0.000	20:29	
Air Blank	0.000	20:29	
Control Test	0.078	20:29	
Air Blank	0.000	20:30	
Diagnostics Check	OK	20:30	

---

Data Download Date/Time: 01/27/2022 15:16



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'DOR,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: O'DOR,ROBERT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 12/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: GONZALEZ GUTIERREZ

DATE/TIME OF TEST: 12/10/2021 22:17:44

---

Diagnostics Check	OK	22:16	
Air Blank	0.000	22:16	
Control Test	0.074	22:17	Control Outside Tolerance
Air Blank	0.000	22:17	

---

Data Download Date/Time: 01/05/2022 11:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'DOR,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: O'DOR,ROBERT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 12/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: GONZALEZ GUTIERREZ

DATE/TIME OF TEST: 12/10/2021 22:48:04

---

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:46	
Control Test	0.078	22:46	
Air Blank	0.000	22:46	
Subject Sample	# 1	22:47	RFI Detect
Breath Volume	0.000	22:47	
Air Blank	0.000	22:47	
		22:47	RFI Detect
Air Blank		22:48	RFI Detect

---

Data Download Date/Time: 01/05/2022 11:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WHITE, EDWARD  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: BOIVERT

DATE/TIME OF TEST: 12/03/2021 22:48:35

---

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.080	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.137	22:43
Breath Volume	1.667	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.139	22:46
Breath Volume	1.753	22:46
Air Blank	0.000	22:47
Control Test	0.079	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

---

Data Download Date/Time: 12/10/2021 14:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CISSON,JACKIE  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: YANICK

DATE/TIME OF TEST: 12/04/2021 00:21:42

---

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.061	00:17
Breath Volume	1.593	00:17
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.060	00:20
Breath Volume	1.218	00:20
Air Blank	0.000	00:20
Control Test	0.080	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

---

Data Download Date/Time: 12/10/2021 14:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WALDKOETTER, ELIZABETH  
ARRESTING OFFICER AGENCY: Lantana PD  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: SIERRA

DATE/TIME OF TEST: 12/04/2021 01:39:35

---

Diagnostics Check		OK	01:32
Air Blank		0.000	01:33
Control Test		0.079	01:33
Air Blank		0.000	01:34
Subject Sample	# 1	0.227	01:34
Breath Volume		1.968	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.232	01:37
Breath Volume		1.949	01:37
Air Blank		0.000	01:38
Control Test		0.079	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

---

Data Download Date/Time: 12/10/2021 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: AMADON, WILL  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: VARGAS RAMIREZ

DATE/TIME OF TEST: 12/04/2021 03:08:27

---

Diagnostics Check		OK	02:59
Air Blank		0.000	03:00
Control Test		0.080	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.213	03:03
Breath Volume		1.695	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:05
Subject Sample	# 2	0.212	03:06
Breath Volume		1.847	03:06
Air Blank		0.000	03:07
Control Test		0.076	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

---

Data Download Date/Time: 12/10/2021 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SOLOWAY,ALAN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ZELAYA CARDONA

DATE/TIME OF TEST: 12/04/2021 03:31:54

---

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.168	03:27
Breath Volume	2.355	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:29
Subject Sample # 2	0.159	03:30
Breath Volume	1.289	03:30
Air Blank	0.000	03:30
Control Test	0.079	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

---

Data Download Date/Time: 12/10/2021 14:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: TRINIDAD,OSCAR  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:14

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 12/04/2021 06:49:18

---

Diagnostics Check	OK	06:42
Air Blank	0.000	06:43
Control Test	0.079	06:43
Air Blank	0.000	06:44
Subject Sample # 1	0.105	06:44
Breath Volume	1.644	06:44
Air Blank	0.000	06:45
Air Blank	0.000	06:47
Subject Sample # 2	0.109	06:47
Breath Volume	3.550	06:47
Air Blank	0.000	06:48
Control Test	0.080	06:48
Air Blank	0.000	06:49
Diagnostics Check	OK	06:49

---

Data Download Date/Time: 12/10/2021 14:23



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: AMADON,WILL  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: AJTUM IXCOY

DATE/TIME OF TEST: 12/04/2021 21:52:43

---

Diagnostics Check	OK	21:46
Air Blank	0.000	21:47
Control Test	0.080	21:47
Air Blank	0.000	21:47
Subject Sample # 1	0.231	21:48
Breath Volume	1.679	21:48
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 2	0.229	21:51
Breath Volume	2.441	21:51
Air Blank	0.000	21:51
Control Test	0.079	21:52
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

---

Data Download Date/Time: 12/10/2021 14:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: TAVERAS,JOSUE  
ARRESTING OFFICER AGENCY: West Palm Beach PD  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: JIMENEZ RAMIREZ

DATE/TIME OF TEST: 12/05/2021 02:37:24

---

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.080	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.189	02:32
Breath Volume		1.976	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.193	02:35
Breath Volume		2.359	02:35
Air Blank		0.000	02:36
Control Test		0.078	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

---

Data Download Date/Time: 12/10/2021 14:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FANDREY,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/05/2021 03:28:00

---

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.079	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.146	03:22
Breath Volume	1.636	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.144	03:26
Breath Volume	1.714	03:26
Air Blank	0.000	03:27
Control Test	0.079	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:28

---

Data Download Date/Time: 12/10/2021 14:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WINDSOR,NICHOLAS  
ARRESTING OFFICER AGENCY: Delray Beach PD  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: ROCHEMONT

DATE/TIME OF TEST: 12/05/2021 04:03:18

---

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.078	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.169	03:58
Breath Volume	2.843	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.161	04:01
Breath Volume	2.617	04:01
Air Blank	0.000	04:02
Control Test	0.078	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

---

Data Download Date/Time: 12/13/2021 14:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: AMADON,WILL  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/05/2021 04:28:31

---

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.080	04:21
Air Blank		0.000	04:22
Subject Sample	# 1	0.177	04:23
Breath Volume		1.605	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:25
Subject Sample	# 2	0.166	04:26
Breath Volume		1.371	04:26
Air Blank		0.000	04:27
Control Test		0.078	04:27
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

---

Data Download Date/Time: 12/13/2021 14:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SAGET,HORLKINS  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 06:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: SAVAGE

DATE/TIME OF TEST: 12/05/2021 07:38:56

---

Diagnostics Check	OK	07:32
Air Blank	0.000	07:32
Control Test	0.079	07:33
Air Blank	0.000	07:33
Subject Sample # 1	0.070	07:34
Breath Volume	1.929	07:34
Air Blank	0.000	07:34
Air Blank	0.000	07:36
Subject Sample # 2	0.074	07:37
Breath Volume	2.457	07:37
Air Blank	0.000	07:38
Control Test	0.080	07:38
Air Blank	0.000	07:38
Diagnostics Check	OK	07:38

---

Data Download Date/Time: 12/13/2021 14:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WINDSOR,NICHOLAS  
ARRESTING OFFICER AGENCY: Delray Beach PD  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: CEDENO

DATE/TIME OF TEST: 12/05/2021 20:32:46

---

Diagnostics Check	OK	20:25
Air Blank	0.000	20:26
Control Test	0.080	20:26
Air Blank	0.000	20:26
Subject Sample # 1	0.171	20:27
Breath Volume	1.343	20:27
Air Blank	0.000	20:28
Air Blank	0.000	20:30
Subject Sample # 2	0.184	20:31
Breath Volume	3.156	20:31
Air Blank	0.000	20:31
Control Test	0.079	20:32
Air Blank	0.000	20:32
Diagnostics Check	OK	20:32

---

Data Download Date/Time: 12/13/2021 14:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: LOVETT,JAMES  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/05/2021 22:08:26

---

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.079	22:02
Air Blank	0.000	22:02
Subject Sample # 1	0.145	22:04
Breath Volume	1.605	22:04
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.153	22:06
Breath Volume	1.593	22:06
Air Blank	0.000	22:07
Control Test	0.080	22:07
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

---

Data Download Date/Time: 12/13/2021 14:05



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: BUTZBACH,WILLIAM  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: OLSON

DATE/TIME OF TEST: 12/06/2021 01:56:37

---

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.080	01:50
Air Blank		0.000	01:50
Subject Sample	# 1	0.085	01:51
Breath Volume		1.628	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.076	01:55
Breath Volume		2.449	01:55
Air Blank		0.000	01:55
Control Test		0.078	01:56
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

---

Data Download Date/Time: 12/13/2021 14:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: HERNANDEZ,EDWIN  
ARRESTING OFFICER AGENCY: Delray Beach PD  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 16:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: STUBBS

DATE/TIME OF TEST: 12/06/2021 18:18:14

---

Diagnostics Check	OK	18:12	
Air Blank	0.000	18:13	
Control Test	0.080	18:13	
Air Blank	0.000	18:14	
Subject Sample	# 1	18:16	Subject Test Refused
Breath Volume	0.000	18:16	
Air Blank	0.000	18:17	
Control Test	0.080	18:17	
Air Blank	0.000	18:18	
Diagnostics Check	OK	18:18	

---

Data Download Date/Time: 12/13/2021 14:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: AMADON,WILL  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: MANSOLILLO

DATE/TIME OF TEST: 12/06/2021 21:02:39

---

Diagnostics Check		OK	20:56
Air Blank		0.000	20:56
Control Test		0.080	20:56
Air Blank		0.000	20:57
Subject Sample	# 1	0.173	20:58
Breath Volume		1.363	20:58
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.159	21:00
Breath Volume		1.351	21:00
Air Blank		0.000	21:01
Control Test		0.079	21:02
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

---

Data Download Date/Time: 12/13/2021 14:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: TEJEDA,ANTHONY  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SACUJ TIAN

DATE/TIME OF TEST: 12/07/2021 03:41:10

---

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.081	03:35
Air Blank		0.000	03:35
Subject Sample	# 1	0.233	03:36
Breath Volume		1.277	03:36
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.233	03:39
Breath Volume		1.875	03:39
Air Blank		0.000	03:40
Control Test		0.079	03:40
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

---

Data Download Date/Time: 12/13/2021 14:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: HERNANDEZ,EDWIN  
ARRESTING OFFICER AGENCY: Delray Beach PD  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 11:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:05

SUBJECT'S LAST NAME: NASSIFF

DATE/TIME OF TEST: 12/07/2021 12:34:46

---

Diagnostics Check	OK	12:28
Air Blank	0.000	12:28
Control Test	0.080	12:28
Air Blank	0.000	12:29
Subject Sample # 1	0.000	12:30
Breath Volume	2.347	12:30
Air Blank	0.000	12:30
Air Blank	0.000	12:32
Subject Sample # 2	0.000	12:33
Breath Volume	2.648	12:33
Air Blank	0.000	12:33
Control Test	0.081	12:34
Air Blank	0.000	12:34
Diagnostics Check	OK	12:34

---

Data Download Date/Time: 12/13/2021 14:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: JOHNSON,TONRICA  
ARRESTING OFFICER AGENCY: Riviera Beach PD  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: PEREIRA

DATE/TIME OF TEST: 12/08/2021 02:19:32

---

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.080	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.109	02:15
Breath Volume	1.578	02:15
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.104	02:17
Breath Volume	1.839	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

---

Data Download Date/Time: 12/13/2021 14:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: NALERIO,LUIS  
ARRESTING OFFICER AGENCY: Boynton Beach PD  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 16:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: BARTALINI

DATE/TIME OF TEST: 12/08/2021 17:44:11

---

Diagnostics Check	OK	17:36
Air Blank	0.000	17:36
Control Test	0.080	17:37
Air Blank	0.000	17:37
Subject Sample # 1	0.000	17:38
Breath Volume	3.226	17:38
Air Blank	0.000	17:38
Air Blank	0.000	17:40
Subject Sample # 2	0.000	17:42
Breath Volume	1.664	17:42
Air Blank	0.000	17:43
Control Test	0.080	17:43
Air Blank	0.000	17:44
Diagnostics Check	OK	17:44

---

Data Download Date/Time: 12/13/2021 14:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FLINK,ANDREW  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: LLOYD

DATE/TIME OF TEST: 12/08/2021 22:35:35

---

Diagnostics Check	OK	22:29
Air Blank	0.000	22:29
Control Test	0.080	22:29
Air Blank	0.000	22:30
Subject Sample # 1	0.000	22:31
Breath Volume	3.191	22:31
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.000	22:34
Breath Volume	3.691	22:34
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

---

Data Download Date/Time: 12/13/2021 14:05



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WINDSOR,NICHOLAS  
ARRESTING OFFICER AGENCY: Delray Beach PD  
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: HERMAN

DATE/TIME OF TEST: 12/09/2021 01:16:51

---

Diagnostics Check		OK	01:05	
Air Blank		0.000	01:05	
Control Test		0.080	01:05	
Air Blank		0.000	01:06	
Subject Sample	# 1	0.129	01:09	Volume Not Met
Breath Volume		0.558	01:09	
Air Blank		0.000	01:10	
Air Blank		0.000	01:12	
Subject Sample	# 2	0.116	01:15	Volume Not Met
Breath Volume		0.312	01:15	
Air Blank		0.000	01:15	
Control Test		0.078	01:16	
Air Blank		0.000	01:16	
Diagnostics Check		OK	01:16	

---

Data Download Date/Time: 12/13/2021 14:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: POINTU,PATRICK  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MEULEMAN

DATE/TIME OF TEST: 12/14/2021 03:20:06

---

Diagnostics Check		OK	03:13
Air Blank		0.000	03:13
Control Test		0.080	03:14
Air Blank		0.000	03:14
Subject Sample	# 1	0.172	03:15
Breath Volume		2.234	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:17
Subject Sample	# 2	0.171	03:18
Breath Volume		1.453	03:18
Air Blank		0.000	03:19
Control Test		0.079	03:19
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

---

Data Download Date/Time: 01/13/2022 13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WHITE,EDWARD  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: VICENTE IXMAY

DATE/TIME OF TEST: 12/15/2021 01:31:54

---

Diagnostics Check		OK	01:24
Air Blank		0.000	01:25
Control Test		0.081	01:25
Air Blank		0.000	01:25
Subject Sample	# 1	0.174	01:27
Breath Volume		1.906	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:29
Subject Sample	# 2	0.164	01:30
Breath Volume		1.121	01:30
Air Blank		0.000	01:30
Control Test		0.079	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

---

Data Download Date/Time: 01/13/2022 13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FLINK,ANDREW  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: CEDENO PICO

DATE/TIME OF TEST: 12/17/2021 04:51:11

---

Diagnostics Check	OK	04:44
Air Blank	0.000	04:44
Control Test	0.080	04:44
Air Blank	0.000	04:45
Subject Sample # 1	0.126	04:46
Breath Volume	2.554	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.126	04:49
Breath Volume	2.839	04:49
Air Blank	0.000	04:50
Control Test	0.080	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

---

Data Download Date/Time: 01/13/2022 13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: POINTU,PATRICK  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/18/2021 01:32:56

---

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.081	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.178	01:28
Breath Volume	2.890	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:30
Subject Sample # 2	0.180	01:31
Breath Volume	3.277	01:31
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

---

Data Download Date/Time: 01/13/2022 13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: POINTU,PATRICK  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: CREARY

DATE/TIME OF TEST: 12/18/2021 04:51:22

---

Diagnostics Check	OK	04:44
Air Blank	0.000	04:45
Control Test	0.080	04:45
Air Blank	0.000	04:46
Subject Sample # 1	0.173	04:46
Breath Volume	1.800	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.174	04:49
Breath Volume	2.140	04:49
Air Blank	0.000	04:50
Control Test	0.080	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

---

Data Download Date/Time: 01/13/2022 13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: POINTU,PATRICK  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: VASQUEZ LOPEZ

DATE/TIME OF TEST: 12/18/2021 22:42:40

---

Diagnostics Check	OK	22:33
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.120	22:37
Breath Volume	1.667	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:39
Subject Sample # 2	0.117	22:41
Breath Volume	1.363	22:41
Air Blank	0.000	22:41
Control Test	0.079	22:42
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

---

Data Download Date/Time: 01/13/2022 13:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: JEFFRIES,MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/19/2021 01:41:09

---

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.000	01:36
Breath Volume	2.664	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.000	01:39
Breath Volume	1.156	01:39
Air Blank	0.000	01:40
Control Test	0.080	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

---

Data Download Date/Time: 01/13/2022 13:29



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: NALERIO,LUIS  
ARRESTING OFFICER AGENCY: Boynton Beach PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: MARCHENA

DATE/TIME OF TEST: 12/19/2021 02:55:30

---

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.079	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.168	02:49
Breath Volume	1.753	02:49
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.170	02:53
Breath Volume	1.238	02:53
Air Blank	0.000	02:54
Control Test	0.078	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

---

Data Download Date/Time: 01/13/2022 13:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CISSON,JACKIE  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: CRISTOBAL

DATE/TIME OF TEST: 12/19/2021 04:19:40

---

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.185	04:15
Breath Volume	3.253	04:15
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.177	04:18
Breath Volume	3.437	04:18
Air Blank	0.000	04:18
Control Test	0.078	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

---

Data Download Date/Time: 01/13/2022 13:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FLINK,ANDREW  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: HOLLOWAY

DATE/TIME OF TEST: 12/21/2021 23:37:12

---

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.080	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.253	23:32
Breath Volume	3.144	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.253	23:35
Breath Volume	3.183	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

---

Data Download Date/Time: 01/13/2022 13:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FLINK,A  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: SCHUCHMANN

DATE/TIME OF TEST: 12/22/2021 04:00:22

---

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.182	03:52
Breath Volume	1.503	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.204	03:55
Breath Volume	2.667	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 3	0.184	03:58
Breath Volume	2.062	03:58
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

---

Data Download Date/Time: 01/13/2022 13:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: ROBINSON,HAROLD  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 16:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: FITZGERALD

DATE/TIME OF TEST: 12/22/2021 18:39:32

---

Diagnostics Check	OK	18:29
Air Blank	0.000	18:30
Control Test	0.080	18:30
Air Blank	0.000	18:31
Subject Sample # 1	0.289	18:31
Breath Volume	2.046	18:31
Air Blank	0.000	18:32
Air Blank	0.000	18:34
Subject Sample # 2	0.259	18:34
Breath Volume	2.242	18:34
Air Blank	0.000	18:35
Air Blank	0.000	18:37
Subject Sample # 3	0.240	18:37
Breath Volume	1.699	18:37
Air Blank	0.000	18:38
Control Test	0.078	18:38
Air Blank	0.000	18:39
Diagnostics Check	OK	18:39

---

Data Download Date/Time: 01/13/2022 13:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: NALERIO,LUIS  
ARRESTING OFFICER AGENCY: Boynton Beach PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 12/23/2021 02:18:31

---

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.080	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.245	02:13
Breath Volume		2.035	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.249	02:16
Breath Volume		1.511	02:16
Air Blank		0.000	02:17
Control Test		0.078	02:17
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

---

Data Download Date/Time: 01/13/2022 13:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: LOPEZ,NATALIE  
ARRESTING OFFICER AGENCY: West Palm Beach PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 17:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: RANCIC

DATE/TIME OF TEST: 12/23/2021 18:46:27

---

Diagnostics Check	OK	18:34	
Air Blank	0.000	18:35	
Control Test	0.080	18:35	
Air Blank	0.000	18:35	
Subject Sample	# 1	18:39	No Sample Provided
Breath Volume	0.000	18:39	
Air Blank	0.000	18:39	
Air Blank	0.000	18:41	
Subject Sample	# 2 0.213	18:44	Volume Not Met
Breath Volume	0.359	18:44	
Air Blank	0.000	18:45	
Control Test	0.079	18:45	
Air Blank	0.000	18:46	
Diagnostics Check	OK	18:46	

---

Data Download Date/Time: 01/13/2022 13:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: TEJEDA,ANTHONY  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/23/2021 23:40:32

---

Diagnostics Check	OK	23:33
Air Blank	0.000	23:34
Control Test	0.079	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.178	23:35
Breath Volume	1.859	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.174	23:38
Breath Volume	1.550	23:38
Air Blank	0.000	23:39
Control Test	0.079	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

---

Data Download Date/Time: 01/13/2022 13:31



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: POINTU,PATRICK  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: LAYNEZ-RAMIREZ

DATE/TIME OF TEST: 12/25/2021 00:19:25

---

Diagnostics Check	OK	00:11
Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:12
Subject Sample # 1	0.324	00:14
Breath Volume	1.824	00:14
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.307	00:17
Breath Volume	1.101	00:17
Air Blank	0.000	00:18
Control Test	0.078	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

---

Data Download Date/Time: 01/13/2022 13:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: DELHOMME,SHAWN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 12/25/2021 03:09:26

---

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.083	03:04
Breath Volume	3.199	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.080	03:07
Breath Volume	3.289	03:07
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

---

Data Download Date/Time: 01/13/2022 13:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: JENNINGL,DAVON  
ARRESTING OFFICER AGENCY: Boynton Beach PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: SCHIFFMAN

DATE/TIME OF TEST: 12/25/2021 03:59:02

---

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.079	03:55	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:58	
Control Test	0.080	03:58	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:59	

---

Data Download Date/Time: 01/13/2022 13:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: HARTY,WARREN  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: WALCOTT

DATE/TIME OF TEST: 12/25/2021 19:57:58

---

Diagnostics Check	OK	19:55	
Air Blank	0.000	19:55	
Control Test	0.080	19:55	
Air Blank	0.000	19:56	
Subject Sample	# 1	19:56	Subject Test Refused
Breath Volume	0.000	19:56	
Air Blank	0.000	19:57	
Control Test	0.079	19:57	
Air Blank	0.000	19:57	
Diagnostics Check	OK	19:57	

---

Data Download Date/Time: 01/13/2022 13:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MCGILLICUDDY,STEVEN  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: CAMPOSECOPEREZ

DATE/TIME OF TEST: 12/25/2021 21:21:13

---

Diagnostics Check	OK	21:14
Air Blank	0.000	21:15
Control Test	0.080	21:15
Air Blank	0.000	21:15
Subject Sample # 1	0.097	21:16
Breath Volume	3.421	21:16
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample # 2	0.094	21:19
Breath Volume	2.695	21:19
Air Blank	0.000	21:20
Control Test	0.079	21:20
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

---

Data Download Date/Time: 01/13/2022 13:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SOLOWAY,ALAN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: VILLAFUERTE

DATE/TIME OF TEST: 12/25/2021 23:50:46

---

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.080	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.238	23:46
Breath Volume	1.988	23:46
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.229	23:49
Breath Volume	1.816	23:49
Air Blank	0.000	23:49
Control Test	0.079	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

---

Data Download Date/Time: 01/13/2022 13:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: BRUNO,SNYDER  
ARRESTING OFFICER AGENCY: Delray Beach PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 12/26/2021 00:41:13

---

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.080	00:35
Air Blank	0.000	00:35
Subject Sample # 1	0.175	00:36
Breath Volume	1.441	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.184	00:39
Breath Volume	1.964	00:39
Air Blank	0.000	00:40
Control Test	0.079	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

---

Data Download Date/Time: 01/13/2022 13:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WHITE,EDWARD  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: COUNIHAN

DATE/TIME OF TEST: 12/26/2021 02:39:07

---

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.079	02:33
Air Blank		0.000	02:34
Subject Sample	# 1	0.226	02:34
Breath Volume		1.429	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.211	02:37
Breath Volume		1.289	02:37
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

---

Data Download Date/Time: 01/13/2022 13:32



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FESTA,JESSI  
ARRESTING OFFICER AGENCY: Boynton Beach PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/27/2021 00:11:48

---

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.080	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.000	00:07
Breath Volume	2.414	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.000	00:10
Breath Volume	1.847	00:10
Air Blank	0.000	00:10
Control Test	0.080	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

---

Data Download Date/Time: 01/13/2022 13:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SOLOWAY,ALAN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: VAZQUEZ PEREZ

DATE/TIME OF TEST: 12/27/2021 01:41:38

---

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.080	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.177	01:37
Breath Volume		2.613	01:37
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.175	01:40
Breath Volume		2.753	01:40
Air Blank		0.000	01:40
Control Test		0.079	01:41
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

---

Data Download Date/Time: 01/13/2022 13:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WARD,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 05:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: METHELUS

DATE/TIME OF TEST: 12/27/2021 06:18:26

---

Diagnostics Check	OK	06:11
Air Blank	0.000	06:11
Control Test	0.080	06:12
Air Blank	0.000	06:12
Subject Sample # 1	0.096	06:13
Breath Volume	1.140	06:13
Air Blank	0.000	06:14
Air Blank	0.000	06:15
Subject Sample # 2	0.094	06:16
Breath Volume	1.167	06:16
Air Blank	0.000	06:17
Control Test	0.080	06:17
Air Blank	0.000	06:18
Diagnostics Check	OK	06:18

---

Data Download Date/Time: 01/13/2022 13:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OWEN,SUE,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: PALMER,GLENN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 09:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: ACKERMAN

DATE/TIME OF TEST: 12/28/2021 10:53:30

---

Diagnostics Check	OK	10:47
Air Blank	0.000	10:47
Control Test	0.080	10:48
Air Blank	0.000	10:48
Subject Sample # 1	0.000	10:49
Breath Volume	2.972	10:49
Air Blank	0.000	10:49
Air Blank	0.000	10:51
Subject Sample # 2	0.000	10:52
Breath Volume	2.343	10:52
Air Blank	0.000	10:52
Control Test	0.080	10:52
Air Blank	0.000	10:53
Diagnostics Check	OK	10:53

---

Data Download Date/Time: 01/13/2022 13:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: LOVETT,JAMES  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: GOLDSTEIN

DATE/TIME OF TEST: 12/30/2021 04:01:22

---

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:49	
Control Test	0.079	03:50	
Air Blank	0.000	03:50	
Subject Sample	# 1	03:53	No Sample Provided
Breath Volume	0.000	03:53	
Air Blank	0.000	03:54	
Air Blank	0.000	03:56	
Subject Sample	# 2	0.176	03:56
Breath Volume	1.328	03:56	
Air Blank	0.000	03:57	
Air Blank	0.000	03:59	
Subject Sample	# 3	0.178	03:59
Breath Volume	1.351	03:59	
Air Blank	0.000	04:00	
Control Test	0.078	04:00	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

---

Data Download Date/Time: 01/13/2022 13:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WHITE,EDWARD  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: POSNICK

DATE/TIME OF TEST: 12/30/2021 23:28:00

---

Diagnostics Check	OK	23:13	
Air Blank	0.000	23:14	
Control Test	0.080	23:14	
Air Blank	0.000	23:15	
Subject Sample	# 1 0.139	23:16	
Breath Volume	1.136	23:16	
Air Blank	0.000	23:17	
Air Blank	0.000	23:19	
Subject Sample	# 2 0.131	23:22	Volume Not Met
Breath Volume	0.906	23:22	
Air Blank	0.000	23:23	
Air Blank	0.000	23:25	
Subject Sample	# 3 0.133	23:26	
Breath Volume	1.386	23:26	
Air Blank	0.000	23:27	
Control Test	0.077	23:27	
Air Blank	0.000	23:27	
Diagnostics Check	OK	23:28	

---

Data Download Date/Time: 01/13/2022 13:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SPRAROW,BRADLEY  
ARRESTING OFFICER AGENCY: Riviera Beach PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: CRUZ PEREZ

DATE/TIME OF TEST: 12/31/2021 01:14:38

---

Diagnostics Check		OK	01:06
Air Blank		0.000	01:07
Control Test		0.079	01:07
Air Blank		0.000	01:07
Subject Sample	# 1	0.224	01:10
Breath Volume		2.156	01:10
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.215	01:12
Breath Volume		3.109	01:12
Air Blank		0.000	01:13
Control Test		0.077	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

---

Data Download Date/Time: 01/13/2022 13:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MARTINEZ,JACOB  
ARRESTING OFFICER AGENCY: West Palm Beach PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ELAM

DATE/TIME OF TEST: 12/31/2021 02:37:06

---

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.079	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.178	02:29
Breath Volume	1.613	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.170	02:35
Breath Volume	1.468	02:35
Air Blank	0.000	02:36
Control Test	0.078	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

---

Data Download Date/Time: 01/13/2022 13:32



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: ROBINSON,HAROLD  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: MARUQUIN CHAVEZ

DATE/TIME OF TEST: 12/31/2021 03:46:57

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.079	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.139	03:42
Breath Volume	2.867	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:44
Subject Sample # 2	0.138	03:45
Breath Volume	2.777	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

---

Data Download Date/Time: 01/13/2022 13:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: HUDSON,DILLAN  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 14:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/31/2021 15:56:49

---

Diagnostics Check	OK	15:50
Air Blank	0.000	15:50
Control Test	0.080	15:50
Air Blank	0.000	15:51
Subject Sample # 1	0.013	15:52
Breath Volume	1.757	15:52
Air Blank	0.000	15:52
Air Blank	0.000	15:54
Subject Sample # 2	0.012	15:55
Breath Volume	2.171	15:55
Air Blank	0.000	15:55
Control Test	0.079	15:56
Air Blank	0.000	15:56
Diagnostics Check	OK	15:56

---

Data Download Date/Time: 01/13/2022 13:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MCGILLICUDDY,STEVEN  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19021080A2  
DRY GAS EXPIRY DATE: 09/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: MUBARAK

DATE/TIME OF TEST: 12/31/2021 23:45:11

---

Diagnostics Check	OK	23:38
Air Blank	0.000	23:39
Control Test	0.079	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.160	23:40
Breath Volume	1.898	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.152	23:43
Breath Volume	1.660	23:43
Air Blank	0.000	23:44
Control Test	0.080	23:44
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

---

Data Download Date/Time: 01/13/2022 13:32