

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESTON, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Winter Springs PD
ARRESTING OFFICER: PRESTON, WILLIAM
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 11/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/01/2021 21:14:08

Diagnostics Check	OK	21:08
Air Blank	0.000	21:08
Control Test	0.080	21:08
Air Blank	0.000	21:09
Subject Sample # 1	0.220	21:09
Breath Volume	1.750	21:09
Air Blank	0.000	21:10
Air Blank	0.000	21:12
Subject Sample # 2	0.222	21:12
Breath Volume	2.070	21:12
Air Blank	0.000	21:13
Control Test	0.080	21:13
Air Blank	0.000	21:14
Diagnostics Check	OK	21:14

Data Download Date/Time: 12/09/2021 02:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CANADY,C
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: KEENER

DATE/TIME OF TEST: 12/02/2021 02:25:18

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.211	02:19
Breath Volume	1.304	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:21
Subject Sample # 2	0.217	02:23
Breath Volume	2.460	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 12/09/2021 02:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: UCELO,BRYAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 12/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 12/09/2021 03:18:41

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.079	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.156	03:12
Breath Volume	2.570	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.138	03:17
Breath Volume	1.433	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 02/02/2022 22:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: UCELO,BRYAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 12/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: JAVIER

DATE/TIME OF TEST: 12/10/2021 02:54:56

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.079	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.104	02:50
Breath Volume	1.714	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.109	02:53
Breath Volume	1.476	02:53
Air Blank	0.000	02:54
Control Test	0.079	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 02/02/2022 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUPA,KYLE,T
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: BELT,WILLIAM
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 12/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: COKE

DATE/TIME OF TEST: 12/10/2021 19:44:13

Diagnostics Check	OK	19:41	
Air Blank	0.000	19:41	
Control Test	0.080	19:41	
Air Blank	0.000	19:42	
Subject Sample	# 1	19:42	Subject Test Refused
Breath Volume	0.000	19:42	
Air Blank	0.000	19:43	
Control Test	0.080	19:43	
Air Blank	0.000	19:44	
Diagnostics Check	OK	19:44	

Data Download Date/Time: 02/02/2022 22:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: NOTARIO,FREDERICK
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 12/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: HITZING

DATE/TIME OF TEST: 12/10/2021 22:56:10

Diagnostics Check	OK	22:47
Air Blank	0.000	22:48
Control Test	0.079	22:48
Air Blank	0.000	22:49
Subject Sample # 1	0.198	22:50
Breath Volume	1.359	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.190	22:54
Breath Volume	1.820	22:54
Air Blank	0.000	22:55
Control Test	0.079	22:55
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 02/02/2022 22:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CHENOWETH,DEREK
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 12/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: PAGAN

DATE/TIME OF TEST: 12/10/2021 23:51:00

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.080	23:42
Air Blank	0.000	23:43
Subject Sample # 1	0.008	23:46
Breath Volume	1.605	23:46
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.000	23:49
Breath Volume	1.289	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:51

Data Download Date/Time: 02/02/2022 22:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DASQUE,DARREN
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 12/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: PALOMAR

DATE/TIME OF TEST: 12/11/2021 03:02:47

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.079	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.224	02:58
Breath Volume	2.160	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.227	03:01
Breath Volume	2.046	03:01
Air Blank	0.000	03:01
Control Test	0.080	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 02/02/2022 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: REYES-CENTENO,SERGIO
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 12/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: MC GRIFF

DATE/TIME OF TEST: 12/11/2021 23:24:39

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.080	23:18
Air Blank	0.000	23:19
Subject Sample # 1	0.135	23:20
Breath Volume	3.226	23:20
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.130	23:23
Breath Volume	2.695	23:23
Air Blank	0.000	23:23
Control Test	0.080	23:24
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 02/02/2022 22:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SCHWARTZ,E
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 12/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: BISCHOF

DATE/TIME OF TEST: 12/12/2021 01:52:52

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.080	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.122	01:48
Breath Volume	2.835	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.124	01:51
Breath Volume	2.542	01:51
Air Blank	0.000	01:51
Control Test	0.079	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 02/02/2022 22:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CHAMBERS,AARON
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 12/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: RUSH JR

DATE/TIME OF TEST: 12/12/2021 04:20:47

Diagnostics Check	OK	04:14
Air Blank	0.000	04:14
Control Test	0.080	04:15
Air Blank	0.000	04:15
Subject Sample # 1	0.157	04:16
Breath Volume	1.972	04:16
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.156	04:19
Breath Volume	1.468	04:19
Air Blank	0.000	04:19
Control Test	0.079	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 02/02/2022 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CANADY,CHRISTINA
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 12/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CAMPOS DELGADO

DATE/TIME OF TEST: 12/12/2021 04:42:47

Diagnostics Check	OK	04:36
Air Blank	0.000	04:37
Control Test	0.079	04:37
Air Blank	0.000	04:37
Subject Sample # 1	0.128	04:38
Breath Volume	2.312	04:38
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.132	04:41
Breath Volume	2.019	04:41
Air Blank	0.000	04:41
Control Test	0.080	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 02/02/2022 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GERVAIS, ROSEMARY, C
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ROSS, J
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 12/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 12/17/2021 19:10:51

Diagnostics Check	OK	19:04
Air Blank	0.000	19:04
Control Test	0.080	19:05
Air Blank	0.000	19:05
Subject Sample # 1	0.251	19:06
Breath Volume	2.750	19:06
Air Blank	0.000	19:06
Air Blank	0.000	19:08
Subject Sample # 2	0.238	19:09
Breath Volume	1.988	19:09
Air Blank	0.000	19:09
Control Test	0.080	19:10
Air Blank	0.000	19:10
Diagnostics Check	OK	19:10

Data Download Date/Time: 02/02/2022 22:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KEE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 12/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 12/18/2021 04:33:24

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.081	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.166	04:29
Breath Volume	2.492	04:29
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.163	04:31
Breath Volume	1.253	04:31
Air Blank	0.000	04:32
Control Test	0.080	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 02/02/2022 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: HARTZELL,E
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 12/15/2021 21:41:23

Diagnostics Check	OK	21:38	
Air Blank	0.000	21:38	
Control Test	0.081	21:39	
Air Blank	0.000	21:39	
Subject Sample	# 1	21:40	Subject Test Refused
Breath Volume	0.000	21:40	
Air Blank	0.000	21:40	
Control Test	0.081	21:40	
Air Blank	0.000	21:41	
Diagnostics Check	OK	21:41	

Data Download Date/Time: 01/06/2022 21:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWENS,STEVE,E
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: WHITE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1327124
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: HERSHMAN

DATE/TIME OF TEST: 12/03/2021 20:45:26

Diagnostics Check	OK	20:39
Air Blank	0.000	20:39
Control Test	0.078	20:39
Air Blank	0.000	20:40
Subject Sample # 1	0.158	20:40
Breath Volume	3.781	20:40
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.155	20:43
Breath Volume	3.843	20:43
Air Blank	0.000	20:44
Control Test	0.077	20:44
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 12/23/2021 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: KANE,NOAH
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1327124
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: ROCKWELL

DATE/TIME OF TEST: 12/06/2021 01:49:34

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.077	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.000	01:45
Breath Volume	2.289	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.000	01:48
Breath Volume	2.355	01:48
Air Blank	0.000	01:48
Control Test	0.076	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 12/23/2021 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,ERIC,G
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: WALLER,ARTHUR
ARRESTING OFFICER AGENCY: Patrick AFB PD
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1327124
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 08:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:15

SUBJECT'S LAST NAME: DELAROSA

DATE/TIME OF TEST: 12/06/2021 09:15:56

Diagnostics Check	OK	09:09
Air Blank	0.000	09:09
Control Test	0.077	09:10
Air Blank	0.000	09:10
Subject Sample # 1	0.233	09:11
Breath Volume	1.597	09:11
Air Blank	0.000	09:12
Air Blank	0.000	09:13
Subject Sample # 2	0.236	09:14
Breath Volume	2.621	09:14
Air Blank	0.000	09:15
Control Test	0.076	09:15
Air Blank	0.000	09:15
Diagnostics Check	OK	09:15

Data Download Date/Time: 12/23/2021 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWENS,STEVE,E
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: LARSON,DREW
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1327124
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: ARSENAULT

DATE/TIME OF TEST: 12/07/2021 00:34:23

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.077	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.186	00:29
Breath Volume	1.929	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.193	00:32
Breath Volume	1.308	00:32
Air Blank	0.000	00:33
Control Test	0.077	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 12/23/2021 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,ERIC,G
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: LARSON,BEN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1327124
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 13:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:56

SUBJECT'S LAST NAME: SHEEKEY

DATE/TIME OF TEST: 12/11/2021 17:23:39

Diagnostics Check	OK	17:16
Air Blank	0.000	17:17
Control Test	0.077	17:17
Air Blank	0.000	17:18
Subject Sample # 1	0.259	17:18
Breath Volume	1.558	17:18
Air Blank	0.000	17:19
Air Blank	0.000	17:21
Subject Sample # 2	0.242	17:22
Breath Volume	1.320	17:22
Air Blank	0.000	17:22
Control Test	0.077	17:23
Air Blank	0.000	17:23
Diagnostics Check	OK	17:23

Data Download Date/Time: 12/23/2021 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOUSMAN,JOHN,D
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: LARSON,DREW
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1327124
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: CRONIN

DATE/TIME OF TEST: 12/13/2021 01:33:15

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.077	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.194	01:28
Breath Volume	2.390	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.190	01:31
Breath Volume	2.800	01:31
Air Blank	0.000	01:32
Control Test	0.076	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 12/23/2021 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRANN,SEAN,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: WHITE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: KOENIG

DATE/TIME OF TEST: 12/24/2021 00:48:03

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.080	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.302	00:43
Breath Volume	1.804	00:43
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.303	00:46
Breath Volume	2.707	00:46
Air Blank	0.000	00:47
Control Test	0.080	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:48

Data Download Date/Time: 01/16/2022 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: YBARRA,AIMEE
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: CHRISTOFFEL

DATE/TIME OF TEST: 12/29/2021 05:09:09

Diagnostics Check	OK	05:02
Air Blank	0.000	05:03
Control Test	0.080	05:03
Air Blank	0.000	05:04
Subject Sample # 1	0.000	05:04
Breath Volume	1.574	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.000	05:07
Breath Volume	2.199	05:07
Air Blank	0.000	05:08
Control Test	0.079	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 01/16/2022 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: FRAZIER,FREDDIE
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 11/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: VILLEGAS

DATE/TIME OF TEST: 12/11/2021 01:27:24

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.078	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.217	01:22
Breath Volume	2.265	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.235	01:25
Breath Volume	2.867	01:25
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 01/30/2022 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: TYLER,ETHAN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 11/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 12:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: ANAYA

DATE/TIME OF TEST: 12/12/2021 02:01:27

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:00	
Control Test	0.064	02:00	Control Outside Tolerance
Air Blank	0.000	02:01	

Data Download Date/Time: 01/30/2022 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: TYLER,ETHAN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 11/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 12:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: ANAYA

DATE/TIME OF TEST: 12/12/2021 02:09:52

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.077	02:04
Air Blank	0.000	02:04
Subject Sample # 1	0.124	02:05
Breath Volume	2.746	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:07
Subject Sample # 2	0.119	02:08
Breath Volume	2.851	02:08
Air Blank	0.000	02:08
Control Test	0.077	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 01/30/2022 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JESSICA,N
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: PEREZ,JONATHAN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 11/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 15:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: PEREZ LOPEZ

DATE/TIME OF TEST: 12/12/2021 17:01:51

Diagnostics Check	OK	16:50	
Air Blank	0.000	16:50	
Control Test	0.079	16:51	
Air Blank	0.000	16:51	
Subject Sample	# 1 0.223	16:54	Volume Not Met
Breath Volume	0.035	16:54	
Air Blank	0.000	16:55	
Air Blank	0.000	16:57	
Subject Sample	# 2 0.265	17:00	Volume Not Met
Breath Volume	0.156	17:00	
Air Blank	0.000	17:00	
Control Test	0.077	17:01	
Air Blank	0.000	17:01	
Diagnostics Check	OK	17:01	

Data Download Date/Time: 01/30/2022 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: RAMQUIST

DATE/TIME OF TEST: 12/03/2021 02:50:40

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:47	
Control Test	0.081	02:48	
Air Blank	0.000	02:48	
Subject Sample	# 1	02:49	Subject Test Refused
Breath Volume	0.000	02:49	
Air Blank	0.000	02:49	
Control Test	0.081	02:50	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

Data Download Date/Time: 12/13/2021 10:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: HOANG

DATE/TIME OF TEST: 12/03/2021 03:00:44

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.153	02:56
Breath Volume	2.085	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.156	02:59
Breath Volume	2.039	02:59
Air Blank	0.000	02:59
Control Test	0.080	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 12/13/2021 10:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: MACDONALD

DATE/TIME OF TEST: 12/03/2021 20:05:17

Diagnostics Check	OK	20:02	
Air Blank	0.000	20:02	
Control Test	0.081	20:02	
Air Blank	0.000	20:03	
Subject Sample	# 1	20:03	Subject Test Refused
Breath Volume	0.000	20:03	
Air Blank	0.000	20:04	
Control Test	0.081	20:04	
Air Blank	0.000	20:05	
Diagnostics Check	OK	20:05	

Data Download Date/Time: 12/13/2021 10:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: DUKES

DATE/TIME OF TEST: 12/05/2021 04:13:19

Diagnostics Check	OK	04:06
Air Blank	0.000	04:07
Control Test	0.081	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.144	04:08
Breath Volume	2.691	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.145	04:11
Breath Volume	2.265	04:11
Air Blank	0.000	04:12
Control Test	0.081	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 12/13/2021 10:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HICKS,JASON
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: DEANDA

DATE/TIME OF TEST: 12/06/2021 01:01:18

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:58	
Control Test	0.081	00:59	
Air Blank	0.000	00:59	
Subject Sample	# 1	00:59	Subject Test Refused
Breath Volume	0.000	00:59	
Air Blank	0.000	01:00	
Control Test	0.081	01:00	
Air Blank	0.000	01:01	
Diagnostics Check	OK	01:01	

Data Download Date/Time: 12/13/2021 10:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRILL,SHANNON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: ZIVIANI GEROLAMO

DATE/TIME OF TEST: 12/06/2021 02:29:42

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:26	
Control Test	0.081	02:26	
Air Blank	0.000	02:27	
Subject Sample	# 1	02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:28	
Control Test	0.081	02:29	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

Data Download Date/Time: 12/13/2021 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: GUTIERREZ JUAREZ

DATE/TIME OF TEST: 12/17/2021 03:11:11

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.081	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.137	03:06
Breath Volume	2.402	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.137	03:09
Breath Volume	2.019	03:09
Air Blank	0.000	03:10
Control Test	0.081	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 01/07/2022 18:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/19/2021 04:46:37

Diagnostics Check	OK	04:39
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:41
Subject Sample # 1	0.126	04:41
Breath Volume	1.769	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.131	04:45
Breath Volume	2.054	04:45
Air Blank	0.000	04:45
Control Test	0.080	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 01/07/2022 18:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 12/26/2021 04:01:06

Diagnostics Check	OK	03:58	
Air Blank	0.000	03:58	
Control Test	0.081	03:59	
Air Blank	0.000	03:59	
Subject Sample	# 1	03:59	Subject Test Refused
Breath Volume	0.000	03:59	
Air Blank	0.000	04:00	
Control Test	0.080	04:00	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

Data Download Date/Time: 01/07/2022 18:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/31/2021 03:35:56

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.081	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.215	03:31
Breath Volume	2.964	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.216	03:34
Breath Volume	2.410	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 01/07/2022 18:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, TYLER, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BAUERNFEIND, BRANDON
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 11/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: GLADNICK

DATE/TIME OF TEST: 12/02/2021 04:20:45

Diagnostics Check	OK	04:17	
Air Blank	0.000	04:18	
Control Test	0.079	04:18	
Air Blank	0.000	04:19	
Subject Sample	# 1	04:19	Subject Test Refused
Breath Volume	0.000	04:19	
Air Blank	0.000	04:19	
Control Test	0.078	04:20	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:20	

Data Download Date/Time: 12/29/2021 11:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GREEN,KATELYN
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 11/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 12/04/2021 20:03:19

Diagnostics Check	OK	20:00	
Air Blank	0.000	20:01	
Control Test	0.079	20:01	
Air Blank	0.000	20:02	
Subject Sample	# 1	20:02	RFI Detect
Breath Volume	0.000	20:02	
Air Blank		20:02	RFI Detect
Air Blank	0.000	20:03	

Data Download Date/Time: 12/29/2021 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GREEN,KATELYN
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 11/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 12/04/2021 20:10:27

Diagnostics Check	OK	20:05	
Air Blank	0.000	20:06	
Control Test	0.075	20:06	
Air Blank	0.000	20:07	
Subject Sample	# 1 0.289	20:09	
Breath Volume	1.121	20:09	
Air Blank		20:09	RFI Detect
Air Blank		20:09	RFI Detect
Air Blank	0.000	20:10	

Data Download Date/Time: 12/29/2021 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GREEN,KATELYN
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 11/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 12/04/2021 20:13:41

Diagnostics Check	OK	20:12	
Air Blank	0.000	20:12	
Control Test	0.074	20:13	Control Outside Tolerance
Air Blank	0.000	20:13	

Data Download Date/Time: 12/29/2021 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GREEN,KATELYN
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 11/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 12/04/2021 20:26:58

Diagnostics Check	OK	20:15	
Air Blank	0.000	20:16	
Control Test	0.075	20:16	
Air Blank	0.000	20:17	
Subject Sample	# 1 0.288	20:18	
Breath Volume	1.132	20:18	
Air Blank	0.000	20:19	
Air Blank	0.000	20:21	
Subject Sample	# 2 0.264	20:24	Volume Not Met
Breath Volume	0.984	20:24	
Air Blank	0.000	20:25	
Air Blank		20:26	RFI Detect
Air Blank		20:26	RFI Detect
Air Blank		20:26	RFI Detect

Data Download Date/Time: 12/29/2021 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GREEN,KATELYN
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 11/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 12/04/2021 20:30:29

Diagnostics Check	OK	20:29	
Air Blank	0.000	20:29	
Control Test	0.074	20:29	Control Outside Tolerance
Air Blank	0.000	20:30	

Data Download Date/Time: 12/29/2021 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GREEN,KATELYN
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 11/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 12/04/2021 20:46:27

Diagnostics Check	OK	20:39
Air Blank	0.000	20:39
Control Test	0.077	20:40
Air Blank	0.000	20:40
Subject Sample # 1	0.299	20:41
Breath Volume	1.535	20:41
Air Blank	0.000	20:42
Air Blank	0.000	20:44
Subject Sample # 2	0.296	20:44
Breath Volume	1.464	20:44
Air Blank	0.000	20:45
Control Test	0.077	20:45
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

Data Download Date/Time: 12/29/2021 11:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLACRES,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: KENNEDY,KELLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ULRICH

DATE/TIME OF TEST: 12/05/2021 04:31:59

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:19	
Control Test	0.079	04:20	
Air Blank	0.000	04:20	
Subject Sample	# 1 0.186	04:23	Volume Not Met
Breath Volume	1.085	04:23	
Air Blank	0.000	04:24	
Air Blank	0.000	04:26	
Subject Sample	# 2 0.194	04:27	
Breath Volume	1.718	04:27	
Air Blank	0.000	04:27	
Air Blank	0.000	04:29	
Subject Sample	# 3 0.184	04:30	
Breath Volume	1.195	04:30	
Air Blank	0.000	04:31	
Control Test	0.077	04:31	
Air Blank	0.000	04:31	
Diagnostics Check	OK	04:31	

Data Download Date/Time: 12/29/2021 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINSTEAD,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: PORTOCARRERO,CHRISTIAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 12/08/2021 23:50:27

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.078	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.215	23:45
Breath Volume	2.398	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.214	23:48
Breath Volume	1.808	23:48
Air Blank	0.000	23:49
Control Test	0.077	23:49
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 12/29/2021 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTERIO,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: REILLY,PAUL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/16/2021 01:28:57

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.078	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.180	01:24
Breath Volume	2.746	01:24
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.180	01:27
Breath Volume	2.957	01:27
Air Blank	0.000	01:28
Control Test	0.078	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 12/29/2021 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,ZACHARY,W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ROBINSON,ZACHARY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: OTHER
ARREST TIME: 11:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 12/21/2021 12:30:46

Diagnostics Check	OK	12:21
Air Blank	0.000	12:21
Control Test	0.079	12:21
Air Blank	0.000	12:22
Subject Sample # 1	0.341	12:23
Breath Volume	1.425	12:23
Air Blank	0.000	12:23
Air Blank	0.000	12:25
Subject Sample # 2	0.320	12:26
Breath Volume	1.406	12:26
Air Blank	0.000	12:26
Air Blank	0.000	12:28
Subject Sample # 3	0.320	12:29
Breath Volume	1.191	12:29
Air Blank	0.000	12:29
Control Test	0.079	12:30
Air Blank	0.000	12:30
Diagnostics Check	OK	12:30

Data Download Date/Time: 12/29/2021 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,F
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WAGNER,ROBERT
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:22

SUBJECT'S LAST NAME: T512101923610

DATE/TIME OF TEST: 12/23/2021 08:04:15

Diagnostics Check	OK	07:49	
Air Blank	0.000	07:49	
Control Test	0.078	07:49	
Air Blank	0.000	07:50	
Subject Sample	# 1 0.111	07:53	Volume Not Met
Breath Volume	0.433	07:53	
Air Blank	0.000	07:54	
Air Blank	0.000	07:56	
Subject Sample	# 2 0.115	07:57	
Breath Volume	1.269	07:57	
Air Blank	0.000	07:57	
Air Blank	0.000	07:59	
Subject Sample	# 3 0.103	08:02	Volume Not Met
Breath Volume	0.312	08:02	
Air Blank	0.000	08:03	
Control Test	0.077	08:03	
Air Blank	0.000	08:04	
Diagnostics Check	OK	08:04	

Data Download Date/Time: 12/29/2021 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREWS,M,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MANCINI,J
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:46

SUBJECT'S LAST NAME: PROVENCHER

DATE/TIME OF TEST: 12/23/2021 15:42:22

Diagnostics Check	OK	15:36
Air Blank	0.000	15:36
Control Test	0.079	15:37
Air Blank	0.000	15:37
Subject Sample # 1	0.000	15:38
Breath Volume	2.011	15:38
Air Blank	0.000	15:38
Air Blank	0.000	15:40
Subject Sample # 2	0.000	15:41
Breath Volume	2.808	15:41
Air Blank	0.000	15:41
Control Test	0.078	15:41
Air Blank	0.000	15:42
Diagnostics Check	OK	15:42

Data Download Date/Time: 12/29/2021 11:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTERIO,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CHRISTOFFEL,JAMES
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: TOOLE

DATE/TIME OF TEST: 12/25/2021 22:22:52

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:20	
Control Test	0.078	22:20	
Air Blank	0.000	22:21	
Subject Sample	# 1	22:21	Subject Test Refused
Breath Volume	0.000	22:21	
Air Blank	0.000	22:21	
Control Test	0.078	22:22	
Air Blank	0.000	22:22	
Diagnostics Check	OK	22:22	

Data Download Date/Time: 12/29/2021 11:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: HEATLEY

DATE/TIME OF TEST: 12/04/2021 02:02:32

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.080	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.219	01:58
Breath Volume	2.156	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.219	02:00
Breath Volume	1.804	02:00
Air Blank	0.000	02:01
Control Test	0.079	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 12/13/2021 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 12/04/2021 03:59:29

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:57	
Control Test	0.080	03:57	
Air Blank	0.000	03:57	
Subject Sample	# 1	03:58	Subject Test Refused
Breath Volume	0.000	03:58	
Air Blank	0.000	03:58	
Control Test	0.079	03:58	
Air Blank	0.000	03:59	
Diagnostics Check	OK	03:59	

Data Download Date/Time: 12/13/2021 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: DUEHRING

DATE/TIME OF TEST: 12/09/2021 01:13:21

Diagnostics Check	OK	01:09	
Air Blank	0.000	01:09	
Control Test	0.079	01:10	
Air Blank	0.000	01:10	
Subject Sample	# 1	01:11	Subject Test Refused
Breath Volume	0.000	01:11	
Air Blank	0.000	01:12	
Control Test	0.079	01:12	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 12/13/2021 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: HOUSLEY

DATE/TIME OF TEST: 12/09/2021 02:12:06

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:09	
Control Test	0.079	02:10	
Air Blank	0.000	02:10	
Subject Sample	# 1	02:10	Subject Test Refused
Breath Volume	0.000	02:10	
Air Blank	0.000	02:11	
Control Test	0.078	02:11	
Air Blank	0.000	02:11	
Diagnostics Check	OK	02:12	

Data Download Date/Time: 12/13/2021 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: RASE

DATE/TIME OF TEST: 12/09/2021 22:54:14

Diagnostics Check	OK	22:48
Air Blank	0.000	22:48
Control Test	0.079	22:48
Air Blank	0.000	22:49
Subject Sample # 1	0.126	22:49
Breath Volume	1.398	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.123	22:52
Breath Volume	1.480	22:52
Air Blank	0.000	22:53
Control Test	0.079	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 12/13/2021 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 12/10/2021 01:08:49

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.080	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.000	01:04
Breath Volume	2.527	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.000	01:07
Breath Volume	2.917	01:07
Air Blank	0.000	01:07
Control Test	0.079	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 12/13/2021 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: MCKAY

DATE/TIME OF TEST: 12/11/2021 01:19:29

Diagnostics Check	OK	01:16	
Air Blank	0.000	01:17	
Control Test	0.080	01:17	
Air Blank	0.000	01:17	
Subject Sample	# 1	01:18	Subject Test Refused
Breath Volume	0.000	01:18	
Air Blank	0.000	01:18	
Control Test	0.079	01:18	
Air Blank	0.000	01:19	
Diagnostics Check	OK	01:19	

Data Download Date/Time: 12/13/2021 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: HOVER

DATE/TIME OF TEST: 12/11/2021 02:01:35

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.206	01:57
Breath Volume	2.039	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.200	02:00
Breath Volume	1.480	02:00
Air Blank	0.000	02:00
Control Test	0.079	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 12/13/2021 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 12/11/2021 02:55:54

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.079	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.024	02:50
Breath Volume	1.394	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.022	02:54
Breath Volume	1.453	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 12/13/2021 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: KOZMIUK

DATE/TIME OF TEST: 12/18/2021 22:28:49

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.079	22:23
Air Blank	0.000	22:23
Subject Sample # 1	0.153	22:24
Breath Volume	1.472	22:24
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.146	22:27
Breath Volume	2.546	22:27
Air Blank	0.000	22:27
Control Test	0.079	22:28
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 01/06/2022 22:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: GONZALEZPEREZ

DATE/TIME OF TEST: 12/18/2021 23:40:08

Diagnostics Check	OK	23:36	
Air Blank	0.000	23:37	
Control Test	0.079	23:37	
Air Blank	0.000	23:37	
Subject Sample	# 1	23:38	Subject Test Refused
Breath Volume	0.000	23:38	
Air Blank	0.000	23:39	
Control Test	0.079	23:39	
Air Blank	0.000	23:40	
Diagnostics Check	OK	23:40	

Data Download Date/Time: 01/06/2022 22:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: BASSEY

DATE/TIME OF TEST: 12/19/2021 01:06:03

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:03	
Control Test	0.079	01:03	
Air Blank	0.000	01:04	
Subject Sample	# 1	01:04	Subject Test Refused
Breath Volume	0.000	01:04	
Air Blank	0.000	01:05	
Control Test	0.079	01:05	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 01/06/2022 22:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 12/19/2021 02:38:49

Diagnostics Check	OK	02:37	
Air Blank		02:38	RFI Detect
Air Blank	0.000	02:38	

Data Download Date/Time: 01/06/2022 22:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 12/19/2021 02:43:12

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:40	
Control Test	0.079	02:41	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:41	Subject Test Refused
Breath Volume	0.000	02:41	
Air Blank	0.000	02:42	
Control Test	0.080	02:42	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 01/06/2022 22:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 12/04/2021 06:36:06

Diagnostics Check	OK	06:28
Air Blank	0.000	06:28
Control Test	0.081	06:29
Air Blank	0.000	06:29
Subject Sample # 1	0.194	06:31
Breath Volume	3.777	06:31
Air Blank	0.000	06:32
Air Blank	0.000	06:34
Subject Sample # 2	0.201	06:34
Breath Volume	2.980	06:34
Air Blank	0.000	06:35
Control Test	0.081	06:35
Air Blank	0.000	06:36
Diagnostics Check	OK	06:36

Data Download Date/Time: 01/01/2022 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:25

SUBJECT'S LAST NAME: SCARBOROUGH

DATE/TIME OF TEST: 12/09/2021 14:02:43

Diagnostics Check	OK	13:56
Air Blank	0.000	13:57
Control Test	0.081	13:57
Air Blank	0.000	13:58
Subject Sample # 1	0.000	13:58
Breath Volume	2.738	13:58
Air Blank	0.000	13:58
Air Blank	0.000	14:00
Subject Sample # 2	0.000	14:01
Breath Volume	2.710	14:01
Air Blank	0.000	14:01
Control Test	0.080	14:02
Air Blank	0.000	14:02
Diagnostics Check	OK	14:02

Data Download Date/Time: 01/01/2022 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: DESIR

DATE/TIME OF TEST: 12/09/2021 14:17:36

Diagnostics Check	OK	14:11
Air Blank	0.000	14:11
Control Test	0.080	14:12
Air Blank	0.000	14:12
Subject Sample # 1	0.000	14:13
Breath Volume	2.406	14:13
Air Blank	0.000	14:13
Air Blank	0.000	14:15
Subject Sample # 2	0.000	14:16
Breath Volume	2.550	14:16
Air Blank	0.000	14:16
Control Test	0.081	14:16
Air Blank	0.000	14:17
Diagnostics Check	OK	14:17

Data Download Date/Time: 01/01/2022 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: PUPO

DATE/TIME OF TEST: 12/09/2021 14:33:39

Diagnostics Check	OK	14:27
Air Blank	0.000	14:27
Control Test	0.081	14:28
Air Blank	0.000	14:28
Subject Sample # 1	0.000	14:29
Breath Volume	2.792	14:29
Air Blank	0.000	14:29
Air Blank	0.000	14:31
Subject Sample # 2	0.000	14:32
Breath Volume	2.710	14:32
Air Blank	0.000	14:32
Control Test	0.080	14:33
Air Blank	0.000	14:33
Diagnostics Check	OK	14:33

Data Download Date/Time: 01/01/2022 21:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:05

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 12/10/2021 10:01:22

Diagnostics Check	OK	09:55
Air Blank	0.000	09:55
Control Test	0.080	09:55
Air Blank	0.000	09:56
Subject Sample # 1	0.000	09:57
Breath Volume	2.738	09:57
Air Blank	0.000	09:57
Air Blank	0.000	09:59
Subject Sample # 2	0.000	09:59
Breath Volume	2.710	09:59
Air Blank	0.000	10:00
Control Test	0.080	10:00
Air Blank	0.000	10:01
Diagnostics Check	OK	10:01

Data Download Date/Time: 01/01/2022 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:25

SUBJECT'S LAST NAME: MC ERLEAN

DATE/TIME OF TEST: 12/10/2021 10:14:57

Diagnostics Check	OK	10:09
Air Blank	0.000	10:09
Control Test	0.081	10:09
Air Blank	0.000	10:10
Subject Sample # 1	0.000	10:10
Breath Volume	1.375	10:10
Air Blank	0.000	10:11
Air Blank	0.000	10:13
Subject Sample # 2	0.000	10:13
Breath Volume	1.335	10:13
Air Blank	0.000	10:14
Control Test	0.080	10:14
Air Blank	0.000	10:14
Diagnostics Check	OK	10:14

Data Download Date/Time: 01/01/2022 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: GRIECO

DATE/TIME OF TEST: 12/18/2021 02:38:09

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.080	02:32
Air Blank	0.000	02:33
Subject Sample # 1	0.217	02:33
Breath Volume	1.343	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.208	02:36
Breath Volume	1.140	02:36
Air Blank	0.000	02:37
Control Test	0.081	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 01/01/2022 21:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: LAFONTANT

DATE/TIME OF TEST: 12/19/2021 23:17:58

Diagnostics Check	OK	23:15	
Air Blank	0.000	23:15	
Control Test	0.081	23:15	
Air Blank	0.000	23:16	
Subject Sample	# 1	23:16	Subject Test Refused
Breath Volume	0.000	23:16	
Air Blank	0.000	23:17	
Control Test	0.080	23:17	
Air Blank	0.000	23:17	
Diagnostics Check	OK	23:17	

Data Download Date/Time: 01/01/2022 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: CARLO

DATE/TIME OF TEST: 12/22/2021 21:06:35

Diagnostics Check	OK	21:00
Air Blank	0.000	21:00
Control Test	0.081	21:00
Air Blank	0.000	21:01
Subject Sample # 1	0.000	21:02
Breath Volume	2.039	21:02
Air Blank	0.000	21:02
Air Blank	0.000	21:04
Subject Sample # 2	0.000	21:05
Breath Volume	1.488	21:05
Air Blank	0.000	21:05
Control Test	0.081	21:05
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

Data Download Date/Time: 01/01/2022 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: GILMAN,DUSTIN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: DAVILA MORALES

DATE/TIME OF TEST: 12/23/2021 00:10:23

Diagnostics Check	OK	00:03
Air Blank	0.000	00:04
Control Test	0.080	00:04
Air Blank	0.000	00:05
Subject Sample # 1	0.180	00:05
Breath Volume	1.511	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.178	00:08
Breath Volume	2.093	00:08
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 01/01/2022 21:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: ESPARRA RIOS

DATE/TIME OF TEST: 12/28/2021 00:30:13

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:27	
Control Test	0.081	00:28	
Air Blank	0.000	00:28	
Subject Sample	# 1	00:28	Subject Test Refused
Breath Volume	0.000	00:28	
Air Blank	0.000	00:29	
Control Test	0.081	00:29	
Air Blank	0.000	00:30	
Diagnostics Check	OK	00:30	

Data Download Date/Time: 01/01/2022 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: HEMPLALL

DATE/TIME OF TEST: 12/28/2021 00:35:08

Diagnostics Check	OK	00:32	
Air Blank	0.000	00:32	
Control Test	0.081	00:33	
Air Blank	0.000	00:33	
Subject Sample	# 1	00:33	Subject Test Refused
Breath Volume	0.000	00:33	
Air Blank	0.000	00:34	
Control Test	0.080	00:34	
Air Blank	0.000	00:35	
Diagnostics Check	OK	00:35	

Data Download Date/Time: 01/01/2022 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,MATTHEW,P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LEE,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 12/28/2021 22:36:28

Diagnostics Check	OK	22:30
Air Blank	0.000	22:30
Control Test	0.080	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.217	22:32
Breath Volume	2.613	22:32
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.202	22:34
Breath Volume	1.515	22:34
Air Blank	0.000	22:35
Control Test	0.079	22:35
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 01/01/2022 21:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: LOPEZ RAMIREZ

DATE/TIME OF TEST: 12/31/2021 19:11:42

Diagnostics Check	OK	19:05
Air Blank	0.000	19:06
Control Test	0.081	19:06
Air Blank	0.000	19:06
Subject Sample # 1	0.223	19:07
Breath Volume	1.875	19:07
Air Blank	0.000	19:07
Air Blank	0.000	19:09
Subject Sample # 2	0.219	19:10
Breath Volume	1.316	19:10
Air Blank	0.000	19:10
Control Test	0.080	19:11
Air Blank	0.000	19:11
Diagnostics Check	OK	19:11

Data Download Date/Time: 01/01/2022 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TRAJBAR,CHRISTOPHER
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: BARRANCA MONTES

DATE/TIME OF TEST: 12/31/2021 19:36:19

Diagnostics Check	OK	19:29
Air Blank	0.000	19:29
Control Test	0.081	19:29
Air Blank	0.000	19:30
Subject Sample # 1	0.158	19:30
Breath Volume	1.593	19:30
Air Blank	0.000	19:31
Air Blank	0.000	19:33
Subject Sample # 2	0.158	19:34
Breath Volume	1.910	19:34
Air Blank	0.000	19:35
Control Test	0.081	19:35
Air Blank	0.000	19:36
Diagnostics Check	OK	19:36

Data Download Date/Time: 01/01/2022 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER, THOMAS, J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: BATCH, JOSEPH
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 11/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: BOSS

DATE/TIME OF TEST: 12/10/2021 05:30:23

Diagnostics Check	OK	05:23
Air Blank	0.000	05:24
Control Test	0.081	05:24
Air Blank	0.000	05:25
Subject Sample # 1	0.184	05:25
Breath Volume	2.171	05:25
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample # 2	0.173	05:28
Breath Volume	2.152	05:28
Air Blank	0.000	05:29
Control Test	0.079	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 12/29/2021 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN,BRADLEY,
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: MCLAUGHLIN,BRADLEY
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 11/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: SODERBERG

DATE/TIME OF TEST: 12/18/2021 03:10:00

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:59	
Control Test	0.081	02:59	
Air Blank	0.000	02:59	
Subject Sample	# 1	03:03	No Sample Provided
Breath Volume	0.000	03:03	
Air Blank	0.000	03:03	
Air Blank	0.000	03:05	
Subject Sample	# 2	03:08	No Sample Provided
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.081	03:09	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 12/29/2021 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JUSTIN,A
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: BAEZ,XZEVIES
ARRESTING OFFICER AGENCY: TITUSVILLE PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: VALERIO

DATE/TIME OF TEST: 12/30/2021 21:00:57

Diagnostics Check	OK	20:54
Air Blank	0.000	20:54
Control Test	0.081	20:55
Air Blank	0.000	20:55
Subject Sample # 1	0.034	20:56
Breath Volume	1.261	20:56
Air Blank	0.000	20:57
Air Blank	0.000	20:58
Subject Sample # 2	0.033	20:59
Breath Volume	1.847	20:59
Air Blank	0.000	21:00
Control Test	0.079	21:00
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time: 01/28/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JUSTIN,A
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: CULP,HEATHER
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/31/2021 02:20:18

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.081	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.062	02:14
Breath Volume	1.988	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.056	02:18
Breath Volume	2.152	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 01/28/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: WITKOSKY

DATE/TIME OF TEST: 12/01/2021 00:04:40

Diagnostics Check	OK	23:57
Air Blank	0.000	23:58
Control Test	0.081	23:58
Air Blank	0.000	23:59
Subject Sample # 2	0.065	23:59
Breath Volume	2.886	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 1	0.063	00:03
Breath Volume	2.175	00:03
Air Blank	0.000	00:03
Control Test	0.080	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 12/14/2021 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: GARAY

DATE/TIME OF TEST: 12/01/2021 01:42:07

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.081	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.196	01:37
Breath Volume	1.773	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:39
Subject Sample # 2	0.177	01:40
Breath Volume	2.050	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 12/14/2021 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TURPIN,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 12/02/2021 20:34:48

Diagnostics Check	OK	20:28
Air Blank	0.000	20:28
Control Test	0.081	20:29
Air Blank	0.000	20:29
Subject Sample # 1	0.117	20:30
Breath Volume	2.453	20:30
Air Blank	0.000	20:30
Air Blank	0.000	20:32
Subject Sample # 2	0.117	20:33
Breath Volume	2.140	20:33
Air Blank	0.000	20:33
Control Test	0.080	20:34
Air Blank	0.000	20:34
Diagnostics Check	OK	20:34

Data Download Date/Time: 12/14/2021 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FOSTER,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/02/2021 21:20:32

Diagnostics Check	OK	21:12
Air Blank	0.000	21:12
Control Test	0.081	21:12
Air Blank	0.000	21:13
Subject Sample # 1	0.000	21:14
Breath Volume	1.554	21:14
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.000	21:19
Breath Volume	1.390	21:19
Air Blank	0.000	21:19
Control Test	0.080	21:19
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

Data Download Date/Time: 12/14/2021 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: WARNER

DATE/TIME OF TEST: 12/02/2021 22:24:59

Diagnostics Check	OK	22:22	
Air Blank	0.000	22:22	
Control Test	0.081	22:22	
Air Blank	0.000	22:23	
Subject Sample	# 1	22:23	Subject Test Refused
Breath Volume	0.000	22:23	
Air Blank	0.000	22:24	
Control Test	0.080	22:24	
Air Blank	0.000	22:24	
Diagnostics Check	OK	22:24	

Data Download Date/Time: 12/14/2021 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 12/03/2021 01:23:15

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:20	
Control Test	0.081	01:21	
Air Blank	0.000	01:21	
Subject Sample	# 1	01:21	Subject Test Refused
Breath Volume	0.000	01:21	
Air Blank	0.000	01:22	
Control Test	0.080	01:22	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 12/14/2021 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: U of North FL PD
ARRESTING OFFICER: CLYBURN,J
ARRESTING OFFICER AGENCY: U of North FL PD
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 12/03/2021 02:51:06

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.081	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.045	02:46
Breath Volume	1.437	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:48
Subject Sample # 2	0.043	02:49
Breath Volume	1.398	02:49
Air Blank	0.000	02:50
Control Test	0.081	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 12/14/2021 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 12/03/2021 23:45:58

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.081	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.124	23:41
Breath Volume	1.910	23:41
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.124	23:44
Breath Volume	1.394	23:44
Air Blank	0.000	23:45
Control Test	0.080	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 12/14/2021 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: SHAVER

DATE/TIME OF TEST: 12/04/2021 01:14:01

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.000	01:08
Breath Volume	1.558	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.000	01:12
Breath Volume	1.875	01:12
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:14

Data Download Date/Time: 12/14/2021 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: MCGHEE

DATE/TIME OF TEST: 12/04/2021 06:01:51

Diagnostics Check	OK	05:58	
Air Blank	0.000	05:59	
Control Test	0.081	05:59	
Air Blank	0.000	06:00	
Subject Sample	# 1	06:00	Subject Test Refused
Breath Volume	0.000	06:00	
Air Blank	0.000	06:00	
Control Test	0.081	06:01	
Air Blank	0.000	06:01	
Diagnostics Check	OK	06:01	

Data Download Date/Time: 12/14/2021 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSES,L,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LEVERENZ,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:01

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 12/04/2021 09:38:03

Diagnostics Check	OK	09:27	
Air Blank	0.000	09:27	
Control Test	0.081	09:27	
Air Blank	0.000	09:28	
Subject Sample	# 1	09:31	No Sample Provided
Breath Volume	0.000	09:31	
Air Blank	0.000	09:31	
Air Blank	0.000	09:33	
Subject Sample	# 2	09:36	Subject Test Refused
Breath Volume	0.035	09:36	
Air Blank	0.000	09:37	
Control Test	0.080	09:37	
Air Blank	0.000	09:37	
Diagnostics Check	OK	09:38	

Data Download Date/Time: 12/14/2021 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGGINS,R,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 12/04/2021 21:32:14

Diagnostics Check	OK	21:29	
Air Blank	0.000	21:29	
Control Test	0.080	21:30	
Air Blank	0.000	21:30	
Subject Sample	# 1	21:30	Subject Test Refused
Breath Volume	0.000	21:30	
Air Blank	0.000	21:31	
Control Test	0.080	21:31	
Air Blank	0.000	21:32	
Diagnostics Check	OK	21:32	

Data Download Date/Time: 12/14/2021 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WILLIAMS,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/05/2021 01:34:58

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:32	
Control Test	0.082	01:32	
Air Blank	0.000	01:33	
Subject Sample	# 1	01:33	Subject Test Refused
Breath Volume	0.000	01:33	
Air Blank	0.000	01:34	
Control Test	0.080	01:34	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 12/14/2021 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: RITCHIE

DATE/TIME OF TEST: 12/05/2021 02:53:06

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.081	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.053	02:48
Breath Volume	1.957	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:50
Subject Sample # 2	0.058	02:51
Breath Volume	1.816	02:51
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 12/14/2021 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: CURRIE

DATE/TIME OF TEST: 12/05/2021 03:49:27

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.081	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.123	03:44
Breath Volume	1.882	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.125	03:47
Breath Volume	2.269	03:47
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 12/14/2021 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: PUSEY

DATE/TIME OF TEST: 12/05/2021 05:59:50

Diagnostics Check	OK	05:51
Air Blank	0.000	05:52
Control Test	0.080	05:52
Air Blank	0.000	05:53
Subject Sample # 1	0.236	05:54
Breath Volume	1.746	05:54
Air Blank	0.000	05:55
Air Blank	0.000	05:57
Subject Sample # 2	0.235	05:58
Breath Volume	1.585	05:58
Air Blank	0.000	05:58
Control Test	0.079	05:59
Air Blank	0.000	05:59
Diagnostics Check	OK	05:59

Data Download Date/Time: 12/14/2021 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/05/2021 23:50:27

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:47	
Control Test	0.081	23:48	
Air Blank	0.000	23:48	
Subject Sample	# 1	23:49	Subject Test Refused
Breath Volume	0.000	23:49	
Air Blank	0.000	23:49	
Control Test	0.080	23:49	
Air Blank	0.000	23:50	
Diagnostics Check	OK	23:50	

Data Download Date/Time: 12/14/2021 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: RIVERO EXPOSITO

DATE/TIME OF TEST: 12/06/2021 04:25:55

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.080	04:19
Air Blank	0.000	04:19
Subject Sample # 1	0.072	04:20
Breath Volume	2.644	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.071	04:24
Breath Volume	2.671	04:24
Air Blank	0.000	04:24
Control Test	0.080	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 12/14/2021 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORRISON,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 12/07/2021 04:16:10

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.159	04:10
Breath Volume	2.007	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.165	04:14
Breath Volume	2.066	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 12/14/2021 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: JULIUS

DATE/TIME OF TEST: 12/08/2021 01:57:47

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.080	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.245	01:49
Breath Volume	1.296	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:51
Subject Sample # 2	0.267	01:52
Breath Volume	1.757	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 3	0.263	01:56
Breath Volume	2.468	01:56
Air Blank	0.000	01:56
Control Test	0.078	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 12/14/2021 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 12/08/2021 04:13:01

Diagnostics Check	OK	04:10	
Air Blank	0.000	04:10	
Control Test	0.080	04:10	
Air Blank	0.000	04:11	
Subject Sample	# 1	04:11	Subject Test Refused
Breath Volume	0.000	04:11	
Air Blank	0.000	04:12	
Control Test	0.080	04:12	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 12/14/2021 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 12/08/2021 23:31:08

Diagnostics Check	OK	23:23
Air Blank	0.000	23:24
Control Test	0.079	23:24
Air Blank	0.000	23:25
Subject Sample # 1	0.208	23:25
Breath Volume	2.710	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.194	23:29
Breath Volume	2.886	23:29
Air Blank	0.000	23:30
Control Test	0.080	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 12/14/2021 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: SMALLS

DATE/TIME OF TEST: 12/09/2021 01:50:45

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:48	
Control Test	0.080	01:48	
Air Blank	0.000	01:49	
Subject Sample	# 1	01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:49	
Control Test	0.080	01:50	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:50	

Data Download Date/Time: 12/14/2021 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARRISH,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 12/09/2021 21:16:18

Diagnostics Check	OK	21:09
Air Blank	0.000	21:10
Control Test	0.081	21:10
Air Blank	0.000	21:10
Subject Sample # 1	0.166	21:11
Breath Volume	1.738	21:11
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.175	21:14
Breath Volume	1.859	21:14
Air Blank	0.000	21:15
Control Test	0.079	21:15
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 12/14/2021 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/10/2021 01:43:24

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.081	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.114	01:38
Breath Volume	1.710	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:40
Subject Sample # 2	0.114	01:41
Breath Volume	1.941	01:41
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 12/14/2021 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 12/10/2021 23:27:32

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.000	23:23
Breath Volume	3.160	23:23
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.000	23:26
Breath Volume	2.804	23:26
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 12/14/2021 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: ACTE ISEM

DATE/TIME OF TEST: 12/11/2021 03:39:00

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.081	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.229	03:33
Breath Volume	2.503	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.221	03:37
Breath Volume	1.480	03:37
Air Blank	0.000	03:38
Control Test	0.079	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:39

Data Download Date/Time: 12/14/2021 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSES,L,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:59

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/11/2021 17:28:05

Diagnostics Check	OK	17:20
Air Blank	0.000	17:20
Control Test	0.080	17:21
Air Blank	0.000	17:21
Subject Sample # 1	0.285	17:23
Breath Volume	1.699	17:23
Air Blank	0.000	17:24
Air Blank	0.000	17:25
Subject Sample # 2	0.273	17:26
Breath Volume	1.496	17:26
Air Blank	0.000	17:27
Control Test	0.076	17:27
Air Blank	0.000	17:27
Diagnostics Check	OK	17:28

Data Download Date/Time: 12/14/2021 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: RYBECK

DATE/TIME OF TEST: 12/12/2021 01:29:57

Diagnostics Check	OK	01:26	
Air Blank	0.000	01:27	
Control Test	0.081	01:27	
Air Blank	0.000	01:28	
Subject Sample	# 1	01:28	Subject Test Refused
Breath Volume	0.000	01:28	
Air Blank	0.000	01:29	
Control Test	0.081	01:29	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

Data Download Date/Time: 12/14/2021 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: ROLLINS

DATE/TIME OF TEST: 12/12/2021 02:29:51

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.169	02:24
Breath Volume	2.148	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.168	02:28
Breath Volume	1.773	02:28
Air Blank	0.000	02:28
Control Test	0.081	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 12/14/2021 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 12/12/2021 05:32:15

Diagnostics Check	OK	05:25
Air Blank	0.000	05:25
Control Test	0.081	05:26
Air Blank	0.000	05:26
Subject Sample # 1	0.109	05:27
Breath Volume	1.476	05:27
Air Blank	0.000	05:28
Air Blank	0.000	05:29
Subject Sample # 2	0.108	05:30
Breath Volume	2.007	05:30
Air Blank	0.000	05:31
Control Test	0.081	05:31
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

Data Download Date/Time: 12/14/2021 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSES,L,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:39

SUBJECT'S LAST NAME: ESTEBAN MANUEL

DATE/TIME OF TEST: 12/12/2021 18:06:27

Diagnostics Check	OK	17:59
Air Blank	0.000	17:59
Control Test	0.081	18:00
Air Blank	0.000	18:00
Subject Sample # 1	0.105	18:01
Breath Volume	2.402	18:01
Air Blank	0.000	18:02
Air Blank	0.000	18:04
Subject Sample # 2	0.104	18:04
Breath Volume	2.628	18:04
Air Blank	0.000	18:05
Control Test	0.081	18:05
Air Blank	0.000	18:06
Diagnostics Check	OK	18:06

Data Download Date/Time: 12/14/2021 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: DENNEY

DATE/TIME OF TEST: 12/13/2021 01:17:23

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.081	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.136	01:12
Breath Volume	2.488	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.136	01:15
Breath Volume	2.429	01:15
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 12/14/2021 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: ASHLEY

DATE/TIME OF TEST: 12/13/2021 23:45:08

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:42	
Control Test	0.080	23:42	
Air Blank	0.000	23:43	
Subject Sample	# 1	23:43	Subject Test Refused
Breath Volume	0.000	23:43	
Air Blank	0.000	23:44	
Control Test	0.081	23:44	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:45	

Data Download Date/Time: 12/14/2021 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: GALLAGHER

DATE/TIME OF TEST: 12/17/2021 03:53:18

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.081	03:51	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:52	
Control Test	0.081	03:52	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 12/29/2021 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RUZANKA,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: ADMIN
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: UNKNOWN

DATE/TIME OF TEST: 12/17/2021 04:08:07

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.067	04:03
Breath Volume	3.191	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:05
Subject Sample # 2	0.067	04:06
Breath Volume	3.031	04:06
Air Blank	0.000	04:07
Control Test	0.081	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 12/29/2021 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RUZANKA,P
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: ADMIN
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: UNKNOWN

DATE/TIME OF TEST: 12/17/2021 04:29:27

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.081	04:24
Air Blank	0.000	04:24
Subject Sample # 1	0.063	04:25
Breath Volume	2.464	04:25
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.062	04:27
Breath Volume	2.695	04:27
Air Blank	0.000	04:28
Control Test	0.081	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 12/29/2021 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: GALLAGHER

DATE/TIME OF TEST: 12/17/2021 04:40:34

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.081	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.057	04:35
Breath Volume	1.933	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.060	04:38
Breath Volume	2.949	04:38
Air Blank	0.000	04:39
Control Test	0.081	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 12/29/2021 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JONES,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 12/17/2021 13:35:23

Diagnostics Check	OK	13:27
Air Blank	0.000	13:28
Control Test	0.080	13:28
Air Blank	0.000	13:29
Subject Sample # 1	0.123	13:30
Breath Volume	1.628	13:30
Air Blank	0.000	13:31
Air Blank	0.000	13:33
Subject Sample # 2	0.132	13:33
Breath Volume	1.539	13:33
Air Blank	0.000	13:34
Control Test	0.079	13:34
Air Blank	0.000	13:35
Diagnostics Check	OK	13:35

Data Download Date/Time: 12/29/2021 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: FIGUEROA RODRIGUEZ

DATE/TIME OF TEST: 12/17/2021 21:08:09

Diagnostics Check	OK	21:01
Air Blank	0.000	21:01
Control Test	0.081	21:02
Air Blank	0.000	21:02
Subject Sample # 1	0.153	21:03
Breath Volume	1.515	21:03
Air Blank	0.000	21:04
Air Blank	0.000	21:05
Subject Sample # 2	0.161	21:06
Breath Volume	2.371	21:06
Air Blank	0.000	21:07
Control Test	0.080	21:07
Air Blank	0.000	21:08
Diagnostics Check	OK	21:08

Data Download Date/Time: 12/29/2021 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 12/17/2021 22:42:17

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:39	
Control Test	0.081	22:40	
Air Blank	0.000	22:40	
Subject Sample	# 1	22:40	Subject Test Refused
Breath Volume	0.000	22:40	
Air Blank	0.000	22:41	
Control Test	0.081	22:41	
Air Blank	0.000	22:42	
Diagnostics Check	OK	22:42	

Data Download Date/Time: 12/29/2021 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 12/17/2021 23:31:46

Diagnostics Check	OK	23:26	
Air Blank	0.000	23:26	
Control Test	0.081	23:26	
Air Blank	0.000	23:27	
Subject Sample	# 1 0.249	23:30	Volume Not Met
Breath Volume	0.261	23:30	
Air Blank		23:31	Purge Fail
Air Blank	0.000	23:31	

Data Download Date/Time: 12/29/2021 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 12/17/2021 23:47:09

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.075	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.276	23:39
Breath Volume	1.292	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:41
Subject Sample # 2	0.302	23:42
Breath Volume	1.398	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:44
Subject Sample # 3	0.307	23:45
Breath Volume	1.535	23:45
Air Blank	0.000	23:46
Control Test	0.078	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 12/29/2021 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 12/18/2021 01:30:33

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:28	
Control Test	0.081	01:28	
Air Blank	0.000	01:28	
Subject Sample	# 1	01:29	Subject Test Refused
Breath Volume	0.000	01:29	
Air Blank	0.000	01:29	
Control Test	0.081	01:29	
Air Blank	0.000	01:30	
Diagnostics Check	OK	01:30	

Data Download Date/Time: 12/29/2021 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: MAZARIEGO PEREZ

DATE/TIME OF TEST: 12/18/2021 02:44:29

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.180	02:39
Breath Volume	1.574	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:41
Subject Sample # 2	0.197	02:42
Breath Volume	1.277	02:42
Air Blank	0.000	02:43
Control Test	0.078	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 12/29/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,TYLER
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SHEPPARD

DATE/TIME OF TEST: 12/18/2021 04:14:50

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:12	
Control Test	0.081	04:12	
Air Blank	0.000	04:13	
Subject Sample	# 1	04:13	Subject Test Refused
Breath Volume	0.000	04:13	
Air Blank	0.000	04:13	
Control Test	0.081	04:14	
Air Blank	0.000	04:14	
Diagnostics Check	OK	04:14	

Data Download Date/Time: 12/29/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOROZOV, S
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: STREETER

DATE/TIME OF TEST: 12/20/2021 02:06:23

Diagnostics Check	OK	01:59
Air Blank	0.000	02:00
Control Test	0.081	02:00
Air Blank	0.000	02:01
Subject Sample # 1	0.160	02:01
Breath Volume	1.335	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.158	02:04
Breath Volume	1.132	02:04
Air Blank	0.000	02:05
Control Test	0.079	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 12/29/2021 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: HERNANDEZ BAUTISTA

DATE/TIME OF TEST: 12/20/2021 04:43:31

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.081	04:37
Air Blank	0.000	04:37
Subject Sample # 1	0.184	04:38
Breath Volume	1.835	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:40
Subject Sample # 2	0.186	04:41
Breath Volume	1.933	04:41
Air Blank	0.000	04:42
Control Test	0.080	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 12/29/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 12/21/2021 02:07:43

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:05	
Control Test	0.081	02:05	
Air Blank	0.000	02:06	
Subject Sample	# 1	02:06	Subject Test Refused
Breath Volume	0.000	02:06	
Air Blank	0.000	02:06	
Control Test	0.081	02:07	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 12/29/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 12/21/2021 22:14:23

Diagnostics Check	OK	22:11	
Air Blank	0.000	22:11	
Control Test	0.081	22:12	
Air Blank	0.000	22:12	
Subject Sample	# 1	22:12	Subject Test Refused
Breath Volume	0.000	22:12	
Air Blank	0.000	22:13	
Control Test	0.081	22:13	
Air Blank	0.000	22:14	
Diagnostics Check	OK	22:14	

Data Download Date/Time: 12/29/2021 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: ALEMAN CRUZ

DATE/TIME OF TEST: 12/23/2021 20:10:02

Diagnostics Check	OK	20:03
Air Blank	0.000	20:03
Control Test	0.081	20:03
Air Blank	0.000	20:04
Subject Sample # 1	0.216	20:04
Breath Volume	1.523	20:04
Air Blank	0.000	20:05
Air Blank	0.000	20:07
Subject Sample # 2	0.216	20:08
Breath Volume	2.375	20:08
Air Blank	0.000	20:09
Control Test	0.081	20:09
Air Blank	0.000	20:09
Diagnostics Check	OK	20:10

Data Download Date/Time: 12/29/2021 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGGINS,R,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DELBROCCO-BLUES,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/24/2021 05:04:40

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.082	04:58
Air Blank	0.000	04:59
Subject Sample # 1	0.119	04:59
Breath Volume	1.121	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.128	05:02
Breath Volume	1.328	05:02
Air Blank	0.000	05:03
Control Test	0.081	05:04
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 12/29/2021 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DE SOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: QUEVEDO AVILA

DATE/TIME OF TEST: 12/24/2021 21:30:52

Diagnostics Check	OK	21:24
Air Blank	0.000	21:24
Control Test	0.081	21:25
Air Blank	0.000	21:25
Subject Sample # 1	0.251	21:26
Breath Volume	2.453	21:26
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.242	21:29
Breath Volume	1.535	21:29
Air Blank	0.000	21:29
Control Test	0.081	21:30
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 12/29/2021 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOCKERS,B
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: HASE GAWA BELINELI

DATE/TIME OF TEST: 12/24/2021 23:57:24

Diagnostics Check	OK	23:49	
Air Blank	0.000	23:49	
Control Test	0.081	23:50	
Air Blank	0.000	23:50	
Subject Sample # 1	0.242	23:51	
Breath Volume	2.042	23:51	
Air Blank	0.000	23:52	
Air Blank	0.000	23:54	
Subject Sample # 2		23:55	Subject Test Refused
Breath Volume	0.000	23:55	
Air Blank	0.000	23:56	
Control Test	0.080	23:56	
Air Blank	0.000	23:57	
Diagnostics Check	OK	23:57	

Data Download Date/Time: 12/29/2021 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MATTIACE,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: WARNER

DATE/TIME OF TEST: 12/25/2021 03:56:12

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.081	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.157	03:51
Breath Volume	3.402	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:53
Subject Sample # 2	0.154	03:54
Breath Volume	3.035	03:54
Air Blank	0.000	03:55
Control Test	0.081	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 12/29/2021 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: TOUCH

DATE/TIME OF TEST: 12/25/2021 23:27:17

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.081	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.124	23:22
Breath Volume	1.945	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:24
Subject Sample # 2	0.125	23:25
Breath Volume	1.980	23:25
Air Blank	0.000	23:26
Control Test	0.081	23:26
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 12/29/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 12/26/2021 02:22:16

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:19	
Control Test	0.081	02:20	
Air Blank	0.000	02:20	
Subject Sample	# 1	02:20	Subject Test Refused
Breath Volume	0.000	02:20	
Air Blank	0.000	02:21	
Control Test	0.081	02:21	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

Data Download Date/Time: 12/29/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: BENT

DATE/TIME OF TEST: 12/27/2021 02:02:47

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.045	01:58
Breath Volume	4.453	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.044	02:01
Breath Volume	3.960	02:01
Air Blank	0.000	02:01
Control Test	0.081	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 12/29/2021 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: BENTON

DATE/TIME OF TEST: 12/28/2021 04:03:36

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.081	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.193	03:58
Breath Volume	3.515	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.188	04:01
Breath Volume	3.503	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 12/29/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: WALIZER

DATE/TIME OF TEST: 12/29/2021 02:29:48

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.080	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.090	02:24
Breath Volume	2.542	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.083	02:28
Breath Volume	1.367	02:28
Air Blank	0.000	02:28
Control Test	0.078	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 12/29/2021 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: JEFFERSON

DATE/TIME OF TEST: 12/29/2021 22:03:08

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.213	21:58
Breath Volume	1.945	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:00
Subject Sample # 2	0.210	22:01
Breath Volume	1.699	22:01
Air Blank	0.000	22:02
Control Test	0.079	22:02
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 01/13/2022 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: TURKNETT

DATE/TIME OF TEST: 12/30/2021 04:39:53

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:37	
Control Test	0.081	04:37	
Air Blank	0.000	04:38	
Subject Sample	# 1	04:38	Subject Test Refused
Breath Volume	0.000	04:38	
Air Blank	0.000	04:38	
Control Test	0.080	04:39	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:39	

Data Download Date/Time: 01/13/2022 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOCKERS,B
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: PANTESCO

DATE/TIME OF TEST: 12/31/2021 02:07:47

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:05	
Control Test	0.080	02:05	
Air Blank	0.000	02:06	
Subject Sample	# 1	02:06	Subject Test Refused
Breath Volume	0.000	02:06	
Air Blank	0.000	02:06	
Control Test	0.080	02:07	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 01/13/2022 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: BALTAZAR

DATE/TIME OF TEST: 12/31/2021 03:13:07

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.081	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.162	03:07
Breath Volume	2.113	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.160	03:11
Breath Volume	2.074	03:11
Air Blank	0.000	03:12
Control Test	0.078	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 01/13/2022 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: HERNANDEZ GOMEZ

DATE/TIME OF TEST: 12/31/2021 23:34:59

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.079	23:30
Breath Volume	2.601	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:32
Subject Sample # 2	0.079	23:33
Breath Volume	1.656	23:33
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 01/13/2022 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 12/09/2021 20:59:06

Diagnostics Check	OK	20:56	
Air Blank	0.000	20:56	
Control Test	0.077	20:57	
Air Blank	0.000	20:57	
Subject Sample	# 1	20:57	Subject Test Refused
Breath Volume	0.000	20:57	
Air Blank	0.000	20:58	
Control Test	0.076	20:58	
Air Blank	0.000	20:58	
Diagnostics Check	OK	20:59	

Data Download Date/Time: 12/15/2021 16:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: RAU

DATE/TIME OF TEST: 12/11/2021 17:41:42

Diagnostics Check	OK	17:38	
Air Blank	0.000	17:39	
Control Test	0.076	17:39	
Air Blank	0.000	17:40	
Subject Sample	# 1	17:40	Subject Test Refused
Breath Volume	0.000	17:40	
Air Blank	0.000	17:40	
Control Test	0.076	17:41	
Air Blank	0.000	17:41	
Diagnostics Check	OK	17:41	

Data Download Date/Time: 12/15/2021 16:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 11/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: SYLVESTER

DATE/TIME OF TEST: 12/11/2021 20:23:18

Diagnostics Check	OK	20:16
Air Blank	0.000	20:17
Control Test	0.076	20:17
Air Blank	0.000	20:17
Subject Sample # 1	0.157	20:18
Breath Volume	1.570	20:18
Air Blank	0.000	20:19
Air Blank	0.000	20:21
Subject Sample # 2	0.153	20:21
Breath Volume	1.312	20:21
Air Blank	0.000	20:22
Control Test	0.077	20:22
Air Blank	0.000	20:23
Diagnostics Check	OK	20:23

Data Download Date/Time: 12/15/2021 16:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MONTES

DATE/TIME OF TEST: 12/16/2021 00:12:07

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.077	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.213	00:07
Breath Volume	3.199	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.210	00:10
Breath Volume	1.359	00:10
Air Blank	0.000	00:11
Control Test	0.079	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 01/19/2022 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS,BENJAMIN,F
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: MIELE,ROBERT
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: RAMADORI

DATE/TIME OF TEST: 12/19/2021 23:58:39

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:46	
Control Test	0.077	23:46	
Air Blank	0.000	23:47	
Subject Sample	# 1 0.033	23:50	Volume Not Met
Breath Volume	0.480	23:50	
Air Blank	0.000	23:50	
Air Blank	0.000	23:52	
Subject Sample	# 2 0.042	23:54	
Breath Volume	2.742	23:54	
Air Blank	0.000	23:54	
Air Blank	0.000	23:56	
Subject Sample	# 3 0.042	23:57	
Breath Volume	2.082	23:57	
Air Blank	0.000	23:57	
Control Test	0.077	23:58	
Air Blank	0.000	23:58	
Diagnostics Check	OK	23:58	

Data Download Date/Time: 01/19/2022 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: RESSLER, GREYSON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 12/21/2021 02:31:07

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.077	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.256	02:26
Breath Volume	1.742	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.261	02:29
Breath Volume	2.722	02:29
Air Blank	0.000	02:30
Control Test	0.076	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 01/19/2022 15:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCORDINO, DOUGLAS, J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS, STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: DENNIS JR

DATE/TIME OF TEST: 12/21/2021 03:03:28

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:01	
Control Test	0.078	03:01	
Air Blank	0.000	03:01	
Subject Sample	# 1	03:02	Subject Test Refused
Breath Volume	0.000	03:02	
Air Blank	0.000	03:02	
Control Test	0.078	03:02	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:03	

Data Download Date/Time: 01/19/2022 15:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCHUK,J,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: CORTES,A
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: MICHELS

DATE/TIME OF TEST: 12/26/2021 19:31:41

Diagnostics Check	OK	19:28	
Air Blank	0.000	19:29	
Control Test	0.076	19:29	
Air Blank	0.000	19:30	
Subject Sample	# 1	19:30	Subject Test Refused
Breath Volume	0.000	19:30	
Air Blank	0.000	19:30	
Control Test	0.077	19:31	
Air Blank	0.000	19:31	
Diagnostics Check	OK	19:31	

Data Download Date/Time: 01/19/2022 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REQUENA,CARLOS,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: MASTERS-SAMRETH,DAVID
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 12/26/2021 23:15:15

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:11	
Control Test	0.078	23:11	
Air Blank	0.000	23:12	
Subject Sample	# 1	23:13	Subject Test Refused
Breath Volume	0.000	23:13	
Air Blank	0.000	23:14	
Control Test	0.078	23:14	
Air Blank	0.000	23:15	
Diagnostics Check	OK	23:15	

Data Download Date/Time: 01/19/2022 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REQUENA,CARLOS,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHORTT,DEREK
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/29/2021 19:46:38

Diagnostics Check	OK	19:40
Air Blank	0.000	19:41
Control Test	0.077	19:41
Air Blank	0.000	19:42
Subject Sample # 1	0.160	19:42
Breath Volume	2.179	19:42
Air Blank	0.000	19:42
Air Blank	0.000	19:44
Subject Sample # 2	0.159	19:45
Breath Volume	1.843	19:45
Air Blank	0.000	19:45
Control Test	0.077	19:46
Air Blank	0.000	19:46
Diagnostics Check	OK	19:46

Data Download Date/Time: 01/19/2022 15:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 11/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094981
DRY GAS EXPIRY DATE: 02/22/2022
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: CHAGASRODRIGUES

DATE/TIME OF TEST: 12/20/2021 20:42:22

Diagnostics Check	OK	20:36
Air Blank	0.000	20:36
Control Test	0.080	20:36
Air Blank	0.000	20:37
Subject Sample # 1	0.196	20:38
Breath Volume	1.945	20:38
Air Blank	0.000	20:38
Air Blank	0.000	20:40
Subject Sample # 2	0.194	20:40
Breath Volume	2.035	20:40
Air Blank	0.000	20:41
Control Test	0.080	20:41
Air Blank	0.000	20:42
Diagnostics Check	OK	20:42

Data Download Date/Time: 01/13/2022 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MICHAEL,C
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: CORTEZ,JONATHAN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 11/17/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1094981
DRY GAS EXPIRY DATE: 02/22/2022
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: STUARDO STUARDO

DATE/TIME OF TEST: 12/21/2021 00:00:49

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:55
Subject Sample # 1	0.156	23:56
Breath Volume	3.113	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.159	23:59
Breath Volume	3.804	23:59
Air Blank	0.000	23:59
Control Test	0.079	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 01/13/2022 09:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: AMONETTE

DATE/TIME OF TEST: 12/02/2021 04:23:04

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:20	
Control Test	0.081	04:20	
Air Blank	0.000	04:21	
Subject Sample	# 1	04:21	Subject Test Refused
Breath Volume	0.000	04:21	
Air Blank	0.000	04:22	
Control Test	0.080	04:22	
Air Blank	0.000	04:22	
Diagnostics Check	OK	04:23	

Data Download Date/Time: 12/16/2021 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 12/02/2021 22:49:38

Diagnostics Check	OK	22:46	
Air Blank	0.000	22:47	
Control Test	0.081	22:47	
Air Blank	0.000	22:47	
Subject Sample	# 1	22:48	Subject Test Refused
Breath Volume	0.000	22:48	
Air Blank	0.000	22:48	
Control Test	0.081	22:49	
Air Blank	0.000	22:49	
Diagnostics Check	OK	22:49	

Data Download Date/Time: 12/16/2021 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 12/03/2021 03:48:40

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.081	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.154	03:44
Breath Volume	3.207	03:44
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.157	03:47
Breath Volume	2.507	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 12/16/2021 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALE,J,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARRISH,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: AGUIRRE

DATE/TIME OF TEST: 12/04/2021 04:35:19

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.081	04:29
Air Blank	0.000	04:29
Subject Sample # 1	0.082	04:30
Breath Volume	1.585	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.083	04:33
Breath Volume	1.433	04:33
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 12/16/2021 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: BROADUS JR

DATE/TIME OF TEST: 12/04/2021 22:27:43

Diagnostics Check	OK	22:21
Air Blank	0.000	22:21
Control Test	0.080	22:21
Air Blank	0.000	22:22
Subject Sample # 1	0.095	22:22
Breath Volume	2.742	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.098	22:26
Breath Volume	2.253	22:26
Air Blank	0.000	22:26
Control Test	0.080	22:27
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 12/16/2021 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/05/2021 01:37:44

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.081	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.127	01:30
Breath Volume	1.152	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.126	01:36
Breath Volume	1.597	01:36
Air Blank	0.000	01:36
Control Test	0.080	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 12/16/2021 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 12/05/2021 03:07:28

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.200	03:02
Breath Volume	1.554	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.186	03:05
Breath Volume	1.144	03:05
Air Blank	0.000	03:06
Control Test	0.079	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 12/16/2021 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: KRISHNAKUMAR

DATE/TIME OF TEST: 12/05/2021 19:33:35

Diagnostics Check	OK	19:27
Air Blank	0.000	19:27
Control Test	0.080	19:27
Air Blank	0.000	19:28
Subject Sample # 1	0.163	19:29
Breath Volume	1.406	19:29
Air Blank	0.000	19:29
Air Blank	0.000	19:31
Subject Sample # 2	0.154	19:32
Breath Volume	1.242	19:32
Air Blank	0.000	19:32
Control Test	0.080	19:32
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

Data Download Date/Time: 12/16/2021 10:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,G
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: LING

DATE/TIME OF TEST: 12/06/2021 04:16:03

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.080	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.133	04:11
Breath Volume	2.070	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:13
Subject Sample # 2	0.127	04:14
Breath Volume	1.960	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:16

Data Download Date/Time: 12/16/2021 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: BUTEAU

DATE/TIME OF TEST: 12/07/2021 02:25:35

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.080	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.158	02:20
Breath Volume	1.335	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.160	02:23
Breath Volume	1.304	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 12/16/2021 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DELBROCCO-BLUES,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: SMART

DATE/TIME OF TEST: 12/07/2021 05:41:05

Diagnostics Check	OK	05:34
Air Blank	0.000	05:34
Control Test	0.080	05:35
Air Blank	0.000	05:35
Subject Sample # 1	0.108	05:36
Breath Volume	1.937	05:36
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.117	05:39
Breath Volume	1.449	05:39
Air Blank	0.000	05:40
Control Test	0.079	05:40
Air Blank	0.000	05:40
Diagnostics Check	OK	05:41

Data Download Date/Time: 12/16/2021 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JORDAN,A
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 12/07/2021 19:24:54

Diagnostics Check	OK	19:18
Air Blank	0.000	19:18
Control Test	0.081	19:19
Air Blank	0.000	19:19
Subject Sample # 1	0.035	19:20
Breath Volume	1.511	19:20
Air Blank	0.000	19:20
Air Blank	0.000	19:22
Subject Sample # 2	0.032	19:23
Breath Volume	1.242	19:23
Air Blank	0.000	19:23
Control Test	0.080	19:24
Air Blank	0.000	19:24
Diagnostics Check	OK	19:24

Data Download Date/Time: 12/16/2021 10:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGGINS,R,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/08/2021 00:02:11

Diagnostics Check	OK	23:59	
Air Blank	0.000	23:59	
Control Test	0.080	00:00	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:00	Subject Test Refused
Breath Volume	0.000	00:00	
Air Blank	0.000	00:01	
Control Test	0.080	00:01	
Air Blank	0.000	00:02	
Diagnostics Check	OK	00:02	

Data Download Date/Time: 12/16/2021 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: OSHODI

DATE/TIME OF TEST: 12/08/2021 01:05:07

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:51	
Control Test	0.080	00:51	
Air Blank	0.000	00:51	
Subject Sample	# 1 0.094	00:55	Volume Not Met
Breath Volume	0.015	00:55	
Air Blank	0.000	00:55	
Air Blank	0.000	00:57	
Subject Sample	# 2 0.090	00:59	
Breath Volume	1.347	00:59	
Air Blank	0.000	00:59	
Air Blank	0.000	01:01	
Subject Sample	# 3 0.092	01:03	
Breath Volume	1.148	01:03	
Air Blank	0.000	01:04	
Control Test	0.079	01:04	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

Data Download Date/Time: 12/16/2021 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: MULLEN

DATE/TIME OF TEST: 12/08/2021 01:52:46

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.079	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.243	01:48
Breath Volume	2.324	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.249	01:51
Breath Volume	2.742	01:51
Air Blank	0.000	01:51
Control Test	0.079	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 12/16/2021 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: SEGNA

DATE/TIME OF TEST: 12/08/2021 03:14:55

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.069	03:10
Breath Volume	2.472	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:12
Subject Sample # 2	0.067	03:13
Breath Volume	2.367	03:13
Air Blank	0.000	03:13
Control Test	0.079	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 12/16/2021 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: GLASS

DATE/TIME OF TEST: 12/08/2021 20:21:37

Diagnostics Check	OK	20:12
Air Blank	0.000	20:13
Control Test	0.080	20:13
Air Blank	0.000	20:14
Subject Sample # 1	0.177	20:16
Breath Volume	1.804	20:16
Air Blank	0.000	20:17
Air Blank	0.000	20:18
Subject Sample # 2	0.173	20:20
Breath Volume	1.207	20:20
Air Blank	0.000	20:20
Control Test	0.078	20:21
Air Blank	0.000	20:21
Diagnostics Check	OK	20:21

Data Download Date/Time: 12/16/2021 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: WHEATON

DATE/TIME OF TEST: 12/09/2021 01:59:04

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:56	
Control Test	0.080	01:57	
Air Blank	0.000	01:57	
Subject Sample	# 1	01:57	Subject Test Refused
Breath Volume	0.000	01:57	
Air Blank	0.000	01:58	
Control Test	0.080	01:58	
Air Blank	0.000	01:58	
Diagnostics Check	OK	01:59	

Data Download Date/Time: 12/16/2021 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 12/09/2021 20:39:24

Diagnostics Check	OK	20:33
Air Blank	0.000	20:33
Control Test	0.081	20:34
Air Blank	0.000	20:34
Subject Sample # 1	0.000	20:35
Breath Volume	3.582	20:35
Air Blank	0.000	20:35
Air Blank	0.000	20:37
Subject Sample # 2	0.000	20:38
Breath Volume	3.058	20:38
Air Blank	0.000	20:38
Control Test	0.080	20:38
Air Blank	0.000	20:39
Diagnostics Check	OK	20:39

Data Download Date/Time: 12/16/2021 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: CRUZ HERNANDEZ

DATE/TIME OF TEST: 12/10/2021 23:46:49

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.080	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.191	23:42
Breath Volume	1.605	23:42
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.177	23:45
Breath Volume	2.210	23:45
Air Blank	0.000	23:45
Control Test	0.080	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 12/16/2021 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: FRESHWATER

DATE/TIME OF TEST: 12/11/2021 03:52:37

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:50	
Control Test	0.081	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:51	
Control Test	0.080	03:52	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:52	

Data Download Date/Time: 12/16/2021 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/12/2021 04:01:08

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:58	
Control Test	0.080	03:58	
Air Blank	0.000	03:59	
Subject Sample	# 1	03:59	Subject Test Refused
Breath Volume	0.000	03:59	
Air Blank	0.000	04:00	
Control Test	0.080	04:00	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

Data Download Date/Time: 12/16/2021 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: MEJIA JIMENEZ

DATE/TIME OF TEST: 12/13/2021 00:22:55

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:20	
Control Test	0.080	00:20	
Air Blank	0.000	00:21	
Subject Sample	# 1	00:21	Subject Test Refused
Breath Volume	0.000	00:21	
Air Blank	0.000	00:21	
Control Test	0.080	00:22	
Air Blank	0.000	00:22	
Diagnostics Check	OK	00:22	

Data Download Date/Time: 12/16/2021 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FLORY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: SMALL

DATE/TIME OF TEST: 12/14/2021 01:24:31

Diagnostics Check	OK	01:17
Air Blank	0.000	01:18
Control Test	0.080	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.186	01:19
Breath Volume	2.300	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.198	01:22
Breath Volume	2.332	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 12/16/2021 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS, A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 12/14/2021 04:33:23

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.175	04:28
Breath Volume	2.628	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.175	04:31
Breath Volume	2.136	04:31
Air Blank	0.000	04:32
Control Test	0.080	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 12/16/2021 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: WESLEY

DATE/TIME OF TEST: 12/14/2021 21:13:44

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:11	
Control Test	0.080	21:11	
Air Blank	0.000	21:12	
Subject Sample	# 1	21:12	Subject Test Refused
Breath Volume	0.000	21:12	
Air Blank	0.000	21:12	
Control Test	0.080	21:13	
Air Blank	0.000	21:13	
Diagnostics Check	OK	21:13	

Data Download Date/Time: 12/16/2021 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: SMILEY

DATE/TIME OF TEST: 12/15/2021 21:24:43

Diagnostics Check	OK	21:18
Air Blank	0.000	21:18
Control Test	0.080	21:18
Air Blank	0.000	21:19
Subject Sample # 1	0.156	21:20
Breath Volume	2.714	21:20
Air Blank	0.000	21:20
Air Blank	0.000	21:22
Subject Sample # 2	0.160	21:23
Breath Volume	2.750	21:23
Air Blank	0.000	21:23
Control Test	0.079	21:24
Air Blank	0.000	21:24
Diagnostics Check	OK	21:24

Data Download Date/Time: 12/16/2021 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: DINMORE

DATE/TIME OF TEST: 12/15/2021 22:19:54

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.080	22:14
Air Blank	0.000	22:14
Subject Sample # 1	0.095	22:15
Breath Volume	2.371	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:17
Subject Sample # 2	0.092	22:18
Breath Volume	2.355	22:18
Air Blank	0.000	22:18
Control Test	0.080	22:19
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 12/16/2021 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: BACHTOLD

DATE/TIME OF TEST: 12/16/2021 15:42:35

Diagnostics Check	OK	15:36
Air Blank	0.000	15:36
Control Test	0.081	15:36
Air Blank	0.000	15:37
Subject Sample # 1	0.230	15:37
Breath Volume	1.703	15:37
Air Blank	0.000	15:38
Air Blank	0.000	15:40
Subject Sample # 2	0.221	15:40
Breath Volume	2.761	15:40
Air Blank	0.000	15:41
Control Test	0.081	15:41
Air Blank	0.000	15:42
Diagnostics Check	OK	15:42

Data Download Date/Time: 12/30/2021 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 12/16/2021 18:43:14

Diagnostics Check	OK	18:40	
Air Blank	0.000	18:40	
Control Test	0.080	18:41	
Air Blank	0.000	18:41	
Subject Sample	# 1	18:41	Subject Test Refused
Breath Volume	0.000	18:41	
Air Blank	0.000	18:42	
Control Test	0.079	18:42	
Air Blank	0.000	18:43	
Diagnostics Check	OK	18:43	

Data Download Date/Time: 12/30/2021 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 12/16/2021 21:07:36

Diagnostics Check	OK	21:01
Air Blank	0.000	21:01
Control Test	0.081	21:02
Air Blank	0.000	21:02
Subject Sample # 1	0.000	21:03
Breath Volume	2.695	21:03
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.000	21:06
Breath Volume	2.320	21:06
Air Blank	0.000	21:06
Control Test	0.080	21:06
Air Blank	0.000	21:07
Diagnostics Check	OK	21:07

Data Download Date/Time: 12/30/2021 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGGINS,R,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: DENNIS

DATE/TIME OF TEST: 12/17/2021 00:03:01

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:00	
Control Test	0.081	00:00	
Air Blank	0.000	00:01	
Subject Sample	# 1	00:01	Subject Test Refused
Breath Volume	0.000	00:01	
Air Blank	0.000	00:02	
Control Test	0.081	00:02	
Air Blank	0.000	00:02	
Diagnostics Check	OK	00:03	

Data Download Date/Time: 12/30/2021 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGGINS,R,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 12/18/2021 03:15:04

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:12	
Control Test	0.081	03:13	
Air Blank	0.000	03:13	
Subject Sample	# 1	03:13	Subject Test Refused
Breath Volume	0.000	03:13	
Air Blank	0.000	03:14	
Control Test	0.080	03:14	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:15	

Data Download Date/Time: 12/30/2021 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGGINS,R,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: VASCONCELOS GOMEZ

DATE/TIME OF TEST: 12/18/2021 04:13:06

Diagnostics Check	OK	04:06
Air Blank	0.000	04:07
Control Test	0.081	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.198	04:08
Breath Volume	2.535	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:10
Subject Sample # 2	0.200	04:11
Breath Volume	2.113	04:11
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 12/30/2021 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: SATCHELL

DATE/TIME OF TEST: 12/19/2021 02:42:37

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:31	
Control Test	0.081	02:31	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:35	No Sample Provided
Breath Volume	0.000	02:35	
Air Blank	0.000	02:36	
Air Blank	0.000	02:38	
Subject Sample	# 2	02:41	No Sample Provided
Breath Volume	0.000	02:41	
Air Blank	0.000	02:41	
Control Test	0.081	02:42	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

Data Download Date/Time: 12/30/2021 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BORCHARDT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/19/2021 08:48:03

Diagnostics Check	OK	08:40
Air Blank	0.000	08:41
Control Test	0.080	08:41
Air Blank	0.000	08:42
Subject Sample # 1	0.124	08:43
Breath Volume	3.429	08:43
Air Blank	0.000	08:44
Air Blank	0.000	08:45
Subject Sample # 2	0.118	08:46
Breath Volume	2.468	08:46
Air Blank	0.000	08:47
Control Test	0.079	08:47
Air Blank	0.000	08:47
Diagnostics Check	OK	08:48

Data Download Date/Time: 12/30/2021 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: VILA

DATE/TIME OF TEST: 12/19/2021 20:54:45

Diagnostics Check	OK	20:51	
Air Blank	0.000	20:52	
Control Test	0.080	20:52	
Air Blank	0.000	20:53	
Subject Sample	# 1	20:53	Subject Test Refused
Breath Volume	0.000	20:53	
Air Blank	0.000	20:53	
Control Test	0.080	20:54	
Air Blank	0.000	20:54	
Diagnostics Check	OK	20:54	

Data Download Date/Time: 12/30/2021 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARRISH,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 12/20/2021 03:08:17

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:05	
Control Test	0.081	03:06	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:06	Subject Test Refused
Breath Volume	0.000	03:06	
Air Blank	0.000	03:07	
Control Test	0.080	03:07	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 12/30/2021 09:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: DAVENPORT,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: TARTER

DATE/TIME OF TEST: 12/21/2021 02:26:42

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.011	02:22
Breath Volume	3.164	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.010	02:25
Breath Volume	3.820	02:25
Air Blank	0.000	02:25
Control Test	0.081	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 12/30/2021 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: PERNELL

DATE/TIME OF TEST: 12/22/2021 22:49:52

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.081	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.172	22:42
Breath Volume	1.230	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.167	22:48
Breath Volume	1.246	22:48
Air Blank	0.000	22:48
Control Test	0.080	22:49
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 12/30/2021 09:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CROUCH,H
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 12/23/2021 00:06:10

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:03	
Control Test	0.080	00:04	
Air Blank	0.000	00:04	
Subject Sample	# 1	00:04	Subject Test Refused
Breath Volume	0.000	00:04	
Air Blank	0.000	00:05	
Control Test	0.080	00:05	
Air Blank	0.000	00:06	
Diagnostics Check	OK	00:06	

Data Download Date/Time: 12/30/2021 09:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville Beach
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/23/2021 03:07:08

Diagnostics Check	OK	03:04	
Air Blank	0.000	03:04	
Control Test	0.080	03:05	
Air Blank	0.000	03:05	
Subject Sample	# 1	03:05	Subject Test Refused
Breath Volume	0.000	03:05	
Air Blank	0.000	03:06	
Control Test	0.080	03:06	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

Data Download Date/Time: 12/30/2021 09:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/24/2021 04:31:54

Diagnostics Check	OK	04:23
Air Blank	0.000	04:24
Control Test	0.081	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.233	04:27
Breath Volume	1.406	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.227	04:30
Breath Volume	2.289	04:30
Air Blank	0.000	04:30
Control Test	0.081	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 12/30/2021 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGGINS,R,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROUNTREE,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: RIVERS

DATE/TIME OF TEST: 12/26/2021 00:41:18

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.081	00:35
Air Blank	0.000	00:35
Subject Sample # 1	0.144	00:36
Breath Volume	1.312	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.140	00:39
Breath Volume	1.304	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 12/30/2021 09:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DELBROCCO-BLUES,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: WERTS

DATE/TIME OF TEST: 12/26/2021 01:22:39

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:20	
Control Test	0.081	01:20	
Air Blank	0.000	01:21	
Subject Sample	# 1	01:21	Subject Test Refused
Breath Volume	0.000	01:21	
Air Blank	0.000	01:21	
Control Test	0.080	01:22	
Air Blank	0.000	01:22	
Diagnostics Check	OK	01:22	

Data Download Date/Time: 12/30/2021 09:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGGINS,R,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 12/26/2021 06:16:59

Diagnostics Check	OK	06:14	
Air Blank	0.000	06:14	
Control Test	0.081	06:14	
Air Blank	0.000	06:15	
Subject Sample	# 1	06:15	Subject Test Refused
Breath Volume	0.000	06:15	
Air Blank	0.000	06:16	
Control Test	0.080	06:16	
Air Blank	0.000	06:16	
Diagnostics Check	OK	06:16	

Data Download Date/Time: 12/30/2021 09:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: DURDEN

DATE/TIME OF TEST: 12/27/2021 01:57:06

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.081	01:43	
Air Blank	0.000	01:44	
Subject Sample	# 1 0.080	01:47	Volume Not Met
Breath Volume	0.687	01:47	
Air Blank	0.000	01:48	
Air Blank	0.000	01:49	
Subject Sample	# 2 0.105	01:50	
Breath Volume	1.410	01:50	
Air Blank	0.000	01:51	
Air Blank	0.000	01:53	
Subject Sample	# 3 0.104	01:55	
Breath Volume	2.363	01:55	
Air Blank	0.000	01:56	
Control Test	0.080	01:56	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

Data Download Date/Time: 12/30/2021 09:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: WROBLEWSKI

DATE/TIME OF TEST: 12/28/2021 00:40:32

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.079	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.265	00:35
Breath Volume	1.859	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.262	00:38
Breath Volume	1.527	00:38
Air Blank	0.000	00:39
Control Test	0.080	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 12/30/2021 09:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGGINS,R,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAINES,J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SWAIN

DATE/TIME OF TEST: 12/28/2021 02:38:33

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.078	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.251	02:33
Breath Volume	2.156	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:35
Subject Sample # 2	0.243	02:36
Breath Volume	1.375	02:36
Air Blank	0.000	02:37
Control Test	0.079	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 12/30/2021 09:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: GOSSELIN

DATE/TIME OF TEST: 12/29/2021 00:07:48

Diagnostics Check	OK	00:00
Air Blank	0.000	00:00
Control Test	0.080	00:01
Air Blank	0.000	00:01
Subject Sample # 1	0.155	00:02
Breath Volume	1.199	00:02
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.138	00:06
Breath Volume	1.695	00:06
Air Blank	0.000	00:06
Control Test	0.080	00:07
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 12/30/2021 09:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/30/2021 03:07:44

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.081	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.105	03:03
Breath Volume	2.398	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.107	03:06
Breath Volume	1.527	03:06
Air Blank	0.000	03:06
Control Test	0.080	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 12/30/2021 09:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: HYAMS

DATE/TIME OF TEST: 12/02/2021 01:46:17

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:38	
Control Test	0.080	01:38	
Air Blank	0.000	01:38	
Subject Sample	# 1	01:42	No Sample Provided
Breath Volume	0.000	01:42	
Air Blank	0.000	01:42	
Air Blank	0.000	01:44	
Subject Sample	# 2	01:44	Subject Test Refused
Breath Volume	0.000	01:44	
Air Blank	0.000	01:45	
Control Test	0.080	01:45	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 12/27/2021 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,MISTY,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: FLOYD

DATE/TIME OF TEST: 12/04/2021 01:46:25

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:35	
Control Test	0.080	01:35	
Air Blank	0.000	01:36	
Subject Sample	# 1	01:39	No Sample Provided
Breath Volume	0.000	01:39	
Air Blank	0.000	01:39	
Air Blank	0.000	01:41	
Subject Sample	# 2	01:45	No Sample Provided
Breath Volume	0.000	01:45	
Air Blank	0.000	01:45	
Control Test	0.080	01:45	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 12/27/2021 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,MISTY,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SIEMINSKI

DATE/TIME OF TEST: 12/05/2021 03:42:48

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.202	03:37
Breath Volume	2.296	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.197	03:41
Breath Volume	2.089	03:41
Air Blank	0.000	03:41
Control Test	0.079	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 12/27/2021 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,MISTY,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: HULLETT

DATE/TIME OF TEST: 12/06/2021 01:43:49

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.174	01:39
Breath Volume	1.343	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.171	01:42
Breath Volume	3.476	01:42
Air Blank	0.000	01:42
Control Test	0.079	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 12/27/2021 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LARAWAY,D
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: AMBURGEY

DATE/TIME OF TEST: 12/06/2021 22:48:23

Diagnostics Check	OK	22:41
Air Blank	0.000	22:42
Control Test	0.079	22:42
Air Blank	0.000	22:42
Subject Sample # 1	0.040	22:43
Breath Volume	4.203	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.041	22:46
Breath Volume	4.113	22:46
Air Blank	0.000	22:47
Control Test	0.079	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 12/27/2021 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,MISTY,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 12/08/2021 19:41:59

Diagnostics Check	OK	19:35
Air Blank	0.000	19:36
Control Test	0.079	19:36
Air Blank	0.000	19:36
Subject Sample # 1	0.111	19:37
Breath Volume	2.027	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:39
Subject Sample # 2	0.105	19:40
Breath Volume	1.640	19:40
Air Blank	0.000	19:41
Control Test	0.079	19:41
Air Blank	0.000	19:41
Diagnostics Check	OK	19:41

Data Download Date/Time: 12/27/2021 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,MISTY,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: HASTON

DATE/TIME OF TEST: 12/09/2021 21:55:45

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:46	
Control Test	0.080	21:47	
Air Blank	0.000	21:47	
Subject Sample	# 1	21:50	No Sample Provided
Breath Volume	0.000	21:50	
Air Blank	0.000	21:51	
Air Blank	0.000	21:53	
Subject Sample	# 2	21:54	Subject Test Refused
Breath Volume	0.000	21:54	
Air Blank	0.000	21:54	
Control Test	0.080	21:55	
Air Blank	0.000	21:55	
Diagnostics Check	OK	21:55	

Data Download Date/Time: 12/27/2021 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,MISTY,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: RAINEY

DATE/TIME OF TEST: 12/10/2021 02:06:17

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:58	
Control Test	0.079	01:58	
Air Blank	0.000	01:58	
Subject Sample	# 1	02:02	No Sample Provided
Breath Volume	0.000	02:02	
Air Blank	0.000	02:02	
Air Blank	0.000	02:04	
Subject Sample	# 2	02:04	Subject Test Refused
Breath Volume	0.000	02:04	
Air Blank	0.000	02:05	
Control Test	0.080	02:05	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:06	

Data Download Date/Time: 12/27/2021 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: SWOGGER

DATE/TIME OF TEST: 12/12/2021 02:40:43

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:32	
Control Test	0.079	02:32	
Air Blank	0.000	02:33	
Subject Sample	# 1	02:36	No Sample Provided
Breath Volume	0.000	02:36	
Air Blank	0.000	02:37	
Air Blank	0.000	02:39	
Subject Sample	# 2	02:39	Subject Test Refused
Breath Volume	0.000	02:39	
Air Blank	0.000	02:39	
Control Test	0.079	02:40	
Air Blank	0.000	02:40	
Diagnostics Check	OK	02:40	

Data Download Date/Time: 12/27/2021 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: ALLUMS JR

DATE/TIME OF TEST: 12/12/2021 03:51:46

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.079	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.131	03:47
Breath Volume	2.308	03:47
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.122	03:50
Breath Volume	1.496	03:50
Air Blank	0.000	03:50
Control Test	0.079	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 12/27/2021 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CAMP,B
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 12/13/2021 00:54:25

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.079	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.247	00:49
Breath Volume	1.812	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.233	00:52
Breath Volume	2.398	00:52
Air Blank	0.000	00:53
Control Test	0.079	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 12/27/2021 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAERTZ, MARK, J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: COUSSENS

DATE/TIME OF TEST: 12/13/2021 02:09:52

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:01	
Control Test	0.080	02:02	
Air Blank	0.000	02:02	
Subject Sample	# 1	02:05	No Sample Provided
Breath Volume	0.000	02:05	
Air Blank	0.000	02:06	
Air Blank	0.000	02:08	
Subject Sample	# 2	02:08	Subject Test Refused
Breath Volume	0.000	02:08	
Air Blank	0.000	02:08	
Control Test	0.080	02:09	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

Data Download Date/Time: 12/27/2021 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,MISTY,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: MAAS

DATE/TIME OF TEST: 12/14/2021 01:26:57

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:18	
Control Test	0.080	01:18	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:22	No Sample Provided
Breath Volume	0.000	01:22	
Air Blank	0.000	01:23	
Air Blank	0.000	01:25	
Subject Sample	# 2	01:25	Subject Test Refused
Breath Volume	0.000	01:25	
Air Blank	0.000	01:26	
Control Test	0.080	01:26	
Air Blank	0.000	01:26	
Diagnostics Check	OK	01:26	

Data Download Date/Time: 12/27/2021 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DYKE,T,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: BILLUPS

DATE/TIME OF TEST: 12/15/2021 01:29:10

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:20	
Control Test	0.080	01:21	
Air Blank	0.000	01:21	
Subject Sample	# 1	01:24	No Sample Provided
Breath Volume	0.000	01:24	
Air Blank	0.000	01:25	
Air Blank	0.000	01:27	
Subject Sample	# 2	01:27	Subject Test Refused
Breath Volume	0.000	01:27	
Air Blank	0.000	01:28	
Control Test	0.080	01:28	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

Data Download Date/Time: 12/27/2021 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,J,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITMAN,W
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:38

SUBJECT'S LAST NAME: PALATO

DATE/TIME OF TEST: 12/15/2021 12:24:25

Diagnostics Check	OK	12:16	
Air Blank	0.000	12:16	
Control Test	0.080	12:16	
Air Blank	0.000	12:17	
Subject Sample	# 1	12:20	No Sample Provided
Breath Volume	0.000	12:20	
Air Blank	0.000	12:20	
Air Blank	0.000	12:22	
Subject Sample	# 2	12:23	Subject Test Refused
Breath Volume	0.000	12:23	
Air Blank	0.000	12:23	
Control Test	0.079	12:23	
Air Blank	0.000	12:24	
Diagnostics Check	OK	12:24	

Data Download Date/Time: 12/27/2021 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CAMP,B
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: WELBES

DATE/TIME OF TEST: 12/16/2021 00:48:49

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:40	
Control Test	0.079	00:41	
Air Blank	0.000	00:41	
Subject Sample	# 1	00:44	No Sample Provided
Breath Volume	0.000	00:44	
Air Blank	0.000	00:45	
Air Blank	0.000	00:47	
Subject Sample	# 2	00:47	Subject Test Refused
Breath Volume	0.000	00:47	
Air Blank	0.000	00:47	
Control Test	0.079	00:48	
Air Blank	0.000	00:48	
Diagnostics Check	OK	00:48	

Data Download Date/Time: 12/27/2021 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAERTZ, MARK, J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS, S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 12/16/2021 02:45:55

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.321	02:41
Breath Volume	1.550	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.327	02:44
Breath Volume	3.933	02:44
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 12/27/2021 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,J,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SMITH,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:58

SUBJECT'S LAST NAME: HICKMAN

DATE/TIME OF TEST: 12/16/2021 11:31:46

Diagnostics Check	OK	11:22
Air Blank	0.000	11:23
Control Test	0.079	11:23
Air Blank	0.000	11:24
Subject Sample # 1	0.233	11:26
Breath Volume	1.207	11:26
Air Blank	0.000	11:27
Air Blank	0.000	11:29
Subject Sample # 2	0.238	11:30
Breath Volume	1.781	11:30
Air Blank	0.000	11:30
Control Test	0.078	11:31
Air Blank	0.000	11:31
Diagnostics Check	OK	11:31

Data Download Date/Time: 12/27/2021 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: AUGUST JR

DATE/TIME OF TEST: 12/17/2021 01:10:47

Diagnostics Check	OK	01:03
Air Blank	0.000	01:04
Control Test	0.080	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.211	01:05
Breath Volume	3.105	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.216	01:09
Breath Volume	2.042	01:09
Air Blank	0.000	01:09
Control Test	0.079	01:10
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 12/27/2021 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,MISTY,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/18/2021 01:24:00

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.080	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.140	01:19
Breath Volume	2.261	01:19
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.132	01:22
Breath Volume	2.222	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:24

Data Download Date/Time: 12/27/2021 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DYKE,T,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WILLIAMS,J
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: LINTON

DATE/TIME OF TEST: 12/18/2021 03:33:24

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.189	03:29
Breath Volume	1.683	03:29
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.191	03:31
Breath Volume	1.542	03:31
Air Blank	0.000	03:32
Control Test	0.080	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 12/27/2021 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,MISTY,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: BARBER

DATE/TIME OF TEST: 12/18/2021 22:15:54

Diagnostics Check	OK	22:06	
Air Blank	0.000	22:07	
Control Test	0.080	22:07	
Air Blank	0.000	22:07	
Subject Sample	# 1	22:11	No Sample Provided
Breath Volume	0.000	22:11	
Air Blank	0.000	22:11	
Air Blank	0.000	22:13	
Subject Sample	# 2	22:14	Subject Test Refused
Breath Volume	0.000	22:14	
Air Blank	0.000	22:15	
Control Test	0.080	22:15	
Air Blank	0.000	22:15	
Diagnostics Check	OK	22:15	

Data Download Date/Time: 12/27/2021 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAERTZ, MARK, J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS, S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: HALTER

DATE/TIME OF TEST: 12/21/2021 19:56:43

Diagnostics Check	OK	19:48	
Air Blank	0.000	19:48	
Control Test	0.080	19:49	
Air Blank	0.000	19:49	
Subject Sample	# 1	19:52	No Sample Provided
Breath Volume	0.000	19:52	
Air Blank	0.000	19:53	
Air Blank	0.000	19:55	
Subject Sample	# 2	19:55	Subject Test Refused
Breath Volume	0.000	19:55	
Air Blank	0.000	19:55	
Control Test	0.081	19:56	
Air Blank	0.000	19:56	
Diagnostics Check	OK	19:56	

Data Download Date/Time: 12/27/2021 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: THOMAS,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: ASSAR

DATE/TIME OF TEST: 12/22/2021 02:57:22

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:49	
Control Test	0.080	02:49	
Air Blank	0.000	02:50	
Subject Sample	# 1	02:53	No Sample Provided
Breath Volume	0.000	02:53	
Air Blank	0.000	02:53	
Air Blank	0.000	02:55	
Subject Sample	# 2	02:56	Subject Test Refused
Breath Volume	0.000	02:56	
Air Blank	0.000	02:56	
Control Test	0.080	02:56	
Air Blank	0.000	02:57	
Diagnostics Check	OK	02:57	

Data Download Date/Time: 12/27/2021 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,MISTY,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CLEMONS,C
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 12/23/2021 01:24:53

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:13	
Control Test	0.081	01:13	
Air Blank	0.000	01:14	
Subject Sample	# 1 0.146	01:17	Volume Not Met
Breath Volume	0.699	01:17	
Air Blank	0.000	01:18	
Air Blank	0.000	01:20	
Subject Sample	# 2 0.154	01:23	Volume Not Met
Breath Volume	0.207	01:23	
Air Blank	0.000	01:23	
Control Test	0.079	01:24	
Air Blank	0.000	01:24	
Diagnostics Check	OK	01:24	

Data Download Date/Time: 12/27/2021 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,MISTY,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: HENSCHKE

DATE/TIME OF TEST: 12/24/2021 00:33:26

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.080	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.000	00:29
Breath Volume	2.160	00:29
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.000	00:32
Breath Volume	3.113	00:32
Air Blank	0.000	00:32
Control Test	0.081	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 12/27/2021 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: FITCH

DATE/TIME OF TEST: 12/25/2021 23:27:55

Diagnostics Check	OK	23:19	
Air Blank	0.000	23:19	
Control Test	0.080	23:20	
Air Blank	0.000	23:20	
Subject Sample	# 1	23:23	No Sample Provided
Breath Volume	0.000	23:23	
Air Blank	0.000	23:24	
Air Blank	0.000	23:26	
Subject Sample	# 2	23:26	Subject Test Refused
Breath Volume	0.000	23:26	
Air Blank	0.000	23:27	
Control Test	0.080	23:27	
Air Blank	0.000	23:27	
Diagnostics Check	OK	23:27	

Data Download Date/Time: 12/27/2021 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,MISTY,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: DUBOVIS

DATE/TIME OF TEST: 12/27/2021 19:16:48

Diagnostics Check	OK	19:09
Air Blank	0.000	19:10
Control Test	0.080	19:10
Air Blank	0.000	19:10
Subject Sample # 1	0.221	19:12
Breath Volume	1.558	19:12
Air Blank	0.000	19:13
Air Blank	0.000	19:14
Subject Sample # 2	0.225	19:15
Breath Volume	1.308	19:15
Air Blank	0.000	19:15
Control Test	0.080	19:16
Air Blank	0.000	19:16
Diagnostics Check	OK	19:16

Data Download Date/Time: 01/27/2022 19:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DYKE,T,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 12/28/2021 00:10:46

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:02	
Control Test	0.080	00:02	
Air Blank	0.000	00:03	
Subject Sample	# 1	00:06	No Sample Provided
Breath Volume	0.000	00:06	
Air Blank	0.000	00:07	
Air Blank	0.000	00:09	
Subject Sample	# 2	00:09	Subject Test Refused
Breath Volume	0.000	00:09	
Air Blank	0.000	00:09	
Control Test	0.080	00:10	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 01/27/2022 19:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANDYKE,T,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: DAY

DATE/TIME OF TEST: 12/29/2021 02:00:52

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:52	
Control Test	0.080	01:53	
Air Blank	0.000	01:53	
Subject Sample	# 1	01:56	No Sample Provided
Breath Volume	0.000	01:56	
Air Blank	0.000	01:57	
Air Blank	0.000	01:59	
Subject Sample	# 2	01:59	Subject Test Refused
Breath Volume	0.000	01:59	
Air Blank	0.000	01:59	
Control Test	0.080	02:00	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

Data Download Date/Time: 01/27/2022 19:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DYKE,T,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 12/29/2021 03:30:01

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.080	03:21	
Air Blank	0.000	03:22	
Subject Sample	# 1	03:25	No Sample Provided
Breath Volume	0.000	03:25	
Air Blank	0.000	03:26	
Air Blank	0.000	03:28	
Subject Sample	# 2	03:28	Subject Test Refused
Breath Volume	0.000	03:28	
Air Blank	0.000	03:29	
Control Test	0.080	03:29	
Air Blank	0.000	03:29	
Diagnostics Check	OK	03:30	

Data Download Date/Time: 01/27/2022 19:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,MISTY,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: HESS

DATE/TIME OF TEST: 12/31/2021 23:36:07

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.080	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.226	23:31
Breath Volume	3.320	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:33
Subject Sample # 2	0.230	23:34
Breath Volume	4.035	23:34
Air Blank	0.000	23:35
Control Test	0.080	23:35
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 01/27/2022 19:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,BRODY,A
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: GIRALDO,E
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 11/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: NORCIUS

DATE/TIME OF TEST: 12/04/2021 08:56:35

Diagnostics Check	OK	08:50
Air Blank	0.000	08:50
Control Test	0.082	08:51
Air Blank	0.000	08:51
Subject Sample # 1	0.121	08:52
Breath Volume	2.328	08:52
Air Blank	0.000	08:52
Air Blank	0.000	08:54
Subject Sample # 2	0.111	08:55
Breath Volume	2.316	08:55
Air Blank	0.000	08:55
Control Test	0.080	08:55
Air Blank	0.000	08:56
Diagnostics Check	OK	08:56

Data Download Date/Time: 12/28/2021 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,ALEXANDER,J
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WAYE,MAKINA
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 11/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: HILDAGO MARQUEZ

DATE/TIME OF TEST: 12/17/2021 22:54:36

Diagnostics Check	OK	22:51	
Air Blank	0.000	22:52	
Control Test	0.080	22:52	
Air Blank	0.000	22:53	
Subject Sample	# 1	22:53	Subject Test Refused
Breath Volume	0.000	22:53	
Air Blank	0.000	22:53	
Control Test	0.081	22:54	
Air Blank	0.000	22:54	
Diagnostics Check	OK	22:54	

Data Download Date/Time: 12/28/2021 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDES,DEREK,
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: KEHR,JOSHUA
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 11/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:46

SUBJECT'S LAST NAME: BARRIGA QUINTERO

DATE/TIME OF TEST: 12/21/2021 09:15:09

Diagnostics Check	OK	09:08
Air Blank	0.000	09:09
Control Test	0.080	09:09
Air Blank	0.000	09:09
Subject Sample # 1	0.159	09:10
Breath Volume	2.195	09:10
Air Blank	0.000	09:11
Air Blank	0.000	09:13
Subject Sample # 2	0.152	09:13
Breath Volume	1.230	09:13
Air Blank	0.000	09:14
Control Test	0.079	09:14
Air Blank	0.000	09:15
Diagnostics Check	OK	09:15

Data Download Date/Time: 12/28/2021 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: ALANIS

DATE/TIME OF TEST: 12/08/2021 23:40:12

Diagnostics Check	OK	23:34
Air Blank	0.000	23:34
Control Test	0.079	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.125	23:35
Breath Volume	2.046	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.123	23:38
Breath Volume	2.511	23:38
Air Blank	0.000	23:39
Control Test	0.079	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 12/23/2021 08:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUSAM,MATTHEW,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,C.
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: IRWIN

DATE/TIME OF TEST: 12/10/2021 23:56:00

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:47	
Control Test	0.079	23:48	
Air Blank	0.000	23:48	
Subject Sample	# 1	23:51	No Sample Provided
Breath Volume	0.000	23:51	
Air Blank	0.000	23:52	
Air Blank	0.000	23:54	
Subject Sample	# 2	23:54	Subject Test Refused
Breath Volume	0.000	23:54	
Air Blank	0.000	23:55	
Control Test	0.078	23:55	
Air Blank	0.000	23:55	
Diagnostics Check	OK	23:56	

Data Download Date/Time: 12/23/2021 08:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: BORRELI

DATE/TIME OF TEST: 12/11/2021 03:09:21

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.131	03:04
Breath Volume	1.421	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.128	03:07
Breath Volume	1.871	03:07
Air Blank	0.000	03:08
Control Test	0.078	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 12/23/2021 08:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: AMOS

DATE/TIME OF TEST: 12/17/2021 00:43:00

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:37	
Control Test	0.079	00:37	
Air Blank	0.000	00:38	
Subject Sample	# 1 0.245	00:38	
Breath Volume	1.218	00:38	
Air Blank	0.000	00:39	
Air Blank	0.000	00:41	
Subject Sample	# 2	00:41	RFI Detect
Breath Volume	1.285	00:41	
Air Blank	0.000	00:42	
Control Test	0.078	00:42	Interferent Detect
Air Blank	0.000	00:43	

Data Download Date/Time: 12/23/2021 08:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: AMOS

DATE/TIME OF TEST: 12/17/2021 00:52:34

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.079	00:47
Air Blank	0.000	00:47
Subject Sample # 1	0.234	00:48
Breath Volume	1.574	00:48
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.230	00:51
Breath Volume	1.433	00:51
Air Blank	0.000	00:51
Control Test	0.078	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 12/23/2021 08:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: HOBBS

DATE/TIME OF TEST: 12/18/2021 00:06:48

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:01
Subject Sample # 1	0.154	00:02
Breath Volume	2.769	00:02
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.155	00:05
Breath Volume	2.316	00:05
Air Blank	0.000	00:05
Control Test	0.079	00:06
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 12/23/2021 08:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: VENTURA

DATE/TIME OF TEST: 12/21/2021 20:37:20

Diagnostics Check	OK	20:30
Air Blank	0.000	20:31
Control Test	0.080	20:31
Air Blank	0.000	20:32
Subject Sample # 1	0.213	20:32
Breath Volume	2.570	20:32
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample # 2	0.206	20:35
Breath Volume	2.531	20:35
Air Blank	0.000	20:36
Control Test	0.079	20:36
Air Blank	0.000	20:37
Diagnostics Check	OK	20:37

Data Download Date/Time: 12/23/2021 08:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: NIEVES

DATE/TIME OF TEST: 12/24/2021 01:26:03

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.078	01:21
Breath Volume	1.574	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.077	01:24
Breath Volume	1.894	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:25
Diagnostics Check	OK	01:26

Data Download Date/Time: 01/07/2022 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 12/24/2021 01:59:13

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.080	01:54
Air Blank	0.000	01:54
Subject Sample # 1	0.000	01:54
Breath Volume	2.156	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.000	01:57
Breath Volume	1.574	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 01/07/2022 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: DUNBAR

DATE/TIME OF TEST: 12/30/2021 23:06:13

Diagnostics Check	OK	23:03	
Air Blank	0.000	23:03	
Control Test	0.077	23:03	
Air Blank	0.000	23:04	
Subject Sample	# 1	23:04	Subject Test Refused
Breath Volume	0.000	23:04	
Air Blank	0.000	23:05	
Control Test	0.077	23:05	
Air Blank	0.000	23:06	
Diagnostics Check	OK	23:06	

Data Download Date/Time: 01/07/2022 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: ROMAN VALDES

DATE/TIME OF TEST: 12/31/2021 16:56:51

Diagnostics Check	OK	16:50
Air Blank	0.000	16:50
Control Test	0.079	16:51
Air Blank	0.000	16:51
Subject Sample # 1	0.000	16:52
Breath Volume	2.261	16:52
Air Blank	0.000	16:52
Air Blank	0.000	16:54
Subject Sample # 2	0.000	16:55
Breath Volume	2.320	16:55
Air Blank	0.000	16:55
Control Test	0.079	16:56
Air Blank	0.000	16:56
Diagnostics Check	OK	16:56

Data Download Date/Time: 01/07/2022 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIPPINS,C,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WOODS,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/04/2021 02:20:59

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:18	
Control Test	0.081	02:18	
Air Blank	0.000	02:19	
Subject Sample	# 1	02:19	Subject Test Refused
Breath Volume	0.000	02:19	
Air Blank	0.000	02:20	
Control Test	0.081	02:20	
Air Blank	0.000	02:20	
Diagnostics Check	OK	02:20	

Data Download Date/Time: 12/22/2021 03:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIPPINS,C,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WOODS,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: DILLON

DATE/TIME OF TEST: 12/10/2021 00:46:23

Diagnostics Check	OK	00:43	
Air Blank	0.000	00:44	
Control Test	0.081	00:44	
Air Blank	0.000	00:44	
Subject Sample	# 1	00:45	Subject Test Refused
Breath Volume	0.000	00:45	
Air Blank	0.000	00:45	
Control Test	0.081	00:45	
Air Blank	0.000	00:46	
Diagnostics Check	OK	00:46	

Data Download Date/Time: 12/22/2021 03:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIPPINS,C,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HOLTON,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: HO

DATE/TIME OF TEST: 12/16/2021 00:55:23

Diagnostics Check	OK	00:49
Air Blank	0.000	00:49
Control Test	0.080	00:50
Air Blank	0.000	00:50
Subject Sample # 1	0.190	00:50
Breath Volume	1.179	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.198	00:53
Breath Volume	1.144	00:53
Air Blank	0.000	00:54
Control Test	0.079	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 12/22/2021 03:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNEAD,CODY,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: CARPENTER,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: JARRELL

DATE/TIME OF TEST: 12/16/2021 20:24:05

Diagnostics Check	OK	20:17
Air Blank	0.000	20:17
Control Test	0.079	20:18
Air Blank	0.000	20:18
Subject Sample # 1	0.227	20:19
Breath Volume	2.500	20:19
Air Blank	0.000	20:19
Air Blank	0.000	20:21
Subject Sample # 2	0.232	20:22
Breath Volume	3.386	20:22
Air Blank	0.000	20:23
Control Test	0.079	20:23
Air Blank	0.000	20:24
Diagnostics Check	OK	20:24

Data Download Date/Time: 12/22/2021 03:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,J,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GOETHE,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: CREWS

DATE/TIME OF TEST: 12/17/2021 23:42:03

Diagnostics Check	OK	23:34
Air Blank	0.000	23:35
Control Test	0.080	23:35
Air Blank	0.000	23:35
Subject Sample # 1	0.000	23:37
Breath Volume	2.535	23:37
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.000	23:40
Breath Volume	2.062	23:40
Air Blank	0.000	23:41
Control Test	0.081	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:42

Data Download Date/Time: 12/22/2021 03:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WOODS,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 12/22/2021 20:26:54

Diagnostics Check	OK	20:20
Air Blank	0.000	20:21
Control Test	0.081	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.124	20:22
Breath Volume	2.453	20:22
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.116	20:25
Breath Volume	2.273	20:25
Air Blank	0.000	20:26
Control Test	0.079	20:26
Air Blank	0.000	20:26
Diagnostics Check	OK	20:26

Data Download Date/Time: 01/25/2022 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIPPINS,C,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WOODS,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 12/23/2021 03:52:32

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.080	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.205	03:48
Breath Volume	2.007	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.206	03:51
Breath Volume	1.531	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 01/25/2022 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DILL,T
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: GRALEY

DATE/TIME OF TEST: 12/24/2021 01:26:23

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:23	
Control Test	0.081	01:23	
Air Blank	0.000	01:24	
Subject Sample	# 1	01:25	Subject Test Refused
Breath Volume	0.000	01:25	
Air Blank	0.000	01:25	
Control Test	0.081	01:25	
Air Blank	0.000	01:26	
Diagnostics Check	OK	01:26	

Data Download Date/Time: 01/25/2022 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIPPINS,C,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WOODS,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 12/24/2021 03:37:22

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.081	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.197	03:33
Breath Volume	1.414	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.201	03:35
Breath Volume	1.121	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 01/25/2022 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIPPINS,C,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GOETHE,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: BOYLAN

DATE/TIME OF TEST: 12/28/2021 01:55:13

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.000	01:50
Breath Volume	1.156	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.000	01:53
Breath Volume	1.312	01:53
Air Blank	0.000	01:54
Control Test	0.081	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 01/25/2022 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PICKETT,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MILLAN,S
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PAGE

DATE/TIME OF TEST: 12/28/2021 03:57:29

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.080	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.193	03:52
Breath Volume	3.238	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.191	03:56
Breath Volume	2.371	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 01/25/2022 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLOUNT,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SPIVEY,D
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 11:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/29/2021 01:24:46

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.080	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.092	01:20
Breath Volume	1.597	01:20
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.106	01:23
Breath Volume	3.472	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:24
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 01/25/2022 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNEAD,CODY,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HOLTON,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: WHITTINGTON

DATE/TIME OF TEST: 12/30/2021 23:45:50

Diagnostics Check	OK	23:38
Air Blank	0.000	23:38
Control Test	0.080	23:39
Air Blank	0.000	23:39
Subject Sample # 1	0.000	23:40
Breath Volume	3.914	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.000	23:44
Breath Volume	2.328	23:44
Air Blank	0.000	23:44
Control Test	0.080	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 01/25/2022 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: ROLDAN-MARTINEZ

DATE/TIME OF TEST: 12/06/2021 20:07:05

Diagnostics Check	OK	19:59
Air Blank	0.000	20:00
Control Test	0.082	20:00
Air Blank	0.000	20:00
Subject Sample # 1	0.134	20:02
Breath Volume	1.507	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:04
Subject Sample # 2	0.135	20:05
Breath Volume	1.324	20:05
Air Blank	0.000	20:06
Control Test	0.081	20:06
Air Blank	0.000	20:06
Diagnostics Check	OK	20:07

Data Download Date/Time: 12/10/2021 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 12/08/2021 02:22:59

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.081	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.000	02:18
Breath Volume	1.722	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.000	02:21
Breath Volume	1.136	02:21
Air Blank	0.000	02:22
Control Test	0.082	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 12/10/2021 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: PLAISIMOND

DATE/TIME OF TEST: 12/08/2021 04:11:01

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.081	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.131	04:05
Breath Volume	1.207	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.136	04:09
Breath Volume	1.429	04:09
Air Blank	0.000	04:10
Control Test	0.081	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:11

Data Download Date/Time: 12/10/2021 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: YARBROUGH

DATE/TIME OF TEST: 12/11/2021 03:24:32

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.080	03:13	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:17	No Sample Provided
Breath Volume	0.000	03:17	
Air Blank	0.000	03:18	
Air Blank	0.000	03:20	
Subject Sample	# 2	03:23	No Sample Provided
Breath Volume	0.000	03:23	
Air Blank	0.000	03:23	
Control Test	0.082	03:23	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

Data Download Date/Time: 01/11/2022 16:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: ARREOLA-OVANDO

DATE/TIME OF TEST: 12/11/2021 22:11:59

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.081	22:05
Air Blank	0.000	22:06
Subject Sample # 1	0.185	22:07
Breath Volume	2.750	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.177	22:10
Breath Volume	1.769	22:10
Air Blank	0.000	22:11
Control Test	0.081	22:11
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 01/11/2022 16:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: GARGUS

DATE/TIME OF TEST: 12/12/2021 03:37:17

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.082	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.165	03:32
Breath Volume	3.109	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.156	03:35
Breath Volume	2.453	03:35
Air Blank	0.000	03:36
Control Test	0.081	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 01/11/2022 16:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: JANEGO

DATE/TIME OF TEST: 12/12/2021 21:54:01

Diagnostics Check	OK	21:47
Air Blank	0.000	21:47
Control Test	0.080	21:48
Air Blank	0.000	21:48
Subject Sample # 1	0.146	21:49
Breath Volume	2.871	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.147	21:52
Breath Volume	2.781	21:52
Air Blank	0.000	21:53
Control Test	0.081	21:53
Air Blank	0.000	21:53
Diagnostics Check	OK	21:54

Data Download Date/Time: 01/11/2022 16:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: GODINEZ

DATE/TIME OF TEST: 12/13/2021 02:12:00

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.081	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.141	02:06
Breath Volume	1.117	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.138	02:10
Breath Volume	1.550	02:10
Air Blank	0.000	02:11
Control Test	0.081	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:12

Data Download Date/Time: 01/11/2022 16:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: PENABADE

DATE/TIME OF TEST: 12/16/2021 20:18:51

Diagnostics Check	OK	20:05	
Air Blank	0.000	20:06	
Control Test	0.080	20:06	
Air Blank	0.000	20:07	
Subject Sample	# 1	20:10	No Sample Provided
Breath Volume	0.000	20:10	
Air Blank	0.000	20:10	
Air Blank	0.000	20:12	
Subject Sample	# 2	0.172	20:14
Breath Volume	2.632	20:14	
Air Blank	0.000	20:15	
Air Blank	0.000	20:16	
Subject Sample	# 3	0.164	20:17
Breath Volume	1.410	20:17	
Air Blank	0.000	20:17	
Control Test	0.081	20:18	
Air Blank	0.000	20:18	
Diagnostics Check	OK	20:18	

Data Download Date/Time: 01/11/2022 16:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE, ABIGAIL, D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PAYNE, ABIGAIL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: HOLLOWAY

DATE/TIME OF TEST: 12/16/2021 21:03:46

Diagnostics Check	OK	20:55
Air Blank	0.000	20:55
Control Test	0.081	20:56
Air Blank	0.000	20:56
Subject Sample # 1	0.090	20:58
Breath Volume	1.335	20:58
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.081	21:02
Breath Volume	1.285	21:02
Air Blank	0.000	21:02
Control Test	0.080	21:03
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 01/11/2022 16:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/17/2021 00:15:59

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:10	
Control Test	0.082	00:10	
Air Blank	0.000	00:11	
Subject Sample	# 1 0.178	00:11	
Breath Volume	2.812	00:11	
Air Blank	0.000	00:12	
Air Blank	0.000	00:14	
Subject Sample	# 2	00:14	RFI Detect
Breath Volume	1.789	00:14	
Air Blank	0.000	00:15	
Control Test	0.082	00:15	
Air Blank	0.000	00:15	
Diagnostics Check	OK	00:15	

Data Download Date/Time: 01/11/2022 16:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/17/2021 00:25:03

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.082	00:19
Air Blank	0.000	00:19
Subject Sample # 1	0.178	00:20
Breath Volume	2.343	00:20
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.178	00:23
Breath Volume	2.683	00:23
Air Blank	0.000	00:24
Control Test	0.081	00:24
Air Blank	0.000	00:24
Diagnostics Check	OK	00:25

Data Download Date/Time: 01/11/2022 16:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: MEGNATH

DATE/TIME OF TEST: 12/21/2021 21:39:50

Diagnostics Check	OK	21:28	
Air Blank	0.000	21:28	
Control Test	0.081	21:29	
Air Blank	0.000	21:29	
Subject Sample	# 1	21:32	No Sample Provided
Breath Volume	0.000	21:32	
Air Blank	0.000	21:33	
Air Blank	0.000	21:35	
Subject Sample	# 2	21:38	No Sample Provided
Breath Volume	0.000	21:38	
Air Blank	0.000	21:38	
Control Test	0.081	21:39	
Air Blank	0.000	21:39	
Diagnostics Check	OK	21:39	

Data Download Date/Time: 01/11/2022 16:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: CELENZA

DATE/TIME OF TEST: 12/22/2021 02:05:36

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:54	
Control Test	0.081	01:55	
Air Blank	0.000	01:55	
Subject Sample	# 1	01:58	No Sample Provided
Breath Volume	0.000	01:58	
Air Blank	0.000	01:59	
Air Blank	0.000	02:01	
Subject Sample	# 2	02:04	No Sample Provided
Breath Volume	0.000	02:04	
Air Blank	0.000	02:04	
Control Test	0.081	02:05	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 01/11/2022 16:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: JEFFERS

DATE/TIME OF TEST: 12/26/2021 01:59:54

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:48	
Control Test	0.081	01:49	
Air Blank	0.000	01:49	
Subject Sample	# 1	01:52	No Sample Provided
Breath Volume	0.000	01:52	
Air Blank	0.000	01:53	
Air Blank	0.000	01:55	
Subject Sample	# 2	01:58	No Sample Provided
Breath Volume	0.000	01:58	
Air Blank	0.000	01:58	
Control Test	0.081	01:59	
Air Blank	0.000	01:59	
Diagnostics Check	OK	01:59	

Data Download Date/Time: 01/11/2022 16:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE, ABIGAIL, D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PAYNE, ABIGAIL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: ARMLIN

DATE/TIME OF TEST: 12/26/2021 04:59:23

Diagnostics Check	OK	04:47	
Air Blank	0.000	04:47	
Control Test	0.080	04:47	
Air Blank	0.000	04:48	
Subject Sample	# 1	04:51	No Sample Provided
Breath Volume	0.000	04:51	
Air Blank	0.000	04:51	
Air Blank	0.000	04:53	
Subject Sample	# 2	0.165	04:54
Breath Volume	1.156	04:54	
Air Blank	0.000	04:55	
Air Blank	0.000	04:57	
Subject Sample	# 3	0.152	04:57
Breath Volume	1.312	04:57	
Air Blank	0.000	04:58	
Control Test	0.080	04:58	
Air Blank	0.000	04:59	
Diagnostics Check	OK	04:59	

Data Download Date/Time: 01/11/2022 16:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 12/27/2021 02:53:39

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.080	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.203	02:48
Breath Volume	2.167	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.213	02:52
Breath Volume	2.113	02:52
Air Blank	0.000	02:52
Control Test	0.081	02:53
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 01/11/2022 16:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE, ABIGAIL, D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PAYNE, ABIGAIL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 12/30/2021 01:06:10

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.080	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.199	01:01
Breath Volume	1.675	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.197	01:04
Breath Volume	1.468	01:04
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 01/11/2022 16:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: UPTHEGROVE

DATE/TIME OF TEST: 12/30/2021 03:21:22

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:10	
Control Test	0.080	03:10	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:14	No Sample Provided
Breath Volume	0.000	03:14	
Air Blank	0.000	03:14	
Air Blank	0.000	03:16	
Subject Sample	# 2	03:20	No Sample Provided
Breath Volume	0.000	03:20	
Air Blank	0.000	03:20	
Control Test	0.081	03:20	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 01/11/2022 16:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE, ABIGAIL, D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PAYNE, ABIGAIL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: AVILES DORANTES

DATE/TIME OF TEST: 12/30/2021 21:59:02

Diagnostics Check	OK	21:52
Air Blank	0.000	21:53
Control Test	0.081	21:53
Air Blank	0.000	21:54
Subject Sample # 1	0.161	21:54
Breath Volume	1.109	21:54
Air Blank	0.000	21:55
Air Blank	0.000	21:57
Subject Sample # 2	0.178	21:57
Breath Volume	1.546	21:57
Air Blank	0.000	21:58
Control Test	0.081	21:58
Air Blank	0.000	21:58
Diagnostics Check	OK	21:59

Data Download Date/Time: 01/11/2022 16:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ALBERT,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WILLIAMS,ALBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: MOLNAR

DATE/TIME OF TEST: 12/30/2021 23:30:19

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.081	23:23
Air Blank	0.000	23:24
Subject Sample # 1	0.165	23:25
Breath Volume	1.343	23:25
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.181	23:28
Breath Volume	1.402	23:28
Air Blank	0.000	23:29
Control Test	0.082	23:29
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 01/11/2022 16:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: MARTIN-RAMIREZ

DATE/TIME OF TEST: 12/30/2021 23:41:36

Diagnostics Check	OK	23:34	
Air Blank	0.000	23:34	
Control Test	0.080	23:35	
Air Blank	0.000	23:35	
Subject Sample # 1	0.117	23:36	
Breath Volume	1.375	23:36	
Air Blank	0.000	23:36	
Air Blank	0.000	23:38	
Subject Sample # 2	0.125	23:39	
Breath Volume	2.398	23:39	
Air Blank	0.000	23:40	
Control Test	0.081	23:40	
Air Blank	0.000	23:41	
Diagnostics Check	Fail	23:41	Temperature Regulation Test
Air Blank	0.000	23:41	

Data Download Date/Time: 01/11/2022 16:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: MARTIN-RAMIREZ

DATE/TIME OF TEST: 12/30/2021 23:43:44

Diagnostics Check	Fail	23:43	Temperature Regulation Test
Air Blank	0.000	23:43	

Data Download Date/Time: 01/11/2022 16:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: MARTIN-RAMIREZ

DATE/TIME OF TEST: 12/30/2021 23:56:06

Diagnostics Check	Fail	23:55	Temperature Regulation Test
Air Blank	0.000	23:56	

Data Download Date/Time: 01/11/2022 16:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: MARTIN-RAMIREZ

DATE/TIME OF TEST: 12/31/2021 00:16:35

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.081	00:10
Air Blank	0.000	00:10
Subject Sample # 1	0.103	00:12
Breath Volume	2.550	00:12
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.100	00:15
Breath Volume	3.617	00:15
Air Blank	0.000	00:15
Control Test	0.081	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 01/11/2022 16:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAINES,DAVID,K
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: MILIAN,ANET
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 11/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: ACEVEDO GIL

DATE/TIME OF TEST: 12/04/2021 01:23:16

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:14
Subject Sample # 1	0.198	01:15
Breath Volume	1.109	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.177	01:18
Breath Volume	1.546	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 3	0.195	01:21
Breath Volume	2.390	01:21
Air Blank	0.000	01:22
Control Test	0.078	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 12/27/2021 12:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CORREA,MATTHEW,C

BREATH TEST OPERATOR AGENCY: Davie PD

ARRESTING OFFICER: RAMIREZ,ANGEL

ARRESTING OFFICER AGENCY: Davie PD

LAST AGENCY INSPECTION: 11/29/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1283438

DRY GAS EXPIRY DATE: 02/25/2023

VIOLATION CODE: DUI

ARREST TIME: 23:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: DEETER

DATE/TIME OF TEST: 12/15/2021 00:42:57

Diagnostics Check	OK	00:36
Air Blank	0.000	00:37
Control Test	0.080	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.292	00:38
Breath Volume	3.007	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.301	00:41
Breath Volume	2.503	00:41
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 12/27/2021 12:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD, RONALD, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD, RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 11/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: CARNES

DATE/TIME OF TEST: 12/03/2021 23:54:29

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:52	
Control Test	0.079	23:52	
Air Blank	0.000	23:52	
Subject Sample	# 1	23:53	Subject Test Refused
Breath Volume	0.000	23:53	
Air Blank	0.000	23:53	
Control Test	0.079	23:53	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:54	

Data Download Date/Time: 12/06/2021 15:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEARS,C,D
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,J
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 12/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/08/2021 01:31:33

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.144	01:26
Breath Volume	1.132	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.148	01:30
Breath Volume	2.445	01:30
Air Blank	0.000	01:30
Control Test	0.080	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 01/19/2022 17:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD, RONALD, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD, RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 12/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: RUSHTON

DATE/TIME OF TEST: 12/12/2021 23:32:01

Diagnostics Check	OK	23:29	
Air Blank	0.000	23:29	
Control Test	0.079	23:29	
Air Blank	0.000	23:30	
Subject Sample	# 1	23:30	Subject Test Refused
Breath Volume	0.000	23:30	
Air Blank	0.000	23:31	
Control Test	0.079	23:31	
Air Blank	0.000	23:31	
Diagnostics Check	OK	23:32	

Data Download Date/Time: 01/19/2022 17:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 12/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: MOONEY

DATE/TIME OF TEST: 12/25/2021 03:35:16

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.079	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.151	03:30
Breath Volume	1.261	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.133	03:33
Breath Volume	2.554	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 01/19/2022 17:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEARS,C,D
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: SEARS,C
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 12/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 12/25/2021 04:04:23

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.246	03:59
Breath Volume	3.976	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.249	04:02
Breath Volume	3.183	04:02
Air Blank	0.000	04:03
Control Test	0.078	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 01/19/2022 17:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEARS,C,D
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: SEARS,C
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 12/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: RICHO JR

DATE/TIME OF TEST: 12/28/2021 00:53:19

Diagnostics Check	OK	00:42
Air Blank	0.000	00:43
Control Test	0.079	00:43
Air Blank	0.000	00:44
Subject Sample # 1	0.259	00:45
Breath Volume	1.183	00:45
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.280	00:48
Breath Volume	1.980	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 3	0.283	00:51
Breath Volume	1.980	00:51
Air Blank	0.000	00:52
Control Test	0.078	00:52
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 01/19/2022 17:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD, RONALD, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD, RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 12/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: HAAS

DATE/TIME OF TEST: 12/30/2021 03:21:41

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.204	03:17
Breath Volume	1.695	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.200	03:20
Breath Volume	1.777	03:20
Air Blank	0.000	03:20
Control Test	0.079	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 01/19/2022 17:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COAKE,SPENCER,M
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: COAKE,SPENCER
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 11/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: DESPATHY

DATE/TIME OF TEST: 12/03/2021 04:03:11

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:00	
Control Test	0.082	04:00	
Air Blank	0.000	04:01	
Subject Sample	# 1	04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:02	
Control Test	0.082	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 12/21/2021 13:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COAKE,SPENCER,M
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: PERRY,DARRIET
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 11/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: DEVORE

DATE/TIME OF TEST: 12/10/2021 19:37:53

Diagnostics Check	OK	19:31
Air Blank	0.000	19:31
Control Test	0.081	19:32
Air Blank	0.000	19:32
Subject Sample # 1	0.239	19:33
Breath Volume	1.207	19:33
Air Blank	0.000	19:33
Air Blank	0.000	19:35
Subject Sample # 2	0.234	19:36
Breath Volume	2.660	19:36
Air Blank	0.000	19:36
Control Test	0.081	19:37
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time: 12/21/2021 13:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMARCO,AUSTIN,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: IVEY,BRIAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A3
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 12/01/2021 01:56:41

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.079	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.147	01:52
Breath Volume	1.656	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.147	01:55
Breath Volume	2.011	01:55
Air Blank	0.000	01:55
Control Test	0.079	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 12/16/2021 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGEL,JASON,S
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: DOWD,TIMOTHY
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 12/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09820080A3
DRY GAS EXPIRY DATE: 06/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:01

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/29/2021 07:28:26

Diagnostics Check	OK	07:22
Air Blank	0.000	07:22
Control Test	0.080	07:22
Air Blank	0.000	07:23
Subject Sample # 1	0.000	07:24
Breath Volume	3.339	07:24
Air Blank	0.000	07:24
Air Blank	0.000	07:26
Subject Sample # 2	0.000	07:27
Breath Volume	1.167	07:27
Air Blank	0.000	07:27
Control Test	0.080	07:27
Air Blank	0.000	07:28
Diagnostics Check	OK	07:28

Data Download Date/Time: 01/13/2022 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: GOODING

DATE/TIME OF TEST: 12/01/2021 03:33:34

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.085	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.193	03:28
Breath Volume	2.421	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.186	03:31
Breath Volume	1.812	03:31
Air Blank	0.000	03:32
Control Test	0.082	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 12/14/2021 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES, ALFONSO,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SMITH, M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/01/2021 20:29:02

Diagnostics Check	OK	20:20	
Air Blank	0.000	20:21	
Control Test	0.082	20:21	
Air Blank	0.000	20:21	
Subject Sample	# 1 0.123	20:25	Volume Not Met
Breath Volume	0.527	20:25	
Air Blank	0.000	20:25	
Air Blank	0.000	20:27	
Subject Sample	# 2	20:27	Subject Test Refused
Breath Volume	0.000	20:27	
Air Blank	0.000	20:28	
Control Test	0.079	20:28	
Air Blank	0.000	20:28	
Diagnostics Check	OK	20:29	

Data Download Date/Time: 12/14/2021 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARBER,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BRAGG

DATE/TIME OF TEST: 12/03/2021 04:12:33

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.083	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.149	04:07
Breath Volume	4.398	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.146	04:10
Breath Volume	4.222	04:10
Air Blank	0.000	04:11
Control Test	0.083	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 12/14/2021 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALE,J,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: RAULERSON

DATE/TIME OF TEST: 12/03/2021 16:53:07

Diagnostics Check	OK	16:45
Air Blank	0.000	16:45
Control Test	0.082	16:46
Air Blank	0.000	16:46
Subject Sample # 1	0.000	16:47
Breath Volume	2.875	16:47
Air Blank	0.000	16:47
Air Blank	0.000	16:49
Subject Sample # 2	0.000	16:51
Breath Volume	4.335	16:51
Air Blank	0.000	16:52
Control Test	0.081	16:52
Air Blank	0.000	16:53
Diagnostics Check	OK	16:53

Data Download Date/Time: 12/14/2021 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MADYS

DATE/TIME OF TEST: 12/04/2021 01:17:25

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:14	
Control Test	0.083	01:15	
Air Blank	0.000	01:15	
Subject Sample	# 1	01:16	Subject Test Refused
Breath Volume	0.000	01:16	
Air Blank	0.000	01:16	
Control Test	0.082	01:16	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 12/14/2021 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville Beach
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: RICH

DATE/TIME OF TEST: 12/04/2021 03:14:29

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:11	
Control Test	0.083	03:11	
Air Blank	0.000	03:12	
Subject Sample	# 1	03:13	Subject Test Refused
Breath Volume	0.000	03:13	
Air Blank	0.000	03:13	
Control Test	0.083	03:13	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 12/14/2021 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Atlantic Beach PD
ARRESTING OFFICER: IKRAM,Y
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: BABBITT

DATE/TIME OF TEST: 12/04/2021 05:05:13

Diagnostics Check	OK	04:56
Air Blank	0.000	04:57
Control Test	0.084	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.216	04:59
Breath Volume	1.613	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.213	05:03
Breath Volume	1.214	05:03
Air Blank	0.000	05:04
Control Test	0.080	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 12/14/2021 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: LANDAAS

DATE/TIME OF TEST: 12/04/2021 21:55:23

Diagnostics Check	OK	21:48
Air Blank	0.000	21:49
Control Test	0.081	21:49
Air Blank	0.000	21:49
Subject Sample # 1	0.136	21:50
Breath Volume	2.519	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.143	21:53
Breath Volume	3.000	21:53
Air Blank	0.000	21:54
Control Test	0.082	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 12/14/2021 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SPURRIER,J
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: MCMANUS

DATE/TIME OF TEST: 12/05/2021 01:06:11

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:03	
Control Test	0.085	01:04	
Air Blank	0.000	01:04	
Subject Sample	# 1	01:04	Subject Test Refused
Breath Volume	0.000	01:04	
Air Blank	0.000	01:05	
Control Test	0.082	01:05	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 12/14/2021 12:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGGINS,R,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: COLON MARTINEZ

DATE/TIME OF TEST: 12/05/2021 02:39:22

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.082	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.116	02:34
Breath Volume	2.074	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.114	02:37
Breath Volume	1.347	02:37
Air Blank	0.000	02:38
Control Test	0.082	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 12/14/2021 12:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: STEWARD

DATE/TIME OF TEST: 12/05/2021 04:29:24

Diagnostics Check	OK	04:21
Air Blank	0.000	04:22
Control Test	0.082	04:22
Air Blank	0.000	04:22
Subject Sample # 1	0.000	04:24
Breath Volume	1.261	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.000	04:28
Breath Volume	1.187	04:28
Air Blank	0.000	04:28
Control Test	0.081	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 12/14/2021 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/05/2021 21:01:12

Diagnostics Check	OK	20:54
Air Blank	0.000	20:55
Control Test	0.082	20:55
Air Blank	0.000	20:55
Subject Sample # 1	0.114	20:56
Breath Volume	1.371	20:56
Air Blank	0.000	20:57
Air Blank	0.000	20:59
Subject Sample # 2	0.115	20:59
Breath Volume	1.429	20:59
Air Blank	0.000	21:00
Control Test	0.081	21:00
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

Data Download Date/Time: 12/16/2021 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGGINS,R,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: GOSSETT

DATE/TIME OF TEST: 12/06/2021 00:21:54

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:19	
Control Test	0.082	00:19	
Air Blank	0.000	00:20	
Subject Sample	# 1	00:20	Subject Test Refused
Breath Volume	0.000	00:20	
Air Blank	0.000	00:20	
Control Test	0.083	00:21	
Air Blank	0.000	00:21	
Diagnostics Check	OK	00:21	

Data Download Date/Time: 12/16/2021 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: WARNOCK

DATE/TIME OF TEST: 12/06/2021 03:39:39

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:37	
Control Test	0.082	03:37	
Air Blank	0.000	03:38	
Subject Sample	# 1	03:38	Subject Test Refused
Breath Volume	0.000	03:38	
Air Blank	0.000	03:38	
Control Test	0.083	03:39	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 12/16/2021 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: MARRERO

DATE/TIME OF TEST: 12/06/2021 21:50:14

Diagnostics Check	OK	21:44
Air Blank	0.000	21:44
Control Test	0.083	21:44
Air Blank	0.000	21:45
Subject Sample # 1	0.000	21:46
Breath Volume	1.914	21:46
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample # 2	0.000	21:48
Breath Volume	2.113	21:48
Air Blank	0.000	21:49
Control Test	0.082	21:49
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 12/16/2021 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: VILLAVICENCIO

DATE/TIME OF TEST: 12/06/2021 23:23:55

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:21	
Control Test	0.081	23:21	
Air Blank	0.000	23:22	
Subject Sample	# 1	23:22	Subject Test Refused
Breath Volume	0.000	23:22	
Air Blank	0.000	23:23	
Control Test	0.082	23:23	
Air Blank	0.000	23:23	
Diagnostics Check	OK	23:23	

Data Download Date/Time: 12/16/2021 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MATTIACE,R
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: BIAGAS

DATE/TIME OF TEST: 12/07/2021 01:49:18

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:46	
Control Test	0.081	01:46	
Air Blank	0.000	01:47	
Subject Sample	# 1	01:47	Subject Test Refused
Breath Volume	0.000	01:47	
Air Blank	0.000	01:47	
Control Test	0.082	01:48	
Air Blank	0.000	01:48	
Diagnostics Check	Fail	01:48	Analytical Stability Test
Air Blank	0.000	01:49	

Data Download Date/Time: 12/16/2021 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGGINS,R,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 12/07/2021 03:08:11

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:05	
Control Test	0.082	03:06	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:06	Subject Test Refused
Breath Volume	0.000	03:06	
Air Blank	0.000	03:07	
Control Test	0.082	03:07	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 12/16/2021 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: CHAMPAGNE

DATE/TIME OF TEST: 12/07/2021 22:39:13

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:36	
Control Test	0.083	22:37	
Air Blank	0.000	22:37	
Subject Sample	# 1	22:37	Subject Test Refused
Breath Volume	0.000	22:37	
Air Blank	0.000	22:38	
Control Test	0.082	22:38	
Air Blank	0.000	22:39	
Diagnostics Check	OK	22:39	

Data Download Date/Time: 12/16/2021 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: KRAMER

DATE/TIME OF TEST: 12/08/2021 00:05:42

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:03	
Control Test	0.082	00:03	
Air Blank	0.000	00:04	
Subject Sample	# 1	00:04	Subject Test Refused
Breath Volume	0.000	00:04	
Air Blank	0.000	00:04	
Control Test	0.081	00:05	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:05	

Data Download Date/Time: 12/16/2021 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGGINS,R,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: POUNCEY

DATE/TIME OF TEST: 12/08/2021 02:23:37

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.083	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.144	02:18
Breath Volume	2.308	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.140	02:22
Breath Volume	2.695	02:22
Air Blank	0.000	02:22
Control Test	0.082	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 12/16/2021 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALE,J,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ADMIN,ADMIN
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 08:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: ADMIN

DATE/TIME OF TEST: 12/08/2021 08:47:30

Diagnostics Check	OK	08:41
Air Blank	0.000	08:41
Control Test	0.083	08:41
Air Blank	0.000	08:42
Subject Sample # 1	0.000	08:43
Breath Volume	2.128	08:43
Air Blank	0.000	08:43
Air Blank	0.000	08:45
Subject Sample # 2	0.000	08:46
Breath Volume	2.218	08:46
Air Blank	0.000	08:46
Control Test	0.080	08:46
Air Blank	0.000	08:47
Diagnostics Check	OK	08:47

Data Download Date/Time: 12/16/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: MATHESON

DATE/TIME OF TEST: 12/08/2021 20:50:02

Diagnostics Check	OK	20:43
Air Blank	0.000	20:43
Control Test	0.082	20:44
Air Blank	0.000	20:44
Subject Sample # 1	0.180	20:45
Breath Volume	1.546	20:45
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.166	20:48
Breath Volume	1.253	20:48
Air Blank	0.000	20:49
Control Test	0.080	20:49
Air Blank	0.000	20:49
Diagnostics Check	OK	20:50

Data Download Date/Time: 12/16/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: ANCELET

DATE/TIME OF TEST: 12/09/2021 00:30:31

Diagnostics Check	OK	00:22
Air Blank	0.000	00:23
Control Test	0.081	00:23
Air Blank	0.000	00:24
Subject Sample # 1	0.205	00:25
Breath Volume	1.722	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:27
Subject Sample # 2	0.200	00:28
Breath Volume	1.195	00:28
Air Blank	0.000	00:29
Control Test	0.080	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 12/16/2021 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: O'QUINN

DATE/TIME OF TEST: 12/09/2021 04:17:44

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:02	
Control Test	0.081	04:02	
Air Blank	0.000	04:03	
Subject Sample	# 1 0.143	04:06	Volume Not Met
Breath Volume	0.359	04:06	
Air Blank	0.000	04:07	
Air Blank	0.000	04:09	
Subject Sample	# 2 0.164	04:10	
Breath Volume	1.726	04:10	
Air Blank	0.000	04:11	
Air Blank	0.000	04:12	
Subject Sample	# 3 0.163	04:16	Volume Not Met
Breath Volume	0.085	04:16	
Air Blank	0.000	04:16	
Control Test	0.082	04:17	
Air Blank	0.000	04:17	
Diagnostics Check	OK	04:17	

Data Download Date/Time: 12/16/2021 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: CARSWELL

DATE/TIME OF TEST: 12/09/2021 20:29:38

Diagnostics Check	OK	20:26	
Air Blank	0.000	20:26	
Control Test	0.082	20:27	
Air Blank	0.000	20:27	
Subject Sample	# 1	20:28	Subject Test Refused
Breath Volume	0.000	20:28	
Air Blank	0.000	20:28	
Control Test	0.082	20:29	
Air Blank	0.000	20:29	
Diagnostics Check	OK	20:29	

Data Download Date/Time: 12/16/2021 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville Beach
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: SUMMERS

DATE/TIME OF TEST: 12/10/2021 02:42:57

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.082	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.196	02:38
Breath Volume	1.406	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.208	02:41
Breath Volume	1.292	02:41
Air Blank	0.000	02:42
Control Test	0.082	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 12/16/2021 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville Beach
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: KHEDKAR

DATE/TIME OF TEST: 12/11/2021 03:49:49

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:47	
Control Test	0.082	03:47	
Air Blank	0.000	03:48	
Subject Sample	# 1	03:48	Subject Test Refused
Breath Volume	0.000	03:48	
Air Blank	0.000	03:48	
Control Test	0.082	03:49	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 12/16/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: PIPKINS

DATE/TIME OF TEST: 12/11/2021 18:17:29

Diagnostics Check	OK	18:10
Air Blank	0.000	18:11
Control Test	0.080	18:11
Air Blank	0.000	18:11
Subject Sample # 1	0.179	18:12
Breath Volume	1.730	18:12
Air Blank	0.000	18:13
Air Blank	0.000	18:15
Subject Sample # 2	0.178	18:15
Breath Volume	1.667	18:15
Air Blank	0.000	18:16
Control Test	0.080	18:16
Air Blank	0.000	18:17
Diagnostics Check	OK	18:17

Data Download Date/Time: 12/16/2021 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWLINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: AMINI HAJIBASHI

DATE/TIME OF TEST: 12/12/2021 02:04:14

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.082	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.098	01:59
Breath Volume	1.429	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:01
Subject Sample # 2	0.101	02:02
Breath Volume	1.695	02:02
Air Blank	0.000	02:03
Control Test	0.080	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 12/16/2021 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 12/12/2021 05:50:38

Diagnostics Check	OK	05:47	
Air Blank	0.000	05:48	
Control Test	0.083	05:48	
Air Blank	0.000	05:48	
Subject Sample	# 1	05:49	Subject Test Refused
Breath Volume	0.000	05:49	
Air Blank	0.000	05:49	
Control Test	0.082	05:50	
Air Blank	0.000	05:50	
Diagnostics Check	OK	05:50	

Data Download Date/Time: 12/16/2021 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSES,L,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CURRY,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:27

SUBJECT'S LAST NAME: GODINEZ DIAZ

DATE/TIME OF TEST: 12/12/2021 08:50:40

Diagnostics Check	OK	08:47	
Air Blank	0.000	08:48	
Control Test	0.083	08:48	
Air Blank	0.000	08:48	
Subject Sample	# 1	08:49	Subject Test Refused
Breath Volume	0.000	08:49	
Air Blank	0.000	08:49	
Control Test	0.083	08:50	
Air Blank	0.000	08:50	
Diagnostics Check	OK	08:50	

Data Download Date/Time: 12/16/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: ZAHRN

DATE/TIME OF TEST: 12/12/2021 21:49:25

Diagnostics Check	OK	21:42
Air Blank	0.000	21:43
Control Test	0.083	21:43
Air Blank	0.000	21:44
Subject Sample # 1	0.206	21:44
Breath Volume	2.687	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.198	21:47
Breath Volume	2.582	21:47
Air Blank	0.000	21:48
Control Test	0.082	21:48
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 12/16/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MATTIACE,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: UZOECHI

DATE/TIME OF TEST: 12/12/2021 22:58:48

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.082	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.155	22:54
Breath Volume	1.960	22:54
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.150	22:57
Breath Volume	1.785	22:57
Air Blank	0.000	22:57
Control Test	0.083	22:58
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 12/16/2021 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARRISH,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: SPOONER

DATE/TIME OF TEST: 12/13/2021 01:18:15

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:15	
Control Test	0.082	01:16	
Air Blank	0.000	01:16	
Subject Sample	# 1	01:16	Subject Test Refused
Breath Volume	0.000	01:16	
Air Blank	0.000	01:17	
Control Test	0.082	01:17	
Air Blank	0.000	01:18	
Diagnostics Check	OK	01:18	

Data Download Date/Time: 12/16/2021 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 12/13/2021 03:39:47

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:37	
Control Test	0.083	03:37	
Air Blank	0.000	03:38	
Subject Sample	# 1	03:38	Subject Test Refused
Breath Volume	0.000	03:38	
Air Blank	0.000	03:38	
Control Test	0.082	03:39	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 12/16/2021 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHIPP,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: BRENNAN

DATE/TIME OF TEST: 12/14/2021 01:17:38

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.082	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.155	01:12
Breath Volume	2.867	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.146	01:16
Breath Volume	1.652	01:16
Air Blank	0.000	01:16
Control Test	0.079	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 12/16/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TAYLOR,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: NOBLES

DATE/TIME OF TEST: 12/16/2021 00:50:59

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.082	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.151	00:46
Breath Volume	1.742	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:48
Subject Sample # 2	0.153	00:49
Breath Volume	2.453	00:49
Air Blank	0.000	00:50
Control Test	0.082	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 12/16/2021 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: ALONZO MORALES

DATE/TIME OF TEST: 12/19/2021 01:18:04

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.083	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.041	01:13
Breath Volume	1.398	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.039	01:16
Breath Volume	2.347	01:16
Air Blank	0.000	01:17
Control Test	0.082	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:18

Data Download Date/Time: 12/29/2021 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: FALLS

DATE/TIME OF TEST: 12/19/2021 04:27:09

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:24	
Control Test	0.083	04:24	
Air Blank	0.000	04:25	
Subject Sample	# 1	04:25	Subject Test Refused
Breath Volume	0.000	04:25	
Air Blank	0.000	04:26	
Control Test	0.083	04:26	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

Data Download Date/Time: 12/29/2021 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: MEA

DATE/TIME OF TEST: 12/19/2021 06:20:30

Diagnostics Check	OK	06:14
Air Blank	0.000	06:14
Control Test	0.084	06:15
Air Blank	0.000	06:15
Subject Sample # 1	0.182	06:16
Breath Volume	1.593	06:16
Air Blank	0.000	06:16
Air Blank	0.000	06:18
Subject Sample # 2	0.174	06:18
Breath Volume	1.414	06:18
Air Blank	0.000	06:19
Control Test	0.083	06:19
Air Blank	0.000	06:20
Diagnostics Check	OK	06:20

Data Download Date/Time: 12/29/2021 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: DESSER

DATE/TIME OF TEST: 12/20/2021 00:15:34

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:13	
Control Test	0.084	00:13	
Air Blank	0.000	00:14	
Subject Sample	# 1	00:14	Subject Test Refused
Breath Volume	0.000	00:14	
Air Blank	0.000	00:14	
Control Test	0.082	00:14	
Air Blank	0.000	00:15	
Diagnostics Check	OK	00:15	

Data Download Date/Time: 12/29/2021 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: SHERMAN

DATE/TIME OF TEST: 12/20/2021 04:20:47

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.084	04:12
Air Blank	0.000	04:13
Subject Sample # 1	0.024	04:15
Breath Volume	1.250	04:15
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.022	04:19
Breath Volume	1.871	04:19
Air Blank	0.000	04:19
Control Test	0.080	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 12/29/2021 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSES,L,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:09

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/20/2021 16:36:26

Diagnostics Check	OK	16:30
Air Blank	0.000	16:30
Control Test	0.083	16:30
Air Blank	0.000	16:31
Subject Sample # 1	0.174	16:31
Breath Volume	1.839	16:31
Air Blank	0.000	16:32
Air Blank	0.000	16:34
Subject Sample # 2	0.175	16:34
Breath Volume	1.902	16:34
Air Blank	0.000	16:35
Control Test	0.082	16:35
Air Blank	0.000	16:36
Diagnostics Check	OK	16:36

Data Download Date/Time: 12/29/2021 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 12/21/2021 01:17:05

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:14	
Control Test	0.083	01:15	
Air Blank	0.000	01:15	
Subject Sample	# 1	01:15	Subject Test Refused
Breath Volume	0.000	01:15	
Air Blank	0.000	01:16	
Control Test	0.082	01:16	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 12/29/2021 14:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 12/21/2021 20:59:34

Diagnostics Check	OK	20:53
Air Blank	0.000	20:53
Control Test	0.083	20:53
Air Blank	0.000	20:54
Subject Sample # 1	0.209	20:54
Breath Volume	2.769	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.201	20:57
Breath Volume	2.785	20:57
Air Blank	0.000	20:58
Control Test	0.081	20:58
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

Data Download Date/Time: 12/29/2021 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: HAMPTON

DATE/TIME OF TEST: 12/22/2021 23:53:47

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.083	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.145	23:48
Breath Volume	1.550	23:48
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.139	23:52
Breath Volume	1.859	23:52
Air Blank	0.000	23:52
Control Test	0.082	23:53
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 12/29/2021 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 12/23/2021 03:33:20

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.084	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.148	03:28
Breath Volume	2.769	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.143	03:31
Breath Volume	2.128	03:31
Air Blank	0.000	03:32
Control Test	0.084	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 12/29/2021 14:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:17

SUBJECT'S LAST NAME: VEREEN

DATE/TIME OF TEST: 12/23/2021 06:44:40

Diagnostics Check	OK	06:37
Air Blank	0.000	06:38
Control Test	0.084	06:38
Air Blank	0.000	06:39
Subject Sample # 1	0.167	06:39
Breath Volume	1.472	06:39
Air Blank	0.000	06:40
Air Blank	0.000	06:42
Subject Sample # 2	0.161	06:43
Breath Volume	1.828	06:43
Air Blank	0.000	06:43
Control Test	0.082	06:44
Air Blank	0.000	06:44
Diagnostics Check	OK	06:44

Data Download Date/Time: 12/29/2021 14:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 12/23/2021 23:11:07

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.083	23:05
Air Blank	0.000	23:05
Subject Sample # 1	0.200	23:06
Breath Volume	2.152	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:08
Subject Sample # 2	0.210	23:09
Breath Volume	1.863	23:09
Air Blank	0.000	23:10
Control Test	0.081	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 12/29/2021 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 12/24/2021 04:20:57

Diagnostics Check	OK	04:04	
Air Blank	0.000	04:04	
Control Test	0.083	04:05	
Air Blank	0.000	04:05	
Subject Sample	# 1 0.147	04:08	Volume Not Met
Breath Volume	0.492	04:08	
Air Blank	0.000	04:09	
Air Blank	0.000	04:11	
Subject Sample	# 2 0.160	04:13	
Breath Volume	1.332	04:13	
Air Blank	0.000	04:14	
Air Blank	0.000	04:15	
Subject Sample	# 3 0.145	04:19	
Breath Volume	1.785	04:19	
Air Blank	0.000	04:20	
Control Test	0.081	04:20	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:20	

Data Download Date/Time: 12/29/2021 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: MOHR

DATE/TIME OF TEST: 12/25/2021 00:11:17

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.082	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.067	00:06
Breath Volume	1.101	00:06
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.068	00:09
Breath Volume	1.910	00:09
Air Blank	0.000	00:10
Control Test	0.083	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 12/29/2021 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: HARMON

DATE/TIME OF TEST: 12/25/2021 04:05:19

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.083	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.037	04:00
Breath Volume	3.812	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.036	04:03
Breath Volume	4.312	04:03
Air Blank	0.000	04:04
Control Test	0.083	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 12/29/2021 14:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 12/26/2021 01:57:50

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.082	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.203	01:53
Breath Volume	2.972	01:53
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.187	01:56
Breath Volume	2.128	01:56
Air Blank	0.000	01:56
Control Test	0.082	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 12/29/2021 14:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: BACON

DATE/TIME OF TEST: 12/26/2021 23:11:32

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:09	
Control Test	0.081	23:09	
Air Blank	0.000	23:09	
Subject Sample	# 1	23:10	Subject Test Refused
Breath Volume	0.000	23:10	
Air Blank	0.000	23:10	
Control Test	0.082	23:10	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:11	

Data Download Date/Time: 12/29/2021 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/27/2021 00:05:50

Diagnostics Check	OK	23:59
Air Blank	0.000	23:59
Control Test	0.081	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.113	00:01
Breath Volume	3.562	00:01
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.111	00:04
Breath Volume	3.171	00:04
Air Blank	0.000	00:04
Control Test	0.081	00:05
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 12/29/2021 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGGINS,R,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/27/2021 02:42:08

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.081	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.083	02:37
Breath Volume	1.457	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:39
Subject Sample # 2	0.085	02:40
Breath Volume	1.125	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 12/29/2021 14:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: COOLMAN

DATE/TIME OF TEST: 12/27/2021 03:48:39

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:46	
Control Test	0.082	03:46	
Air Blank	0.000	03:47	
Subject Sample	# 1	03:47	Subject Test Refused
Breath Volume	0.000	03:47	
Air Blank	0.000	03:47	
Control Test	0.082	03:48	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 12/29/2021 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL, EDWARD, A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS, B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: MCELWEE

DATE/TIME OF TEST: 12/27/2021 21:20:57

Diagnostics Check	OK	21:14
Air Blank	0.000	21:15
Control Test	0.081	21:15
Air Blank	0.000	21:16
Subject Sample # 1	0.050	21:16
Breath Volume	1.160	21:16
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample # 2	0.047	21:19
Breath Volume	1.796	21:19
Air Blank	0.000	21:20
Control Test	0.081	21:20
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

Data Download Date/Time: 12/29/2021 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAINRIGHT, DAVID, T
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: CHANDLER, ROBERT
ARRESTING OFFICER AGENCY: Zephyrhills PD
LAST AGENCY INSPECTION: 12/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: FAHY

DATE/TIME OF TEST: 12/19/2021 04:58:05

Diagnostics Check	OK	04:51
Air Blank	0.000	04:52
Control Test	0.080	04:52
Air Blank	0.000	04:53
Subject Sample # 1	0.213	04:53
Breath Volume	2.101	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample # 2	0.202	04:56
Breath Volume	1.101	04:56
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 02/07/2022 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 12/02/2021 03:38:22

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.248	03:33
Breath Volume	1.558	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.255	03:36
Breath Volume	1.925	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 12/13/2021 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 12/02/2021 22:37:32

Diagnostics Check	OK	22:26	
Air Blank	0.000	22:26	
Control Test	0.081	22:26	
Air Blank	0.000	22:27	
Subject Sample	# 1	22:30	No Sample Provided
Breath Volume	0.000	22:30	
Air Blank	0.000	22:30	
Air Blank	0.000	22:32	
Subject Sample	# 2	0.044	22:36
Breath Volume	0.480	22:36	Volume Not Met
Air Blank	0.000	22:36	
Control Test	0.081	22:36	
Air Blank	0.000	22:37	
Diagnostics Check	OK	22:37	

Data Download Date/Time: 12/13/2021 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/25/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 07220080A2

DRY GAS EXPIRY DATE: 05/05/2022

VIOLATION CODE: DUI

ARREST TIME: 21:38

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 12/02/2021 22:46:05

Diagnostics Check	OK	22:39
Air Blank	0.000	22:39
Control Test	0.081	22:40
Air Blank	0.000	22:40
Subject Sample # 1	0.048	22:41
Breath Volume	1.917	22:41
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.045	22:44
Breath Volume	2.421	22:44
Air Blank	0.000	22:45
Control Test	0.081	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 12/13/2021 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/25/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 07220080A2

DRY GAS EXPIRY DATE: 05/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: CIAMPOLILLO

DATE/TIME OF TEST: 12/03/2021 03:11:20

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:08	
Control Test	0.080	03:09	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:10	
Control Test	0.080	03:10	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

Data Download Date/Time: 12/13/2021 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER, MATTHEW, D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER, MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 12/04/2021 02:08:09

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:53	
Control Test	0.080	01:53	
Air Blank	0.000	01:53	
Subject Sample	# 1 0.122	01:57	Volume Not Met
Breath Volume	0.257	01:57	
Air Blank	0.000	01:57	
Air Blank	0.000	01:59	
Subject Sample	# 2 0.114	02:02	
Breath Volume	1.281	02:02	
Air Blank	0.000	02:02	
Air Blank	0.000	02:04	
Subject Sample	# 3 0.122	02:06	
Breath Volume	1.667	02:06	
Air Blank	0.000	02:07	
Control Test	0.079	02:07	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

Data Download Date/Time: 12/13/2021 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/25/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 07220080A2

DRY GAS EXPIRY DATE: 05/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:36

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: MUNOZ ARIZA

DATE/TIME OF TEST: 12/05/2021 03:25:23

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.081	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.089	03:20
Breath Volume	2.121	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.088	03:23
Breath Volume	1.550	03:23
Air Blank	0.000	03:24
Control Test	0.081	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 12/13/2021 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CALDERONE, STEVEN, A

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: CALDERONE, STEVEN

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/25/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 07220080A2

DRY GAS EXPIRY DATE: 05/05/2022

VIOLATION CODE: DUI

ARREST TIME: 10:15

GENDER: Female

OBSERVATION PERIOD BEGAN: 10:59

SUBJECT'S LAST NAME: VITA

DATE/TIME OF TEST: 12/07/2021 11:27:17

Diagnostics Check	OK	11:21
Air Blank	0.000	11:21
Control Test	0.081	11:22
Air Blank	0.000	11:22
Subject Sample # 1	0.000	11:23
Breath Volume	1.160	11:23
Air Blank	0.000	11:23
Air Blank	0.000	11:25
Subject Sample # 2	0.000	11:25
Breath Volume	1.222	11:25
Air Blank	0.000	11:26
Control Test	0.081	11:26
Air Blank	0.000	11:27
Diagnostics Check	OK	11:27

Data Download Date/Time: 12/13/2021 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 12/11/2021 03:00:54

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.161	02:56
Breath Volume	2.148	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.159	02:59
Breath Volume	3.785	02:59
Air Blank	0.000	02:59
Control Test	0.080	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 12/13/2021 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/25/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 07220080A2

DRY GAS EXPIRY DATE: 05/05/2022

VIOLATION CODE: DUI

ARREST TIME: 03:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: KADNER

DATE/TIME OF TEST: 12/11/2021 04:00:21

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.081	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.000	03:56
Breath Volume	1.957	03:56
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.000	03:58
Breath Volume	2.242	03:58
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 12/13/2021 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER, MATTHEW, D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER, MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/12/2021 04:44:27

Diagnostics Check	OK	04:41	
Air Blank	0.000	04:41	
Control Test	0.081	04:42	
Air Blank	0.000	04:42	
Subject Sample	# 1	04:43	Subject Test Refused
Breath Volume	0.000	04:43	
Air Blank	0.000	04:43	
Control Test	0.080	04:43	
Air Blank	0.000	04:44	
Diagnostics Check	OK	04:44	

Data Download Date/Time: 12/13/2021 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,JOHN,B
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: LIGHTFOOT,LANE
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: CHEK

DATE/TIME OF TEST: 12/15/2021 01:21:15

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.081	01:15
Air Blank	0.000	01:15
Subject Sample # 1	0.211	01:16
Breath Volume	1.515	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.212	01:19
Breath Volume	1.503	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 01/05/2022 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: SCHWARTZ SAFKO

DATE/TIME OF TEST: 12/16/2021 00:01:10

Diagnostics Check	OK	23:49	
Air Blank	0.000	23:50	
Control Test	0.081	23:50	
Air Blank	0.000	23:51	
Subject Sample	# 1	23:54	No Sample Provided
Breath Volume	0.000	23:54	
Air Blank	0.000	23:54	
Air Blank	0.000	23:56	
Subject Sample	# 2	23:59	No Sample Provided
Breath Volume	0.000	23:59	
Air Blank	0.000	00:00	
Control Test	0.081	00:00	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 01/05/2022 13:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 07220080A2

DRY GAS EXPIRY DATE: 05/05/2022

VIOLATION CODE: DUI

ARREST TIME: 03:30

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: WALTER

DATE/TIME OF TEST: 12/16/2021 04:17:19

Diagnostics Check	OK	04:14	
Air Blank	0.000	04:14	
Control Test	0.080	04:15	
Air Blank	0.000	04:15	
Subject Sample	# 1	04:15	Subject Test Refused
Breath Volume	0.000	04:15	
Air Blank	0.000	04:16	
Control Test	0.080	04:16	
Air Blank	0.000	04:17	
Diagnostics Check	OK	04:17	

Data Download Date/Time: 01/05/2022 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: STANICH

DATE/TIME OF TEST: 12/16/2021 22:31:44

Diagnostics Check	OK	22:25
Air Blank	0.000	22:25
Control Test	0.081	22:26
Air Blank	0.000	22:26
Subject Sample # 1	0.126	22:27
Breath Volume	2.554	22:27
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.125	22:30
Breath Volume	2.023	22:30
Air Blank	0.000	22:30
Control Test	0.080	22:31
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 01/05/2022 13:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 07220080A2

DRY GAS EXPIRY DATE: 05/05/2022

VIOLATION CODE: DUI

ARREST TIME: 01:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: PHOEUK

DATE/TIME OF TEST: 12/18/2021 01:49:23

Diagnostics Check	OK	01:42
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:44
Subject Sample # 1	0.164	01:44
Breath Volume	2.246	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.167	01:47
Breath Volume	2.070	01:47
Air Blank	0.000	01:48
Control Test	0.081	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 01/05/2022 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: WILLETS

DATE/TIME OF TEST: 12/18/2021 03:00:05

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:57	
Control Test	0.080	02:57	
Air Blank	0.000	02:57	
Subject Sample	# 1	02:58	Subject Test Refused
Breath Volume	0.000	02:58	
Air Blank	0.000	02:58	
Control Test	0.081	02:58	
Air Blank	0.000	02:59	
Diagnostics Check	Fail	02:59	Temperature Regulation Test
Air Blank	0.000	03:00	

Data Download Date/Time: 01/05/2022 13:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON, WILLIAM, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NIXON, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: DASSEL

DATE/TIME OF TEST: 12/03/2021 16:51:00

Diagnostics Check	OK	16:44
Air Blank	0.000	16:45
Control Test	0.081	16:45
Air Blank	0.000	16:45
Subject Sample # 1	0.255	16:46
Breath Volume	2.171	16:46
Air Blank	0.000	16:47
Air Blank	0.000	16:48
Subject Sample # 2	0.239	16:49
Breath Volume	2.082	16:49
Air Blank	0.000	16:50
Control Test	0.080	16:50
Air Blank	0.000	16:50
Diagnostics Check	OK	16:51

Data Download Date/Time: 12/20/2021 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: TOOHEY

DATE/TIME OF TEST: 12/04/2021 01:39:01

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:36	
Control Test	0.081	01:36	
Air Blank	0.000	01:37	
Subject Sample	# 1	01:37	RFI Detect
Breath Volume	0.000	01:37	
Air Blank	0.000	01:38	
Control Test	0.081	01:38	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:39	

Data Download Date/Time: 12/20/2021 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: TOOHEY

DATE/TIME OF TEST: 12/04/2021 01:47:49

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.266	01:42
Breath Volume	1.308	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.274	01:46
Breath Volume	1.472	01:46
Air Blank	0.000	01:46
Control Test	0.079	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 12/20/2021 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: PLEUS

DATE/TIME OF TEST: 12/05/2021 00:12:10

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:08	
Control Test	0.079	00:08	
Air Blank	0.000	00:09	
Subject Sample	# 1	00:10	Subject Test Refused
Breath Volume	0.000	00:10	
Air Blank	0.000	00:11	
Control Test	0.081	00:11	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:12	

Data Download Date/Time: 12/20/2021 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: DUPOUX

DATE/TIME OF TEST: 12/06/2021 03:05:42

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.081	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.034	03:01
Breath Volume	1.148	03:01
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.032	03:04
Breath Volume	1.117	03:04
Air Blank	0.000	03:04
Control Test	0.081	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 12/20/2021 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 12/06/2021 19:14:14

Diagnostics Check	OK	19:08
Air Blank	0.000	19:08
Control Test	0.080	19:08
Air Blank	0.000	19:09
Subject Sample # 1	0.228	19:09
Breath Volume	1.718	19:09
Air Blank	0.000	19:10
Air Blank	0.000	19:12
Subject Sample # 2	0.225	19:12
Breath Volume	1.750	19:12
Air Blank	0.000	19:13
Control Test	0.081	19:13
Air Blank	0.000	19:14
Diagnostics Check	OK	19:14

Data Download Date/Time: 12/20/2021 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HEMMINGS,DWIGHT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 12/07/2021 04:49:21

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.081	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.295	04:44
Breath Volume	2.562	04:44
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.295	04:47
Breath Volume	2.929	04:47
Air Blank	0.000	04:48
Control Test	0.079	04:48
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 12/20/2021 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: VILLSR SUAREZ

DATE/TIME OF TEST: 12/07/2021 23:49:37

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.224	23:45
Breath Volume	2.335	23:45
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.223	23:48
Breath Volume	1.906	23:48
Air Blank	0.000	23:48
Control Test	0.080	23:49
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 12/20/2021 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUNNER,JOSEPH,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RUNNER,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 11:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: PETTY

DATE/TIME OF TEST: 12/08/2021 12:32:31

Diagnostics Check	OK	12:28	
Air Blank	0.000	12:29	
Control Test	0.081	12:29	
Air Blank	0.000	12:30	
Subject Sample	# 1	12:31	Subject Test Refused
Breath Volume	0.000	12:31	
Air Blank	0.000	12:31	
Control Test	0.081	12:31	
Air Blank	0.000	12:32	
Diagnostics Check	OK	12:32	

Data Download Date/Time: 12/20/2021 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: BRIDGER

DATE/TIME OF TEST: 12/16/2021 05:38:46

Diagnostics Check	OK	05:31
Air Blank	0.000	05:32
Control Test	0.081	05:32
Air Blank	0.000	05:33
Subject Sample # 1	0.122	05:33
Breath Volume	1.949	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.117	05:37
Breath Volume	1.808	05:37
Air Blank	0.000	05:37
Control Test	0.080	05:38
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 12/20/2021 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 12/16/2021 22:03:26

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.081	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.000	21:58
Breath Volume	2.312	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.000	22:02
Breath Volume	2.152	22:02
Air Blank	0.000	22:02
Control Test	0.081	22:02
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 12/20/2021 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:11

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/18/2021 06:58:52

Diagnostics Check	OK	06:52
Air Blank	0.000	06:52
Control Test	0.080	06:52
Air Blank	0.000	06:53
Subject Sample # 1	0.152	06:54
Breath Volume	2.191	06:54
Air Blank	0.000	06:55
Air Blank	0.000	06:56
Subject Sample # 2	0.156	06:57
Breath Volume	1.480	06:57
Air Blank	0.000	06:57
Control Test	0.079	06:58
Air Blank	0.000	06:58
Diagnostics Check	OK	06:58

Data Download Date/Time: 12/20/2021 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COMSTOCK,M,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GRAY,A
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 12/19/2021 17:36:47

Diagnostics Check	OK	17:30
Air Blank	0.000	17:30
Control Test	0.080	17:31
Air Blank	0.000	17:31
Subject Sample # 1	0.210	17:32
Breath Volume	2.046	17:32
Air Blank	0.000	17:32
Air Blank	0.000	17:34
Subject Sample # 2	0.225	17:35
Breath Volume	2.183	17:35
Air Blank	0.000	17:35
Control Test	0.081	17:36
Air Blank	0.000	17:36
Diagnostics Check	OK	17:36

Data Download Date/Time: 12/20/2021 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 12/19/2021 22:05:40

Diagnostics Check	OK	21:58
Air Blank	0.000	21:59
Control Test	0.081	21:59
Air Blank	0.000	22:00
Subject Sample # 1	0.340	22:00
Breath Volume	1.273	22:00
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.325	22:04
Breath Volume	1.457	22:04
Air Blank	0.000	22:04
Control Test	0.081	22:05
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 12/20/2021 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: MARENGHI

DATE/TIME OF TEST: 12/23/2021 23:47:36

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.081	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.148	23:43
Breath Volume	1.871	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.152	23:46
Breath Volume	3.179	23:46
Air Blank	0.000	23:46
Control Test	0.080	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 01/07/2022 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONNELL,JACOB,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CONNELL,JACOB
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:02

SUBJECT'S LAST NAME: TORRES MUNIZ

DATE/TIME OF TEST: 12/24/2021 12:30:23

Diagnostics Check	OK	12:24
Air Blank	0.000	12:24
Control Test	0.080	12:25
Air Blank	0.000	12:25
Subject Sample # 1	0.265	12:25
Breath Volume	2.343	12:25
Air Blank	0.000	12:26
Air Blank	0.000	12:28
Subject Sample # 2	0.262	12:28
Breath Volume	2.347	12:28
Air Blank	0.000	12:29
Control Test	0.080	12:29
Air Blank	0.000	12:30
Diagnostics Check	OK	12:30

Data Download Date/Time: 01/07/2022 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 12/25/2021 00:19:13

Diagnostics Check	OK	00:07
Air Blank	0.000	00:07
Control Test	0.081	00:08
Air Blank	0.000	00:08
Subject Sample # 1	0.175	00:10
Breath Volume	1.160	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:12
Subject Sample # 2	0.154	00:14
Breath Volume	1.191	00:14
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 3	0.168	00:17
Breath Volume	1.425	00:17
Air Blank	0.000	00:18
Control Test	0.080	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 01/07/2022 15:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: E215284931331

DATE/TIME OF TEST: 12/25/2021 02:17:15

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.081	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.092	02:12
Breath Volume	1.535	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.093	02:15
Breath Volume	1.699	02:15
Air Blank	0.000	02:16
Control Test	0.082	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 01/07/2022 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: CRISSON

DATE/TIME OF TEST: 12/26/2021 04:48:23

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:40	
Control Test	0.081	04:40	
Air Blank	0.000	04:41	
Subject Sample	# 1	04:44	No Sample Provided
Breath Volume	0.000	04:44	
Air Blank	0.000	04:44	
Air Blank	0.000	04:46	
Subject Sample	# 2	04:47	Subject Test Refused
Breath Volume	0.000	04:47	
Air Blank	0.000	04:47	
Control Test	0.081	04:47	
Air Blank	0.000	04:48	
Diagnostics Check	OK	04:48	

Data Download Date/Time: 01/07/2022 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUNNER,JOSEPH,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RUNNER,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/28/2021 17:47:05

Diagnostics Check	OK	17:40
Air Blank	0.000	17:41
Control Test	0.081	17:41
Air Blank	0.000	17:41
Subject Sample # 1	0.000	17:42
Breath Volume	2.562	17:42
Air Blank	0.000	17:43
Air Blank	0.000	17:45
Subject Sample # 2	0.000	17:45
Breath Volume	1.609	17:45
Air Blank	0.000	17:46
Control Test	0.082	17:46
Air Blank	0.000	17:46
Diagnostics Check	OK	17:47

Data Download Date/Time: 01/07/2022 15:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MINTON

DATE/TIME OF TEST: 12/29/2021 03:33:24

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:22	
Control Test	0.082	03:22	
Air Blank	0.000	03:23	
Subject Sample	# 1 0.186	03:26	Volume Not Met
Breath Volume	0.753	03:26	
Air Blank	0.000	03:27	
Air Blank	0.000	03:28	
Subject Sample	# 2 0.186	03:31	Volume Not Met
Breath Volume	0.667	03:31	
Air Blank	0.000	03:32	
Control Test	0.077	03:32	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 01/07/2022 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MINTON

DATE/TIME OF TEST: 12/29/2021 03:46:58

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.079	03:36	
Air Blank	0.000	03:36	
Subject Sample	# 1 0.205	03:39	Volume Not Met
Breath Volume	0.191	03:39	
Air Blank	0.000	03:40	
Air Blank	0.000	03:42	
Subject Sample	# 2 0.195	03:45	Volume Not Met
Breath Volume	0.289	03:45	
Air Blank	0.000	03:46	
Control Test	0.080	03:46	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 01/07/2022 15:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: MCMILES

DATE/TIME OF TEST: 12/30/2021 00:47:32

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.081	00:41
Air Blank	0.000	00:42
Subject Sample # 1	0.146	00:42
Breath Volume	2.136	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.146	00:46
Breath Volume	2.300	00:46
Air Blank	0.000	00:46
Control Test	0.080	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 01/07/2022 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIZZO,HAYLEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RIZZO,HAYLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: SIZEMORE

DATE/TIME OF TEST: 12/30/2021 02:06:29

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.081	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.000	01:59
Breath Volume	1.550	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.000	02:05
Breath Volume	1.535	02:05
Air Blank	0.000	02:05
Control Test	0.083	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 01/07/2022 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIZZO,HAYLEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RIZZO,HAYLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: EDMON

DATE/TIME OF TEST: 12/30/2021 22:44:37

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.080	22:36
Air Blank	0.000	22:37
Subject Sample # 1	0.084	22:40
Breath Volume	1.398	22:40
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.085	22:43
Breath Volume	1.644	22:43
Air Blank	0.000	22:43
Control Test	0.080	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 01/07/2022 15:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: GRAUER

DATE/TIME OF TEST: 12/31/2021 04:22:57

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:08	
Control Test	0.081	04:08	
Air Blank	0.000	04:09	
Subject Sample	# 1	04:12	No Sample Provided
Breath Volume	0.000	04:12	
Air Blank	0.000	04:12	
Air Blank	0.000	04:14	
Subject Sample	# 2	0.136	04:17
Breath Volume	1.132	04:17	
Air Blank	0.000	04:18	
Air Blank	0.000	04:20	
Subject Sample	# 3	0.136	04:21
Breath Volume	2.171	04:21	
Air Blank	0.000	04:22	
Control Test	0.081	04:22	
Air Blank	0.000	04:22	
Diagnostics Check	OK	04:22	

Data Download Date/Time: 01/07/2022 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGINS, TYLER, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ACQUAFREDDA, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/29/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 14:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: RUSH

DATE/TIME OF TEST: 12/05/2021 16:36:18

Diagnostics Check	OK	16:30
Air Blank	0.000	16:30
Control Test	0.082	16:31
Air Blank	0.000	16:31
Subject Sample # 1	0.000	16:32
Breath Volume	1.371	16:32
Air Blank	0.000	16:32
Air Blank	0.000	16:34
Subject Sample # 2	0.000	16:34
Breath Volume	1.156	16:34
Air Blank	0.000	16:35
Control Test	0.081	16:35
Air Blank	0.000	16:36
Diagnostics Check	OK	16:36

Data Download Date/Time: 12/09/2021 05:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUDAY,JOSHUA,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BUDAY,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: ESPINOZA UGARTE

DATE/TIME OF TEST: 12/08/2021 16:49:23

Diagnostics Check	OK	16:44	
Air Blank	0.000	16:44	
Control Test	0.081	16:45	
Air Blank	0.000	16:45	
Subject Sample	# 1 0.155	16:47	
Breath Volume	1.539	16:47	
Air Blank		16:48	Purge Fail
Air Blank	0.000	16:49	

Data Download Date/Time: 12/23/2021 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUDAY,JOSHUA,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BUDAY,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: ESPINOZA UGARTE

DATE/TIME OF TEST: 12/08/2021 17:05:18

Diagnostics Check	OK	16:50	
Air Blank	0.000	16:51	
Control Test	0.080	16:51	
Air Blank	0.000	16:52	
Subject Sample	# 1 0.163	16:55	Volume Not Met
Breath Volume	0.195	16:55	
Air Blank	0.000	16:55	
Air Blank	0.000	16:57	
Subject Sample	# 2 0.174	16:59	
Breath Volume	1.261	16:59	
Air Blank	0.000	17:00	
Air Blank	0.000	17:01	
Subject Sample	# 3 0.169	17:03	
Breath Volume	1.515	17:03	
Air Blank	0.000	17:04	
Control Test	0.079	17:04	
Air Blank	0.000	17:05	
Diagnostics Check	OK	17:05	

Data Download Date/Time: 12/23/2021 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARTHY, EVAN, V
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MCCARTHY, EVAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: LOIACONO

DATE/TIME OF TEST: 12/18/2021 01:26:30

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.082	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.160	01:21
Breath Volume	3.078	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.163	01:25
Breath Volume	3.125	01:25
Air Blank	0.000	01:25
Control Test	0.081	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 12/23/2021 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARTHY, EVAN, V
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MCKINNON, SIR FRANCIS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 12/18/2021 04:18:54

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.081	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.177	04:12
Breath Volume	1.394	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.176	04:17
Breath Volume	1.753	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 12/23/2021 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,LUIS,E
BREATH TEST OPERATOR AGENCY: FHP TROOP K
ARRESTING OFFICER: LOPEZ,LUIS
ARRESTING OFFICER AGENCY: FHP TROOP K
LAST AGENCY INSPECTION: 11/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 07:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:25

SUBJECT'S LAST NAME: HIDALGO

DATE/TIME OF TEST: 12/05/2021 09:01:51

Diagnostics Check	OK	08:53
Air Blank	0.000	08:54
Control Test	0.082	08:54
Air Blank	0.000	08:55
Subject Sample # 1	0.159	08:56
Breath Volume	1.789	08:56
Air Blank	0.000	08:57
Air Blank	0.000	08:59
Subject Sample # 2	0.164	09:00
Breath Volume	1.242	09:00
Air Blank	0.000	09:00
Control Test	0.081	09:01
Air Blank	0.000	09:01
Diagnostics Check	OK	09:01

Data Download Date/Time: 12/31/2021 00:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 12/01/2021 00:28:56

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.080	00:21
Air Blank	0.000	00:22
Subject Sample # 1	0.176	00:23
Breath Volume	1.164	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:25
Subject Sample # 2	0.161	00:27
Breath Volume	1.128	00:27
Air Blank	0.000	00:28
Control Test	0.079	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 12/10/2021 08:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE, ABIGAIL, D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PAYNE, ABIGAIL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 12/01/2021 22:30:24

Diagnostics Check	OK	22:27	
Air Blank	0.000	22:28	
Control Test	0.081	22:28	
Air Blank	0.000	22:28	
Subject Sample	# 1	22:29	Subject Test Refused
Breath Volume	0.000	22:29	
Air Blank	0.000	22:29	
Control Test	0.080	22:29	
Air Blank	0.000	22:30	
Diagnostics Check	OK	22:30	

Data Download Date/Time: 12/10/2021 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 12/02/2021 00:47:33

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.081	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.000	00:43
Breath Volume	2.578	00:43
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.000	00:46
Breath Volume	2.816	00:46
Air Blank	0.000	00:46
Control Test	0.080	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 12/10/2021 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: GLISSON

DATE/TIME OF TEST: 12/03/2021 23:57:46

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.081	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.080	23:53
Breath Volume	1.636	23:53
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.083	23:56
Breath Volume	1.949	23:56
Air Blank	0.000	23:56
Control Test	0.081	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 12/10/2021 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: GRIMALDO

DATE/TIME OF TEST: 12/04/2021 03:34:22

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.081	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.150	03:29
Breath Volume	1.925	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.143	03:32
Breath Volume	1.601	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 12/10/2021 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES, STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER, BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HERNANDEZ RUIZ

DATE/TIME OF TEST: 12/04/2021 04:21:41

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.156	04:13
Breath Volume	1.113	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.181	04:17
Breath Volume	2.183	04:17
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 3	0.179	04:20
Breath Volume	1.859	04:20
Air Blank	0.000	04:20
Control Test	0.080	04:21
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 12/10/2021 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE, MATTHEW, P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LEE, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: ALLEYNE

DATE/TIME OF TEST: 12/04/2021 05:57:21

Diagnostics Check	OK	05:51
Air Blank	0.000	05:51
Control Test	0.080	05:51
Air Blank	0.000	05:52
Subject Sample # 1	0.160	05:52
Breath Volume	2.167	05:52
Air Blank	0.000	05:53
Air Blank	0.000	05:55
Subject Sample # 2	0.165	05:55
Breath Volume	2.656	05:55
Air Blank	0.000	05:56
Control Test	0.081	05:56
Air Blank	0.000	05:57
Diagnostics Check	OK	05:57

Data Download Date/Time: 12/10/2021 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,RICHARD,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CRUZ,EDGARDO
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: SOACHE DIAZ

DATE/TIME OF TEST: 12/04/2021 19:50:54

Diagnostics Check	OK	19:43
Air Blank	0.000	19:44
Control Test	0.081	19:44
Air Blank	0.000	19:45
Subject Sample # 1	0.260	19:46
Breath Volume	2.039	19:46
Air Blank	0.000	19:46
Air Blank	0.000	19:48
Subject Sample # 2	0.246	19:49
Breath Volume	2.253	19:49
Air Blank	0.000	19:50
Control Test	0.080	19:50
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

Data Download Date/Time: 12/10/2021 08:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MORALES,ORLANDO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 12/04/2021 21:07:37

Diagnostics Check	OK	21:03	
Air Blank	0.000	21:04	
Control Test	0.081	21:04	
Air Blank	0.000	21:05	
Subject Sample	# 1	21:06	Subject Test Refused
Breath Volume	0.382	21:06	
Air Blank	0.000	21:06	
Control Test	0.081	21:07	
Air Blank	0.000	21:07	
Diagnostics Check	OK	21:07	

Data Download Date/Time: 12/10/2021 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAFONTANT,SEDONA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: BRACKETT

DATE/TIME OF TEST: 12/04/2021 23:57:46

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.196	23:52
Breath Volume	1.140	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.195	23:56
Breath Volume	1.968	23:56
Air Blank	0.000	23:56
Control Test	0.080	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 12/10/2021 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: SPURLOCK

DATE/TIME OF TEST: 12/05/2021 00:30:08

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:19	
Control Test	0.079	00:19	
Air Blank	0.000	00:20	
Subject Sample	# 1	00:23	No Sample Provided
Breath Volume	0.000	00:23	
Air Blank	0.000	00:23	
Air Blank	0.000	00:25	
Subject Sample	# 2	00:28	No Sample Provided
Breath Volume	0.000	00:28	
Air Blank	0.000	00:29	
Control Test	0.080	00:29	
Air Blank	0.000	00:30	
Diagnostics Check	OK	00:30	

Data Download Date/Time: 12/10/2021 08:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE, MATTHEW, P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MORALES, ORLANDO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: LEAL FIGUEROA

DATE/TIME OF TEST: 12/05/2021 02:06:38

Diagnostics Check	OK	01:53	
Air Blank	0.000	01:53	
Control Test	0.080	01:54	
Air Blank	0.000	01:54	
Subject Sample	# 1	01:57	No Sample Provided
Breath Volume	0.000	01:57	
Air Blank	0.000	01:58	
Air Blank	0.000	02:00	
Subject Sample	# 2	0.110	02:01
Breath Volume	1.808	02:01	
Air Blank	0.000	02:02	
Air Blank	0.000	02:04	
Subject Sample	# 3	0.112	02:05
Breath Volume	2.257	02:05	
Air Blank	0.000	02:05	
Control Test	0.080	02:06	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:06	

Data Download Date/Time: 12/10/2021 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES, STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES, STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: VEGA VEGA

DATE/TIME OF TEST: 12/05/2021 03:08:29

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:54	
Control Test	0.081	02:54	
Air Blank	0.000	02:55	
Subject Sample # 1	0.000	02:55	
Breath Volume	1.703	02:55	
Air Blank	0.000	02:56	
Air Blank	0.000	02:58	
Subject Sample # 2		03:01	No Sample Provided
Breath Volume	0.000	03:01	
Air Blank	0.000	03:01	
Air Blank	0.000	03:04	
Subject Sample # 3		03:07	No Sample Provided
Breath Volume	0.000	03:07	
Air Blank	0.000	03:07	
Control Test	0.080	03:07	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 12/10/2021 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 12/05/2021 03:19:53

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.189	03:14
Breath Volume	1.527	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.190	03:18
Breath Volume	1.914	03:18
Air Blank	0.000	03:18
Control Test	0.079	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 12/10/2021 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BAGGETT,TYLER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: WEBSTER

DATE/TIME OF TEST: 12/05/2021 04:18:18

Diagnostics Check	OK	04:11
Air Blank	0.000	04:12
Control Test	0.079	04:12
Air Blank	0.000	04:13
Subject Sample # 1	0.212	04:13
Breath Volume	2.394	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.207	04:16
Breath Volume	2.285	04:16
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 12/10/2021 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARBANAS,SHANE,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ARBANAS,SHANE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: REASER

DATE/TIME OF TEST: 12/05/2021 07:53:05

Diagnostics Check	OK	07:45
Air Blank	0.000	07:46
Control Test	0.081	07:46
Air Blank	0.000	07:47
Subject Sample # 1	0.231	07:47
Breath Volume	1.515	07:47
Air Blank	0.000	07:48
Air Blank	0.000	07:50
Subject Sample # 2	0.235	07:51
Breath Volume	2.781	07:51
Air Blank	0.000	07:52
Control Test	0.081	07:52
Air Blank	0.000	07:53
Diagnostics Check	OK	07:53

Data Download Date/Time: 12/10/2021 08:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDRON, DAVID, A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: COLLINS, JOSHUA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:02

SUBJECT'S LAST NAME: MCCORMICK

DATE/TIME OF TEST: 12/05/2021 15:02:43

Diagnostics Check	OK	14:56
Air Blank	0.000	14:56
Control Test	0.081	14:57
Air Blank	0.000	14:57
Subject Sample # 1	0.143	14:58
Breath Volume	1.464	14:58
Air Blank	0.000	14:58
Air Blank	0.000	15:00
Subject Sample # 2	0.155	15:01
Breath Volume	1.179	15:01
Air Blank	0.000	15:01
Control Test	0.081	15:02
Air Blank	0.000	15:02
Diagnostics Check	OK	15:02

Data Download Date/Time: 12/10/2021 08:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: BRANCH

DATE/TIME OF TEST: 12/05/2021 18:56:18

Diagnostics Check	OK	18:52	
Air Blank	0.000	18:52	
Control Test	0.080	18:53	
Air Blank	0.000	18:53	
Subject Sample	# 1	18:54	Subject Test Refused
Breath Volume	0.000	18:54	
Air Blank	0.000	18:55	
Control Test	0.080	18:55	
Air Blank	0.000	18:56	
Diagnostics Check	OK	18:56	

Data Download Date/Time: 12/10/2021 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ, LEONARDO, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DE LA CRUZ, LEONARDO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: HILLHOUSE

DATE/TIME OF TEST: 12/06/2021 18:09:04

Diagnostics Check	OK	18:01
Air Blank	0.000	18:01
Control Test	0.080	18:02
Air Blank	0.000	18:02
Subject Sample # 1	0.000	18:04
Breath Volume	1.750	18:04
Air Blank	0.000	18:05
Air Blank	0.000	18:07
Subject Sample # 2	0.000	18:07
Breath Volume	1.199	18:07
Air Blank	0.000	18:08
Control Test	0.079	18:08
Air Blank	0.000	18:08
Diagnostics Check	OK	18:09

Data Download Date/Time: 12/10/2021 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ, LEONARDO, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BINET, RICKEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ROUSCH

DATE/TIME OF TEST: 12/07/2021 22:58:22

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.080	22:52
Air Blank	0.000	22:52
Subject Sample # 1	0.000	22:54
Breath Volume	1.453	22:54
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.000	22:57
Breath Volume	1.171	22:57
Air Blank	0.000	22:57
Control Test	0.079	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 12/10/2021 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAWSON JR, MICHAEL, E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HODGES, MARCO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: MENDEZ-VELAQUEZ

DATE/TIME OF TEST: 12/08/2021 03:36:46

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.134	03:32
Breath Volume	1.695	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.141	03:35
Breath Volume	1.753	03:35
Air Blank	0.000	03:35
Control Test	0.078	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 12/10/2021 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 12/08/2021 21:15:50

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.079	21:08
Air Blank	0.000	21:08
Subject Sample # 1	0.135	21:10
Breath Volume	1.347	21:10
Air Blank	0.000	21:11
Air Blank	0.000	21:13
Subject Sample # 2	0.146	21:14
Breath Volume	2.519	21:14
Air Blank	0.000	21:14
Control Test	0.079	21:15
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

Data Download Date/Time: 12/10/2021 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 12/09/2021 21:16:49

Diagnostics Check	OK	21:01	
Air Blank	0.000	21:01	
Control Test	0.080	21:01	
Air Blank	0.000	21:02	
Subject Sample	# 1 0.235	21:04	
Breath Volume	2.765	21:04	
Air Blank	0.000	21:04	
Air Blank	0.000	21:06	
Subject Sample	# 2	21:09	No Sample Provided
Breath Volume	0.000	21:09	
Air Blank	0.000	21:10	
Air Blank	0.000	21:12	
Subject Sample	# 3	21:15	No Sample Provided
Breath Volume	0.000	21:15	
Air Blank	0.000	21:15	
Control Test	0.080	21:16	
Air Blank	0.000	21:16	
Diagnostics Check	OK	21:16	

Data Download Date/Time: 12/10/2021 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE, MATTHEW, P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LEE, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: IRIZARRY

DATE/TIME OF TEST: 12/10/2021 03:02:31

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:51	
Control Test	0.080	02:51	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:55	No Sample Provided
Breath Volume	0.000	02:55	
Air Blank	0.000	02:56	
Air Blank	0.000	02:58	
Subject Sample	# 2	03:01	No Sample Provided
Breath Volume	0.000	03:01	
Air Blank	0.000	03:01	
Control Test	0.081	03:01	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 12/10/2021 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTS,JAMES,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HODGES,MARCO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:52

SUBJECT'S LAST NAME: EASTERBROOK

DATE/TIME OF TEST: 12/10/2021 14:23:59

Diagnostics Check	OK	14:17
Air Blank	0.000	14:18
Control Test	0.080	14:18
Air Blank	0.000	14:19
Subject Sample # 1	0.208	14:19
Breath Volume	2.394	14:19
Air Blank	0.000	14:20
Air Blank	0.000	14:22
Subject Sample # 2	0.217	14:22
Breath Volume	2.183	14:22
Air Blank	0.000	14:23
Control Test	0.079	14:23
Air Blank	0.000	14:23
Diagnostics Check	OK	14:23

Data Download Date/Time: 01/11/2022 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: GILROY

DATE/TIME OF TEST: 12/10/2021 21:29:08

Diagnostics Check	OK	21:22
Air Blank	0.000	21:23
Control Test	0.081	21:23
Air Blank	0.000	21:24
Subject Sample # 1	0.000	21:24
Breath Volume	2.093	21:24
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.000	21:27
Breath Volume	1.671	21:27
Air Blank	0.000	21:28
Control Test	0.080	21:28
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 01/11/2022 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: WILLCOCK

DATE/TIME OF TEST: 12/11/2021 00:21:35

Diagnostics Check	OK	00:14
Air Blank	0.000	00:14
Control Test	0.080	00:15
Air Blank	0.000	00:15
Subject Sample # 1	0.190	00:16
Breath Volume	2.472	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.191	00:20
Breath Volume	2.210	00:20
Air Blank	0.000	00:20
Control Test	0.079	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 01/11/2022 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE, MATTHEW, P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MORALES, ORLANDO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: SCANNELL

DATE/TIME OF TEST: 12/11/2021 00:58:06

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.079	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.092	00:52
Breath Volume	1.273	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:54
Subject Sample # 2	0.095	00:56
Breath Volume	2.535	00:56
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 01/11/2022 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGAN, LAKEBRA, S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LEDFORD, ANN
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: NICKERSON

DATE/TIME OF TEST: 12/11/2021 03:17:02

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:06	
Control Test	0.081	03:06	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:10	No Sample Provided
Breath Volume	0.000	03:10	
Air Blank	0.000	03:10	
Air Blank	0.000	03:12	
Subject Sample	# 2	03:15	No Sample Provided
Breath Volume	0.000	03:15	
Air Blank	0.000	03:16	
Control Test	0.080	03:16	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 01/11/2022 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ, LEONARDO, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DE LA CRUZ, LEONARDO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: CALDERON RESTREPO

DATE/TIME OF TEST: 12/11/2021 13:12:52

Diagnostics Check	OK	13:05
Air Blank	0.000	13:05
Control Test	0.080	13:06
Air Blank	0.000	13:06
Subject Sample # 1	0.183	13:08
Breath Volume	1.484	13:08
Air Blank	0.000	13:08
Air Blank	0.000	13:10
Subject Sample # 2	0.182	13:11
Breath Volume	1.292	13:11
Air Blank	0.000	13:11
Control Test	0.078	13:12
Air Blank	0.000	13:12
Diagnostics Check	OK	13:12

Data Download Date/Time: 01/11/2022 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRELL,AUSTIN,G
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MARINO,KEVIN
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/12/2021 00:34:19

Diagnostics Check	OK	00:22	
Air Blank	0.000	00:23	
Control Test	0.081	00:23	
Air Blank	0.000	00:24	
Subject Sample	# 1	00:27	No Sample Provided
Breath Volume	0.000	00:27	
Air Blank	0.000	00:27	
Air Blank	0.000	00:29	
Subject Sample	# 2	00:32	No Sample Provided
Breath Volume	0.000	00:32	
Air Blank	0.000	00:33	
Control Test	0.081	00:33	
Air Blank	0.000	00:34	
Diagnostics Check	OK	00:34	

Data Download Date/Time: 01/11/2022 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: DANDRIDGE

DATE/TIME OF TEST: 12/12/2021 04:09:35

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.081	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.000	04:04
Breath Volume	2.609	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.000	04:08
Breath Volume	1.921	04:08
Air Blank	0.000	04:08
Control Test	0.081	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 01/11/2022 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ, LEONARDO, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DE LA CRUZ, LEONARDO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: MENDEZ PEREZ

DATE/TIME OF TEST: 12/12/2021 13:02:00

Diagnostics Check	OK	12:54
Air Blank	0.000	12:54
Control Test	0.080	12:54
Air Blank	0.000	12:55
Subject Sample # 1	0.152	12:57
Breath Volume	1.289	12:57
Air Blank	0.000	12:57
Air Blank	0.000	12:59
Subject Sample # 2	0.146	13:00
Breath Volume	1.324	13:00
Air Blank	0.000	13:01
Control Test	0.078	13:01
Air Blank	0.000	13:01
Diagnostics Check	OK	13:02

Data Download Date/Time: 01/11/2022 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: CASTONEDA ARGUMEDO

DATE/TIME OF TEST: 12/14/2021 21:29:34

Diagnostics Check	OK	21:22
Air Blank	0.000	21:22
Control Test	0.081	21:23
Air Blank	0.000	21:23
Subject Sample # 1	0.060	21:24
Breath Volume	1.171	21:24
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.064	21:28
Breath Volume	2.312	21:28
Air Blank	0.000	21:28
Control Test	0.081	21:28
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 01/11/2022 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 12/15/2021 03:13:35

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.081	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.131	03:09
Breath Volume	1.312	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.132	03:12
Breath Volume	1.265	03:12
Air Blank	0.000	03:12
Control Test	0.080	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 01/11/2022 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BAGGETT,TYLER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: TYLER

DATE/TIME OF TEST: 12/15/2021 05:58:21

Diagnostics Check	OK	05:51
Air Blank	0.000	05:51
Control Test	0.081	05:52
Air Blank	0.000	05:52
Subject Sample # 1	0.094	05:53
Breath Volume	1.222	05:53
Air Blank	0.000	05:54
Air Blank	0.000	05:55
Subject Sample # 2	0.083	05:56
Breath Volume	2.167	05:56
Air Blank	0.000	05:57
Control Test	0.081	05:57
Air Blank	0.000	05:58
Diagnostics Check	OK	05:58

Data Download Date/Time: 01/11/2022 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAWSON JR, MICHAEL, E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HODGES, MARCO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: VARGAS HERNANDEZ

DATE/TIME OF TEST: 12/17/2021 00:28:29

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.081	00:22
Air Blank	0.000	00:23
Subject Sample # 1	0.165	00:23
Breath Volume	2.488	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.162	00:26
Breath Volume	2.640	00:26
Air Blank	0.000	00:27
Control Test	0.077	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 01/11/2022 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RAGUSA,JOEY
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: ANTHONY

DATE/TIME OF TEST: 12/18/2021 04:11:16

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.081	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.167	04:06
Breath Volume	2.339	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.165	04:09
Breath Volume	1.320	04:09
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 01/11/2022 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: HERMITT

DATE/TIME OF TEST: 12/18/2021 04:55:25

Diagnostics Check	OK	04:49
Air Blank	0.000	04:49
Control Test	0.081	04:49
Air Blank	0.000	04:50
Subject Sample # 1	0.108	04:50
Breath Volume	2.437	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.106	04:53
Breath Volume	1.953	04:53
Air Blank	0.000	04:54
Control Test	0.081	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 01/11/2022 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES, STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RIGGALL, NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 12/18/2021 05:18:43

Diagnostics Check	OK	05:11
Air Blank	0.000	05:12
Control Test	0.081	05:12
Air Blank	0.000	05:13
Subject Sample # 1	0.077	05:14
Breath Volume	1.957	05:14
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.078	05:17
Breath Volume	1.640	05:17
Air Blank	0.000	05:17
Control Test	0.080	05:18
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 01/11/2022 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: STUBL

DATE/TIME OF TEST: 12/18/2021 20:53:43

Diagnostics Check	OK	20:46
Air Blank	0.000	20:46
Control Test	0.079	20:46
Air Blank	0.000	20:47
Subject Sample # 1	0.354	20:49
Breath Volume	1.800	20:49
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.354	20:52
Breath Volume	1.226	20:52
Air Blank	0.000	20:52
Control Test	0.078	20:53
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

Data Download Date/Time: 01/11/2022 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES, STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MORALES, ORLANDO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: SIMILIEN

DATE/TIME OF TEST: 12/18/2021 21:15:59

Diagnostics Check	OK	21:07	
Air Blank	0.000	21:08	
Control Test	0.079	21:08	
Air Blank	0.000	21:08	
Subject Sample	# 1	21:12	No Sample Provided
Breath Volume	0.000	21:12	
Air Blank	0.000	21:12	
Air Blank	0.000	21:14	
Subject Sample	# 2	21:14	Subject Test Refused
Breath Volume	0.000	21:14	
Air Blank	0.000	21:15	
Control Test	0.080	21:15	
Air Blank	0.000	21:15	
Diagnostics Check	OK	21:15	

Data Download Date/Time: 01/11/2022 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES, STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES, STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: SHAFFER

DATE/TIME OF TEST: 12/19/2021 00:45:15

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:37	
Control Test	0.081	00:37	
Air Blank	0.000	00:38	
Subject Sample # 1	0.179	00:39	
Breath Volume	1.171	00:39	
Air Blank	0.000	00:40	
Air Blank	0.000	00:42	
Subject Sample # 2	0.181	00:44	
Breath Volume	1.242	00:44	
Air Blank		00:44	Improper Sample
Air Blank	0.000	00:45	

Data Download Date/Time: 01/11/2022 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 12/20/2021 02:25:19

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.081	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.053	02:20
Breath Volume	1.390	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.050	02:23
Breath Volume	1.476	02:23
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 01/11/2022 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSTOS,ALEJANDRO,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BUSTOS,ALEJANDRO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: HO

DATE/TIME OF TEST: 12/20/2021 05:17:33

Diagnostics Check	OK	05:06	
Air Blank	0.000	05:06	
Control Test	0.080	05:06	
Air Blank	0.000	05:07	
Subject Sample	# 1	05:10	No Sample Provided
Breath Volume	0.000	05:10	
Air Blank	0.000	05:11	
Air Blank	0.000	05:13	
Subject Sample	# 2	05:16	No Sample Provided
Breath Volume	0.000	05:16	
Air Blank	0.000	05:16	
Control Test	0.081	05:16	
Air Blank	0.000	05:17	
Diagnostics Check	OK	05:17	

Data Download Date/Time: 01/11/2022 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSTOS,ALEJANDRO,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BUSTOS,ALEJANDRO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: HO

DATE/TIME OF TEST: 12/20/2021 05:22:59

Diagnostics Check	OK	05:19	
Air Blank	0.000	05:20	
Control Test	0.081	05:20	
Air Blank	0.000	05:21	
Subject Sample	# 1	05:21	Subject Test Refused
Breath Volume	0.000	05:21	
Air Blank	0.000	05:22	
Control Test	0.080	05:22	
Air Blank	0.000	05:22	
Diagnostics Check	OK	05:22	

Data Download Date/Time: 01/11/2022 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: KEAHL

DATE/TIME OF TEST: 12/22/2021 22:55:57

Diagnostics Check	OK	22:48
Air Blank	0.000	22:48
Control Test	0.080	22:49
Air Blank	0.000	22:49
Subject Sample # 1	0.000	22:51
Breath Volume	1.394	22:51
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.000	22:54
Breath Volume	1.613	22:54
Air Blank	0.000	22:55
Control Test	0.079	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 01/11/2022 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: ARENAS

DATE/TIME OF TEST: 12/23/2021 01:22:26

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.081	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.211	01:17
Breath Volume	2.316	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.197	01:20
Breath Volume	1.843	01:20
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 01/11/2022 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROPST,JOHN,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: COLLINS,JAMES
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 09:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:44

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 12/23/2021 11:29:09

Diagnostics Check	OK	11:21
Air Blank	0.000	11:22
Control Test	0.081	11:22
Air Blank	0.000	11:23
Subject Sample # 1	0.133	11:24
Breath Volume	1.882	11:24
Air Blank	0.000	11:25
Air Blank	0.000	11:27
Subject Sample # 2	0.122	11:27
Breath Volume	2.234	11:27
Air Blank	0.000	11:28
Control Test	0.079	11:28
Air Blank	0.000	11:29
Diagnostics Check	OK	11:29

Data Download Date/Time: 01/11/2022 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES, STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES, STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 12/23/2021 22:42:28

Diagnostics Check	OK	22:31	
Air Blank	0.000	22:31	
Control Test	0.081	22:31	
Air Blank	0.000	22:32	
Subject Sample	# 1	22:35	No Sample Provided
Breath Volume	0.000	22:35	
Air Blank	0.000	22:35	
Air Blank	0.000	22:37	
Subject Sample	# 2	22:41	No Sample Provided
Breath Volume	0.000	22:41	
Air Blank	0.000	22:41	
Control Test	0.081	22:41	
Air Blank	0.000	22:42	
Diagnostics Check	OK	22:42	

Data Download Date/Time: 01/11/2022 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ROQUE,ALEXANDER
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: PALALA JUAREZ

DATE/TIME OF TEST: 12/23/2021 23:50:57

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.156	23:45
Breath Volume	2.070	23:45
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.149	23:49
Breath Volume	1.871	23:49
Air Blank	0.000	23:50
Control Test	0.080	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 01/11/2022 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RAINES,JACOB
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 12/24/2021 02:00:36

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:58	
Control Test	0.081	01:58	
Air Blank	0.000	01:58	
Subject Sample	# 1	01:59	Subject Test Refused
Breath Volume	0.000	01:59	
Air Blank	0.000	01:59	
Control Test	0.081	02:00	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

Data Download Date/Time: 01/11/2022 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: LIGHTSEY

DATE/TIME OF TEST: 12/24/2021 03:04:30

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:53	
Control Test	0.081	02:53	
Air Blank	0.000	02:54	
Subject Sample	# 1	02:57	No Sample Provided
Breath Volume	0.000	02:57	
Air Blank	0.000	02:58	
Air Blank	0.000	03:00	
Subject Sample	# 2	03:03	No Sample Provided
Breath Volume	0.000	03:03	
Air Blank	0.000	03:03	
Control Test	0.080	03:03	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:04	

Data Download Date/Time: 01/11/2022 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,MELISSA,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DOUGLAS,WILLIAM
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: ARMENTA

DATE/TIME OF TEST: 12/25/2021 01:23:13

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.081	01:17
Air Blank	0.000	01:17
Subject Sample # 1	0.186	01:18
Breath Volume	1.308	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.171	01:21
Breath Volume	1.390	01:21
Air Blank	0.000	01:22
Control Test	0.080	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 01/11/2022 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,MELISSA,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HODGES,MARCO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 12/25/2021 01:40:26

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:29	
Control Test	0.080	01:29	
Air Blank	0.000	01:30	
Subject Sample	# 1	01:33	No Sample Provided
Breath Volume	0.000	01:33	
Air Blank	0.000	01:33	
Air Blank	0.000	01:35	
Subject Sample	# 2	01:39	No Sample Provided
Breath Volume	0.000	01:39	
Air Blank	0.000	01:39	
Control Test	0.080	01:39	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 01/11/2022 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/25/2021 02:40:46

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:29	
Control Test	0.080	02:29	
Air Blank	0.000	02:29	
Subject Sample	# 1 0.152	02:33	Volume Not Met
Breath Volume	0.328	02:33	
Air Blank	0.000	02:33	
Air Blank	0.000	02:35	
Subject Sample	# 2 0.166	02:36	
Breath Volume	2.726	02:36	
Air Blank	0.000	02:37	
Air Blank	0.000	02:38	
Subject Sample	# 3 0.162	02:39	
Breath Volume	1.757	02:39	
Air Blank	0.000	02:39	
Control Test	0.079	02:40	
Air Blank	0.000	02:40	
Diagnostics Check	OK	02:40	

Data Download Date/Time: 01/11/2022 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ, LEONARDO, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DE LA CRUZ, LEONARDO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CARBO ALVARADO

DATE/TIME OF TEST: 12/25/2021 03:54:28

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.192	03:49
Breath Volume	1.195	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.185	03:52
Breath Volume	1.496	03:52
Air Blank	0.000	03:53
Control Test	0.078	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 01/11/2022 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE, ABIGAIL, D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PAYNE, ABIGAIL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: GALAN

DATE/TIME OF TEST: 12/25/2021 04:19:12

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.080	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.132	04:14
Breath Volume	1.585	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:16
Subject Sample # 2	0.132	04:17
Breath Volume	1.898	04:17
Air Blank	0.000	04:18
Control Test	0.080	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 01/11/2022 16:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRELL,AUSTIN,G
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HODGES,MARCO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: BANUELOS

DATE/TIME OF TEST: 12/26/2021 04:19:05

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:08	
Control Test	0.081	04:08	
Air Blank	0.000	04:08	
Subject Sample	# 1	04:12	No Sample Provided
Breath Volume	0.000	04:12	
Air Blank	0.000	04:12	
Air Blank	0.000	04:14	
Subject Sample	# 2	04:17	No Sample Provided
Breath Volume	0.000	04:17	
Air Blank	0.000	04:18	
Control Test	0.081	04:18	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:19	

Data Download Date/Time: 01/11/2022 16:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 12/26/2021 05:55:27

Diagnostics Check	OK	05:48
Air Blank	0.000	05:49
Control Test	0.081	05:49
Air Blank	0.000	05:49
Subject Sample # 1	0.211	05:50
Breath Volume	1.496	05:50
Air Blank	0.000	05:51
Air Blank	0.000	05:53
Subject Sample # 2	0.206	05:53
Breath Volume	2.222	05:53
Air Blank	0.000	05:54
Control Test	0.081	05:54
Air Blank	0.000	05:55
Diagnostics Check	OK	05:55

Data Download Date/Time: 01/11/2022 16:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROPST,JOHN,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PROPST,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:35

SUBJECT'S LAST NAME: TULLIS

DATE/TIME OF TEST: 12/26/2021 13:07:27

Diagnostics Check	OK	13:00
Air Blank	0.000	13:00
Control Test	0.081	13:01
Air Blank	0.000	13:01
Subject Sample # 1	0.000	13:02
Breath Volume	1.433	13:02
Air Blank	0.000	13:03
Air Blank	0.000	13:05
Subject Sample # 2	0.000	13:06
Breath Volume	1.257	13:06
Air Blank	0.000	13:06
Control Test	0.081	13:06
Air Blank	0.000	13:07
Diagnostics Check	OK	13:07

Data Download Date/Time: 01/11/2022 16:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRELL,AUSTIN,G
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HODGES,MARCO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: TALLON

DATE/TIME OF TEST: 12/27/2021 00:04:07

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.081	23:58
Air Blank	0.000	23:59
Subject Sample # 2	0.000	23:59
Breath Volume	3.574	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 1	0.000	00:02
Breath Volume	2.546	00:02
Air Blank	0.000	00:03
Control Test	0.080	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 01/11/2022 16:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE, ABIGAIL, D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PAYNE, ABIGAIL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: BARADA RIVERA

DATE/TIME OF TEST: 12/27/2021 00:15:08

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.081	00:09
Air Blank	0.000	00:09
Subject Sample # 1	0.000	00:10
Breath Volume	2.191	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.000	00:13
Breath Volume	2.175	00:13
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 01/11/2022 16:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: HATFIELD

DATE/TIME OF TEST: 12/27/2021 21:17:39

Diagnostics Check	OK	21:10
Air Blank	0.000	21:11
Control Test	0.080	21:11
Air Blank	0.000	21:12
Subject Sample # 1	0.134	21:12
Breath Volume	2.332	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.132	21:16
Breath Volume	2.406	21:16
Air Blank	0.000	21:16
Control Test	0.080	21:17
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 01/11/2022 16:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: VALERIO,ANTHONY
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: MONGOLD

DATE/TIME OF TEST: 12/27/2021 21:33:42

Diagnostics Check	OK	21:29	
Air Blank	0.000	21:29	
Control Test	0.080	21:29	
Air Blank	0.000	21:30	
Subject Sample	# 1	21:32	Subject Test Refused
Breath Volume	0.000	21:32	
Air Blank	0.000	21:32	
Control Test	0.080	21:33	
Air Blank	0.000	21:33	
Diagnostics Check	OK	21:33	

Data Download Date/Time: 01/11/2022 16:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES, STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES, STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/28/2021 00:50:33

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.080	00:43
Air Blank	0.000	00:43
Subject Sample # 1	0.119	00:45
Breath Volume	1.226	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.122	00:49
Breath Volume	1.320	00:49
Air Blank	0.000	00:49
Control Test	0.080	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 01/11/2022 16:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROPST,JOHN,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PROPST,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 09:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:47

SUBJECT'S LAST NAME: WATERS

DATE/TIME OF TEST: 12/28/2021 11:31:07

Diagnostics Check	OK	11:19	
Air Blank	0.000	11:20	
Control Test	0.080	11:20	
Air Blank	0.000	11:21	
Subject Sample	# 1	11:24	No Sample Provided
Breath Volume	0.000	11:24	
Air Blank	0.000	11:24	
Air Blank	0.000	11:26	
Subject Sample	# 2	11:29	No Sample Provided
Breath Volume	0.000	11:29	
Air Blank	0.000	11:30	
Control Test	0.080	11:30	
Air Blank	0.000	11:31	
Diagnostics Check	OK	11:31	

Data Download Date/Time: 01/11/2022 16:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCH,JOHN,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHURCH,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 12/28/2021 21:21:19

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:19	
Control Test	0.081	21:19	
Air Blank	0.000	21:19	
Subject Sample	# 1	21:19	Subject Test Refused
Breath Volume	0.000	21:19	
Air Blank	0.000	21:20	
Control Test	0.081	21:20	
Air Blank	0.000	21:21	
Diagnostics Check	OK	21:21	

Data Download Date/Time: 01/11/2022 16:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:28

SUBJECT'S LAST NAME: KVETON

DATE/TIME OF TEST: 12/29/2021 08:01:32

Diagnostics Check	OK	07:55
Air Blank	0.000	07:55
Control Test	0.080	07:56
Air Blank	0.000	07:56
Subject Sample # 1	0.161	07:57
Breath Volume	1.187	07:57
Air Blank	0.000	07:57
Air Blank	0.000	07:59
Subject Sample # 2	0.164	08:00
Breath Volume	1.558	08:00
Air Blank	0.000	08:00
Control Test	0.080	08:00
Air Blank	0.000	08:01
Diagnostics Check	OK	08:01

Data Download Date/Time: 01/11/2022 16:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ALBERT,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HODGES,MARCO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: CASE

DATE/TIME OF TEST: 12/31/2021 01:10:02

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.080	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.146	01:05
Breath Volume	1.542	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.145	01:08
Breath Volume	2.285	01:08
Air Blank	0.000	01:09
Control Test	0.080	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:10

Data Download Date/Time: 01/11/2022 16:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,MELISSA,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DOYON,LUC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: QAMAR

DATE/TIME OF TEST: 12/31/2021 02:45:56

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:34	
Control Test	0.081	02:35	
Air Blank	0.000	02:35	
Subject Sample	# 1 0.146	02:38	Volume Not Met
Breath Volume	0.101	02:38	
Air Blank	0.000	02:39	
Air Blank	0.000	02:41	
Subject Sample	# 2 0.039	02:44	Volume Not Met
Breath Volume	0.375	02:44	
Air Blank	0.000	02:45	
Control Test	0.077	02:45	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 01/11/2022 16:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: HOLLENBERG

DATE/TIME OF TEST: 12/31/2021 21:39:53

Diagnostics Check	OK	21:33
Air Blank	0.000	21:34
Control Test	0.080	21:34
Air Blank	0.000	21:34
Subject Sample # 1	0.000	21:35
Breath Volume	2.714	21:35
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.000	21:38
Breath Volume	2.687	21:38
Air Blank	0.000	21:38
Control Test	0.080	21:39
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 01/11/2022 16:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITSON,GEOFFREY,B
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WHITSON,GEOFFREY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 13:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 12/14/2021 14:45:39

Diagnostics Check	OK	14:39
Air Blank	0.000	14:39
Control Test	0.078	14:40
Air Blank	0.000	14:40
Subject Sample # 1	0.248	14:41
Breath Volume	2.332	14:41
Air Blank	0.000	14:41
Air Blank	0.000	14:43
Subject Sample # 2	0.246	14:44
Breath Volume	2.898	14:44
Air Blank	0.000	14:44
Control Test	0.078	14:45
Air Blank	0.000	14:45
Diagnostics Check	OK	14:45

Data Download Date/Time: 01/21/2022 09:50