

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLSON, TRAVON, L
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: RODRIGUEZ, YADIEL
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 02/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: MARTINEZ TORRES

DATE/TIME OF TEST: 03/01/2021 03:33:50

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:31	
Control Test	0.080	03:31	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Control Test	0.080	03:33	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 03/31/2021 17:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 02/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: LOZOYA

DATE/TIME OF TEST: 03/04/2021 22:41:46

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.080	22:35
Air Blank	0.000	22:36
Subject Sample # 1	0.243	22:36
Breath Volume	2.304	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.226	22:39
Breath Volume	2.820	22:39
Air Blank	0.000	22:40
Control Test	0.079	22:41
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 03/31/2021 17:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: BOWENS,BRANDON
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 02/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: ROLLE

DATE/TIME OF TEST: 03/05/2021 00:50:51

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:42	
Control Test	0.082	00:43	
Air Blank	0.000	00:43	
Subject Sample	# 1	00:46	No Sample Provided
Breath Volume	0.000	00:46	
Air Blank	0.000	00:47	
Air Blank	0.000	00:49	
Subject Sample	# 2	00:49	Subject Test Refused
Breath Volume	0.000	00:49	
Air Blank	0.000	00:49	
Control Test	0.080	00:50	
Air Blank	0.000	00:50	
Diagnostics Check	OK	00:50	

Data Download Date/Time: 03/31/2021 17:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLSON, TRAVON, L
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: NICHOLSON, TRAVON
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 02/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: GODINEZ GARCIA

DATE/TIME OF TEST: 03/05/2021 23:03:22

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.079	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.209	22:58
Breath Volume	1.648	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.208	23:01
Breath Volume	2.042	23:01
Air Blank	0.000	23:02
Control Test	0.079	23:02
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 03/31/2021 17:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINCORE, MICHAEL, A
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: HERNANDEZ, XAVIER
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 02/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1237851
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: RAYMUNDO CORIO

DATE/TIME OF TEST: 03/11/2021 00:07:33

Diagnostics Check		OK	23:53	
Air Blank		0.000	23:53	
Control Test		0.080	23:53	
Air Blank		0.000	23:54	
Subject Sample	# 3	0.198	23:57	Volume Not Met
Breath Volume		0.234	23:57	
Air Blank		0.000	23:58	
Air Blank		0.000	00:00	
Subject Sample	# 1	0.216	00:01	No .020 agreement
Breath Volume		1.296	00:01	
Air Blank		0.000	00:01	
Air Blank		0.000	00:03	
Subject Sample	# 2	0.238	00:05	No .020 agreement
Breath Volume		1.933	00:05	
Air Blank		0.000	00:06	
Control Test		0.079	00:06	
Air Blank		0.000	00:07	
Diagnostics Check		OK	00:07	

Data Download Date/Time: 03/31/2021 17:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GIBSON,NICHOLAS
ARRESTING OFFICER AGENCY: FLORIDA CITY PD
LAST AGENCY INSPECTION: 02/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/13/2021 03:40:19

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.160	03:35
Breath Volume	2.218	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.162	03:38
Breath Volume	1.574	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 03/31/2021 17:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLSON, TRAVON, L
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: PEREZ, RAUL
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 02/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: LALUZ JR

DATE/TIME OF TEST: 03/19/2021 01:26:30

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:23	
Control Test	0.080	01:24	
Air Blank	0.000	01:24	
Subject Sample	# 1	01:25	Subject Test Refused
Breath Volume	0.000	01:25	
Air Blank	0.000	01:25	
Control Test	0.080	01:25	
Air Blank	0.000	01:26	
Diagnostics Check	OK	01:26	

Data Download Date/Time: 03/31/2021 17:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: SMART,C
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 02/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: MORALES MENDEZ

DATE/TIME OF TEST: 03/26/2021 21:25:35

Diagnostics Check	OK	21:17	
Air Blank	0.000	21:17	
Control Test	0.081	21:17	
Air Blank	0.000	21:18	
Subject Sample	# 1	21:21	No Sample Provided
Breath Volume	0.000	21:21	
Air Blank	0.000	21:21	
Air Blank	0.000	21:23	
Subject Sample	# 2	21:24	Subject Test Refused
Breath Volume	0.000	21:24	
Air Blank	0.000	21:24	
Control Test	0.080	21:24	
Air Blank	0.000	21:25	
Diagnostics Check	OK	21:25	

Data Download Date/Time: 03/31/2021 17:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE, KRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KRUSE, KRISTOPHER
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 03/11/2021 23:51:36

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.078	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.141	23:46
Breath Volume	3.441	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:48
Subject Sample # 2	0.133	23:49
Breath Volume	3.300	23:49
Air Blank	0.000	23:50
Control Test	0.077	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 04/26/2021 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: MERCER,JOSH
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 03/13/2021 00:31:40

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.078	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.137	00:26
Breath Volume	1.996	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.137	00:29
Breath Volume	2.375	00:29
Air Blank	0.000	00:30
Control Test	0.078	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 04/26/2021 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,NICHOLAS,L
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: PEREZ,NICHOLAS
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 03/13/2021 03:18:19

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.144	03:12
Breath Volume	1.312	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.140	03:16
Breath Volume	2.679	03:16
Air Blank	0.000	03:17
Control Test	0.077	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 04/26/2021 15:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,BROOKE,M
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: PARISH,CLINT
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 03/14/2021 04:47:05

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.078	04:41
Air Blank	0.000	04:41
Subject Sample # 1	0.190	04:42
Breath Volume	2.335	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:44
Subject Sample # 2	0.183	04:45
Breath Volume	1.503	04:45
Air Blank	0.000	04:46
Control Test	0.077	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 04/26/2021 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADE,CODY,L
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: RICKETTS,J
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: ALVAREZ GUTIERREZ

DATE/TIME OF TEST: 03/14/2021 23:03:39

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.079	22:58
Air Blank	0.000	22:58
Subject Sample # 1	0.209	22:59
Breath Volume	1.914	22:59
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.201	23:01
Breath Volume	1.183	23:01
Air Blank	0.000	23:02
Control Test	0.075	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 04/26/2021 15:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: DANIELS,DEVIN
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: STANDLEY

DATE/TIME OF TEST: 03/15/2021 21:25:00

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:20	
Control Test	0.078	21:21	
Air Blank	0.000	21:21	
Subject Sample	# 1	21:23	Subject Test Refused
Breath Volume	0.000	21:23	
Air Blank	0.000	21:24	
Control Test	0.078	21:24	
Air Blank	0.000	21:24	
Diagnostics Check	OK	21:25	

Data Download Date/Time: 04/26/2021 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERRY, WAYNE, G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: PERRY, WAYNE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: GRAZIANO

DATE/TIME OF TEST: 03/19/2021 10:30:43

Diagnostics Check	OK	10:22	
Air Blank	0.000	10:23	
Control Test	0.079	10:23	
Air Blank	0.000	10:23	
Subject Sample # 1	0.208	10:25	
Breath Volume	2.171	10:25	
Air Blank	0.000	10:26	
Air Blank	0.000	10:28	
Subject Sample # 2	0.212	10:29	
Breath Volume	2.132	10:29	
Air Blank	0.000	10:29	
Control Test	0.072	10:30	Control Outside Tolerance
Air Blank	0.000	10:30	

Data Download Date/Time: 04/26/2021 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERRY,WAYNE,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: PERRY,PERRY
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: GRAZIANO

DATE/TIME OF TEST: 03/19/2021 10:46:16

Diagnostics Check	OK	10:38
Air Blank	0.000	10:39
Control Test	0.077	10:39
Air Blank	0.000	10:40
Subject Sample # 1	0.213	10:41
Breath Volume	1.847	10:41
Air Blank	0.000	10:41
Air Blank	0.000	10:43
Subject Sample # 2	0.199	10:44
Breath Volume	2.121	10:44
Air Blank	0.000	10:45
Control Test	0.077	10:45
Air Blank	0.000	10:46
Diagnostics Check	OK	10:46

Data Download Date/Time: 04/26/2021 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE, KRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: PEREZ, CALEB
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 03/20/2021 01:45:20

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.093	01:40
Breath Volume	1.207	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:42
Subject Sample # 2	0.092	01:43
Breath Volume	1.265	01:43
Air Blank	0.000	01:44
Control Test	0.079	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 04/26/2021 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PARISH,C
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: CLASS

DATE/TIME OF TEST: 03/20/2021 02:17:43

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.079	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.145	02:13
Breath Volume	1.789	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.164	02:16
Breath Volume	2.074	02:16
Air Blank	0.000	02:16
Control Test	0.077	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 04/26/2021 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PARISH,C
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 03/20/2021 18:25:37

Diagnostics Check	OK	18:19	
Air Blank	0.000	18:20	
Control Test	0.079	18:20	
Air Blank	0.000	18:21	
Subject Sample # 1	0.207	18:21	
Breath Volume	1.601	18:21	
Air Blank	0.000	18:22	
Air Blank	0.000	18:24	
Subject Sample # 2	0.212	18:25	
Breath Volume	1.273	18:25	
Air Blank		18:25	RFI Detect
Air Blank		18:25	RFI Detect

Data Download Date/Time: 04/26/2021 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: OREILLY

DATE/TIME OF TEST: 03/20/2021 21:21:25

Diagnostics Check	OK	21:14
Air Blank	0.000	21:15
Control Test	0.079	21:15
Air Blank	0.000	21:15
Subject Sample # 1	0.218	21:16
Breath Volume	1.964	21:16
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample # 2	0.234	21:19
Breath Volume	2.082	21:19
Air Blank	0.000	21:20
Control Test	0.077	21:20
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

Data Download Date/Time: 04/26/2021 15:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: BENTIL,GERALD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: TEELE

DATE/TIME OF TEST: 03/21/2021 03:17:32

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.079	03:13	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:16	Subject Test Refused
Breath Volume	0.000	03:16	
Air Blank	0.000	03:16	
Control Test	0.079	03:16	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 04/26/2021 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,BROOKE,M
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: WHEELER,BROOKE
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BIEVENOUR

DATE/TIME OF TEST: 03/23/2021 00:28:12

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.079	00:22
Air Blank	0.000	00:22
Subject Sample # 1	0.000	00:23
Breath Volume	2.894	00:23
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.000	00:26
Breath Volume	2.257	00:26
Air Blank	0.000	00:27
Control Test	0.078	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 04/26/2021 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,BROOKE,M
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: PARDO,JUAN
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: VIVKERS

DATE/TIME OF TEST: 03/23/2021 04:35:33

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.078	04:29
Air Blank	0.000	04:30
Subject Sample # 1	0.147	04:30
Breath Volume	1.906	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:32
Subject Sample # 2	0.140	04:33
Breath Volume	2.664	04:33
Air Blank	0.000	04:34
Control Test	0.077	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 04/26/2021 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: RICKETTS,J
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 03/24/2021 04:26:54

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.078	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.178	04:22
Breath Volume	2.406	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:24
Subject Sample # 2	0.166	04:25
Breath Volume	2.210	04:25
Air Blank	0.000	04:25
Control Test	0.077	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 04/26/2021 15:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 03/24/2021 04:37:45

Diagnostics Check		OK	04:30
Air Blank		0.000	04:30
Control Test		0.077	04:31
Air Blank		0.000	04:31
Subject Sample	# 1	0.162	04:32
Breath Volume		1.679	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:34
Subject Sample	# 2	0.169	04:36
Breath Volume		2.332	04:36
Air Blank		0.000	04:36
Control Test		0.077	04:37
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 04/26/2021 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADE,CODY,L
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: WADE,CODY
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 03/26/2021 23:37:36

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.079	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.327	23:32
Breath Volume	3.023	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.339	23:35
Breath Volume	3.828	23:35
Air Blank	0.000	23:36
Control Test	0.077	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 04/26/2021 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,BROOKE,M
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: RICKETTS,JERRRMIAH
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: FREND

DATE/TIME OF TEST: 03/27/2021 20:12:30

Diagnostics Check	OK	20:04	
Air Blank	0.000	20:04	
Control Test	0.078	20:04	
Air Blank	0.000	20:05	
Subject Sample	# 1	20:08	No Sample Provided
Breath Volume	0.000	20:08	
Air Blank	0.000	20:08	
Air Blank	0.000	20:10	
Subject Sample	# 2	20:11	Subject Test Refused
Breath Volume	0.000	20:11	
Air Blank	0.000	20:11	
Control Test	0.078	20:11	
Air Blank	0.000	20:12	
Diagnostics Check	OK	20:12	

Data Download Date/Time: 04/26/2021 15:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY,KEVIN,J
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: GUY,KEVIN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 03/28/2021 05:58:55

Diagnostics Check	OK	05:56	
Air Blank	0.000	05:56	
Control Test	0.079	05:56	
Air Blank	0.000	05:57	
Subject Sample	# 1	05:57	Subject Test Refused
Breath Volume	0.000	05:57	
Air Blank	0.000	05:57	
Control Test	0.078	05:58	
Air Blank	0.000	05:58	
Diagnostics Check	OK	05:58	

Data Download Date/Time: 04/26/2021 15:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BILBERY,M
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 03/29/2021 19:38:10

Diagnostics Check	OK	19:30
Air Blank	0.000	19:31
Control Test	0.078	19:31
Air Blank	0.000	19:32
Subject Sample # 1	0.213	19:33
Breath Volume	2.140	19:33
Air Blank	0.000	19:34
Air Blank	0.000	19:35
Subject Sample # 2	0.204	19:36
Breath Volume	1.515	19:36
Air Blank	0.000	19:37
Control Test	0.076	19:37
Air Blank	0.000	19:38
Diagnostics Check	OK	19:38

Data Download Date/Time: 04/26/2021 15:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYNN,JOSEPH,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CAPROIZZO,KELLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 03/01/2021 10:30:58

Diagnostics Check	OK	10:25
Air Blank	0.000	10:25
Control Test	0.079	10:25
Air Blank	0.000	10:26
Subject Sample # 1	0.000	10:26
Breath Volume	2.875	10:26
Air Blank	0.000	10:27
Air Blank	0.000	10:29
Subject Sample # 2	0.000	10:29
Breath Volume	3.042	10:29
Air Blank	0.000	10:30
Control Test	0.076	10:30
Air Blank	0.000	10:30
Diagnostics Check	OK	10:30

Data Download Date/Time: 04/02/2021 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JUSTIN,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ADAMS,JUSTIN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 03/07/2021 20:47:56

Diagnostics Check	OK	20:40
Air Blank	0.000	20:41
Control Test	0.079	20:41
Air Blank	0.000	20:42
Subject Sample # 1	0.278	20:43
Breath Volume	2.121	20:43
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.283	20:46
Breath Volume	2.261	20:46
Air Blank	0.000	20:47
Control Test	0.078	20:47
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 04/02/2021 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TREVOR,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LEWIS,TREVOR
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:25

SUBJECT'S LAST NAME: TARICHE

DATE/TIME OF TEST: 03/10/2021 10:55:05

Diagnostics Check	OK	10:46	
Air Blank	0.000	10:46	
Control Test	0.080	10:47	
Air Blank	0.000	10:47	
Subject Sample	# 1	10:50	No Sample Provided
Breath Volume	0.000	10:50	
Air Blank	0.000	10:51	
Air Blank	0.000	10:53	
Subject Sample	# 2	10:53	Subject Test Refused
Breath Volume	0.000	10:53	
Air Blank	0.000	10:54	
Control Test	0.080	10:54	
Air Blank	0.000	10:55	
Diagnostics Check	OK	10:55	

Data Download Date/Time: 04/02/2021 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TREVOR,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LEWIS,TREVOR
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:11

SUBJECT'S LAST NAME: TARICHE

DATE/TIME OF TEST: 03/10/2021 11:46:25

Diagnostics Check		OK	11:33	
Air Blank		0.000	11:33	
Control Test		0.080	11:33	
Air Blank		0.000	11:34	
Subject Sample	# 1	0.014	11:37	Volume Not Met
Breath Volume		0.199	11:37	
Air Blank		0.000	11:38	
Air Blank		0.000	11:39	
Subject Sample	# 2	0.015	11:41	
Breath Volume		1.640	11:41	
Air Blank		0.000	11:42	
Air Blank		0.000	11:44	
Subject Sample	# 3	0.014	11:44	
Breath Volume		1.335	11:44	
Air Blank		0.000	11:45	
Control Test		0.079	11:45	
Air Blank		0.000	11:46	
Diagnostics Check		OK	11:46	

Data Download Date/Time: 04/02/2021 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN, BRADLEY,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MCLAUGHLIN, BRADLEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 03/18/2021 04:05:43

Diagnostics Check		OK	03:58
Air Blank		0.000	03:59
Control Test		0.080	03:59
Air Blank		0.000	04:00
Subject Sample	# 1	0.205	04:01
Breath Volume		4.363	04:01
Air Blank		0.000	04:01
Air Blank		0.000	04:03
Subject Sample	# 2	0.201	04:03
Breath Volume		3.269	04:03
Air Blank		0.000	04:04
Control Test		0.077	04:05
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 04/02/2021 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CODDING,JOSHUA,S
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CODDING,JOSHUA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/18/2021 21:56:51

Diagnostics Check	OK	21:49
Air Blank	0.000	21:49
Control Test	0.080	21:50
Air Blank	0.000	21:50
Subject Sample # 1	0.299	21:52
Breath Volume	3.433	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:54
Subject Sample # 2	0.294	21:55
Breath Volume	3.296	21:55
Air Blank	0.000	21:55
Control Test	0.078	21:56
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 04/02/2021 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN, BRADLEY,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FARRAGHER, THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: GULLIGE

DATE/TIME OF TEST: 03/19/2021 03:06:53

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.080	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.143	03:02
Breath Volume	3.023	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:04
Subject Sample # 2	0.141	03:05
Breath Volume	3.187	03:05
Air Blank	0.000	03:05
Control Test	0.077	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 04/02/2021 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JUSTIN,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: JONES,HOLLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: CIEKOT

DATE/TIME OF TEST: 03/21/2021 02:22:31

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.080	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.249	02:17
Breath Volume	1.402	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.246	02:20
Breath Volume	1.371	02:20
Air Blank	0.000	02:21
Control Test	0.077	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 04/02/2021 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN, BRADLEY,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MATTHEWS, GREGORY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: AGUNDEZ

DATE/TIME OF TEST: 03/23/2021 23:20:03

Diagnostics Check		OK	23:13
Air Blank		0.000	23:13
Control Test		0.079	23:13
Air Blank		0.000	23:14
Subject Sample	# 1	0.180	23:15
Breath Volume		2.601	23:15
Air Blank		0.000	23:16
Air Blank		0.000	23:17
Subject Sample	# 2	0.163	23:18
Breath Volume		2.074	23:18
Air Blank		0.000	23:19
Control Test		0.077	23:19
Air Blank		0.000	23:19
Diagnostics Check		OK	23:20

Data Download Date/Time: 04/02/2021 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON, RAYMOND, M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: SCelfo, ZACHARY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 03/05/2021 03:11:29

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:09	
Control Test	0.081	03:09	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:10	Subject Test Refused
Breath Volume	0.000	03:10	
Air Blank	0.000	03:10	
Control Test	0.081	03:10	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

Data Download Date/Time: 03/30/2021 18:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON, RAYMOND, M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: COLON, RAYMOND
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: MCVAY

DATE/TIME OF TEST: 03/10/2021 03:34:47

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:32	
Control Test	0.082	03:32	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:33	Subject Test Refused
Breath Volume	0.000	03:33	
Air Blank	0.000	03:33	
Control Test	0.081	03:34	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 03/30/2021 18:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON, RAYMOND, M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: SCELFO, ZACHARY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 03/13/2021 03:24:34

Diagnostics Check		OK	03:17
Air Blank		0.000	03:17
Control Test		0.082	03:17
Air Blank		0.000	03:18
Subject Sample	# 1	0.102	03:19
Breath Volume		2.261	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.096	03:22
Breath Volume		2.382	03:22
Air Blank		0.000	03:23
Control Test		0.081	03:23
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 03/30/2021 18:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON, RAYMOND, M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: COLON, RAYMOND
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:16

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 03/13/2021 06:49:46

Diagnostics Check		OK	06:41
Air Blank		0.000	06:42
Control Test		0.081	06:42
Air Blank		0.000	06:43
Subject Sample	# 1	0.179	06:44
Breath Volume		1.761	06:44
Air Blank		0.000	06:45
Air Blank		0.000	06:47
Subject Sample	# 2	0.165	06:48
Breath Volume		1.742	06:48
Air Blank		0.000	06:48
Control Test		0.080	06:49
Air Blank		0.000	06:49
Diagnostics Check		OK	06:49

Data Download Date/Time: 03/30/2021 18:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON, RAYMOND, M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: GERACI, CHRIS
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 03/13/2021 19:46:38

Diagnostics Check	OK	19:43	
Air Blank	0.000	19:43	
Control Test	0.082	19:44	
Air Blank	0.000	19:44	
Subject Sample	# 1	19:45	Subject Test Refused
Breath Volume	0.000	19:45	
Air Blank	0.000	19:45	
Control Test	0.081	19:45	
Air Blank	0.000	19:46	
Diagnostics Check	OK	19:46	

Data Download Date/Time: 03/30/2021 18:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON, RAYMOND, M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CORONILLA, DANIEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: FLORES SANCHEZ

DATE/TIME OF TEST: 03/13/2021 23:28:42

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.082	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.078	23:23
Breath Volume	2.199	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.072	23:27
Breath Volume	2.363	23:27
Air Blank	0.000	23:27
Control Test	0.082	23:28
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 03/30/2021 18:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COREY,GAVIN,T
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: SCELFO,ZACHARY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 03/15/2021 21:28:27

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:26	
Control Test	0.081	21:26	
Air Blank	0.000	21:26	
Subject Sample	# 1	21:27	Subject Test Refused
Breath Volume	0.000	21:27	
Air Blank	0.000	21:27	
Control Test	0.082	21:27	
Air Blank	0.000	21:28	
Diagnostics Check	OK	21:28	

Data Download Date/Time: 03/30/2021 18:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,EVEN,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: DREW,LIAM
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: SARMIENTO PRIETO

DATE/TIME OF TEST: 03/20/2021 19:37:44

Diagnostics Check	OK	19:30
Air Blank	0.000	19:30
Control Test	0.081	19:31
Air Blank	0.000	19:31
Subject Sample # 1	0.128	19:33
Breath Volume	1.546	19:33
Air Blank	0.000	19:33
Air Blank	0.000	19:35
Subject Sample # 2	0.122	19:36
Breath Volume	2.371	19:36
Air Blank	0.000	19:36
Control Test	0.081	19:37
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time: 03/30/2021 18:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIMMERLY,GREGORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ZIMMERLY,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: FIGUEROA LOPEZ

DATE/TIME OF TEST: 03/27/2021 03:22:49

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.083	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.165	03:18
Breath Volume	2.734	03:18
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.159	03:20
Breath Volume	3.566	03:20
Air Blank	0.000	03:21
Control Test	0.078	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 03/30/2021 18:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAMES,R
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: HESTER,JAMES
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: SCHAEFFER

DATE/TIME OF TEST: 03/28/2021 00:31:07

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:28	
Control Test	0.081	00:28	
Air Blank	0.000	00:29	
Subject Sample	# 1	00:29	Subject Test Refused
Breath Volume	0.000	00:29	
Air Blank	0.000	00:30	
Control Test	0.082	00:30	
Air Blank	0.000	00:31	
Diagnostics Check	OK	00:31	

Data Download Date/Time: 03/30/2021 18:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON, RAYMOND, M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: SCELFO, ZACHARY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: BILLIE

DATE/TIME OF TEST: 03/28/2021 03:24:13

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.081	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.263	03:19
Breath Volume	3.046	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.253	03:22
Breath Volume	2.761	03:22
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 03/30/2021 18:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: AGUILAR ARANGO

DATE/TIME OF TEST: 03/06/2021 04:59:33

Diagnostics Check	OK	04:51
Air Blank	0.000	04:52
Control Test	0.080	04:52
Air Blank	0.000	04:53
Subject Sample # 1	0.214	04:54
Breath Volume	2.593	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:56
Subject Sample # 2	0.216	04:57
Breath Volume	1.925	04:57
Air Blank	0.000	04:58
Control Test	0.076	04:58
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 03/25/2021 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: AGUILAR COLLAZO

DATE/TIME OF TEST: 03/07/2021 04:09:22

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.079	04:04
Breath Volume	1.535	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.085	04:07
Breath Volume	2.062	04:07
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 03/25/2021 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: MOONEY

DATE/TIME OF TEST: 03/08/2021 04:28:57

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.080	04:20
Air Blank		0.000	04:21
Subject Sample	# 1	0.214	04:23
Breath Volume		2.660	04:23
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.206	04:27
Breath Volume		2.675	04:27
Air Blank		0.000	04:28
Control Test		0.079	04:28
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 03/25/2021 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALLUCCA,MARIO,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WENTLAND,ALEX
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 03/10/2021 02:48:54

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.081	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.000	02:44
Breath Volume	2.046	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.000	02:47
Breath Volume	1.863	02:47
Air Blank	0.000	02:47
Control Test	0.077	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 03/25/2021 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,RYAN,
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: VAUGHN,RYAN
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 16:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 03/12/2021 18:29:14

Diagnostics Check	OK	18:23
Air Blank	0.000	18:23
Control Test	0.079	18:23
Air Blank	0.000	18:24
Subject Sample # 1	0.000	18:24
Breath Volume	1.835	18:24
Air Blank	0.000	18:25
Air Blank	0.000	18:27
Subject Sample # 2	0.000	18:27
Breath Volume	3.031	18:27
Air Blank	0.000	18:28
Control Test	0.079	18:28
Air Blank	0.000	18:29
Diagnostics Check	OK	18:29

Data Download Date/Time: 03/25/2021 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALLUCCA,MARIO,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FALLUCCA,MARIO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 03/13/2021 01:21:02

Diagnostics Check		OK	01:14
Air Blank		0.000	01:14
Control Test		0.079	01:15
Air Blank		0.000	01:15
Subject Sample	# 1	0.204	01:16
Breath Volume		2.660	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:18
Subject Sample	# 2	0.190	01:19
Breath Volume		2.589	01:19
Air Blank		0.000	01:20
Control Test		0.075	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:21

Data Download Date/Time: 03/25/2021 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: NAVARRO ARANDA

DATE/TIME OF TEST: 03/13/2021 03:36:03

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.186	03:27
Breath Volume	1.238	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.219	03:30
Breath Volume	2.921	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 3	0.223	03:34
Breath Volume	1.863	03:34
Air Blank	0.000	03:35
Control Test	0.078	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:36

Data Download Date/Time: 03/25/2021 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JASON,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BROWN,JASON
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 12:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: PLAZA

DATE/TIME OF TEST: 03/14/2021 01:56:55

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.217	01:51
Breath Volume	1.769	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.207	01:55
Breath Volume	1.210	01:55
Air Blank	0.000	01:55
Control Test	0.078	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 03/25/2021 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GWODZ,OFC
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/15/2021 02:49:17

Diagnostics Check		OK	02:37
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:39
Subject Sample	# 1	0.302	02:40
Breath Volume		1.589	02:40
Air Blank		0.000	02:41
Air Blank		0.000	02:42
Subject Sample	# 2	0.276	02:43
Breath Volume		1.300	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 3	0.290	02:47
Breath Volume		2.175	02:47
Air Blank		0.000	02:48
Control Test		0.075	02:48
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 03/25/2021 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JASON,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BROWN,JASON
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MORRELL

DATE/TIME OF TEST: 03/15/2021 03:10:39

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.077	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.135	03:05
Breath Volume	3.714	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.128	03:08
Breath Volume	2.597	03:08
Air Blank	0.000	03:09
Control Test	0.078	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 03/25/2021 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: DYCHE

DATE/TIME OF TEST: 03/20/2021 22:51:56

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.080	22:45
Air Blank	0.000	22:46
Subject Sample # 1	0.126	22:47
Breath Volume	1.832	22:47
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample # 2	0.125	22:50
Breath Volume	2.222	22:50
Air Blank	0.000	22:50
Control Test	0.079	22:51
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 03/25/2021 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,ROBERT,J
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GWODZ,OFC
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: ARTHUR

DATE/TIME OF TEST: 03/22/2021 23:23:24

Diagnostics Check	OK	23:19	
Air Blank	0.000	23:20	
Control Test	0.079	23:20	
Air Blank	0.000	23:20	
Subject Sample # 1	0.222	23:21	
Breath Volume	1.593	23:21	
Air Blank		23:21	Improper Sample
Air Blank	0.000	23:22	
Control Test	0.080	23:22	
Air Blank	0.000	23:23	
Diagnostics Check	OK	23:23	

Data Download Date/Time: 03/25/2021 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,ROBERT,J
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GWODZ,OFC
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: JEFFREY

DATE/TIME OF TEST: 03/22/2021 23:32:26

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.077	23:26
Air Blank	0.000	23:26
Subject Sample # 1	0.238	23:27
Breath Volume	1.781	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.229	23:30
Breath Volume	2.683	23:30
Air Blank	0.000	23:31
Control Test	0.076	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 03/25/2021 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTY, JOSHLYNN, C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: PROFFITT

DATE/TIME OF TEST: 03/02/2021 21:13:37

Diagnostics Check	OK	21:07
Air Blank	0.000	21:08
Control Test	0.080	21:08
Air Blank	0.000	21:08
Subject Sample # 1	0.000	21:09
Breath Volume	2.421	21:09
Air Blank	0.000	21:09
Air Blank	0.000	21:11
Subject Sample # 2	0.000	21:12
Breath Volume	2.148	21:12
Air Blank	0.000	21:12
Control Test	0.080	21:13
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

Data Download Date/Time: 04/05/2021 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTY, JOSHLYNN, C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: WOOLSEY

DATE/TIME OF TEST: 03/03/2021 03:27:12

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.080	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.163	03:22
Breath Volume		2.066	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.156	03:25
Breath Volume		1.226	03:25
Air Blank		0.000	03:26
Control Test		0.079	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 04/05/2021 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSSMAN,R,K
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: PIZUR

DATE/TIME OF TEST: 03/04/2021 02:05:55

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.080	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.205	02:00
Breath Volume		2.460	02:00
Air Blank		0.000	02:01
Air Blank		0.000	02:02
Subject Sample	# 2	0.197	02:04
Breath Volume		1.621	02:04
Air Blank		0.000	02:04
Control Test		0.081	02:05
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 04/05/2021 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,J,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 03/04/2021 19:33:07

Diagnostics Check	OK	19:22
Air Blank	0.000	19:22
Control Test	0.080	19:22
Air Blank	0.000	19:23
Subject Sample # 1	0.367	19:26
Breath Volume	2.414	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:28
Subject Sample # 2	0.363	19:31
Breath Volume	1.609	19:31
Air Blank	0.000	19:32
Control Test	0.079	19:32
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

Data Download Date/Time: 04/05/2021 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,J,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: COTTON

DATE/TIME OF TEST: 03/05/2021 01:12:18

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.080	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.098	01:07
Breath Volume	1.546	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.097	01:10
Breath Volume	1.746	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 04/05/2021 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTY, JOSHLYNN, C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RICHARDSON, N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 03/05/2021 21:28:16

Diagnostics Check	OK	21:20
Air Blank	0.000	21:21
Control Test	0.080	21:21
Air Blank	0.000	21:22
Subject Sample # 1	0.164	21:23
Breath Volume	2.988	21:23
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.156	21:26
Breath Volume	2.058	21:26
Air Blank	0.000	21:27
Control Test	0.079	21:27
Air Blank	0.000	21:28
Diagnostics Check	OK	21:28

Data Download Date/Time: 04/05/2021 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTY, JOSHLYNN, C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS, S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: LAWLEY

DATE/TIME OF TEST: 03/05/2021 22:08:39

Diagnostics Check	OK	22:01
Air Blank	0.000	22:02
Control Test	0.080	22:02
Air Blank	0.000	22:03
Subject Sample # 1	0.235	22:04
Breath Volume	1.664	22:04
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.249	22:07
Breath Volume	1.445	22:07
Air Blank	0.000	22:07
Control Test	0.079	22:08
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 04/05/2021 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTY, JOSHLYNN, C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: ZURVALEC

DATE/TIME OF TEST: 03/06/2021 02:16:56

Diagnostics Check		OK	02:03	
Air Blank		0.000	02:04	
Control Test		0.081	02:04	
Air Blank		0.000	02:05	
Subject Sample	# 1	0.198	02:06	
Breath Volume		2.027	02:06	
Air Blank		0.000	02:06	
Air Blank		0.000	02:08	
Subject Sample	# 2	0.179	02:11	Volume Not Met
Breath Volume		0.437	02:11	
Air Blank		0.000	02:12	
Air Blank		0.000	02:14	
Subject Sample	# 3	0.194	02:15	
Breath Volume		1.378	02:15	
Air Blank		0.000	02:15	
Control Test		0.080	02:16	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 04/05/2021 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTY,JOSHALYNN,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GRACKL,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: GILL

DATE/TIME OF TEST: 03/06/2021 04:00:58

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.080	03:53
Air Blank	0.000	03:54
Subject Sample # 1	0.147	03:55
Breath Volume	3.023	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.142	03:59
Breath Volume	4.824	03:59
Air Blank	0.000	04:00
Control Test	0.078	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 04/05/2021 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOD,CHARLES,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: DOTSETH

DATE/TIME OF TEST: 03/06/2021 23:22:33

Diagnostics Check	OK	23:13	
Air Blank	0.000	23:14	
Control Test	0.080	23:14	
Air Blank	0.000	23:15	
Subject Sample	# 1	23:18	No Sample Provided
Breath Volume	0.000	23:18	
Air Blank	0.000	23:18	
Air Blank	0.000	23:20	
Subject Sample	# 2	23:21	Subject Test Refused
Breath Volume	0.000	23:21	
Air Blank	0.000	23:21	
Control Test	0.080	23:21	
Air Blank	0.000	23:22	
Diagnostics Check	OK	23:22	

Data Download Date/Time: 04/05/2021 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTY, JOSHLYNN, C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CAMP, B
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 03/07/2021 01:07:32

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.081	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.245	01:02
Breath Volume	2.156	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.238	01:05
Breath Volume	2.726	01:05
Air Blank	0.000	01:06
Control Test	0.079	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 04/05/2021 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTY, JOSHLYNN, C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RICHARDSON, N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 03/16/2021 02:55:34

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:47	
Control Test	0.080	02:47	
Air Blank	0.000	02:48	
Subject Sample	# 1	02:51	No Sample Provided
Breath Volume	0.000	02:51	
Air Blank	0.000	02:51	
Air Blank	0.000	02:53	
Subject Sample	# 2	02:54	Subject Test Refused
Breath Volume	0.000	02:54	
Air Blank	0.000	02:54	
Control Test	0.080	02:54	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 04/05/2021 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAUMGARTNER,NATHAN
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 03/17/2021 02:50:26

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.130	02:45
Breath Volume	1.554	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.121	02:48
Breath Volume	1.519	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 04/05/2021 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTY, JOSHLYNN, C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAUMGARTNER, N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 03/20/2021 00:35:30

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.026	00:29
Breath Volume	1.343	00:29
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.026	00:33
Breath Volume	1.582	00:33
Air Blank	0.000	00:34
Control Test	0.080	00:34
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 04/05/2021 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARCHER,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAUMGARTNER,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: GLASS

DATE/TIME OF TEST: 03/20/2021 16:47:24

Diagnostics Check	OK	16:40
Air Blank	0.000	16:41
Control Test	0.080	16:41
Air Blank	0.000	16:42
Subject Sample # 1	0.000	16:42
Breath Volume	2.007	16:42
Air Blank	0.000	16:43
Air Blank	0.000	16:45
Subject Sample # 2	0.000	16:46
Breath Volume	3.554	16:46
Air Blank	0.000	16:46
Control Test	0.080	16:46
Air Blank	0.000	16:47
Diagnostics Check	OK	16:47

Data Download Date/Time: 04/05/2021 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTY, JOSHALYNN, C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: HENRIKSON

DATE/TIME OF TEST: 03/20/2021 23:16:41

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.081	23:11
Air Blank	0.000	23:11
Subject Sample # 1	0.089	23:12
Breath Volume	2.050	23:12
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.083	23:15
Breath Volume	1.488	23:15
Air Blank	0.000	23:15
Control Test	0.081	23:16
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 04/05/2021 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: CUYUCH LOPEZ

DATE/TIME OF TEST: 03/01/2021 02:00:16

Diagnostics Check		OK	01:54
Air Blank		0.000	01:54
Control Test		0.078	01:54
Air Blank		0.000	01:55
Subject Sample	# 1	0.239	01:55
Breath Volume		2.437	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.241	01:58
Breath Volume		2.695	01:58
Air Blank		0.000	01:59
Control Test		0.078	01:59
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

Data Download Date/Time: 03/12/2021 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROBERTS,ROGER
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: HOUX

DATE/TIME OF TEST: 03/01/2021 18:03:19

Diagnostics Check	OK	17:52
Air Blank	0.000	17:53
Control Test	0.078	17:53
Air Blank	0.000	17:54
Subject Sample # 1	0.293	17:55
Breath Volume	2.218	17:55
Air Blank	0.000	17:56
Air Blank	0.000	17:57
Subject Sample # 2	0.260	17:58
Breath Volume	1.808	17:58
Air Blank	0.000	17:59
Air Blank	0.000	18:01
Subject Sample # 3	0.282	18:01
Breath Volume	3.691	18:01
Air Blank	0.000	18:02
Control Test	0.076	18:02
Air Blank	0.000	18:03
Diagnostics Check	OK	18:03

Data Download Date/Time: 03/12/2021 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STILES, MEGAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: MANGUIA

DATE/TIME OF TEST: 03/02/2021 04:28:55

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.078	04:23
Air Blank	0.000	04:23
Subject Sample # 1	0.000	04:24
Breath Volume	1.191	04:24
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.000	04:27
Breath Volume	1.578	04:27
Air Blank	0.000	04:28
Control Test	0.078	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 03/12/2021 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TRENTMAN,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: PELTON

DATE/TIME OF TEST: 03/02/2021 15:59:39

Diagnostics Check	OK	15:52
Air Blank	0.000	15:52
Control Test	0.078	15:53
Air Blank	0.000	15:53
Subject Sample # 1	0.352	15:54
Breath Volume	1.312	15:54
Air Blank	0.000	15:55
Air Blank	0.000	15:56
Subject Sample # 2	0.364	15:58
Breath Volume	1.882	15:58
Air Blank	0.000	15:58
Control Test	0.077	15:59
Air Blank	0.000	15:59
Diagnostics Check	OK	15:59

Data Download Date/Time: 03/12/2021 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: A'ANDREA,ADAM
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:04

SUBJECT'S LAST NAME: BERRIS

DATE/TIME OF TEST: 03/02/2021 16:44:03

Diagnostics Check	OK	16:36
Air Blank	0.000	16:36
Control Test	0.077	16:36
Air Blank	0.000	16:37
Subject Sample # 1	0.000	16:38
Breath Volume	1.320	16:38
Air Blank	0.000	16:38
Air Blank	0.000	16:40
Subject Sample # 2	0.000	16:42
Breath Volume	1.257	16:42
Air Blank	0.000	16:43
Control Test	0.077	16:43
Air Blank	0.000	16:43
Diagnostics Check	OK	16:44

Data Download Date/Time: 03/12/2021 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 03/02/2021 23:17:57

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.080	23:12
Air Blank	0.000	23:13
Subject Sample # 1	0.000	23:13
Breath Volume	1.394	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.000	23:16
Breath Volume	1.781	23:16
Air Blank	0.000	23:17
Control Test	0.078	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 03/12/2021 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SMITH, MARK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: RECK II

DATE/TIME OF TEST: 03/03/2021 00:12:47

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.078	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.045	00:07
Breath Volume	1.921	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.046	00:11
Breath Volume	2.285	00:11
Air Blank	0.000	00:11
Control Test	0.077	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 03/12/2021 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHARELUS,OBED
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 03/03/2021 05:43:42

Diagnostics Check	OK	05:36
Air Blank	0.000	05:37
Control Test	0.079	05:37
Air Blank	0.000	05:37
Subject Sample # 1	0.153	05:39
Breath Volume	1.578	05:39
Air Blank	0.000	05:39
Air Blank	0.000	05:41
Subject Sample # 2	0.156	05:42
Breath Volume	1.343	05:42
Air Blank	0.000	05:42
Control Test	0.078	05:43
Air Blank	0.000	05:43
Diagnostics Check	OK	05:43

Data Download Date/Time: 03/12/2021 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: USON

DATE/TIME OF TEST: 03/06/2021 00:08:49

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.078	00:03
Air Blank	0.000	00:03
Subject Sample # 1	0.049	00:04
Breath Volume	1.136	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.049	00:07
Breath Volume	1.281	00:07
Air Blank	0.000	00:07
Control Test	0.078	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 03/12/2021 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: GODINEZ PEREZ

DATE/TIME OF TEST: 03/06/2021 01:35:28

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.077	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.130	01:31
Breath Volume	1.539	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.127	01:33
Breath Volume	2.296	01:33
Air Blank	0.000	01:34
Control Test	0.077	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 03/12/2021 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: IRAHETA,RAFAEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: RODLITZ

DATE/TIME OF TEST: 03/07/2021 08:43:19

Diagnostics Check	OK	08:37
Air Blank	0.000	08:37
Control Test	0.078	08:37
Air Blank	0.000	08:38
Subject Sample # 1	0.000	08:38
Breath Volume	1.917	08:38
Air Blank	0.000	08:39
Air Blank	0.000	08:41
Subject Sample # 2	0.000	08:41
Breath Volume	2.597	08:41
Air Blank	0.000	08:42
Control Test	0.079	08:42
Air Blank	0.000	08:43
Diagnostics Check	OK	08:43

Data Download Date/Time: 03/12/2021 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: DEZARD

DATE/TIME OF TEST: 03/07/2021 20:52:53

Diagnostics Check	OK	20:44
Air Blank	0.000	20:45
Control Test	0.078	20:45
Air Blank	0.000	20:45
Subject Sample # 1	0.123	20:48
Breath Volume	1.785	20:48
Air Blank	0.000	20:49
Air Blank	0.000	20:50
Subject Sample # 2	0.126	20:51
Breath Volume	1.546	20:51
Air Blank	0.000	20:51
Control Test	0.077	20:52
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 03/12/2021 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MOREA,DEAN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: DAYAN

DATE/TIME OF TEST: 03/07/2021 21:20:04

Diagnostics Check	OK	21:13
Air Blank	0.000	21:13
Control Test	0.079	21:13
Air Blank	0.000	21:14
Subject Sample # 1	0.177	21:15
Breath Volume	2.300	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:17
Subject Sample # 2	0.182	21:18
Breath Volume	2.710	21:18
Air Blank	0.000	21:19
Control Test	0.079	21:19
Air Blank	0.000	21:19
Diagnostics Check	OK	21:20

Data Download Date/Time: 03/12/2021 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROSENBER,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: PAZ RIVERA

DATE/TIME OF TEST: 03/08/2021 10:34:44

Diagnostics Check		OK	10:20	
Air Blank		0.000	10:20	
Control Test		0.079	10:21	
Air Blank		0.000	10:21	
Subject Sample	# 1	0.077	10:24	Volume Not Met
Breath Volume		0.207	10:24	
Air Blank		0.000	10:25	
Air Blank		0.000	10:27	
Subject Sample	# 2	0.092	10:29	
Breath Volume		1.246	10:29	
Air Blank		0.000	10:29	
Air Blank		0.000	10:31	
Subject Sample	# 3	0.093	10:33	
Breath Volume		1.609	10:33	
Air Blank		0.000	10:33	
Control Test		0.077	10:34	
Air Blank		0.000	10:34	
Diagnostics Check		OK	10:34	

Data Download Date/Time: 03/12/2021 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LASSO,JAVIER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: RUIZ MENDEZ

DATE/TIME OF TEST: 03/08/2021 11:18:02

Diagnostics Check		OK	11:08
Air Blank		0.000	11:08
Control Test		0.079	11:09
Air Blank		0.000	11:09
Subject Sample	# 1	0.202	11:10
Breath Volume		1.429	11:10
Air Blank		0.000	11:10
Air Blank		0.000	11:12
Subject Sample	# 2	0.175	11:13
Breath Volume		1.425	11:13
Air Blank		0.000	11:13
Air Blank		0.000	11:15
Subject Sample	# 3	0.183	11:16
Breath Volume		2.816	11:16
Air Blank		0.000	11:17
Control Test		0.078	11:17
Air Blank		0.000	11:17
Diagnostics Check		OK	11:18

Data Download Date/Time: 03/12/2021 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STILES,MEGAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: MANNERY

DATE/TIME OF TEST: 03/08/2021 22:04:27

Diagnostics Check	OK	21:57
Air Blank	0.000	21:58
Control Test	0.080	21:58
Air Blank	0.000	21:58
Subject Sample # 1	0.186	21:59
Breath Volume	2.195	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.184	22:02
Breath Volume	2.468	22:02
Air Blank	0.000	22:03
Control Test	0.078	22:03
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 03/12/2021 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SMITH,MARK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: OSBOURNE

DATE/TIME OF TEST: 03/08/2021 23:37:27

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.079	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.268	23:32
Breath Volume	1.632	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.267	23:35
Breath Volume	2.078	23:35
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 03/12/2021 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREG
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: PEREZ-VICENTE

DATE/TIME OF TEST: 03/09/2021 20:00:42

Diagnostics Check	OK	19:53
Air Blank	0.000	19:53
Control Test	0.079	19:54
Air Blank	0.000	19:54
Subject Sample # 1	0.223	19:56
Breath Volume	2.085	19:56
Air Blank	0.000	19:56
Air Blank	0.000	19:58
Subject Sample # 2	0.222	19:59
Breath Volume	2.187	19:59
Air Blank	0.000	19:59
Control Test	0.077	20:00
Air Blank	0.000	20:00
Diagnostics Check	OK	20:00

Data Download Date/Time: 03/12/2021 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PEREZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: North Palm Beach
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: YUNCK

DATE/TIME OF TEST: 03/09/2021 22:42:59

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.079	22:37
Air Blank	0.000	22:37
Subject Sample # 1	0.147	22:38
Breath Volume	1.843	22:38
Air Blank	0.000	22:39
Air Blank	0.000	22:40
Subject Sample # 2	0.150	22:41
Breath Volume	1.671	22:41
Air Blank	0.000	22:42
Control Test	0.079	22:42
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 03/12/2021 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCGILLICUDDY,STEVEN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: LIANA

DATE/TIME OF TEST: 03/10/2021 00:31:58

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.078	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.207	00:27
Breath Volume	1.652	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.210	00:30
Breath Volume	2.171	00:30
Air Blank	0.000	00:31
Control Test	0.078	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 03/12/2021 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HOWARD,EDWARD
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: VANDER WAL

DATE/TIME OF TEST: 03/10/2021 22:29:41

Diagnostics Check	OK	22:22
Air Blank	0.000	22:23
Control Test	0.078	22:23
Air Blank	0.000	22:24
Subject Sample # 1	0.133	22:24
Breath Volume	1.898	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.147	22:28
Breath Volume	2.019	22:28
Air Blank	0.000	22:28
Control Test	0.077	22:29
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 03/12/2021 13:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALKER,KIANA
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: BARBAGALLO

DATE/TIME OF TEST: 03/11/2021 23:30:47

Diagnostics Check	OK	23:23
Air Blank	0.000	23:24
Control Test	0.079	23:24
Air Blank	0.000	23:25
Subject Sample # 1	0.021	23:25
Breath Volume	1.406	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.020	23:29
Breath Volume	1.964	23:29
Air Blank	0.000	23:29
Control Test	0.079	23:30
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 03/12/2021 13:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: BISSOON

DATE/TIME OF TEST: 03/12/2021 03:06:25

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.078	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.257	03:01
Breath Volume		2.953	03:01
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.238	03:04
Breath Volume		3.101	03:04
Air Blank		0.000	03:05
Control Test		0.078	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 03/12/2021 13:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH, GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: NEVILLE

DATE/TIME OF TEST: 03/13/2021 01:10:03

Diagnostics Check	OK	01:03
Air Blank	0.000	01:04
Control Test	0.079	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.093	01:05
Breath Volume	1.933	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.098	01:08
Breath Volume	2.640	01:08
Air Blank	0.000	01:09
Control Test	0.078	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:10

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR, NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: GRANADENO

DATE/TIME OF TEST: 03/13/2021 02:41:28

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.177	02:36
Breath Volume	2.527	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.178	02:39
Breath Volume	2.546	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NALERIO, LUIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: LIBERTUCCI

DATE/TIME OF TEST: 03/14/2021 00:52:59

Diagnostics Check		OK	00:46
Air Blank		0.000	00:47
Control Test		0.079	00:47
Air Blank		0.000	00:47
Subject Sample	# 1	0.096	00:48
Breath Volume		1.769	00:48
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.093	00:51
Breath Volume		1.218	00:51
Air Blank		0.000	00:52
Control Test		0.079	00:52
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCCLUSKEY, RYAN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: PEREZ GUERRERO

DATE/TIME OF TEST: 03/14/2021 08:16:33

Diagnostics Check	OK	08:10
Air Blank	0.000	08:10
Control Test	0.079	08:11
Air Blank	0.000	08:11
Subject Sample # 1	0.103	08:12
Breath Volume	3.261	08:12
Air Blank	0.000	08:12
Air Blank	0.000	08:14
Subject Sample # 2	0.102	08:15
Breath Volume	3.683	08:15
Air Blank	0.000	08:15
Control Test	0.078	08:15
Air Blank	0.000	08:16
Diagnostics Check	OK	08:16

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STOKER,RONALD
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: DUDLEY

DATE/TIME OF TEST: 03/14/2021 21:17:34

Diagnostics Check	OK	21:10
Air Blank	0.000	21:11
Control Test	0.079	21:11
Air Blank	0.000	21:12
Subject Sample # 1	0.065	21:13
Breath Volume	4.222	21:13
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.065	21:16
Breath Volume	4.105	21:16
Air Blank	0.000	21:16
Control Test	0.079	21:16
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SMITH, MARK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: XIVIR

DATE/TIME OF TEST: 03/15/2021 00:02:39

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.079	23:55
Air Blank	0.000	23:56
Subject Sample # 2	0.160	23:58
Breath Volume	1.441	23:58
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.154	00:01
Breath Volume	1.480	00:01
Air Blank	0.000	00:01
Control Test	0.078	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANDREY, CHRISTOPHER
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: GUIDO

DATE/TIME OF TEST: 03/15/2021 00:47:03

Diagnostics Check		OK	00:32	
Air Blank		0.000	00:33	
Control Test		0.079	00:33	
Air Blank		0.000	00:33	
Subject Sample	# 1	0.191	00:37	Volume Not Met
Breath Volume		0.964	00:37	
Air Blank		0.000	00:37	
Air Blank		0.000	00:39	
Subject Sample	# 2	0.196	00:41	
Breath Volume		1.386	00:41	
Air Blank		0.000	00:42	
Air Blank		0.000	00:44	
Subject Sample	# 3	0.203	00:45	
Breath Volume		1.285	00:45	
Air Blank		0.000	00:46	
Control Test		0.076	00:46	
Air Blank		0.000	00:46	
Diagnostics Check		OK	00:47	

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: IV,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 03/15/2021 19:49:39

Diagnostics Check	OK	19:43
Air Blank	0.000	19:43
Control Test	0.079	19:43
Air Blank	0.000	19:44
Subject Sample # 1	0.000	19:45
Breath Volume	4.765	19:45
Air Blank	0.000	19:45
Air Blank	0.000	19:47
Subject Sample # 2	0.000	19:48
Breath Volume	4.785	19:48
Air Blank	0.000	19:48
Control Test	0.079	19:49
Air Blank	0.000	19:49
Diagnostics Check	OK	19:49

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SOLOWAY,ALLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/17/2021 00:23:14

Diagnostics Check		OK	00:16
Air Blank		0.000	00:16
Control Test		0.079	00:17
Air Blank		0.000	00:17
Subject Sample	# 1	0.104	00:18
Breath Volume		1.777	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.103	00:21
Breath Volume		1.503	00:21
Air Blank		0.000	00:22
Control Test		0.077	00:22
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: THOMAS,MICAH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/17/2021 01:02:59

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.079	00:56
Air Blank	0.000	00:57
Subject Sample # 1	0.169	00:58
Breath Volume	2.195	00:58
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.150	01:01
Breath Volume	1.402	01:01
Air Blank	0.000	01:02
Control Test	0.078	01:02
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR,NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: ENGLERT

DATE/TIME OF TEST: 03/17/2021 22:31:08

Diagnostics Check	OK	22:24
Air Blank	0.000	22:25
Control Test	0.080	22:25
Air Blank	0.000	22:25
Subject Sample # 1	0.130	22:26
Breath Volume	1.359	22:26
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.131	22:29
Breath Volume	1.464	22:29
Air Blank	0.000	22:30
Control Test	0.079	22:30
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 03/17/2021 23:18:36

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.079	23:12
Air Blank	0.000	23:12
Subject Sample # 1	0.065	23:14
Breath Volume	1.617	23:14
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.063	23:17
Breath Volume	1.324	23:17
Air Blank	0.000	23:17
Control Test	0.079	23:17
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MARTINEZ,JACOB
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: EASON

DATE/TIME OF TEST: 03/18/2021 00:07:51

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:05	
Control Test	0.079	00:05	
Air Blank	0.000	00:05	
Subject Sample	# 1	00:06	Subject Test Refused
Breath Volume	0.000	00:06	
Air Blank	0.000	00:06	
Control Test	0.079	00:07	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:07	

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALKER,JODI
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: SILIEN

DATE/TIME OF TEST: 03/18/2021 01:24:20

Diagnostics Check	OK	01:17
Air Blank	0.000	01:18
Control Test	0.078	01:18
Air Blank	0.000	01:18
Subject Sample # 1	0.198	01:19
Breath Volume	1.183	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.193	01:22
Breath Volume	1.929	01:22
Air Blank	0.000	01:23
Control Test	0.078	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NALERIO,LUIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: RAUX

DATE/TIME OF TEST: 03/19/2021 06:03:15

Diagnostics Check	OK	05:57
Air Blank	0.000	05:57
Control Test	0.079	05:58
Air Blank	0.000	05:58
Subject Sample # 1	0.000	05:58
Breath Volume	1.855	05:58
Air Blank	0.000	05:59
Air Blank	0.000	06:01
Subject Sample # 2	0.000	06:01
Breath Volume	2.535	06:01
Air Blank	0.000	06:02
Control Test	0.079	06:02
Air Blank	0.000	06:03
Diagnostics Check	OK	06:03

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VARGAS, WILBERT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: HARADA

DATE/TIME OF TEST: 03/19/2021 23:05:29

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.079	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.212	23:00
Breath Volume	2.015	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.211	23:03
Breath Volume	1.292	23:03
Air Blank	0.000	23:04
Control Test	0.079	23:04
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR, NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 03/20/2021 03:14:56

Diagnostics Check		OK	03:02	
Air Blank		0.000	03:03	
Control Test		0.079	03:03	
Air Blank		0.000	03:04	
Subject Sample	# 1	0.173	03:07	Volume Not Met
Breath Volume		0.507	03:07	
Air Blank		0.000	03:07	
Air Blank		0.000	03:09	
Subject Sample	# 2	0.199	03:10	
Breath Volume		1.765	03:10	
Air Blank		0.000	03:10	
Air Blank		0.000	03:12	
Subject Sample	# 3	0.197	03:13	
Breath Volume		1.476	03:13	
Air Blank		0.000	03:14	
Control Test		0.078	03:14	
Air Blank		0.000	03:14	
Diagnostics Check		OK	03:14	

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHARELUS, OBED
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: HARTIGAN

DATE/TIME OF TEST: 03/20/2021 23:11:04

Diagnostics Check	OK	23:04
Air Blank	0.000	23:05
Control Test	0.080	23:05
Air Blank	0.000	23:06
Subject Sample # 1	0.000	23:06
Breath Volume	2.800	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.000	23:09
Breath Volume	2.414	23:09
Air Blank	0.000	23:10
Control Test	0.079	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:11

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: RINGELSTEIN

DATE/TIME OF TEST: 03/21/2021 02:16:21

Diagnostics Check		OK	02:09
Air Blank		0.000	02:10
Control Test		0.079	02:10
Air Blank		0.000	02:11
Subject Sample	# 1	0.256	02:11
Breath Volume		2.132	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.249	02:14
Breath Volume		2.050	02:14
Air Blank		0.000	02:15
Control Test		0.079	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: IVORY, BRELAN
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: VAIL VELASQUEZ

DATE/TIME OF TEST: 03/21/2021 03:20:26

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.189	03:16
Breath Volume	1.851	03:16
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.188	03:18
Breath Volume	1.195	03:18
Air Blank	0.000	03:19
Control Test	0.078	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: IV,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: RAMIREZ FELIPE

DATE/TIME OF TEST: 03/21/2021 20:25:56

Diagnostics Check	OK	20:18
Air Blank	0.000	20:19
Control Test	0.080	20:19
Air Blank	0.000	20:20
Subject Sample # 1	0.163	20:21
Breath Volume	2.015	20:21
Air Blank	0.000	20:22
Air Blank	0.000	20:23
Subject Sample # 2	0.169	20:24
Breath Volume	2.164	20:24
Air Blank	0.000	20:25
Control Test	0.078	20:25
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: ROJAS AGUILAR

DATE/TIME OF TEST: 03/21/2021 22:22:30

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.079	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.127	22:18
Breath Volume	1.496	22:18
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.130	22:20
Breath Volume	1.570	22:20
Air Blank	0.000	22:21
Control Test	0.079	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCGILLICUDDY, STEVEN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MERCER

DATE/TIME OF TEST: 03/22/2021 03:06:43

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.079	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.098	03:02
Breath Volume		2.605	03:02
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.093	03:05
Breath Volume		2.113	03:05
Air Blank		0.000	03:05
Control Test		0.078	03:06
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ESPOSITO,DAVID
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: DEBLASIS

DATE/TIME OF TEST: 03/22/2021 12:11:05

Diagnostics Check		OK	12:04
Air Blank		0.000	12:05
Control Test		0.080	12:05
Air Blank		0.000	12:06
Subject Sample	# 1	0.000	12:06
Breath Volume		1.312	12:06
Air Blank		0.000	12:07
Air Blank		0.000	12:09
Subject Sample	# 2	0.000	12:09
Breath Volume		1.894	12:09
Air Blank		0.000	12:10
Control Test		0.079	12:10
Air Blank		0.000	12:10
Diagnostics Check		OK	12:11

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STREIGOLD,DAVID
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: MARTIRENA

DATE/TIME OF TEST: 03/23/2021 13:40:38

Diagnostics Check	OK	13:33
Air Blank	0.000	13:34
Control Test	0.079	13:34
Air Blank	0.000	13:35
Subject Sample # 1	0.248	13:36
Breath Volume	2.144	13:36
Air Blank	0.000	13:36
Air Blank	0.000	13:38
Subject Sample # 2	0.238	13:39
Breath Volume	1.789	13:39
Air Blank	0.000	13:39
Control Test	0.079	13:40
Air Blank	0.000	13:40
Diagnostics Check	OK	13:40

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STREIGOLD,DAVID
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 03/23/2021 18:02:22

Diagnostics Check	OK	17:50	
Air Blank	0.000	17:50	
Control Test	0.078	17:50	
Air Blank	0.000	17:51	
Subject Sample # 1	0.423	17:52	
Breath Volume	3.066	17:52	
Air Blank	0.000	17:53	
Air Blank	0.000	17:55	
Subject Sample # 2		17:58	No Sample Provided
Breath Volume	0.000	17:58	
Air Blank	0.000	17:58	
Air Blank	0.000	18:00	
Subject Sample # 3		18:01	Subject Test Refused
Breath Volume	0.000	18:01	
Air Blank	0.000	18:01	
Control Test	0.077	18:01	
Air Blank	0.000	18:02	
Diagnostics Check	OK	18:02	

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RODRIGUEZ,ELIAN
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: MEDCRAFT

DATE/TIME OF TEST: 03/23/2021 20:30:58

Diagnostics Check	OK	20:27	
Air Blank	0.000	20:28	
Control Test	0.079	20:28	
Air Blank	0.000	20:28	
Subject Sample	# 1	20:29	Subject Test Refused
Breath Volume	0.000	20:29	
Air Blank	0.000	20:30	
Control Test	0.079	20:30	
Air Blank	0.000	20:30	
Diagnostics Check	OK	20:30	

Data Download Date/Time: 04/09/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SOLOWAY,ALAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: DICKERSON

DATE/TIME OF TEST: 03/23/2021 23:48:24

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.079	23:42
Air Blank	0.000	23:43
Subject Sample # 1	0.197	23:43
Breath Volume	3.312	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.198	23:46
Breath Volume	3.054	23:46
Air Blank	0.000	23:47
Control Test	0.079	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 04/09/2021 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NALERIO,LUIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 03/24/2021 01:11:16

Diagnostics Check	OK	01:04
Air Blank	0.000	01:05
Control Test	0.078	01:05
Air Blank	0.000	01:06
Subject Sample # 1	0.203	01:06
Breath Volume	1.105	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.205	01:09
Breath Volume	2.253	01:09
Air Blank	0.000	01:10
Control Test	0.078	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 04/09/2021 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MITCHELL,COLE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: ST AMAND

DATE/TIME OF TEST: 03/24/2021 08:36:07

Diagnostics Check		OK	08:29
Air Blank		0.000	08:30
Control Test		0.078	08:30
Air Blank		0.000	08:31
Subject Sample	# 1	0.156	08:31
Breath Volume		2.699	08:31
Air Blank		0.000	08:32
Air Blank		0.000	08:34
Subject Sample	# 2	0.158	08:34
Breath Volume		2.628	08:34
Air Blank		0.000	08:35
Control Test		0.078	08:35
Air Blank		0.000	08:36
Diagnostics Check		OK	08:36

Data Download Date/Time: 04/09/2021 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: MONROY

DATE/TIME OF TEST: 03/24/2021 20:53:34

Diagnostics Check	OK	20:46
Air Blank	0.000	20:47
Control Test	0.079	20:47
Air Blank	0.000	20:47
Subject Sample # 1	0.158	20:49
Breath Volume	1.925	20:49
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.153	20:52
Breath Volume	2.375	20:52
Air Blank	0.000	20:52
Control Test	0.079	20:52
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

Data Download Date/Time: 04/09/2021 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: PEREZ SONTAY

DATE/TIME OF TEST: 03/25/2021 00:40:27

Diagnostics Check	OK	00:33
Air Blank	0.000	00:34
Control Test	0.078	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.239	00:35
Breath Volume	1.562	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.226	00:38
Breath Volume	2.164	00:38
Air Blank	0.000	00:39
Control Test	0.078	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 04/09/2021 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 03/25/2021 01:54:33

Diagnostics Check		OK	01:48
Air Blank		0.000	01:48
Control Test		0.078	01:49
Air Blank		0.000	01:49
Subject Sample	# 1	0.165	01:50
Breath Volume		1.417	01:50
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.170	01:52
Breath Volume		1.992	01:52
Air Blank		0.000	01:53
Control Test		0.079	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 04/09/2021 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR,NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: STEEN

DATE/TIME OF TEST: 03/25/2021 03:56:55

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.000	03:52
Breath Volume	2.906	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.000	03:55
Breath Volume	2.976	03:55
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 04/09/2021 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: LEBEJKO

DATE/TIME OF TEST: 03/25/2021 04:15:13

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.079	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.000	04:10
Breath Volume	1.597	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.000	04:13
Breath Volume	2.085	04:13
Air Blank	0.000	04:14
Control Test	0.078	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 04/09/2021 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STILES, MEGAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: GRZECH

DATE/TIME OF TEST: 03/27/2021 01:10:19

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.081	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.126	01:05
Breath Volume	2.511	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.119	01:08
Breath Volume	1.386	01:08
Air Blank	0.000	01:09
Control Test	0.079	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 04/09/2021 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MARCH, THOMAS
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:47

SUBJECT'S LAST NAME: GALLAGHER

DATE/TIME OF TEST: 03/27/2021 07:18:42

Diagnostics Check	OK	07:12
Air Blank	0.000	07:12
Control Test	0.080	07:13
Air Blank	0.000	07:13
Subject Sample # 1	0.152	07:14
Breath Volume	1.593	07:14
Air Blank	0.000	07:14
Air Blank	0.000	07:16
Subject Sample # 2	0.148	07:17
Breath Volume	2.285	07:17
Air Blank	0.000	07:17
Control Test	0.080	07:18
Air Blank	0.000	07:18
Diagnostics Check	OK	07:18

Data Download Date/Time: 04/09/2021 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PRUDHOMME, DANIEL
ARRESTING OFFICER AGENCY: North Palm Beach
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: AUXIER

DATE/TIME OF TEST: 03/27/2021 23:40:04

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:37	
Control Test	0.080	23:38	
Air Blank	0.000	23:38	
Subject Sample	# 1	23:38	Subject Test Refused
Breath Volume	0.000	23:38	
Air Blank	0.000	23:39	
Control Test	0.080	23:39	
Air Blank	0.000	23:39	
Diagnostics Check	OK	23:40	

Data Download Date/Time: 04/09/2021 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGUIRRE, PETER
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: GRANDOIT

DATE/TIME OF TEST: 03/28/2021 04:57:51

Diagnostics Check	OK	04:51
Air Blank	0.000	04:52
Control Test	0.080	04:52
Air Blank	0.000	04:52
Subject Sample # 1	0.158	04:53
Breath Volume	1.343	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:55
Subject Sample # 2	0.148	04:56
Breath Volume	1.875	04:56
Air Blank	0.000	04:56
Control Test	0.080	04:57
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 04/09/2021 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SAYIH, ROBERT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:27

SUBJECT'S LAST NAME: POPPELL

DATE/TIME OF TEST: 03/28/2021 08:57:59

Diagnostics Check		OK	08:51
Air Blank		0.000	08:52
Control Test		0.080	08:52
Air Blank		0.000	08:52
Subject Sample	# 1	0.183	08:53
Breath Volume		2.593	08:53
Air Blank		0.000	08:54
Air Blank		0.000	08:55
Subject Sample	# 2	0.185	08:56
Breath Volume		4.054	08:56
Air Blank		0.000	08:57
Control Test		0.080	08:57
Air Blank		0.000	08:57
Diagnostics Check		OK	08:57

Data Download Date/Time: 04/09/2021 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEHENY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: AUGUSTO

DATE/TIME OF TEST: 03/29/2021 02:17:30

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.144	02:12
Breath Volume	2.304	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.154	02:15
Breath Volume	2.617	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 04/09/2021 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH, GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: PEREZ GUTIERREZ

DATE/TIME OF TEST: 03/29/2021 03:39:57

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.081	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.204	03:35
Breath Volume		2.789	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:37
Subject Sample	# 2	0.200	03:38
Breath Volume		2.949	03:38
Air Blank		0.000	03:39
Control Test		0.080	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 04/09/2021 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: VIRVETE

DATE/TIME OF TEST: 03/29/2021 05:29:21

Diagnostics Check	OK	05:21
Air Blank	0.000	05:22
Control Test	0.081	05:22
Air Blank	0.000	05:23
Subject Sample # 1	0.174	05:23
Breath Volume	1.429	05:23
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.179	05:27
Breath Volume	2.214	05:27
Air Blank	0.000	05:28
Control Test	0.081	05:28
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 04/09/2021 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: GERIBON

DATE/TIME OF TEST: 03/29/2021 07:01:48

Diagnostics Check	OK	06:55
Air Blank	0.000	06:55
Control Test	0.080	06:55
Air Blank	0.000	06:56
Subject Sample # 1	0.136	06:57
Breath Volume	3.363	06:57
Air Blank	0.000	06:58
Air Blank	0.000	06:59
Subject Sample # 2	0.138	07:00
Breath Volume	3.039	07:00
Air Blank	0.000	07:00
Control Test	0.080	07:01
Air Blank	0.000	07:01
Diagnostics Check	OK	07:01

Data Download Date/Time: 04/09/2021 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALLEN,RYAN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: GODINEZ

DATE/TIME OF TEST: 03/31/2021 01:47:29

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.080	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.189	01:43
Breath Volume	1.812	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.200	01:45
Breath Volume	2.480	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 04/09/2021 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HENDERSON,MARVIN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: WALTHOUR

DATE/TIME OF TEST: 03/31/2021 09:36:46

Diagnostics Check	OK	09:30
Air Blank	0.000	09:30
Control Test	0.080	09:31
Air Blank	0.000	09:31
Subject Sample # 1	0.085	09:32
Breath Volume	1.113	09:32
Air Blank	0.000	09:32
Air Blank	0.000	09:34
Subject Sample # 2	0.085	09:35
Breath Volume	1.835	09:35
Air Blank	0.000	09:35
Control Test	0.080	09:36
Air Blank	0.000	09:36
Diagnostics Check	OK	09:36

Data Download Date/Time: 04/09/2021 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODENZI,CAMERON,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: GODENZI,CAMERON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 02/08/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 15:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:02

SUBJECT'S LAST NAME: LELINSKI

DATE/TIME OF TEST: 03/09/2021 17:30:36

Diagnostics Check	OK	17:27	
Air Blank	0.000	17:27	
Control Test	0.079	17:27	
Air Blank	0.000	17:28	
Subject Sample	# 1	17:29	Interferent Detect
Breath Volume	2.089	17:29	
Air Blank	0.000	17:29	
Control Test	0.079	17:29	
Air Blank	0.000	17:30	
Diagnostics Check	OK	17:30	

Data Download Date/Time: 03/29/2021 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISSONNETTE,JACOB,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MAXWELL,TIMOTHY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 02/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: WIDMAIER

DATE/TIME OF TEST: 03/13/2021 23:56:47

Diagnostics Check		OK	23:44	
Air Blank		0.000	23:44	
Control Test		0.080	23:45	
Air Blank		0.000	23:45	
Subject Sample	# 1	0.140	23:48	Volume Not Met
Breath Volume		0.132	23:48	
Air Blank		0.000	23:49	
Air Blank		0.000	23:51	
Subject Sample	# 2	0.147	23:51	
Breath Volume		1.562	23:51	
Air Blank		0.000	23:52	
Air Blank		0.000	23:54	
Subject Sample	# 3	0.143	23:55	
Breath Volume		1.558	23:55	
Air Blank		0.000	23:55	
Control Test		0.078	23:56	
Air Blank		0.000	23:56	
Diagnostics Check		OK	23:56	

Data Download Date/Time: 03/29/2021 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CARABALLO

DATE/TIME OF TEST: 03/01/2021 02:17:48

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.119	02:13
Breath Volume	1.496	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:15
Subject Sample # 2	0.116	02:16
Breath Volume	2.097	02:16
Air Blank	0.000	02:16
Control Test	0.079	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 03/10/2021 14:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: PEREZ BETANCOURT

DATE/TIME OF TEST: 03/01/2021 03:59:23

Diagnostics Check		OK	03:47	
Air Blank		0.000	03:47	
Control Test		0.080	03:47	
Air Blank		0.000	03:48	
Subject Sample	# 1	0.145	03:51	Volume Not Met
Breath Volume		0.285	03:51	
Air Blank		0.000	03:52	
Air Blank		0.000	03:54	
Subject Sample	# 2	0.161	03:54	No .020 agreement
Breath Volume		1.527	03:54	
Air Blank		0.000	03:55	
Air Blank		0.000	03:57	
Subject Sample	# 3	0.140	03:57	No .020 agreement
Breath Volume		1.347	03:57	
Air Blank		0.000	03:58	
Control Test		0.078	03:58	
Air Blank		0.000	03:59	
Diagnostics Check		OK	03:59	

Data Download Date/Time: 03/10/2021 14:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: PEREZ BETANCOURT

DATE/TIME OF TEST: 03/01/2021 04:08:59

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.079	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.143	04:03
Breath Volume	1.246	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.153	04:07
Breath Volume	1.351	04:07
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 03/10/2021 14:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: PEREZ GONZALEZ

DATE/TIME OF TEST: 03/02/2021 01:09:04

Diagnostics Check		OK	01:02
Air Blank		0.000	01:02
Control Test		0.080	01:03
Air Blank		0.000	01:03
Subject Sample	# 1	0.256	01:04
Breath Volume		2.464	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:06
Subject Sample	# 2	0.252	01:07
Breath Volume		1.937	01:07
Air Blank		0.000	01:08
Control Test		0.078	01:08
Air Blank		0.000	01:08
Diagnostics Check		OK	01:09

Data Download Date/Time: 03/10/2021 14:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: FRICKE

DATE/TIME OF TEST: 03/02/2021 05:39:17

Diagnostics Check	OK	05:32
Air Blank	0.000	05:33
Control Test	0.079	05:33
Air Blank	0.000	05:34
Subject Sample # 1	0.000	05:34
Breath Volume	1.972	05:34
Air Blank	0.000	05:35
Air Blank	0.000	05:37
Subject Sample # 2	0.000	05:37
Breath Volume	1.664	05:37
Air Blank	0.000	05:38
Control Test	0.080	05:38
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

Data Download Date/Time: 03/10/2021 14:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KENNEDY,D 129956
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:09

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/02/2021 14:38:34

Diagnostics Check	OK	14:32
Air Blank	0.000	14:32
Control Test	0.080	14:32
Air Blank	0.000	14:33
Subject Sample # 1	0.000	14:34
Breath Volume	2.882	14:34
Air Blank	0.000	14:34
Air Blank	0.000	14:36
Subject Sample # 2	0.000	14:37
Breath Volume	3.093	14:37
Air Blank	0.000	14:37
Control Test	0.080	14:37
Air Blank	0.000	14:38
Diagnostics Check	OK	14:38

Data Download Date/Time: 03/10/2021 14:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARRELL,Q
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/03/2021 00:19:10

Diagnostics Check	OK	00:11
Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:13
Subject Sample # 1	0.091	00:14
Breath Volume	1.371	00:14
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.103	00:17
Breath Volume	1.691	00:17
Air Blank	0.000	00:18
Control Test	0.079	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 03/10/2021 14:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY, ANIA, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: SOLOMON JR

DATE/TIME OF TEST: 03/04/2021 04:25:09

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.081	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.163	04:20
Breath Volume	1.824	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:22
Subject Sample # 2	0.163	04:23
Breath Volume	2.355	04:23
Air Blank	0.000	04:24
Control Test	0.079	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 03/10/2021 14:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: ENNIS

DATE/TIME OF TEST: 03/05/2021 01:14:45

Diagnostics Check		OK	01:07
Air Blank		0.000	01:08
Control Test		0.081	01:08
Air Blank		0.000	01:08
Subject Sample	# 1	0.194	01:09
Breath Volume		2.402	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.203	01:13
Breath Volume		1.167	01:13
Air Blank		0.000	01:13
Control Test		0.079	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 03/10/2021 14:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: NOLASCO LUNA

DATE/TIME OF TEST: 03/05/2021 03:52:51

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.081	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.134	03:48
Breath Volume	3.128	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:50
Subject Sample # 2	0.138	03:51
Breath Volume	3.230	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 03/10/2021 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 03/05/2021 23:08:16

Diagnostics Check	OK	23:03	
Air Blank	0.000	23:04	
Control Test	0.080	23:04	
Air Blank	0.000	23:05	
Subject Sample # 1	0.215	23:06	
Breath Volume	1.683	23:06	
Air Blank		23:06	RFI Detect
Air Blank	0.000	23:07	
Control Test	0.080	23:07	
Air Blank	0.000	23:08	
Diagnostics Check	OK	23:08	

Data Download Date/Time: 03/10/2021 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 03/05/2021 23:18:56

Diagnostics Check		OK	23:12
Air Blank		0.000	23:12
Control Test		0.080	23:12
Air Blank		0.000	23:13
Subject Sample	# 1	0.214	23:14
Breath Volume		1.683	23:14
Air Blank		0.000	23:14
Air Blank		0.000	23:16
Subject Sample	# 2	0.204	23:17
Breath Volume		1.507	23:17
Air Blank		0.000	23:18
Control Test		0.079	23:18
Air Blank		0.000	23:18
Diagnostics Check		OK	23:18

Data Download Date/Time: 03/10/2021 14:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: JOHN

DATE/TIME OF TEST: 03/06/2021 03:21:49

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:09	
Control Test		0.080	03:09	
Air Blank		0.000	03:09	
Subject Sample	# 1	0.127	03:13	Volume Not Met
Breath Volume		0.992	03:13	
Air Blank		0.000	03:13	
Air Blank		0.000	03:15	
Subject Sample	# 2	0.136	03:16	
Breath Volume		2.027	03:16	
Air Blank		0.000	03:17	
Air Blank		0.000	03:18	
Subject Sample	# 3	0.138	03:20	
Breath Volume		1.660	03:20	
Air Blank		0.000	03:20	
Control Test		0.079	03:21	
Air Blank		0.000	03:21	
Diagnostics Check		OK	03:21	

Data Download Date/Time: 03/10/2021 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: SALEMI

DATE/TIME OF TEST: 03/06/2021 03:49:10

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.106	03:44
Breath Volume	1.187	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.094	03:47
Breath Volume	1.136	03:47
Air Blank	0.000	03:48
Control Test	0.077	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 03/10/2021 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROSALES,F
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: HERNANDEZ HERNANDEZ

DATE/TIME OF TEST: 03/07/2021 01:34:36

Diagnostics Check		OK	01:27
Air Blank		0.000	01:27
Control Test		0.080	01:28
Air Blank		0.000	01:28
Subject Sample	# 1	0.251	01:29
Breath Volume		1.328	01:29
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.250	01:32
Breath Volume		1.230	01:32
Air Blank		0.000	01:33
Control Test		0.079	01:33
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 03/10/2021 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: AGUILA JR

DATE/TIME OF TEST: 03/07/2021 02:45:48

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.080	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.213	02:40
Breath Volume	1.855	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:42
Subject Sample # 2	0.198	02:44
Breath Volume	2.980	02:44
Air Blank	0.000	02:44
Control Test	0.079	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 03/10/2021 14:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY, ANIA, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON, M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: GALLEGOS JR

DATE/TIME OF TEST: 03/07/2021 03:09:50

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.080	03:06	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:08	
Control Test	0.080	03:09	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

Data Download Date/Time: 03/10/2021 14:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY, ANIA, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: DAVID CERPA

DATE/TIME OF TEST: 03/07/2021 05:32:27

Diagnostics Check		OK	05:19	
Air Blank		0.000	05:20	
Control Test		0.081	05:20	
Air Blank		0.000	05:21	
Subject Sample	# 1		05:24	No Sample Provided
Breath Volume		0.000	05:24	
Air Blank		0.000	05:24	
Air Blank		0.000	05:26	
Subject Sample	# 2	0.151	05:27	
Breath Volume		1.511	05:27	
Air Blank		0.000	05:28	
Air Blank		0.000	05:29	
Subject Sample	# 3	0.164	05:30	
Breath Volume		2.625	05:30	
Air Blank		0.000	05:31	
Control Test		0.080	05:31	
Air Blank		0.000	05:32	
Diagnostics Check		OK	05:32	

Data Download Date/Time: 03/10/2021 14:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: MARRERO GANDARILLAS

DATE/TIME OF TEST: 03/07/2021 23:44:10

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.081	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.165	23:39
Breath Volume	2.339	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:41
Subject Sample # 2	0.166	23:42
Breath Volume	2.585	23:42
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 03/10/2021 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 03/08/2021 04:52:35

Diagnostics Check		OK	04:40	
Air Blank		0.000	04:41	
Control Test		0.082	04:41	
Air Blank		0.000	04:42	
Subject Sample	# 1	0.157	04:45	Volume Not Met
Breath Volume		0.558	04:45	
Air Blank		0.000	04:46	
Air Blank		0.000	04:47	
Subject Sample	# 2	0.155	04:50	Volume Not Met
Breath Volume		0.203	04:50	
Air Blank		0.000	04:51	
Control Test		0.079	04:51	
Air Blank		0.000	04:52	
Diagnostics Check		OK	04:52	

Data Download Date/Time: 03/10/2021 14:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 03/08/2021 05:37:27

Diagnostics Check		OK	05:20	
Air Blank		0.000	05:20	
Control Test		0.079	05:20	
Air Blank		0.000	05:21	
Subject Sample	# 1	0.165	05:24	Volume Not Met
Breath Volume		0.183	05:24	
Air Blank		0.000	05:25	
Air Blank		0.000	05:27	
Subject Sample	# 2	0.153	05:30	
Breath Volume		1.179	05:30	
Air Blank		0.000	05:30	
Air Blank		0.000	05:32	
Subject Sample	# 3	0.145	05:35	Volume Not Met
Breath Volume		0.394	05:35	
Air Blank		0.000	05:36	
Control Test		0.077	05:36	
Air Blank		0.000	05:37	
Diagnostics Check		OK	05:37	

Data Download Date/Time: 03/10/2021 14:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 03/08/2021 05:56:24

Diagnostics Check		OK	05:39	
Air Blank		0.000	05:39	
Control Test		0.078	05:40	
Air Blank		0.000	05:40	
Subject Sample	# 1	0.153	05:43	Volume Not Met
Breath Volume		0.789	05:43	
Air Blank		0.000	05:44	
Air Blank		0.000	05:46	
Subject Sample	# 2	0.160	05:48	
Breath Volume		1.324	05:48	
Air Blank		0.000	05:49	
Air Blank		0.000	05:51	
Subject Sample	# 3	0.162	05:54	Volume Not Met
Breath Volume		0.453	05:54	
Air Blank		0.000	05:55	
Control Test		0.077	05:55	
Air Blank		0.000	05:56	
Diagnostics Check		OK	05:56	

Data Download Date/Time: 03/10/2021 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: MCKINNEY

DATE/TIME OF TEST: 03/11/2021 04:06:22

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.081	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.117	04:01
Breath Volume	3.546	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.114	04:04
Breath Volume	1.476	04:04
Air Blank	0.000	04:05
Control Test	0.080	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 04/06/2021 16:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY, ANIA, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND, M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ORTA GONELL

DATE/TIME OF TEST: 03/12/2021 03:32:41

Diagnostics Check		OK	03:25
Air Blank		0.000	03:26
Control Test		0.081	03:26
Air Blank		0.000	03:27
Subject Sample	# 1	0.104	03:27
Breath Volume		1.648	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.106	03:30
Breath Volume		1.289	03:30
Air Blank		0.000	03:31
Control Test		0.079	03:32
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 04/06/2021 16:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HARRIS,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: SOTO CARRILLO

DATE/TIME OF TEST: 03/13/2021 02:08:48

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.081	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.136	02:03
Breath Volume	1.375	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.150	02:07
Breath Volume	1.246	02:07
Air Blank	0.000	02:07
Control Test	0.080	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 04/06/2021 16:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 03/13/2021 05:06:26

Diagnostics Check	OK	04:59
Air Blank	0.000	05:00
Control Test	0.080	05:00
Air Blank	0.000	05:01
Subject Sample # 1	0.139	05:01
Breath Volume	2.792	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.147	05:04
Breath Volume	2.808	05:04
Air Blank	0.000	05:05
Control Test	0.079	05:05
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 04/06/2021 16:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: VELAZQUEZ ONTIVEROS

DATE/TIME OF TEST: 03/13/2021 05:41:53

Diagnostics Check	OK	05:34
Air Blank	0.000	05:35
Control Test	0.079	05:35
Air Blank	0.000	05:36
Subject Sample # 1	0.193	05:37
Breath Volume	2.398	05:37
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample # 2	0.199	05:40
Breath Volume	1.507	05:40
Air Blank	0.000	05:40
Control Test	0.079	05:41
Air Blank	0.000	05:41
Diagnostics Check	OK	05:41

Data Download Date/Time: 04/06/2021 16:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HOPPE,Z
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: BUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: ENRIGHT

DATE/TIME OF TEST: 03/13/2021 18:17:32

Diagnostics Check	OK	18:06	
Air Blank	0.000	18:07	
Control Test	0.081	18:07	
Air Blank	0.000	18:08	
Subject Sample	# 1	18:11	No Sample Provided
Breath Volume	0.000	18:11	
Air Blank	0.000	18:11	
Air Blank	0.000	18:13	
Subject Sample	# 2	18:16	Subject Test Refused
Breath Volume	0.000	18:16	
Air Blank	0.000	18:16	
Control Test	0.081	18:16	
Air Blank	0.000	18:17	
Diagnostics Check	OK	18:17	

Data Download Date/Time: 04/06/2021 16:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 03/13/2021 23:33:31

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.081	23:27
Air Blank	0.000	23:27
Subject Sample # 1	0.087	23:28
Breath Volume	1.179	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:30
Subject Sample # 2	0.088	23:31
Breath Volume	1.695	23:31
Air Blank	0.000	23:32
Control Test	0.080	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 04/06/2021 16:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 03/15/2021 05:30:42

Diagnostics Check	OK	05:23
Air Blank	0.000	05:24
Control Test	0.081	05:24
Air Blank	0.000	05:25
Subject Sample # 1	0.223	05:25
Breath Volume	2.050	05:25
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample # 2	0.232	05:28
Breath Volume	2.140	05:28
Air Blank	0.000	05:29
Control Test	0.079	05:30
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 04/06/2021 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 03/17/2021 22:50:33

Diagnostics Check	OK	22:43
Air Blank	0.000	22:43
Control Test	0.081	22:43
Air Blank	0.000	22:44
Subject Sample # 1	0.161	22:45
Breath Volume	1.328	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.165	22:48
Breath Volume	2.316	22:48
Air Blank	0.000	22:49
Control Test	0.080	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 04/06/2021 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY, ANIA, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: TITO

DATE/TIME OF TEST: 03/18/2021 02:38:28

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.080	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.121	02:33
Breath Volume		2.617	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.118	02:36
Breath Volume		2.226	02:36
Air Blank		0.000	02:37
Control Test		0.080	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 04/06/2021 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: MARR JR

DATE/TIME OF TEST: 03/18/2021 20:00:02

Diagnostics Check		OK	19:53
Air Blank		0.000	19:53
Control Test		0.079	19:54
Air Blank		0.000	19:54
Subject Sample	# 1	0.212	19:55
Breath Volume		2.687	19:55
Air Blank		0.000	19:56
Air Blank		0.000	19:57
Subject Sample	# 2	0.203	19:58
Breath Volume		1.796	19:58
Air Blank		0.000	19:59
Control Test		0.079	19:59
Air Blank		0.000	19:59
Diagnostics Check		OK	20:00

Data Download Date/Time: 04/06/2021 16:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: YUKNIEWICZ,B
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: LAMOUNTAIN

DATE/TIME OF TEST: 03/19/2021 04:04:52

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.164	03:59
Breath Volume	1.347	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.184	04:03
Breath Volume	1.500	04:03
Air Blank	0.000	04:03
Control Test	0.078	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 04/06/2021 16:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 03/20/2021 02:21:28

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.081	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.119	02:16
Breath Volume	1.125	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.119	02:19
Breath Volume	1.890	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 04/06/2021 16:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STABINS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:24

SUBJECT'S LAST NAME: BUZZARD

DATE/TIME OF TEST: 03/20/2021 08:51:54

Diagnostics Check	OK	08:44
Air Blank	0.000	08:44
Control Test	0.081	08:45
Air Blank	0.000	08:45
Subject Sample # 1	0.291	08:46
Breath Volume	3.453	08:46
Air Blank	0.000	08:47
Air Blank	0.000	08:49
Subject Sample # 2	0.294	08:50
Breath Volume	3.730	08:50
Air Blank	0.000	08:50
Control Test	0.079	08:51
Air Blank	0.000	08:51
Diagnostics Check	OK	08:51

Data Download Date/Time: 04/06/2021 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 03/20/2021 23:28:11

Diagnostics Check		OK	23:20
Air Blank		0.000	23:20
Control Test		0.081	23:20
Air Blank		0.000	23:21
Subject Sample	# 1	0.240	23:23
Breath Volume		3.265	23:23
Air Blank		0.000	23:23
Air Blank		0.000	23:25
Subject Sample	# 2	0.231	23:26
Breath Volume		4.003	23:26
Air Blank		0.000	23:27
Control Test		0.079	23:27
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 04/06/2021 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY, ANIA, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROSALES, F
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/21/2021 03:09:43

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.080	03:03
Air Blank		0.000	03:04
Subject Sample	# 1	0.145	03:04
Breath Volume		2.964	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.145	03:08
Breath Volume		2.539	03:08
Air Blank		0.000	03:08
Control Test		0.080	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 04/06/2021 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: ULLOA LANDA

DATE/TIME OF TEST: 03/21/2021 04:40:02

Diagnostics Check		OK	04:28	
Air Blank		0.000	04:28	
Control Test		0.081	04:29	
Air Blank		0.000	04:29	
Subject Sample	# 1	0.087	04:32	Volume Not Met
Breath Volume		0.601	04:32	
Air Blank		0.000	04:33	
Air Blank		0.000	04:35	
Subject Sample	# 2	0.088	04:38	Volume Not Met
Breath Volume		0.343	04:38	
Air Blank		0.000	04:39	
Control Test		0.080	04:39	
Air Blank		0.000	04:39	
Diagnostics Check		OK	04:40	

Data Download Date/Time: 04/06/2021 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: ULLOA LANDA

DATE/TIME OF TEST: 03/21/2021 04:53:54

Diagnostics Check		OK	04:42	
Air Blank		0.000	04:42	
Control Test		0.080	04:43	
Air Blank		0.000	04:43	
Subject Sample	# 1	0.087	04:46	Volume Not Met
Breath Volume		0.335	04:46	
Air Blank		0.000	04:47	
Air Blank		0.000	04:49	
Subject Sample	# 2		04:52	No Sample Provided
Breath Volume		0.000	04:52	
Air Blank		0.000	04:52	
Control Test		0.079	04:53	
Air Blank		0.000	04:53	
Diagnostics Check		OK	04:53	

Data Download Date/Time: 04/06/2021 16:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY, ANIA, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: OATES

DATE/TIME OF TEST: 03/21/2021 06:04:34

Diagnostics Check	OK	05:58
Air Blank	0.000	05:58
Control Test	0.081	05:58
Air Blank	0.000	05:59
Subject Sample # 1	0.141	05:59
Breath Volume	1.433	05:59
Air Blank	0.000	06:00
Air Blank	0.000	06:02
Subject Sample # 2	0.148	06:02
Breath Volume	2.210	06:02
Air Blank	0.000	06:03
Control Test	0.081	06:03
Air Blank	0.000	06:04
Diagnostics Check	OK	06:04

Data Download Date/Time: 04/06/2021 16:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY, ANIA, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: LOLLIS

DATE/TIME OF TEST: 03/23/2021 01:40:53

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.081	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.000	01:36
Breath Volume	1.234	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.000	01:39
Breath Volume	1.289	01:39
Air Blank	0.000	01:39
Control Test	0.081	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 04/06/2021 16:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VANTREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MOREAU

DATE/TIME OF TEST: 03/24/2021 03:18:57

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.081	03:12
Air Blank		0.000	03:12
Subject Sample	# 1	0.198	03:14
Breath Volume		2.167	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:16
Subject Sample	# 2	0.191	03:17
Breath Volume		2.449	03:17
Air Blank		0.000	03:17
Control Test		0.079	03:18
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 04/06/2021 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: SMITH III

DATE/TIME OF TEST: 03/25/2021 05:22:53

Diagnostics Check	OK	05:15
Air Blank	0.000	05:16
Control Test	0.080	05:16
Air Blank	0.000	05:17
Subject Sample # 1	0.189	05:18
Breath Volume	1.593	05:18
Air Blank	0.000	05:19
Air Blank	0.000	05:20
Subject Sample # 2	0.198	05:21
Breath Volume	2.542	05:21
Air Blank	0.000	05:21
Control Test	0.077	05:22
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 04/06/2021 16:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY, ANIA, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY, M
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/26/2021 02:05:00

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.081	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.000	02:00
Breath Volume	1.886	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.000	02:03
Breath Volume	1.613	02:03
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:05

Data Download Date/Time: 04/06/2021 16:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY, ANIA, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: PILLAINAYAGAM

DATE/TIME OF TEST: 03/27/2021 02:31:46

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.081	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.169	02:23
Breath Volume	1.152	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.191	02:27
Breath Volume	2.531	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 3	0.187	02:30
Breath Volume	1.835	02:30
Air Blank	0.000	02:30
Control Test	0.079	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 04/06/2021 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY, ANIA, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VANTREESE, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: CAMPOSANO

DATE/TIME OF TEST: 03/27/2021 03:07:22

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:59	
Control Test	0.080	02:59	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:03	No Sample Provided
Breath Volume	0.000	03:03	
Air Blank	0.000	03:03	
Air Blank	0.000	03:05	
Subject Sample	# 2	03:05	Subject Test Refused
Breath Volume	0.000	03:05	
Air Blank	0.000	03:06	
Control Test	0.080	03:06	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

Data Download Date/Time: 04/06/2021 16:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: BORENSTEIN

DATE/TIME OF TEST: 03/27/2021 03:26:14

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.171	03:21
Breath Volume	2.957	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:23
Subject Sample # 2	0.163	03:24
Breath Volume	2.441	03:24
Air Blank	0.000	03:25
Control Test	0.078	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 04/06/2021 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: ZAMARRIPA

DATE/TIME OF TEST: 03/27/2021 05:19:37

Diagnostics Check		OK	05:12
Air Blank		0.000	05:13
Control Test		0.080	05:13
Air Blank		0.000	05:13
Subject Sample	# 1	0.186	05:14
Breath Volume		1.503	05:14
Air Blank		0.000	05:15
Air Blank		0.000	05:17
Subject Sample	# 2	0.186	05:17
Breath Volume		1.191	05:17
Air Blank		0.000	05:18
Control Test		0.079	05:19
Air Blank		0.000	05:19
Diagnostics Check		OK	05:19

Data Download Date/Time: 04/06/2021 16:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KROUSE,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: PARIS

DATE/TIME OF TEST: 03/27/2021 18:42:32

Diagnostics Check	OK	18:36
Air Blank	0.000	18:36
Control Test	0.081	18:37
Air Blank	0.000	18:37
Subject Sample # 1	0.000	18:38
Breath Volume	2.640	18:38
Air Blank	0.000	18:38
Air Blank	0.000	18:40
Subject Sample # 2	0.000	18:41
Breath Volume	1.675	18:41
Air Blank	0.000	18:41
Control Test	0.081	18:41
Air Blank	0.000	18:42
Diagnostics Check	OK	18:42

Data Download Date/Time: 04/06/2021 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/28/2021 03:51:31

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.081	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.201	03:46
Breath Volume	1.851	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:48
Subject Sample # 2	0.214	03:49
Breath Volume	2.460	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 04/06/2021 16:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: GOODRICH

DATE/TIME OF TEST: 03/29/2021 02:25:08

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.081	02:18
Air Blank		0.000	02:19
Subject Sample	# 1	0.179	02:20
Breath Volume		1.726	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:22
Subject Sample	# 2	0.169	02:23
Breath Volume		1.406	02:23
Air Blank		0.000	02:24
Control Test		0.080	02:24
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 04/06/2021 16:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 03/29/2021 04:34:03

Diagnostics Check		OK	04:21	
Air Blank		0.000	04:21	
Control Test		0.081	04:22	
Air Blank		0.000	04:22	
Subject Sample	# 1	0.114	04:25	Volume Not Met
Breath Volume		0.960	04:25	
Air Blank		0.000	04:26	
Air Blank		0.000	04:28	
Subject Sample	# 2	0.132	04:28	
Breath Volume		2.054	04:28	
Air Blank		0.000	04:29	
Air Blank		0.000	04:31	
Subject Sample	# 3	0.130	04:32	
Breath Volume		2.062	04:32	
Air Blank		0.000	04:33	
Control Test		0.079	04:33	
Air Blank		0.000	04:33	
Diagnostics Check		OK	04:34	

Data Download Date/Time: 04/06/2021 16:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOOTHE,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOOTHE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 182793
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: ROUTH

DATE/TIME OF TEST: 03/11/2021 22:59:32

Diagnostics Check	OK	22:56	
Air Blank	0.000	22:57	
Control Test	0.083	22:57	
Air Blank	0.000	22:57	
Subject Sample	# 1	22:58	Subject Test Refused
Breath Volume	0.000	22:58	
Air Blank	0.000	22:58	
Control Test	0.082	22:58	
Air Blank	0.000	22:59	
Diagnostics Check	OK	22:59	

Data Download Date/Time: 04/14/2021 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KITTRELL,WILLIAM,G
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KITTRELL,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: CHRISTENOT

DATE/TIME OF TEST: 03/14/2021 00:48:32

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.241	00:43
Breath Volume	1.113	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.243	00:46
Breath Volume	1.433	00:46
Air Blank	0.000	00:47
Control Test	0.081	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 04/14/2021 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORELLANA,ADALBERTO,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ORELLANA,ADALBERTO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 07:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 03/14/2021 08:35:43

Diagnostics Check	OK	08:29
Air Blank	0.000	08:29
Control Test	0.082	08:30
Air Blank	0.000	08:30
Subject Sample # 1	0.000	08:31
Breath Volume	3.355	08:31
Air Blank	0.000	08:31
Air Blank	0.000	08:33
Subject Sample # 2	0.000	08:34
Breath Volume	2.781	08:34
Air Blank	0.000	08:34
Control Test	0.081	08:35
Air Blank	0.000	08:35
Diagnostics Check	OK	08:35

Data Download Date/Time: 04/14/2021 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORELLANA, ADALBERTO, L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ORELLANA, ADALBERTO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 09:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 03/26/2021 23:31:52

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.082	23:25
Air Blank	0.000	23:25
Subject Sample # 1	0.101	23:27
Breath Volume	2.554	23:27
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.096	23:30
Breath Volume	2.527	23:30
Air Blank	0.000	23:30
Control Test	0.081	23:31
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 04/14/2021 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLIVA,RAUL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: OLIVA,RAUL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: LAZARO

DATE/TIME OF TEST: 03/01/2021 00:31:36

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:30	
Control Test	0.081	00:30	
Air Blank		00:31	RFI Detect
Air Blank	0.000	00:31	

Data Download Date/Time: 03/22/2021 15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,ALEXIA,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: REYES,ALEXIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:29

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/04/2021 18:29:52

Diagnostics Check	OK	18:27	
Air Blank	0.000	18:27	
Control Test	0.080	18:27	
Air Blank	0.000	18:28	
Subject Sample	# 1	18:28	Subject Test Refused
Breath Volume	0.000	18:28	
Air Blank	0.000	18:28	
Control Test	0.080	18:29	
Air Blank	0.000	18:29	
Diagnostics Check	OK	18:29	

Data Download Date/Time: 03/22/2021 15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MONTROYA,STEVEN,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: GUERRIER,CLERVENSCIA

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 18:45

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 03/05/2021 20:38:42

Diagnostics Check	OK	20:35	
Air Blank	0.000	20:36	
Control Test	0.081	20:36	
Air Blank	0.000	20:37	
Subject Sample	# 1	20:37	Subject Test Refused
Breath Volume	0.000	20:37	
Air Blank	0.000	20:37	
Control Test	0.082	20:38	
Air Blank	0.000	20:38	
Diagnostics Check	OK	20:38	

Data Download Date/Time: 03/22/2021 15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PILOTO,F,E
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RODRIGUEZ,A
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 03/14/2021 22:09:34

Diagnostics Check	OK	22:06	
Air Blank	0.000	22:07	
Control Test	0.080	22:07	
Air Blank	0.000	22:07	
Subject Sample	# 1	22:08	Subject Test Refused
Breath Volume	0.000	22:08	
Air Blank	0.000	22:08	
Control Test	0.081	22:08	
Air Blank	0.000	22:09	
Diagnostics Check	OK	22:09	

Data Download Date/Time: 03/22/2021 15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: OQUENDO,JESSICA,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: OQUENDO,JESSICA

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 08:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 08:51

SUBJECT'S LAST NAME: WILLIS KAYN

DATE/TIME OF TEST: 03/19/2021 09:16:23

Diagnostics Check	OK	09:12	
Air Blank	0.000	09:13	
Control Test	0.080	09:13	
Air Blank	0.000	09:13	
Subject Sample	# 1	09:15	Subject Test Refused
Breath Volume	0.000	09:15	
Air Blank	0.000	09:15	
Control Test	0.081	09:15	
Air Blank	0.000	09:16	
Diagnostics Check	OK	09:16	

Data Download Date/Time: 03/22/2021 15:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTAYA, GUILLERMO,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BOLEY, CHRISTOPHER
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: BUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: VIOLA

DATE/TIME OF TEST: 03/26/2021 23:41:15

Diagnostics Check	OK	23:34
Air Blank	0.000	23:35
Control Test	0.082	23:35
Air Blank	0.000	23:36
Subject Sample # 1	0.117	23:36
Breath Volume	1.753	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.118	23:39
Breath Volume	2.425	23:39
Air Blank	0.000	23:40
Control Test	0.081	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 05/03/2021 14:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTAYA,GUILLERMO,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: WHITT,JESSIE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: BUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: LUGO AGRAPIDAKIS

DATE/TIME OF TEST: 03/20/2021 23:42:37

Diagnostics Check	OK	23:39	
Air Blank	0.000	23:40	
Control Test	0.080	23:40	
Air Blank	0.000	23:40	
Subject Sample	# 1	23:41	Subject Test Refused
Breath Volume	0.000	23:41	
Air Blank	0.000	23:41	
Control Test	0.072	23:42	Control Outside Tolerance
Air Blank	0.000	23:42	

Data Download Date/Time: 03/22/2021 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,RUBEN,G
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: CARTAYA,GUILLERMO
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: BUI
ARREST TIME: 17:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: FERRANDO

DATE/TIME OF TEST: 03/27/2021 19:02:09

Diagnostics Check	OK	18:58	
Air Blank	0.000	18:59	
Control Test	0.079	18:59	
Air Blank	0.000	19:00	
Subject Sample	# 1	19:00	Subject Test Refused
Breath Volume	0.000	19:00	
Air Blank	0.000	19:01	
Control Test	0.083	19:01	
Air Blank	0.000	19:02	
Diagnostics Check	OK	19:02	

Data Download Date/Time: 05/03/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTAYA,GUILLERMO,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: CARTAYA,GUILLERMO
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: BUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 03/27/2021 20:57:34

Diagnostics Check	OK	20:56	
Air Blank	0.000	20:56	
Control Test	0.090	20:56	Control Outside Tolerance
Air Blank	0.000	20:57	

Data Download Date/Time: 05/03/2021 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTAYA,GUILLERMO,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: CARTAYA,GUILLERMO
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: BUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 03/27/2021 21:03:03

Diagnostics Check	OK	21:00	
Air Blank	0.000	21:00	
Control Test	0.076	21:00	
Air Blank	0.000	21:01	
Subject Sample	# 1	21:01	Subject Test Refused
Breath Volume	0.000	21:01	
Air Blank	0.000	21:02	
Control Test	0.083	21:02	
Air Blank	0.000	21:02	
Diagnostics Check	OK	21:03	

Data Download Date/Time: 05/03/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTAYA,GUILLERMO,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SAMPLE,CARSON
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: BUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 03/27/2021 23:40:52

Diagnostics Check	OK	23:39	
Air Blank	0.000	23:39	
Control Test	0.086	23:40	Control Outside Tolerance
Air Blank	0.000	23:40	

Data Download Date/Time: 05/03/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTAYA,GUILLERMO,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SAMPLE,CARSON
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: BUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 03/27/2021 23:45:04

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:44	
Control Test	0.087	23:44	Control Outside Tolerance
Air Blank	0.000	23:45	

Data Download Date/Time: 05/03/2021 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTAYA,GUILLERMO,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: YUREWITCH,KYLE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: BUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/28/2021 20:32:55

Diagnostics Check	OK	20:26	
Air Blank	0.000	20:27	
Control Test	0.081	20:27	
Air Blank	0.000	20:28	
Subject Sample # 1	0.023	20:28	
Breath Volume	1.324	20:28	
Air Blank	0.000	20:29	
Air Blank	0.000	20:31	
Subject Sample # 2		20:31	Slope Not Met
Breath Volume	0.839	20:31	
Air Blank	0.000	20:32	
Control Test	0.082	20:32	
Air Blank	0.000	20:32	
Diagnostics Check	OK	20:32	

Data Download Date/Time: 05/03/2021 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTAYA, GUILLERMO,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: YUREWITCH, KYLE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: BUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/28/2021 20:49:12

Diagnostics Check	OK	20:46	
Air Blank	0.000	20:46	
Control Test	0.084	20:46	
Air Blank	0.000	20:47	
Subject Sample	# 1	20:47	Slope Not Met
Breath Volume	2.347	20:47	
Air Blank	0.000	20:48	
Control Test	0.074	20:48	Control Outside Tolerance
Air Blank	0.000	20:49	

Data Download Date/Time: 05/03/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTAYA,GUILLERMO,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: YUREWITCH,KYLE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: BUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/28/2021 20:52:18

Diagnostics Check	OK	20:50	
Air Blank	0.000	20:51	
Control Test	0.073	20:51	Control Outside Tolerance
Air Blank	0.000	20:52	

Data Download Date/Time: 05/03/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CARTAYA,GUILLERMO,

BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)

ARRESTING OFFICER: YUREWITCH,KYLE

ARRESTING OFFICER AGENCY: FFWCC (West Palm)

LAST AGENCY INSPECTION: 03/22/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: BUI

ARREST TIME: 19:49

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/28/2021 21:07:19

Diagnostics Check	OK	21:05	
Air Blank	0.000	21:06	
Control Test	0.087	21:06	Control Outside Tolerance
Air Blank	0.000	21:07	

Data Download Date/Time: 05/03/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ESTAPE,ZACHARY,C

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: HEREDIA RUBIO,7L26

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 03/22/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 02:37

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: NUNEZ ORELLANA

DATE/TIME OF TEST: 03/31/2021 03:08:34

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:06	
Control Test	0.082	03:06	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:07	
Control Test	0.083	03:07	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 05/03/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MARANO,JORGE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ROE

DATE/TIME OF TEST: 03/05/2021 02:44:00

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:41	
Control Test	0.078	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:43	
Control Test	0.079	02:43	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:44	

Data Download Date/Time: 04/07/2021 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: GROESCHEN

DATE/TIME OF TEST: 03/06/2021 02:21:13

Diagnostics Check		OK	02:14
Air Blank		0.000	02:15
Control Test		0.079	02:15
Air Blank		0.000	02:15
Subject Sample	# 1	0.248	02:16
Breath Volume		1.429	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.259	02:19
Breath Volume		2.175	02:19
Air Blank		0.000	02:20
Control Test		0.077	02:20
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 04/07/2021 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SLATTERY

DATE/TIME OF TEST: 03/06/2021 03:39:07

Diagnostics Check		OK	03:29
Air Blank		0.000	03:30
Control Test		0.079	03:30
Air Blank		0.000	03:31
Subject Sample	# 1	0.228	03:33
Breath Volume		1.214	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.225	03:37
Breath Volume		1.660	03:37
Air Blank		0.000	03:38
Control Test		0.077	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 04/07/2021 09:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: THEODORE

DATE/TIME OF TEST: 03/07/2021 01:27:42

Diagnostics Check	OK	01:24	
Air Blank	0.000	01:24	
Control Test	0.078	01:25	
Air Blank	0.000	01:25	
Subject Sample	# 1	01:26	Subject Test Refused
Breath Volume	0.000	01:26	
Air Blank	0.000	01:26	
Control Test	0.078	01:27	
Air Blank	0.000	01:27	
Diagnostics Check	OK	01:27	

Data Download Date/Time: 04/07/2021 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: ANDREW

DATE/TIME OF TEST: 03/07/2021 03:37:43

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.078	03:32
Air Blank		0.000	03:32
Subject Sample	# 1	0.000	03:33
Breath Volume		2.691	03:33
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.000	03:36
Breath Volume		2.839	03:36
Air Blank		0.000	03:36
Control Test		0.079	03:37
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 04/07/2021 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: BANHAN

DATE/TIME OF TEST: 03/07/2021 23:29:55

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:21	
Control Test	0.079	23:22	
Air Blank	0.000	23:22	
Subject Sample	# 1	23:25	No Sample Provided
Breath Volume	0.000	23:25	
Air Blank	0.000	23:26	
Air Blank	0.000	23:28	
Subject Sample	# 2	23:28	Subject Test Refused
Breath Volume	0.000	23:28	
Air Blank	0.000	23:28	
Control Test	0.079	23:29	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

Data Download Date/Time: 04/07/2021 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 03/08/2021 04:55:12

Diagnostics Check		OK	04:48
Air Blank		0.000	04:49
Control Test		0.078	04:49
Air Blank		0.000	04:49
Subject Sample	# 1	0.201	04:50
Breath Volume		1.445	04:50
Air Blank		0.000	04:50
Air Blank		0.000	04:52
Subject Sample	# 2	0.196	04:53
Breath Volume		1.527	04:53
Air Blank		0.000	04:54
Control Test		0.078	04:54
Air Blank		0.000	04:55
Diagnostics Check		OK	04:55

Data Download Date/Time: 04/07/2021 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,ALEXANDRIA,N
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MATTHEWS,CASEY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: FRODGE

DATE/TIME OF TEST: 03/09/2021 21:50:21

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:46	
Control Test	0.079	21:47	
Air Blank	0.000	21:47	
Subject Sample	# 1	21:48	RFI Detect
Breath Volume	0.328	21:48	
Air Blank	0.000	21:49	
Control Test	0.079	21:49	
Air Blank	0.000	21:50	
Diagnostics Check	OK	21:50	

Data Download Date/Time: 04/07/2021 09:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,ALEXANDRIA,N
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MATTHEWS,CASEY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: FRODGE

DATE/TIME OF TEST: 03/09/2021 21:54:48

Diagnostics Check	OK	21:51	
Air Blank	0.000	21:52	
Control Test	0.078	21:52	
Air Blank	0.000	21:53	
Subject Sample	# 1	21:53	RFI Detect
Breath Volume	0.000	21:53	
Air Blank	0.000	21:53	
Control Test	0.077	21:54	
Air Blank	0.000	21:54	
Diagnostics Check	OK	21:54	

Data Download Date/Time: 04/07/2021 09:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,ALEXANDRIA,N
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MATTHEWS,CASEY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: FRODGE

DATE/TIME OF TEST: 03/09/2021 21:59:14

Diagnostics Check	OK	21:56	
Air Blank	0.000	21:56	
Control Test	0.078	21:57	
Air Blank	0.000	21:57	
Subject Sample	# 1	21:57	RFI Detect
Breath Volume	0.000	21:57	
Air Blank	0.000	21:58	
Control Test	0.078	21:58	
Air Blank	0.000	21:59	
Diagnostics Check	OK	21:59	

Data Download Date/Time: 04/07/2021 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,ALEXANDRIA,N
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MATTHEW,CASEY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: FRODGE

DATE/TIME OF TEST: 03/09/2021 22:22:00

Diagnostics Check		OK	22:10	
Air Blank		0.000	22:10	
Control Test		0.078	22:10	
Air Blank		0.000	22:11	
Subject Sample	# 1	0.155	22:14	Volume Not Met
Breath Volume		0.203	22:14	
Air Blank		0.000	22:15	
Air Blank		0.000	22:17	
Subject Sample	# 2	0.156	22:20	Volume Not Met
Breath Volume		0.191	22:20	
Air Blank		0.000	22:21	
Control Test		0.077	22:21	
Air Blank		0.000	22:21	
Diagnostics Check		OK	22:22	

Data Download Date/Time: 04/07/2021 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: WEBBER

DATE/TIME OF TEST: 03/11/2021 03:17:03

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.132	03:12
Breath Volume		1.324	03:12
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.131	03:15
Breath Volume		1.160	03:15
Air Blank		0.000	03:16
Control Test		0.079	03:16
Air Blank		0.000	03:16
Diagnostics Check		OK	03:17

Data Download Date/Time: 04/07/2021 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIDAL,NERELYS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MORENO,JORGE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 03/12/2021 20:24:23

Diagnostics Check	OK	20:20	
Air Blank	0.000	20:21	
Control Test	0.078	20:21	
Air Blank	0.000	20:21	
Subject Sample	# 1	20:23	Subject Test Refused
Breath Volume	0.000	20:23	
Air Blank	0.000	20:23	
Control Test	0.078	20:23	
Air Blank	0.000	20:24	
Diagnostics Check	OK	20:24	

Data Download Date/Time: 04/07/2021 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIDAL,NERELYS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: VIDAL,NERELYS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 03/13/2021 00:33:50

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:32	
Control Test	0.080	00:32	
Air Blank	0.000	00:32	
Subject Sample	# 1	00:33	Sequence Aborted
Breath Volume	0.000	00:33	
Air Blank	0.000	00:33	

Data Download Date/Time: 04/07/2021 09:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIDAL,NERELYS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: VIDAL,NERELYS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 03/13/2021 00:52:54

Diagnostics Check		OK	00:45
Air Blank		0.000	00:45
Control Test		0.080	00:46
Air Blank		0.000	00:46
Subject Sample	# 1	0.075	00:47
Breath Volume		2.011	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.082	00:51
Breath Volume		2.664	00:51
Air Blank		0.000	00:51
Control Test		0.079	00:52
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 04/07/2021 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,ALEXANDRIA,N
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HRADECKY,ALEXANDRIA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 03/13/2021 21:33:50

Diagnostics Check	OK	21:30	
Air Blank	0.000	21:31	
Control Test	0.079	21:31	
Air Blank	0.000	21:32	
Subject Sample	# 1	21:32	Subject Test Refused
Breath Volume	0.000	21:32	
Air Blank	0.000	21:32	
Control Test	0.079	21:33	
Air Blank	0.000	21:33	
Diagnostics Check	OK	21:33	

Data Download Date/Time: 04/07/2021 09:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,ALEXANDRIA,N
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MATTHEWS,CASEY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: LLUFRIO PEREZ

DATE/TIME OF TEST: 03/14/2021 00:50:02

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.079	00:42
Air Blank		0.000	00:42
Subject Sample	# 1	0.109	00:44
Breath Volume		1.761	00:44
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.129	00:48
Breath Volume		1.515	00:48
Air Blank		0.000	00:49
Control Test		0.078	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:50

Data Download Date/Time: 04/07/2021 09:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIDAL,NERELYS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MORENO,JORGE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: ORTEGA

DATE/TIME OF TEST: 03/14/2021 04:48:21

Diagnostics Check		OK	04:38
Air Blank		0.000	04:39
Control Test		0.078	04:39
Air Blank		0.000	04:39
Subject Sample	# 1	0.198	04:42
Breath Volume		1.910	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:44
Subject Sample	# 2	0.200	04:46
Breath Volume		1.207	04:46
Air Blank		0.000	04:47
Control Test		0.077	04:47
Air Blank		0.000	04:48
Diagnostics Check		OK	04:48

Data Download Date/Time: 04/07/2021 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,ALEXANDRIA,N
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MATTHEWS,CASEY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: BESZE

DATE/TIME OF TEST: 03/15/2021 03:46:09

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:43	
Control Test	0.079	03:43	
Air Blank	0.000	03:44	
Subject Sample	# 1	03:44	Subject Test Refused
Breath Volume	0.000	03:44	
Air Blank	0.000	03:45	
Control Test	0.079	03:45	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 04/07/2021 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: HUSTON

DATE/TIME OF TEST: 03/17/2021 00:11:01

Diagnostics Check		OK	00:00
Air Blank		0.000	00:01
Control Test		0.079	00:01
Air Blank		0.000	00:02
Subject Sample	# 1	0.166	00:04
Breath Volume		1.207	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:06
Subject Sample	# 2	0.172	00:09
Breath Volume		1.617	00:09
Air Blank		0.000	00:10
Control Test		0.078	00:10
Air Blank		0.000	00:10
Diagnostics Check		OK	00:11

Data Download Date/Time: 04/07/2021 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIDAL,NERELYS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MORENO,JORGE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: MORA

DATE/TIME OF TEST: 03/18/2021 00:21:22

Diagnostics Check		OK	00:05	
Air Blank		0.000	00:05	
Control Test		0.078	00:06	
Air Blank		0.000	00:06	
Subject Sample	# 1	0.226	00:09	Volume Not Met
Breath Volume		0.687	00:09	
Air Blank		0.000	00:10	
Air Blank		0.000	00:12	
Subject Sample	# 2	0.217	00:13	
Breath Volume		1.742	00:13	
Air Blank		0.000	00:14	
Air Blank		0.000	00:16	
Subject Sample	# 3	0.198	00:19	Volume Not Met
Breath Volume		0.453	00:19	
Air Blank		0.000	00:20	
Control Test		0.076	00:20	
Air Blank		0.000	00:21	
Diagnostics Check		OK	00:21	

Data Download Date/Time: 04/07/2021 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: GILLER

DATE/TIME OF TEST: 03/19/2021 19:19:19

Diagnostics Check	OK	19:15	
Air Blank	0.000	19:16	
Control Test	0.078	19:16	
Air Blank	0.000	19:16	
Subject Sample	# 1	19:17	Subject Test Refused
Breath Volume	0.000	19:17	
Air Blank	0.000	19:18	
Control Test	0.078	19:18	
Air Blank	0.000	19:19	
Diagnostics Check	OK	19:19	

Data Download Date/Time: 04/07/2021 09:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 03/20/2021 03:30:03

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.197	03:25
Breath Volume	2.242	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:27
Subject Sample # 2	0.189	03:28
Breath Volume	1.847	03:28
Air Blank	0.000	03:29
Control Test	0.078	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:30

Data Download Date/Time: 04/07/2021 09:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: DURHAM,TOMMY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: CAULK

DATE/TIME OF TEST: 03/20/2021 18:50:18

Diagnostics Check	OK	18:43
Air Blank	0.000	18:44
Control Test	0.079	18:44
Air Blank	0.000	18:45
Subject Sample # 1	0.000	18:45
Breath Volume	2.003	18:45
Air Blank	0.000	18:45
Air Blank	0.000	18:47
Subject Sample # 2	0.000	18:48
Breath Volume	1.917	18:48
Air Blank	0.000	18:49
Control Test	0.078	18:49
Air Blank	0.000	18:50
Diagnostics Check	OK	18:50

Data Download Date/Time: 04/07/2021 09:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: LOMONACO

DATE/TIME OF TEST: 03/20/2021 21:11:53

Diagnostics Check	OK	21:05
Air Blank	0.000	21:05
Control Test	0.078	21:06
Air Blank	0.000	21:06
Subject Sample # 1	0.079	21:07
Breath Volume	1.906	21:07
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.076	21:10
Breath Volume	2.039	21:10
Air Blank	0.000	21:10
Control Test	0.078	21:11
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

Data Download Date/Time: 04/07/2021 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 03/21/2021 02:57:16

Diagnostics Check		OK	02:51
Air Blank		0.000	02:51
Control Test		0.079	02:51
Air Blank		0.000	02:52
Subject Sample	# 1	0.154	02:52
Breath Volume		2.480	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.152	02:55
Breath Volume		2.406	02:55
Air Blank		0.000	02:56
Control Test		0.078	02:56
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 04/07/2021 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 05:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:23

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 03/21/2021 06:50:06

Diagnostics Check	OK	06:43
Air Blank	0.000	06:43
Control Test	0.078	06:43
Air Blank	0.000	06:44
Subject Sample # 1	0.116	06:45
Breath Volume	1.792	06:45
Air Blank	0.000	06:46
Air Blank	0.000	06:48
Subject Sample # 2	0.105	06:48
Breath Volume	1.480	06:48
Air Blank	0.000	06:49
Control Test	0.078	06:49
Air Blank	0.000	06:50
Diagnostics Check	OK	06:50

Data Download Date/Time: 04/07/2021 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CARROLL,COURTNEY
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 09:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: DYKAS

DATE/TIME OF TEST: 03/23/2021 23:08:54

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.078	23:03
Air Blank	0.000	23:03
Subject Sample # 1	0.118	23:04
Breath Volume	2.816	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:06
Subject Sample # 2	0.108	23:07
Breath Volume	3.757	23:07
Air Blank	0.000	23:07
Control Test	0.076	23:08
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 04/07/2021 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,ALEXANDRIA,N
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: SANTOS,CRISTIAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ALDERETE

DATE/TIME OF TEST: 03/24/2021 02:24:56

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.078	02:21	
Air Blank	0.000	02:22	
Subject Sample	# 1	02:23	Subject Test Refused
Breath Volume	0.000	02:23	
Air Blank	0.000	02:24	
Control Test	0.079	02:24	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

Data Download Date/Time: 04/07/2021 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,JOSHUA,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HARRILL,ANDERSEN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: NERESON

DATE/TIME OF TEST: 03/27/2021 03:19:04

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.078	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.166	03:13
Breath Volume	1.152	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.166	03:17
Breath Volume	1.441	03:17
Air Blank	0.000	03:18
Control Test	0.078	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:19

Data Download Date/Time: 04/07/2021 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIDAL,NERELYS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: VIDAL,NERELYS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: MOTWANI

DATE/TIME OF TEST: 03/27/2021 20:50:50

Diagnostics Check	OK	20:47	
Air Blank	0.000	20:48	
Control Test	0.078	20:48	
Air Blank	0.000	20:49	
Subject Sample	# 1	20:49	Subject Test Refused
Breath Volume	0.000	20:49	
Air Blank	0.000	20:49	
Control Test	0.078	20:50	
Air Blank	0.000	20:50	
Diagnostics Check	OK	20:50	

Data Download Date/Time: 04/07/2021 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIDAL,NERELYS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: VIDAL,NERELYS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 03/27/2021 22:39:24

Diagnostics Check	OK	22:34	
Air Blank	0.000	22:34	
Control Test	0.078	22:35	
Air Blank	0.000	22:35	
Subject Sample	# 1	22:38	Subject Test Refused
Breath Volume	0.000	22:38	
Air Blank	0.000	22:38	
Control Test	0.079	22:38	
Air Blank	0.000	22:39	
Diagnostics Check	OK	22:39	

Data Download Date/Time: 04/07/2021 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CARROLL,COURTNEY
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: BUI
ARREST TIME: 08:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: MACKLIN

DATE/TIME OF TEST: 03/27/2021 23:10:51

Diagnostics Check	OK	23:04
Air Blank	0.000	23:05
Control Test	0.079	23:05
Air Blank	0.000	23:06
Subject Sample # 1	0.000	23:06
Breath Volume	3.023	23:06
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.000	23:09
Breath Volume	3.117	23:09
Air Blank	0.000	23:09
Control Test	0.078	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 04/07/2021 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,ALEXANDRIA,N
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HRADECKY,ALEXANDRIA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 03/29/2021 03:44:58

Diagnostics Check		OK	03:37
Air Blank		0.000	03:38
Control Test		0.078	03:38
Air Blank		0.000	03:39
Subject Sample	# 1	0.201	03:40
Breath Volume		1.917	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:42
Subject Sample	# 2	0.202	03:43
Breath Volume		1.750	03:43
Air Blank		0.000	03:44
Control Test		0.077	03:44
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 04/07/2021 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: JULA

DATE/TIME OF TEST: 03/29/2021 22:50:34

Diagnostics Check	OK	22:46	
Air Blank	0.000	22:47	
Control Test	0.080	22:47	
Air Blank	0.000	22:47	
Subject Sample	# 1	22:49	Subject Test Refused
Breath Volume	0.000	22:49	
Air Blank	0.000	22:49	
Control Test	0.079	22:49	
Air Blank	0.000	22:50	
Diagnostics Check	OK	22:50	

Data Download Date/Time: 04/07/2021 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/30/2021 00:55:25

Diagnostics Check		OK	00:49
Air Blank		0.000	00:49
Control Test		0.078	00:49
Air Blank		0.000	00:50
Subject Sample	# 1	0.191	00:50
Breath Volume		2.203	00:50
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.203	00:53
Breath Volume		2.054	00:53
Air Blank		0.000	00:54
Control Test		0.078	00:54
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 04/07/2021 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: NEHRBASS

DATE/TIME OF TEST: 03/31/2021 02:37:01

Diagnostics Check		OK	02:30
Air Blank		0.000	02:30
Control Test		0.078	02:31
Air Blank		0.000	02:31
Subject Sample	# 1	0.098	02:32
Breath Volume		1.457	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:34
Subject Sample	# 2	0.101	02:35
Breath Volume		1.824	02:35
Air Blank		0.000	02:36
Control Test		0.078	02:36
Air Blank		0.000	02:36
Diagnostics Check		OK	02:37

Data Download Date/Time: 04/07/2021 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGGAN,MARLON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOGGAN,MARLON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 03/02/2021 03:07:44

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.079	03:01
Air Blank		0.000	03:01
Subject Sample	# 1	0.262	03:03
Breath Volume		1.750	03:03
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.272	03:06
Breath Volume		2.109	03:06
Air Blank		0.000	03:06
Control Test		0.079	03:07
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 03/10/2021 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGGAN,MARLON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MARCUM,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: JOHANSEN

DATE/TIME OF TEST: 03/12/2021 01:41:12

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.179	01:36
Breath Volume	1.917	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.188	01:39
Breath Volume	2.101	01:39
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 04/17/2021 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGGAN,MARLON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOGGAN,MARLON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: K523804961090

DATE/TIME OF TEST: 03/12/2021 23:31:09

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.078	23:25
Air Blank	0.000	23:25
Subject Sample # 1	0.116	23:26
Breath Volume	2.082	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:28
Subject Sample # 2	0.111	23:29
Breath Volume	1.289	23:29
Air Blank	0.000	23:30
Control Test	0.078	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 04/17/2021 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGGAN,MARLON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOGGAN,MARLON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: ROBARDS

DATE/TIME OF TEST: 03/15/2021 23:29:54

Diagnostics Check	OK	23:23
Air Blank	0.000	23:24
Control Test	0.078	23:24
Air Blank	0.000	23:24
Subject Sample # 1	0.187	23:25
Breath Volume	4.531	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:27
Subject Sample # 2	0.192	23:28
Breath Volume	4.281	23:28
Air Blank	0.000	23:28
Control Test	0.077	23:29
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 04/17/2021 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FERGUSON,RYAN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: RAMIREZ CHAVEZ

DATE/TIME OF TEST: 03/01/2021 00:08:15

Diagnostics Check	OK	00:01
Air Blank	0.000	00:02
Control Test	0.080	00:02
Air Blank	0.000	00:02
Subject Sample # 1	0.217	00:03
Breath Volume	3.148	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.212	00:06
Breath Volume	3.214	00:06
Air Blank	0.000	00:07
Control Test	0.077	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 03/12/2021 17:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: PLATZER

DATE/TIME OF TEST: 03/01/2021 00:55:32

Diagnostics Check	OK	00:49
Air Blank	0.000	00:49
Control Test	0.078	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.000	00:50
Breath Volume	1.910	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.000	00:54
Breath Volume	3.042	00:54
Air Blank	0.000	00:54
Control Test	0.078	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 03/12/2021 17:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SUPER,MICHAEL
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 03/02/2021 03:59:20

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.078	03:52
Air Blank		0.000	03:53
Subject Sample	# 1	0.105	03:54
Breath Volume		2.222	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:56
Subject Sample	# 2	0.096	03:57
Breath Volume		1.316	03:57
Air Blank		0.000	03:58
Control Test		0.077	03:58
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 03/12/2021 17:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HOWARD,EDWARD
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: ARNONE

DATE/TIME OF TEST: 03/03/2021 23:25:54

Diagnostics Check		OK	23:18
Air Blank		0.000	23:18
Control Test		0.079	23:19
Air Blank		0.000	23:19
Subject Sample	# 1	0.000	23:20
Breath Volume		1.277	23:20
Air Blank		0.000	23:21
Air Blank		0.000	23:23
Subject Sample	# 2	0.000	23:24
Breath Volume		1.597	23:24
Air Blank		0.000	23:24
Control Test		0.079	23:25
Air Blank		0.000	23:25
Diagnostics Check		OK	23:25

Data Download Date/Time: 03/12/2021 17:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SMITH,MARK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: WOLTERS

DATE/TIME OF TEST: 03/04/2021 00:17:59

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.079	00:12
Air Blank	0.000	00:12
Subject Sample # 1	0.192	00:13
Breath Volume	2.082	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:15
Subject Sample # 2	0.192	00:16
Breath Volume	1.875	00:16
Air Blank	0.000	00:17
Control Test	0.078	00:17
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 03/12/2021 17:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RODRIGUEZ,ELIAN
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: CURD

DATE/TIME OF TEST: 03/04/2021 00:43:16

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:39	
Control Test	0.078	00:40	
Air Blank	0.000	00:40	
Subject Sample	# 1	00:41	Subject Test Refused
Breath Volume	0.000	00:41	
Air Blank	0.000	00:42	
Control Test	0.078	00:42	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

Data Download Date/Time: 03/12/2021 17:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: GONZALEZ PABLO

DATE/TIME OF TEST: 03/05/2021 01:04:58

Diagnostics Check		OK	00:57
Air Blank		0.000	00:58
Control Test		0.079	00:58
Air Blank		0.000	00:59
Subject Sample	# 1	0.125	01:00
Breath Volume		3.632	01:00
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.124	01:03
Breath Volume		2.683	01:03
Air Blank		0.000	01:04
Control Test		0.078	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 03/12/2021 17:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DANDREA,ADAM
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: LAMAR

DATE/TIME OF TEST: 03/05/2021 17:47:06

Diagnostics Check	OK	17:37	
Air Blank	0.000	17:37	
Control Test	0.079	17:38	
Air Blank	0.000	17:38	
Subject Sample	# 1 0.000	17:41	Volume Not Met
Breath Volume	0.453	17:41	
Air Blank	0.000	17:42	
Air Blank	0.000	17:44	
Subject Sample	# 2	17:45	Subject Test Refused
Breath Volume	0.105	17:45	
Air Blank	0.000	17:46	
Control Test	0.078	17:46	
Air Blank	0.000	17:47	
Diagnostics Check	OK	17:47	

Data Download Date/Time: 03/12/2021 17:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOCHUM,CRAIG
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: LLOYD

DATE/TIME OF TEST: 03/06/2021 00:41:03

Diagnostics Check		OK	00:34
Air Blank		0.000	00:34
Control Test		0.079	00:35
Air Blank		0.000	00:35
Subject Sample	# 1	0.374	00:36
Breath Volume		1.906	00:36
Air Blank		0.000	00:37
Air Blank		0.000	00:38
Subject Sample	# 2	0.370	00:39
Breath Volume		1.675	00:39
Air Blank		0.000	00:40
Control Test		0.077	00:40
Air Blank		0.000	00:40
Diagnostics Check		OK	00:41

Data Download Date/Time: 03/12/2021 17:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: PINZON MEDINA

DATE/TIME OF TEST: 03/06/2021 02:27:48

Diagnostics Check		OK	02:19
Air Blank		0.000	02:20
Control Test		0.079	02:20
Air Blank		0.000	02:21
Subject Sample	# 1	0.193	02:23
Breath Volume		2.000	02:23
Air Blank		0.000	02:24
Air Blank		0.000	02:25
Subject Sample	# 2	0.176	02:26
Breath Volume		1.984	02:26
Air Blank		0.000	02:26
Control Test		0.076	02:27
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 03/12/2021 17:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KUEBLER,COLE
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: MATERA

DATE/TIME OF TEST: 03/06/2021 19:57:58

Diagnostics Check		OK	19:50
Air Blank		0.000	19:51
Control Test		0.079	19:51
Air Blank		0.000	19:52
Subject Sample	# 1	0.189	19:52
Breath Volume		2.210	19:52
Air Blank		0.000	19:53
Air Blank		0.000	19:55
Subject Sample	# 2	0.186	19:56
Breath Volume		2.660	19:56
Air Blank		0.000	19:57
Control Test		0.078	19:57
Air Blank		0.000	19:57
Diagnostics Check		OK	19:57

Data Download Date/Time: 03/12/2021 17:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SOLOWAY,ALAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: GULINO

DATE/TIME OF TEST: 03/07/2021 04:12:17

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.079	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.296	04:07
Breath Volume		2.640	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.297	04:10
Breath Volume		2.609	04:10
Air Blank		0.000	04:11
Control Test		0.079	04:11
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 03/12/2021 17:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WEBER,RONALD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:43

SUBJECT'S LAST NAME: POTTS

DATE/TIME OF TEST: 03/07/2021 08:22:30

Diagnostics Check	OK	08:15
Air Blank	0.000	08:15
Control Test	0.079	08:15
Air Blank	0.000	08:16
Subject Sample # 1	0.102	08:17
Breath Volume	2.335	08:17
Air Blank	0.000	08:18
Air Blank	0.000	08:19
Subject Sample # 2	0.099	08:20
Breath Volume	2.335	08:20
Air Blank	0.000	08:21
Control Test	0.078	08:21
Air Blank	0.000	08:22
Diagnostics Check	OK	08:22

Data Download Date/Time: 03/12/2021 17:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXANDRE,DOMINIC
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: PELCHAT

DATE/TIME OF TEST: 03/07/2021 23:39:28

Diagnostics Check	OK	23:32
Air Blank	0.000	23:33
Control Test	0.080	23:33
Air Blank	0.000	23:34
Subject Sample # 1	0.174	23:34
Breath Volume	1.203	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.172	23:37
Breath Volume	1.410	23:37
Air Blank	0.000	23:38
Control Test	0.079	23:38
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 03/12/2021 17:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: ALVAREZ RAMIREZ

DATE/TIME OF TEST: 03/08/2021 01:55:48

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.079	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.168	01:50
Breath Volume	1.335	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:52
Subject Sample # 2	0.163	01:54
Breath Volume	1.328	01:54
Air Blank	0.000	01:54
Control Test	0.077	01:55
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 03/16/2021 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: MCCLOSKEY

DATE/TIME OF TEST: 03/08/2021 22:15:19

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.080	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.000	22:10
Breath Volume	2.394	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.000	22:13
Breath Volume	2.078	22:13
Air Blank	0.000	22:14
Control Test	0.079	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 03/16/2021 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/09/2021 00:50:42

Diagnostics Check		OK	00:42	
Air Blank		0.000	00:42	
Control Test		0.078	00:42	
Air Blank		0.000	00:43	
Subject Sample	# 1	0.234	00:46	Volume Not Met
Breath Volume		0.128	00:46	
Air Blank		0.000	00:47	
Air Blank		0.000	00:49	
Subject Sample	# 2		00:49	Subject Test Refused
Breath Volume		0.000	00:49	
Air Blank		0.000	00:49	
Control Test		0.075	00:50	
Air Blank		0.000	00:50	
Diagnostics Check		OK	00:50	

Data Download Date/Time: 03/16/2021 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BALDINO,GEORGE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 11:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:10

SUBJECT'S LAST NAME: GERMANN

DATE/TIME OF TEST: 03/11/2021 12:52:09

Diagnostics Check	OK	12:45
Air Blank	0.000	12:45
Control Test	0.079	12:46
Air Blank	0.000	12:46
Subject Sample # 1	0.172	12:47
Breath Volume	2.574	12:47
Air Blank	0.000	12:48
Air Blank	0.000	12:49
Subject Sample # 2	0.174	12:50
Breath Volume	2.449	12:50
Air Blank	0.000	12:51
Control Test	0.078	12:51
Air Blank	0.000	12:52
Diagnostics Check	OK	12:52

Data Download Date/Time: 03/16/2021 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCGILLICUDDY,STEVEN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: HUFF

DATE/TIME OF TEST: 03/11/2021 21:05:32

Diagnostics Check	OK	20:58
Air Blank	0.000	20:59
Control Test	0.079	20:59
Air Blank	0.000	21:00
Subject Sample # 1	0.000	21:01
Breath Volume	2.113	21:01
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample # 2	0.000	21:04
Breath Volume	2.046	21:04
Air Blank	0.000	21:04
Control Test	0.078	21:04
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 03/16/2021 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WOOD,DERRICK
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: CROCE

DATE/TIME OF TEST: 03/11/2021 22:42:30

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.079	22:37
Air Blank	0.000	22:37
Subject Sample # 1	0.000	22:38
Breath Volume	2.886	22:38
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.000	22:41
Breath Volume	1.812	22:41
Air Blank	0.000	22:41
Control Test	0.078	22:41
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 03/16/2021 14:46