

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,JUSTIN,J
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: GONZALEZ,JUSTIN
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 03/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/11/2021 21:02:56

Diagnostics Check	OK	20:53
Air Blank	0.000	20:53
Control Test	0.080	20:54
Air Blank	0.000	20:54
Subject Sample # 1	0.181	20:55
Breath Volume	1.406	20:55
Air Blank	0.000	20:56
Air Blank	0.000	20:57
Subject Sample # 2	0.207	20:58
Breath Volume	1.574	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:00
Subject Sample # 3	0.203	21:01
Breath Volume	1.636	21:01
Air Blank	0.000	21:02
Control Test	0.079	21:02
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 04/21/2021 15:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDGEPEETH,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: HEDGEPEETH,TIMOTHY
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: BUI
ARREST TIME: 19:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 04/15/2021 21:14:49

Diagnostics Check	OK	21:11	
Air Blank	0.000	21:12	
Control Test	0.081	21:12	
Air Blank	0.000	21:13	
Subject Sample	# 1	21:13	Subject Test Refused
Breath Volume	0.000	21:13	
Air Blank	0.000	21:13	
Control Test	0.081	21:14	
Air Blank	0.000	21:14	
Diagnostics Check	OK	21:14	

Data Download Date/Time: 04/21/2021 15:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,CRISTINO,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: CABRAL,EDUARDO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 03/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: MONTELONGO

DATE/TIME OF TEST: 04/17/2021 00:36:46

Diagnostics Check	OK	00:33	
Air Blank	0.000	00:34	
Control Test	0.080	00:34	
Air Blank	0.000	00:35	
Subject Sample	# 1	00:35	Subject Test Refused
Breath Volume	0.000	00:35	
Air Blank	0.000	00:35	
Control Test	0.080	00:36	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time: 04/21/2021 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,CRISTINO,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: CABELLO,OSWALDO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 04/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: MUSTAFOYSKI

DATE/TIME OF TEST: 04/21/2021 19:36:23

Diagnostics Check	OK	19:29
Air Blank	0.000	19:30
Control Test	0.081	19:30
Air Blank	0.000	19:31
Subject Sample # 1	0.000	19:31
Breath Volume	3.441	19:31
Air Blank	0.000	19:32
Air Blank	0.000	19:34
Subject Sample # 2	0.000	19:35
Breath Volume	3.351	19:35
Air Blank	0.000	19:35
Control Test	0.080	19:35
Air Blank	0.000	19:36
Diagnostics Check	OK	19:36

Data Download Date/Time: 05/24/2021 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,CRISTINO,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: SISNEY,WHITNEY
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 04/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 14:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:37

SUBJECT'S LAST NAME: ESCOBAR ORTEZ

DATE/TIME OF TEST: 04/23/2021 16:09:09

Diagnostics Check	OK	16:00	
Air Blank	0.000	16:01	
Control Test	0.080	16:01	
Air Blank	0.000	16:02	
Subject Sample # 1	0.222	16:03	
Breath Volume	2.347	16:03	
Air Blank	0.000	16:04	
Air Blank	0.000	16:05	
Subject Sample # 2		16:07	Subject Test Refused
Breath Volume	0.839	16:07	
Air Blank	0.000	16:08	
Control Test	0.078	16:08	
Air Blank	0.000	16:09	
Diagnostics Check	OK	16:09	

Data Download Date/Time: 05/24/2021 15:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,RYAN,M
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: PINO,A
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 04/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: RODRIGUEZ MIGUEL

DATE/TIME OF TEST: 04/25/2021 05:17:08

Diagnostics Check	OK	05:14	
Air Blank	0.000	05:14	
Control Test	0.080	05:15	
Air Blank	0.000	05:15	
Subject Sample	# 1	05:15	Subject Test Refused
Breath Volume	0.000	05:15	
Air Blank	0.000	05:16	
Control Test	0.081	05:16	
Air Blank	0.000	05:17	
Diagnostics Check	OK	05:17	

Data Download Date/Time: 05/24/2021 15:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OPRY, WILLIAM, J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DILLARD, PAUL
ARRESTING OFFICER AGENCY: Calhoun County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: BOZEMAN

DATE/TIME OF TEST: 04/02/2021 00:59:55

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:58	
Control Test	0.081	00:58	
Air Blank	0.000	00:59	
Subject Sample	# 1	00:59	RFI Detect
Breath Volume	0.000	00:59	
Air Blank		00:59	RFI Detect
Air Blank		00:59	RFI Detect

Data Download Date/Time: 04/13/2021 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OPRY, WILLIAM, J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DILLARD, PAUL
ARRESTING OFFICER AGENCY: Calhoun County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: BOZEMAN

DATE/TIME OF TEST: 04/02/2021 01:09:28

Diagnostics Check		OK	01:03
Air Blank		0.000	01:03
Control Test		0.081	01:03
Air Blank		0.000	01:04
Subject Sample	# 1	0.143	01:05
Breath Volume		2.441	01:05
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.133	01:07
Breath Volume		1.472	01:07
Air Blank		0.000	01:08
Control Test		0.080	01:08
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 04/13/2021 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: MCLENDON

DATE/TIME OF TEST: 04/18/2021 04:20:19

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:12	
Control Test	0.081	04:12	
Air Blank	0.000	04:13	
Subject Sample	# 1	04:16	No Sample Provided
Breath Volume	0.000	04:16	
Air Blank	0.000	04:16	
Air Blank	0.000	04:18	
Subject Sample	# 2	04:18	Subject Test Refused
Breath Volume	0.000	04:18	
Air Blank	0.000	04:19	
Control Test	0.080	04:19	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:20	

Data Download Date/Time: 05/12/2021 14:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CASTILLO,CHRISTIAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:34

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 04/11/2021 18:03:32

Diagnostics Check	OK	17:54
Air Blank	0.000	17:54
Control Test	0.080	17:55
Air Blank	0.000	17:55
Subject Sample # 1	0.211	17:58
Breath Volume	2.953	17:58
Air Blank	0.000	17:59
Air Blank	0.000	18:01
Subject Sample # 2	0.215	18:01
Breath Volume	1.371	18:01
Air Blank	0.000	18:02
Control Test	0.079	18:02
Air Blank	0.000	18:03
Diagnostics Check	OK	18:03

Data Download Date/Time: 04/26/2021 16:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CASTILLO,CHRISTIAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 04/14/2021 23:51:19

Diagnostics Check	OK	23:39	
Air Blank	0.000	23:40	
Control Test	0.079	23:40	
Air Blank	0.000	23:41	
Subject Sample	# 1	23:44	No Sample Provided
Breath Volume	0.000	23:44	
Air Blank	0.000	23:44	
Air Blank	0.000	23:46	
Subject Sample	# 2	23:49	No Sample Provided
Breath Volume	0.000	23:49	
Air Blank	0.000	23:50	
Control Test	0.079	23:50	
Air Blank	0.000	23:51	
Diagnostics Check	OK	23:51	

Data Download Date/Time: 04/26/2021 16:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CASTILLO,CHRISTIAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: CUADRA

DATE/TIME OF TEST: 04/18/2021 20:15:54

Diagnostics Check	OK	20:10	
Air Blank	0.000	20:10	
Control Test	0.080	20:10	
Air Blank	0.000	20:11	
Subject Sample	# 1 0.237	20:14	Volume Not Met
Breath Volume	0.070	20:14	
Air Blank		20:15	Purge Fail
Air Blank	0.000	20:15	

Data Download Date/Time: 04/26/2021 16:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CASTILLO,CHRISTIAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: CUADRA

DATE/TIME OF TEST: 04/18/2021 20:28:33

Diagnostics Check		OK	20:18
Air Blank		0.000	20:18
Control Test		0.079	20:18
Air Blank		0.000	20:19
Subject Sample	# 1	0.257	20:20
Breath Volume		1.695	20:20
Air Blank		0.000	20:21
Air Blank		0.000	20:22
Subject Sample	# 2	0.227	20:23
Breath Volume		1.929	20:23
Air Blank		0.000	20:24
Air Blank		0.000	20:26
Subject Sample	# 3	0.233	20:26
Breath Volume		1.292	20:26
Air Blank		0.000	20:27
Control Test		0.079	20:27
Air Blank		0.000	20:28
Diagnostics Check		OK	20:28

Data Download Date/Time: 04/26/2021 16:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NADAL,JOSE,V
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 04/30/2021 15:33:07

Diagnostics Check	OK	15:25
Air Blank	0.000	15:26
Control Test	0.078	15:26
Air Blank	0.000	15:27
Subject Sample # 1	0.267	15:27
Breath Volume	1.828	15:27
Air Blank	0.000	15:28
Air Blank	0.000	15:30
Subject Sample # 2	0.268	15:31
Breath Volume	2.488	15:31
Air Blank	0.000	15:32
Control Test	0.079	15:32
Air Blank	0.000	15:33
Diagnostics Check	OK	15:33

Data Download Date/Time: 05/18/2021 06:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWNING,CADEN,T
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DOWNING,CADEN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: LIAINA

DATE/TIME OF TEST: 04/03/2021 19:38:42

Diagnostics Check	OK	19:31
Air Blank	0.000	19:32
Control Test	0.081	19:32
Air Blank	0.000	19:33
Subject Sample # 1	0.035	19:34
Breath Volume	1.496	19:34
Air Blank	0.000	19:34
Air Blank	0.000	19:36
Subject Sample # 2	0.037	19:37
Breath Volume	1.714	19:37
Air Blank	0.000	19:37
Control Test	0.082	19:38
Air Blank	0.000	19:38
Diagnostics Check	OK	19:38

Data Download Date/Time: 05/25/2021 19:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1206004
BREATH TEST OPERATOR: CORBITT,KYLE,L	DRY GAS EXPIRY DATE: 09/20/2022
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: BUI
ARRESTING OFFICER: WILLIAMS,DANIEL	ARREST TIME: 18:07
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 04/01/2021	OBSERVATION PERIOD BEGAN: 18:07

SUBJECT'S LAST NAME: PERDOMO VELASQUEZ DATE/TIME OF TEST: 04/11/2021 18:42:13

Diagnostics Check	OK	18:35
Air Blank	0.000	18:35
Control Test	0.080	18:35
Air Blank	0.000	18:36
Subject Sample # 1	0.193	18:36
Breath Volume	1.996	18:36
Air Blank	0.000	18:37
Air Blank	0.000	18:39
Subject Sample # 2	0.196	18:40
Breath Volume	1.722	18:40
Air Blank	0.000	18:41
Control Test	0.080	18:41
Air Blank	0.000	18:42
Diagnostics Check	OK	18:42

Data Download Date/Time: 05/25/2021 19:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALTAIS,ANDREW,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DOWNING,CADEN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 16:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: HASTY

DATE/TIME OF TEST: 04/18/2021 16:52:44

Diagnostics Check	OK	16:49	
Air Blank	0.000	16:50	
Control Test	0.081	16:50	
Air Blank	0.000	16:50	
Subject Sample	# 1	16:51	Subject Test Refused
Breath Volume	0.000	16:51	
Air Blank	0.000	16:51	
Control Test	0.081	16:52	
Air Blank	0.000	16:52	
Diagnostics Check	OK	16:52	

Data Download Date/Time: 05/25/2021 19:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 17:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: BABAUTA

DATE/TIME OF TEST: 04/23/2021 18:06:58

Diagnostics Check	OK	18:00
Air Blank	0.000	18:01
Control Test	0.081	18:01
Air Blank	0.000	18:02
Subject Sample # 1	0.000	18:02
Breath Volume	3.863	18:02
Air Blank	0.000	18:03
Air Blank	0.000	18:05
Subject Sample # 2	0.000	18:05
Breath Volume	2.609	18:05
Air Blank	0.000	18:06
Control Test	0.079	18:06
Air Blank	0.000	18:06
Diagnostics Check	OK	18:06

Data Download Date/Time: 05/25/2021 19:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONOVAN,CHARLES,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DONOVAN,CHARLES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 18:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: TOUPS
DATE/TIME OF TEST: 04/25/2021 18:47:20

Diagnostics Check	OK	18:41
Air Blank	0.000	18:41
Control Test	0.080	18:41
Air Blank	0.000	18:42
Subject Sample # 1	0.111	18:42
Breath Volume	2.257	18:42
Air Blank	0.000	18:43
Air Blank	0.000	18:45
Subject Sample # 2	0.111	18:45
Breath Volume	1.945	18:45
Air Blank	0.000	18:46
Control Test	0.080	18:46
Air Blank	0.000	18:47
Diagnostics Check	OK	18:47

Data Download Date/Time: 05/25/2021 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DOWNING,CADEN,T

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: DOWNING,CADEN

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1206004

DRY GAS EXPIRY DATE: 09/20/2022

VIOLATION CODE: BUI

ARREST TIME: 14:48

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:38

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/26/2021 17:09:30

Diagnostics Check	OK	17:01
Air Blank	0.000	17:02
Control Test	0.081	17:02
Air Blank	0.000	17:03
Subject Sample # 1	0.000	17:04
Breath Volume	4.589	17:04
Air Blank	0.000	17:05
Air Blank	0.000	17:07
Subject Sample # 2	0.000	17:08
Breath Volume	4.351	17:08
Air Blank	0.000	17:08
Control Test	0.079	17:08
Air Blank	0.000	17:09
Diagnostics Check	OK	17:09

Data Download Date/Time: 05/25/2021 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANEY, MICHAEL, P
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: HANEY, MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 17:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: MCKINNEY
DATE/TIME OF TEST: 04/30/2021 17:54:26

Diagnostics Check	OK	17:51	
Air Blank	0.000	17:52	
Control Test	0.081	17:52	
Air Blank	0.000	17:52	
Subject Sample	# 1	17:53	Subject Test Refused
Breath Volume	0.000	17:53	
Air Blank	0.000	17:53	
Control Test	0.080	17:53	
Air Blank	0.000	17:54	
Diagnostics Check	OK	17:54	

Data Download Date/Time: 05/25/2021 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAHR,DANIEL,E
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: ARBOGAST,JOSH
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 17:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: PEERSON

DATE/TIME OF TEST: 04/30/2021 18:05:18

Diagnostics Check	OK	18:01	
Air Blank	0.000	18:01	
Control Test	0.080	18:02	
Air Blank	0.000	18:02	
Subject Sample	# 1	18:03	Subject Test Refused
Breath Volume	0.000	18:03	
Air Blank	0.000	18:04	
Control Test	0.081	18:04	
Air Blank	0.000	18:05	
Diagnostics Check	OK	18:05	

Data Download Date/Time: 05/25/2021 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MURSCH,THEODORE,R

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: MURSCH,THEODORE

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1206004

DRY GAS EXPIRY DATE: 09/20/2022

VIOLATION CODE: BUI

ARREST TIME: 17:55

GENDER: Female

OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: CASTRO-GARCIA

DATE/TIME OF TEST: 04/30/2021 19:07:35

Diagnostics Check	OK	18:55		
Air Blank	0.000	18:56		
Control Test	0.080	18:56		
Air Blank	0.000	18:57		
Subject Sample	# 1	19:00	No Sample Provided	
Breath Volume	0.000	19:00		
Air Blank	0.000	19:00		
Air Blank	0.000	19:02		
Subject Sample	# 2	0.102	19:05	Volume Not Met
Breath Volume	0.082	19:05		
Air Blank	0.000	19:06		
Control Test	0.081	19:06		
Air Blank	0.000	19:07		
Diagnostics Check	OK	19:07		

Data Download Date/Time: 05/25/2021 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURSCH,THEODORE,R
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MURSCH,THEODORE
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 17:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 04/30/2021 19:25:32

Diagnostics Check	OK	19:14
Air Blank	0.000	19:14
Control Test	0.080	19:15
Air Blank	0.000	19:15
Subject Sample # 1	0.117	19:18
Breath Volume	1.359	19:18
Air Blank	0.000	19:19
Air Blank	0.000	19:21
Subject Sample # 2	0.105	19:23
Breath Volume	1.855	19:23
Air Blank	0.000	19:24
Control Test	0.076	19:24
Air Blank	0.000	19:25
Diagnostics Check	OK	19:25

Data Download Date/Time: 05/25/2021 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARROQUIN,RICHARD,J
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MARROQUIN,RICHARD
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 04/30/2021 19:36:51

Diagnostics Check	OK	19:30
Air Blank	0.000	19:30
Control Test	0.079	19:31
Air Blank	0.000	19:31
Subject Sample # 1	0.100	19:32
Breath Volume	2.402	19:32
Air Blank	0.000	19:32
Air Blank	0.000	19:34
Subject Sample # 2	0.097	19:35
Breath Volume	2.035	19:35
Air Blank	0.000	19:35
Control Test	0.079	19:36
Air Blank	0.000	19:36
Diagnostics Check	OK	19:36

Data Download Date/Time: 05/25/2021 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO:	FWC	Date or Time Changed:	No
SOFTWARE VERSION:	8100.27	DRY GAS STANDARD LOT #:	1060911
BREATH TEST OPERATOR:	ALSOBROOKS,KEELIN,G	DRY GAS EXPIRY DATE:	12/21/2021
BREATH TEST OPERATOR AGENCY:	FFWCC (Panama City)	VIOLATION CODE:	DUI
ARRESTING OFFICER:	ALSOBROOKS,KEELIN	ARREST TIME:	18:30
ARRESTING OFFICER AGENCY:	FFWCC (Panama City)	GENDER:	Male
LAST AGENCY INSPECTION:	04/01/2021	OBSERVATION PERIOD BEGAN:	18:35

SUBJECT'S LAST NAME: GARCIA DATE/TIME OF TEST: 04/03/2021 19:02:22

Diagnostics Check	OK	18:56
Air Blank	0.000	18:56
Control Test	0.081	18:56
Air Blank	0.000	18:57
Subject Sample # 1	0.000	18:57
Breath Volume	3.441	18:57
Air Blank	0.000	18:58
Air Blank	0.000	19:00
Subject Sample # 2	0.000	19:01
Breath Volume	1.496	19:01
Air Blank	0.000	19:01
Control Test	0.080	19:01
Air Blank	0.000	19:02
Diagnostics Check	OK	19:02

Data Download Date/Time: 05/25/2021 19:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,MALACHI,C
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: ROTHCHILD,FOREST
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 03/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: BUI
ARREST TIME: 14:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 04/17/2021 14:57:52

Diagnostics Check	OK	14:55	
Air Blank	0.000	14:55	
Control Test	0.079	14:55	
Air Blank	0.000	14:56	
Subject Sample	# 1	14:56	Subject Test Refused
Breath Volume	0.000	14:56	
Air Blank	0.000	14:56	
Control Test	0.077	14:57	
Air Blank	0.000	14:57	
Diagnostics Check	OK	14:57	

Data Download Date/Time: 04/21/2021 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,MALACHI,C
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: HOPPE,AUSTIN
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 03/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: BUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:52

SUBJECT'S LAST NAME: CASSIDY

DATE/TIME OF TEST: 04/17/2021 18:34:15

Diagnostics Check	OK	18:23	
Air Blank	0.000	18:23	
Control Test	0.079	18:24	
Air Blank	0.000	18:24	
Subject Sample	# 1	18:27	No Sample Provided
Breath Volume	0.000	18:27	
Air Blank	0.000	18:28	
Air Blank	0.000	18:30	
Subject Sample	# 2	18:33	No Sample Provided
Breath Volume	0.000	18:33	
Air Blank		18:33	RFI Detect
Air Blank	0.000	18:34	

Data Download Date/Time: 04/21/2021 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,MALACHI,C
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: HOPPE,ZACH
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 03/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: BUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: KEEL

DATE/TIME OF TEST: 04/17/2021 19:35:09

Diagnostics Check	OK	19:23	
Air Blank	0.000	19:24	
Control Test	0.077	19:24	
Air Blank	0.000	19:25	
Subject Sample	# 1	19:28	No Sample Provided
Breath Volume	0.000	19:28	
Air Blank	0.000	19:28	
Air Blank	0.000	19:30	
Subject Sample	# 2	19:33	No Sample Provided
Breath Volume	0.000	19:33	
Air Blank	0.000	19:34	
Control Test	0.078	19:34	
Air Blank	0.000	19:35	
Diagnostics Check	OK	19:35	

Data Download Date/Time: 04/21/2021 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,MALACHI,C
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: WILKINS,MALACHI
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 03/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: BUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: ORTEGA

DATE/TIME OF TEST: 04/17/2021 21:46:08

Diagnostics Check	OK	21:39
Air Blank	0.000	21:39
Control Test	0.077	21:40
Air Blank	0.000	21:40
Subject Sample # 1	0.025	21:41
Breath Volume	1.390	21:41
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.027	21:44
Breath Volume	3.140	21:44
Air Blank	0.000	21:45
Control Test	0.077	21:45
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 04/21/2021 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,MALACHI,C
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: HOPPE,ZACH
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 03/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: BUI
ARREST TIME: 16:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: ROARK

DATE/TIME OF TEST: 04/18/2021 17:09:25

Diagnostics Check	OK	16:58	
Air Blank	0.000	16:58	
Control Test	0.078	16:58	
Air Blank	0.000	16:59	
Subject Sample	# 1	17:02	No Sample Provided
Breath Volume	0.000	17:02	
Air Blank	0.000	17:02	
Air Blank	0.000	17:04	
Subject Sample	# 2	17:08	No Sample Provided
Breath Volume	0.000	17:08	
Air Blank	0.000	17:08	
Control Test	0.077	17:08	
Air Blank	0.000	17:09	
Diagnostics Check	OK	17:09	

Data Download Date/Time: 04/21/2021 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,MALACHI,C
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: BERG,TIM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: BUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: WHITING

DATE/TIME OF TEST: 04/18/2021 20:44:09

Diagnostics Check	OK	20:40	
Air Blank	0.000	20:40	
Control Test	0.078	20:40	
Air Blank	0.000	20:41	
Subject Sample	# 1	20:42	Subject Test Refused
Breath Volume	0.000	20:42	
Air Blank	0.000	20:43	
Control Test	0.077	20:43	
Air Blank	0.000	20:44	
Diagnostics Check	OK	20:44	

Data Download Date/Time: 04/21/2021 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHT,JAMES,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GONZALEZ,JOEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: DE PAZ RIZO

DATE/TIME OF TEST: 04/20/2021 05:03:17

Diagnostics Check		OK	04:56
Air Blank		0.000	04:56
Control Test		0.079	04:56
Air Blank		0.000	04:57
Subject Sample	# 1	0.171	04:58
Breath Volume		1.453	04:58
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.158	05:01
Breath Volume		1.785	05:01
Air Blank		0.000	05:02
Control Test		0.077	05:02
Air Blank		0.000	05:03
Diagnostics Check		OK	05:03

Data Download Date/Time: 05/13/2021 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 03/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: AVILA

DATE/TIME OF TEST: 04/23/2021 06:07:56

Diagnostics Check	OK	06:01
Air Blank	0.000	06:02
Control Test	0.079	06:02
Air Blank	0.000	06:03
Subject Sample # 1	0.212	06:03
Breath Volume	1.160	06:03
Air Blank	0.000	06:04
Air Blank	0.000	06:05
Subject Sample # 2	0.220	06:06
Breath Volume	1.972	06:06
Air Blank	0.000	06:07
Control Test	0.078	06:07
Air Blank	0.000	06:07
Diagnostics Check	OK	06:07

Data Download Date/Time: 04/29/2021 19:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 03/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: MURDOCK

DATE/TIME OF TEST: 04/25/2021 02:48:03

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:39	
Control Test	0.078	02:39	
Air Blank	0.000	02:40	
Subject Sample	# 1	02:43	No Sample Provided
Breath Volume	0.000	02:43	
Air Blank	0.000	02:43	
Air Blank	0.000	02:45	
Subject Sample	# 2	02:46	Subject Test Refused
Breath Volume	0.000	02:46	
Air Blank	0.000	02:47	
Control Test	0.079	02:47	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:48	

Data Download Date/Time: 04/29/2021 19:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,JULIAN,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: HARVEY,CRAIG
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 03/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:28

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 04/25/2021 06:52:02

Diagnostics Check	OK	06:49	
Air Blank	0.000	06:49	
Control Test	0.078	06:49	
Air Blank	0.000	06:50	
Subject Sample	# 1	06:50	Subject Test Refused
Breath Volume	0.000	06:50	
Air Blank	0.000	06:51	
Control Test	0.078	06:51	
Air Blank	0.000	06:51	
Diagnostics Check	OK	06:52	

Data Download Date/Time: 04/29/2021 19:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/30/2021 03:14:27

Diagnostics Check		OK	03:08
Air Blank		0.000	03:08
Control Test		0.079	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.196	03:09
Breath Volume		2.195	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.185	03:12
Breath Volume		1.937	03:12
Air Blank		0.000	03:13
Control Test		0.078	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 05/30/2021 20:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MILLER,K
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: ESPOSITO

DATE/TIME OF TEST: 04/01/2021 01:09:59

Diagnostics Check	OK	01:03
Air Blank	0.000	01:04
Control Test	0.079	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.193	01:05
Breath Volume	3.902	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.191	01:08
Breath Volume	3.812	01:08
Air Blank	0.000	01:09
Control Test	0.079	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 04/13/2021 08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATSON,ANDRAE,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BORCHERS,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: ZALETEL

DATE/TIME OF TEST: 04/01/2021 18:09:22

Diagnostics Check	OK	18:03	
Air Blank	0.000	18:04	
Control Test	0.079	18:04	
Air Blank	0.000	18:05	
Subject Sample # 1	0.000	18:05	
Breath Volume	2.605	18:05	
Air Blank	0.000	18:06	
Air Blank	0.000	18:08	
Subject Sample # 2	0.000	18:08	
Breath Volume	2.808	18:08	
Air Blank		18:08	RFI Detect
Air Blank	0.000	18:09	

Data Download Date/Time: 04/13/2021 08:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER JR,SCOTT,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: FOLEY,K
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: RESTO JR

DATE/TIME OF TEST: 04/01/2021 23:43:00

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.079	23:37
Air Blank	0.000	23:37
Subject Sample # 1	0.237	23:38
Breath Volume	2.812	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:40
Subject Sample # 2	0.230	23:41
Breath Volume	3.089	23:41
Air Blank	0.000	23:42
Control Test	0.078	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:43

Data Download Date/Time: 04/13/2021 08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,M,C
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: WILLIAMS,ZACHARY
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: MEDEROGONZALEZ

DATE/TIME OF TEST: 04/04/2021 04:01:04

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.079	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.167	03:52
Breath Volume	1.519	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.189	03:56
Breath Volume	1.382	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 3	0.184	03:59
Breath Volume	2.054	03:59
Air Blank	0.000	04:00
Control Test	0.078	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:01

Data Download Date/Time: 04/13/2021 08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: KENDALL

DATE/TIME OF TEST: 04/04/2021 04:12:26

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.124	04:07
Breath Volume	2.878	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.122	04:10
Breath Volume	2.285	04:10
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 04/13/2021 08:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLMSTEAD,J,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ZYNDA,T
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BARRAVECCHIO

DATE/TIME OF TEST: 04/05/2021 00:50:37

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.079	00:43
Air Blank	0.000	00:44
Subject Sample # 1	0.048	00:45
Breath Volume	1.414	00:45
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.043	00:49
Breath Volume	1.472	00:49
Air Blank	0.000	00:49
Control Test	0.079	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 04/13/2021 08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,THOMAS,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: WALBERT,A
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: CESARIO

DATE/TIME OF TEST: 04/07/2021 01:35:16

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.079	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.230	01:30
Breath Volume	3.382	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.228	01:33
Breath Volume	2.507	01:33
Air Blank	0.000	01:34
Control Test	0.078	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 04/13/2021 08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JOSE,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CISNERO,JOHN
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:39

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 04/14/2021 16:08:26

Diagnostics Check	OK	16:01
Air Blank	0.000	16:02
Control Test	0.079	16:02
Air Blank	0.000	16:03
Subject Sample # 1	0.217	16:03
Breath Volume	2.378	16:03
Air Blank	0.000	16:04
Air Blank	0.000	16:06
Subject Sample # 2	0.206	16:06
Breath Volume	2.406	16:06
Air Blank	0.000	16:07
Control Test	0.079	16:07
Air Blank	0.000	16:08
Diagnostics Check	OK	16:08

Data Download Date/Time: 05/13/2021 09:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRICE,MARCUS,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: WOODS,H
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: VELEZ DURAN

DATE/TIME OF TEST: 04/17/2021 23:11:01

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.079	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.217	23:06
Breath Volume	4.386	23:06
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.222	23:09
Breath Volume	4.898	23:09
Air Blank	0.000	23:10
Control Test	0.080	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:11

Data Download Date/Time: 05/13/2021 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLELLAND, WILLIAM, R
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MENDOZA, U
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: MATEOLUCAS

DATE/TIME OF TEST: 04/24/2021 22:44:58

Diagnostics Check	OK	22:37
Air Blank	0.000	22:38
Control Test	0.080	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.135	22:39
Breath Volume	1.742	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.134	22:43
Breath Volume	1.917	22:43
Air Blank	0.000	22:44
Control Test	0.081	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 05/13/2021 09:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARVIN,ANGELINA,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: WOODS,H
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/29/2021 19:44:47

Diagnostics Check	OK	19:39
Air Blank	0.000	19:39
Control Test	0.079	19:39
Air Blank	0.000	19:40
Subject Sample # 1	0.000	19:40
Breath Volume	3.265	19:40
Air Blank	0.000	19:41
Air Blank	0.000	19:43
Subject Sample # 2	0.000	19:43
Breath Volume	3.339	19:43
Air Blank	0.000	19:43
Control Test	0.079	19:44
Air Blank	0.000	19:44
Diagnostics Check	OK	19:44

Data Download Date/Time: 05/13/2021 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THROESCH,PATRICK,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: TERUZ,G
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: SIERRA

DATE/TIME OF TEST: 04/18/2021 02:22:12

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.173	02:17
Breath Volume	1.406	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.169	02:20
Breath Volume	2.128	02:20
Air Blank	0.000	02:21
Control Test	0.077	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 05/06/2021 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EPPS,QUINTELLA,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: OLEUPSK,R
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 03/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: GAYLE

DATE/TIME OF TEST: 04/17/2021 23:07:30

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:05	
Control Test	0.078	23:05	
Air Blank	0.000	23:05	
Subject Sample	# 1	23:06	Subject Test Refused
Breath Volume	0.000	23:06	
Air Blank	0.000	23:06	
Control Test	0.078	23:06	
Air Blank	0.000	23:07	
Diagnostics Check	OK	23:07	

Data Download Date/Time: 05/01/2021 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000950

INSTRUMENT REGISTERED TO: HAVANA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMALL,BRIAN,A
BREATH TEST OPERATOR AGENCY: Havana PD
ARRESTING OFFICER: SMALL,BRIAN
ARRESTING OFFICER AGENCY: Havana PD
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:58

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 04/05/2021 01:02:14

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:58	
Control Test	0.079	00:59	
Air Blank	0.000	00:59	
Subject Sample	# 1	01:00	Subject Test Refused
Breath Volume	0.000	01:00	
Air Blank	0.000	01:01	
Control Test	0.080	01:01	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 04/26/2021 14:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORDON,ROLF,E
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: GORDON,ROLF
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: HUTCHINSON

DATE/TIME OF TEST: 04/14/2021 22:32:47

Diagnostics Check	OK	22:31	
Air Blank	0.000	22:32	
		22:32	RFI Detect
Air Blank	0.000	22:32	

Data Download Date/Time: 04/28/2021 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORDON,ROLF,E
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: GORDON,ROLF
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: H325478871080

DATE/TIME OF TEST: 04/14/2021 22:39:06

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:36	
Control Test	0.079	22:37	
Air Blank	0.000	22:37	
Subject Sample	# 1	22:37	Subject Test Refused
Breath Volume	0.000	22:37	
Air Blank	0.000	22:38	
Control Test	0.080	22:38	
Air Blank	0.000	22:39	
Diagnostics Check	OK	22:39	

Data Download Date/Time: 04/28/2021 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT, JOSHUA, N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ADAMCZYK, ANDREW
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: BURNARD

DATE/TIME OF TEST: 04/06/2021 17:29:26

Diagnostics Check	OK	17:17	
Air Blank	0.000	17:18	
Control Test	0.080	17:18	
Air Blank	0.000	17:19	
Subject Sample	# 1 0.225	17:22	Volume Not Met
Breath Volume	0.140	17:22	
Air Blank	0.000	17:23	
Air Blank	0.000	17:24	
Subject Sample	# 2 0.224	17:27	Volume Not Met
Breath Volume	0.117	17:27	
Air Blank	0.000	17:28	
Control Test	0.077	17:28	
Air Blank	0.000	17:29	
Diagnostics Check	OK	17:29	

Data Download Date/Time: 04/26/2021 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BRIAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WARREN,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: PETTWAY

DATE/TIME OF TEST: 04/24/2021 03:12:30

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.079	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.000	03:08
Breath Volume	2.542	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.000	03:11
Breath Volume	2.785	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 04/26/2021 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,JAMES,D
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: TUCKER,CEDRIC
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 03/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: PESCE

DATE/TIME OF TEST: 04/01/2021 04:25:21

Diagnostics Check		OK	04:16
Air Blank		0.000	04:16
Control Test		0.081	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.253	04:17
Breath Volume		2.078	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.226	04:20
Breath Volume		1.109	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 3	0.230	04:23
Breath Volume		1.335	04:23
Air Blank		0.000	04:24
Control Test		0.081	04:24
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 04/12/2021 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CURVEY

DATE/TIME OF TEST: 04/03/2021 01:59:00

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:51	
Control Test	0.082	01:51	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:54	No Sample Provided
Breath Volume	0.000	01:54	
Air Blank	0.000	01:55	
Air Blank	0.000	01:57	
Subject Sample	# 2	01:57	Subject Test Refused
Breath Volume	0.000	01:57	
Air Blank	0.000	01:58	
Control Test	0.081	01:58	
Air Blank	0.000	01:58	
Diagnostics Check	OK	01:59	

Data Download Date/Time: 04/12/2021 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD,LEE,M
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: HARWOOD,LEE
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 03/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 04/04/2021 02:12:35

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:09	
Control Test	0.081	02:09	
Air Blank	0.000	02:10	
Subject Sample	# 1	02:11	Subject Test Refused
Breath Volume	0.000	02:11	
Air Blank	0.000	02:11	
Control Test	0.081	02:11	
Air Blank	0.000	02:12	
Diagnostics Check	OK	02:12	

Data Download Date/Time: 04/12/2021 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 04/06/2021 23:33:05

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.080	23:27
Air Blank	0.000	23:27
Subject Sample # 1	0.032	23:28
Breath Volume	1.773	23:28
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.029	23:31
Breath Volume	1.683	23:31
Air Blank	0.000	23:32
Control Test	0.080	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 04/12/2021 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,BENJAMIN,Z
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: DANIELS,BENJAMIN
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 03/04/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: DOGAN

DATE/TIME OF TEST: 04/09/2021 04:28:04

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:23	
Control Test	0.081	04:24	
Air Blank	0.000	04:24	
Subject Sample	# 1	04:26	Subject Test Refused
Breath Volume	0.000	04:26	
Air Blank	0.000	04:27	
Control Test	0.080	04:27	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:28	

Data Download Date/Time: 04/12/2021 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEFFIELD,BRANDON,S
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: SHEFFIELD,BRANDON
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 03/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: COPELAND

DATE/TIME OF TEST: 04/09/2021 05:56:47

Diagnostics Check	OK	05:50
Air Blank	0.000	05:50
Control Test	0.081	05:51
Air Blank	0.000	05:51
Subject Sample # 1	0.138	05:52
Breath Volume	1.394	05:52
Air Blank	0.000	05:52
Air Blank	0.000	05:54
Subject Sample # 2	0.138	05:55
Breath Volume	1.246	05:55
Air Blank	0.000	05:55
Control Test	0.081	05:56
Air Blank	0.000	05:56
Diagnostics Check	OK	05:56

Data Download Date/Time: 04/12/2021 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SODERHOLM,CHELSEA,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SODERHOLM,CHELSEA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 04/10/2021 02:46:38

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:43	
Control Test	0.081	02:43	
Air Blank	0.000	02:43	
Subject Sample	# 1	02:46	Sequence Aborted
Breath Volume	0.000	02:46	
Air Blank	0.000	02:46	

Data Download Date/Time: 04/12/2021 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: DEMOTT

DATE/TIME OF TEST: 04/13/2021 05:33:43

Diagnostics Check	OK	05:27
Air Blank	0.000	05:28
Control Test	0.081	05:28
Air Blank	0.000	05:28
Subject Sample # 1	0.000	05:29
Breath Volume	4.433	05:29
Air Blank	0.000	05:29
Air Blank	0.000	05:31
Subject Sample # 2	0.000	05:32
Breath Volume	4.257	05:32
Air Blank	0.000	05:32
Control Test	0.080	05:33
Air Blank	0.000	05:33
Diagnostics Check	OK	05:33

Data Download Date/Time: 04/27/2021 16:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: THOMPSON,JEFFREY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: ODER

DATE/TIME OF TEST: 04/13/2021 19:45:00

Diagnostics Check	OK	19:38
Air Blank	0.000	19:39
Control Test	0.081	19:39
Air Blank	0.000	19:40
Subject Sample # 1	0.209	19:40
Breath Volume	1.347	19:40
Air Blank	0.000	19:41
Air Blank	0.000	19:43
Subject Sample # 2	0.216	19:43
Breath Volume	1.781	19:43
Air Blank	0.000	19:44
Control Test	0.081	19:44
Air Blank	0.000	19:44
Diagnostics Check	OK	19:45

Data Download Date/Time: 04/27/2021 16:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WONSCH,J,W
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SHEATS,C
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: GRIFFEN

DATE/TIME OF TEST: 04/16/2021 04:07:26

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.082	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.090	04:03
Breath Volume	2.042	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.095	04:05
Breath Volume	2.113	04:05
Air Blank	0.000	04:06
Control Test	0.081	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 04/27/2021 16:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: GRANADOS

DATE/TIME OF TEST: 04/21/2021 10:41:40

Diagnostics Check	OK	10:35
Air Blank	0.000	10:35
Control Test	0.081	10:36
Air Blank	0.000	10:36
Subject Sample # 1	0.077	10:37
Breath Volume	5.042	10:37
Air Blank	0.000	10:37
Air Blank	0.000	10:39
Subject Sample # 2	0.079	10:40
Breath Volume	4.886	10:40
Air Blank	0.000	10:40
Control Test	0.082	10:41
Air Blank	0.000	10:41
Diagnostics Check	OK	10:41

Data Download Date/Time: 04/27/2021 16:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MCMILLAN

DATE/TIME OF TEST: 04/22/2021 00:05:32

Diagnostics Check		OK	23:58
Air Blank		0.000	23:58
Control Test		0.081	23:59
Air Blank		0.000	23:59
Subject Sample	# 1	0.256	00:00
Breath Volume		2.953	00:00
Air Blank		0.000	00:01
Air Blank		0.000	00:03
Subject Sample	# 2	0.252	00:04
Breath Volume		2.546	00:04
Air Blank		0.000	00:04
Control Test		0.081	00:04
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 04/27/2021 16:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: HELM

DATE/TIME OF TEST: 04/25/2021 05:08:21

Diagnostics Check	OK	05:01
Air Blank	0.000	05:02
Control Test	0.081	05:02
Air Blank	0.000	05:03
Subject Sample # 1	0.092	05:03
Breath Volume	1.726	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.091	05:06
Breath Volume	1.746	05:06
Air Blank	0.000	05:07
Control Test	0.080	05:07
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 04/27/2021 16:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,TREVOR,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: GOEB

DATE/TIME OF TEST: 04/26/2021 18:39:21

Diagnostics Check	OK	18:32
Air Blank	0.000	18:33
Control Test	0.081	18:33
Air Blank	0.000	18:34
Subject Sample # 1	0.153	18:34
Breath Volume	4.121	18:34
Air Blank	0.000	18:35
Air Blank	0.000	18:37
Subject Sample # 2	0.142	18:37
Breath Volume	4.109	18:37
Air Blank	0.000	18:38
Control Test	0.082	18:38
Air Blank	0.000	18:39
Diagnostics Check	OK	18:39

Data Download Date/Time: 04/27/2021 16:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: NIX

DATE/TIME OF TEST: 04/27/2021 10:33:01

Diagnostics Check	OK	10:26
Air Blank	0.000	10:26
Control Test	0.081	10:26
Air Blank	0.000	10:27
Subject Sample # 1	0.163	10:28
Breath Volume	1.382	10:28
Air Blank	0.000	10:29
Air Blank	0.000	10:30
Subject Sample # 2	0.158	10:31
Breath Volume	1.375	10:31
Air Blank	0.000	10:32
Control Test	0.080	10:32
Air Blank	0.000	10:32
Diagnostics Check	OK	10:33

Data Download Date/Time: 04/27/2021 16:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/12/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: BAREFIELD

DATE/TIME OF TEST: 04/13/2021 20:13:48

Diagnostics Check	OK	20:06
Air Blank	0.000	20:06
Control Test	0.080	20:07
Air Blank	0.000	20:07
Subject Sample # 1	0.287	20:08
Breath Volume	2.964	20:08
Air Blank	0.000	20:09
Air Blank	0.000	20:11
Subject Sample # 2	0.289	20:12
Breath Volume	2.234	20:12
Air Blank	0.000	20:12
Control Test	0.079	20:13
Air Blank	0.000	20:13
Diagnostics Check	OK	20:13

Data Download Date/Time: 04/27/2021 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 04/17/2021 04:58:22

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.081	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.190	04:53
Breath Volume	1.718	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:55
Subject Sample # 2	0.190	04:56
Breath Volume	1.472	04:56
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 04/27/2021 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,BENJAMIN,Z
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: HILL,HEATHER
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 04/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: STEEN

DATE/TIME OF TEST: 04/22/2021 02:50:41

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.081	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.189	02:46
Breath Volume	1.410	02:46
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.205	02:49
Breath Volume	2.335	02:49
Air Blank	0.000	02:49
Control Test	0.078	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 04/27/2021 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: FRANKLIN,M
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: OSBORNE

DATE/TIME OF TEST: 04/26/2021 03:36:53

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.166	03:32
Breath Volume	1.535	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.158	03:35
Breath Volume	1.527	03:35
Air Blank	0.000	03:35
Control Test	0.078	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 04/27/2021 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIDDLEBROOKS,ALAN,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MIDDLEBROOKS,ALAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 04/03/2021 00:00:51

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:57	
Control Test	0.080	23:58	
Air Blank	0.000	23:58	
Subject Sample	# 1 0.126	23:59	
Breath Volume	2.839	23:59	
Air Blank		00:00	Purge Fail
Air Blank	0.000	00:00	

Data Download Date/Time: 04/28/2021 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIDDLEBROOKS,ALAN,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MIDDLEBROOKS,ALAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 04/03/2021 00:08:43

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.080	00:03
Air Blank	0.000	00:03
Subject Sample # 1	0.123	00:04
Breath Volume	2.515	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.122	00:07
Breath Volume	2.503	00:07
Air Blank	0.000	00:07
Control Test	0.081	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 04/28/2021 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,GEORGE,H
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: EDWARDS,GEORGE
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: GOFF

DATE/TIME OF TEST: 04/03/2021 02:06:06

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:02	
Control Test	0.081	02:03	
Air Blank	0.000	02:03	
Subject Sample	# 1 0.139	02:04	
Breath Volume	2.324	02:04	
Air Blank		02:05	Purge Fail
Air Blank		02:06	Purge Fail

Data Download Date/Time: 04/28/2021 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,GEORGE,H
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: EDWARDS,GEORGE
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/25/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: GOFF

DATE/TIME OF TEST: 04/03/2021 02:15:18

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:09	
Control Test	0.080	02:10	
Air Blank	0.000	02:10	
Subject Sample	# 1 0.137	02:11	
Breath Volume	1.855	02:11	
Air Blank	0.000	02:11	
Air Blank	0.000	02:13	
Subject Sample	# 2	02:13	RFI Detect
Breath Volume	0.000	02:13	
Air Blank	0.000	02:14	
Control Test	0.080	02:14	
Air Blank	0.000	02:15	
Diagnostics Check	OK	02:15	

Data Download Date/Time: 04/28/2021 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,GEORGE,H
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: EDWARDS,GEORGE
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: GOFF

DATE/TIME OF TEST: 04/03/2021 02:23:14

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.080	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.134	02:18
Breath Volume	1.746	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.134	02:21
Breath Volume	2.238	02:21
Air Blank	0.000	02:22
Control Test	0.081	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 04/28/2021 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS, SR.,MICHAEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: CROSS, SR.,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 04/04/2021 00:59:27

Diagnostics Check	OK	00:56	
Air Blank	0.000	00:56	
Control Test	0.080	00:56	
Air Blank	0.000	00:57	
Subject Sample # 1	0.206	00:58	
Breath Volume	1.289	00:58	
Air Blank		00:58	Purge Fail
Air Blank	0.000	00:59	

Data Download Date/Time: 04/28/2021 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS, SR.,MICHAEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: CROSS, SR.,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 04/04/2021 01:16:06

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:01	
Control Test	0.078	01:02	
Air Blank	0.000	01:02	
Subject Sample	# 1	01:05	No Sample Provided
Breath Volume	0.000	01:05	
Air Blank	0.000	01:06	
Air Blank	0.000	01:08	
Subject Sample	# 2	0.176	01:09
Breath Volume	1.152	01:09	
Air Blank	0.000	01:09	
Air Blank	0.000	01:11	
Subject Sample	# 3	01:14	No Sample Provided
Breath Volume	0.000	01:14	
Air Blank	0.000	01:15	
Control Test	0.077	01:15	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

Data Download Date/Time: 04/28/2021 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAUFFMAN, MITCHELL, P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: KAUFFMAN, MITCHELL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: GRIEGER

DATE/TIME OF TEST: 04/19/2021 04:03:50

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.080	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.140	03:58
Breath Volume	3.441	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:00
Subject Sample # 2	0.122	04:02
Breath Volume	1.949	04:02
Air Blank	0.000	04:02
Control Test	0.080	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 04/28/2021 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEAVER,WILLIAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HARDIN

DATE/TIME OF TEST: 04/17/2021 04:33:15

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.172	04:28
Breath Volume	2.617	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.171	04:31
Breath Volume	3.011	04:31
Air Blank	0.000	04:32
Control Test	0.079	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 04/22/2021 16:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GILLESPIE,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 04/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: ENRIQUEZ

DATE/TIME OF TEST: 04/18/2021 21:41:17

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.080	21:34
Air Blank	0.000	21:34
Subject Sample # 1	0.000	21:36
Breath Volume	2.132	21:36
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.000	21:39
Breath Volume	1.542	21:39
Air Blank	0.000	21:40
Control Test	0.079	21:40
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 04/22/2021 16:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEWAN,JANICE,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WILKES,DAVID
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 04/01/2021 03:04:20

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.080	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.000	03:00
Breath Volume	2.082	03:00
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.000	03:02
Breath Volume	1.714	03:02
Air Blank	0.000	03:03
Control Test	0.080	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 04/13/2021 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEWAN,JANICE,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 04/02/2021 04:32:13

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.080	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.089	04:27
Breath Volume	1.105	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.089	04:30
Breath Volume	1.359	04:30
Air Blank	0.000	04:31
Control Test	0.080	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 04/13/2021 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEWAN,JANICE,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIDEAUX,ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: RAPP

DATE/TIME OF TEST: 04/03/2021 05:00:13

Diagnostics Check		OK	04:52
Air Blank		0.000	04:52
Control Test		0.081	04:53
Air Blank		0.000	04:53
Subject Sample	# 1	0.205	04:54
Breath Volume		1.660	04:54
Air Blank		0.000	04:54
Air Blank		0.000	04:56
Subject Sample	# 2	0.209	04:58
Breath Volume		1.992	04:58
Air Blank		0.000	04:59
Control Test		0.080	04:59
Air Blank		0.000	05:00
Diagnostics Check		OK	05:00

Data Download Date/Time: 04/13/2021 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEWAN,JANICE,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIAZ,JASON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: BATISTA

DATE/TIME OF TEST: 04/03/2021 23:20:52

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.080	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.000	23:16
Breath Volume	3.144	23:16
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.000	23:19
Breath Volume	3.449	23:19
Air Blank	0.000	23:19
Control Test	0.081	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 04/13/2021 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JACKSON,ALLEN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: DOMINGUEZGONZALEZ

DATE/TIME OF TEST: 04/04/2021 02:41:41

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.080	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.151	02:37
Breath Volume	1.339	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.155	02:40
Breath Volume	3.222	02:40
Air Blank	0.000	02:40
Control Test	0.080	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 04/13/2021 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,CRAIG
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: MANAV

DATE/TIME OF TEST: 04/05/2021 05:28:49

Diagnostics Check	OK	05:20
Air Blank	0.000	05:21
Control Test	0.081	05:21
Air Blank	0.000	05:22
Subject Sample # 1	0.100	05:24
Breath Volume	1.582	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:26
Subject Sample # 2	0.099	05:27
Breath Volume	1.523	05:27
Air Blank	0.000	05:27
Control Test	0.080	05:28
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 04/13/2021 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:08

SUBJECT'S LAST NAME: CUBAS

DATE/TIME OF TEST: 04/05/2021 06:38:33

Diagnostics Check	OK	06:32
Air Blank	0.000	06:33
Control Test	0.080	06:33
Air Blank	0.000	06:33
Subject Sample # 1	0.067	06:34
Breath Volume	2.250	06:34
Air Blank	0.000	06:34
Air Blank	0.000	06:36
Subject Sample # 2	0.068	06:37
Breath Volume	2.707	06:37
Air Blank	0.000	06:37
Control Test	0.080	06:37
Air Blank	0.000	06:38
Diagnostics Check	OK	06:38

Data Download Date/Time: 04/13/2021 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 04/07/2021 02:20:46

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:09	
Control Test	0.079	02:10	
Air Blank	0.000	02:10	
Subject Sample	# 1	02:13	No Sample Provided
Breath Volume	0.000	02:13	
Air Blank	0.000	02:14	
Air Blank	0.000	02:16	
Subject Sample	# 2	02:19	No Sample Provided
Breath Volume	0.000	02:19	
Air Blank	0.000	02:19	
Control Test	0.079	02:20	
Air Blank	0.000	02:20	
Diagnostics Check	OK	02:20	

Data Download Date/Time: 04/13/2021 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ESPAILLAT,LIANA
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 04/07/2021 04:45:21

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.081	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.000	04:41
Breath Volume	1.882	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.000	04:44
Breath Volume	3.367	04:44
Air Blank	0.000	04:44
Control Test	0.080	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 04/13/2021 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEWAN,JANICE,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: OGLETREE,JERRELL
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: BIGLEY

DATE/TIME OF TEST: 04/08/2021 03:50:56

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.080	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.162	03:41
Breath Volume		1.269	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.194	03:44
Breath Volume		1.527	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 3	0.184	03:49
Breath Volume		2.343	03:49
Air Blank		0.000	03:50
Control Test		0.080	03:50
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 04/13/2021 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEWAN,JANICE,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ESPAILLAT,LIANA
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: BUGASCH

DATE/TIME OF TEST: 04/09/2021 02:46:43

Diagnostics Check		OK	02:34	
Air Blank		0.000	02:35	
Control Test		0.080	02:35	
Air Blank		0.000	02:36	
Subject Sample	# 1	0.032	02:39	Volume Not Met
Breath Volume		0.449	02:39	
Air Blank		0.000	02:39	
Air Blank		0.000	02:41	
Subject Sample	# 2	0.033	02:42	
Breath Volume		1.832	02:42	
Air Blank		0.000	02:42	
Air Blank		0.000	02:44	
Subject Sample	# 3	0.032	02:45	
Breath Volume		1.593	02:45	
Air Blank		0.000	02:45	
Control Test		0.080	02:46	
Air Blank		0.000	02:46	
Diagnostics Check		OK	02:46	

Data Download Date/Time: 04/13/2021 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMPSON, JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MUNOZ VELAZQUEZ

DATE/TIME OF TEST: 04/09/2021 04:28:38

Diagnostics Check	OK	04:14	
Air Blank	0.000	04:14	
Control Test	0.079	04:15	
Air Blank	0.000	04:15	
Subject Sample # 1	0.181	04:16	
Breath Volume	2.667	04:16	
Air Blank	0.000	04:16	
Air Blank	0.000	04:18	
Subject Sample # 2		04:21	No Sample Provided
Breath Volume	0.000	04:21	
Air Blank	0.000	04:22	
Air Blank	0.000	04:24	
Subject Sample # 3		04:27	No Sample Provided
Breath Volume	0.000	04:27	
Air Blank	0.000	04:27	
Control Test	0.081	04:28	
Air Blank	0.000	04:28	
Diagnostics Check	OK	04:28	

Data Download Date/Time: 04/13/2021 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS, CHRISTOPHER
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 04/10/2021 00:00:43

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:55
Subject Sample # 1	0.000	23:56
Breath Volume	3.011	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.000	23:59
Breath Volume	2.027	23:59
Air Blank	0.000	23:59
Control Test	0.080	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 04/13/2021 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GILZEN,JACOB
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: ANGRY

DATE/TIME OF TEST: 04/10/2021 02:45:44

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.107	02:41
Breath Volume	1.820	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:43
Subject Sample # 2	0.111	02:44
Breath Volume	1.527	02:44
Air Blank	0.000	02:44
Control Test	0.079	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 04/13/2021 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALL,JEFF
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: ISHAQ

DATE/TIME OF TEST: 04/10/2021 03:39:30

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.080	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.212	03:34
Breath Volume		1.601	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.213	03:38
Breath Volume		1.250	03:38
Air Blank		0.000	03:38
Control Test		0.080	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 04/13/2021 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VENTURA, CHRISTOPHER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: STAMM

DATE/TIME OF TEST: 04/10/2021 04:26:42

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.080	04:20
Air Blank		0.000	04:21
Subject Sample	# 1	0.234	04:21
Breath Volume		1.515	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.230	04:25
Breath Volume		1.460	04:25
Air Blank		0.000	04:25
Control Test		0.080	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 04/13/2021 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WONGSHUE, ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 04/10/2021 12:17:26

Diagnostics Check	OK	12:11
Air Blank	0.000	12:12
Control Test	0.080	12:12
Air Blank	0.000	12:12
Subject Sample # 1	0.151	12:13
Breath Volume	1.578	12:13
Air Blank	0.000	12:13
Air Blank	0.000	12:15
Subject Sample # 2	0.150	12:15
Breath Volume	1.144	12:15
Air Blank	0.000	12:16
Control Test	0.080	12:16
Air Blank	0.000	12:17
Diagnostics Check	OK	12:17

Data Download Date/Time: 04/13/2021 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMPSON, JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: BEAUX

DATE/TIME OF TEST: 04/11/2021 02:52:01

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.079	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.210	02:47
Breath Volume	1.167	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.190	02:50
Breath Volume	1.503	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:52

Data Download Date/Time: 04/13/2021 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LIESENFELT,JASON
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: CRIST

DATE/TIME OF TEST: 04/11/2021 03:31:41

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.133	03:25
Breath Volume	1.160	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.136	03:30
Breath Volume	1.332	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 04/13/2021 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JEAN BAPTISTE,ALBENS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:08

SUBJECT'S LAST NAME: LINDSAY

DATE/TIME OF TEST: 04/11/2021 06:43:10

Diagnostics Check	OK	06:35
Air Blank	0.000	06:36
Control Test	0.079	06:36
Air Blank	0.000	06:37
Subject Sample # 1	0.127	06:38
Breath Volume	1.457	06:38
Air Blank	0.000	06:38
Air Blank	0.000	06:40
Subject Sample # 2	0.131	06:41
Breath Volume	2.562	06:41
Air Blank	0.000	06:42
Control Test	0.079	06:42
Air Blank	0.000	06:43
Diagnostics Check	OK	06:43

Data Download Date/Time: 04/13/2021 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: REYNANTE,PORCHIA
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:54

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 04/11/2021 07:29:01

Diagnostics Check	OK	07:21
Air Blank	0.000	07:22
Control Test	0.079	07:22
Air Blank	0.000	07:23
Subject Sample # 1	0.130	07:24
Breath Volume	2.539	07:24
Air Blank	0.000	07:24
Air Blank	0.000	07:26
Subject Sample # 2	0.132	07:27
Breath Volume	1.843	07:27
Air Blank	0.000	07:28
Control Test	0.079	07:28
Air Blank	0.000	07:28
Diagnostics Check	OK	07:29

Data Download Date/Time: 04/13/2021 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOHNS,MARK
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 04/12/2021 10:36:10

Diagnostics Check	OK	10:24	
Air Blank	0.000	10:25	
Control Test	0.080	10:25	
Air Blank	0.000	10:26	
Subject Sample	# 1	10:29	No Sample Provided
Breath Volume	0.000	10:29	
Air Blank	0.000	10:29	
Air Blank	0.000	10:31	
Subject Sample	# 2	10:34	No Sample Provided
Breath Volume	0.000	10:34	
Air Blank	0.000	10:35	
Control Test	0.081	10:35	
Air Blank	0.000	10:36	
Diagnostics Check	OK	10:36	

Data Download Date/Time: 04/13/2021 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PEARDON,MATTHEW
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: DEL RIO

DATE/TIME OF TEST: 04/12/2021 23:09:20

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.080	23:02
Air Blank	0.000	23:02
Subject Sample # 1	0.108	23:04
Breath Volume	1.171	23:04
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.102	23:07
Breath Volume	1.402	23:07
Air Blank	0.000	23:08
Control Test	0.080	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 04/13/2021 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 11:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: REINKEMEYER

DATE/TIME OF TEST: 04/02/2021 12:42:42

Diagnostics Check	OK	12:36
Air Blank	0.000	12:37
Control Test	0.081	12:37
Air Blank	0.000	12:37
Subject Sample # 1	0.000	12:38
Breath Volume	2.156	12:38
Air Blank	0.000	12:38
Air Blank	0.000	12:40
Subject Sample # 2	0.000	12:41
Breath Volume	1.699	12:41
Air Blank	0.000	12:41
Control Test	0.082	12:42
Air Blank	0.000	12:42
Diagnostics Check	OK	12:42

Data Download Date/Time: 04/19/2021 09:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: RESENDIZ MEJIA

DATE/TIME OF TEST: 04/09/2021 21:03:26

Diagnostics Check	OK	20:56
Air Blank	0.000	20:57
Control Test	0.079	20:57
Air Blank	0.000	20:57
Subject Sample # 1	0.071	20:58
Breath Volume	1.667	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.074	21:01
Breath Volume	3.335	21:01
Air Blank	0.000	21:02
Control Test	0.081	21:02
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 04/19/2021 09:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDENAS YANEZ, JUAN, E
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: PAULINO, JOSE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: CARDENAS YANEZ

DATE/TIME OF TEST: 04/02/2021 03:38:56

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.080	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.154	03:34
Breath Volume		1.992	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:36
Subject Sample	# 2	0.154	03:37
Breath Volume		3.855	03:37
Air Blank		0.000	03:38
Control Test		0.078	03:38
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 04/30/2021 16:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: COWLEY,DUSTIN
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: LOOR

DATE/TIME OF TEST: 04/02/2021 03:51:57

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.147	03:47
Breath Volume	3.074	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:49
Subject Sample # 2	0.131	03:50
Breath Volume	1.554	03:50
Air Blank	0.000	03:51
Control Test	0.078	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 04/30/2021 16:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHOBREY,SPENCER,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WHOBREY,SPENCER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: DONOVAN

DATE/TIME OF TEST: 04/03/2021 00:02:38

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.078	23:54
Air Blank	0.000	23:54
Subject Sample # 2	0.110	23:57
Breath Volume	2.296	23:57
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 1	0.109	00:01
Breath Volume	2.210	00:01
Air Blank	0.000	00:01
Control Test	0.078	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 04/30/2021 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ, MONICA, M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RUSSAK, KEVIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 04/04/2021 03:47:13

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:36	
Control Test	0.080	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 1	03:40	No Sample Provided
Breath Volume	0.000	03:40	
Air Blank	0.000	03:40	
Air Blank	0.000	03:42	
Subject Sample	# 2	03:45	No Sample Provided
Breath Volume	0.000	03:45	
Air Blank	0.000	03:46	
Control Test	0.080	03:46	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

Data Download Date/Time: 04/30/2021 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REDDITT,MATTHEW,A
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: HELLER,MICHAEL
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: GOPAUL

DATE/TIME OF TEST: 04/11/2021 02:41:56

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.138	02:37
Breath Volume	1.636	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:39
Subject Sample # 2	0.136	02:40
Breath Volume	1.425	02:40
Air Blank	0.000	02:41
Control Test	0.078	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 04/30/2021 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: GIRON GOMEZ

DATE/TIME OF TEST: 04/11/2021 04:04:01

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.079	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.180	03:59
Breath Volume	2.898	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.172	04:02
Breath Volume	2.472	04:02
Air Blank	0.000	04:03
Control Test	0.079	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:04

Data Download Date/Time: 04/30/2021 16:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: GREENWALL,ROBERT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 04/11/2021 04:29:05

Diagnostics Check	OK	04:19
Air Blank	0.000	04:20
Control Test	0.079	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.115	04:22
Breath Volume	1.214	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.120	04:27
Breath Volume	2.109	04:27
Air Blank	0.000	04:28
Control Test	0.077	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 04/30/2021 16:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEDOYA,HERNANDO,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BEHRENSHOUSER,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: LARACUENTE

DATE/TIME OF TEST: 04/16/2021 03:40:26

Diagnostics Check	OK	03:40	
Air Blank		03:40	RFI Detect
Air Blank		03:40	RFI Detect

Data Download Date/Time: 04/30/2021 16:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEDOYA,HERNANDO,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BEHRENSHOUSER,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: LARACUENTE

DATE/TIME OF TEST: 04/16/2021 03:44:23

Diagnostics Check	OK	03:44	
Air Blank		03:44	RFI Detect
Air Blank		03:44	RFI Detect

Data Download Date/Time: 04/30/2021 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEDOYA,HERNANDO,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BEHRENSHOUSER,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: LARACUENTE

DATE/TIME OF TEST: 04/16/2021 03:59:21

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.116	03:54
Breath Volume	1.328	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.113	03:57
Breath Volume	1.632	03:57
Air Blank	0.000	03:58
Control Test	0.078	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 04/30/2021 16:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTONE,LACEY,M
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: SANTONE,LACEY
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: OSORIO MIRANDA

DATE/TIME OF TEST: 04/18/2021 02:46:28

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:37	
Control Test	0.079	02:37	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:41	No Sample Provided
Breath Volume	0.000	02:41	
Air Blank	0.000	02:41	
Air Blank	0.000	02:43	
Subject Sample	# 2	02:46	Sequence Aborted
Breath Volume	0.000	02:46	
Air Blank	0.000	02:46	

Data Download Date/Time: 04/30/2021 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 06:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:53

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 04/18/2021 07:24:29

Diagnostics Check	OK	07:18
Air Blank	0.000	07:18
Control Test	0.078	07:18
Air Blank	0.000	07:19
Subject Sample # 1	0.168	07:19
Breath Volume	3.296	07:19
Air Blank	0.000	07:20
Air Blank	0.000	07:22
Subject Sample # 2	0.170	07:22
Breath Volume	3.246	07:22
Air Blank	0.000	07:23
Control Test	0.079	07:23
Air Blank	0.000	07:24
Diagnostics Check	OK	07:24

Data Download Date/Time: 04/30/2021 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CALDERON,RAFAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: GUZMAN ANTONMARCHI

DATE/TIME OF TEST: 04/24/2021 04:10:47

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.080	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.162	04:06
Breath Volume	2.574	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.162	04:09
Breath Volume	2.535	04:09
Air Blank	0.000	04:09
Control Test	0.078	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 04/30/2021 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: IRIZARRY RUIZ

DATE/TIME OF TEST: 04/25/2021 00:34:54

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.080	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.135	00:29
Breath Volume	1.375	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.133	00:33
Breath Volume	1.554	00:33
Air Blank	0.000	00:33
Control Test	0.079	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 04/30/2021 16:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUNEZ,ISAI
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: CLEVINGER

DATE/TIME OF TEST: 04/25/2021 00:52:34

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.079	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.229	00:44
Breath Volume	1.191	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.207	00:47
Breath Volume	1.789	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 3	0.204	00:50
Breath Volume	1.781	00:50
Air Blank	0.000	00:51
Control Test	0.077	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 04/30/2021 16:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HANSELL,ROBERT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: MIESES

DATE/TIME OF TEST: 04/30/2021 01:44:02

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.081	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.128	01:38
Breath Volume	1.148	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.142	01:42
Breath Volume	2.207	01:42
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:44

Data Download Date/Time: 04/30/2021 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCHBACH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: GIVANS,GARRETH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:17

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 04/02/2021 16:57:17

Diagnostics Check	OK	16:50
Air Blank	0.000	16:50
Control Test	0.080	16:51
Air Blank	0.000	16:51
Subject Sample # 1	0.279	16:52
Breath Volume	2.117	16:52
Air Blank	0.000	16:53
Air Blank	0.000	16:54
Subject Sample # 2	0.277	16:55
Breath Volume	3.902	16:55
Air Blank	0.000	16:56
Control Test	0.080	16:56
Air Blank	0.000	16:57
Diagnostics Check	OK	16:57

Data Download Date/Time: 05/07/2021 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LLEWELLYN,A,
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: STANLEY,W
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: GERSTENBERGER

DATE/TIME OF TEST: 04/02/2021 23:28:39

Diagnostics Check	OK	23:22
Air Blank	0.000	23:22
Control Test	0.082	23:22
Air Blank	0.000	23:23
Subject Sample # 1	0.161	23:23
Breath Volume	2.101	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.165	23:27
Breath Volume	2.339	23:27
Air Blank	0.000	23:27
Control Test	0.081	23:28
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 05/07/2021 11:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INGERSOLL,NICHOLAS,F
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: FARRAGHER,THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: CRUZ RODRIGUEZ

DATE/TIME OF TEST: 04/09/2021 20:36:42

Diagnostics Check	OK	20:30
Air Blank	0.000	20:30
Control Test	0.080	20:30
Air Blank	0.000	20:31
Subject Sample # 1	0.192	20:32
Breath Volume	4.164	20:32
Air Blank	0.000	20:32
Air Blank	0.000	20:34
Subject Sample # 2	0.189	20:35
Breath Volume	4.289	20:35
Air Blank	0.000	20:35
Control Test	0.080	20:36
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

Data Download Date/Time: 05/07/2021 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INGERSOLL,NICHOLAS,F
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: OLSEN,IAN
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: FONTANEZ VARGAS

DATE/TIME OF TEST: 04/11/2021 03:00:05

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.081	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.225	02:55
Breath Volume	1.589	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.224	02:58
Breath Volume	1.511	02:58
Air Blank	0.000	02:59
Control Test	0.081	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 05/07/2021 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTMAN, MATTHEW, R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HATZEL, BRANDON
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: LOYD

DATE/TIME OF TEST: 04/11/2021 23:29:08

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.081	23:23
Air Blank	0.000	23:24
Subject Sample # 1	0.000	23:24
Breath Volume	2.203	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.000	23:27
Breath Volume	1.417	23:27
Air Blank	0.000	23:28
Control Test	0.081	23:28
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 05/07/2021 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: EASTMAN,MATTHEW
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: GEBREHIWOT

DATE/TIME OF TEST: 04/12/2021 01:30:57

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.082	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.000	01:26
Breath Volume	2.453	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.000	01:29
Breath Volume	3.832	01:29
Air Blank	0.000	01:30
Control Test	0.081	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 05/07/2021 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,JAMIE,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MENDEZ,JAMIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DAVILA

DATE/TIME OF TEST: 04/24/2021 03:42:47

Diagnostics Check		OK	03:35
Air Blank		0.000	03:35
Control Test		0.082	03:36
Air Blank		0.000	03:36
Subject Sample	# 1	0.210	03:38
Breath Volume		1.269	03:38
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.215	03:41
Breath Volume		1.804	03:41
Air Blank		0.000	03:41
Control Test		0.080	03:42
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 05/07/2021 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDERMAN,CASEY,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: WRIGHT,TYLER
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: ALACORN

DATE/TIME OF TEST: 04/07/2021 00:40:05

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.080	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.045	00:35
Breath Volume	3.316	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:37
Subject Sample # 2	0.045	00:38
Breath Volume	3.609	00:38
Air Blank	0.000	00:39
Control Test	0.081	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 05/27/2021 19:08