

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 04/03/2021 03:33:39

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 03:30 |                    |
| Air Blank         | 0.000 | 03:30 |                    |
| Control Test      | 0.084 | 03:31 |                    |
| Air Blank         | 0.000 | 03:31 |                    |
| Subject Sample    | # 1   | 03:32 | Interferent Detect |
| Breath Volume     | 1.906 | 03:32 |                    |
| Air Blank         |       | 03:32 | Purge Fail         |
| Air Blank         | 0.000 | 03:33 |                    |

---

Data Download Date/Time: 04/16/2021 13:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 04/03/2021 03:38:50

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 03:35 |            |
| Air Blank         | 0.000     | 03:35 |            |
| Control Test      | 0.084     | 03:36 |            |
| Air Blank         | 0.000     | 03:36 |            |
| Subject Sample    | # 1 0.075 | 03:37 |            |
| Breath Volume     | 2.546     | 03:37 |            |
| Air Blank         |           | 03:37 | Purge Fail |
| Air Blank         | 0.000     | 03:38 |            |

---

Data Download Date/Time: 04/16/2021 13:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROWE,BRANDON,D  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: ROWE,BRANDON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 04/03/2021 03:47:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:40 |
| Air Blank          | 0.000 | 03:41 |
| Control Test       | 0.084 | 03:41 |
| Air Blank          | 0.000 | 03:41 |
| Subject Sample # 1 | 0.069 | 03:42 |
| Breath Volume      | 2.867 | 03:42 |
| Air Blank          | 0.000 | 03:43 |
| Air Blank          | 0.000 | 03:44 |
| Subject Sample # 2 | 0.071 | 03:45 |
| Breath Volume      | 2.957 | 03:45 |
| Air Blank          | 0.000 | 03:46 |
| Control Test       | 0.085 | 03:46 |
| Air Blank          | 0.000 | 03:46 |
| Diagnostics Check  | OK    | 03:47 |

---

Data Download Date/Time: 04/16/2021 13:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 04/03/2021 03:50:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:43 |
| Air Blank          | 0.000 | 03:43 |
| Control Test       | 0.081 | 03:44 |
| Air Blank          | 0.000 | 03:44 |
| Subject Sample # 1 | 0.203 | 03:45 |
| Breath Volume      | 2.113 | 03:45 |
| Air Blank          | 0.000 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Subject Sample # 2 | 0.204 | 03:48 |
| Breath Volume      | 2.488 | 03:48 |
| Air Blank          | 0.000 | 03:49 |
| Control Test       | 0.080 | 03:49 |
| Air Blank          | 0.000 | 03:50 |
| Diagnostics Check  | OK    | 03:50 |

---

Data Download Date/Time: 05/13/2021 15:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER III,RICHARD,F  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: SNYDER III,RICHARD  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: MARTINEZ JR

DATE/TIME OF TEST: 04/03/2021 05:21:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:15 |
| Air Blank          | 0.000 | 05:15 |
| Control Test       | 0.081 | 05:15 |
| Air Blank          | 0.000 | 05:16 |
| Subject Sample # 1 | 0.000 | 05:16 |
| Breath Volume      | 2.886 | 05:16 |
| Air Blank          | 0.000 | 05:17 |
| Air Blank          | 0.000 | 05:19 |
| Subject Sample # 2 | 0.000 | 05:19 |
| Breath Volume      | 2.769 | 05:19 |
| Air Blank          | 0.000 | 05:20 |
| Control Test       | 0.081 | 05:20 |
| Air Blank          | 0.000 | 05:21 |
| Diagnostics Check  | OK    | 05:21 |

---

Data Download Date/Time: 05/13/2021 15:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: GONZALEZ GREER

DATE/TIME OF TEST: 04/03/2021 21:42:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:36 |
| Air Blank          | 0.000 | 21:36 |
| Control Test       | 0.080 | 21:37 |
| Air Blank          | 0.000 | 21:37 |
| Subject Sample # 1 | 0.137 | 21:38 |
| Breath Volume      | 1.714 | 21:38 |
| Air Blank          | 0.000 | 21:38 |
| Air Blank          | 0.000 | 21:40 |
| Subject Sample # 2 | 0.140 | 21:41 |
| Breath Volume      | 1.503 | 21:41 |
| Air Blank          | 0.000 | 21:42 |
| Control Test       | 0.080 | 21:42 |
| Air Blank          | 0.000 | 21:42 |
| Diagnostics Check  | OK    | 21:42 |

---

Data Download Date/Time: 05/13/2021 15:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: TREJO GUTIERREZ

DATE/TIME OF TEST: 04/04/2021 04:22:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:15 |
| Air Blank          | 0.000 | 04:15 |
| Control Test       | 0.081 | 04:15 |
| Air Blank          | 0.000 | 04:16 |
| Subject Sample # 1 | 0.141 | 04:17 |
| Breath Volume      | 1.785 | 04:17 |
| Air Blank          | 0.000 | 04:18 |
| Air Blank          | 0.000 | 04:19 |
| Subject Sample # 2 | 0.139 | 04:20 |
| Breath Volume      | 2.644 | 04:20 |
| Air Blank          | 0.000 | 04:21 |
| Control Test       | 0.080 | 04:21 |
| Air Blank          | 0.000 | 04:22 |
| Diagnostics Check  | OK    | 04:22 |

---

Data Download Date/Time: 05/13/2021 15:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: US TIQUIRAM

DATE/TIME OF TEST: 04/04/2021 04:55:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:48 |
| Air Blank          | 0.000 | 04:48 |
| Control Test       | 0.080 | 04:48 |
| Air Blank          | 0.000 | 04:49 |
| Subject Sample # 1 | 0.179 | 04:50 |
| Breath Volume      | 1.535 | 04:50 |
| Air Blank          | 0.000 | 04:51 |
| Air Blank          | 0.000 | 04:52 |
| Subject Sample # 2 | 0.177 | 04:53 |
| Breath Volume      | 2.394 | 04:53 |
| Air Blank          | 0.000 | 04:54 |
| Control Test       | 0.079 | 04:54 |
| Air Blank          | 0.000 | 04:54 |
| Diagnostics Check  | OK    | 04:55 |

---

Data Download Date/Time: 05/13/2021 15:47



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: KELLS,BRANDON  
ARRESTING OFFICER AGENCY: Bradenton PD  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: ALFONSO-TREJO

DATE/TIME OF TEST: 04/04/2021 05:46:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:39 |
| Air Blank          | 0.000 | 05:39 |
| Control Test       | 0.081 | 05:39 |
| Air Blank          | 0.000 | 05:40 |
| Subject Sample # 1 | 0.136 | 05:41 |
| Breath Volume      | 1.835 | 05:41 |
| Air Blank          | 0.000 | 05:42 |
| Air Blank          | 0.000 | 05:43 |
| Subject Sample # 2 | 0.143 | 05:44 |
| Breath Volume      | 3.046 | 05:44 |
| Air Blank          | 0.000 | 05:45 |
| Control Test       | 0.079 | 05:45 |
| Air Blank          | 0.000 | 05:46 |
| Diagnostics Check  | OK    | 05:46 |

---

Data Download Date/Time: 05/13/2021 15:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORRAO,MICHAEL,W  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: KOPE,MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: W325419991460

DATE/TIME OF TEST: 04/08/2021 03:17:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:10 |
| Air Blank          | 0.000 | 03:11 |
| Control Test       | 0.080 | 03:11 |
| Air Blank          | 0.000 | 03:12 |
| Subject Sample # 1 | 0.149 | 03:12 |
| Breath Volume      | 3.664 | 03:12 |
| Air Blank          | 0.000 | 03:13 |
| Air Blank          | 0.000 | 03:15 |
| Subject Sample # 2 | 0.143 | 03:16 |
| Breath Volume      | 3.191 | 03:16 |
| Air Blank          | 0.000 | 03:16 |
| Control Test       | 0.078 | 03:17 |
| Air Blank          | 0.000 | 03:17 |
| Diagnostics Check  | OK    | 03:17 |

---

Data Download Date/Time: 05/13/2021 15:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIDSON,CHRISTOPHER,E  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: DAVIDSON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 04/09/2021 19:58:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:51 |
| Air Blank          | 0.000 | 19:51 |
| Control Test       | 0.080 | 19:52 |
| Air Blank          | 0.000 | 19:52 |
| Subject Sample # 1 | 0.220 | 19:53 |
| Breath Volume      | 2.015 | 19:53 |
| Air Blank          | 0.000 | 19:54 |
| Air Blank          | 0.000 | 19:55 |
| Subject Sample # 2 | 0.223 | 19:56 |
| Breath Volume      | 1.562 | 19:56 |
| Air Blank          | 0.000 | 19:57 |
| Control Test       | 0.079 | 19:57 |
| Air Blank          | 0.000 | 19:58 |
| Diagnostics Check  | OK    | 19:58 |

---

Data Download Date/Time: 05/13/2021 15:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRERO,FRAICOR,  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: TERRERO,FRAICOR  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 04/09/2021 23:13:53

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:11 |                      |
| Air Blank         | 0.000 | 23:11 |                      |
| Control Test      | 0.080 | 23:11 |                      |
| Air Blank         | 0.000 | 23:12 |                      |
| Subject Sample    | # 1   | 23:12 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:12 |                      |
| Air Blank         | 0.000 | 23:12 |                      |
| Control Test      | 0.080 | 23:13 |                      |
| Air Blank         | 0.000 | 23:13 |                      |
| Diagnostics Check | OK    | 23:13 |                      |

---

Data Download Date/Time: 05/13/2021 15:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER III,RICHARD,F  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: SNYDER III,RICHARD  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: TOMAS BARTOLON

DATE/TIME OF TEST: 04/11/2021 03:05:37

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 02:51 |                |
| Air Blank         | 0.000     | 02:52 |                |
| Control Test      | 0.081     | 02:52 |                |
| Air Blank         | 0.000     | 02:53 |                |
| Subject Sample    | # 1 0.133 | 02:56 | Volume Not Met |
| Breath Volume     | 0.160     | 02:56 |                |
| Air Blank         | 0.000     | 02:56 |                |
| Air Blank         | 0.000     | 02:58 |                |
| Subject Sample    | # 2 0.135 | 02:59 |                |
| Breath Volume     | 1.769     | 02:59 |                |
| Air Blank         | 0.000     | 03:00 |                |
| Air Blank         | 0.000     | 03:01 |                |
| Subject Sample    | # 3 0.146 | 03:03 |                |
| Breath Volume     | 2.035     | 03:03 |                |
| Air Blank         | 0.000     | 03:04 |                |
| Control Test      | 0.080     | 03:04 |                |
| Air Blank         | 0.000     | 03:05 |                |
| Diagnostics Check | OK        | 03:05 |                |

---

Data Download Date/Time: 05/13/2021 15:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: WALDREP

DATE/TIME OF TEST: 04/12/2021 23:14:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:06 |
| Air Blank          | 0.000 | 23:06 |
| Control Test       | 0.080 | 23:07 |
| Air Blank          | 0.000 | 23:07 |
| Subject Sample # 1 | 0.243 | 23:08 |
| Breath Volume      | 1.281 | 23:08 |
| Air Blank          | 0.000 | 23:09 |
| Air Blank          | 0.000 | 23:10 |
| Subject Sample # 2 | 0.223 | 23:12 |
| Breath Volume      | 1.332 | 23:12 |
| Air Blank          | 0.000 | 23:13 |
| Control Test       | 0.079 | 23:13 |
| Air Blank          | 0.000 | 23:13 |
| Diagnostics Check  | OK    | 23:14 |

---

Data Download Date/Time: 05/13/2021 15:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRETT,CODY,R  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BARRETT,CODY  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: FONSECA

DATE/TIME OF TEST: 04/13/2021 03:58:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:51 |
| Air Blank          | 0.000 | 03:51 |
| Control Test       | 0.080 | 03:51 |
| Air Blank          | 0.000 | 03:52 |
| Subject Sample # 1 | 0.085 | 03:53 |
| Breath Volume      | 2.585 | 03:53 |
| Air Blank          | 0.000 | 03:54 |
| Air Blank          | 0.000 | 03:55 |
| Subject Sample # 2 | 0.075 | 03:56 |
| Breath Volume      | 1.652 | 03:56 |
| Air Blank          | 0.000 | 03:57 |
| Control Test       | 0.079 | 03:57 |
| Air Blank          | 0.000 | 03:57 |
| Diagnostics Check  | OK    | 03:58 |

---

Data Download Date/Time: 05/13/2021 15:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JIMENEZ, WILLIAM,  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: JIMENEZ, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 12:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: BENITEZ

DATE/TIME OF TEST: 04/17/2021 13:53:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:44 |
| Air Blank          | 0.000 | 13:44 |
| Control Test       | 0.080 | 13:45 |
| Air Blank          | 0.000 | 13:45 |
| Subject Sample # 1 | 0.262 | 13:46 |
| Breath Volume      | 1.402 | 13:46 |
| Air Blank          | 0.000 | 13:46 |
| Air Blank          | 0.000 | 13:48 |
| Subject Sample # 2 | 0.237 | 13:49 |
| Breath Volume      | 1.589 | 13:49 |
| Air Blank          | 0.000 | 13:50 |
| Air Blank          | 0.000 | 13:51 |
| Subject Sample # 3 | 0.245 | 13:52 |
| Breath Volume      | 1.570 | 13:52 |
| Air Blank          | 0.000 | 13:52 |
| Control Test       | 0.078 | 13:53 |
| Air Blank          | 0.000 | 13:53 |
| Diagnostics Check  | OK    | 13:53 |

---

Data Download Date/Time: 05/13/2021 15:48



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRERO,FRAICOR,  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: TERRERO,FRAICOR  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: DIAZ PEREZ

DATE/TIME OF TEST: 04/17/2021 22:31:19

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 22:22 |                |
| Air Blank         | 0.000     | 22:23 |                |
| Control Test      | 0.079     | 22:23 |                |
| Air Blank         | 0.000     | 22:23 |                |
| Subject Sample    | # 1 0.290 | 22:24 |                |
| Breath Volume     | 1.843     | 22:24 |                |
| Air Blank         | 0.000     | 22:25 |                |
| Air Blank         | 0.000     | 22:26 |                |
| Subject Sample    | # 2 0.288 | 22:29 | Volume Not Met |
| Breath Volume     | 0.070     | 22:29 |                |
| Air Blank         |           | 22:30 | Purge Fail     |
| Air Blank         | 0.000     | 22:31 |                |

---

Data Download Date/Time: 05/13/2021 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRERO,FRAICOR,  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: KELLS,BRANDON  
ARRESTING OFFICER AGENCY: Bradenton PD  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: POUTRE

DATE/TIME OF TEST: 04/18/2021 00:08:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:01 |
| Air Blank          | 0.000 | 00:02 |
| Control Test       | 0.079 | 00:02 |
| Air Blank          | 0.000 | 00:03 |
| Subject Sample # 1 | 0.074 | 00:03 |
| Breath Volume      | 2.820 | 00:03 |
| Air Blank          | 0.000 | 00:04 |
| Air Blank          | 0.000 | 00:06 |
| Subject Sample # 2 | 0.073 | 00:06 |
| Breath Volume      | 2.863 | 00:06 |
| Air Blank          | 0.000 | 00:07 |
| Control Test       | 0.079 | 00:07 |
| Air Blank          | 0.000 | 00:08 |
| Diagnostics Check  | OK    | 00:08 |

---

Data Download Date/Time: 05/13/2021 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: ENYEART

DATE/TIME OF TEST: 04/18/2021 01:45:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:39 |
| Air Blank          | 0.000 | 01:39 |
| Control Test       | 0.080 | 01:39 |
| Air Blank          | 0.000 | 01:40 |
| Subject Sample # 1 | 0.153 | 01:41 |
| Breath Volume      | 1.820 | 01:41 |
| Air Blank          | 0.000 | 01:41 |
| Air Blank          | 0.000 | 01:43 |
| Subject Sample # 2 | 0.146 | 01:44 |
| Breath Volume      | 1.746 | 01:44 |
| Air Blank          | 0.000 | 01:44 |
| Control Test       | 0.079 | 01:45 |
| Air Blank          | 0.000 | 01:45 |
| Diagnostics Check  | OK    | 01:45 |

---

Data Download Date/Time: 05/13/2021 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: KOPE,MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 04/18/2021 03:09:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:02 |
| Air Blank          | 0.000 | 03:02 |
| Control Test       | 0.080 | 03:02 |
| Air Blank          | 0.000 | 03:03 |
| Subject Sample # 1 | 0.263 | 03:04 |
| Breath Volume      | 2.496 | 03:04 |
| Air Blank          | 0.000 | 03:05 |
| Air Blank          | 0.000 | 03:06 |
| Subject Sample # 2 | 0.259 | 03:07 |
| Breath Volume      | 2.414 | 03:07 |
| Air Blank          | 0.000 | 03:08 |
| Control Test       | 0.079 | 03:08 |
| Air Blank          | 0.000 | 03:09 |
| Diagnostics Check  | OK    | 03:09 |

---

Data Download Date/Time: 05/13/2021 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRETT,CODY,R  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BARRETT,CODY  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/19/2021 01:31:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:24 |
| Air Blank          | 0.000 | 01:24 |
| Control Test       | 0.079 | 01:25 |
| Air Blank          | 0.000 | 01:25 |
| Subject Sample # 1 | 0.251 | 01:27 |
| Breath Volume      | 1.371 | 01:27 |
| Air Blank          | 0.000 | 01:27 |
| Air Blank          | 0.000 | 01:29 |
| Subject Sample # 2 | 0.253 | 01:30 |
| Breath Volume      | 1.335 | 01:30 |
| Air Blank          | 0.000 | 01:30 |
| Control Test       | 0.078 | 01:31 |
| Air Blank          | 0.000 | 01:31 |
| Diagnostics Check  | OK    | 01:31 |

---

Data Download Date/Time: 05/13/2021 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: MOLEN

DATE/TIME OF TEST: 04/19/2021 03:07:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:57 |
| Air Blank          | 0.000 | 02:58 |
| Control Test       | 0.081 | 02:58 |
| Air Blank          | 0.000 | 02:58 |
| Subject Sample # 1 | 0.167 | 02:59 |
| Breath Volume      | 1.121 | 02:59 |
| Air Blank          | 0.000 | 03:00 |
| Air Blank          | 0.000 | 03:02 |
| Subject Sample # 2 | 0.146 | 03:02 |
| Breath Volume      | 1.257 | 03:02 |
| Air Blank          | 0.000 | 03:03 |
| Air Blank          | 0.000 | 03:05 |
| Subject Sample # 3 | 0.149 | 03:05 |
| Breath Volume      | 1.671 | 03:05 |
| Air Blank          | 0.000 | 03:06 |
| Control Test       | 0.080 | 03:06 |
| Air Blank          | 0.000 | 03:07 |
| Diagnostics Check  | OK    | 03:07 |

---

Data Download Date/Time: 05/13/2021 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER III,RICHARD,F  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: SNYDER III,RICHARD  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 04/20/2021 00:14:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:08 |
| Air Blank          | 0.000 | 00:08 |
| Control Test       | 0.080 | 00:09 |
| Air Blank          | 0.000 | 00:09 |
| Subject Sample # 1 | 0.198 | 00:10 |
| Breath Volume      | 3.476 | 00:10 |
| Air Blank          | 0.000 | 00:10 |
| Air Blank          | 0.000 | 00:12 |
| Subject Sample # 2 | 0.205 | 00:12 |
| Breath Volume      | 2.976 | 00:12 |
| Air Blank          | 0.000 | 00:13 |
| Control Test       | 0.080 | 00:14 |
| Air Blank          | 0.000 | 00:14 |
| Diagnostics Check  | OK    | 00:14 |

---

Data Download Date/Time: 05/13/2021 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER III,RICHARD,F  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: SNYDER III,RICHARD  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 04/20/2021 20:38:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:31 |
| Air Blank          | 0.000 | 20:31 |
| Control Test       | 0.080 | 20:32 |
| Air Blank          | 0.000 | 20:32 |
| Subject Sample # 1 | 0.186 | 20:34 |
| Breath Volume      | 2.000 | 20:34 |
| Air Blank          | 0.000 | 20:34 |
| Air Blank          | 0.000 | 20:36 |
| Subject Sample # 2 | 0.180 | 20:37 |
| Breath Volume      | 1.957 | 20:37 |
| Air Blank          | 0.000 | 20:37 |
| Control Test       | 0.079 | 20:38 |
| Air Blank          | 0.000 | 20:38 |
| Diagnostics Check  | OK    | 20:38 |

---

Data Download Date/Time: 05/13/2021 15:48



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRETT,CODY,R  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: KOPE,MICHAEL  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: STEEN

DATE/TIME OF TEST: 04/21/2021 23:03:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:57 |
| Air Blank          | 0.000 | 22:57 |
| Control Test       | 0.081 | 22:58 |
| Air Blank          | 0.000 | 22:58 |
| Subject Sample # 1 | 0.174 | 22:59 |
| Breath Volume      | 2.281 | 22:59 |
| Air Blank          | 0.000 | 22:59 |
| Air Blank          | 0.000 | 23:01 |
| Subject Sample # 2 | 0.171 | 23:02 |
| Breath Volume      | 2.078 | 23:02 |
| Air Blank          | 0.000 | 23:02 |
| Control Test       | 0.080 | 23:03 |
| Air Blank          | 0.000 | 23:03 |
| Diagnostics Check  | OK    | 23:03 |

---

Data Download Date/Time: 05/13/2021 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRETT,CODY,R  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BARRETT,CODY  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: GANT

DATE/TIME OF TEST: 04/22/2021 01:26:30

---

|                   |           |       |                   |
|-------------------|-----------|-------|-------------------|
| Diagnostics Check | OK        | 01:13 |                   |
| Air Blank         | 0.000     | 01:14 |                   |
| Control Test      | 0.081     | 01:14 |                   |
| Air Blank         | 0.000     | 01:15 |                   |
| Subject Sample    | # 1 0.208 | 01:15 | No .020 agreement |
| Breath Volume     | 1.882     | 01:15 |                   |
| Air Blank         | 0.000     | 01:16 |                   |
| Air Blank         | 0.000     | 01:18 |                   |
| Subject Sample    | # 2 0.165 | 01:21 | Volume Not Met    |
| Breath Volume     | 0.363     | 01:21 |                   |
| Air Blank         | 0.000     | 01:22 |                   |
| Air Blank         | 0.000     | 01:23 |                   |
| Subject Sample    | # 3 0.187 | 01:24 | No .020 agreement |
| Breath Volume     | 1.238     | 01:24 |                   |
| Air Blank         | 0.000     | 01:25 |                   |
| Control Test      | 0.078     | 01:25 |                   |
| Air Blank         | 0.000     | 01:26 |                   |
| Diagnostics Check | OK        | 01:26 |                   |

---

Data Download Date/Time: 05/13/2021 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: WELBORN

DATE/TIME OF TEST: 04/22/2021 02:15:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:08 |
| Air Blank          | 0.000 | 02:09 |
| Control Test       | 0.080 | 02:09 |
| Air Blank          | 0.000 | 02:10 |
| Subject Sample # 1 | 0.173 | 02:10 |
| Breath Volume      | 2.148 | 02:10 |
| Air Blank          | 0.000 | 02:11 |
| Air Blank          | 0.000 | 02:13 |
| Subject Sample # 2 | 0.175 | 02:13 |
| Breath Volume      | 1.765 | 02:13 |
| Air Blank          | 0.000 | 02:14 |
| Control Test       | 0.080 | 02:14 |
| Air Blank          | 0.000 | 02:15 |
| Diagnostics Check  | OK    | 02:15 |

---

Data Download Date/Time: 05/13/2021 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: TREJADA RODRIGUEZ

DATE/TIME OF TEST: 04/25/2021 04:19:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:12 |
| Air Blank          | 0.000 | 04:13 |
| Control Test       | 0.081 | 04:13 |
| Air Blank          | 0.000 | 04:13 |
| Subject Sample # 1 | 0.196 | 04:14 |
| Breath Volume      | 2.332 | 04:14 |
| Air Blank          | 0.000 | 04:15 |
| Air Blank          | 0.000 | 04:16 |
| Subject Sample # 2 | 0.194 | 04:17 |
| Breath Volume      | 1.792 | 04:17 |
| Air Blank          | 0.000 | 04:18 |
| Control Test       | 0.080 | 04:18 |
| Air Blank          | 0.000 | 04:18 |
| Diagnostics Check  | OK    | 04:19 |

---

Data Download Date/Time: 05/13/2021 15:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GWODZ,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Bradenton PD  
ARRESTING OFFICER: GWODZ,ROBERT  
ARRESTING OFFICER AGENCY: Bradenton PD  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 04/25/2021 04:40:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:33 |
| Air Blank          | 0.000 | 04:33 |
| Control Test       | 0.080 | 04:34 |
| Air Blank          | 0.000 | 04:34 |
| Subject Sample # 1 | 0.141 | 04:35 |
| Breath Volume      | 1.804 | 04:35 |
| Air Blank          | 0.000 | 04:36 |
| Air Blank          | 0.000 | 04:37 |
| Subject Sample # 2 | 0.148 | 04:38 |
| Breath Volume      | 1.449 | 04:38 |
| Air Blank          | 0.000 | 04:39 |
| Control Test       | 0.080 | 04:39 |
| Air Blank          | 0.000 | 04:40 |
| Diagnostics Check  | OK    | 04:40 |

---

Data Download Date/Time: 05/13/2021 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: DICHRISTOFARO

DATE/TIME OF TEST: 04/27/2021 04:40:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:32 |
| Air Blank          | 0.000 | 04:32 |
| Control Test       | 0.081 | 04:32 |
| Air Blank          | 0.000 | 04:33 |
| Subject Sample # 1 | 0.188 | 04:34 |
| Breath Volume      | 2.957 | 04:34 |
| Air Blank          | 0.000 | 04:35 |
| Air Blank          | 0.000 | 04:37 |
| Subject Sample # 2 | 0.199 | 04:39 |
| Breath Volume      | 3.339 | 04:39 |
| Air Blank          | 0.000 | 04:39 |
| Control Test       | 0.081 | 04:40 |
| Air Blank          | 0.000 | 04:40 |
| Diagnostics Check  | OK    | 04:40 |

---

Data Download Date/Time: 05/13/2021 15:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIGBY,CHARLES,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BIGBY,CHARLES  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 04/28/2021 02:40:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:33 |
| Air Blank          | 0.000 | 02:34 |
| Control Test       | 0.081 | 02:34 |
| Air Blank          | 0.000 | 02:35 |
| Subject Sample # 1 | 0.118 | 02:35 |
| Breath Volume      | 2.460 | 02:35 |
| Air Blank          | 0.000 | 02:36 |
| Air Blank          | 0.000 | 02:38 |
| Subject Sample # 2 | 0.115 | 02:38 |
| Breath Volume      | 2.550 | 02:38 |
| Air Blank          | 0.000 | 02:39 |
| Control Test       | 0.080 | 02:39 |
| Air Blank          | 0.000 | 02:40 |
| Diagnostics Check  | OK    | 02:40 |

---

Data Download Date/Time: 05/13/2021 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: NIEVES

DATE/TIME OF TEST: 04/28/2021 03:27:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:19 |
| Air Blank          | 0.000 | 03:20 |
| Control Test       | 0.081 | 03:20 |
| Air Blank          | 0.000 | 03:21 |
| Subject Sample # 1 | 0.161 | 03:21 |
| Breath Volume      | 2.351 | 03:21 |
| Air Blank          | 0.000 | 03:22 |
| Air Blank          | 0.000 | 03:24 |
| Subject Sample # 2 | 0.162 | 03:25 |
| Breath Volume      | 2.464 | 03:25 |
| Air Blank          | 0.000 | 03:26 |
| Control Test       | 0.081 | 03:27 |
| Air Blank          | 0.000 | 03:27 |
| Diagnostics Check  | OK    | 03:27 |

---

Data Download Date/Time: 05/13/2021 15:48



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER III,RICHARD,F  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: SNYDER III,RICHARD  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BARROW

DATE/TIME OF TEST: 04/30/2021 01:49:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:42 |
| Air Blank          | 0.000 | 01:42 |
| Control Test       | 0.080 | 01:43 |
| Air Blank          | 0.000 | 01:43 |
| Subject Sample # 1 | 0.071 | 01:44 |
| Breath Volume      | 1.335 | 01:44 |
| Air Blank          | 0.000 | 01:45 |
| Air Blank          | 0.000 | 01:46 |
| Subject Sample # 2 | 0.077 | 01:47 |
| Breath Volume      | 1.386 | 01:47 |
| Air Blank          | 0.000 | 01:48 |
| Control Test       | 0.081 | 01:48 |
| Air Blank          | 0.000 | 01:48 |
| Diagnostics Check  | OK    | 01:49 |

---

Data Download Date/Time: 05/13/2021 15:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREENLEE,NANCY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LANDRY,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: CLERVIL

DATE/TIME OF TEST: 04/03/2021 04:21:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:13 |
| Air Blank          | 0.000 | 04:13 |
| Control Test       | 0.081 | 04:13 |
| Air Blank          | 0.000 | 04:14 |
| Subject Sample # 1 | 0.088 | 04:16 |
| Breath Volume      | 1.390 | 04:16 |
| Air Blank          | 0.000 | 04:16 |
| Air Blank          | 0.000 | 04:18 |
| Subject Sample # 2 | 0.084 | 04:19 |
| Breath Volume      | 1.722 | 04:19 |
| Air Blank          | 0.000 | 04:20 |
| Control Test       | 0.081 | 04:20 |
| Air Blank          | 0.000 | 04:21 |
| Diagnostics Check  | OK    | 04:21 |

---

Data Download Date/Time: 04/06/2021 16:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: VANDENBERG III

DATE/TIME OF TEST: 04/05/2021 03:46:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:39 |
| Air Blank          | 0.000 | 03:39 |
| Control Test       | 0.080 | 03:39 |
| Air Blank          | 0.000 | 03:40 |
| Subject Sample # 1 | 0.083 | 03:41 |
| Breath Volume      | 3.335 | 03:41 |
| Air Blank          | 0.000 | 03:42 |
| Air Blank          | 0.000 | 03:44 |
| Subject Sample # 2 | 0.084 | 03:44 |
| Breath Volume      | 3.695 | 03:44 |
| Air Blank          | 0.000 | 03:45 |
| Control Test       | 0.080 | 03:45 |
| Air Blank          | 0.000 | 03:46 |
| Diagnostics Check  | OK    | 03:46 |

---

Data Download Date/Time: 04/06/2021 16:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONEY, ANIA, L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VANTREESE, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: EISINGER

DATE/TIME OF TEST: 04/09/2021 01:22:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:15 |
| Air Blank          | 0.000 | 01:15 |
| Control Test       | 0.081 | 01:15 |
| Air Blank          | 0.000 | 01:16 |
| Subject Sample # 1 | 0.058 | 01:17 |
| Breath Volume      | 1.250 | 01:17 |
| Air Blank          | 0.000 | 01:18 |
| Air Blank          | 0.000 | 01:19 |
| Subject Sample # 2 | 0.060 | 01:20 |
| Breath Volume      | 2.187 | 01:20 |
| Air Blank          | 0.000 | 01:21 |
| Control Test       | 0.080 | 01:21 |
| Air Blank          | 0.000 | 01:22 |
| Diagnostics Check  | OK    | 01:22 |

---

Data Download Date/Time: 05/04/2021 15:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONEY,ANIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: O'MALLEY

DATE/TIME OF TEST: 04/10/2021 03:43:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:36 |
| Air Blank          | 0.000 | 03:37 |
| Control Test       | 0.081 | 03:37 |
| Air Blank          | 0.000 | 03:38 |
| Subject Sample # 1 | 0.084 | 03:39 |
| Breath Volume      | 1.230 | 03:39 |
| Air Blank          | 0.000 | 03:39 |
| Air Blank          | 0.000 | 03:41 |
| Subject Sample # 2 | 0.076 | 03:42 |
| Breath Volume      | 1.488 | 03:42 |
| Air Blank          | 0.000 | 03:42 |
| Control Test       | 0.080 | 03:43 |
| Air Blank          | 0.000 | 03:43 |
| Diagnostics Check  | OK    | 03:43 |

---

Data Download Date/Time: 05/04/2021 15:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LANDRY,M  
ARRESTING OFFICER AGENCY: Temple Terrace PD  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 04/10/2021 04:10:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:03 |
| Air Blank          | 0.000 | 04:03 |
| Control Test       | 0.081 | 04:03 |
| Air Blank          | 0.000 | 04:04 |
| Subject Sample # 1 | 0.100 | 04:05 |
| Breath Volume      | 3.039 | 04:05 |
| Air Blank          | 0.000 | 04:06 |
| Air Blank          | 0.000 | 04:08 |
| Subject Sample # 2 | 0.098 | 04:08 |
| Breath Volume      | 1.339 | 04:08 |
| Air Blank          | 0.000 | 04:09 |
| Control Test       | 0.080 | 04:09 |
| Air Blank          | 0.000 | 04:10 |
| Diagnostics Check  | OK    | 04:10 |

---

Data Download Date/Time: 05/04/2021 15:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KNOX,STEPHANIE,R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BRABAND,M  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: ROMERO SANDOVAL

DATE/TIME OF TEST: 04/15/2021 02:46:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:37 |
| Air Blank          | 0.000 | 02:38 |
| Control Test       | 0.079 | 02:38 |
| Air Blank          | 0.000 | 02:38 |
| Subject Sample # 1 | 0.169 | 02:40 |
| Breath Volume      | 2.890 | 02:40 |
| Air Blank          | 0.000 | 02:41 |
| Air Blank          | 0.000 | 02:42 |
| Subject Sample # 2 | 0.154 | 02:44 |
| Breath Volume      | 1.390 | 02:44 |
| Air Blank          | 0.000 | 02:45 |
| Control Test       | 0.079 | 02:45 |
| Air Blank          | 0.000 | 02:46 |
| Diagnostics Check  | OK    | 02:46 |

---

Data Download Date/Time: 05/04/2021 15:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GARNER,W  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: ORTEGA

DATE/TIME OF TEST: 04/17/2021 04:44:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:38 |
| Air Blank          | 0.000 | 04:38 |
| Control Test       | 0.081 | 04:38 |
| Air Blank          | 0.000 | 04:39 |
| Subject Sample # 1 | 0.060 | 04:40 |
| Breath Volume      | 2.859 | 04:40 |
| Air Blank          | 0.000 | 04:40 |
| Air Blank          | 0.000 | 04:42 |
| Subject Sample # 2 | 0.058 | 04:43 |
| Breath Volume      | 2.863 | 04:43 |
| Air Blank          | 0.000 | 04:43 |
| Control Test       | 0.080 | 04:44 |
| Air Blank          | 0.000 | 04:44 |
| Diagnostics Check  | OK    | 04:44 |

---

Data Download Date/Time: 05/04/2021 15:22



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS,ANGELA,V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HALL,L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:22

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 04/17/2021 07:53:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:43 |
| Air Blank          | 0.000 | 07:44 |
| Control Test       | 0.080 | 07:44 |
| Air Blank          | 0.000 | 07:45 |
| Subject Sample # 1 | 0.175 | 07:46 |
| Breath Volume      | 2.148 | 07:46 |
| Air Blank          | 0.000 | 07:46 |
| Air Blank          | 0.000 | 07:48 |
| Subject Sample # 2 | 0.151 | 07:49 |
| Breath Volume      | 1.515 | 07:49 |
| Air Blank          | 0.000 | 07:49 |
| Air Blank          | 0.000 | 07:51 |
| Subject Sample # 3 | 0.164 | 07:52 |
| Breath Volume      | 1.250 | 07:52 |
| Air Blank          | 0.000 | 07:52 |
| Control Test       | 0.080 | 07:53 |
| Air Blank          | 0.000 | 07:53 |
| Diagnostics Check  | OK    | 07:53 |

---

Data Download Date/Time: 05/04/2021 15:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONEY,ANIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: RIVERA-ROJAS

DATE/TIME OF TEST: 04/18/2021 00:49:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:42 |
| Air Blank          | 0.000 | 00:42 |
| Control Test       | 0.081 | 00:43 |
| Air Blank          | 0.000 | 00:43 |
| Subject Sample # 1 | 0.202 | 00:44 |
| Breath Volume      | 1.214 | 00:44 |
| Air Blank          | 0.000 | 00:45 |
| Air Blank          | 0.000 | 00:46 |
| Subject Sample # 2 | 0.194 | 00:47 |
| Breath Volume      | 1.808 | 00:47 |
| Air Blank          | 0.000 | 00:48 |
| Control Test       | 0.079 | 00:48 |
| Air Blank          | 0.000 | 00:49 |
| Diagnostics Check  | OK    | 00:49 |

---

Data Download Date/Time: 05/04/2021 15:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KNOX,STEPHANIE,R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/18/2021 03:37:19

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 03:24 |                |
| Air Blank         | 0.000     | 03:25 |                |
| Control Test      | 0.080     | 03:25 |                |
| Air Blank         | 0.000     | 03:25 |                |
| Subject Sample    | # 1 0.090 | 03:28 | Volume Not Met |
| Breath Volume     | 0.050     | 03:28 |                |
| Air Blank         | 0.000     | 03:29 |                |
| Air Blank         | 0.000     | 03:31 |                |
| Subject Sample    | # 2 0.081 | 03:31 |                |
| Breath Volume     | 1.972     | 03:31 |                |
| Air Blank         | 0.000     | 03:32 |                |
| Air Blank         | 0.000     | 03:34 |                |
| Subject Sample    | # 3 0.081 | 03:35 |                |
| Breath Volume     | 1.601     | 03:35 |                |
| Air Blank         | 0.000     | 03:36 |                |
| Control Test      | 0.080     | 03:36 |                |
| Air Blank         | 0.000     | 03:37 |                |
| Diagnostics Check | OK        | 03:37 |                |

---

Data Download Date/Time: 05/04/2021 15:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENDY,HALEY,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GARRELL,Q  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: SHERMAN

DATE/TIME OF TEST: 04/18/2021 23:56:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:48 |
| Air Blank          | 0.000 | 23:49 |
| Control Test       | 0.080 | 23:49 |
| Air Blank          | 0.000 | 23:50 |
| Subject Sample # 1 | 0.355 | 23:51 |
| Breath Volume      | 1.179 | 23:51 |
| Air Blank          | 0.000 | 23:51 |
| Air Blank          | 0.000 | 23:53 |
| Subject Sample # 2 | 0.346 | 23:54 |
| Breath Volume      | 1.390 | 23:54 |
| Air Blank          | 0.000 | 23:55 |
| Control Test       | 0.079 | 23:55 |
| Air Blank          | 0.000 | 23:56 |
| Diagnostics Check  | OK    | 23:56 |

---

Data Download Date/Time: 05/04/2021 15:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAIL,JULIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HAUGSTAD,G  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 14:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: PAGAN

DATE/TIME OF TEST: 04/21/2021 15:07:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:00 |
| Air Blank          | 0.000 | 15:01 |
| Control Test       | 0.081 | 15:01 |
| Air Blank          | 0.000 | 15:02 |
| Subject Sample # 1 | 0.124 | 15:02 |
| Breath Volume      | 1.343 | 15:02 |
| Air Blank          | 0.000 | 15:03 |
| Air Blank          | 0.000 | 15:05 |
| Subject Sample # 2 | 0.107 | 15:05 |
| Breath Volume      | 1.179 | 15:05 |
| Air Blank          | 0.000 | 15:06 |
| Control Test       | 0.080 | 15:06 |
| Air Blank          | 0.000 | 15:07 |
| Diagnostics Check  | OK    | 15:07 |

---

Data Download Date/Time: 05/04/2021 15:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DOBOSIEWICZ,A  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 04/21/2021 22:45:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:38 |
| Air Blank          | 0.000 | 22:38 |
| Control Test       | 0.081 | 22:38 |
| Air Blank          | 0.000 | 22:39 |
| Subject Sample # 1 | 0.123 | 22:40 |
| Breath Volume      | 2.429 | 22:40 |
| Air Blank          | 0.000 | 22:41 |
| Air Blank          | 0.000 | 22:43 |
| Subject Sample # 2 | 0.132 | 22:43 |
| Breath Volume      | 2.652 | 22:43 |
| Air Blank          | 0.000 | 22:44 |
| Control Test       | 0.081 | 22:44 |
| Air Blank          | 0.000 | 22:45 |
| Diagnostics Check  | OK    | 22:45 |

---

Data Download Date/Time: 05/04/2021 15:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMONS,ANGELA,V  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BRUCE,KEATON  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: MARRINER

DATE/TIME OF TEST: 04/22/2021 17:49:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:43 |
| Air Blank          | 0.000 | 17:43 |
| Control Test       | 0.080 | 17:44 |
| Air Blank          | 0.000 | 17:44 |
| Subject Sample # 1 | 0.219 | 17:45 |
| Breath Volume      | 1.886 | 17:45 |
| Air Blank          | 0.000 | 17:45 |
| Air Blank          | 0.000 | 17:47 |
| Subject Sample # 2 | 0.226 | 17:47 |
| Breath Volume      | 1.695 | 17:47 |
| Air Blank          | 0.000 | 17:48 |
| Control Test       | 0.080 | 17:49 |
| Air Blank          | 0.000 | 17:49 |
| Diagnostics Check  | OK    | 17:49 |

---

Data Download Date/Time: 05/04/2021 15:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KNOX,STEPHANIE,R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GARNER,W  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: VERA RINCON

DATE/TIME OF TEST: 04/23/2021 05:19:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:12 |
| Air Blank          | 0.000 | 05:13 |
| Control Test       | 0.081 | 05:13 |
| Air Blank          | 0.000 | 05:14 |
| Subject Sample # 1 | 0.092 | 05:15 |
| Breath Volume      | 1.824 | 05:15 |
| Air Blank          | 0.000 | 05:15 |
| Air Blank          | 0.000 | 05:17 |
| Subject Sample # 2 | 0.093 | 05:18 |
| Breath Volume      | 2.765 | 05:18 |
| Air Blank          | 0.000 | 05:18 |
| Control Test       | 0.080 | 05:19 |
| Air Blank          | 0.000 | 05:19 |
| Diagnostics Check  | OK    | 05:19 |

---

Data Download Date/Time: 05/04/2021 15:22



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONEY,ANIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BRABAND,M  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: COLE JR.

DATE/TIME OF TEST: 04/23/2021 21:07:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:01 |
| Air Blank          | 0.000 | 21:01 |
| Control Test       | 0.081 | 21:02 |
| Air Blank          | 0.000 | 21:02 |
| Subject Sample # 1 | 0.183 | 21:03 |
| Breath Volume      | 3.242 | 21:03 |
| Air Blank          | 0.000 | 21:04 |
| Air Blank          | 0.000 | 21:05 |
| Subject Sample # 2 | 0.176 | 21:06 |
| Breath Volume      | 2.855 | 21:06 |
| Air Blank          | 0.000 | 21:07 |
| Control Test       | 0.080 | 21:07 |
| Air Blank          | 0.000 | 21:07 |
| Diagnostics Check  | OK    | 21:07 |

---

Data Download Date/Time: 05/04/2021 15:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BARLAUG,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: GOMES BARBOSA

DATE/TIME OF TEST: 04/24/2021 05:05:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:57 |
| Air Blank          | 0.000 | 04:58 |
| Control Test       | 0.080 | 04:58 |
| Air Blank          | 0.000 | 04:59 |
| Subject Sample # 1 | 0.129 | 05:00 |
| Breath Volume      | 2.472 | 05:00 |
| Air Blank          | 0.000 | 05:01 |
| Air Blank          | 0.000 | 05:02 |
| Subject Sample # 2 | 0.131 | 05:03 |
| Breath Volume      | 2.128 | 05:03 |
| Air Blank          | 0.000 | 05:04 |
| Control Test       | 0.079 | 05:04 |
| Air Blank          | 0.000 | 05:05 |
| Diagnostics Check  | OK    | 05:05 |

---

Data Download Date/Time: 05/04/2021 15:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KNOX,STEPHANIE,R  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HUTCHESON,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: OCONNOR

DATE/TIME OF TEST: 04/28/2021 20:56:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:49 |
| Air Blank          | 0.000 | 20:50 |
| Control Test       | 0.080 | 20:50 |
| Air Blank          | 0.000 | 20:51 |
| Subject Sample # 1 | 0.130 | 20:51 |
| Breath Volume      | 1.277 | 20:51 |
| Air Blank          | 0.000 | 20:52 |
| Air Blank          | 0.000 | 20:54 |
| Subject Sample # 2 | 0.149 | 20:55 |
| Breath Volume      | 1.437 | 20:55 |
| Air Blank          | 0.000 | 20:55 |
| Control Test       | 0.079 | 20:56 |
| Air Blank          | 0.000 | 20:56 |
| Diagnostics Check  | OK    | 20:56 |

---

Data Download Date/Time: 05/04/2021 15:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONEY,ANIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: JORDAN CANADA

DATE/TIME OF TEST: 04/29/2021 00:31:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:23 |
| Air Blank          | 0.000 | 00:23 |
| Control Test       | 0.080 | 00:23 |
| Air Blank          | 0.000 | 00:24 |
| Subject Sample # 1 | 0.093 | 00:26 |
| Breath Volume      | 1.234 | 00:26 |
| Air Blank          | 0.000 | 00:27 |
| Air Blank          | 0.000 | 00:28 |
| Subject Sample # 2 | 0.093 | 00:29 |
| Breath Volume      | 1.375 | 00:29 |
| Air Blank          | 0.000 | 00:30 |
| Control Test       | 0.080 | 00:30 |
| Air Blank          | 0.000 | 00:31 |
| Diagnostics Check  | OK    | 00:31 |

---

Data Download Date/Time: 05/04/2021 15:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASON,FRANCESCA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ANDREWS,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: KIRKLAND

DATE/TIME OF TEST: 04/30/2021 05:57:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:50 |
| Air Blank          | 0.000 | 05:51 |
| Control Test       | 0.081 | 05:51 |
| Air Blank          | 0.000 | 05:52 |
| Subject Sample # 1 | 0.206 | 05:52 |
| Breath Volume      | 1.910 | 05:52 |
| Air Blank          | 0.000 | 05:53 |
| Air Blank          | 0.000 | 05:55 |
| Subject Sample # 2 | 0.209 | 05:55 |
| Breath Volume      | 2.292 | 05:55 |
| Air Blank          | 0.000 | 05:56 |
| Control Test       | 0.081 | 05:56 |
| Air Blank          | 0.000 | 05:57 |
| Diagnostics Check  | OK    | 05:57 |

---

Data Download Date/Time: 05/04/2021 15:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAINEY,CRISTA,N  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: GUERRERO,MARIA  
ARRESTING OFFICER AGENCY: Flagler Beach PD  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: O'CONNOR

DATE/TIME OF TEST: 04/01/2021 23:30:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:23 |
| Air Blank          | 0.000 | 23:24 |
| Control Test       | 0.081 | 23:24 |
| Air Blank          | 0.000 | 23:25 |
| Subject Sample # 1 | 0.262 | 23:25 |
| Breath Volume      | 3.652 | 23:25 |
| Air Blank          | 0.000 | 23:26 |
| Air Blank          | 0.000 | 23:28 |
| Subject Sample # 2 | 0.251 | 23:28 |
| Breath Volume      | 1.488 | 23:28 |
| Air Blank          | 0.000 | 23:29 |
| Control Test       | 0.080 | 23:29 |
| Air Blank          | 0.000 | 23:30 |
| Diagnostics Check  | OK    | 23:30 |

---

Data Download Date/Time: 04/02/2021 09:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAINEY,CRISTA,N  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: GUERRERO,MARIA  
ARRESTING OFFICER AGENCY: Flagler Beach PD  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: O'CONNOR

DATE/TIME OF TEST: 04/02/2021 00:02:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:52 |
| Air Blank          | 0.000 | 23:53 |
| Control Test       | 0.080 | 23:53 |
| Air Blank          | 0.000 | 23:54 |
| Subject Sample # 2 | 0.241 | 23:54 |
| Breath Volume      | 1.906 | 23:54 |
| Air Blank          | 0.000 | 23:55 |
| Air Blank          | 0.000 | 23:57 |
| Subject Sample # 3 | 0.265 | 23:57 |
| Breath Volume      | 2.308 | 23:57 |
| Air Blank          | 0.000 | 23:58 |
| Air Blank          | 0.000 | 00:00 |
| Subject Sample # 1 | 0.221 | 00:00 |
| Breath Volume      | 1.796 | 00:00 |
| Air Blank          | 0.000 | 00:01 |
| Control Test       | 0.078 | 00:01 |
| Air Blank          | 0.000 | 00:02 |
| Diagnostics Check  | OK    | 00:02 |

---

Data Download Date/Time: 04/02/2021 09:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,CARL,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PARKER,CARL  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: SOUSA

DATE/TIME OF TEST: 04/03/2021 05:38:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:30 |
| Air Blank          | 0.000 | 05:30 |
| Control Test       | 0.081 | 05:31 |
| Air Blank          | 0.000 | 05:31 |
| Subject Sample # 1 | 0.174 | 05:32 |
| Breath Volume      | 1.246 | 05:32 |
| Air Blank          | 0.000 | 05:32 |
| Air Blank          | 0.000 | 05:34 |
| Subject Sample # 2 | 0.184 | 05:36 |
| Breath Volume      | 2.964 | 05:36 |
| Air Blank          | 0.000 | 05:37 |
| Control Test       | 0.080 | 05:37 |
| Air Blank          | 0.000 | 05:38 |
| Diagnostics Check  | OK    | 05:38 |

---

Data Download Date/Time: 05/03/2021 16:52



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ST. JOHN,J,M  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: ST. JOHN,J  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 13:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: TARTER

DATE/TIME OF TEST: 04/03/2021 14:35:07

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 14:29 |                |
| Air Blank         | 0.000     | 14:29 |                |
| Control Test      | 0.081     | 14:30 |                |
| Air Blank         | 0.000     | 14:30 |                |
| Subject Sample    | # 1 0.173 | 14:33 | Volume Not Met |
| Breath Volume     | 0.828     | 14:33 |                |
| Air Blank         |           | 14:34 | Purge Fail     |
| Air Blank         | 0.000     | 14:35 |                |

---

Data Download Date/Time: 05/03/2021 16:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONG,JOSHUA,A  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: SCHERR,EVAN  
ARRESTING OFFICER AGENCY: Flagler Beach PD  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: RUBRIGHT

DATE/TIME OF TEST: 04/03/2021 18:24:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:16 |
| Air Blank          | 0.000 | 18:17 |
| Control Test       | 0.080 | 18:17 |
| Air Blank          | 0.000 | 18:18 |
| Subject Sample # 1 | 0.217 | 18:18 |
| Breath Volume      | 1.382 | 18:18 |
| Air Blank          | 0.000 | 18:19 |
| Air Blank          | 0.000 | 18:21 |
| Subject Sample # 2 | 0.225 | 18:22 |
| Breath Volume      | 1.171 | 18:22 |
| Air Blank          | 0.000 | 18:23 |
| Control Test       | 0.079 | 18:23 |
| Air Blank          | 0.000 | 18:24 |
| Diagnostics Check  | OK    | 18:24 |

---

Data Download Date/Time: 05/03/2021 16:52

# **Florida Department of Law Enforcement**

## **Alcohol Testing Program**

### **Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESTEP,J,D  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: CAPELA,J  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: COPE

DATE/TIME OF TEST: 04/04/2021 04:49:49

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 04:48 |            |
| Air Blank         |       | 04:49 | RFI Detect |
| Air Blank         | 0.000 | 04:49 |            |

---

Data Download Date/Time: 05/03/2021 16:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESTEP,J,D  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: CAPELA,J  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: COPE

DATE/TIME OF TEST: 04/04/2021 04:59:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:53 |
| Air Blank          | 0.000 | 04:53 |
| Control Test       | 0.081 | 04:54 |
| Air Blank          | 0.000 | 04:54 |
| Subject Sample # 1 | 0.185 | 04:55 |
| Breath Volume      | 2.917 | 04:55 |
| Air Blank          | 0.000 | 04:56 |
| Air Blank          | 0.000 | 04:57 |
| Subject Sample # 2 | 0.175 | 04:58 |
| Breath Volume      | 2.601 | 04:58 |
| Air Blank          | 0.000 | 04:59 |
| Control Test       | 0.078 | 04:59 |
| Air Blank          | 0.000 | 04:59 |
| Diagnostics Check  | OK    | 04:59 |

---

Data Download Date/Time: 05/03/2021 16:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NICOLAS,MITA,  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: MUSIAL,C  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/07/2021 00:00:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:53 |
| Air Blank          | 0.000 | 23:54 |
| Control Test       | 0.080 | 23:54 |
| Air Blank          | 0.000 | 23:55 |
| Subject Sample # 1 | 0.182 | 23:55 |
| Breath Volume      | 1.328 | 23:55 |
| Air Blank          | 0.000 | 23:56 |
| Air Blank          | 0.000 | 23:58 |
| Subject Sample # 2 | 0.177 | 23:58 |
| Breath Volume      | 2.273 | 23:58 |
| Air Blank          | 0.000 | 23:59 |
| Control Test       | 0.079 | 23:59 |
| Air Blank          | 0.000 | 00:00 |
| Diagnostics Check  | OK    | 00:00 |

---

Data Download Date/Time: 05/03/2021 16:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STOGDON,BRAD,M  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: ZALAK,JONATHAN  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:58

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 04/10/2021 16:24:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:18 |
| Air Blank          | 0.000 | 16:19 |
| Control Test       | 0.080 | 16:19 |
| Air Blank          | 0.000 | 16:19 |
| Subject Sample # 1 | 0.205 | 16:20 |
| Breath Volume      | 1.210 | 16:20 |
| Air Blank          | 0.000 | 16:20 |
| Air Blank          | 0.000 | 16:22 |
| Subject Sample # 2 | 0.215 | 16:23 |
| Breath Volume      | 1.457 | 16:23 |
| Air Blank          | 0.000 | 16:23 |
| Control Test       | 0.080 | 16:24 |
| Air Blank          | 0.000 | 16:24 |
| Diagnostics Check  | OK    | 16:24 |

---

Data Download Date/Time: 05/03/2021 16:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,BRYAN,M  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: HARPER,DARIN  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 04/11/2021 02:24:02

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:19 |                      |
| Air Blank         | 0.000 | 02:19 |                      |
| Control Test      | 0.080 | 02:20 |                      |
| Air Blank         | 0.000 | 02:20 |                      |
| Subject Sample    | # 1   | 02:22 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:22 |                      |
| Air Blank         | 0.000 | 02:23 |                      |
| Control Test      | 0.080 | 02:23 |                      |
| Air Blank         | 0.000 | 02:23 |                      |
| Diagnostics Check | OK    | 02:24 |                      |

---

Data Download Date/Time: 05/03/2021 16:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,BRYAN,M  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: CLAY,AARON  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: DE MERS

DATE/TIME OF TEST: 04/12/2021 02:35:11

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:26 |                      |
| Air Blank         | 0.000 | 02:26 |                      |
| Control Test      | 0.080 | 02:26 |                      |
| Air Blank         | 0.000 | 02:27 |                      |
| Subject Sample    | # 1   | 02:30 | No Sample Provided   |
| Breath Volume     | 0.000 | 02:30 |                      |
| Air Blank         | 0.000 | 02:30 |                      |
| Air Blank         | 0.000 | 02:32 |                      |
| Subject Sample    | # 2   | 02:33 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:33 |                      |
| Air Blank         | 0.000 | 02:34 |                      |
| Control Test      | 0.081 | 02:34 |                      |
| Air Blank         | 0.000 | 02:35 |                      |
| Diagnostics Check | OK    | 02:35 |                      |

---

Data Download Date/Time: 05/03/2021 16:52



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAPELA,JORDAN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: CAPELA,JORDAN  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 04/14/2021 02:38:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:32 |
| Air Blank          | 0.000 | 02:32 |
| Control Test       | 0.081 | 02:32 |
| Air Blank          | 0.000 | 02:33 |
| Subject Sample # 1 | 0.154 | 02:33 |
| Breath Volume      | 3.085 | 02:33 |
| Air Blank          | 0.000 | 02:34 |
| Air Blank          | 0.000 | 02:36 |
| Subject Sample # 2 | 0.151 | 02:36 |
| Breath Volume      | 3.781 | 02:36 |
| Air Blank          | 0.000 | 02:37 |
| Control Test       | 0.079 | 02:38 |
| Air Blank          | 0.000 | 02:38 |
| Diagnostics Check  | OK    | 02:38 |

---

Data Download Date/Time: 05/03/2021 16:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAUN,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: ZALAK,JONATHAN  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:59

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 04/14/2021 17:26:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:20 |
| Air Blank          | 0.000 | 17:20 |
| Control Test       | 0.080 | 17:20 |
| Air Blank          | 0.000 | 17:21 |
| Subject Sample # 1 | 0.231 | 17:21 |
| Breath Volume      | 3.132 | 17:21 |
| Air Blank          | 0.000 | 17:22 |
| Air Blank          | 0.000 | 17:24 |
| Subject Sample # 2 | 0.237 | 17:24 |
| Breath Volume      | 2.203 | 17:24 |
| Air Blank          | 0.000 | 17:25 |
| Control Test       | 0.080 | 17:25 |
| Air Blank          | 0.000 | 17:26 |
| Diagnostics Check  | OK    | 17:26 |

---

Data Download Date/Time: 05/03/2021 16:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAPELA,JORDAN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: CAPELA,JORDAN  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 04/18/2021 03:10:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:04 |
| Air Blank          | 0.000 | 03:04 |
| Control Test       | 0.080 | 03:05 |
| Air Blank          | 0.000 | 03:05 |
| Subject Sample # 1 | 0.027 | 03:06 |
| Breath Volume      | 1.472 | 03:06 |
| Air Blank          | 0.000 | 03:06 |
| Air Blank          | 0.000 | 03:08 |
| Subject Sample # 2 | 0.027 | 03:09 |
| Breath Volume      | 1.628 | 03:09 |
| Air Blank          | 0.000 | 03:09 |
| Control Test       | 0.079 | 03:10 |
| Air Blank          | 0.000 | 03:10 |
| Diagnostics Check  | OK    | 03:10 |

---

Data Download Date/Time: 05/03/2021 16:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STRACK,ROBERT,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: STRACK,ROBERT  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 10:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: STRUBLE

DATE/TIME OF TEST: 04/18/2021 11:41:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 11:35 |
| Air Blank          | 0.000 | 11:36 |
| Control Test       | 0.080 | 11:36 |
| Air Blank          | 0.000 | 11:36 |
| Subject Sample # 1 | 0.000 | 11:37 |
| Breath Volume      | 1.328 | 11:37 |
| Air Blank          | 0.000 | 11:37 |
| Air Blank          | 0.000 | 11:39 |
| Subject Sample # 2 | 0.000 | 11:40 |
| Breath Volume      | 1.726 | 11:40 |
| Air Blank          | 0.000 | 11:40 |
| Control Test       | 0.078 | 11:40 |
| Air Blank          | 0.000 | 11:41 |
| Diagnostics Check  | OK    | 11:41 |

---

Data Download Date/Time: 05/03/2021 16:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY, KENNETH, R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: GADDIE, K  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: WYATT

DATE/TIME OF TEST: 04/18/2021 20:55:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:48 |
| Air Blank          | 0.000 | 20:48 |
| Control Test       | 0.080 | 20:49 |
| Air Blank          | 0.000 | 20:49 |
| Subject Sample # 1 | 0.130 | 20:50 |
| Breath Volume      | 2.566 | 20:50 |
| Air Blank          | 0.000 | 20:50 |
| Air Blank          | 0.000 | 20:52 |
| Subject Sample # 2 | 0.116 | 20:53 |
| Breath Volume      | 1.273 | 20:53 |
| Air Blank          | 0.000 | 20:54 |
| Control Test       | 0.079 | 20:54 |
| Air Blank          | 0.000 | 20:55 |
| Diagnostics Check  | OK    | 20:55 |

---

Data Download Date/Time: 05/03/2021 16:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,CARL,L  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PARKER,CARL  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: DANLOVICH

DATE/TIME OF TEST: 04/19/2021 01:23:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:17 |
| Air Blank          | 0.000 | 01:17 |
| Control Test       | 0.080 | 01:17 |
| Air Blank          | 0.000 | 01:18 |
| Subject Sample # 1 | 0.081 | 01:18 |
| Breath Volume      | 3.175 | 01:18 |
| Air Blank          | 0.000 | 01:19 |
| Air Blank          | 0.000 | 01:21 |
| Subject Sample # 2 | 0.078 | 01:21 |
| Breath Volume      | 2.636 | 01:21 |
| Air Blank          | 0.000 | 01:22 |
| Control Test       | 0.080 | 01:22 |
| Air Blank          | 0.000 | 01:23 |
| Diagnostics Check  | OK    | 01:23 |

---

Data Download Date/Time: 05/03/2021 16:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOTOWSKI, PHILIP, P  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: SLAGA, JASMINE  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:06

SUBJECT'S LAST NAME: NOBLE

DATE/TIME OF TEST: 04/20/2021 16:33:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:26 |
| Air Blank          | 0.000 | 16:27 |
| Control Test       | 0.080 | 16:27 |
| Air Blank          | 0.000 | 16:27 |
| Subject Sample # 1 | 0.130 | 16:28 |
| Breath Volume      | 1.441 | 16:28 |
| Air Blank          | 0.000 | 16:29 |
| Air Blank          | 0.000 | 16:31 |
| Subject Sample # 2 | 0.131 | 16:31 |
| Breath Volume      | 1.953 | 16:31 |
| Air Blank          | 0.000 | 16:32 |
| Control Test       | 0.080 | 16:32 |
| Air Blank          | 0.000 | 16:33 |
| Diagnostics Check  | OK    | 16:33 |

---

Data Download Date/Time: 05/03/2021 16:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOTOWSKI, PHILIP, P  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: WICHMAN, MICHELLE  
ARRESTING OFFICER AGENCY: Bunnell PD  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 04/20/2021 17:07:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:01 |
| Air Blank          | 0.000 | 17:01 |
| Control Test       | 0.080 | 17:01 |
| Air Blank          | 0.000 | 17:02 |
| Subject Sample # 1 | 0.099 | 17:03 |
| Breath Volume      | 1.437 | 17:03 |
| Air Blank          | 0.000 | 17:03 |
| Air Blank          | 0.000 | 17:05 |
| Subject Sample # 2 | 0.092 | 17:06 |
| Breath Volume      | 1.484 | 17:06 |
| Air Blank          | 0.000 | 17:07 |
| Control Test       | 0.080 | 17:07 |
| Air Blank          | 0.000 | 17:07 |
| Diagnostics Check  | OK    | 17:07 |

---

Data Download Date/Time: 05/03/2021 16:53



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOTOWSKI, PHILIP, P  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: KOTOWSKI, PHILIP  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: DARE

DATE/TIME OF TEST: 04/23/2021 16:40:48

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:33 |
| Air Blank          | 0.000 | 16:34 |
| Control Test       | 0.081 | 16:34 |
| Air Blank          | 0.000 | 16:35 |
| Subject Sample # 1 | 0.231 | 16:35 |
| Breath Volume      | 2.875 | 16:35 |
| Air Blank          | 0.000 | 16:36 |
| Air Blank          | 0.000 | 16:38 |
| Subject Sample # 2 | 0.236 | 16:39 |
| Breath Volume      | 2.792 | 16:39 |
| Air Blank          | 0.000 | 16:39 |
| Control Test       | 0.079 | 16:40 |
| Air Blank          | 0.000 | 16:40 |
| Diagnostics Check  | OK    | 16:40 |

---

Data Download Date/Time: 05/03/2021 16:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,BRYAN,M  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: CARTER,BRYAN  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: GOEHRING

DATE/TIME OF TEST: 04/25/2021 22:05:36

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 21:53 |                |
| Air Blank         | 0.000     | 21:54 |                |
| Control Test      | 0.080     | 21:54 |                |
| Air Blank         | 0.000     | 21:55 |                |
| Subject Sample    | # 1 0.144 | 21:58 | Volume Not Met |
| Breath Volume     | 0.097     | 21:58 |                |
| Air Blank         | 0.000     | 21:58 |                |
| Air Blank         | 0.000     | 22:00 |                |
| Subject Sample    | # 2 0.143 | 22:03 | Volume Not Met |
| Breath Volume     | 0.179     | 22:03 |                |
| Air Blank         | 0.000     | 22:04 |                |
| Control Test      | 0.079     | 22:04 |                |
| Air Blank         | 0.000     | 22:05 |                |
| Diagnostics Check | OK        | 22:05 |                |

---

Data Download Date/Time: 05/03/2021 16:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAPELA,JORDAN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: CAPELA,JORDAN  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 04/27/2021 02:17:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:10 |
| Air Blank          | 0.000 | 02:11 |
| Control Test       | 0.081 | 02:11 |
| Air Blank          | 0.000 | 02:12 |
| Subject Sample # 1 | 0.084 | 02:12 |
| Breath Volume      | 1.449 | 02:12 |
| Air Blank          | 0.000 | 02:13 |
| Air Blank          | 0.000 | 02:15 |
| Subject Sample # 2 | 0.086 | 02:16 |
| Breath Volume      | 1.632 | 02:16 |
| Air Blank          | 0.000 | 02:16 |
| Control Test       | 0.081 | 02:17 |
| Air Blank          | 0.000 | 02:17 |
| Diagnostics Check  | OK    | 02:17 |

---

Data Download Date/Time: 05/03/2021 16:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONG,JOSHUA,A  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: SCHERR,EVAN  
ARRESTING OFFICER AGENCY: Flagler Beach PD  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: ALCANTAR

DATE/TIME OF TEST: 04/27/2021 17:26:59

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 17:24 |                      |
| Air Blank         | 0.000 | 17:24 |                      |
| Control Test      | 0.081 | 17:24 |                      |
| Air Blank         | 0.000 | 17:25 |                      |
| Subject Sample    | # 1   | 17:25 | Subject Test Refused |
| Breath Volume     | 0.000 | 17:25 |                      |
| Air Blank         | 0.000 | 17:26 |                      |
| Control Test      | 0.080 | 17:26 |                      |
| Air Blank         | 0.000 | 17:26 |                      |
| Diagnostics Check | OK    | 17:26 |                      |

---

Data Download Date/Time: 05/03/2021 16:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAPELA,JORDAN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: CAPELA,JORDAN  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: OCONNOR

DATE/TIME OF TEST: 04/28/2021 01:38:25

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 01:27 |                      |
| Air Blank         | 0.000     | 01:27 |                      |
| Control Test      | 0.080     | 01:28 |                      |
| Air Blank         | 0.000     | 01:28 |                      |
| Subject Sample    | # 1 0.223 | 01:31 | Volume Not Met       |
| Breath Volume     | 0.238     | 01:31 |                      |
| Air Blank         | 0.000     | 01:32 |                      |
| Air Blank         | 0.000     | 01:34 |                      |
| Subject Sample    | # 2       | 01:36 | Subject Test Refused |
| Breath Volume     | 0.039     | 01:36 |                      |
| Air Blank         | 0.000     | 01:37 |                      |
| Control Test      | 0.075     | 01:37 |                      |
| Air Blank         | 0.000     | 01:38 |                      |
| Diagnostics Check | OK        | 01:38 |                      |

---

Data Download Date/Time: 05/03/2021 16:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,BRYAN,M  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: RAINEY,CRISTA  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: STREIB

DATE/TIME OF TEST: 04/29/2021 20:12:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:05 |
| Air Blank          | 0.000 | 20:05 |
| Control Test       | 0.081 | 20:06 |
| Air Blank          | 0.000 | 20:06 |
| Subject Sample # 1 | 0.164 | 20:08 |
| Breath Volume      | 1.367 | 20:08 |
| Air Blank          | 0.000 | 20:08 |
| Air Blank          | 0.000 | 20:10 |
| Subject Sample # 2 | 0.152 | 20:11 |
| Breath Volume      | 1.804 | 20:11 |
| Air Blank          | 0.000 | 20:11 |
| Control Test       | 0.080 | 20:12 |
| Air Blank          | 0.000 | 20:12 |
| Diagnostics Check  | OK    | 20:12 |

---

Data Download Date/Time: 05/03/2021 16:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRIEMER,AMANDA,M  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: MORGANTE,T  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 04/29/2021 21:28:21

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 21:18 |                |
| Air Blank         | 0.000     | 21:18 |                |
| Control Test      | 0.080     | 21:19 |                |
| Air Blank         | 0.000     | 21:19 |                |
| Subject Sample    | # 1 0.268 | 21:22 | Volume Not Met |
| Breath Volume     | 0.609     | 21:22 |                |
| Air Blank         | 0.000     | 21:23 |                |
| Air Blank         | 0.000     | 21:25 |                |
| Subject Sample    | # 2 0.307 | 21:26 |                |
| Breath Volume     | 2.003     | 21:26 |                |
| Air Blank         |           | 21:27 | Purge Fail     |
| Air Blank         |           | 21:28 | Purge Fail     |

---

Data Download Date/Time: 05/03/2021 16:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,BRYAN,M  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: MORGANTE,TALOR  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 04/30/2021 00:34:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:25 |
| Air Blank          | 0.000 | 00:25 |
| Control Test       | 0.079 | 00:26 |
| Air Blank          | 0.000 | 00:26 |
| Subject Sample # 1 | 0.273 | 00:27 |
| Breath Volume      | 1.843 | 00:27 |
| Air Blank          | 0.000 | 00:28 |
| Air Blank          | 0.000 | 00:30 |
| Subject Sample # 2 | 0.260 | 00:32 |
| Breath Volume      | 1.890 | 00:32 |
| Air Blank          | 0.000 | 00:33 |
| Control Test       | 0.079 | 00:33 |
| Air Blank          | 0.000 | 00:33 |
| Diagnostics Check  | OK    | 00:34 |

---

Data Download Date/Time: 05/03/2021 16:53



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,BRYAN,M  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: CLAY,AARON  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: KESSLER

DATE/TIME OF TEST: 04/30/2021 05:32:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:25 |
| Air Blank          | 0.000 | 05:26 |
| Control Test       | 0.080 | 05:26 |
| Air Blank          | 0.000 | 05:27 |
| Subject Sample # 1 | 0.104 | 05:27 |
| Breath Volume      | 2.937 | 05:27 |
| Air Blank          | 0.000 | 05:28 |
| Air Blank          | 0.000 | 05:30 |
| Subject Sample # 2 | 0.103 | 05:31 |
| Breath Volume      | 2.121 | 05:31 |
| Air Blank          | 0.000 | 05:31 |
| Control Test       | 0.080 | 05:32 |
| Air Blank          | 0.000 | 05:32 |
| Diagnostics Check  | OK    | 05:32 |

---

Data Download Date/Time: 05/03/2021 16:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VAN CAMP,JEFFERY,A  
BREATH TEST OPERATOR AGENCY: Boca Raton PD  
ARRESTING OFFICER: HENDERSON,MARVIN  
ARRESTING OFFICER AGENCY: Boca Raton PD  
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A3  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: MONTEIRO

DATE/TIME OF TEST: 04/02/2021 18:10:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:03 |
| Air Blank          | 0.000 | 18:03 |
| Control Test       | 0.078 | 18:04 |
| Air Blank          | 0.000 | 18:04 |
| Subject Sample # 1 | 0.141 | 18:05 |
| Breath Volume      | 1.687 | 18:05 |
| Air Blank          | 0.000 | 18:06 |
| Air Blank          | 0.000 | 18:07 |
| Subject Sample # 2 | 0.140 | 18:08 |
| Breath Volume      | 1.660 | 18:08 |
| Air Blank          | 0.000 | 18:09 |
| Control Test       | 0.079 | 18:09 |
| Air Blank          | 0.000 | 18:10 |
| Diagnostics Check  | OK    | 18:10 |

---

Data Download Date/Time: 04/28/2021 09:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTILLO,CARLOS,M  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: DOMINGUEZ,JOSE  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: PRISINZANO

DATE/TIME OF TEST: 04/01/2021 01:34:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:27 |
| Air Blank          | 0.000 | 01:28 |
| Control Test       | 0.081 | 01:28 |
| Air Blank          | 0.000 | 01:28 |
| Subject Sample # 1 | 0.000 | 01:30 |
| Breath Volume      | 2.652 | 01:30 |
| Air Blank          | 0.000 | 01:31 |
| Air Blank          | 0.000 | 01:33 |
| Subject Sample # 2 | 0.000 | 01:33 |
| Breath Volume      | 2.656 | 01:33 |
| Air Blank          | 0.000 | 01:33 |
| Control Test       | 0.079 | 01:34 |
| Air Blank          | 0.000 | 01:34 |
| Diagnostics Check  | OK    | 01:34 |

---

Data Download Date/Time: 04/05/2021 00:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALBO,JOSHUA,D  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: PEREZ,ADRIAN  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: ELIAS

DATE/TIME OF TEST: 04/03/2021 19:01:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:55 |
| Air Blank          | 0.000 | 18:55 |
| Control Test       | 0.082 | 18:55 |
| Air Blank          | 0.000 | 18:56 |
| Subject Sample # 1 | 0.193 | 18:57 |
| Breath Volume      | 1.386 | 18:57 |
| Air Blank          | 0.000 | 18:57 |
| Air Blank          | 0.000 | 18:59 |
| Subject Sample # 2 | 0.174 | 18:59 |
| Breath Volume      | 3.382 | 18:59 |
| Air Blank          | 0.000 | 19:00 |
| Control Test       | 0.080 | 19:00 |
| Air Blank          | 0.000 | 19:01 |
| Diagnostics Check  | OK    | 19:01 |

---

Data Download Date/Time: 04/05/2021 00:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ARIAS,ELOY,  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: SNOW,REX  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 04/04/2021 08:44:13

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 08:32 |                    |
| Air Blank         | 0.000 | 08:33 |                    |
| Control Test      | 0.082 | 08:33 |                    |
| Air Blank         | 0.000 | 08:34 |                    |
| Subject Sample    | # 1   | 08:37 | No Sample Provided |
| Breath Volume     | 0.000 | 08:37 |                    |
| Air Blank         | 0.000 | 08:37 |                    |
| Air Blank         | 0.000 | 08:39 |                    |
| Subject Sample    | # 2   | 08:42 | No Sample Provided |
| Breath Volume     | 0.000 | 08:42 |                    |
| Air Blank         | 0.000 | 08:43 |                    |
| Control Test      | 0.082 | 08:43 |                    |
| Air Blank         | 0.000 | 08:44 |                    |
| Diagnostics Check | OK    | 08:44 |                    |

---

Data Download Date/Time: 04/05/2021 00:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESTEVEZ,JOSE,M  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: ESTEVEZ,JOSE  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: DURANCASTELLANOS

DATE/TIME OF TEST: 04/14/2021 05:58:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:51 |
| Air Blank          | 0.000 | 05:52 |
| Control Test       | 0.082 | 05:52 |
| Air Blank          | 0.000 | 05:53 |
| Subject Sample # 1 | 0.179 | 05:53 |
| Breath Volume      | 2.863 | 05:53 |
| Air Blank          | 0.000 | 05:54 |
| Air Blank          | 0.000 | 05:56 |
| Subject Sample # 2 | 0.171 | 05:56 |
| Breath Volume      | 1.468 | 05:56 |
| Air Blank          | 0.000 | 05:57 |
| Control Test       | 0.080 | 05:57 |
| Air Blank          | 0.000 | 05:58 |
| Diagnostics Check  | OK    | 05:58 |

---

Data Download Date/Time: 06/02/2021 21:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LLANIO,PIERRE,  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: LLANIO,PIERRE  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 04/17/2021 00:32:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:25 |
| Air Blank          | 0.000 | 00:26 |
| Control Test       | 0.081 | 00:26 |
| Air Blank          | 0.000 | 00:27 |
| Subject Sample # 1 | 0.214 | 00:27 |
| Breath Volume      | 1.519 | 00:27 |
| Air Blank          | 0.000 | 00:28 |
| Air Blank          | 0.000 | 00:30 |
| Subject Sample # 2 | 0.200 | 00:31 |
| Breath Volume      | 1.582 | 00:31 |
| Air Blank          | 0.000 | 00:31 |
| Control Test       | 0.081 | 00:32 |
| Air Blank          | 0.000 | 00:32 |
| Diagnostics Check  | OK    | 00:32 |

---

Data Download Date/Time: 06/02/2021 21:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELENDEZ-BONILLA, RONALD,  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: MELENDEZ-BONILLA, RONALD  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 04/24/2021 18:03:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:57 |
| Air Blank          | 0.000 | 17:57 |
| Control Test       | 0.082 | 17:58 |
| Air Blank          | 0.000 | 17:58 |
| Subject Sample # 1 | 0.188 | 17:59 |
| Breath Volume      | 2.441 | 17:59 |
| Air Blank          | 0.000 | 17:59 |
| Air Blank          | 0.000 | 18:01 |
| Subject Sample # 2 | 0.189 | 18:01 |
| Breath Volume      | 2.433 | 18:01 |
| Air Blank          | 0.000 | 18:02 |
| Control Test       | 0.081 | 18:02 |
| Air Blank          | 0.000 | 18:03 |
| Diagnostics Check  | OK    | 18:03 |

---

Data Download Date/Time: 06/02/2021 21:54



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CANTELAR,JOAN,  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: CANTELAR,JOAN  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: RUIZ VASQUEZ

DATE/TIME OF TEST: 04/25/2021 04:54:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:46 |
| Air Blank          | 0.000 | 04:47 |
| Control Test       | 0.082 | 04:47 |
| Air Blank          | 0.000 | 04:47 |
| Subject Sample # 1 | 0.203 | 04:48 |
| Breath Volume      | 2.621 | 04:48 |
| Air Blank          | 0.000 | 04:49 |
| Air Blank          | 0.000 | 04:51 |
| Subject Sample # 2 | 0.193 | 04:52 |
| Breath Volume      | 2.535 | 04:52 |
| Air Blank          | 0.000 | 04:53 |
| Control Test       | 0.079 | 04:53 |
| Air Blank          | 0.000 | 04:53 |
| Diagnostics Check  | OK    | 04:54 |

---

Data Download Date/Time: 06/02/2021 21:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESTEVEZ,BRYAN,  
BREATH TEST OPERATOR AGENCY: FHP TROOP K  
ARRESTING OFFICER: JOSEPH,JEAN  
ARRESTING OFFICER AGENCY: FHP TROOP E  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: BERMUDEZ GARCIA

DATE/TIME OF TEST: 04/25/2021 08:25:28

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 08:22 |                      |
| Air Blank         | 0.000 | 08:22 |                      |
| Control Test      | 0.082 | 08:23 |                      |
| Air Blank         | 0.000 | 08:23 |                      |
| Subject Sample    | # 1   | 08:24 | Subject Test Refused |
| Breath Volume     | 0.000 | 08:24 |                      |
| Air Blank         | 0.000 | 08:24 |                      |
| Control Test      | 0.082 | 08:24 |                      |
| Air Blank         | 0.000 | 08:25 |                      |
| Diagnostics Check | OK    | 08:25 |                      |

---

Data Download Date/Time: 06/02/2021 21:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ JR,RAUL,  
BREATH TEST OPERATOR AGENCY: FHP TROOP E  
ARRESTING OFFICER: RODRIGUEZ,GIOVANY  
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1114319  
DRY GAS EXPIRY DATE: 03/27/2022  
VIOLATION CODE: DUI  
ARREST TIME: 14:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: SCHAW

DATE/TIME OF TEST: 04/25/2021 17:11:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:02 |
| Air Blank          | 0.000 | 17:02 |
| Control Test       | 0.081 | 17:02 |
| Air Blank          | 0.000 | 17:03 |
| Subject Sample # 1 | 0.262 | 17:04 |
| Breath Volume      | 2.371 | 17:04 |
| Air Blank          | 0.000 | 17:05 |
| Air Blank          | 0.000 | 17:06 |
| Subject Sample # 2 | 0.238 | 17:07 |
| Breath Volume      | 1.492 | 17:07 |
| Air Blank          | 0.000 | 17:07 |
| Air Blank          | 0.000 | 17:09 |
| Subject Sample # 3 | 0.233 | 17:10 |
| Breath Volume      | 1.261 | 17:10 |
| Air Blank          | 0.000 | 17:10 |
| Control Test       | 0.080 | 17:11 |
| Air Blank          | 0.000 | 17:11 |
| Diagnostics Check  | OK    | 17:11 |

---

Data Download Date/Time: 06/02/2021 21:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAWKINS,DEBORAH,  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HAWKINS,DEBORAH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 03:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: SHEFFIELD

DATE/TIME OF TEST: 04/04/2021 04:12:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:05 |
| Air Blank          | 0.000 | 04:05 |
| Control Test       | 0.079 | 04:05 |
| Air Blank          | 0.000 | 04:06 |
| Subject Sample # 1 | 0.198 | 04:07 |
| Breath Volume      | 4.246 | 04:07 |
| Air Blank          | 0.000 | 04:08 |
| Air Blank          | 0.000 | 04:10 |
| Subject Sample # 2 | 0.198 | 04:11 |
| Breath Volume      | 3.140 | 04:11 |
| Air Blank          | 0.000 | 04:12 |
| Control Test       | 0.077 | 04:12 |
| Air Blank          | 0.000 | 04:12 |
| Diagnostics Check  | OK    | 04:12 |

---

Data Download Date/Time: 04/29/2021 12:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRERO,FRAICOR,  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: TERRERO,FRAICOR  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: FAIRLEY

DATE/TIME OF TEST: 04/10/2021 03:22:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:16 |
| Air Blank          | 0.000 | 03:16 |
| Control Test       | 0.080 | 03:16 |
| Air Blank          | 0.000 | 03:17 |
| Subject Sample # 1 | 0.197 | 03:17 |
| Breath Volume      | 2.507 | 03:17 |
| Air Blank          | 0.000 | 03:18 |
| Air Blank          | 0.000 | 03:20 |
| Subject Sample # 2 | 0.197 | 03:20 |
| Breath Volume      | 2.132 | 03:20 |
| Air Blank          | 0.000 | 03:21 |
| Control Test       | 0.079 | 03:21 |
| Air Blank          | 0.000 | 03:22 |
| Diagnostics Check  | OK    | 03:22 |

---

Data Download Date/Time: 05/17/2021 11:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRERO,FRAICOR,  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: TERRERO,FRAICOR  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: BUI  
ARREST TIME: 18:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: PICKWOAD

DATE/TIME OF TEST: 04/16/2021 19:08:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:02 |
| Air Blank          | 0.000 | 19:03 |
| Control Test       | 0.078 | 19:03 |
| Air Blank          | 0.000 | 19:03 |
| Subject Sample # 1 | 0.146 | 19:04 |
| Breath Volume      | 2.308 | 19:04 |
| Air Blank          | 0.000 | 19:05 |
| Air Blank          | 0.000 | 19:06 |
| Subject Sample # 2 | 0.142 | 19:07 |
| Breath Volume      | 2.156 | 19:07 |
| Air Blank          | 0.000 | 19:07 |
| Control Test       | 0.078 | 19:08 |
| Air Blank          | 0.000 | 19:08 |
| Diagnostics Check  | OK    | 19:08 |

---

Data Download Date/Time: 05/17/2021 11:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRERO,FRAICOR,  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: TERRERO,FRAICOR  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: DIAZ PEREZ

DATE/TIME OF TEST: 04/17/2021 21:12:21

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 20:57 |                |
| Air Blank         | 0.000     | 20:58 |                |
| Control Test      | 0.078     | 20:58 |                |
| Air Blank         | 0.000     | 20:59 |                |
| Subject Sample    | # 1 0.301 | 21:02 | Volume Not Met |
| Breath Volume     | 1.070     | 21:02 |                |
| Air Blank         | 0.000     | 21:03 |                |
| Air Blank         | 0.000     | 21:04 |                |
| Subject Sample    | # 2 0.294 | 21:05 |                |
| Breath Volume     | 1.179     | 21:05 |                |
| Air Blank         | 0.000     | 21:06 |                |
| Air Blank         | 0.000     | 21:08 |                |
| Subject Sample    | # 3 0.295 | 21:11 | Volume Not Met |
| Breath Volume     | 0.687     | 21:11 |                |
| Air Blank         |           | 21:11 | RFI Detect     |
| Air Blank         | 0.000     | 21:12 |                |

---

Data Download Date/Time: 05/17/2021 11:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRERO,FRAICOR,  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BARRETT,CODY  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: GRUSHCO

DATE/TIME OF TEST: 04/18/2021 20:26:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:20 |
| Air Blank          | 0.000 | 20:21 |
| Control Test       | 0.078 | 20:21 |
| Air Blank          | 0.000 | 20:22 |
| Subject Sample # 1 | 0.095 | 20:22 |
| Breath Volume      | 2.347 | 20:22 |
| Air Blank          | 0.000 | 20:23 |
| Air Blank          | 0.000 | 20:24 |
| Subject Sample # 2 | 0.092 | 20:25 |
| Breath Volume      | 2.023 | 20:25 |
| Air Blank          | 0.000 | 20:25 |
| Control Test       | 0.079 | 20:26 |
| Air Blank          | 0.000 | 20:26 |
| Diagnostics Check  | OK    | 20:26 |

---

Data Download Date/Time: 05/17/2021 11:21



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRERO,FRAICOR,  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: TERRERO,FRAICOR  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 17:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:34

SUBJECT'S LAST NAME: GARNER

DATE/TIME OF TEST: 04/30/2021 18:01:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:54 |
| Air Blank          | 0.000 | 17:55 |
| Control Test       | 0.080 | 17:55 |
| Air Blank          | 0.000 | 17:56 |
| Subject Sample # 1 | 0.000 | 17:56 |
| Breath Volume      | 4.164 | 17:56 |
| Air Blank          | 0.000 | 17:57 |
| Air Blank          | 0.000 | 17:59 |
| Subject Sample # 2 | 0.000 | 17:59 |
| Breath Volume      | 3.339 | 17:59 |
| Air Blank          | 0.000 | 18:00 |
| Control Test       | 0.080 | 18:00 |
| Air Blank          | 0.000 | 18:01 |
| Diagnostics Check  | OK    | 18:01 |

---

Data Download Date/Time: 05/17/2021 11:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,WILLIAM,T  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SMITH,WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 972301  
DRY GAS EXPIRY DATE: 06/15/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: CHILDS

DATE/TIME OF TEST: 04/04/2021 22:13:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:05 |
| Air Blank          | 0.000 | 22:06 |
| Control Test       | 0.079 | 22:06 |
| Air Blank          | 0.000 | 22:07 |
| Subject Sample # 1 | 0.174 | 22:09 |
| Breath Volume      | 1.671 | 22:09 |
| Air Blank          | 0.000 | 22:09 |
| Air Blank          | 0.000 | 22:11 |
| Subject Sample # 2 | 0.174 | 22:12 |
| Breath Volume      | 1.136 | 22:12 |
| Air Blank          | 0.000 | 22:12 |
| Control Test       | 0.078 | 22:13 |
| Air Blank          | 0.000 | 22:13 |
| Diagnostics Check  | OK    | 22:13 |

---

Data Download Date/Time: 05/28/2021 16:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,WILLIAM,T  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SMITH,WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 972301  
DRY GAS EXPIRY DATE: 06/15/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: PITTAWAY

DATE/TIME OF TEST: 04/18/2021 19:56:26

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:53 |                      |
| Air Blank         | 0.000 | 19:53 |                      |
| Control Test      | 0.078 | 19:54 |                      |
| Air Blank         | 0.000 | 19:54 |                      |
| Subject Sample    | # 1   | 19:55 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:55 |                      |
| Air Blank         | 0.000 | 19:55 |                      |
| Control Test      | 0.078 | 19:55 |                      |
| Air Blank         | 0.000 | 19:56 |                      |
| Diagnostics Check | OK    | 19:56 |                      |

---

Data Download Date/Time: 05/28/2021 16:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,WILLIAM,T  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: SMITH,WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 972301  
DRY GAS EXPIRY DATE: 06/15/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 04/21/2021 22:35:07

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:32 |                      |
| Air Blank         | 0.000 | 22:32 |                      |
| Control Test      | 0.079 | 22:32 |                      |
| Air Blank         | 0.000 | 22:33 |                      |
| Subject Sample    | # 1   | 22:33 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:33 |                      |
| Air Blank         | 0.000 | 22:34 |                      |
| Control Test      | 0.079 | 22:34 |                      |
| Air Blank         | 0.000 | 22:35 |                      |
| Diagnostics Check | OK    | 22:35 |                      |

---

Data Download Date/Time: 05/28/2021 16:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/01/2021 22:14:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:08 |
| Air Blank          | 0.000 | 22:08 |
| Control Test       | 0.083 | 22:08 |
| Air Blank          | 0.000 | 22:09 |
| Subject Sample # 1 | 0.000 | 22:09 |
| Breath Volume      | 1.550 | 22:09 |
| Air Blank          | 0.000 | 22:10 |
| Air Blank          | 0.000 | 22:12 |
| Subject Sample # 2 | 0.000 | 22:12 |
| Breath Volume      | 3.226 | 22:12 |
| Air Blank          | 0.000 | 22:13 |
| Control Test       | 0.083 | 22:13 |
| Air Blank          | 0.000 | 22:14 |
| Diagnostics Check  | OK    | 22:14 |

---

Data Download Date/Time: 04/20/2021 09:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: GULLETT

DATE/TIME OF TEST: 04/03/2021 01:09:39

---

|                    |       |       |            |
|--------------------|-------|-------|------------|
| Diagnostics Check  | OK    | 01:06 |            |
| Air Blank          | 0.000 | 01:06 |            |
| Control Test       | 0.083 | 01:06 |            |
| Air Blank          | 0.000 | 01:07 |            |
| Subject Sample # 1 | 0.114 | 01:07 |            |
| Breath Volume      | 5.253 | 01:07 |            |
| Air Blank          |       | 01:08 | Purge Fail |
| Air Blank          | 0.000 | 01:09 |            |

---

Data Download Date/Time: 04/20/2021 09:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: GULLETT

DATE/TIME OF TEST: 04/03/2021 01:17:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:10 |
| Air Blank          | 0.000 | 01:11 |
| Control Test       | 0.081 | 01:11 |
| Air Blank          | 0.000 | 01:12 |
| Subject Sample # 1 | 0.110 | 01:12 |
| Breath Volume      | 5.367 | 01:12 |
| Air Blank          | 0.000 | 01:13 |
| Air Blank          | 0.000 | 01:15 |
| Subject Sample # 2 | 0.114 | 01:15 |
| Breath Volume      | 5.347 | 01:15 |
| Air Blank          | 0.000 | 01:16 |
| Control Test       | 0.079 | 01:16 |
| Air Blank          | 0.000 | 01:17 |
| Diagnostics Check  | OK    | 01:17 |

---

Data Download Date/Time: 04/20/2021 09:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER III,RICHARD,F  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: SNYDER III,RICHARD  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/20/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: GONZALEZ OTERO

DATE/TIME OF TEST: 04/26/2021 01:51:41

---

|                    |       |       |            |
|--------------------|-------|-------|------------|
| Diagnostics Check  | OK    | 01:45 |            |
| Air Blank          | 0.000 | 01:46 |            |
| Control Test       | 0.081 | 01:46 |            |
| Air Blank          | 0.000 | 01:46 |            |
| Subject Sample # 1 | 0.187 | 01:47 |            |
| Breath Volume      | 2.011 | 01:47 |            |
| Air Blank          | 0.000 | 01:48 |            |
| Air Blank          | 0.000 | 01:49 |            |
| Subject Sample # 2 | 0.182 | 01:50 |            |
| Breath Volume      | 1.363 | 01:50 |            |
| Air Blank          |       | 01:51 | Purge Fail |
| Air Blank          | 0.000 | 01:51 |            |

---

Data Download Date/Time: 05/17/2021 11:10



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER III,RICHARD,F  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: SNYDER III,RICHARD  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/20/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: GONZALEZ OTERO

DATE/TIME OF TEST: 04/26/2021 02:01:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:54 |
| Air Blank          | 0.000 | 01:54 |
| Control Test       | 0.078 | 01:55 |
| Air Blank          | 0.000 | 01:55 |
| Subject Sample # 1 | 0.177 | 01:56 |
| Breath Volume      | 1.488 | 01:56 |
| Air Blank          | 0.000 | 01:57 |
| Air Blank          | 0.000 | 01:59 |
| Subject Sample # 2 | 0.176 | 02:00 |
| Breath Volume      | 1.667 | 02:00 |
| Air Blank          | 0.000 | 02:00 |
| Control Test       | 0.078 | 02:01 |
| Air Blank          | 0.000 | 02:01 |
| Diagnostics Check  | OK    | 02:01 |

---

Data Download Date/Time: 05/17/2021 11:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/20/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: RAS

DATE/TIME OF TEST: 04/30/2021 00:18:16

---

|                    |       |       |            |
|--------------------|-------|-------|------------|
| Diagnostics Check  | OK    | 00:11 |            |
| Air Blank          | 0.000 | 00:11 |            |
| Control Test       | 0.078 | 00:12 |            |
| Air Blank          | 0.000 | 00:12 |            |
| Subject Sample # 1 | 0.264 | 00:13 |            |
| Breath Volume      | 1.113 | 00:13 |            |
| Air Blank          | 0.000 | 00:14 |            |
| Air Blank          | 0.000 | 00:15 |            |
| Subject Sample # 2 | 0.306 | 00:16 |            |
| Breath Volume      | 1.296 | 00:16 |            |
| Air Blank          |       | 00:17 | Purge Fail |
| Air Blank          | 0.000 | 00:18 |            |

---

Data Download Date/Time: 05/17/2021 11:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/20/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1303107  
DRY GAS EXPIRY DATE: 03/28/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: R200181780510

DATE/TIME OF TEST: 04/30/2021 00:26:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:19 |
| Air Blank          | 0.000 | 00:19 |
| Control Test       | 0.075 | 00:20 |
| Air Blank          | 0.000 | 00:20 |
| Subject Sample # 1 | 0.290 | 00:21 |
| Breath Volume      | 1.132 | 00:21 |
| Air Blank          | 0.000 | 00:22 |
| Air Blank          | 0.000 | 00:23 |
| Subject Sample # 2 | 0.272 | 00:24 |
| Breath Volume      | 1.453 | 00:24 |
| Air Blank          | 0.000 | 00:25 |
| Control Test       | 0.079 | 00:26 |
| Air Blank          | 0.000 | 00:26 |
| Diagnostics Check  | OK    | 00:26 |

---

Data Download Date/Time: 05/17/2021 11:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HITCHCOCK, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/25/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: ERHART

DATE/TIME OF TEST: 04/03/2021 03:00:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:54 |
| Air Blank          | 0.000 | 02:55 |
| Control Test       | 0.081 | 02:55 |
| Air Blank          | 0.000 | 02:56 |
| Subject Sample # 1 | 0.145 | 02:56 |
| Breath Volume      | 1.500 | 02:56 |
| Air Blank          | 0.000 | 02:57 |
| Air Blank          | 0.000 | 02:58 |
| Subject Sample # 2 | 0.158 | 02:59 |
| Breath Volume      | 2.550 | 02:59 |
| Air Blank          | 0.000 | 03:00 |
| Control Test       | 0.080 | 03:00 |
| Air Blank          | 0.000 | 03:00 |
| Diagnostics Check  | OK    | 03:00 |

---

Data Download Date/Time: 04/30/2021 21:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HITCHCOCK, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/25/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: KVEREK

DATE/TIME OF TEST: 04/04/2021 00:32:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:25 |
| Air Blank          | 0.000 | 00:26 |
| Control Test       | 0.080 | 00:26 |
| Air Blank          | 0.000 | 00:27 |
| Subject Sample # 1 | 0.141 | 00:27 |
| Breath Volume      | 1.718 | 00:27 |
| Air Blank          | 0.000 | 00:28 |
| Air Blank          | 0.000 | 00:30 |
| Subject Sample # 2 | 0.134 | 00:30 |
| Breath Volume      | 1.675 | 00:30 |
| Air Blank          | 0.000 | 00:31 |
| Control Test       | 0.080 | 00:31 |
| Air Blank          | 0.000 | 00:32 |
| Diagnostics Check  | OK    | 00:32 |

---

Data Download Date/Time: 04/30/2021 21:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HITCHCOCK, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/25/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ISAAC

DATE/TIME OF TEST: 04/09/2021 02:34:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:28 |
| Air Blank          | 0.000 | 02:28 |
| Control Test       | 0.080 | 02:29 |
| Air Blank          | 0.000 | 02:29 |
| Subject Sample # 1 | 0.129 | 02:29 |
| Breath Volume      | 2.820 | 02:29 |
| Air Blank          | 0.000 | 02:30 |
| Air Blank          | 0.000 | 02:32 |
| Subject Sample # 2 | 0.124 | 02:32 |
| Breath Volume      | 2.535 | 02:32 |
| Air Blank          | 0.000 | 02:33 |
| Control Test       | 0.077 | 02:33 |
| Air Blank          | 0.000 | 02:34 |
| Diagnostics Check  | OK    | 02:34 |

---

Data Download Date/Time: 04/30/2021 21:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HITCHCOCK, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/25/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 04/21/2021 02:23:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:17 |
| Air Blank          | 0.000 | 02:17 |
| Control Test       | 0.076 | 02:17 |
| Air Blank          | 0.000 | 02:18 |
| Subject Sample # 1 | 0.060 | 02:18 |
| Breath Volume      | 3.308 | 02:18 |
| Air Blank          | 0.000 | 02:19 |
| Air Blank          | 0.000 | 02:21 |
| Subject Sample # 2 | 0.059 | 02:21 |
| Breath Volume      | 2.664 | 02:21 |
| Air Blank          | 0.000 | 02:22 |
| Control Test       | 0.080 | 02:22 |
| Air Blank          | 0.000 | 02:23 |
| Diagnostics Check  | OK    | 02:23 |

---

Data Download Date/Time: 04/30/2021 21:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: HITCHCOCK, WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/25/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: VALENCIA

DATE/TIME OF TEST: 04/26/2021 00:50:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:43 |
| Air Blank          | 0.000 | 00:44 |
| Control Test       | 0.080 | 00:44 |
| Air Blank          | 0.000 | 00:45 |
| Subject Sample # 1 | 0.183 | 00:45 |
| Breath Volume      | 1.281 | 00:45 |
| Air Blank          | 0.000 | 00:46 |
| Air Blank          | 0.000 | 00:48 |
| Subject Sample # 2 | 0.202 | 00:48 |
| Breath Volume      | 3.273 | 00:48 |
| Air Blank          | 0.000 | 00:49 |
| Control Test       | 0.080 | 00:49 |
| Air Blank          | 0.000 | 00:50 |
| Diagnostics Check  | OK    | 00:50 |

---

Data Download Date/Time: 04/30/2021 21:07



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HURT,ALEXANDER,C  
BREATH TEST OPERATOR AGENCY: Holmes Beach PD  
ARRESTING OFFICER: KELLS,BRANDON  
ARRESTING OFFICER AGENCY: Bradenton PD  
LAST AGENCY INSPECTION: 04/18/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1175970  
DRY GAS EXPIRY DATE: 08/02/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: CAPEN

DATE/TIME OF TEST: 04/19/2021 04:14:59

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:12 |                      |
| Air Blank         | 0.000 | 04:12 |                      |
| Control Test      | 0.076 | 04:12 |                      |
| Air Blank         | 0.000 | 04:13 |                      |
| Subject Sample    | # 1   | 04:13 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:13 |                      |
| Air Blank         | 0.000 | 04:14 |                      |
| Control Test      | 0.077 | 04:14 |                      |
| Air Blank         | 0.000 | 04:14 |                      |
| Diagnostics Check | OK    | 04:14 |                      |

---

Data Download Date/Time: 05/24/2021 16:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: CHLEBANOWSKI,RYAN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: MCCAFFREY

DATE/TIME OF TEST: 04/03/2021 00:05:59

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:03 |                      |
| Air Blank         | 0.000 | 00:03 |                      |
| Control Test      | 0.081 | 00:03 |                      |
| Air Blank         | 0.000 | 00:04 |                      |
| Subject Sample    | # 1   | 00:04 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:04 |                      |
| Air Blank         | 0.000 | 00:05 |                      |
| Control Test      | 0.080 | 00:05 |                      |
| Air Blank         | 0.000 | 00:05 |                      |
| Diagnostics Check | OK    | 00:05 |                      |

---

Data Download Date/Time: 04/07/2021 16:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: WAITE,KENNETH  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 04/04/2021 02:14:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:07 |
| Air Blank          | 0.000 | 02:08 |
| Control Test       | 0.081 | 02:08 |
| Air Blank          | 0.000 | 02:09 |
| Subject Sample # 1 | 0.262 | 02:09 |
| Breath Volume      | 2.105 | 02:09 |
| Air Blank          | 0.000 | 02:10 |
| Air Blank          | 0.000 | 02:12 |
| Subject Sample # 2 | 0.264 | 02:12 |
| Breath Volume      | 2.261 | 02:12 |
| Air Blank          | 0.000 | 02:13 |
| Control Test       | 0.080 | 02:13 |
| Air Blank          | 0.000 | 02:14 |
| Diagnostics Check  | OK    | 02:14 |

---

Data Download Date/Time: 04/07/2021 16:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FIERRO,ALEXANDER,D  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: THORNBRUGH,ELAM  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: ALBINES

DATE/TIME OF TEST: 04/04/2021 05:10:05

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 04:56 |                |
| Air Blank         | 0.000     | 04:56 |                |
| Control Test      | 0.081     | 04:57 |                |
| Air Blank         | 0.000     | 04:57 |                |
| Subject Sample    | # 1 0.078 | 05:00 | Volume Not Met |
| Breath Volume     | 0.015     | 05:00 |                |
| Air Blank         | 0.000     | 05:01 |                |
| Air Blank         | 0.000     | 05:03 |                |
| Subject Sample    | # 2 0.070 | 05:04 |                |
| Breath Volume     | 1.812     | 05:04 |                |
| Air Blank         | 0.000     | 05:05 |                |
| Air Blank         | 0.000     | 05:07 |                |
| Subject Sample    | # 3 0.075 | 05:08 |                |
| Breath Volume     | 1.671     | 05:08 |                |
| Air Blank         | 0.000     | 05:09 |                |
| Control Test      | 0.080     | 05:09 |                |
| Air Blank         | 0.000     | 05:10 |                |
| Diagnostics Check | OK        | 05:10 |                |

---

Data Download Date/Time: 04/07/2021 16:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WATKINS,MICHAEL,W  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: MALAK,T  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: VALERO-SIRVENT

DATE/TIME OF TEST: 04/05/2021 03:07:59

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:05 |                      |
| Air Blank         | 0.000 | 03:05 |                      |
| Control Test      | 0.081 | 03:05 |                      |
| Air Blank         | 0.000 | 03:06 |                      |
| Subject Sample    | # 1   | 03:06 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:06 |                      |
| Air Blank         | 0.000 | 03:07 |                      |
| Control Test      | 0.081 | 03:07 |                      |
| Air Blank         | 0.000 | 03:07 |                      |
| Diagnostics Check | OK    | 03:07 |                      |

---

Data Download Date/Time: 04/07/2021 16:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANE,JONATHAN,M  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: SCHWARTZ,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: CLINE

DATE/TIME OF TEST: 04/06/2021 01:16:46

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:13 |                      |
| Air Blank         | 0.000 | 01:14 |                      |
| Control Test      | 0.080 | 01:14 |                      |
| Air Blank         | 0.000 | 01:15 |                      |
| Subject Sample    | # 1   | 01:15 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:15 |                      |
| Air Blank         | 0.000 | 01:15 |                      |
| Control Test      | 0.079 | 01:16 |                      |
| Air Blank         | 0.000 | 01:16 |                      |
| Diagnostics Check | OK    | 01:16 |                      |

---

Data Download Date/Time: 04/07/2021 16:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLEN,JOHN,M  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: MALAK,TIMOTHY  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: CARMONA

DATE/TIME OF TEST: 04/07/2021 21:57:56

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:54 |                      |
| Air Blank         | 0.000 | 21:54 |                      |
| Control Test      | 0.080 | 21:55 |                      |
| Air Blank         | 0.000 | 21:55 |                      |
| Subject Sample    | # 1   | 21:56 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:56 |                      |
| Air Blank         | 0.000 | 21:57 |                      |
| Control Test      | 0.079 | 21:57 |                      |
| Air Blank         | 0.000 | 21:57 |                      |
| Diagnostics Check | OK    | 21:57 |                      |

---

Data Download Date/Time: 05/11/2021 12:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: CHLEBANOWSKI,RYAN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: KRASNOVA

DATE/TIME OF TEST: 04/08/2021 02:11:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:01 |
| Air Blank          | 0.000 | 02:02 |
| Control Test       | 0.079 | 02:02 |
| Air Blank          | 0.000 | 02:03 |
| Subject Sample # 1 | 0.241 | 02:05 |
| Breath Volume      | 1.621 | 02:05 |
| Air Blank          | 0.000 | 02:05 |
| Air Blank          | 0.000 | 02:07 |
| Subject Sample # 2 | 0.234 | 02:09 |
| Breath Volume      | 1.601 | 02:09 |
| Air Blank          | 0.000 | 02:10 |
| Control Test       | 0.078 | 02:10 |
| Air Blank          | 0.000 | 02:11 |
| Diagnostics Check  | OK    | 02:11 |

---

Data Download Date/Time: 05/11/2021 12:39



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODDY,AARON,A  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: RODDY,AARON  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 10:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:08

SUBJECT'S LAST NAME: RIOS RAMIREZ

DATE/TIME OF TEST: 04/11/2021 11:47:08

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 11:38 |                      |
| Air Blank         | 0.000 | 11:39 |                      |
| Control Test      | 0.079 | 11:39 |                      |
| Air Blank         | 0.000 | 11:39 |                      |
| Subject Sample    | # 1   | 11:43 | No Sample Provided   |
| Breath Volume     | 0.000 | 11:43 |                      |
| Air Blank         | 0.000 | 11:43 |                      |
| Air Blank         | 0.000 | 11:45 |                      |
| Subject Sample    | # 2   | 11:45 | Subject Test Refused |
| Breath Volume     | 0.000 | 11:45 |                      |
| Air Blank         | 0.000 | 11:46 |                      |
| Control Test      | 0.079 | 11:46 |                      |
| Air Blank         | 0.000 | 11:47 |                      |
| Diagnostics Check | OK    | 11:47 |                      |

Data Download Date/Time: 05/11/2021 12:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUDSON,ANDREW,J  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: HUDSON,ANDREW  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 04/15/2021 22:34:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:27 |
| Air Blank          | 0.000 | 22:28 |
| Control Test       | 0.080 | 22:28 |
| Air Blank          | 0.000 | 22:29 |
| Subject Sample # 1 | 0.222 | 22:29 |
| Breath Volume      | 2.859 | 22:29 |
| Air Blank          | 0.000 | 22:30 |
| Air Blank          | 0.000 | 22:32 |
| Subject Sample # 2 | 0.225 | 22:32 |
| Breath Volume      | 2.773 | 22:32 |
| Air Blank          | 0.000 | 22:33 |
| Control Test       | 0.079 | 22:33 |
| Air Blank          | 0.000 | 22:34 |
| Diagnostics Check  | OK    | 22:34 |

---

Data Download Date/Time: 05/11/2021 12:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAPP,BRIAN,L  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: SAPP,BRIAN  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: BUI  
ARREST TIME: 21:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 04/16/2021 22:35:13

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:30 |                      |
| Air Blank         | 0.000 | 22:31 |                      |
| Control Test      | 0.080 | 22:31 |                      |
| Air Blank         | 0.000 | 22:32 |                      |
| Subject Sample    | # 1   | 22:33 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:33 |                      |
| Air Blank         | 0.000 | 22:34 |                      |
| Control Test      | 0.080 | 22:34 |                      |
| Air Blank         | 0.000 | 22:35 |                      |
| Diagnostics Check | OK    | 22:35 |                      |

---

Data Download Date/Time: 05/11/2021 12:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: CHLEBANOWSKI,RYAN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: LEWANDOWSKI

DATE/TIME OF TEST: 04/17/2021 03:52:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:46 |
| Air Blank          | 0.000 | 03:46 |
| Control Test       | 0.080 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Subject Sample # 1 | 0.223 | 03:48 |
| Breath Volume      | 1.843 | 03:48 |
| Air Blank          | 0.000 | 03:48 |
| Air Blank          | 0.000 | 03:50 |
| Subject Sample # 2 | 0.216 | 03:51 |
| Breath Volume      | 1.921 | 03:51 |
| Air Blank          | 0.000 | 03:51 |
| Control Test       | 0.079 | 03:52 |
| Air Blank          | 0.000 | 03:52 |
| Diagnostics Check  | OK    | 03:52 |

---

Data Download Date/Time: 05/11/2021 12:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLEN,JOHN,M  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: MALAK,TIM  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: BIRD

DATE/TIME OF TEST: 04/17/2021 21:58:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:52 |
| Air Blank          | 0.000 | 21:52 |
| Control Test       | 0.080 | 21:52 |
| Air Blank          | 0.000 | 21:53 |
| Subject Sample # 1 | 0.000 | 21:53 |
| Breath Volume      | 1.929 | 21:53 |
| Air Blank          | 0.000 | 21:54 |
| Air Blank          | 0.000 | 21:56 |
| Subject Sample # 2 | 0.000 | 21:56 |
| Breath Volume      | 1.812 | 21:56 |
| Air Blank          | 0.000 | 21:57 |
| Control Test       | 0.080 | 21:57 |
| Air Blank          | 0.000 | 21:58 |
| Diagnostics Check  | OK    | 21:58 |

---

Data Download Date/Time: 05/11/2021 12:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAPP,BRIAN,L  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: BOLEY,CHRISTOPHER  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: BUI  
ARREST TIME: 21:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: BUSBY

DATE/TIME OF TEST: 04/17/2021 22:37:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:30 |
| Air Blank          | 0.000 | 22:31 |
| Control Test       | 0.079 | 22:31 |
| Air Blank          | 0.000 | 22:32 |
| Subject Sample # 1 | 0.222 | 22:32 |
| Breath Volume      | 2.269 | 22:32 |
| Air Blank          | 0.000 | 22:33 |
| Air Blank          | 0.000 | 22:35 |
| Subject Sample # 2 | 0.204 | 22:35 |
| Breath Volume      | 1.960 | 22:35 |
| Air Blank          | 0.000 | 22:36 |
| Control Test       | 0.076 | 22:36 |
| Air Blank          | 0.000 | 22:37 |
| Diagnostics Check  | OK    | 22:37 |

---

Data Download Date/Time: 05/11/2021 12:40

# **Florida Department of Law Enforcement**

## **Alcohol Testing Program**

### **Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WATKINS,MICHAEL,W  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: ALLEN,JOHN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: SPRINGER

DATE/TIME OF TEST: 04/28/2021 02:44:58

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 02:43 |            |
| Air Blank         | 0.000 | 02:43 |            |
| Control Test      | 0.080 | 02:43 |            |
| Air Blank         | 0.000 | 02:44 |            |
|                   |       | 02:44 | RFI Detect |
| Air Blank         | 0.000 | 02:44 |            |

---

Data Download Date/Time: 05/11/2021 12:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WATKINS,MICHAEL,W  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: ALLEN,JOHN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: SPRINGER

DATE/TIME OF TEST: 04/28/2021 02:50:20

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:47 |                      |
| Air Blank         | 0.000 | 02:47 |                      |
| Control Test      | 0.079 | 02:48 |                      |
| Air Blank         | 0.000 | 02:48 |                      |
| Subject Sample    | # 1   | 02:48 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:48 |                      |
| Air Blank         | 0.000 | 02:49 |                      |
| Control Test      | 0.080 | 02:49 |                      |
| Air Blank         | 0.000 | 02:50 |                      |
| Diagnostics Check | OK    | 02:50 |                      |

---

Data Download Date/Time: 05/11/2021 12:40



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRERO,FRAICOR,  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: TERRERO,FRAICOR  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: O'HARA

DATE/TIME OF TEST: 04/04/2021 23:10:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:03 |
| Air Blank          | 0.000 | 23:03 |
| Control Test       | 0.080 | 23:03 |
| Air Blank          | 0.000 | 23:04 |
| Subject Sample # 1 | 0.169 | 23:04 |
| Breath Volume      | 1.445 | 23:04 |
| Air Blank          | 0.000 | 23:05 |
| Air Blank          | 0.000 | 23:07 |
| Subject Sample # 2 | 0.168 | 23:08 |
| Breath Volume      | 2.617 | 23:08 |
| Air Blank          | 0.000 | 23:09 |
| Control Test       | 0.080 | 23:09 |
| Air Blank          | 0.000 | 23:10 |
| Diagnostics Check  | OK    | 23:10 |

---

Data Download Date/Time: 04/16/2021 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRERO,FRAICOR,  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: TERRERO,FRAICOR  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: MANCILLAS

DATE/TIME OF TEST: 04/05/2021 03:30:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:24 |
| Air Blank          | 0.000 | 03:24 |
| Control Test       | 0.080 | 03:24 |
| Air Blank          | 0.000 | 03:25 |
| Subject Sample # 1 | 0.158 | 03:25 |
| Breath Volume      | 1.941 | 03:25 |
| Air Blank          | 0.000 | 03:26 |
| Air Blank          | 0.000 | 03:28 |
| Subject Sample # 2 | 0.158 | 03:28 |
| Breath Volume      | 1.804 | 03:28 |
| Air Blank          | 0.000 | 03:29 |
| Control Test       | 0.079 | 03:29 |
| Air Blank          | 0.000 | 03:30 |
| Diagnostics Check  | OK    | 03:30 |

---

Data Download Date/Time: 04/16/2021 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1111579  
DRY GAS EXPIRY DATE: 03/19/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: GONZALEZ PINEDA

DATE/TIME OF TEST: 04/03/2021 22:34:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:26 |
| Air Blank          | 0.000 | 22:27 |
| Control Test       | 0.081 | 22:27 |
| Air Blank          | 0.000 | 22:28 |
| Subject Sample # 1 | 0.172 | 22:29 |
| Breath Volume      | 4.015 | 22:29 |
| Air Blank          | 0.000 | 22:29 |
| Air Blank          | 0.000 | 22:31 |
| Subject Sample # 2 | 0.152 | 22:32 |
| Breath Volume      | 1.230 | 22:32 |
| Air Blank          | 0.000 | 22:33 |
| Control Test       | 0.081 | 22:33 |
| Air Blank          | 0.000 | 22:34 |
| Diagnostics Check  | OK    | 22:34 |

---

Data Download Date/Time: 04/19/2021 08:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

|                                          |                                 |
|------------------------------------------|---------------------------------|
| INSTRUMENT REGISTERED TO: FHP            | Date or Time Changed: No        |
| SOFTWARE VERSION: 8100.27                | DRY GAS STANDARD LOT #: 1076482 |
| BREATH TEST OPERATOR: RUBEL,NEIL,F       | DRY GAS EXPIRY DATE: 01/15/2022 |
| BREATH TEST OPERATOR AGENCY: FHP Troop A | VIOLATION CODE: DUI             |
| ARRESTING OFFICER: SISK,CARLOS           | ARREST TIME: 18:34              |
| ARRESTING OFFICER AGENCY: FHP Troop A    | GENDER: Male                    |
| LAST AGENCY INSPECTION: 04/09/2021       | OBSERVATION PERIOD BEGAN: 18:34 |

SUBJECT'S LAST NAME: ROBERTS      DATE/TIME OF TEST: 04/20/2021      18:58:44

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 18:54 |                      |
| Air Blank         | 0.000 | 18:54 |                      |
| Control Test      | 0.079 | 18:55 |                      |
| Air Blank         | 0.000 | 18:55 |                      |
| Subject Sample    | # 1   | 18:57 | Subject Test Refused |
| Breath Volume     | 0.000 | 18:57 |                      |
| Air Blank         | 0.000 | 18:57 |                      |
| Control Test      | 0.080 | 18:58 |                      |
| Air Blank         | 0.000 | 18:58 |                      |
| Diagnostics Check | OK    | 18:58 |                      |

---

Data Download Date/Time: 05/06/2021 12:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY, KENNETH, R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: CAPELA, J  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35019080A3  
DRY GAS EXPIRY DATE: 02/25/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: TEN EYCK

DATE/TIME OF TEST: 04/02/2021 19:00:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:53 |
| Air Blank          | 0.000 | 18:54 |
| Control Test       | 0.079 | 18:54 |
| Air Blank          | 0.000 | 18:55 |
| Subject Sample # 1 | 0.344 | 18:55 |
| Breath Volume      | 1.152 | 18:55 |
| Air Blank          | 0.000 | 18:56 |
| Air Blank          | 0.000 | 18:58 |
| Subject Sample # 2 | 0.352 | 18:59 |
| Breath Volume      | 1.550 | 18:59 |
| Air Blank          | 0.000 | 18:59 |
| Control Test       | 0.080 | 19:00 |
| Air Blank          | 0.000 | 19:00 |
| Diagnostics Check  | OK    | 19:00 |

---

Data Download Date/Time: 04/07/2021 19:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY, KENNETH, R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY, KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35019080A3  
DRY GAS EXPIRY DATE: 02/25/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: FREDERICK

DATE/TIME OF TEST: 04/03/2021 03:27:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:20 |
| Air Blank          | 0.000 | 03:21 |
| Control Test       | 0.079 | 03:21 |
| Air Blank          | 0.000 | 03:22 |
| Subject Sample # 1 | 0.122 | 03:22 |
| Breath Volume      | 1.128 | 03:22 |
| Air Blank          | 0.000 | 03:23 |
| Air Blank          | 0.000 | 03:25 |
| Subject Sample # 2 | 0.120 | 03:25 |
| Breath Volume      | 2.339 | 03:25 |
| Air Blank          | 0.000 | 03:26 |
| Control Test       | 0.079 | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Diagnostics Check  | OK    | 03:27 |

---

Data Download Date/Time: 04/07/2021 19:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY, KENNETH, R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY, KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35019080A3  
DRY GAS EXPIRY DATE: 02/25/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: DOLAN

DATE/TIME OF TEST: 04/04/2021 01:43:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:35 |
| Air Blank          | 0.000 | 01:36 |
| Control Test       | 0.079 | 01:36 |
| Air Blank          | 0.000 | 01:37 |
| Subject Sample # 1 | 0.015 | 01:39 |
| Breath Volume      | 2.378 | 01:39 |
| Air Blank          | 0.000 | 01:40 |
| Air Blank          | 0.000 | 01:41 |
| Subject Sample # 2 | 0.011 | 01:42 |
| Breath Volume      | 1.199 | 01:42 |
| Air Blank          | 0.000 | 01:42 |
| Control Test       | 0.076 | 01:43 |
| Air Blank          | 0.000 | 01:43 |
| Diagnostics Check  | OK    | 01:43 |

---

Data Download Date/Time: 04/07/2021 19:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35019080A3  
DRY GAS EXPIRY DATE: 02/25/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SHOEMAKER

DATE/TIME OF TEST: 04/13/2021 03:26:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:20 |
| Air Blank          | 0.000 | 03:20 |
| Control Test       | 0.078 | 03:21 |
| Air Blank          | 0.000 | 03:21 |
| Subject Sample # 1 | 0.070 | 03:22 |
| Breath Volume      | 3.539 | 03:22 |
| Air Blank          | 0.000 | 03:22 |
| Air Blank          | 0.000 | 03:24 |
| Subject Sample # 2 | 0.067 | 03:24 |
| Breath Volume      | 4.054 | 03:24 |
| Air Blank          | 0.000 | 03:25 |
| Control Test       | 0.077 | 03:25 |
| Air Blank          | 0.000 | 03:26 |
| Diagnostics Check  | OK    | 03:26 |

---

Data Download Date/Time: 05/06/2021 19:31



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35019080A3  
DRY GAS EXPIRY DATE: 02/25/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: SNEAD

DATE/TIME OF TEST: 04/19/2021 01:04:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:55 |
| Air Blank          | 0.000 | 00:55 |
| Control Test       | 0.078 | 00:56 |
| Air Blank          | 0.000 | 00:56 |
| Subject Sample # 1 | 0.158 | 00:59 |
| Breath Volume      | 1.828 | 00:59 |
| Air Blank          | 0.000 | 01:00 |
| Air Blank          | 0.000 | 01:02 |
| Subject Sample # 2 | 0.168 | 01:03 |
| Breath Volume      | 2.183 | 01:03 |
| Air Blank          | 0.000 | 01:03 |
| Control Test       | 0.078 | 01:04 |
| Air Blank          | 0.000 | 01:04 |
| Diagnostics Check  | OK    | 01:04 |

---

Data Download Date/Time: 05/06/2021 19:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35019080A3  
DRY GAS EXPIRY DATE: 02/25/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: DALESSIO

DATE/TIME OF TEST: 04/21/2021 19:33:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:27 |
| Air Blank          | 0.000 | 19:28 |
| Control Test       | 0.078 | 19:28 |
| Air Blank          | 0.000 | 19:29 |
| Subject Sample # 1 | 0.000 | 19:29 |
| Breath Volume      | 2.925 | 19:29 |
| Air Blank          | 0.000 | 19:29 |
| Air Blank          | 0.000 | 19:31 |
| Subject Sample # 2 | 0.000 | 19:32 |
| Breath Volume      | 2.484 | 19:32 |
| Air Blank          | 0.000 | 19:32 |
| Control Test       | 0.077 | 19:33 |
| Air Blank          | 0.000 | 19:33 |
| Diagnostics Check  | OK    | 19:33 |

---

Data Download Date/Time: 05/06/2021 19:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35019080A3  
DRY GAS EXPIRY DATE: 02/25/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: KOBELEV

DATE/TIME OF TEST: 04/22/2021 00:47:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:40 |
| Air Blank          | 0.000 | 00:41 |
| Control Test       | 0.077 | 00:41 |
| Air Blank          | 0.000 | 00:42 |
| Subject Sample # 1 | 0.130 | 00:42 |
| Breath Volume      | 1.199 | 00:42 |
| Air Blank          | 0.000 | 00:43 |
| Air Blank          | 0.000 | 00:45 |
| Subject Sample # 2 | 0.127 | 00:46 |
| Breath Volume      | 1.406 | 00:46 |
| Air Blank          | 0.000 | 00:46 |
| Control Test       | 0.078 | 00:47 |
| Air Blank          | 0.000 | 00:47 |
| Diagnostics Check  | OK    | 00:47 |

---

Data Download Date/Time: 05/06/2021 19:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35019080A3  
DRY GAS EXPIRY DATE: 02/25/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: SCHINDLER

DATE/TIME OF TEST: 04/22/2021 22:41:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:35 |
| Air Blank          | 0.000 | 22:35 |
| Control Test       | 0.078 | 22:36 |
| Air Blank          | 0.000 | 22:36 |
| Subject Sample # 1 | 0.166 | 22:37 |
| Breath Volume      | 2.964 | 22:37 |
| Air Blank          | 0.000 | 22:37 |
| Air Blank          | 0.000 | 22:39 |
| Subject Sample # 2 | 0.162 | 22:40 |
| Breath Volume      | 2.390 | 22:40 |
| Air Blank          | 0.000 | 22:40 |
| Control Test       | 0.077 | 22:41 |
| Air Blank          | 0.000 | 22:41 |
| Diagnostics Check  | OK    | 22:41 |

---

Data Download Date/Time: 05/06/2021 19:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY, KENNETH, R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY, KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 35019080A3  
DRY GAS EXPIRY DATE: 02/25/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: DULANEY

DATE/TIME OF TEST: 04/23/2021 02:22:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:15 |
| Air Blank          | 0.000 | 02:16 |
| Control Test       | 0.079 | 02:16 |
| Air Blank          | 0.000 | 02:16 |
| Subject Sample # 1 | 0.140 | 02:17 |
| Breath Volume      | 3.121 | 02:17 |
| Air Blank          | 0.000 | 02:18 |
| Air Blank          | 0.000 | 02:20 |
| Subject Sample # 2 | 0.135 | 02:20 |
| Breath Volume      | 2.292 | 02:20 |
| Air Blank          | 0.000 | 02:21 |
| Control Test       | 0.077 | 02:21 |
| Air Blank          | 0.000 | 02:22 |
| Diagnostics Check  | OK    | 02:22 |

---

Data Download Date/Time: 05/06/2021 19:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35019080A3

DRY GAS EXPIRY DATE: 02/05/2022

VIOLATION CODE: DUI

ARREST TIME: 01:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 04/10/2021 01:30:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:23 |
| Air Blank          | 0.000 | 01:23 |
| Control Test       | 0.080 | 01:23 |
| Air Blank          | 0.000 | 01:24 |
| Subject Sample # 1 | 0.114 | 01:24 |
| Breath Volume      | 1.621 | 01:24 |
| Air Blank          | 0.000 | 01:25 |
| Air Blank          | 0.000 | 01:27 |
| Subject Sample # 2 | 0.115 | 01:29 |
| Breath Volume      | 2.003 | 01:29 |
| Air Blank          | 0.000 | 01:30 |
| Control Test       | 0.080 | 01:30 |
| Air Blank          | 0.000 | 01:30 |
| Diagnostics Check  | OK    | 01:30 |

---

Data Download Date/Time: 04/15/2021 19:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUINTANA,JORGE,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 03/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PERDOMO

DATE/TIME OF TEST: 04/08/2021 02:02:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:54 |
| Air Blank          | 0.000 | 01:55 |
| Control Test       | 0.079 | 01:55 |
| Air Blank          | 0.000 | 01:56 |
| Subject Sample # 1 | 0.152 | 01:57 |
| Breath Volume      | 1.312 | 01:57 |
| Air Blank          | 0.000 | 01:58 |
| Air Blank          | 0.000 | 02:00 |
| Subject Sample # 2 | 0.167 | 02:00 |
| Breath Volume      | 1.652 | 02:00 |
| Air Blank          | 0.000 | 02:01 |
| Control Test       | 0.078 | 02:01 |
| Air Blank          | 0.000 | 02:02 |
| Diagnostics Check  | OK    | 02:02 |

---

Data Download Date/Time: 04/26/2021 17:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 03/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 04/09/2021 03:15:53

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:12 |                      |
| Air Blank         | 0.000 | 03:13 |                      |
| Control Test      | 0.079 | 03:13 |                      |
| Air Blank         | 0.000 | 03:14 |                      |
| Subject Sample    | # 1   | 03:14 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:14 |                      |
| Air Blank         | 0.000 | 03:14 |                      |
| Control Test      | 0.077 | 03:15 |                      |
| Air Blank         | 0.000 | 03:15 |                      |
| Diagnostics Check | OK    | 03:15 |                      |

---

Data Download Date/Time: 04/26/2021 17:09



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUINTANA,JORGE,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: QUINTANA,JORGE  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 03/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ESTEVES INFANTE

DATE/TIME OF TEST: 04/09/2021 04:11:29

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 03:56 |                      |
| Air Blank         | 0.000     | 03:56 |                      |
| Control Test      | 0.079     | 03:57 |                      |
| Air Blank         | 0.000     | 03:57 |                      |
| Subject Sample    | # 1 0.122 | 03:59 |                      |
| Breath Volume     | 1.566     | 03:59 |                      |
| Air Blank         | 0.000     | 04:00 |                      |
| Air Blank         | 0.000     | 04:02 |                      |
| Subject Sample    | # 2 0.117 | 04:05 | Volume Not Met       |
| Breath Volume     | 0.757     | 04:05 |                      |
| Air Blank         | 0.000     | 04:06 |                      |
| Air Blank         | 0.000     | 04:08 |                      |
| Subject Sample    | # 3       | 04:09 | Subject Test Refused |
| Breath Volume     | 0.656     | 04:09 |                      |
| Air Blank         | 0.000     | 04:10 |                      |
| Control Test      | 0.075     | 04:10 |                      |
| Air Blank         | 0.000     | 04:11 |                      |
| Diagnostics Check | OK        | 04:11 |                      |

---

Data Download Date/Time: 04/26/2021 17:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 03/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: CONDE ARENCIBIA

DATE/TIME OF TEST: 04/10/2021 03:29:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:22 |
| Air Blank          | 0.000 | 03:22 |
| Control Test       | 0.078 | 03:22 |
| Air Blank          | 0.000 | 03:23 |
| Subject Sample # 1 | 0.000 | 03:25 |
| Breath Volume      | 2.917 | 03:25 |
| Air Blank          | 0.000 | 03:25 |
| Air Blank          | 0.000 | 03:27 |
| Subject Sample # 2 | 0.000 | 03:27 |
| Breath Volume      | 2.250 | 03:27 |
| Air Blank          | 0.000 | 03:28 |
| Control Test       | 0.078 | 03:28 |
| Air Blank          | 0.000 | 03:29 |
| Diagnostics Check  | OK    | 03:29 |

---

Data Download Date/Time: 04/26/2021 17:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: QUINTANA,J  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 03/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: CHACON CESPEDES

DATE/TIME OF TEST: 04/11/2021 02:31:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:24 |
| Air Blank          | 0.000 | 02:25 |
| Control Test       | 0.078 | 02:25 |
| Air Blank          | 0.000 | 02:25 |
| Subject Sample # 1 | 0.217 | 02:26 |
| Breath Volume      | 2.289 | 02:26 |
| Air Blank          | 0.000 | 02:27 |
| Air Blank          | 0.000 | 02:29 |
| Subject Sample # 2 | 0.215 | 02:30 |
| Breath Volume      | 2.398 | 02:30 |
| Air Blank          | 0.000 | 02:30 |
| Control Test       | 0.078 | 02:31 |
| Air Blank          | 0.000 | 02:31 |
| Diagnostics Check  | OK    | 02:31 |

---

Data Download Date/Time: 04/26/2021 17:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 03/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: AJTUM HERNANDEZ

DATE/TIME OF TEST: 04/11/2021 04:38:48

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:31 |
| Air Blank          | 0.000 | 04:32 |
| Control Test       | 0.078 | 04:32 |
| Air Blank          | 0.000 | 04:32 |
| Subject Sample # 1 | 0.197 | 04:33 |
| Breath Volume      | 2.085 | 04:33 |
| Air Blank          | 0.000 | 04:34 |
| Air Blank          | 0.000 | 04:36 |
| Subject Sample # 2 | 0.190 | 04:37 |
| Breath Volume      | 2.031 | 04:37 |
| Air Blank          | 0.000 | 04:37 |
| Control Test       | 0.078 | 04:38 |
| Air Blank          | 0.000 | 04:38 |
| Diagnostics Check  | OK    | 04:38 |

---

Data Download Date/Time: 04/26/2021 17:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: SMITH,T  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 03/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 07:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:14

SUBJECT'S LAST NAME: MONTENEGRO MEZA

DATE/TIME OF TEST: 04/11/2021 08:48:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:40 |
| Air Blank          | 0.000 | 08:40 |
| Control Test       | 0.078 | 08:40 |
| Air Blank          | 0.000 | 08:41 |
| Subject Sample # 1 | 0.149 | 08:43 |
| Breath Volume      | 2.796 | 08:43 |
| Air Blank          | 0.000 | 08:44 |
| Air Blank          | 0.000 | 08:45 |
| Subject Sample # 2 | 0.155 | 08:46 |
| Breath Volume      | 3.050 | 08:46 |
| Air Blank          | 0.000 | 08:47 |
| Control Test       | 0.076 | 08:47 |
| Air Blank          | 0.000 | 08:47 |
| Diagnostics Check  | OK    | 08:48 |

---

Data Download Date/Time: 04/26/2021 17:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 03/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:54

SUBJECT'S LAST NAME: PEREZ LUNA

DATE/TIME OF TEST: 04/14/2021 07:46:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:38 |
| Air Blank          | 0.000 | 07:38 |
| Control Test       | 0.078 | 07:39 |
| Air Blank          | 0.000 | 07:39 |
| Subject Sample # 1 | 0.177 | 07:40 |
| Breath Volume      | 1.964 | 07:40 |
| Air Blank          | 0.000 | 07:41 |
| Air Blank          | 0.000 | 07:43 |
| Subject Sample # 2 | 0.179 | 07:44 |
| Breath Volume      | 1.539 | 07:44 |
| Air Blank          | 0.000 | 07:45 |
| Control Test       | 0.078 | 07:45 |
| Air Blank          | 0.000 | 07:46 |
| Diagnostics Check  | OK    | 07:46 |

---

Data Download Date/Time: 04/26/2021 17:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 03/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: SACCHETTONI GERSTL

DATE/TIME OF TEST: 04/17/2021 05:17:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:10 |
| Air Blank          | 0.000 | 05:10 |
| Control Test       | 0.077 | 05:11 |
| Air Blank          | 0.000 | 05:11 |
| Subject Sample # 1 | 0.140 | 05:12 |
| Breath Volume      | 2.480 | 05:12 |
| Air Blank          | 0.000 | 05:13 |
| Air Blank          | 0.000 | 05:15 |
| Subject Sample # 2 | 0.140 | 05:15 |
| Breath Volume      | 2.128 | 05:15 |
| Air Blank          | 0.000 | 05:16 |
| Control Test       | 0.078 | 05:16 |
| Air Blank          | 0.000 | 05:17 |
| Diagnostics Check  | OK    | 05:17 |

---

Data Download Date/Time: 04/26/2021 17:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 04/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: LEGIS

DATE/TIME OF TEST: 04/28/2021 04:02:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:55 |
| Air Blank          | 0.000 | 03:56 |
| Control Test       | 0.079 | 03:56 |
| Air Blank          | 0.000 | 03:57 |
| Subject Sample # 1 | 0.045 | 03:57 |
| Breath Volume      | 2.453 | 03:57 |
| Air Blank          | 0.000 | 03:58 |
| Air Blank          | 0.000 | 03:59 |
| Subject Sample # 2 | 0.042 | 04:00 |
| Breath Volume      | 2.214 | 04:00 |
| Air Blank          | 0.000 | 04:01 |
| Control Test       | 0.079 | 04:01 |
| Air Blank          | 0.000 | 04:01 |
| Diagnostics Check  | OK    | 04:02 |

---

Data Download Date/Time: 05/18/2021 05:56



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: FLEMING,H  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 04/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: BARRIOS

DATE/TIME OF TEST: 04/28/2021 07:21:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:15 |
| Air Blank          | 0.000 | 07:15 |
| Control Test       | 0.078 | 07:16 |
| Air Blank          | 0.000 | 07:16 |
| Subject Sample # 1 | 0.218 | 07:17 |
| Breath Volume      | 1.886 | 07:17 |
| Air Blank          | 0.000 | 07:17 |
| Air Blank          | 0.000 | 07:19 |
| Subject Sample # 2 | 0.200 | 07:20 |
| Breath Volume      | 1.289 | 07:20 |
| Air Blank          | 0.000 | 07:20 |
| Control Test       | 0.080 | 07:21 |
| Air Blank          | 0.000 | 07:21 |
| Diagnostics Check  | OK    | 07:21 |

---

Data Download Date/Time: 05/18/2021 05:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ODOR,ROBERT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: BAYON HERNANDEZ

DATE/TIME OF TEST: 04/03/2021 05:14:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:08 |
| Air Blank          | 0.000 | 05:08 |
| Control Test       | 0.079 | 05:08 |
| Air Blank          | 0.000 | 05:09 |
| Subject Sample # 1 | 0.108 | 05:10 |
| Breath Volume      | 2.101 | 05:10 |
| Air Blank          | 0.000 | 05:10 |
| Air Blank          | 0.000 | 05:12 |
| Subject Sample # 2 | 0.104 | 05:13 |
| Breath Volume      | 1.769 | 05:13 |
| Air Blank          | 0.000 | 05:13 |
| Control Test       | 0.078 | 05:14 |
| Air Blank          | 0.000 | 05:14 |
| Diagnostics Check  | OK    | 05:14 |

---

Data Download Date/Time: 04/07/2021 11:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: PETTIGNY,VANEL  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: QUINTANILLA

DATE/TIME OF TEST: 04/03/2021 22:59:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:52 |
| Air Blank          | 0.000 | 22:52 |
| Control Test       | 0.078 | 22:53 |
| Air Blank          | 0.000 | 22:53 |
| Subject Sample # 1 | 0.138 | 22:55 |
| Breath Volume      | 1.113 | 22:55 |
| Air Blank          | 0.000 | 22:55 |
| Air Blank          | 0.000 | 22:57 |
| Subject Sample # 2 | 0.151 | 22:57 |
| Breath Volume      | 1.480 | 22:57 |
| Air Blank          | 0.000 | 22:58 |
| Control Test       | 0.077 | 22:58 |
| Air Blank          | 0.000 | 22:59 |
| Diagnostics Check  | OK    | 22:59 |

---

Data Download Date/Time: 04/07/2021 11:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LUNA,KRISTINA  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: LOUIS

DATE/TIME OF TEST: 04/04/2021 19:58:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:49 |
| Air Blank          | 0.000 | 19:50 |
| Control Test       | 0.079 | 19:50 |
| Air Blank          | 0.000 | 19:51 |
| Subject Sample # 1 | 0.000 | 19:52 |
| Breath Volume      | 1.792 | 19:52 |
| Air Blank          | 0.000 | 19:52 |
| Air Blank          | 0.000 | 19:54 |
| Subject Sample # 2 | 0.000 | 19:57 |
| Breath Volume      | 1.539 | 19:57 |
| Air Blank          | 0.000 | 19:57 |
| Control Test       | 0.081 | 19:57 |
| Air Blank          | 0.000 | 19:58 |
| Diagnostics Check  | OK    | 19:58 |

---

Data Download Date/Time: 04/07/2021 11:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: MAGANA,ELIAS  
ARRESTING OFFICER AGENCY: Plantation PD  
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: HANIFF-STOUTE

DATE/TIME OF TEST: 04/05/2021 02:17:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:06 |
| Air Blank          | 0.000 | 02:07 |
| Control Test       | 0.079 | 02:07 |
| Air Blank          | 0.000 | 02:08 |
| Subject Sample # 1 | 0.213 | 02:09 |
| Breath Volume      | 1.160 | 02:09 |
| Air Blank          | 0.000 | 02:09 |
| Air Blank          | 0.000 | 02:11 |
| Subject Sample # 2 | 0.187 | 02:12 |
| Breath Volume      | 1.964 | 02:12 |
| Air Blank          | 0.000 | 02:13 |
| Air Blank          | 0.000 | 02:14 |
| Subject Sample # 3 | 0.200 | 02:15 |
| Breath Volume      | 1.175 | 02:15 |
| Air Blank          | 0.000 | 02:16 |
| Control Test       | 0.077 | 02:16 |
| Air Blank          | 0.000 | 02:17 |
| Diagnostics Check  | OK    | 02:17 |

---

Data Download Date/Time: 04/07/2021 11:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: GUE,R  
ARRESTING OFFICER AGENCY: Plantation PD  
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SANDOVAL

DATE/TIME OF TEST: 04/05/2021 04:04:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:58 |
| Air Blank          | 0.000 | 03:59 |
| Control Test       | 0.079 | 03:59 |
| Air Blank          | 0.000 | 03:59 |
| Subject Sample # 1 | 0.154 | 04:00 |
| Breath Volume      | 3.472 | 04:00 |
| Air Blank          | 0.000 | 04:01 |
| Air Blank          | 0.000 | 04:02 |
| Subject Sample # 2 | 0.148 | 04:03 |
| Breath Volume      | 3.542 | 04:03 |
| Air Blank          | 0.000 | 04:03 |
| Control Test       | 0.078 | 04:04 |
| Air Blank          | 0.000 | 04:04 |
| Diagnostics Check  | OK    | 04:04 |

---

Data Download Date/Time: 04/07/2021 11:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: HALES,ANDREW  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 11:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:49

SUBJECT'S LAST NAME: CHAMPAGNE

DATE/TIME OF TEST: 04/06/2021 13:20:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:13 |
| Air Blank          | 0.000 | 13:14 |
| Control Test       | 0.079 | 13:14 |
| Air Blank          | 0.000 | 13:15 |
| Subject Sample # 1 | 0.267 | 13:16 |
| Breath Volume      | 1.703 | 13:16 |
| Air Blank          | 0.000 | 13:16 |
| Air Blank          | 0.000 | 13:18 |
| Subject Sample # 2 | 0.260 | 13:19 |
| Breath Volume      | 1.542 | 13:19 |
| Air Blank          | 0.000 | 13:19 |
| Control Test       | 0.077 | 13:20 |
| Air Blank          | 0.000 | 13:20 |
| Diagnostics Check  | OK    | 13:20 |

---

Data Download Date/Time: 04/07/2021 11:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: FACCINI,DAVID  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: KIRKMAN

DATE/TIME OF TEST: 04/12/2021 00:24:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:17 |
| Air Blank          | 0.000 | 00:18 |
| Control Test       | 0.078 | 00:18 |
| Air Blank          | 0.000 | 00:19 |
| Subject Sample # 1 | 0.000 | 00:19 |
| Breath Volume      | 2.359 | 00:19 |
| Air Blank          | 0.000 | 00:20 |
| Air Blank          | 0.000 | 00:22 |
| Subject Sample # 2 | 0.000 | 00:22 |
| Breath Volume      | 1.816 | 00:22 |
| Air Blank          | 0.000 | 00:23 |
| Control Test       | 0.078 | 00:23 |
| Air Blank          | 0.000 | 00:24 |
| Diagnostics Check  | OK    | 00:24 |

---

Data Download Date/Time: 05/12/2021 11:03



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: MAGANA,ELIAS  
ARRESTING OFFICER AGENCY: Plantation PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: III BLUE

DATE/TIME OF TEST: 04/15/2021 05:31:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:25 |
| Air Blank          | 0.000 | 05:25 |
| Control Test       | 0.079 | 05:25 |
| Air Blank          | 0.000 | 05:26 |
| Subject Sample # 1 | 0.157 | 05:27 |
| Breath Volume      | 1.933 | 05:27 |
| Air Blank          | 0.000 | 05:27 |
| Air Blank          | 0.000 | 05:29 |
| Subject Sample # 2 | 0.168 | 05:30 |
| Breath Volume      | 2.640 | 05:30 |
| Air Blank          | 0.000 | 05:30 |
| Control Test       | 0.078 | 05:31 |
| Air Blank          | 0.000 | 05:31 |
| Diagnostics Check  | OK    | 05:31 |

---

Data Download Date/Time: 05/12/2021 11:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: PARRISH,MATTHEW  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 07:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 04/22/2021 09:10:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:03 |
| Air Blank          | 0.000 | 09:04 |
| Control Test       | 0.080 | 09:04 |
| Air Blank          | 0.000 | 09:05 |
| Subject Sample # 1 | 0.000 | 09:05 |
| Breath Volume      | 1.550 | 09:05 |
| Air Blank          | 0.000 | 09:06 |
| Air Blank          | 0.000 | 09:08 |
| Subject Sample # 2 | 0.000 | 09:08 |
| Breath Volume      | 1.261 | 09:08 |
| Air Blank          | 0.000 | 09:09 |
| Control Test       | 0.078 | 09:09 |
| Air Blank          | 0.000 | 09:10 |
| Diagnostics Check  | OK    | 09:10 |

---

Data Download Date/Time: 05/12/2021 11:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUZMAN,JOSE,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: GUZMAN,JOSE  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: DAOULY

DATE/TIME OF TEST: 04/25/2021 03:10:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:01 |
| Air Blank          | 0.000 | 03:02 |
| Control Test       | 0.078 | 03:02 |
| Air Blank          | 0.000 | 03:02 |
| Subject Sample # 1 | 0.174 | 03:05 |
| Breath Volume      | 1.179 | 03:05 |
| Air Blank          | 0.000 | 03:06 |
| Air Blank          | 0.000 | 03:07 |
| Subject Sample # 2 | 0.172 | 03:09 |
| Breath Volume      | 1.359 | 03:09 |
| Air Blank          | 0.000 | 03:10 |
| Control Test       | 0.075 | 03:10 |
| Air Blank          | 0.000 | 03:10 |
| Diagnostics Check  | OK    | 03:10 |

---

Data Download Date/Time: 05/12/2021 11:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS  
ARRESTING OFFICER AGENCY: Miramar PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: COVINGTON

DATE/TIME OF TEST: 04/28/2021 23:37:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:30 |
| Air Blank          | 0.000 | 23:31 |
| Control Test       | 0.079 | 23:31 |
| Air Blank          | 0.000 | 23:32 |
| Subject Sample # 1 | 0.064 | 23:32 |
| Breath Volume      | 1.652 | 23:32 |
| Air Blank          | 0.000 | 23:33 |
| Air Blank          | 0.000 | 23:35 |
| Subject Sample # 2 | 0.054 | 23:35 |
| Breath Volume      | 1.753 | 23:35 |
| Air Blank          | 0.000 | 23:36 |
| Control Test       | 0.079 | 23:36 |
| Air Blank          | 0.000 | 23:37 |
| Diagnostics Check  | OK    | 23:37 |

---

Data Download Date/Time: 05/12/2021 11:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: MALAVE,WILLIAM  
ARRESTING OFFICER AGENCY: Miramar PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 04/29/2021 05:16:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:10 |
| Air Blank          | 0.000 | 05:10 |
| Control Test       | 0.078 | 05:11 |
| Air Blank          | 0.000 | 05:11 |
| Subject Sample # 1 | 0.139 | 05:12 |
| Breath Volume      | 1.347 | 05:12 |
| Air Blank          | 0.000 | 05:12 |
| Air Blank          | 0.000 | 05:14 |
| Subject Sample # 2 | 0.154 | 05:15 |
| Breath Volume      | 2.363 | 05:15 |
| Air Blank          | 0.000 | 05:15 |
| Control Test       | 0.077 | 05:16 |
| Air Blank          | 0.000 | 05:16 |
| Diagnostics Check  | OK    | 05:16 |

---

Data Download Date/Time: 05/12/2021 11:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOHLIN,MARK,T  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: SWENSON,BRIAN  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 04/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: P623779873820

DATE/TIME OF TEST: 04/14/2021 21:04:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:58 |
| Air Blank          | 0.000 | 20:58 |
| Control Test       | 0.081 | 20:59 |
| Air Blank          | 0.000 | 20:59 |
| Subject Sample # 1 | 0.329 | 21:00 |
| Breath Volume      | 2.769 | 21:00 |
| Air Blank          | 0.000 | 21:00 |
| Air Blank          | 0.000 | 21:02 |
| Subject Sample # 2 | 0.321 | 21:03 |
| Breath Volume      | 2.718 | 21:03 |
| Air Blank          | 0.000 | 21:04 |
| Control Test       | 0.078 | 21:04 |
| Air Blank          | 0.000 | 21:04 |
| Diagnostics Check  | OK    | 21:04 |

---

Data Download Date/Time: 04/19/2021 04:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOHLIN,MARK,T  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: SCHREINER,CORY  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 04/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:23

SUBJECT'S LAST NAME: LLEWELLYN

DATE/TIME OF TEST: 04/21/2021 21:33:14

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 21:27 |                    |
| Air Blank         | 0.000 | 21:28 |                    |
| Control Test      | 0.083 | 21:28 |                    |
| Air Blank         | 0.000 | 21:29 |                    |
| Subject Sample    | # 1   | 21:32 | No Sample Provided |
| Breath Volume     | 0.000 | 21:32 |                    |
| Air Blank         | 0.000 | 21:32 |                    |
| Air Blank         | 0.000 | 21:33 |                    |

---

Data Download Date/Time: 04/30/2021 22:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOHLIN,MARK,T  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: SCHREINER,CORY  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 04/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: LLEWELLYN

DATE/TIME OF TEST: 04/21/2021 21:49:28

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:46 |                      |
| Air Blank         | 0.000 | 21:47 |                      |
| Control Test      | 0.083 | 21:47 |                      |
| Air Blank         | 0.000 | 21:47 |                      |
| Subject Sample    | # 1   | 21:48 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:48 |                      |
| Air Blank         | 0.000 | 21:48 |                      |
| Control Test      | 0.083 | 21:48 |                      |
| Air Blank         | 0.000 | 21:49 |                      |
| Diagnostics Check | OK    | 21:49 |                      |

---

Data Download Date/Time: 04/30/2021 22:45



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,JAMES,E  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: MARTIN,JAMES  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 04/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202323  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: BRITTON

DATE/TIME OF TEST: 04/23/2021 22:53:49

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:44 |                      |
| Air Blank         | 0.000 | 22:44 |                      |
| Control Test      | 0.082 | 22:44 |                      |
| Air Blank         | 0.000 | 22:45 |                      |
| Subject Sample    | # 1   | 22:48 | No Sample Provided   |
| Breath Volume     | 0.000 | 22:48 |                      |
| Air Blank         | 0.000 | 22:48 |                      |
| Air Blank         | 0.000 | 22:50 |                      |
| Subject Sample    | # 2   | 22:52 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:52 |                      |
| Air Blank         | 0.000 | 22:52 |                      |
| Control Test      | 0.081 | 22:53 |                      |
| Air Blank         | 0.000 | 22:53 |                      |
| Diagnostics Check | OK    | 22:53 |                      |

---

Data Download Date/Time: 04/30/2021 22:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,JAMES,W  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: SIVORE

DATE/TIME OF TEST: 04/08/2021 22:13:28

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:10 |                      |
| Air Blank         | 0.000 | 22:10 |                      |
| Control Test      | 0.078 | 22:10 |                      |
| Air Blank         | 0.000 | 22:11 |                      |
| Subject Sample    | # 1   | 22:12 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:12 |                      |
| Air Blank         | 0.000 | 22:12 |                      |
| Control Test      | 0.078 | 22:12 |                      |
| Air Blank         | 0.000 | 22:13 |                      |
| Diagnostics Check | OK    | 22:13 |                      |

---

Data Download Date/Time: 05/12/2021 10:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,JAMES,W  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: COLUMBO

DATE/TIME OF TEST: 04/14/2021 21:36:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:27 |
| Air Blank          | 0.000 | 21:27 |
| Control Test       | 0.078 | 21:28 |
| Air Blank          | 0.000 | 21:28 |
| Subject Sample # 1 | 0.161 | 21:31 |
| Breath Volume      | 1.148 | 21:31 |
| Air Blank          | 0.000 | 21:31 |
| Air Blank          | 0.000 | 21:33 |
| Subject Sample # 2 | 0.176 | 21:34 |
| Breath Volume      | 1.148 | 21:34 |
| Air Blank          | 0.000 | 21:35 |
| Control Test       | 0.077 | 21:35 |
| Air Blank          | 0.000 | 21:36 |
| Diagnostics Check  | OK    | 21:36 |

---

Data Download Date/Time: 05/12/2021 10:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,JAMES,W  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 04/16/2021 00:13:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:04 |
| Air Blank          | 0.000 | 00:05 |
| Control Test       | 0.078 | 00:05 |
| Air Blank          | 0.000 | 00:06 |
| Subject Sample # 1 | 0.202 | 00:08 |
| Breath Volume      | 2.574 | 00:08 |
| Air Blank          | 0.000 | 00:08 |
| Air Blank          | 0.000 | 00:10 |
| Subject Sample # 2 | 0.207 | 00:11 |
| Breath Volume      | 2.558 | 00:11 |
| Air Blank          | 0.000 | 00:12 |
| Control Test       | 0.076 | 00:12 |
| Air Blank          | 0.000 | 00:12 |
| Diagnostics Check  | OK    | 00:13 |

---

Data Download Date/Time: 05/12/2021 10:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,JAMES,W  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: LA CROIX

DATE/TIME OF TEST: 04/16/2021 03:54:43

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 03:40 |                |
| Air Blank         | 0.000     | 03:41 |                |
| Control Test      | 0.078     | 03:41 |                |
| Air Blank         | 0.000     | 03:42 |                |
| Subject Sample    | # 1 0.054 | 03:45 | Volume Not Met |
| Breath Volume     | 0.070     | 03:45 |                |
| Air Blank         | 0.000     | 03:45 |                |
| Air Blank         | 0.000     | 03:47 |                |
| Subject Sample    | # 2 0.062 | 03:48 |                |
| Breath Volume     | 1.617     | 03:48 |                |
| Air Blank         | 0.000     | 03:49 |                |
| Air Blank         | 0.000     | 03:51 |                |
| Subject Sample    | # 3 0.058 | 03:53 |                |
| Breath Volume     | 1.460     | 03:53 |                |
| Air Blank         | 0.000     | 03:53 |                |
| Control Test      | 0.078     | 03:54 |                |
| Air Blank         | 0.000     | 03:54 |                |
| Diagnostics Check | OK        | 03:54 |                |

---

Data Download Date/Time: 05/12/2021 10:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STONER,NATHAN,C  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 04/20/2021 23:29:14

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:25 |                      |
| Air Blank         | 0.000 | 23:25 |                      |
| Control Test      | 0.077 | 23:26 |                      |
| Air Blank         | 0.000 | 23:26 |                      |
| Subject Sample    | # 1   | 23:27 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:27 |                      |
| Air Blank         | 0.000 | 23:28 |                      |
| Control Test      | 0.078 | 23:28 |                      |
| Air Blank         | 0.000 | 23:29 |                      |
| Diagnostics Check | OK    | 23:29 |                      |

---

Data Download Date/Time: 05/12/2021 10:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,JAMES,W  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 04/22/2021 01:32:00

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:27 |                      |
| Air Blank         | 0.000 | 01:28 |                      |
| Control Test      | 0.077 | 01:28 |                      |
| Air Blank         | 0.000 | 01:29 |                      |
| Subject Sample    | # 1   | 01:30 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:30 |                      |
| Air Blank         | 0.000 | 01:31 |                      |
| Control Test      | 0.078 | 01:31 |                      |
| Air Blank         | 0.000 | 01:31 |                      |
| Diagnostics Check | OK    | 01:32 |                      |

---

Data Download Date/Time: 05/12/2021 10:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,JAMES,W  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: GOLDBERG

DATE/TIME OF TEST: 04/22/2021 23:34:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:27 |
| Air Blank          | 0.000 | 23:28 |
| Control Test       | 0.079 | 23:28 |
| Air Blank          | 0.000 | 23:28 |
| Subject Sample # 1 | 0.213 | 23:29 |
| Breath Volume      | 2.699 | 23:29 |
| Air Blank          | 0.000 | 23:30 |
| Air Blank          | 0.000 | 23:32 |
| Subject Sample # 2 | 0.212 | 23:32 |
| Breath Volume      | 1.273 | 23:32 |
| Air Blank          | 0.000 | 23:33 |
| Control Test       | 0.077 | 23:33 |
| Air Blank          | 0.000 | 23:33 |
| Diagnostics Check  | OK    | 23:34 |

---

Data Download Date/Time: 05/12/2021 10:21



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STONER,NATHAN,C  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: LAINEZ

DATE/TIME OF TEST: 04/28/2021 02:20:11

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:16 |                      |
| Air Blank         | 0.000 | 02:17 |                      |
| Control Test      | 0.079 | 02:17 |                      |
| Air Blank         | 0.000 | 02:18 |                      |
| Subject Sample    | # 1   | 02:18 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:18 |                      |
| Air Blank         | 0.000 | 02:19 |                      |
| Control Test      | 0.078 | 02:19 |                      |
| Air Blank         | 0.000 | 02:20 |                      |
| Diagnostics Check | OK    | 02:20 |                      |

---

Data Download Date/Time: 05/12/2021 10:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTER,JAMES,W  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: RODRIGUEZ GAINZA

DATE/TIME OF TEST: 04/30/2021 03:29:37

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:26 |                      |
| Air Blank         | 0.000 | 03:26 |                      |
| Control Test      | 0.078 | 03:27 |                      |
| Air Blank         | 0.000 | 03:27 |                      |
| Subject Sample    | # 1   | 03:28 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:28 |                      |
| Air Blank         | 0.000 | 03:28 |                      |
| Control Test      | 0.078 | 03:29 |                      |
| Air Blank         | 0.000 | 03:29 |                      |
| Diagnostics Check | OK    | 03:29 |                      |

---

Data Download Date/Time: 05/12/2021 10:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STONER,NATHAN,C  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: STONER,NATHAN  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: GIL

DATE/TIME OF TEST: 04/26/2021 02:37:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:30 |
| Air Blank          | 0.000 | 02:30 |
| Control Test       | 0.077 | 02:30 |
| Air Blank          | 0.000 | 02:31 |
| Subject Sample # 1 | 0.178 | 02:32 |
| Breath Volume      | 1.464 | 02:32 |
| Air Blank          | 0.000 | 02:33 |
| Air Blank          | 0.000 | 02:34 |
| Subject Sample # 2 | 0.179 | 02:35 |
| Breath Volume      | 1.203 | 02:35 |
| Air Blank          | 0.000 | 02:36 |
| Control Test       | 0.078 | 02:36 |
| Air Blank          | 0.000 | 02:37 |
| Diagnostics Check  | OK    | 02:37 |

---

Data Download Date/Time: 05/12/2021 11:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STONER,NATHAN,C  
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD  
ARRESTING OFFICER: STONER,NATHAN  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: KIRTON

DATE/TIME OF TEST: 04/27/2021 00:27:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:21 |
| Air Blank          | 0.000 | 00:21 |
| Control Test       | 0.078 | 00:21 |
| Air Blank          | 0.000 | 00:22 |
| Subject Sample # 1 | 0.268 | 00:23 |
| Breath Volume      | 4.605 | 00:23 |
| Air Blank          | 0.000 | 00:23 |
| Air Blank          | 0.000 | 00:25 |
| Subject Sample # 2 | 0.272 | 00:26 |
| Breath Volume      | 4.792 | 00:26 |
| Air Blank          | 0.000 | 00:27 |
| Control Test       | 0.078 | 00:27 |
| Air Blank          | 0.000 | 00:27 |
| Diagnostics Check  | OK    | 00:27 |

---

Data Download Date/Time: 05/12/2021 11:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLOPS,C,M  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: BAXTER,J  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119070A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: GARCIA MARTINEZ

DATE/TIME OF TEST: 04/10/2021 00:23:01

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 00:21 |            |
| Air Blank         | 0.000 | 00:22 |            |
|                   |       | 00:22 | RFI Detect |
| Air Blank         | 0.000 | 00:23 |            |

---

Data Download Date/Time: 05/11/2021 12:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLOPS,C,M  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: BAXTER,J  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119070A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: GARCIA MARTINEZ

DATE/TIME OF TEST: 04/10/2021 00:38:47

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 00:25 |                      |
| Air Blank         | 0.000     | 00:25 |                      |
| Control Test      | 0.078     | 00:26 |                      |
| Air Blank         | 0.000     | 00:26 |                      |
| Subject Sample    | # 1 0.192 | 00:28 |                      |
| Breath Volume     | 1.171     | 00:28 |                      |
| Air Blank         | 0.000     | 00:29 |                      |
| Air Blank         | 0.000     | 00:31 |                      |
| Subject Sample    | # 2       | 00:34 | No Sample Provided   |
| Breath Volume     | 0.000     | 00:34 |                      |
| Air Blank         | 0.000     | 00:35 |                      |
| Air Blank         | 0.000     | 00:37 |                      |
| Subject Sample    | # 3       | 00:37 | Subject Test Refused |
| Breath Volume     | 0.000     | 00:37 |                      |
| Air Blank         | 0.000     | 00:37 |                      |
| Control Test      | 0.077     | 00:38 |                      |
| Air Blank         | 0.000     | 00:38 |                      |
| Diagnostics Check | OK        | 00:38 |                      |

---

Data Download Date/Time: 05/11/2021 12:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TURNER,S,C  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: GONZALEZ,C  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 24119070A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: THACKTHAY

DATE/TIME OF TEST: 04/10/2021 01:03:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:56 |
| Air Blank          | 0.000 | 00:57 |
| Control Test       | 0.077 | 00:57 |
| Air Blank          | 0.000 | 00:58 |
| Subject Sample # 1 | 0.175 | 00:58 |
| Breath Volume      | 1.972 | 00:58 |
| Air Blank          | 0.000 | 00:59 |
| Air Blank          | 0.000 | 01:00 |
| Subject Sample # 2 | 0.165 | 01:01 |
| Breath Volume      | 1.925 | 01:01 |
| Air Blank          | 0.000 | 01:02 |
| Control Test       | 0.077 | 01:02 |
| Air Blank          | 0.000 | 01:03 |
| Diagnostics Check  | OK    | 01:03 |

---

Data Download Date/Time: 05/11/2021 12:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PLUSA,KYLE,A  
BREATH TEST OPERATOR AGENCY: FFWCC (SW REGION)  
ARRESTING OFFICER: PLUSA,KYLE  
ARRESTING OFFICER AGENCY: FFWCC (SW REGION)  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119070A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: BUI  
ARREST TIME: 20:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 04/10/2021 22:52:33

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:49 |                      |
| Air Blank         | 0.000 | 22:50 |                      |
| Control Test      | 0.078 | 22:50 |                      |
| Air Blank         | 0.000 | 22:50 |                      |
| Subject Sample    | # 1   | 22:51 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:51 |                      |
| Air Blank         | 0.000 | 22:51 |                      |
| Control Test      | 0.078 | 22:51 |                      |
| Air Blank         | 0.000 | 22:52 |                      |
| Diagnostics Check | OK    | 22:52 |                      |

---

Data Download Date/Time: 05/11/2021 12:34



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELIAS,D,  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: ELIAS,D  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119070A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: VIECHEC

DATE/TIME OF TEST: 04/11/2021 01:58:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:52 |
| Air Blank          | 0.000 | 01:52 |
| Control Test       | 0.079 | 01:53 |
| Air Blank          | 0.000 | 01:53 |
| Subject Sample # 1 | 0.123 | 01:54 |
| Breath Volume      | 2.628 | 01:54 |
| Air Blank          | 0.000 | 01:54 |
| Air Blank          | 0.000 | 01:56 |
| Subject Sample # 2 | 0.119 | 01:56 |
| Breath Volume      | 1.578 | 01:56 |
| Air Blank          | 0.000 | 01:57 |
| Control Test       | 0.078 | 01:57 |
| Air Blank          | 0.000 | 01:58 |
| Diagnostics Check  | OK    | 01:58 |

---

Data Download Date/Time: 05/11/2021 12:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOEL,Y,  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: GONZALEZ,C  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119070A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 04/11/2021 04:48:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:42 |
| Air Blank          | 0.000 | 04:42 |
| Control Test       | 0.079 | 04:42 |
| Air Blank          | 0.000 | 04:43 |
| Subject Sample # 1 | 0.067 | 04:43 |
| Breath Volume      | 1.332 | 04:43 |
| Air Blank          | 0.000 | 04:44 |
| Air Blank          | 0.000 | 04:46 |
| Subject Sample # 2 | 0.065 | 04:46 |
| Breath Volume      | 2.394 | 04:46 |
| Air Blank          | 0.000 | 04:47 |
| Control Test       | 0.077 | 04:47 |
| Air Blank          | 0.000 | 04:48 |
| Diagnostics Check  | OK    | 04:48 |

---

Data Download Date/Time: 05/11/2021 12:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANNORAT,ADELY,  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: LEONI,S  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119070A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: SOLIS

DATE/TIME OF TEST: 04/14/2021 01:07:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:00 |
| Air Blank          | 0.000 | 01:01 |
| Control Test       | 0.079 | 01:01 |
| Air Blank          | 0.000 | 01:02 |
| Subject Sample # 1 | 0.188 | 01:02 |
| Breath Volume      | 3.082 | 01:02 |
| Air Blank          | 0.000 | 01:03 |
| Air Blank          | 0.000 | 01:05 |
| Subject Sample # 2 | 0.176 | 01:05 |
| Breath Volume      | 3.785 | 01:05 |
| Air Blank          | 0.000 | 01:06 |
| Control Test       | 0.078 | 01:06 |
| Air Blank          | 0.000 | 01:06 |
| Diagnostics Check  | OK    | 01:07 |

---

Data Download Date/Time: 05/11/2021 12:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUTLER JR,SCOTT,A  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: ANTONINO,C  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119070A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/15/2021 05:23:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:16 |
| Air Blank          | 0.000 | 05:16 |
| Control Test       | 0.077 | 05:17 |
| Air Blank          | 0.000 | 05:17 |
| Subject Sample # 1 | 0.125 | 05:18 |
| Breath Volume      | 1.976 | 05:18 |
| Air Blank          | 0.000 | 05:18 |
| Air Blank          | 0.000 | 05:20 |
| Subject Sample # 2 | 0.138 | 05:21 |
| Breath Volume      | 2.734 | 05:21 |
| Air Blank          | 0.000 | 05:22 |
| Control Test       | 0.078 | 05:22 |
| Air Blank          | 0.000 | 05:22 |
| Diagnostics Check  | OK    | 05:23 |

---

Data Download Date/Time: 05/11/2021 12:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARRILLO,XAVIER,R  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: MCCOY,J  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119070A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:28

SUBJECT'S LAST NAME: RILEA

DATE/TIME OF TEST: 04/16/2021 15:55:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:48 |
| Air Blank          | 0.000 | 15:48 |
| Control Test       | 0.079 | 15:48 |
| Air Blank          | 0.000 | 15:49 |
| Subject Sample # 1 | 0.000 | 15:50 |
| Breath Volume      | 1.132 | 15:50 |
| Air Blank          | 0.000 | 15:50 |
| Air Blank          | 0.000 | 15:52 |
| Subject Sample # 2 | 0.000 | 15:53 |
| Breath Volume      | 1.542 | 15:53 |
| Air Blank          | 0.000 | 15:54 |
| Control Test       | 0.077 | 15:54 |
| Air Blank          | 0.000 | 15:55 |
| Diagnostics Check  | OK    | 15:55 |

---

Data Download Date/Time: 05/11/2021 12:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,ZACHARY,H  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: WILLIAMS,ZACHARY  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119070A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 04/17/2021 02:56:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:50 |
| Air Blank          | 0.000 | 02:50 |
| Control Test       | 0.078 | 02:51 |
| Air Blank          | 0.000 | 02:51 |
| Subject Sample # 1 | 0.000 | 02:52 |
| Breath Volume      | 1.589 | 02:52 |
| Air Blank          | 0.000 | 02:52 |
| Air Blank          | 0.000 | 02:54 |
| Subject Sample # 2 | 0.000 | 02:55 |
| Breath Volume      | 1.710 | 02:55 |
| Air Blank          | 0.000 | 02:55 |
| Control Test       | 0.077 | 02:56 |
| Air Blank          | 0.000 | 02:56 |
| Diagnostics Check  | OK    | 02:56 |

---

Data Download Date/Time: 05/11/2021 12:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GASH,C,  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: GASH,C  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: CARMAN

DATE/TIME OF TEST: 04/17/2021 21:37:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:31 |
| Air Blank          | 0.000 | 21:31 |
| Control Test       | 0.078 | 21:31 |
| Air Blank          | 0.000 | 21:32 |
| Subject Sample # 1 | 0.084 | 21:33 |
| Breath Volume      | 4.203 | 21:33 |
| Air Blank          | 0.000 | 21:33 |
| Air Blank          | 0.000 | 21:35 |
| Subject Sample # 2 | 0.076 | 21:36 |
| Breath Volume      | 2.199 | 21:36 |
| Air Blank          | 0.000 | 21:36 |
| Control Test       | 0.078 | 21:37 |
| Air Blank          | 0.000 | 21:37 |
| Diagnostics Check  | OK    | 21:37 |

---

Data Download Date/Time: 05/11/2021 12:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THROESCH,PATRICK,J  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: TERUZ,G  
ARRESTING OFFICER AGENCY: FHP TROOP F  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: SIERRA

DATE/TIME OF TEST: 04/18/2021 02:08:56

---

|                   |           |       |                           |
|-------------------|-----------|-------|---------------------------|
| Diagnostics Check | OK        | 01:57 |                           |
| Air Blank         | 0.000     | 01:58 |                           |
| Control Test      | 0.078     | 01:58 |                           |
| Air Blank         | 0.000     | 01:59 |                           |
| Subject Sample    | # 1 0.193 | 01:59 |                           |
| Breath Volume     | 1.539     | 01:59 |                           |
| Air Blank         | 0.000     | 02:00 |                           |
| Air Blank         | 0.000     | 02:02 |                           |
| Subject Sample    | # 2 0.167 | 02:04 |                           |
| Breath Volume     | 1.210     | 02:04 |                           |
| Air Blank         | 0.000     | 02:05 |                           |
| Air Blank         | 0.000     | 02:06 |                           |
| Subject Sample    | # 3 0.156 | 02:07 |                           |
| Breath Volume     | 1.117     | 02:07 |                           |
| Air Blank         | 0.000     | 02:08 |                           |
| Control Test      | 0.074     | 02:08 | Control Outside Tolerance |
| Air Blank         | 0.000     | 02:08 |                           |

---

Data Download Date/Time: 05/11/2021 12:34



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROGERS,MATTHEW,L  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: JEAN,P  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 04/18/2021 19:58:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:52 |
| Air Blank          | 0.000 | 19:52 |
| Control Test       | 0.079 | 19:53 |
| Air Blank          | 0.000 | 19:53 |
| Subject Sample # 1 | 0.153 | 19:54 |
| Breath Volume      | 3.328 | 19:54 |
| Air Blank          | 0.000 | 19:54 |
| Air Blank          | 0.000 | 19:56 |
| Subject Sample # 2 | 0.155 | 19:57 |
| Breath Volume      | 3.535 | 19:57 |
| Air Blank          | 0.000 | 19:57 |
| Control Test       | 0.078 | 19:58 |
| Air Blank          | 0.000 | 19:58 |
| Diagnostics Check  | OK    | 19:58 |

---

Data Download Date/Time: 05/11/2021 12:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,S,  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: TERCERO,K  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119070A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 18:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: BRADSHAW

DATE/TIME OF TEST: 04/18/2021 20:37:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:31 |
| Air Blank          | 0.000 | 20:31 |
| Control Test       | 0.078 | 20:31 |
| Air Blank          | 0.000 | 20:32 |
| Subject Sample # 1 | 0.181 | 20:32 |
| Breath Volume      | 1.238 | 20:32 |
| Air Blank          | 0.000 | 20:33 |
| Air Blank          | 0.000 | 20:35 |
| Subject Sample # 2 | 0.182 | 20:35 |
| Breath Volume      | 1.238 | 20:35 |
| Air Blank          | 0.000 | 20:36 |
| Control Test       | 0.077 | 20:36 |
| Air Blank          | 0.000 | 20:37 |
| Diagnostics Check  | OK    | 20:37 |

---

Data Download Date/Time: 05/11/2021 12:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,ZACHARY,H  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: WILLIAMS,ZACHARY  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119070A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 04/18/2021 22:53:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:47 |
| Air Blank          | 0.000 | 22:47 |
| Control Test       | 0.078 | 22:48 |
| Air Blank          | 0.000 | 22:48 |
| Subject Sample # 1 | 0.171 | 22:49 |
| Breath Volume      | 3.832 | 22:49 |
| Air Blank          | 0.000 | 22:49 |
| Air Blank          | 0.000 | 22:51 |
| Subject Sample # 2 | 0.167 | 22:52 |
| Breath Volume      | 3.417 | 22:52 |
| Air Blank          | 0.000 | 22:52 |
| Control Test       | 0.076 | 22:53 |
| Air Blank          | 0.000 | 22:53 |
| Diagnostics Check  | OK    | 22:53 |

---

Data Download Date/Time: 05/11/2021 12:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAYLOR,E,W  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: RODRIGUES,V  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119070A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 04/18/2021 23:18:19

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 23:17 |            |
| Air Blank         |       | 23:17 | RFI Detect |
| Air Blank         | 0.000 | 23:18 |            |

---

Data Download Date/Time: 05/11/2021 12:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LECLERC,A,  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: PEREZ,M  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 24119070A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/20/2021 04:03:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:55 |
| Air Blank          | 0.000 | 03:56 |
| Control Test       | 0.078 | 03:56 |
| Air Blank          | 0.000 | 03:57 |
| Subject Sample # 1 | 0.000 | 03:58 |
| Breath Volume      | 1.101 | 03:58 |
| Air Blank          | 0.000 | 03:58 |
| Air Blank          | 0.000 | 04:00 |
| Subject Sample # 2 | 0.000 | 04:02 |
| Breath Volume      | 2.093 | 04:02 |
| Air Blank          | 0.000 | 04:02 |
| Control Test       | 0.078 | 04:03 |
| Air Blank          | 0.000 | 04:03 |
| Diagnostics Check  | OK    | 04:03 |

---

Data Download Date/Time: 05/11/2021 12:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NATSON,ANDRAE,M  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: ANTONINO,C  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119070A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 12:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: ZOELLNER

DATE/TIME OF TEST: 04/21/2021 01:51:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:41 |
| Air Blank          | 0.000 | 01:41 |
| Control Test       | 0.078 | 01:41 |
| Air Blank          | 0.000 | 01:42 |
| Subject Sample # 1 | 0.171 | 01:44 |
| Breath Volume      | 2.515 | 01:44 |
| Air Blank          | 0.000 | 01:44 |
| Air Blank          | 0.000 | 01:46 |
| Subject Sample # 2 | 0.145 | 01:46 |
| Breath Volume      | 1.257 | 01:46 |
| Air Blank          | 0.000 | 01:47 |
| Air Blank          | 0.000 | 01:49 |
| Subject Sample # 3 | 0.147 | 01:49 |
| Breath Volume      | 1.445 | 01:49 |
| Air Blank          | 0.000 | 01:50 |
| Control Test       | 0.078 | 01:50 |
| Air Blank          | 0.000 | 01:51 |
| Diagnostics Check  | OK    | 01:51 |

---

Data Download Date/Time: 05/11/2021 12:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORRIS,THOMAS,L  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: SELLER,D  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 04/24/2021 03:39:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:31 |
| Air Blank          | 0.000 | 03:32 |
| Control Test       | 0.079 | 03:32 |
| Air Blank          | 0.000 | 03:33 |
| Subject Sample # 1 | 0.143 | 03:34 |
| Breath Volume      | 1.292 | 03:34 |
| Air Blank          | 0.000 | 03:35 |
| Air Blank          | 0.000 | 03:36 |
| Subject Sample # 2 | 0.145 | 03:38 |
| Breath Volume      | 1.425 | 03:38 |
| Air Blank          | 0.000 | 03:38 |
| Control Test       | 0.077 | 03:39 |
| Air Blank          | 0.000 | 03:39 |
| Diagnostics Check  | OK    | 03:39 |

---

Data Download Date/Time: 05/11/2021 12:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVAS,WALTHER,  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: MORRIS,H  
ARRESTING OFFICER AGENCY: FHP TROOP F  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: ASPRILLA MOSQUERA

DATE/TIME OF TEST: 04/25/2021 03:06:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:59 |
| Air Blank          | 0.000 | 03:00 |
| Control Test       | 0.079 | 03:00 |
| Air Blank          | 0.000 | 03:01 |
| Subject Sample # 1 | 0.193 | 03:01 |
| Breath Volume      | 1.804 | 03:01 |
| Air Blank          | 0.000 | 03:02 |
| Air Blank          | 0.000 | 03:04 |
| Subject Sample # 2 | 0.194 | 03:05 |
| Breath Volume      | 1.765 | 03:05 |
| Air Blank          | 0.000 | 03:06 |
| Control Test       | 0.076 | 03:06 |
| Air Blank          | 0.000 | 03:06 |
| Diagnostics Check  | OK    | 03:06 |

---

Data Download Date/Time: 05/11/2021 12:34



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EMANS,A,  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: EMANS,A  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119070A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: RINCORUIZ

DATE/TIME OF TEST: 04/25/2021 21:54:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:48 |
| Air Blank          | 0.000 | 21:48 |
| Control Test       | 0.078 | 21:48 |
| Air Blank          | 0.000 | 21:49 |
| Subject Sample # 1 | 0.202 | 21:50 |
| Breath Volume      | 1.195 | 21:50 |
| Air Blank          | 0.000 | 21:51 |
| Air Blank          | 0.000 | 21:52 |
| Subject Sample # 2 | 0.212 | 21:53 |
| Breath Volume      | 2.207 | 21:53 |
| Air Blank          | 0.000 | 21:53 |
| Control Test       | 0.077 | 21:54 |
| Air Blank          | 0.000 | 21:54 |
| Diagnostics Check  | OK    | 21:54 |

---

Data Download Date/Time: 05/11/2021 12:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLOPS,C,M  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: HERNANDEZ,K  
ARRESTING OFFICER AGENCY: FHP TROOP F  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 20:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 04/25/2021 23:11:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:04 |
| Air Blank          | 0.000 | 23:05 |
| Control Test       | 0.078 | 23:05 |
| Air Blank          | 0.000 | 23:06 |
| Subject Sample # 1 | 0.163 | 23:06 |
| Breath Volume      | 2.640 | 23:06 |
| Air Blank          | 0.000 | 23:07 |
| Air Blank          | 0.000 | 23:09 |
| Subject Sample # 2 | 0.151 | 23:09 |
| Breath Volume      | 2.218 | 23:09 |
| Air Blank          | 0.000 | 23:10 |
| Control Test       | 0.077 | 23:10 |
| Air Blank          | 0.000 | 23:11 |
| Diagnostics Check  | OK    | 23:11 |

---

Data Download Date/Time: 05/11/2021 12:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LECLERC,A,  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: SUERO-LANTIGUA,J  
ARRESTING OFFICER AGENCY: FHP TROOP F  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119070A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: SIEVERT

DATE/TIME OF TEST: 04/26/2021 00:32:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:25 |
| Air Blank          | 0.000 | 00:25 |
| Control Test       | 0.078 | 00:25 |
| Air Blank          | 0.000 | 00:26 |
| Subject Sample # 1 | 0.109 | 00:27 |
| Breath Volume      | 1.117 | 00:27 |
| Air Blank          | 0.000 | 00:27 |
| Air Blank          | 0.000 | 00:29 |
| Subject Sample # 2 | 0.124 | 00:30 |
| Breath Volume      | 1.363 | 00:30 |
| Air Blank          | 0.000 | 00:31 |
| Control Test       | 0.078 | 00:31 |
| Air Blank          | 0.000 | 00:31 |
| Diagnostics Check  | OK    | 00:32 |

---

Data Download Date/Time: 05/11/2021 12:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOMEZ,JERRY,  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: LANCIA,0  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119070A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 15:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 04/26/2021 16:02:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:56 |
| Air Blank          | 0.000 | 15:57 |
| Control Test       | 0.078 | 15:57 |
| Air Blank          | 0.000 | 15:58 |
| Subject Sample # 1 | 0.000 | 15:58 |
| Breath Volume      | 2.632 | 15:58 |
| Air Blank          | 0.000 | 15:59 |
| Air Blank          | 0.000 | 16:01 |
| Subject Sample # 2 | 0.000 | 16:01 |
| Breath Volume      | 3.082 | 16:01 |
| Air Blank          | 0.000 | 16:01 |
| Control Test       | 0.078 | 16:02 |
| Air Blank          | 0.000 | 16:02 |
| Diagnostics Check  | OK    | 16:02 |

---

Data Download Date/Time: 05/11/2021 12:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OLMSTEAD,J,  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: GROSSKLAS,D  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119070A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: DONALD

DATE/TIME OF TEST: 04/26/2021 23:23:57

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 23:11 |                |
| Air Blank         | 0.000     | 23:11 |                |
| Control Test      | 0.079     | 23:12 |                |
| Air Blank         | 0.000     | 23:12 |                |
| Subject Sample    | # 1 0.000 | 23:15 | Volume Not Met |
| Breath Volume     | 0.351     | 23:15 |                |
| Air Blank         | 0.000     | 23:16 |                |
| Air Blank         | 0.000     | 23:18 |                |
| Subject Sample    | # 2 0.000 | 23:19 |                |
| Breath Volume     | 1.386     | 23:19 |                |
| Air Blank         | 0.000     | 23:19 |                |
| Air Blank         | 0.000     | 23:21 |                |
| Subject Sample    | # 3 0.000 | 23:22 |                |
| Breath Volume     | 1.628     | 23:22 |                |
| Air Blank         | 0.000     | 23:23 |                |
| Control Test      | 0.077     | 23:23 |                |
| Air Blank         | 0.000     | 23:23 |                |
| Diagnostics Check | OK        | 23:23 |                |

---

Data Download Date/Time: 05/11/2021 12:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANNORAT,ADELY,  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: CAUDILL,A  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MIMS

DATE/TIME OF TEST: 04/28/2021 01:09:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:03 |
| Air Blank          | 0.000 | 01:03 |
| Control Test       | 0.078 | 01:03 |
| Air Blank          | 0.000 | 01:04 |
| Subject Sample # 1 | 0.109 | 01:04 |
| Breath Volume      | 2.503 | 01:04 |
| Air Blank          | 0.000 | 01:05 |
| Air Blank          | 0.000 | 01:07 |
| Subject Sample # 2 | 0.105 | 01:07 |
| Breath Volume      | 1.675 | 01:07 |
| Air Blank          | 0.000 | 01:08 |
| Control Test       | 0.077 | 01:08 |
| Air Blank          | 0.000 | 01:09 |
| Diagnostics Check  | OK    | 01:09 |

---

Data Download Date/Time: 05/11/2021 12:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAFFARAS,CELINA,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: JOSEPH,C  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: ROLLINS

DATE/TIME OF TEST: 04/01/2021 04:02:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:56 |
| Air Blank          | 0.000 | 03:56 |
| Control Test       | 0.080 | 03:57 |
| Air Blank          | 0.000 | 03:57 |
| Subject Sample # 1 | 0.138 | 03:58 |
| Breath Volume      | 3.042 | 03:58 |
| Air Blank          | 0.000 | 03:58 |
| Air Blank          | 0.000 | 04:00 |
| Subject Sample # 2 | 0.148 | 04:01 |
| Breath Volume      | 3.542 | 04:01 |
| Air Blank          | 0.000 | 04:01 |
| Control Test       | 0.079 | 04:02 |
| Air Blank          | 0.000 | 04:02 |
| Diagnostics Check  | OK    | 04:02 |

---

Data Download Date/Time: 04/29/2021 07:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RISI,ALEYA,D  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: GIRRARD,JOSHUA  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: FERDINAND

DATE/TIME OF TEST: 04/03/2021 22:15:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:09 |
| Air Blank          | 0.000 | 22:09 |
| Control Test       | 0.080 | 22:09 |
| Air Blank          | 0.000 | 22:10 |
| Subject Sample # 1 | 0.045 | 22:10 |
| Breath Volume      | 2.261 | 22:10 |
| Air Blank          | 0.000 | 22:11 |
| Air Blank          | 0.000 | 22:13 |
| Subject Sample # 2 | 0.044 | 22:14 |
| Breath Volume      | 1.816 | 22:14 |
| Air Blank          | 0.000 | 22:14 |
| Control Test       | 0.080 | 22:15 |
| Air Blank          | 0.000 | 22:15 |
| Diagnostics Check  | OK    | 22:15 |

---

Data Download Date/Time: 04/29/2021 07:53



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDROP,P,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: HARTMANSTORFER,DFC  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 04/07/2021 18:31:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:24 |
| Air Blank          | 0.000 | 18:24 |
| Control Test       | 0.079 | 18:24 |
| Air Blank          | 0.000 | 18:25 |
| Subject Sample # 1 | 0.000 | 18:26 |
| Breath Volume      | 3.945 | 18:26 |
| Air Blank          | 0.000 | 18:27 |
| Air Blank          | 0.000 | 18:29 |
| Subject Sample # 2 | 0.000 | 18:29 |
| Breath Volume      | 2.828 | 18:29 |
| Air Blank          | 0.000 | 18:30 |
| Control Test       | 0.079 | 18:30 |
| Air Blank          | 0.000 | 18:31 |
| Diagnostics Check  | OK    | 18:31 |

---

Data Download Date/Time: 04/29/2021 07:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDROP,P,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: REILLY,OFC  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 04/08/2021 02:04:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:57 |
| Air Blank          | 0.000 | 01:57 |
| Control Test       | 0.080 | 01:57 |
| Air Blank          | 0.000 | 01:58 |
| Subject Sample # 1 | 0.379 | 01:59 |
| Breath Volume      | 2.707 | 01:59 |
| Air Blank          | 0.000 | 02:00 |
| Air Blank          | 0.000 | 02:02 |
| Subject Sample # 2 | 0.369 | 02:02 |
| Breath Volume      | 2.546 | 02:02 |
| Air Blank          | 0.000 | 02:03 |
| Control Test       | 0.077 | 02:03 |
| Air Blank          | 0.000 | 02:04 |
| Diagnostics Check  | OK    | 02:04 |

---

Data Download Date/Time: 04/29/2021 07:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDROP,P,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: GORMAN,OFC  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:59

SUBJECT'S LAST NAME: WEGER

DATE/TIME OF TEST: 04/08/2021 19:36:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:29 |
| Air Blank          | 0.000 | 19:29 |
| Control Test       | 0.079 | 19:30 |
| Air Blank          | 0.000 | 19:30 |
| Subject Sample # 1 | 0.111 | 19:31 |
| Breath Volume      | 2.953 | 19:31 |
| Air Blank          | 0.000 | 19:32 |
| Air Blank          | 0.000 | 19:33 |
| Subject Sample # 2 | 0.108 | 19:34 |
| Breath Volume      | 2.953 | 19:34 |
| Air Blank          | 0.000 | 19:35 |
| Control Test       | 0.078 | 19:35 |
| Air Blank          | 0.000 | 19:36 |
| Diagnostics Check  | OK    | 19:36 |

---

Data Download Date/Time: 04/29/2021 07:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANCILLA,RICHARD,P  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: JOSEPH,CARL  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ABELE JR.

DATE/TIME OF TEST: 04/09/2021 23:02:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:53 |
| Air Blank          | 0.000 | 22:54 |
| Control Test       | 0.079 | 22:54 |
| Air Blank          | 0.000 | 22:54 |
| Subject Sample # 1 | 0.224 | 22:55 |
| Breath Volume      | 2.914 | 22:55 |
| Air Blank          | 0.000 | 22:55 |
| Air Blank          | 0.000 | 22:57 |
| Subject Sample # 2 | 0.221 | 23:00 |
| Breath Volume      | 3.464 | 23:00 |
| Air Blank          | 0.000 | 23:01 |
| Control Test       | 0.080 | 23:01 |
| Air Blank          | 0.000 | 23:02 |
| Diagnostics Check  | OK    | 23:02 |

---

Data Download Date/Time: 04/29/2021 07:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAFFARAS,CELINA,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: MOTZ,D  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: ROUTON

DATE/TIME OF TEST: 04/10/2021 01:10:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:03 |
| Air Blank          | 0.000 | 01:03 |
| Control Test       | 0.079 | 01:04 |
| Air Blank          | 0.000 | 01:04 |
| Subject Sample # 1 | 0.111 | 01:05 |
| Breath Volume      | 1.843 | 01:05 |
| Air Blank          | 0.000 | 01:06 |
| Air Blank          | 0.000 | 01:08 |
| Subject Sample # 2 | 0.112 | 01:09 |
| Breath Volume      | 2.738 | 01:09 |
| Air Blank          | 0.000 | 01:09 |
| Control Test       | 0.077 | 01:10 |
| Air Blank          | 0.000 | 01:10 |
| Diagnostics Check  | OK    | 01:10 |

---

Data Download Date/Time: 04/29/2021 07:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANCILLA,RICHARD,P  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: JANKOWSKI,JOSEPH  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: GUIDO

DATE/TIME OF TEST: 04/10/2021 01:34:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:28 |
| Air Blank          | 0.000 | 01:28 |
| Control Test       | 0.078 | 01:29 |
| Air Blank          | 0.000 | 01:29 |
| Subject Sample # 1 | 0.000 | 01:30 |
| Breath Volume      | 1.730 | 01:30 |
| Air Blank          | 0.000 | 01:30 |
| Air Blank          | 0.000 | 01:32 |
| Subject Sample # 2 | 0.000 | 01:32 |
| Breath Volume      | 2.792 | 01:32 |
| Air Blank          | 0.000 | 01:33 |
| Control Test       | 0.079 | 01:33 |
| Air Blank          | 0.000 | 01:34 |
| Diagnostics Check  | OK    | 01:34 |

---

Data Download Date/Time: 04/29/2021 07:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAWSON,BRIAN,E  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: MEDDAUGH,BRANDON  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: KIPFER

DATE/TIME OF TEST: 04/13/2021 04:44:52

---

|                   |       |       |                 |
|-------------------|-------|-------|-----------------|
| Diagnostics Check | OK    | 04:44 |                 |
| Air Blank         |       | 04:44 | Improper Sample |
| Air Blank         | 0.000 | 04:44 |                 |

---

Data Download Date/Time: 04/29/2021 07:53

# **Florida Department of Law Enforcement**

## **Alcohol Testing Program**

### **Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAWSON,BRIAN,E  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: MEDDAUGH,BRANDON  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: KIPFER

DATE/TIME OF TEST: 04/13/2021 04:47:48

---

|                   |    |       |                 |
|-------------------|----|-------|-----------------|
| Diagnostics Check | OK | 04:47 |                 |
| Air Blank         |    | 04:47 | Improper Sample |
| Air Blank         |    | 04:47 | Improper Sample |

---

Data Download Date/Time: 04/29/2021 07:53



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RONGA,DEPUTY,R  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: ANDERSON,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 14:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:44

SUBJECT'S LAST NAME: MYSLINSKI

DATE/TIME OF TEST: 04/13/2021 16:15:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:08 |
| Air Blank          | 0.000 | 16:09 |
| Control Test       | 0.080 | 16:09 |
| Air Blank          | 0.000 | 16:09 |
| Subject Sample # 1 | 0.063 | 16:10 |
| Breath Volume      | 1.671 | 16:10 |
| Air Blank          | 0.000 | 16:11 |
| Air Blank          | 0.000 | 16:13 |
| Subject Sample # 2 | 0.061 | 16:14 |
| Breath Volume      | 2.285 | 16:14 |
| Air Blank          | 0.000 | 16:14 |
| Control Test       | 0.079 | 16:15 |
| Air Blank          | 0.000 | 16:15 |
| Diagnostics Check  | OK    | 16:15 |

---

Data Download Date/Time: 04/29/2021 07:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDROP,P,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: WALDROP,P  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 17:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 04/13/2021 19:55:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:49 |
| Air Blank          | 0.000 | 19:50 |
| Control Test       | 0.080 | 19:50 |
| Air Blank          | 0.000 | 19:51 |
| Subject Sample # 1 | 0.000 | 19:51 |
| Breath Volume      | 3.929 | 19:51 |
| Air Blank          | 0.000 | 19:52 |
| Air Blank          | 0.000 | 19:54 |
| Subject Sample # 2 | 0.000 | 19:54 |
| Breath Volume      | 4.281 | 19:54 |
| Air Blank          | 0.000 | 19:54 |
| Control Test       | 0.080 | 19:55 |
| Air Blank          | 0.000 | 19:55 |
| Diagnostics Check  | OK    | 19:55 |

---

Data Download Date/Time: 04/29/2021 07:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDROP,P,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: ERICKSON,DFC  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: GRIFFY-SHATTUCK

DATE/TIME OF TEST: 04/14/2021 04:26:30

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 04:14 |                |
| Air Blank         | 0.000     | 04:15 |                |
| Control Test      | 0.080     | 04:15 |                |
| Air Blank         | 0.000     | 04:16 |                |
| Subject Sample    | # 1 0.159 | 04:19 | Volume Not Met |
| Breath Volume     | 0.253     | 04:19 |                |
| Air Blank         | 0.000     | 04:19 |                |
| Air Blank         | 0.000     | 04:21 |                |
| Subject Sample    | # 2 0.152 | 04:24 | Volume Not Met |
| Breath Volume     | 0.550     | 04:24 |                |
| Air Blank         | 0.000     | 04:25 |                |
| Control Test      | 0.078     | 04:25 |                |
| Air Blank         | 0.000     | 04:26 |                |
| Diagnostics Check | OK        | 04:26 |                |

---

Data Download Date/Time: 04/29/2021 07:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EGAN,MELISSA,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: EGAN,M  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 14:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: FORREST

DATE/TIME OF TEST: 04/14/2021 14:34:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:28 |
| Air Blank          | 0.000 | 14:29 |
| Control Test       | 0.079 | 14:29 |
| Air Blank          | 0.000 | 14:29 |
| Subject Sample # 1 | 0.000 | 14:30 |
| Breath Volume      | 1.636 | 14:30 |
| Air Blank          | 0.000 | 14:30 |
| Air Blank          | 0.000 | 14:32 |
| Subject Sample # 2 | 0.000 | 14:33 |
| Breath Volume      | 1.289 | 14:33 |
| Air Blank          | 0.000 | 14:33 |
| Control Test       | 0.080 | 14:34 |
| Air Blank          | 0.000 | 14:34 |
| Diagnostics Check  | OK    | 14:34 |

---

Data Download Date/Time: 04/29/2021 07:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHWOCHO,ALAN'  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: PERO,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 00:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 04/15/2021 01:01:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:55 |
| Air Blank          | 0.000 | 00:55 |
| Control Test       | 0.080 | 00:56 |
| Air Blank          | 0.000 | 00:56 |
| Subject Sample # 1 | 0.000 | 00:57 |
| Breath Volume      | 3.933 | 00:57 |
| Air Blank          | 0.000 | 00:57 |
| Air Blank          | 0.000 | 00:59 |
| Subject Sample # 2 | 0.000 | 00:59 |
| Breath Volume      | 3.964 | 00:59 |
| Air Blank          | 0.000 | 01:00 |
| Control Test       | 0.079 | 01:00 |
| Air Blank          | 0.000 | 01:01 |
| Diagnostics Check  | OK    | 01:01 |

---

Data Download Date/Time: 04/29/2021 07:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COURTNEY,PATRICK,R  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: PERO,J  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 04/15/2021 23:08:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:01 |
| Air Blank          | 0.000 | 23:01 |
| Control Test       | 0.080 | 23:02 |
| Air Blank          | 0.000 | 23:02 |
| Subject Sample # 1 | 0.157 | 23:03 |
| Breath Volume      | 1.910 | 23:03 |
| Air Blank          | 0.000 | 23:04 |
| Air Blank          | 0.000 | 23:06 |
| Subject Sample # 2 | 0.163 | 23:07 |
| Breath Volume      | 2.234 | 23:07 |
| Air Blank          | 0.000 | 23:08 |
| Control Test       | 0.079 | 23:08 |
| Air Blank          | 0.000 | 23:08 |
| Diagnostics Check  | OK    | 23:08 |

---

Data Download Date/Time: 04/29/2021 07:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANCILLA,RICHARD,P  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: JANKOWSKI,JOSEPH  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: CONLEY

DATE/TIME OF TEST: 04/15/2021 23:37:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:31 |
| Air Blank          | 0.000 | 23:32 |
| Control Test       | 0.080 | 23:32 |
| Air Blank          | 0.000 | 23:32 |
| Subject Sample # 1 | 0.156 | 23:33 |
| Breath Volume      | 2.195 | 23:33 |
| Air Blank          | 0.000 | 23:33 |
| Air Blank          | 0.000 | 23:35 |
| Subject Sample # 2 | 0.143 | 23:36 |
| Breath Volume      | 1.273 | 23:36 |
| Air Blank          | 0.000 | 23:36 |
| Control Test       | 0.079 | 23:37 |
| Air Blank          | 0.000 | 23:37 |
| Diagnostics Check  | OK    | 23:37 |

---

Data Download Date/Time: 04/29/2021 07:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANCILLA,RICHARD,P  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: TUCK,BILLY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: GADWAY

DATE/TIME OF TEST: 04/16/2021 03:21:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:11 |
| Air Blank          | 0.000 | 03:12 |
| Control Test       | 0.080 | 03:12 |
| Air Blank          | 0.000 | 03:13 |
| Subject Sample # 1 | 0.342 | 03:13 |
| Breath Volume      | 1.988 | 03:13 |
| Air Blank          | 0.000 | 03:14 |
| Air Blank          | 0.000 | 03:16 |
| Subject Sample # 2 | 0.315 | 03:16 |
| Breath Volume      | 2.007 | 03:16 |
| Air Blank          | 0.000 | 03:17 |
| Air Blank          | 0.000 | 03:19 |
| Subject Sample # 3 | 0.339 | 03:19 |
| Breath Volume      | 1.992 | 03:19 |
| Air Blank          | 0.000 | 03:20 |
| Control Test       | 0.078 | 03:20 |
| Air Blank          | 0.000 | 03:21 |
| Diagnostics Check  | OK    | 03:21 |

---

Data Download Date/Time: 04/29/2021 07:54



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COURTNEY,PATRICK,R  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: BRASCH,D  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BERNARD

DATE/TIME OF TEST: 04/16/2021 04:08:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:01 |
| Air Blank          | 0.000 | 04:02 |
| Control Test       | 0.080 | 04:02 |
| Air Blank          | 0.000 | 04:03 |
| Subject Sample # 1 | 0.000 | 04:04 |
| Breath Volume      | 1.843 | 04:04 |
| Air Blank          | 0.000 | 04:04 |
| Air Blank          | 0.000 | 04:06 |
| Subject Sample # 2 | 0.000 | 04:07 |
| Breath Volume      | 4.734 | 04:07 |
| Air Blank          | 0.000 | 04:07 |
| Control Test       | 0.079 | 04:07 |
| Air Blank          | 0.000 | 04:08 |
| Diagnostics Check  | OK    | 04:08 |

---

Data Download Date/Time: 04/29/2021 07:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DZUGAN,MICHAEL,A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: DZUGAN,MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 11:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 04/16/2021 06:33:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:26 |
| Air Blank          | 0.000 | 06:27 |
| Control Test       | 0.080 | 06:27 |
| Air Blank          | 0.000 | 06:28 |
| Subject Sample # 1 | 0.084 | 06:29 |
| Breath Volume      | 4.246 | 06:29 |
| Air Blank          | 0.000 | 06:29 |
| Air Blank          | 0.000 | 06:31 |
| Subject Sample # 2 | 0.071 | 06:32 |
| Breath Volume      | 1.734 | 06:32 |
| Air Blank          | 0.000 | 06:32 |
| Control Test       | 0.078 | 06:33 |
| Air Blank          | 0.000 | 06:33 |
| Diagnostics Check  | OK    | 06:33 |

---

Data Download Date/Time: 04/29/2021 07:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRITES,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: CRITES,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 21:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 04/16/2021 22:34:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:27 |
| Air Blank          | 0.000 | 22:28 |
| Control Test       | 0.079 | 22:28 |
| Air Blank          | 0.000 | 22:29 |
| Subject Sample # 1 | 0.116 | 22:30 |
| Breath Volume      | 1.980 | 22:30 |
| Air Blank          | 0.000 | 22:31 |
| Air Blank          | 0.000 | 22:32 |
| Subject Sample # 2 | 0.118 | 22:33 |
| Breath Volume      | 2.132 | 22:33 |
| Air Blank          | 0.000 | 22:34 |
| Control Test       | 0.078 | 22:34 |
| Air Blank          | 0.000 | 22:34 |
| Diagnostics Check  | OK    | 22:34 |

---

Data Download Date/Time: 04/29/2021 07:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUNN,CDFC,C  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: DUNN,D  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 06:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 04/17/2021 06:33:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:26 |
| Air Blank          | 0.000 | 06:26 |
| Control Test       | 0.079 | 06:27 |
| Air Blank          | 0.000 | 06:27 |
| Subject Sample # 1 | 0.000 | 06:28 |
| Breath Volume      | 1.910 | 06:28 |
| Air Blank          | 0.000 | 06:29 |
| Air Blank          | 0.000 | 06:31 |
| Subject Sample # 2 | 0.000 | 06:32 |
| Breath Volume      | 1.929 | 06:32 |
| Air Blank          | 0.000 | 06:32 |
| Control Test       | 0.080 | 06:32 |
| Air Blank          | 0.000 | 06:33 |
| Diagnostics Check  | OK    | 06:33 |

---

Data Download Date/Time: 04/29/2021 07:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDROP,P,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: CUADRA,DFC  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MARCHMAN1

DATE/TIME OF TEST: 04/18/2021 01:12:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:03 |
| Air Blank          | 0.000 | 01:04 |
| Control Test       | 0.079 | 01:04 |
| Air Blank          | 0.000 | 01:04 |
| Subject Sample # 1 | 0.304 | 01:07 |
| Breath Volume      | 1.550 | 01:07 |
| Air Blank          | 0.000 | 01:08 |
| Air Blank          | 0.000 | 01:09 |
| Subject Sample # 2 | 0.296 | 01:11 |
| Breath Volume      | 3.363 | 01:11 |
| Air Blank          | 0.000 | 01:12 |
| Control Test       | 0.079 | 01:12 |
| Air Blank          | 0.000 | 01:12 |
| Diagnostics Check  | OK    | 01:12 |

---

Data Download Date/Time: 04/29/2021 07:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUNN,D,C  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: DUNN,D  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 09:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 04/18/2021 10:21:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:14 |
| Air Blank          | 0.000 | 10:15 |
| Control Test       | 0.079 | 10:15 |
| Air Blank          | 0.000 | 10:16 |
| Subject Sample # 1 | 0.087 | 10:17 |
| Breath Volume      | 1.542 | 10:17 |
| Air Blank          | 0.000 | 10:17 |
| Air Blank          | 0.000 | 10:19 |
| Subject Sample # 2 | 0.099 | 10:20 |
| Breath Volume      | 1.656 | 10:20 |
| Air Blank          | 0.000 | 10:20 |
| Control Test       | 0.079 | 10:21 |
| Air Blank          | 0.000 | 10:21 |
| Diagnostics Check  | OK    | 10:21 |

---

Data Download Date/Time: 04/29/2021 07:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SERRANO, MONICA, S  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: UNKNOWN, CCSO  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: REARDON

DATE/TIME OF TEST: 04/18/2021 14:17:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:11 |
| Air Blank          | 0.000 | 14:11 |
| Control Test       | 0.079 | 14:11 |
| Air Blank          | 0.000 | 14:12 |
| Subject Sample # 1 | 0.016 | 14:12 |
| Breath Volume      | 2.113 | 14:12 |
| Air Blank          | 0.000 | 14:13 |
| Air Blank          | 0.000 | 14:15 |
| Subject Sample # 2 | 0.015 | 14:15 |
| Breath Volume      | 2.191 | 14:15 |
| Air Blank          | 0.000 | 14:16 |
| Control Test       | 0.079 | 14:16 |
| Air Blank          | 0.000 | 14:17 |
| Diagnostics Check  | OK    | 14:17 |

---

Data Download Date/Time: 04/29/2021 07:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDROP,P,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: OWENS,DFC  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 21:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 04/18/2021 22:04:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:57 |
| Air Blank          | 0.000 | 21:57 |
| Control Test       | 0.080 | 21:57 |
| Air Blank          | 0.000 | 21:58 |
| Subject Sample # 1 | 0.000 | 21:59 |
| Breath Volume      | 3.023 | 21:59 |
| Air Blank          | 0.000 | 22:00 |
| Air Blank          | 0.000 | 22:02 |
| Subject Sample # 2 | 0.000 | 22:03 |
| Breath Volume      | 2.703 | 22:03 |
| Air Blank          | 0.000 | 22:03 |
| Control Test       | 0.080 | 22:03 |
| Air Blank          | 0.000 | 22:04 |
| Diagnostics Check  | OK    | 22:04 |

---

Data Download Date/Time: 04/29/2021 07:55



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EGAN,MELISSA,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: MATZICAFF,A  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 08:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: BOUTALL HARPER

DATE/TIME OF TEST: 04/19/2021 09:04:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:58 |
| Air Blank          | 0.000 | 08:59 |
| Control Test       | 0.080 | 08:59 |
| Air Blank          | 0.000 | 08:59 |
| Subject Sample # 1 | 0.000 | 09:00 |
| Breath Volume      | 1.523 | 09:00 |
| Air Blank          | 0.000 | 09:00 |
| Air Blank          | 0.000 | 09:02 |
| Subject Sample # 2 | 0.000 | 09:03 |
| Breath Volume      | 1.613 | 09:03 |
| Air Blank          | 0.000 | 09:03 |
| Control Test       | 0.079 | 09:04 |
| Air Blank          | 0.000 | 09:04 |
| Diagnostics Check  | OK    | 09:04 |

---

Data Download Date/Time: 04/29/2021 07:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RISI,ALEYA,D  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: MEDDAUGH,BRANDON  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: BRODEUR III

DATE/TIME OF TEST: 04/21/2021 23:17:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:10 |
| Air Blank          | 0.000 | 23:10 |
| Control Test       | 0.079 | 23:11 |
| Air Blank          | 0.000 | 23:11 |
| Subject Sample # 1 | 0.124 | 23:12 |
| Breath Volume      | 4.191 | 23:12 |
| Air Blank          | 0.000 | 23:13 |
| Air Blank          | 0.000 | 23:14 |
| Subject Sample # 2 | 0.112 | 23:15 |
| Breath Volume      | 2.617 | 23:15 |
| Air Blank          | 0.000 | 23:16 |
| Control Test       | 0.079 | 23:16 |
| Air Blank          | 0.000 | 23:17 |
| Diagnostics Check  | OK    | 23:17 |

---

Data Download Date/Time: 04/29/2021 07:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDROP,P,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: ERICKSON,CHRISTINA  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: AUMAN

DATE/TIME OF TEST: 04/22/2021 01:42:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:35 |
| Air Blank          | 0.000 | 01:36 |
| Control Test       | 0.079 | 01:36 |
| Air Blank          | 0.000 | 01:37 |
| Subject Sample # 1 | 0.156 | 01:38 |
| Breath Volume      | 2.492 | 01:38 |
| Air Blank          | 0.000 | 01:38 |
| Air Blank          | 0.000 | 01:40 |
| Subject Sample # 2 | 0.159 | 01:41 |
| Breath Volume      | 2.800 | 01:41 |
| Air Blank          | 0.000 | 01:42 |
| Control Test       | 0.079 | 01:42 |
| Air Blank          | 0.000 | 01:42 |
| Diagnostics Check  | OK    | 01:42 |

---

Data Download Date/Time: 04/29/2021 07:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAWSON,BRIAN,E  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: DAVIDSON,MIKE  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: VANRAVENSWAAY

DATE/TIME OF TEST: 04/23/2021 00:32:24

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 00:16 |                |
| Air Blank         | 0.000     | 00:17 |                |
| Control Test      | 0.079     | 00:17 |                |
| Air Blank         | 0.000     | 00:18 |                |
| Subject Sample    | # 1 0.155 | 00:19 |                |
| Breath Volume     | 1.777     | 00:19 |                |
| Air Blank         | 0.000     | 00:20 |                |
| Air Blank         | 0.000     | 00:21 |                |
| Subject Sample    | # 2 0.156 | 00:24 | Volume Not Met |
| Breath Volume     | 0.082     | 00:24 |                |
| Air Blank         | 0.000     | 00:25 |                |
| Air Blank         | 0.000     | 00:27 |                |
| Subject Sample    | # 3 0.142 | 00:30 | Volume Not Met |
| Breath Volume     | 0.558     | 00:30 |                |
| Air Blank         | 0.000     | 00:31 |                |
| Control Test      | 0.077     | 00:31 |                |
| Air Blank         | 0.000     | 00:32 |                |
| Diagnostics Check | OK        | 00:32 |                |

---

Data Download Date/Time: 04/29/2021 07:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAWSON,BRIAN,E  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: DAVIDSON,MIKE  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: VANRAVENSWAAY

DATE/TIME OF TEST: 04/23/2021 00:43:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:34 |
| Air Blank          | 0.000 | 00:35 |
| Control Test       | 0.077 | 00:35 |
| Air Blank          | 0.000 | 00:35 |
| Subject Sample # 1 | 0.135 | 00:38 |
| Breath Volume      | 1.941 | 00:38 |
| Air Blank          | 0.000 | 00:39 |
| Air Blank          | 0.000 | 00:40 |
| Subject Sample # 2 | 0.145 | 00:41 |
| Breath Volume      | 1.773 | 00:41 |
| Air Blank          | 0.000 | 00:42 |
| Control Test       | 0.077 | 00:42 |
| Air Blank          | 0.000 | 00:43 |
| Diagnostics Check  | OK    | 00:43 |

---

Data Download Date/Time: 04/29/2021 07:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RISI,ALEYA,D  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: SCHWARTZ,JONATHAN  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: PETROCK

DATE/TIME OF TEST: 04/23/2021 01:37:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:30 |
| Air Blank          | 0.000 | 01:31 |
| Control Test       | 0.079 | 01:31 |
| Air Blank          | 0.000 | 01:32 |
| Subject Sample # 1 | 0.220 | 01:32 |
| Breath Volume      | 2.960 | 01:32 |
| Air Blank          | 0.000 | 01:33 |
| Air Blank          | 0.000 | 01:35 |
| Subject Sample # 2 | 0.222 | 01:36 |
| Breath Volume      | 2.953 | 01:36 |
| Air Blank          | 0.000 | 01:36 |
| Control Test       | 0.078 | 01:37 |
| Air Blank          | 0.000 | 01:37 |
| Diagnostics Check  | OK    | 01:37 |

---

Data Download Date/Time: 04/29/2021 07:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RISI,ALEYA,D  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: NONE,NONE  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 00:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 04/23/2021 05:17:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:11 |
| Air Blank          | 0.000 | 05:11 |
| Control Test       | 0.079 | 05:12 |
| Air Blank          | 0.000 | 05:12 |
| Subject Sample # 1 | 0.139 | 05:13 |
| Breath Volume      | 3.296 | 05:13 |
| Air Blank          | 0.000 | 05:13 |
| Air Blank          | 0.000 | 05:15 |
| Subject Sample # 2 | 0.127 | 05:16 |
| Breath Volume      | 2.871 | 05:16 |
| Air Blank          | 0.000 | 05:16 |
| Control Test       | 0.078 | 05:17 |
| Air Blank          | 0.000 | 05:17 |
| Diagnostics Check  | OK    | 05:17 |

---

Data Download Date/Time: 04/29/2021 07:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANO,MARDEE,G  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: NONE,NONE  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 05:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: TIMMONS

DATE/TIME OF TEST: 04/23/2021 08:51:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:44 |
| Air Blank          | 0.000 | 08:45 |
| Control Test       | 0.080 | 08:45 |
| Air Blank          | 0.000 | 08:46 |
| Subject Sample # 1 | 0.067 | 08:46 |
| Breath Volume      | 2.605 | 08:46 |
| Air Blank          | 0.000 | 08:47 |
| Air Blank          | 0.000 | 08:49 |
| Subject Sample # 2 | 0.060 | 08:49 |
| Breath Volume      | 2.039 | 08:49 |
| Air Blank          | 0.000 | 08:50 |
| Control Test       | 0.077 | 08:50 |
| Air Blank          | 0.000 | 08:50 |
| Diagnostics Check  | OK    | 08:51 |

---

Data Download Date/Time: 04/29/2021 07:56



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANO,MARDEE,G  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: NONE,NONE  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 04:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: TIMMONS

DATE/TIME OF TEST: 04/23/2021 12:29:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 12:23 |
| Air Blank          | 0.000 | 12:23 |
| Control Test       | 0.080 | 12:23 |
| Air Blank          | 0.000 | 12:24 |
| Subject Sample # 1 | 0.000 | 12:24 |
| Breath Volume      | 2.242 | 12:24 |
| Air Blank          | 0.000 | 12:25 |
| Air Blank          | 0.000 | 12:27 |
| Subject Sample # 2 | 0.000 | 12:27 |
| Breath Volume      | 2.929 | 12:27 |
| Air Blank          | 0.000 | 12:28 |
| Control Test       | 0.079 | 12:28 |
| Air Blank          | 0.000 | 12:29 |
| Diagnostics Check  | OK    | 12:29 |

---

Data Download Date/Time: 04/29/2021 07:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANCILLA,RICHARD,P  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: PERO,JOSIAH  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 04/23/2021 23:31:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:23 |
| Air Blank          | 0.000 | 23:23 |
| Control Test       | 0.080 | 23:24 |
| Air Blank          | 0.000 | 23:24 |
| Subject Sample # 1 | 0.076 | 23:25 |
| Breath Volume      | 1.484 | 23:25 |
| Air Blank          | 0.000 | 23:25 |
| Air Blank          | 0.000 | 23:27 |
| Subject Sample # 2 | 0.080 | 23:29 |
| Breath Volume      | 1.238 | 23:29 |
| Air Blank          | 0.000 | 23:30 |
| Control Test       | 0.080 | 23:30 |
| Air Blank          | 0.000 | 23:31 |
| Diagnostics Check  | OK    | 23:31 |

---

Data Download Date/Time: 04/29/2021 07:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHWOCHO,ALAN,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: ANDERSON,J  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: EADDY

DATE/TIME OF TEST: 04/24/2021 19:15:49

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 19:04 |                    |
| Air Blank         | 0.000 | 19:04 |                    |
| Control Test      | 0.080 | 19:05 |                    |
| Air Blank         | 0.000 | 19:05 |                    |
| Subject Sample    | # 1   | 19:08 | No Sample Provided |
| Breath Volume     | 0.000 | 19:08 |                    |
| Air Blank         | 0.000 | 19:09 |                    |
| Air Blank         | 0.000 | 19:11 |                    |
| Subject Sample    | # 2   | 19:14 | No Sample Provided |
| Breath Volume     | 0.000 | 19:14 |                    |
| Air Blank         | 0.000 | 19:14 |                    |
| Control Test      | 0.080 | 19:15 |                    |
| Air Blank         | 0.000 | 19:15 |                    |
| Diagnostics Check | OK    | 19:15 |                    |

---

Data Download Date/Time: 04/29/2021 07:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDROP,P,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: CUADRA,DFC  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: RONICK

DATE/TIME OF TEST: 04/26/2021 21:50:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:41 |
| Air Blank          | 0.000 | 21:42 |
| Control Test       | 0.080 | 21:42 |
| Air Blank          | 0.000 | 21:42 |
| Subject Sample # 1 | 0.147 | 21:45 |
| Breath Volume      | 1.292 | 21:45 |
| Air Blank          | 0.000 | 21:45 |
| Air Blank          | 0.000 | 21:47 |
| Subject Sample # 2 | 0.137 | 21:48 |
| Breath Volume      | 2.171 | 21:48 |
| Air Blank          | 0.000 | 21:49 |
| Control Test       | 0.078 | 21:49 |
| Air Blank          | 0.000 | 21:50 |
| Diagnostics Check  | OK    | 21:50 |

---

Data Download Date/Time: 04/29/2021 07:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRITES,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: POORMON,COURTNEY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: GOODRIDGE

DATE/TIME OF TEST: 04/28/2021 03:03:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:56 |
| Air Blank          | 0.000 | 02:57 |
| Control Test       | 0.079 | 02:57 |
| Air Blank          | 0.000 | 02:58 |
| Subject Sample # 1 | 0.092 | 02:58 |
| Breath Volume      | 2.152 | 02:58 |
| Air Blank          | 0.000 | 02:59 |
| Air Blank          | 0.000 | 03:01 |
| Subject Sample # 2 | 0.093 | 03:02 |
| Breath Volume      | 3.789 | 03:02 |
| Air Blank          | 0.000 | 03:02 |
| Control Test       | 0.079 | 03:03 |
| Air Blank          | 0.000 | 03:03 |
| Diagnostics Check  | OK    | 03:03 |

---

Data Download Date/Time: 04/29/2021 07:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANCILLA,RICHARD,P  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: WELCH,KERI  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: WAGENER

DATE/TIME OF TEST: 04/29/2021 22:57:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:50 |
| Air Blank          | 0.000 | 22:51 |
| Control Test       | 0.079 | 22:51 |
| Air Blank          | 0.000 | 22:51 |
| Subject Sample # 1 | 0.174 | 22:52 |
| Breath Volume      | 1.253 | 22:52 |
| Air Blank          | 0.000 | 22:53 |
| Air Blank          | 0.000 | 22:55 |
| Subject Sample # 2 | 0.165 | 22:56 |
| Breath Volume      | 1.375 | 22:56 |
| Air Blank          | 0.000 | 22:56 |
| Control Test       | 0.079 | 22:57 |
| Air Blank          | 0.000 | 22:57 |
| Diagnostics Check  | OK    | 22:57 |

---

Data Download Date/Time: 05/22/2021 23:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EGAN,MELISSA,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: BALABAN,P  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 14:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: GARCIA CANTU

DATE/TIME OF TEST: 04/30/2021 15:57:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:50 |
| Air Blank          | 0.000 | 15:50 |
| Control Test       | 0.079 | 15:51 |
| Air Blank          | 0.000 | 15:51 |
| Subject Sample # 1 | 0.263 | 15:52 |
| Breath Volume      | 3.003 | 15:52 |
| Air Blank          | 0.000 | 15:53 |
| Air Blank          | 0.000 | 15:55 |
| Subject Sample # 2 | 0.264 | 15:55 |
| Breath Volume      | 3.121 | 15:55 |
| Air Blank          | 0.000 | 15:56 |
| Control Test       | 0.079 | 15:56 |
| Air Blank          | 0.000 | 15:57 |
| Diagnostics Check  | OK    | 15:57 |

---

Data Download Date/Time: 05/22/2021 23:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDROP,P,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: HARTMANSTORFER,DFC  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 04/30/2021 18:18:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:11 |
| Air Blank          | 0.000 | 18:12 |
| Control Test       | 0.079 | 18:12 |
| Air Blank          | 0.000 | 18:13 |
| Subject Sample # 1 | 0.000 | 18:14 |
| Breath Volume      | 1.457 | 18:14 |
| Air Blank          | 0.000 | 18:14 |
| Air Blank          | 0.000 | 18:16 |
| Subject Sample # 2 | 0.000 | 18:17 |
| Breath Volume      | 2.082 | 18:17 |
| Air Blank          | 0.000 | 18:17 |
| Control Test       | 0.079 | 18:18 |
| Air Blank          | 0.000 | 18:18 |
| Diagnostics Check  | OK    | 18:18 |

---

Data Download Date/Time: 05/22/2021 23:52



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OCEGUERA,YASMANI,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: OCEGUERA,YASMANI  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 03/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: HERNANDEZVALLADARES

DATE/TIME OF TEST: 04/08/2021 01:13:36

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:08 |                      |
| Air Blank         | 0.000 | 01:09 |                      |
| Control Test      | 0.080 | 01:09 |                      |
| Air Blank         | 0.000 | 01:09 |                      |
| Subject Sample    | # 1   | 01:12 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:12 |                      |
| Air Blank         | 0.000 | 01:12 |                      |
| Control Test      | 0.081 | 01:12 |                      |
| Air Blank         | 0.000 | 01:13 |                      |
| Diagnostics Check | OK    | 01:13 |                      |

---

Data Download Date/Time: 04/26/2021 16:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OCEGUERA,YASMANI,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: HERES,CARLOS  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 03/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 04/18/2021 21:43:33

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:35 |                      |
| Air Blank         | 0.000 | 21:35 |                      |
| Control Test      | 0.079 | 21:35 |                      |
| Air Blank         | 0.000 | 21:36 |                      |
| Subject Sample    | # 1   | 21:39 | No Sample Provided   |
| Breath Volume     | 0.000 | 21:39 |                      |
| Air Blank         | 0.000 | 21:39 |                      |
| Air Blank         | 0.000 | 21:41 |                      |
| Subject Sample    | # 2   | 21:42 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:42 |                      |
| Air Blank         | 0.000 | 21:42 |                      |
| Control Test      | 0.080 | 21:42 |                      |
| Air Blank         | 0.000 | 21:43 |                      |
| Diagnostics Check | OK    | 21:43 |                      |

---

Data Download Date/Time: 04/26/2021 16:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POPICK,SCOTT,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: POPICK,SCOTT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: W436544744631

DATE/TIME OF TEST: 04/10/2021 05:06:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:59 |
| Air Blank          | 0.000 | 05:00 |
| Control Test       | 0.079 | 05:00 |
| Air Blank          | 0.000 | 05:01 |
| Subject Sample # 1 | 0.000 | 05:02 |
| Breath Volume      | 1.664 | 05:02 |
| Air Blank          | 0.000 | 05:02 |
| Air Blank          | 0.000 | 05:04 |
| Subject Sample # 2 | 0.000 | 05:05 |
| Breath Volume      | 1.281 | 05:05 |
| Air Blank          | 0.000 | 05:05 |
| Control Test       | 0.079 | 05:06 |
| Air Blank          | 0.000 | 05:06 |
| Diagnostics Check  | OK    | 05:06 |

---

Data Download Date/Time: 05/12/2021 11:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POPICK,SCOTT,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: POPICK,SCOTT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: PACHECO PADILLA

DATE/TIME OF TEST: 04/16/2021 01:36:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:30 |
| Air Blank          | 0.000 | 01:30 |
| Control Test       | 0.079 | 01:31 |
| Air Blank          | 0.000 | 01:31 |
| Subject Sample # 1 | 0.184 | 01:32 |
| Breath Volume      | 1.203 | 01:32 |
| Air Blank          | 0.000 | 01:32 |
| Air Blank          | 0.000 | 01:34 |
| Subject Sample # 2 | 0.180 | 01:35 |
| Breath Volume      | 2.125 | 01:35 |
| Air Blank          | 0.000 | 01:35 |
| Control Test       | 0.079 | 01:36 |
| Air Blank          | 0.000 | 01:36 |
| Diagnostics Check  | OK    | 01:36 |

---

Data Download Date/Time: 05/12/2021 11:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POPICK,SCOTT,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: POPICK,SCOTT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: LUARCA

DATE/TIME OF TEST: 04/18/2021 05:05:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:58 |
| Air Blank          | 0.000 | 04:58 |
| Control Test       | 0.078 | 04:59 |
| Air Blank          | 0.000 | 04:59 |
| Subject Sample # 1 | 0.045 | 05:00 |
| Breath Volume      | 1.605 | 05:00 |
| Air Blank          | 0.000 | 05:01 |
| Air Blank          | 0.000 | 05:03 |
| Subject Sample # 2 | 0.041 | 05:03 |
| Breath Volume      | 1.476 | 05:03 |
| Air Blank          | 0.000 | 05:04 |
| Control Test       | 0.079 | 05:04 |
| Air Blank          | 0.000 | 05:05 |
| Diagnostics Check  | OK    | 05:05 |

---

Data Download Date/Time: 05/12/2021 11:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PATRON,A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: SOLOMON

DATE/TIME OF TEST: 04/03/2021 04:44:49

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 04:43 |                           |
| Air Blank         | 0.000 | 04:43 |                           |
| Control Test      | 0.000 | 04:44 | Control Outside Tolerance |
| Air Blank         | 0.000 | 04:44 |                           |

---

Data Download Date/Time: 04/06/2021 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THORNE,S,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: GRACIA,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: JWANOUSKOS

DATE/TIME OF TEST: 04/10/2021 20:55:48

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:49 |
| Air Blank          | 0.000 | 20:49 |
| Control Test       | 0.079 | 20:50 |
| Air Blank          | 0.000 | 20:50 |
| Subject Sample # 1 | 0.203 | 20:51 |
| Breath Volume      | 3.781 | 20:51 |
| Air Blank          | 0.000 | 20:51 |
| Air Blank          | 0.000 | 20:53 |
| Subject Sample # 2 | 0.198 | 20:54 |
| Breath Volume      | 3.417 | 20:54 |
| Air Blank          | 0.000 | 20:54 |
| Control Test       | 0.078 | 20:55 |
| Air Blank          | 0.000 | 20:55 |
| Diagnostics Check  | OK    | 20:55 |

---

Data Download Date/Time: 05/04/2021 15:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THORNE,S,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: THORNE,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: PEREZ HERNANDEZ

DATE/TIME OF TEST: 04/11/2021 01:16:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:08 |
| Air Blank          | 0.000 | 01:08 |
| Control Test       | 0.078 | 01:09 |
| Air Blank          | 0.000 | 01:09 |
| Subject Sample # 1 | 0.149 | 01:11 |
| Breath Volume      | 2.593 | 01:11 |
| Air Blank          | 0.000 | 01:12 |
| Air Blank          | 0.000 | 01:14 |
| Subject Sample # 2 | 0.130 | 01:14 |
| Breath Volume      | 1.937 | 01:14 |
| Air Blank          | 0.000 | 01:15 |
| Control Test       | 0.077 | 01:15 |
| Air Blank          | 0.000 | 01:16 |
| Diagnostics Check  | OK    | 01:16 |

---

Data Download Date/Time: 05/04/2021 15:28



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WASHINGTON, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:34

SUBJECT'S LAST NAME: ROCHA

DATE/TIME OF TEST: 04/13/2021 07:05:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:57 |
| Air Blank          | 0.000 | 06:57 |
| Control Test       | 0.079 | 06:58 |
| Air Blank          | 0.000 | 06:58 |
| Subject Sample # 1 | 0.195 | 07:00 |
| Breath Volume      | 2.539 | 07:00 |
| Air Blank          | 0.000 | 07:01 |
| Air Blank          | 0.000 | 07:02 |
| Subject Sample # 2 | 0.182 | 07:03 |
| Breath Volume      | 1.953 | 07:03 |
| Air Blank          | 0.000 | 07:04 |
| Control Test       | 0.079 | 07:04 |
| Air Blank          | 0.000 | 07:05 |
| Diagnostics Check  | OK    | 07:05 |

---

Data Download Date/Time: 05/04/2021 15:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THORNE,S,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURKE,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: LEHMAN

DATE/TIME OF TEST: 04/16/2021 02:56:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:50 |
| Air Blank          | 0.000 | 02:51 |
| Control Test       | 0.078 | 02:51 |
| Air Blank          | 0.000 | 02:51 |
| Subject Sample # 1 | 0.135 | 02:52 |
| Breath Volume      | 2.933 | 02:52 |
| Air Blank          | 0.000 | 02:53 |
| Air Blank          | 0.000 | 02:54 |
| Subject Sample # 2 | 0.133 | 02:55 |
| Breath Volume      | 2.875 | 02:55 |
| Air Blank          | 0.000 | 02:56 |
| Control Test       | 0.078 | 02:56 |
| Air Blank          | 0.000 | 02:56 |
| Diagnostics Check  | OK    | 02:56 |

---

Data Download Date/Time: 05/04/2021 15:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THORNE,S,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RIVERA,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 04/16/2021 03:15:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:08 |
| Air Blank          | 0.000 | 03:08 |
| Control Test       | 0.078 | 03:09 |
| Air Blank          | 0.000 | 03:09 |
| Subject Sample # 1 | 0.219 | 03:10 |
| Breath Volume      | 2.476 | 03:10 |
| Air Blank          | 0.000 | 03:11 |
| Air Blank          | 0.000 | 03:12 |
| Subject Sample # 2 | 0.203 | 03:13 |
| Breath Volume      | 1.742 | 03:13 |
| Air Blank          | 0.000 | 03:14 |
| Control Test       | 0.077 | 03:14 |
| Air Blank          | 0.000 | 03:14 |
| Diagnostics Check  | OK    | 03:15 |

---

Data Download Date/Time: 05/04/2021 15:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WASHINGTON, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: GARCIA-PEREZ

DATE/TIME OF TEST: 04/18/2021 04:10:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:01 |
| Air Blank          | 0.000 | 04:02 |
| Control Test       | 0.079 | 04:02 |
| Air Blank          | 0.000 | 04:02 |
| Subject Sample # 1 | 0.274 | 04:04 |
| Breath Volume      | 1.437 | 04:04 |
| Air Blank          | 0.000 | 04:05 |
| Air Blank          | 0.000 | 04:07 |
| Subject Sample # 2 | 0.273 | 04:08 |
| Breath Volume      | 1.140 | 04:08 |
| Air Blank          | 0.000 | 04:09 |
| Control Test       | 0.077 | 04:09 |
| Air Blank          | 0.000 | 04:10 |
| Diagnostics Check  | OK    | 04:10 |

---

Data Download Date/Time: 05/04/2021 15:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PATRON, A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: PRYCE JR

DATE/TIME OF TEST: 04/19/2021 00:11:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:03 |
| Air Blank          | 0.000 | 00:03 |
| Control Test       | 0.078 | 00:03 |
| Air Blank          | 0.000 | 00:04 |
| Subject Sample # 1 | 0.106 | 00:06 |
| Breath Volume      | 1.175 | 00:06 |
| Air Blank          | 0.000 | 00:06 |
| Air Blank          | 0.000 | 00:08 |
| Subject Sample # 2 | 0.105 | 00:09 |
| Breath Volume      | 1.187 | 00:09 |
| Air Blank          | 0.000 | 00:10 |
| Control Test       | 0.079 | 00:10 |
| Air Blank          | 0.000 | 00:11 |
| Diagnostics Check  | OK    | 00:11 |

---

Data Download Date/Time: 05/04/2021 15:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WASHINGTON, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: LAROSE

DATE/TIME OF TEST: 04/21/2021 22:06:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:59 |
| Air Blank          | 0.000 | 21:59 |
| Control Test       | 0.078 | 22:00 |
| Air Blank          | 0.000 | 22:00 |
| Subject Sample # 1 | 0.000 | 22:01 |
| Breath Volume      | 3.164 | 22:01 |
| Air Blank          | 0.000 | 22:02 |
| Air Blank          | 0.000 | 22:04 |
| Subject Sample # 2 | 0.000 | 22:05 |
| Breath Volume      | 3.187 | 22:05 |
| Air Blank          | 0.000 | 22:05 |
| Control Test       | 0.081 | 22:06 |
| Air Blank          | 0.000 | 22:06 |
| Diagnostics Check  | OK    | 22:06 |

---

Data Download Date/Time: 05/04/2021 15:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WASHINGTON, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: PICHARDO MARMOLEJOS

DATE/TIME OF TEST: 04/23/2021 02:47:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:39 |
| Air Blank          | 0.000 | 02:40 |
| Control Test       | 0.079 | 02:40 |
| Air Blank          | 0.000 | 02:40 |
| Subject Sample # 1 | 0.000 | 02:42 |
| Breath Volume      | 1.363 | 02:42 |
| Air Blank          | 0.000 | 02:42 |
| Air Blank          | 0.000 | 02:44 |
| Subject Sample # 2 | 0.000 | 02:46 |
| Breath Volume      | 2.136 | 02:46 |
| Air Blank          | 0.000 | 02:46 |
| Control Test       | 0.080 | 02:47 |
| Air Blank          | 0.000 | 02:47 |
| Diagnostics Check  | OK    | 02:47 |

---

Data Download Date/Time: 05/04/2021 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THORNE,S,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RIVERA,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 04/24/2021 00:10:34

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:07 |                      |
| Air Blank         | 0.000 | 00:08 |                      |
| Control Test      | 0.079 | 00:08 |                      |
| Air Blank         | 0.000 | 00:08 |                      |
| Subject Sample    | # 1   | 00:09 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:09 |                      |
| Air Blank         | 0.000 | 00:09 |                      |
| Control Test      | 0.079 | 00:09 |                      |
| Air Blank         | 0.000 | 00:10 |                      |
| Diagnostics Check | OK    | 00:10 |                      |

---

Data Download Date/Time: 05/04/2021 15:29



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THORNE,S,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FISHER,T  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: REPOZA

DATE/TIME OF TEST: 04/25/2021 00:20:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:14 |
| Air Blank          | 0.000 | 00:14 |
| Control Test       | 0.078 | 00:14 |
| Air Blank          | 0.000 | 00:15 |
| Subject Sample # 1 | 0.213 | 00:15 |
| Breath Volume      | 2.277 | 00:15 |
| Air Blank          | 0.000 | 00:16 |
| Air Blank          | 0.000 | 00:18 |
| Subject Sample # 2 | 0.195 | 00:18 |
| Breath Volume      | 1.507 | 00:18 |
| Air Blank          | 0.000 | 00:19 |
| Control Test       | 0.077 | 00:19 |
| Air Blank          | 0.000 | 00:20 |
| Diagnostics Check  | OK    | 00:20 |

---

Data Download Date/Time: 05/04/2021 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THORNE,S,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: THORNE,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: BARTON

DATE/TIME OF TEST: 04/25/2021 02:17:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:09 |
| Air Blank          | 0.000 | 02:09 |
| Control Test       | 0.078 | 02:10 |
| Air Blank          | 0.000 | 02:10 |
| Subject Sample # 1 | 0.000 | 02:13 |
| Breath Volume      | 1.265 | 02:13 |
| Air Blank          | 0.000 | 02:13 |
| Air Blank          | 0.000 | 02:15 |
| Subject Sample # 2 | 0.000 | 02:16 |
| Breath Volume      | 1.253 | 02:16 |
| Air Blank          | 0.000 | 02:16 |
| Control Test       | 0.078 | 02:16 |
| Air Blank          | 0.000 | 02:17 |
| Diagnostics Check  | OK    | 02:17 |

---

Data Download Date/Time: 05/04/2021 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THORNE,S,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RIVERA,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: VIERA JR

DATE/TIME OF TEST: 04/25/2021 05:01:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:54 |
| Air Blank          | 0.000 | 04:55 |
| Control Test       | 0.080 | 04:55 |
| Air Blank          | 0.000 | 04:56 |
| Subject Sample # 1 | 0.070 | 04:56 |
| Breath Volume      | 2.410 | 04:56 |
| Air Blank          | 0.000 | 04:57 |
| Air Blank          | 0.000 | 04:59 |
| Subject Sample # 2 | 0.071 | 04:59 |
| Breath Volume      | 2.484 | 04:59 |
| Air Blank          | 0.000 | 05:00 |
| Control Test       | 0.079 | 05:00 |
| Air Blank          | 0.000 | 05:01 |
| Diagnostics Check  | OK    | 05:01 |

---

Data Download Date/Time: 05/04/2021 15:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WASHINGTON,DESMND,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WASHINGTON,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: CHARTER

DATE/TIME OF TEST: 04/27/2021 04:39:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:33 |
| Air Blank          | 0.000 | 04:33 |
| Control Test       | 0.079 | 04:34 |
| Air Blank          | 0.000 | 04:34 |
| Subject Sample # 1 | 0.000 | 04:35 |
| Breath Volume      | 4.261 | 04:35 |
| Air Blank          | 0.000 | 04:35 |
| Air Blank          | 0.000 | 04:37 |
| Subject Sample # 2 | 0.000 | 04:38 |
| Breath Volume      | 4.156 | 04:38 |
| Air Blank          | 0.000 | 04:38 |
| Control Test       | 0.078 | 04:38 |
| Air Blank          | 0.000 | 04:39 |
| Diagnostics Check  | OK    | 04:39 |

---

Data Download Date/Time: 05/04/2021 15:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WASHINGTON, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: RAJAB

DATE/TIME OF TEST: 04/28/2021 05:42:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:34 |
| Air Blank          | 0.000 | 05:35 |
| Control Test       | 0.080 | 05:35 |
| Air Blank          | 0.000 | 05:36 |
| Subject Sample # 1 | 0.000 | 05:37 |
| Breath Volume      | 2.214 | 05:37 |
| Air Blank          | 0.000 | 05:37 |
| Air Blank          | 0.000 | 05:39 |
| Subject Sample # 2 | 0.000 | 05:41 |
| Breath Volume      | 1.476 | 05:41 |
| Air Blank          | 0.000 | 05:42 |
| Control Test       | 0.080 | 05:42 |
| Air Blank          | 0.000 | 05:42 |
| Diagnostics Check  | OK    | 05:42 |

---

Data Download Date/Time: 05/04/2021 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THORNE,S,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: THORNE,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 18220080A2  
DRY GAS EXPIRY DATE: 08/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: PEREIRA GONZALEZ

DATE/TIME OF TEST: 04/30/2021 03:22:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:15 |
| Air Blank          | 0.000 | 03:16 |
| Control Test       | 0.079 | 03:16 |
| Air Blank          | 0.000 | 03:16 |
| Subject Sample # 1 | 0.193 | 03:17 |
| Breath Volume      | 1.328 | 03:17 |
| Air Blank          | 0.000 | 03:18 |
| Air Blank          | 0.000 | 03:20 |
| Subject Sample # 2 | 0.188 | 03:20 |
| Breath Volume      | 1.320 | 03:20 |
| Air Blank          | 0.000 | 03:21 |
| Control Test       | 0.077 | 03:21 |
| Air Blank          | 0.000 | 03:22 |
| Diagnostics Check  | OK    | 03:22 |

---

Data Download Date/Time: 05/04/2021 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

INSTRUMENT REGISTERED TO: FWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BLANCO,JUAN,S

BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)

ARRESTING OFFICER: SAAVEDRA,OSCAR

ARRESTING OFFICER AGENCY: FFWCC (West Palm)

LAST AGENCY INSPECTION: 03/30/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1217042

DRY GAS EXPIRY DATE: 10/10/2022

VIOLATION CODE: BUI

ARREST TIME: 19:33

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: NIETO

DATE/TIME OF TEST: 04/30/2021 20:25:30

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:21 |                      |
| Air Blank         | 0.000 | 20:21 |                      |
| Control Test      | 0.080 | 20:22 |                      |
| Air Blank         | 0.000 | 20:22 |                      |
| Subject Sample    | # 1   | 20:24 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:24 |                      |
| Air Blank         | 0.000 | 20:24 |                      |
| Control Test      | 0.079 | 20:24 |                      |
| Air Blank         | 0.000 | 20:25 |                      |
| Diagnostics Check | OK    | 20:25 |                      |

---

Data Download Date/Time: 05/26/2021 17:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILKINS,MALACHI,C

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: WILKINS,MALACHI

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 31420080A2

DRY GAS EXPIRY DATE: 01/05/2023

VIOLATION CODE: BUI

ARREST TIME: 17:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: FALONE

DATE/TIME OF TEST: 04/23/2021 18:58:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:51 |
| Air Blank          | 0.000 | 18:52 |
| Control Test       | 0.079 | 18:52 |
| Air Blank          | 0.000 | 18:53 |
| Subject Sample # 1 | 0.098 | 18:53 |
| Breath Volume      | 1.820 | 18:53 |
| Air Blank          | 0.000 | 18:54 |
| Air Blank          | 0.000 | 18:56 |
| Subject Sample # 2 | 0.099 | 18:56 |
| Breath Volume      | 1.914 | 18:56 |
| Air Blank          | 0.000 | 18:57 |
| Control Test       | 0.078 | 18:57 |
| Air Blank          | 0.000 | 18:58 |
| Diagnostics Check  | OK    | 18:58 |

---

Data Download Date/Time: 05/21/2021 12:58



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILKINS,MALACHI,C

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: PETRU,MIROS

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 31420080A2

DRY GAS EXPIRY DATE: 01/05/2023

VIOLATION CODE: BUI

ARREST TIME: 16:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:19

SUBJECT'S LAST NAME: JURGENS

DATE/TIME OF TEST: 04/24/2021 16:56:54

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 16:45 |                    |
| Air Blank         | 0.000 | 16:45 |                    |
| Control Test      | 0.080 | 16:46 |                    |
| Air Blank         | 0.000 | 16:46 |                    |
| Subject Sample    | # 1   | 16:49 | No Sample Provided |
| Breath Volume     | 0.000 | 16:49 |                    |
| Air Blank         | 0.000 | 16:50 |                    |
| Air Blank         | 0.000 | 16:52 |                    |
| Subject Sample    | # 2   | 16:55 | No Sample Provided |
| Breath Volume     | 0.000 | 16:55 |                    |
| Air Blank         | 0.000 | 16:55 |                    |
| Control Test      | 0.080 | 16:56 |                    |
| Air Blank         | 0.000 | 16:56 |                    |
| Diagnostics Check | OK    | 16:56 |                    |

---

Data Download Date/Time: 05/21/2021 12:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILKINS,MALACHI,C

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: PASS,JUSTIN

ARRESTING OFFICER AGENCY: Hillsborough Co SO

LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 31420080A2

DRY GAS EXPIRY DATE: 01/05/2023

VIOLATION CODE: BUI

ARREST TIME: 17:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 04/24/2021 17:45:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:38 |
| Air Blank          | 0.000 | 17:39 |
| Control Test       | 0.079 | 17:39 |
| Air Blank          | 0.000 | 17:40 |
| Subject Sample # 1 | 0.102 | 17:40 |
| Breath Volume      | 1.597 | 17:40 |
| Air Blank          | 0.000 | 17:41 |
| Air Blank          | 0.000 | 17:43 |
| Subject Sample # 2 | 0.098 | 17:43 |
| Breath Volume      | 2.000 | 17:43 |
| Air Blank          | 0.000 | 17:44 |
| Control Test       | 0.078 | 17:44 |
| Air Blank          | 0.000 | 17:45 |
| Diagnostics Check  | OK    | 17:45 |

---

Data Download Date/Time: 05/21/2021 12:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILKINS,MALACHI,C

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: PROUTY,EDWARD

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 31420080A2

DRY GAS EXPIRY DATE: 01/05/2023

VIOLATION CODE: BUI

ARREST TIME: 20:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/24/2021 21:09:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:03 |
| Air Blank          | 0.000 | 21:03 |
| Control Test       | 0.078 | 21:03 |
| Air Blank          | 0.000 | 21:04 |
| Subject Sample # 1 | 0.112 | 21:05 |
| Breath Volume      | 4.097 | 21:05 |
| Air Blank          | 0.000 | 21:05 |
| Air Blank          | 0.000 | 21:07 |
| Subject Sample # 2 | 0.108 | 21:08 |
| Breath Volume      | 3.679 | 21:08 |
| Air Blank          | 0.000 | 21:08 |
| Control Test       | 0.079 | 21:09 |
| Air Blank          | 0.000 | 21:09 |
| Diagnostics Check  | OK    | 21:09 |

---

Data Download Date/Time: 05/21/2021 12:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RORER,JESSICA,L  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: HOPPE,ZACH  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: BUI  
ARREST TIME: 22:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: ALBACH

DATE/TIME OF TEST: 04/24/2021 23:10:55

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 22:59 |                    |
| Air Blank         | 0.000 | 22:59 |                    |
| Control Test      | 0.079 | 23:00 |                    |
| Air Blank         | 0.000 | 23:00 |                    |
| Subject Sample    | # 1   | 23:03 | No Sample Provided |
| Breath Volume     | 0.000 | 23:03 |                    |
| Air Blank         | 0.000 | 23:04 |                    |
| Air Blank         | 0.000 | 23:06 |                    |
| Subject Sample    | # 2   | 23:09 | No Sample Provided |
| Breath Volume     | 0.000 | 23:09 |                    |
| Air Blank         | 0.000 | 23:09 |                    |
| Control Test      | 0.079 | 23:10 |                    |
| Air Blank         | 0.000 | 23:10 |                    |
| Diagnostics Check | OK    | 23:10 |                    |

---

Data Download Date/Time: 05/21/2021 12:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILKINS,MALACHI,C

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: RORER,JESSICA

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 31420080A2

DRY GAS EXPIRY DATE: 01/05/2023

VIOLATION CODE: BUI

ARREST TIME: 17:22

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: LLANES

DATE/TIME OF TEST: 04/25/2021 18:49:25

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 18:46 |            |
| Air Blank         | 0.000 | 18:46 |            |
| Control Test      | 0.078 | 18:46 |            |
| Air Blank         | 0.000 | 18:47 |            |
| Subject Sample    | # 1   | 18:48 | RFI Detect |
| Breath Volume     | 0.000 | 18:48 |            |
| Air Blank         |       | 18:48 | RFI Detect |
| Air Blank         | 0.000 | 18:49 |            |

---

Data Download Date/Time: 05/21/2021 12:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,MALACHI,C  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: RORER,JESSICA  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 31420080A2  
DRY GAS EXPIRY DATE: 01/05/2023  
VIOLATION CODE: BUI  
ARREST TIME: 17:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: LLANES

DATE/TIME OF TEST: 04/25/2021 19:03:41

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 18:52 |                    |
| Air Blank         | 0.000 | 18:52 |                    |
| Control Test      | 0.079 | 18:53 |                    |
| Air Blank         | 0.000 | 18:53 |                    |
| Subject Sample    | # 1   | 18:56 | No Sample Provided |
| Breath Volume     | 0.000 | 18:56 |                    |
| Air Blank         | 0.000 | 18:57 |                    |
| Air Blank         | 0.000 | 18:59 |                    |
| Subject Sample    | # 2   | 19:02 | No Sample Provided |
| Breath Volume     | 0.000 | 19:02 |                    |
| Air Blank         | 0.000 | 19:02 |                    |
| Control Test      | 0.080 | 19:03 |                    |
| Air Blank         | 0.000 | 19:03 |                    |
| Diagnostics Check | OK    | 19:03 |                    |

---

Data Download Date/Time: 05/21/2021 12:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HRADECKY,CORBIN,J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: GUERRA,WILLIAM  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433249  
DRY GAS EXPIRY DATE: 01/08/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 04/01/2021 00:39:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:33 |
| Air Blank          | 0.000 | 00:33 |
| Control Test       | 0.080 | 00:33 |
| Air Blank          | 0.000 | 00:34 |
| Subject Sample # 1 | 0.116 | 00:35 |
| Breath Volume      | 1.691 | 00:35 |
| Air Blank          | 0.000 | 00:35 |
| Air Blank          | 0.000 | 00:37 |
| Subject Sample # 2 | 0.114 | 00:37 |
| Breath Volume      | 1.277 | 00:37 |
| Air Blank          | 0.000 | 00:38 |
| Control Test       | 0.080 | 00:38 |
| Air Blank          | 0.000 | 00:39 |
| Diagnostics Check  | OK    | 00:39 |

---

Data Download Date/Time: 04/08/2021 10:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUERRA, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: FOSTER, MEGAN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433249  
DRY GAS EXPIRY DATE: 01/08/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: LAMATTINA-DAVIS

DATE/TIME OF TEST: 04/01/2021 02:45:54

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:42 |                      |
| Air Blank         | 0.000 | 02:43 |                      |
| Control Test      | 0.080 | 02:43 |                      |
| Air Blank         | 0.000 | 02:44 |                      |
| Subject Sample    | # 1   | 02:44 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:44 |                      |
| Air Blank         | 0.000 | 02:44 |                      |
| Control Test      | 0.080 | 02:45 |                      |
| Air Blank         | 0.000 | 02:45 |                      |
| Diagnostics Check | OK    | 02:45 |                      |

---

Data Download Date/Time: 04/08/2021 10:49



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WESTERBAND,FRANK,M  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: AGUANNO,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433249  
DRY GAS EXPIRY DATE: 01/08/2024  
VIOLATION CODE: DUI  
ARREST TIME: 08:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:46

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 04/03/2021 09:22:06

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 09:10 |                    |
| Air Blank         | 0.000 | 09:11 |                    |
| Control Test      | 0.082 | 09:11 |                    |
| Air Blank         | 0.000 | 09:11 |                    |
| Subject Sample    | # 1   | 09:15 | No Sample Provided |
| Breath Volume     | 0.000 | 09:15 |                    |
| Air Blank         | 0.000 | 09:15 |                    |
| Air Blank         | 0.000 | 09:17 |                    |
| Subject Sample    | # 2   | 09:20 | No Sample Provided |
| Breath Volume     | 0.000 | 09:20 |                    |
| Air Blank         | 0.000 | 09:21 |                    |
| Control Test      | 0.081 | 09:21 |                    |
| Air Blank         | 0.000 | 09:22 |                    |
| Diagnostics Check | OK    | 09:22 |                    |

---

Data Download Date/Time: 04/08/2021 10:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WESTERBAND,FRANK,M  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: AGUANNO,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433249  
DRY GAS EXPIRY DATE: 01/08/2024  
VIOLATION CODE: DUI  
ARREST TIME: 08:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:46

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 04/03/2021 09:41:49

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 09:30 |                    |
| Air Blank         | 0.000 | 09:30 |                    |
| Control Test      | 0.081 | 09:31 |                    |
| Air Blank         | 0.000 | 09:31 |                    |
| Subject Sample    | # 1   | 09:34 | No Sample Provided |
| Breath Volume     | 0.000 | 09:34 |                    |
| Air Blank         | 0.000 | 09:35 |                    |
| Air Blank         | 0.000 | 09:37 |                    |
| Subject Sample    | # 2   | 09:40 | No Sample Provided |
| Breath Volume     | 0.000 | 09:40 |                    |
| Air Blank         | 0.000 | 09:40 |                    |
| Control Test      | 0.081 | 09:41 |                    |
| Air Blank         | 0.000 | 09:41 |                    |
| Diagnostics Check | OK    | 09:41 |                    |

---

Data Download Date/Time: 04/08/2021 10:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOURCIER, LEON, M  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: CHANDLER, KYLE  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433249  
DRY GAS EXPIRY DATE: 01/08/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: GONGORA DIAZ

DATE/TIME OF TEST: 04/04/2021 01:06:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:00 |
| Air Blank          | 0.000 | 01:00 |
| Control Test       | 0.081 | 01:01 |
| Air Blank          | 0.000 | 01:01 |
| Subject Sample # 1 | 0.234 | 01:02 |
| Breath Volume      | 2.945 | 01:02 |
| Air Blank          | 0.000 | 01:03 |
| Air Blank          | 0.000 | 01:04 |
| Subject Sample # 2 | 0.234 | 01:05 |
| Breath Volume      | 3.070 | 01:05 |
| Air Blank          | 0.000 | 01:05 |
| Control Test       | 0.080 | 01:06 |
| Air Blank          | 0.000 | 01:06 |
| Diagnostics Check  | OK    | 01:06 |

---

Data Download Date/Time: 04/08/2021 10:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUERRA, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: MARINO, JUAN-CARLOS  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433249  
DRY GAS EXPIRY DATE: 01/08/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: BOGGS

DATE/TIME OF TEST: 04/05/2021 22:44:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:37 |
| Air Blank          | 0.000 | 22:38 |
| Control Test       | 0.081 | 22:38 |
| Air Blank          | 0.000 | 22:39 |
| Subject Sample # 1 | 0.000 | 22:39 |
| Breath Volume      | 2.261 | 22:39 |
| Air Blank          | 0.000 | 22:40 |
| Air Blank          | 0.000 | 22:42 |
| Subject Sample # 2 | 0.000 | 22:42 |
| Breath Volume      | 2.593 | 22:42 |
| Air Blank          | 0.000 | 22:43 |
| Control Test       | 0.080 | 22:43 |
| Air Blank          | 0.000 | 22:44 |
| Diagnostics Check  | OK    | 22:44 |

---

Data Download Date/Time: 04/08/2021 10:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONES,SHAUN,G  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: LONES,SHAUN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433249  
DRY GAS EXPIRY DATE: 01/08/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 04/10/2021 21:53:16

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:49 |                      |
| Air Blank         | 0.000 | 21:50 |                      |
| Control Test      | 0.080 | 21:50 |                      |
| Air Blank         | 0.000 | 21:51 |                      |
| Subject Sample    | # 1   | 21:51 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:51 |                      |
| Air Blank         | 0.000 | 21:52 |                      |
| Control Test      | 0.080 | 21:52 |                      |
| Air Blank         | 0.000 | 21:53 |                      |
| Diagnostics Check | OK    | 21:53 |                      |

---

Data Download Date/Time: 05/12/2021 13:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUERRA, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HRADECKY, CORBIN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433249  
DRY GAS EXPIRY DATE: 01/08/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: WELLBORN

DATE/TIME OF TEST: 04/11/2021 01:43:58

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:41 |                      |
| Air Blank         | 0.000 | 01:41 |                      |
| Control Test      | 0.080 | 01:41 |                      |
| Air Blank         | 0.000 | 01:42 |                      |
| Subject Sample    | # 1   | 01:42 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:42 |                      |
| Air Blank         | 0.000 | 01:43 |                      |
| Control Test      | 0.080 | 01:43 |                      |
| Air Blank         | 0.000 | 01:43 |                      |
| Diagnostics Check | OK    | 01:43 |                      |

---

Data Download Date/Time: 05/12/2021 13:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HRADECKY,CORBIN,J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HRADECKY,CORBIN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433249  
DRY GAS EXPIRY DATE: 01/08/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: PLA

DATE/TIME OF TEST: 04/11/2021 22:06:51

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:03 |                      |
| Air Blank         | 0.000 | 22:04 |                      |
| Control Test      | 0.080 | 22:04 |                      |
| Air Blank         | 0.000 | 22:05 |                      |
| Subject Sample    | # 1   | 22:05 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:05 |                      |
| Air Blank         | 0.000 | 22:05 |                      |
| Control Test      | 0.080 | 22:06 |                      |
| Air Blank         | 0.000 | 22:06 |                      |
| Diagnostics Check | OK    | 22:06 |                      |

---

Data Download Date/Time: 05/12/2021 13:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUERRA, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HRADECKY, CORBIN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433249  
DRY GAS EXPIRY DATE: 01/08/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: ALLMAN

DATE/TIME OF TEST: 04/12/2021 02:27:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:20 |
| Air Blank          | 0.000 | 02:20 |
| Control Test       | 0.080 | 02:21 |
| Air Blank          | 0.000 | 02:21 |
| Subject Sample # 1 | 0.219 | 02:22 |
| Breath Volume      | 1.792 | 02:22 |
| Air Blank          | 0.000 | 02:23 |
| Air Blank          | 0.000 | 02:24 |
| Subject Sample # 2 | 0.221 | 02:25 |
| Breath Volume      | 1.222 | 02:25 |
| Air Blank          | 0.000 | 02:26 |
| Control Test       | 0.079 | 02:26 |
| Air Blank          | 0.000 | 02:27 |
| Diagnostics Check  | OK    | 02:27 |

---

Data Download Date/Time: 05/12/2021 13:02



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUERRA, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: GUERRA, WILLIAM  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433249  
DRY GAS EXPIRY DATE: 01/08/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 04/14/2021 23:53:46

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:50 |                      |
| Air Blank         | 0.000 | 23:51 |                      |
| Control Test      | 0.081 | 23:51 |                      |
| Air Blank         | 0.000 | 23:52 |                      |
| Subject Sample    | # 1   | 23:52 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:52 |                      |
| Air Blank         | 0.000 | 23:52 |                      |
| Control Test      | 0.080 | 23:53 |                      |
| Air Blank         | 0.000 | 23:53 |                      |
| Diagnostics Check | OK    | 23:53 |                      |

---

Data Download Date/Time: 05/12/2021 13:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOURCIER, LEON, M  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: MARINO, JUAN CARLOS  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433249  
DRY GAS EXPIRY DATE: 01/08/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: HAY

DATE/TIME OF TEST: 04/19/2021 01:29:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:22 |
| Air Blank          | 0.000 | 01:22 |
| Control Test       | 0.080 | 01:23 |
| Air Blank          | 0.000 | 01:23 |
| Subject Sample # 1 | 0.209 | 01:24 |
| Breath Volume      | 2.179 | 01:24 |
| Air Blank          | 0.000 | 01:25 |
| Air Blank          | 0.000 | 01:27 |
| Subject Sample # 2 | 0.216 | 01:27 |
| Breath Volume      | 1.187 | 01:27 |
| Air Blank          | 0.000 | 01:28 |
| Control Test       | 0.079 | 01:28 |
| Air Blank          | 0.000 | 01:29 |
| Diagnostics Check  | OK    | 01:29 |

---

Data Download Date/Time: 05/12/2021 13:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUERRA, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: FOSTER, MEGAN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433249  
DRY GAS EXPIRY DATE: 01/08/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: BRILL

DATE/TIME OF TEST: 04/21/2021 01:42:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:34 |
| Air Blank          | 0.000 | 01:34 |
| Control Test       | 0.080 | 01:35 |
| Air Blank          | 0.000 | 01:35 |
| Subject Sample # 1 | 0.207 | 01:37 |
| Breath Volume      | 3.058 | 01:37 |
| Air Blank          | 0.000 | 01:38 |
| Air Blank          | 0.000 | 01:40 |
| Subject Sample # 2 | 0.219 | 01:40 |
| Breath Volume      | 3.371 | 01:40 |
| Air Blank          | 0.000 | 01:41 |
| Control Test       | 0.079 | 01:41 |
| Air Blank          | 0.000 | 01:42 |
| Diagnostics Check  | OK    | 01:42 |

---

Data Download Date/Time: 05/12/2021 13:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUERRA, WILLIAM, J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HRADECKY, CORBIN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433249  
DRY GAS EXPIRY DATE: 01/08/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/24/2021 04:28:05

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:25 |                      |
| Air Blank         | 0.000 | 04:25 |                      |
| Control Test      | 0.080 | 04:25 |                      |
| Air Blank         | 0.000 | 04:26 |                      |
| Subject Sample    | # 1   | 04:26 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:26 |                      |
| Air Blank         | 0.000 | 04:27 |                      |
| Control Test      | 0.080 | 04:27 |                      |
| Air Blank         | 0.000 | 04:28 |                      |
| Diagnostics Check | OK    | 04:28 |                      |

---

Data Download Date/Time: 05/12/2021 13:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELMORE,BENJAMIN,R  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: ONSGARD,MICHAEL  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433249  
DRY GAS EXPIRY DATE: 01/08/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 04/24/2021 20:10:41

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:07 |                      |
| Air Blank         | 0.000 | 20:08 |                      |
| Control Test      | 0.081 | 20:08 |                      |
| Air Blank         | 0.000 | 20:09 |                      |
| Subject Sample    | # 1   | 20:09 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:09 |                      |
| Air Blank         | 0.000 | 20:09 |                      |
| Control Test      | 0.080 | 20:10 |                      |
| Air Blank         | 0.000 | 20:10 |                      |
| Diagnostics Check | OK    | 20:10 |                      |

---

Data Download Date/Time: 05/12/2021 13:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: LONEY,J  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: STAHN

DATE/TIME OF TEST: 04/01/2021 21:12:51

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 21:02 |                      |
| Air Blank         | 0.000     | 21:02 |                      |
| Control Test      | 0.081     | 21:03 |                      |
| Air Blank         | 0.000     | 21:03 |                      |
| Subject Sample    | # 1 0.116 | 21:06 | Volume Not Met       |
| Breath Volume     | 0.812     | 21:06 |                      |
| Air Blank         | 0.000     | 21:07 |                      |
| Air Blank         | 0.000     | 21:09 |                      |
| Subject Sample    | # 2       | 21:11 | Subject Test Refused |
| Breath Volume     | 0.210     | 21:11 |                      |
| Air Blank         | 0.000     | 21:11 |                      |
| Control Test      | 0.079     | 21:12 |                      |
| Air Blank         | 0.000     | 21:12 |                      |
| Diagnostics Check | OK        | 21:12 |                      |

---

Data Download Date/Time: 04/27/2021 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOEDICKE, WILLIAM, R  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: JOEDICKE, WILLIAM  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: SAMANIEGO

DATE/TIME OF TEST: 04/03/2021 22:57:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:50 |
| Air Blank          | 0.000 | 22:51 |
| Control Test       | 0.081 | 22:51 |
| Air Blank          | 0.000 | 22:51 |
| Subject Sample # 1 | 0.000 | 22:52 |
| Breath Volume      | 2.269 | 22:52 |
| Air Blank          | 0.000 | 22:53 |
| Air Blank          | 0.000 | 22:55 |
| Subject Sample # 2 | 0.000 | 22:55 |
| Breath Volume      | 2.390 | 22:55 |
| Air Blank          | 0.000 | 22:56 |
| Control Test       | 0.081 | 22:56 |
| Air Blank          | 0.000 | 22:57 |
| Diagnostics Check  | OK    | 22:57 |

---

Data Download Date/Time: 04/27/2021 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PETERSON,KIP,J  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: PETERSON,KIP  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: MCCORD

DATE/TIME OF TEST: 04/04/2021 02:24:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:15 |
| Air Blank          | 0.000 | 02:16 |
| Control Test       | 0.080 | 02:16 |
| Air Blank          | 0.000 | 02:16 |
| Subject Sample # 1 | 0.072 | 02:19 |
| Breath Volume      | 1.347 | 02:19 |
| Air Blank          | 0.000 | 02:19 |
| Air Blank          | 0.000 | 02:21 |
| Subject Sample # 2 | 0.070 | 02:22 |
| Breath Volume      | 1.406 | 02:22 |
| Air Blank          | 0.000 | 02:23 |
| Control Test       | 0.079 | 02:23 |
| Air Blank          | 0.000 | 02:24 |
| Diagnostics Check  | OK    | 02:24 |

---

Data Download Date/Time: 04/27/2021 14:23



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOEDICKE, WILLIAM, R  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: JOEDICKE, WILLIAM  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 04/07/2021 23:19:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:12 |
| Air Blank          | 0.000 | 23:12 |
| Control Test       | 0.080 | 23:13 |
| Air Blank          | 0.000 | 23:13 |
| Subject Sample # 1 | 0.175 | 23:14 |
| Breath Volume      | 1.652 | 23:14 |
| Air Blank          | 0.000 | 23:14 |
| Air Blank          | 0.000 | 23:16 |
| Subject Sample # 2 | 0.169 | 23:17 |
| Breath Volume      | 1.992 | 23:17 |
| Air Blank          | 0.000 | 23:18 |
| Control Test       | 0.080 | 23:18 |
| Air Blank          | 0.000 | 23:19 |
| Diagnostics Check  | OK    | 23:19 |

---

Data Download Date/Time: 04/27/2021 14:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORCHARD,DUSTIN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: ORCHARD,DUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 04/07/2021 23:52:51

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:49 |                      |
| Air Blank         | 0.000 | 23:50 |                      |
| Control Test      | 0.079 | 23:50 |                      |
| Air Blank         | 0.000 | 23:51 |                      |
| Subject Sample    | # 1   | 23:51 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:51 |                      |
| Air Blank         | 0.000 | 23:51 |                      |
| Control Test      | 0.080 | 23:52 |                      |
| Air Blank         | 0.000 | 23:52 |                      |
| Diagnostics Check | OK    | 23:52 |                      |

---

Data Download Date/Time: 04/27/2021 14:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORCHARD,DUSTIN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: ORCHARD,DUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: CURTIS

DATE/TIME OF TEST: 04/08/2021 21:45:26

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:42 |                      |
| Air Blank         | 0.000 | 21:42 |                      |
| Control Test      | 0.080 | 21:43 |                      |
| Air Blank         | 0.000 | 21:43 |                      |
| Subject Sample    | # 1   | 21:44 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:44 |                      |
| Air Blank         | 0.000 | 21:44 |                      |
| Control Test      | 0.080 | 21:44 |                      |
| Air Blank         | 0.000 | 21:45 |                      |
| Diagnostics Check | OK    | 21:45 |                      |

---

Data Download Date/Time: 04/27/2021 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOEDICKE, WILLIAM, R  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: JOEDICKE, WILLIAM  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 04/08/2021 22:17:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:10 |
| Air Blank          | 0.000 | 22:11 |
| Control Test       | 0.080 | 22:11 |
| Air Blank          | 0.000 | 22:12 |
| Subject Sample # 1 | 0.172 | 22:12 |
| Breath Volume      | 3.023 | 22:12 |
| Air Blank          | 0.000 | 22:13 |
| Air Blank          | 0.000 | 22:15 |
| Subject Sample # 2 | 0.170 | 22:15 |
| Breath Volume      | 3.425 | 22:15 |
| Air Blank          | 0.000 | 22:16 |
| Control Test       | 0.079 | 22:16 |
| Air Blank          | 0.000 | 22:17 |
| Diagnostics Check  | OK    | 22:17 |

---

Data Download Date/Time: 04/27/2021 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: YOUNG,ROBERT  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: VERNON

DATE/TIME OF TEST: 04/10/2021 03:25:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:19 |
| Air Blank          | 0.000 | 03:19 |
| Control Test       | 0.080 | 03:20 |
| Air Blank          | 0.000 | 03:20 |
| Subject Sample # 1 | 0.092 | 03:21 |
| Breath Volume      | 1.699 | 03:21 |
| Air Blank          | 0.000 | 03:22 |
| Air Blank          | 0.000 | 03:23 |
| Subject Sample # 2 | 0.091 | 03:24 |
| Breath Volume      | 1.437 | 03:24 |
| Air Blank          | 0.000 | 03:25 |
| Control Test       | 0.080 | 03:25 |
| Air Blank          | 0.000 | 03:25 |
| Diagnostics Check  | OK    | 03:25 |

---

Data Download Date/Time: 04/27/2021 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUSSO,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: CESPEDES,ROXANA  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HERRERA RAMOS

DATE/TIME OF TEST: 04/11/2021 03:15:27

---

|                   |       |       |                 |
|-------------------|-------|-------|-----------------|
| Diagnostics Check | OK    | 03:13 |                 |
| Air Blank         | 0.000 | 03:14 |                 |
| Control Test      | 0.080 | 03:14 |                 |
| Air Blank         | 0.000 | 03:15 |                 |
|                   |       | 03:15 | Improper Sample |
| Air Blank         |       | 03:15 | Improper Sample |

---

Data Download Date/Time: 04/27/2021 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUSSO,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: CESPEDES,ROXANA  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HERRERA RAMOS

DATE/TIME OF TEST: 04/11/2021 03:18:07

---

|                   |       |       |              |
|-------------------|-------|-------|--------------|
| Diagnostics Check | OK    | 03:17 |              |
| Air Blank         |       | 03:17 | Ambient Fail |
| Air Blank         | 0.000 | 03:18 |              |

---

Data Download Date/Time: 04/27/2021 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUSSO,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: CESPEDES,ROXANA  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HERRERA RAMOS

DATE/TIME OF TEST: 04/11/2021 03:25:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:19 |
| Air Blank          | 0.000 | 03:19 |
| Control Test       | 0.080 | 03:20 |
| Air Blank          | 0.000 | 03:20 |
| Subject Sample # 1 | 0.256 | 03:21 |
| Breath Volume      | 1.312 | 03:21 |
| Air Blank          | 0.000 | 03:21 |
| Air Blank          | 0.000 | 03:23 |
| Subject Sample # 2 | 0.251 | 03:24 |
| Breath Volume      | 1.472 | 03:24 |
| Air Blank          | 0.000 | 03:24 |
| Control Test       | 0.077 | 03:25 |
| Air Blank          | 0.000 | 03:25 |
| Diagnostics Check  | OK    | 03:25 |

---

Data Download Date/Time: 04/27/2021 14:23



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BALL,CODY,W  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: BALL,CODY  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 04/11/2021 23:57:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:50 |
| Air Blank          | 0.000 | 23:51 |
| Control Test       | 0.080 | 23:51 |
| Air Blank          | 0.000 | 23:52 |
| Subject Sample # 1 | 0.131 | 23:52 |
| Breath Volume      | 2.671 | 23:52 |
| Air Blank          | 0.000 | 23:53 |
| Air Blank          | 0.000 | 23:55 |
| Subject Sample # 2 | 0.136 | 23:55 |
| Breath Volume      | 2.250 | 23:55 |
| Air Blank          | 0.000 | 23:56 |
| Control Test       | 0.079 | 23:56 |
| Air Blank          | 0.000 | 23:57 |
| Diagnostics Check  | OK    | 23:57 |

---

Data Download Date/Time: 04/27/2021 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSACI, NEIL, R  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: ROSACI, NEIL  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 04/13/2021 20:19:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:13 |
| Air Blank          | 0.000 | 20:13 |
| Control Test       | 0.080 | 20:13 |
| Air Blank          | 0.000 | 20:14 |
| Subject Sample # 1 | 0.000 | 20:15 |
| Breath Volume      | 2.226 | 20:15 |
| Air Blank          | 0.000 | 20:15 |
| Air Blank          | 0.000 | 20:17 |
| Subject Sample # 2 | 0.000 | 20:18 |
| Breath Volume      | 2.386 | 20:18 |
| Air Blank          | 0.000 | 20:18 |
| Control Test       | 0.079 | 20:18 |
| Air Blank          | 0.000 | 20:19 |
| Diagnostics Check  | OK    | 20:19 |

---

Data Download Date/Time: 04/27/2021 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DETWEILER,BRETT,R  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: WILLIAMS,NIA  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 17:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: VAN

DATE/TIME OF TEST: 04/14/2021 18:30:15

---

|                    |       |       |            |
|--------------------|-------|-------|------------|
| Diagnostics Check  | OK    | 18:24 |            |
| Air Blank          | 0.000 | 18:25 |            |
| Control Test       | 0.079 | 18:25 |            |
| Air Blank          | 0.000 | 18:26 |            |
| Subject Sample # 1 | 0.408 | 18:26 |            |
| Breath Volume      | 2.585 | 18:26 |            |
| Air Blank          | 0.000 | 18:27 |            |
| Air Blank          |       | 18:28 | RFI Detect |
| Air Blank          | 0.000 | 18:29 |            |
| Control Test       | 0.080 | 18:29 |            |
| Air Blank          | 0.000 | 18:30 |            |
| Diagnostics Check  | OK    | 18:30 |            |

---

Data Download Date/Time: 04/27/2021 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DETWEILER,BRETT,R  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: WILLIAMS,NIA  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 15:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: VAN

DATE/TIME OF TEST: 04/14/2021 18:40:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:33 |
| Air Blank          | 0.000 | 18:33 |
| Control Test       | 0.079 | 18:33 |
| Air Blank          | 0.000 | 18:34 |
| Subject Sample # 1 | 0.391 | 18:34 |
| Breath Volume      | 2.394 | 18:34 |
| Air Blank          | 0.000 | 18:35 |
| Air Blank          | 0.000 | 18:37 |
| Subject Sample # 2 | 0.388 | 18:38 |
| Breath Volume      | 1.367 | 18:38 |
| Air Blank          | 0.000 | 18:39 |
| Control Test       | 0.078 | 18:39 |
| Air Blank          | 0.000 | 18:40 |
| Diagnostics Check  | OK    | 18:40 |

---

Data Download Date/Time: 04/27/2021 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRIGGS,AMBER,L  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: GRIGGS,AMBER  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 04/15/2021 00:41:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:33 |
| Air Blank          | 0.000 | 00:33 |
| Control Test       | 0.079 | 00:34 |
| Air Blank          | 0.000 | 00:34 |
| Subject Sample # 1 | 0.135 | 00:35 |
| Breath Volume      | 1.492 | 00:35 |
| Air Blank          | 0.000 | 00:36 |
| Air Blank          | 0.000 | 00:38 |
| Subject Sample # 2 | 0.129 | 00:39 |
| Breath Volume      | 1.789 | 00:39 |
| Air Blank          | 0.000 | 00:40 |
| Control Test       | 0.077 | 00:40 |
| Air Blank          | 0.000 | 00:40 |
| Diagnostics Check  | OK    | 00:41 |

---

Data Download Date/Time: 04/27/2021 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESQUIVEL,CHRISTINA,T  
BREATH TEST OPERATOR AGENCY: Belleview PD  
ARRESTING OFFICER: ESQUIVEL,CHRISTINA  
ARRESTING OFFICER AGENCY: Belleview PD  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 07:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:06

SUBJECT'S LAST NAME: ASAMOAH

DATE/TIME OF TEST: 04/18/2021 08:37:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:31 |
| Air Blank          | 0.000 | 08:31 |
| Control Test       | 0.080 | 08:32 |
| Air Blank          | 0.000 | 08:32 |
| Subject Sample # 1 | 0.000 | 08:33 |
| Breath Volume      | 1.937 | 08:33 |
| Air Blank          | 0.000 | 08:33 |
| Air Blank          | 0.000 | 08:35 |
| Subject Sample # 2 | 0.000 | 08:36 |
| Breath Volume      | 1.117 | 08:36 |
| Air Blank          | 0.000 | 08:36 |
| Control Test       | 0.080 | 08:37 |
| Air Blank          | 0.000 | 08:37 |
| Diagnostics Check  | OK    | 08:37 |

---

Data Download Date/Time: 04/27/2021 14:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: WILDEMAN-STUART,J  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: VELASCO

DATE/TIME OF TEST: 04/21/2021 01:56:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:50 |
| Air Blank          | 0.000 | 01:50 |
| Control Test       | 0.080 | 01:51 |
| Air Blank          | 0.000 | 01:51 |
| Subject Sample # 1 | 0.251 | 01:52 |
| Breath Volume      | 2.792 | 01:52 |
| Air Blank          | 0.000 | 01:53 |
| Air Blank          | 0.000 | 01:54 |
| Subject Sample # 2 | 0.239 | 01:55 |
| Breath Volume      | 2.000 | 01:55 |
| Air Blank          | 0.000 | 01:55 |
| Control Test       | 0.079 | 01:56 |
| Air Blank          | 0.000 | 01:56 |
| Diagnostics Check  | OK    | 01:56 |

---

Data Download Date/Time: 04/27/2021 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: MOOREHEAD,DAVID  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 04/21/2021 21:32:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:25 |
| Air Blank          | 0.000 | 21:25 |
| Control Test       | 0.080 | 21:26 |
| Air Blank          | 0.000 | 21:26 |
| Subject Sample # 1 | 0.221 | 21:27 |
| Breath Volume      | 3.082 | 21:27 |
| Air Blank          | 0.000 | 21:28 |
| Air Blank          | 0.000 | 21:30 |
| Subject Sample # 2 | 0.218 | 21:30 |
| Breath Volume      | 5.054 | 21:30 |
| Air Blank          | 0.000 | 21:31 |
| Control Test       | 0.079 | 21:31 |
| Air Blank          | 0.000 | 21:32 |
| Diagnostics Check  | OK    | 21:32 |

---

Data Download Date/Time: 04/27/2021 14:23



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAYS,MATTHEW,W  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: MAYS,MATTHEW  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 04/23/2021 00:08:34

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 23:56 |                |
| Air Blank         | 0.000     | 23:57 |                |
| Control Test      | 0.080     | 23:57 |                |
| Air Blank         | 0.000     | 23:58 |                |
| Subject Sample    | # 1 0.182 | 00:01 | Volume Not Met |
| Breath Volume     | 0.097     | 00:01 |                |
| Air Blank         | 0.000     | 00:02 |                |
| Air Blank         | 0.000     | 00:03 |                |
| Subject Sample    | # 2 0.178 | 00:06 | Volume Not Met |
| Breath Volume     | 0.503     | 00:06 |                |
| Air Blank         | 0.000     | 00:07 |                |
| Control Test      | 0.079     | 00:07 |                |
| Air Blank         | 0.000     | 00:08 |                |
| Diagnostics Check | OK        | 00:08 |                |

---

Data Download Date/Time: 04/27/2021 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FISHER,BRITTANY,N  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: FISHER,BRITTANY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 04/23/2021 01:57:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:51 |
| Air Blank          | 0.000 | 01:51 |
| Control Test       | 0.080 | 01:52 |
| Air Blank          | 0.000 | 01:52 |
| Subject Sample # 1 | 0.101 | 01:53 |
| Breath Volume      | 1.328 | 01:53 |
| Air Blank          | 0.000 | 01:53 |
| Air Blank          | 0.000 | 01:55 |
| Subject Sample # 2 | 0.100 | 01:55 |
| Breath Volume      | 1.429 | 01:55 |
| Air Blank          | 0.000 | 01:56 |
| Control Test       | 0.081 | 01:56 |
| Air Blank          | 0.000 | 01:57 |
| Diagnostics Check  | OK    | 01:57 |

---

Data Download Date/Time: 04/27/2021 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EDWARDS,ALAN,J  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: VASQUEZ,C  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: BONDI

DATE/TIME OF TEST: 04/24/2021 03:21:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:14 |
| Air Blank          | 0.000 | 03:15 |
| Control Test       | 0.080 | 03:15 |
| Air Blank          | 0.000 | 03:16 |
| Subject Sample # 1 | 0.000 | 03:17 |
| Breath Volume      | 1.671 | 03:17 |
| Air Blank          | 0.000 | 03:17 |
| Air Blank          | 0.000 | 03:19 |
| Subject Sample # 2 | 0.000 | 03:20 |
| Breath Volume      | 1.695 | 03:20 |
| Air Blank          | 0.000 | 03:20 |
| Control Test       | 0.079 | 03:20 |
| Air Blank          | 0.000 | 03:21 |
| Diagnostics Check  | OK    | 03:21 |

---

Data Download Date/Time: 04/27/2021 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRENCH,DUSTIN,S  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: FRENCH,DUSTIN  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 17:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: CONNOR

DATE/TIME OF TEST: 04/24/2021 19:36:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:28 |
| Air Blank          | 0.000 | 19:29 |
| Control Test       | 0.080 | 19:29 |
| Air Blank          | 0.000 | 19:29 |
| Subject Sample # 1 | 0.129 | 19:31 |
| Breath Volume      | 1.335 | 19:31 |
| Air Blank          | 0.000 | 19:32 |
| Air Blank          | 0.000 | 19:34 |
| Subject Sample # 2 | 0.125 | 19:34 |
| Breath Volume      | 1.441 | 19:34 |
| Air Blank          | 0.000 | 19:35 |
| Control Test       | 0.077 | 19:35 |
| Air Blank          | 0.000 | 19:36 |
| Diagnostics Check  | OK    | 19:36 |

---

Data Download Date/Time: 04/27/2021 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: YOUNG,ROBERT  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1298240  
DRY GAS EXPIRY DATE: 03/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: FOCKLER

DATE/TIME OF TEST: 04/25/2021 02:54:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:46 |
| Air Blank          | 0.000 | 02:46 |
| Control Test       | 0.080 | 02:47 |
| Air Blank          | 0.000 | 02:47 |
| Subject Sample # 1 | 0.205 | 02:48 |
| Breath Volume      | 2.226 | 02:48 |
| Air Blank          | 0.000 | 02:49 |
| Air Blank          | 0.000 | 02:51 |
| Subject Sample # 2 | 0.198 | 02:52 |
| Breath Volume      | 2.207 | 02:52 |
| Air Blank          | 0.000 | 02:53 |
| Control Test       | 0.079 | 02:53 |
| Air Blank          | 0.000 | 02:54 |
| Diagnostics Check  | OK    | 02:54 |

---

Data Download Date/Time: 04/27/2021 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOYER,ZACHARY,M  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: FELICIEN,PHILLIP  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1305520  
DRY GAS EXPIRY DATE: 04/01/2023  
VIOLATION CODE: DUI  
ARREST TIME: 15:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:43

SUBJECT'S LAST NAME: LITTON

DATE/TIME OF TEST: 04/28/2021 16:14:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:08 |
| Air Blank          | 0.000 | 16:08 |
| Control Test       | 0.080 | 16:09 |
| Air Blank          | 0.000 | 16:09 |
| Subject Sample # 1 | 0.267 | 16:10 |
| Breath Volume      | 1.445 | 16:10 |
| Air Blank          | 0.000 | 16:11 |
| Air Blank          | 0.000 | 16:12 |
| Subject Sample # 2 | 0.271 | 16:13 |
| Breath Volume      | 1.828 | 16:13 |
| Air Blank          | 0.000 | 16:13 |
| Control Test       | 0.078 | 16:14 |
| Air Blank          | 0.000 | 16:14 |
| Diagnostics Check  | OK    | 16:14 |

---

Data Download Date/Time: 05/26/2021 13:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: YOUONG,ROBERT  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1305520  
DRY GAS EXPIRY DATE: 04/01/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: IRIZARRY PLAZA

DATE/TIME OF TEST: 04/29/2021 03:47:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:40 |
| Air Blank          | 0.000 | 03:40 |
| Control Test       | 0.080 | 03:41 |
| Air Blank          | 0.000 | 03:41 |
| Subject Sample # 1 | 0.155 | 03:42 |
| Breath Volume      | 1.882 | 03:42 |
| Air Blank          | 0.000 | 03:43 |
| Air Blank          | 0.000 | 03:45 |
| Subject Sample # 2 | 0.142 | 03:45 |
| Breath Volume      | 2.265 | 03:45 |
| Air Blank          | 0.000 | 03:46 |
| Control Test       | 0.079 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Diagnostics Check  | OK    | 03:47 |

---

Data Download Date/Time: 05/26/2021 13:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLEMAN, JONATHAN, M  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: COLEMAN, JONATHAN  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1305520  
DRY GAS EXPIRY DATE: 04/01/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: SHOFFNER

DATE/TIME OF TEST: 04/30/2021 03:59:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:51 |
| Air Blank          | 0.000 | 03:52 |
| Control Test       | 0.079 | 03:52 |
| Air Blank          | 0.000 | 03:52 |
| Subject Sample # 1 | 0.302 | 03:53 |
| Breath Volume      | 1.218 | 03:53 |
| Air Blank          | 0.000 | 03:54 |
| Air Blank          | 0.000 | 03:56 |
| Subject Sample # 2 | 0.284 | 03:58 |
| Breath Volume      | 1.625 | 03:58 |
| Air Blank          | 0.000 | 03:58 |
| Control Test       | 0.078 | 03:59 |
| Air Blank          | 0.000 | 03:59 |
| Diagnostics Check  | OK    | 03:59 |

---

Data Download Date/Time: 05/26/2021 13:05



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAY,WILLIAM,P  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: MAY,WILLIAM  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 03/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/02/2021 00:48:12

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:44 |                      |
| Air Blank         | 0.000 | 00:44 |                      |
| Control Test      | 0.079 | 00:44 |                      |
| Air Blank         | 0.000 | 00:45 |                      |
| Subject Sample    | # 1   | 00:46 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:46 |                      |
| Air Blank         | 0.000 | 00:47 |                      |
| Control Test      | 0.081 | 00:47 |                      |
| Air Blank         | 0.000 | 00:48 |                      |
| Diagnostics Check | OK    | 00:48 |                      |

---

Data Download Date/Time: 04/17/2021 13:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOST,CRAIG,R  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: KRIMM,KOLE  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 03/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: HERMELINDO

DATE/TIME OF TEST: 04/04/2021 03:09:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:02 |
| Air Blank          | 0.000 | 03:03 |
| Control Test       | 0.081 | 03:03 |
| Air Blank          | 0.000 | 03:04 |
| Subject Sample # 1 | 0.174 | 03:04 |
| Breath Volume      | 2.015 | 03:04 |
| Air Blank          | 0.000 | 03:05 |
| Air Blank          | 0.000 | 03:07 |
| Subject Sample # 2 | 0.170 | 03:08 |
| Breath Volume      | 1.976 | 03:08 |
| Air Blank          | 0.000 | 03:08 |
| Control Test       | 0.083 | 03:09 |
| Air Blank          | 0.000 | 03:09 |
| Diagnostics Check  | OK    | 03:09 |

---

Data Download Date/Time: 04/17/2021 13:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAY,JODY,L  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: PHILIPS,HAILEY  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 03/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 04/05/2021 01:39:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:32 |
| Air Blank          | 0.000 | 01:32 |
| Control Test       | 0.079 | 01:32 |
| Air Blank          | 0.000 | 01:33 |
| Subject Sample # 1 | 0.178 | 01:34 |
| Breath Volume      | 1.718 | 01:34 |
| Air Blank          | 0.000 | 01:35 |
| Air Blank          | 0.000 | 01:36 |
| Subject Sample # 2 | 0.177 | 01:37 |
| Breath Volume      | 1.589 | 01:37 |
| Air Blank          | 0.000 | 01:38 |
| Control Test       | 0.081 | 01:38 |
| Air Blank          | 0.000 | 01:39 |
| Diagnostics Check  | OK    | 01:39 |

---

Data Download Date/Time: 04/17/2021 13:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRANE,RICHARD,B  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: DRANE,RICHARD  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 03/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: DEE

DATE/TIME OF TEST: 04/05/2021 19:22:27

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 19:10 |                    |
| Air Blank         | 0.000 | 19:11 |                    |
| Control Test      | 0.077 | 19:11 |                    |
| Air Blank         | 0.000 | 19:12 |                    |
| Subject Sample    | # 1   | 19:15 | No Sample Provided |
| Breath Volume     | 0.000 | 19:15 |                    |
| Air Blank         | 0.000 | 19:15 |                    |
| Air Blank         | 0.000 | 19:17 |                    |
| Subject Sample    | # 2   | 19:21 | No Sample Provided |
| Breath Volume     | 0.000 | 19:21 |                    |
| Air Blank         | 0.000 | 19:21 |                    |
| Control Test      | 0.078 | 19:21 |                    |
| Air Blank         | 0.000 | 19:22 |                    |
| Diagnostics Check | OK    | 19:22 |                    |

---

Data Download Date/Time: 04/17/2021 13:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOUCHNIR,EUGENE,V  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: KOUCHNIR,EUGENE  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 03/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 07:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: GIDDENS

DATE/TIME OF TEST: 04/09/2021 09:19:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:13 |
| Air Blank          | 0.000 | 09:13 |
| Control Test       | 0.079 | 09:13 |
| Air Blank          | 0.000 | 09:14 |
| Subject Sample # 1 | 0.148 | 09:15 |
| Breath Volume      | 3.289 | 09:15 |
| Air Blank          | 0.000 | 09:16 |
| Air Blank          | 0.000 | 09:17 |
| Subject Sample # 2 | 0.148 | 09:18 |
| Breath Volume      | 1.273 | 09:18 |
| Air Blank          | 0.000 | 09:18 |
| Control Test       | 0.080 | 09:19 |
| Air Blank          | 0.000 | 09:19 |
| Diagnostics Check  | OK    | 09:19 |

---

Data Download Date/Time: 04/17/2021 13:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAY,WILLIAM,P  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: MAY,WILLIAM  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 03/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 21:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: SPARKS

DATE/TIME OF TEST: 04/10/2021 22:34:57

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:29 |                      |
| Air Blank         | 0.000 | 22:30 |                      |
| Control Test      | 0.079 | 22:30 |                      |
| Air Blank         | 0.000 | 22:31 |                      |
| Subject Sample    | # 1   | 22:33 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:33 |                      |
| Air Blank         | 0.000 | 22:33 |                      |
| Control Test      | 0.081 | 22:34 |                      |
| Air Blank         | 0.000 | 22:34 |                      |
| Diagnostics Check | OK    | 22:34 |                      |

---

Data Download Date/Time: 04/17/2021 13:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAY,WILLIAM,P  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: MAY,WILLIAM  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 03/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 04:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 04/11/2021 05:28:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:21 |
| Air Blank          | 0.000 | 05:21 |
| Control Test       | 0.078 | 05:22 |
| Air Blank          | 0.000 | 05:22 |
| Subject Sample # 1 | 0.163 | 05:23 |
| Breath Volume      | 2.218 | 05:23 |
| Air Blank          | 0.000 | 05:24 |
| Air Blank          | 0.000 | 05:26 |
| Subject Sample # 2 | 0.150 | 05:26 |
| Breath Volume      | 2.285 | 05:26 |
| Air Blank          | 0.000 | 05:27 |
| Control Test       | 0.079 | 05:27 |
| Air Blank          | 0.000 | 05:28 |
| Diagnostics Check  | OK    | 05:28 |

---

Data Download Date/Time: 04/17/2021 13:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUSTER,CORY,J  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: GUSTER,CORY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 03/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 17:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: KENYON

DATE/TIME OF TEST: 04/11/2021 19:56:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:44 |
| Air Blank          | 0.000 | 19:45 |
| Control Test       | 0.078 | 19:45 |
| Air Blank          | 0.000 | 19:46 |
| Subject Sample # 1 | 0.154 | 19:47 |
| Breath Volume      | 1.179 | 19:47 |
| Air Blank          | 0.000 | 19:48 |
| Air Blank          | 0.000 | 19:49 |
| Subject Sample # 2 | 0.175 | 19:51 |
| Breath Volume      | 1.250 | 19:51 |
| Air Blank          | 0.000 | 19:52 |
| Air Blank          | 0.000 | 19:53 |
| Subject Sample # 3 | 0.158 | 19:54 |
| Breath Volume      | 1.781 | 19:54 |
| Air Blank          | 0.000 | 19:55 |
| Control Test       | 0.082 | 19:55 |
| Air Blank          | 0.000 | 19:56 |
| Diagnostics Check  | OK    | 19:56 |

---

Data Download Date/Time: 04/17/2021 13:44



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSE,ALEXIS,M  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: CREAMER,REMINGTON  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 03/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: CHRISTENSEN

DATE/TIME OF TEST: 04/13/2021 15:27:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:21 |
| Air Blank          | 0.000 | 15:21 |
| Control Test       | 0.078 | 15:21 |
| Air Blank          | 0.000 | 15:22 |
| Subject Sample # 1 | 0.147 | 15:23 |
| Breath Volume      | 3.929 | 15:23 |
| Air Blank          | 0.000 | 15:23 |
| Air Blank          | 0.000 | 15:25 |
| Subject Sample # 2 | 0.155 | 15:26 |
| Breath Volume      | 2.621 | 15:26 |
| Air Blank          | 0.000 | 15:26 |
| Control Test       | 0.082 | 15:27 |
| Air Blank          | 0.000 | 15:27 |
| Diagnostics Check  | OK    | 15:27 |

---

Data Download Date/Time: 04/17/2021 13:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BREEDLOVE,TROY,J  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: CREAMER,REMINGTON  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 03/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 07:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:26

SUBJECT'S LAST NAME: LYONS

DATE/TIME OF TEST: 04/17/2021 09:10:25

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 09:07 |                      |
| Air Blank         | 0.000 | 09:07 |                      |
| Control Test      | 0.078 | 09:08 |                      |
| Air Blank         | 0.000 | 09:08 |                      |
| Subject Sample    | # 1   | 09:08 | Subject Test Refused |
| Breath Volume     | 0.000 | 09:08 |                      |
| Air Blank         | 0.000 | 09:09 |                      |
| Control Test      | 0.079 | 09:09 |                      |
| Air Blank         | 0.000 | 09:10 |                      |
| Diagnostics Check | OK    | 09:10 |                      |

---

Data Download Date/Time: 04/17/2021 13:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: DAVIS,ALEXANDER  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 04/17/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 22:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BERNAL BERNAL

DATE/TIME OF TEST: 04/23/2021 23:33:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:26 |
| Air Blank          | 0.000 | 23:26 |
| Control Test       | 0.079 | 23:26 |
| Air Blank          | 0.000 | 23:27 |
| Subject Sample # 1 | 0.133 | 23:29 |
| Breath Volume      | 2.730 | 23:29 |
| Air Blank          | 0.000 | 23:29 |
| Air Blank          | 0.000 | 23:31 |
| Subject Sample # 2 | 0.132 | 23:32 |
| Breath Volume      | 2.132 | 23:32 |
| Air Blank          | 0.000 | 23:32 |
| Control Test       | 0.082 | 23:33 |
| Air Blank          | 0.000 | 23:33 |
| Diagnostics Check  | OK    | 23:33 |

---

Data Download Date/Time: 05/15/2021 01:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: VENTRY,WILLIS  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 04/17/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: HEMMER

DATE/TIME OF TEST: 04/30/2021 03:03:34

---

|                   |           |       |                           |
|-------------------|-----------|-------|---------------------------|
| Diagnostics Check | OK        | 02:53 |                           |
| Air Blank         | 0.000     | 02:53 |                           |
| Control Test      | 0.078     | 02:53 |                           |
| Air Blank         | 0.000     | 02:54 |                           |
| Subject Sample    | # 1 0.189 | 02:56 |                           |
| Breath Volume     | 1.117     | 02:56 |                           |
| Air Blank         | 0.000     | 02:57 |                           |
| Air Blank         | 0.000     | 02:59 |                           |
| Subject Sample    | # 2 0.189 | 03:01 |                           |
| Breath Volume     | 2.804     | 03:01 |                           |
| Air Blank         | 0.000     | 03:02 |                           |
| Control Test      | 0.086     | 03:02 | Control Outside Tolerance |
| Air Blank         | 0.000     | 03:03 |                           |

---

Data Download Date/Time: 05/15/2021 01:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: VENTRY,WILLIS  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 04/17/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: HEMMER

DATE/TIME OF TEST: 04/30/2021 03:21:04

---

|                   |           |       |                   |
|-------------------|-----------|-------|-------------------|
| Diagnostics Check | OK        | 03:06 |                   |
| Air Blank         | 0.000     | 03:07 |                   |
| Control Test      | 0.080     | 03:07 |                   |
| Air Blank         | 0.000     | 03:07 |                   |
| Subject Sample    | # 1 0.155 | 03:11 | Volume Not Met    |
| Breath Volume     | 0.484     | 03:11 |                   |
| Air Blank         | 0.000     | 03:11 |                   |
| Air Blank         | 0.000     | 03:13 |                   |
| Subject Sample    | # 2 0.185 | 03:14 | No .020 agreement |
| Breath Volume     | 1.789     | 03:14 |                   |
| Air Blank         | 0.000     | 03:15 |                   |
| Air Blank         | 0.000     | 03:16 |                   |
| Subject Sample    | # 3 0.156 | 03:19 | No .020 agreement |
| Breath Volume     | 1.804     | 03:19 |                   |
| Air Blank         | 0.000     | 03:20 |                   |
| Control Test      | 0.079     | 03:20 |                   |
| Air Blank         | 0.000     | 03:20 |                   |
| Diagnostics Check | OK        | 03:21 |                   |

---

Data Download Date/Time: 05/15/2021 01:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: VENTRY,WILLIS  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 04/17/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1012059  
DRY GAS EXPIRY DATE: 09/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: HEMMER

DATE/TIME OF TEST: 04/30/2021 03:39:13

---

|                   |           |       |                           |
|-------------------|-----------|-------|---------------------------|
| Diagnostics Check | OK        | 03:27 |                           |
| Air Blank         | 0.000     | 03:27 |                           |
| Control Test      | 0.076     | 03:27 |                           |
| Air Blank         | 0.000     | 03:28 |                           |
| Subject Sample    | # 1 0.181 | 03:31 | Volume Not Met            |
| Breath Volume     | 0.917     | 03:31 |                           |
| Air Blank         | 0.000     | 03:32 |                           |
| Air Blank         | 0.000     | 03:33 |                           |
| Subject Sample    | # 2 0.160 | 03:37 | Volume Not Met            |
| Breath Volume     | 0.015     | 03:37 |                           |
| Air Blank         | 0.000     | 03:38 |                           |
| Control Test      | 0.074     | 03:38 | Control Outside Tolerance |
| Air Blank         | 0.000     | 03:39 |                           |

---

Data Download Date/Time: 05/15/2021 01:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,SEAN,C  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: BAKER,JERRY  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A2  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 04/05/2021 20:40:18

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:36 |                      |
| Air Blank         | 0.000 | 20:37 |                      |
| Control Test      | 0.079 | 20:37 |                      |
| Air Blank         | 0.000 | 20:38 |                      |
| Subject Sample    | # 1   | 20:38 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:38 |                      |
| Air Blank         | 0.000 | 20:39 |                      |
| Control Test      | 0.079 | 20:39 |                      |
| Air Blank         | 0.000 | 20:40 |                      |
| Diagnostics Check | OK    | 20:40 |                      |

---

Data Download Date/Time: 04/16/2021 11:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAYWARD,KYLE,M  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: WARWIN,K  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 14020080A2  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: MOREHOUSE

DATE/TIME OF TEST: 04/06/2021 19:22:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:16 |
| Air Blank          | 0.000 | 19:16 |
| Control Test       | 0.078 | 19:16 |
| Air Blank          | 0.000 | 19:17 |
| Subject Sample # 1 | 0.200 | 19:17 |
| Breath Volume      | 3.125 | 19:17 |
| Air Blank          | 0.000 | 19:18 |
| Air Blank          | 0.000 | 19:20 |
| Subject Sample # 2 | 0.211 | 19:20 |
| Breath Volume      | 3.480 | 19:20 |
| Air Blank          | 0.000 | 19:21 |
| Control Test       | 0.077 | 19:21 |
| Air Blank          | 0.000 | 19:22 |
| Diagnostics Check  | OK    | 19:22 |

---

Data Download Date/Time: 04/16/2021 11:17



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAYWARD,KYLE,M

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: GORDON,JOSH

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 14020080A2

DRY GAS EXPIRY DATE: 07/05/2022

VIOLATION CODE: DUI

ARREST TIME: 19:32

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: TENORIO

DATE/TIME OF TEST: 04/06/2021 20:20:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:13 |
| Air Blank          | 0.000 | 20:14 |
| Control Test       | 0.078 | 20:14 |
| Air Blank          | 0.000 | 20:14 |
| Subject Sample # 1 | 0.171 | 20:15 |
| Breath Volume      | 1.304 | 20:15 |
| Air Blank          | 0.000 | 20:16 |
| Air Blank          | 0.000 | 20:18 |
| Subject Sample # 2 | 0.172 | 20:18 |
| Breath Volume      | 1.589 | 20:18 |
| Air Blank          | 0.000 | 20:19 |
| Control Test       | 0.076 | 20:19 |
| Air Blank          | 0.000 | 20:20 |
| Diagnostics Check  | OK    | 20:20 |

---

Data Download Date/Time: 04/16/2021 11:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMS,WALKER,L  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: WARWIN,PATRICK  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A2  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: GOODMAN

DATE/TIME OF TEST: 04/10/2021 18:57:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:50 |
| Air Blank          | 0.000 | 18:50 |
| Control Test       | 0.079 | 18:51 |
| Air Blank          | 0.000 | 18:51 |
| Subject Sample # 1 | 0.163 | 18:52 |
| Breath Volume      | 1.238 | 18:52 |
| Air Blank          | 0.000 | 18:53 |
| Air Blank          | 0.000 | 18:55 |
| Subject Sample # 2 | 0.169 | 18:56 |
| Breath Volume      | 1.285 | 18:56 |
| Air Blank          | 0.000 | 18:56 |
| Control Test       | 0.079 | 18:57 |
| Air Blank          | 0.000 | 18:57 |
| Diagnostics Check  | OK    | 18:57 |

---

Data Download Date/Time: 04/16/2021 11:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,DARBY,J  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: BROWN,JASON  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A2  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: PRUNET

DATE/TIME OF TEST: 04/18/2021 01:06:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:00 |
| Air Blank          | 0.000 | 01:01 |
| Control Test       | 0.079 | 01:01 |
| Air Blank          | 0.000 | 01:02 |
| Subject Sample # 1 | 0.200 | 01:02 |
| Breath Volume      | 3.652 | 01:02 |
| Air Blank          | 0.000 | 01:03 |
| Air Blank          | 0.000 | 01:04 |
| Subject Sample # 2 | 0.197 | 01:05 |
| Breath Volume      | 3.593 | 01:05 |
| Air Blank          | 0.000 | 01:06 |
| Control Test       | 0.077 | 01:06 |
| Air Blank          | 0.000 | 01:06 |
| Diagnostics Check  | OK    | 01:06 |

---

Data Download Date/Time: 05/26/2021 20:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUSIAL,CHRISTOPHER,S  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MUSIAL,CHRISTOPHER  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 14020080A2  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: DENNIS

DATE/TIME OF TEST: 04/19/2021 19:13:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:06 |
| Air Blank          | 0.000 | 19:06 |
| Control Test       | 0.078 | 19:07 |
| Air Blank          | 0.000 | 19:07 |
| Subject Sample # 1 | 0.192 | 19:08 |
| Breath Volume      | 3.464 | 19:08 |
| Air Blank          | 0.000 | 19:09 |
| Air Blank          | 0.000 | 19:10 |
| Subject Sample # 2 | 0.175 | 19:12 |
| Breath Volume      | 2.441 | 19:12 |
| Air Blank          | 0.000 | 19:12 |
| Control Test       | 0.078 | 19:13 |
| Air Blank          | 0.000 | 19:13 |
| Diagnostics Check  | OK    | 19:13 |

---

Data Download Date/Time: 05/26/2021 20:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,JOHNATHAN,R  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: TUSHA,JYLTON  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A2  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: CHRISTIANSON

DATE/TIME OF TEST: 04/20/2021 00:02:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:55 |
| Air Blank          | 0.000 | 23:56 |
| Control Test       | 0.078 | 23:56 |
| Air Blank          | 0.000 | 23:57 |
| Subject Sample # 2 | 0.010 | 23:57 |
| Breath Volume      | 2.089 | 23:57 |
| Air Blank          | 0.000 | 23:58 |
| Air Blank          | 0.000 | 00:00 |
| Subject Sample # 1 | 0.010 | 00:00 |
| Breath Volume      | 2.113 | 00:00 |
| Air Blank          | 0.000 | 00:01 |
| Control Test       | 0.078 | 00:01 |
| Air Blank          | 0.000 | 00:02 |
| Diagnostics Check  | OK    | 00:02 |

---

Data Download Date/Time: 05/26/2021 20:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAYWARD,KYLE,M  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: HAYWARD,KYLE  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A2  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:56

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 04/20/2021 17:24:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:16 |
| Air Blank          | 0.000 | 17:16 |
| Control Test       | 0.078 | 17:17 |
| Air Blank          | 0.000 | 17:17 |
| Subject Sample # 1 | 0.258 | 17:19 |
| Breath Volume      | 2.296 | 17:19 |
| Air Blank          | 0.000 | 17:20 |
| Air Blank          | 0.000 | 17:22 |
| Subject Sample # 2 | 0.253 | 17:22 |
| Breath Volume      | 2.234 | 17:22 |
| Air Blank          | 0.000 | 17:23 |
| Control Test       | 0.075 | 17:23 |
| Air Blank          | 0.000 | 17:24 |
| Diagnostics Check  | OK    | 17:24 |

---

Data Download Date/Time: 05/26/2021 20:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MILLS,SCOTT,T

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: MILLS,SCOTT

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 14020080A2

DRY GAS EXPIRY DATE: 07/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:49

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: NESBIT

DATE/TIME OF TEST: 04/21/2021 03:52:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:45 |
| Air Blank          | 0.000 | 03:45 |
| Control Test       | 0.078 | 03:46 |
| Air Blank          | 0.000 | 03:46 |
| Subject Sample # 1 | 0.072 | 03:47 |
| Breath Volume      | 1.269 | 03:47 |
| Air Blank          | 0.000 | 03:48 |
| Air Blank          | 0.000 | 03:49 |
| Subject Sample # 2 | 0.064 | 03:50 |
| Breath Volume      | 1.636 | 03:50 |
| Air Blank          | 0.000 | 03:51 |
| Control Test       | 0.078 | 03:51 |
| Air Blank          | 0.000 | 03:52 |
| Diagnostics Check  | OK    | 03:52 |

---

Data Download Date/Time: 05/26/2021 20:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: COMBS,BRADLEY  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A2  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:59

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/23/2021 18:28:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:21 |
| Air Blank          | 0.000 | 18:22 |
| Control Test       | 0.079 | 18:22 |
| Air Blank          | 0.000 | 18:22 |
| Subject Sample # 1 | 0.132 | 18:23 |
| Breath Volume      | 1.335 | 18:23 |
| Air Blank          | 0.000 | 18:24 |
| Air Blank          | 0.000 | 18:26 |
| Subject Sample # 2 | 0.133 | 18:27 |
| Breath Volume      | 1.648 | 18:27 |
| Air Blank          | 0.000 | 18:28 |
| Control Test       | 0.078 | 18:28 |
| Air Blank          | 0.000 | 18:28 |
| Diagnostics Check  | OK    | 18:28 |

---

Data Download Date/Time: 05/26/2021 20:29



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DA SILVA,ANDRE,F

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: COMBS,A

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 14020080A2

DRY GAS EXPIRY DATE: 07/05/2022

VIOLATION CODE: DUI

ARREST TIME: 16:41

GENDER: Female

OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: PITRE

DATE/TIME OF TEST: 04/25/2021 17:57:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:49 |
| Air Blank          | 0.000 | 17:50 |
| Control Test       | 0.078 | 17:50 |
| Air Blank          | 0.000 | 17:50 |
| Subject Sample # 1 | 0.125 | 17:51 |
| Breath Volume      | 1.847 | 17:51 |
| Air Blank          | 0.000 | 17:52 |
| Air Blank          | 0.000 | 17:54 |
| Subject Sample # 2 | 0.122 | 17:55 |
| Breath Volume      | 2.093 | 17:55 |
| Air Blank          | 0.000 | 17:56 |
| Control Test       | 0.079 | 17:56 |
| Air Blank          | 0.000 | 17:57 |
| Diagnostics Check  | OK    | 17:57 |

---

Data Download Date/Time: 05/26/2021 20:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: BARRERA,ALEXANDER  
ARRESTING OFFICER AGENCY: St. Augustine PD  
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A2  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/27/2021 00:58:14

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:52 |                      |
| Air Blank         | 0.000 | 00:53 |                      |
| Control Test      | 0.078 | 00:53 |                      |
| Air Blank         | 0.000 | 00:54 |                      |
| Subject Sample    | # 1   | 00:56 | Subject Test Refused |
| Breath Volume     | 0.277 | 00:56 |                      |
| Air Blank         | 0.000 | 00:57 |                      |
| Control Test      | 0.078 | 00:57 |                      |
| Air Blank         | 0.000 | 00:58 |                      |
| Diagnostics Check | OK    | 00:58 |                      |

---

Data Download Date/Time: 05/26/2021 20:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: BARRERA,ALEXANDER  
ARRESTING OFFICER AGENCY: St. Augustine PD  
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A2  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: SANFORD

DATE/TIME OF TEST: 04/27/2021 20:02:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:56 |
| Air Blank          | 0.000 | 19:56 |
| Control Test       | 0.078 | 19:57 |
| Air Blank          | 0.000 | 19:57 |
| Subject Sample # 1 | 0.299 | 19:58 |
| Breath Volume      | 1.960 | 19:58 |
| Air Blank          | 0.000 | 19:58 |
| Air Blank          | 0.000 | 20:00 |
| Subject Sample # 2 | 0.283 | 20:01 |
| Breath Volume      | 2.011 | 20:01 |
| Air Blank          | 0.000 | 20:01 |
| Control Test       | 0.078 | 20:02 |
| Air Blank          | 0.000 | 20:02 |
| Diagnostics Check  | OK    | 20:02 |

---

Data Download Date/Time: 05/26/2021 20:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREEN,J,A

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: FIGUEROA,K

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 14020080A2

DRY GAS EXPIRY DATE: 07/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:27

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/30/2021 03:37:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:31 |
| Air Blank          | 0.000 | 03:31 |
| Control Test       | 0.078 | 03:31 |
| Air Blank          | 0.000 | 03:32 |
| Subject Sample # 1 | 0.137 | 03:32 |
| Breath Volume      | 2.062 | 03:32 |
| Air Blank          | 0.000 | 03:33 |
| Air Blank          | 0.000 | 03:35 |
| Subject Sample # 2 | 0.138 | 03:35 |
| Breath Volume      | 1.617 | 03:35 |
| Air Blank          | 0.000 | 03:36 |
| Control Test       | 0.077 | 03:36 |
| Air Blank          | 0.000 | 03:37 |
| Diagnostics Check  | OK    | 03:37 |

---

Data Download Date/Time: 05/26/2021 20:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NASH,JOSHUA,D  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: NASH,JOSHUA  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 03/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: LUCIUS

DATE/TIME OF TEST: 04/03/2021 01:24:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:17 |
| Air Blank          | 0.000 | 01:18 |
| Control Test       | 0.078 | 01:18 |
| Air Blank          | 0.000 | 01:19 |
| Subject Sample # 1 | 0.167 | 01:19 |
| Breath Volume      | 2.101 | 01:19 |
| Air Blank          | 0.000 | 01:20 |
| Air Blank          | 0.000 | 01:22 |
| Subject Sample # 2 | 0.174 | 01:22 |
| Breath Volume      | 1.828 | 01:22 |
| Air Blank          | 0.000 | 01:23 |
| Control Test       | 0.077 | 01:23 |
| Air Blank          | 0.000 | 01:24 |
| Diagnostics Check  | OK    | 01:24 |

---

Data Download Date/Time: 04/17/2021 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NASH,JOSHUA,D  
BREATH TEST OPERATOR AGENCY: Walton County SO  
ARRESTING OFFICER: SYVERSON,PAUL  
ARRESTING OFFICER AGENCY: Walton County SO  
LAST AGENCY INSPECTION: 03/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 04/03/2021 05:04:17

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 05:01 |                      |
| Air Blank         | 0.000 | 05:01 |                      |
| Control Test      | 0.078 | 05:02 |                      |
| Air Blank         | 0.000 | 05:02 |                      |
| Subject Sample    | # 1   | 05:02 | Subject Test Refused |
| Breath Volume     | 0.000 | 05:02 |                      |
| Air Blank         | 0.000 | 05:03 |                      |
| Control Test      | 0.079 | 05:03 |                      |
| Air Blank         | 0.000 | 05:04 |                      |
| Diagnostics Check | OK    | 05:04 |                      |

---

Data Download Date/Time: 04/17/2021 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYBRAND,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: LYBRAND,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 03/15/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 30219080A3  
DRY GAS EXPIRY DATE: 01/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 04/09/2021 02:19:27

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:16 |                      |
| Air Blank         | 0.000 | 02:16 |                      |
| Control Test      | 0.077 | 02:17 |                      |
| Air Blank         | 0.000 | 02:17 |                      |
| Subject Sample    | # 1   | 02:18 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:18 |                      |
| Air Blank         | 0.000 | 02:18 |                      |
| Control Test      | 0.075 | 02:18 |                      |
| Air Blank         | 0.000 | 02:19 |                      |
| Diagnostics Check | OK    | 02:19 |                      |

---

Data Download Date/Time: 04/17/2021 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: DIAZ,JASIEL  
ARRESTING OFFICER AGENCY: Hallandale Beach PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: FENTON

DATE/TIME OF TEST: 04/10/2021 04:24:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:17 |
| Air Blank          | 0.000 | 04:18 |
| Control Test       | 0.081 | 04:18 |
| Air Blank          | 0.000 | 04:19 |
| Subject Sample # 1 | 0.044 | 04:20 |
| Breath Volume      | 1.425 | 04:20 |
| Air Blank          | 0.000 | 04:20 |
| Air Blank          | 0.000 | 04:22 |
| Subject Sample # 2 | 0.042 | 04:23 |
| Breath Volume      | 1.308 | 04:23 |
| Air Blank          | 0.000 | 04:23 |
| Control Test       | 0.080 | 04:24 |
| Air Blank          | 0.000 | 04:24 |
| Diagnostics Check  | OK    | 04:24 |

---

Data Download Date/Time: 04/13/2021 11:00



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: HALES,ANDREW  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: ARANGO GIRALDO

DATE/TIME OF TEST: 04/10/2021 04:44:13

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 04:30 |                    |
| Air Blank         | 0.000 | 04:30 |                    |
| Control Test      | 0.081 | 04:30 |                    |
| Air Blank         | 0.000 | 04:31 |                    |
| Subject Sample    | # 1   | 04:34 | No Sample Provided |
| Breath Volume     | 0.000 | 04:34 |                    |
| Air Blank         | 0.000 | 04:34 |                    |
| Air Blank         | 0.000 | 04:36 |                    |
| Subject Sample    | # 2   | 0.106 | 04:38              |
| Breath Volume     | 1.281 | 04:38 |                    |
| Air Blank         | 0.000 | 04:39 |                    |
| Air Blank         | 0.000 | 04:40 |                    |
| Subject Sample    | # 3   | 0.106 | 04:42              |
| Breath Volume     | 1.390 | 04:42 |                    |
| Air Blank         | 0.000 | 04:43 |                    |
| Control Test      | 0.079 | 04:43 |                    |
| Air Blank         | 0.000 | 04:44 |                    |
| Diagnostics Check | OK    | 04:44 |                    |

---

Data Download Date/Time: 04/13/2021 11:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ARMSTRONG,VICTOR  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: ABAD

DATE/TIME OF TEST: 04/10/2021 05:03:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:56 |
| Air Blank          | 0.000 | 04:57 |
| Control Test       | 0.080 | 04:57 |
| Air Blank          | 0.000 | 04:58 |
| Subject Sample # 1 | 0.147 | 04:59 |
| Breath Volume      | 1.441 | 04:59 |
| Air Blank          | 0.000 | 04:59 |
| Air Blank          | 0.000 | 05:01 |
| Subject Sample # 2 | 0.141 | 05:02 |
| Breath Volume      | 1.183 | 05:02 |
| Air Blank          | 0.000 | 05:02 |
| Control Test       | 0.080 | 05:03 |
| Air Blank          | 0.000 | 05:03 |
| Diagnostics Check  | OK    | 05:03 |

---

Data Download Date/Time: 04/13/2021 11:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POPICK,SCOTT,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LYDEN,C  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: BETETA SILVA

DATE/TIME OF TEST: 04/10/2021 07:15:33

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 07:02 |                |
| Air Blank         | 0.000     | 07:03 |                |
| Control Test      | 0.081     | 07:03 |                |
| Air Blank         | 0.000     | 07:04 |                |
| Subject Sample    | # 1 0.184 | 07:07 | Volume Not Met |
| Breath Volume     | 0.945     | 07:07 |                |
| Air Blank         | 0.000     | 07:07 |                |
| Air Blank         | 0.000     | 07:09 |                |
| Subject Sample    | # 2 0.184 | 07:10 |                |
| Breath Volume     | 1.589     | 07:10 |                |
| Air Blank         | 0.000     | 07:10 |                |
| Air Blank         | 0.000     | 07:12 |                |
| Subject Sample    | # 3 0.195 | 07:13 |                |
| Breath Volume     | 1.945     | 07:13 |                |
| Air Blank         | 0.000     | 07:14 |                |
| Control Test      | 0.081     | 07:14 |                |
| Air Blank         | 0.000     | 07:15 |                |
| Diagnostics Check | OK        | 07:15 |                |

---

Data Download Date/Time: 04/13/2021 11:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: GUZMAN,JOSE  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: ARAUJO DE OLIVEIRA

DATE/TIME OF TEST: 04/11/2021 03:57:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:51 |
| Air Blank          | 0.000 | 03:51 |
| Control Test       | 0.081 | 03:51 |
| Air Blank          | 0.000 | 03:52 |
| Subject Sample # 1 | 0.155 | 03:52 |
| Breath Volume      | 1.210 | 03:52 |
| Air Blank          | 0.000 | 03:53 |
| Air Blank          | 0.000 | 03:55 |
| Subject Sample # 2 | 0.157 | 03:56 |
| Breath Volume      | 1.589 | 03:56 |
| Air Blank          | 0.000 | 03:57 |
| Control Test       | 0.082 | 03:57 |
| Air Blank          | 0.000 | 03:57 |
| Diagnostics Check  | OK    | 03:57 |

---

Data Download Date/Time: 04/13/2021 11:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: HALES,ANDREW  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 04/16/2021 23:32:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:26 |
| Air Blank          | 0.000 | 23:26 |
| Control Test       | 0.082 | 23:27 |
| Air Blank          | 0.000 | 23:27 |
| Subject Sample # 1 | 0.000 | 23:28 |
| Breath Volume      | 1.792 | 23:28 |
| Air Blank          | 0.000 | 23:29 |
| Air Blank          | 0.000 | 23:31 |
| Subject Sample # 2 | 0.000 | 23:31 |
| Breath Volume      | 1.933 | 23:31 |
| Air Blank          | 0.000 | 23:31 |
| Control Test       | 0.081 | 23:32 |
| Air Blank          | 0.000 | 23:32 |
| Diagnostics Check  | OK    | 23:32 |

---

Data Download Date/Time: 05/12/2021 10:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,LA'TECE,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: SELF,DAULTON  
ARRESTING OFFICER AGENCY: Seminole Dept of LE  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: ST. LOUIS

DATE/TIME OF TEST: 04/18/2021 05:40:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:34 |
| Air Blank          | 0.000 | 05:34 |
| Control Test       | 0.081 | 05:35 |
| Air Blank          | 0.000 | 05:35 |
| Subject Sample # 1 | 0.217 | 05:36 |
| Breath Volume      | 1.738 | 05:36 |
| Air Blank          | 0.000 | 05:36 |
| Air Blank          | 0.000 | 05:38 |
| Subject Sample # 2 | 0.218 | 05:39 |
| Breath Volume      | 1.980 | 05:39 |
| Air Blank          | 0.000 | 05:39 |
| Control Test       | 0.081 | 05:40 |
| Air Blank          | 0.000 | 05:40 |
| Diagnostics Check  | OK    | 05:40 |

---

Data Download Date/Time: 05/12/2021 10:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LEZAMA,JOMARIE  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: HAHN

DATE/TIME OF TEST: 04/21/2021 21:03:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:57 |
| Air Blank          | 0.000 | 20:58 |
| Control Test       | 0.081 | 20:58 |
| Air Blank          | 0.000 | 20:58 |
| Subject Sample # 1 | 0.000 | 20:59 |
| Breath Volume      | 2.136 | 20:59 |
| Air Blank          | 0.000 | 20:59 |
| Air Blank          | 0.000 | 21:01 |
| Subject Sample # 2 | 0.000 | 21:02 |
| Breath Volume      | 1.910 | 21:02 |
| Air Blank          | 0.000 | 21:03 |
| Control Test       | 0.080 | 21:03 |
| Air Blank          | 0.000 | 21:03 |
| Diagnostics Check  | OK    | 21:03 |

---

Data Download Date/Time: 05/12/2021 10:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ARMSTRONG,BRITTANY  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/22/2021 01:36:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:29 |
| Air Blank          | 0.000 | 01:30 |
| Control Test       | 0.081 | 01:30 |
| Air Blank          | 0.000 | 01:30 |
| Subject Sample # 1 | 0.000 | 01:31 |
| Breath Volume      | 2.445 | 01:31 |
| Air Blank          | 0.000 | 01:32 |
| Air Blank          | 0.000 | 01:34 |
| Subject Sample # 2 | 0.000 | 01:35 |
| Breath Volume      | 1.597 | 01:35 |
| Air Blank          | 0.000 | 01:35 |
| Control Test       | 0.080 | 01:35 |
| Air Blank          | 0.000 | 01:36 |
| Diagnostics Check  | OK    | 01:36 |

---

Data Download Date/Time: 05/12/2021 10:15



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: DUKE,CASEY  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: GUERRIER

DATE/TIME OF TEST: 04/22/2021 04:04:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:56 |
| Air Blank          | 0.000 | 03:57 |
| Control Test       | 0.081 | 03:57 |
| Air Blank          | 0.000 | 03:57 |
| Subject Sample # 1 | 0.159 | 03:59 |
| Breath Volume      | 1.210 | 03:59 |
| Air Blank          | 0.000 | 04:00 |
| Air Blank          | 0.000 | 04:01 |
| Subject Sample # 2 | 0.157 | 04:03 |
| Breath Volume      | 1.550 | 04:03 |
| Air Blank          | 0.000 | 04:03 |
| Control Test       | 0.079 | 04:04 |
| Air Blank          | 0.000 | 04:04 |
| Diagnostics Check  | OK    | 04:04 |

---

Data Download Date/Time: 05/12/2021 10:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: DIAZ,DANNY  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 13:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:06

SUBJECT'S LAST NAME: CATALDO

DATE/TIME OF TEST: 04/22/2021 14:43:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:37 |
| Air Blank          | 0.000 | 14:37 |
| Control Test       | 0.082 | 14:38 |
| Air Blank          | 0.000 | 14:38 |
| Subject Sample # 1 | 0.000 | 14:39 |
| Breath Volume      | 2.078 | 14:39 |
| Air Blank          | 0.000 | 14:39 |
| Air Blank          | 0.000 | 14:41 |
| Subject Sample # 2 | 0.000 | 14:42 |
| Breath Volume      | 1.109 | 14:42 |
| Air Blank          | 0.000 | 14:42 |
| Control Test       | 0.081 | 14:42 |
| Air Blank          | 0.000 | 14:43 |
| Diagnostics Check  | OK    | 14:43 |

---

Data Download Date/Time: 05/12/2021 10:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: MARLEY,KEVIN  
ARRESTING OFFICER AGENCY: Plantation PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: HASKINS

DATE/TIME OF TEST: 04/23/2021 23:59:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:52 |
| Air Blank          | 0.000 | 23:53 |
| Control Test       | 0.081 | 23:53 |
| Air Blank          | 0.000 | 23:53 |
| Subject Sample # 1 | 0.196 | 23:54 |
| Breath Volume      | 1.625 | 23:54 |
| Air Blank          | 0.000 | 23:55 |
| Air Blank          | 0.000 | 23:57 |
| Subject Sample # 2 | 0.211 | 23:57 |
| Breath Volume      | 2.640 | 23:57 |
| Air Blank          | 0.000 | 23:58 |
| Control Test       | 0.079 | 23:59 |
| Air Blank          | 0.000 | 23:59 |
| Diagnostics Check  | OK    | 23:59 |

---

Data Download Date/Time: 05/12/2021 10:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ARMSTRONG,BRITTANY  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: ALMEIDA

DATE/TIME OF TEST: 04/24/2021 00:42:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:35 |
| Air Blank          | 0.000 | 00:36 |
| Control Test       | 0.081 | 00:36 |
| Air Blank          | 0.000 | 00:37 |
| Subject Sample # 1 | 0.000 | 00:38 |
| Breath Volume      | 1.945 | 00:38 |
| Air Blank          | 0.000 | 00:38 |
| Air Blank          | 0.000 | 00:40 |
| Subject Sample # 2 | 0.000 | 00:40 |
| Breath Volume      | 2.437 | 00:40 |
| Air Blank          | 0.000 | 00:41 |
| Control Test       | 0.081 | 00:41 |
| Air Blank          | 0.000 | 00:42 |
| Diagnostics Check  | OK    | 00:42 |

---

Data Download Date/Time: 05/12/2021 10:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: ROMERO,ELISA  
ARRESTING OFFICER AGENCY: Miramar PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: ZEPHIRIN

DATE/TIME OF TEST: 04/24/2021 02:44:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:38 |
| Air Blank          | 0.000 | 02:38 |
| Control Test       | 0.081 | 02:39 |
| Air Blank          | 0.000 | 02:39 |
| Subject Sample # 1 | 0.074 | 02:40 |
| Breath Volume      | 1.582 | 02:40 |
| Air Blank          | 0.000 | 02:40 |
| Air Blank          | 0.000 | 02:42 |
| Subject Sample # 2 | 0.075 | 02:43 |
| Breath Volume      | 2.410 | 02:43 |
| Air Blank          | 0.000 | 02:43 |
| Control Test       | 0.081 | 02:44 |
| Air Blank          | 0.000 | 02:44 |
| Diagnostics Check  | OK    | 02:44 |

---

Data Download Date/Time: 05/12/2021 10:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: HALES,ANDREW  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: HOYT

DATE/TIME OF TEST: 04/24/2021 04:31:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:25 |
| Air Blank          | 0.000 | 04:25 |
| Control Test       | 0.082 | 04:26 |
| Air Blank          | 0.000 | 04:26 |
| Subject Sample # 1 | 0.000 | 04:27 |
| Breath Volume      | 2.738 | 04:27 |
| Air Blank          | 0.000 | 04:27 |
| Air Blank          | 0.000 | 04:29 |
| Subject Sample # 2 | 0.000 | 04:29 |
| Breath Volume      | 2.832 | 04:29 |
| Air Blank          | 0.000 | 04:30 |
| Control Test       | 0.080 | 04:30 |
| Air Blank          | 0.000 | 04:31 |
| Diagnostics Check  | OK    | 04:31 |

---

Data Download Date/Time: 05/12/2021 10:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,LA'TECE,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LUC,NORBENS  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:37

SUBJECT'S LAST NAME: HONORE

DATE/TIME OF TEST: 04/24/2021 18:11:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:03 |
| Air Blank          | 0.000 | 18:04 |
| Control Test       | 0.082 | 18:04 |
| Air Blank          | 0.000 | 18:04 |
| Subject Sample # 1 | 0.244 | 18:05 |
| Breath Volume      | 1.835 | 18:05 |
| Air Blank          | 0.000 | 18:06 |
| Air Blank          | 0.000 | 18:08 |
| Subject Sample # 2 | 0.251 | 18:09 |
| Breath Volume      | 1.828 | 18:09 |
| Air Blank          | 0.000 | 18:10 |
| Control Test       | 0.082 | 18:10 |
| Air Blank          | 0.000 | 18:11 |
| Diagnostics Check  | OK    | 18:11 |

---

Data Download Date/Time: 05/12/2021 10:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: FANFAN,KUSHNER  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 04/25/2021 00:01:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:54 |
| Air Blank          | 0.000 | 23:54 |
| Control Test       | 0.081 | 23:55 |
| Air Blank          | 0.000 | 23:55 |
| Subject Sample # 1 | 0.203 | 23:56 |
| Breath Volume      | 1.304 | 23:56 |
| Air Blank          | 0.000 | 23:57 |
| Air Blank          | 0.000 | 23:59 |
| Subject Sample # 2 | 0.198 | 23:59 |
| Breath Volume      | 1.117 | 23:59 |
| Air Blank          | 0.000 | 00:00 |
| Control Test       | 0.081 | 00:00 |
| Air Blank          | 0.000 | 00:01 |
| Diagnostics Check  | OK    | 00:01 |

---

Data Download Date/Time: 05/12/2021 10:16



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,LA'TECE,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: HARO,DANNY  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 14:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:11

SUBJECT'S LAST NAME: CLEMENTE

DATE/TIME OF TEST: 04/26/2021 16:38:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:31 |
| Air Blank          | 0.000 | 16:32 |
| Control Test       | 0.081 | 16:32 |
| Air Blank          | 0.000 | 16:32 |
| Subject Sample # 1 | 0.200 | 16:33 |
| Breath Volume      | 2.402 | 16:33 |
| Air Blank          | 0.000 | 16:34 |
| Air Blank          | 0.000 | 16:35 |
| Subject Sample # 2 | 0.194 | 16:36 |
| Breath Volume      | 2.257 | 16:36 |
| Air Blank          | 0.000 | 16:37 |
| Control Test       | 0.081 | 16:37 |
| Air Blank          | 0.000 | 16:38 |
| Diagnostics Check  | OK    | 16:38 |

---

Data Download Date/Time: 05/12/2021 10:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: HALES,ANDREW  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/29/2021 23:52:30

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 23:40 |                |
| Air Blank         | 0.000     | 23:41 |                |
| Control Test      | 0.082     | 23:41 |                |
| Air Blank         | 0.000     | 23:41 |                |
| Subject Sample    | # 1 0.196 | 23:45 | Volume Not Met |
| Breath Volume     | 0.894     | 23:45 |                |
| Air Blank         | 0.000     | 23:46 |                |
| Air Blank         | 0.000     | 23:47 |                |
| Subject Sample    | # 2 0.188 | 23:50 | Volume Not Met |
| Breath Volume     | 0.921     | 23:50 |                |
| Air Blank         | 0.000     | 23:51 |                |
| Control Test      | 0.079     | 23:51 |                |
| Air Blank         | 0.000     | 23:52 |                |
| Diagnostics Check | OK        | 23:52 |                |

---

Data Download Date/Time: 05/12/2021 10:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: MUSCILLO,GINA  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: MONNAR

DATE/TIME OF TEST: 04/30/2021 04:47:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:38 |
| Air Blank          | 0.000 | 04:38 |
| Control Test       | 0.080 | 04:39 |
| Air Blank          | 0.000 | 04:39 |
| Subject Sample # 1 | 0.210 | 04:42 |
| Breath Volume      | 2.554 | 04:42 |
| Air Blank          | 0.000 | 04:43 |
| Air Blank          | 0.000 | 04:44 |
| Subject Sample # 2 | 0.208 | 04:45 |
| Breath Volume      | 2.328 | 04:45 |
| Air Blank          | 0.000 | 04:46 |
| Control Test       | 0.078 | 04:46 |
| Air Blank          | 0.000 | 04:47 |
| Diagnostics Check  | OK    | 04:47 |

---

Data Download Date/Time: 05/12/2021 10:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: FARQUHARSON,KEVIN  
ARRESTING OFFICER AGENCY: Plantation PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: EGGLESTON

DATE/TIME OF TEST: 04/30/2021 06:28:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:19 |
| Air Blank          | 0.000 | 06:20 |
| Control Test       | 0.081 | 06:20 |
| Air Blank          | 0.000 | 06:21 |
| Subject Sample # 1 | 0.175 | 06:22 |
| Breath Volume      | 1.464 | 06:22 |
| Air Blank          | 0.000 | 06:23 |
| Air Blank          | 0.000 | 06:25 |
| Subject Sample # 2 | 0.182 | 06:26 |
| Breath Volume      | 1.578 | 06:26 |
| Air Blank          | 0.000 | 06:27 |
| Control Test       | 0.081 | 06:27 |
| Air Blank          | 0.000 | 06:28 |
| Diagnostics Check  | OK    | 06:28 |

---

Data Download Date/Time: 05/12/2021 10:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS  
ARRESTING OFFICER AGENCY: Miramar PD  
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: MCLEAN

DATE/TIME OF TEST: 04/01/2021 05:17:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:10 |
| Air Blank          | 0.000 | 05:11 |
| Control Test       | 0.081 | 05:11 |
| Air Blank          | 0.000 | 05:12 |
| Subject Sample # 1 | 0.011 | 05:12 |
| Breath Volume      | 3.156 | 05:12 |
| Air Blank          | 0.000 | 05:13 |
| Air Blank          | 0.000 | 05:15 |
| Subject Sample # 2 | 0.009 | 05:15 |
| Breath Volume      | 2.902 | 05:15 |
| Air Blank          | 0.000 | 05:16 |
| Control Test       | 0.080 | 05:16 |
| Air Blank          | 0.000 | 05:17 |
| Diagnostics Check  | OK    | 05:17 |

---

Data Download Date/Time: 04/13/2021 11:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: O'DOR,ROBERT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: NEMETH

DATE/TIME OF TEST: 04/02/2021 20:25:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:18 |
| Air Blank          | 0.000 | 20:18 |
| Control Test       | 0.081 | 20:18 |
| Air Blank          | 0.000 | 20:19 |
| Subject Sample # 1 | 0.138 | 20:21 |
| Breath Volume      | 2.539 | 20:21 |
| Air Blank          | 0.000 | 20:21 |
| Air Blank          | 0.000 | 20:23 |
| Subject Sample # 2 | 0.146 | 20:24 |
| Breath Volume      | 2.921 | 20:24 |
| Air Blank          | 0.000 | 20:24 |
| Control Test       | 0.079 | 20:25 |
| Air Blank          | 0.000 | 20:25 |
| Diagnostics Check  | OK    | 20:25 |

---

Data Download Date/Time: 04/13/2021 11:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: POPICK,SCOTT  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BARBOSA JOAZEIRO

DATE/TIME OF TEST: 04/03/2021 03:40:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:32 |
| Air Blank          | 0.000 | 03:32 |
| Control Test       | 0.081 | 03:33 |
| Air Blank          | 0.000 | 03:33 |
| Subject Sample # 1 | 0.166 | 03:35 |
| Breath Volume      | 1.832 | 03:35 |
| Air Blank          | 0.000 | 03:36 |
| Air Blank          | 0.000 | 03:38 |
| Subject Sample # 2 | 0.161 | 03:39 |
| Breath Volume      | 1.707 | 03:39 |
| Air Blank          | 0.000 | 03:39 |
| Control Test       | 0.079 | 03:40 |
| Air Blank          | 0.000 | 03:40 |
| Diagnostics Check  | OK    | 03:40 |

---

Data Download Date/Time: 04/13/2021 11:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: MASEILLE,BRYANT  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: GANI

DATE/TIME OF TEST: 04/03/2021 05:08:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:02 |
| Air Blank          | 0.000 | 05:02 |
| Control Test       | 0.081 | 05:02 |
| Air Blank          | 0.000 | 05:03 |
| Subject Sample # 1 | 0.197 | 05:03 |
| Breath Volume      | 2.003 | 05:03 |
| Air Blank          | 0.000 | 05:04 |
| Air Blank          | 0.000 | 05:06 |
| Subject Sample # 2 | 0.199 | 05:06 |
| Breath Volume      | 1.898 | 05:06 |
| Air Blank          | 0.000 | 05:07 |
| Control Test       | 0.081 | 05:07 |
| Air Blank          | 0.000 | 05:08 |
| Diagnostics Check  | OK    | 05:08 |

---

Data Download Date/Time: 04/13/2021 11:15



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: LOPEZ MUNOZ,ANDRES  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/04/2021 17:13:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:05 |
| Air Blank          | 0.000 | 17:06 |
| Control Test       | 0.081 | 17:06 |
| Air Blank          | 0.000 | 17:07 |
| Subject Sample # 1 | 0.185 | 17:08 |
| Breath Volume      | 2.003 | 17:08 |
| Air Blank          | 0.000 | 17:08 |
| Air Blank          | 0.000 | 17:10 |
| Subject Sample # 2 | 0.182 | 17:11 |
| Breath Volume      | 1.671 | 17:11 |
| Air Blank          | 0.000 | 17:12 |
| Control Test       | 0.081 | 17:12 |
| Air Blank          | 0.000 | 17:12 |
| Diagnostics Check  | OK    | 17:13 |

---

Data Download Date/Time: 04/13/2021 11:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAZIER,ANAYA,S  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: GUZMAN,JOSE  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ABELLO

DATE/TIME OF TEST: 04/06/2021 01:29:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:22 |
| Air Blank          | 0.000 | 01:22 |
| Control Test       | 0.081 | 01:23 |
| Air Blank          | 0.000 | 01:23 |
| Subject Sample # 1 | 0.221 | 01:24 |
| Breath Volume      | 2.886 | 01:24 |
| Air Blank          | 0.000 | 01:25 |
| Air Blank          | 0.000 | 01:26 |
| Subject Sample # 2 | 0.205 | 01:27 |
| Breath Volume      | 2.898 | 01:27 |
| Air Blank          | 0.000 | 01:28 |
| Control Test       | 0.080 | 01:28 |
| Air Blank          | 0.000 | 01:29 |
| Diagnostics Check  | OK    | 01:29 |

---

Data Download Date/Time: 04/13/2021 11:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: DIAZ,JASIEL  
ARRESTING OFFICER AGENCY: Hallandale Beach PD  
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: NOVIKOV

DATE/TIME OF TEST: 04/06/2021 22:25:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:17 |
| Air Blank          | 0.000 | 22:17 |
| Control Test       | 0.081 | 22:17 |
| Air Blank          | 0.000 | 22:18 |
| Subject Sample # 1 | 0.245 | 22:19 |
| Breath Volume      | 2.011 | 22:19 |
| Air Blank          | 0.000 | 22:20 |
| Air Blank          | 0.000 | 22:22 |
| Subject Sample # 2 | 0.249 | 22:23 |
| Breath Volume      | 1.148 | 22:23 |
| Air Blank          | 0.000 | 22:24 |
| Control Test       | 0.078 | 22:24 |
| Air Blank          | 0.000 | 22:25 |
| Diagnostics Check  | OK    | 22:25 |

---

Data Download Date/Time: 04/13/2021 11:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: GUZMAN,JOSE  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: MAPP

DATE/TIME OF TEST: 04/07/2021 01:49:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:42 |
| Air Blank          | 0.000 | 01:42 |
| Control Test       | 0.080 | 01:42 |
| Air Blank          | 0.000 | 01:43 |
| Subject Sample # 1 | 0.080 | 01:43 |
| Breath Volume      | 2.625 | 01:43 |
| Air Blank          | 0.000 | 01:44 |
| Air Blank          | 0.000 | 01:46 |
| Subject Sample # 2 | 0.081 | 01:47 |
| Breath Volume      | 2.832 | 01:47 |
| Air Blank          | 0.000 | 01:48 |
| Control Test       | 0.081 | 01:48 |
| Air Blank          | 0.000 | 01:49 |
| Diagnostics Check  | OK    | 01:49 |

---

Data Download Date/Time: 04/13/2021 11:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: CARTER,JAMES  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: SHECK

DATE/TIME OF TEST: 04/07/2021 21:44:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:37 |
| Air Blank          | 0.000 | 21:38 |
| Control Test       | 0.080 | 21:38 |
| Air Blank          | 0.000 | 21:39 |
| Subject Sample # 1 | 0.000 | 21:39 |
| Breath Volume      | 2.644 | 21:39 |
| Air Blank          | 0.000 | 21:40 |
| Air Blank          | 0.000 | 21:42 |
| Subject Sample # 2 | 0.000 | 21:42 |
| Breath Volume      | 1.179 | 21:42 |
| Air Blank          | 0.000 | 21:43 |
| Control Test       | 0.080 | 21:43 |
| Air Blank          | 0.000 | 21:44 |
| Diagnostics Check  | OK    | 21:44 |

---

Data Download Date/Time: 04/13/2021 11:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: BOHM,JONATHAN  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 04/08/2021 23:17:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:10 |
| Air Blank          | 0.000 | 23:10 |
| Control Test       | 0.080 | 23:11 |
| Air Blank          | 0.000 | 23:11 |
| Subject Sample # 1 | 0.169 | 23:12 |
| Breath Volume      | 1.746 | 23:12 |
| Air Blank          | 0.000 | 23:13 |
| Air Blank          | 0.000 | 23:15 |
| Subject Sample # 2 | 0.164 | 23:16 |
| Breath Volume      | 1.957 | 23:16 |
| Air Blank          | 0.000 | 23:16 |
| Control Test       | 0.078 | 23:17 |
| Air Blank          | 0.000 | 23:17 |
| Diagnostics Check  | OK    | 23:17 |

---

Data Download Date/Time: 04/13/2021 11:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TILLMAN,SHEILA,L  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: FARQUHARSON,KEVIN  
ARRESTING OFFICER AGENCY: Plantation PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: ZAMBRANO

DATE/TIME OF TEST: 04/10/2021 01:00:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:52 |
| Air Blank          | 0.000 | 00:52 |
| Control Test       | 0.081 | 00:53 |
| Air Blank          | 0.000 | 00:53 |
| Subject Sample # 1 | 0.193 | 00:56 |
| Breath Volume      | 1.480 | 00:56 |
| Air Blank          | 0.000 | 00:57 |
| Air Blank          | 0.000 | 00:58 |
| Subject Sample # 2 | 0.186 | 00:59 |
| Breath Volume      | 1.628 | 00:59 |
| Air Blank          | 0.000 | 00:59 |
| Control Test       | 0.078 | 01:00 |
| Air Blank          | 0.000 | 01:00 |
| Diagnostics Check  | OK    | 01:00 |

---

Data Download Date/Time: 04/13/2021 11:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: BOHM,JONATHAN  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: MOJICA RODRIGUEZ

DATE/TIME OF TEST: 04/10/2021 03:23:24

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 03:11 |                    |
| Air Blank         | 0.000 | 03:12 |                    |
| Control Test      | 0.077 | 03:12 |                    |
| Air Blank         | 0.000 | 03:13 |                    |
| Subject Sample    | # 1   | 03:16 | No Sample Provided |
| Breath Volume     | 0.000 | 03:16 |                    |
| Air Blank         | 0.000 | 03:16 |                    |
| Air Blank         | 0.000 | 03:18 |                    |
| Subject Sample    | # 2   | 03:22 | No Sample Provided |
| Breath Volume     | 0.000 | 03:22 |                    |
| Air Blank         | 0.000 | 03:22 |                    |
| Control Test      | 0.077 | 03:22 |                    |
| Air Blank         | 0.000 | 03:23 |                    |
| Diagnostics Check | OK    | 03:23 |                    |

---

Data Download Date/Time: 04/13/2021 11:16



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,SANDRA,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: BOHM,JONATHAN  
ARRESTING OFFICER AGENCY: Fort Lauderdale PD  
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1157961  
DRY GAS EXPIRY DATE: 06/17/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: MOJICA RODRIGUEZ

DATE/TIME OF TEST: 04/10/2021 03:32:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:26 |
| Air Blank          | 0.000 | 03:26 |
| Control Test       | 0.078 | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Subject Sample # 1 | 0.233 | 03:27 |
| Breath Volume      | 1.343 | 03:27 |
| Air Blank          | 0.000 | 03:28 |
| Air Blank          | 0.000 | 03:30 |
| Subject Sample # 2 | 0.220 | 03:30 |
| Breath Volume      | 1.546 | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Control Test       | 0.078 | 03:31 |
| Air Blank          | 0.000 | 03:32 |
| Diagnostics Check  | OK    | 03:32 |

---

Data Download Date/Time: 04/13/2021 11:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAIG,SHAWN,T  
BREATH TEST OPERATOR AGENCY: Live Oak PD  
ARRESTING OFFICER: CRAIG,SHAWN  
ARRESTING OFFICER AGENCY: Live Oak PD  
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: DIAZ ALTAMIRANO

DATE/TIME OF TEST: 04/17/2021 19:25:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:19 |
| Air Blank          | 0.000 | 19:19 |
| Control Test       | 0.079 | 19:19 |
| Air Blank          | 0.000 | 19:20 |
| Subject Sample # 1 | 0.233 | 19:20 |
| Breath Volume      | 2.570 | 19:20 |
| Air Blank          | 0.000 | 19:21 |
| Air Blank          | 0.000 | 19:23 |
| Subject Sample # 2 | 0.229 | 19:23 |
| Breath Volume      | 2.679 | 19:23 |
| Air Blank          | 0.000 | 19:24 |
| Control Test       | 0.077 | 19:25 |
| Air Blank          | 0.000 | 19:25 |
| Diagnostics Check  | OK    | 19:25 |

---

Data Download Date/Time: 05/06/2021 11:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARRS,RUSSELL,B  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: HOWARD,ROGER  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/18/2021 00:23:36

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:20 |                      |
| Air Blank         | 0.000 | 00:21 |                      |
| Control Test      | 0.080 | 00:21 |                      |
| Air Blank         | 0.000 | 00:22 |                      |
| Subject Sample    | # 1   | 00:22 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:22 |                      |
| Air Blank         | 0.000 | 00:22 |                      |
| Control Test      | 0.080 | 00:22 |                      |
| Air Blank         | 0.000 | 00:23 |                      |
| Diagnostics Check | OK    | 00:23 |                      |

---

Data Download Date/Time: 05/06/2021 11:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,DANIEL,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MARTIN,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/18/2021 04:51:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:45 |
| Air Blank          | 0.000 | 04:46 |
| Control Test       | 0.080 | 04:46 |
| Air Blank          | 0.000 | 04:46 |
| Subject Sample # 1 | 0.000 | 04:47 |
| Breath Volume      | 1.925 | 04:47 |
| Air Blank          | 0.000 | 04:47 |
| Air Blank          | 0.000 | 04:49 |
| Subject Sample # 2 | 0.000 | 04:50 |
| Breath Volume      | 2.472 | 04:50 |
| Air Blank          | 0.000 | 04:50 |
| Control Test       | 0.080 | 04:51 |
| Air Blank          | 0.000 | 04:51 |
| Diagnostics Check  | OK    | 04:51 |

---

Data Download Date/Time: 05/06/2021 11:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAIG,SHAWN,T  
BREATH TEST OPERATOR AGENCY: Live Oak PD  
ARRESTING OFFICER: CRAIG,SHAWN  
ARRESTING OFFICER AGENCY: Live Oak PD  
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: FOLEY

DATE/TIME OF TEST: 04/19/2021 02:03:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:56 |
| Air Blank          | 0.000 | 01:57 |
| Control Test       | 0.080 | 01:57 |
| Air Blank          | 0.000 | 01:58 |
| Subject Sample # 1 | 0.262 | 01:58 |
| Breath Volume      | 2.195 | 01:58 |
| Air Blank          | 0.000 | 01:59 |
| Air Blank          | 0.000 | 02:00 |
| Subject Sample # 2 | 0.257 | 02:01 |
| Breath Volume      | 1.925 | 02:01 |
| Air Blank          | 0.000 | 02:02 |
| Control Test       | 0.078 | 02:02 |
| Air Blank          | 0.000 | 02:03 |
| Diagnostics Check  | OK    | 02:03 |

---

Data Download Date/Time: 05/06/2021 11:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MACFADDEN,RYAN,E  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MACFADDEN,RYAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: VANALLER

DATE/TIME OF TEST: 04/20/2021 00:31:33

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 00:27 |            |
| Air Blank         | 0.000     | 00:28 |            |
| Control Test      | 0.081     | 00:28 |            |
| Air Blank         | 0.000     | 00:29 |            |
| Subject Sample    | # 1 0.180 | 00:30 |            |
| Breath Volume     | 2.687     | 00:30 |            |
| Air Blank         |           | 00:31 | Purge Fail |
| Air Blank         | 0.000     | 00:31 |            |

---

Data Download Date/Time: 05/06/2021 11:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MACFADDEN,RYAN,E  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MACFADDEN,RYAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: VANALLER

DATE/TIME OF TEST: 04/20/2021 00:40:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:34 |
| Air Blank          | 0.000 | 00:34 |
| Control Test       | 0.076 | 00:35 |
| Air Blank          | 0.000 | 00:35 |
| Subject Sample # 1 | 0.168 | 00:36 |
| Breath Volume      | 1.582 | 00:36 |
| Air Blank          | 0.000 | 00:36 |
| Air Blank          | 0.000 | 00:38 |
| Subject Sample # 2 | 0.164 | 00:38 |
| Breath Volume      | 1.375 | 00:38 |
| Air Blank          | 0.000 | 00:39 |
| Control Test       | 0.076 | 00:39 |
| Air Blank          | 0.000 | 00:40 |
| Diagnostics Check  | OK    | 00:40 |

---

Data Download Date/Time: 05/06/2021 11:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAIG,SHAWN,T  
BREATH TEST OPERATOR AGENCY: Live Oak PD  
ARRESTING OFFICER: CRAIG,SHAWN  
ARRESTING OFFICER AGENCY: Live Oak PD  
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: NERIA

DATE/TIME OF TEST: 04/30/2021 23:52:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:45 |
| Air Blank          | 0.000 | 23:45 |
| Control Test       | 0.080 | 23:46 |
| Air Blank          | 0.000 | 23:46 |
| Subject Sample # 1 | 0.199 | 23:47 |
| Breath Volume      | 3.304 | 23:47 |
| Air Blank          | 0.000 | 23:48 |
| Air Blank          | 0.000 | 23:50 |
| Subject Sample # 2 | 0.189 | 23:50 |
| Breath Volume      | 2.988 | 23:50 |
| Air Blank          | 0.000 | 23:51 |
| Control Test       | 0.075 | 23:51 |
| Air Blank          | 0.000 | 23:52 |
| Diagnostics Check  | OK    | 23:52 |

---

Data Download Date/Time: 05/06/2021 11:54