

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: MARTIN DIAZ

DATE/TIME OF TEST: 05/01/2021 21:19:28

Diagnostics Check	OK	21:11
Air Blank	0.000	21:11
Control Test	0.078	21:12
Air Blank	0.000	21:12
Subject Sample # 1	0.262	21:13
Breath Volume	2.003	21:13
Air Blank	0.000	21:14
Air Blank	0.000	21:16
Subject Sample # 2	0.260	21:17
Breath Volume	1.453	21:17
Air Blank	0.000	21:18
Control Test	0.077	21:18
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

Data Download Date/Time: 05/26/2021 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: PEREZ LOPEZ

DATE/TIME OF TEST: 05/06/2021 01:50:55

Diagnostics Check		OK	01:44
Air Blank		0.000	01:45
Control Test		0.080	01:45
Air Blank		0.000	01:45
Subject Sample	# 1	0.234	01:46
Breath Volume		2.417	01:46
Air Blank		0.000	01:47
Air Blank		0.000	01:48
Subject Sample	# 2	0.227	01:49
Breath Volume		2.449	01:49
Air Blank		0.000	01:50
Control Test		0.080	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 05/26/2021 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: MARTIN GODINEZ

DATE/TIME OF TEST: 05/07/2021 03:04:45

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:02	
Control Test	0.082	03:02	
Air Blank	0.000	03:03	
Subject Sample	# 1	03:03	Subject Test Refused
Breath Volume	0.000	03:03	
Air Blank	0.000	03:03	
Control Test	0.079	03:04	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:04	

Data Download Date/Time: 05/26/2021 14:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: RODRIGUEZ MENESES

DATE/TIME OF TEST: 05/08/2021 03:22:51

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:20	
Control Test	0.080	03:20	
Air Blank	0.000	03:21	
Subject Sample	# 1	03:21	Subject Test Refused
Breath Volume	0.000	03:21	
Air Blank	0.000	03:21	
Control Test	0.080	03:22	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

Data Download Date/Time: 05/26/2021 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: MATILLAS CRUZ

DATE/TIME OF TEST: 05/08/2021 18:53:48

Diagnostics Check	OK	18:46
Air Blank	0.000	18:47
Control Test	0.080	18:47
Air Blank	0.000	18:48
Subject Sample # 1	0.241	18:49
Breath Volume	1.617	18:49
Air Blank	0.000	18:49
Air Blank	0.000	18:51
Subject Sample # 2	0.245	18:52
Breath Volume	1.976	18:52
Air Blank	0.000	18:52
Control Test	0.078	18:53
Air Blank	0.000	18:53
Diagnostics Check	OK	18:53

Data Download Date/Time: 05/26/2021 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURGUIDO,ALEX,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: IRBY,DOMINIC
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: BERNAL

DATE/TIME OF TEST: 05/11/2021 22:28:40

Diagnostics Check	OK	22:22	
Air Blank	0.000	22:22	
Control Test	0.081	22:23	
Air Blank	0.000	22:23	
Subject Sample	# 1 0.156	22:24	
Breath Volume	3.058	22:24	
Air Blank	0.000	22:25	
Air Blank	0.000	22:27	
Subject Sample	# 2	22:27	RFI Detect
Breath Volume	0.000	22:27	
Air Blank	0.000	22:27	
Control Test	0.077	22:28	
Air Blank	0.000	22:28	
Diagnostics Check	OK	22:28	

Data Download Date/Time: 05/26/2021 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURGUIDO,ALEX,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: IRBY,DOMINIC
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: BERNAL

DATE/TIME OF TEST: 05/11/2021 22:37:34

Diagnostics Check	OK	22:30
Air Blank	0.000	22:31
Control Test	0.078	22:31
Air Blank	0.000	22:32
Subject Sample # 1	0.152	22:32
Breath Volume	2.945	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.142	22:35
Breath Volume	2.496	22:35
Air Blank	0.000	22:36
Control Test	0.076	22:36
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 05/26/2021 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 05/13/2021 04:00:46

Diagnostics Check	OK	03:54
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.085	03:56
Breath Volume	3.058	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:58
Subject Sample # 2	0.082	03:59
Breath Volume	2.597	03:59
Air Blank	0.000	03:59
Control Test	0.079	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 05/26/2021 14:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: ANDINO BAUTISTA

DATE/TIME OF TEST: 05/14/2021 01:51:10

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:48	
Control Test	0.080	01:49	
Air Blank	0.000	01:49	
Subject Sample	# 1	01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:50	
Control Test	0.080	01:50	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 05/26/2021 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: HERNANDEZ GREGORIO

DATE/TIME OF TEST: 05/15/2021 02:05:24

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.080	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.168	02:00
Breath Volume	1.414	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.167	02:03
Breath Volume	1.222	02:03
Air Blank	0.000	02:04
Control Test	0.078	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 05/26/2021 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,DARRYL,A
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: MAYS,DARRYL
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 14:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: CHILEL NIZ

DATE/TIME OF TEST: 05/16/2021 15:18:57

Diagnostics Check	OK	15:09
Air Blank	0.000	15:10
Control Test	0.079	15:10
Air Blank	0.000	15:11
Subject Sample # 1	0.336	15:11
Breath Volume	2.453	15:11
Air Blank	0.000	15:12
Air Blank	0.000	15:13
Subject Sample # 2	0.315	15:14
Breath Volume	1.687	15:14
Air Blank	0.000	15:15
Air Blank	0.000	15:16
Subject Sample # 3	0.356	15:17
Breath Volume	1.640	15:17
Air Blank	0.000	15:18
Control Test	0.078	15:18
Air Blank	0.000	15:18
Diagnostics Check	OK	15:18

Data Download Date/Time: 05/26/2021 14:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: YEATTS

DATE/TIME OF TEST: 05/18/2021 06:03:11

Diagnostics Check	OK	06:00	
Air Blank	0.000	06:00	
Control Test	0.080	06:01	
Air Blank	0.000	06:01	
Subject Sample	# 1	06:01	Subject Test Refused
Breath Volume	0.000	06:01	
Air Blank	0.000	06:02	
Control Test	0.079	06:02	
Air Blank	0.000	06:03	
Diagnostics Check	OK	06:03	

Data Download Date/Time: 05/26/2021 14:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: PERDO

DATE/TIME OF TEST: 05/19/2021 03:19:23

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:17	
Control Test	0.080	03:17	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:18	Subject Test Refused
Breath Volume	0.000	03:18	
Air Blank	0.000	03:18	
Control Test	0.080	03:18	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 05/26/2021 14:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: LOPEZ DENIS

DATE/TIME OF TEST: 05/20/2021 04:54:51

Diagnostics Check	OK	04:52	
Air Blank	0.000	04:52	
Control Test	0.080	04:52	
Air Blank	0.000	04:53	
Subject Sample	# 1	04:53	Subject Test Refused
Breath Volume	0.000	04:53	
Air Blank	0.000	04:53	
Control Test	0.080	04:54	
Air Blank	0.000	04:54	
Diagnostics Check	OK	04:54	

Data Download Date/Time: 05/26/2021 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: TROCONIS

DATE/TIME OF TEST: 05/21/2021 01:56:07

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:54	
Control Test	0.080	01:55	
Air Blank		01:55	Ambient Fail
Air Blank	0.000	01:56	

Data Download Date/Time: 05/26/2021 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: TROCONIS

DATE/TIME OF TEST: 05/21/2021 02:03:59

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.000	01:59
Breath Volume	1.214	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.000	02:02
Breath Volume	1.968	02:02
Air Blank	0.000	02:03
Control Test	0.081	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 05/26/2021 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: ROMERO DIAZ

DATE/TIME OF TEST: 05/21/2021 22:13:35

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.080	22:05
Air Blank	0.000	22:06
Subject Sample # 1	0.216	22:08
Breath Volume	3.132	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:11
Subject Sample # 2	0.213	22:11
Breath Volume	2.148	22:11
Air Blank	0.000	22:12
Control Test	0.079	22:12
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

Data Download Date/Time: 05/26/2021 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MORIARTY

DATE/TIME OF TEST: 05/22/2021 02:52:56

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:50	
Control Test	0.081	02:50	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:52	
Control Test	0.081	02:52	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 05/26/2021 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 05:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 05/22/2021 06:09:57

Diagnostics Check	OK	06:07	
Air Blank	0.000	06:07	
Control Test	0.079	06:07	
Air Blank	0.000	06:08	
Subject Sample	# 1	06:08	Subject Test Refused
Breath Volume	0.000	06:08	
Air Blank	0.000	06:09	
Control Test	0.079	06:09	
Air Blank	0.000	06:09	
Diagnostics Check	OK	06:09	

Data Download Date/Time: 05/26/2021 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: ARREOLA MORENO

DATE/TIME OF TEST: 05/22/2021 20:02:39

Diagnostics Check	OK	19:54	
Air Blank	0.000	19:54	
Control Test	0.080	19:54	
Air Blank	0.000	19:55	
Subject Sample	# 1	19:58	No Sample Provided
Breath Volume	0.000	19:58	
Air Blank	0.000	19:59	
Air Blank	0.000	20:01	
Subject Sample	# 2	20:01	Subject Test Refused
Breath Volume	0.000	20:01	
Air Blank	0.000	20:01	
Control Test	0.080	20:02	
Air Blank	0.000	20:02	
Diagnostics Check	OK	20:02	

Data Download Date/Time: 05/26/2021 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: RIVERA ENRRIQUEZ

DATE/TIME OF TEST: 05/27/2021 02:08:16

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:05	
Control Test	0.081	02:06	
Air Blank	0.000	02:06	
Subject Sample	# 1	02:06	Subject Test Refused
Breath Volume	0.000	02:06	
Air Blank	0.000	02:07	
Control Test	0.081	02:07	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

Data Download Date/Time: 06/30/2021 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: CIRIANO CARDONA

DATE/TIME OF TEST: 05/27/2021 04:53:22

Diagnostics Check		OK	04:47
Air Blank		0.000	04:47
Control Test		0.080	04:47
Air Blank		0.000	04:48
Subject Sample	# 1	0.231	04:48
Breath Volume		1.824	04:48
Air Blank		0.000	04:49
Air Blank		0.000	04:51
Subject Sample	# 2	0.244	04:51
Breath Volume		1.671	04:51
Air Blank		0.000	04:52
Control Test		0.078	04:52
Air Blank		0.000	04:53
Diagnostics Check		OK	04:53

Data Download Date/Time: 06/30/2021 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: MACEDO ROMAN

DATE/TIME OF TEST: 05/28/2021 01:03:31

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.080	00:57
Air Blank	0.000	00:57
Subject Sample # 1	0.176	00:58
Breath Volume	2.824	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.170	01:01
Breath Volume	2.445	01:01
Air Blank	0.000	01:02
Control Test	0.080	01:02
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 06/30/2021 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: SANTACRUZ MARTINEZ

DATE/TIME OF TEST: 05/28/2021 02:26:19

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.080	02:20
Air Blank		0.000	02:21
Subject Sample	# 1	0.232	02:21
Breath Volume		4.132	02:21
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.237	02:24
Breath Volume		4.277	02:24
Air Blank		0.000	02:25
Control Test		0.078	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 06/30/2021 14:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: VASQUEZ VASQUEZ

DATE/TIME OF TEST: 05/29/2021 02:05:06

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:02	
Control Test	0.079	02:03	
Air Blank	0.000	02:03	
Subject Sample	# 1	02:03	Subject Test Refused
Breath Volume	0.000	02:03	
Air Blank	0.000	02:04	
Control Test	0.080	02:04	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 06/30/2021 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TREVOR,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MATTHEWS,GREGORY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: VAZQUEZ DIAZ

DATE/TIME OF TEST: 05/29/2021 01:29:14

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:23	
Control Test	0.079	01:23	
Air Blank	0.000	01:24	
Subject Sample	# 1 0.214	01:24	
Breath Volume	1.425	01:24	
Air Blank	0.000	01:25	
Air Blank	0.000	01:27	
Subject Sample	# 2	01:27	Subject Test Refused
Breath Volume	0.000	01:27	
Air Blank	0.000	01:28	
Control Test	0.077	01:28	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

Data Download Date/Time: 07/23/2021 22:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JUSITN,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ADAMS,JUSTIN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 05/30/2021 03:07:14

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.079	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.229	03:02
Breath Volume		1.363	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.225	03:05
Breath Volume		2.445	03:05
Air Blank		0.000	03:06
Control Test		0.076	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 07/23/2021 22:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON, RAYMOND, M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: PARKS, TRAVIS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:13

SUBJECT'S LAST NAME: DANESAN

DATE/TIME OF TEST: 05/23/2021 19:49:22

Diagnostics Check	OK	19:40
Air Blank	0.000	19:40
Control Test	0.082	19:41
Air Blank	0.000	19:41
Subject Sample # 1	0.271	19:43
Breath Volume	1.855	19:43
Air Blank	0.000	19:44
Air Blank	0.000	19:46
Subject Sample # 2	0.271	19:47
Breath Volume	2.683	19:47
Air Blank	0.000	19:48
Control Test	0.082	19:48
Air Blank	0.000	19:49
Diagnostics Check	OK	19:49

Data Download Date/Time: 06/03/2021 15:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON, RAYMOND, M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: ROSADO, KIARA
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 04/24/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 11:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: NEAGLE

DATE/TIME OF TEST: 05/27/2021 23:57:19

Diagnostics Check	OK	23:54	
Air Blank	0.000	23:54	
Control Test	0.082	23:55	
Air Blank	0.000	23:55	
Subject Sample	# 1	23:55	Subject Test Refused
Breath Volume	0.000	23:55	
Air Blank	0.000	23:56	
Control Test	0.082	23:56	
Air Blank	0.000	23:57	
Diagnostics Check	OK	23:57	

Data Download Date/Time: 06/03/2021 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AFONSO,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: AFONSO,MICHAEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 04/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 10:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: MORALES GARCIA

DATE/TIME OF TEST: 05/31/2021 11:02:37

Diagnostics Check	OK	10:55
Air Blank	0.000	10:56
Control Test	0.082	10:56
Air Blank	0.000	10:57
Subject Sample # 1	0.248	10:57
Breath Volume	2.425	10:57
Air Blank	0.000	10:58
Air Blank	0.000	11:00
Subject Sample # 2	0.240	11:00
Breath Volume	1.902	11:00
Air Blank	0.000	11:01
Control Test	0.081	11:01
Air Blank	0.000	11:02
Diagnostics Check	OK	11:02

Data Download Date/Time: 06/03/2021 15:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/16/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/07/2021 20:21:37

Diagnostics Check	OK	20:15	
Air Blank	0.000	20:16	
Control Test	0.078	20:16	
Air Blank	0.000	20:17	
Subject Sample	# 1	20:19	Interferent Detect
Breath Volume	1.210	20:19	
Air Blank	0.000	20:20	
Control Test	0.080	20:20	
Air Blank	0.000	20:21	
Diagnostics Check	OK	20:21	

Data Download Date/Time: 05/17/2021 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/07/2021 20:30:28

Diagnostics Check	OK	20:22	
Air Blank	0.000	20:23	
Control Test	0.077	20:23	
Air Blank	0.000	20:24	
Subject Sample # 1	0.125	20:24	
Breath Volume	1.207	20:24	
Air Blank	0.000	20:25	
Air Blank	0.000	20:27	
Subject Sample # 2	0.126	20:29	
Breath Volume	1.402	20:29	
Air Blank		20:29	Purge Fail
Air Blank	0.000	20:30	

Data Download Date/Time: 05/17/2021 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: SYS CHECK
ARREST TIME: 15:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 05/13/2021 16:18:29

Diagnostics Check	OK	16:12	
Air Blank	0.000	16:12	
Control Test	0.080	16:13	
Air Blank	0.000	16:13	
Subject Sample # 1	0.000	16:14	
Breath Volume	2.511	16:14	
Air Blank	0.000	16:14	
Air Blank	0.000	16:16	
Subject Sample # 2		16:17	Interferent Detect
Breath Volume	1.261	16:17	
Air Blank	0.000	16:17	
Control Test	0.079	16:17	
Air Blank	0.000	16:18	
Diagnostics Check	OK	16:18	

Data Download Date/Time: 05/17/2021 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 15:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:01

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 05/13/2021 16:34:09

Diagnostics Check	OK	16:28
Air Blank	0.000	16:28
Control Test	0.080	16:28
Air Blank	0.000	16:29
Subject Sample # 1	0.000	16:29
Breath Volume	2.441	16:29
Air Blank	0.000	16:30
Air Blank	0.000	16:32
Subject Sample # 2	0.000	16:32
Breath Volume	2.007	16:32
Air Blank	0.000	16:33
Control Test	0.080	16:33
Air Blank	0.000	16:34
Diagnostics Check	OK	16:34

Data Download Date/Time: 05/17/2021 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: EBLIN

DATE/TIME OF TEST: 05/21/2021 22:26:29

Diagnostics Check	OK	22:19
Air Blank	0.000	22:20
Control Test	0.081	22:20
Air Blank	0.000	22:21
Subject Sample # 1	0.083	22:21
Breath Volume	1.632	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.089	22:24
Breath Volume	1.792	22:24
Air Blank	0.000	22:25
Control Test	0.081	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 06/09/2021 10:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 05/25/2021 02:37:39

Diagnostics Check		OK	02:30
Air Blank		0.000	02:30
Control Test		0.080	02:31
Air Blank		0.000	02:31
Subject Sample	# 1	0.100	02:32
Breath Volume		1.898	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:34
Subject Sample	# 2	0.091	02:35
Breath Volume		1.960	02:35
Air Blank		0.000	02:36
Control Test		0.077	02:37
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 06/09/2021 10:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAUMGARTNER,NATHAN
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/01/2021 03:30:43

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.078	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.171	03:25
Breath Volume	2.445	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.175	03:29
Breath Volume	1.617	03:29
Air Blank	0.000	03:29
Control Test	0.078	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 05/28/2021 20:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RICHARDSON,NATHAN
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 05/01/2021 04:06:32

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:54	
Control Test	0.078	03:54	
Air Blank	0.000	03:55	
Subject Sample	# 1 0.142	03:55	
Breath Volume	1.863	03:55	
Air Blank	0.000	03:56	
Air Blank	0.000	03:58	
Subject Sample	# 2	04:01	No Sample Provided
Breath Volume	0.000	04:01	
Air Blank	0.000	04:02	
Air Blank	0.000	04:04	
Subject Sample	# 3 0.135	04:04	
Breath Volume	1.691	04:04	
Air Blank	0.000	04:05	
Control Test	0.078	04:05	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 05/28/2021 20:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD,C,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAUMGARTNER,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: SCHWINDT

DATE/TIME OF TEST: 05/01/2021 20:21:55

Diagnostics Check	OK	20:15
Air Blank	0.000	20:16
Control Test	0.080	20:16
Air Blank	0.000	20:17
Subject Sample # 1	0.000	20:17
Breath Volume	2.328	20:17
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample # 2	0.000	20:20
Breath Volume	2.027	20:20
Air Blank	0.000	20:21
Control Test	0.080	20:21
Air Blank	0.000	20:21
Diagnostics Check	OK	20:21

Data Download Date/Time: 05/28/2021 20:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD, CHARLES, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HARWOOD, CHARLES
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: MOUNTAIN

DATE/TIME OF TEST: 05/01/2021 23:10:02

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.080	23:03
Air Blank	0.000	23:04
Subject Sample # 1	0.297	23:05
Breath Volume	2.328	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:07
Subject Sample # 2	0.300	23:08
Breath Volume	2.132	23:08
Air Blank	0.000	23:09
Control Test	0.078	23:09
Air Blank	0.000	23:09
Diagnostics Check	OK	23:10

Data Download Date/Time: 05/28/2021 20:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD,CHARELS,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAUMGARTNER,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 05/02/2021 01:22:58

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.079	01:17
Air Blank	0.000	01:17
Subject Sample # 1	0.192	01:18
Breath Volume	4.324	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:20
Subject Sample # 2	0.193	01:21
Breath Volume	4.445	01:21
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 05/28/2021 20:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD, CHARLES, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SENAK, K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/22/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: JOBE

DATE/TIME OF TEST: 05/02/2021 02:43:33

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.224	02:38
Breath Volume	3.175	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.222	02:41
Breath Volume	2.234	02:41
Air Blank	0.000	02:42
Control Test	0.077	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 05/28/2021 20:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD, CHARLES, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAUMGARTNER, N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: SAITO

DATE/TIME OF TEST: 05/02/2021 03:52:10

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.080	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.056	03:47
Breath Volume	2.199	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:49
Subject Sample # 2	0.050	03:50
Breath Volume	1.175	03:50
Air Blank	0.000	03:51
Control Test	0.078	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 05/28/2021 20:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDER,S,R,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: BAILY

DATE/TIME OF TEST: 05/04/2021 19:04:07

Diagnostics Check	OK	18:52	
Air Blank	0.000	18:53	
Control Test	0.079	18:53	
Air Blank	0.000	18:54	
Subject Sample	# 1	18:57	No Sample Provided
Breath Volume	0.000	18:57	
Air Blank	0.000	18:57	
Air Blank	0.000	18:59	
Subject Sample	# 2	19:02	No Sample Provided
Breath Volume	0.000	19:02	
Air Blank	0.000	19:03	
Control Test	0.078	19:03	
Air Blank	0.000	19:04	
Diagnostics Check	OK	19:04	

Data Download Date/Time: 05/28/2021 20:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTY, JOSHALYNN, C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GEELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 05/06/2021 03:28:52

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.080	03:22
Air Blank		0.000	03:22
Subject Sample	# 1	0.000	03:24
Breath Volume		1.222	03:24
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.000	03:27
Breath Volume		1.410	03:27
Air Blank		0.000	03:27
Control Test		0.079	03:28
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 05/28/2021 20:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDER,S,R,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SPAULDING,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CADE

DATE/TIME OF TEST: 05/08/2021 03:09:33

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.080	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.154	03:04
Breath Volume		2.179	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.152	03:07
Breath Volume		2.394	03:07
Air Blank		0.000	03:08
Control Test		0.080	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 05/28/2021 20:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,MATTHEW
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 05/12/2021 03:25:41

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.080	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.000	03:21
Breath Volume	1.503	03:21
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.000	03:24
Breath Volume	1.503	03:24
Air Blank	0.000	03:24
Control Test	0.079	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 05/28/2021 20:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,JEREMY,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LOVETT,K
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 04/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 05/15/2021 02:14:46

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.156	02:09
Breath Volume	1.472	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:11
Subject Sample # 2	0.170	02:13
Breath Volume	2.281	02:13
Air Blank	0.000	02:13
Control Test	0.079	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 05/28/2021 20:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD, CHARLES, M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS, S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MONGE

DATE/TIME OF TEST: 05/15/2021 02:34:45

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.079	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.243	02:29
Breath Volume		1.964	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.255	02:33
Breath Volume		4.097	02:33
Air Blank		0.000	02:33
Control Test		0.079	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 05/28/2021 20:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,JEREMY,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAUMGARTNER,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: PEREZ AVILES

DATE/TIME OF TEST: 05/16/2021 03:41:58

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:35
Subject Sample # 1	0.107	03:36
Breath Volume	2.234	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.102	03:40
Breath Volume	1.300	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 05/30/2021 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,JEREMY,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: CROWELL

DATE/TIME OF TEST: 05/16/2021 05:20:39

Diagnostics Check	OK	05:12
Air Blank	0.000	05:12
Control Test	0.080	05:13
Air Blank	0.000	05:13
Subject Sample # 1	0.113	05:15
Breath Volume	1.667	05:15
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.108	05:18
Breath Volume	1.335	05:18
Air Blank	0.000	05:19
Control Test	0.078	05:20
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

Data Download Date/Time: 05/30/2021 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,JEREMY,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LANGE,A
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 04/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: PICKMAN

DATE/TIME OF TEST: 05/16/2021 19:22:49

Diagnostics Check	OK	19:16
Air Blank	0.000	19:16
Control Test	0.080	19:16
Air Blank	0.000	19:17
Subject Sample # 1	0.117	19:18
Breath Volume	1.242	19:18
Air Blank	0.000	19:18
Air Blank	0.000	19:20
Subject Sample # 2	0.109	19:21
Breath Volume	1.519	19:21
Air Blank	0.000	19:21
Control Test	0.079	19:22
Air Blank	0.000	19:22
Diagnostics Check	OK	19:22

Data Download Date/Time: 05/30/2021 09:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,J,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BARBER,B
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 04/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: BUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: DIGIACOMO

DATE/TIME OF TEST: 05/22/2021 22:32:27

Diagnostics Check	OK	22:22	
Air Blank	0.000	22:23	
Control Test	0.080	22:23	
Air Blank	0.000	22:24	
Subject Sample	# 1	22:27	No Sample Provided
Breath Volume	0.000	22:27	
Air Blank	0.000	22:27	
Air Blank	0.000	22:29	
Subject Sample	# 2	22:31	Subject Test Refused
Breath Volume	0.000	22:31	
Air Blank	0.000	22:31	
Control Test	0.080	22:31	
Air Blank	0.000	22:32	
Diagnostics Check	OK	22:32	

Data Download Date/Time: 05/30/2021 09:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,J,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HOULE,A
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 04/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: MISHOE

DATE/TIME OF TEST: 05/23/2021 01:20:08

Diagnostics Check		OK	01:13
Air Blank		0.000	01:14
Control Test		0.080	01:14
Air Blank		0.000	01:14
Subject Sample	# 1	0.185	01:15
Breath Volume		2.085	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:17
Subject Sample	# 2	0.182	01:18
Breath Volume		1.691	01:18
Air Blank		0.000	01:19
Control Test		0.080	01:19
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 05/30/2021 09:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDER,S,R,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: PFALGRAF

DATE/TIME OF TEST: 05/26/2021 23:54:17

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.080	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.171	23:48
Breath Volume	1.117	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.178	23:52
Breath Volume	1.511	23:52
Air Blank	0.000	23:53
Control Test	0.079	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 05/30/2021 09:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR,NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 04/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: KREMER

DATE/TIME OF TEST: 05/01/2021 02:10:00

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:04
Subject Sample # 1	0.177	02:05
Breath Volume	1.343	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.187	02:08
Breath Volume	4.023	02:08
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:10

Data Download Date/Time: 05/14/2021 17:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: IV,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: FREDERICKS

DATE/TIME OF TEST: 05/01/2021 03:17:15

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.000	03:12
Breath Volume	3.492	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.000	03:15
Breath Volume	2.390	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 05/14/2021 17:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHARELUS, OBED
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 04/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: SELLITTI

DATE/TIME OF TEST: 05/02/2021 00:07:56

Diagnostics Check	OK	00:01
Air Blank	0.000	00:02
Control Test	0.080	00:02
Air Blank	0.000	00:02
Subject Sample # 1	0.009	00:03
Breath Volume	2.714	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.009	00:06
Breath Volume	3.179	00:06
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 05/14/2021 17:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MORALES,WILLIAM
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 04/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: TYNER

DATE/TIME OF TEST: 05/02/2021 05:06:23

Diagnostics Check	OK	05:00
Air Blank	0.000	05:00
Control Test	0.080	05:01
Air Blank	0.000	05:01
Subject Sample # 1	0.000	05:02
Breath Volume	1.382	05:02
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.000	05:05
Breath Volume	1.671	05:05
Air Blank	0.000	05:05
Control Test	0.080	05:05
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 05/14/2021 17:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STILES,MEGAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: ROSAMBERT

DATE/TIME OF TEST: 05/03/2021 02:56:07

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.037	02:51
Breath Volume	1.445	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.038	02:54
Breath Volume	2.210	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 05/14/2021 17:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BUTZBACH,WILLIAM
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 04/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 05/05/2021 02:17:36

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.081	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.000	02:13
Breath Volume	3.332	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.000	02:16
Breath Volume	3.535	02:16
Air Blank	0.000	02:16
Control Test	0.081	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 05/14/2021 17:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: D'ANDREA,ADAM
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:53

SUBJECT'S LAST NAME: WALDEN

DATE/TIME OF TEST: 05/05/2021 08:26:46

Diagnostics Check	OK	08:20
Air Blank	0.000	08:20
Control Test	0.080	08:20
Air Blank	0.000	08:21
Subject Sample # 1	0.342	08:22
Breath Volume	1.781	08:22
Air Blank	0.000	08:22
Air Blank	0.000	08:24
Subject Sample # 2	0.335	08:25
Breath Volume	1.492	08:25
Air Blank	0.000	08:25
Control Test	0.080	08:26
Air Blank	0.000	08:26
Diagnostics Check	OK	08:26

Data Download Date/Time: 05/14/2021 17:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: IV,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: BERRIOS

DATE/TIME OF TEST: 05/06/2021 00:13:52

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.080	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.157	00:07
Breath Volume	1.101	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.173	00:12
Breath Volume	1.640	00:12
Air Blank	0.000	00:12
Control Test	0.080	00:13
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 05/14/2021 17:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VAN CAMP,JEFFERY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: CORAL

DATE/TIME OF TEST: 05/06/2021 01:51:50

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.080	01:44
Air Blank		0.000	01:44
Subject Sample	# 1	0.201	01:45
Breath Volume		2.070	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.189	01:50
Breath Volume		1.347	01:50
Air Blank		0.000	01:50
Control Test		0.080	01:51
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 05/14/2021 17:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: D'ANDREA,ADAM
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 05/06/2021 17:38:59

Diagnostics Check	OK	17:32
Air Blank	0.000	17:33
Control Test	0.079	17:33
Air Blank	0.000	17:34
Subject Sample # 1	0.000	17:34
Breath Volume	3.597	17:34
Air Blank	0.000	17:35
Air Blank	0.000	17:37
Subject Sample # 2	0.000	17:37
Breath Volume	3.488	17:37
Air Blank	0.000	17:38
Control Test	0.079	17:38
Air Blank	0.000	17:38
Diagnostics Check	OK	17:38

Data Download Date/Time: 05/14/2021 17:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR,NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 04/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: DENEUS

DATE/TIME OF TEST: 05/07/2021 04:53:18

Diagnostics Check	OK	04:46
Air Blank	0.000	04:47
Control Test	0.081	04:47
Air Blank	0.000	04:48
Subject Sample # 1	0.139	04:48
Breath Volume	2.058	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.139	04:51
Breath Volume	2.617	04:51
Air Blank	0.000	04:52
Control Test	0.080	04:52
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 05/14/2021 17:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCGILLICUDDY,STEVEN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 04/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: WEINENHAMER

DATE/TIME OF TEST: 05/07/2021 22:00:16

Diagnostics Check	OK	21:53
Air Blank	0.000	21:53
Control Test	0.081	21:53
Air Blank	0.000	21:54
Subject Sample # 1	0.194	21:55
Breath Volume	2.546	21:55
Air Blank	0.000	21:55
Air Blank	0.000	21:57
Subject Sample # 2	0.194	21:58
Breath Volume	2.500	21:58
Air Blank	0.000	21:59
Control Test	0.080	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 05/14/2021 17:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HOWARD,EDWARD
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 04/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: VASQUEZ LOPEZ

DATE/TIME OF TEST: 05/08/2021 02:27:05

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.081	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.171	02:22
Breath Volume	2.753	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:24
Subject Sample # 2	0.166	02:25
Breath Volume	3.363	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 05/14/2021 17:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR, NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 04/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: PRESSLEY

DATE/TIME OF TEST: 05/08/2021 04:28:36

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.080	04:22
Air Blank	0.000	04:22
Subject Sample # 1	0.181	04:24
Breath Volume	1.722	04:24
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.187	04:27
Breath Volume	1.781	04:27
Air Blank	0.000	04:27
Control Test	0.079	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 05/14/2021 17:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: ORTEGA LNTURIAS

DATE/TIME OF TEST: 05/10/2021 03:03:10

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.081	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.165	02:58
Breath Volume	1.195	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.160	03:01
Breath Volume	2.316	03:01
Air Blank	0.000	03:02
Control Test	0.080	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 05/14/2021 17:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TRINIDAD,OSCAR
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: BARBUTO

DATE/TIME OF TEST: 05/11/2021 03:56:38

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.080	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.218	03:52
Breath Volume		2.492	03:52
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.215	03:55
Breath Volume		1.589	03:55
Air Blank		0.000	03:55
Control Test		0.079	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 05/14/2021 17:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FRIEDMAN,RICHARD
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 04/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 05/11/2021 22:58:38

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.081	22:52
Air Blank	0.000	22:52
Subject Sample # 1	0.259	22:53
Breath Volume	2.964	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.259	22:57
Breath Volume	2.929	22:57
Air Blank	0.000	22:57
Control Test	0.079	22:58
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 05/14/2021 17:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALEZ,MARQUIS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: CYPRIAN

DATE/TIME OF TEST: 05/12/2021 02:20:49

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.080	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.000	02:16
Breath Volume	2.074	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.000	02:19
Breath Volume	1.300	02:19
Air Blank	0.000	02:19
Control Test	0.081	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 05/14/2021 17:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCGILLICUDDY,STEVEN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 04/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: YORKE

DATE/TIME OF TEST: 05/13/2021 01:33:27

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.081	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.111	01:28
Breath Volume	3.125	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.106	01:31
Breath Volume	3.164	01:31
Air Blank	0.000	01:32
Control Test	0.080	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 05/14/2021 17:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HALL,HAROLD
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 04/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:18

SUBJECT'S LAST NAME: MIODINI

DATE/TIME OF TEST: 05/14/2021 13:51:53

Diagnostics Check	OK	13:45
Air Blank	0.000	13:45
Control Test	0.080	13:46
Air Blank	0.000	13:46
Subject Sample # 1	0.201	13:47
Breath Volume	3.035	13:47
Air Blank	0.000	13:48
Air Blank	0.000	13:49
Subject Sample # 2	0.190	13:50
Breath Volume	2.214	13:50
Air Blank	0.000	13:50
Control Test	0.079	13:51
Air Blank	0.000	13:51
Diagnostics Check	OK	13:51

Data Download Date/Time: 05/14/2021 17:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ALFONSO

DATE/TIME OF TEST: 05/06/2021 00:14:48

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:13	
Control Test	0.078	00:13	
Air Blank	0.000	00:14	
Subject Sample	# 1	00:14	RFI Detect
Breath Volume	0.000	00:14	
Air Blank		00:14	RFI Detect
Air Blank		00:14	RFI Detect

Data Download Date/Time: 05/06/2021 20:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ALFONSO

DATE/TIME OF TEST: 05/06/2021 00:22:14

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:18	
Control Test	0.077	00:18	
Air Blank	0.000	00:19	
Subject Sample	# 1 0.159	00:19	
Breath Volume	4.273	00:19	
Air Blank	0.000	00:20	
Air Blank		00:21	RFI Detect
Air Blank		00:22	RFI Detect
Air Blank		00:22	RFI Detect

Data Download Date/Time: 05/06/2021 20:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: ALFONSO

DATE/TIME OF TEST: 05/06/2021 00:57:26

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.141	00:52
Breath Volume	4.542	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.141	00:55
Breath Volume	4.296	00:55
Air Blank	0.000	00:56
Control Test	0.077	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 05/06/2021 20:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: POWELL,AARON
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 04/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 05/06/2021 01:52:37

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.070	01:52	RFI Detect
Air Blank	0.000	01:52	

Data Download Date/Time: 05/06/2021 20:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: POWELL,AARON
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 04/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 05/06/2021 02:01:48

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.076	01:55
Air Blank		0.000	01:56
Subject Sample	# 1	0.223	01:57
Breath Volume		2.035	01:57
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.218	02:00
Breath Volume		3.035	02:00
Air Blank		0.000	02:00
Control Test		0.076	02:01
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 05/06/2021 20:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 05/06/2021 02:34:36

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:32	
Control Test	0.077	02:32	
Air Blank	0.000	02:33	
Subject Sample	# 1	02:33	RFI Detect
Breath Volume	0.000	02:33	
Air Blank		02:34	RFI Detect
Air Blank	0.000	02:34	

Data Download Date/Time: 05/06/2021 20:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 05/06/2021 02:40:01

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:37	
Control Test	0.076	02:37	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:38	RFI Detect
Breath Volume	0.000	02:38	
Air Blank	0.000	02:39	
Control Test	0.078	02:39	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:40	

Data Download Date/Time: 05/06/2021 20:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: CORTEZ

DATE/TIME OF TEST: 05/08/2021 03:00:48

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:57	
Control Test	0.080	02:58	
Air Blank	0.000	02:58	
Subject Sample # 1	0.148	02:58	
Breath Volume	4.695	02:58	
Air Blank		02:59	Improper Sample
Air Blank	0.000	02:59	
Control Test	0.081	03:00	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 06/22/2021 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: CORTEZ

DATE/TIME OF TEST: 05/08/2021 03:09:00

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.079	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.151	03:04
Breath Volume		2.269	03:04
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.168	03:07
Breath Volume		2.714	03:07
Air Blank		0.000	03:08
Control Test		0.079	03:08
Air Blank		0.000	03:08
Diagnostics Check		OK	03:09

Data Download Date/Time: 06/22/2021 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 05/11/2021 01:47:24

Diagnostics Check		OK	01:40
Air Blank		0.000	01:41
Control Test		0.078	01:41
Air Blank		0.000	01:42
Subject Sample	# 1	0.095	01:42
Breath Volume		2.703	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.096	01:45
Breath Volume		2.132	01:45
Air Blank		0.000	01:46
Control Test		0.077	01:46
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 06/22/2021 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: MCCONNELL

DATE/TIME OF TEST: 05/15/2021 03:18:43

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.080	03:12
Air Blank		0.000	03:12
Subject Sample	# 1	0.104	03:13
Breath Volume		2.265	03:13
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.101	03:17
Breath Volume		1.351	03:17
Air Blank		0.000	03:17
Control Test		0.082	03:18
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 06/22/2021 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: CAPALLIA

DATE/TIME OF TEST: 05/18/2021 00:23:26

Diagnostics Check		OK	00:14
Air Blank		0.000	00:15
Control Test		0.079	00:15
Air Blank		0.000	00:16
Subject Sample	# 1	0.173	00:18
Breath Volume		1.429	00:18
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.169	00:21
Breath Volume		1.492	00:21
Air Blank		0.000	00:22
Control Test		0.077	00:22
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 06/22/2021 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: POLSTON

DATE/TIME OF TEST: 05/22/2021 23:00:57

Diagnostics Check	OK	23:00	
Air Blank		23:00	RFI Detect
Air Blank		23:00	RFI Detect

Data Download Date/Time: 06/22/2021 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: POLSTON

DATE/TIME OF TEST: 05/22/2021 23:09:37

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.078	23:03
Air Blank		0.000	23:03
Subject Sample	# 1	0.093	23:04
Breath Volume		2.609	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.087	23:08
Breath Volume		2.195	23:08
Air Blank		0.000	23:08
Control Test		0.078	23:08
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 06/22/2021 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: RUSSANO

DATE/TIME OF TEST: 05/23/2021 23:02:36

Diagnostics Check	OK	22:55	
Air Blank	0.000	22:56	
Control Test	0.080	22:56	
Air Blank	0.000	22:57	
Subject Sample # 1	0.007	22:57	
Breath Volume	3.296	22:57	
Air Blank	0.000	22:58	
Air Blank	0.000	23:00	
Subject Sample # 2	0.000	23:00	
Breath Volume	2.945	23:00	
Air Blank	0.000	23:01	
Control Test	0.079	23:01	
Air Blank	0.000	23:02	
Diagnostics Check	Fail	23:02	Analytical Stability Test
Air Blank	0.000	23:02	

Data Download Date/Time: 06/22/2021 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,MILLS,S
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: RUSSANO

DATE/TIME OF TEST: 05/23/2021 23:13:43

Diagnostics Check	OK	23:06
Air Blank	0.000	23:07
Control Test	0.080	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.000	23:09
Breath Volume	3.320	23:09
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.000	23:12
Breath Volume	2.566	23:12
Air Blank	0.000	23:12
Control Test	0.079	23:13
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 06/22/2021 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 05/30/2021 01:22:29

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.077	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.110	01:17
Breath Volume		2.804	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.108	01:20
Breath Volume		4.195	01:20
Air Blank		0.000	01:21
Control Test		0.077	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 06/22/2021 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTENOUR,COREY,T
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: RITTENOUR,COREY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: MATSON

DATE/TIME OF TEST: 05/15/2021 03:30:00

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.079	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.345	03:25
Breath Volume		2.406	03:25
Air Blank		0.000	03:26
Air Blank		0.000	03:27
Subject Sample	# 2	0.345	03:28
Breath Volume		3.187	03:28
Air Blank		0.000	03:29
Control Test		0.078	03:29
Air Blank		0.000	03:29
Diagnostics Check		OK	03:30

Data Download Date/Time: 05/28/2021 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTENOUR,COREY,T
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: RITTENOUR,COREY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: SITES

DATE/TIME OF TEST: 05/15/2021 21:40:48

Diagnostics Check		OK	21:29	
Air Blank		0.000	21:29	
Control Test		0.080	21:29	
Air Blank		0.000	21:30	
Subject Sample	# 1	0.188	21:33	Volume Not Met
Breath Volume		0.035	21:33	
Air Blank		0.000	21:34	
Air Blank		0.000	21:35	
Subject Sample	# 2	0.208	21:39	Volume Not Met
Breath Volume		0.054	21:39	
Air Blank		0.000	21:39	
Control Test		0.078	21:40	
Air Blank		0.000	21:40	
Diagnostics Check		OK	21:40	

Data Download Date/Time: 05/28/2021 15:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTENOUR,COREY,T
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: RITTENOUR,COREY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 05/17/2021 04:01:08

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.179	03:56
Breath Volume	2.179	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:58
Subject Sample # 2	0.185	03:59
Breath Volume	1.960	03:59
Air Blank	0.000	04:00
Control Test	0.078	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 05/28/2021 15:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTENOUR,COREY,T
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: SPELLER,BENJAMIN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: DRAPER

DATE/TIME OF TEST: 05/21/2021 02:50:52

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.080	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.251	02:46
Breath Volume	3.312	02:46
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.256	02:49
Breath Volume	2.113	02:49
Air Blank	0.000	02:49
Control Test	0.078	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 05/28/2021 15:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTENOUR,COREY,T
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: RITTENOUR,COREY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: FINK

DATE/TIME OF TEST: 05/25/2021 01:38:17

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.126	01:33
Breath Volume	3.238	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.117	01:36
Breath Volume	1.964	01:36
Air Blank	0.000	01:37
Control Test	0.078	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 05/28/2021 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: SCHUCH

DATE/TIME OF TEST: 05/01/2021 03:25:51

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.122	03:18
Breath Volume	1.261	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.128	03:24
Breath Volume	2.468	03:24
Air Blank	0.000	03:24
Control Test	0.080	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 05/04/2021 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: COLLINS JR

DATE/TIME OF TEST: 05/01/2021 04:11:07

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:57	
Control Test	0.080	03:57	
Air Blank	0.000	03:58	
Subject Sample	# 1	04:01	No Sample Provided
Breath Volume	0.000	04:01	
Air Blank	0.000	04:01	
Air Blank	0.000	04:03	
Subject Sample	# 2	0.164	04:05
Breath Volume	1.402	04:05	
Air Blank	0.000	04:06	
Air Blank	0.000	04:08	
Subject Sample	# 3	0.165	04:09
Breath Volume	1.957	04:09	
Air Blank	0.000	04:10	
Control Test	0.078	04:10	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:11	

Data Download Date/Time: 05/04/2021 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: GONZALEZ ANTEQUERA

DATE/TIME OF TEST: 05/01/2021 23:42:02

Diagnostics Check	OK	23:33
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.038	23:36
Breath Volume	1.296	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.036	23:40
Breath Volume	1.589	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:42

Data Download Date/Time: 05/04/2021 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 05/02/2021 00:26:06

Diagnostics Check	OK	00:18
Air Blank	0.000	00:18
Control Test	0.080	00:18
Air Blank	0.000	00:19
Subject Sample # 1	0.090	00:21
Breath Volume	2.972	00:21
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.086	00:24
Breath Volume	3.250	00:24
Air Blank	0.000	00:25
Control Test	0.079	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 05/04/2021 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY, ANIA, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARRELL, Q
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/02/2021 05:33:40

Diagnostics Check	OK	05:27
Air Blank	0.000	05:27
Control Test	0.080	05:27
Air Blank	0.000	05:28
Subject Sample # 1	0.211	05:28
Breath Volume	1.445	05:28
Air Blank	0.000	05:29
Air Blank	0.000	05:31
Subject Sample # 2	0.223	05:31
Breath Volume	2.023	05:31
Air Blank	0.000	05:32
Control Test	0.080	05:33
Air Blank	0.000	05:33
Diagnostics Check	OK	05:33

Data Download Date/Time: 05/04/2021 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DOBSON,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: ESCALANTE PEREZ

DATE/TIME OF TEST: 05/02/2021 19:55:43

Diagnostics Check	OK	19:52	
Air Blank	0.000	19:52	
Control Test	0.079	19:52	
Air Blank	0.000	19:53	
Subject Sample	# 1	19:54	Subject Test Refused
Breath Volume	0.000	19:54	
Air Blank	0.000	19:54	
Control Test	0.080	19:55	
Air Blank	0.000	19:55	
Diagnostics Check	OK	19:55	

Data Download Date/Time: 05/04/2021 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: KINIRY

DATE/TIME OF TEST: 05/03/2021 01:13:25

Diagnostics Check		OK	01:07
Air Blank		0.000	01:07
Control Test		0.081	01:07
Air Blank		0.000	01:08
Subject Sample	# 1	0.054	01:08
Breath Volume		1.644	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.055	01:11
Breath Volume		1.441	01:11
Air Blank		0.000	01:12
Control Test		0.080	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 05/04/2021 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 05/04/2021 20:37:54

Diagnostics Check	OK	20:35	
Air Blank	0.000	20:36	
Control Test	0.080	20:36	
Air Blank	0.000	20:37	
Subject Sample	# 1	20:37	Sequence Aborted
Breath Volume	0.000	20:37	
Air Blank	0.000	20:37	

Data Download Date/Time: 06/02/2021 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 05/19/2021 05:04:18

Diagnostics Check	OK	04:55
Air Blank	0.000	04:56
Control Test	0.081	04:56
Air Blank	0.000	04:57
Subject Sample # 1	0.252	04:59
Breath Volume	1.664	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:01
Subject Sample # 2	0.238	05:02
Breath Volume	1.281	05:02
Air Blank	0.000	05:03
Control Test	0.077	05:03
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 06/02/2021 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HARRIS,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 05/27/2021 23:47:23

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.080	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.158	23:42
Breath Volume	1.269	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.177	23:45
Breath Volume	2.476	23:45
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 06/02/2021 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: RAMOS RODRIGUEZ

DATE/TIME OF TEST: 05/30/2021 02:49:22

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.189	02:44
Breath Volume	1.800	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:46
Subject Sample # 2	0.194	02:47
Breath Volume	3.453	02:47
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 06/02/2021 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SWEENEY,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: KAFALENOS

DATE/TIME OF TEST: 05/31/2021 03:34:02

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.081	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.098	03:29
Breath Volume	2.511	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:31
Subject Sample # 2	0.099	03:32
Breath Volume	2.343	03:32
Air Blank	0.000	03:33
Control Test	0.081	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:34

Data Download Date/Time: 06/02/2021 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,REGINALD,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROGERS,REGINALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 12:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: PARDUE

DATE/TIME OF TEST: 05/16/2021 12:58:00

Diagnostics Check	OK	12:51
Air Blank	0.000	12:52
Control Test	0.081	12:52
Air Blank	0.000	12:52
Subject Sample # 1	0.201	12:53
Breath Volume	3.250	12:53
Air Blank	0.000	12:54
Air Blank	0.000	12:55
Subject Sample # 2	0.193	12:56
Breath Volume	4.023	12:56
Air Blank	0.000	12:57
Control Test	0.079	12:57
Air Blank	0.000	12:57
Diagnostics Check	OK	12:58

Data Download Date/Time: 06/10/2021 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,ROBERT,
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: RAMOS,ROBERT
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: BUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 05/31/2021 19:14:40

Diagnostics Check	OK	19:08
Air Blank	0.000	19:08
Control Test	0.082	19:08
Air Blank	0.000	19:09
Subject Sample # 1	0.151	19:10
Breath Volume	2.722	19:10
Air Blank	0.000	19:10
Air Blank	0.000	19:12
Subject Sample # 2	0.146	19:13
Breath Volume	2.746	19:13
Air Blank	0.000	19:13
Control Test	0.079	19:14
Air Blank	0.000	19:14
Diagnostics Check	OK	19:14

Data Download Date/Time: 06/10/2021 19:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CAMILA,P
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CARRERO,B
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: BETHEL

DATE/TIME OF TEST: 05/01/2021 04:25:34

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:23	
Control Test	0.081	04:23	
Air Blank	0.000	04:23	
Subject Sample	# 1	04:24	Subject Test Refused
Breath Volume	0.000	04:24	
Air Blank	0.000	04:24	
Control Test	0.081	04:24	
Air Blank	0.000	04:25	
Diagnostics Check	OK	04:25	

Data Download Date/Time: 05/03/2021 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODOLPH,JOEL,J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: REVERE,RODARI
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: LINDOR

DATE/TIME OF TEST: 05/29/2021 20:45:35

Diagnostics Check	OK	20:38
Air Blank	0.000	20:39
Control Test	0.081	20:39
Air Blank	0.000	20:40
Subject Sample # 1	0.225	20:41
Breath Volume	3.125	20:41
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.222	20:43
Breath Volume	3.375	20:43
Air Blank	0.000	20:44
Control Test	0.080	20:44
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 06/01/2021 07:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,WILLIAM,G
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,WILLIAMS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 05/01/2021 03:59:55

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.079	03:53
Air Blank	0.000	03:54
Subject Sample # 1	0.104	03:54
Breath Volume	2.250	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.090	03:58
Breath Volume	1.816	03:58
Air Blank	0.000	03:58
Control Test	0.077	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 05/03/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ESTAPE,ZACHARY,C

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: TARGIA,F

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 04/30/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 21:02

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 05/01/2021 21:57:11

Diagnostics Check	OK	21:54	
Air Blank	0.000	21:54	
Control Test	0.080	21:55	
Air Blank	0.000	21:55	
Subject Sample	# 1	21:55	Subject Test Refused
Breath Volume	0.000	21:55	
Air Blank	0.000	21:56	
Control Test	0.079	21:56	
Air Blank	0.000	21:57	
Diagnostics Check	OK	21:57	

Data Download Date/Time: 05/03/2021 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARR,JASON,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MACHADO AVENDANO

DATE/TIME OF TEST: 05/02/2021 01:32:41

Diagnostics Check		OK	01:20	
Air Blank		0.000	01:21	
Control Test		0.078	01:21	
Air Blank		0.000	01:22	
Subject Sample	# 1	0.185	01:25	Volume Not Met
Breath Volume		0.066	01:25	
Air Blank		0.000	01:26	
Air Blank		0.000	01:27	
Subject Sample	# 2	0.208	01:30	Volume Not Met
Breath Volume		0.050	01:30	
Air Blank		0.000	01:31	
Control Test		0.077	01:32	
Air Blank		0.000	01:32	
Diagnostics Check		OK	01:32	

Data Download Date/Time: 05/13/2021 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARR,JASON,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: MACHADO AVENDANO

DATE/TIME OF TEST: 05/02/2021 01:59:06

Diagnostics Check	OK	01:53	
Air Blank	0.000	01:54	
Control Test	0.076	01:54	
Air Blank	0.000	01:55	
Subject Sample	# 1	01:57	Subject Test Refused
Breath Volume	0.035	01:57	
Air Blank	0.000	01:58	
Control Test	0.078	01:58	
Air Blank	0.000	01:59	
Diagnostics Check	OK	01:59	

Data Download Date/Time: 05/13/2021 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MORENO,JORGE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 11:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: KOEPPPEL

DATE/TIME OF TEST: 05/05/2021 00:26:14

Diagnostics Check	OK	00:23	
Air Blank	0.000	00:23	
Control Test	0.079	00:24	
Air Blank	0.000	00:24	
Subject Sample	# 1	00:24	Subject Test Refused
Breath Volume	0.000	00:24	
Air Blank	0.000	00:25	
Control Test	0.078	00:25	
Air Blank	0.000	00:26	
Diagnostics Check	OK	00:26	

Data Download Date/Time: 05/13/2021 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARR,JASON,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: GUARNERO

DATE/TIME OF TEST: 05/06/2021 04:35:51

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.078	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.146	04:28
Breath Volume	3.320	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.123	04:31
Breath Volume	1.250	04:31
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 3	0.145	04:34
Breath Volume	3.214	04:34
Air Blank	0.000	04:34
Control Test	0.078	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 05/13/2021 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: VICUNA GUEVARA

DATE/TIME OF TEST: 05/08/2021 01:10:43

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:07	
Control Test	0.079	01:07	
Air Blank	0.000	01:07	
Subject Sample	# 1	01:09	Subject Test Refused
Breath Volume	0.000	01:09	
Air Blank	0.000	01:09	
Control Test	0.079	01:10	
Air Blank	0.000	01:10	
Diagnostics Check	OK	01:10	

Data Download Date/Time: 05/13/2021 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 05/08/2021 02:20:35

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:16	
Control Test	0.078	02:16	
Air Blank	0.000	02:16	
Subject Sample	# 1	02:19	Subject Test Refused
Breath Volume	0.000	02:19	
Air Blank	0.000	02:19	
Control Test	0.078	02:19	
Air Blank	0.000	02:20	
Diagnostics Check	OK	02:20	

Data Download Date/Time: 05/13/2021 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MORENO,JORGE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 05/08/2021 02:38:10

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:35	
Control Test	0.078	02:35	
Air Blank	0.000	02:36	
Subject Sample	# 1	02:36	Subject Test Refused
Breath Volume	0.000	02:36	
Air Blank	0.000	02:37	
Control Test	0.078	02:37	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 05/13/2021 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIDAL,NERELYS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: VIDAL,NERELYS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: BELLISSIMO

DATE/TIME OF TEST: 05/08/2021 23:10:13

Diagnostics Check	OK	23:06	
Air Blank	0.000	23:06	
Control Test	0.079	23:06	
Air Blank	0.000	23:07	
Subject Sample	# 1	23:08	Subject Test Refused
Breath Volume	0.000	23:08	
Air Blank	0.000	23:09	
Control Test	0.079	23:09	
Air Blank	0.000	23:10	
Diagnostics Check	OK	23:10	

Data Download Date/Time: 05/13/2021 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HANNA,WARDELL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 11:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: RENEAU-RICH

DATE/TIME OF TEST: 05/09/2021 00:10:03

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:01	
Control Test	0.079	00:02	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:05	No Sample Provided
Breath Volume	0.000	00:05	
Air Blank	0.000	00:06	
Air Blank	0.000	00:08	
Subject Sample	# 2	00:08	Subject Test Refused
Breath Volume	0.000	00:08	
Air Blank	0.000	00:09	
Control Test	0.078	00:09	
Air Blank	0.000	00:09	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 05/13/2021 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIDAL,NERELYS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HANNA,WARDELL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: RAMIA

DATE/TIME OF TEST: 05/12/2021 21:27:05

Diagnostics Check	OK	21:20
Air Blank	0.000	21:20
Control Test	0.078	21:20
Air Blank	0.000	21:21
Subject Sample # 1	0.128	21:22
Breath Volume	2.324	21:22
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample # 2	0.127	21:25
Breath Volume	2.460	21:25
Air Blank	0.000	21:26
Control Test	0.078	21:26
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 05/13/2021 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,ALEXANDRIA,N
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HRADECKY,ALEXANDRIA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: SHAUGHNESSY

DATE/TIME OF TEST: 05/13/2021 00:58:19

Diagnostics Check		OK	00:51
Air Blank		0.000	00:51
Control Test		0.078	00:52
Air Blank		0.000	00:52
Subject Sample	# 1	0.156	00:53
Breath Volume		2.765	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.151	00:56
Breath Volume		2.007	00:56
Air Blank		0.000	00:57
Control Test		0.076	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 05/13/2021 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CRISTEA,VICTOR
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: BARTOLOMEO

DATE/TIME OF TEST: 05/16/2021 03:45:39

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.078	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.194	03:41
Breath Volume		1.425	03:41
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.187	03:44
Breath Volume		1.917	03:44
Air Blank		0.000	03:44
Control Test		0.078	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 06/10/2021 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIDAL,NERELYS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: VIDAL,NERELYS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 05/18/2021 00:05:03

Diagnostics Check		OK	23:57
Air Blank		0.000	23:58
Control Test		0.078	23:58
Air Blank		0.000	23:59
Subject Sample	# 1	0.273	00:00
Breath Volume		1.757	00:00
Air Blank		0.000	00:00
Air Blank		0.000	00:02
Subject Sample	# 2	0.269	00:03
Breath Volume		1.601	00:03
Air Blank		0.000	00:04
Control Test		0.078	00:04
Air Blank		0.000	00:04
Diagnostics Check		OK	00:05

Data Download Date/Time: 06/10/2021 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIDAL,NERELYS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MATTHEWS,CASEY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: MANNINO

DATE/TIME OF TEST: 05/18/2021 00:27:26

Diagnostics Check	OK	00:23	
Air Blank	0.000	00:24	
Control Test	0.078	00:24	
Air Blank	0.000	00:25	
Subject Sample	# 1	00:26	Subject Test Refused
Breath Volume	0.000	00:26	
Air Blank	0.000	00:26	
Control Test	0.078	00:26	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

Data Download Date/Time: 06/10/2021 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: DENNINGHOFF

DATE/TIME OF TEST: 05/22/2021 02:09:11

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:06	
Control Test	0.078	02:06	
Air Blank	0.000	02:07	
Subject Sample	# 1	02:07	Subject Test Refused
Breath Volume	0.000	02:07	
Air Blank	0.000	02:08	
Control Test	0.078	02:08	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

Data Download Date/Time: 06/10/2021 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: DONOVAN

DATE/TIME OF TEST: 05/23/2021 00:21:09

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:18	
Control Test	0.079	00:19	
Air Blank	0.000	00:19	
Subject Sample	# 1	00:19	Subject Test Refused
Breath Volume	0.000	00:19	
Air Blank	0.000	00:20	
Control Test	0.080	00:20	
Air Blank	0.000	00:21	
Diagnostics Check	OK	00:21	

Data Download Date/Time: 06/10/2021 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSEN,DYLLON,E
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HANSEN,DYLLON
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 14:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:38

SUBJECT'S LAST NAME: MAUS

DATE/TIME OF TEST: 05/25/2021 15:01:49

Diagnostics Check	OK	14:59	
Air Blank	0.000	14:59	
Control Test	0.077	14:59	
Air Blank	0.000	15:00	
Subject Sample	# 1	15:01	Sequence Aborted
Breath Volume	0.000	15:01	
Air Blank	0.000	15:01	

Data Download Date/Time: 06/10/2021 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANGEL,ROBERT,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: RANGEL,ROBERT
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: LEYTON

DATE/TIME OF TEST: 05/26/2021 01:37:04

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:35	
Control Test	0.079	01:36	
Air Blank	0.000	01:36	
Subject Sample	# 1	01:36	RFI Detect
Breath Volume	0.000	01:36	
Air Blank		01:36	RFI Detect
Air Blank		01:37	RFI Detect

Data Download Date/Time: 06/10/2021 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANGEL,ROBERT,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: RANGEL,ROBERT
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: LEYTON

DATE/TIME OF TEST: 05/26/2021 01:47:45

Diagnostics Check	OK	01:44	
Air Blank	0.000	01:44	
Control Test	0.079	01:44	
Air Blank	0.000	01:45	
Subject Sample # 1	0.203	01:46	
Breath Volume	1.722	01:46	
Air Blank		01:46	RFI Detect
Air Blank	0.000	01:46	
Control Test	0.078	01:47	
Air Blank	0.000	01:47	
Diagnostics Check	OK	01:47	

Data Download Date/Time: 06/10/2021 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANGEL,ROBERT,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: RANGEL,ROBERT
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: LEYTON

DATE/TIME OF TEST: 05/26/2021 02:08:53

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:57	
Control Test	0.078	01:57	
Air Blank	0.000	01:58	
Subject Sample	# 1 0.176	01:59	
Breath Volume	1.414	01:59	
Air Blank	0.000	01:59	
Air Blank	0.000	02:01	
Subject Sample	# 2	02:04	No Sample Provided
Breath Volume	0.000	02:04	
Air Blank	0.000	02:05	
Air Blank	0.000	02:07	
Subject Sample	# 3	02:07	Subject Test Refused
Breath Volume	0.000	02:07	
Air Blank	0.000	02:07	
Control Test	0.079	02:08	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

Data Download Date/Time: 06/10/2021 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIDAL,NERELYS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MORENO,JORGE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: TREJO

DATE/TIME OF TEST: 05/26/2021 23:46:03

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:43	
Control Test	0.078	23:43	
Air Blank	0.000	23:44	
Subject Sample	# 1	23:44	Subject Test Refused
Breath Volume	0.000	23:44	
Air Blank	0.000	23:45	
Control Test	0.079	23:45	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:46	

Data Download Date/Time: 06/10/2021 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: TRUJILLO

DATE/TIME OF TEST: 05/29/2021 03:15:48

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:13	
Control Test	0.078	03:13	
Air Blank	0.000	03:13	
Subject Sample	# 1	03:14	Subject Test Refused
Breath Volume	0.000	03:14	
Air Blank	0.000	03:14	
Control Test	0.078	03:15	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:15	

Data Download Date/Time: 06/10/2021 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANGEL,ROBERT,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: BURTON

DATE/TIME OF TEST: 05/30/2021 02:07:20

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.078	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.200	02:02
Breath Volume		2.042	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.181	02:05
Breath Volume		1.734	02:05
Air Blank		0.000	02:06
Control Test		0.078	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 06/10/2021 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 05/30/2021 02:39:13

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.078	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.133	02:34
Breath Volume	1.250	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.132	02:37
Breath Volume	1.558	02:37
Air Blank	0.000	02:38
Control Test	0.078	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 06/10/2021 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: ARISTIL

DATE/TIME OF TEST: 05/30/2021 04:44:38

Diagnostics Check		OK	04:34
Air Blank		0.000	04:34
Control Test		0.078	04:35
Air Blank		0.000	04:35
Subject Sample	# 1	0.220	04:36
Breath Volume		1.722	04:36
Air Blank		0.000	04:37
Air Blank		0.000	04:39
Subject Sample	# 2	0.196	04:40
Breath Volume		1.359	04:40
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 3	0.217	04:43
Breath Volume		1.656	04:43
Air Blank		0.000	04:43
Control Test		0.077	04:44
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 06/10/2021 10:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTSON,CHRIS,J
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: CONRAD,LUCAS
ARRESTING OFFICER AGENCY: BISCAYNE NATL PARK
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: BUI
ARREST TIME: 17:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: MATRANCO

DATE/TIME OF TEST: 05/30/2021 18:34:50

Diagnostics Check	OK	18:31	
Air Blank	0.000	18:32	
Control Test	0.078	18:32	
Air Blank	0.000	18:33	
Subject Sample	# 1	18:33	Subject Test Refused
Breath Volume	0.000	18:33	
Air Blank	0.000	18:33	
Control Test	0.079	18:34	
Air Blank	0.000	18:34	
Diagnostics Check	OK	18:34	

Data Download Date/Time: 06/10/2021 10:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: MONTENEGRO

DATE/TIME OF TEST: 05/31/2021 02:18:01

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.078	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.185	02:13
Breath Volume	1.648	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.182	02:16
Breath Volume	1.382	02:16
Air Blank	0.000	02:17
Control Test	0.077	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:18

Data Download Date/Time: 06/10/2021 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONES,SHAUN,G
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: LONES,SHAUN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 11:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:10

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/31/2021 12:39:19

Diagnostics Check	OK	12:31
Air Blank	0.000	12:31
Control Test	0.079	12:31
Air Blank	0.000	12:32
Subject Sample # 1	0.193	12:34
Breath Volume	1.644	12:34
Air Blank	0.000	12:35
Air Blank	0.000	12:37
Subject Sample # 2	0.191	12:37
Breath Volume	2.789	12:37
Air Blank	0.000	12:38
Control Test	0.076	12:38
Air Blank	0.000	12:39
Diagnostics Check	OK	12:39

Data Download Date/Time: 06/10/2021 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,LANCE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:58

SUBJECT'S LAST NAME: SERRANO CORDERO

DATE/TIME OF TEST: 05/31/2021 17:24:59

Diagnostics Check	OK	17:18
Air Blank	0.000	17:18
Control Test	0.078	17:19
Air Blank	0.000	17:19
Subject Sample # 1	0.270	17:20
Breath Volume	3.023	17:20
Air Blank	0.000	17:21
Air Blank	0.000	17:22
Subject Sample # 2	0.276	17:23
Breath Volume	2.617	17:23
Air Blank	0.000	17:24
Control Test	0.078	17:24
Air Blank	0.000	17:24
Diagnostics Check	OK	17:24

Data Download Date/Time: 06/10/2021 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,ALEXANDRIA,N
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HRADECKY,ALEXANDRIA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: FULIN

DATE/TIME OF TEST: 05/31/2021 20:46:37

Diagnostics Check	OK	20:37
Air Blank	0.000	20:38
Control Test	0.078	20:38
Air Blank	0.000	20:39
Subject Sample # 1	0.188	20:40
Breath Volume	2.144	20:40
Air Blank	0.000	20:40
Air Blank	0.000	20:42
Subject Sample # 2	0.192	20:44
Breath Volume	1.390	20:44
Air Blank	0.000	20:45
Control Test	0.078	20:45
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

Data Download Date/Time: 06/10/2021 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGGAN,MARLON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: COCHRAN,CHRISTIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 05/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 10:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:52

SUBJECT'S LAST NAME: MILANESI-MROZ

DATE/TIME OF TEST: 05/15/2021 11:20:16

Diagnostics Check	OK	11:13
Air Blank	0.000	11:14
Control Test	0.080	11:14
Air Blank	0.000	11:15
Subject Sample # 1	0.000	11:15
Breath Volume	2.285	11:15
Air Blank	0.000	11:16
Air Blank	0.000	11:18
Subject Sample # 2	0.000	11:18
Breath Volume	2.210	11:18
Air Blank	0.000	11:19
Control Test	0.079	11:19
Air Blank	0.000	11:20
Diagnostics Check	OK	11:20

Data Download Date/Time: 06/25/2021 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,EDISON,B
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: PEREZ,JULIAN
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: CLAYPOOL

DATE/TIME OF TEST: 05/06/2021 11:50:55

Diagnostics Check		OK	11:43
Air Blank		0.000	11:44
Control Test		0.077	11:44
Air Blank		0.000	11:44
Subject Sample	# 1	0.265	11:45
Breath Volume		1.195	11:45
Air Blank		0.000	11:46
Air Blank		0.000	11:48
Subject Sample	# 2	0.277	11:49
Breath Volume		1.714	11:49
Air Blank		0.000	11:49
Control Test		0.077	11:50
Air Blank		0.000	11:50
Diagnostics Check		OK	11:50

Data Download Date/Time: 05/30/2021 20:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,EDISON,B
BREATH TEST OPERATOR AGENCY: SOUTH MIAMI PD
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: VILLAPINZON

DATE/TIME OF TEST: 05/14/2021 05:12:47

Diagnostics Check	OK	05:04	
Air Blank	0.000	05:04	
Control Test	0.076	05:05	
Air Blank	0.000	05:05	
Subject Sample	# 1	05:08	No Sample Provided
Breath Volume	0.000	05:08	
Air Blank	0.000	05:09	
Air Blank	0.000	05:11	
Subject Sample	# 2	05:11	Subject Test Refused
Breath Volume	0.000	05:11	
Air Blank	0.000	05:11	
Control Test	0.077	05:12	
Air Blank	0.000	05:12	
Diagnostics Check	OK	05:12	

Data Download Date/Time: 05/30/2021 20:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNOZ,ROBERT,A
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: MUNOZ,ROBERT
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: ARDILA RUEDA

DATE/TIME OF TEST: 05/23/2021 20:31:04

Diagnostics Check	OK	20:24
Air Blank	0.000	20:24
Control Test	0.077	20:25
Air Blank	0.000	20:25
Subject Sample # 1	0.143	20:26
Breath Volume	1.996	20:26
Air Blank	0.000	20:26
Air Blank	0.000	20:28
Subject Sample # 2	0.141	20:29
Breath Volume	1.320	20:29
Air Blank	0.000	20:30
Control Test	0.078	20:30
Air Blank	0.000	20:30
Diagnostics Check	OK	20:31

Data Download Date/Time: 05/30/2021 20:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,EDISON,B
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: HARVEY,CRAIG
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: CRESPO

DATE/TIME OF TEST: 05/29/2021 05:20:13

Diagnostics Check		OK	05:10
Air Blank		0.000	05:10
Control Test		0.077	05:11
Air Blank		0.000	05:11
Subject Sample	# 1	0.062	05:14
Breath Volume		2.011	05:14
Air Blank		0.000	05:14
Air Blank		0.000	05:16
Subject Sample	# 2	0.055	05:18
Breath Volume		1.363	05:18
Air Blank		0.000	05:19
Control Test		0.075	05:19
Air Blank		0.000	05:20
Diagnostics Check		OK	05:20

Data Download Date/Time: 05/30/2021 20:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'DOR,ROBERT,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: COJON ORTIZ

DATE/TIME OF TEST: 05/17/2021 05:31:17

Diagnostics Check	OK	05:22
Air Blank	0.000	05:23
Control Test	0.078	05:23
Air Blank	0.000	05:24
Subject Sample # 1	0.198	05:26
Breath Volume	3.472	05:26
Air Blank	0.000	05:27
Air Blank	0.000	05:28
Subject Sample # 2	0.195	05:29
Breath Volume	1.652	05:29
Air Blank	0.000	05:30
Control Test	0.078	05:30
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

Data Download Date/Time: 06/05/2021 22:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'DOR,ROBERT,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: SIUGZDINIENE

DATE/TIME OF TEST: 05/27/2021 23:28:55

Diagnostics Check	OK	23:18	
Air Blank	0.000	23:18	
Control Test	0.080	23:18	
Air Blank	0.000	23:19	
Subject Sample	# 1 0.280	23:20	
Breath Volume	2.203	23:20	
Air Blank	0.000	23:21	
Air Blank	0.000	23:22	
Subject Sample	# 2 0.235	23:24	
Breath Volume	1.804	23:24	
Air Blank	0.000	23:24	
Air Blank	0.000	23:26	
Subject Sample	# 3 0.268	23:27	
Breath Volume	2.257	23:27	
Air Blank	0.000	23:27	
Control Test	0.074	23:28	Control Outside Tolerance
Air Blank	0.000	23:28	

Data Download Date/Time: 06/05/2021 22:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: ROSALES

DATE/TIME OF TEST: 05/15/2021 01:39:07

Diagnostics Check		OK	01:32
Air Blank		0.000	01:33
Control Test		0.079	01:33
Air Blank		0.000	01:34
Subject Sample	# 1	0.213	01:34
Breath Volume		2.093	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:36
Subject Sample	# 2	0.212	01:37
Breath Volume		2.128	01:37
Air Blank		0.000	01:38
Control Test		0.077	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 06/11/2021 14:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SOLOWAY,ALAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: ROSIER

DATE/TIME OF TEST: 05/15/2021 03:52:54

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:48	
Control Test	0.079	03:49	
Air Blank	0.000	03:49	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:51	
Control Test	0.078	03:52	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:52	

Data Download Date/Time: 06/11/2021 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HARTY, WARREN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: SIMILORME

DATE/TIME OF TEST: 05/15/2021 05:48:00

Diagnostics Check	OK	05:41
Air Blank	0.000	05:41
Control Test	0.080	05:41
Air Blank	0.000	05:42
Subject Sample # 1	0.151	05:43
Breath Volume	1.863	05:43
Air Blank	0.000	05:44
Air Blank	0.000	05:45
Subject Sample # 2	0.154	05:46
Breath Volume	1.742	05:46
Air Blank	0.000	05:47
Control Test	0.078	05:47
Air Blank	0.000	05:47
Diagnostics Check	OK	05:48

Data Download Date/Time: 06/11/2021 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER, CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: DAYAN

DATE/TIME OF TEST: 05/15/2021 22:38:46

Diagnostics Check	OK	22:32
Air Blank	0.000	22:32
Control Test	0.079	22:32
Air Blank	0.000	22:33
Subject Sample # 1	0.110	22:33
Breath Volume	1.398	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.105	22:36
Breath Volume	1.136	22:36
Air Blank	0.000	22:37
Control Test	0.079	22:38
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 06/11/2021 14:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: VALENTE GONZALEZ

DATE/TIME OF TEST: 05/16/2021 19:46:20

Diagnostics Check	OK	19:35
Air Blank	0.000	19:36
Control Test	0.080	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.224	19:38
Breath Volume	1.785	19:38
Air Blank	0.000	19:39
Air Blank	0.000	19:40
Subject Sample # 2	0.203	19:41
Breath Volume	1.871	19:41
Air Blank	0.000	19:42
Air Blank	0.000	19:43
Subject Sample # 3	0.211	19:44
Breath Volume	2.167	19:44
Air Blank	0.000	19:45
Control Test	0.078	19:45
Air Blank	0.000	19:46
Diagnostics Check	OK	19:46

Data Download Date/Time: 06/11/2021 14:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STILES,MEGAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 05/17/2021 00:31:55

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.079	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.149	00:27
Breath Volume	2.156	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:29
Subject Sample # 2	0.151	00:30
Breath Volume	1.277	00:30
Air Blank	0.000	00:30
Control Test	0.076	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 06/11/2021 14:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HERNANDEZ,EDWIN
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 05/17/2021 20:08:59

Diagnostics Check	OK	20:02
Air Blank	0.000	20:02
Control Test	0.080	20:02
Air Blank	0.000	20:03
Subject Sample # 1	0.000	20:04
Breath Volume	3.277	20:04
Air Blank	0.000	20:04
Air Blank	0.000	20:06
Subject Sample # 2	0.000	20:07
Breath Volume	3.730	20:07
Air Blank	0.000	20:08
Control Test	0.079	20:08
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

Data Download Date/Time: 06/11/2021 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NALERIO,LUIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 05/17/2021 22:52:08

Diagnostics Check		OK	22:45
Air Blank		0.000	22:45
Control Test		0.080	22:45
Air Blank		0.000	22:46
Subject Sample	# 1	0.208	22:47
Breath Volume		2.515	22:47
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.203	22:50
Breath Volume		2.679	22:50
Air Blank		0.000	22:51
Control Test		0.078	22:51
Air Blank		0.000	22:52
Diagnostics Check		OK	22:52

Data Download Date/Time: 06/11/2021 14:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VAN CAMP,JEFFERY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: SHELLY

DATE/TIME OF TEST: 05/19/2021 17:40:30

Diagnostics Check	OK	17:34
Air Blank	0.000	17:34
Control Test	0.080	17:34
Air Blank	0.000	17:35
Subject Sample # 1	0.000	17:36
Breath Volume	2.812	17:36
Air Blank	0.000	17:36
Air Blank	0.000	17:38
Subject Sample # 2	0.000	17:39
Breath Volume	2.750	17:39
Air Blank	0.000	17:39
Control Test	0.080	17:39
Air Blank	0.000	17:40
Diagnostics Check	OK	17:40

Data Download Date/Time: 06/11/2021 14:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: THOMAS, MICAH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 05/19/2021 23:21:45

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.080	23:16
Air Blank	0.000	23:16
Subject Sample # 1	0.116	23:17
Breath Volume	2.011	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.120	23:19
Breath Volume	3.847	23:19
Air Blank	0.000	23:20
Control Test	0.079	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 06/11/2021 14:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: VAIL VELASQUEZ

DATE/TIME OF TEST: 05/20/2021 23:22:05

Diagnostics Check	OK	23:14
Air Blank	0.000	23:15
Control Test	0.081	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.190	23:17
Breath Volume	1.234	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:19
Subject Sample # 2	0.189	23:20
Breath Volume	2.593	23:20
Air Blank	0.000	23:21
Control Test	0.077	23:21
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 06/11/2021 14:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH, GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: VELASQUEZ VASQUEZ

DATE/TIME OF TEST: 05/22/2021 00:24:55

Diagnostics Check	OK	00:18
Air Blank	0.000	00:18
Control Test	0.081	00:18
Air Blank	0.000	00:19
Subject Sample # 1	0.262	00:19
Breath Volume	3.843	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.254	00:22
Breath Volume	2.902	00:22
Air Blank	0.000	00:23
Control Test	0.078	00:24
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 06/11/2021 14:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: FRANCISCO MANUEL

DATE/TIME OF TEST: 05/22/2021 01:19:04

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.081	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.170	01:13
Breath Volume	2.175	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.169	01:17
Breath Volume	3.019	01:17
Air Blank	0.000	01:18
Control Test	0.078	01:18
Air Blank	0.000	01:18
Diagnostics Check	OK	01:19

Data Download Date/Time: 06/11/2021 14:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HARTY, WARREN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:19

SUBJECT'S LAST NAME: POTEAU

DATE/TIME OF TEST: 05/22/2021 06:50:50

Diagnostics Check	OK	06:44
Air Blank	0.000	06:44
Control Test	0.080	06:45
Air Blank	0.000	06:45
Subject Sample # 1	0.146	06:46
Breath Volume	1.589	06:46
Air Blank	0.000	06:46
Air Blank	0.000	06:48
Subject Sample # 2	0.161	06:49
Breath Volume	2.742	06:49
Air Blank	0.000	06:49
Control Test	0.079	06:50
Air Blank	0.000	06:50
Diagnostics Check	OK	06:50

Data Download Date/Time: 06/11/2021 14:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ANDRE, GAETGENS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:23

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/22/2021 08:58:55

Diagnostics Check	OK	08:50
Air Blank	0.000	08:50
Control Test	0.080	08:51
Air Blank	0.000	08:51
Subject Sample # 1	0.058	08:52
Breath Volume	1.339	08:52
Air Blank	0.000	08:53
Air Blank	0.000	08:55
Subject Sample # 2	0.061	08:57
Breath Volume	2.789	08:57
Air Blank	0.000	08:57
Control Test	0.080	08:58
Air Blank	0.000	08:58
Diagnostics Check	OK	08:58

Data Download Date/Time: 06/11/2021 14:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PRUD'HOMME,DANIEL
ARRESTING OFFICER AGENCY: North Palm Beach
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: BEAVINER

DATE/TIME OF TEST: 05/23/2021 00:43:37

Diagnostics Check	OK	00:36
Air Blank	0.000	00:37
Control Test	0.081	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.000	00:39
Breath Volume	3.769	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.000	00:42
Breath Volume	3.914	00:42
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 06/11/2021 14:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR, NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 05/23/2021 02:58:36

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.132	02:53
Breath Volume	1.777	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.132	02:56
Breath Volume	1.945	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 06/11/2021 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MACHIM, NICODEMUS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: CHAVES WICHES

DATE/TIME OF TEST: 05/23/2021 05:14:40

Diagnostics Check	OK	05:07
Air Blank	0.000	05:07
Control Test	0.081	05:08
Air Blank	0.000	05:08
Subject Sample # 1	0.172	05:09
Breath Volume	1.800	05:09
Air Blank	0.000	05:10
Air Blank	0.000	05:11
Subject Sample # 2	0.170	05:12
Breath Volume	2.535	05:12
Air Blank	0.000	05:13
Control Test	0.079	05:14
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

Data Download Date/Time: 06/11/2021 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DILLARD,DANIEL
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: FICKEN

DATE/TIME OF TEST: 05/23/2021 13:19:52

Diagnostics Check	OK	13:13
Air Blank	0.000	13:13
Control Test	0.080	13:14
Air Blank	0.000	13:14
Subject Sample # 1	0.000	13:15
Breath Volume	2.421	13:15
Air Blank	0.000	13:15
Air Blank	0.000	13:17
Subject Sample # 2	0.000	13:18
Breath Volume	1.601	13:18
Air Blank	0.000	13:18
Control Test	0.080	13:19
Air Blank	0.000	13:19
Diagnostics Check	OK	13:19

Data Download Date/Time: 06/11/2021 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JONES,KENNETH
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: ROLLINS

DATE/TIME OF TEST: 05/23/2021 14:32:59

Diagnostics Check	OK	14:26
Air Blank	0.000	14:26
Control Test	0.080	14:26
Air Blank	0.000	14:27
Subject Sample # 1	0.153	14:28
Breath Volume	1.445	14:28
Air Blank	0.000	14:29
Air Blank	0.000	14:30
Subject Sample # 2	0.158	14:31
Breath Volume	1.957	14:31
Air Blank	0.000	14:32
Control Test	0.079	14:32
Air Blank	0.000	14:32
Diagnostics Check	OK	14:32

Data Download Date/Time: 06/11/2021 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: SCHMIDT

DATE/TIME OF TEST: 05/23/2021 19:48:20

Diagnostics Check	OK	19:41
Air Blank	0.000	19:42
Control Test	0.081	19:42
Air Blank	0.000	19:43
Subject Sample # 1	0.000	19:43
Breath Volume	2.976	19:43
Air Blank	0.000	19:44
Air Blank	0.000	19:46
Subject Sample # 2	0.000	19:46
Breath Volume	3.460	19:46
Air Blank	0.000	19:47
Control Test	0.080	19:47
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 06/11/2021 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RAMIREZ,DAVID
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 11:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 05/24/2021 12:41:44

Diagnostics Check	OK	12:31
Air Blank	0.000	12:31
Control Test	0.081	12:31
Air Blank	0.000	12:32
Subject Sample # 1	0.397	12:33
Breath Volume	3.253	12:33
Air Blank	0.000	12:34
Air Blank	0.000	12:36
Subject Sample # 2	0.371	12:36
Breath Volume	2.988	12:36
Air Blank	0.000	12:37
Air Blank	0.000	12:39
Subject Sample # 3	0.356	12:39
Breath Volume	3.195	12:39
Air Blank	0.000	12:40
Control Test	0.078	12:41
Air Blank	0.000	12:41
Diagnostics Check	OK	12:41

Data Download Date/Time: 06/11/2021 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROBINSON,HAROLD
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: MORAN

DATE/TIME OF TEST: 05/24/2021 19:58:36

Diagnostics Check	OK	19:51
Air Blank	0.000	19:52
Control Test	0.079	19:52
Air Blank	0.000	19:53
Subject Sample # 1	0.159	19:53
Breath Volume	3.585	19:53
Air Blank	0.000	19:54
Air Blank	0.000	19:56
Subject Sample # 2	0.170	19:56
Breath Volume	1.835	19:56
Air Blank	0.000	19:57
Control Test	0.079	19:57
Air Blank	0.000	19:58
Diagnostics Check	OK	19:58

Data Download Date/Time: 06/11/2021 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DANDREA, ADAM
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:16

SUBJECT'S LAST NAME: WOMELSDORF

DATE/TIME OF TEST: 05/25/2021 10:46:59

Diagnostics Check	OK	10:40
Air Blank	0.000	10:41
Control Test	0.080	10:41
Air Blank	0.000	10:42
Subject Sample # 1	0.000	10:42
Breath Volume	1.816	10:42
Air Blank	0.000	10:43
Air Blank	0.000	10:45
Subject Sample # 2	0.000	10:45
Breath Volume	1.699	10:45
Air Blank	0.000	10:46
Control Test	0.079	10:46
Air Blank	0.000	10:46
Diagnostics Check	OK	10:46

Data Download Date/Time: 06/11/2021 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHARELUS,OBED
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: AUGUSTIN

DATE/TIME OF TEST: 05/25/2021 21:26:43

Diagnostics Check		OK	21:19
Air Blank		0.000	21:20
Control Test		0.081	21:20
Air Blank		0.000	21:21
Subject Sample	# 1	0.234	21:21
Breath Volume		2.707	21:21
Air Blank		0.000	21:22
Air Blank		0.000	21:24
Subject Sample	# 2	0.241	21:24
Breath Volume		2.875	21:24
Air Blank		0.000	21:25
Control Test		0.079	21:26
Air Blank		0.000	21:26
Diagnostics Check		OK	21:26

Data Download Date/Time: 06/11/2021 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: OPORTO,KYLE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: MALDONADO TORRES

DATE/TIME OF TEST: 05/26/2021 18:50:14

Diagnostics Check	OK	18:42
Air Blank	0.000	18:42
Control Test	0.080	18:42
Air Blank	0.000	18:43
Subject Sample # 1	0.000	18:45
Breath Volume	2.933	18:45
Air Blank	0.000	18:46
Air Blank	0.000	18:48
Subject Sample # 2	0.000	18:48
Breath Volume	2.851	18:48
Air Blank	0.000	18:49
Control Test	0.080	18:49
Air Blank	0.000	18:50
Diagnostics Check	OK	18:50

Data Download Date/Time: 06/11/2021 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SOLOWAY, ALAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: SIMEON

DATE/TIME OF TEST: 05/27/2021 01:19:44

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.080	01:14
Air Blank	0.000	01:14
Subject Sample # 1	0.162	01:14
Breath Volume	1.527	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.168	01:17
Breath Volume	1.914	01:17
Air Blank	0.000	01:18
Control Test	0.079	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 06/11/2021 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH, GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: BARR

DATE/TIME OF TEST: 05/27/2021 23:37:19

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.080	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.000	23:32
Breath Volume	3.175	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.000	23:35
Breath Volume	2.347	23:35
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 06/11/2021 14:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR, NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: JIMENEZ JIMENEZ

DATE/TIME OF TEST: 05/28/2021 22:52:59

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.080	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.224	22:48
Breath Volume	1.785	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.232	22:51
Breath Volume	2.832	22:51
Air Blank	0.000	22:51
Control Test	0.077	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 06/11/2021 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CASAS, JAVIER
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: HASAN

DATE/TIME OF TEST: 05/29/2021 01:49:25

Diagnostics Check		OK	01:40
Air Blank		0.000	01:41
Control Test		0.080	01:41
Air Blank		0.000	01:42
Subject Sample	# 1	0.123	01:44
Breath Volume		1.160	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:46
Subject Sample	# 2	0.126	01:47
Breath Volume		1.468	01:47
Air Blank		0.000	01:48
Control Test		0.076	01:48
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 06/11/2021 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KYSOR, MITCHELL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: EARSLEY

DATE/TIME OF TEST: 05/29/2021 03:06:00

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.125	02:58
Breath Volume	1.128	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:00
Subject Sample # 2	0.114	03:03
Breath Volume	1.292	03:03
Air Blank	0.000	03:05
Control Test	0.078	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:06

Data Download Date/Time: 06/11/2021 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 05/29/2021 03:33:00

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.076	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.116	03:27
Breath Volume	2.222	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.115	03:31
Breath Volume	1.839	03:31
Air Blank	0.000	03:32
Control Test	0.077	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:33

Data Download Date/Time: 06/11/2021 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: PEREZ GAMEZ

DATE/TIME OF TEST: 05/29/2021 20:26:45

Diagnostics Check		OK	20:19
Air Blank		0.000	20:19
Control Test		0.080	20:20
Air Blank		0.000	20:20
Subject Sample	# 1	0.226	20:21
Breath Volume		2.285	20:21
Air Blank		0.000	20:23
Air Blank		0.000	20:24
Subject Sample	# 2	0.220	20:24
Breath Volume		2.082	20:24
Air Blank		0.000	20:25
Control Test		0.076	20:26
Air Blank		0.000	20:26
Diagnostics Check		OK	20:26

Data Download Date/Time: 06/11/2021 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: ELIGIO ANDRACA

DATE/TIME OF TEST: 05/30/2021 00:49:57

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.079	00:44
Air Blank	0.000	00:44
Subject Sample # 1	0.142	00:45
Breath Volume	1.523	00:45
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.140	00:48
Breath Volume	2.156	00:48
Air Blank	0.000	00:48
Control Test	0.078	00:49
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

Data Download Date/Time: 06/11/2021 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DEPEINE,KENS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: PIERRE LOUIS

DATE/TIME OF TEST: 05/30/2021 20:05:12

Diagnostics Check		OK	19:53	
Air Blank		0.000	19:53	
Control Test		0.080	19:54	
Air Blank		0.000	19:54	
Subject Sample	# 1	0.053	19:57	Volume Not Met
Breath Volume		0.035	19:57	
Air Blank		0.000	19:58	
Air Blank		0.000	20:00	
Subject Sample	# 2	0.075	20:03	Volume Not Met
Breath Volume		0.019	20:03	
Air Blank		0.000	20:04	
Control Test		0.079	20:04	
Air Blank		0.000	20:05	
Diagnostics Check		OK	20:05	

Data Download Date/Time: 06/11/2021 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DEPEINE,KENS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: PIERRE-LOUIS

DATE/TIME OF TEST: 05/30/2021 20:20:05

Diagnostics Check		OK	20:08	
Air Blank		0.000	20:08	
Control Test		0.078	20:08	
Air Blank		0.000	20:09	
Subject Sample	# 1	0.073	20:12	Volume Not Met
Breath Volume		0.039	20:12	
Air Blank		0.000	20:13	
Air Blank		0.000	20:15	
Subject Sample	# 2	0.066	20:18	Volume Not Met
Breath Volume		0.121	20:18	
Air Blank		0.000	20:19	
Control Test		0.077	20:19	
Air Blank		0.000	20:20	
Diagnostics Check		OK	20:20	

Data Download Date/Time: 06/11/2021 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: LOPEZAGUILAR

DATE/TIME OF TEST: 05/30/2021 23:48:07

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.182	23:43
Breath Volume	2.031	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:45
Subject Sample # 2	0.188	23:46
Breath Volume	2.339	23:46
Air Blank	0.000	23:47
Control Test	0.077	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 06/11/2021 14:54