

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYA,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: GONZALEZ,FERNANDO
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 04/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CAUDLE

DATE/TIME OF TEST: 05/05/2021 23:28:07

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.151	23:23
Breath Volume	1.144	23:23
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.149	23:26
Breath Volume	1.148	23:26
Air Blank	0.000	23:27
Control Test	0.078	23:27
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 05/24/2021 15:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIMES, DENISE, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DANLEY, JEDIDIAH
ARRESTING OFFICER AGENCY: Calhoun County SO
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 08:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:32

SUBJECT'S LAST NAME: HOBBY

DATE/TIME OF TEST: 05/10/2021 10:04:30

Diagnostics Check	OK	09:52	
Air Blank	0.000	09:53	
Control Test	0.081	09:53	
Air Blank	0.000	09:53	
Subject Sample	# 1	09:57	No Sample Provided
Breath Volume	0.000	09:57	
Air Blank	0.000	09:57	
Air Blank	0.000	09:59	
Subject Sample	# 2	0.000	10:00
Breath Volume	2.945	10:00	
Air Blank	0.000	10:00	
Air Blank	0.000	10:02	
Subject Sample	# 3	0.000	10:03
Breath Volume	3.796	10:03	
Air Blank	0.000	10:03	
Control Test	0.080	10:03	
Air Blank	0.000	10:04	
Diagnostics Check	OK	10:04	

Data Download Date/Time: 05/12/2021 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUBOSE,ROGER,C
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: SCOTT,TRAVIS
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: SCHOUST

DATE/TIME OF TEST: 05/30/2021 19:48:13

Diagnostics Check	OK	19:37
Air Blank	0.000	19:37
Control Test	0.080	19:38
Air Blank	0.000	19:38
Subject Sample # 1	0.159	19:40
Breath Volume	1.226	19:40
Air Blank	0.000	19:41
Air Blank	0.000	19:43
Subject Sample # 2	0.182	19:43
Breath Volume	2.121	19:43
Air Blank	0.000	19:44
Air Blank	0.000	19:46
Subject Sample # 3	0.179	19:46
Breath Volume	1.816	19:46
Air Blank	0.000	19:47
Control Test	0.078	19:47
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 06/09/2021 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: SMALL

DATE/TIME OF TEST: 05/12/2021 03:57:50

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:54	
Control Test	0.081	03:55	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank	0.000	03:56	
Control Test	0.080	03:57	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

Data Download Date/Time: 05/20/2021 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONETZ,JONATHAN,
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: KONETZ,JONATHAN
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD
LAST AGENCY INSPECTION: 04/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: LEGNANI

DATE/TIME OF TEST: 05/18/2021 00:33:43

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:31	
Control Test	0.080	00:31	
Air Blank	0.000	00:32	
Subject Sample	# 1	00:32	Subject Test Refused
Breath Volume	0.000	00:32	
Air Blank	0.000	00:32	
Control Test	0.079	00:33	
Air Blank	0.000	00:33	
Diagnostics Check	OK	00:33	

Data Download Date/Time: 05/20/2021 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: SANTOS CAMACHO

DATE/TIME OF TEST: 05/29/2021 21:45:37

Diagnostics Check	OK	21:39
Air Blank	0.000	21:39
Control Test	0.081	21:39
Air Blank	0.000	21:40
Subject Sample # 1	0.080	21:40
Breath Volume	1.421	21:40
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.090	21:43
Breath Volume	1.500	21:43
Air Blank	0.000	21:44
Control Test	0.081	21:45
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 06/08/2021 05:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROIG,DAWN,M
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: ROIG,DAWN
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 05/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:36

SUBJECT'S LAST NAME: AGUILUZ MANCIA

DATE/TIME OF TEST: 05/30/2021 07:04:39

Diagnostics Check	OK	07:01	
Air Blank	0.000	07:02	
Control Test	0.081	07:02	
Air Blank	0.000	07:03	
Subject Sample	# 1	07:03	Subject Test Refused
Breath Volume	0.000	07:03	
Air Blank	0.000	07:03	
Control Test	0.080	07:04	
Air Blank	0.000	07:04	
Diagnostics Check	OK	07:04	

Data Download Date/Time: 06/08/2021 05:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,RUBEN,G
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: GOMEZ,RUBEN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: BUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: VILLALOBOS

DATE/TIME OF TEST: 05/30/2021 22:21:39

Diagnostics Check	OK	22:17	
Air Blank	0.000	22:17	
Control Test	0.082	22:17	
Air Blank	0.000	22:18	
Subject Sample	# 1	22:20	Subject Test Refused
Breath Volume	0.000	22:20	
Air Blank	0.000	22:20	
Control Test	0.081	22:21	
Air Blank	0.000	22:21	
Diagnostics Check	OK	22:21	

Data Download Date/Time: 06/08/2021 05:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 05/08/2021 03:41:33

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.079	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.178	03:36
Breath Volume	2.390	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.180	03:39
Breath Volume	2.359	03:39
Air Blank	0.000	03:40
Control Test	0.079	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 05/18/2021 06:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 05/11/2021 23:34:42

Diagnostics Check	OK	23:24	
Air Blank	0.000	23:24	
Control Test	0.079	23:24	
Air Blank	0.000	23:25	
Subject Sample	# 1	23:28	No Sample Provided
Breath Volume	0.000	23:28	
Air Blank	0.000	23:28	
Air Blank	0.000	23:30	
Subject Sample	# 2	23:33	Subject Test Refused
Breath Volume	0.000	23:33	
Air Blank	0.000	23:33	
Control Test	0.080	23:34	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

Data Download Date/Time: 05/18/2021 06:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: WALLER

DATE/TIME OF TEST: 05/23/2021 00:51:26

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.080	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.000	00:47
Breath Volume	2.027	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.000	00:50
Breath Volume	1.488	00:50
Air Blank	0.000	00:50
Control Test	0.079	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 06/08/2021 05:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCEGUERA,YASMANI,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: OCEGUERA,YASMANI
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: PACHON GONZALEZ

DATE/TIME OF TEST: 05/23/2021 05:51:12

Diagnostics Check	OK	05:44
Air Blank	0.000	05:44
Control Test	0.079	05:45
Air Blank	0.000	05:45
Subject Sample # 1	0.238	05:45
Breath Volume	1.308	05:45
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample # 2	0.240	05:49
Breath Volume	2.277	05:49
Air Blank	0.000	05:50
Control Test	0.079	05:50
Air Blank	0.000	05:51
Diagnostics Check	OK	05:51

Data Download Date/Time: 06/08/2021 05:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARNEY,ANGELA,Y
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KEARNEY,ANGELA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: WILCOX

DATE/TIME OF TEST: 05/28/2021 00:47:43

Diagnostics Check	OK	00:44	
Air Blank	0.000	00:45	
Control Test	0.079	00:45	
Air Blank	0.000	00:46	
Subject Sample	# 1	00:46	Subject Test Refused
Breath Volume	0.000	00:46	
Air Blank	0.000	00:46	
Control Test	0.079	00:47	
Air Blank	0.000	00:47	
Diagnostics Check	OK	00:47	

Data Download Date/Time: 06/08/2021 05:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FHP TROOP A
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONEY,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MALONEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1327124
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: BATYE

DATE/TIME OF TEST: 05/28/2021 05:15:47

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.080	05:10
Air Blank	0.000	05:10
Subject Sample # 1	0.000	05:11
Breath Volume	3.410	05:11
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.000	05:14
Breath Volume	3.375	05:14
Air Blank	0.000	05:14
Control Test	0.076	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 06/17/2021 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE, MICHAEL,
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: RICE, MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 17:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: IRVINE

DATE/TIME OF TEST: 05/01/2021 18:47:31

Diagnostics Check	OK	18:44	
Air Blank	0.000	18:45	
Control Test	0.081	18:45	
Air Blank	0.000	18:45	
Subject Sample	# 1	18:46	Subject Test Refused
Breath Volume	0.000	18:46	
Air Blank	0.000	18:46	
Control Test	0.081	18:46	
Air Blank	0.000	18:47	
Diagnostics Check	OK	18:47	

Data Download Date/Time: 05/25/2021 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALTAIS,ANDREW,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DOWNING,CADEN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 12:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:59

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 05/02/2021 13:51:21

Diagnostics Check	OK	13:44
Air Blank	0.000	13:45
Control Test	0.076	13:45
Air Blank	0.000	13:46
Subject Sample # 1	0.243	13:46
Breath Volume	5.207	13:46
Air Blank	0.000	13:47
Air Blank	0.000	13:49
Subject Sample # 2	0.241	13:49
Breath Volume	4.882	13:49
Air Blank	0.000	13:50
Control Test	0.077	13:50
Air Blank	0.000	13:51
Diagnostics Check	OK	13:51

Data Download Date/Time: 05/25/2021 21:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONOVAN,CHARLES,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: HETHERINGTON

DATE/TIME OF TEST: 05/08/2021 17:59:37

Diagnostics Check	OK	17:51	
Air Blank	0.000	17:51	
Control Test	0.082	17:51	
Air Blank	0.000	17:52	
Subject Sample	# 1	17:55	No Sample Provided
Breath Volume	0.000	17:55	
Air Blank	0.000	17:55	
Air Blank	0.000	17:57	
Subject Sample	# 2	17:58	Subject Test Refused
Breath Volume	0.000	17:58	
Air Blank	0.000	17:58	
Control Test	0.080	17:59	
Air Blank	0.000	17:59	
Diagnostics Check	OK	17:59	

Data Download Date/Time: 05/25/2021 21:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONOVAN,CHARLES,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 15:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: KNUCKLES

DATE/TIME OF TEST: 05/09/2021 16:21:23

Diagnostics Check	OK	16:15
Air Blank	0.000	16:15
Control Test	0.081	16:16
Air Blank	0.000	16:16
Subject Sample # 1	0.046	16:17
Breath Volume	2.816	16:17
Air Blank	0.000	16:17
Air Blank	0.000	16:19
Subject Sample # 2	0.041	16:19
Breath Volume	1.457	16:19
Air Blank	0.000	16:20
Control Test	0.081	16:20
Air Blank	0.000	16:21
Diagnostics Check	OK	16:21

Data Download Date/Time: 05/25/2021 21:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 18:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:19

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 05/12/2021 18:14:48

Diagnostics Check	OK	18:14	
Air Blank		18:14	RFI Detect
Air Blank		18:14	RFI Detect

Data Download Date/Time: 05/25/2021 21:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 18:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: JIMENEZ
DATE/TIME OF TEST: 05/12/2021 18:20:07

Diagnostics Check	OK	18:16	
Air Blank	0.000	18:17	
Control Test	0.081	18:17	
Air Blank	0.000	18:18	
Subject Sample	# 1	18:18	RFI Detect
Breath Volume	0.121	18:18	
Air Blank	0.000	18:19	
Control Test	0.080	18:19	
Air Blank	0.000	18:20	
Diagnostics Check	OK	18:20	

Data Download Date/Time: 05/25/2021 21:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORBITT,KYLE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DOWNING,CADEN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 14:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: LINDQUIST
DATE/TIME OF TEST: 05/14/2021 15:03:05

Diagnostics Check	OK	15:00	
Air Blank	0.000	15:00	
Control Test	0.081	15:01	
Air Blank	0.000	15:01	
Subject Sample	# 1	15:01	Subject Test Refused
Breath Volume	0.000	15:01	
Air Blank	0.000	15:02	
Control Test	0.081	15:02	
Air Blank	0.000	15:03	
Diagnostics Check	OK	15:03	

Data Download Date/Time: 05/25/2021 21:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORBIN,BRITTON,L
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: CORBIN,BRITTON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 17:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 05/14/2021 18:05:18

Diagnostics Check	OK	18:01	
Air Blank	0.000	18:01	
Control Test	0.081	18:02	
Air Blank	0.000	18:02	
Subject Sample	# 1	18:03	Subject Test Refused
Breath Volume	0.000	18:03	
Air Blank	0.000	18:04	
Control Test	0.081	18:04	
Air Blank	0.000	18:05	
Diagnostics Check	OK	18:05	

Data Download Date/Time: 05/25/2021 21:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWNING,CADEN,T
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DOWNING,CADEN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 05/15/2021 17:20:32

Diagnostics Check		OK	17:09	
Air Blank		0.000	17:09	
Control Test		0.081	17:09	
Air Blank		0.000	17:10	
Subject Sample	# 1	0.017	17:13	Volume Not Met
Breath Volume		0.566	17:13	
Air Blank		0.000	17:14	
Air Blank		0.000	17:15	
Subject Sample	# 2	0.016	17:19	Volume Not Met
Breath Volume		0.359	17:19	
Air Blank		0.000	17:19	
Control Test		0.080	17:19	
Air Blank		0.000	17:20	
Diagnostics Check		OK	17:20	

Data Download Date/Time: 05/25/2021 21:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWNING,CADEN,T
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DOWNING,CADEN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/01/2021
Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: PATEL
DATE/TIME OF TEST: 05/15/2021 17:30:05

Diagnostics Check	OK	17:23
Air Blank	0.000	17:24
Control Test	0.080	17:24
Air Blank	0.000	17:25
Subject Sample # 1	0.014	17:25
Breath Volume	1.988	17:25
Air Blank	0.000	17:26
Air Blank	0.000	17:28
Subject Sample # 2	0.014	17:28
Breath Volume	2.335	17:28
Air Blank	0.000	17:29
Control Test	0.081	17:29
Air Blank	0.000	17:30
Diagnostics Check	OK	17:30

Data Download Date/Time: 05/25/2021 21:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALTAIS,ANDREW,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MALTAIS,ANDREW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:04

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 05/16/2021 18:42:34

Diagnostics Check		OK	18:36
Air Blank		0.000	18:36
Control Test		0.082	18:36
Air Blank		0.000	18:37
Subject Sample	# 1	0.068	18:38
Breath Volume		2.382	18:38
Air Blank		0.000	18:38
Air Blank		0.000	18:40
Subject Sample	# 2	0.068	18:40
Breath Volume		2.519	18:40
Air Blank		0.000	18:41
Control Test		0.079	18:41
Air Blank		0.000	18:42
Diagnostics Check		OK	18:42

Data Download Date/Time: 05/25/2021 21:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALTAIS,ANDREW,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MALTAIS,ANDREW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 17:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:57

SUBJECT'S LAST NAME: BLANK

DATE/TIME OF TEST: 05/20/2021 18:32:44

Diagnostics Check	OK	18:24	
Air Blank	0.000	18:24	
Control Test	0.081	18:25	
Air Blank	0.000	18:25	
Subject Sample	# 1	18:28	No Sample Provided
Breath Volume	0.000	18:28	
Air Blank	0.000	18:29	
Air Blank	0.000	18:31	
Subject Sample	# 2	18:31	Subject Test Refused
Breath Volume	0.000	18:31	
Air Blank	0.000	18:31	
Control Test	0.081	18:32	
Air Blank	0.000	18:32	
Diagnostics Check	OK	18:32	

Data Download Date/Time: 05/25/2021 21:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DONOVAN,CHARLES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 05/23/2021 15:15:11

Diagnostics Check	OK	15:08
Air Blank	0.000	15:09
Control Test	0.080	15:09
Air Blank	0.000	15:10
Subject Sample # 1	0.035	15:10
Breath Volume	2.464	15:10
Air Blank	0.000	15:11
Air Blank	0.000	15:13
Subject Sample # 2	0.033	15:13
Breath Volume	3.074	15:13
Air Blank	0.000	15:14
Control Test	0.078	15:14
Air Blank	0.000	15:15
Diagnostics Check	OK	15:15

Data Download Date/Time: 05/25/2021 21:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAHR,DANIEL,E
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: CORBIN,BRITTON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: MINER

DATE/TIME OF TEST: 05/23/2021 21:30:05

Diagnostics Check	OK	21:22
Air Blank	0.000	21:22
Control Test	0.081	21:23
Air Blank	0.000	21:23
Subject Sample # 1	0.109	21:25
Breath Volume	1.546	21:25
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.118	21:28
Breath Volume	1.210	21:28
Air Blank	0.000	21:29
Control Test	0.079	21:29
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 05/25/2021 21:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONOVAN,CHARLES,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FULGHUM,JOSEPH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:38

SUBJECT'S LAST NAME: KING
DATE/TIME OF TEST: 05/29/2021 18:03:12

Diagnostics Check	OK	18:00	
Air Blank	0.000	18:00	
Control Test	0.079	18:01	
Air Blank	0.000	18:01	
Subject Sample	# 1	18:01	Subject Test Refused
Breath Volume	0.000	18:01	
Air Blank	0.000	18:02	
Control Test	0.079	18:02	
Air Blank	0.000	18:03	
Diagnostics Check	OK	18:03	

Data Download Date/Time: 06/22/2021 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALTAIS,ANDREW,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CORBITT,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: HOOD

DATE/TIME OF TEST: 05/29/2021 18:51:21

Diagnostics Check	OK	18:44
Air Blank	0.000	18:45
Control Test	0.080	18:45
Air Blank	0.000	18:46
Subject Sample # 1	0.162	18:46
Breath Volume	5.214	18:46
Air Blank	0.000	18:47
Air Blank	0.000	18:49
Subject Sample # 2	0.163	18:49
Breath Volume	4.628	18:49
Air Blank	0.000	18:50
Control Test	0.078	18:50
Air Blank	0.000	18:51
Diagnostics Check	OK	18:51

Data Download Date/Time: 06/22/2021 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTLETT,SHELTON,T
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: HAHR,DANIEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 18:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 05/29/2021 19:08:52

Diagnostics Check	OK	19:06	
Air Blank	0.000	19:06	
Control Test	0.080	19:06	
Air Blank	0.000	19:07	
Subject Sample	# 1	19:07	Subject Test Refused
Breath Volume	0.000	19:07	
Air Blank	0.000	19:07	
Control Test	0.079	19:08	
Air Blank	0.000	19:08	
Diagnostics Check	OK	19:08	

Data Download Date/Time: 06/22/2021 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWNING,CADEN,T
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:47

SUBJECT'S LAST NAME: PRYOR

DATE/TIME OF TEST: 05/30/2021 18:26:43

Diagnostics Check	OK	18:18	
Air Blank	0.000	18:18	
Control Test	0.082	18:18	
Air Blank	0.000	18:19	
Subject Sample	# 1	18:22	No Sample Provided
Breath Volume	0.000	18:22	
Air Blank	0.000	18:23	
Air Blank	0.000	18:25	
Subject Sample	# 2	18:25	Subject Test Refused
Breath Volume	0.000	18:25	
Air Blank	0.000	18:25	
Control Test	0.081	18:26	
Air Blank	0.000	18:26	
Diagnostics Check	OK	18:26	

Data Download Date/Time: 06/22/2021 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORBITT,KYLE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CORBITT,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: JERNIGAN

DATE/TIME OF TEST: 05/30/2021 19:21:00

Diagnostics Check	OK	19:14
Air Blank	0.000	19:15
Control Test	0.080	19:15
Air Blank	0.000	19:15
Subject Sample # 1	0.128	19:16
Breath Volume	3.187	19:16
Air Blank	0.000	19:17
Air Blank	0.000	19:18
Subject Sample # 2	0.120	19:19
Breath Volume	2.738	19:19
Air Blank	0.000	19:20
Control Test	0.077	19:20
Air Blank	0.000	19:20
Diagnostics Check	OK	19:21

Data Download Date/Time: 06/22/2021 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,GEORGE,N
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: POSAR,ADAM
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/21/2021
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 05/17/2021 07:29:23

Diagnostics Check	OK	07:23
Air Blank	0.000	07:23
Control Test	0.081	07:23
Air Blank	0.000	07:24
Subject Sample # 1	0.000	07:25
Breath Volume	3.175	07:25
Air Blank	0.000	07:25
Air Blank	0.000	07:27
Subject Sample # 2	0.000	07:27
Breath Volume	4.417	07:27
Air Blank	0.000	07:28
Control Test	0.079	07:28
Air Blank	0.000	07:29
Diagnostics Check	OK	07:29

Data Download Date/Time: 05/25/2021 19:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD,TRAVIS,L
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BASFORD,TRAVIS
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/21/2021
VIOLATION CODE: BUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/22/2021 20:26:00

Diagnostics Check	OK	20:19
Air Blank	0.000	20:19
Control Test	0.081	20:20
Air Blank	0.000	20:20
Subject Sample # 1	0.210	20:21
Breath Volume	4.082	20:21
Air Blank	0.000	20:22
Air Blank	0.000	20:23
Subject Sample # 2	0.219	20:24
Breath Volume	4.410	20:24
Air Blank	0.000	20:25
Control Test	0.079	20:25
Air Blank	0.000	20:25
Diagnostics Check	OK	20:26

Data Download Date/Time: 05/25/2021 19:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKER,JOHN,W
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: COKER,JOHN
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/21/2021
VIOLATION CODE: BUI
ARREST TIME: 16:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: KRUCZEK
DATE/TIME OF TEST: 05/29/2021 16:59:13

Diagnostics Check	OK	16:50	
Air Blank	0.000	16:50	
Control Test	0.080	16:51	
Air Blank	0.000	16:51	
Subject Sample	# 1	16:54	No Sample Provided
Breath Volume	0.000	16:54	
Air Blank	0.000	16:55	
Air Blank	0.000	16:57	
Subject Sample	# 2	16:57	Subject Test Refused
Breath Volume	0.000	16:57	
Air Blank	0.000	16:58	
Control Test	0.079	16:58	
Air Blank	0.000	16:59	
Diagnostics Check	OK	16:59	

Data Download Date/Time: 06/22/2021 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD,NICOLE,M
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BASFORD,NICOLE
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/21/2021
VIOLATION CODE: BUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 05/29/2021 17:09:00

Diagnostics Check	OK	17:01
Air Blank	0.000	17:01
Control Test	0.081	17:02
Air Blank	0.000	17:02
Subject Sample # 1	0.118	17:04
Breath Volume	2.160	17:04
Air Blank	0.000	17:05
Air Blank	0.000	17:07
Subject Sample # 2	0.105	17:07
Breath Volume	1.203	17:07
Air Blank	0.000	17:08
Control Test	0.078	17:08
Air Blank	0.000	17:08
Diagnostics Check	OK	17:09

Data Download Date/Time: 06/22/2021 11:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD,TRAVIS,L
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BASFORD,TRAVIS
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/21/2021
VIOLATION CODE: BUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: DEJESUS

DATE/TIME OF TEST: 05/31/2021 19:55:50

Diagnostics Check	OK	19:50	
Air Blank	0.000	19:50	
Control Test	0.081	19:51	
Air Blank	0.000	19:51	
Subject Sample	# 1	19:54	Subject Test Refused
Breath Volume	0.000	19:54	
Air Blank	0.000	19:54	
Control Test	0.080	19:55	
Air Blank	0.000	19:55	
Diagnostics Check	OK	19:55	

Data Download Date/Time: 06/22/2021 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULASKI,DAMON,J
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PULASKI,DAMON
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: BUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 05/30/2021 19:46:27

Diagnostics Check	OK	19:40
Air Blank	0.000	19:40
Control Test	0.078	19:40
Air Blank	0.000	19:41
Subject Sample # 1	0.000	19:42
Breath Volume	3.089	19:42
Air Blank	0.000	19:42
Air Blank	0.000	19:44
Subject Sample # 2	0.000	19:45
Breath Volume	3.203	19:45
Air Blank	0.000	19:45
Control Test	0.078	19:45
Air Blank	0.000	19:46
Diagnostics Check	OK	19:46

Data Download Date/Time: 06/26/2021 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPPE,ZACHARY,D
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: HOPPE,ZACHARY
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: BUI
ARREST TIME: 17:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:03

SUBJECT'S LAST NAME: KIBBY

DATE/TIME OF TEST: 05/29/2021 17:46:30

Diagnostics Check	OK	17:43	
Air Blank	0.000	17:43	
Control Test	0.080	17:43	
Air Blank	0.000	17:44	
Subject Sample	# 1	17:45	Subject Test Refused
Breath Volume	0.000	17:45	
Air Blank	0.000	17:45	
Control Test	0.081	17:45	
Air Blank	0.000	17:46	
Diagnostics Check	OK	17:46	

Data Download Date/Time: 06/25/2021 15:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPPE,ZACHARY,D
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PASS,JUSTIN
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: BUI
ARREST TIME: 18:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 05/30/2021 18:46:23

Diagnostics Check	OK	18:39
Air Blank	0.000	18:39
Control Test	0.081	18:40
Air Blank	0.000	18:40
Subject Sample # 1	0.015	18:42
Breath Volume	2.468	18:42
Air Blank	0.000	18:42
Air Blank	0.000	18:44
Subject Sample # 2	0.013	18:44
Breath Volume	2.136	18:44
Air Blank	0.000	18:45
Control Test	0.081	18:45
Air Blank	0.000	18:46
Diagnostics Check	OK	18:46

Data Download Date/Time: 06/25/2021 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPPE,ZACHARY,D
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: BERG,TIM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: BUI
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:26

SUBJECT'S LAST NAME: WHARTON

DATE/TIME OF TEST: 05/31/2021 16:38:14

Diagnostics Check	OK	16:36	
Air Blank	0.000	16:37	
Control Test	0.000	16:37	Control Outside Tolerance
Air Blank	0.000	16:38	

Data Download Date/Time: 06/25/2021 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPPE,ZACHARY,D
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: BERG,TIM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: BUI
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:26

SUBJECT'S LAST NAME: WHARTON

DATE/TIME OF TEST: 05/31/2021 16:41:21

Diagnostics Check	OK	16:40	
Air Blank	0.000	16:40	
Control Test	0.000	16:40	Control Outside Tolerance
Air Blank	0.000	16:41	

Data Download Date/Time: 06/25/2021 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPPE,ZACHARY,D
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: BERG,TIM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: BUI
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:26

SUBJECT'S LAST NAME: WHARTON

DATE/TIME OF TEST: 05/31/2021 16:44:14

Diagnostics Check	OK	16:42	
Air Blank	0.000	16:43	
Control Test	0.000	16:43	Control Outside Tolerance
Air Blank	0.000	16:44	

Data Download Date/Time: 06/25/2021 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPPE,ZACHARY,D
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: BERG,TIM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: BUI
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:26

SUBJECT'S LAST NAME: WHARTON

DATE/TIME OF TEST: 05/31/2021 16:47:11

Diagnostics Check	OK	16:45	
Air Blank	0.000	16:46	
Control Test	0.000	16:46	Control Outside Tolerance
Air Blank	0.000	16:47	

Data Download Date/Time: 06/25/2021 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPPE,ZACHARY,D
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PASS,JUSTIN
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: BUI
ARREST TIME: 16:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:44

SUBJECT'S LAST NAME: BARTOLOMEI

DATE/TIME OF TEST: 05/31/2021 17:25:05

Diagnostics Check	OK	17:23	
Air Blank	0.000	17:24	
Control Test	0.000	17:24	Control Outside Tolerance
Air Blank	0.000	17:25	

Data Download Date/Time: 06/25/2021 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,MALACHI,C
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PASS,JUSTIN
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: BUI
ARREST TIME: 16:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:44

SUBJECT'S LAST NAME: BARTOLOMEI

DATE/TIME OF TEST: 05/31/2021 17:44:45

Diagnostics Check	OK	17:33	
Air Blank	0.000	17:33	
Control Test	0.081	17:34	
Air Blank	0.000	17:34	
Subject Sample	# 1	17:37	No Sample Provided
Breath Volume	0.000	17:37	
Air Blank	0.000	17:38	
Air Blank	0.000	17:40	
Subject Sample	# 2	17:43	No Sample Provided
Breath Volume	0.000	17:43	
Air Blank	0.000	17:43	
Control Test	0.081	17:44	
Air Blank	0.000	17:44	
Diagnostics Check	OK	17:44	

Data Download Date/Time: 06/25/2021 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,MALACHI,C
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: WILKINS,MALACHI
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: BUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 05/31/2021 21:02:22

Diagnostics Check	OK	20:51	
Air Blank	0.000	20:51	
Control Test	0.081	20:52	
Air Blank	0.000	20:52	
Subject Sample	# 1	20:55	No Sample Provided
Breath Volume	0.000	20:55	
Air Blank	0.000	20:56	
Air Blank	0.000	20:58	
Subject Sample	# 2	21:01	RFI Detect
Breath Volume	0.000	21:01	
Air Blank	0.000	21:01	
Control Test	0.082	21:01	
Air Blank	0.000	21:02	
Diagnostics Check	OK	21:02	

Data Download Date/Time: 06/25/2021 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOBLEY,BRITTANY,R
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: ZAMBO,BRIEANNA
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: GRANDA

DATE/TIME OF TEST: 05/29/2021 18:25:19

Diagnostics Check	OK	18:22	
Air Blank	0.000	18:23	
Control Test	0.081	18:23	
Air Blank	0.000	18:23	
Subject Sample	# 1	18:23	Subject Test Refused
Breath Volume	0.000	18:23	
Air Blank	0.000	18:24	
Control Test	0.081	18:24	
Air Blank	0.000	18:25	
Diagnostics Check	OK	18:25	

Data Download Date/Time: 06/30/2021 19:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,WILLIAM,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: CARROLL,COURTNEY
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: BUI
ARREST TIME: 16:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: PEREZ ZAMORA

DATE/TIME OF TEST: 05/30/2021 17:31:21

Diagnostics Check	OK	17:24
Air Blank	0.000	17:24
Control Test	0.080	17:25
Air Blank	0.000	17:25
Subject Sample # 1	0.064	17:26
Breath Volume	1.476	17:26
Air Blank	0.000	17:27
Air Blank	0.000	17:29
Subject Sample # 2	0.065	17:29
Breath Volume	2.164	17:29
Air Blank	0.000	17:30
Control Test	0.081	17:30
Air Blank	0.000	17:31
Diagnostics Check	OK	17:31

Data Download Date/Time: 06/30/2021 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000905

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTH,DUANE,S
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: SEXTON,JACK
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG928002
DRY GAS EXPIRY DATE: 10/07/2021
VIOLATION CODE: BUI
ARREST TIME: 15:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:44

SUBJECT'S LAST NAME: PANICO

DATE/TIME OF TEST: 05/30/2021 16:47:27

Diagnostics Check	OK	16:33	
Air Blank	0.000	16:33	
Control Test	0.078	16:33	
Air Blank	0.000	16:34	
Subject Sample # 1	0.145	16:37	
Breath Volume	1.976	16:37	
Air Blank	0.000	16:38	
Air Blank	0.000	16:39	
Subject Sample # 2	0.125	16:43	Volume Not Met
Breath Volume	0.304	16:43	
Air Blank	0.000	16:43	
Air Blank	0.000	16:45	
Subject Sample # 3	0.143	16:45	
Breath Volume	1.804	16:45	
Air Blank	0.000	16:46	
Control Test	0.077	16:46	
Air Blank	0.000	16:47	
Diagnostics Check	OK	16:47	

Data Download Date/Time: 06/30/2021 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000905

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTH,DUANE,S
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: VANAS,EZEKIEL
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG928002
DRY GAS EXPIRY DATE: 10/07/2021
VIOLATION CODE: BUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: YEOMANS

DATE/TIME OF TEST: 05/30/2021 18:42:48

Diagnostics Check	OK	18:35
Air Blank	0.000	18:35
Control Test	0.078	18:36
Air Blank	0.000	18:36
Subject Sample # 1	0.142	18:38
Breath Volume	3.210	18:38
Air Blank	0.000	18:38
Air Blank	0.000	18:40
Subject Sample # 2	0.132	18:41
Breath Volume	2.218	18:41
Air Blank	0.000	18:41
Control Test	0.078	18:42
Air Blank	0.000	18:42
Diagnostics Check	OK	18:42

Data Download Date/Time: 06/30/2021 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000906

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUSA,KYLE,A
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PLUSA,KYLE
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 04/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: BUI
ARREST TIME: 16:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: GUEDES

DATE/TIME OF TEST: 05/08/2021 17:14:57

Diagnostics Check	OK	17:08	
Air Blank	0.000	17:09	
Control Test	0.079	17:09	
Air Blank	0.000	17:09	
Subject Sample	# 1 0.071	17:10	
Breath Volume	1.871	17:10	
Air Blank	0.000	17:10	
Air Blank	0.000	17:12	
Subject Sample	# 2 0.073	17:13	
Breath Volume	3.648	17:13	
Air Blank	0.000	17:13	
Control Test	0.079	17:14	
Air Blank		17:14	RFI Detect
Air Blank	0.000	17:14	

Data Download Date/Time: 05/11/2021 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000906

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUSSA,KYLE,A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: PLUSSA,KYLE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 04/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: BUI
ARREST TIME: 16:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: GUEDES

DATE/TIME OF TEST: 05/08/2021 17:26:16

Diagnostics Check	OK	17:19	
Air Blank	0.000	17:20	
Control Test	0.079	17:20	
Air Blank	0.000	17:20	
Subject Sample # 1	0.067	17:21	
Breath Volume	2.531	17:21	
Air Blank	0.000	17:22	
Air Blank	0.000	17:24	
Subject Sample # 2	0.068	17:24	
Breath Volume	3.703	17:24	
Air Blank	0.000	17:25	
Control Test	0.079	17:25	
Air Blank		17:25	RFI Detect
Air Blank	0.000	17:26	

Data Download Date/Time: 05/11/2021 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: ROSOTO

DATE/TIME OF TEST: 05/01/2021 03:09:09

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.078	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.180	03:04
Breath Volume	1.988	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.187	03:07
Breath Volume	2.023	03:07
Air Blank	0.000	03:08
Control Test	0.078	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 05/13/2021 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: SALGUERO

DATE/TIME OF TEST: 05/01/2021 04:28:34

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.078	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.000	04:24
Breath Volume	3.679	04:24
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.000	04:27
Breath Volume	3.070	04:27
Air Blank	0.000	04:27
Control Test	0.079	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 05/13/2021 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MORRISSEY

DATE/TIME OF TEST: 05/03/2021 02:52:13

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.225	02:47
Breath Volume	1.441	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.244	02:50
Breath Volume	2.328	02:50
Air Blank	0.000	02:51
Control Test	0.078	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 05/13/2021 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: VELASQUEZ PEREZ

DATE/TIME OF TEST: 05/03/2021 03:22:06

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.079	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.124	03:17
Breath Volume	1.367	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:19
Subject Sample # 2	0.142	03:20
Breath Volume	1.320	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 05/13/2021 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: SANDBROOK

DATE/TIME OF TEST: 05/04/2021 21:36:58

Diagnostics Check	OK	21:30
Air Blank	0.000	21:31
Control Test	0.078	21:31
Air Blank	0.000	21:31
Subject Sample # 1	0.393	21:32
Breath Volume	2.828	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:35
Subject Sample # 2	0.382	21:35
Breath Volume	3.093	21:35
Air Blank	0.000	21:36
Control Test	0.079	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 05/13/2021 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER III,RICHARD,F
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SNYDER III,RICHARD
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SAMUEL

DATE/TIME OF TEST: 05/06/2021 03:38:07

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.078	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.157	03:33
Breath Volume	1.859	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.158	03:36
Breath Volume	1.875	03:36
Air Blank	0.000	03:37
Control Test	0.078	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 05/13/2021 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: BENTON

DATE/TIME OF TEST: 05/06/2021 19:33:39

Diagnostics Check	OK	19:27
Air Blank	0.000	19:28
Control Test	0.078	19:28
Air Blank	0.000	19:28
Subject Sample # 1	0.194	19:29
Breath Volume	1.792	19:29
Air Blank	0.000	19:29
Air Blank	0.000	19:31
Subject Sample # 2	0.176	19:32
Breath Volume	2.496	19:32
Air Blank	0.000	19:32
Control Test	0.078	19:33
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

Data Download Date/Time: 05/13/2021 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MURILLO RUIZ

DATE/TIME OF TEST: 05/16/2021 04:40:47

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.079	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.119	04:35
Breath Volume	1.585	04:35
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.114	04:39
Breath Volume	1.250	04:39
Air Blank	0.000	04:39
Control Test	0.079	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 06/10/2021 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JIMENEZ,WILLIAM,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JIMENEZ,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 14:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: AHEARN

DATE/TIME OF TEST: 05/19/2021 15:54:08

Diagnostics Check	OK	15:47
Air Blank	0.000	15:48
Control Test	0.080	15:48
Air Blank	0.000	15:49
Subject Sample # 1	0.000	15:49
Breath Volume	3.710	15:49
Air Blank	0.000	15:50
Air Blank	0.000	15:52
Subject Sample # 2	0.000	15:52
Breath Volume	4.828	15:52
Air Blank	0.000	15:53
Control Test	0.078	15:53
Air Blank	0.000	15:54
Diagnostics Check	OK	15:54

Data Download Date/Time: 06/10/2021 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: OSTERTAG,SHELBY
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: WEBER

DATE/TIME OF TEST: 05/27/2021 00:32:47

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.079	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.197	00:27
Breath Volume	1.425	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.195	00:31
Breath Volume	2.480	00:31
Air Blank	0.000	00:31
Control Test	0.078	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 06/10/2021 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: CRUZ MORALES

DATE/TIME OF TEST: 05/29/2021 03:26:27

Diagnostics Check	OK	03:19
Air Blank	0.000	03:20
Control Test	0.078	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.178	03:21
Breath Volume	1.472	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.183	03:25
Breath Volume	1.441	03:25
Air Blank	0.000	03:25
Control Test	0.078	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 06/10/2021 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIDSON,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: KLOBUCHAR,RICHARD
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: BUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: YAHNKE

DATE/TIME OF TEST: 05/29/2021 20:51:48

Diagnostics Check	OK	20:45
Air Blank	0.000	20:45
Control Test	0.078	20:46
Air Blank	0.000	20:46
Subject Sample # 1	0.096	20:47
Breath Volume	1.222	20:47
Air Blank	0.000	20:47
Air Blank	0.000	20:49
Subject Sample # 2	0.093	20:50
Breath Volume	1.953	20:50
Air Blank	0.000	20:50
Control Test	0.079	20:51
Air Blank	0.000	20:51
Diagnostics Check	OK	20:51

Data Download Date/Time: 06/10/2021 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMBERLIN,KEVIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TOMBERLIN,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 10:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:07

SUBJECT'S LAST NAME: IVES

DATE/TIME OF TEST: 05/30/2021 12:08:31

Diagnostics Check	OK	12:02
Air Blank	0.000	12:02
Control Test	0.079	12:03
Air Blank	0.000	12:03
Subject Sample # 1	0.000	12:04
Breath Volume	4.707	12:04
Air Blank	0.000	12:04
Air Blank	0.000	12:06
Subject Sample # 2	0.000	12:07
Breath Volume	3.292	12:07
Air Blank	0.000	12:07
Control Test	0.078	12:07
Air Blank	0.000	12:08
Diagnostics Check	OK	12:08

Data Download Date/Time: 06/10/2021 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: MARTINEZ RAMIREZ

DATE/TIME OF TEST: 05/31/2021 03:35:00

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.129	03:30
Breath Volume	1.882	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.122	03:33
Breath Volume	1.113	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:35

Data Download Date/Time: 06/10/2021 08:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JABLONSKI,MARK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 867139
DRY GAS EXPIRY DATE: 11/15/2020
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: MARTINEZ ALVARADO

DATE/TIME OF TEST: 05/31/2021 05:40:46

Diagnostics Check	OK	05:33
Air Blank	0.000	05:34
Control Test	0.079	05:34
Air Blank	0.000	05:35
Subject Sample # 1	0.176	05:36
Breath Volume	3.167	05:36
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.174	05:39
Breath Volume	3.445	05:39
Air Blank	0.000	05:39
Control Test	0.079	05:40
Air Blank	0.000	05:40
Diagnostics Check	OK	05:40

Data Download Date/Time: 06/10/2021 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICKEY, WILLIAM, Q
BREATH TEST OPERATOR AGENCY: PORT ST. JOE PD
ARRESTING OFFICER: DICKEY, WILLIAM
ARRESTING OFFICER AGENCY: PORT ST. JOE PD
LAST AGENCY INSPECTION: 05/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 05/02/2021 02:47:18

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.078	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.228	02:42
Breath Volume		1.277	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.226	02:45
Breath Volume		1.238	02:45
Air Blank		0.000	02:46
Control Test		0.079	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 06/11/2021 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,PAUL,G
BREATH TEST OPERATOR AGENCY: GULF COUNTY SO
ARRESTING OFFICER: YOUNG,PAUL
ARRESTING OFFICER AGENCY: GULF COUNTY SO
LAST AGENCY INSPECTION: 05/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 05/09/2021 00:14:52

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:12	
Control Test	0.080	00:12	
Air Blank	0.000	00:13	
Subject Sample	# 1	00:13	Subject Test Refused
Breath Volume	0.000	00:13	
Air Blank	0.000	00:13	
Control Test	0.080	00:14	
Air Blank	0.000	00:14	
Diagnostics Check	OK	00:14	

Data Download Date/Time: 06/11/2021 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURCH,RUSSELL,L
BREATH TEST OPERATOR AGENCY: PORT ST. JOE PD
ARRESTING OFFICER: BURKETT,JESSE
ARRESTING OFFICER AGENCY: PORT ST. JOE PD
LAST AGENCY INSPECTION: 05/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: MINNICK

DATE/TIME OF TEST: 05/21/2021 23:39:43

Diagnostics Check	OK	23:32
Air Blank	0.000	23:33
Control Test	0.079	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.239	23:35
Breath Volume	2.984	23:35
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.222	23:38
Breath Volume	2.269	23:38
Air Blank	0.000	23:38
Control Test	0.080	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 06/11/2021 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: South Miami PD
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: BASILIADES

DATE/TIME OF TEST: 05/04/2021 05:15:49

Diagnostics Check	OK	05:08
Air Blank	0.000	05:09
Control Test	0.080	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.122	05:11
Breath Volume	2.046	05:11
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.117	05:14
Breath Volume	1.730	05:14
Air Blank	0.000	05:14
Control Test	0.079	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 05/30/2021 20:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: South Miami PD
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: ARNAIZ

DATE/TIME OF TEST: 05/08/2021 04:15:10

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:12	
Control Test	0.080	04:13	
Air Blank	0.000	04:13	
Subject Sample	# 1	04:13	RFI Detect
Breath Volume	0.000	04:13	
Air Blank	0.000	04:14	
Control Test	0.080	04:14	
Air Blank	0.000	04:15	
Diagnostics Check	OK	04:15	

Data Download Date/Time: 05/30/2021 20:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: South Miami PD
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: ARNAIZ

DATE/TIME OF TEST: 05/08/2021 04:23:17

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.178	04:18
Breath Volume	1.503	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.197	04:21
Breath Volume	1.992	04:21
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 05/30/2021 20:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: South Miami PD
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/15/2021 05:46:19

Diagnostics Check		OK	05:39
Air Blank		0.000	05:39
Control Test		0.080	05:39
Air Blank		0.000	05:40
Subject Sample	# 1	0.230	05:41
Breath Volume		2.246	05:41
Air Blank		0.000	05:42
Air Blank		0.000	05:43
Subject Sample	# 2	0.225	05:44
Breath Volume		2.101	05:44
Air Blank		0.000	05:45
Control Test		0.079	05:45
Air Blank		0.000	05:46
Diagnostics Check		OK	05:46

Data Download Date/Time: 05/30/2021 20:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: South Miami PD
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 05/30/2021 04:25:34

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.079	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.122	04:20
Breath Volume	2.710	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.123	04:24
Breath Volume	2.902	04:24
Air Blank	0.000	04:24
Control Test	0.079	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 05/30/2021 20:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARR,L,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: TARR,L
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: LOPEZGARCIA

DATE/TIME OF TEST: 05/11/2021 22:25:51

Diagnostics Check		OK	22:13	
Air Blank		0.000	22:13	
Control Test		0.079	22:13	
Air Blank		0.000	22:14	
Subject Sample	# 1	0.218	22:17	Volume Not Met
Breath Volume		0.000	22:17	
Air Blank		0.000	22:18	
Air Blank		0.000	22:20	
Subject Sample	# 2	0.218	22:21	
Breath Volume		1.835	22:21	
Air Blank		0.000	22:21	
Air Blank		0.000	22:23	
Subject Sample	# 3	0.236	22:24	
Breath Volume		2.132	22:24	
Air Blank		0.000	22:24	
Control Test		0.078	22:25	
Air Blank		0.000	22:25	
Diagnostics Check		OK	22:25	

Data Download Date/Time: 05/13/2021 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EPPS,QUINTELLA,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: SEMPLE,W
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: EILLISON

DATE/TIME OF TEST: 05/14/2021 00:28:05

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.079	00:20
Air Blank		0.000	00:21
Subject Sample	# 1	0.172	00:22
Breath Volume		1.453	00:22
Air Blank		0.000	00:23
Air Blank		0.000	00:24
Subject Sample	# 2	0.188	00:26
Breath Volume		1.828	00:26
Air Blank		0.000	00:27
Control Test		0.077	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 05/24/2021 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EPPS,QUINTELLA,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: DIAZ,R
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: MARINTEZ

DATE/TIME OF TEST: 05/17/2021 23:02:44

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.079	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.103	22:57
Breath Volume	2.191	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.100	23:01
Breath Volume	2.187	23:01
Air Blank	0.000	23:01
Control Test	0.079	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 05/24/2021 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,CHRISTINA,L
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: DIAZ,ROBIN
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/18/2021 06:37:39

Diagnostics Check	OK	06:31
Air Blank	0.000	06:32
Control Test	0.078	06:32
Air Blank	0.000	06:33
Subject Sample # 1	0.000	06:33
Breath Volume	1.632	06:33
Air Blank	0.000	06:33
Air Blank	0.000	06:35
Subject Sample # 2	0.000	06:36
Breath Volume	1.566	06:36
Air Blank	0.000	06:36
Control Test	0.078	06:37
Air Blank	0.000	06:37
Diagnostics Check	OK	06:37

Data Download Date/Time: 05/24/2021 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EPPS,QUINTELLA,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CASTANEDA,R
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: OSCEOLA

DATE/TIME OF TEST: 05/19/2021 05:00:26

Diagnostics Check	OK	04:51
Air Blank	0.000	04:52
Control Test	0.078	04:52
Air Blank	0.000	04:53
Subject Sample # 1	0.000	04:56
Breath Volume	2.378	04:56
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.000	04:59
Breath Volume	2.070	04:59
Air Blank	0.000	04:59
Control Test	0.078	04:59
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 05/24/2021 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EPPS,QUINTELLA,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: OKUPSKI,R
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 05/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: COY

DATE/TIME OF TEST: 05/29/2021 01:35:23

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.188	01:29
Breath Volume	1.378	01:29
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.179	01:33
Breath Volume	1.441	01:33
Air Blank	0.000	01:34
Control Test	0.078	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 06/16/2021 06:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONDRAGON,EDWIN,F
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: OKUPSKI,R
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 05/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: COY

DATE/TIME OF TEST: 05/29/2021 08:56:44

Diagnostics Check	OK	08:44	
Air Blank	0.000	08:45	
Control Test	0.078	08:45	
Air Blank	0.000	08:46	
Subject Sample	# 1 0.058	08:46	
Breath Volume	1.531	08:46	
Air Blank	0.000	08:47	
Air Blank	0.000	08:49	
Subject Sample	# 2 0.056	08:52	Volume Not Met
Breath Volume	0.367	08:52	
Air Blank	0.000	08:53	
Air Blank	0.000	08:54	
Subject Sample	# 3 0.054	08:55	
Breath Volume	1.562	08:55	
Air Blank	0.000	08:55	
Control Test	0.077	08:56	
Air Blank	0.000	08:56	
Diagnostics Check	OK	08:56	

Data Download Date/Time: 06/16/2021 06:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EPPS,Q,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: LOUIS-PIERRE,R
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 05/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 05/29/2021 19:33:35

Diagnostics Check		OK	19:27
Air Blank		0.000	19:27
Control Test		0.079	19:27
Air Blank		0.000	19:28
Subject Sample	# 1	0.294	19:28
Breath Volume		1.992	19:28
Air Blank		0.000	19:29
Air Blank		0.000	19:31
Subject Sample	# 2	0.291	19:31
Breath Volume		3.187	19:31
Air Blank		0.000	19:32
Control Test		0.078	19:32
Air Blank		0.000	19:33
Diagnostics Check		OK	19:33

Data Download Date/Time: 06/16/2021 06:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EPPS,QUINTELLA,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: LOUIS-PIERRE,R
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 05/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 05/30/2021 04:45:09

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.078	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.103	04:40
Breath Volume	2.089	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.101	04:43
Breath Volume	2.761	04:43
Air Blank	0.000	04:44
Control Test	0.078	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 06/16/2021 06:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONDRAGON, EDWIN, F
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: LOUIS-PIERRE, RUBEN
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 05/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 07:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:02

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 05/30/2021 09:07:24

Diagnostics Check		OK	08:59
Air Blank		0.000	09:00
Control Test		0.078	09:00
Air Blank		0.000	09:01
Subject Sample	# 1	0.007	09:03
Breath Volume		3.015	09:03
Air Blank		0.000	09:03
Air Blank		0.000	09:05
Subject Sample	# 2	0.000	09:06
Breath Volume		3.144	09:06
Air Blank		0.000	09:06
Control Test		0.078	09:06
Air Blank		0.000	09:07
Diagnostics Check		OK	09:07

Data Download Date/Time: 06/16/2021 06:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEWITT,JARED,S
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: HEWITT,JARED
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: HAWTHORNE

DATE/TIME OF TEST: 05/28/2021 00:06:31

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:55	
Control Test	0.080	23:55	
Air Blank	0.000	23:56	
Subject Sample	# 2 0.186	23:59	Volume Not Met
Breath Volume	0.097	23:59	
Air Blank	0.000	00:00	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:05	No Sample Provided
Breath Volume	0.000	00:05	
Air Blank	0.000	00:05	
Control Test	0.076	00:05	
Air Blank	0.000	00:06	
Diagnostics Check	OK	00:06	

Data Download Date/Time: 06/29/2021 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BILLINGSLEY,HEATHER,D
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: DAMBROSIO,JOHN
ARRESTING OFFICER AGENCY: Carrabelle PD
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: BILLINGSLEY

DATE/TIME OF TEST: 05/20/2021 20:38:41

Diagnostics Check	OK	20:27	
Air Blank	0.000	20:27	
Control Test	0.078	20:27	
Air Blank	0.000	20:28	
Subject Sample	# 1	20:31	No Sample Provided
Breath Volume	0.000	20:31	
Air Blank	0.000	20:32	
Air Blank	0.000	20:34	
Subject Sample	# 2	20:37	No Sample Provided
Breath Volume	0.000	20:37	
Air Blank	0.000	20:37	
Control Test	0.079	20:38	
Air Blank	0.000	20:38	
Diagnostics Check	OK	20:38	

Data Download Date/Time: 05/25/2021 11:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORDON,ROLF,E
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: GORDON,ROLF
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: GODWIN

DATE/TIME OF TEST: 05/26/2021 04:21:34

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:10	
Control Test	0.079	04:10	
Air Blank	0.000	04:10	
Subject Sample	# 1	04:14	No Sample Provided
Breath Volume	0.000	04:14	
Air Blank	0.000	04:14	
Air Blank	0.000	04:16	
Subject Sample	# 2	0.000	04:17
Breath Volume	2.554	04:17	
Air Blank	0.000	04:17	
Air Blank	0.000	04:19	
Subject Sample	# 3	0.000	04:20
Breath Volume	2.183	04:20	
Air Blank	0.000	04:20	
Control Test	0.079	04:20	
Air Blank	0.000	04:21	
Diagnostics Check	OK	04:21	

Data Download Date/Time: 06/29/2021 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEWITT,JARED,S
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: HEWITT,JARED
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: HAWTHORNE

DATE/TIME OF TEST: 05/27/2021 23:43:53

Diagnostics Check	OK	23:43	
Air Blank		23:43	RFI Detect
Air Blank	0.000	23:43	

Data Download Date/Time: 06/29/2021 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEWITT,JARED,S
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: HEWITT,JARED
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: HAWTHORNE

DATE/TIME OF TEST: 05/27/2021 23:53:33

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:48	
Control Test	0.079	23:48	
Air Blank	0.000	23:49	
Subject Sample # 1	0.149	23:52	Volume Not Met
Breath Volume	0.554	23:52	
Air Blank		23:53	Purge Fail
Air Blank	0.000	23:53	

Data Download Date/Time: 06/29/2021 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAKER,RUEL,W
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: RAKER,RUEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: BUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: MASKEVICH

DATE/TIME OF TEST: 05/30/2021 20:43:26

Diagnostics Check	OK	20:40	
Air Blank	0.000	20:41	
Control Test	0.079	20:41	
Air Blank	0.000	20:41	
Subject Sample	# 1	20:42	Subject Test Refused
Breath Volume	0.000	20:42	
Air Blank	0.000	20:42	
Control Test	0.079	20:42	
Air Blank	0.000	20:43	
Diagnostics Check	OK	20:43	

Data Download Date/Time: 06/29/2021 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELSCH, KENNETH, P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ADAMCZYK, ANDREW
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 05/10/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 05/23/2021 06:58:09

Diagnostics Check	OK	06:50
Air Blank	0.000	06:51
Control Test	0.079	06:51
Air Blank	0.000	06:51
Subject Sample # 1	0.183	06:53
Breath Volume	2.386	06:53
Air Blank	0.000	06:54
Air Blank	0.000	06:55
Subject Sample # 2	0.181	06:56
Breath Volume	1.433	06:56
Air Blank	0.000	06:57
Control Test	0.075	06:57
Air Blank	0.000	06:58
Diagnostics Check	OK	06:58

Data Download Date/Time: 06/16/2021 14:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINA,RYAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MARTINA,RYAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 05/05/2021 00:58:02

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.080	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.040	00:52
Breath Volume		1.121	00:52
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.037	00:56
Breath Volume		2.414	00:56
Air Blank		0.000	00:57
Control Test		0.082	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:58

Data Download Date/Time: 06/02/2021 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DANIEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WAGNER,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: JEPSEN

DATE/TIME OF TEST: 05/06/2021 00:00:53

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:52	
Control Test	0.080	23:53	
Air Blank	0.000	23:53	
Subject Sample	# 1	23:56	No Sample Provided
Breath Volume	0.000	23:56	
Air Blank	0.000	23:57	
Air Blank	0.000	23:59	
Subject Sample	# 2	23:59	Subject Test Refused
Breath Volume	0.000	23:59	
Air Blank	0.000	23:59	
Control Test	0.079	00:00	
Air Blank	0.000	00:00	
Diagnostics Check	OK	00:00	

Data Download Date/Time: 06/02/2021 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORIGA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: NEAL-BALL

DATE/TIME OF TEST: 05/09/2021 01:54:25

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.081	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.105	01:49
Breath Volume	2.332	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.097	01:52
Breath Volume	1.238	01:52
Air Blank	0.000	01:53
Control Test	0.081	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 06/02/2021 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,C,D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CROUCH,CLIFFORD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: BENTON

DATE/TIME OF TEST: 05/09/2021 04:03:06

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.080	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.143	03:58
Breath Volume	1.128	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.143	04:01
Breath Volume	1.253	04:01
Air Blank	0.000	04:02
Control Test	0.081	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 06/02/2021 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,R,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARVEY,R
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: BARRINGTON

DATE/TIME OF TEST: 05/14/2021 03:34:12

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.081	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.167	03:29
Breath Volume	1.335	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.149	03:32
Breath Volume	1.113	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 06/02/2021 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:53

SUBJECT'S LAST NAME: GASKIN ROSAS

DATE/TIME OF TEST: 05/16/2021 07:26:14

Diagnostics Check	OK	07:19
Air Blank	0.000	07:20
Control Test	0.081	07:20
Air Blank	0.000	07:21
Subject Sample # 1	0.147	07:21
Breath Volume	3.398	07:21
Air Blank	0.000	07:22
Air Blank	0.000	07:24
Subject Sample # 2	0.133	07:24
Breath Volume	2.105	07:24
Air Blank	0.000	07:25
Control Test	0.081	07:25
Air Blank	0.000	07:26
Diagnostics Check	OK	07:26

Data Download Date/Time: 06/02/2021 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 05/17/2021 05:26:22

Diagnostics Check	OK	05:19
Air Blank	0.000	05:20
Control Test	0.081	05:20
Air Blank	0.000	05:21
Subject Sample # 1	0.000	05:21
Breath Volume	1.375	05:21
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 2	0.000	05:24
Breath Volume	1.410	05:24
Air Blank	0.000	05:25
Control Test	0.081	05:25
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 06/02/2021 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTON,JEREMY,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BARTON,JEREMY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/22/2021 05:31:38

Diagnostics Check	OK	05:25
Air Blank	0.000	05:26
Control Test	0.080	05:26
Air Blank	0.000	05:26
Subject Sample # 1	0.145	05:27
Breath Volume	1.265	05:27
Air Blank	0.000	05:27
Air Blank	0.000	05:29
Subject Sample # 2	0.143	05:30
Breath Volume	1.417	05:30
Air Blank	0.000	05:30
Control Test	0.081	05:31
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

Data Download Date/Time: 06/02/2021 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CHARLES,B
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HANSFORD,CHARLES
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:46

SUBJECT'S LAST NAME: BRADFORD

DATE/TIME OF TEST: 05/27/2021 09:16:43

Diagnostics Check	OK	09:08
Air Blank	0.000	09:09
Control Test	0.080	09:09
Air Blank	0.000	09:10
Subject Sample # 1	0.000	09:10
Breath Volume	1.437	09:10
Air Blank	0.000	09:11
Air Blank	0.000	09:13
Subject Sample # 2	0.000	09:15
Breath Volume	2.417	09:15
Air Blank	0.000	09:15
Control Test	0.083	09:16
Air Blank	0.000	09:16
Diagnostics Check	OK	09:16

Data Download Date/Time: 06/02/2021 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTON,JEREMY,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BARTON,JEREMY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: BENSON

DATE/TIME OF TEST: 05/27/2021 23:01:27

Diagnostics Check	OK	22:54
Air Blank	0.000	22:55
Control Test	0.079	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.169	22:56
Breath Volume	2.464	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.161	22:59
Breath Volume	2.199	22:59
Air Blank	0.000	23:00
Control Test	0.079	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 06/02/2021 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WONSCH,J,W
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: YORK,S
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: SANTIAGUE

DATE/TIME OF TEST: 05/29/2021 04:51:29

Diagnostics Check		OK	04:42
Air Blank		0.000	04:42
Control Test		0.081	04:42
Air Blank		0.000	04:43
Subject Sample	# 1	0.204	04:45
Breath Volume		2.593	04:45
Air Blank		0.000	04:46
Air Blank		0.000	04:48
Subject Sample	# 2	0.191	04:49
Breath Volume		1.835	04:49
Air Blank		0.000	04:50
Control Test		0.080	04:50
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 06/02/2021 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEON,L
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: PARSONS BUCKHALT, TIFFAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: HUAPILLA

DATE/TIME OF TEST: 05/02/2021 03:55:12

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.082	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.175	03:50
Breath Volume	1.382	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.177	03:53
Breath Volume	1.632	03:53
Air Blank	0.000	03:54
Control Test	0.083	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 05/11/2021 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: MESSENGALE

DATE/TIME OF TEST: 05/02/2021 19:13:33

Diagnostics Check	OK	19:06
Air Blank	0.000	19:07
Control Test	0.083	19:07
Air Blank	0.000	19:08
Subject Sample # 1	0.000	19:09
Breath Volume	3.425	19:09
Air Blank	0.000	19:09
Air Blank	0.000	19:11
Subject Sample # 2	0.000	19:12
Breath Volume	3.863	19:12
Air Blank	0.000	19:12
Control Test	0.083	19:12
Air Blank	0.000	19:13
Diagnostics Check	OK	19:13

Data Download Date/Time: 05/11/2021 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,JOSEPH,B
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FRANKLIN,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 05/05/2021 08:37:33

Diagnostics Check		OK	08:30
Air Blank		0.000	08:30
Control Test		0.082	08:31
Air Blank		0.000	08:31
Subject Sample	# 1	0.110	08:32
Breath Volume		2.492	08:32
Air Blank		0.000	08:32
Air Blank		0.000	08:34
Subject Sample	# 2	0.106	08:36
Breath Volume		1.640	08:36
Air Blank		0.000	08:36
Control Test		0.083	08:36
Air Blank		0.000	08:37
Diagnostics Check		OK	08:37

Data Download Date/Time: 05/11/2021 15:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DANIEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WAGNER,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: BURDESHAW

DATE/TIME OF TEST: 05/01/2021 20:24:18

Diagnostics Check	OK	20:15	
Air Blank	0.000	20:16	
Control Test	0.079	20:16	
Air Blank	0.000	20:17	
Subject Sample	# 1	20:20	No Sample Provided
Breath Volume	0.000	20:20	
Air Blank	0.000	20:20	
Air Blank	0.000	20:22	
Subject Sample	# 2	20:22	Subject Test Refused
Breath Volume	0.000	20:22	
Air Blank	0.000	20:23	
Control Test	0.079	20:23	
Air Blank	0.000	20:24	
Diagnostics Check	OK	20:24	

Data Download Date/Time: 05/26/2021 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,KENYON,L
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: CAUSSEAUX,DYLAN
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:59

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/13/2021 19:30:51

Diagnostics Check	OK	19:28	
Air Blank	0.000	19:28	
Control Test	0.080	19:28	
Air Blank	0.000	19:29	
Subject Sample	# 1	19:29	Subject Test Refused
Breath Volume	0.000	19:29	
Air Blank	0.000	19:29	
Control Test	0.080	19:30	
Air Blank	0.000	19:30	
Diagnostics Check	OK	19:30	

Data Download Date/Time: 05/26/2021 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,CURTIS,D
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HUNTER,CURTIS
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: PAPINEAU

DATE/TIME OF TEST: 05/17/2021 23:08:59

Diagnostics Check	OK	23:02
Air Blank	0.000	23:03
Control Test	0.080	23:03
Air Blank	0.000	23:04
Subject Sample # 1	0.000	23:04
Breath Volume	2.917	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.000	23:07
Breath Volume	3.902	23:07
Air Blank	0.000	23:08
Control Test	0.080	23:08
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 05/26/2021 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,KENYON,L
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MILLER,KENYON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:14

SUBJECT'S LAST NAME: MARRA

DATE/TIME OF TEST: 05/28/2021 17:51:37

Diagnostics Check	OK	17:45
Air Blank	0.000	17:45
Control Test	0.080	17:45
Air Blank	0.000	17:46
Subject Sample # 1	0.118	17:47
Breath Volume	1.398	17:47
Air Blank	0.000	17:47
Air Blank	0.000	17:49
Subject Sample # 2	0.128	17:50
Breath Volume	2.820	17:50
Air Blank	0.000	17:50
Control Test	0.079	17:51
Air Blank	0.000	17:51
Diagnostics Check	OK	17:51

Data Download Date/Time: 06/30/2021 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,BOBBY,M
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MORALES,BOBBY
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: BUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:08

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 05/29/2021 19:26:28

Diagnostics Check	OK	19:23	
Air Blank	0.000	19:24	
Control Test	0.081	19:24	
Air Blank	0.000	19:24	
Subject Sample	# 1	19:25	Subject Test Refused
Breath Volume	0.000	19:25	
Air Blank	0.000	19:25	
Control Test	0.080	19:25	
Air Blank		19:26	RFI Detect
Air Blank	0.000	19:26	

Data Download Date/Time: 06/30/2021 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SETH,M
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: GRAY,SETH
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: BUI
ARREST TIME: 18:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: TRAYLOR

DATE/TIME OF TEST: 05/31/2021 19:12:29

Diagnostics Check	OK	19:05
Air Blank	0.000	19:06
Control Test	0.081	19:06
Air Blank	0.000	19:07
Subject Sample # 1	0.233	19:07
Breath Volume	3.351	19:07
Air Blank	0.000	19:08
Air Blank	0.000	19:10
Subject Sample # 2	0.233	19:10
Breath Volume	3.015	19:10
Air Blank	0.000	19:11
Control Test	0.080	19:11
Air Blank	0.000	19:12
Diagnostics Check	OK	19:12

Data Download Date/Time: 06/30/2021 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 05/01/2021 04:20:45

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:09	
Control Test	0.080	04:10	
Air Blank	0.000	04:10	
Subject Sample	# 1	04:13	No Sample Provided
Breath Volume	0.000	04:13	
Air Blank	0.000	04:14	
Air Blank	0.000	04:16	
Subject Sample	# 2	04:19	No Sample Provided
Breath Volume	0.000	04:19	
Air Blank	0.000	04:19	
Control Test	0.079	04:20	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:20	

Data Download Date/Time: 05/05/2021 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MELLA, MELISSA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: TORRES-SERRANO

DATE/TIME OF TEST: 05/02/2021 04:54:10

Diagnostics Check	OK	04:47
Air Blank	0.000	04:47
Control Test	0.081	04:48
Air Blank	0.000	04:48
Subject Sample # 1	0.109	04:49
Breath Volume	4.875	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:51
Subject Sample # 2	0.109	04:52
Breath Volume	5.332	04:52
Air Blank	0.000	04:53
Control Test	0.079	04:53
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 05/05/2021 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CARBONE,ANDRES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: FLEURANT

DATE/TIME OF TEST: 05/06/2021 02:33:43

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.020	02:29
Breath Volume	1.148	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.017	02:32
Breath Volume	1.156	02:32
Air Blank	0.000	02:32
Control Test	0.079	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 06/08/2021 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IMPARATO,FRANK
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HAWS

DATE/TIME OF TEST: 05/06/2021 04:33:47

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.138	04:29
Breath Volume	1.714	04:29
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.138	04:32
Breath Volume	1.582	04:32
Air Blank	0.000	04:32
Control Test	0.079	04:33
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 06/08/2021 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALL,ANDREW
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BLANDIN

DATE/TIME OF TEST: 05/07/2021 02:45:08

Diagnostics Check		OK	02:39
Air Blank		0.000	02:39
Control Test		0.080	02:39
Air Blank		0.000	02:40
Subject Sample	# 1	0.106	02:40
Breath Volume		2.937	02:40
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.107	02:43
Breath Volume		2.050	02:43
Air Blank		0.000	02:44
Control Test		0.078	02:44
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 06/08/2021 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KANE,ALEXANDER
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: MEARES

DATE/TIME OF TEST: 05/08/2021 04:03:21

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.081	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.180	03:58
Breath Volume	1.355	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.186	04:01
Breath Volume	1.574	04:01
Air Blank	0.000	04:02
Control Test	0.078	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 06/08/2021 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IRELAND,DAVID
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: MOJZAK

DATE/TIME OF TEST: 05/12/2021 01:58:42

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.080	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.105	01:54
Breath Volume	1.894	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:56
Subject Sample # 2	0.107	01:57
Breath Volume	1.765	01:57
Air Blank	0.000	01:57
Control Test	0.078	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 06/08/2021 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IMPARATO,FRANK
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: SALINAS

DATE/TIME OF TEST: 05/12/2021 03:14:10

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.080	03:09
Air Blank	0.000	03:09
Subject Sample # 1	0.000	03:10
Breath Volume	1.816	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.000	03:12
Breath Volume	2.191	03:12
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 06/08/2021 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: HERAI

DATE/TIME OF TEST: 05/13/2021 05:36:49

Diagnostics Check	OK	05:25	
Air Blank	0.000	05:25	
Control Test	0.080	05:26	
Air Blank	0.000	05:26	
Subject Sample	# 1	05:29	No Sample Provided
Breath Volume	0.000	05:29	
Air Blank	0.000	05:30	
Air Blank	0.000	05:32	
Subject Sample	# 2	05:35	No Sample Provided
Breath Volume	0.000	05:35	
Air Blank	0.000	05:35	
Control Test	0.080	05:36	
Air Blank	0.000	05:36	
Diagnostics Check	OK	05:36	

Data Download Date/Time: 06/08/2021 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KIPP,ALEXANDER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 05/14/2021 02:00:08

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.080	01:54
Air Blank	0.000	01:54
Subject Sample # 1	0.178	01:55
Breath Volume	1.183	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.185	01:58
Breath Volume	1.441	01:58
Air Blank	0.000	01:59
Control Test	0.078	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 06/08/2021 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NELSON,GREGG
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 05/15/2021 01:06:06

Diagnostics Check		OK	01:00
Air Blank		0.000	01:00
Control Test		0.079	01:00
Air Blank		0.000	01:01
Subject Sample	# 1	0.189	01:01
Breath Volume		2.406	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.186	01:04
Breath Volume		3.398	01:04
Air Blank		0.000	01:05
Control Test		0.079	01:05
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 06/08/2021 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SINGSON,ROLAND
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: KNOFF

DATE/TIME OF TEST: 05/21/2021 01:17:19

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.080	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.000	01:13
Breath Volume	2.828	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.000	01:15
Breath Volume	2.367	01:15
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 06/08/2021 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KUZMA, TIM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: DONALD

DATE/TIME OF TEST: 05/23/2021 03:39:00

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.081	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.223	03:34
Breath Volume		1.421	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.229	03:37
Breath Volume		1.367	03:37
Air Blank		0.000	03:38
Control Test		0.079	03:38
Air Blank		0.000	03:38
Diagnostics Check		OK	03:39

Data Download Date/Time: 06/08/2021 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PACE, AUSTIN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 05/23/2021 20:54:15

Diagnostics Check	OK	20:48
Air Blank	0.000	20:48
Control Test	0.080	20:48
Air Blank	0.000	20:49
Subject Sample # 1	0.230	20:49
Breath Volume	2.148	20:49
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.224	20:52
Breath Volume	2.082	20:52
Air Blank	0.000	20:53
Control Test	0.078	20:53
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 06/08/2021 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRIGGS,DERIENNE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: SCHEIBEL

DATE/TIME OF TEST: 05/27/2021 01:43:22

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.079	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.108	01:38
Breath Volume	1.355	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.106	01:41
Breath Volume	1.726	01:41
Air Blank	0.000	01:42
Control Test	0.078	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 06/08/2021 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRAZEE,COLE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: ALCARAZ PULIDO

DATE/TIME OF TEST: 05/28/2021 03:11:03

Diagnostics Check	OK	03:10	
Air Blank		03:10	RFI Detect
Air Blank	0.000	03:11	

Data Download Date/Time: 06/08/2021 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRAZEE,COLE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: ALCARAZ PULIDO

DATE/TIME OF TEST: 05/28/2021 03:19:08

Diagnostics Check		OK	03:12
Air Blank		0.000	03:13
Control Test		0.080	03:13
Air Blank		0.000	03:14
Subject Sample	# 1	0.198	03:14
Breath Volume		4.003	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.212	03:17
Breath Volume		3.933	03:17
Air Blank		0.000	03:18
Control Test		0.079	03:18
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 06/08/2021 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEWAN,JANICE,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SAUNDERS,HUNTER
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 05/29/2021 02:35:37

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.079	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.169	02:31
Breath Volume	1.406	02:31
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.157	02:34
Breath Volume	2.160	02:34
Air Blank	0.000	02:34
Control Test	0.079	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 06/08/2021 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: EDWARDS,ANTHONY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: BODLEY

DATE/TIME OF TEST: 05/29/2021 02:53:58

Diagnostics Check		OK	02:44
Air Blank		0.000	02:45
Control Test		0.079	02:45
Air Blank		0.000	02:46
Subject Sample	# 1	0.158	02:49
Breath Volume		1.226	02:49
Air Blank		0.000	02:49
Air Blank		0.000	02:51
Subject Sample	# 2	0.160	02:52
Breath Volume		1.222	02:52
Air Blank		0.000	02:53
Control Test		0.076	02:53
Air Blank		0.000	02:53
Diagnostics Check		OK	02:53

Data Download Date/Time: 06/08/2021 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PARADIS,GEORGE
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: SANTANA RODRIGUEZ

DATE/TIME OF TEST: 05/29/2021 04:34:20

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.106	04:29
Breath Volume	1.125	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.102	04:32
Breath Volume	1.265	04:32
Air Blank	0.000	04:33
Control Test	0.079	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 06/08/2021 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCLEAN, NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 05/30/2021 01:06:00

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.102	01:01
Breath Volume	3.101	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.097	01:04
Breath Volume	2.546	01:04
Air Blank	0.000	01:05
Control Test	0.079	01:05
Air Blank	0.000	01:05
Diagnostics Check	OK	01:06

Data Download Date/Time: 06/08/2021 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BLACKWELL, ROBERT
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: FORRESTER

DATE/TIME OF TEST: 05/30/2021 02:42:52

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.201	02:38
Breath Volume	1.265	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.191	02:41
Breath Volume	3.445	02:41
Air Blank	0.000	02:41
Control Test	0.079	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 06/08/2021 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LITTLE, ELAINA
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 05/30/2021 03:31:06

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.076	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.123	03:26
Breath Volume	2.011	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.123	03:29
Breath Volume	2.398	03:29
Air Blank	0.000	03:30
Control Test	0.075	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 06/08/2021 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: STARNES,JACOB
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: CANINO

DATE/TIME OF TEST: 05/30/2021 05:08:32

Diagnostics Check	OK	05:02
Air Blank	0.000	05:02
Control Test	0.079	05:03
Air Blank	0.000	05:03
Subject Sample # 1	0.000	05:04
Breath Volume	2.402	05:04
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.000	05:07
Breath Volume	1.992	05:07
Air Blank	0.000	05:07
Control Test	0.079	05:07
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 06/08/2021 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: DEBLASE

DATE/TIME OF TEST: 05/31/2021 00:26:10

Diagnostics Check		OK	00:12	
Air Blank		0.000	00:12	
Control Test		0.080	00:13	
Air Blank		0.000	00:13	
Subject Sample	# 1	0.117	00:16	Volume Not Met
Breath Volume		0.339	00:16	
Air Blank		0.000	00:17	
Air Blank		0.000	00:19	
Subject Sample	# 2	0.094	00:21	
Breath Volume		1.628	00:21	
Air Blank		0.000	00:22	
Air Blank		0.000	00:24	
Subject Sample	# 3	0.104	00:24	
Breath Volume		1.429	00:24	
Air Blank		0.000	00:25	
Control Test		0.076	00:25	
Air Blank		0.000	00:26	
Diagnostics Check		OK	00:26	

Data Download Date/Time: 06/08/2021 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CALISTE,VIXON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: WIEST

DATE/TIME OF TEST: 05/31/2021 20:07:05

Diagnostics Check	OK	20:00
Air Blank	0.000	20:01
Control Test	0.079	20:01
Air Blank	0.000	20:02
Subject Sample # 1	0.178	20:02
Breath Volume	2.125	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.175	20:05
Breath Volume	2.355	20:05
Air Blank	0.000	20:06
Control Test	0.078	20:06
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 06/08/2021 10:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN,CHARLES,E

BREATH TEST OPERATOR AGENCY: Gretna PD

ARRESTING OFFICER: BARBER,HAROLD

ARRESTING OFFICER AGENCY: Quincy PD

LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 04120080A1

DRY GAS EXPIRY DATE: 03/05/2022

VIOLATION CODE: DUI

ARREST TIME: 22:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: HAVERSTRAW

DATE/TIME OF TEST: 05/11/2021 23:17:18

Diagnostics Check	OK	23:14	
Air Blank	0.000	23:14	
Control Test	0.079	23:15	
Air Blank	0.000	23:15	
Subject Sample	# 1	23:15	Subject Test Refused
Breath Volume	0.000	23:15	
Air Blank	0.000	23:16	
Control Test	0.080	23:16	
Air Blank	0.000	23:17	
Diagnostics Check	OK	23:17	

Data Download Date/Time: 06/28/2021 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,JOSEPH,B
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FRANKLIN,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:58

SUBJECT'S LAST NAME: PETTUS

DATE/TIME OF TEST: 05/13/2021 08:24:37

Diagnostics Check	OK	08:18
Air Blank	0.000	08:18
Control Test	0.080	08:19
Air Blank	0.000	08:19
Subject Sample # 1	0.000	08:20
Breath Volume	2.250	08:20
Air Blank	0.000	08:20
Air Blank	0.000	08:22
Subject Sample # 2	0.000	08:23
Breath Volume	1.488	08:23
Air Blank	0.000	08:23
Control Test	0.081	08:24
Air Blank	0.000	08:24
Diagnostics Check	OK	08:24

Data Download Date/Time: 06/28/2021 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REAGLE,ASHLEY,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: REAGLE,ASHLEY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MOWBRAY

DATE/TIME OF TEST: 05/21/2021 02:19:44

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.079	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.267	02:12
Breath Volume		4.429	02:12
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.222	02:15
Breath Volume		1.785	02:15
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 3	0.255	02:18
Breath Volume		3.691	02:18
Air Blank		0.000	02:18
Control Test		0.078	02:19
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 06/28/2021 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: COLKMIRE

DATE/TIME OF TEST: 05/30/2021 05:27:57

Diagnostics Check	OK	05:21
Air Blank	0.000	05:22
Control Test	0.079	05:22
Air Blank	0.000	05:23
Subject Sample # 1	0.121	05:23
Breath Volume	3.847	05:23
Air Blank	0.000	05:24
Air Blank	0.000	05:25
Subject Sample # 2	0.118	05:26
Breath Volume	4.582	05:26
Air Blank	0.000	05:27
Control Test	0.080	05:27
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 06/28/2021 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORAN,ETHAN,L
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FORAN,ETHAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: BESSIGANO

DATE/TIME OF TEST: 05/31/2021 09:11:01

Diagnostics Check	OK	09:07	
Air Blank	0.000	09:08	
Control Test	0.079	09:08	
Air Blank	0.000	09:09	
Subject Sample	# 1	09:09	Subject Test Refused
Breath Volume	0.000	09:09	
Air Blank	0.000	09:10	
Control Test	0.079	09:10	
Air Blank	0.000	09:10	
Diagnostics Check	OK	09:11	

Data Download Date/Time: 06/28/2021 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHOBREY,SPENCER,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WHOBREY,SPENCER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: AYERS

DATE/TIME OF TEST: 05/06/2021 21:33:21

Diagnostics Check	OK	21:26
Air Blank	0.000	21:27
Control Test	0.080	21:27
Air Blank	0.000	21:27
Subject Sample # 1	0.083	21:28
Breath Volume	1.750	21:28
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.076	21:31
Breath Volume	3.062	21:31
Air Blank	0.000	21:32
Control Test	0.081	21:32
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 05/11/2021 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HELLER,MICHAEL
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 09:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:27

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 05/13/2021 10:54:19

Diagnostics Check	OK	10:47
Air Blank	0.000	10:48
Control Test	0.081	10:48
Air Blank	0.000	10:48
Subject Sample # 1	0.000	10:49
Breath Volume	2.445	10:49
Air Blank	0.000	10:50
Air Blank	0.000	10:52
Subject Sample # 2	0.000	10:52
Breath Volume	2.757	10:52
Air Blank	0.000	10:53
Control Test	0.080	10:53
Air Blank	0.000	10:54
Diagnostics Check	OK	10:54

Data Download Date/Time: 06/03/2021 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MILAM,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:14

SUBJECT'S LAST NAME: ALLI

DATE/TIME OF TEST: 05/22/2021 06:36:56

Diagnostics Check	OK	06:35	
Air Blank	0.000	06:35	
Control Test	0.082	06:36	
Air Blank		06:36	RFI Detect
Air Blank	0.000	06:36	

Data Download Date/Time: 06/03/2021 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MILAN,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: MAHAMAD

DATE/TIME OF TEST: 05/22/2021 06:53:34

Diagnostics Check		OK	06:39	
Air Blank		0.000	06:39	
Control Test		0.081	06:40	
Air Blank		0.000	06:40	
Subject Sample	# 1	0.133	06:43	Volume Not Met
Breath Volume		0.976	06:43	
Air Blank		0.000	06:44	
Air Blank		0.000	06:46	
Subject Sample	# 2	0.127	06:48	
Breath Volume		1.566	06:48	
Air Blank		0.000	06:49	
Air Blank		0.000	06:51	
Subject Sample	# 3	0.136	06:51	
Breath Volume		1.933	06:51	
Air Blank		0.000	06:52	
Control Test		0.078	06:52	
Air Blank		0.000	06:53	
Diagnostics Check		OK	06:53	

Data Download Date/Time: 06/03/2021 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COODY,TASHAHNDA,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: COODY,TASHAHNDA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 11:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: BAUGH

DATE/TIME OF TEST: 05/23/2021 00:52:03

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.081	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.085	00:47
Breath Volume	1.464	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:49
Subject Sample # 2	0.083	00:50
Breath Volume	1.597	00:50
Air Blank	0.000	00:51
Control Test	0.081	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:52

Data Download Date/Time: 06/03/2021 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRISWOLD,DAVID,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: GREENWALL,ROBERT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: SYNOWIEC

DATE/TIME OF TEST: 05/27/2021 01:06:15

Diagnostics Check		OK	00:59
Air Blank		0.000	01:00
Control Test		0.081	01:00
Air Blank		0.000	01:00
Subject Sample	# 1	0.124	01:01
Breath Volume		1.332	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.123	01:04
Breath Volume		2.101	01:04
Air Blank		0.000	01:05
Control Test		0.080	01:05
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 06/03/2021 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COODY,TASHAHNDA,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MILAN,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 05/29/2021 05:20:03

Diagnostics Check		OK	05:13
Air Blank		0.000	05:13
Control Test		0.081	05:14
Air Blank		0.000	05:14
Subject Sample	# 1	0.121	05:15
Breath Volume		2.804	05:15
Air Blank		0.000	05:16
Air Blank		0.000	05:18
Subject Sample	# 2	0.114	05:18
Breath Volume		1.964	05:18
Air Blank		0.000	05:19
Control Test		0.081	05:19
Air Blank		0.000	05:19
Diagnostics Check		OK	05:20

Data Download Date/Time: 06/03/2021 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLIER,THOMAS,L
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MARTINEZ-
IRIZARRY,CHARLES
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 09:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:01

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 05/30/2021 10:25:11

Diagnostics Check	OK	10:22	
Air Blank	0.000	10:22	
Control Test	0.080	10:23	
Air Blank	0.000	10:23	
Subject Sample	# 1	10:23	Subject Test Refused
Breath Volume	0.000	10:23	
Air Blank	0.000	10:24	
Control Test	0.080	10:24	
Air Blank	0.000	10:25	
Diagnostics Check	OK	10:25	

Data Download Date/Time: 06/03/2021 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HILL,ROBERT
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/04/2021 23:19:42

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.081	23:14
Air Blank	0.000	23:14
Subject Sample # 1	0.147	23:15
Breath Volume	1.492	23:15
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.143	23:18
Breath Volume	2.292	23:18
Air Blank	0.000	23:18
Control Test	0.081	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 05/07/2021 11:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INGERSOLL,NICHOLAS,F
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HATZEL,BRANDON
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: MCALISTER

DATE/TIME OF TEST: 05/08/2021 00:19:55

Diagnostics Check		OK	00:12
Air Blank		0.000	00:12
Control Test		0.081	00:13
Air Blank		0.000	00:13
Subject Sample	# 1	0.211	00:14
Breath Volume		1.316	00:14
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.205	00:18
Breath Volume		1.605	00:18
Air Blank		0.000	00:19
Control Test		0.081	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 05/18/2021 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INGERSOLL,NICHOLAS,F
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: ADAMES,RAUL
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: HARTLE

DATE/TIME OF TEST: 05/09/2021 02:30:33

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.081	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.181	02:25
Breath Volume		1.882	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.175	02:29
Breath Volume		1.687	02:29
Air Blank		0.000	02:29
Control Test		0.082	02:29
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 05/18/2021 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INGERSOLL,NICHOLAS,F
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HATZEL,BRANDON
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: BERTRAND

DATE/TIME OF TEST: 05/31/2021 22:07:05

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:02
Subject Sample # 1	0.000	22:02
Breath Volume	2.417	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.000	22:05
Breath Volume	2.097	22:05
Air Blank	0.000	22:06
Control Test	0.079	22:06
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 06/04/2021 09:54