

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 06/25/2021 02:14:55

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 02:08 |
| Air Blank | | 0.000 | 02:09 |
| Control Test | | 0.080 | 02:09 |
| Air Blank | | 0.000 | 02:10 |
| Subject Sample | # 1 | 0.254 | 02:10 |
| Breath Volume | | 3.292 | 02:10 |
| Air Blank | | 0.000 | 02:11 |
| Air Blank | | 0.000 | 02:12 |
| Subject Sample | # 2 | 0.261 | 02:13 |
| Breath Volume | | 2.914 | 02:13 |
| Air Blank | | 0.000 | 02:13 |
| Control Test | | 0.079 | 02:14 |
| Air Blank | | 0.000 | 02:14 |
| Diagnostics Check | | OK | 02:14 |

Data Download Date/Time: 07/23/2021 00:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: PEREZ AGUILAR

DATE/TIME OF TEST: 06/09/2021 02:26:30

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:20 |
| Air Blank | 0.000 | 02:20 |
| Control Test | 0.080 | 02:21 |
| Air Blank | 0.000 | 02:21 |
| Subject Sample # 1 | 0.197 | 02:22 |
| Breath Volume | 1.847 | 02:22 |
| Air Blank | 0.000 | 02:22 |
| Air Blank | 0.000 | 02:24 |
| Subject Sample # 2 | 0.209 | 02:24 |
| Breath Volume | 2.117 | 02:24 |
| Air Blank | 0.000 | 02:25 |
| Control Test | 0.080 | 02:25 |
| Air Blank | 0.000 | 02:26 |
| Diagnostics Check | OK | 02:26 |

Data Download Date/Time: 06/30/2021 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DIAZ JIMENEZ

DATE/TIME OF TEST: 06/10/2021 02:04:36

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:58 |
| Air Blank | 0.000 | 01:58 |
| Control Test | 0.080 | 01:59 |
| Air Blank | 0.000 | 01:59 |
| Subject Sample # 1 | 0.209 | 02:00 |
| Breath Volume | 1.886 | 02:00 |
| Air Blank | 0.000 | 02:00 |
| Air Blank | 0.000 | 02:02 |
| Subject Sample # 2 | 0.194 | 02:02 |
| Breath Volume | 2.265 | 02:02 |
| Air Blank | 0.000 | 02:03 |
| Control Test | 0.079 | 02:04 |
| Air Blank | 0.000 | 02:04 |
| Diagnostics Check | OK | 02:04 |

Data Download Date/Time: 06/30/2021 14:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: TOMAS DOMINGUES

DATE/TIME OF TEST: 06/10/2021 05:24:53

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 05:22 | |
| Air Blank | 0.000 | 05:22 | |
| Control Test | 0.080 | 05:22 | |
| Air Blank | 0.000 | 05:23 | |
| Subject Sample | # 1 | 05:23 | Subject Test Refused |
| Breath Volume | 0.000 | 05:23 | |
| Air Blank | 0.000 | 05:23 | |
| Control Test | 0.080 | 05:24 | |
| Air Blank | 0.000 | 05:24 | |
| Diagnostics Check | OK | 05:24 | |

Data Download Date/Time: 06/30/2021 14:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 06/11/2021 03:58:36

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 03:55 | |
| Air Blank | 0.000 | 03:56 | |
| Control Test | 0.080 | 03:56 | |
| Air Blank | 0.000 | 03:56 | |
| Subject Sample | # 1 | 03:57 | Subject Test Refused |
| Breath Volume | 0.000 | 03:57 | |
| Air Blank | 0.000 | 03:57 | |
| Control Test | 0.080 | 03:57 | |
| Air Blank | 0.000 | 03:58 | |
| Diagnostics Check | OK | 03:58 | |

Data Download Date/Time: 06/30/2021 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: GONZALEZ PEREZ

DATE/TIME OF TEST: 06/12/2021 01:19:06

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 01:09 |
| Air Blank | | 0.000 | 01:10 |
| Control Test | | 0.080 | 01:10 |
| Air Blank | | 0.000 | 01:11 |
| Subject Sample | # 1 | 0.213 | 01:13 |
| Breath Volume | | 1.265 | 01:13 |
| Air Blank | | 0.000 | 01:14 |
| Air Blank | | 0.000 | 01:16 |
| Subject Sample | # 2 | 0.215 | 01:17 |
| Breath Volume | | 1.300 | 01:17 |
| Air Blank | | 0.000 | 01:18 |
| Control Test | | 0.078 | 01:18 |
| Air Blank | | 0.000 | 01:19 |
| Diagnostics Check | | OK | 01:19 |

Data Download Date/Time: 06/30/2021 14:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: RAYMUNDO

DATE/TIME OF TEST: 06/12/2021 02:37:12

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 02:34 | |
| Air Blank | 0.000 | 02:34 | |
| Control Test | 0.080 | 02:35 | |
| Air Blank | 0.000 | 02:35 | |
| Subject Sample | # 1 | 02:35 | Subject Test Refused |
| Breath Volume | 0.000 | 02:35 | |
| Air Blank | 0.000 | 02:36 | |
| Control Test | 0.080 | 02:36 | |
| Air Blank | 0.000 | 02:37 | |
| Diagnostics Check | OK | 02:37 | |

Data Download Date/Time: 06/30/2021 14:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: MONTOYA VILLACRES

DATE/TIME OF TEST: 06/12/2021 05:14:55

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 05:12 | |
| Air Blank | 0.000 | 05:12 | |
| Control Test | 0.078 | 05:12 | |
| Air Blank | 0.000 | 05:13 | |
| Subject Sample | # 1 | 05:13 | Subject Test Refused |
| Breath Volume | 0.000 | 05:13 | |
| Air Blank | 0.000 | 05:14 | |
| Control Test | 0.077 | 05:14 | |
| Air Blank | 0.000 | 05:14 | |
| Diagnostics Check | OK | 05:14 | |

Data Download Date/Time: 06/30/2021 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: EGGERT

DATE/TIME OF TEST: 06/17/2021 04:15:43

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 04:09 |
| Air Blank | | 0.000 | 04:10 |
| Control Test | | 0.080 | 04:10 |
| Air Blank | | 0.000 | 04:11 |
| Subject Sample | # 1 | 0.000 | 04:11 |
| Breath Volume | | 2.007 | 04:11 |
| Air Blank | | 0.000 | 04:12 |
| Air Blank | | 0.000 | 04:14 |
| Subject Sample | # 2 | 0.000 | 04:14 |
| Breath Volume | | 2.414 | 04:14 |
| Air Blank | | 0.000 | 04:14 |
| Control Test | | 0.081 | 04:15 |
| Air Blank | | 0.000 | 04:15 |
| Diagnostics Check | | OK | 04:15 |

Data Download Date/Time: 06/30/2021 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINCORE,MICHAEL,A
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: EDWARDS,SHAWN
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1237851
DRY GAS EXPIRY DATE: 11/21/2022
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: NUNEZRODRIGUEZ

DATE/TIME OF TEST: 06/17/2021 22:30:11

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:22 |
| Air Blank | 0.000 | 22:22 |
| Control Test | 0.080 | 22:23 |
| Air Blank | 0.000 | 22:23 |
| Subject Sample # 1 | 0.210 | 22:25 |
| Breath Volume | 2.503 | 22:25 |
| Air Blank | 0.000 | 22:26 |
| Air Blank | 0.000 | 22:27 |
| Subject Sample # 2 | 0.207 | 22:28 |
| Breath Volume | 1.980 | 22:28 |
| Air Blank | 0.000 | 22:29 |
| Control Test | 0.078 | 22:29 |
| Air Blank | 0.000 | 22:30 |
| Diagnostics Check | OK | 22:30 |

Data Download Date/Time: 06/30/2021 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 06/18/2021 03:24:40

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 03:16 |
| Air Blank | | 0.000 | 03:17 |
| Control Test | | 0.080 | 03:17 |
| Air Blank | | 0.000 | 03:17 |
| Subject Sample | # 1 | 0.268 | 03:20 |
| Breath Volume | | 1.933 | 03:20 |
| Air Blank | | 0.000 | 03:21 |
| Air Blank | | 0.000 | 03:22 |
| Subject Sample | # 2 | 0.272 | 03:22 |
| Breath Volume | | 2.933 | 03:22 |
| Air Blank | | 0.000 | 03:23 |
| Control Test | | 0.075 | 03:24 |
| Air Blank | | 0.000 | 03:24 |
| Diagnostics Check | | OK | 03:24 |

Data Download Date/Time: 06/30/2021 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: JIMENEZ PEREZ

DATE/TIME OF TEST: 06/19/2021 02:13:15

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:07 |
| Air Blank | 0.000 | 02:07 |
| Control Test | 0.079 | 02:07 |
| Air Blank | 0.000 | 02:08 |
| Subject Sample # 1 | 0.147 | 02:08 |
| Breath Volume | 2.710 | 02:08 |
| Air Blank | 0.000 | 02:09 |
| Air Blank | 0.000 | 02:11 |
| Subject Sample # 2 | 0.154 | 02:11 |
| Breath Volume | 3.214 | 02:11 |
| Air Blank | 0.000 | 02:12 |
| Control Test | 0.080 | 02:12 |
| Air Blank | 0.000 | 02:13 |
| Diagnostics Check | OK | 02:13 |

Data Download Date/Time: 06/30/2021 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINCORE,MICHAEL,A
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: SINCORE,MICHAEL
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1237851
DRY GAS EXPIRY DATE: 11/21/2022
VIOLATION CODE: DUI
ARREST TIME: 15:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:32

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 06/19/2021 16:01:03

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 15:52 | |
| Air Blank | 0.000 | 15:53 | |
| Control Test | 0.079 | 15:53 | |
| Air Blank | 0.000 | 15:53 | |
| Subject Sample | # 1 | 15:57 | No Sample Provided |
| Breath Volume | 0.000 | 15:57 | |
| Air Blank | 0.000 | 15:57 | |
| Air Blank | 0.000 | 15:59 | |
| Subject Sample | # 2 | 15:59 | Subject Test Refused |
| Breath Volume | 0.000 | 15:59 | |
| Air Blank | 0.000 | 16:00 | |
| Control Test | 0.081 | 16:00 | |
| Air Blank | 0.000 | 16:00 | |
| Diagnostics Check | OK | 16:01 | |

Data Download Date/Time: 06/30/2021 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: FARFAN OTZIN

DATE/TIME OF TEST: 06/24/2021 02:44:01

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 02:37 |
| Air Blank | | 0.000 | 02:38 |
| Control Test | | 0.080 | 02:38 |
| Air Blank | | 0.000 | 02:39 |
| Subject Sample | # 1 | 0.210 | 02:39 |
| Breath Volume | | 2.460 | 02:39 |
| Air Blank | | 0.000 | 02:40 |
| Air Blank | | 0.000 | 02:41 |
| Subject Sample | # 2 | 0.212 | 02:42 |
| Breath Volume | | 1.531 | 02:42 |
| Air Blank | | 0.000 | 02:43 |
| Control Test | | 0.079 | 02:43 |
| Air Blank | | 0.000 | 02:43 |
| Diagnostics Check | | OK | 02:44 |

Data Download Date/Time: 06/30/2021 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: PEREZ MENDEZ

DATE/TIME OF TEST: 06/24/2021 04:12:36

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 04:09 | |
| Air Blank | 0.000 | 04:10 | |
| Control Test | 0.078 | 04:10 | |
| Air Blank | 0.000 | 04:11 | |
| Subject Sample | # 1 | 04:11 | Subject Test Refused |
| Breath Volume | 0.000 | 04:11 | |
| Air Blank | 0.000 | 04:11 | |
| Control Test | 0.078 | 04:12 | |
| Air Blank | 0.000 | 04:12 | |
| Diagnostics Check | OK | 04:12 | |

Data Download Date/Time: 06/30/2021 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINCORE,MICHAEL,A
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: SINCORE,MICHAEL
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1237851
DRY GAS EXPIRY DATE: 11/21/2022
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/25/2021 22:20:50

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:14 |
| Air Blank | 0.000 | 22:14 |
| Control Test | 0.080 | 22:15 |
| Air Blank | 0.000 | 22:15 |
| Subject Sample # 1 | 0.296 | 22:16 |
| Breath Volume | 2.121 | 22:16 |
| Air Blank | 0.000 | 22:17 |
| Air Blank | 0.000 | 22:18 |
| Subject Sample # 2 | 0.300 | 22:19 |
| Breath Volume | 1.835 | 22:19 |
| Air Blank | 0.000 | 22:19 |
| Control Test | 0.079 | 22:20 |
| Air Blank | 0.000 | 22:20 |
| Diagnostics Check | OK | 22:20 |

Data Download Date/Time: 06/30/2021 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURGUIDO,ALEX,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: MURGUIDO,ALEX
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 09:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: EASTON

DATE/TIME OF TEST: 06/26/2021 10:37:35

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 10:31 |
| Air Blank | 0.000 | 10:31 |
| Control Test | 0.079 | 10:31 |
| Air Blank | 0.000 | 10:32 |
| Subject Sample # 1 | 0.240 | 10:32 |
| Breath Volume | 2.429 | 10:32 |
| Air Blank | 0.000 | 10:33 |
| Air Blank | 0.000 | 10:35 |
| Subject Sample # 2 | 0.234 | 10:35 |
| Breath Volume | 1.226 | 10:35 |
| Air Blank | 0.000 | 10:36 |
| Control Test | 0.077 | 10:36 |
| Air Blank | 0.000 | 10:37 |
| Diagnostics Check | OK | 10:37 |

Data Download Date/Time: 06/30/2021 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER, BLAKE, E
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CARPENTER, BLAKE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/16/2021 02:37:09

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 02:25 | |
| Air Blank | | 0.000 | 02:25 | |
| Control Test | | 0.078 | 02:25 | |
| Air Blank | | 0.000 | 02:26 | |
| Subject Sample | # 1 | 0.140 | 02:29 | Volume Not Met |
| Breath Volume | | 0.527 | 02:29 | |
| Air Blank | | 0.000 | 02:30 | |
| Air Blank | | 0.000 | 02:32 | |
| Subject Sample | # 2 | 0.134 | 02:35 | Volume Not Met |
| Breath Volume | | 0.539 | 02:35 | |
| Air Blank | | 0.000 | 02:36 | |
| Control Test | | 0.077 | 02:36 | |
| Air Blank | | 0.000 | 02:37 | |
| Diagnostics Check | | OK | 02:37 | |

Data Download Date/Time: 07/23/2021 22:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER, BLAKE, E
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CARPENTER, BLAKE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/16/2021 02:53:17

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:39 |
| Air Blank | 0.000 | 02:39 |
| Control Test | 0.077 | 02:40 |
| Air Blank | 0.000 | 02:40 |
| Subject Sample # 1 | 0.111 | 02:44 |
| Breath Volume | 1.816 | 02:44 |
| Air Blank | 0.000 | 02:44 |
| Air Blank | 0.000 | 02:46 |
| Subject Sample # 2 | 0.139 | 02:48 |
| Breath Volume | 2.007 | 02:48 |
| Air Blank | 0.000 | 02:49 |
| Air Blank | 0.000 | 02:50 |
| Subject Sample # 3 | 0.140 | 02:51 |
| Breath Volume | 1.785 | 02:51 |
| Air Blank | 0.000 | 02:52 |
| Control Test | 0.076 | 02:52 |
| Air Blank | 0.000 | 02:53 |
| Diagnostics Check | OK | 02:53 |

Data Download Date/Time: 07/23/2021 22:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN, BRADLEY,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MCLAUGHLIN, BRADLEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 06/19/2021 03:54:40

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:48 |
| Air Blank | 0.000 | 03:48 |
| Control Test | 0.078 | 03:49 |
| Air Blank | 0.000 | 03:49 |
| Subject Sample # 1 | 0.113 | 03:50 |
| Breath Volume | 2.921 | 03:50 |
| Air Blank | 0.000 | 03:50 |
| Air Blank | 0.000 | 03:52 |
| Subject Sample # 2 | 0.107 | 03:52 |
| Breath Volume | 2.445 | 03:52 |
| Air Blank | 0.000 | 03:53 |
| Control Test | 0.076 | 03:54 |
| Air Blank | 0.000 | 03:54 |
| Diagnostics Check | OK | 03:54 |

Data Download Date/Time: 07/23/2021 22:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JUSTIN,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WIMER,STANTON
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: LUPO

DATE/TIME OF TEST: 06/23/2021 04:04:14

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 03:57 |
| Air Blank | | 0.000 | 03:58 |
| Control Test | | 0.078 | 03:58 |
| Air Blank | | 0.000 | 03:58 |
| Subject Sample | # 1 | 0.250 | 03:59 |
| Breath Volume | | 1.964 | 03:59 |
| Air Blank | | 0.000 | 04:00 |
| Air Blank | | 0.000 | 04:02 |
| Subject Sample | # 2 | 0.246 | 04:02 |
| Breath Volume | | 1.851 | 04:02 |
| Air Blank | | 0.000 | 04:03 |
| Control Test | | 0.077 | 04:03 |
| Air Blank | | 0.000 | 04:04 |
| Diagnostics Check | | OK | 04:04 |

Data Download Date/Time: 07/23/2021 22:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAN, TYLER, M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: THOMAN, TYLER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:53

SUBJECT'S LAST NAME: SMILING

DATE/TIME OF TEST: 06/24/2021 11:17:09

| | | | |
|--------------------|-------|-------|------------|
| Diagnostics Check | OK | 11:13 | |
| Air Blank | 0.000 | 11:13 | |
| Control Test | 0.078 | 11:14 | |
| Air Blank | 0.000 | 11:14 | |
| Subject Sample # 1 | 0.019 | 11:15 | |
| Breath Volume | 1.773 | 11:15 | |
| Air Blank | | 11:15 | RFI Detect |
| Air Blank | 0.000 | 11:16 | |
| Control Test | 0.078 | 11:16 | |
| Air Blank | 0.000 | 11:17 | |
| Diagnostics Check | OK | 11:17 | |

Data Download Date/Time: 07/23/2021 22:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAN, TYLER, M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: THOMAN, TYLER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:53

SUBJECT'S LAST NAME: SMILING

DATE/TIME OF TEST: 06/24/2021 11:25:55

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 11:18 |
| Air Blank | | 0.000 | 11:19 |
| Control Test | | 0.079 | 11:19 |
| Air Blank | | 0.000 | 11:20 |
| Subject Sample | # 1 | 0.018 | 11:21 |
| Breath Volume | | 2.187 | 11:21 |
| Air Blank | | 0.000 | 11:21 |
| Air Blank | | 0.000 | 11:23 |
| Subject Sample | # 2 | 0.017 | 11:24 |
| Breath Volume | | 1.160 | 11:24 |
| Air Blank | | 0.000 | 11:24 |
| Control Test | | 0.079 | 11:25 |
| Air Blank | | 0.000 | 11:25 |
| Diagnostics Check | | OK | 11:25 |

Data Download Date/Time: 07/23/2021 22:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JUSTIN,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WIMER,STANTON
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:06

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 06/25/2021 23:10:52

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 23:07 | |
| Air Blank | 0.000 | 23:08 | |
| Control Test | 0.079 | 23:08 | |
| Air Blank | 0.000 | 23:09 | |
| Subject Sample | # 1 | 23:09 | Subject Test Refused |
| Breath Volume | 0.000 | 23:09 | |
| Air Blank | 0.000 | 23:09 | |
| Control Test | 0.079 | 23:10 | |
| Air Blank | 0.000 | 23:10 | |
| Diagnostics Check | OK | 23:10 | |

Data Download Date/Time: 07/23/2021 22:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JUSTIN,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WIMER,STANTON
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 06/25/2021 23:32:23

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:26 |
| Air Blank | 0.000 | 23:26 |
| Control Test | 0.079 | 23:26 |
| Air Blank | 0.000 | 23:27 |
| Subject Sample # 1 | 0.000 | 23:28 |
| Breath Volume | 2.683 | 23:28 |
| Air Blank | 0.000 | 23:28 |
| Air Blank | 0.000 | 23:30 |
| Subject Sample # 2 | 0.000 | 23:31 |
| Breath Volume | 1.441 | 23:31 |
| Air Blank | 0.000 | 23:31 |
| Control Test | 0.076 | 23:31 |
| Air Blank | 0.000 | 23:32 |
| Diagnostics Check | OK | 23:32 |

Data Download Date/Time: 07/23/2021 22:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER, BLAKE, E
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CARPENTER, BLAKE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 06/29/2021 03:41:37

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 03:34 |
| Air Blank | | 0.000 | 03:34 |
| Control Test | | 0.079 | 03:35 |
| Air Blank | | 0.000 | 03:35 |
| Subject Sample | # 1 | 0.177 | 03:36 |
| Breath Volume | | 2.460 | 03:36 |
| Air Blank | | 0.000 | 03:37 |
| Air Blank | | 0.000 | 03:39 |
| Subject Sample | # 2 | 0.175 | 03:39 |
| Breath Volume | | 2.691 | 03:39 |
| Air Blank | | 0.000 | 03:40 |
| Control Test | | 0.078 | 03:40 |
| Air Blank | | 0.000 | 03:41 |
| Diagnostics Check | | OK | 03:41 |

Data Download Date/Time: 07/23/2021 22:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JUSTIN,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ADAMS,JUSTIN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: ZWOLENSKY

DATE/TIME OF TEST: 06/29/2021 23:47:28

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:40 |
| Air Blank | 0.000 | 23:41 |
| Control Test | 0.079 | 23:41 |
| Air Blank | 0.000 | 23:41 |
| Subject Sample # 1 | 0.187 | 23:42 |
| Breath Volume | 2.617 | 23:42 |
| Air Blank | 0.000 | 23:43 |
| Air Blank | 0.000 | 23:45 |
| Subject Sample # 2 | 0.175 | 23:45 |
| Breath Volume | 1.570 | 23:45 |
| Air Blank | 0.000 | 23:46 |
| Control Test | 0.078 | 23:46 |
| Air Blank | 0.000 | 23:47 |
| Diagnostics Check | OK | 23:47 |

Data Download Date/Time: 07/23/2021 22:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON, RAYMOND, M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: COLON, RAYMOND
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: PEREZ AMADOR

DATE/TIME OF TEST: 06/05/2021 04:55:58

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 04:53 | |
| Air Blank | 0.000 | 04:53 | |
| Control Test | 0.082 | 04:53 | |
| Air Blank | 0.000 | 04:54 | |
| Subject Sample | # 1 | 04:54 | Subject Test Refused |
| Breath Volume | 0.000 | 04:54 | |
| Air Blank | 0.000 | 04:55 | |
| Control Test | 0.082 | 04:55 | |
| Air Blank | 0.000 | 04:55 | |
| Diagnostics Check | OK | 04:55 | |

Data Download Date/Time: 06/28/2021 16:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON, RAYMOND, M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: ROSADO, KIARA
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 06/05/2021 21:16:02

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 21:13 | |
| Air Blank | 0.000 | 21:13 | |
| Control Test | 0.082 | 21:13 | |
| Air Blank | 0.000 | 21:14 | |
| Subject Sample | # 1 | 21:14 | Subject Test Refused |
| Breath Volume | 0.000 | 21:14 | |
| Air Blank | 0.000 | 21:15 | |
| Control Test | 0.082 | 21:15 | |
| Air Blank | 0.000 | 21:15 | |
| Diagnostics Check | OK | 21:16 | |

Data Download Date/Time: 06/28/2021 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: ROSADO,KIARA
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 06:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: HURTADO

DATE/TIME OF TEST: 06/06/2021 06:53:30

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 06:50 | |
| Air Blank | 0.000 | 06:50 | |
| Control Test | 0.082 | 06:51 | |
| Air Blank | 0.000 | 06:51 | |
| Subject Sample | # 1 | 06:52 | Subject Test Refused |
| Breath Volume | 0.000 | 06:52 | |
| Air Blank | 0.000 | 06:52 | |
| Control Test | 0.082 | 06:52 | |
| Air Blank | 0.000 | 06:53 | |
| Diagnostics Check | OK | 06:53 | |

Data Download Date/Time: 06/28/2021 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,EVEN,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: DREW,LIAM
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: TELLEZ MORA

DATE/TIME OF TEST: 06/12/2021 20:48:05

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:40 |
| Air Blank | 0.000 | 20:40 |
| Control Test | 0.082 | 20:41 |
| Air Blank | 0.000 | 20:41 |
| Subject Sample # 1 | 0.222 | 20:43 |
| Breath Volume | 1.949 | 20:43 |
| Air Blank | 0.000 | 20:43 |
| Air Blank | 0.000 | 20:45 |
| Subject Sample # 2 | 0.231 | 20:46 |
| Breath Volume | 3.070 | 20:46 |
| Air Blank | 0.000 | 20:47 |
| Control Test | 0.080 | 20:47 |
| Air Blank | 0.000 | 20:47 |
| Diagnostics Check | OK | 20:48 |

Data Download Date/Time: 06/28/2021 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON, RAYMOND, M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CORONILLA, DANIEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/31/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: GALINDO

DATE/TIME OF TEST: 06/19/2021 20:51:25

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:44 |
| Air Blank | 0.000 | 20:44 |
| Control Test | 0.083 | 20:45 |
| Air Blank | 0.000 | 20:45 |
| Subject Sample # 1 | 0.140 | 20:46 |
| Breath Volume | 1.609 | 20:46 |
| Air Blank | 0.000 | 20:47 |
| Air Blank | 0.000 | 20:49 |
| Subject Sample # 2 | 0.142 | 20:49 |
| Breath Volume | 1.859 | 20:49 |
| Air Blank | 0.000 | 20:50 |
| Control Test | 0.083 | 20:50 |
| Air Blank | 0.000 | 20:51 |
| Diagnostics Check | OK | 20:51 |

Data Download Date/Time: 06/28/2021 16:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON, RAYMOND, M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: KRASINSKAS, RICHARD
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/31/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 06/20/2021 03:10:40

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:03 |
| Air Blank | 0.000 | 03:03 |
| Control Test | 0.083 | 03:04 |
| Air Blank | 0.000 | 03:04 |
| Subject Sample # 1 | 0.113 | 03:05 |
| Breath Volume | 2.457 | 03:05 |
| Air Blank | 0.000 | 03:06 |
| Air Blank | 0.000 | 03:08 |
| Subject Sample # 2 | 0.109 | 03:09 |
| Breath Volume | 2.394 | 03:09 |
| Air Blank | 0.000 | 03:09 |
| Control Test | 0.082 | 03:10 |
| Air Blank | 0.000 | 03:10 |
| Diagnostics Check | OK | 03:10 |

Data Download Date/Time: 06/28/2021 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON, RAYMOND, M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CORONILLA, DANIEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: YARBER

DATE/TIME OF TEST: 06/20/2021 21:07:49

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:00 |
| Air Blank | 0.000 | 21:01 |
| Control Test | 0.082 | 21:01 |
| Air Blank | 0.000 | 21:02 |
| Subject Sample # 1 | 0.235 | 21:03 |
| Breath Volume | 1.960 | 21:03 |
| Air Blank | 0.000 | 21:03 |
| Air Blank | 0.000 | 21:05 |
| Subject Sample # 2 | 0.221 | 21:06 |
| Breath Volume | 1.339 | 21:06 |
| Air Blank | 0.000 | 21:06 |
| Control Test | 0.081 | 21:07 |
| Air Blank | 0.000 | 21:07 |
| Diagnostics Check | OK | 21:07 |

Data Download Date/Time: 06/28/2021 16:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON, RAYMOND, M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: ROSADO, KIARA
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 06/24/2021 22:40:28

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 22:37 | |
| Air Blank | 0.000 | 22:37 | |
| Control Test | 0.082 | 22:38 | |
| Air Blank | 0.000 | 22:38 | |
| Subject Sample | # 1 | 22:39 | Subject Test Refused |
| Breath Volume | 0.000 | 22:39 | |
| Air Blank | 0.000 | 22:39 | |
| Control Test | 0.082 | 22:39 | |
| Air Blank | 0.000 | 22:40 | |
| Diagnostics Check | OK | 22:40 | |

Data Download Date/Time: 06/28/2021 16:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER III,RICHARD,F
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SNYDER III,RICHARD
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: SIMILIEN

DATE/TIME OF TEST: 06/11/2021 00:52:20

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:45 |
| Air Blank | 0.000 | 00:46 |
| Control Test | 0.080 | 00:46 |
| Air Blank | 0.000 | 00:46 |
| Subject Sample # 1 | 0.181 | 00:47 |
| Breath Volume | 1.726 | 00:47 |
| Air Blank | 0.000 | 00:48 |
| Air Blank | 0.000 | 00:50 |
| Subject Sample # 2 | 0.178 | 00:50 |
| Breath Volume | 2.234 | 00:50 |
| Air Blank | 0.000 | 00:51 |
| Control Test | 0.079 | 00:51 |
| Air Blank | 0.000 | 00:52 |
| Diagnostics Check | OK | 00:52 |

Data Download Date/Time: 07/12/2021 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: QUAGLIERI

DATE/TIME OF TEST: 06/21/2021 00:05:05

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:58 |
| Air Blank | 0.000 | 23:59 |
| Control Test | 0.081 | 23:59 |
| Air Blank | 0.000 | 00:00 |
| Subject Sample # 1 | 0.000 | 00:00 |
| Breath Volume | 1.210 | 00:00 |
| Air Blank | 0.000 | 00:01 |
| Air Blank | 0.000 | 00:03 |
| Subject Sample # 2 | 0.000 | 00:03 |
| Breath Volume | 1.851 | 00:03 |
| Air Blank | 0.000 | 00:04 |
| Control Test | 0.080 | 00:04 |
| Air Blank | 0.000 | 00:04 |
| Diagnostics Check | OK | 00:05 |

Data Download Date/Time: 07/12/2021 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,J,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: FUSSELL

DATE/TIME OF TEST: 06/07/2021 01:18:26

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:11 |
| Air Blank | 0.000 | 01:11 |
| Control Test | 0.079 | 01:11 |
| Air Blank | 0.000 | 01:12 |
| Subject Sample # 1 | 0.164 | 01:13 |
| Breath Volume | 1.675 | 01:13 |
| Air Blank | 0.000 | 01:14 |
| Air Blank | 0.000 | 01:15 |
| Subject Sample # 2 | 0.168 | 01:16 |
| Breath Volume | 2.429 | 01:16 |
| Air Blank | 0.000 | 01:17 |
| Control Test | 0.078 | 01:17 |
| Air Blank | 0.000 | 01:18 |
| Diagnostics Check | OK | 01:18 |

Data Download Date/Time: 06/26/2021 23:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAUMGARTNER,NATHAN
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: COLON ROCHET

DATE/TIME OF TEST: 06/13/2021 01:11:42

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:01 |
| Air Blank | 0.000 | 01:02 |
| Control Test | 0.080 | 01:02 |
| Air Blank | 0.000 | 01:02 |
| Subject Sample # 1 | 0.142 | 01:04 |
| Breath Volume | 1.906 | 01:04 |
| Air Blank | 0.000 | 01:05 |
| Air Blank | 0.000 | 01:07 |
| Subject Sample # 2 | 0.135 | 01:09 |
| Breath Volume | 1.195 | 01:09 |
| Air Blank | 0.000 | 01:10 |
| Control Test | 0.079 | 01:11 |
| Air Blank | 0.000 | 01:11 |
| Diagnostics Check | OK | 01:11 |

Data Download Date/Time: 06/26/2021 23:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,J,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MARTINS

DATE/TIME OF TEST: 06/15/2021 03:04:25

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 02:55 | |
| Air Blank | 0.000 | 02:56 | |
| Control Test | 0.079 | 02:56 | |
| Air Blank | 0.000 | 02:57 | |
| Subject Sample | # 1 | 03:00 | No Sample Provided |
| Breath Volume | 0.000 | 03:00 | |
| Air Blank | 0.000 | 03:00 | |
| Air Blank | 0.000 | 03:02 | |
| Subject Sample | # 2 | 03:03 | Subject Test Refused |
| Breath Volume | 0.000 | 03:03 | |
| Air Blank | 0.000 | 03:03 | |
| Control Test | 0.079 | 03:03 | |
| Air Blank | 0.000 | 03:04 | |
| Diagnostics Check | OK | 03:04 | |

Data Download Date/Time: 06/26/2021 23:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAERTZ,MARK,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/17/2021 01:10:25

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:03 |
| Air Blank | 0.000 | 01:03 |
| Control Test | 0.080 | 01:04 |
| Air Blank | 0.000 | 01:04 |
| Subject Sample # 1 | 0.195 | 01:05 |
| Breath Volume | 2.105 | 01:05 |
| Air Blank | 0.000 | 01:06 |
| Air Blank | 0.000 | 01:07 |
| Subject Sample # 2 | 0.193 | 01:08 |
| Breath Volume | 2.242 | 01:08 |
| Air Blank | 0.000 | 01:09 |
| Control Test | 0.078 | 01:09 |
| Air Blank | 0.000 | 01:10 |
| Diagnostics Check | OK | 01:10 |

Data Download Date/Time: 06/26/2021 23:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDER,S,R,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: LABRAYERE

DATE/TIME OF TEST: 06/29/2021 22:17:12

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:09 |
| Air Blank | 0.000 | 22:10 |
| Control Test | 0.079 | 22:10 |
| Air Blank | 0.000 | 22:10 |
| Subject Sample # 1 | 0.186 | 22:12 |
| Breath Volume | 1.746 | 22:12 |
| Air Blank | 0.000 | 22:12 |
| Air Blank | 0.000 | 22:14 |
| Subject Sample # 2 | 0.180 | 22:15 |
| Breath Volume | 1.660 | 22:15 |
| Air Blank | 0.000 | 22:16 |
| Control Test | 0.077 | 22:16 |
| Air Blank | 0.000 | 22:17 |
| Diagnostics Check | OK | 22:17 |

Data Download Date/Time: 07/27/2021 09:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,J,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: VATELLE

DATE/TIME OF TEST: 06/30/2021 01:07:19

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 00:59 |
| Air Blank | | 0.000 | 00:59 |
| Control Test | | 0.079 | 01:00 |
| Air Blank | | 0.000 | 01:00 |
| Subject Sample | # 1 | 0.145 | 01:02 |
| Breath Volume | | 2.984 | 01:02 |
| Air Blank | | 0.000 | 01:03 |
| Air Blank | | 0.000 | 01:05 |
| Subject Sample | # 2 | 0.137 | 01:05 |
| Breath Volume | | 1.421 | 01:05 |
| Air Blank | | 0.000 | 01:06 |
| Control Test | | 0.079 | 01:06 |
| Air Blank | | 0.000 | 01:07 |
| Diagnostics Check | | OK | 01:07 |

Data Download Date/Time: 07/27/2021 09:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KOBITKA,NICHOLAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: VILLATORO

DATE/TIME OF TEST: 06/12/2021 03:59:19

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:52 |
| Air Blank | 0.000 | 03:53 |
| Control Test | 0.080 | 03:53 |
| Air Blank | 0.000 | 03:53 |
| Subject Sample # 1 | 0.132 | 03:54 |
| Breath Volume | 1.699 | 03:54 |
| Air Blank | 0.000 | 03:55 |
| Air Blank | 0.000 | 03:57 |
| Subject Sample # 2 | 0.144 | 03:57 |
| Breath Volume | 2.253 | 03:57 |
| Air Blank | 0.000 | 03:58 |
| Control Test | 0.079 | 03:58 |
| Air Blank | 0.000 | 03:59 |
| Diagnostics Check | OK | 03:59 |

Data Download Date/Time: 07/16/2021 14:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DANDREA, ADAM
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:58

SUBJECT'S LAST NAME: SANCHEZ JR

DATE/TIME OF TEST: 06/12/2021 08:30:39

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 08:24 |
| Air Blank | 0.000 | 08:24 |
| Control Test | 0.080 | 08:24 |
| Air Blank | 0.000 | 08:25 |
| Subject Sample # 1 | 0.122 | 08:26 |
| Breath Volume | 3.472 | 08:26 |
| Air Blank | 0.000 | 08:26 |
| Air Blank | 0.000 | 08:28 |
| Subject Sample # 2 | 0.119 | 08:29 |
| Breath Volume | 4.328 | 08:29 |
| Air Blank | 0.000 | 08:29 |
| Control Test | 0.079 | 08:30 |
| Air Blank | 0.000 | 08:30 |
| Diagnostics Check | OK | 08:30 |

Data Download Date/Time: 07/16/2021 14:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: RUTHERFORD

DATE/TIME OF TEST: 06/12/2021 20:53:46

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:45 |
| Air Blank | 0.000 | 20:46 |
| Control Test | 0.080 | 20:46 |
| Air Blank | 0.000 | 20:46 |
| Subject Sample # 1 | 0.195 | 20:47 |
| Breath Volume | 2.734 | 20:47 |
| Air Blank | 0.000 | 20:48 |
| Air Blank | 0.000 | 20:50 |
| Subject Sample # 2 | 0.179 | 20:52 |
| Breath Volume | 2.457 | 20:52 |
| Air Blank | 0.000 | 20:52 |
| Control Test | 0.079 | 20:53 |
| Air Blank | 0.000 | 20:53 |
| Diagnostics Check | OK | 20:53 |

Data Download Date/Time: 07/16/2021 14:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KYSOR,M
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: GASKINS

DATE/TIME OF TEST: 06/12/2021 22:14:04

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:07 |
| Air Blank | 0.000 | 22:08 |
| Control Test | 0.079 | 22:08 |
| Air Blank | 0.000 | 22:08 |
| Subject Sample # 1 | 0.187 | 22:09 |
| Breath Volume | 3.343 | 22:09 |
| Air Blank | 0.000 | 22:10 |
| Air Blank | 0.000 | 22:11 |
| Subject Sample # 2 | 0.189 | 22:12 |
| Breath Volume | 3.007 | 22:12 |
| Air Blank | 0.000 | 22:13 |
| Control Test | 0.079 | 22:13 |
| Air Blank | 0.000 | 22:13 |
| Diagnostics Check | OK | 22:14 |

Data Download Date/Time: 07/16/2021 14:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TEJEDA,ANTHONY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: JUAN ANDRES

DATE/TIME OF TEST: 06/13/2021 00:40:09

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 00:33 |
| Air Blank | | 0.000 | 00:33 |
| Control Test | | 0.079 | 00:34 |
| Air Blank | | 0.000 | 00:34 |
| Subject Sample | # 1 | 0.171 | 00:35 |
| Breath Volume | | 2.230 | 00:35 |
| Air Blank | | 0.000 | 00:36 |
| Air Blank | | 0.000 | 00:38 |
| Subject Sample | # 2 | 0.165 | 00:38 |
| Breath Volume | | 1.726 | 00:38 |
| Air Blank | | 0.000 | 00:39 |
| Control Test | | 0.077 | 00:39 |
| Air Blank | | 0.000 | 00:40 |
| Diagnostics Check | | OK | 00:40 |

Data Download Date/Time: 07/16/2021 14:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: LOPEZ MATOM

DATE/TIME OF TEST: 06/13/2021 01:04:39

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 00:58 |
| Air Blank | | 0.000 | 00:58 |
| Control Test | | 0.078 | 00:59 |
| Air Blank | | 0.000 | 00:59 |
| Subject Sample | # 1 | 0.170 | 01:00 |
| Breath Volume | | 2.378 | 01:00 |
| Air Blank | | 0.000 | 01:00 |
| Air Blank | | 0.000 | 01:02 |
| Subject Sample | # 2 | 0.170 | 01:03 |
| Breath Volume | | 2.175 | 01:03 |
| Air Blank | | 0.000 | 01:03 |
| Control Test | | 0.079 | 01:04 |
| Air Blank | | 0.000 | 01:04 |
| Diagnostics Check | | OK | 01:04 |

Data Download Date/Time: 07/16/2021 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: MARKS

DATE/TIME OF TEST: 06/13/2021 01:38:39

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 01:32 |
| Air Blank | | 0.000 | 01:32 |
| Control Test | | 0.078 | 01:33 |
| Air Blank | | 0.000 | 01:33 |
| Subject Sample | # 1 | 0.000 | 01:34 |
| Breath Volume | | 3.000 | 01:34 |
| Air Blank | | 0.000 | 01:34 |
| Air Blank | | 0.000 | 01:36 |
| Subject Sample | # 2 | 0.000 | 01:37 |
| Breath Volume | | 3.417 | 01:37 |
| Air Blank | | 0.000 | 01:37 |
| Control Test | | 0.079 | 01:38 |
| Air Blank | | 0.000 | 01:38 |
| Diagnostics Check | | OK | 01:38 |

Data Download Date/Time: 07/16/2021 14:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: DORSEY

DATE/TIME OF TEST: 06/13/2021 04:47:50

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 04:40 |
| Air Blank | 0.000 | 04:40 |
| Control Test | 0.079 | 04:41 |
| Air Blank | 0.000 | 04:41 |
| Subject Sample # 1 | 0.187 | 04:42 |
| Breath Volume | 1.105 | 04:42 |
| Air Blank | 0.000 | 04:43 |
| Air Blank | 0.000 | 04:44 |
| Subject Sample # 2 | 0.182 | 04:46 |
| Breath Volume | 1.371 | 04:46 |
| Air Blank | 0.000 | 04:46 |
| Control Test | 0.079 | 04:47 |
| Air Blank | 0.000 | 04:47 |
| Diagnostics Check | OK | 04:47 |

Data Download Date/Time: 07/16/2021 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: CIFELI

DATE/TIME OF TEST: 06/13/2021 18:33:03

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 18:23 |
| Air Blank | 0.000 | 18:23 |
| Control Test | 0.080 | 18:24 |
| Air Blank | 0.000 | 18:24 |
| Subject Sample # 1 | 0.147 | 18:25 |
| Breath Volume | 1.390 | 18:25 |
| Air Blank | 0.000 | 18:26 |
| Air Blank | 0.000 | 18:28 |
| Subject Sample # 2 | 0.151 | 18:31 |
| Breath Volume | 2.125 | 18:31 |
| Air Blank | 0.000 | 18:32 |
| Control Test | 0.079 | 18:32 |
| Air Blank | 0.000 | 18:32 |
| Diagnostics Check | OK | 18:33 |

Data Download Date/Time: 07/16/2021 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: NAST

DATE/TIME OF TEST: 06/13/2021 22:01:02

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:54 |
| Air Blank | 0.000 | 21:54 |
| Control Test | 0.079 | 21:55 |
| Air Blank | 0.000 | 21:55 |
| Subject Sample # 1 | 0.176 | 21:56 |
| Breath Volume | 4.023 | 21:56 |
| Air Blank | 0.000 | 21:57 |
| Air Blank | 0.000 | 21:58 |
| Subject Sample # 2 | 0.156 | 21:59 |
| Breath Volume | 2.121 | 21:59 |
| Air Blank | 0.000 | 22:00 |
| Control Test | 0.078 | 22:00 |
| Air Blank | 0.000 | 22:00 |
| Diagnostics Check | OK | 22:01 |

Data Download Date/Time: 07/16/2021 14:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/14/2021 02:57:14

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:50 |
| Air Blank | 0.000 | 02:51 |
| Control Test | 0.079 | 02:51 |
| Air Blank | 0.000 | 02:52 |
| Subject Sample # 1 | 0.000 | 02:52 |
| Breath Volume | 2.128 | 02:52 |
| Air Blank | 0.000 | 02:53 |
| Air Blank | 0.000 | 02:55 |
| Subject Sample # 2 | 0.000 | 02:55 |
| Breath Volume | 2.050 | 02:55 |
| Air Blank | 0.000 | 02:56 |
| Control Test | 0.079 | 02:56 |
| Air Blank | 0.000 | 02:57 |
| Diagnostics Check | OK | 02:57 |

Data Download Date/Time: 07/16/2021 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHUMER,JARED
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:21

SUBJECT'S LAST NAME: BAHAR

DATE/TIME OF TEST: 06/14/2021 07:02:13

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 06:55 |
| Air Blank | | 0.000 | 06:55 |
| Control Test | | 0.079 | 06:56 |
| Air Blank | | 0.000 | 06:56 |
| Subject Sample | # 1 | 0.201 | 06:57 |
| Breath Volume | | 1.617 | 06:57 |
| Air Blank | | 0.000 | 06:58 |
| Air Blank | | 0.000 | 06:59 |
| Subject Sample | # 2 | 0.208 | 07:00 |
| Breath Volume | | 2.472 | 07:00 |
| Air Blank | | 0.000 | 07:01 |
| Control Test | | 0.079 | 07:01 |
| Air Blank | | 0.000 | 07:02 |
| Diagnostics Check | | OK | 07:02 |

Data Download Date/Time: 07/16/2021 14:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NALERIO,LUIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: SENEXANT

DATE/TIME OF TEST: 06/15/2021 01:02:38

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:55 |
| Air Blank | 0.000 | 00:56 |
| Control Test | 0.080 | 00:56 |
| Air Blank | 0.000 | 00:57 |
| Subject Sample # 1 | 0.274 | 00:57 |
| Breath Volume | 1.281 | 00:57 |
| Air Blank | 0.000 | 00:58 |
| Air Blank | 0.000 | 01:00 |
| Subject Sample # 2 | 0.266 | 01:01 |
| Breath Volume | 1.769 | 01:01 |
| Air Blank | 0.000 | 01:01 |
| Control Test | 0.078 | 01:02 |
| Air Blank | 0.000 | 01:02 |
| Diagnostics Check | OK | 01:02 |

Data Download Date/Time: 07/16/2021 14:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: CONALES EUSED S

DATE/TIME OF TEST: 06/15/2021 19:02:41

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 18:53 |
| Air Blank | 0.000 | 18:54 |
| Control Test | 0.078 | 18:54 |
| Air Blank | 0.000 | 18:55 |
| Subject Sample # 1 | 0.043 | 18:57 |
| Breath Volume | 1.546 | 18:57 |
| Air Blank | 0.000 | 18:58 |
| Air Blank | 0.000 | 19:00 |
| Subject Sample # 2 | 0.047 | 19:01 |
| Breath Volume | 2.621 | 19:01 |
| Air Blank | 0.000 | 19:01 |
| Control Test | 0.079 | 19:02 |
| Air Blank | 0.000 | 19:02 |
| Diagnostics Check | OK | 19:02 |

Data Download Date/Time: 07/16/2021 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NALERIO,LUIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: DELOA DAVIS

DATE/TIME OF TEST: 06/15/2021 21:58:55

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 21:47 | |
| Air Blank | | 0.000 | 21:48 | |
| Control Test | | 0.079 | 21:48 | |
| Air Blank | | 0.000 | 21:49 | |
| Subject Sample | # 1 | 0.210 | 21:52 | Volume Not Met |
| Breath Volume | | 1.058 | 21:52 | |
| Air Blank | | 0.000 | 21:53 | |
| Air Blank | | 0.000 | 21:54 | |
| Subject Sample | # 2 | | 21:57 | Slope Not Met |
| Breath Volume | | 0.746 | 21:57 | |
| Air Blank | | 0.000 | 21:57 | |
| Control Test | | 0.077 | 21:58 | |
| Air Blank | | 0.000 | 21:58 | |
| Diagnostics Check | | OK | 21:58 | |

Data Download Date/Time: 07/16/2021 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FORDELLA,BLAKE
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: FAIRFAX-TURNER

DATE/TIME OF TEST: 06/16/2021 00:15:07

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:08 |
| Air Blank | 0.000 | 00:08 |
| Control Test | 0.080 | 00:09 |
| Air Blank | 0.000 | 00:09 |
| Subject Sample # 1 | 0.198 | 00:10 |
| Breath Volume | 2.132 | 00:10 |
| Air Blank | 0.000 | 00:11 |
| Air Blank | 0.000 | 00:13 |
| Subject Sample # 2 | 0.183 | 00:13 |
| Breath Volume | 1.265 | 00:13 |
| Air Blank | 0.000 | 00:14 |
| Control Test | 0.078 | 00:14 |
| Air Blank | 0.000 | 00:15 |
| Diagnostics Check | OK | 00:15 |

Data Download Date/Time: 07/16/2021 14:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MARTINEZ,JACOB
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: MATTHEW

DATE/TIME OF TEST: 06/17/2021 04:04:36

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:57 |
| Air Blank | 0.000 | 03:58 |
| Control Test | 0.080 | 03:58 |
| Air Blank | 0.000 | 03:59 |
| Subject Sample # 1 | 0.190 | 04:00 |
| Breath Volume | 1.234 | 04:00 |
| Air Blank | 0.000 | 04:00 |
| Air Blank | 0.000 | 04:02 |
| Subject Sample # 2 | 0.194 | 04:03 |
| Breath Volume | 1.441 | 04:03 |
| Air Blank | 0.000 | 04:03 |
| Control Test | 0.078 | 04:03 |
| Air Blank | 0.000 | 04:04 |
| Diagnostics Check | OK | 04:04 |

Data Download Date/Time: 07/16/2021 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER,CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 06/18/2021 04:20:54

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 04:14 |
| Air Blank | 0.000 | 04:14 |
| Control Test | 0.080 | 04:15 |
| Air Blank | 0.000 | 04:15 |
| Subject Sample # 1 | 0.074 | 04:16 |
| Breath Volume | 2.558 | 04:16 |
| Air Blank | 0.000 | 04:17 |
| Air Blank | 0.000 | 04:18 |
| Subject Sample # 2 | 0.074 | 04:19 |
| Breath Volume | 2.820 | 04:19 |
| Air Blank | 0.000 | 04:19 |
| Control Test | 0.080 | 04:20 |
| Air Blank | 0.000 | 04:20 |
| Diagnostics Check | OK | 04:20 |

Data Download Date/Time: 07/16/2021 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR,NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: DESTY

DATE/TIME OF TEST: 06/19/2021 00:49:55

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 00:41 |
| Air Blank | | 0.000 | 00:42 |
| Control Test | | 0.080 | 00:42 |
| Air Blank | | 0.000 | 00:43 |
| Subject Sample | # 1 | 0.205 | 00:44 |
| Breath Volume | | 1.324 | 00:44 |
| Air Blank | | 0.000 | 00:45 |
| Air Blank | | 0.000 | 00:46 |
| Subject Sample | # 2 | 0.190 | 00:48 |
| Breath Volume | | 1.601 | 00:48 |
| Air Blank | | 0.000 | 00:48 |
| Control Test | | 0.078 | 00:49 |
| Air Blank | | 0.000 | 00:49 |
| Diagnostics Check | | OK | 00:49 |

Data Download Date/Time: 07/16/2021 14:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: PEDRO MATEO

DATE/TIME OF TEST: 06/19/2021 02:02:17

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 01:50 | |
| Air Blank | | 0.000 | 01:50 | |
| Control Test | | 0.079 | 01:50 | |
| Air Blank | | 0.000 | 01:51 | |
| Subject Sample | # 1 | 0.113 | 01:54 | Volume Not Met |
| Breath Volume | | 0.035 | 01:54 | |
| Air Blank | | 0.000 | 01:55 | |
| Air Blank | | 0.000 | 01:56 | |
| Subject Sample | # 2 | 0.160 | 01:57 | |
| Breath Volume | | 2.875 | 01:57 | |
| Air Blank | | 0.000 | 01:58 | |
| Air Blank | | 0.000 | 01:59 | |
| Subject Sample | # 3 | 0.160 | 02:00 | |
| Breath Volume | | 3.000 | 02:00 | |
| Air Blank | | 0.000 | 02:01 | |
| Control Test | | 0.077 | 02:01 | |
| Air Blank | | 0.000 | 02:02 | |
| Diagnostics Check | | OK | 02:02 | |

Data Download Date/Time: 07/16/2021 14:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MARGOTTA

DATE/TIME OF TEST: 06/19/2021 03:12:24

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:06 |
| Air Blank | 0.000 | 03:06 |
| Control Test | 0.079 | 03:07 |
| Air Blank | 0.000 | 03:07 |
| Subject Sample # 1 | 0.067 | 03:08 |
| Breath Volume | 2.167 | 03:08 |
| Air Blank | 0.000 | 03:08 |
| Air Blank | 0.000 | 03:10 |
| Subject Sample # 2 | 0.068 | 03:10 |
| Breath Volume | 2.109 | 03:10 |
| Air Blank | 0.000 | 03:11 |
| Control Test | 0.080 | 03:11 |
| Air Blank | 0.000 | 03:12 |
| Diagnostics Check | OK | 03:12 |

Data Download Date/Time: 07/16/2021 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TEJEDA, ANTHONY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: DORCIN

DATE/TIME OF TEST: 06/19/2021 04:17:29

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 04:11 |
| Air Blank | 0.000 | 04:11 |
| Control Test | 0.079 | 04:11 |
| Air Blank | 0.000 | 04:12 |
| Subject Sample # 1 | 0.112 | 04:13 |
| Breath Volume | 1.644 | 04:13 |
| Air Blank | 0.000 | 04:13 |
| Air Blank | 0.000 | 04:15 |
| Subject Sample # 2 | 0.113 | 04:15 |
| Breath Volume | 1.746 | 04:15 |
| Air Blank | 0.000 | 04:16 |
| Control Test | 0.079 | 04:16 |
| Air Blank | 0.000 | 04:17 |
| Diagnostics Check | OK | 04:17 |

Data Download Date/Time: 07/16/2021 14:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KYSOR, MITCHELL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/20/2021 04:22:19

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 04:18 | |
| Air Blank | 0.000 | 04:19 | |
| Control Test | 0.080 | 04:19 | |
| Air Blank | 0.000 | 04:19 | |
| Subject Sample | # 1 | 04:20 | Subject Test Refused |
| Breath Volume | 0.000 | 04:20 | |
| Air Blank | 0.000 | 04:21 | |
| Control Test | 0.079 | 04:21 | |
| Air Blank | 0.000 | 04:22 | |
| Diagnostics Check | OK | 04:22 | |

Data Download Date/Time: 07/16/2021 14:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/20/2021 05:12:02

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 05:05 |
| Air Blank | | 0.000 | 05:06 |
| Control Test | | 0.080 | 05:06 |
| Air Blank | | 0.000 | 05:07 |
| Subject Sample | # 1 | 0.157 | 05:07 |
| Breath Volume | | 2.207 | 05:07 |
| Air Blank | | 0.000 | 05:08 |
| Air Blank | | 0.000 | 05:10 |
| Subject Sample | # 2 | 0.157 | 05:10 |
| Breath Volume | | 2.550 | 05:10 |
| Air Blank | | 0.000 | 05:11 |
| Control Test | | 0.079 | 05:11 |
| Air Blank | | 0.000 | 05:11 |
| Diagnostics Check | | OK | 05:12 |

Data Download Date/Time: 07/16/2021 14:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/20/2021 23:02:21

| | | | |
|--------------------|-------|-------|----------------|
| Diagnostics Check | OK | 22:45 | |
| Air Blank | 0.000 | 22:45 | |
| Control Test | 0.080 | 22:45 | |
| Air Blank | 0.000 | 22:46 | |
| Subject Sample # 1 | 0.121 | 22:49 | |
| Breath Volume | 1.378 | 22:49 | |
| Air Blank | 0.000 | 22:50 | |
| Air Blank | 0.000 | 22:51 | |
| Subject Sample # 2 | 0.099 | 22:55 | Volume Not Met |
| Breath Volume | 0.550 | 22:55 | |
| Air Blank | 0.000 | 22:55 | |
| Air Blank | 0.000 | 22:57 | |
| Subject Sample # 3 | 0.110 | 23:00 | Volume Not Met |
| Breath Volume | 0.769 | 23:00 | |
| Air Blank | 0.000 | 23:01 | |
| Control Test | 0.079 | 23:01 | |
| Air Blank | 0.000 | 23:02 | |
| Diagnostics Check | OK | 23:02 | |

Data Download Date/Time: 07/16/2021 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LOVETT,JAMES
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: HOGAN

DATE/TIME OF TEST: 06/21/2021 23:16:33

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:09 |
| Air Blank | 0.000 | 23:10 |
| Control Test | 0.080 | 23:10 |
| Air Blank | 0.000 | 23:11 |
| Subject Sample # 1 | 0.168 | 23:11 |
| Breath Volume | 1.164 | 23:11 |
| Air Blank | 0.000 | 23:12 |
| Air Blank | 0.000 | 23:14 |
| Subject Sample # 2 | 0.168 | 23:14 |
| Breath Volume | 1.339 | 23:14 |
| Air Blank | 0.000 | 23:15 |
| Control Test | 0.079 | 23:15 |
| Air Blank | 0.000 | 23:16 |
| Diagnostics Check | OK | 23:16 |

Data Download Date/Time: 07/16/2021 14:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BRANDT,NICKOLAS
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: ADLEN

DATE/TIME OF TEST: 06/22/2021 17:46:12

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 17:39 |
| Air Blank | 0.000 | 17:40 |
| Control Test | 0.079 | 17:40 |
| Air Blank | 0.000 | 17:40 |
| Subject Sample # 1 | 0.267 | 17:41 |
| Breath Volume | 3.742 | 17:41 |
| Air Blank | 0.000 | 17:42 |
| Air Blank | 0.000 | 17:44 |
| Subject Sample # 2 | 0.265 | 17:44 |
| Breath Volume | 3.457 | 17:44 |
| Air Blank | 0.000 | 17:45 |
| Control Test | 0.078 | 17:45 |
| Air Blank | 0.000 | 17:46 |
| Diagnostics Check | OK | 17:46 |

Data Download Date/Time: 07/16/2021 14:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR,NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: REDMAN

DATE/TIME OF TEST: 06/24/2021 03:45:01

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:38 |
| Air Blank | 0.000 | 03:38 |
| Control Test | 0.080 | 03:39 |
| Air Blank | 0.000 | 03:39 |
| Subject Sample # 1 | 0.163 | 03:40 |
| Breath Volume | 2.550 | 03:40 |
| Air Blank | 0.000 | 03:41 |
| Air Blank | 0.000 | 03:42 |
| Subject Sample # 2 | 0.161 | 03:43 |
| Breath Volume | 1.722 | 03:43 |
| Air Blank | 0.000 | 03:44 |
| Control Test | 0.079 | 03:44 |
| Air Blank | 0.000 | 03:44 |
| Diagnostics Check | OK | 03:45 |

Data Download Date/Time: 07/16/2021 14:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HERNANDEZ, EDWIN
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: TIRADO

DATE/TIME OF TEST: 06/24/2021 05:55:09

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 05:48 |
| Air Blank | 0.000 | 05:49 |
| Control Test | 0.080 | 05:49 |
| Air Blank | 0.000 | 05:49 |
| Subject Sample # 1 | 0.078 | 05:50 |
| Breath Volume | 2.140 | 05:50 |
| Air Blank | 0.000 | 05:51 |
| Air Blank | 0.000 | 05:53 |
| Subject Sample # 2 | 0.078 | 05:53 |
| Breath Volume | 2.609 | 05:53 |
| Air Blank | 0.000 | 05:54 |
| Control Test | 0.080 | 05:54 |
| Air Blank | 0.000 | 05:55 |
| Diagnostics Check | OK | 05:55 |

Data Download Date/Time: 07/16/2021 14:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: GRAEFE

DATE/TIME OF TEST: 06/24/2021 23:41:02

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:34 |
| Air Blank | 0.000 | 23:34 |
| Control Test | 0.080 | 23:35 |
| Air Blank | 0.000 | 23:35 |
| Subject Sample # 1 | 0.107 | 23:36 |
| Breath Volume | 2.671 | 23:36 |
| Air Blank | 0.000 | 23:36 |
| Air Blank | 0.000 | 23:38 |
| Subject Sample # 2 | 0.098 | 23:39 |
| Breath Volume | 2.921 | 23:39 |
| Air Blank | 0.000 | 23:40 |
| Control Test | 0.079 | 23:40 |
| Air Blank | 0.000 | 23:40 |
| Diagnostics Check | OK | 23:41 |

Data Download Date/Time: 07/16/2021 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PENAGOS,CARLOS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: CHUN MATUL

DATE/TIME OF TEST: 06/25/2021 03:07:43

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 03:01 |
| Air Blank | | 0.000 | 03:01 |
| Control Test | | 0.080 | 03:01 |
| Air Blank | | 0.000 | 03:02 |
| Subject Sample | # 1 | 0.295 | 03:03 |
| Breath Volume | | 2.671 | 03:03 |
| Air Blank | | 0.000 | 03:03 |
| Air Blank | | 0.000 | 03:05 |
| Subject Sample | # 2 | 0.299 | 03:06 |
| Breath Volume | | 2.125 | 03:06 |
| Air Blank | | 0.000 | 03:06 |
| Control Test | | 0.079 | 03:07 |
| Air Blank | | 0.000 | 03:07 |
| Diagnostics Check | | OK | 03:07 |

Data Download Date/Time: 07/16/2021 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MARCH,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: MCFADDEN

DATE/TIME OF TEST: 06/25/2021 06:40:23

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 06:33 |
| Air Blank | 0.000 | 06:34 |
| Control Test | 0.080 | 06:34 |
| Air Blank | 0.000 | 06:35 |
| Subject Sample # 1 | 0.144 | 06:35 |
| Breath Volume | 2.054 | 06:35 |
| Air Blank | 0.000 | 06:36 |
| Air Blank | 0.000 | 06:38 |
| Subject Sample # 2 | 0.143 | 06:38 |
| Breath Volume | 3.093 | 06:38 |
| Air Blank | 0.000 | 06:39 |
| Control Test | 0.079 | 06:39 |
| Air Blank | 0.000 | 06:40 |
| Diagnostics Check | OK | 06:40 |

Data Download Date/Time: 07/16/2021 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SMITH,MARK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: STEVENSON

DATE/TIME OF TEST: 06/25/2021 16:41:48

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 16:35 |
| Air Blank | 0.000 | 16:35 |
| Control Test | 0.080 | 16:36 |
| Air Blank | 0.000 | 16:36 |
| Subject Sample # 1 | 0.000 | 16:37 |
| Breath Volume | 1.511 | 16:37 |
| Air Blank | 0.000 | 16:37 |
| Air Blank | 0.000 | 16:39 |
| Subject Sample # 2 | 0.000 | 16:40 |
| Breath Volume | 1.820 | 16:40 |
| Air Blank | 0.000 | 16:40 |
| Control Test | 0.080 | 16:41 |
| Air Blank | 0.000 | 16:41 |
| Diagnostics Check | OK | 16:41 |

Data Download Date/Time: 07/16/2021 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FERGUSON,RYAN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: ONEAL

DATE/TIME OF TEST: 06/25/2021 22:45:26

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:39 |
| Air Blank | 0.000 | 22:39 |
| Control Test | 0.080 | 22:39 |
| Air Blank | 0.000 | 22:40 |
| Subject Sample # 1 | 0.000 | 22:41 |
| Breath Volume | 1.308 | 22:41 |
| Air Blank | 0.000 | 22:41 |
| Air Blank | 0.000 | 22:43 |
| Subject Sample # 2 | 0.000 | 22:44 |
| Breath Volume | 1.632 | 22:44 |
| Air Blank | 0.000 | 22:44 |
| Control Test | 0.080 | 22:44 |
| Air Blank | 0.000 | 22:45 |
| Diagnostics Check | OK | 22:45 |

Data Download Date/Time: 07/16/2021 14:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: NAVARRETE GARCIA

DATE/TIME OF TEST: 06/26/2021 03:44:06

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:36 |
| Air Blank | 0.000 | 03:36 |
| Control Test | 0.080 | 03:36 |
| Air Blank | 0.000 | 03:37 |
| Subject Sample # 1 | 0.167 | 03:39 |
| Breath Volume | 1.750 | 03:39 |
| Air Blank | 0.000 | 03:39 |
| Air Blank | 0.000 | 03:41 |
| Subject Sample # 2 | 0.158 | 03:42 |
| Breath Volume | 1.785 | 03:42 |
| Air Blank | 0.000 | 03:43 |
| Control Test | 0.079 | 03:43 |
| Air Blank | 0.000 | 03:44 |
| Diagnostics Check | OK | 03:44 |

Data Download Date/Time: 07/16/2021 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 06/26/2021 20:37:54

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:30 |
| Air Blank | 0.000 | 20:31 |
| Control Test | 0.080 | 20:31 |
| Air Blank | 0.000 | 20:32 |
| Subject Sample # 1 | 0.113 | 20:33 |
| Breath Volume | 4.453 | 20:33 |
| Air Blank | 0.000 | 20:34 |
| Air Blank | 0.000 | 20:35 |
| Subject Sample # 2 | 0.111 | 20:36 |
| Breath Volume | 4.382 | 20:36 |
| Air Blank | 0.000 | 20:36 |
| Control Test | 0.079 | 20:37 |
| Air Blank | 0.000 | 20:37 |
| Diagnostics Check | OK | 20:37 |

Data Download Date/Time: 07/16/2021 14:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 06/26/2021 21:57:54

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:50 |
| Air Blank | 0.000 | 21:51 |
| Control Test | 0.079 | 21:51 |
| Air Blank | 0.000 | 21:52 |
| Subject Sample # 1 | 0.243 | 21:53 |
| Breath Volume | 1.125 | 21:53 |
| Air Blank | 0.000 | 21:53 |
| Air Blank | 0.000 | 21:55 |
| Subject Sample # 2 | 0.242 | 21:56 |
| Breath Volume | 2.730 | 21:56 |
| Air Blank | 0.000 | 21:56 |
| Control Test | 0.079 | 21:57 |
| Air Blank | 0.000 | 21:57 |
| Diagnostics Check | OK | 21:57 |

Data Download Date/Time: 07/16/2021 14:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PAGE, KAYLEE
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: PALESE

DATE/TIME OF TEST: 06/26/2021 22:27:15

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:18 |
| Air Blank | 0.000 | 22:18 |
| Control Test | 0.078 | 22:18 |
| Air Blank | 0.000 | 22:19 |
| Subject Sample # 1 | 0.152 | 22:19 |
| Breath Volume | 1.707 | 22:19 |
| Air Blank | 0.000 | 22:20 |
| Air Blank | 0.000 | 22:22 |
| Subject Sample # 2 | 0.178 | 22:22 |
| Breath Volume | 1.835 | 22:22 |
| Air Blank | 0.000 | 22:23 |
| Air Blank | 0.000 | 22:25 |
| Subject Sample # 3 | 0.176 | 22:25 |
| Breath Volume | 1.929 | 22:25 |
| Air Blank | 0.000 | 22:26 |
| Control Test | 0.078 | 22:26 |
| Air Blank | 0.000 | 22:27 |
| Diagnostics Check | OK | 22:27 |

Data Download Date/Time: 07/16/2021 14:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: METCALF

DATE/TIME OF TEST: 06/27/2021 00:02:09

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:55 |
| Air Blank | 0.000 | 23:56 |
| Control Test | 0.080 | 23:56 |
| Air Blank | 0.000 | 23:56 |
| Subject Sample # 2 | 0.173 | 23:57 |
| Breath Volume | 2.546 | 23:57 |
| Air Blank | 0.000 | 23:58 |
| Air Blank | 0.000 | 00:00 |
| Subject Sample # 1 | 0.176 | 00:00 |
| Breath Volume | 3.687 | 00:00 |
| Air Blank | 0.000 | 00:01 |
| Control Test | 0.079 | 00:01 |
| Air Blank | 0.000 | 00:02 |
| Diagnostics Check | OK | 00:02 |

Data Download Date/Time: 07/16/2021 14:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TEJEDA,ANTHONY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MARBAN-CHAVEZ

DATE/TIME OF TEST: 06/27/2021 01:45:41

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:39 |
| Air Blank | 0.000 | 01:39 |
| Control Test | 0.079 | 01:39 |
| Air Blank | 0.000 | 01:40 |
| Subject Sample # 1 | 0.158 | 01:41 |
| Breath Volume | 3.234 | 01:41 |
| Air Blank | 0.000 | 01:41 |
| Air Blank | 0.000 | 01:43 |
| Subject Sample # 2 | 0.160 | 01:44 |
| Breath Volume | 3.238 | 01:44 |
| Air Blank | 0.000 | 01:44 |
| Control Test | 0.080 | 01:45 |
| Air Blank | 0.000 | 01:45 |
| Diagnostics Check | OK | 01:45 |

Data Download Date/Time: 07/16/2021 14:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: MACET

DATE/TIME OF TEST: 06/27/2021 03:29:44

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:23 |
| Air Blank | 0.000 | 03:23 |
| Control Test | 0.079 | 03:24 |
| Air Blank | 0.000 | 03:24 |
| Subject Sample # 1 | 0.166 | 03:25 |
| Breath Volume | 1.808 | 03:25 |
| Air Blank | 0.000 | 03:25 |
| Air Blank | 0.000 | 03:27 |
| Subject Sample # 2 | 0.168 | 03:28 |
| Breath Volume | 2.351 | 03:28 |
| Air Blank | 0.000 | 03:28 |
| Control Test | 0.078 | 03:29 |
| Air Blank | 0.000 | 03:29 |
| Diagnostics Check | OK | 03:29 |

Data Download Date/Time: 07/16/2021 14:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SKOTZKE, HANNAH
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: LOISEAU

DATE/TIME OF TEST: 06/27/2021 06:02:53

| | | | | |
|-------------------|-----|-------|-------|----------------------|
| Diagnostics Check | | OK | 05:52 | |
| Air Blank | | 0.000 | 05:52 | |
| Control Test | | 0.079 | 05:52 | |
| Air Blank | | 0.000 | 05:53 | |
| Subject Sample | # 1 | 0.146 | 05:56 | Volume Not Met |
| Breath Volume | | 0.250 | 05:56 | |
| Air Blank | | 0.000 | 05:57 | |
| Air Blank | | 0.000 | 05:58 | |
| Subject Sample | # 2 | | 06:01 | Subject Test Refused |
| Breath Volume | | 1.039 | 06:01 | |
| Air Blank | | 0.000 | 06:01 | |
| Control Test | | 0.078 | 06:02 | |
| Air Blank | | 0.000 | 06:02 | |
| Diagnostics Check | | OK | 06:02 | |

Data Download Date/Time: 07/16/2021 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TEJEDA,ANTHONY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: LANZA FIGUEROA

DATE/TIME OF TEST: 06/27/2021 22:23:00

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:16 |
| Air Blank | 0.000 | 22:16 |
| Control Test | 0.080 | 22:17 |
| Air Blank | 0.000 | 22:17 |
| Subject Sample # 1 | 0.218 | 22:18 |
| Breath Volume | 2.589 | 22:18 |
| Air Blank | 0.000 | 22:19 |
| Air Blank | 0.000 | 22:20 |
| Subject Sample # 2 | 0.216 | 22:21 |
| Breath Volume | 2.835 | 22:21 |
| Air Blank | 0.000 | 22:22 |
| Control Test | 0.081 | 22:22 |
| Air Blank | 0.000 | 22:22 |
| Diagnostics Check | OK | 22:23 |

Data Download Date/Time: 07/16/2021 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 06/27/2021 23:37:20

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 23:24 | |
| Air Blank | | 0.000 | 23:25 | |
| Control Test | | 0.079 | 23:25 | |
| Air Blank | | 0.000 | 23:26 | |
| Subject Sample | # 1 | 0.171 | 23:29 | Volume Not Met |
| Breath Volume | | 0.363 | 23:29 | |
| Air Blank | | 0.000 | 23:30 | |
| Air Blank | | 0.000 | 23:31 | |
| Subject Sample | # 2 | 0.190 | 23:32 | |
| Breath Volume | | 2.812 | 23:32 | |
| Air Blank | | 0.000 | 23:33 | |
| Air Blank | | 0.000 | 23:35 | |
| Subject Sample | # 3 | 0.201 | 23:35 | |
| Breath Volume | | 3.140 | 23:35 | |
| Air Blank | | 0.000 | 23:36 | |
| Control Test | | 0.078 | 23:36 | |
| Air Blank | | 0.000 | 23:37 | |
| Diagnostics Check | | OK | 23:37 | |

Data Download Date/Time: 07/16/2021 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: ALONSO

DATE/TIME OF TEST: 06/28/2021 00:58:09

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:52 |
| Air Blank | 0.000 | 00:52 |
| Control Test | 0.079 | 00:52 |
| Air Blank | 0.000 | 00:53 |
| Subject Sample # 1 | 0.231 | 00:53 |
| Breath Volume | 1.867 | 00:53 |
| Air Blank | 0.000 | 00:54 |
| Air Blank | 0.000 | 00:56 |
| Subject Sample # 2 | 0.226 | 00:56 |
| Breath Volume | 2.324 | 00:56 |
| Air Blank | 0.000 | 00:57 |
| Control Test | 0.078 | 00:57 |
| Air Blank | 0.000 | 00:58 |
| Diagnostics Check | OK | 00:58 |

Data Download Date/Time: 07/16/2021 14:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TEJEDA,ANTHONY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: FALCON VASQUEZ

DATE/TIME OF TEST: 06/28/2021 02:53:54

| | | | |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK | 02:45 | |
| Air Blank | 0.000 | 02:45 | |
| Control Test | 0.080 | 02:46 | |
| Air Blank | 0.000 | 02:46 | |
| Subject Sample | # 1 0.147 | 02:49 | Volume Not Met |
| Breath Volume | 0.351 | 02:49 | |
| Air Blank | 0.000 | 02:50 | |
| Air Blank | 0.000 | 02:52 | |
| Subject Sample | # 2 | 02:52 | Subject Test Refused |
| Breath Volume | 0.000 | 02:52 | |
| Air Blank | 0.000 | 02:52 | |
| Control Test | 0.077 | 02:53 | |
| Air Blank | 0.000 | 02:53 | |
| Diagnostics Check | OK | 02:53 | |

Data Download Date/Time: 07/16/2021 14:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JONES,KENNETH
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:25

SUBJECT'S LAST NAME: MONROE

DATE/TIME OF TEST: 06/28/2021 11:08:14

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 11:01 |
| Air Blank | 0.000 | 11:02 |
| Control Test | 0.079 | 11:02 |
| Air Blank | 0.000 | 11:02 |
| Subject Sample # 1 | 0.167 | 11:03 |
| Breath Volume | 1.164 | 11:03 |
| Air Blank | 0.000 | 11:04 |
| Air Blank | 0.000 | 11:06 |
| Subject Sample # 2 | 0.161 | 11:06 |
| Breath Volume | 2.363 | 11:06 |
| Air Blank | 0.000 | 11:07 |
| Control Test | 0.079 | 11:07 |
| Air Blank | 0.000 | 11:08 |
| Diagnostics Check | OK | 11:08 |

Data Download Date/Time: 07/16/2021 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: CHAVEZ ARCON

DATE/TIME OF TEST: 06/29/2021 00:26:13

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:19 |
| Air Blank | 0.000 | 00:19 |
| Control Test | 0.080 | 00:20 |
| Air Blank | 0.000 | 00:20 |
| Subject Sample # 1 | 0.129 | 00:21 |
| Breath Volume | 2.960 | 00:21 |
| Air Blank | 0.000 | 00:22 |
| Air Blank | 0.000 | 00:24 |
| Subject Sample # 2 | 0.120 | 00:24 |
| Breath Volume | 1.648 | 00:24 |
| Air Blank | 0.000 | 00:25 |
| Control Test | 0.080 | 00:25 |
| Air Blank | 0.000 | 00:26 |
| Diagnostics Check | OK | 00:26 |

Data Download Date/Time: 07/16/2021 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: BECKER

DATE/TIME OF TEST: 06/29/2021 04:27:11

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 04:21 |
| Air Blank | 0.000 | 04:21 |
| Control Test | 0.080 | 04:21 |
| Air Blank | 0.000 | 04:22 |
| Subject Sample # 1 | 0.014 | 04:22 |
| Breath Volume | 1.339 | 04:22 |
| Air Blank | 0.000 | 04:23 |
| Air Blank | 0.000 | 04:25 |
| Subject Sample # 2 | 0.015 | 04:25 |
| Breath Volume | 2.457 | 04:25 |
| Air Blank | 0.000 | 04:26 |
| Control Test | 0.080 | 04:26 |
| Air Blank | 0.000 | 04:27 |
| Diagnostics Check | OK | 04:27 |

Data Download Date/Time: 07/16/2021 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALDKOETTER,ELIZABETH
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 09:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: HOFFMAN-DOOLEY

DATE/TIME OF TEST: 06/29/2021 12:55:21

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 12:47 |
| Air Blank | 0.000 | 12:47 |
| Control Test | 0.080 | 12:48 |
| Air Blank | 0.000 | 12:48 |
| Subject Sample # 1 | 0.182 | 12:50 |
| Breath Volume | 1.527 | 12:50 |
| Air Blank | 0.000 | 12:51 |
| Air Blank | 0.000 | 12:52 |
| Subject Sample # 2 | 0.191 | 12:53 |
| Breath Volume | 2.164 | 12:53 |
| Air Blank | 0.000 | 12:54 |
| Control Test | 0.079 | 12:54 |
| Air Blank | 0.000 | 12:55 |
| Diagnostics Check | OK | 12:55 |

Data Download Date/Time: 07/16/2021 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: INFINGER

DATE/TIME OF TEST: 06/29/2021 23:23:06

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:17 |
| Air Blank | 0.000 | 23:17 |
| Control Test | 0.080 | 23:18 |
| Air Blank | 0.000 | 23:18 |
| Subject Sample # 1 | 0.000 | 23:18 |
| Breath Volume | 1.808 | 23:18 |
| Air Blank | 0.000 | 23:19 |
| Air Blank | 0.000 | 23:21 |
| Subject Sample # 2 | 0.000 | 23:21 |
| Breath Volume | 2.078 | 23:21 |
| Air Blank | 0.000 | 23:22 |
| Control Test | 0.080 | 23:22 |
| Air Blank | 0.000 | 23:23 |
| Diagnostics Check | OK | 23:23 |

Data Download Date/Time: 07/16/2021 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCGILLICUDDY,STEVEN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: BENOZICH

DATE/TIME OF TEST: 06/30/2021 21:06:38

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:59 |
| Air Blank | 0.000 | 20:59 |
| Control Test | 0.080 | 21:00 |
| Air Blank | 0.000 | 21:00 |
| Subject Sample # 1 | 0.189 | 21:02 |
| Breath Volume | 2.925 | 21:02 |
| Air Blank | 0.000 | 21:02 |
| Air Blank | 0.000 | 21:04 |
| Subject Sample # 2 | 0.187 | 21:05 |
| Breath Volume | 2.675 | 21:05 |
| Air Blank | 0.000 | 21:05 |
| Control Test | 0.079 | 21:06 |
| Air Blank | 0.000 | 21:06 |
| Diagnostics Check | OK | 21:06 |

Data Download Date/Time: 07/16/2021 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR,NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: JURICH

DATE/TIME OF TEST: 06/30/2021 22:51:55

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:45 |
| Air Blank | 0.000 | 22:45 |
| Control Test | 0.080 | 22:46 |
| Air Blank | 0.000 | 22:46 |
| Subject Sample # 1 | 0.062 | 22:47 |
| Breath Volume | 2.261 | 22:47 |
| Air Blank | 0.000 | 22:47 |
| Air Blank | 0.000 | 22:49 |
| Subject Sample # 2 | 0.063 | 22:50 |
| Breath Volume | 3.000 | 22:50 |
| Air Blank | 0.000 | 22:51 |
| Control Test | 0.080 | 22:51 |
| Air Blank | 0.000 | 22:51 |
| Diagnostics Check | OK | 22:51 |

Data Download Date/Time: 07/16/2021 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY,THOMAS,H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEHENY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 06/30/2021 23:16:57

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:10 |
| Air Blank | 0.000 | 23:11 |
| Control Test | 0.079 | 23:11 |
| Air Blank | 0.000 | 23:11 |
| Subject Sample # 1 | 0.000 | 23:12 |
| Breath Volume | 1.636 | 23:12 |
| Air Blank | 0.000 | 23:12 |
| Air Blank | 0.000 | 23:14 |
| Subject Sample # 2 | 0.000 | 23:15 |
| Breath Volume | 1.468 | 23:15 |
| Air Blank | 0.000 | 23:16 |
| Control Test | 0.080 | 23:16 |
| Air Blank | 0.000 | 23:16 |
| Diagnostics Check | OK | 23:16 |

Data Download Date/Time: 07/16/2021 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

INSTRUMENT REGISTERED TO: FHP TROOP F

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LEMEN,LOGAN,C

BREATH TEST OPERATOR AGENCY: FHP Troop F

ARRESTING OFFICER: LEMEN,LOGAN

ARRESTING OFFICER AGENCY: FHP Troop F

LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1167545

DRY GAS EXPIRY DATE: 07/02/2022

VIOLATION CODE: DUI

ARREST TIME: 01:36

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MEZGER

DATE/TIME OF TEST: 06/26/2021 02:17:50

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 02:10 |
| Air Blank | | 0.000 | 02:11 |
| Control Test | | 0.079 | 02:11 |
| Air Blank | | 0.000 | 02:12 |
| Subject Sample | # 1 | 0.280 | 02:13 |
| Breath Volume | | 1.503 | 02:13 |
| Air Blank | | 0.000 | 02:13 |
| Air Blank | | 0.000 | 02:15 |
| Subject Sample | # 2 | 0.289 | 02:16 |
| Breath Volume | | 1.714 | 02:16 |
| Air Blank | | 0.000 | 02:16 |
| Control Test | | 0.080 | 02:17 |
| Air Blank | | 0.000 | 02:17 |
| Diagnostics Check | | OK | 02:17 |

Data Download Date/Time: 07/12/2021 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

INSTRUMENT REGISTERED TO: FHP TROOP F

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LEMEN,LOGAN,C

BREATH TEST OPERATOR AGENCY: FHP Troop F

ARRESTING OFFICER: LEMEN, LGAN

ARRESTING OFFICER AGENCY: FHP Troop F

LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1167545

DRY GAS EXPIRY DATE: 07/02/2022

VIOLATION CODE: DUI

ARREST TIME: 04:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 06/27/2021 05:22:37

| | | | |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK | 05:14 | |
| Air Blank | 0.000 | 05:14 | |
| Control Test | 0.079 | 05:14 | |
| Air Blank | 0.000 | 05:15 | |
| Subject Sample | # 1 0.317 | 05:16 | |
| Breath Volume | 1.703 | 05:16 | |
| Air Blank | 0.000 | 05:16 | |
| Air Blank | 0.000 | 05:18 | |
| Subject Sample | # 2 | 05:20 | Subject Test Refused |
| Breath Volume | 0.089 | 05:20 | |
| Air Blank | 0.000 | 05:21 | |
| Control Test | 0.079 | 05:21 | |
| Air Blank | 0.000 | 05:22 | |
| Diagnostics Check | OK | 05:22 | |

Data Download Date/Time: 07/12/2021 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: PHAIR

DATE/TIME OF TEST: 06/01/2021 00:42:43

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 00:36 |
| Air Blank | | 0.000 | 00:36 |
| Control Test | | 0.078 | 00:36 |
| Air Blank | | 0.000 | 00:37 |
| Subject Sample | # 1 | 0.159 | 00:38 |
| Breath Volume | | 4.468 | 00:38 |
| Air Blank | | 0.000 | 00:38 |
| Air Blank | | 0.000 | 00:40 |
| Subject Sample | # 2 | 0.155 | 00:41 |
| Breath Volume | | 4.765 | 00:41 |
| Air Blank | | 0.000 | 00:41 |
| Control Test | | 0.080 | 00:42 |
| Air Blank | | 0.000 | 00:42 |
| Diagnostics Check | | OK | 00:42 |

Data Download Date/Time: 06/22/2021 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/02/2021 02:48:22

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:41 |
| Air Blank | 0.000 | 02:42 |
| Control Test | 0.080 | 02:42 |
| Air Blank | 0.000 | 02:43 |
| Subject Sample # 1 | 0.090 | 02:43 |
| Breath Volume | 1.996 | 02:43 |
| Air Blank | 0.000 | 02:44 |
| Air Blank | 0.000 | 02:46 |
| Subject Sample # 2 | 0.090 | 02:46 |
| Breath Volume | 3.335 | 02:46 |
| Air Blank | 0.000 | 02:47 |
| Control Test | 0.079 | 02:47 |
| Air Blank | 0.000 | 02:48 |
| Diagnostics Check | OK | 02:48 |

Data Download Date/Time: 06/22/2021 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/10/2021 00:30:07

| | | | |
|--------------------|-------|-------|---------------------------|
| Diagnostics Check | OK | 00:23 | |
| Air Blank | 0.000 | 00:23 | |
| Control Test | 0.078 | 00:23 | |
| Air Blank | 0.000 | 00:24 | |
| Subject Sample # 1 | 0.000 | 00:25 | |
| Breath Volume | 2.156 | 00:25 | |
| Air Blank | 0.000 | 00:25 | |
| Air Blank | 0.000 | 00:27 | |
| Subject Sample # 2 | 0.000 | 00:28 | |
| Breath Volume | 2.023 | 00:28 | |
| Air Blank | 0.000 | 00:28 | |
| Control Test | 0.077 | 00:29 | |
| Air Blank | 0.000 | 00:29 | |
| Diagnostics Check | Fail | 00:29 | Analytical Stability Test |
| Air Blank | 0.000 | 00:30 | |

Data Download Date/Time: 06/22/2021 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: MILLECAN

DATE/TIME OF TEST: 06/16/2021 01:20:51

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:14 |
| Air Blank | 0.000 | 01:15 |
| Control Test | 0.077 | 01:15 |
| Air Blank | 0.000 | 01:16 |
| Subject Sample # 1 | 0.000 | 01:16 |
| Breath Volume | 1.679 | 01:16 |
| Air Blank | 0.000 | 01:17 |
| Air Blank | 0.000 | 01:19 |
| Subject Sample # 2 | 0.000 | 01:19 |
| Breath Volume | 2.601 | 01:19 |
| Air Blank | 0.000 | 01:19 |
| Control Test | 0.077 | 01:20 |
| Air Blank | 0.000 | 01:20 |
| Diagnostics Check | OK | 01:20 |

Data Download Date/Time: 06/22/2021 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ROMEIN,CASEY
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 06/19/2021 00:17:22

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:11 |
| Air Blank | 0.000 | 00:11 |
| Control Test | 0.076 | 00:11 |
| Air Blank | 0.000 | 00:12 |
| Subject Sample # 1 | 0.043 | 00:12 |
| Breath Volume | 3.250 | 00:12 |
| Air Blank | 0.000 | 00:13 |
| Air Blank | 0.000 | 00:15 |
| Subject Sample # 2 | 0.042 | 00:15 |
| Breath Volume | 2.953 | 00:15 |
| Air Blank | 0.000 | 00:16 |
| Control Test | 0.077 | 00:16 |
| Air Blank | 0.000 | 00:17 |
| Diagnostics Check | OK | 00:17 |

Data Download Date/Time: 06/22/2021 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: HOLLOWAY

DATE/TIME OF TEST: 06/28/2021 20:23:05

| | | | |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK | 20:22 | |
| Air Blank | 0.000 | 20:22 | |
| | | 20:22 | RFI Detect |
| Air Blank | | 20:23 | RFI Detect |

Data Download Date/Time: 07/27/2021 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: HOLLOWAY

DATE/TIME OF TEST: 06/28/2021 20:29:33

| | | | |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK | 20:24 | |
| Air Blank | 0.000 | 20:25 | |
| Control Test | 0.077 | 20:25 | |
| Air Blank | 0.000 | 20:25 | |
| Subject Sample | # 1 0.178 | 20:26 | |
| Breath Volume | 3.082 | 20:26 | |
| Air Blank | 0.000 | 20:27 | |
| Air Blank | | 20:28 | RFI Detect |
| Air Blank | | 20:29 | RFI Detect |
| Air Blank | | 20:29 | RFI Detect |

Data Download Date/Time: 07/27/2021 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: HOLLOWAY

DATE/TIME OF TEST: 06/28/2021 20:36:39

| | | | |
|--------------------|-------|-------|------------|
| Diagnostics Check | OK | 20:30 | |
| Air Blank | 0.000 | 20:31 | |
| Control Test | 0.077 | 20:31 | |
| Air Blank | 0.000 | 20:32 | |
| Subject Sample # 1 | 0.181 | 20:32 | |
| Breath Volume | 3.710 | 20:32 | |
| Air Blank | 0.000 | 20:33 | |
| Air Blank | 0.000 | 20:35 | |
| Subject Sample # 2 | | 20:35 | RFI Detect |
| Breath Volume | 0.000 | 20:35 | |
| Air Blank | 0.000 | 20:36 | |
| | | 20:36 | RFI Detect |
| Air Blank | 0.000 | 20:36 | |

Data Download Date/Time: 07/27/2021 20:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: HOLLOWAY

DATE/TIME OF TEST: 06/28/2021 21:07:02

| | | | |
|-------------------|----|-------|------------|
| Diagnostics Check | OK | 21:06 | |
| Air Blank | | 21:06 | RFI Detect |
| Air Blank | | 21:07 | RFI Detect |

Data Download Date/Time: 07/27/2021 20:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: HOLLOWAY

DATE/TIME OF TEST: 06/28/2021 21:13:20

| | | | |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK | 21:12 | |
| Air Blank | | 21:12 | RFI Detect |
| Air Blank | 0.000 | 21:13 | |

Data Download Date/Time: 07/27/2021 20:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 06/29/2021 02:48:19

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:41 |
| Air Blank | 0.000 | 02:41 |
| Control Test | 0.078 | 02:42 |
| Air Blank | 0.000 | 02:42 |
| Subject Sample # 1 | 0.239 | 02:43 |
| Breath Volume | 2.203 | 02:43 |
| Air Blank | 0.000 | 02:44 |
| Air Blank | 0.000 | 02:46 |
| Subject Sample # 2 | 0.238 | 02:46 |
| Breath Volume | 2.289 | 02:46 |
| Air Blank | 0.000 | 02:47 |
| Control Test | 0.076 | 02:47 |
| Air Blank | 0.000 | 02:48 |
| Diagnostics Check | OK | 02:48 |

Data Download Date/Time: 07/27/2021 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,SHANNON,N
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: SPELLER,BENJAMIN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/06/2021 04:13:16

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 04:05 |
| Air Blank | 0.000 | 04:05 |
| Control Test | 0.079 | 04:06 |
| Air Blank | 0.000 | 04:06 |
| Subject Sample # 1 | 0.170 | 04:08 |
| Breath Volume | 1.843 | 04:08 |
| Air Blank | 0.000 | 04:09 |
| Air Blank | 0.000 | 04:10 |
| Subject Sample # 2 | 0.180 | 04:11 |
| Breath Volume | 1.335 | 04:11 |
| Air Blank | 0.000 | 04:12 |
| Control Test | 0.079 | 04:12 |
| Air Blank | 0.000 | 04:13 |
| Diagnostics Check | OK | 04:13 |

Data Download Date/Time: 06/30/2021 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: FIDLER

DATE/TIME OF TEST: 06/04/2021 02:20:24

| | | | |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK | 02:08 | |
| Air Blank | 0.000 | 02:09 | |
| Control Test | 0.080 | 02:09 | |
| Air Blank | 0.000 | 02:10 | |
| Subject Sample | # 1 | 02:13 | No Sample Provided |
| Breath Volume | 0.000 | 02:13 | |
| Air Blank | 0.000 | 02:13 | |
| Air Blank | 0.000 | 02:15 | |
| Subject Sample | # 2 | 02:19 | No Sample Provided |
| Breath Volume | 0.000 | 02:19 | |
| Air Blank | 0.000 | 02:19 | |
| Control Test | 0.081 | 02:19 | |
| Air Blank | 0.000 | 02:20 | |
| Diagnostics Check | OK | 02:20 | |

Data Download Date/Time: 07/16/2021 14:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: KHATIB

DATE/TIME OF TEST: 06/04/2021 03:34:35

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 03:22 | |
| Air Blank | | 0.000 | 03:22 | |
| Control Test | | 0.081 | 03:23 | |
| Air Blank | | 0.000 | 03:23 | |
| Subject Sample | # 1 | 0.136 | 03:26 | Volume Not Met |
| Breath Volume | | 0.832 | 03:26 | |
| Air Blank | | 0.000 | 03:27 | |
| Air Blank | | 0.000 | 03:29 | |
| Subject Sample | # 2 | 0.137 | 03:29 | |
| Breath Volume | | 1.382 | 03:29 | |
| Air Blank | | 0.000 | 03:30 | |
| Air Blank | | 0.000 | 03:32 | |
| Subject Sample | # 3 | 0.134 | 03:32 | |
| Breath Volume | | 1.242 | 03:32 | |
| Air Blank | | 0.000 | 03:33 | |
| Control Test | | 0.079 | 03:33 | |
| Air Blank | | 0.000 | 03:34 | |
| Diagnostics Check | | OK | 03:34 | |

Data Download Date/Time: 07/16/2021 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY, ANIA, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: FAITH

DATE/TIME OF TEST: 06/05/2021 00:11:40

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 00:02 |
| Air Blank | | 0.000 | 00:03 |
| Control Test | | 0.080 | 00:03 |
| Air Blank | | 0.000 | 00:04 |
| Subject Sample | # 1 | 0.208 | 00:05 |
| Breath Volume | | 1.511 | 00:05 |
| Air Blank | | 0.000 | 00:06 |
| Air Blank | | 0.000 | 00:08 |
| Subject Sample | # 2 | 0.207 | 00:09 |
| Breath Volume | | 1.738 | 00:09 |
| Air Blank | | 0.000 | 00:10 |
| Control Test | | 0.080 | 00:11 |
| Air Blank | | 0.000 | 00:11 |
| Diagnostics Check | | OK | 00:11 |

Data Download Date/Time: 07/16/2021 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: RETENELLER

DATE/TIME OF TEST: 06/05/2021 03:35:49

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:29 |
| Air Blank | 0.000 | 03:29 |
| Control Test | 0.080 | 03:29 |
| Air Blank | 0.000 | 03:30 |
| Subject Sample # 1 | 0.000 | 03:31 |
| Breath Volume | 3.132 | 03:31 |
| Air Blank | 0.000 | 03:31 |
| Air Blank | 0.000 | 03:33 |
| Subject Sample # 2 | 0.000 | 03:34 |
| Breath Volume | 3.046 | 03:34 |
| Air Blank | 0.000 | 03:34 |
| Control Test | 0.080 | 03:35 |
| Air Blank | 0.000 | 03:35 |
| Diagnostics Check | OK | 03:35 |

Data Download Date/Time: 07/16/2021 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PATINO ROLDAN

DATE/TIME OF TEST: 06/06/2021 03:43:40

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 03:36 |
| Air Blank | | 0.000 | 03:37 |
| Control Test | | 0.081 | 03:37 |
| Air Blank | | 0.000 | 03:38 |
| Subject Sample | # 1 | 0.000 | 03:38 |
| Breath Volume | | 1.320 | 03:38 |
| Air Blank | | 0.000 | 03:39 |
| Air Blank | | 0.000 | 03:41 |
| Subject Sample | # 2 | 0.000 | 03:42 |
| Breath Volume | | 1.507 | 03:42 |
| Air Blank | | 0.000 | 03:42 |
| Control Test | | 0.081 | 03:43 |
| Air Blank | | 0.000 | 03:43 |
| Diagnostics Check | | OK | 03:43 |

Data Download Date/Time: 07/16/2021 14:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WINTER,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: MONTES JR

DATE/TIME OF TEST: 06/06/2021 06:14:01

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 06:07 |
| Air Blank | 0.000 | 06:07 |
| Control Test | 0.080 | 06:08 |
| Air Blank | 0.000 | 06:08 |
| Subject Sample # 1 | 0.166 | 06:09 |
| Breath Volume | 2.574 | 06:09 |
| Air Blank | 0.000 | 06:10 |
| Air Blank | 0.000 | 06:11 |
| Subject Sample # 2 | 0.172 | 06:12 |
| Breath Volume | 2.140 | 06:12 |
| Air Blank | 0.000 | 06:13 |
| Control Test | 0.080 | 06:13 |
| Air Blank | 0.000 | 06:13 |
| Diagnostics Check | OK | 06:14 |

Data Download Date/Time: 07/16/2021 14:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: TATE

DATE/TIME OF TEST: 06/11/2021 03:14:20

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 03:07 |
| Air Blank | | 0.000 | 03:07 |
| Control Test | | 0.081 | 03:07 |
| Air Blank | | 0.000 | 03:08 |
| Subject Sample | # 1 | 0.210 | 03:09 |
| Breath Volume | | 2.496 | 03:09 |
| Air Blank | | 0.000 | 03:10 |
| Air Blank | | 0.000 | 03:11 |
| Subject Sample | # 2 | 0.213 | 03:12 |
| Breath Volume | | 2.445 | 03:12 |
| Air Blank | | 0.000 | 03:13 |
| Control Test | | 0.079 | 03:13 |
| Air Blank | | 0.000 | 03:14 |
| Diagnostics Check | | OK | 03:14 |

Data Download Date/Time: 07/16/2021 14:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: RAMIREZ JR.

DATE/TIME OF TEST: 06/13/2021 06:18:27

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 06:06 |
| Air Blank | 0.000 | 06:06 |
| Control Test | 0.081 | 06:07 |
| Air Blank | 0.000 | 06:07 |
| Subject Sample # 1 | 0.231 | 06:08 |
| Breath Volume | 3.414 | 06:08 |
| Air Blank | 0.000 | 06:09 |
| Air Blank | 0.000 | 06:11 |
| Subject Sample # 2 | 0.203 | 06:13 |
| Breath Volume | 2.046 | 06:13 |
| Air Blank | 0.000 | 06:14 |
| Air Blank | 0.000 | 06:16 |
| Subject Sample # 3 | 0.203 | 06:16 |
| Breath Volume | 3.207 | 06:16 |
| Air Blank | 0.000 | 06:17 |
| Control Test | 0.079 | 06:17 |
| Air Blank | 0.000 | 06:18 |
| Diagnostics Check | OK | 06:18 |

Data Download Date/Time: 07/16/2021 14:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: CANSLER

DATE/TIME OF TEST: 06/13/2021 23:37:55

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:30 |
| Air Blank | 0.000 | 23:31 |
| Control Test | 0.080 | 23:31 |
| Air Blank | 0.000 | 23:31 |
| Subject Sample # 1 | 0.124 | 23:33 |
| Breath Volume | 1.835 | 23:33 |
| Air Blank | 0.000 | 23:33 |
| Air Blank | 0.000 | 23:35 |
| Subject Sample # 2 | 0.124 | 23:36 |
| Breath Volume | 1.964 | 23:36 |
| Air Blank | 0.000 | 23:36 |
| Control Test | 0.081 | 23:37 |
| Air Blank | 0.000 | 23:37 |
| Diagnostics Check | OK | 23:37 |

Data Download Date/Time: 07/16/2021 14:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 06/14/2021 01:41:16

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 01:27 | |
| Air Blank | | 0.000 | 01:28 | |
| Control Test | | 0.081 | 01:28 | |
| Air Blank | | 0.000 | 01:29 | |
| Subject Sample | # 1 | 0.154 | 01:32 | Volume Not Met |
| Breath Volume | | 1.261 | 01:32 | |
| Air Blank | | 0.000 | 01:33 | |
| Air Blank | | 0.000 | 01:34 | |
| Subject Sample | # 2 | 0.176 | 01:35 | |
| Breath Volume | | 1.480 | 01:35 | |
| Air Blank | | 0.000 | 01:36 | |
| Air Blank | | 0.000 | 01:38 | |
| Subject Sample | # 3 | 0.172 | 01:39 | |
| Breath Volume | | 1.214 | 01:39 | |
| Air Blank | | 0.000 | 01:40 | |
| Control Test | | 0.079 | 01:40 | |
| Air Blank | | 0.000 | 01:41 | |
| Diagnostics Check | | OK | 01:41 | |

Data Download Date/Time: 07/16/2021 14:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: FALETTE

DATE/TIME OF TEST: 06/14/2021 02:45:19

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:38 |
| Air Blank | 0.000 | 02:38 |
| Control Test | 0.080 | 02:39 |
| Air Blank | 0.000 | 02:39 |
| Subject Sample # 1 | 0.239 | 02:40 |
| Breath Volume | 2.449 | 02:40 |
| Air Blank | 0.000 | 02:41 |
| Air Blank | 0.000 | 02:43 |
| Subject Sample # 2 | 0.221 | 02:43 |
| Breath Volume | 1.648 | 02:43 |
| Air Blank | 0.000 | 02:44 |
| Control Test | 0.080 | 02:44 |
| Air Blank | 0.000 | 02:45 |
| Diagnostics Check | OK | 02:45 |

Data Download Date/Time: 07/16/2021 14:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY, ANIA, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 06/15/2021 00:30:58

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:24 |
| Air Blank | 0.000 | 00:24 |
| Control Test | 0.081 | 00:25 |
| Air Blank | 0.000 | 00:25 |
| Subject Sample # 1 | 0.162 | 00:26 |
| Breath Volume | 1.628 | 00:26 |
| Air Blank | 0.000 | 00:27 |
| Air Blank | 0.000 | 00:28 |
| Subject Sample # 2 | 0.161 | 00:29 |
| Breath Volume | 1.300 | 00:29 |
| Air Blank | 0.000 | 00:30 |
| Control Test | 0.080 | 00:30 |
| Air Blank | 0.000 | 00:30 |
| Diagnostics Check | OK | 00:30 |

Data Download Date/Time: 07/16/2021 14:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: AVANT JR

DATE/TIME OF TEST: 06/17/2021 03:26:17

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 03:18 |
| Air Blank | | 0.000 | 03:18 |
| Control Test | | 0.081 | 03:19 |
| Air Blank | | 0.000 | 03:19 |
| Subject Sample | # 1 | 0.173 | 03:20 |
| Breath Volume | | 2.117 | 03:20 |
| Air Blank | | 0.000 | 03:21 |
| Air Blank | | 0.000 | 03:23 |
| Subject Sample | # 2 | 0.164 | 03:24 |
| Breath Volume | | 1.343 | 03:24 |
| Air Blank | | 0.000 | 03:25 |
| Control Test | | 0.079 | 03:25 |
| Air Blank | | 0.000 | 03:26 |
| Diagnostics Check | | OK | 03:26 |

Data Download Date/Time: 07/16/2021 14:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: COLES

DATE/TIME OF TEST: 06/17/2021 03:44:44

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:38 |
| Air Blank | 0.000 | 03:38 |
| Control Test | 0.079 | 03:38 |
| Air Blank | 0.000 | 03:39 |
| Subject Sample # 1 | 0.139 | 03:40 |
| Breath Volume | 2.652 | 03:40 |
| Air Blank | 0.000 | 03:40 |
| Air Blank | 0.000 | 03:42 |
| Subject Sample # 2 | 0.132 | 03:43 |
| Breath Volume | 2.640 | 03:43 |
| Air Blank | 0.000 | 03:43 |
| Control Test | 0.080 | 03:44 |
| Air Blank | 0.000 | 03:44 |
| Diagnostics Check | OK | 03:44 |

Data Download Date/Time: 07/16/2021 14:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: POHLMEIER

DATE/TIME OF TEST: 06/18/2021 01:53:45

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 01:46 |
| Air Blank | | 0.000 | 01:46 |
| Control Test | | 0.080 | 01:47 |
| Air Blank | | 0.000 | 01:47 |
| Subject Sample | # 1 | 0.097 | 01:48 |
| Breath Volume | | 2.667 | 01:48 |
| Air Blank | | 0.000 | 01:49 |
| Air Blank | | 0.000 | 01:51 |
| Subject Sample | # 2 | 0.093 | 01:52 |
| Breath Volume | | 2.496 | 01:52 |
| Air Blank | | 0.000 | 01:52 |
| Control Test | | 0.079 | 01:53 |
| Air Blank | | 0.000 | 01:53 |
| Diagnostics Check | | OK | 01:53 |

Data Download Date/Time: 07/16/2021 14:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY, ANIA, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: PODD

DATE/TIME OF TEST: 06/19/2021 00:43:16

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 00:36 |
| Air Blank | | 0.000 | 00:37 |
| Control Test | | 0.081 | 00:37 |
| Air Blank | | 0.000 | 00:37 |
| Subject Sample | # 1 | 0.266 | 00:38 |
| Breath Volume | | 4.269 | 00:38 |
| Air Blank | | 0.000 | 00:39 |
| Air Blank | | 0.000 | 00:40 |
| Subject Sample | # 2 | 0.253 | 00:41 |
| Breath Volume | | 3.699 | 00:41 |
| Air Blank | | 0.000 | 00:42 |
| Control Test | | 0.080 | 00:42 |
| Air Blank | | 0.000 | 00:43 |
| Diagnostics Check | | OK | 00:43 |

Data Download Date/Time: 07/16/2021 14:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: PEREZ GUERRA

DATE/TIME OF TEST: 06/21/2021 04:04:24

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 03:48 | |
| Air Blank | | 0.000 | 03:48 | |
| Control Test | | 0.080 | 03:49 | |
| Air Blank | | 0.000 | 03:49 | |
| Subject Sample | # 1 | 0.121 | 03:52 | Volume Not Met |
| Breath Volume | | 0.191 | 03:52 | |
| Air Blank | | 0.000 | 03:53 | |
| Air Blank | | 0.000 | 03:55 | |
| Subject Sample | # 2 | 0.108 | 03:56 | |
| Breath Volume | | 1.410 | 03:56 | |
| Air Blank | | 0.000 | 03:57 | |
| Air Blank | | 0.000 | 03:59 | |
| Subject Sample | # 3 | 0.118 | 04:02 | |
| Breath Volume | | 1.437 | 04:02 | |
| Air Blank | | 0.000 | 04:03 | |
| Control Test | | 0.078 | 04:03 | |
| Air Blank | | 0.000 | 04:04 | |
| Diagnostics Check | | OK | 04:04 | |

Data Download Date/Time: 07/16/2021 14:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DURANCEAU,R
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/25/2021 04:25:34

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 04:13 | |
| Air Blank | | 0.000 | 04:14 | |
| Control Test | | 0.081 | 04:14 | |
| Air Blank | | 0.000 | 04:14 | |
| Subject Sample | # 1 | 0.146 | 04:18 | Volume Not Met |
| Breath Volume | | 0.792 | 04:18 | |
| Air Blank | | 0.000 | 04:19 | |
| Air Blank | | 0.000 | 04:20 | |
| Subject Sample | # 2 | 0.151 | 04:23 | Volume Not Met |
| Breath Volume | | 0.671 | 04:23 | |
| Air Blank | | 0.000 | 04:24 | |
| Control Test | | 0.078 | 04:24 | |
| Air Blank | | 0.000 | 04:25 | |
| Diagnostics Check | | OK | 04:25 | |

Data Download Date/Time: 07/16/2021 14:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DURANCEAU,R
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/25/2021 05:06:11

| | | | |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK | 04:51 | |
| Air Blank | 0.000 | 04:52 | |
| Control Test | 0.079 | 04:52 | |
| Air Blank | 0.000 | 04:52 | |
| Subject Sample | # 1 0.145 | 04:54 | |
| Breath Volume | 1.277 | 04:54 | |
| Air Blank | 0.000 | 04:55 | |
| Air Blank | 0.000 | 04:57 | |
| Subject Sample | # 2 0.130 | 05:00 | Volume Not Met |
| Breath Volume | 0.605 | 05:00 | |
| Air Blank | 0.000 | 05:01 | |
| Air Blank | 0.000 | 05:02 | |
| Subject Sample | # 3 0.142 | 05:04 | |
| Breath Volume | 1.121 | 05:04 | |
| Air Blank | 0.000 | 05:05 | |
| Control Test | 0.076 | 05:05 | |
| Air Blank | 0.000 | 05:06 | |
| Diagnostics Check | OK | 05:06 | |

Data Download Date/Time: 07/16/2021 14:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: WINTER

DATE/TIME OF TEST: 06/26/2021 04:03:38

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:54 |
| Air Blank | 0.000 | 03:54 |
| Control Test | 0.081 | 03:54 |
| Air Blank | 0.000 | 03:55 |
| Subject Sample # 1 | 0.130 | 03:58 |
| Breath Volume | 1.671 | 03:58 |
| Air Blank | 0.000 | 03:59 |
| Air Blank | 0.000 | 04:00 |
| Subject Sample # 2 | 0.134 | 04:01 |
| Breath Volume | 1.250 | 04:01 |
| Air Blank | 0.000 | 04:02 |
| Control Test | 0.079 | 04:03 |
| Air Blank | 0.000 | 04:03 |
| Diagnostics Check | OK | 04:03 |

Data Download Date/Time: 07/16/2021 14:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: WAINIO

DATE/TIME OF TEST: 06/27/2021 00:23:38

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:16 |
| Air Blank | 0.000 | 00:16 |
| Control Test | 0.080 | 00:17 |
| Air Blank | 0.000 | 00:17 |
| Subject Sample # 1 | 0.165 | 00:18 |
| Breath Volume | 1.781 | 00:18 |
| Air Blank | 0.000 | 00:19 |
| Air Blank | 0.000 | 00:21 |
| Subject Sample # 2 | 0.178 | 00:21 |
| Breath Volume | 3.375 | 00:21 |
| Air Blank | 0.000 | 00:22 |
| Control Test | 0.080 | 00:23 |
| Air Blank | 0.000 | 00:23 |
| Diagnostics Check | OK | 00:23 |

Data Download Date/Time: 07/16/2021 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: LARGEL

DATE/TIME OF TEST: 06/27/2021 01:12:53

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:05 |
| Air Blank | 0.000 | 01:06 |
| Control Test | 0.081 | 01:06 |
| Air Blank | 0.000 | 01:07 |
| Subject Sample # 1 | 0.056 | 01:08 |
| Breath Volume | 1.152 | 01:08 |
| Air Blank | 0.000 | 01:08 |
| Air Blank | 0.000 | 01:10 |
| Subject Sample # 2 | 0.059 | 01:11 |
| Breath Volume | 1.382 | 01:11 |
| Air Blank | 0.000 | 01:11 |
| Control Test | 0.081 | 01:12 |
| Air Blank | 0.000 | 01:12 |
| Diagnostics Check | OK | 01:12 |

Data Download Date/Time: 07/16/2021 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DACHS,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:49

SUBJECT'S LAST NAME: OJEDA

DATE/TIME OF TEST: 06/28/2021 17:18:34

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 17:10 |
| Air Blank | 0.000 | 17:10 |
| Control Test | 0.081 | 17:11 |
| Air Blank | 0.000 | 17:11 |
| Subject Sample # 1 | 0.000 | 17:13 |
| Breath Volume | 1.128 | 17:13 |
| Air Blank | 0.000 | 17:14 |
| Air Blank | 0.000 | 17:16 |
| Subject Sample # 2 | 0.000 | 17:17 |
| Breath Volume | 1.671 | 17:17 |
| Air Blank | 0.000 | 17:17 |
| Control Test | 0.080 | 17:17 |
| Air Blank | 0.000 | 17:18 |
| Diagnostics Check | OK | 17:18 |

Data Download Date/Time: 07/16/2021 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:19

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/30/2021 06:45:47

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 06:39 |
| Air Blank | 0.000 | 06:39 |
| Control Test | 0.081 | 06:40 |
| Air Blank | 0.000 | 06:40 |
| Subject Sample # 1 | 0.211 | 06:41 |
| Breath Volume | 2.765 | 06:41 |
| Air Blank | 0.000 | 06:41 |
| Air Blank | 0.000 | 06:43 |
| Subject Sample # 2 | 0.210 | 06:44 |
| Breath Volume | 3.417 | 06:44 |
| Air Blank | 0.000 | 06:44 |
| Control Test | 0.079 | 06:45 |
| Air Blank | 0.000 | 06:45 |
| Diagnostics Check | OK | 06:45 |

Data Download Date/Time: 07/16/2021 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNETT,JAMES,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: BARNETT,JAMES
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 05/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 06/09/2021 21:06:25

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 20:52 | |
| Air Blank | | 0.000 | 20:52 | |
| Control Test | | 0.081 | 20:52 | |
| Air Blank | | 0.000 | 20:53 | |
| Subject Sample | # 1 | 0.000 | 20:56 | Volume Not Met |
| Breath Volume | | 0.066 | 20:56 | |
| Air Blank | | 0.000 | 20:57 | |
| Air Blank | | 0.000 | 20:59 | |
| Subject Sample | # 2 | 0.317 | 21:01 | |
| Breath Volume | | 1.953 | 21:01 | |
| Air Blank | | 0.000 | 21:01 | |
| Air Blank | | 0.000 | 21:03 | |
| Subject Sample | # 3 | 0.310 | 21:04 | |
| Breath Volume | | 1.691 | 21:04 | |
| Air Blank | | 0.000 | 21:05 | |
| Control Test | | 0.078 | 21:05 | |
| Air Blank | | 0.000 | 21:06 | |
| Diagnostics Check | | OK | 21:06 | |

Data Download Date/Time: 06/10/2021 19:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/15/2021 01:11:46

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 01:08 | |
| Air Blank | 0.000 | 01:09 | |
| Control Test | 0.082 | 01:09 | |
| Air Blank | 0.000 | 01:10 | |
| Subject Sample | # 1 | 01:10 | Subject Test Refused |
| Breath Volume | 0.000 | 01:10 | |
| Air Blank | 0.000 | 01:10 | |
| Control Test | 0.082 | 01:11 | |
| Air Blank | 0.000 | 01:11 | |
| Diagnostics Check | OK | 01:11 | |

Data Download Date/Time: 07/13/2021 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIRES, DAVID,
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SMITH, MORGAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: OLSON

DATE/TIME OF TEST: 06/17/2021 03:43:21

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:36 |
| Air Blank | 0.000 | 03:36 |
| Control Test | 0.082 | 03:36 |
| Air Blank | 0.000 | 03:37 |
| Subject Sample # 1 | 0.217 | 03:38 |
| Breath Volume | 1.656 | 03:38 |
| Air Blank | 0.000 | 03:39 |
| Air Blank | 0.000 | 03:40 |
| Subject Sample # 2 | 0.215 | 03:41 |
| Breath Volume | 1.320 | 03:41 |
| Air Blank | 0.000 | 03:42 |
| Control Test | 0.080 | 03:42 |
| Air Blank | 0.000 | 03:43 |
| Diagnostics Check | OK | 03:43 |

Data Download Date/Time: 07/13/2021 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KITTRELL,WILLIAM,G
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KITTRELL,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: ELVIR

DATE/TIME OF TEST: 06/19/2021 03:47:51

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 03:41 |
| Air Blank | | 0.000 | 03:41 |
| Control Test | | 0.081 | 03:42 |
| Air Blank | | 0.000 | 03:42 |
| Subject Sample | # 1 | 0.198 | 03:43 |
| Breath Volume | | 1.996 | 03:43 |
| Air Blank | | 0.000 | 03:44 |
| Air Blank | | 0.000 | 03:45 |
| Subject Sample | # 2 | 0.206 | 03:46 |
| Breath Volume | | 3.558 | 03:46 |
| Air Blank | | 0.000 | 03:46 |
| Control Test | | 0.081 | 03:47 |
| Air Blank | | 0.000 | 03:47 |
| Diagnostics Check | | OK | 03:47 |

Data Download Date/Time: 07/13/2021 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 14:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:52

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/20/2021 15:28:31

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 15:21 |
| Air Blank | 0.000 | 15:22 |
| Control Test | 0.083 | 15:22 |
| Air Blank | 0.000 | 15:23 |
| Subject Sample # 1 | 0.153 | 15:23 |
| Breath Volume | 1.664 | 15:23 |
| Air Blank | 0.000 | 15:24 |
| Air Blank | 0.000 | 15:26 |
| Subject Sample # 2 | 0.161 | 15:26 |
| Breath Volume | 1.328 | 15:26 |
| Air Blank | 0.000 | 15:27 |
| Control Test | 0.080 | 15:27 |
| Air Blank | 0.000 | 15:28 |
| Diagnostics Check | OK | 15:28 |

Data Download Date/Time: 07/13/2021 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOOTHE,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOOTHE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: PROSEN

DATE/TIME OF TEST: 06/22/2021 00:42:22

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:35 |
| Air Blank | 0.000 | 00:36 |
| Control Test | 0.080 | 00:36 |
| Air Blank | 0.000 | 00:36 |
| Subject Sample # 1 | 0.193 | 00:37 |
| Breath Volume | 4.324 | 00:37 |
| Air Blank | 0.000 | 00:38 |
| Air Blank | 0.000 | 00:40 |
| Subject Sample # 2 | 0.188 | 00:40 |
| Breath Volume | 4.242 | 00:40 |
| Air Blank | 0.000 | 00:41 |
| Control Test | 0.079 | 00:41 |
| Air Blank | 0.000 | 00:42 |
| Diagnostics Check | OK | 00:42 |

Data Download Date/Time: 07/13/2021 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN PELT, COLTEN, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VAN PELT, COLTEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 16:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/23/2021 17:41:07

| | | | | |
|-------------------|-----|-------|-------|-------------------|
| Diagnostics Check | | OK | 17:30 | |
| Air Blank | | 0.000 | 17:30 | |
| Control Test | | 0.081 | 17:31 | |
| Air Blank | | 0.000 | 17:31 | |
| Subject Sample | # 1 | 0.250 | 17:32 | No .020 agreement |
| Breath Volume | | 2.523 | 17:32 | |
| Air Blank | | 0.000 | 17:33 | |
| Air Blank | | 0.000 | 17:35 | |
| Subject Sample | # 2 | 0.229 | 17:35 | No .020 agreement |
| Breath Volume | | 1.406 | 17:35 | |
| Air Blank | | 0.000 | 17:36 | |
| Air Blank | | 0.000 | 17:38 | |
| Subject Sample | # 3 | 0.204 | 17:39 | No .020 agreement |
| Breath Volume | | 1.175 | 17:39 | |
| Air Blank | | 0.000 | 17:40 | |
| Control Test | | 0.079 | 17:40 | |
| Air Blank | | 0.000 | 17:41 | |
| Diagnostics Check | | OK | 17:41 | |

Data Download Date/Time: 07/13/2021 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGALADO,JORGE,I
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: REGALADO,JORGE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/26/2021 19:24:14

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 19:17 |
| Air Blank | | 0.000 | 19:18 |
| Control Test | | 0.081 | 19:18 |
| Air Blank | | 0.000 | 19:19 |
| Subject Sample | # 1 | 0.253 | 19:19 |
| Breath Volume | | 1.664 | 19:19 |
| Air Blank | | 0.000 | 19:20 |
| Air Blank | | 0.000 | 19:22 |
| Subject Sample | # 2 | 0.244 | 19:22 |
| Breath Volume | | 1.296 | 19:22 |
| Air Blank | | 0.000 | 19:23 |
| Control Test | | 0.081 | 19:23 |
| Air Blank | | 0.000 | 19:24 |
| Diagnostics Check | | OK | 19:24 |

Data Download Date/Time: 07/13/2021 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGALADO,JORGE,I
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PURVIS,BOBBY
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: COUNTS

DATE/TIME OF TEST: 06/26/2021 21:57:17

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 21:50 |
| Air Blank | | 0.000 | 21:51 |
| Control Test | | 0.081 | 21:51 |
| Air Blank | | 0.000 | 21:51 |
| Subject Sample | # 1 | 0.214 | 21:52 |
| Breath Volume | | 1.695 | 21:52 |
| Air Blank | | 0.000 | 21:53 |
| Air Blank | | 0.000 | 21:55 |
| Subject Sample | # 2 | 0.202 | 21:55 |
| Breath Volume | | 1.789 | 21:55 |
| Air Blank | | 0.000 | 21:56 |
| Control Test | | 0.081 | 21:56 |
| Air Blank | | 0.000 | 21:57 |
| Diagnostics Check | | OK | 21:57 |

Data Download Date/Time: 07/13/2021 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINH,TAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DINH,TAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 12:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:46

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 06/27/2021 13:19:33

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 13:06 | |
| Air Blank | | 0.000 | 13:06 | |
| Control Test | | 0.082 | 13:07 | |
| Air Blank | | 0.000 | 13:07 | |
| Subject Sample | # 1 | 0.059 | 13:10 | Volume Not Met |
| Breath Volume | | 0.863 | 13:10 | |
| Air Blank | | 0.000 | 13:11 | |
| Air Blank | | 0.000 | 13:13 | |
| Subject Sample | # 2 | 0.063 | 13:13 | |
| Breath Volume | | 1.664 | 13:13 | |
| Air Blank | | 0.000 | 13:14 | |
| Air Blank | | 0.000 | 13:16 | |
| Subject Sample | # 3 | 0.067 | 13:17 | |
| Breath Volume | | 2.125 | 13:17 | |
| Air Blank | | 0.000 | 13:18 | |
| Control Test | | 0.079 | 13:18 | |
| Air Blank | | 0.000 | 13:19 | |
| Diagnostics Check | | OK | 13:19 | |

Data Download Date/Time: 07/13/2021 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNCH,LOGAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LYNCH,CHAD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/30/2021 21:56:17

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 21:52 | |
| Air Blank | 0.000 | 21:53 | |
| Control Test | 0.082 | 21:53 | |
| Air Blank | 0.000 | 21:53 | |
| Subject Sample | # 1 | 21:54 | Subject Test Refused |
| Breath Volume | 0.000 | 21:54 | |
| Air Blank | 0.000 | 21:55 | |
| Control Test | 0.082 | 21:55 | |
| Air Blank | 0.000 | 21:56 | |
| Diagnostics Check | OK | 21:56 | |

Data Download Date/Time: 07/13/2021 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNCH,CHAD,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LYNCH,CHAD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/30/2021 22:17:42

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:10 |
| Air Blank | 0.000 | 22:11 |
| Control Test | 0.082 | 22:11 |
| Air Blank | 0.000 | 22:12 |
| Subject Sample # 1 | 0.000 | 22:13 |
| Breath Volume | 2.109 | 22:13 |
| Air Blank | 0.000 | 22:13 |
| Air Blank | 0.000 | 22:15 |
| Subject Sample # 2 | 0.000 | 22:16 |
| Breath Volume | 1.988 | 22:16 |
| Air Blank | 0.000 | 22:16 |
| Control Test | 0.082 | 22:17 |
| Air Blank | 0.000 | 22:17 |
| Diagnostics Check | OK | 22:17 |

Data Download Date/Time: 07/13/2021 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROQUE,EUGENE,H
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ROQUE,EUGENE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 10:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:18

SUBJECT'S LAST NAME: BELLO

DATE/TIME OF TEST: 06/04/2021 11:47:19

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 11:44 | |
| Air Blank | 0.000 | 11:44 | |
| Control Test | 0.081 | 11:45 | |
| Air Blank | 0.000 | 11:45 | |
| Subject Sample | # 1 | 11:45 | Subject Test Refused |
| Breath Volume | 0.000 | 11:45 | |
| Air Blank | 0.000 | 11:46 | |
| Control Test | 0.081 | 11:46 | |
| Air Blank | 0.000 | 11:47 | |
| Diagnostics Check | OK | 11:47 | |

Data Download Date/Time: 06/28/2021 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CHARLES,BYANKA,S

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: RODRIGUEZ,GABRIELA

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 04:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 06/06/2021 05:18:48

| | | | |
|-------------------|----|-------|------------|
| Diagnostics Check | OK | 05:18 | |
| Air Blank | | 05:18 | RFI Detect |
| Air Blank | | 05:18 | RFI Detect |

Data Download Date/Time: 06/28/2021 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CHARLES,BYANKA,S

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: RODRIGUEZ,GABRIELA

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 04:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 06/06/2021 05:29:08

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 05:26 | |
| Air Blank | 0.000 | 05:26 | |
| Control Test | 0.081 | 05:27 | |
| Air Blank | 0.000 | 05:27 | |
| Subject Sample | # 1 | 05:27 | Subject Test Refused |
| Breath Volume | 0.000 | 05:27 | |
| Air Blank | 0.000 | 05:28 | |
| Control Test | 0.082 | 05:28 | |
| Air Blank | 0.000 | 05:29 | |
| Diagnostics Check | OK | 05:29 | |

Data Download Date/Time: 06/28/2021 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,BRIAN,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,YUNIER
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: ABEL

DATE/TIME OF TEST: 06/08/2021 21:26:44

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 21:23 | |
| Air Blank | 0.000 | 21:24 | |
| Control Test | 0.082 | 21:24 | |
| Air Blank | 0.000 | 21:25 | |
| Subject Sample | # 1 | 21:25 | Subject Test Refused |
| Breath Volume | 0.000 | 21:25 | |
| Air Blank | 0.000 | 21:25 | |
| Control Test | 0.081 | 21:26 | |
| Air Blank | 0.000 | 21:26 | |
| Diagnostics Check | OK | 21:26 | |

Data Download Date/Time: 06/28/2021 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDEROS,BRYAN,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: LLOVET,GERALDO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 06/12/2021 05:43:22

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 05:40 | |
| Air Blank | 0.000 | 05:41 | |
| Control Test | 0.082 | 05:41 | |
| Air Blank | 0.000 | 05:41 | |
| Subject Sample | # 1 | 05:42 | Subject Test Refused |
| Breath Volume | 0.000 | 05:42 | |
| Air Blank | 0.000 | 05:42 | |
| Control Test | 0.081 | 05:42 | |
| Air Blank | 0.000 | 05:43 | |
| Diagnostics Check | OK | 05:43 | |

Data Download Date/Time: 06/28/2021 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ACEVEDO,CRISTIAN,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CARRA,ANTHONY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 18:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: ACEVEDO

DATE/TIME OF TEST: 06/18/2021 18:54:11

| | | | |
|-------------------|----|-------|-----------------|
| Diagnostics Check | OK | 18:53 | |
| Air Blank | | 18:53 | Improper Sample |
| Air Blank | | 18:54 | Improper Sample |

Data Download Date/Time: 06/28/2021 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ACEVEDO,CRISTIAN,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CARRA,ANTHONY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 18:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: ACEVEDO

DATE/TIME OF TEST: 06/18/2021 18:57:27

| | | | |
|-------------------|-------|-------|--------------|
| Diagnostics Check | OK | 18:56 | |
| Air Blank | | 18:56 | Ambient Fail |
| Air Blank | 0.000 | 18:57 | |

Data Download Date/Time: 06/28/2021 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ACEVEDO,CRISTIAN,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CARRA,ANTHONY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 18:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: ACEVEDO

DATE/TIME OF TEST: 06/18/2021 18:59:50

| | | | |
|-------------------|----|-------|-----------------|
| Diagnostics Check | OK | 18:59 | |
| Air Blank | | 18:59 | Improper Sample |
| Air Blank | | 18:59 | Improper Sample |

Data Download Date/Time: 06/28/2021 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ACEVEDO,CRISTIAN,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CARRA,ANTHONY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 18:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: D6665788

DATE/TIME OF TEST: 06/18/2021 19:02:53

| | | | |
|-------------------|-------|-------|--------------|
| Diagnostics Check | OK | 19:01 | |
| Air Blank | | 19:02 | Ambient Fail |
| Air Blank | 0.000 | 19:02 | |

Data Download Date/Time: 06/28/2021 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRA,ANTHONY,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CARRA,ANTHONY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: ACEVEDO

DATE/TIME OF TEST: 06/18/2021 19:10:31

| | | | |
|-------------------|-------|-------|-----------------|
| Diagnostics Check | OK | 19:09 | |
| Air Blank | | 19:09 | Improper Sample |
| Air Blank | 0.000 | 19:10 | |

Data Download Date/Time: 06/28/2021 16:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CARRA, ANTHONY,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CARRA, ANTHONY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 18:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: ACEVEDO

DATE/TIME OF TEST: 06/18/2021 19:12:48

| | | | |
|-------------------|----|-------|-----------------|
| Diagnostics Check | OK | 19:12 | |
| Air Blank | | 19:12 | Improper Sample |
| Air Blank | | 19:12 | Improper Sample |

Data Download Date/Time: 06/28/2021 16:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CARRA, ANTHONY,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CARRA, ANTHONY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 18:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: ACEVEDO

DATE/TIME OF TEST: 06/18/2021 19:18:37

| | | | |
|-------------------|----|-------|-----------------|
| Diagnostics Check | OK | 19:18 | |
| Air Blank | | 19:18 | Improper Sample |
| Air Blank | | 19:18 | Improper Sample |

Data Download Date/Time: 06/28/2021 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NARDONI,J,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: NARDONI,J
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 09:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:44

SUBJECT'S LAST NAME: FISH JR

DATE/TIME OF TEST: 06/19/2021 10:11:35

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 10:05 |
| Air Blank | 0.000 | 10:05 |
| Control Test | 0.080 | 10:05 |
| Air Blank | 0.000 | 10:06 |
| Subject Sample # 1 | 0.065 | 10:07 |
| Breath Volume | 2.640 | 10:07 |
| Air Blank | 0.000 | 10:07 |
| Air Blank | 0.000 | 10:09 |
| Subject Sample # 2 | 0.058 | 10:10 |
| Breath Volume | 1.445 | 10:10 |
| Air Blank | 0.000 | 10:10 |
| Control Test | 0.075 | 10:10 |
| Air Blank | 0.000 | 10:11 |
| Diagnostics Check | OK | 10:11 |

Data Download Date/Time: 06/28/2021 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRERO,MARCOS,E
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ROJAS,ROBERT
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 06:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:18

SUBJECT'S LAST NAME: ZEPEDA

DATE/TIME OF TEST: 06/27/2021 07:44:35

| | | | |
|--------------------|-------|-------|----------------|
| Diagnostics Check | OK | 07:38 | |
| Air Blank | 0.000 | 07:39 | |
| Control Test | 0.081 | 07:39 | |
| Air Blank | 0.000 | 07:40 | |
| Subject Sample # 1 | 0.271 | 07:43 | Volume Not Met |
| Breath Volume | 0.832 | 07:43 | |
| Air Blank | | 07:44 | Purge Fail |
| Air Blank | 0.000 | 07:44 | |

Data Download Date/Time: 06/28/2021 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRERO,MARCOS,E
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ROJAS,ROBERT
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 06:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:18

SUBJECT'S LAST NAME: ZEPEDA

DATE/TIME OF TEST: 06/27/2021 07:58:59

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 07:47 | |
| Air Blank | | 0.000 | 07:48 | |
| Control Test | | 0.079 | 07:48 | |
| Air Blank | | 0.000 | 07:49 | |
| Subject Sample | # 1 | 0.258 | 07:52 | Volume Not Met |
| Breath Volume | | 0.886 | 07:52 | |
| Air Blank | | 0.000 | 07:53 | |
| Air Blank | | 0.000 | 07:54 | |
| Subject Sample | # 2 | 0.262 | 07:57 | Volume Not Met |
| Breath Volume | | 0.519 | 07:57 | |
| Air Blank | | | 07:58 | Purge Fail |
| Air Blank | | 0.000 | 07:58 | |

Data Download Date/Time: 06/28/2021 16:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MARRERO,MARCOS,E

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: ROJAS,ROBERT

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 06:46

GENDER: Female

OBSERVATION PERIOD BEGAN: 07:18

SUBJECT'S LAST NAME: ZEPEDA

DATE/TIME OF TEST: 06/27/2021 08:12:20

| | | | | |
|-------------------|-----|-------|-------|----------------------|
| Diagnostics Check | | OK | 08:01 | |
| Air Blank | | 0.000 | 08:02 | |
| Control Test | | 0.078 | 08:02 | |
| Air Blank | | 0.000 | 08:03 | |
| Subject Sample | # 1 | 0.270 | 08:06 | Volume Not Met |
| Breath Volume | | 0.906 | 08:06 | |
| Air Blank | | 0.000 | 08:07 | |
| Air Blank | | 0.000 | 08:08 | |
| Subject Sample | # 2 | | 08:10 | Subject Test Refused |
| Breath Volume | | 0.570 | 08:10 | |
| Air Blank | | 0.000 | 08:11 | |
| Control Test | | 0.077 | 08:11 | |
| Air Blank | | 0.000 | 08:12 | |
| Diagnostics Check | | OK | 08:12 | |

Data Download Date/Time: 06/28/2021 16:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDGEPEETH,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: HEDGEPEETH,TIMOTHY
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: BUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: DISOTUAR

DATE/TIME OF TEST: 06/28/2021 22:21:02

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 22:17 | |
| Air Blank | 0.000 | 22:18 | |
| Control Test | 0.081 | 22:18 | |
| Air Blank | 0.000 | 22:19 | |
| Subject Sample | # 1 | 22:19 | Subject Test Refused |
| Breath Volume | 0.000 | 22:19 | |
| Air Blank | 0.000 | 22:20 | |
| Control Test | 0.081 | 22:20 | |
| Air Blank | 0.000 | 22:20 | |
| Diagnostics Check | OK | 22:21 | |

Data Download Date/Time: 07/27/2021 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PASTOR,JOSE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: COURT ORD
ARREST TIME: 11:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/29/2021 11:56:51

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 11:50 |
| Air Blank | 0.000 | 11:51 |
| Control Test | 0.081 | 11:51 |
| Air Blank | 0.000 | 11:52 |
| Subject Sample # 1 | 0.000 | 11:52 |
| Breath Volume | 1.812 | 11:52 |
| Air Blank | 0.000 | 11:53 |
| Air Blank | 0.000 | 11:55 |
| Subject Sample # 2 | 0.000 | 11:55 |
| Breath Volume | 1.609 | 11:55 |
| Air Blank | 0.000 | 11:55 |
| Control Test | 0.082 | 11:56 |
| Air Blank | 0.000 | 11:56 |
| Diagnostics Check | OK | 11:56 |

Data Download Date/Time: 07/27/2021 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,ALEXANDRIA,N
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HRADECKY,ALEXANDRIA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/01/2021 23:31:45

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:25 |
| Air Blank | 0.000 | 23:25 |
| Control Test | 0.078 | 23:25 |
| Air Blank | 0.000 | 23:26 |
| Subject Sample # 1 | 0.000 | 23:27 |
| Breath Volume | 2.609 | 23:27 |
| Air Blank | 0.000 | 23:27 |
| Air Blank | 0.000 | 23:29 |
| Subject Sample # 2 | 0.000 | 23:30 |
| Breath Volume | 2.437 | 23:30 |
| Air Blank | 0.000 | 23:30 |
| Control Test | 0.078 | 23:31 |
| Air Blank | 0.000 | 23:31 |
| Diagnostics Check | OK | 23:31 |

Data Download Date/Time: 06/10/2021 10:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HANNA,WARDELL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 11:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: CAMPANELLA

DATE/TIME OF TEST: 06/05/2021 00:03:13

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 00:00 | |
| Air Blank | 0.000 | 00:00 | |
| Control Test | 0.079 | 00:00 | |
| Air Blank | 0.000 | 00:01 | |
| Subject Sample | # 1 | 00:01 | Subject Test Refused |
| Breath Volume | 0.000 | 00:01 | |
| Air Blank | 0.000 | 00:02 | |
| Control Test | 0.079 | 00:02 | |
| Air Blank | 0.000 | 00:03 | |
| Diagnostics Check | OK | 00:03 | |

Data Download Date/Time: 06/10/2021 10:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MORENO,JORGE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: KRAUS

DATE/TIME OF TEST: 06/05/2021 02:13:59

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 02:10 | |
| Air Blank | 0.000 | 02:11 | |
| Control Test | 0.078 | 02:11 | |
| Air Blank | 0.000 | 02:12 | |
| Subject Sample | # 1 | 02:12 | Subject Test Refused |
| Breath Volume | 0.000 | 02:12 | |
| Air Blank | 0.000 | 02:13 | |
| Control Test | 0.079 | 02:13 | |
| Air Blank | 0.000 | 02:13 | |
| Diagnostics Check | OK | 02:13 | |

Data Download Date/Time: 06/10/2021 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MORENO,JORGE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 12:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:56

SUBJECT'S LAST NAME: KULKA

DATE/TIME OF TEST: 06/06/2021 01:10:21

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 00:59 |
| Air Blank | | 0.000 | 00:59 |
| Control Test | | 0.078 | 01:00 |
| Air Blank | | 0.000 | 01:00 |
| Subject Sample | # 1 | 0.238 | 01:02 |
| Breath Volume | | 1.488 | 01:02 |
| Air Blank | | 0.000 | 01:03 |
| Air Blank | | 0.000 | 01:04 |
| Subject Sample | # 2 | 0.266 | 01:05 |
| Breath Volume | | 3.199 | 01:05 |
| Air Blank | | 0.000 | 01:06 |
| Air Blank | | 0.000 | 01:08 |
| Subject Sample | # 3 | 0.257 | 01:08 |
| Breath Volume | | 2.464 | 01:08 |
| Air Blank | | 0.000 | 01:09 |
| Control Test | | 0.076 | 01:09 |
| Air Blank | | 0.000 | 01:10 |
| Diagnostics Check | | OK | 01:10 |

Data Download Date/Time: 06/10/2021 10:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,ALEXANDRIA,N
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HRADECKY,ALEXANDRIA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 06/06/2021 23:13:08

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 23:10 | |
| Air Blank | 0.000 | 23:10 | |
| Control Test | 0.078 | 23:10 | |
| Air Blank | 0.000 | 23:11 | |
| Subject Sample | # 1 | 23:11 | Subject Test Refused |
| Breath Volume | 0.000 | 23:11 | |
| Air Blank | 0.000 | 23:12 | |
| Control Test | 0.078 | 23:12 | |
| Air Blank | 0.000 | 23:13 | |
| Diagnostics Check | OK | 23:13 | |

Data Download Date/Time: 06/10/2021 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,JOSHUA,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: BRADY,JOSHUA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/09/2021 03:09:11

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 03:02 |
| Air Blank | | 0.000 | 03:03 |
| Control Test | | 0.078 | 03:03 |
| Air Blank | | 0.000 | 03:04 |
| Subject Sample | # 1 | 0.174 | 03:04 |
| Breath Volume | | 2.191 | 03:04 |
| Air Blank | | 0.000 | 03:05 |
| Air Blank | | 0.000 | 03:07 |
| Subject Sample | # 2 | 0.170 | 03:07 |
| Breath Volume | | 1.316 | 03:07 |
| Air Blank | | 0.000 | 03:08 |
| Control Test | | 0.078 | 03:08 |
| Air Blank | | 0.000 | 03:09 |
| Diagnostics Check | | OK | 03:09 |

Data Download Date/Time: 06/10/2021 10:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANGEL,ROBERT,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: LINDSEY

DATE/TIME OF TEST: 06/09/2021 04:21:34

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 04:18 | |
| Air Blank | 0.000 | 04:19 | |
| Control Test | 0.078 | 04:19 | |
| Air Blank | 0.000 | 04:20 | |
| Subject Sample | # 1 | 04:20 | Subject Test Refused |
| Breath Volume | 0.000 | 04:20 | |
| Air Blank | 0.000 | 04:20 | |
| Control Test | 0.078 | 04:20 | |
| Air Blank | 0.000 | 04:21 | |
| Diagnostics Check | OK | 04:21 | |

Data Download Date/Time: 06/10/2021 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,ALEXANDRIA,N
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HRADECKY,ALEXANDRIA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 07:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: VINCENT

DATE/TIME OF TEST: 06/10/2021 08:47:23

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 08:40 |
| Air Blank | 0.000 | 08:41 |
| Control Test | 0.078 | 08:41 |
| Air Blank | 0.000 | 08:42 |
| Subject Sample # 1 | 0.121 | 08:42 |
| Breath Volume | 1.984 | 08:42 |
| Air Blank | 0.000 | 08:43 |
| Air Blank | 0.000 | 08:45 |
| Subject Sample # 2 | 0.118 | 08:45 |
| Breath Volume | 1.824 | 08:45 |
| Air Blank | 0.000 | 08:46 |
| Control Test | 0.078 | 08:46 |
| Air Blank | 0.000 | 08:47 |
| Diagnostics Check | OK | 08:47 |

Data Download Date/Time: 06/10/2021 10:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: GENIE

DATE/TIME OF TEST: 06/11/2021 22:29:54

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 22:25 | |
| Air Blank | 0.000 | 22:26 | |
| Control Test | 0.077 | 22:26 | |
| Air Blank | 0.000 | 22:27 | |
| Subject Sample | # 1 | 22:28 | Subject Test Refused |
| Breath Volume | 0.000 | 22:28 | |
| Air Blank | 0.000 | 22:28 | |
| Control Test | 0.078 | 22:29 | |
| Air Blank | 0.000 | 22:29 | |
| Diagnostics Check | OK | 22:29 | |

Data Download Date/Time: 07/01/2021 17:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: MCGARTHWAITE

DATE/TIME OF TEST: 06/12/2021 23:13:34

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 23:07 |
| Air Blank | | 0.000 | 23:07 |
| Control Test | | 0.078 | 23:07 |
| Air Blank | | 0.000 | 23:08 |
| Subject Sample | # 1 | 0.115 | 23:08 |
| Breath Volume | | 1.730 | 23:08 |
| Air Blank | | 0.000 | 23:09 |
| Air Blank | | 0.000 | 23:11 |
| Subject Sample | # 2 | 0.122 | 23:12 |
| Breath Volume | | 1.527 | 23:12 |
| Air Blank | | 0.000 | 23:12 |
| Control Test | | 0.078 | 23:12 |
| Air Blank | | 0.000 | 23:13 |
| Diagnostics Check | | OK | 23:13 |

Data Download Date/Time: 07/01/2021 17:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 06/13/2021 01:19:59

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 01:16 | |
| Air Blank | 0.000 | 01:17 | |
| Control Test | 0.078 | 01:17 | |
| Air Blank | 0.000 | 01:18 | |
| Subject Sample | # 1 | 01:18 | Subject Test Refused |
| Breath Volume | 0.000 | 01:18 | |
| Air Blank | 0.000 | 01:19 | |
| Control Test | 0.078 | 01:19 | |
| Air Blank | 0.000 | 01:19 | |
| Diagnostics Check | OK | 01:19 | |

Data Download Date/Time: 07/01/2021 17:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/13/2021 21:37:51

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 21:34 | |
| Air Blank | 0.000 | 21:35 | |
| Control Test | 0.078 | 21:35 | |
| Air Blank | 0.000 | 21:36 | |
| Subject Sample | # 1 | 21:36 | Subject Test Refused |
| Breath Volume | 0.000 | 21:36 | |
| Air Blank | 0.000 | 21:36 | |
| Control Test | 0.079 | 21:37 | |
| Air Blank | 0.000 | 21:37 | |
| Diagnostics Check | OK | 21:37 | |

Data Download Date/Time: 07/01/2021 17:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUNA,ANGEL,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: MORENO,JORGE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: MASSIE

DATE/TIME OF TEST: 06/15/2021 20:17:16

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 20:13 | |
| Air Blank | 0.000 | 20:13 | |
| Control Test | 0.079 | 20:13 | |
| Air Blank | 0.000 | 20:14 | |
| Subject Sample | # 1 | 20:15 | Subject Test Refused |
| Breath Volume | 0.000 | 20:15 | |
| Air Blank | 0.000 | 20:16 | |
| Control Test | 0.079 | 20:16 | |
| Air Blank | 0.000 | 20:17 | |
| Diagnostics Check | OK | 20:17 | |

Data Download Date/Time: 07/01/2021 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: UDINK

DATE/TIME OF TEST: 06/17/2021 00:50:14

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 00:46 | |
| Air Blank | 0.000 | 00:47 | |
| Control Test | 0.078 | 00:47 | |
| Air Blank | 0.000 | 00:47 | |
| Subject Sample | # 1 | 00:48 | Subject Test Refused |
| Breath Volume | 0.000 | 00:48 | |
| Air Blank | 0.000 | 00:49 | |
| Control Test | 0.079 | 00:49 | |
| Air Blank | 0.000 | 00:50 | |
| Diagnostics Check | OK | 00:50 | |

Data Download Date/Time: 07/01/2021 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: LANDA,LAUREN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/17/2021 01:56:07

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:48 |
| Air Blank | 0.000 | 01:49 |
| Control Test | 0.079 | 01:49 |
| Air Blank | 0.000 | 01:49 |
| Subject Sample # 1 | 0.235 | 01:50 |
| Breath Volume | 1.585 | 01:50 |
| Air Blank | 0.000 | 01:51 |
| Air Blank | 0.000 | 01:53 |
| Subject Sample # 2 | 0.217 | 01:54 |
| Breath Volume | 1.207 | 01:54 |
| Air Blank | 0.000 | 01:55 |
| Control Test | 0.076 | 01:55 |
| Air Blank | 0.000 | 01:56 |
| Diagnostics Check | OK | 01:56 |

Data Download Date/Time: 07/01/2021 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,ALEXANDRIA,N
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HRADECKY,ALEXANDRIA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 07:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:01

SUBJECT'S LAST NAME: HASSAN

DATE/TIME OF TEST: 06/19/2021 08:27:56

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 08:21 |
| Air Blank | 0.000 | 08:21 |
| Control Test | 0.079 | 08:22 |
| Air Blank | 0.000 | 08:22 |
| Subject Sample # 1 | 0.000 | 08:23 |
| Breath Volume | 3.347 | 08:23 |
| Air Blank | 0.000 | 08:23 |
| Air Blank | 0.000 | 08:25 |
| Subject Sample # 2 | 0.000 | 08:26 |
| Breath Volume | 3.031 | 08:26 |
| Air Blank | 0.000 | 08:27 |
| Control Test | 0.079 | 08:27 |
| Air Blank | 0.000 | 08:27 |
| Diagnostics Check | OK | 08:27 |

Data Download Date/Time: 07/01/2021 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTHEWS,CASEY,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: ALFARO

DATE/TIME OF TEST: 06/20/2021 04:41:55

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 04:38 | |
| Air Blank | 0.000 | 04:39 | |
| Control Test | 0.078 | 04:39 | |
| Air Blank | 0.000 | 04:40 | |
| Subject Sample | # 1 | 04:40 | Subject Test Refused |
| Breath Volume | 0.000 | 04:40 | |
| Air Blank | 0.000 | 04:40 | |
| Control Test | 0.078 | 04:41 | |
| Air Blank | 0.000 | 04:41 | |
| Diagnostics Check | OK | 04:41 | |

Data Download Date/Time: 07/01/2021 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTHEWS,CASEY,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: HARTLEY

DATE/TIME OF TEST: 06/25/2021 02:17:39

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 02:14 | |
| Air Blank | 0.000 | 02:15 | |
| Control Test | 0.079 | 02:15 | |
| Air Blank | 0.000 | 02:16 | |
| Subject Sample | # 1 | 02:16 | Subject Test Refused |
| Breath Volume | 0.000 | 02:16 | |
| Air Blank | 0.000 | 02:16 | |
| Control Test | 0.078 | 02:17 | |
| Air Blank | 0.000 | 02:17 | |
| Diagnostics Check | OK | 02:17 | |

Data Download Date/Time: 07/01/2021 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HANNA,WARDELL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: BODKIN

DATE/TIME OF TEST: 06/25/2021 05:06:45

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 05:03 | |
| Air Blank | 0.000 | 05:04 | |
| Control Test | 0.079 | 05:04 | |
| Air Blank | 0.000 | 05:05 | |
| Subject Sample | # 1 | 05:05 | Subject Test Refused |
| Breath Volume | 0.000 | 05:05 | |
| Air Blank | 0.000 | 05:05 | |
| Control Test | 0.078 | 05:06 | |
| Air Blank | 0.000 | 05:06 | |
| Diagnostics Check | OK | 05:06 | |

Data Download Date/Time: 07/01/2021 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,CRISTIAN,X
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: SANTOS,CRISTIAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: ROSENWALD

DATE/TIME OF TEST: 06/26/2021 05:47:41

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 05:40 |
| Air Blank | | 0.000 | 05:41 |
| Control Test | | 0.079 | 05:41 |
| Air Blank | | 0.000 | 05:42 |
| Subject Sample | # 1 | 0.104 | 05:43 |
| Breath Volume | | 2.289 | 05:43 |
| Air Blank | | 0.000 | 05:43 |
| Air Blank | | 0.000 | 05:45 |
| Subject Sample | # 2 | 0.099 | 05:46 |
| Breath Volume | | 1.597 | 05:46 |
| Air Blank | | 0.000 | 05:46 |
| Control Test | | 0.078 | 05:47 |
| Air Blank | | 0.000 | 05:47 |
| Diagnostics Check | | OK | 05:47 |

Data Download Date/Time: 07/01/2021 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: TAMAYO

DATE/TIME OF TEST: 06/27/2021 02:23:44

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 02:20 | |
| Air Blank | 0.000 | 02:21 | |
| Control Test | 0.079 | 02:21 | |
| Air Blank | 0.000 | 02:21 | |
| Subject Sample | # 1 | 02:22 | Subject Test Refused |
| Breath Volume | 0.000 | 02:22 | |
| Air Blank | 0.000 | 02:22 | |
| Control Test | 0.079 | 02:23 | |
| Air Blank | 0.000 | 02:23 | |
| Diagnostics Check | OK | 02:23 | |

Data Download Date/Time: 07/01/2021 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,L
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: LUDGATE

DATE/TIME OF TEST: 06/27/2021 18:07:31

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 18:01 |
| Air Blank | 0.000 | 18:01 |
| Control Test | 0.078 | 18:01 |
| Air Blank | 0.000 | 18:02 |
| Subject Sample # 1 | 0.146 | 18:02 |
| Breath Volume | 2.328 | 18:02 |
| Air Blank | 0.000 | 18:03 |
| Air Blank | 0.000 | 18:05 |
| Subject Sample # 2 | 0.145 | 18:05 |
| Breath Volume | 1.523 | 18:05 |
| Air Blank | 0.000 | 18:06 |
| Control Test | 0.079 | 18:06 |
| Air Blank | 0.000 | 18:07 |
| Diagnostics Check | OK | 18:07 |

Data Download Date/Time: 07/01/2021 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 11:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: DE LA VEGA

DATE/TIME OF TEST: 06/28/2021 00:41:07

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:34 |
| Air Blank | 0.000 | 00:35 |
| Control Test | 0.078 | 00:35 |
| Air Blank | 0.000 | 00:35 |
| Subject Sample # 1 | 0.169 | 00:36 |
| Breath Volume | 1.347 | 00:36 |
| Air Blank | 0.000 | 00:37 |
| Air Blank | 0.000 | 00:38 |
| Subject Sample # 2 | 0.164 | 00:39 |
| Breath Volume | 1.722 | 00:39 |
| Air Blank | 0.000 | 00:40 |
| Control Test | 0.077 | 00:40 |
| Air Blank | 0.000 | 00:41 |
| Diagnostics Check | OK | 00:41 |

Data Download Date/Time: 07/01/2021 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MORENO,JORGE
ARRESTING OFFICER AGENCY: BAY COUNTY SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 06:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: KMAK

DATE/TIME OF TEST: 06/28/2021 19:36:22

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 19:30 |
| Air Blank | 0.000 | 19:30 |
| Control Test | 0.079 | 19:30 |
| Air Blank | 0.000 | 19:31 |
| Subject Sample # 1 | 0.016 | 19:31 |
| Breath Volume | 2.046 | 19:31 |
| Air Blank | 0.000 | 19:32 |
| Air Blank | 0.000 | 19:34 |
| Subject Sample # 2 | 0.015 | 19:34 |
| Breath Volume | 1.625 | 19:34 |
| Air Blank | 0.000 | 19:35 |
| Control Test | 0.078 | 19:35 |
| Air Blank | 0.000 | 19:36 |
| Diagnostics Check | OK | 19:36 |

Data Download Date/Time: 07/01/2021 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUNA,ANGEL,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: LUNA,ANGEL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: BAEZ

DATE/TIME OF TEST: 06/28/2021 19:52:46

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 19:49 | |
| Air Blank | 0.000 | 19:50 | |
| Control Test | 0.079 | 19:50 | |
| Air Blank | 0.000 | 19:51 | |
| Subject Sample | # 1 | 19:51 | Subject Test Refused |
| Breath Volume | 0.000 | 19:51 | |
| Air Blank | 0.000 | 19:51 | |
| Control Test | 0.079 | 19:52 | |
| Air Blank | 0.000 | 19:52 | |
| Diagnostics Check | OK | 19:52 | |

Data Download Date/Time: 07/01/2021 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HANNA,WARDELL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 12:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: NICADO HERNANDEZ

DATE/TIME OF TEST: 06/29/2021 00:54:37

| | | | |
|-------------------|----|-------|------------|
| Diagnostics Check | OK | 00:54 | |
| Air Blank | | 00:54 | RFI Detect |
| Air Blank | | 00:54 | RFI Detect |

Data Download Date/Time: 07/01/2021 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HANNA,WARDELL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: NICADO HERNANDEZ

DATE/TIME OF TEST: 06/29/2021 01:04:30

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 00:57 |
| Air Blank | | 0.000 | 00:57 |
| Control Test | | 0.079 | 00:57 |
| Air Blank | | 0.000 | 00:58 |
| Subject Sample | # 1 | 0.087 | 01:00 |
| Breath Volume | | 1.792 | 01:00 |
| Air Blank | | 0.000 | 01:00 |
| Air Blank | | 0.000 | 01:02 |
| Subject Sample | # 2 | 0.086 | 01:02 |
| Breath Volume | | 1.800 | 01:02 |
| Air Blank | | 0.000 | 01:03 |
| Control Test | | 0.078 | 01:03 |
| Air Blank | | 0.000 | 01:04 |
| Diagnostics Check | | OK | 01:04 |

Data Download Date/Time: 07/01/2021 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,JORGE,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: CARILLO JIMENEZ

DATE/TIME OF TEST: 06/29/2021 19:09:15

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 19:06 | |
| Air Blank | 0.000 | 19:06 | |
| Control Test | 0.079 | 19:07 | |
| Air Blank | 0.000 | 19:07 | |
| Subject Sample | # 1 | 19:07 | Subject Test Refused |
| Breath Volume | 0.000 | 19:07 | |
| Air Blank | 0.000 | 19:08 | |
| Control Test | 0.078 | 19:08 | |
| Air Blank | 0.000 | 19:09 | |
| Diagnostics Check | OK | 19:09 | |

Data Download Date/Time: 07/01/2021 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MATTHEWS,CASEY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: NOLAN

DATE/TIME OF TEST: 06/29/2021 23:02:36

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:55 |
| Air Blank | 0.000 | 22:55 |
| Control Test | 0.079 | 22:56 |
| Air Blank | 0.000 | 22:56 |
| Subject Sample # 1 | 0.000 | 22:58 |
| Breath Volume | 1.699 | 22:58 |
| Air Blank | 0.000 | 22:58 |
| Air Blank | 0.000 | 23:00 |
| Subject Sample # 2 | 0.000 | 23:01 |
| Breath Volume | 1.355 | 23:01 |
| Air Blank | 0.000 | 23:01 |
| Control Test | 0.079 | 23:01 |
| Air Blank | 0.000 | 23:02 |
| Diagnostics Check | OK | 23:02 |

Data Download Date/Time: 07/01/2021 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGGAN,MARLON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOGGAN,MARLON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: ELCOOK GOBER

DATE/TIME OF TEST: 06/12/2021 18:51:54

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 18:45 |
| Air Blank | 0.000 | 18:45 |
| Control Test | 0.078 | 18:46 |
| Air Blank | 0.000 | 18:46 |
| Subject Sample # 1 | 0.159 | 18:47 |
| Breath Volume | 1.828 | 18:47 |
| Air Blank | 0.000 | 18:48 |
| Air Blank | 0.000 | 18:49 |
| Subject Sample # 2 | 0.156 | 18:50 |
| Breath Volume | 1.492 | 18:50 |
| Air Blank | 0.000 | 18:50 |
| Control Test | 0.077 | 18:51 |
| Air Blank | 0.000 | 18:51 |
| Diagnostics Check | OK | 18:51 |

Data Download Date/Time: 06/25/2021 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'DOR,ROBERT,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: SANSONE

DATE/TIME OF TEST: 06/26/2021 06:00:09

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 05:48 | |
| Air Blank | | 0.000 | 05:49 | |
| Control Test | | 0.078 | 05:49 | |
| Air Blank | | 0.000 | 05:50 | |
| Subject Sample | # 1 | 0.000 | 05:53 | Volume Not Met |
| Breath Volume | | 0.542 | 05:53 | |
| Air Blank | | 0.000 | 05:53 | |
| Air Blank | | 0.000 | 05:55 | |
| Subject Sample | # 2 | 0.000 | 05:58 | Volume Not Met |
| Breath Volume | | 0.347 | 05:58 | |
| Air Blank | | 0.000 | 05:59 | |
| Control Test | | 0.077 | 05:59 | |
| Air Blank | | 0.000 | 06:00 | |
| Diagnostics Check | | OK | 06:00 | |

Data Download Date/Time: 07/14/2021 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GIBSON,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 06/01/2021 16:57:43

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 16:50 |
| Air Blank | 0.000 | 16:51 |
| Control Test | 0.080 | 16:51 |
| Air Blank | 0.000 | 16:52 |
| Subject Sample # 1 | 0.141 | 16:52 |
| Breath Volume | 2.414 | 16:52 |
| Air Blank | 0.000 | 16:53 |
| Air Blank | 0.000 | 16:55 |
| Subject Sample # 2 | 0.134 | 16:55 |
| Breath Volume | 2.949 | 16:55 |
| Air Blank | 0.000 | 16:56 |
| Control Test | 0.078 | 16:57 |
| Air Blank | 0.000 | 16:57 |
| Diagnostics Check | OK | 16:57 |

Data Download Date/Time: 06/11/2021 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KUBIAK, KRISTOPHER
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: YURKUS

DATE/TIME OF TEST: 06/02/2021 02:56:13

| | | | | |
|-------------------|-----|-------|-------|---------------------------|
| Diagnostics Check | | OK | 02:44 | |
| Air Blank | | 0.000 | 02:45 | |
| Control Test | | 0.080 | 02:45 | |
| Air Blank | | 0.000 | 02:46 | |
| Subject Sample | # 1 | 0.233 | 02:49 | Volume Not Met |
| Breath Volume | | 0.050 | 02:49 | |
| Air Blank | | 0.000 | 02:50 | |
| Air Blank | | 0.000 | 02:51 | |
| Subject Sample | # 2 | | 02:54 | Subject Test Refused |
| Breath Volume | | 0.132 | 02:54 | |
| Air Blank | | 0.000 | 02:55 | |
| Control Test | | 0.073 | 02:55 | Control Outside Tolerance |
| Air Blank | | 0.000 | 02:56 | |

Data Download Date/Time: 06/11/2021 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR,NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: FACTOR

DATE/TIME OF TEST: 06/03/2021 23:17:00

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:09 |
| Air Blank | 0.000 | 23:09 |
| Control Test | 0.081 | 23:10 |
| Air Blank | 0.000 | 23:10 |
| Subject Sample # 1 | 0.225 | 23:11 |
| Breath Volume | 1.296 | 23:11 |
| Air Blank | 0.000 | 23:12 |
| Air Blank | 0.000 | 23:14 |
| Subject Sample # 2 | 0.219 | 23:15 |
| Breath Volume | 1.609 | 23:15 |
| Air Blank | 0.000 | 23:16 |
| Control Test | 0.077 | 23:16 |
| Air Blank | 0.000 | 23:16 |
| Diagnostics Check | OK | 23:17 |

Data Download Date/Time: 06/11/2021 14:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: ANDRES

DATE/TIME OF TEST: 06/04/2021 23:46:21

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:39 |
| Air Blank | 0.000 | 23:40 |
| Control Test | 0.080 | 23:40 |
| Air Blank | 0.000 | 23:41 |
| Subject Sample # 1 | 0.000 | 23:41 |
| Breath Volume | 3.332 | 23:41 |
| Air Blank | 0.000 | 23:42 |
| Air Blank | 0.000 | 23:44 |
| Subject Sample # 2 | 0.000 | 23:44 |
| Breath Volume | 2.867 | 23:44 |
| Air Blank | 0.000 | 23:45 |
| Control Test | 0.079 | 23:45 |
| Air Blank | 0.000 | 23:46 |
| Diagnostics Check | OK | 23:46 |

Data Download Date/Time: 06/11/2021 14:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TEJEDA,ANTHONY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: HORN

DATE/TIME OF TEST: 06/05/2021 02:19:13

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 02:08 |
| Air Blank | | 0.000 | 02:08 |
| Control Test | | 0.080 | 02:08 |
| Air Blank | | 0.000 | 02:09 |
| Subject Sample | # 1 | 0.229 | 02:10 |
| Breath Volume | | 2.402 | 02:10 |
| Air Blank | | 0.000 | 02:11 |
| Air Blank | | 0.000 | 02:12 |
| Subject Sample | # 2 | 0.192 | 02:13 |
| Breath Volume | | 1.265 | 02:13 |
| Air Blank | | 0.000 | 02:14 |
| Air Blank | | 0.000 | 02:15 |
| Subject Sample | # 3 | 0.232 | 02:17 |
| Breath Volume | | 2.339 | 02:17 |
| Air Blank | | 0.000 | 02:18 |
| Control Test | | 0.078 | 02:18 |
| Air Blank | | 0.000 | 02:19 |
| Diagnostics Check | | OK | 02:19 |

Data Download Date/Time: 06/11/2021 14:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KYSOR,MITCHELL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: ZABALA

DATE/TIME OF TEST: 06/06/2021 00:53:14

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:46 |
| Air Blank | 0.000 | 00:46 |
| Control Test | 0.080 | 00:47 |
| Air Blank | 0.000 | 00:47 |
| Subject Sample # 1 | 0.241 | 00:48 |
| Breath Volume | 1.867 | 00:48 |
| Air Blank | 0.000 | 00:49 |
| Air Blank | 0.000 | 00:50 |
| Subject Sample # 2 | 0.246 | 00:51 |
| Breath Volume | 2.214 | 00:51 |
| Air Blank | 0.000 | 00:52 |
| Control Test | 0.078 | 00:52 |
| Air Blank | 0.000 | 00:53 |
| Diagnostics Check | OK | 00:53 |

Data Download Date/Time: 06/11/2021 14:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR,NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: BRAVO

DATE/TIME OF TEST: 06/06/2021 01:26:50

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:16 |
| Air Blank | 0.000 | 01:16 |
| Control Test | 0.078 | 01:17 |
| Air Blank | 0.000 | 01:17 |
| Subject Sample # 1 | 0.126 | 01:20 |
| Breath Volume | 1.433 | 01:20 |
| Air Blank | 0.000 | 01:21 |
| Air Blank | 0.000 | 01:22 |
| Subject Sample # 2 | 0.137 | 01:24 |
| Breath Volume | 2.179 | 01:24 |
| Air Blank | 0.000 | 01:25 |
| Control Test | 0.077 | 01:26 |
| Air Blank | 0.000 | 01:26 |
| Diagnostics Check | OK | 01:26 |

Data Download Date/Time: 06/11/2021 14:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PARENT,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 11:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: BARR

DATE/TIME OF TEST: 06/06/2021 13:08:25

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 13:02 |
| Air Blank | 0.000 | 13:02 |
| Control Test | 0.080 | 13:02 |
| Air Blank | 0.000 | 13:03 |
| Subject Sample # 1 | 0.000 | 13:04 |
| Breath Volume | 1.281 | 13:04 |
| Air Blank | 0.000 | 13:04 |
| Air Blank | 0.000 | 13:06 |
| Subject Sample # 2 | 0.000 | 13:06 |
| Breath Volume | 2.203 | 13:06 |
| Air Blank | 0.000 | 13:07 |
| Control Test | 0.080 | 13:07 |
| Air Blank | 0.000 | 13:08 |
| Diagnostics Check | OK | 13:08 |

Data Download Date/Time: 06/11/2021 14:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DARBY,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 06/06/2021 15:04:12

| | | | |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK | 14:59 | |
| Air Blank | 0.000 | 15:00 | |
| Control Test | 0.079 | 15:00 | |
| Air Blank | 0.000 | 15:01 | |
| Subject Sample | # 1 | 15:02 | Interferent Detect |
| Breath Volume | 3.953 | 15:02 | |
| Air Blank | 0.000 | 15:03 | |
| Control Test | 0.079 | 15:03 | |
| Air Blank | 0.000 | 15:04 | |
| Diagnostics Check | OK | 15:04 | |

Data Download Date/Time: 06/11/2021 14:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DARBY,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 06/06/2021 15:11:39

| | | | |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK | 15:07 | |
| Air Blank | 0.000 | 15:08 | |
| Control Test | 0.079 | 15:08 | |
| Air Blank | 0.000 | 15:08 | |
| Subject Sample | # 1 | 15:09 | Interferent Detect |
| Breath Volume | 3.726 | 15:09 | |
| Air Blank | 0.000 | 15:10 | |
| Control Test | 0.079 | 15:10 | |
| Air Blank | 0.000 | 15:11 | |
| Diagnostics Check | OK | 15:11 | |

Data Download Date/Time: 06/11/2021 14:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BUTZBACH,WILLIAM
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 06/06/2021 18:10:10

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 18:01 |
| Air Blank | 0.000 | 18:02 |
| Control Test | 0.080 | 18:02 |
| Air Blank | 0.000 | 18:02 |
| Subject Sample # 1 | 0.344 | 18:04 |
| Breath Volume | 1.343 | 18:04 |
| Air Blank | 0.000 | 18:05 |
| Air Blank | 0.000 | 18:06 |
| Subject Sample # 2 | 0.336 | 18:08 |
| Breath Volume | 1.953 | 18:08 |
| Air Blank | 0.000 | 18:09 |
| Control Test | 0.076 | 18:09 |
| Air Blank | 0.000 | 18:10 |
| Diagnostics Check | OK | 18:10 |

Data Download Date/Time: 06/11/2021 14:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TEJEDA,ANTHONY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: TIBBETTS

DATE/TIME OF TEST: 06/06/2021 21:39:27

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:30 |
| Air Blank | 0.000 | 21:30 |
| Control Test | 0.079 | 21:31 |
| Air Blank | 0.000 | 21:31 |
| Subject Sample # 1 | 0.137 | 21:34 |
| Breath Volume | 1.527 | 21:34 |
| Air Blank | 0.000 | 21:35 |
| Air Blank | 0.000 | 21:37 |
| Subject Sample # 2 | 0.137 | 21:37 |
| Breath Volume | 2.332 | 21:37 |
| Air Blank | 0.000 | 21:38 |
| Control Test | 0.077 | 21:38 |
| Air Blank | 0.000 | 21:39 |
| Diagnostics Check | OK | 21:39 |

Data Download Date/Time: 06/11/2021 14:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CASTINEIRAS,MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: HOSFELD

DATE/TIME OF TEST: 06/07/2021 03:52:10

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:43 |
| Air Blank | 0.000 | 03:44 |
| Control Test | 0.079 | 03:44 |
| Air Blank | 0.000 | 03:45 |
| Subject Sample # 1 | 0.183 | 03:47 |
| Breath Volume | 2.394 | 03:47 |
| Air Blank | 0.000 | 03:48 |
| Air Blank | 0.000 | 03:49 |
| Subject Sample # 2 | 0.184 | 03:50 |
| Breath Volume | 3.117 | 03:50 |
| Air Blank | 0.000 | 03:51 |
| Control Test | 0.079 | 03:51 |
| Air Blank | 0.000 | 03:52 |
| Diagnostics Check | OK | 03:52 |

Data Download Date/Time: 06/11/2021 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: HIGGINS

DATE/TIME OF TEST: 06/07/2021 04:09:50

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 04:03 |
| Air Blank | 0.000 | 04:04 |
| Control Test | 0.078 | 04:04 |
| Air Blank | 0.000 | 04:04 |
| Subject Sample # 1 | 0.000 | 04:05 |
| Breath Volume | 2.316 | 04:05 |
| Air Blank | 0.000 | 04:05 |
| Air Blank | 0.000 | 04:07 |
| Subject Sample # 2 | 0.000 | 04:08 |
| Breath Volume | 2.707 | 04:08 |
| Air Blank | 0.000 | 04:08 |
| Control Test | 0.078 | 04:09 |
| Air Blank | 0.000 | 04:09 |
| Diagnostics Check | OK | 04:09 |

Data Download Date/Time: 06/11/2021 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: KURYNKA

DATE/TIME OF TEST: 06/07/2021 05:55:50

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 05:49 |
| Air Blank | 0.000 | 05:50 |
| Control Test | 0.079 | 05:50 |
| Air Blank | 0.000 | 05:50 |
| Subject Sample # 1 | 0.000 | 05:51 |
| Breath Volume | 1.804 | 05:51 |
| Air Blank | 0.000 | 05:51 |
| Air Blank | 0.000 | 05:53 |
| Subject Sample # 2 | 0.000 | 05:54 |
| Breath Volume | 2.476 | 05:54 |
| Air Blank | 0.000 | 05:54 |
| Control Test | 0.079 | 05:55 |
| Air Blank | 0.000 | 05:55 |
| Diagnostics Check | OK | 05:55 |

Data Download Date/Time: 06/11/2021 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VARGAS,CHRISTIAN
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: KOPRA

DATE/TIME OF TEST: 06/07/2021 13:38:43

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 13:30 |
| Air Blank | 0.000 | 13:31 |
| Control Test | 0.080 | 13:31 |
| Air Blank | 0.000 | 13:32 |
| Subject Sample # 1 | 0.124 | 13:33 |
| Breath Volume | 1.449 | 13:33 |
| Air Blank | 0.000 | 13:34 |
| Air Blank | 0.000 | 13:35 |
| Subject Sample # 2 | 0.128 | 13:36 |
| Breath Volume | 2.230 | 13:36 |
| Air Blank | 0.000 | 13:37 |
| Control Test | 0.079 | 13:38 |
| Air Blank | 0.000 | 13:38 |
| Diagnostics Check | OK | 13:38 |

Data Download Date/Time: 06/11/2021 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: THOMAS,MICAH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/08/2021 00:27:11

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:20 |
| Air Blank | 0.000 | 00:21 |
| Control Test | 0.080 | 00:21 |
| Air Blank | 0.000 | 00:21 |
| Subject Sample # 1 | 0.163 | 00:22 |
| Breath Volume | 2.050 | 00:22 |
| Air Blank | 0.000 | 00:23 |
| Air Blank | 0.000 | 00:24 |
| Subject Sample # 2 | 0.165 | 00:25 |
| Breath Volume | 2.277 | 00:25 |
| Air Blank | 0.000 | 00:26 |
| Control Test | 0.079 | 00:26 |
| Air Blank | 0.000 | 00:27 |
| Diagnostics Check | OK | 00:27 |

Data Download Date/Time: 06/11/2021 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MARCHESI II

DATE/TIME OF TEST: 06/08/2021 02:16:27

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:10 |
| Air Blank | 0.000 | 02:10 |
| Control Test | 0.080 | 02:11 |
| Air Blank | 0.000 | 02:11 |
| Subject Sample # 1 | 0.000 | 02:12 |
| Breath Volume | 3.285 | 02:12 |
| Air Blank | 0.000 | 02:12 |
| Air Blank | 0.000 | 02:14 |
| Subject Sample # 2 | 0.000 | 02:15 |
| Breath Volume | 2.648 | 02:15 |
| Air Blank | 0.000 | 02:15 |
| Control Test | 0.080 | 02:15 |
| Air Blank | 0.000 | 02:16 |
| Diagnostics Check | OK | 02:16 |

Data Download Date/Time: 06/11/2021 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALEZ,MARQUIS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: SCHUMACHER

DATE/TIME OF TEST: 06/08/2021 05:09:02

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 05:01 |
| Air Blank | | 0.000 | 05:02 |
| Control Test | | 0.079 | 05:02 |
| Air Blank | | 0.000 | 05:03 |
| Subject Sample | # 1 | 0.211 | 05:03 |
| Breath Volume | | 1.152 | 05:03 |
| Air Blank | | 0.000 | 05:04 |
| Air Blank | | 0.000 | 05:06 |
| Subject Sample | # 2 | 0.199 | 05:07 |
| Breath Volume | | 1.718 | 05:07 |
| Air Blank | | 0.000 | 05:08 |
| Control Test | | 0.079 | 05:08 |
| Air Blank | | 0.000 | 05:08 |
| Diagnostics Check | | OK | 05:09 |

Data Download Date/Time: 06/11/2021 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESANTIS,BRIAN
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 11:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:55

SUBJECT'S LAST NAME: GULIANO

DATE/TIME OF TEST: 06/08/2021 12:30:12

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 12:23 |
| Air Blank | 0.000 | 12:23 |
| Control Test | 0.080 | 12:24 |
| Air Blank | 0.000 | 12:24 |
| Subject Sample # 1 | 0.109 | 12:25 |
| Breath Volume | 1.652 | 12:25 |
| Air Blank | 0.000 | 12:26 |
| Air Blank | 0.000 | 12:27 |
| Subject Sample # 2 | 0.111 | 12:28 |
| Breath Volume | 1.617 | 12:28 |
| Air Blank | 0.000 | 12:29 |
| Control Test | 0.079 | 12:29 |
| Air Blank | 0.000 | 12:30 |
| Diagnostics Check | OK | 12:30 |

Data Download Date/Time: 06/11/2021 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MARTINEZ,JACOB
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: VELASQUEZ LOPEZ

DATE/TIME OF TEST: 06/08/2021 20:19:29

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:10 |
| Air Blank | 0.000 | 20:11 |
| Control Test | 0.081 | 20:11 |
| Air Blank | 0.000 | 20:11 |
| Subject Sample # 1 | 0.247 | 20:14 |
| Breath Volume | 1.964 | 20:14 |
| Air Blank | 0.000 | 20:15 |
| Air Blank | 0.000 | 20:16 |
| Subject Sample # 2 | 0.243 | 20:17 |
| Breath Volume | 2.839 | 20:17 |
| Air Blank | 0.000 | 20:18 |
| Control Test | 0.076 | 20:18 |
| Air Blank | 0.000 | 20:19 |
| Diagnostics Check | OK | 20:19 |

Data Download Date/Time: 06/11/2021 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GOLDEN,LISA
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: MCALLISTER

DATE/TIME OF TEST: 06/08/2021 20:49:50

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:43 |
| Air Blank | 0.000 | 20:43 |
| Control Test | 0.078 | 20:43 |
| Air Blank | 0.000 | 20:44 |
| Subject Sample # 1 | 0.000 | 20:45 |
| Breath Volume | 3.617 | 20:45 |
| Air Blank | 0.000 | 20:45 |
| Air Blank | 0.000 | 20:47 |
| Subject Sample # 2 | 0.000 | 20:48 |
| Breath Volume | 3.933 | 20:48 |
| Air Blank | 0.000 | 20:48 |
| Control Test | 0.079 | 20:49 |
| Air Blank | 0.000 | 20:49 |
| Diagnostics Check | OK | 20:49 |

Data Download Date/Time: 06/11/2021 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCGILLICUDDY, STEVEN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: MERLIN

DATE/TIME OF TEST: 06/09/2021 01:41:48

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:34 |
| Air Blank | 0.000 | 01:35 |
| Control Test | 0.080 | 01:35 |
| Air Blank | 0.000 | 01:36 |
| Subject Sample # 1 | 0.000 | 01:37 |
| Breath Volume | 2.742 | 01:37 |
| Air Blank | 0.000 | 01:37 |
| Air Blank | 0.000 | 01:39 |
| Subject Sample # 2 | 0.000 | 01:40 |
| Breath Volume | 1.535 | 01:40 |
| Air Blank | 0.000 | 01:40 |
| Control Test | 0.079 | 01:41 |
| Air Blank | 0.000 | 01:41 |
| Diagnostics Check | OK | 01:41 |

Data Download Date/Time: 06/11/2021 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR, NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PEEK

DATE/TIME OF TEST: 06/10/2021 02:03:31

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:55 |
| Air Blank | 0.000 | 01:55 |
| Control Test | 0.080 | 01:56 |
| Air Blank | 0.000 | 01:56 |
| Subject Sample # 1 | 0.203 | 01:57 |
| Breath Volume | 1.445 | 01:57 |
| Air Blank | 0.000 | 01:58 |
| Air Blank | 0.000 | 02:00 |
| Subject Sample # 2 | 0.188 | 02:01 |
| Breath Volume | 1.160 | 02:01 |
| Air Blank | 0.000 | 02:02 |
| Control Test | 0.077 | 02:02 |
| Air Blank | 0.000 | 02:03 |
| Diagnostics Check | OK | 02:03 |

Data Download Date/Time: 06/11/2021 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCCOURT, GREGORY
ARRESTING OFFICER AGENCY: Atlantis PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: PIZARRO JR

DATE/TIME OF TEST: 06/10/2021 05:58:04

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 05:50 |
| Air Blank | 0.000 | 05:50 |
| Control Test | 0.078 | 05:51 |
| Air Blank | 0.000 | 05:51 |
| Subject Sample # 1 | 0.284 | 05:53 |
| Breath Volume | 2.574 | 05:53 |
| Air Blank | 0.000 | 05:54 |
| Air Blank | 0.000 | 05:55 |
| Subject Sample # 2 | 0.276 | 05:56 |
| Breath Volume | 2.687 | 05:56 |
| Air Blank | 0.000 | 05:57 |
| Control Test | 0.076 | 05:57 |
| Air Blank | 0.000 | 05:57 |
| Diagnostics Check | OK | 05:58 |

Data Download Date/Time: 06/11/2021 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROCHA,LUIS
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: CULLEN

DATE/TIME OF TEST: 06/10/2021 21:22:02

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:15 |
| Air Blank | 0.000 | 21:16 |
| Control Test | 0.079 | 21:16 |
| Air Blank | 0.000 | 21:17 |
| Subject Sample # 1 | 0.011 | 21:17 |
| Breath Volume | 1.878 | 21:17 |
| Air Blank | 0.000 | 21:18 |
| Air Blank | 0.000 | 21:20 |
| Subject Sample # 2 | 0.011 | 21:20 |
| Breath Volume | 4.003 | 21:20 |
| Air Blank | 0.000 | 21:21 |
| Control Test | 0.079 | 21:21 |
| Air Blank | 0.000 | 21:21 |
| Diagnostics Check | OK | 21:22 |

Data Download Date/Time: 06/11/2021 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANTON, MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: KEMP

DATE/TIME OF TEST: 06/10/2021 22:59:49

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:51 |
| Air Blank | 0.000 | 22:51 |
| Control Test | 0.080 | 22:52 |
| Air Blank | 0.000 | 22:52 |
| Subject Sample # 1 | 0.142 | 22:54 |
| Breath Volume | 3.437 | 22:54 |
| Air Blank | 0.000 | 22:55 |
| Air Blank | 0.000 | 22:57 |
| Subject Sample # 2 | 0.136 | 22:57 |
| Breath Volume | 3.019 | 22:57 |
| Air Blank | 0.000 | 22:58 |
| Control Test | 0.075 | 22:59 |
| Air Blank | 0.000 | 22:59 |
| Diagnostics Check | OK | 22:59 |

Data Download Date/Time: 06/11/2021 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAMPIASI,JAKE
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: GUINERA SEGREA

DATE/TIME OF TEST: 06/11/2021 21:29:36

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:21 |
| Air Blank | 0.000 | 21:22 |
| Control Test | 0.080 | 21:22 |
| Air Blank | 0.000 | 21:23 |
| Subject Sample # 1 | 0.187 | 21:24 |
| Breath Volume | 1.789 | 21:24 |
| Air Blank | 0.000 | 21:25 |
| Air Blank | 0.000 | 21:27 |
| Subject Sample # 2 | 0.193 | 21:27 |
| Breath Volume | 1.746 | 21:27 |
| Air Blank | 0.000 | 21:28 |
| Control Test | 0.079 | 21:28 |
| Air Blank | 0.000 | 21:29 |
| Diagnostics Check | OK | 21:29 |

Data Download Date/Time: 07/16/2021 17:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CASAS, JAVIER
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 06/12/2021 02:23:29

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:16 |
| Air Blank | 0.000 | 02:17 |
| Control Test | 0.079 | 02:17 |
| Air Blank | 0.000 | 02:18 |
| Subject Sample # 1 | 0.000 | 02:19 |
| Breath Volume | 1.320 | 02:19 |
| Air Blank | 0.000 | 02:19 |
| Air Blank | 0.000 | 02:21 |
| Subject Sample # 2 | 0.000 | 02:22 |
| Breath Volume | 1.238 | 02:22 |
| Air Blank | 0.000 | 02:22 |
| Control Test | 0.079 | 02:22 |
| Air Blank | 0.000 | 02:23 |
| Diagnostics Check | OK | 02:23 |

Data Download Date/Time: 07/16/2021 17:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR,NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: DIAZ-ALVAREZ

DATE/TIME OF TEST: 06/12/2021 04:08:01

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 04:01 |
| Air Blank | | 0.000 | 04:01 |
| Control Test | | 0.080 | 04:02 |
| Air Blank | | 0.000 | 04:02 |
| Subject Sample | # 1 | 0.204 | 04:03 |
| Breath Volume | | 1.605 | 04:03 |
| Air Blank | | 0.000 | 04:04 |
| Air Blank | | 0.000 | 04:05 |
| Subject Sample | # 2 | 0.216 | 04:06 |
| Breath Volume | | 1.964 | 04:06 |
| Air Blank | | 0.000 | 04:07 |
| Control Test | | 0.078 | 04:07 |
| Air Blank | | 0.000 | 04:07 |
| Diagnostics Check | | OK | 04:08 |

Data Download Date/Time: 07/16/2021 17:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TEJEDA,ANTHONY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 06/14/2021 03:01:55

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 02:54 |
| Air Blank | | 0.000 | 02:54 |
| Control Test | | 0.080 | 02:54 |
| Air Blank | | 0.000 | 02:55 |
| Subject Sample | # 1 | 0.208 | 02:56 |
| Breath Volume | | 1.316 | 02:56 |
| Air Blank | | 0.000 | 02:57 |
| Air Blank | | 0.000 | 02:58 |
| Subject Sample | # 2 | 0.195 | 02:59 |
| Breath Volume | | 1.371 | 02:59 |
| Air Blank | | 0.000 | 03:00 |
| Control Test | | 0.079 | 03:01 |
| Air Blank | | 0.000 | 03:01 |
| Diagnostics Check | | OK | 03:01 |

Data Download Date/Time: 07/16/2021 17:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: SACALXOT GONZALEZ

DATE/TIME OF TEST: 06/14/2021 22:35:10

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:28 |
| Air Blank | 0.000 | 22:28 |
| Control Test | 0.080 | 22:29 |
| Air Blank | 0.000 | 22:29 |
| Subject Sample # 1 | 0.310 | 22:30 |
| Breath Volume | 1.562 | 22:30 |
| Air Blank | 0.000 | 22:31 |
| Air Blank | 0.000 | 22:32 |
| Subject Sample # 2 | 0.308 | 22:33 |
| Breath Volume | 1.652 | 22:33 |
| Air Blank | 0.000 | 22:34 |
| Control Test | 0.077 | 22:34 |
| Air Blank | 0.000 | 22:35 |
| Diagnostics Check | OK | 22:35 |

Data Download Date/Time: 07/16/2021 17:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: GODINEZ SANCHEZ

DATE/TIME OF TEST: 06/15/2021 01:22:46

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:15 |
| Air Blank | 0.000 | 01:16 |
| Control Test | 0.079 | 01:16 |
| Air Blank | 0.000 | 01:17 |
| Subject Sample # 1 | 0.156 | 01:18 |
| Breath Volume | 1.515 | 01:18 |
| Air Blank | 0.000 | 01:18 |
| Air Blank | 0.000 | 01:20 |
| Subject Sample # 2 | 0.159 | 01:20 |
| Breath Volume | 1.824 | 01:20 |
| Air Blank | 0.000 | 01:21 |
| Control Test | 0.078 | 01:22 |
| Air Blank | 0.000 | 01:22 |
| Diagnostics Check | OK | 01:22 |

Data Download Date/Time: 07/16/2021 17:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BUTZBACH,WILLIAM
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: ORTIZ DIAZ

DATE/TIME OF TEST: 06/16/2021 00:29:33

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 00:22 |
| Air Blank | | 0.000 | 00:22 |
| Control Test | | 0.080 | 00:23 |
| Air Blank | | 0.000 | 00:23 |
| Subject Sample | # 1 | 0.253 | 00:24 |
| Breath Volume | | 2.246 | 00:24 |
| Air Blank | | 0.000 | 00:25 |
| Air Blank | | 0.000 | 00:27 |
| Subject Sample | # 2 | 0.240 | 00:27 |
| Breath Volume | | 1.351 | 00:27 |
| Air Blank | | 0.000 | 00:28 |
| Control Test | | 0.077 | 00:28 |
| Air Blank | | 0.000 | 00:29 |
| Diagnostics Check | | OK | 00:29 |

Data Download Date/Time: 07/16/2021 17:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: CHEATHAM

DATE/TIME OF TEST: 06/18/2021 04:43:37

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 04:37 |
| Air Blank | 0.000 | 04:37 |
| Control Test | 0.080 | 04:37 |
| Air Blank | 0.000 | 04:38 |
| Subject Sample # 1 | 0.000 | 04:39 |
| Breath Volume | 1.972 | 04:39 |
| Air Blank | 0.000 | 04:39 |
| Air Blank | 0.000 | 04:41 |
| Subject Sample # 2 | 0.000 | 04:42 |
| Breath Volume | 1.800 | 04:42 |
| Air Blank | 0.000 | 04:42 |
| Control Test | 0.078 | 04:42 |
| Air Blank | 0.000 | 04:43 |
| Diagnostics Check | OK | 04:43 |

Data Download Date/Time: 07/16/2021 17:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: GOMEZ RAMIREZ

DATE/TIME OF TEST: 06/20/2021 02:11:56

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:05 |
| Air Blank | 0.000 | 02:05 |
| Control Test | 0.080 | 02:05 |
| Air Blank | 0.000 | 02:06 |
| Subject Sample # 1 | 0.210 | 02:07 |
| Breath Volume | 2.855 | 02:07 |
| Air Blank | 0.000 | 02:08 |
| Air Blank | 0.000 | 02:09 |
| Subject Sample # 2 | 0.205 | 02:09 |
| Breath Volume | 1.855 | 02:09 |
| Air Blank | 0.000 | 02:10 |
| Control Test | 0.076 | 02:11 |
| Air Blank | 0.000 | 02:11 |
| Diagnostics Check | OK | 02:11 |

Data Download Date/Time: 07/16/2021 17:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HOLLIGAN,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: SEPTIMUS

DATE/TIME OF TEST: 06/20/2021 04:28:04

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 04:20 |
| Air Blank | 0.000 | 04:21 |
| Control Test | 0.079 | 04:21 |
| Air Blank | 0.000 | 04:22 |
| Subject Sample # 1 | 0.149 | 04:22 |
| Breath Volume | 1.632 | 04:22 |
| Air Blank | 0.000 | 04:23 |
| Air Blank | 0.000 | 04:25 |
| Subject Sample # 2 | 0.149 | 04:26 |
| Breath Volume | 2.085 | 04:26 |
| Air Blank | 0.000 | 04:27 |
| Control Test | 0.075 | 04:27 |
| Air Blank | 0.000 | 04:27 |
| Diagnostics Check | OK | 04:28 |

Data Download Date/Time: 07/16/2021 17:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TEJEDA,ANTHONY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: AGUILAR LOPEZ

DATE/TIME OF TEST: 06/20/2021 22:54:08

| | | | |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK | 22:48 | |
| Air Blank | 0.000 | 22:48 | |
| Control Test | 0.080 | 22:48 | |
| Air Blank | 0.000 | 22:49 | |
| Subject Sample | # 1 0.218 | 22:52 | Volume Not Met |
| Breath Volume | 0.285 | 22:52 | |
| Air Blank | | 22:53 | Purge Fail |
| Air Blank | 0.000 | 22:54 | |

Data Download Date/Time: 07/16/2021 17:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TEJEDA,ANTHONY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: AGUILAR LOPEZ

DATE/TIME OF TEST: 06/20/2021 23:20:45

| | | | |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK | 23:19 | |
| Air Blank | 0.000 | 23:19 | |
| Control Test | 0.073 | 23:20 | Control Outside Tolerance |
| Air Blank | 0.000 | 23:20 | |

Data Download Date/Time: 07/16/2021 17:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TEJEDA,ANTHONY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 06/20/2021 23:26:28

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 23:22 | |
| Air Blank | 0.000 | 23:22 | |
| Control Test | 0.077 | 23:23 | |
| Air Blank | 0.000 | 23:23 | |
| Subject Sample | # 1 | 23:24 | Subject Test Refused |
| Breath Volume | 0.320 | 23:24 | |
| Air Blank | 0.000 | 23:25 | |
| Control Test | 0.079 | 23:25 | |
| Air Blank | 0.000 | 23:26 | |
| Diagnostics Check | OK | 23:26 | |

Data Download Date/Time: 07/16/2021 17:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WERTMAN,GREGORY
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: LOCKEY

DATE/TIME OF TEST: 06/21/2021 01:40:16

| | | | |
|--------------------|-------|-------|------------|
| Diagnostics Check | OK | 01:34 | |
| Air Blank | 0.000 | 01:34 | |
| Control Test | 0.075 | 01:34 | |
| Air Blank | 0.000 | 01:35 | |
| Subject Sample # 1 | 0.335 | 01:35 | |
| Breath Volume | 2.292 | 01:35 | |
| Air Blank | 0.000 | 01:36 | |
| Air Blank | 0.000 | 01:38 | |
| Subject Sample # 2 | 0.308 | 01:38 | |
| Breath Volume | 1.640 | 01:38 | |
| Air Blank | | 01:39 | Purge Fail |
| Air Blank | 0.000 | 01:40 | |

Data Download Date/Time: 07/16/2021 17:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WERTMAN,GREGORY
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: LOCKEY

DATE/TIME OF TEST: 06/21/2021 02:13:22

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 02:06 |
| Air Blank | | 0.000 | 02:06 |
| Control Test | | 0.076 | 02:07 |
| Air Blank | | 0.000 | 02:07 |
| Subject Sample | # 1 | 0.303 | 02:08 |
| Breath Volume | | 2.375 | 02:08 |
| Air Blank | | 0.000 | 02:09 |
| Air Blank | | 0.000 | 02:10 |
| Subject Sample | # 2 | 0.290 | 02:11 |
| Breath Volume | | 3.058 | 02:11 |
| Air Blank | | 0.000 | 02:12 |
| Control Test | | 0.076 | 02:12 |
| Air Blank | | 0.000 | 02:13 |
| Diagnostics Check | | OK | 02:13 |

Data Download Date/Time: 07/16/2021 17:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: BOTHE

DATE/TIME OF TEST: 06/24/2021 00:20:02

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 00:08 |
| Air Blank | | 0.000 | 00:08 |
| Control Test | | 0.080 | 00:08 |
| Air Blank | | 0.000 | 00:09 |
| Subject Sample | # 1 | 0.244 | 00:10 |
| Breath Volume | | 1.355 | 00:10 |
| Air Blank | | 0.000 | 00:11 |
| Air Blank | | 0.000 | 00:13 |
| Subject Sample | # 2 | 0.222 | 00:14 |
| Breath Volume | | 1.632 | 00:14 |
| Air Blank | | 0.000 | 00:15 |
| Air Blank | | 0.000 | 00:16 |
| Subject Sample | # 3 | 0.237 | 00:18 |
| Breath Volume | | 1.238 | 00:18 |
| Air Blank | | 0.000 | 00:19 |
| Control Test | | 0.077 | 00:19 |
| Air Blank | | 0.000 | 00:19 |
| Diagnostics Check | | OK | 00:20 |

Data Download Date/Time: 07/16/2021 17:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NALERIOL,LUIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: CHERILUS

DATE/TIME OF TEST: 06/24/2021 04:53:00

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 04:46 |
| Air Blank | 0.000 | 04:46 |
| Control Test | 0.079 | 04:47 |
| Air Blank | 0.000 | 04:47 |
| Subject Sample # 1 | 0.152 | 04:48 |
| Breath Volume | 3.011 | 04:48 |
| Air Blank | 0.000 | 04:48 |
| Air Blank | 0.000 | 04:50 |
| Subject Sample # 2 | 0.151 | 04:51 |
| Breath Volume | 3.128 | 04:51 |
| Air Blank | 0.000 | 04:52 |
| Control Test | 0.077 | 04:52 |
| Air Blank | 0.000 | 04:52 |
| Diagnostics Check | OK | 04:53 |

Data Download Date/Time: 07/16/2021 17:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CASAS,JAVIER
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: BRITTLE

DATE/TIME OF TEST: 06/24/2021 20:07:27

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 19:55 | |
| Air Blank | | 0.000 | 19:56 | |
| Control Test | | 0.080 | 19:56 | |
| Air Blank | | 0.000 | 19:56 | |
| Subject Sample | # 1 | 0.000 | 20:00 | Volume Not Met |
| Breath Volume | | 0.601 | 20:00 | |
| Air Blank | | 0.000 | 20:00 | |
| Air Blank | | 0.000 | 20:02 | |
| Subject Sample | # 2 | 0.000 | 20:03 | |
| Breath Volume | | 2.218 | 20:03 | |
| Air Blank | | 0.000 | 20:03 | |
| Air Blank | | 0.000 | 20:05 | |
| Subject Sample | # 3 | 0.000 | 20:06 | |
| Breath Volume | | 1.238 | 20:06 | |
| Air Blank | | 0.000 | 20:06 | |
| Control Test | | 0.078 | 20:06 | |
| Air Blank | | 0.000 | 20:07 | |
| Diagnostics Check | | OK | 20:07 | |

Data Download Date/Time: 07/16/2021 17:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PAGE, KAYLEE
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: PALESE

DATE/TIME OF TEST: 06/26/2021 22:11:36

| | | | |
|--------------------|-------|-------|------------|
| Diagnostics Check | OK | 22:06 | |
| Air Blank | 0.000 | 22:07 | |
| Control Test | 0.079 | 22:07 | |
| Air Blank | 0.000 | 22:08 | |
| Subject Sample # 1 | 0.183 | 22:10 | |
| Breath Volume | 2.503 | 22:10 | |
| Air Blank | | 22:11 | Purge Fail |
| Air Blank | 0.000 | 22:11 | |

Data Download Date/Time: 07/16/2021 17:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VALDMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: WATTS

DATE/TIME OF TEST: 06/27/2021 22:26:37

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:19 |
| Air Blank | 0.000 | 22:20 |
| Control Test | 0.079 | 22:20 |
| Air Blank | 0.000 | 22:21 |
| Subject Sample # 1 | 0.239 | 22:21 |
| Breath Volume | 1.855 | 22:21 |
| Air Blank | 0.000 | 22:22 |
| Air Blank | 0.000 | 22:24 |
| Subject Sample # 2 | 0.239 | 22:24 |
| Breath Volume | 2.406 | 22:24 |
| Air Blank | 0.000 | 22:25 |
| Control Test | 0.075 | 22:25 |
| Air Blank | 0.000 | 22:26 |
| Diagnostics Check | OK | 22:26 |

Data Download Date/Time: 07/16/2021 17:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: LOPEZ GABRIEL

DATE/TIME OF TEST: 06/29/2021 00:48:08

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 00:41 |
| Air Blank | | 0.000 | 00:41 |
| Control Test | | 0.080 | 00:41 |
| Air Blank | | 0.000 | 00:42 |
| Subject Sample | # 1 | 0.186 | 00:43 |
| Breath Volume | | 2.640 | 00:43 |
| Air Blank | | 0.000 | 00:44 |
| Air Blank | | 0.000 | 00:45 |
| Subject Sample | # 2 | 0.167 | 00:46 |
| Breath Volume | | 1.554 | 00:46 |
| Air Blank | | 0.000 | 00:47 |
| Control Test | | 0.076 | 00:47 |
| Air Blank | | 0.000 | 00:48 |
| Diagnostics Check | | OK | 00:48 |

Data Download Date/Time: 07/16/2021 17:19