

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BECKER,ALEXA,C  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: ALZATE,MELENNI  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 07/05/2021 20:12:31

---

Diagnostics Check	OK	20:09	
Air Blank	0.000	20:10	
Control Test	0.078	20:10	
Air Blank	0.000	20:10	
Subject Sample	# 1	20:11	Subject Test Refused
Breath Volume	0.000	20:11	
Air Blank	0.000	20:11	
Control Test	0.078	20:11	
Air Blank	0.000	20:12	
Diagnostics Check	OK	20:12	

---

Data Download Date/Time: 07/22/2021 13:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERA QUINONES,IGNACIO,  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: CABRAL,EDUARDO  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 16:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: GARCIA SANCHEZ

DATE/TIME OF TEST: 07/08/2021 18:08:46

---

Diagnostics Check	OK	18:05	
Air Blank	0.000	18:06	
Control Test	0.080	18:06	
Air Blank	0.000	18:07	
Subject Sample	# 1	18:07	Subject Test Refused
Breath Volume	0.000	18:07	
Air Blank	0.000	18:07	
Control Test	0.079	18:08	
Air Blank	0.000	18:08	
Diagnostics Check	OK	18:08	

---

Data Download Date/Time: 07/22/2021 13:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,DANIEL,A  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: SMITH,DANIEL  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 08:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: BARBERENA

DATE/TIME OF TEST: 07/16/2021 10:10:48

---

Diagnostics Check	OK	10:07	
Air Blank	0.000	10:08	
Control Test	0.079	10:08	
Air Blank	0.000	10:09	
Subject Sample	# 1	10:09	Subject Test Refused
Breath Volume	0.000	10:09	
Air Blank	0.000	10:09	
Control Test	0.080	10:10	
Air Blank	0.000	10:10	
Diagnostics Check	OK	10:10	

---

Data Download Date/Time: 07/22/2021 13:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SISNEY,WHITNEY,L  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: PINO,ALEJANDRO  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 07/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 07/29/2021 00:46:28

---

Diagnostics Check	OK	00:43	
Air Blank	0.000	00:44	
Control Test	0.079	00:44	
Air Blank	0.000	00:44	
Subject Sample	# 1	00:45	Subject Test Refused
Breath Volume	0.000	00:45	
Air Blank	0.000	00:45	
Control Test	0.080	00:45	
Air Blank	0.000	00:46	
Diagnostics Check	OK	00:46	

---

Data Download Date/Time: 08/26/2021 12:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SISNEY,WHITNEY,L  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: ALEMAN,ANTONIO  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 07/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 202303  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 07/29/2021 02:43:46

---

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.080	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.000	02:39
Breath Volume	2.710	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.000	02:42
Breath Volume	2.320	02:42
Air Blank	0.000	02:42
Control Test	0.079	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

---

Data Download Date/Time: 08/26/2021 12:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTAYA,GUILLERMO,  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: BRUDER,ALEXANDER  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: BUI  
ARREST TIME: 16:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:29

SUBJECT'S LAST NAME: LOPEZ ROJAS

DATE/TIME OF TEST: 07/04/2021 16:56:19

---

Diagnostics Check	OK	16:50
Air Blank	0.000	16:50
Control Test	0.080	16:51
Air Blank	0.000	16:51
Subject Sample # 1	0.025	16:52
Breath Volume	2.089	16:52
Air Blank	0.000	16:52
Air Blank	0.000	16:54
Subject Sample # 2	0.022	16:54
Breath Volume	2.226	16:54
Air Blank	0.000	16:55
Control Test	0.079	16:55
Air Blank	0.000	16:56
Diagnostics Check	OK	16:56

---

Data Download Date/Time: 07/29/2021 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PASTOR,JOSE,F  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: PASTOR,JOSE  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: PEREZ GARCIA

DATE/TIME OF TEST: 07/11/2021 01:24:13

---

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.079	01:18
Air Blank	0.000	01:18
Subject Sample # 1	0.253	01:19
Breath Volume	2.878	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.256	01:22
Breath Volume	2.308	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

---

Data Download Date/Time: 07/29/2021 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PASTOR,JOSE,F

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: HERNANDEZ,RIGOBERTO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 02:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SORIANO

DATE/TIME OF TEST: 07/11/2021 03:16:13

---

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.144	03:10
Breath Volume	1.226	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.153	03:14
Breath Volume	1.585	03:14
Air Blank	0.000	03:15
Control Test	0.080	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

---

Data Download Date/Time: 07/29/2021 09:42



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PASTOR,JOSE,F  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: PASTOR,JOSE  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 989917  
DRY GAS EXPIRY DATE: 07/25/2021  
VIOLATION CODE: DUI  
ARREST TIME: 23:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: ZAVALA

DATE/TIME OF TEST: 07/24/2021 00:08:14

---

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.079	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.113	00:03
Breath Volume	1.199	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:05
Subject Sample # 2	0.117	00:06
Breath Volume	1.343	00:06
Air Blank	0.000	00:07
Control Test	0.078	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

---

Data Download Date/Time: 07/29/2021 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PASTOR,JOSE,F  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: PASTOR,JOSE  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1506449  
DRY GAS EXPIRY DATE: 06/10/2024  
VIOLATION CODE: DUI  
ARREST TIME: 12:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 07/25/2021 01:22:18

---

Diagnostics Check	OK	01:12
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.171	01:15
Breath Volume	1.328	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.166	01:20
Breath Volume	1.449	01:20
Air Blank	0.000	01:21
Control Test	0.082	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

---

Data Download Date/Time: 07/29/2021 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PASTOR,JOSE,F  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: PASTOR,JOSE  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1506449  
DRY GAS EXPIRY DATE: 06/10/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/25/2021 03:20:40

---

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.078	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.145	03:16
Breath Volume	1.703	03:16
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.139	03:19
Breath Volume	1.429	03:19
Air Blank	0.000	03:19
Control Test	0.078	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

---

Data Download Date/Time: 07/29/2021 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PASTOR,JOSE,F  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: PASTOR,JOSE  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1506449  
DRY GAS EXPIRY DATE: 06/10/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MUNOZ PORTILLO

DATE/TIME OF TEST: 07/26/2021 00:22:04

---

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:17	
Control Test	0.080	00:17	
Air Blank	0.000	00:18	
Subject Sample	# 1 0.310	00:20	
Breath Volume	2.082	00:20	
Air Blank		00:21	Purge Fail
Air Blank	0.000	00:22	

---

Data Download Date/Time: 07/29/2021 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PASTOR,JOSE,F  
BREATH TEST OPERATOR AGENCY: Miami PD  
ARRESTING OFFICER: PASTOR,JOSE  
ARRESTING OFFICER AGENCY: Miami PD  
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1506449  
DRY GAS EXPIRY DATE: 06/10/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MUNOZ PORTILLO

DATE/TIME OF TEST: 07/26/2021 00:35:55

---

Diagnostics Check		OK	00:28
Air Blank		0.000	00:28
Control Test		0.077	00:29
Air Blank		0.000	00:29
Subject Sample	# 1	0.309	00:30
Breath Volume		1.777	00:30
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.311	00:34
Breath Volume		1.914	00:34
Air Blank		0.000	00:34
Control Test		0.079	00:35
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

---

Data Download Date/Time: 07/29/2021 09:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINCOLN,TREY,B  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: LINCOLN,TREY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 07/19/2021 19:59:12

---

Diagnostics Check	OK	19:52
Air Blank	0.000	19:52
Control Test	0.080	19:52
Air Blank	0.000	19:53
Subject Sample # 1	0.140	19:54
Breath Volume	2.683	19:54
Air Blank	0.000	19:55
Air Blank	0.000	19:57
Subject Sample # 2	0.137	19:57
Breath Volume	1.500	19:57
Air Blank	0.000	19:58
Control Test	0.080	19:58
Air Blank	0.000	19:59
Diagnostics Check	OK	19:59

---

Data Download Date/Time: 08/08/2021 12:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLEYNE, DAVANO, L  
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD  
ARRESTING OFFICER: ALLEYNE, DAVANO  
ARRESTING OFFICER AGENCY: MIAMI BEACH PD  
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 13:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: PARRONDO

DATE/TIME OF TEST: 07/03/2021 15:40:01

---

Diagnostics Check	OK	15:37	
Air Blank	0.000	15:37	
Control Test	0.080	15:37	
Air Blank	0.000	15:38	
Subject Sample	# 1	15:38	Subject Test Refused
Breath Volume	0.000	15:38	
Air Blank	0.000	15:39	
Control Test	0.080	15:39	
Air Blank	0.000	15:39	
Diagnostics Check	OK	15:40	

---

Data Download Date/Time: 07/22/2021 13:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: BAEZ,RICHARD  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BENAVIDES-VASQUEZ

DATE/TIME OF TEST: 07/05/2021 03:43:44

---

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:41	
Control Test	0.079	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:42	
Control Test	0.080	03:43	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

---

Data Download Date/Time: 08/19/2021 03:45



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICHARDSON,IVETTE,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: HERRERA,CARLOS  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/06/2021 00:02:01

---

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.079	23:54
Air Blank	0.000	23:55
Subject Sample # 2	0.259	23:56
Breath Volume	2.648	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 1	0.258	00:00
Breath Volume	2.191	00:00
Air Blank	0.000	00:01
Control Test	0.079	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:02

---

Data Download Date/Time: 08/19/2021 03:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HELFNER,ISAAC,S  
BREATH TEST OPERATOR AGENCY: Aventura PD  
ARRESTING OFFICER: HELFNER,ISAAC  
ARRESTING OFFICER AGENCY: Aventura PD  
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: ATTLEY

DATE/TIME OF TEST: 07/03/2021 21:23:19

---

Diagnostics Check	OK	21:17
Air Blank	0.000	21:17
Control Test	0.081	21:17
Air Blank	0.000	21:18
Subject Sample # 1	0.179	21:18
Breath Volume	2.839	21:18
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample # 2	0.167	21:21
Breath Volume	2.101	21:21
Air Blank	0.000	21:22
Control Test	0.080	21:22
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

---

Data Download Date/Time: 08/19/2021 03:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KONETZ,JONATHAN,D  
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD  
ARRESTING OFFICER: KONETZ,JONATHAN  
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD  
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: RAKOVSKY

DATE/TIME OF TEST: 07/07/2021 00:52:33

---

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:49	
Control Test	0.080	00:49	
Air Blank	0.000	00:50	
Subject Sample	# 1	00:51	Subject Test Refused
Breath Volume	0.000	00:51	
Air Blank	0.000	00:51	
Control Test	0.080	00:51	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

---

Data Download Date/Time: 08/19/2021 03:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANTELO,JOSEPH,M  
BREATH TEST OPERATOR AGENCY: Aventura PD  
ARRESTING OFFICER: ANTELO,JOSEPH  
ARRESTING OFFICER AGENCY: Aventura PD  
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 08:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: KHOEE

DATE/TIME OF TEST: 07/31/2021 09:32:48

---

Diagnostics Check	OK	09:28	
Air Blank	0.000	09:28	
Control Test	0.081	09:29	
Air Blank	0.000	09:29	
Subject Sample	# 1	09:31	Subject Test Refused
Breath Volume	0.000	09:31	
Air Blank	0.000	09:31	
Control Test	0.081	09:32	
Air Blank	0.000	09:32	
Diagnostics Check	OK	09:32	

---

Data Download Date/Time: 08/19/2021 03:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHANG,BRADLEY,M  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CHANG,B  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CURBELO HORTA

DATE/TIME OF TEST: 07/12/2021 02:00:43

---

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:58	
Control Test	0.079	01:58	
Air Blank	0.000	01:59	
Subject Sample	# 1	01:59	Subject Test Refused
Breath Volume	0.000	01:59	
Air Blank	0.000	01:59	
Control Test	0.079	02:00	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

---

Data Download Date/Time: 07/14/2021 10:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PUGA,JORGE,A  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PUGA,JORGE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: RIVERA GUZARO

DATE/TIME OF TEST: 07/17/2021 21:16:44

---

Diagnostics Check	OK	21:13	
Air Blank	0.000	21:14	
Control Test	0.079	21:14	
Air Blank	0.000	21:15	
Subject Sample	# 1	21:15	Subject Test Refused
Breath Volume	0.000	21:15	
Air Blank	0.000	21:15	
Control Test	0.079	21:16	
Air Blank	0.000	21:16	
Diagnostics Check	OK	21:16	

---

Data Download Date/Time: 08/06/2021 10:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEANDRE,GERALD,L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: LEANDRE,GERALD  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/24/2021 05:14:41

---

Diagnostics Check	OK	05:05
Air Blank	0.000	05:06
Control Test	0.078	05:06
Air Blank	0.000	05:07
Subject Sample # 1	0.199	05:10
Breath Volume	2.238	05:10
Air Blank	0.000	05:11
Air Blank	0.000	05:12
Subject Sample # 2	0.205	05:13
Breath Volume	1.664	05:13
Air Blank	0.000	05:13
Control Test	0.079	05:14
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

---

Data Download Date/Time: 08/06/2021 10:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOMEZ,RUBEN,G  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: SAMPLE,CARSON  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: BUI  
ARREST TIME: 15:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:29

SUBJECT'S LAST NAME: ALMEIDA

DATE/TIME OF TEST: 07/03/2021 16:02:11

---

Diagnostics Check	OK	16:01	
Air Blank		16:01	RFI Detect
Air Blank	0.000	16:02	

---

Data Download Date/Time: 07/14/2021 09:58



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOMEZ,RUBEN,G  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: SAMPLE,CARSON  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: BUI  
ARREST TIME: 15:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:29

SUBJECT'S LAST NAME: ALMEIDA

DATE/TIME OF TEST: 07/03/2021 16:06:11

---

Diagnostics Check	Fail	16:05	Temperature Regulation Test
Air Blank		16:06	RFI Detect

---

Data Download Date/Time: 07/14/2021 09:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOMEZ,RUBEN,G  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: SAMPLE,CARSON  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: BUI  
ARREST TIME: 15:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:29

SUBJECT'S LAST NAME: ALMEIDA

DATE/TIME OF TEST: 07/03/2021 16:09:24

---

Diagnostics Check	Fail	16:08	Temperature Regulation Test
Air Blank	0.000	16:09	

---

Data Download Date/Time: 07/14/2021 09:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOMEZ,RUBEN,G  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: SAMPLE,CARSON  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: BUI  
ARREST TIME: 15:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:29

SUBJECT'S LAST NAME: ALMEIDA

DATE/TIME OF TEST: 07/03/2021 16:13:03

---

Diagnostics Check	Fail	16:12	Temperature Regulation Test
Air Blank		16:13	RFI Detect

---

Data Download Date/Time: 07/14/2021 09:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOMEZ,RUBEN,G  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: SAMPLE,CARSON  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: BUI  
ARREST TIME: 15:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:29

SUBJECT'S LAST NAME: ALMEIDA

DATE/TIME OF TEST: 07/03/2021 16:44:53

---

Diagnostics Check	OK	16:41	
Air Blank	0.000	16:42	
Control Test	0.082	16:42	
Air Blank	0.000	16:43	
Subject Sample	# 1	16:43	Subject Test Refused
Breath Volume	0.000	16:43	
Air Blank	0.000	16:43	
Control Test	0.080	16:44	
Air Blank	0.000	16:44	
Diagnostics Check	OK	16:44	

---

Data Download Date/Time: 07/14/2021 09:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARTAYA,GUILLERMO,  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: CARTAYA,GUILLERMO  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: BUI  
ARREST TIME: 17:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/03/2021 18:25:07

---

Diagnostics Check	OK	18:19	
Air Blank	0.000	18:20	
Control Test	0.081	18:20	
Air Blank	0.000	18:20	
Subject Sample	# 1	18:23	Subject Test Refused
Breath Volume	0.000	18:23	
Air Blank	0.000	18:24	
Control Test	0.082	18:24	
Air Blank	0.000	18:25	
Diagnostics Check	OK	18:25	

---

Data Download Date/Time: 07/14/2021 09:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAWSON, GLADSTON,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: OROL, MANUEL  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: WICKS

DATE/TIME OF TEST: 07/26/2021 18:55:02

---

Diagnostics Check		OK	18:47
Air Blank		0.000	18:48
Control Test		0.080	18:48
Air Blank		0.000	18:49
Subject Sample	# 1	0.238	18:50
Breath Volume		2.078	18:50
Air Blank		0.000	18:50
Air Blank		0.000	18:52
Subject Sample	# 2	0.233	18:53
Breath Volume		1.539	18:53
Air Blank		0.000	18:54
Control Test		0.079	18:54
Air Blank		0.000	18:54
Diagnostics Check		OK	18:55

---

Data Download Date/Time: 08/05/2021 11:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PACHECO,GEORGE,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PACHECO,MICHELLE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/31/2021 01:29:50

---

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:27	
Control Test	0.081	01:27	
Air Blank	0.000	01:28	
Subject Sample	# 1	01:28	Subject Test Refused
Breath Volume	0.000	01:28	
Air Blank	0.000	01:28	
Control Test	0.081	01:29	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

---

Data Download Date/Time: 08/05/2021 11:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUINTANA,JORGE,F  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: QUINTANA,JORGE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 07/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05620080A1  
DRY GAS EXPIRY DATE: 04/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MCNEAL

DATE/TIME OF TEST: 07/31/2021 01:43:14

---

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.105	01:38
Breath Volume	1.765	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:40
Subject Sample # 2	0.110	01:41
Breath Volume	2.085	01:41
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

---

Data Download Date/Time: 08/05/2021 11:00



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANEY,DAMON,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LANEY,DAMON  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/10/2021 05:09:08

---

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.077	05:03
Air Blank	0.000	05:04
Subject Sample # 1	0.000	05:04
Breath Volume	1.601	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.000	05:07
Breath Volume	1.449	05:07
Air Blank	0.000	05:08
Control Test	0.078	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

---

Data Download Date/Time: 08/20/2021 19:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LANEY,DAMON,A  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: LANEY,DAMON  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: OPHAUG-HENRY

DATE/TIME OF TEST: 07/17/2021 00:30:09

---

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.077	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.000	00:25
Breath Volume	3.609	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.000	00:28
Breath Volume	3.390	00:28
Air Blank	0.000	00:29
Control Test	0.079	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

---

Data Download Date/Time: 08/20/2021 19:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MULLINS,JOHN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG011102  
DRY GAS EXPIRY DATE: 04/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: BOISVENUE

DATE/TIME OF TEST: 07/24/2021 05:04:53

---

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.078	04:59
Air Blank	0.000	04:59
Subject Sample # 1	0.159	05:00
Breath Volume	1.421	05:00
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.157	05:03
Breath Volume	1.296	05:03
Air Blank	0.000	05:03
Control Test	0.079	05:04
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

---

Data Download Date/Time: 08/20/2021 19:15

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1206004
BREATH TEST OPERATOR: MALTAIS,ANDREW,W	DRY GAS EXPIRY DATE: 09/20/2022
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: BUI
ARRESTING OFFICER: MALTAIS,ANDREW	ARREST TIME: 18:33
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2021	OBSERVATION PERIOD BEGAN: 18:33
SUBJECT'S LAST NAME: HALLEY	DATE/TIME OF TEST: 07/01/2021 19:07:13

---

Diagnostics Check	OK	19:04	
Air Blank	0.000	19:04	
Control Test	0.081	19:05	
Air Blank	0.000	19:05	
Subject Sample	# 1	19:05	Subject Test Refused
Breath Volume	0.000	19:05	
Air Blank	0.000	19:06	
Control Test	0.081	19:06	
Air Blank	0.000	19:07	
Diagnostics Check	OK	19:07	

---

Data Download Date/Time: 07/29/2021 14:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: DOWNING,CADEN  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: BUI  
ARREST TIME: 14:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:58

SUBJECT'S LAST NAME: CORDOVA  
DATE/TIME OF TEST: 07/03/2021 15:36:14

---

Diagnostics Check	OK	15:30
Air Blank	0.000	15:30
Control Test	0.080	15:30
Air Blank	0.000	15:31
Subject Sample # 1	0.061	15:31
Breath Volume	2.222	15:31
Air Blank	0.000	15:32
Air Blank	0.000	15:34
Subject Sample # 2	0.061	15:34
Breath Volume	2.835	15:34
Air Blank	0.000	15:35
Control Test	0.078	15:35
Air Blank	0.000	15:36
Diagnostics Check	OK	15:36

---

Data Download Date/Time: 07/29/2021 14:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: DONOVAN,CHARLES  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: BUI  
ARREST TIME: 16:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:54

SUBJECT'S LAST NAME: CREAN

DATE/TIME OF TEST: 07/03/2021 17:22:52

---

Diagnostics Check	OK	17:16
Air Blank	0.000	17:17
Control Test	0.079	17:17
Air Blank	0.000	17:18
Subject Sample # 1	0.000	17:18
Breath Volume	2.281	17:18
Air Blank	0.000	17:19
Air Blank	0.000	17:21
Subject Sample # 2	0.000	17:21
Breath Volume	3.351	17:21
Air Blank	0.000	17:21
Control Test	0.080	17:22
Air Blank	0.000	17:22
Diagnostics Check	OK	17:22

---

Data Download Date/Time: 07/29/2021 14:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIFER,JUSTIN,S  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: MATECHIK,GREGORY  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 07/03/2021 20:22:58

---

Diagnostics Check	OK	20:16
Air Blank	0.000	20:16
Control Test	0.080	20:17
Air Blank	0.000	20:17
Subject Sample	# 1 0.198	20:18
Breath Volume	3.050	20:18
Air Blank	0.000	20:19
Air Blank	0.000	20:20
Subject Sample	# 2 0.202	20:21
Breath Volume	3.906	20:21
Air Blank	0.000	20:22
Control Test	0.079	20:22
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

---

Data Download Date/Time: 07/29/2021 14:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1206004
BREATH TEST OPERATOR: CORBITT,KYLE,L	DRY GAS EXPIRY DATE: 09/20/2022
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: BUI
ARRESTING OFFICER: DOWNING,CADEN	ARREST TIME: 20:24
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2021	OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: MARTIN	DATE/TIME OF TEST: 07/03/2021	21:13:35
-----------------------------	-------------------------------	----------

---

Diagnostics Check	OK	21:04
Air Blank	0.000	21:04
Control Test	0.080	21:04
Air Blank	0.000	21:05
Subject Sample # 1	0.208	21:08
Breath Volume	1.613	21:08
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample # 2	0.225	21:11
Breath Volume	2.496	21:11
Air Blank	0.000	21:12
Control Test	0.079	21:12
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

---

Data Download Date/Time: 07/29/2021 14:22



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: DONOVAN,CHARLES  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: BUI  
ARREST TIME: 17:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: AVEN

DATE/TIME OF TEST: 07/04/2021 17:34:38

---

Diagnostics Check		OK	17:28
Air Blank		0.000	17:28
Control Test		0.080	17:29
Air Blank		0.000	17:29
Subject Sample	# 1	0.156	17:30
Breath Volume		1.480	17:30
Air Blank		0.000	17:30
Air Blank		0.000	17:32
Subject Sample	# 2	0.163	17:33
Breath Volume		3.351	17:33
Air Blank		0.000	17:33
Control Test		0.080	17:34
Air Blank		0.000	17:34
Diagnostics Check		OK	17:34

---

Data Download Date/Time: 07/29/2021 14:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORBITT,KYLE,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: POWELL,MICHAEL  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: BUI  
ARREST TIME: 17:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: TUCKER  
DATE/TIME OF TEST: 07/04/2021 18:38:17

---

Diagnostics Check	OK	18:35	
Air Blank	0.000	18:35	
Control Test	0.080	18:36	
Air Blank	0.000	18:36	
Subject Sample	# 1	18:36	Subject Test Refused
Breath Volume	0.000	18:36	
Air Blank	0.000	18:37	
Control Test	0.093	18:37	Control Outside Tolerance
Air Blank	0.000	18:38	

---

Data Download Date/Time: 07/29/2021 14:22

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORBITT,KYLE,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: POWELL,MICHAEL  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: BUI  
ARREST TIME: 17:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 07/04/2021 18:43:00

---

Diagnostics Check	OK	18:40	
Air Blank	0.000	18:40	
Control Test	0.080	18:40	
Air Blank	0.000	18:41	
Subject Sample	# 1	18:41	Subject Test Refused
Breath Volume	0.000	18:41	
Air Blank	0.000	18:42	
Control Test	0.080	18:42	
Air Blank	0.000	18:42	
Diagnostics Check	OK	18:43	

---

Data Download Date/Time: 07/29/2021 14:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1206004
BREATH TEST OPERATOR: CORBITT,KYLE,L	DRY GAS EXPIRY DATE: 09/20/2022
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: BUI
ARRESTING OFFICER: WILLIAMS,DANIEL	ARREST TIME: 19:44
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2021	OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: BASS                      DATE/TIME OF TEST: 07/04/2021                      20:10:38

---

Diagnostics Check	OK	20:07	
Air Blank	0.000	20:08	
Control Test	0.080	20:08	
Air Blank	0.000	20:08	
Subject Sample	# 1	20:09	Subject Test Refused
Breath Volume	0.000	20:09	
Air Blank	0.000	20:09	
Control Test	0.080	20:10	
Air Blank	0.000	20:10	
Diagnostics Check	OK	20:10	

---

Data Download Date/Time: 07/29/2021 14:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MALTAIS,ANDREW,W  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: MALTAIS,ANDREW  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: BUI  
ARREST TIME: 20:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: PIKE

DATE/TIME OF TEST: 07/04/2021 20:58:29

---

Diagnostics Check	OK	20:52
Air Blank	0.000	20:52
Control Test	0.080	20:53
Air Blank	0.000	20:53
Subject Sample # 1	0.149	20:53
Breath Volume	1.144	20:53
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.163	20:56
Breath Volume	1.390	20:56
Air Blank	0.000	20:57
Control Test	0.079	20:57
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

---

Data Download Date/Time: 07/29/2021 14:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1206004
BREATH TEST OPERATOR: CORBIN,BRITTON,L	DRY GAS EXPIRY DATE: 09/20/2022
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: CORBIN,BRITTON	ARREST TIME: 20:52
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2021	OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: FUENTES CASILLAS                      DATE/TIME OF TEST: 07/04/2021                      21:17:30

---

Diagnostics Check	OK	21:13	
Air Blank	0.000	21:13	
Control Test	0.080	21:13	
Air Blank	0.000	21:14	
Subject Sample	# 1	21:16	RFI Detect
Breath Volume	0.000	21:16	
Air Blank	0.000	21:16	
Control Test	0.080	21:16	
Air Blank	0.000	21:17	
Diagnostics Check	OK	21:17	

---

Data Download Date/Time: 07/29/2021 14:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1206004
BREATH TEST OPERATOR: CORBIN,BRITTON,L	DRY GAS EXPIRY DATE: 09/20/2022
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: CORBIN,BRITTON	ARREST TIME: 20:52
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2021	OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: FUENTES CASILLAS      DATE/TIME OF TEST: 07/04/2021      21:32:56

---

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:21	
Control Test	0.080	21:21	
Air Blank	0.000	21:21	
Subject Sample	# 1 0.121	21:25	Volume Not Met
Breath Volume	0.078	21:25	
Air Blank	0.000	21:25	
Air Blank	0.000	21:27	
Subject Sample	# 2 0.135	21:28	
Breath Volume	1.988	21:28	
Air Blank	0.000	21:29	
Air Blank	0.000	21:30	
Subject Sample	# 3 0.137	21:31	
Breath Volume	1.691	21:31	
Air Blank	0.000	21:32	
Control Test	0.079	21:32	
Air Blank	0.000	21:32	
Diagnostics Check	OK	21:32	

---

Data Download Date/Time: 07/29/2021 14:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAHR,DANIEL,E  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: HAHR,DANIEL  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: BUI  
ARREST TIME: 20:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: SALTZMANN

DATE/TIME OF TEST: 07/04/2021 21:41:35

---

Diagnostics Check	OK	21:38	
Air Blank	0.000	21:39	
Control Test	0.080	21:39	
Air Blank	0.000	21:39	
Subject Sample	# 1	21:40	Subject Test Refused
Breath Volume	0.000	21:40	
Air Blank	0.000	21:40	
Control Test	0.080	21:41	
Air Blank	0.000	21:41	
Diagnostics Check	OK	21:41	

---

Data Download Date/Time: 07/29/2021 14:58



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1206004
BREATH TEST OPERATOR: CORBITT,KYLE,L	DRY GAS EXPIRY DATE: 09/20/2022
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: BUI
ARRESTING OFFICER: DOWNING,CADEN	ARREST TIME: 22:04
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2021	OBSERVATION PERIOD BEGAN: 22:04
 SUBJECT'S LAST NAME: JOHNSON	 DATE/TIME OF TEST: 07/04/2021 22:40:08

---

Diagnostics Check	OK	22:33
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.096	22:35
Breath Volume	2.101	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.095	22:38
Breath Volume	2.289	22:38
Air Blank	0.000	22:39
Control Test	0.079	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

---

Data Download Date/Time: 07/29/2021 14:58

**Florida Department of Law Enforcement**

**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: WILLIAMS,DANIEL  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: BUI  
ARREST TIME: 22:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: TIMBLIN

DATE/TIME OF TEST: 07/04/2021 22:51:18

---

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:43	
Control Test	0.079	22:43	
Air Blank	0.000	22:44	
Subject Sample	# 1	22:47	No Sample Provided
Breath Volume	0.000	22:47	
Air Blank	0.000	22:47	
Air Blank	0.000	22:49	
Subject Sample	# 2	22:49	Subject Test Refused
Breath Volume	0.000	22:49	
Air Blank	0.000	22:50	
Control Test	0.080	22:50	
Air Blank	0.000	22:51	
Diagnostics Check	OK	22:51	

---

Data Download Date/Time: 07/29/2021 14:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1206004
BREATH TEST OPERATOR: MOLNAR,JARROD,T	DRY GAS EXPIRY DATE: 09/20/2022
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: MOLNAR,JARROD	ARREST TIME: 22:08
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2021	OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: NUNEZ-CARMONA      DATE/TIME OF TEST: 07/04/2021      23:01:49

---

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:59	
Control Test	0.080	22:59	
Air Blank	0.000	23:00	
Subject Sample	# 1	23:00	Subject Test Refused
Breath Volume	0.000	23:00	
Air Blank	0.000	23:00	
Control Test	0.080	23:01	
Air Blank	0.000	23:01	
Diagnostics Check	OK	23:01	

---

Data Download Date/Time: 07/29/2021 14:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANEY,MICHAEL,P  
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)  
ARRESTING OFFICER: HANEY,MICHAEL  
ARRESTING OFFICER AGENCY: FFWCC (Lake City)  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: BUI  
ARREST TIME: 15:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: LANDRUM

DATE/TIME OF TEST: 07/09/2021 16:25:52

---

Diagnostics Check		OK	16:19
Air Blank		0.000	16:19
Control Test		0.079	16:20
Air Blank		0.000	16:20
Subject Sample	# 1	0.155	16:21
Breath Volume		2.382	16:21
Air Blank		0.000	16:22
Air Blank		0.000	16:23
Subject Sample	# 2	0.156	16:24
Breath Volume		2.222	16:24
Air Blank		0.000	16:24
Control Test		0.079	16:25
Air Blank		0.000	16:25
Diagnostics Check		OK	16:25

---

Data Download Date/Time: 07/29/2021 14:59



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1206004
BREATH TEST OPERATOR: HANEY,MICHAEL,P	DRY GAS EXPIRY DATE: 09/20/2022
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)	VIOLATION CODE: DUI
ARRESTING OFFICER: GRAY,SETH	ARREST TIME: 16:11
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2021	OBSERVATION PERIOD BEGAN: 16:30
<hr/>	
SUBJECT'S LAST NAME: MCDONNELL	DATE/TIME OF TEST: 07/10/2021 17:08:28

---

Diagnostics Check	OK	17:05	
Air Blank	0.000	17:06	
Control Test	0.080	17:06	
Air Blank	0.000	17:06	
Subject Sample	# 1	17:07	Subject Test Refused
Breath Volume	0.000	17:07	
Air Blank	0.000	17:07	
Control Test	0.080	17:07	
Air Blank	0.000	17:08	
Diagnostics Check	OK	17:08	

---

Data Download Date/Time: 07/29/2021 14:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANEY, MICHAEL, P  
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)  
ARRESTING OFFICER: CONROY, JESSICA  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: BUI  
ARREST TIME: 16:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:33

SUBJECT'S LAST NAME: BOOTH

DATE/TIME OF TEST: 07/10/2021 17:59:42

---

Diagnostics Check	OK	17:53
Air Blank	0.000	17:53
Control Test	0.079	17:54
Air Blank	0.000	17:54
Subject Sample # 1	0.169	17:55
Breath Volume	3.625	17:55
Air Blank	0.000	17:55
Air Blank	0.000	17:57
Subject Sample # 2	0.158	17:58
Breath Volume	3.726	17:58
Air Blank	0.000	17:58
Control Test	0.076	17:59
Air Blank	0.000	17:59
Diagnostics Check	OK	17:59

---

Data Download Date/Time: 07/29/2021 14:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HANEY, MICHAEL, P

BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)

ARRESTING OFFICER: RICE, MICHAEL

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1206004

DRY GAS EXPIRY DATE: 09/20/2022

VIOLATION CODE: BUI

ARREST TIME: 16:58

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 07/10/2021 18:08:38

---

Diagnostics Check		OK	18:02
Air Blank		0.000	18:02
Control Test		0.079	18:03
Air Blank		0.000	18:03
Subject Sample	# 1	0.221	18:04
Breath Volume		2.109	18:04
Air Blank		0.000	18:04
Air Blank		0.000	18:06
Subject Sample	# 2	0.227	18:07
Breath Volume		2.503	18:07
Air Blank		0.000	18:07
Control Test		0.080	18:08
Air Blank		0.000	18:08
Diagnostics Check		OK	18:08

---

Data Download Date/Time: 07/29/2021 14:59



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANEY, MICHAEL, P  
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)  
ARRESTING OFFICER: CUSHING, MATTHEW  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: BUI  
ARREST TIME: 17:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: MCCOOL JR

DATE/TIME OF TEST: 07/10/2021 18:58:55

---

Diagnostics Check		OK	18:47	
Air Blank		0.000	18:47	
Control Test		0.079	18:48	
Air Blank		0.000	18:48	
Subject Sample	# 1	0.137	18:51	Volume Not Met
Breath Volume		0.222	18:51	
Air Blank		0.000	18:52	
Air Blank		0.000	18:54	
Subject Sample	# 2	0.127	18:57	Volume Not Met
Breath Volume		0.230	18:57	
Air Blank		0.000	18:58	
Control Test		0.077	18:58	
Air Blank		0.000	18:58	
Diagnostics Check		OK	18:58	

---

Data Download Date/Time: 07/29/2021 14:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANEY, MICHAEL, P  
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)  
ARRESTING OFFICER: CUSHING, MATTHEW  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: BUI  
ARREST TIME: 17:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: MCCOOL JR

DATE/TIME OF TEST: 07/10/2021 19:07:07

---

Diagnostics Check	OK	19:00
Air Blank	0.000	19:01
Control Test	0.078	19:01
Air Blank	0.000	19:02
Subject Sample # 1	0.133	19:02
Breath Volume	2.371	19:02
Air Blank	0.000	19:03
Air Blank	0.000	19:05
Subject Sample # 2	0.125	19:05
Breath Volume	1.335	19:05
Air Blank	0.000	19:06
Control Test	0.080	19:06
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

---

Data Download Date/Time: 07/29/2021 14:59



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OVERBAY,WILLIAM,M  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: WILLIAMS,DANIEL  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: BUI  
ARREST TIME: 15:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:51

SUBJECT'S LAST NAME: SHIPPY

DATE/TIME OF TEST: 07/13/2021 16:44:29

---

Diagnostics Check	OK	16:35	
Air Blank	0.000	16:36	
Control Test	0.080	16:36	
Air Blank	0.000	16:37	
Subject Sample	# 1	16:40	No Sample Provided
Breath Volume	0.000	16:40	
Air Blank	0.000	16:40	
Air Blank	0.000	16:42	
Subject Sample	# 2	16:43	Subject Test Refused
Breath Volume	0.000	16:43	
Air Blank	0.000	16:43	
Control Test	0.080	16:43	
Air Blank	0.000	16:44	
Diagnostics Check	OK	16:44	

---

Data Download Date/Time: 07/29/2021 14:59

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: WILLIAMS,DANIEL  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/21/2021  
Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: BUI  
ARREST TIME: 17:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: GARZA  
DATE/TIME OF TEST: 07/15/2021 17:51:39

---

Diagnostics Check	OK	17:43	
Air Blank	0.000	17:43	
Control Test	0.081	17:44	
Air Blank	0.000	17:44	
Subject Sample	# 1	17:47	No Sample Provided
Breath Volume	0.000	17:47	
Air Blank	0.000	17:48	
Air Blank	0.000	17:50	
Subject Sample	# 2	17:50	Subject Test Refused
Breath Volume	0.000	17:50	
Air Blank	0.000	17:50	
Control Test	0.081	17:51	
Air Blank	0.000	17:51	
Diagnostics Check	OK	17:51	

---

Data Download Date/Time: 07/29/2021 14:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: WILLIAMS,DANIEL  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: BUI  
ARREST TIME: 18:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: FRITSCHÉ DATE/TIME OF TEST: 07/16/2021 18:26:21

---

Diagnostics Check	OK	18:20
Air Blank	0.000	18:20
Control Test	0.080	18:21
Air Blank	0.000	18:21
Subject Sample # 1	0.096	18:21
Breath Volume	1.382	18:21
Air Blank	0.000	18:22
Air Blank	0.000	18:24
Subject Sample # 2	0.091	18:24
Breath Volume	2.292	18:24
Air Blank	0.000	18:25
Control Test	0.078	18:25
Air Blank	0.000	18:26
Diagnostics Check	OK	18:26

---

Data Download Date/Time: 07/29/2021 14:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DONOVAN,CHARLES,L  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: WILLIAMS,DANIEL  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: BUI  
ARREST TIME: 15:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:18

SUBJECT'S LAST NAME: BARENTINE

DATE/TIME OF TEST: 07/17/2021 15:44:38

---

Diagnostics Check	OK	15:38
Air Blank	0.000	15:38
Control Test	0.081	15:39
Air Blank	0.000	15:39
Subject Sample # 1	0.163	15:40
Breath Volume	2.917	15:40
Air Blank	0.000	15:40
Air Blank	0.000	15:42
Subject Sample # 2	0.143	15:43
Breath Volume	1.250	15:43
Air Blank	0.000	15:43
Control Test	0.076	15:44
Air Blank	0.000	15:44
Diagnostics Check	OK	15:44

---

Data Download Date/Time: 07/29/2021 14:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,JASON,E  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: HAHR,DANIEL  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: BUI  
ARREST TIME: 17:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: AVELLA

DATE/TIME OF TEST: 07/17/2021 17:42:46

---

Diagnostics Check	OK	17:35
Air Blank	0.000	17:36
Control Test	0.080	17:36
Air Blank	0.000	17:37
Subject Sample # 1	0.193	17:38
Breath Volume	1.695	17:38
Air Blank	0.000	17:38
Air Blank	0.000	17:40
Subject Sample # 2	0.183	17:41
Breath Volume	1.613	17:41
Air Blank	0.000	17:41
Control Test	0.081	17:42
Air Blank	0.000	17:42
Diagnostics Check	OK	17:42

---

Data Download Date/Time: 07/29/2021 14:59





# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: DONOVAN,CHARLES  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: BUI  
ARREST TIME: 16:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:28

SUBJECT'S LAST NAME: SIEVERT  
DATE/TIME OF TEST: 07/18/2021 17:04:08

---

Diagnostics Check	OK	16:56
Air Blank	0.000	16:57
Control Test	0.081	16:57
Air Blank	0.000	16:58
Subject Sample # 1	0.043	16:59
Breath Volume	1.660	16:59
Air Blank	0.000	17:00
Air Blank	0.000	17:02
Subject Sample # 2	0.040	17:02
Breath Volume	1.574	17:02
Air Blank	0.000	17:03
Control Test	0.079	17:03
Air Blank	0.000	17:04
Diagnostics Check	OK	17:04

---

Data Download Date/Time: 07/29/2021 14:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: WILLIAMS,DANIEL  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: BUI  
ARREST TIME: 16:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:46

SUBJECT'S LAST NAME: DUNNAVANT

DATE/TIME OF TEST: 07/22/2021 17:15:11

---

Diagnostics Check	OK	17:07
Air Blank	0.000	17:07
Control Test	0.081	17:08
Air Blank	0.000	17:08
Subject Sample # 1	0.160	17:10
Breath Volume	2.667	17:10
Air Blank	0.000	17:11
Air Blank	0.000	17:13
Subject Sample # 2	0.150	17:13
Breath Volume	1.851	17:13
Air Blank	0.000	17:14
Control Test	0.078	17:14
Air Blank	0.000	17:15
Diagnostics Check	OK	17:15

---

Data Download Date/Time: 07/29/2021 15:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,JASON,E  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: HAHR,DANIEL  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: BUI  
ARREST TIME: 19:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: FISCHER

DATE/TIME OF TEST: 07/24/2021 20:04:21

---

Diagnostics Check	OK	19:59	
Air Blank	0.000	19:59	
Control Test	0.081	20:00	
Air Blank	0.000	20:00	
Subject Sample	# 1	20:02	Subject Test Refused
Breath Volume	0.000	20:02	
Air Blank	0.000	20:03	
Control Test	0.081	20:03	
Air Blank	0.000	20:04	
Diagnostics Check	OK	20:04	

---

Data Download Date/Time: 07/29/2021 15:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MALTAIS,ANDREW,W  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: MALTAIS,ANDREW  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: BUI  
ARREST TIME: 16:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: L450623800690      DATE/TIME OF TEST: 07/25/2021      16:54:23

---

Diagnostics Check	OK	16:51	
Air Blank	0.000	16:51	
Control Test	0.081	16:52	
Air Blank	0.000	16:52	
Subject Sample	# 1	16:53	Subject Test Refused
Breath Volume	0.000	16:53	
Air Blank	0.000	16:53	
Control Test	0.080	16:53	
Air Blank	0.000	16:54	
Diagnostics Check	OK	16:54	

---

Data Download Date/Time: 07/29/2021 15:00

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BASFORD,TRAVIS,L  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: BASFORD,TRAVIS  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/21/2021  
VIOLATION CODE: BUI  
ARREST TIME: 21:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: SHACKLETON

DATE/TIME OF TEST: 07/02/2021 21:58:17

---

Diagnostics Check	OK	21:54	
Air Blank	0.000	21:54	
Control Test	0.081	21:54	
Air Blank	0.000	21:55	
Subject Sample	# 1	21:56	Subject Test Refused
Breath Volume	0.000	21:56	
Air Blank	0.000	21:57	
Control Test	0.080	21:57	
Air Blank	0.000	21:58	
Diagnostics Check	OK	21:58	

---

Data Download Date/Time: 07/29/2021 14:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BASFORD,NICOLE,M  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: BASFORD,NICOLE  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/21/2021  
VIOLATION CODE: DUI  
ARREST TIME: 19:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: CALDERO JUAREZ

DATE/TIME OF TEST: 07/04/2021 19:56:59

---

Diagnostics Check	OK	19:50
Air Blank	0.000	19:50
Control Test	0.081	19:51
Air Blank	0.000	19:51
Subject Sample # 1	0.239	19:52
Breath Volume	1.636	19:52
Air Blank	0.000	19:52
Air Blank	0.000	19:54
Subject Sample # 2	0.229	19:55
Breath Volume	2.171	19:55
Air Blank	0.000	19:56
Control Test	0.080	19:56
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

---

Data Download Date/Time: 07/29/2021 14:51

**Florida Department of Law Enforcement**

**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BASFORD, TRAVIS, L  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: BASFORD, TRAVIS  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/21/2021  
VIOLATION CODE: BUI  
ARREST TIME: 22:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 07/04/2021 23:29:12

---

Diagnostics Check	OK	23:26	
Air Blank	0.000	23:26	
Control Test	0.079	23:26	
Air Blank	0.000	23:27	
Subject Sample	# 1	23:27	Subject Test Refused
Breath Volume	0.000	23:27	
Air Blank	0.000	23:28	
Control Test	0.080	23:28	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

---

Data Download Date/Time: 07/29/2021 14:51





**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BASFORD,NICOLE,M  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: BASFORD,NICOLE  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1060911  
DRY GAS EXPIRY DATE: 12/21/2021  
VIOLATION CODE: BUI  
ARREST TIME: 16:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:46

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 07/25/2021 17:14:31

---

Diagnostics Check	OK	17:08
Air Blank	0.000	17:08
Control Test	0.081	17:09
Air Blank	0.000	17:09
Subject Sample # 1	0.000	17:10
Breath Volume	4.019	17:10
Air Blank	0.000	17:10
Air Blank	0.000	17:12
Subject Sample # 2	0.000	17:13
Breath Volume	1.468	17:13
Air Blank	0.000	17:13
Control Test	0.080	17:13
Air Blank	0.000	17:14
Diagnostics Check	OK	17:14

---

Data Download Date/Time: 07/29/2021 14:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

INSTRUMENT REGISTERED TO: FFWCC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAYCHOFF, COLLIN, T  
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)  
ARRESTING OFFICER: WAYCHOFF, COLLIN  
ARRESTING OFFICER AGENCY: FFWCC (Lake City)  
LAST AGENCY INSPECTION: 06/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: BUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: WORLEY

DATE/TIME OF TEST: 07/04/2021 22:34:22

---

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.079	22:28
Air Blank	0.000	22:28
Subject Sample # 1	0.057	22:29
Breath Volume	3.085	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.055	22:32
Breath Volume	3.078	22:32
Air Blank	0.000	22:33
Control Test	0.077	22:33
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

---

Data Download Date/Time: 09/08/2021 09:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

INSTRUMENT REGISTERED TO: FFWCC TAMPA  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PULASKI,DAMON,J  
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)  
ARRESTING OFFICER: PULASKI,DAMON  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 07/13/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22419080A2  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: BUI  
ARREST TIME: 18:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:21

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 07/16/2021 18:53:00

---

Diagnostics Check	OK	18:45
Air Blank	0.000	18:45
Control Test	0.077	18:46
Air Blank	0.000	18:46
Subject Sample # 1	0.087	18:48
Breath Volume	3.898	18:48
Air Blank	0.000	18:48
Air Blank	0.000	18:50
Subject Sample # 2	0.084	18:51
Breath Volume	3.386	18:51
Air Blank	0.000	18:52
Control Test	0.077	18:52
Air Blank	0.000	18:52
Diagnostics Check	OK	18:53

---

Data Download Date/Time: 08/30/2021 11:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000905

INSTRUMENT REGISTERED TO: FFWCC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICE, MICHAEL,  
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)  
ARRESTING OFFICER: DIAS, ASA  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG928002  
DRY GAS EXPIRY DATE: 10/07/2021  
VIOLATION CODE: BUI  
ARREST TIME: 21:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 07/16/2021 21:46:48

---

Diagnostics Check	OK	21:42	
Air Blank	0.000	21:42	
Control Test	0.078	21:43	
Air Blank	0.000	21:43	
Subject Sample	# 1	21:45	Subject Test Refused
Breath Volume	0.000	21:45	
Air Blank	0.000	21:45	
Control Test	0.079	21:46	
Air Blank	0.000	21:46	
Diagnostics Check	OK	21:46	

---

Data Download Date/Time: 08/19/2021 09:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GWODZ,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Bradenton Beach PD  
ARRESTING OFFICER: WENTLAND,ALEX  
ARRESTING OFFICER AGENCY: Bradenton PD  
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BRECHUN

DATE/TIME OF TEST: 07/14/2021 02:43:33

---

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.077	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.212	02:38
Breath Volume	3.656	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.211	02:42
Breath Volume	3.109	02:42
Air Blank	0.000	02:42
Control Test	0.077	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

---

Data Download Date/Time: 08/09/2021 14:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JIMENEZ,WILLIAM,  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: JIMENEZ,WILLIAM  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 17:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 07/14/2021 21:26:15

---

Diagnostics Check	OK	21:19
Air Blank	0.000	21:20
Control Test	0.079	21:20
Air Blank	0.000	21:21
Subject Sample # 1	0.078	21:22
Breath Volume	1.210	21:22
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample # 2	0.066	21:24
Breath Volume	1.207	21:24
Air Blank	0.000	21:25
Control Test	0.078	21:25
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

---

Data Download Date/Time: 08/09/2021 14:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: CID BERISTAIN

DATE/TIME OF TEST: 07/15/2021 04:22:51

---

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.079	04:16
Air Blank	0.000	04:17
Subject Sample # 1	0.191	04:18
Breath Volume	3.000	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.181	04:21
Breath Volume	2.417	04:21
Air Blank	0.000	04:21
Control Test	0.079	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

---

Data Download Date/Time: 08/09/2021 14:45



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LORD,JOSHUA,N  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LORD,JOSHUA  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 13:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: MEJIA VELASQUEZ

DATE/TIME OF TEST: 07/18/2021 14:33:07

---

Diagnostics Check		OK	14:21	
Air Blank		0.000	14:21	
Control Test		0.079	14:21	
Air Blank		0.000	14:22	
Subject Sample	# 1	0.177	14:25	Volume Not Met
Breath Volume		0.722	14:25	
Air Blank		0.000	14:26	
Air Blank		0.000	14:28	
Subject Sample	# 2	0.180	14:28	
Breath Volume		1.808	14:28	
Air Blank		0.000	14:29	
Air Blank		0.000	14:31	
Subject Sample	# 3	0.187	14:31	
Breath Volume		2.855	14:31	
Air Blank		0.000	14:32	
Control Test		0.079	14:32	
Air Blank		0.000	14:33	
Diagnostics Check		OK	14:33	

---

Data Download Date/Time: 08/09/2021 14:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OSTERTAG,SHELBY,R  
BREATH TEST OPERATOR AGENCY: Bradenton PD  
ARRESTING OFFICER: OSTERTAG,SHELBY  
ARRESTING OFFICER AGENCY: Bradenton PD  
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/22/2021 20:26:31

---

Diagnostics Check	OK	20:15
Air Blank	0.000	20:16
Control Test	0.079	20:16
Air Blank	0.000	20:17
Subject Sample # 1	0.131	20:19
Breath Volume	1.128	20:19
Air Blank	0.000	20:20
Air Blank	0.000	20:21
Subject Sample # 2	0.116	20:25
Breath Volume	1.414	20:25
Air Blank	0.000	20:25
Control Test	0.077	20:25
Air Blank	0.000	20:26
Diagnostics Check	OK	20:26

---

Data Download Date/Time: 08/09/2021 14:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: NUNEZ SANCHEZ

DATE/TIME OF TEST: 07/24/2021 02:14:45

---

Diagnostics Check		OK	02:08
Air Blank		0.000	02:09
Control Test		0.078	02:09
Air Blank		0.000	02:09
Subject Sample	# 1	0.208	02:10
Breath Volume		1.660	02:10
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.205	02:13
Breath Volume		2.937	02:13
Air Blank		0.000	02:13
Control Test		0.078	02:14
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

---

Data Download Date/Time: 08/09/2021 14:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LEDFORD,TAYLOR  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1408275  
DRY GAS EXPIRY DATE: 11/19/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: BERMUDEZ PEREYRA

DATE/TIME OF TEST: 07/25/2021 02:15:59

---

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.078	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.182	02:11
Breath Volume	1.343	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.197	02:14
Breath Volume	2.187	02:14
Air Blank	0.000	02:15
Control Test	0.078	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

---

Data Download Date/Time: 08/09/2021 14:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VAN HEERDEN, TY, C  
BREATH TEST OPERATOR AGENCY: GULF COUNTY SO  
ARRESTING OFFICER: CHASE, JEREMY  
ARRESTING OFFICER AGENCY: PORT ST. JOE PD  
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A2  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ADKISON

DATE/TIME OF TEST: 07/25/2021 02:59:29

---

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:58	
Control Test	0.073	02:58	Control Outside Tolerance
Air Blank	0.000	02:59	

---

Data Download Date/Time: 09/03/2021 08:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORBETA,JORGE,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: OROL,MANUEL  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 06/05/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1380119  
DRY GAS EXPIRY DATE: 09/22/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: PIERRE-LOUIS

DATE/TIME OF TEST: 07/03/2021 23:05:14

---

Diagnostics Check	OK	22:59	
Air Blank	0.000	23:00	
Control Test	0.079	23:00	
Air Blank	0.000	23:00	
Subject Sample	# 1	23:03	Subject Test Refused
Breath Volume	0.000	23:03	
Air Blank	0.000	23:04	
Control Test	0.080	23:04	
Air Blank	0.000	23:05	
Diagnostics Check	OK	23:05	

---

Data Download Date/Time: 07/29/2021 03:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORBETA,JORGE,  
BREATH TEST OPERATOR AGENCY: South Miami PD  
ARRESTING OFFICER: ORBETA,JORGE  
ARRESTING OFFICER AGENCY: South Miami PD  
LAST AGENCY INSPECTION: 06/05/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1380119  
DRY GAS EXPIRY DATE: 09/22/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: HERSKOWITZ

DATE/TIME OF TEST: 07/09/2021 04:14:49

---

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.081	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.165	04:09
Breath Volume	1.718	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.168	04:13
Breath Volume	1.832	04:13
Air Blank	0.000	04:13
Control Test	0.079	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

---

Data Download Date/Time: 07/29/2021 03:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIOS,SEAN,O  
BREATH TEST OPERATOR AGENCY: Village of Pinecrest  
ARRESTING OFFICER: MARIN,STEPHANIE  
ARRESTING OFFICER AGENCY: Village of Pinecrest  
LAST AGENCY INSPECTION: 06/05/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1380119  
DRY GAS EXPIRY DATE: 09/22/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: CENTENO

DATE/TIME OF TEST: 07/11/2021 01:14:35

---

Diagnostics Check		OK	01:07
Air Blank		0.000	01:08
Control Test		0.081	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.000	01:10
Breath Volume		1.257	01:10
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.000	01:13
Breath Volume		1.824	01:13
Air Blank		0.000	01:13
Control Test		0.079	01:13
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

---

Data Download Date/Time: 07/29/2021 03:26



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORBETA,JORGE,  
BREATH TEST OPERATOR AGENCY: South Miami PD  
ARRESTING OFFICER: ORBETA,JORGE  
ARRESTING OFFICER AGENCY: South Miami PD  
LAST AGENCY INSPECTION: 06/05/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1380119  
DRY GAS EXPIRY DATE: 09/22/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: GONZALEZMIRANDA

DATE/TIME OF TEST: 07/12/2021 03:11:10

---

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:07	
Control Test	0.080	03:08	
Air Blank	0.000	03:08	
Subject Sample # 1	0.204	03:09	
Breath Volume	3.246	03:09	
Air Blank		03:09	RFI Detect
Air Blank	0.000	03:10	
Control Test	0.080	03:10	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

---

Data Download Date/Time: 07/29/2021 03:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORBETA,JORGE,  
BREATH TEST OPERATOR AGENCY: South Miami PD  
ARRESTING OFFICER: ORBETA,JORGE  
ARRESTING OFFICER AGENCY: South Miami PD  
LAST AGENCY INSPECTION: 06/05/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1380119  
DRY GAS EXPIRY DATE: 09/22/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: GONZALEZ MIRANDA

DATE/TIME OF TEST: 07/12/2021 03:19:10

---

Diagnostics Check		OK	03:12
Air Blank		0.000	03:13
Control Test		0.079	03:13
Air Blank		0.000	03:14
Subject Sample	# 1	0.224	03:14
Breath Volume		3.570	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.226	03:17
Breath Volume		3.144	03:17
Air Blank		0.000	03:18
Control Test		0.079	03:18
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

---

Data Download Date/Time: 07/29/2021 03:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORBETA,JORGE,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: OROL,MANUEL  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 06/05/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1380119  
DRY GAS EXPIRY DATE: 09/22/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: PHIPPS

DATE/TIME OF TEST: 07/12/2021 06:11:04

---

Diagnostics Check	OK	06:04
Air Blank	0.000	06:05
Control Test	0.079	06:05
Air Blank	0.000	06:05
Subject Sample # 1	0.096	06:06
Breath Volume	3.203	06:06
Air Blank	0.000	06:07
Air Blank	0.000	06:08
Subject Sample # 2	0.097	06:09
Breath Volume	3.222	06:09
Air Blank	0.000	06:10
Control Test	0.080	06:10
Air Blank	0.000	06:10
Diagnostics Check	OK	06:11

---

Data Download Date/Time: 07/29/2021 03:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIOS,SEAN,O  
BREATH TEST OPERATOR AGENCY: Village of Pinecrest  
ARRESTING OFFICER: RIOS,SEAN  
ARRESTING OFFICER AGENCY: Village of Pinecrest  
LAST AGENCY INSPECTION: 06/05/2021

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1380119  
DRY GAS EXPIRY DATE: 09/22/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: FUNDORA

DATE/TIME OF TEST: 07/16/2021 05:02:26

---

Diagnostics Check	OK	04:59	
Air Blank	0.000	04:59	
Control Test	0.079	04:59	
Air Blank	0.000	05:00	
Subject Sample	# 1	05:01	Subject Test Refused
Breath Volume	0.000	05:01	
Air Blank	0.000	05:01	
Control Test	0.080	05:01	
Air Blank	0.000	05:02	
Diagnostics Check	OK	05:02	

---

Data Download Date/Time: 07/29/2021 03:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORBETA,JORGE,  
BREATH TEST OPERATOR AGENCY: South Miami PD  
ARRESTING OFFICER: ORBETA,JORGE  
ARRESTING OFFICER AGENCY: South Miami PD  
LAST AGENCY INSPECTION: 06/05/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1380119  
DRY GAS EXPIRY DATE: 09/22/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: ACOSTA CARMONA

DATE/TIME OF TEST: 07/17/2021 04:45:09

---

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.079	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.128	04:40
Breath Volume	4.160	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:42
Subject Sample # 2	0.127	04:43
Breath Volume	3.929	04:43
Air Blank	0.000	04:44
Control Test	0.080	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

---

Data Download Date/Time: 07/29/2021 03:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORBETA,JORGE,  
BREATH TEST OPERATOR AGENCY: South Miami PD  
ARRESTING OFFICER: ORBETA,JORGE  
ARRESTING OFFICER AGENCY: South Miami PD  
LAST AGENCY INSPECTION: 06/05/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1380119  
DRY GAS EXPIRY DATE: 09/22/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: BARROSO RODRIGUEZ

DATE/TIME OF TEST: 07/19/2021 03:49:06

---

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.233	03:43
Breath Volume	2.214	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.221	03:47
Breath Volume	3.093	03:47
Air Blank	0.000	03:48
Control Test	0.079	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

---

Data Download Date/Time: 07/29/2021 03:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORBETA,JORGE,  
BREATH TEST OPERATOR AGENCY: South Miami PD  
ARRESTING OFFICER: ORBETA,JORGE  
ARRESTING OFFICER AGENCY: South Miami PD  
LAST AGENCY INSPECTION: 06/05/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1380119  
DRY GAS EXPIRY DATE: 09/22/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: BALLESTAS

DATE/TIME OF TEST: 07/23/2021 04:39:26

---

Diagnostics Check	OK	04:35	
Air Blank	0.000	04:35	
Control Test	0.081	04:36	
Air Blank	0.000	04:36	
Subject Sample	# 1	04:38	Subject Test Refused
Breath Volume	0.000	04:38	
Air Blank	0.000	04:38	
Control Test	0.079	04:38	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:39	

---

Data Download Date/Time: 07/29/2021 03:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MUNOZ,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Village of Pinecrest  
ARRESTING OFFICER: RODRIGUEZ,ROBERTO  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 07/29/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 181622  
DRY GAS EXPIRY DATE: 10/05/2021  
VIOLATION CODE: BUI  
ARREST TIME: 21:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: GIL

DATE/TIME OF TEST: 07/29/2021 22:57:33

---

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:52	
Control Test	0.080	22:52	
Air Blank	0.000	22:53	
Subject Sample	# 1	22:56	Subject Test Refused
Breath Volume	0.015	22:56	
Air Blank	0.000	22:56	
Control Test	0.081	22:56	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

---

Data Download Date/Time: 08/31/2021 20:32



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUMLEY,JOSEPH,A  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: CISNERO,JOHNNY  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 08:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:32

SUBJECT'S LAST NAME: TOSCANOFERNANDEZ

DATE/TIME OF TEST: 07/02/2021 11:19:26

---

Diagnostics Check	OK	11:05	
Air Blank	0.000	11:05	
Control Test	0.078	11:05	
Air Blank	0.000	11:06	
Subject Sample # 1	0.129	11:06	
Breath Volume	2.945	11:06	
Air Blank	0.000	11:07	
Air Blank	0.000	11:09	
Subject Sample # 2		11:12	No Sample Provided
Breath Volume	0.000	11:12	
Air Blank	0.000	11:12	
Air Blank	0.000	11:14	
Subject Sample # 3		11:18	No Sample Provided
Breath Volume	0.000	11:18	
Air Blank	0.000	11:18	
Control Test	0.079	11:18	
Air Blank	0.000	11:19	
Diagnostics Check	OK	11:19	

---

Data Download Date/Time: 07/14/2021 11:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: ZYNDA,T  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: CARRIE

DATE/TIME OF TEST: 07/15/2021 03:14:49

---

Diagnostics Check		OK	03:04
Air Blank		0.000	03:04
Control Test		0.078	03:04
Air Blank		0.000	03:05
Subject Sample	# 1	0.177	03:05
Breath Volume		1.300	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.148	03:10
Breath Volume		1.843	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:12
Subject Sample	# 3	0.163	03:13
Breath Volume		1.328	03:13
Air Blank		0.000	03:13
Control Test		0.078	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

---

Data Download Date/Time: 08/02/2021 15:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GILLIARD,M,  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: HENDERSON,K  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 14:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:23

SUBJECT'S LAST NAME: ITZKOWITZ

DATE/TIME OF TEST: 07/15/2021 16:01:52

---

Diagnostics Check	OK	15:53
Air Blank	0.000	15:54
Control Test	0.078	15:54
Air Blank	0.000	15:54
Subject Sample # 1	0.178	15:56
Breath Volume	1.132	15:56
Air Blank	0.000	15:56
Air Blank	0.000	15:58
Subject Sample # 2	0.179	16:00
Breath Volume	1.687	16:00
Air Blank	0.000	16:00
Control Test	0.077	16:01
Air Blank	0.000	16:01
Diagnostics Check	OK	16:01

---

Data Download Date/Time: 08/02/2021 15:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SAINT-FORT,P,  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: WADDELL,N  
ARRESTING OFFICER AGENCY: FHP TROOP F  
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 13:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:24

SUBJECT'S LAST NAME: BALZORA

DATE/TIME OF TEST: 07/17/2021 16:07:17

---

Diagnostics Check		OK	15:52	
Air Blank		0.000	15:53	
Control Test		0.078	15:53	
Air Blank		0.000	15:54	
Subject Sample	# 1	0.067	15:57	Volume Not Met
Breath Volume		0.535	15:57	
Air Blank		0.000	15:57	
Air Blank		0.000	15:59	
Subject Sample	# 2	0.081	16:02	
Breath Volume		1.644	16:02	
Air Blank		0.000	16:03	
Air Blank		0.000	16:05	
Subject Sample	# 3	0.076	16:05	
Breath Volume		1.195	16:05	
Air Blank		0.000	16:06	
Control Test		0.077	16:06	
Air Blank		0.000	16:07	
Diagnostics Check		OK	16:07	

---

Data Download Date/Time: 08/02/2021 15:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUGO,J,B  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: DIKEMAN,H  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 07/18/2021 02:52:39

---

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.078	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.155	02:48
Breath Volume	1.597	02:48
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.162	02:51
Breath Volume	1.304	02:51
Air Blank	0.000	02:51
Control Test	0.079	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

---

Data Download Date/Time: 08/02/2021 15:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TARR,L,  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: TARR,L  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/19/2021 00:35:41

---

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.079	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.137	00:31
Breath Volume	1.277	00:31
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.144	00:34
Breath Volume	1.773	00:34
Air Blank	0.000	00:34
Control Test	0.078	00:35
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

---

Data Download Date/Time: 08/02/2021 15:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,ZACHARY,H  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: WILLIAMS,ZACHARY  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24119080A1  
DRY GAS EXPIRY DATE: 11/05/2021  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ZWART

DATE/TIME OF TEST: 07/20/2021 02:20:05

---

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.079	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.000	02:15
Breath Volume	3.050	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.000	02:18
Breath Volume	2.589	02:18
Air Blank	0.000	02:19
Control Test	0.078	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

---

Data Download Date/Time: 08/02/2021 15:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORRIS,THOMAS,L  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: TURNER,J  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/23/2021 03:48:27

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.079	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.198	03:42
Breath Volume	1.414	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:44
Subject Sample # 2	0.201	03:46
Breath Volume	1.195	03:46
Air Blank	0.000	03:47
Control Test	0.079	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

---

Data Download Date/Time: 08/02/2021 15:19



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENNEDY,DANIEL,L  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: KENNEDY,DANIEL  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: KRES

DATE/TIME OF TEST: 07/24/2021 02:44:02

---

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.079	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.042	02:39
Breath Volume	1.703	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.041	02:42
Breath Volume	2.117	02:42
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:44

---

Data Download Date/Time: 08/02/2021 15:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAWSON,MAXWELL,C  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: CRAIG,JOE  
ARRESTING OFFICER AGENCY: NAPLES PD  
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: CARMINE

DATE/TIME OF TEST: 07/24/2021 03:27:13

---

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.181	03:22
Breath Volume	2.453	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.173	03:25
Breath Volume	2.269	03:25
Air Blank	0.000	03:26
Control Test	0.077	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

---

Data Download Date/Time: 08/02/2021 15:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENNEDY,DANIEL,L  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: KENNEDY,DANIEL  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: CHILDS

DATE/TIME OF TEST: 07/25/2021 00:36:08

---

Diagnostics Check	OK	00:29
Air Blank	0.000	00:29
Control Test	0.079	00:30
Air Blank	0.000	00:30
Subject Sample # 1	0.159	00:31
Breath Volume	2.585	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:33
Subject Sample # 2	0.162	00:34
Breath Volume	1.539	00:34
Air Blank	0.000	00:35
Control Test	0.079	00:35
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

---

Data Download Date/Time: 08/02/2021 15:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,S,  
BREATH TEST OPERATOR AGENCY: NAPLES PD  
ARRESTING OFFICER: HARRINGTON,J  
ARRESTING OFFICER AGENCY: NAPLES PD  
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 07/25/2021 22:50:10

---

Diagnostics Check	OK	22:44
Air Blank	0.000	22:44
Control Test	0.079	22:44
Air Blank	0.000	22:45
Subject Sample # 1	0.177	22:45
Breath Volume	2.816	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.171	22:48
Breath Volume	2.542	22:48
Air Blank	0.000	22:49
Control Test	0.079	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

---

Data Download Date/Time: 08/02/2021 15:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLOPS,C,  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: BELARDO,S  
ARRESTING OFFICER AGENCY: MARCO ISLAND PD  
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: WIGDAHL

DATE/TIME OF TEST: 07/28/2021 01:22:18

---

Diagnostics Check		OK	01:16
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.323	01:17
Breath Volume		2.804	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.309	01:20
Breath Volume		2.218	01:20
Air Blank		0.000	01:21
Control Test		0.079	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

---

Data Download Date/Time: 08/02/2021 15:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUTLER JR,S,A  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: ENDELMANN,K  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: ROANHORSE

DATE/TIME OF TEST: 07/28/2021 03:49:40

---

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.079	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.252	03:45
Breath Volume	2.171	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.252	03:48
Breath Volume	1.179	03:48
Air Blank	0.000	03:48
Control Test	0.079	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

---

Data Download Date/Time: 08/02/2021 15:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VALDEZ,D,R  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: YANEZ,R  
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)  
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: MAJOR

DATE/TIME OF TEST: 07/30/2021 23:54:43

---

Diagnostics Check	OK	23:48
Air Blank	0.000	23:48
Control Test	0.079	23:49
Air Blank	0.000	23:49
Subject Sample # 1	0.152	23:50
Breath Volume	1.496	23:50
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.155	23:53
Breath Volume	2.175	23:53
Air Blank	0.000	23:53
Control Test	0.079	23:54
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

---

Data Download Date/Time: 08/02/2021 15:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOEL,YVES,  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: PEREZ,M  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CISNEROS

DATE/TIME OF TEST: 07/31/2021 04:31:03

---

Diagnostics Check		OK	04:17	
Air Blank		0.000	04:18	
Control Test		0.078	04:18	
Air Blank		0.000	04:19	
Subject Sample	# 1	0.193	04:22	Volume Not Met
Breath Volume		0.078	04:22	
Air Blank		0.000	04:22	
Air Blank		0.000	04:24	
Subject Sample	# 2	0.190	04:25	
Breath Volume		1.246	04:25	
Air Blank		0.000	04:26	
Air Blank		0.000	04:28	
Subject Sample	# 3	0.189	04:29	
Breath Volume		1.445	04:29	
Air Blank		0.000	04:30	
Control Test		0.078	04:30	
Air Blank		0.000	04:30	
Diagnostics Check		OK	04:31	

---

Data Download Date/Time: 08/02/2021 15:20



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUGO,JONATHAN,B  
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO  
ARRESTING OFFICER: MILLER,K  
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO  
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/31/2021 23:57:36

---

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.079	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.043	23:53
Breath Volume	1.605	23:53
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.042	23:56
Breath Volume	1.242	23:56
Air Blank	0.000	23:56
Control Test	0.079	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

---

Data Download Date/Time: 08/02/2021 15:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANCILLA,RICHARD,P  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: SMITH,TAYLOR  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 03:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: KEENE

DATE/TIME OF TEST: 07/18/2021 04:03:29

---

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.079	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.199	03:59
Breath Volume	1.628	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.198	04:02
Breath Volume	1.605	04:02
Air Blank	0.000	04:02
Control Test	0.078	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

---

Data Download Date/Time: 08/02/2021 09:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZAMORA,DEP,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: CCSO,CCSO  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 03:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: KEENE

DATE/TIME OF TEST: 07/18/2021 08:41:28

---

Diagnostics Check	OK	08:35
Air Blank	0.000	08:35
Control Test	0.080	08:35
Air Blank	0.000	08:36
Subject Sample # 1	0.076	08:37
Breath Volume	1.722	08:37
Air Blank	0.000	08:37
Air Blank	0.000	08:39
Subject Sample # 2	0.080	08:40
Breath Volume	2.195	08:40
Air Blank	0.000	08:40
Control Test	0.079	08:40
Air Blank	0.000	08:41
Diagnostics Check	OK	08:41

---

Data Download Date/Time: 08/02/2021 09:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZAMORA,DEP,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: CCSO,CCSO  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 03:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: KEENE

DATE/TIME OF TEST: 07/18/2021 11:01:16

---

Diagnostics Check	OK	10:55
Air Blank	0.000	10:55
Control Test	0.079	10:55
Air Blank	0.000	10:56
Subject Sample # 1	0.028	10:56
Breath Volume	2.480	10:56
Air Blank	0.000	10:57
Air Blank	0.000	10:59
Subject Sample # 2	0.026	10:59
Breath Volume	1.539	10:59
Air Blank	0.000	11:00
Control Test	0.080	11:00
Air Blank	0.000	11:01
Diagnostics Check	OK	11:01

---

Data Download Date/Time: 08/02/2021 09:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANO,MARDEE,G  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: HARTMANSTORFER,CHRIS  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 13:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:44

SUBJECT'S LAST NAME: PETTIT

DATE/TIME OF TEST: 07/18/2021 15:10:17

---

Diagnostics Check	OK	15:04
Air Blank	0.000	15:04
Control Test	0.078	15:04
Air Blank	0.000	15:05
Subject Sample # 1	0.000	15:05
Breath Volume	3.242	15:05
Air Blank	0.000	15:06
Air Blank	0.000	15:08
Subject Sample # 2	0.000	15:08
Breath Volume	4.992	15:08
Air Blank	0.000	15:09
Control Test	0.080	15:09
Air Blank	0.000	15:10
Diagnostics Check	OK	15:10

---

Data Download Date/Time: 08/02/2021 09:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDROP,P,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: FARLEY,OFC  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: WILMES

DATE/TIME OF TEST: 07/19/2021 22:08:51

---

Diagnostics Check	OK	22:01
Air Blank	0.000	22:02
Control Test	0.080	22:02
Air Blank	0.000	22:03
Subject Sample # 1	0.060	22:04
Breath Volume	1.804	22:04
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.057	22:07
Breath Volume	1.796	22:07
Air Blank	0.000	22:07
Control Test	0.079	22:08
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

---

Data Download Date/Time: 08/02/2021 09:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZAMORA,DEP,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: CCSO,DEP  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/26/2021 10:50:04

---

Diagnostics Check	OK	10:44
Air Blank	0.000	10:44
Control Test	0.078	10:44
Air Blank	0.000	10:45
Subject Sample # 1	0.033	10:45
Breath Volume	3.312	10:45
Air Blank	0.000	10:46
Air Blank	0.000	10:48
Subject Sample # 2	0.031	10:48
Breath Volume	3.945	10:48
Air Blank	0.000	10:49
Control Test	0.078	10:49
Air Blank	0.000	10:49
Diagnostics Check	OK	10:50

---

Data Download Date/Time: 08/02/2021 09:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COURTNEY,PATRICK,R  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: GANK,T  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 07/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: DAGG

DATE/TIME OF TEST: 07/31/2021 04:16:38

---

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:09	
Control Test	0.079	04:09	
Air Blank	0.000	04:10	
Subject Sample	# 1 0.188	04:11	
Breath Volume	2.425	04:11	
Air Blank	0.000	04:11	
Air Blank	0.000	04:13	
Subject Sample	# 2	04:15	Slope Not Met
Breath Volume	0.945	04:15	
Air Blank	0.000	04:15	
Control Test	0.079	04:16	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

---

Data Download Date/Time: 08/02/2021 09:59



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COURTNEY,PATRICK,R  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: GANK,T  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 07/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: DAGG

DATE/TIME OF TEST: 07/31/2021 04:34:14

---

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.079	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.192	04:25
Breath Volume	2.242	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.168	04:29
Breath Volume	1.734	04:29
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 3	0.178	04:32
Breath Volume	2.058	04:32
Air Blank	0.000	04:33
Control Test	0.078	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

---

Data Download Date/Time: 08/02/2021 09:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCOTT, BRADLEY, W  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: ANDERSON, JACOB  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 07/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A3  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 12:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:03

SUBJECT'S LAST NAME: KUJAWA

DATE/TIME OF TEST: 07/31/2021 13:32:55

---

Diagnostics Check	OK	13:26
Air Blank	0.000	13:26
Control Test	0.079	13:26
Air Blank	0.000	13:27
Subject Sample # 1	0.000	13:28
Breath Volume	1.718	13:28
Air Blank	0.000	13:28
Air Blank	0.000	13:30
Subject Sample # 2	0.000	13:31
Breath Volume	1.687	13:31
Air Blank	0.000	13:32
Control Test	0.080	13:32
Air Blank	0.000	13:32
Diagnostics Check	OK	13:32

---

Data Download Date/Time: 08/02/2021 09:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIVINGSTON,JESSICA,L  
BREATH TEST OPERATOR AGENCY: Glades County SO  
ARRESTING OFFICER: YOUNGMAN,TALON  
ARRESTING OFFICER AGENCY: Glades County SO  
LAST AGENCY INSPECTION: 06/12/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: ADMIN  
ARREST TIME: 17:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: CREWS

DATE/TIME OF TEST: 07/16/2021 18:39:00

---

Diagnostics Check	OK	18:32
Air Blank	0.000	18:33
Control Test	0.077	18:33
Air Blank	0.000	18:34
Subject Sample # 1	0.000	18:34
Breath Volume	2.230	18:34
Air Blank	0.000	18:35
Air Blank	0.000	18:37
Subject Sample # 2	0.000	18:37
Breath Volume	2.757	18:37
Air Blank	0.000	18:38
Control Test	0.077	18:38
Air Blank	0.000	18:38
Diagnostics Check	OK	18:39

---

Data Download Date/Time: 07/20/2021 15:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EPPS,QUINTELLA,  
BREATH TEST OPERATOR AGENCY: Glades County SO  
ARRESTING OFFICER: HUNTER,MITCHELL  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 07/20/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MOLINERO

DATE/TIME OF TEST: 07/28/2021 22:19:53

---

Diagnostics Check		OK	22:12
Air Blank		0.000	22:12
Control Test		0.077	22:12
Air Blank		0.000	22:13
Subject Sample	# 1	0.239	22:15
Breath Volume		1.718	22:15
Air Blank		0.000	22:16
Air Blank		0.000	22:17
Subject Sample	# 2	0.229	22:18
Breath Volume		1.695	22:18
Air Blank		0.000	22:18
Control Test		0.077	22:19
Air Blank		0.000	22:19
Diagnostics Check		OK	22:19

---

Data Download Date/Time: 08/26/2021 10:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000950

INSTRUMENT REGISTERED TO: HAVANA PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENING,FRED,C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: HENING,FRED  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: PROB/PAROL  
ARREST TIME: 00:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:17

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/08/2021 11:50:16

---

Diagnostics Check	OK	11:44
Air Blank	0.000	11:44
Control Test	0.078	11:45
Air Blank	0.000	11:45
Subject Sample # 1	0.000	11:46
Breath Volume	1.890	11:46
Air Blank	0.000	11:46
Air Blank	0.000	11:48
Subject Sample # 2	0.000	11:48
Breath Volume	1.601	11:48
Air Blank	0.000	11:49
Control Test	0.077	11:49
Air Blank	0.000	11:50
Diagnostics Check	OK	11:50

---

Data Download Date/Time: 07/08/2021 12:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000950

INSTRUMENT REGISTERED TO: HAVANA PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMALL,BRIAN,A  
BREATH TEST OPERATOR AGENCY: Havana PD  
ARRESTING OFFICER: SMALL,BRIAN  
ARRESTING OFFICER AGENCY: Havana PD  
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: ZERO TOL  
ARREST TIME: 18:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/23/2021 18:37:46

---

Diagnostics Check	OK	18:31
Air Blank	0.000	18:32
Control Test	0.079	18:32
Air Blank	0.000	18:33
Subject Sample # 1	0.000	18:33
Breath Volume	3.046	18:33
Air Blank	0.000	18:34
Air Blank	0.000	18:36
Subject Sample # 2	0.000	18:36
Breath Volume	2.878	18:36
Air Blank	0.000	18:36
Control Test	0.078	18:37
Air Blank	0.000	18:37
Diagnostics Check	OK	18:37

---

Data Download Date/Time: 08/09/2021 13:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000950

INSTRUMENT REGISTERED TO: HAVANA PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMALL,BRIAN,A  
BREATH TEST OPERATOR AGENCY: Havana PD  
ARRESTING OFFICER: SMALL,BRIAN  
ARRESTING OFFICER AGENCY: Havana PD  
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: ZERO TOL  
ARREST TIME: 18:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/28/2021 19:58:51

---

Diagnostics Check	OK	19:53
Air Blank	0.000	19:53
Control Test	0.078	19:53
Air Blank	0.000	19:54
Subject Sample # 1	0.000	19:54
Breath Volume	1.687	19:54
Air Blank	0.000	19:55
Air Blank	0.000	19:57
Subject Sample # 2	0.000	19:57
Breath Volume	2.289	19:57
Air Blank	0.000	19:57
Control Test	0.079	19:58
Air Blank	0.000	19:58
Diagnostics Check	OK	19:58

---

Data Download Date/Time: 08/09/2021 13:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SWINDLE, CHARLES,  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: SWINDLE, CHARLES  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 07/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34520080A1  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: LAMPERT

DATE/TIME OF TEST: 07/10/2021 23:16:20

---

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.080	23:10
Air Blank	0.000	23:11
Subject Sample # 1	0.000	23:11
Breath Volume	2.003	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.000	23:14
Breath Volume	2.156	23:14
Air Blank	0.000	23:15
Control Test	0.079	23:15
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

---

Data Download Date/Time: 08/11/2021 11:11



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON,TREVOR,A  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: KALINOWSKI,ADAM  
ARRESTING OFFICER AGENCY: Jefferson County SO  
LAST AGENCY INSPECTION: 07/07/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34520080A1  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:04

SUBJECT'S LAST NAME: AKERS

DATE/TIME OF TEST: 07/28/2021 07:31:33

---

Diagnostics Check	OK	07:25
Air Blank	0.000	07:25
Control Test	0.078	07:25
Air Blank	0.000	07:26
Subject Sample # 1	0.091	07:26
Breath Volume	3.609	07:26
Air Blank	0.000	07:27
Air Blank	0.000	07:29
Subject Sample # 2	0.088	07:30
Breath Volume	3.511	07:30
Air Blank	0.000	07:30
Control Test	0.079	07:30
Air Blank	0.000	07:31
Diagnostics Check	OK	07:31

---

Data Download Date/Time: 08/11/2021 11:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROY,GARY,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: ROY,GARY  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/07/2021 23:10:30

---

Diagnostics Check	OK	23:01
Air Blank	0.000	23:02
Control Test	0.081	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.159	23:05
Breath Volume	1.496	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.155	23:09
Breath Volume	1.753	23:09
Air Blank	0.000	23:09
Control Test	0.081	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

---

Data Download Date/Time: 08/09/2021 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: VOLLERTSEN,CRAIG  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 07/10/2021 01:37:10

---

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.081	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.000	01:33
Breath Volume	1.507	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.000	01:35
Breath Volume	1.933	01:35
Air Blank	0.000	01:36
Control Test	0.081	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

---

Data Download Date/Time: 08/09/2021 11:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REAGLE,ASHLEY,R  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: REAGLE,ASHLEY  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: VARGAS TEJEDA

DATE/TIME OF TEST: 07/11/2021 03:26:39

---

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.080	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.144	03:22
Breath Volume	1.632	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.143	03:25
Breath Volume	1.171	03:25
Air Blank	0.000	03:25
Control Test	0.083	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

---

Data Download Date/Time: 08/09/2021 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINA,RYAN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MARTINA,RYAN  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/11/2021 23:56:23

---

Diagnostics Check	OK	23:49
Air Blank	0.000	23:49
Control Test	0.080	23:50
Air Blank	0.000	23:50
Subject Sample # 1	0.302	23:51
Breath Volume	1.648	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:53
Subject Sample # 2	0.297	23:54
Breath Volume	1.410	23:54
Air Blank	0.000	23:55
Control Test	0.079	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

---

Data Download Date/Time: 08/09/2021 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: NORTHWAY,GEORGIA  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: BLAKE

DATE/TIME OF TEST: 07/16/2021 02:26:30

---

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:20	
Control Test	0.080	02:21	
Air Blank	0.000	02:21	
Subject Sample	# 1 0.232	02:24	Volume Not Met
Breath Volume	0.140	02:24	
Air Blank		02:25	Purge Fail
Air Blank		02:26	Purge Fail

---

Data Download Date/Time: 08/09/2021 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: NORTHWAY,GEORGIA  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: BLAKE

DATE/TIME OF TEST: 07/16/2021 02:42:34

---

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.242	02:36
Breath Volume	1.531	02:36
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.225	02:40
Breath Volume	1.468	02:40
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

---

Data Download Date/Time: 08/09/2021 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,JEFFREY,A  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: BLACKBURN,TAYLOR  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/16/2021 22:42:33

---

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.081	22:36
Air Blank	0.000	22:37
Subject Sample # 1	0.170	22:38
Breath Volume	1.894	22:38
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.173	22:41
Breath Volume	2.207	22:41
Air Blank	0.000	22:41
Control Test	0.081	22:41
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

---

Data Download Date/Time: 08/09/2021 11:35



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRIEN,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: BRIEN,JOSHUA  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: HYMAN

DATE/TIME OF TEST: 07/16/2021 23:46:43

---

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.080	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.269	23:41
Breath Volume	1.101	23:41
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.262	23:45
Breath Volume	1.203	23:45
Air Blank	0.000	23:45
Control Test	0.081	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

---

Data Download Date/Time: 08/09/2021 11:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: NORTHWAY,GEORGIA  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 07/17/2021 01:10:18

---

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:05	
Control Test	0.080	01:06	
Air Blank	0.000	01:06	
Subject Sample	# 1 0.177	01:09	
Breath Volume	1.269	01:09	
Air Blank		01:09	Purge Fail
Air Blank	0.000	01:10	

---

Data Download Date/Time: 08/09/2021 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: NORTHWAY,GEORGIA  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 07/17/2021 01:18:12

---

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.080	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.189	01:13
Breath Volume		2.109	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.195	01:16
Breath Volume		1.558	01:16
Air Blank		0.000	01:17
Control Test		0.081	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

---

Data Download Date/Time: 08/09/2021 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRAXTON,D,L  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: BRAXTON,D  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 07/17/2021 04:34:52

---

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.080	04:29
Air Blank	0.000	04:29
Subject Sample # 1	0.153	04:30
Breath Volume	2.074	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.138	04:33
Breath Volume	1.148	04:33
Air Blank	0.000	04:33
Control Test	0.079	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

---

Data Download Date/Time: 08/09/2021 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: NORTHWAY,GEORGIA  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:29

SUBJECT'S LAST NAME: GOMEZ-GIRON

DATE/TIME OF TEST: 07/23/2021 06:59:17

---

Diagnostics Check		OK	06:52
Air Blank		0.000	06:53
Control Test		0.081	06:53
Air Blank		0.000	06:53
Subject Sample	# 1	0.158	06:54
Breath Volume		1.105	06:54
Air Blank		0.000	06:55
Air Blank		0.000	06:56
Subject Sample	# 2	0.160	06:57
Breath Volume		1.332	06:57
Air Blank		0.000	06:58
Control Test		0.083	06:58
Air Blank		0.000	06:59
Diagnostics Check		OK	06:59

---

Data Download Date/Time: 08/09/2021 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROY,GARY,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: HINDS,LOGAN  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: HATCHER

DATE/TIME OF TEST: 07/24/2021 00:22:22

---

Diagnostics Check		OK	00:15
Air Blank		0.000	00:16
Control Test		0.080	00:16
Air Blank		0.000	00:16
Subject Sample	# 1	0.146	00:17
Breath Volume		2.000	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.144	00:20
Breath Volume		1.292	00:20
Air Blank		0.000	00:21
Control Test		0.081	00:21
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

---

Data Download Date/Time: 08/09/2021 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: VOLLERTSEN,CRAIG  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: WOODRUFF

DATE/TIME OF TEST: 07/24/2021 02:00:22

---

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.122	01:55
Breath Volume	2.039	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.118	01:58
Breath Volume	2.066	01:58
Air Blank	0.000	01:59
Control Test	0.082	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

---

Data Download Date/Time: 08/09/2021 11:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REAGLE,ASHLEY,R  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: REAGLE,ASHLEY  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 07/25/2021 04:20:38

---

Diagnostics Check	OK	04:14
Air Blank	0.000	04:14
Control Test	0.081	04:15
Air Blank	0.000	04:15
Subject Sample # 1	0.161	04:16
Breath Volume	3.757	04:16
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.142	04:19
Breath Volume	3.863	04:19
Air Blank	0.000	04:19
Control Test	0.081	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

---

Data Download Date/Time: 08/09/2021 11:35



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MATLOCK,L,R  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: MATLOCK,L  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: NETTLES

DATE/TIME OF TEST: 07/25/2021 23:36:34

---

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.082	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.164	23:32
Breath Volume	1.500	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.178	23:35
Breath Volume	2.167	23:35
Air Blank	0.000	23:35
Control Test	0.082	23:35
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

---

Data Download Date/Time: 08/09/2021 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FINUFF,DANIEL,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: VOLLERTSEN,CRAIG  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:43

SUBJECT'S LAST NAME: LUNA ARROYO

DATE/TIME OF TEST: 07/26/2021 17:19:24

---

Diagnostics Check	OK	17:13
Air Blank	0.000	17:13
Control Test	0.081	17:14
Air Blank	0.000	17:14
Subject Sample # 1	0.000	17:15
Breath Volume	2.699	17:15
Air Blank	0.000	17:15
Air Blank	0.000	17:17
Subject Sample # 2	0.000	17:18
Breath Volume	2.796	17:18
Air Blank	0.000	17:18
Control Test	0.081	17:18
Air Blank	0.000	17:19
Diagnostics Check	OK	17:19

---

Data Download Date/Time: 08/09/2021 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAXEY,JORDAN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MAXEY,JORDAN  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 13:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: SCHROEDER

DATE/TIME OF TEST: 07/29/2021 16:47:00

---

Diagnostics Check	OK	16:40
Air Blank	0.000	16:41
Control Test	0.080	16:41
Air Blank	0.000	16:42
Subject Sample # 1	0.399	16:42
Breath Volume	1.523	16:42
Air Blank	0.000	16:43
Air Blank	0.000	16:44
Subject Sample # 2	0.397	16:45
Breath Volume	3.671	16:45
Air Blank	0.000	16:46
Control Test	0.081	16:46
Air Blank	0.000	16:46
Diagnostics Check	OK	16:47

---

Data Download Date/Time: 08/09/2021 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRANKLIN,JOSEPH,B  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: FRANKLIN,JOSEPH  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: SAMPSON

DATE/TIME OF TEST: 07/01/2021 03:02:39

---

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.082	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.000	02:58
Breath Volume	3.390	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.000	03:01
Breath Volume	3.363	03:01
Air Blank	0.000	03:01
Control Test	0.082	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

---

Data Download Date/Time: 07/12/2021 12:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FORAN,ETHAN,L  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: FORAN,ETHAN  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: SURA SARAVIA

DATE/TIME OF TEST: 07/03/2021 00:43:48

---

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.082	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.240	00:39
Breath Volume	1.589	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.245	00:42
Breath Volume	1.789	00:42
Air Blank	0.000	00:42
Control Test	0.082	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

---

Data Download Date/Time: 07/12/2021 12:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STROHECKER,KYNDAL,C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: STROHECKER,KYNDAL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 07/03/2021 21:41:42

---

Diagnostics Check	OK	21:35
Air Blank	0.000	21:36
Control Test	0.083	21:36
Air Blank	0.000	21:36
Subject Sample # 1	0.274	21:37
Breath Volume	1.425	21:37
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.272	21:40
Breath Volume	1.269	21:40
Air Blank	0.000	21:40
Control Test	0.081	21:41
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

---

Data Download Date/Time: 07/12/2021 12:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MOSCHETTO, ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: PATZEN AGIN

DATE/TIME OF TEST: 07/04/2021 21:34:59

---

Diagnostics Check	OK	21:27
Air Blank	0.000	21:28
Control Test	0.081	21:28
Air Blank	0.000	21:29
Subject Sample # 1	0.298	21:30
Breath Volume	1.886	21:30
Air Blank	0.000	21:31
Air Blank	0.000	21:32
Subject Sample # 2	0.279	21:33
Breath Volume	1.402	21:33
Air Blank	0.000	21:34
Control Test	0.081	21:34
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

---

Data Download Date/Time: 07/12/2021 12:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MOSCHETTO, ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04120080A1  
DRY GAS EXPIRY DATE: 03/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 07/10/2021 22:48:25

---

Diagnostics Check	OK	22:42
Air Blank	0.000	22:42
Control Test	0.083	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.082	22:44
Breath Volume	2.105	22:44
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.080	22:46
Breath Volume	1.917	22:46
Air Blank	0.000	22:47
Control Test	0.082	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

---

Data Download Date/Time: 07/12/2021 12:43



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STROHECKER,KYNDAL,C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: STROHECKER,KYNDAL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34520080A1  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: PARKS

DATE/TIME OF TEST: 07/02/2021 23:01:52

---

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.080	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.143	22:57
Breath Volume	2.671	22:57
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.145	23:00
Breath Volume	2.707	23:00
Air Blank	0.000	23:00
Control Test	0.081	23:01
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

---

Data Download Date/Time: 07/28/2021 12:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: VOLLERTSEN,CRAIG  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34520080A1  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: KEITH

DATE/TIME OF TEST: 07/14/2021 20:24:10

---

Diagnostics Check	OK	20:18
Air Blank	0.000	20:18
Control Test	0.079	20:18
Air Blank	0.000	20:19
Subject Sample # 1	0.136	20:19
Breath Volume	4.132	20:19
Air Blank	0.000	20:20
Air Blank	0.000	20:22
Subject Sample # 2	0.137	20:22
Breath Volume	3.074	20:22
Air Blank	0.000	20:23
Control Test	0.080	20:23
Air Blank	0.000	20:24
Diagnostics Check	OK	20:24

---

Data Download Date/Time: 07/28/2021 12:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: VOLLERTSEN,CRAIG  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34520080A1  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: VAUSE

DATE/TIME OF TEST: 07/25/2021 02:20:12

---

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:11	
Control Test	0.079	02:12	
Air Blank	0.000	02:12	
Subject Sample	# 1 0.095	02:15	
Breath Volume	1.257	02:15	
Air Blank	0.000	02:16	
Air Blank	0.000	02:18	
Subject Sample	# 2 0.099	02:18	
Breath Volume	2.500	02:18	
Air Blank	0.000	02:19	
Control Test	0.079	02:19	
Air Blank		02:20	RFI Detect
Air Blank		02:20	RFI Detect

---

Data Download Date/Time: 07/28/2021 12:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: VOLLERTSEN,CRAIG  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34520080A1  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: VAUSE

DATE/TIME OF TEST: 07/25/2021 02:28:37

---

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.077	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.097	02:24
Breath Volume	1.882	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.097	02:27
Breath Volume	1.878	02:27
Air Blank	0.000	02:27
Control Test	0.079	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

---

Data Download Date/Time: 07/28/2021 12:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MOORE,ANDREW  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 07/03/2021 02:12:02

---

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.079	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.007	02:07
Breath Volume	3.378	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.000	02:10
Breath Volume	3.171	02:10
Air Blank	0.000	02:11
Control Test	0.079	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:12

---

Data Download Date/Time: 07/06/2021 14:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HAYLES,SHANE  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: GALINDO

DATE/TIME OF TEST: 07/03/2021 02:42:27

---

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.169	02:37
Breath Volume	1.609	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.173	02:40
Breath Volume	2.808	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

---

Data Download Date/Time: 07/06/2021 14:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: LIEBKNECHT,KEVIN  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:03

SUBJECT'S LAST NAME: MADRIGAL COUTINO

DATE/TIME OF TEST: 07/03/2021 06:33:07

---

Diagnostics Check	OK	06:26
Air Blank	0.000	06:27
Control Test	0.079	06:27
Air Blank	0.000	06:28
Subject Sample # 1	0.179	06:28
Breath Volume	1.914	06:28
Air Blank	0.000	06:29
Air Blank	0.000	06:31
Subject Sample # 2	0.178	06:31
Breath Volume	2.171	06:31
Air Blank	0.000	06:32
Control Test	0.080	06:32
Air Blank	0.000	06:33
Diagnostics Check	OK	06:33

---

Data Download Date/Time: 07/06/2021 14:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILKINS,CONNOR,R  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DEL CASTILLO,BRET  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: VALENCIA

DATE/TIME OF TEST: 07/04/2021 03:32:43

---

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.080	03:22	
Air Blank	0.000	03:22	
Subject Sample	# 1	03:25	No Sample Provided
Breath Volume	0.000	03:25	
Air Blank	0.000	03:26	
Air Blank	0.000	03:28	
Subject Sample	# 2	03:31	No Sample Provided
Breath Volume	0.000	03:31	
Air Blank	0.000	03:31	
Control Test	0.080	03:32	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:32	

---

Data Download Date/Time: 07/06/2021 14:09



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHEWAN,JANICE,  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HALL,ANDREW  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: CONTRERAS

DATE/TIME OF TEST: 07/05/2021 00:29:50

---

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:27	
Control Test	0.079	00:27	
Air Blank	0.000	00:28	
Subject Sample	# 1	00:28	Subject Test Refused
Breath Volume	0.000	00:28	
Air Blank	0.000	00:28	
Control Test	0.079	00:29	
Air Blank	0.000	00:29	
Diagnostics Check	OK	00:29	

---

Data Download Date/Time: 07/06/2021 14:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHEWAN,JANICE,  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HALL,JEFFREY  
ARRESTING OFFICER AGENCY: Ocoee PD  
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: BISSESSAR

DATE/TIME OF TEST: 07/05/2021 02:30:33

---

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:15	
Control Test		0.080	02:16	
Air Blank		0.000	02:16	
Subject Sample	# 1	0.237	02:17	
Breath Volume		1.261	02:17	
Air Blank		0.000	02:18	
Air Blank		0.000	02:20	
Subject Sample	# 2	0.212	02:23	Volume Not Met
Breath Volume		0.070	02:23	
Air Blank		0.000	02:24	
Air Blank		0.000	02:25	
Subject Sample	# 3	0.218	02:28	Volume Not Met
Breath Volume		0.082	02:28	
Air Blank		0.000	02:29	
Control Test		0.076	02:29	
Air Blank		0.000	02:30	
Diagnostics Check		OK	02:30	

---

Data Download Date/Time: 07/06/2021 14:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHEWAN,JANICE,  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: TORRES,JOSUE  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: CORRAL

DATE/TIME OF TEST: 07/05/2021 04:14:53

---

Diagnostics Check	OK	04:08
Air Blank	0.000	04:09
Control Test	0.079	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.217	04:10
Breath Volume	2.156	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.225	04:13
Breath Volume	2.250	04:13
Air Blank	0.000	04:13
Control Test	0.078	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

---

Data Download Date/Time: 07/06/2021 14:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRISWOLD,DAVID,J  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: BARBERI,FRANCIS  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 17:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: GEORGE

DATE/TIME OF TEST: 07/03/2021 18:52:00

---

Diagnostics Check	OK	18:45
Air Blank	0.000	18:45
Control Test	0.081	18:45
Air Blank	0.000	18:46
Subject Sample # 1	0.037	18:47
Breath Volume	3.183	18:47
Air Blank	0.000	18:48
Air Blank	0.000	18:49
Subject Sample # 2	0.038	18:50
Breath Volume	3.027	18:50
Air Blank	0.000	18:51
Control Test	0.081	18:51
Air Blank	0.000	18:51
Diagnostics Check	OK	18:52

---

Data Download Date/Time: 07/30/2021 11:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: HANSELL,ROBERT  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: CABRERA-PERALTA

DATE/TIME OF TEST: 07/05/2021 02:24:11

---

Diagnostics Check		OK	02:16
Air Blank		0.000	02:17
Control Test		0.080	02:17
Air Blank		0.000	02:18
Subject Sample	# 1	0.209	02:19
Breath Volume		2.277	02:19
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.200	02:22
Breath Volume		2.195	02:22
Air Blank		0.000	02:23
Control Test		0.077	02:23
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

---

Data Download Date/Time: 07/30/2021 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: RILEY,BRANDON  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 07/05/2021 02:35:05

---

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.077	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.188	02:30
Breath Volume		4.023	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:32
Subject Sample	# 2	0.187	02:33
Breath Volume		3.972	02:33
Air Blank		0.000	02:34
Control Test		0.078	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:35

---

Data Download Date/Time: 07/30/2021 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: SOTO,BRANDON  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: GONZALEZ-VELAZQUEZ

DATE/TIME OF TEST: 07/05/2021 02:58:34

---

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.080	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.021	02:54
Breath Volume	3.234	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.021	02:57
Breath Volume	3.179	02:57
Air Blank	0.000	02:57
Control Test	0.078	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

---

Data Download Date/Time: 07/30/2021 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: COWLEY,DUSTIN  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: DUBRA

DATE/TIME OF TEST: 07/05/2021 04:41:58

---

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.080	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.000	04:36
Breath Volume	4.281	04:36
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.000	04:40
Breath Volume	3.257	04:40
Air Blank	0.000	04:41
Control Test	0.078	04:41
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

---

Data Download Date/Time: 07/30/2021 11:49



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAWFORD,DAVID,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: CRAWFORD,DAVID  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 18:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 07/07/2021 18:41:39

---

Diagnostics Check	OK	18:35
Air Blank	0.000	18:36
Control Test	0.080	18:36
Air Blank	0.000	18:36
Subject Sample # 1	0.000	18:37
Breath Volume	2.488	18:37
Air Blank	0.000	18:37
Air Blank	0.000	18:39
Subject Sample # 2	0.000	18:40
Breath Volume	1.214	18:40
Air Blank	0.000	18:40
Control Test	0.078	18:41
Air Blank	0.000	18:41
Diagnostics Check	OK	18:41

---

Data Download Date/Time: 07/30/2021 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: HANSELL,ROBERT  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: COLLEY

DATE/TIME OF TEST: 07/12/2021 21:36:49

---

Diagnostics Check	OK	21:30
Air Blank	0.000	21:31
Control Test	0.081	21:31
Air Blank	0.000	21:32
Subject Sample # 1	0.000	21:32
Breath Volume	3.503	21:32
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.000	21:35
Breath Volume	3.511	21:35
Air Blank	0.000	21:35
Control Test	0.079	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

---

Data Download Date/Time: 07/30/2021 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHOBREY,SPENCER,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: TURNER,RODNEY  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: ESTEVEZ

DATE/TIME OF TEST: 07/15/2021 02:57:58

---

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.178	02:53
Breath Volume	2.539	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.176	02:56
Breath Volume	2.117	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

---

Data Download Date/Time: 07/30/2021 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: HANSELL,ROBERT  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: SUBNAIK

DATE/TIME OF TEST: 07/17/2021 01:53:59

---

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.079	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.150	01:49
Breath Volume	3.066	01:49
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.146	01:52
Breath Volume	3.007	01:52
Air Blank	0.000	01:53
Control Test	0.079	01:53
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

---

Data Download Date/Time: 07/30/2021 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHOBREY,SPENCER,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: WHOBREY,SPENCER  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1472328  
DRY GAS EXPIRY DATE: 04/02/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: ACOSTA-GUERRA

DATE/TIME OF TEST: 07/26/2021 00:06:27

---

Diagnostics Check		OK	23:59
Air Blank		0.000	23:59
Control Test		0.081	00:00
Air Blank		0.000	00:00
Subject Sample	# 1	0.173	00:01
Breath Volume		2.007	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.179	00:04
Breath Volume		2.171	00:04
Air Blank		0.000	00:05
Control Test		0.081	00:05
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

---

Data Download Date/Time: 07/30/2021 11:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: HANSELL,ROBERT  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1190213  
DRY GAS EXPIRY DATE: 08/28/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: KANE

DATE/TIME OF TEST: 07/27/2021 01:23:39

---

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:17	
Control Test	0.079	01:18	
Air Blank	0.000	01:18	
Subject Sample # 1	0.246	01:19	
Breath Volume	3.089	01:19	
Air Blank	0.000	01:19	
Air Blank	0.000	01:21	
Subject Sample # 2	0.234	01:22	
Breath Volume	1.285	01:22	
Air Blank	0.000	01:22	
Control Test	0.073	01:23	Interferent Detect
Air Blank	0.000	01:23	

---

Data Download Date/Time: 08/18/2021 14:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: INGERSOLL,NICHOLAS,F  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: BELL,ERIC  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A2  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: HALLMARK

DATE/TIME OF TEST: 07/04/2021 01:22:12

---

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.080	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.153	01:17
Breath Volume		2.097	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.143	01:20
Breath Volume		1.781	01:20
Air Blank		0.000	01:21
Control Test		0.080	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

---

Data Download Date/Time: 08/02/2021 10:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,JOSEPH,A  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: STANLEY,WILLIAM  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A2  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: WILLIAMS MURPHY

DATE/TIME OF TEST: 07/11/2021 03:18:53

---

Diagnostics Check		OK	03:07	
Air Blank		0.000	03:07	
Control Test		0.081	03:08	
Air Blank		0.000	03:08	
Subject Sample	# 1	0.195	03:11	Volume Not Met
Breath Volume		0.070	03:11	
Air Blank		0.000	03:12	
Air Blank		0.000	03:14	
Subject Sample	# 2	0.000	03:17	Volume Not Met
Breath Volume		0.046	03:17	
Air Blank		0.000	03:17	
Control Test		0.078	03:18	
Air Blank		0.000	03:18	
Diagnostics Check		OK	03:18	

---

Data Download Date/Time: 08/02/2021 10:00



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: INGERSOLL,NICHOLAS,F  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: WHITE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A2  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 07/13/2021 20:14:33

---

Diagnostics Check	OK	20:07
Air Blank	0.000	20:08
Control Test	0.081	20:08
Air Blank	0.000	20:09
Subject Sample # 1	0.185	20:10
Breath Volume	3.871	20:10
Air Blank	0.000	20:10
Air Blank	0.000	20:12
Subject Sample # 2	0.193	20:13
Breath Volume	3.457	20:13
Air Blank	0.000	20:13
Control Test	0.081	20:13
Air Blank	0.000	20:14
Diagnostics Check	OK	20:14

---

Data Download Date/Time: 08/02/2021 10:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORTIZ,DAVID,  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: ORTIZ,DAVID  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A2  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 18:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: BEAR

DATE/TIME OF TEST: 07/17/2021 18:53:09

---

Diagnostics Check	OK	18:45
Air Blank	0.000	18:46
Control Test	0.081	18:46
Air Blank	0.000	18:46
Subject Sample # 1	0.168	18:48
Breath Volume	2.074	18:48
Air Blank	0.000	18:49
Air Blank	0.000	18:51
Subject Sample # 2	0.165	18:51
Breath Volume	1.421	18:51
Air Blank	0.000	18:52
Control Test	0.079	18:52
Air Blank	0.000	18:53
Diagnostics Check	OK	18:53

---

Data Download Date/Time: 08/02/2021 10:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANTERIO,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: MALOY,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 07/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A2  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 07/24/2021 07:44:37

---

Diagnostics Check	OK	07:38
Air Blank	0.000	07:38
Control Test	0.080	07:38
Air Blank	0.000	07:39
Subject Sample # 1	0.000	07:40
Breath Volume	3.300	07:40
Air Blank	0.000	07:40
Air Blank	0.000	07:42
Subject Sample # 2	0.000	07:43
Breath Volume	3.000	07:43
Air Blank	0.000	07:43
Control Test	0.081	07:44
Air Blank	0.000	07:44
Diagnostics Check	OK	07:44

---

Data Download Date/Time: 08/02/2021 10:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCLAUGHLIN, BRADLEY,  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: MCLAUGHLIN, BRADLEY  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 07/23/2021

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A2  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: RUTH

DATE/TIME OF TEST: 07/27/2021 00:21:35

---

Diagnostics Check		OK	00:12
Air Blank		0.000	00:12
Control Test		0.081	00:13
Air Blank		0.000	00:13
Subject Sample	# 1	0.198	00:14
Breath Volume		1.531	00:14
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.176	00:17
Breath Volume		1.511	00:17
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 3	0.178	00:20
Breath Volume		1.355	00:20
Air Blank		0.000	00:20
Control Test		0.079	00:21
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

---

Data Download Date/Time: 08/02/2021 10:00