

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: GASKILL

DATE/TIME OF TEST: 09/03/2021 01:52:57

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.079	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.208	01:48
Breath Volume	1.667	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.220	01:51
Breath Volume	2.769	01:51
Air Blank	0.000	01:52
Control Test	0.078	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 09/30/2021 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: GAST

DATE/TIME OF TEST: 09/08/2021 00:32:10

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:18	
Control Test	0.079	00:18	
Air Blank	0.000	00:19	
Subject Sample	# 1	00:22	No Sample Provided
Breath Volume	0.000	00:22	
Air Blank	0.000	00:22	
Air Blank	0.000	00:24	
Subject Sample	# 2	0.208	00:27
Breath Volume	1.261	00:27	
Air Blank	0.000	00:28	
Air Blank	0.000	00:29	
Subject Sample	# 3	0.216	00:30
Breath Volume	1.164	00:30	
Air Blank	0.000	00:31	
Control Test	0.077	00:31	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 09/30/2021 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/08/2021 00:51:51

Diagnostics Check	OK	00:40
Air Blank	0.000	00:40
Control Test	0.077	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.277	00:43
Breath Volume	2.234	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:45
Subject Sample # 2	0.241	00:46
Breath Volume	1.164	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 3	0.253	00:50
Breath Volume	1.824	00:50
Air Blank	0.000	00:50
Control Test	0.078	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 09/30/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 09/09/2021 21:29:56

Diagnostics Check	OK	21:23
Air Blank	0.000	21:24
Control Test	0.078	21:24
Air Blank	0.000	21:25
Subject Sample # 1	0.236	21:25
Breath Volume	2.027	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:27
Subject Sample # 2	0.238	21:28
Breath Volume	2.339	21:28
Air Blank	0.000	21:29
Control Test	0.079	21:29
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 09/30/2021 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: KAMMER,ROBERT
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: ORTA

DATE/TIME OF TEST: 09/10/2021 20:35:17

Diagnostics Check	OK	20:27
Air Blank	0.000	20:28
Control Test	0.080	20:28
Air Blank	0.000	20:29
Subject Sample # 1	0.015	20:29
Breath Volume	1.699	20:29
Air Blank	0.000	20:30
Air Blank	0.000	20:32
Subject Sample # 2	0.014	20:33
Breath Volume	1.113	20:33
Air Blank	0.000	20:34
Control Test	0.079	20:34
Air Blank	0.000	20:35
Diagnostics Check	OK	20:35

Data Download Date/Time: 09/30/2021 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/12/2021 03:10:03

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:54	
Control Test	0.079	02:54	
Air Blank	0.000	02:55	
Subject Sample	# 1 0.185	02:58	Volume Not Met
Breath Volume	0.203	02:58	
Air Blank	0.000	02:59	
Air Blank	0.000	03:00	
Subject Sample	# 2 0.173	03:03	
Breath Volume	1.476	03:03	
Air Blank	0.000	03:04	
Air Blank	0.000	03:06	
Subject Sample	# 3 0.162	03:08	
Breath Volume	2.144	03:08	
Air Blank	0.000	03:09	
Control Test	0.076	03:09	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 09/30/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: JAROUSSE,DANIELLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: GIBBY

DATE/TIME OF TEST: 09/12/2021 03:55:42

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.079	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.128	03:51
Breath Volume	2.257	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.130	03:54
Breath Volume	1.441	03:54
Air Blank	0.000	03:54
Control Test	0.079	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 09/30/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: TOLER

DATE/TIME OF TEST: 09/13/2021 01:34:55

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.079	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.021	01:29
Breath Volume	2.019	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.016	01:33
Breath Volume	1.472	01:33
Air Blank	0.000	01:33
Control Test	0.077	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 09/30/2021 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADO,RYAN,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: RADO,RYAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BEDOYA VILLEGAS

DATE/TIME OF TEST: 09/13/2021 03:07:44

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:55	
Control Test	0.079	02:56	
Air Blank	0.000	02:56	
Subject Sample	# 1 0.166	02:59	Volume Not Met
Breath Volume	0.156	02:59	
Air Blank	0.000	03:00	
Air Blank	0.000	03:02	
Subject Sample	# 2 0.189	03:02	
Breath Volume	2.250	03:02	
Air Blank	0.000	03:03	
Air Blank	0.000	03:05	
Subject Sample	# 3 0.194	03:06	
Breath Volume	2.042	03:06	
Air Blank	0.000	03:06	
Control Test	0.078	03:07	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

Data Download Date/Time: 09/30/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMDEN,DENNIS,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CAMDEN,DENNIS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 06:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:37

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/19/2021 07:14:03

Diagnostics Check	OK	07:07
Air Blank	0.000	07:08
Control Test	0.079	07:08
Air Blank	0.000	07:08
Subject Sample # 1	0.000	07:09
Breath Volume	1.417	07:09
Air Blank	0.000	07:09
Air Blank	0.000	07:11
Subject Sample # 2	0.000	07:12
Breath Volume	2.015	07:12
Air Blank	0.000	07:13
Control Test	0.079	07:13
Air Blank	0.000	07:13
Diagnostics Check	OK	07:14

Data Download Date/Time: 09/30/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORALUZZO,MICHAEL,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: COALUZZO,MICHAEL
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: STRATTON

DATE/TIME OF TEST: 09/19/2021 11:00:18

Diagnostics Check	OK	10:53
Air Blank	0.000	10:54
Control Test	0.080	10:54
Air Blank	0.000	10:54
Subject Sample # 1	0.122	10:55
Breath Volume	1.941	10:55
Air Blank	0.000	10:56
Air Blank	0.000	10:58
Subject Sample # 2	0.122	10:58
Breath Volume	1.238	10:58
Air Blank	0.000	10:59
Control Test	0.080	10:59
Air Blank	0.000	11:00
Diagnostics Check	OK	11:00

Data Download Date/Time: 09/30/2021 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUSCIO,JESSE,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TRUSCIO,JESSE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 16:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:11

SUBJECT'S LAST NAME: KUCHTA

DATE/TIME OF TEST: 09/19/2021 17:40:55

Diagnostics Check	OK	17:34
Air Blank	0.000	17:34
Control Test	0.080	17:34
Air Blank	0.000	17:35
Subject Sample # 1	0.000	17:36
Breath Volume	1.324	17:36
Air Blank	0.000	17:36
Air Blank	0.000	17:38
Subject Sample # 2	0.000	17:39
Breath Volume	1.375	17:39
Air Blank	0.000	17:40
Control Test	0.078	17:40
Air Blank	0.000	17:40
Diagnostics Check	OK	17:40

Data Download Date/Time: 09/30/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: DE LA FLOR

DATE/TIME OF TEST: 09/22/2021 23:45:29

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.079	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.103	23:40
Breath Volume	2.585	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.103	23:43
Breath Volume	2.531	23:43
Air Blank	0.000	23:44
Control Test	0.078	23:44
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 09/30/2021 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNING,SEAN,D
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 09:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:46

SUBJECT'S LAST NAME: KELTNER

DATE/TIME OF TEST: 09/23/2021 11:09:09

Diagnostics Check	OK	11:06	
Air Blank	0.000	11:06	
Control Test	0.079	11:07	
Air Blank	0.000	11:07	
Subject Sample	# 1	11:07	Subject Test Refused
Breath Volume	0.000	11:07	
Air Blank	0.000	11:08	
Control Test	0.079	11:08	
Air Blank	0.000	11:09	
Diagnostics Check	OK	11:09	

Data Download Date/Time: 09/30/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: KALACZYNSKI

DATE/TIME OF TEST: 09/23/2021 21:46:17

Diagnostics Check	OK	21:40
Air Blank	0.000	21:40
Control Test	0.079	21:40
Air Blank	0.000	21:41
Subject Sample # 1	0.000	21:41
Breath Volume	2.886	21:41
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.000	21:44
Breath Volume	3.304	21:44
Air Blank	0.000	21:45
Control Test	0.078	21:45
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 09/30/2021 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 08:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:28

SUBJECT'S LAST NAME: RUGGERI

DATE/TIME OF TEST: 09/26/2021 08:56:23

Diagnostics Check	OK	08:49
Air Blank	0.000	08:49
Control Test	0.080	08:50
Air Blank	0.000	08:50
Subject Sample # 1	0.116	08:51
Breath Volume	3.128	08:51
Air Blank	0.000	08:52
Air Blank	0.000	08:54
Subject Sample # 2	0.109	08:54
Breath Volume	1.164	08:54
Air Blank	0.000	08:55
Control Test	0.078	08:55
Air Blank	0.000	08:56
Diagnostics Check	OK	08:56

Data Download Date/Time: 09/30/2021 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: AMBER

DATE/TIME OF TEST: 09/03/2021 22:01:31

Diagnostics Check	OK	21:54
Air Blank	0.000	21:55
Control Test	0.079	21:55
Air Blank	0.000	21:55
Subject Sample # 1	0.180	21:57
Breath Volume	1.992	21:57
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.165	21:59
Breath Volume	1.207	21:59
Air Blank	0.000	22:00
Control Test	0.077	22:00
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 09/30/2021 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 09/04/2021 02:12:41

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.080	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.125	02:08
Breath Volume	2.652	02:08
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.127	02:11
Breath Volume	1.972	02:11
Air Blank	0.000	02:11
Control Test	0.078	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 09/30/2021 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: RACZAK

DATE/TIME OF TEST: 09/09/2021 00:05:35

Diagnostics Check	OK	23:59
Air Blank	0.000	23:59
Control Test	0.079	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.168	00:01
Breath Volume	3.769	00:01
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.178	00:04
Breath Volume	3.171	00:04
Air Blank	0.000	00:04
Control Test	0.080	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 09/30/2021 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: GERBER

DATE/TIME OF TEST: 09/09/2021 02:42:59

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.081	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.124	02:38
Breath Volume	3.250	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.120	02:41
Breath Volume	2.355	02:41
Air Blank	0.000	02:42
Control Test	0.077	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 09/30/2021 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 09/10/2021 01:09:36

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.080	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.000	01:05
Breath Volume	2.480	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.000	01:08
Breath Volume	2.273	01:08
Air Blank	0.000	01:08
Control Test	0.079	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 09/30/2021 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,JASON,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: BAINES

DATE/TIME OF TEST: 09/10/2021 20:34:03

Diagnostics Check	OK	20:27
Air Blank	0.000	20:28
Control Test	0.080	20:28
Air Blank	0.000	20:28
Subject Sample # 1	0.273	20:29
Breath Volume	2.695	20:29
Air Blank	0.000	20:30
Air Blank	0.000	20:32
Subject Sample # 2	0.261	20:32
Breath Volume	2.132	20:32
Air Blank	0.000	20:33
Control Test	0.079	20:33
Air Blank	0.000	20:33
Diagnostics Check	OK	20:34

Data Download Date/Time: 09/30/2021 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEX
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 09/11/2021 02:24:42

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.081	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.116	02:20
Breath Volume	2.417	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.111	02:23
Breath Volume	2.195	02:23
Air Blank	0.000	02:23
Control Test	0.079	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 09/30/2021 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: ESTBERG

DATE/TIME OF TEST: 09/11/2021 21:18:08

Diagnostics Check	OK	21:11
Air Blank	0.000	21:12
Control Test	0.078	21:12
Air Blank	0.000	21:13
Subject Sample # 1	0.071	21:13
Breath Volume	2.589	21:13
Air Blank	0.000	21:14
Air Blank	0.000	21:16
Subject Sample # 2	0.069	21:16
Breath Volume	4.007	21:16
Air Blank	0.000	21:17
Control Test	0.080	21:17
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

Data Download Date/Time: 09/30/2021 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMA,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LIMA,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: MATTHIAS

DATE/TIME OF TEST: 09/12/2021 00:49:58

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:48	
Control Test	0.080	00:48	
Air Blank		00:49	Purge Fail
Air Blank	0.000	00:49	

Data Download Date/Time: 09/30/2021 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMA,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LIMA,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: MATTHIAS

DATE/TIME OF TEST: 09/12/2021 00:59:21

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.080	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.169	00:54
Breath Volume	2.574	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.169	00:57
Breath Volume	2.859	00:57
Air Blank	0.000	00:58
Control Test	0.080	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 09/30/2021 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEX,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 09/12/2021 02:03:44

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.131	01:59
Breath Volume	1.667	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.133	02:02
Breath Volume	1.742	02:02
Air Blank	0.000	02:02
Control Test	0.079	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 09/30/2021 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 09/15/2021 23:14:01

Diagnostics Check	OK	23:09	
Air Blank	0.000	23:09	
Control Test	0.080	23:09	
Air Blank	0.000	23:10	
Subject Sample	# 1	23:12	Subject Test Refused
Breath Volume	0.000	23:12	
Air Blank	0.000	23:13	
Control Test	0.081	23:13	
Air Blank	0.000	23:13	
Diagnostics Check	OK	23:14	

Data Download Date/Time: 09/30/2021 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: CORTESE

DATE/TIME OF TEST: 09/16/2021 03:05:27

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.017	03:01
Breath Volume	2.480	03:01
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.015	03:03
Breath Volume	2.769	03:03
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 09/30/2021 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: FULGHAM

DATE/TIME OF TEST: 09/17/2021 21:41:53

Diagnostics Check	OK	21:35
Air Blank	0.000	21:36
Control Test	0.078	21:36
Air Blank	0.000	21:36
Subject Sample # 1	0.190	21:37
Breath Volume	2.074	21:37
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.196	21:40
Breath Volume	1.906	21:40
Air Blank	0.000	21:40
Control Test	0.080	21:41
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 09/30/2021 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: DOAK

DATE/TIME OF TEST: 09/18/2021 00:02:22

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.079	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.184	23:57
Breath Volume	2.898	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.184	00:00
Breath Volume	2.679	00:00
Air Blank	0.000	00:01
Control Test	0.079	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 09/30/2021 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: DOURADO

DATE/TIME OF TEST: 09/23/2021 00:37:59

Diagnostics Check	OK	00:31
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:33
Subject Sample # 1	0.102	00:33
Breath Volume	2.859	00:33
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.104	00:36
Breath Volume	3.218	00:36
Air Blank	0.000	00:37
Control Test	0.081	00:37
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 09/30/2021 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BEHLING

DATE/TIME OF TEST: 09/24/2021 02:43:18

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:32	
Control Test	0.080	02:32	
Air Blank	0.000	02:33	
Subject Sample	# 1 0.152	02:36	Volume Not Met
Breath Volume	0.816	02:36	
Air Blank	0.000	02:36	
Air Blank	0.000	02:38	
Subject Sample	# 2	02:41	No Sample Provided
Breath Volume	0.000	02:41	
Air Blank	0.000	02:42	
Control Test	0.081	02:42	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 09/30/2021 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO
ARRESTING OFFICER: TINFINA,LOGAN
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 08/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: POPE

DATE/TIME OF TEST: 09/10/2021 23:25:56

Diagnostics Check	OK	23:18
Air Blank	0.000	23:18
Control Test	0.080	23:19
Air Blank	0.000	23:19
Subject Sample # 1	0.282	23:20
Breath Volume	2.792	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:22
Subject Sample # 2	0.279	23:24
Breath Volume	1.175	23:24
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 09/13/2021 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO
LAST AGENCY INSPECTION: 08/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MEJIA-LOPEZ

DATE/TIME OF TEST: 09/13/2021 01:03:59

Diagnostics Check	OK	00:56
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.145	00:59
Breath Volume	3.433	00:59
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.148	01:02
Breath Volume	3.230	01:02
Air Blank	0.000	01:03
Control Test	0.079	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 09/13/2021 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO
LAST AGENCY INSPECTION: 09/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: CIEMNIECKI

DATE/TIME OF TEST: 09/21/2021 01:04:17

Diagnostics Check	OK	01:03	
Air Blank		01:04	RFI Detect
Air Blank		01:04	RFI Detect

Data Download Date/Time: 10/18/2021 14:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO
LAST AGENCY INSPECTION: 09/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: CIEMNIECKI

DATE/TIME OF TEST: 09/21/2021 01:06:51

Diagnostics Check	OK	01:06	
Air Blank		01:06	RFI Detect
Air Blank		01:06	RFI Detect

Data Download Date/Time: 10/18/2021 14:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO
LAST AGENCY INSPECTION: 09/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: RAMIREZ MEJIA

DATE/TIME OF TEST: 09/24/2021 23:23:32

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.080	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.145	23:18
Breath Volume	2.308	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.150	23:21
Breath Volume	3.050	23:21
Air Blank	0.000	23:22
Control Test	0.079	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 10/18/2021 14:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO
LAST AGENCY INSPECTION: 09/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:17

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 09/26/2021 18:42:52

Diagnostics Check	OK	18:41	
Air Blank	0.000	18:41	
Control Test	0.080	18:42	
Air Blank		18:42	RFI Detect
Air Blank	0.000	18:42	

Data Download Date/Time: 10/18/2021 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO
LAST AGENCY INSPECTION: 09/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:17

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 09/26/2021 18:51:10

Diagnostics Check	OK	18:44
Air Blank	0.000	18:44
Control Test	0.079	18:45
Air Blank	0.000	18:45
Subject Sample # 1	0.196	18:46
Breath Volume	1.593	18:46
Air Blank	0.000	18:47
Air Blank	0.000	18:48
Subject Sample # 2	0.180	18:49
Breath Volume	1.554	18:49
Air Blank	0.000	18:50
Control Test	0.079	18:50
Air Blank	0.000	18:51
Diagnostics Check	OK	18:51

Data Download Date/Time: 10/18/2021 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELMS,WILLIAM,A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: GHICA,VLAD
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 09/16/2021 04:00:56

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.079	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.108	03:55
Breath Volume	1.160	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.111	03:59
Breath Volume	2.191	03:59
Air Blank	0.000	03:59
Control Test	0.080	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 09/22/2021 08:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELMS,WILLIAM,A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: GHICA,VLAD
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 09/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: LANOUE

DATE/TIME OF TEST: 09/26/2021 02:57:37

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.079	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.199	02:53
Breath Volume	2.000	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.200	02:56
Breath Volume	2.277	02:56
Air Blank	0.000	02:56
Control Test	0.078	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 10/22/2021 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERESSE, DAVID, A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: PEARSON, DAMIEN
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:57

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 09/19/2021 14:31:35

Diagnostics Check	Fail	14:31	Temperature Regulation Test
Air Blank	0.000	14:31	

Data Download Date/Time: 10/07/2021 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WONSCH,J,W
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HINDS,L
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: MAMAN

DATE/TIME OF TEST: 09/20/2021 03:10:40

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.081	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.162	03:06
Breath Volume	1.835	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.165	03:09
Breath Volume	1.644	03:09
Air Blank	0.000	03:09
Control Test	0.082	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 10/07/2021 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,JAMES,D
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WEST,JAMES
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: URAIN

DATE/TIME OF TEST: 09/25/2021 02:58:17

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.082	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.100	02:53
Breath Volume	2.460	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.094	02:56
Breath Volume	1.921	02:56
Air Blank	0.000	02:57
Control Test	0.081	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 10/07/2021 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,JAMES,D
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WEST,JAMES
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: URAIN

DATE/TIME OF TEST: 09/25/2021 03:07:50

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.082	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.086	03:03
Breath Volume	1.746	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:05
Subject Sample # 2	0.088	03:06
Breath Volume	2.023	03:06
Air Blank	0.000	03:06
Control Test	0.081	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 10/07/2021 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISCHER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: FISCHER,ANDREW
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 09/25/2021 18:11:11

Diagnostics Check	OK	18:04
Air Blank	0.000	18:04
Control Test	0.082	18:05
Air Blank	0.000	18:05
Subject Sample # 1	0.223	18:06
Breath Volume	2.035	18:06
Air Blank	0.000	18:07
Air Blank	0.000	18:09
Subject Sample # 2	0.220	18:09
Breath Volume	1.218	18:09
Air Blank	0.000	18:10
Control Test	0.080	18:10
Air Blank	0.000	18:11
Diagnostics Check	OK	18:11

Data Download Date/Time: 10/07/2021 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: CANO

DATE/TIME OF TEST: 09/25/2021 19:26:50

Diagnostics Check	OK	19:20
Air Blank	0.000	19:21
Control Test	0.081	19:21
Air Blank	0.000	19:22
Subject Sample # 1	0.236	19:22
Breath Volume	1.535	19:22
Air Blank	0.000	19:23
Air Blank	0.000	19:24
Subject Sample # 2	0.227	19:25
Breath Volume	2.464	19:25
Air Blank	0.000	19:25
Control Test	0.081	19:26
Air Blank	0.000	19:26
Diagnostics Check	OK	19:26

Data Download Date/Time: 10/07/2021 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGLE,SHAWN,T
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: OGLE,SHAWN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: SAINZ

DATE/TIME OF TEST: 09/06/2021 06:28:44

Diagnostics Check	OK	06:20
Air Blank	0.000	06:21
Control Test	0.079	06:21
Air Blank	0.000	06:22
Subject Sample # 1	0.117	06:23
Breath Volume	1.851	06:23
Air Blank	0.000	06:23
Air Blank	0.000	06:25
Subject Sample # 2	0.116	06:27
Breath Volume	2.144	06:27
Air Blank	0.000	06:27
Control Test	0.080	06:28
Air Blank	0.000	06:28
Diagnostics Check	OK	06:28

Data Download Date/Time: 09/29/2021 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,RICHARD,L
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: BRAUN,CALEB
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: MCCRANEY

DATE/TIME OF TEST: 09/13/2021 23:01:18

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:58	
Control Test	0.080	22:59	
Air Blank	0.000	22:59	
Subject Sample	# 1	22:59	Subject Test Refused
Breath Volume	0.000	22:59	
Air Blank	0.000	23:00	
Control Test	0.081	23:00	
Air Blank	0.000	23:01	
Diagnostics Check	OK	23:01	

Data Download Date/Time: 09/29/2021 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,RICHARD,L
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: HOWARD,BRANDON
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: JACKSON IV

DATE/TIME OF TEST: 09/18/2021 01:46:46

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:34	
Control Test	0.080	01:35	
Air Blank	0.000	01:35	
Subject Sample	# 1	01:38	No Sample Provided
Breath Volume	0.000	01:38	
Air Blank	0.000	01:39	
Air Blank	0.000	01:41	
Subject Sample	# 2	0.245	01:41
Breath Volume	1.242	01:41	
Air Blank	0.000	01:42	
Air Blank	0.000	01:44	
Subject Sample	# 3	01:45	Subject Test Refused
Breath Volume	0.000	01:45	
Air Blank	0.000	01:45	
Control Test	0.080	01:46	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 09/29/2021 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,RICHARD,L
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: GRIMES,COURTNEY
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: DILLON

DATE/TIME OF TEST: 09/24/2021 19:11:45

Diagnostics Check	OK	19:07	
Air Blank	0.000	19:08	
Control Test	0.079	19:08	
Air Blank	0.000	19:08	
Subject Sample	# 1	19:10	Subject Test Refused
Breath Volume	0.000	19:10	
Air Blank	0.000	19:10	
Control Test	0.080	19:11	
Air Blank	0.000	19:11	
Diagnostics Check	OK	19:11	

Data Download Date/Time: 09/29/2021 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KRUSE, KRISTOPHER, D

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: CHAN, KAM-WING

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 1149031

DRY GAS EXPIRY DATE: 05/30/2022

VIOLATION CODE: DUI

ARREST TIME: 04:16

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: GARCIA ARIAS

DATE/TIME OF TEST: 09/05/2021 04:54:39

Diagnostics Check	OK	04:47
Air Blank	0.000	04:48
Control Test	0.080	04:48
Air Blank	0.000	04:48
Subject Sample # 1	0.086	04:50
Breath Volume	2.000	04:50
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.084	04:53
Breath Volume	2.089	04:53
Air Blank	0.000	04:53
Control Test	0.080	04:54
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 10/30/2021 18:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KRUSE, KRISTOPHER, D

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: KRUSE, KRISTOPHER

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1149031

DRY GAS EXPIRY DATE: 05/30/2022

VIOLATION CODE: DUI

ARREST TIME: 03:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: GUNTER

DATE/TIME OF TEST: 09/06/2021 04:48:54

Diagnostics Check	OK	04:40
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:41
Subject Sample # 1	0.079	04:44
Breath Volume	1.414	04:44
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.088	04:47
Breath Volume	2.960	04:47
Air Blank	0.000	04:47
Control Test	0.080	04:48
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 10/30/2021 18:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GUY,KEVIN,

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: CHAN,KAM

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1149031

DRY GAS EXPIRY DATE: 05/30/2022

VIOLATION CODE: DUI

ARREST TIME: 22:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: GOSSETT

DATE/TIME OF TEST: 09/13/2021 23:45:43

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:43	
Control Test	0.080	23:43	
Air Blank	0.000	23:44	
Subject Sample	# 1	23:44	Subject Test Refused
Breath Volume	0.000	23:44	
Air Blank	0.000	23:44	
Control Test	0.081	23:45	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:45	

Data Download Date/Time: 10/30/2021 18:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARPER,JIMMIE,L
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: CAMACHO,N
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1149031
DRY GAS EXPIRY DATE: 05/30/2022
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 09/18/2021 22:15:37

Diagnostics Check	OK	22:08
Air Blank	0.000	22:09
Control Test	0.080	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.205	22:10
Breath Volume	1.273	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.196	22:14
Breath Volume	1.195	22:14
Air Blank	0.000	22:14
Control Test	0.080	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 10/30/2021 18:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARPER,JIMMIE,L

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: MORALES,K

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1149031

DRY GAS EXPIRY DATE: 05/30/2022

VIOLATION CODE: DUI

ARREST TIME: 03:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SMART

DATE/TIME OF TEST: 09/20/2021 04:18:07

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:15	
Control Test	0.080	04:16	
Air Blank	0.000	04:16	
Subject Sample	# 1	04:16	Subject Test Refused
Breath Volume	0.000	04:16	
Air Blank	0.000	04:17	
Control Test	0.081	04:17	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

Data Download Date/Time: 10/30/2021 18:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE, KRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: CAMACHO, NICOLE
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1149031
DRY GAS EXPIRY DATE: 05/30/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 09/29/2021 03:26:41

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:23	
Control Test	0.080	03:24	
Air Blank	0.000	03:24	
Subject Sample	# 1	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:25	
Control Test	0.080	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 10/30/2021 18:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: GERIER

DATE/TIME OF TEST: 09/01/2021 21:02:12

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.080	20:56
Air Blank	0.000	20:57
Subject Sample # 1	0.283	20:57
Breath Volume	1.988	20:57
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.276	21:00
Breath Volume	1.617	21:00
Air Blank	0.000	21:01
Control Test	0.080	21:01
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 10/01/2021 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 09/02/2021 05:17:54

Diagnostics Check	OK	05:11
Air Blank	0.000	05:11
Control Test	0.081	05:11
Air Blank	0.000	05:12
Subject Sample # 1	0.172	05:12
Breath Volume	1.554	05:12
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.181	05:16
Breath Volume	1.414	05:16
Air Blank	0.000	05:16
Control Test	0.081	05:17
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 10/01/2021 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PELFREY,PAUL,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: KYPROS

DATE/TIME OF TEST: 09/02/2021 22:52:24

Diagnostics Check	OK	22:40	
Air Blank	0.000	22:41	
Control Test	0.080	22:41	
Air Blank	0.000	22:42	
Subject Sample	# 1	22:45	No Sample Provided
Breath Volume	0.000	22:45	
Air Blank	0.000	22:45	
Air Blank	0.000	22:47	
Subject Sample	# 2	0.235	22:50
Breath Volume	0.160	22:50	Volume Not Met
Air Blank	0.000	22:51	
Control Test	0.081	22:51	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 10/01/2021 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: THOMPSON,LATHAM
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 09/03/2021 02:16:57

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.139	02:12
Breath Volume	3.289	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.134	02:15
Breath Volume	2.781	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 10/01/2021 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 09/05/2021 03:30:33

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.081	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.148	03:26
Breath Volume	3.328	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.146	03:29
Breath Volume	3.558	03:29
Air Blank	0.000	03:29
Control Test	0.081	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 10/01/2021 11:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGHT,BRANDON,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: MCMULLEN

DATE/TIME OF TEST: 09/05/2021 20:57:11

Diagnostics Check	OK	20:50
Air Blank	0.000	20:50
Control Test	0.081	20:51
Air Blank	0.000	20:51
Subject Sample # 1	0.177	20:52
Breath Volume	2.093	20:52
Air Blank	0.000	20:52
Air Blank	0.000	20:54
Subject Sample # 2	0.170	20:55
Breath Volume	2.550	20:55
Air Blank	0.000	20:56
Control Test	0.081	20:56
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

Data Download Date/Time: 10/01/2021 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGHT,BRANDON,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: LEFTWICH

DATE/TIME OF TEST: 09/05/2021 22:19:21

Diagnostics Check	OK	22:10
Air Blank	0.000	22:10
Control Test	0.080	22:11
Air Blank	0.000	22:11
Subject Sample # 1	0.134	22:13
Breath Volume	1.335	22:13
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.138	22:17
Breath Volume	1.656	22:17
Air Blank	0.000	22:18
Control Test	0.081	22:18
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 10/01/2021 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PELFREY,PAUL,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: YOUNG,JOHNATHAN
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: HINSON

DATE/TIME OF TEST: 09/11/2021 02:25:23

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.081	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.153	02:20
Breath Volume	1.417	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.150	02:23
Breath Volume	2.085	02:23
Air Blank	0.000	02:24
Control Test	0.082	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 10/01/2021 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PELFREY,PAUL,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: PELFREY,PAUL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: HANLEY

DATE/TIME OF TEST: 09/12/2021 21:44:12

Diagnostics Check	OK	21:35
Air Blank	0.000	21:35
Control Test	0.081	21:35
Air Blank	0.000	21:36
Subject Sample # 1	0.241	21:38
Breath Volume	1.210	21:38
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.225	21:42
Breath Volume	1.289	21:42
Air Blank	0.000	21:43
Control Test	0.081	21:43
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 10/01/2021 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 09/13/2021 02:02:18

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:01	
Control Test	0.073	02:01	Control Outside Tolerance
Air Blank	0.000	02:02	

Data Download Date/Time: 10/01/2021 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 09/13/2021 02:13:14

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.076	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.221	02:08
Breath Volume	1.269	02:08
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.225	02:11
Breath Volume	1.484	02:11
Air Blank	0.000	02:12
Control Test	0.078	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 10/01/2021 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGHT,BRANDON,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/01/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: BRUNO

DATE/TIME OF TEST: 09/18/2021 04:46:37

Diagnostics Check	OK	04:40
Air Blank	0.000	04:40
Control Test	0.081	04:41
Air Blank	0.000	04:41
Subject Sample # 1	0.180	04:42
Breath Volume	1.828	04:42
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.190	04:45
Breath Volume	1.500	04:45
Air Blank	0.000	04:45
Control Test	0.082	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 10/01/2021 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGHT,BRANDON,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: KIGHT,BRANDON
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: MION

DATE/TIME OF TEST: 09/19/2021 05:27:07

Diagnostics Check	OK	05:20
Air Blank	0.000	05:20
Control Test	0.081	05:21
Air Blank	0.000	05:21
Subject Sample # 1	0.192	05:22
Breath Volume	2.355	05:22
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 2	0.176	05:25
Breath Volume	2.023	05:25
Air Blank	0.000	05:26
Control Test	0.082	05:26
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 10/01/2021 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: FORTESCUE

DATE/TIME OF TEST: 09/22/2021 03:56:24

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:53	
Control Test	0.081	03:54	
Air Blank	0.000	03:54	
Subject Sample	# 1	03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:55	
Control Test	0.081	03:55	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 10/01/2021 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 09/30/2021 03:36:01

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:20	
Control Test	0.081	03:21	
Air Blank	0.000	03:21	
Subject Sample	# 1 0.220	03:24	Volume Not Met
Breath Volume	0.378	03:24	
Air Blank	0.000	03:25	
Air Blank	0.000	03:27	
Subject Sample	# 2 0.233	03:28	
Breath Volume	1.171	03:28	
Air Blank	0.000	03:29	
Air Blank	0.000	03:31	
Subject Sample	# 3 0.204	03:34	Volume Not Met
Breath Volume	0.187	03:34	
Air Blank	0.000	03:35	
Control Test	0.080	03:35	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:36	

Data Download Date/Time: 10/01/2021 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,SHANNON,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ENDELMAN,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: CAVALIERE

DATE/TIME OF TEST: 09/17/2021 03:43:26

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.107	03:38
Breath Volume	2.425	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.105	03:41
Breath Volume	1.414	03:41
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 10/13/2021 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 09/19/2021 01:13:18

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.080	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.199	01:08
Breath Volume	2.101	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.210	01:11
Breath Volume	1.753	01:11
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 10/13/2021 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAVEZ CABRERA,A,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HARRINGTON,J
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: PALACIO

DATE/TIME OF TEST: 09/19/2021 01:27:56

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:22	
Control Test	0.079	01:22	
Air Blank	0.000	01:22	
Subject Sample	# 1 0.081	01:24	
Breath Volume	2.152	01:24	
Air Blank	0.000	01:24	
Air Blank	0.000	01:26	
Subject Sample	# 2	01:27	RFI Detect
Breath Volume	0.000	01:27	
Air Blank		01:27	RFI Detect
Air Blank	0.000	01:27	

Data Download Date/Time: 10/13/2021 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAVEZ CABRERA,A,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HARRINGTON,J
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: PALACIO

DATE/TIME OF TEST: 09/19/2021 01:39:23

Diagnostics Check	OK	01:32
Air Blank	0.000	01:33
Control Test	0.080	01:33
Air Blank	0.000	01:34
Subject Sample # 1	0.079	01:34
Breath Volume	1.960	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.078	01:37
Breath Volume	1.292	01:37
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 10/13/2021 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA,R,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: JOHNSON,T
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: HAZZARD

DATE/TIME OF TEST: 09/19/2021 03:14:23

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.219	03:09
Breath Volume	1.625	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.203	03:12
Breath Volume	1.324	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 10/13/2021 09:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYRNIOS,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CAUDILL,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: LEMBO

DATE/TIME OF TEST: 09/19/2021 21:45:15

Diagnostics Check	OK	21:38
Air Blank	0.000	21:39
Control Test	0.080	21:39
Air Blank	0.000	21:40
Subject Sample # 1	0.000	21:40
Breath Volume	2.152	21:40
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.000	21:43
Breath Volume	2.796	21:43
Air Blank	0.000	21:44
Control Test	0.079	21:44
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 10/13/2021 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CURRY,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: DWELLE

DATE/TIME OF TEST: 09/20/2021 01:21:07

Diagnostics Check	OK	01:13
Air Blank	0.000	01:14
Control Test	0.080	01:14
Air Blank	0.000	01:14
Subject Sample # 1	0.040	01:15
Breath Volume	1.464	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.036	01:19
Breath Volume	1.699	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 10/13/2021 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HOEL,JEFF
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: POLCE

DATE/TIME OF TEST: 09/22/2021 22:26:29

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.081	22:20
Air Blank	0.000	22:21
Subject Sample # 1	0.052	22:21
Breath Volume	2.414	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.049	22:25
Breath Volume	2.273	22:25
Air Blank	0.000	22:25
Control Test	0.080	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 10/13/2021 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,MAXWELL,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SOLIS,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: RUBINO

DATE/TIME OF TEST: 09/23/2021 00:45:48

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.137	00:41
Breath Volume	2.558	00:41
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.130	00:44
Breath Volume	2.140	00:44
Air Blank	0.000	00:44
Control Test	0.079	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 10/13/2021 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ZACHARY,H
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WILLIAMS,ZACHARY
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: VALDEZ

DATE/TIME OF TEST: 09/23/2021 02:47:13

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:36	
Control Test	0.080	02:36	
Air Blank	0.000	02:36	
Subject Sample	# 1 0.133	02:40	Volume Not Met
Breath Volume	0.152	02:40	
Air Blank	0.000	02:40	
Air Blank	0.000	02:42	
Subject Sample	# 2 0.148	02:45	Volume Not Met
Breath Volume	0.082	02:45	
Air Blank	0.000	02:46	
Control Test	0.078	02:46	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 10/13/2021 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ZACHARY,H
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WILLIAMS,ZACHARY
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: VALDEZ

DATE/TIME OF TEST: 09/23/2021 02:59:39

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:50	
Control Test	0.078	02:50	
Air Blank	0.000	02:51	
Subject Sample	# 1 0.139	02:54	Volume Not Met
Breath Volume	0.089	02:54	
Air Blank	0.000	02:54	
Air Blank	0.000	02:56	
Subject Sample	# 2	02:58	RFI Detect
Breath Volume	0.234	02:58	
Air Blank	0.000	02:58	
Control Test	0.076	02:59	
Air Blank	0.000	02:59	
Diagnostics Check	OK	02:59	

Data Download Date/Time: 10/13/2021 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ZACHARY,H
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WILLIAMS,ZACHARY
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: VALDEZ

DATE/TIME OF TEST: 09/23/2021 03:13:29

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:02	
Control Test	0.080	03:03	
Air Blank	0.000	03:03	
Subject Sample	# 1 0.133	03:06	Volume Not Met
Breath Volume	0.597	03:06	
Air Blank	0.000	03:07	
Air Blank	0.000	03:09	
Subject Sample	# 2	03:11	Subject Test Refused
Breath Volume	0.015	03:11	
Air Blank	0.000	03:12	
Control Test	0.075	03:12	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 10/13/2021 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ZACHARY,H
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WILLIAMS,ZACHARY
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: ARBAUGH

DATE/TIME OF TEST: 09/24/2021 05:04:01

Diagnostics Check	OK	04:57
Air Blank	0.000	04:58
Control Test	0.080	04:58
Air Blank	0.000	04:59
Subject Sample # 1	0.270	04:59
Breath Volume	2.527	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.282	05:02
Breath Volume	2.656	05:02
Air Blank	0.000	05:03
Control Test	0.078	05:03
Air Blank	0.000	05:03
Diagnostics Check	OK	05:04

Data Download Date/Time: 10/13/2021 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: TURNER,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 09/25/2021 22:00:03

Diagnostics Check	OK	21:54
Air Blank	0.000	21:54
Control Test	0.080	21:54
Air Blank	0.000	21:55
Subject Sample # 1	0.193	21:55
Breath Volume	2.425	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.192	21:58
Breath Volume	2.179	21:58
Air Blank	0.000	21:59
Control Test	0.080	21:59
Air Blank	0.000	21:59
Diagnostics Check	OK	22:00

Data Download Date/Time: 10/13/2021 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS, THOMAS, L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PEREZ, M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: ECKHARDT

DATE/TIME OF TEST: 09/27/2021 03:29:45

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.186	03:25
Breath Volume	2.125	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.190	03:28
Breath Volume	2.320	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 10/13/2021 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LASALA,J,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KIRSH,Z
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: LANDERS

DATE/TIME OF TEST: 09/27/2021 22:06:56

Diagnostics Check	OK	21:59
Air Blank	0.000	21:59
Control Test	0.079	21:59
Air Blank	0.000	22:00
Subject Sample # 1	0.109	22:02
Breath Volume	1.570	22:02
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.108	22:05
Breath Volume	1.257	22:05
Air Blank	0.000	22:06
Control Test	0.079	22:06
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 10/13/2021 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,E,W
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GROSSKLAS,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: TRNKA

DATE/TIME OF TEST: 09/27/2021 22:44:35

Diagnostics Check	OK	22:34
Air Blank	0.000	22:35
Control Test	0.079	22:35
Air Blank	0.000	22:36
Subject Sample # 1	0.238	22:38
Breath Volume	1.578	22:38
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.244	22:42
Breath Volume	1.683	22:42
Air Blank	0.000	22:43
Control Test	0.078	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 10/13/2021 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GASH,C,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GASH,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 09/29/2021 02:27:48

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.080	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.203	02:20
Breath Volume	2.457	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.239	02:23
Breath Volume	2.539	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 3	0.244	02:26
Breath Volume	3.480	02:26
Air Blank	0.000	02:26
Control Test	0.079	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 10/13/2021 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,S,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GONZALEZ,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/30/2021 04:13:39

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:58	
Control Test	0.080	03:58	
Air Blank	0.000	03:59	
Subject Sample	# 1 0.191	04:02	Volume Not Met
Breath Volume	0.210	04:02	
Air Blank	0.000	04:02	
Air Blank	0.000	04:04	
Subject Sample	# 2 0.190	04:07	
Breath Volume	1.199	04:07	
Air Blank	0.000	04:08	
Air Blank	0.000	04:09	
Subject Sample	# 3 0.189	04:12	
Breath Volume	1.191	04:12	
Air Blank	0.000	04:12	
Control Test	0.078	04:13	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 10/13/2021 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GASH,C,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GASH,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 09/05/2021 00:45:11

Diagnostics Check	OK	00:38
Air Blank	0.000	00:39
Control Test	0.080	00:39
Air Blank	0.000	00:40
Subject Sample # 1	0.258	00:40
Breath Volume	2.238	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.246	00:43
Breath Volume	1.753	00:43
Air Blank	0.000	00:44
Control Test	0.077	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 09/16/2021 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUONPANE,V,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WHITE,T
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: SOWELL

DATE/TIME OF TEST: 09/18/2021 21:37:11

Diagnostics Check	OK	21:24	
Air Blank	0.000	21:25	
Control Test	0.078	21:25	
Air Blank	0.000	21:26	
Subject Sample	# 1 0.137	21:29	Volume Not Met
Breath Volume	0.968	21:29	
Air Blank	0.000	21:30	
Air Blank	0.000	21:31	
Subject Sample	# 2 0.160	21:32	
Breath Volume	1.476	21:32	
Air Blank	0.000	21:33	
Air Blank	0.000	21:34	
Subject Sample	# 3 0.162	21:35	
Breath Volume	1.214	21:35	
Air Blank	0.000	21:36	
Control Test	0.078	21:36	
Air Blank	0.000	21:37	
Diagnostics Check	OK	21:37	

Data Download Date/Time: 10/14/2021 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,U,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ROWDEN,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 09/25/2021 03:41:51

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.080	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.191	03:37
Breath Volume	1.867	03:37
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.189	03:40
Breath Volume	2.890	03:40
Air Blank	0.000	03:40
Control Test	0.080	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 10/14/2021 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,ANDREW,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ZERPA,EDWIN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: ZIEGLER

DATE/TIME OF TEST: 09/29/2021 01:44:19

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.080	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.179	01:39
Breath Volume	1.617	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:41
Subject Sample # 2	0.188	01:42
Breath Volume	2.417	01:42
Air Blank	0.000	01:43
Control Test	0.078	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 10/14/2021 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUTCHINSON,JOSEPH,B
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HUTCHINSON,JOSEPH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 09:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: DRAHER

DATE/TIME OF TEST: 09/01/2021 09:47:59

Diagnostics Check	OK	09:41
Air Blank	0.000	09:42
Control Test	0.081	09:42
Air Blank	0.000	09:42
Subject Sample # 1	0.000	09:43
Breath Volume	1.527	09:43
Air Blank	0.000	09:43
Air Blank	0.000	09:45
Subject Sample # 2	0.000	09:46
Breath Volume	2.554	09:46
Air Blank	0.000	09:47
Control Test	0.080	09:47
Air Blank	0.000	09:47
Diagnostics Check	OK	09:47

Data Download Date/Time: 09/21/2021 23:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: VALDEZ,CINDY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 08/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: FLATT

DATE/TIME OF TEST: 09/07/2021 04:40:13

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.081	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.161	04:35
Breath Volume	2.019	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.161	04:38
Breath Volume	1.597	04:38
Air Blank	0.000	04:39
Control Test	0.081	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 09/21/2021 23:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,GARRETT,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: KINGRY,HOBIE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 15:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:11

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 09/16/2021 16:42:27

Diagnostics Check	OK	16:32	
Air Blank	0.000	16:32	
Control Test	0.081	16:32	
Air Blank	0.000	16:33	
Subject Sample	# 1 0.260	16:36	Volume Not Met
Breath Volume	0.261	16:36	
Air Blank	0.000	16:37	
Air Blank	0.000	16:38	
Subject Sample	# 2	16:41	Subject Test Refused
Breath Volume	0.000	16:41	
Air Blank	0.000	16:41	
Control Test	0.079	16:41	
Air Blank	0.000	16:42	
Diagnostics Check	OK	16:42	

Data Download Date/Time: 09/21/2021 23:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,EVEN,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: SCelfo,ZACHARY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 09/26/2021 01:20:10

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:15	
Control Test	0.082	01:15	
Air Blank	0.000	01:16	
Subject Sample	# 1	01:18	Subject Test Refused
Breath Volume	0.066	01:18	
Air Blank	0.000	01:19	
Control Test	0.083	01:19	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 09/27/2021 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AFONSO,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: SCelfo,ZACHARY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 09/30/2021 23:09:49

Diagnostics Check	OK	23:06	
Air Blank	0.000	23:07	
Control Test	0.081	23:07	
Air Blank	0.000	23:07	
Subject Sample	# 1	23:08	Subject Test Refused
Breath Volume	0.000	23:08	
Air Blank	0.000	23:08	
Control Test	0.081	23:09	
Air Blank	0.000	23:09	
Diagnostics Check	OK	23:09	

Data Download Date/Time: 10/07/2021 09:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWLESS, ABIGAIL, D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CARTER, JACOB
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: BUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: CASTLEBERRY

DATE/TIME OF TEST: 09/04/2021 00:08:36

Diagnostics Check	OK	00:05	
Air Blank	0.000	00:06	
Control Test	0.080	00:06	
Air Blank	0.000	00:07	
Subject Sample	# 1	00:07	Subject Test Refused
Breath Volume	0.000	00:07	
Air Blank	0.000	00:07	
Control Test	0.080	00:08	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:08	

Data Download Date/Time: 09/16/2021 17:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,MELISSA,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SILLS,MARK
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: MUNOZ-BORGES

DATE/TIME OF TEST: 09/04/2021 23:07:29

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.080	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.170	23:03
Breath Volume	2.660	23:03
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.162	23:06
Breath Volume	1.386	23:06
Air Blank	0.000	23:06
Control Test	0.080	23:06
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 09/16/2021 17:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: CRUZ-GARCIA

DATE/TIME OF TEST: 09/05/2021 03:52:15

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.080	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.197	03:47
Breath Volume	2.453	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.199	03:50
Breath Volume	1.117	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 09/16/2021 17:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWAN,DANNY,A
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: SWAN,DANNY
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: JERNIGAN

DATE/TIME OF TEST: 09/05/2021 04:49:17

Diagnostics Check	OK	04:43	
Air Blank	0.000	04:43	
Control Test	0.080	04:43	
Air Blank	0.000	04:44	
Subject Sample # 1	0.146	04:44	
Breath Volume	2.679	04:44	
Air Blank	0.000	04:45	
Air Blank	0.000	04:47	
Subject Sample # 2	0.149	04:47	
Breath Volume	2.734	04:47	
Air Blank	0.000	04:48	
Control Test	0.086	04:48	Control Outside Tolerance
Air Blank	0.000	04:49	

Data Download Date/Time: 09/16/2021 17:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWAN,DANNY,A
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: SWAN,DANNY
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: JERNIGAN

DATE/TIME OF TEST: 09/05/2021 05:00:19

Diagnostics Check	OK	04:53
Air Blank	0.000	04:54
Control Test	0.080	04:54
Air Blank	0.000	04:54
Subject Sample # 1	0.133	04:55
Breath Volume	2.554	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.139	04:58
Breath Volume	2.371	04:58
Air Blank	0.000	04:59
Control Test	0.079	04:59
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 09/16/2021 17:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAPPIER,PAUL,A
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: NAPPIER,PAUL
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: EVERETTS

DATE/TIME OF TEST: 09/06/2021 00:08:30

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:07	
Control Test	0.080	00:08	
Air Blank		00:08	Improper Sample
Air Blank		00:08	Improper Sample

Data Download Date/Time: 09/16/2021 17:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAPPIER,PAUL,A
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: NAPPIER,PAUL
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: EVERETTS

DATE/TIME OF TEST: 09/06/2021 00:11:06

Diagnostics Check	OK	00:10	
Air Blank		00:10	Improper Sample
Air Blank		00:11	Improper Sample

Data Download Date/Time: 09/16/2021 17:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAPPIER,PAUL,A
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: NAPPIER,PAUL
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 09:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: EVERETTS

DATE/TIME OF TEST: 09/06/2021 00:19:57

Diagnostics Check	OK	00:19	
Air Blank		00:19	Improper Sample
Air Blank		00:19	Improper Sample

Data Download Date/Time: 09/16/2021 17:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAPPIER,PAUL,A
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: NAPPIER,PAUL
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 09:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: EVERETTS

DATE/TIME OF TEST: 09/06/2021 00:23:22

Diagnostics Check	OK	00:22	
Air Blank		00:22	Ambient Fail
Air Blank	0.000	00:23	

Data Download Date/Time: 09/16/2021 17:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONCION,CHRISTIAN,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MONCION,CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: CARACHEO FRIAS

DATE/TIME OF TEST: 09/06/2021 03:05:00

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:58
Subject Sample # 1	0.215	03:00
Breath Volume	1.328	03:00
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.224	03:03
Breath Volume	1.105	03:03
Air Blank	0.000	03:04
Control Test	0.078	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:05

Data Download Date/Time: 09/16/2021 17:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES, STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES, STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 09/07/2021 18:17:09

Diagnostics Check	OK	18:10
Air Blank	0.000	18:11
Control Test	0.080	18:11
Air Blank	0.000	18:12
Subject Sample # 1	0.000	18:12
Breath Volume	2.421	18:12
Air Blank	0.000	18:13
Air Blank	0.000	18:15
Subject Sample # 2	0.000	18:15
Breath Volume	1.691	18:15
Air Blank	0.000	18:16
Control Test	0.079	18:16
Air Blank	0.000	18:17
Diagnostics Check	OK	18:17

Data Download Date/Time: 09/16/2021 17:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,MELISSA,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SPENCE,QUAMISHA
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 09/08/2021 20:48:02

Diagnostics Check	OK	20:39
Air Blank	0.000	20:39
Control Test	0.080	20:40
Air Blank	0.000	20:40
Subject Sample # 1	0.159	20:42
Breath Volume	2.148	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:44
Subject Sample # 2	0.157	20:46
Breath Volume	2.128	20:46
Air Blank	0.000	20:47
Control Test	0.080	20:47
Air Blank	0.000	20:47
Diagnostics Check	OK	20:48

Data Download Date/Time: 09/16/2021 17:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: DONATELLI

DATE/TIME OF TEST: 09/11/2021 00:08:29

Diagnostics Check	OK	00:01
Air Blank	0.000	00:02
Control Test	0.079	00:02
Air Blank	0.000	00:03
Subject Sample # 1	0.150	00:03
Breath Volume	1.632	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.148	00:07
Breath Volume	1.683	00:07
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 09/16/2021 17:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/12/2021 00:38:36

Diagnostics Check	OK	00:31
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:33
Subject Sample # 1	0.183	00:34
Breath Volume	1.257	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.175	00:37
Breath Volume	1.250	00:37
Air Blank	0.000	00:37
Control Test	0.080	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 09/16/2021 17:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWSON,ASHLEY
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: MERIDA

DATE/TIME OF TEST: 09/12/2021 01:09:42

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.080	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.267	01:05
Breath Volume	2.582	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.254	01:08
Breath Volume	2.136	01:08
Air Blank	0.000	01:08
Control Test	0.080	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 09/16/2021 17:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SELLERS,JAMES
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: RESENDEZ

DATE/TIME OF TEST: 09/13/2021 03:29:54

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.311	03:25
Breath Volume	3.195	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.304	03:28
Breath Volume	3.121	03:28
Air Blank	0.000	03:28
Control Test	0.079	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 09/16/2021 17:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONCION,CHRISTIAN,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MONCION,CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: RIVERA FONTANEZ

DATE/TIME OF TEST: 09/17/2021 02:16:01

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:12	
Control Test	0.081	02:13	
Air Blank	0.000	02:13	
Subject Sample	# 1	02:14	RFI Detect
Breath Volume	0.000	02:14	
Air Blank	0.000	02:15	
Control Test	0.081	02:15	
Air Blank	0.000	02:15	
Diagnostics Check	OK	02:16	

Data Download Date/Time: 10/18/2021 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONCION,CHRISTIAN,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MONCION,CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: RIVERA FONTANEZ

DATE/TIME OF TEST: 09/17/2021 02:21:11

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:18	
Control Test	0.080	02:19	
Air Blank	0.000	02:19	
Subject Sample	# 1	02:19	Subject Test Refused
Breath Volume	0.000	02:19	
Air Blank	0.000	02:20	
Control Test	0.080	02:20	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time: 10/18/2021 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN, GERALD, E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PARANOUSKES, D
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: CAREL

DATE/TIME OF TEST: 09/15/2021 00:31:23

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.135	00:26
Breath Volume	2.234	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.140	00:29
Breath Volume	1.488	00:29
Air Blank	0.000	00:30
Control Test	0.079	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 09/29/2021 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTELLANO,VINCENT,J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: CASTELLANO,VINCENT
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 08/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: SANTIAGO RIVERA

DATE/TIME OF TEST: 09/05/2021 22:36:21

Diagnostics Check	OK	22:28
Air Blank	0.000	22:29
Control Test	0.080	22:29
Air Blank	0.000	22:30
Subject Sample # 1	0.184	22:31
Breath Volume	1.968	22:31
Air Blank	0.000	22:32
Air Blank	0.000	22:33
Subject Sample # 2	0.187	22:34
Breath Volume	2.429	22:34
Air Blank	0.000	22:35
Control Test	0.080	22:35
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 09/15/2021 01:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTELLANO,VINCENT,J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: CASTELLANO,VINCENT
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 08/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: OLIVEIRA

DATE/TIME OF TEST: 09/06/2021 03:26:43

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.080	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.207	03:21
Breath Volume	1.359	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.210	03:25
Breath Volume	1.539	03:25
Air Blank	0.000	03:25
Control Test	0.078	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 09/15/2021 01:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIEDESCH, LAURA, R
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: GAFFNEY, CASEY
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 08/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 09/11/2021 04:24:44

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.081	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.000	04:20
Breath Volume	3.382	04:20
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.000	04:23
Breath Volume	3.074	04:23
Air Blank	0.000	04:23
Control Test	0.079	04:24
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 09/15/2021 01:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTZ,KATHRYN,N
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: WOOLEYHAN,HEATHER
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 08/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 07:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:31

SUBJECT'S LAST NAME: LAMOND

DATE/TIME OF TEST: 09/13/2021 07:59:14

Diagnostics Check	OK	07:52
Air Blank	0.000	07:52
Control Test	0.080	07:52
Air Blank	0.000	07:53
Subject Sample # 1	0.144	07:53
Breath Volume	1.507	07:53
Air Blank	0.000	07:54
Air Blank	0.000	07:56
Subject Sample # 2	0.140	07:57
Breath Volume	1.289	07:57
Air Blank	0.000	07:58
Control Test	0.080	07:58
Air Blank	0.000	07:59
Diagnostics Check	OK	07:59

Data Download Date/Time: 09/15/2021 01:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVILA,NICOLE,M
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MORRIS,LANCE
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 13:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:24

SUBJECT'S LAST NAME: MCCAULEY

DATE/TIME OF TEST: 09/13/2021 14:51:58

Diagnostics Check	OK	14:45
Air Blank	0.000	14:45
Control Test	0.081	14:46
Air Blank	0.000	14:46
Subject Sample # 1	0.000	14:47
Breath Volume	2.250	14:47
Air Blank	0.000	14:47
Air Blank	0.000	14:49
Subject Sample # 2	0.000	14:50
Breath Volume	1.195	14:50
Air Blank	0.000	14:51
Control Test	0.083	14:51
Air Blank	0.000	14:51
Diagnostics Check	OK	14:51

Data Download Date/Time: 10/04/2021 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMBSH,CARLYTON,R
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: CALENDAN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 14:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: HENNESSEY

DATE/TIME OF TEST: 09/15/2021 15:35:35

Diagnostics Check	OK	15:29
Air Blank	0.000	15:29
Control Test	0.082	15:30
Air Blank	0.000	15:30
Subject Sample # 1	0.000	15:31
Breath Volume	3.757	15:31
Air Blank	0.000	15:31
Air Blank	0.000	15:33
Subject Sample # 2	0.000	15:34
Breath Volume	2.503	15:34
Air Blank	0.000	15:34
Control Test	0.082	15:35
Air Blank	0.000	15:35
Diagnostics Check	OK	15:35

Data Download Date/Time: 10/04/2021 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAINES,CHRISTOPHER,G
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: CHUMITA,DOUGLAS
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: HOLMGREN

DATE/TIME OF TEST: 09/20/2021 00:54:35

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.082	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.190	00:50
Breath Volume	2.089	00:50
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.199	00:53
Breath Volume	2.085	00:53
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 10/04/2021 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOVANOVICH, THOMAS, E
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MANCHESTER, K
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: OMALLEY

DATE/TIME OF TEST: 09/27/2021 01:48:43

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.082	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.184	01:41
Breath Volume	3.972	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.211	01:44
Breath Volume	2.203	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 3	0.193	01:47
Breath Volume	3.140	01:47
Air Blank	0.000	01:47
Control Test	0.081	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 10/04/2021 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVILA,NICOLE,M
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: AVILA,NICOLE
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 11:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:16

SUBJECT'S LAST NAME: CHACE

DATE/TIME OF TEST: 09/27/2021 12:50:33

Diagnostics Check	OK	12:44
Air Blank	0.000	12:44
Control Test	0.082	12:45
Air Blank	0.000	12:45
Subject Sample # 1	0.000	12:46
Breath Volume	2.945	12:46
Air Blank	0.000	12:46
Air Blank	0.000	12:48
Subject Sample # 2	0.000	12:49
Breath Volume	1.621	12:49
Air Blank	0.000	12:49
Control Test	0.082	12:49
Air Blank	0.000	12:50
Diagnostics Check	OK	12:50

Data Download Date/Time: 10/04/2021 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN,BRADLEY,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FARRAGHER,THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: WENNER

DATE/TIME OF TEST: 09/02/2021 01:51:05

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:45	
Control Test	0.080	01:46	
Air Blank	0.000	01:46	
Subject Sample	# 1 0.219	01:47	
Breath Volume	3.996	01:47	
Air Blank	0.000	01:47	
Air Blank	0.000	01:49	
Subject Sample	# 2 0.214	01:50	
Breath Volume	3.851	01:50	
Air Blank		01:50	RFI Detect
Air Blank	0.000	01:51	

Data Download Date/Time: 09/21/2021 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN,BRADLEY,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WESTON,ANDREW
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: PENDELL

DATE/TIME OF TEST: 09/06/2021 23:22:28

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.080	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.228	23:18
Breath Volume	2.144	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.234	23:20
Breath Volume	2.031	23:20
Air Blank	0.000	23:21
Control Test	0.079	23:21
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 09/21/2021 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN,BRADLEY,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: KING,BRIAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: PARTEE

DATE/TIME OF TEST: 09/11/2021 20:01:12

Diagnostics Check	OK	19:49	
Air Blank	0.000	19:50	
Control Test	0.081	19:50	
Air Blank	0.000	19:51	
Subject Sample	# 1	19:54	No Sample Provided
Breath Volume	0.000	19:54	
Air Blank	0.000	19:54	
Air Blank	0.000	19:56	
Subject Sample	# 2	19:59	No Sample Provided
Breath Volume	0.000	19:59	
Air Blank	0.000	20:00	
Control Test	0.080	20:00	
Air Blank	0.000	20:01	
Diagnostics Check	OK	20:01	

Data Download Date/Time: 09/21/2021 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN, BRADLEY,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FARRAGHER, THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: CANADA

DATE/TIME OF TEST: 09/12/2021 02:54:47

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.080	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.176	02:50
Breath Volume	1.804	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.184	02:53
Breath Volume	2.417	02:53
Air Blank	0.000	02:53
Control Test	0.080	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 09/21/2021 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JUSTIN,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ITRI,ANGELA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 09/28/2021 02:44:29

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.080	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.073	02:39
Breath Volume	1.132	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.076	02:43
Breath Volume	1.101	02:43
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 10/28/2021 09:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARARD,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 09/22/2021 23:17:38

Diagnostics Check	OK	23:11
Air Blank	0.000	23:11
Control Test	0.081	23:12
Air Blank	0.000	23:12
Subject Sample # 1	0.149	23:13
Breath Volume	3.429	23:13
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.141	23:16
Breath Volume	3.386	23:16
Air Blank	0.000	23:16
Control Test	0.080	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 10/14/2021 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HERRERABAQUERO,DANNY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 09/25/2021 03:39:23

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.080	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.142	03:34
Breath Volume	2.394	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.140	03:37
Breath Volume	2.652	03:37
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 10/14/2021 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: VANCLEEF

DATE/TIME OF TEST: 09/25/2021 04:58:59

Diagnostics Check	OK	04:47	
Air Blank	0.000	04:48	
Control Test	0.081	04:48	
Air Blank	0.000	04:48	
Subject Sample	# 1	04:52	No Sample Provided
Breath Volume	0.000	04:52	
Air Blank	0.000	04:52	
Air Blank	0.000	04:54	
Subject Sample	# 2	04:57	No Sample Provided
Breath Volume	0.000	04:57	
Air Blank	0.000	04:58	
Control Test	0.081	04:58	
Air Blank	0.000	04:58	
Diagnostics Check	OK	04:58	

Data Download Date/Time: 10/14/2021 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARARD,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: OQUENDOBETANCOURT

DATE/TIME OF TEST: 09/30/2021 03:18:13

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.081	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.191	03:13
Breath Volume	3.023	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.186	03:16
Breath Volume	3.218	03:16
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 10/14/2021 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: AUGUSTIN

DATE/TIME OF TEST: 09/30/2021 05:16:27

Diagnostics Check	OK	05:09
Air Blank	0.000	05:10
Control Test	0.082	05:10
Air Blank	0.000	05:11
Subject Sample # 1	0.000	05:11
Breath Volume	1.910	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.000	05:15
Breath Volume	2.066	05:15
Air Blank	0.000	05:15
Control Test	0.081	05:15
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 10/14/2021 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CARBONE,ANDRES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: MATHIS

DATE/TIME OF TEST: 09/30/2021 22:08:26

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.081	22:03
Air Blank	0.000	22:03
Subject Sample # 1	0.000	22:04
Breath Volume	1.429	22:04
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.000	22:07
Breath Volume	1.148	22:07
Air Blank	0.000	22:07
Control Test	0.080	22:07
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 10/14/2021 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JACKSON,ALLEN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: SUBRAN

DATE/TIME OF TEST: 09/02/2021 01:20:31

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.080	01:15
Air Blank	0.000	01:15
Subject Sample # 1	0.177	01:16
Breath Volume	2.042	01:16
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.183	01:19
Breath Volume	2.195	01:19
Air Blank	0.000	01:19
Control Test	0.080	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 09/08/2021 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS, GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 09/02/2021 01:43:00

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.188	01:38
Breath Volume	3.011	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.184	01:41
Breath Volume	2.332	01:41
Air Blank	0.000	01:42
Control Test	0.079	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:43

Data Download Date/Time: 09/08/2021 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIDOMENICO,GINO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: INACIODOSSANTOS

DATE/TIME OF TEST: 09/02/2021 03:26:07

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.080	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.127	03:21
Breath Volume	1.617	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.126	03:24
Breath Volume	2.550	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 09/08/2021 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WILBORN,ARTHUR
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 09/02/2021 04:31:16

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.080	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.177	04:26
Breath Volume	1.765	04:26
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.191	04:29
Breath Volume	1.777	04:29
Air Blank	0.000	04:30
Control Test	0.080	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 09/08/2021 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KING, ARTHUR
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 09/02/2021 19:51:59

Diagnostics Check	OK	19:45
Air Blank	0.000	19:46
Control Test	0.080	19:46
Air Blank	0.000	19:47
Subject Sample # 1	0.128	19:47
Breath Volume	2.761	19:47
Air Blank	0.000	19:48
Air Blank	0.000	19:50
Subject Sample # 2	0.129	19:50
Breath Volume	2.265	19:50
Air Blank	0.000	19:51
Control Test	0.079	19:51
Air Blank	0.000	19:51
Diagnostics Check	OK	19:51

Data Download Date/Time: 09/08/2021 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALL,ANDREW
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: ALBA

DATE/TIME OF TEST: 09/03/2021 00:58:49

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.080	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.135	00:54
Breath Volume	2.210	00:54
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.124	00:57
Breath Volume	1.425	00:57
Air Blank	0.000	00:57
Control Test	0.080	00:58
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 09/08/2021 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCOTT, DAVE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: ROXTON

DATE/TIME OF TEST: 09/03/2021 23:39:56

Diagnostics Check	OK	23:33
Air Blank	0.000	23:34
Control Test	0.081	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.116	23:35
Breath Volume	2.007	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.117	23:38
Breath Volume	2.054	23:38
Air Blank	0.000	23:39
Control Test	0.080	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 09/08/2021 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JEAN BAPTISTE,ALBENS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: WATTS

DATE/TIME OF TEST: 09/04/2021 00:05:36

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.079	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.123	00:00
Breath Volume	1.464	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.118	00:04
Breath Volume	1.585	00:04
Air Blank	0.000	00:04
Control Test	0.079	00:05
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 09/08/2021 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: SALVAVALENTIN

DATE/TIME OF TEST: 09/04/2021 02:59:09

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.165	02:54
Breath Volume	2.546	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.167	02:57
Breath Volume	3.140	02:57
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 09/08/2021 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: OQUENDO KUILAN

DATE/TIME OF TEST: 09/04/2021 04:19:26

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.080	04:14
Air Blank	0.000	04:14
Subject Sample # 1	0.125	04:15
Breath Volume	2.664	04:15
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.117	04:17
Breath Volume	2.234	04:17
Air Blank	0.000	04:18
Control Test	0.080	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 09/08/2021 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 09/05/2021 02:55:33

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.112	02:51
Breath Volume	2.343	02:51
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.117	02:54
Breath Volume	2.199	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 09/08/2021 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH, STEFAN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: AMERTIL

DATE/TIME OF TEST: 09/05/2021 09:30:02

Diagnostics Check	OK	09:24
Air Blank	0.000	09:24
Control Test	0.081	09:24
Air Blank	0.000	09:25
Subject Sample # 1	0.096	09:25
Breath Volume	1.695	09:25
Air Blank	0.000	09:26
Air Blank	0.000	09:28
Subject Sample # 2	0.092	09:28
Breath Volume	2.062	09:28
Air Blank	0.000	09:29
Control Test	0.080	09:29
Air Blank	0.000	09:29
Diagnostics Check	OK	09:30

Data Download Date/Time: 09/08/2021 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WIEBER,DANIEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 09/05/2021 17:15:03

Diagnostics Check	OK	17:08
Air Blank	0.000	17:09
Control Test	0.080	17:09
Air Blank	0.000	17:09
Subject Sample # 1	0.190	17:10
Breath Volume	3.753	17:10
Air Blank	0.000	17:11
Air Blank	0.000	17:13
Subject Sample # 2	0.184	17:13
Breath Volume	2.929	17:13
Air Blank	0.000	17:14
Control Test	0.080	17:14
Air Blank	0.000	17:14
Diagnostics Check	OK	17:15

Data Download Date/Time: 09/08/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ARAGON,LESLIE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: DALMEUS

DATE/TIME OF TEST: 09/06/2021 04:41:26

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.079	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.287	04:36
Breath Volume	1.394	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.286	04:39
Breath Volume	1.492	04:39
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 09/08/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: DOBSON

DATE/TIME OF TEST: 09/07/2021 03:56:46

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.081	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.011	03:52
Breath Volume	2.203	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.010	03:55
Breath Volume	2.734	03:55
Air Blank	0.000	03:55
Control Test	0.080	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 09/08/2021 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ELLIS, JOSHUA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: ALFRED

DATE/TIME OF TEST: 09/07/2021 22:47:10

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.081	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.147	22:42
Breath Volume	3.171	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.149	22:45
Breath Volume	3.207	22:45
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 09/08/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINCENT, JUSTIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: FRYE

DATE/TIME OF TEST: 09/10/2021 05:12:37

Diagnostics Check	OK	05:03
Air Blank	0.000	05:04
Control Test	0.082	05:04
Air Blank	0.000	05:04
Subject Sample # 1	0.140	05:06
Breath Volume	2.285	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.135	05:11
Breath Volume	2.125	05:11
Air Blank	0.000	05:11
Control Test	0.081	05:12
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 10/14/2021 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL, CRAIG
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: MARAIS

DATE/TIME OF TEST: 09/28/2021 21:05:26

Diagnostics Check	OK	20:58
Air Blank	0.000	20:59
Control Test	0.083	20:59
Air Blank	0.000	21:00
Subject Sample # 1	0.043	21:00
Breath Volume	2.546	21:00
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample # 2	0.039	21:03
Breath Volume	1.574	21:03
Air Blank	0.000	21:04
Control Test	0.082	21:04
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 10/14/2021 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GONZALEZ, ASHLEY
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: GONZALEZ-AGUILAR

DATE/TIME OF TEST: 09/30/2021 03:46:21

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.082	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.223	03:41
Breath Volume	1.953	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.224	03:44
Breath Volume	1.433	03:44
Air Blank	0.000	03:45
Control Test	0.082	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 10/14/2021 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LIEBKNECHT,KEVIN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 09/04/2021 02:52:30

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.198	02:47
Breath Volume	1.406	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:49
Subject Sample # 2	0.215	02:50
Breath Volume	2.636	02:50
Air Blank	0.000	02:51
Control Test	0.078	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 09/08/2021 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK, STEPHEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/04/2021 03:24:18

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.143	03:19
Breath Volume	1.226	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:21
Subject Sample # 2	0.132	03:22
Breath Volume	1.171	03:22
Air Blank	0.000	03:23
Control Test	0.078	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 09/08/2021 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SALAZAR,JUAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: DOSSANTOS

DATE/TIME OF TEST: 09/10/2021 04:34:30

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.080	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.236	04:29
Breath Volume	1.820	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.224	04:32
Breath Volume	1.667	04:32
Air Blank	0.000	04:33
Control Test	0.080	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 10/14/2021 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER, DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: BEDWARD

DATE/TIME OF TEST: 09/11/2021 01:42:24

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.081	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.029	01:37
Breath Volume	1.199	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.028	01:40
Breath Volume	1.367	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 10/14/2021 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIAZ,RICHARD
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: DIONICIO AQUILAR

DATE/TIME OF TEST: 09/12/2021 02:54:26

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.079	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.163	02:50
Breath Volume	1.660	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.154	02:52
Breath Volume	1.535	02:52
Air Blank	0.000	02:53
Control Test	0.078	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 10/14/2021 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: BERRIOS VAZQUEZ

DATE/TIME OF TEST: 09/12/2021 03:49:12

Diagnostics Check	OK	03:42
Air Blank	0.000	03:43
Control Test	0.079	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.127	03:44
Breath Volume	2.585	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.121	03:47
Breath Volume	2.050	03:47
Air Blank	0.000	03:48
Control Test	0.079	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 10/14/2021 09:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 09/12/2021 04:31:05

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.079	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.208	04:26
Breath Volume	1.296	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.190	04:29
Breath Volume	2.007	04:29
Air Blank	0.000	04:30
Control Test	0.079	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 10/14/2021 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WONG SHUE, ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: HAFIANI

DATE/TIME OF TEST: 09/12/2021 08:00:50

Diagnostics Check	OK	07:54
Air Blank	0.000	07:55
Control Test	0.079	07:55
Air Blank	0.000	07:56
Subject Sample # 1	0.009	07:56
Breath Volume	3.847	07:56
Air Blank	0.000	07:57
Air Blank	0.000	07:59
Subject Sample # 2	0.009	07:59
Breath Volume	4.394	07:59
Air Blank	0.000	07:59
Control Test	0.080	08:00
Air Blank	0.000	08:00
Diagnostics Check	OK	08:00

Data Download Date/Time: 10/14/2021 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ARCE,BRUNO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: ALVARADO MANUEL

DATE/TIME OF TEST: 09/12/2021 23:53:20

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.079	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.186	23:48
Breath Volume	1.917	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.184	23:51
Breath Volume	2.773	23:51
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 10/14/2021 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GANESH,BRANDON
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: HOLLEY

DATE/TIME OF TEST: 09/13/2021 01:40:49

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:38	
Control Test	0.080	01:38	
Air Blank	0.000	01:39	
Subject Sample	# 1	01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:39	
Control Test	0.080	01:40	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 10/14/2021 09:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PACE,AUSTIN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: SEEPERSAD

DATE/TIME OF TEST: 09/13/2021 02:23:39

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.079	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.000	02:19
Breath Volume	2.191	02:19
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.000	02:22
Breath Volume	2.011	02:22
Air Blank	0.000	02:22
Control Test	0.080	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 10/14/2021 09:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADRON,MARCIE,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LIESENFELT,JASON
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: MUNIZ

DATE/TIME OF TEST: 09/13/2021 06:56:22

Diagnostics Check	OK	06:44	
Air Blank	0.000	06:45	
Control Test	0.080	06:45	
Air Blank	0.000	06:46	
Subject Sample	# 1 0.042	06:49	Volume Not Met
Breath Volume	0.292	06:49	
Air Blank	0.000	06:49	
Air Blank	0.000	06:51	
Subject Sample	# 2 0.044	06:54	Volume Not Met
Breath Volume	0.523	06:54	
Air Blank	0.000	06:55	
Control Test	0.080	06:55	
Air Blank	0.000	06:56	
Diagnostics Check	OK	06:56	

Data Download Date/Time: 10/14/2021 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADRON,MARCIE,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LIESENFELT,JASON
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: MUNIZ

DATE/TIME OF TEST: 09/13/2021 07:02:03

Diagnostics Check	OK	06:59	
Air Blank	0.000	06:59	
Control Test	0.079	06:59	
Air Blank	0.000	07:00	
Subject Sample	# 1	07:00	Subject Test Refused
Breath Volume	0.000	07:00	
Air Blank	0.000	07:01	
Control Test	0.079	07:01	
Air Blank	0.000	07:01	
Diagnostics Check	OK	07:02	

Data Download Date/Time: 10/14/2021 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JEANBAPTISTE,ALBENS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: METZLER

DATE/TIME OF TEST: 09/16/2021 03:19:04

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.000	03:14
Breath Volume	2.109	03:14
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.000	03:17
Breath Volume	1.820	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:19

Data Download Date/Time: 10/14/2021 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SAENGCHOTE,PHANUDEJ
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: LUGG

DATE/TIME OF TEST: 09/16/2021 05:58:10

Diagnostics Check	OK	05:51
Air Blank	0.000	05:52
Control Test	0.080	05:52
Air Blank	0.000	05:53
Subject Sample # 1	0.147	05:53
Breath Volume	2.476	05:53
Air Blank	0.000	05:54
Air Blank	0.000	05:56
Subject Sample # 2	0.149	05:56
Breath Volume	2.464	05:56
Air Blank	0.000	05:57
Control Test	0.079	05:57
Air Blank	0.000	05:58
Diagnostics Check	OK	05:58

Data Download Date/Time: 10/14/2021 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 09/17/2021 03:58:37

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.079	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.168	03:54
Breath Volume	2.246	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.163	03:57
Breath Volume	1.937	03:57
Air Blank	0.000	03:57
Control Test	0.080	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 10/14/2021 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JEAN BAPTISTE,ALBENS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: ALFONSOMALAVE

DATE/TIME OF TEST: 09/17/2021 06:08:16

Diagnostics Check	OK	06:01
Air Blank	0.000	06:01
Control Test	0.080	06:02
Air Blank	0.000	06:02
Subject Sample # 1	0.202	06:03
Breath Volume	1.566	06:03
Air Blank	0.000	06:04
Air Blank	0.000	06:06
Subject Sample # 2	0.197	06:06
Breath Volume	1.488	06:06
Air Blank	0.000	06:07
Control Test	0.079	06:07
Air Blank	0.000	06:08
Diagnostics Check	OK	06:08

Data Download Date/Time: 10/14/2021 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NESBITT,STEPHEN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: DEVANE

DATE/TIME OF TEST: 09/18/2021 22:44:29

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.080	22:39
Air Blank	0.000	22:39
Subject Sample # 1	0.189	22:40
Breath Volume	1.312	22:40
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.197	22:42
Breath Volume	1.917	22:42
Air Blank	0.000	22:43
Control Test	0.079	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 10/14/2021 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RUIZ,ANDRES
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: FAULKNER

DATE/TIME OF TEST: 09/19/2021 23:31:56

Diagnostics Check	OK	23:29	
Air Blank	0.000	23:29	
Control Test	0.080	23:29	
Air Blank	0.000	23:30	
Subject Sample	# 1	23:30	Subject Test Refused
Breath Volume	0.000	23:30	
Air Blank	0.000	23:31	
Control Test	0.080	23:31	
Air Blank	0.000	23:31	
Diagnostics Check	OK	23:31	

Data Download Date/Time: 10/14/2021 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CRUZ,DAVID
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: RIGGS

DATE/TIME OF TEST: 09/20/2021 17:37:39

Diagnostics Check	OK	17:31
Air Blank	0.000	17:32
Control Test	0.079	17:32
Air Blank	0.000	17:32
Subject Sample # 1	0.231	17:33
Breath Volume	1.144	17:33
Air Blank	0.000	17:33
Air Blank	0.000	17:35
Subject Sample # 2	0.232	17:36
Breath Volume	2.796	17:36
Air Blank	0.000	17:36
Control Test	0.079	17:37
Air Blank	0.000	17:37
Diagnostics Check	OK	17:37

Data Download Date/Time: 10/14/2021 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VENTURA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: FREDERICK

DATE/TIME OF TEST: 09/20/2021 20:11:35

Diagnostics Check	OK	20:05
Air Blank	0.000	20:06
Control Test	0.080	20:06
Air Blank	0.000	20:06
Subject Sample # 1	0.000	20:07
Breath Volume	4.257	20:07
Air Blank	0.000	20:07
Air Blank	0.000	20:09
Subject Sample # 2	0.000	20:10
Breath Volume	4.765	20:10
Air Blank	0.000	20:10
Control Test	0.080	20:10
Air Blank	0.000	20:11
Diagnostics Check	OK	20:11

Data Download Date/Time: 10/14/2021 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WISECUP,AUSTIN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: WELDON

DATE/TIME OF TEST: 09/20/2021 21:36:31

Diagnostics Check	OK	21:30
Air Blank	0.000	21:30
Control Test	0.081	21:30
Air Blank	0.000	21:31
Subject Sample # 1	0.117	21:32
Breath Volume	2.425	21:32
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.115	21:35
Breath Volume	2.718	21:35
Air Blank	0.000	21:35
Control Test	0.080	21:35
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 10/14/2021 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MONCAYO, MASON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: MELEAR

DATE/TIME OF TEST: 09/22/2021 01:02:25

Diagnostics Check	OK	00:52
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:54
Subject Sample # 1	0.229	00:54
Breath Volume	3.328	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.254	00:57
Breath Volume	3.988	00:57
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 3	0.234	01:00
Breath Volume	2.550	01:00
Air Blank	0.000	01:01
Control Test	0.079	01:01
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 10/14/2021 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SLAFER,GAGE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: ANSBRO

DATE/TIME OF TEST: 09/22/2021 01:37:10

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.201	01:32
Breath Volume	1.281	01:32
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.195	01:35
Breath Volume	1.644	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 10/14/2021 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOYE, JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: GARCIAORTIZ

DATE/TIME OF TEST: 09/22/2021 02:25:30

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.079	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.169	02:21
Breath Volume	3.730	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.163	02:23
Breath Volume	1.742	02:23
Air Blank	0.000	02:24
Control Test	0.080	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 10/14/2021 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS, GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 09/22/2021 22:58:36

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.079	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.215	22:53
Breath Volume	2.062	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.223	22:57
Breath Volume	1.937	22:57
Air Blank	0.000	22:57
Control Test	0.077	22:58
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 10/14/2021 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JACKSON,ALLEN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: SEYMOUR

DATE/TIME OF TEST: 09/24/2021 23:39:52

Diagnostics Check	OK	23:28	
Air Blank	0.000	23:28	
Control Test	0.080	23:29	
Air Blank	0.000	23:29	
Subject Sample	# 1	23:32	No Sample Provided
Breath Volume	0.000	23:32	
Air Blank	0.000	23:33	
Air Blank	0.000	23:35	
Subject Sample	# 2	23:38	No Sample Provided
Breath Volume	0.000	23:38	
Air Blank	0.000	23:38	
Control Test	0.080	23:39	
Air Blank	0.000	23:39	
Diagnostics Check	OK	23:39	

Data Download Date/Time: 10/14/2021 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIDODOMENICO,GINO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 09/25/2021 02:18:07

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.285	02:12
Breath Volume	2.386	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.288	02:16
Breath Volume	1.195	02:16
Air Blank	0.000	02:17
Control Test	0.079	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 10/14/2021 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER,DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 09/25/2021 04:19:20

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.151	04:14
Breath Volume	1.964	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.151	04:17
Breath Volume	1.160	04:17
Air Blank	0.000	04:18
Control Test	0.079	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 10/14/2021 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PENA CARRION,MARLON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: ZAYAS

DATE/TIME OF TEST: 09/25/2021 05:40:02

Diagnostics Check	OK	05:28	
Air Blank	0.000	05:29	
Control Test	0.079	05:29	
Air Blank	0.000	05:29	
Subject Sample	# 1	05:33	No Sample Provided
Breath Volume	0.000	05:33	
Air Blank	0.000	05:33	
Air Blank	0.000	05:35	
Subject Sample	# 2	05:38	No Sample Provided
Breath Volume	0.000	05:38	
Air Blank	0.000	05:39	
Control Test	0.079	05:39	
Air Blank	0.000	05:39	
Diagnostics Check	OK	05:40	

Data Download Date/Time: 10/14/2021 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WILLS,DARREN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: KIPFINGER

DATE/TIME OF TEST: 09/26/2021 02:24:25

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.367	02:19
Breath Volume	1.929	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.375	02:22
Breath Volume	2.925	02:22
Air Blank	0.000	02:23
Control Test	0.079	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 10/14/2021 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: LAZAROHERRERA

DATE/TIME OF TEST: 09/26/2021 02:58:20

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.079	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.244	02:52
Breath Volume	1.925	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.251	02:56
Breath Volume	3.003	02:56
Air Blank	0.000	02:57
Control Test	0.079	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 10/14/2021 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GRIFFIN,CODY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: NICHOLS JR

DATE/TIME OF TEST: 09/26/2021 04:13:44

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.301	04:09
Breath Volume	2.507	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.296	04:12
Breath Volume	2.363	04:12
Air Blank	0.000	04:12
Control Test	0.078	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 10/14/2021 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,CHRIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: RIVERA TORRES

DATE/TIME OF TEST: 09/26/2021 04:25:26

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.079	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.167	04:20
Breath Volume	1.863	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.166	04:23
Breath Volume	2.191	04:23
Air Blank	0.000	04:24
Control Test	0.079	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 10/14/2021 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALFRED,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: UMANA PRUDENCIO

DATE/TIME OF TEST: 09/26/2021 05:40:03

Diagnostics Check	OK	05:28	
Air Blank	0.000	05:29	
Control Test	0.079	05:29	
Air Blank	0.000	05:29	
Subject Sample	# 1	05:33	No Sample Provided
Breath Volume	0.000	05:33	
Air Blank	0.000	05:33	
Air Blank	0.000	05:35	
Subject Sample	# 2	05:38	No Sample Provided
Breath Volume	0.000	05:38	
Air Blank	0.000	05:39	
Control Test	0.080	05:39	
Air Blank	0.000	05:39	
Diagnostics Check	OK	05:40	

Data Download Date/Time: 10/14/2021 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DELCASTILLO,BRET
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: FRANCO

DATE/TIME OF TEST: 09/26/2021 05:53:10

Diagnostics Check	OK	05:46
Air Blank	0.000	05:46
Control Test	0.080	05:47
Air Blank	0.000	05:47
Subject Sample # 1	0.120	05:48
Breath Volume	1.832	05:48
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.122	05:51
Breath Volume	1.886	05:51
Air Blank	0.000	05:52
Control Test	0.079	05:52
Air Blank	0.000	05:53
Diagnostics Check	OK	05:53

Data Download Date/Time: 10/14/2021 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SAENGCHOTE,PHANUDEJ
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: PASTAL

DATE/TIME OF TEST: 09/26/2021 07:03:55

Diagnostics Check	OK	06:54
Air Blank	0.000	06:54
Control Test	0.080	06:54
Air Blank	0.000	06:55
Subject Sample # 1	0.143	06:56
Breath Volume	2.429	06:56
Air Blank	0.000	06:56
Air Blank	0.000	06:58
Subject Sample # 2	0.122	06:59
Breath Volume	1.722	06:59
Air Blank	0.000	07:00
Air Blank	0.000	07:02
Subject Sample # 3	0.126	07:02
Breath Volume	1.472	07:02
Air Blank	0.000	07:03
Control Test	0.080	07:03
Air Blank	0.000	07:03
Diagnostics Check	OK	07:03

Data Download Date/Time: 10/14/2021 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOUGHERTY,JOSEPH
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 09/27/2021 02:52:08

Diagnostics Check	OK	02:46	
Air Blank	0.000	02:46	
Control Test	0.080	02:46	
Air Blank	0.000	02:47	
Subject Sample	# 1 0.177	02:50	Volume Not Met
Breath Volume	0.035	02:50	
Air Blank		02:50	RFI Detect
Air Blank	0.000	02:51	
Control Test	0.081	02:51	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 10/14/2021 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOUGHERTY,JOSEPH
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 09/27/2021 03:00:28

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.079	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.189	02:56
Breath Volume	1.816	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.194	02:58
Breath Volume	1.601	02:58
Air Blank	0.000	02:59
Control Test	0.078	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 10/14/2021 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DELCASTILLO, BRET
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:17

SUBJECT'S LAST NAME: RIVERADIAZ

DATE/TIME OF TEST: 09/27/2021 08:57:58

Diagnostics Check	OK	08:46	
Air Blank	0.000	08:46	
Control Test	0.081	08:47	
Air Blank	0.000	08:47	
Subject Sample	# 1 0.181	08:50	Volume Not Met
Breath Volume	0.476	08:50	
Air Blank	0.000	08:51	
Air Blank	0.000	08:53	
Subject Sample	# 2 0.201	08:56	Volume Not Met
Breath Volume	0.015	08:56	
Air Blank	0.000	08:57	
Control Test	0.080	08:57	
Air Blank	0.000	08:57	
Diagnostics Check	OK	08:57	

Data Download Date/Time: 10/14/2021 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DELCASTILLO, BRET
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:17

SUBJECT'S LAST NAME: RIVERADIAZ

DATE/TIME OF TEST: 09/27/2021 09:07:57

Diagnostics Check	OK	08:59	
Air Blank	0.000	08:59	
Control Test	0.080	09:00	
Air Blank	0.000	09:00	
Subject Sample	# 1 0.222	09:03	Volume Not Met
Breath Volume	0.082	09:03	
Air Blank	0.000	09:04	
Air Blank	0.000	09:06	
Subject Sample	# 2	09:06	Subject Test Refused
Breath Volume	0.000	09:06	
Air Blank	0.000	09:07	
Control Test	0.078	09:07	
Air Blank	0.000	09:07	
Diagnostics Check	OK	09:07	

Data Download Date/Time: 10/14/2021 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VALLE,OMAR
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: JIMENEZMIRANDA

DATE/TIME OF TEST: 09/28/2021 00:47:43

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.079	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.146	00:43
Breath Volume	2.664	00:43
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.141	00:46
Breath Volume	2.152	00:46
Air Blank	0.000	00:46
Control Test	0.080	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 10/14/2021 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ, MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: MATEO LOPEZ

DATE/TIME OF TEST: 09/30/2021 00:44:07

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.133	00:38
Breath Volume	1.597	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:40
Subject Sample # 2	0.134	00:42
Breath Volume	1.355	00:42
Air Blank	0.000	00:43
Control Test	0.079	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 10/14/2021 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADRON,MARCIE,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MITCHELL,CEDRIC
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:28

SUBJECT'S LAST NAME: AVERY

DATE/TIME OF TEST: 09/30/2021 07:01:42

Diagnostics Check	OK	06:55
Air Blank	0.000	06:55
Control Test	0.080	06:56
Air Blank	0.000	06:56
Subject Sample # 1	0.000	06:57
Breath Volume	1.855	06:57
Air Blank	0.000	06:57
Air Blank	0.000	06:59
Subject Sample # 2	0.000	07:00
Breath Volume	2.175	07:00
Air Blank	0.000	07:00
Control Test	0.080	07:01
Air Blank	0.000	07:01
Diagnostics Check	OK	07:01

Data Download Date/Time: 10/14/2021 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FAIRBANKS, ANTHONY
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: DOBBS

DATE/TIME OF TEST: 09/30/2021 17:45:12

Diagnostics Check	OK	17:37
Air Blank	0.000	17:37
Control Test	0.080	17:37
Air Blank	0.000	17:38
Subject Sample # 1	0.101	17:40
Breath Volume	1.535	17:40
Air Blank	0.000	17:41
Air Blank	0.000	17:42
Subject Sample # 2	0.100	17:43
Breath Volume	1.316	17:43
Air Blank	0.000	17:44
Control Test	0.078	17:44
Air Blank	0.000	17:45
Diagnostics Check	OK	17:45

Data Download Date/Time: 10/14/2021 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALL, ANDREW
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: HOLTZ

DATE/TIME OF TEST: 09/30/2021 19:52:02

Diagnostics Check	OK	19:45
Air Blank	0.000	19:46
Control Test	0.080	19:46
Air Blank	0.000	19:47
Subject Sample # 1	0.138	19:47
Breath Volume	3.671	19:47
Air Blank	0.000	19:48
Air Blank	0.000	19:49
Subject Sample # 2	0.136	19:50
Breath Volume	3.765	19:50
Air Blank	0.000	19:51
Control Test	0.079	19:51
Air Blank	0.000	19:51
Diagnostics Check	OK	19:52

Data Download Date/Time: 10/14/2021 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JACKSON, ALLEN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: BERRIOS

DATE/TIME OF TEST: 09/30/2021 22:21:16

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.080	22:15
Air Blank	0.000	22:15
Subject Sample # 1	0.164	22:16
Breath Volume	1.539	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.164	22:19
Breath Volume	1.996	22:19
Air Blank	0.000	22:20
Control Test	0.079	22:20
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 10/14/2021 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADSHAW,CHRISTOPHER,B
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: PROVENCHER,J
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: ROCHETTE

DATE/TIME OF TEST: 09/09/2021 21:07:29

Diagnostics Check	OK	21:00
Air Blank	0.000	21:01
Control Test	0.079	21:01
Air Blank	0.000	21:01
Subject Sample # 1	0.239	21:02
Breath Volume	2.156	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.234	21:05
Breath Volume	1.921	21:05
Air Blank	0.000	21:06
Control Test	0.077	21:06
Air Blank	0.000	21:07
Diagnostics Check	OK	21:07

Data Download Date/Time: 09/18/2021 22:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: RIS,D
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: ESTUDILLO

DATE/TIME OF TEST: 09/23/2021 20:37:41

Diagnostics Check	OK	20:31
Air Blank	0.000	20:32
Control Test	0.080	20:32
Air Blank	0.000	20:32
Subject Sample # 1	0.145	20:33
Breath Volume	2.644	20:33
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample # 2	0.145	20:36
Breath Volume	2.804	20:36
Air Blank	0.000	20:36
Control Test	0.078	20:37
Air Blank	0.000	20:37
Diagnostics Check	OK	20:37

Data Download Date/Time: 10/13/2021 18:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PICHECO,JOHN,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: WHITFIELD,CIARA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 09/30/2021 03:50:40

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.079	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.118	03:46
Breath Volume	4.253	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.119	03:49
Breath Volume	4.242	03:49
Air Blank	0.000	03:49
Control Test	0.079	03:50
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 10/13/2021 18:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1175968
BREATH TEST OPERATOR: CLARK,JAMES,P	DRY GAS EXPIRY DATE: 07/22/2022
BREATH TEST OPERATOR AGENCY: Rockledge PD	VIOLATION CODE: DUI
ARRESTING OFFICER: COOK,MITCHELL	ARREST TIME: 14:59
ARRESTING OFFICER AGENCY: Rockledge PD	GENDER: Male
LAST AGENCY INSPECTION: 08/18/2021	OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: DRAGO	DATE/TIME OF TEST: 09/02/2021 15:35:15
----------------------------	--

Diagnostics Check	OK	15:28
Air Blank	0.000	15:29
Control Test	0.079	15:29
Air Blank	0.000	15:30
Subject Sample # 1	0.304	15:30
Breath Volume	1.609	15:30
Air Blank	0.000	15:31
Air Blank	0.000	15:33
Subject Sample # 2	0.306	15:33
Breath Volume	1.464	15:33
Air Blank	0.000	15:34
Control Test	0.079	15:34
Air Blank	0.000	15:35
Diagnostics Check	OK	15:35

Data Download Date/Time: 10/05/2021 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1175968
BREATH TEST OPERATOR: CLARK,JAMES,P	DRY GAS EXPIRY DATE: 07/22/2022
BREATH TEST OPERATOR AGENCY: Rockledge PD	VIOLATION CODE: DUI
ARRESTING OFFICER: COOK,MITCHELL	ARREST TIME: 17:35
ARRESTING OFFICER AGENCY: Rockledge PD	GENDER: Female
LAST AGENCY INSPECTION: 08/18/2021	OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: STEVENS DATE/TIME OF TEST: 09/06/2021 18:07:02

Diagnostics Check	OK	17:59
Air Blank	0.000	17:59
Control Test	0.079	18:00
Air Blank	0.000	18:00
Subject Sample # 1	0.181	18:02
Breath Volume	1.191	18:02
Air Blank	0.000	18:02
Air Blank	0.000	18:04
Subject Sample # 2	0.192	18:05
Breath Volume	1.617	18:05
Air Blank	0.000	18:06
Control Test	0.077	18:06
Air Blank	0.000	18:06
Diagnostics Check	OK	18:07

Data Download Date/Time: 10/05/2021 17:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,JAMES,P
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: CLARK,JAMES
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/22/2022
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 09/25/2021 20:01:02

Diagnostics Check	OK	19:58	
Air Blank	0.000	19:58	
Control Test	0.079	19:58	
Air Blank	0.000	19:59	
Subject Sample	# 1	19:59	Subject Test Refused
Breath Volume	0.000	19:59	
Air Blank	0.000	20:00	
Control Test	0.079	20:00	
Air Blank	0.000	20:00	
Diagnostics Check	OK	20:01	

Data Download Date/Time: 10/05/2021 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBROOK,NICHOLAS,B
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: LORAIN,CHARLES
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/22/2022
VIOLATION CODE: DUI
ARREST TIME: 15:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 09/28/2021 16:01:14

Diagnostics Check	OK	15:55
Air Blank	0.000	15:55
Control Test	0.079	15:55
Air Blank	0.000	15:56
Subject Sample # 1	0.249	15:56
Breath Volume	1.660	15:56
Air Blank	0.000	15:57
Air Blank	0.000	15:59
Subject Sample # 2	0.250	15:59
Breath Volume	1.910	15:59
Air Blank	0.000	16:00
Control Test	0.079	16:00
Air Blank	0.000	16:01
Diagnostics Check	OK	16:01

Data Download Date/Time: 10/06/2021 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,A,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: MOFFATT,D
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 09/01/2021 00:12:03

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:58	
Control Test	0.082	23:59	
Air Blank	0.000	23:59	
Subject Sample	# 1 0.043	00:02	Volume Not Met
Breath Volume	0.078	00:02	
Air Blank	0.000	00:03	
Air Blank	0.000	00:05	
Subject Sample	# 2 0.041	00:06	
Breath Volume	3.007	00:06	
Air Blank	0.000	00:07	
Air Blank	0.000	00:09	
Subject Sample	# 3 0.034	00:10	
Breath Volume	2.019	00:10	
Air Blank	0.000	00:11	
Control Test	0.080	00:11	
Air Blank	0.000	00:11	
Diagnostics Check	OK	00:12	

Data Download Date/Time: 09/03/2021 07:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAZEWICZ,ALISON,E
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: BLAZEWICZ,ALISON
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: LAND

DATE/TIME OF TEST: 09/02/2021 23:28:58

Diagnostics Check	OK	23:22
Air Blank	0.000	23:23
Control Test	0.081	23:23
Air Blank	0.000	23:23
Subject Sample # 1	0.000	23:24
Breath Volume	1.171	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.000	23:27
Breath Volume	1.714	23:27
Air Blank	0.000	23:28
Control Test	0.080	23:28
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 09/03/2021 07:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUBRANO, OLIVIA, C
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: MOFFATT, DANIEL
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: VUONG

DATE/TIME OF TEST: 09/05/2021 22:56:14

Diagnostics Check	OK	22:49
Air Blank	0.000	22:50
Control Test	0.081	22:50
Air Blank	0.000	22:51
Subject Sample # 1	0.097	22:51
Breath Volume	2.492	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.094	22:54
Breath Volume	2.597	22:54
Air Blank	0.000	22:55
Control Test	0.081	22:55
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 10/11/2021 07:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: MURPHY,KEVIN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: HICKMAN

DATE/TIME OF TEST: 09/06/2021 22:31:51

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.080	22:26
Air Blank	0.000	22:26
Subject Sample # 1	0.059	22:27
Breath Volume	2.300	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:29
Subject Sample # 2	0.057	22:30
Breath Volume	2.031	22:30
Air Blank	0.000	22:30
Control Test	0.079	22:31
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 10/11/2021 14:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: VAZQUEZ,ANDREW
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: MUHLENBERG

DATE/TIME OF TEST: 09/07/2021 00:23:23

Diagnostics Check	OK	00:17
Air Blank	0.000	00:17
Control Test	0.081	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.000	00:19
Breath Volume	1.625	00:19
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.000	00:22
Breath Volume	1.527	00:22
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 10/11/2021 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,A,M
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: MOFFATT,D
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 09/11/2021 23:28:35

Diagnostics Check	OK	23:22
Air Blank	0.000	23:22
Control Test	0.082	23:23
Air Blank	0.000	23:23
Subject Sample # 1	0.148	23:24
Breath Volume	2.574	23:24
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.145	23:26
Breath Volume	2.402	23:26
Air Blank	0.000	23:27
Control Test	0.081	23:27
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 10/11/2021 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: WALKER,JAMES
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HUGHART

DATE/TIME OF TEST: 09/12/2021 04:17:46

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.081	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.119	04:13
Breath Volume	1.140	04:13
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.114	04:16
Breath Volume	1.648	04:16
Air Blank	0.000	04:16
Control Test	0.081	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 10/11/2021 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: VAZQUEZ,ANDREW
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: MALDANER

DATE/TIME OF TEST: 09/16/2021 22:27:57

Diagnostics Check	OK	22:20
Air Blank	0.000	22:21
Control Test	0.081	22:21
Air Blank	0.000	22:22
Subject Sample # 1	0.184	22:23
Breath Volume	1.312	22:23
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.187	22:26
Breath Volume	1.457	22:26
Air Blank	0.000	22:27
Control Test	0.080	22:27
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 10/11/2021 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,A,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: COLVIN,C
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: MULLINGS

DATE/TIME OF TEST: 09/20/2021 01:45:28

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.081	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.102	01:41
Breath Volume	2.371	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.100	01:43
Breath Volume	1.820	01:43
Air Blank	0.000	01:44
Control Test	0.081	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 10/11/2021 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: VAZQUEZ,ANDREW
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: REDDING

DATE/TIME OF TEST: 09/20/2021 22:57:58

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.082	22:52
Air Blank	0.000	22:52
Subject Sample # 1	0.067	22:53
Breath Volume	1.296	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:55
Subject Sample # 2	0.076	22:56
Breath Volume	1.261	22:56
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 10/11/2021 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUBRANO, OLIVIA, C
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: SANTANA, OSCAR
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: SYLVESTER

DATE/TIME OF TEST: 09/23/2021 02:55:23

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.082	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.000	02:51
Breath Volume	3.175	02:51
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.000	02:54
Breath Volume	2.386	02:54
Air Blank	0.000	02:54
Control Test	0.081	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 10/11/2021 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,A,M
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: PEARSON,K
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 09/26/2021 04:25:39

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.081	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.034	04:21
Breath Volume	1.378	04:21
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.031	04:24
Breath Volume	1.472	04:24
Air Blank	0.000	04:24
Control Test	0.080	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 10/11/2021 14:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: VAZQUEZ,ANDREW
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 09/29/2021 20:26:45

Diagnostics Check	OK	20:20
Air Blank	0.000	20:21
Control Test	0.082	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.000	20:22
Breath Volume	1.667	20:22
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.000	20:25
Breath Volume	1.433	20:25
Air Blank	0.000	20:25
Control Test	0.080	20:26
Air Blank	0.000	20:26
Diagnostics Check	OK	20:26

Data Download Date/Time: 10/11/2021 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: MURPHY,KEVIN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: MILES

DATE/TIME OF TEST: 09/30/2021 00:45:36

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.081	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.000	00:41
Breath Volume	1.660	00:41
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.000	00:44
Breath Volume	1.296	00:44
Air Blank	0.000	00:44
Control Test	0.081	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 10/11/2021 14:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,JEFFERY,L
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: BLAIR,JEFFERY
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 14:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:29

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 09/30/2021 15:01:28

Diagnostics Check	OK	14:53
Air Blank	0.000	14:54
Control Test	0.081	14:54
Air Blank	0.000	14:54
Subject Sample # 1	0.206	14:55
Breath Volume	1.265	14:55
Air Blank	0.000	14:56
Air Blank	0.000	14:58
Subject Sample # 2	0.199	14:59
Breath Volume	1.925	14:59
Air Blank	0.000	15:00
Control Test	0.081	15:00
Air Blank	0.000	15:01
Diagnostics Check	OK	15:01

Data Download Date/Time: 10/11/2021 14:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: VAZQUEZ,ANDREW
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 09/30/2021 20:34:11

Diagnostics Check	OK	20:24
Air Blank	0.000	20:25
Control Test	0.081	20:25
Air Blank	0.000	20:26
Subject Sample # 1	0.012	20:27
Breath Volume	1.125	20:27
Air Blank	0.000	20:28
Air Blank	0.000	20:30
Subject Sample # 2	0.010	20:32
Breath Volume	1.445	20:32
Air Blank	0.000	20:33
Control Test	0.082	20:33
Air Blank	0.000	20:34
Diagnostics Check	OK	20:34

Data Download Date/Time: 10/11/2021 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYNOLDS,JANICE,A
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: HOUSMAN,JOHN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: AGHA

DATE/TIME OF TEST: 09/02/2021 00:51:55

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:42	
Control Test	0.080	00:42	
Air Blank	0.000	00:43	
Subject Sample	# 1 0.184	00:46	Volume Not Met
Breath Volume	0.019	00:46	
Air Blank	0.000	00:47	
Air Blank	0.000	00:49	
Subject Sample	# 2	00:50	Subject Test Refused
Breath Volume	0.058	00:50	
Air Blank	0.000	00:51	
Control Test	0.079	00:51	
Air Blank	0.000	00:51	
Diagnostics Check	OK	00:51	

Data Download Date/Time: 10/01/2021 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYNOLDS,JANICE,A
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: WRIGHT,TIMOTHY
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: VILLAR

DATE/TIME OF TEST: 09/25/2021 23:45:00

Diagnostics Check	OK	23:38
Air Blank	0.000	23:39
Control Test	0.079	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.143	23:40
Breath Volume	2.777	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.145	23:43
Breath Volume	1.996	23:43
Air Blank	0.000	23:44
Control Test	0.080	23:44
Air Blank	0.000	23:44
Diagnostics Check	OK	23:45

Data Download Date/Time: 10/01/2021 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYNOLDS,JANICE,A
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: PINSKER,JOSEPH
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: HARNEY

DATE/TIME OF TEST: 09/29/2021 21:54:02

Diagnostics Check	OK	21:46
Air Blank	0.000	21:47
Control Test	0.079	21:47
Air Blank	0.000	21:48
Subject Sample # 1	0.264	21:49
Breath Volume	1.535	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.274	21:52
Breath Volume	1.703	21:52
Air Blank	0.000	21:53
Control Test	0.079	21:53
Air Blank	0.000	21:53
Diagnostics Check	OK	21:54

Data Download Date/Time: 10/01/2021 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,ERIC,B
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: WRIGHT,TIMMOTHY
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: WEGRZYNOWICZ

DATE/TIME OF TEST: 09/30/2021 02:18:09

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.079	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.237	02:13
Breath Volume	2.632	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.237	02:16
Breath Volume	2.125	02:16
Air Blank	0.000	02:17
Control Test	0.077	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 10/01/2021 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 22:55

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 09/09/2021 23:22:35

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.079	23:16
Air Blank	0.000	23:16
Subject Sample # 1	0.000	23:17
Breath Volume	1.601	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.000	23:21
Breath Volume	1.593	23:21
Air Blank	0.000	23:21
Control Test	0.079	23:21
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 10/04/2021 18:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: SCHWARTZ,Z

ARRESTING OFFICER AGENCY: Melbourne PD

LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 09/10/2021 03:37:23

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.078	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.131	03:33
Breath Volume	3.531	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.132	03:35
Breath Volume	3.550	03:35
Air Blank	0.000	03:36
Control Test	0.078	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 10/04/2021 18:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 21:10

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: RADWAY

DATE/TIME OF TEST: 09/10/2021 21:20:02

Diagnostics Check	OK	21:14
Air Blank	0.000	21:14
Control Test	0.079	21:14
Air Blank	0.000	21:15
Subject Sample # 1	0.000	21:15
Breath Volume	1.832	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:18
Subject Sample # 2	0.000	21:18
Breath Volume	2.195	21:18
Air Blank	0.000	21:19
Control Test	0.079	21:19
Air Blank	0.000	21:19
Diagnostics Check	OK	21:20

Data Download Date/Time: 10/04/2021 18:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 03:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: GLASCO

DATE/TIME OF TEST: 09/12/2021 04:07:41

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:03
Subject Sample # 1	0.016	04:03
Breath Volume	3.031	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.015	04:06
Breath Volume	3.078	04:06
Air Blank	0.000	04:06
Control Test	0.079	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 10/04/2021 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 21:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 09/12/2021 22:11:24

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.079	22:06
Air Blank	0.000	22:06
Subject Sample # 1	0.057	22:06
Breath Volume	1.617	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.061	22:09
Breath Volume	1.503	22:09
Air Blank	0.000	22:10
Control Test	0.080	22:10
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 10/04/2021 18:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEILL,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: NEILL,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 09/19/2021 03:58:20

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.145	03:53
Breath Volume	2.796	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.143	03:56
Breath Volume	2.664	03:56
Air Blank	0.000	03:57
Control Test	0.078	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 10/04/2021 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 22:38

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/24/2021 22:49:30

Diagnostics Check	OK	22:43
Air Blank	0.000	22:44
Control Test	0.079	22:44
Air Blank	0.000	22:44
Subject Sample # 1	0.000	22:45
Breath Volume	3.636	22:45
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.000	22:48
Breath Volume	3.292	22:48
Air Blank	0.000	22:48
Control Test	0.079	22:48
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 10/04/2021 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 09/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 19:12

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: TUIITE

DATE/TIME OF TEST: 09/30/2021 19:34:56

Diagnostics Check	OK	19:20	
Air Blank	0.000	19:20	
Control Test	0.080	19:21	
Air Blank	0.000	19:21	
Subject Sample	# 1 0.197	19:22	
Breath Volume	1.171	19:22	
Air Blank	0.000	19:22	
Air Blank	0.000	19:24	
Subject Sample	# 2 0.164	19:27	Volume Not Met
Breath Volume	0.582	19:27	
Air Blank	0.000	19:28	
Air Blank	0.000	19:30	
Subject Sample	# 3 0.184	19:33	Volume Not Met
Breath Volume	0.136	19:33	
Air Blank	0.000	19:33	
Control Test	0.079	19:34	
Air Blank	0.000	19:34	
Diagnostics Check	OK	19:34	

Data Download Date/Time: 10/04/2021 18:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 09/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 19:12

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: TUIITE

DATE/TIME OF TEST: 09/30/2021 19:47:17

Diagnostics Check	OK	19:36	
Air Blank	0.000	19:36	
Control Test	0.079	19:37	
Air Blank	0.000	19:37	
Subject Sample	# 1 0.184	19:40	Volume Not Met
Breath Volume	0.093	19:40	
Air Blank	0.000	19:41	
Air Blank	0.000	19:43	
Subject Sample	# 2 0.179	19:46	Volume Not Met
Breath Volume	1.003	19:46	
Air Blank		19:46	RFI Detect
Air Blank	0.000	19:47	

Data Download Date/Time: 10/04/2021 18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GRUNAT,J
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BAUTISTA

DATE/TIME OF TEST: 09/01/2021 00:51:46

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.080	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.000	00:47
Breath Volume	2.136	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.000	00:50
Breath Volume	2.230	00:50
Air Blank	0.000	00:50
Control Test	0.080	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 09/01/2021 23:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ESPARRA,RAFAEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: PETRIE

DATE/TIME OF TEST: 09/02/2021 02:48:05

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.078	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.156	02:43
Breath Volume	2.914	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.156	02:46
Breath Volume	3.019	02:46
Air Blank	0.000	02:47
Control Test	0.079	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 10/03/2021 22:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CASSELL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: DIAS

DATE/TIME OF TEST: 09/03/2021 05:47:07

Diagnostics Check	OK	05:38
Air Blank	0.000	05:38
Control Test	0.080	05:39
Air Blank	0.000	05:39
Subject Sample # 1	0.244	05:42
Breath Volume	1.503	05:42
Air Blank	0.000	05:42
Air Blank	0.000	05:44
Subject Sample # 2	0.247	05:45
Breath Volume	1.406	05:45
Air Blank	0.000	05:46
Control Test	0.079	05:46
Air Blank	0.000	05:47
Diagnostics Check	OK	05:47

Data Download Date/Time: 10/03/2021 22:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MASSIP,LEONARDO
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: BUFORD

DATE/TIME OF TEST: 09/05/2021 01:23:11

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.078	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.123	01:18
Breath Volume	3.789	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:20
Subject Sample # 2	0.119	01:21
Breath Volume	3.457	01:21
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 10/03/2021 22:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SANCHEZ,JUBILEE
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: KENNY

DATE/TIME OF TEST: 09/05/2021 01:47:41

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.079	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.147	01:43
Breath Volume	4.144	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.150	01:46
Breath Volume	4.105	01:46
Air Blank	0.000	01:46
Control Test	0.079	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 10/03/2021 22:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: VALENTINE

DATE/TIME OF TEST: 09/05/2021 04:12:44

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.078	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.178	04:08
Breath Volume	2.058	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.182	04:11
Breath Volume	1.703	04:11
Air Blank	0.000	04:11
Control Test	0.079	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 10/03/2021 22:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: VANROMPA,VITO
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 09/05/2021 04:59:18

Diagnostics Check	OK	04:49
Air Blank	0.000	04:50
Control Test	0.080	04:50
Air Blank	0.000	04:50
Subject Sample # 1	0.091	04:52
Breath Volume	1.589	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.088	04:57
Breath Volume	1.886	04:57
Air Blank	0.000	04:58
Control Test	0.079	04:58
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 10/03/2021 22:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BURKE,ROBERT
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: PROBST

DATE/TIME OF TEST: 09/05/2021 05:10:10

Diagnostics Check	OK	05:04
Air Blank	0.000	05:04
Control Test	0.079	05:04
Air Blank	0.000	05:05
Subject Sample # 1	0.194	05:05
Breath Volume	2.402	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.188	05:08
Breath Volume	1.714	05:08
Air Blank	0.000	05:09
Control Test	0.079	05:09
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 10/03/2021 22:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENCH,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: FRENCH,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: ZAPATA

DATE/TIME OF TEST: 09/06/2021 21:42:05

Diagnostics Check	OK	21:35
Air Blank	0.000	21:35
Control Test	0.080	21:35
Air Blank	0.000	21:36
Subject Sample # 1	0.108	21:37
Breath Volume	2.140	21:37
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.106	21:40
Breath Volume	2.335	21:40
Air Blank	0.000	21:41
Control Test	0.079	21:41
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time: 10/03/2021 22:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: UCELO,BRYAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: HHONEYCUTT

DATE/TIME OF TEST: 09/10/2021 22:50:28

Diagnostics Check	OK	22:43
Air Blank	0.000	22:44
Control Test	0.078	22:44
Air Blank	0.000	22:44
Subject Sample # 1	0.099	22:45
Breath Volume	2.480	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.087	22:48
Breath Volume	1.335	22:48
Air Blank	0.000	22:49
Control Test	0.079	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 10/03/2021 22:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BURK,MATTHEW
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: SARVARI

DATE/TIME OF TEST: 09/11/2021 03:10:19

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:07	
Control Test	0.079	03:07	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.079	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 10/03/2021 22:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BURK,MATTHEW
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: SARVARI

DATE/TIME OF TEST: 09/11/2021 03:23:44

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:12	
Control Test	0.079	03:12	
Air Blank	0.000	03:13	
Subject Sample	# 1 0.292	03:13	
Breath Volume	1.195	03:13	
Air Blank	0.000	03:14	
Air Blank	0.000	03:16	
Subject Sample	# 2 0.244	03:19	Volume Not Met
Breath Volume	0.722	03:19	
Air Blank	0.000	03:19	
Air Blank	0.000	03:21	
Subject Sample	# 3	03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:22	
Control Test	0.077	03:23	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 10/03/2021 22:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: REHDER,PATRICK
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: SHPIGELMAN

DATE/TIME OF TEST: 09/12/2021 03:15:00

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.078	03:09
Air Blank	0.000	03:09
Subject Sample # 1	0.282	03:10
Breath Volume	3.929	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:12
Subject Sample # 2	0.276	03:13
Breath Volume	3.566	03:13
Air Blank	0.000	03:14
Control Test	0.079	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:15

Data Download Date/Time: 10/03/2021 22:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KEE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: CAMBARA

DATE/TIME OF TEST: 09/12/2021 04:18:43

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.080	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.168	04:14
Breath Volume	1.601	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.161	04:17
Breath Volume	1.546	04:17
Air Blank	0.000	04:17
Control Test	0.079	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 10/03/2021 22:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SAAVEDRA,JUAN
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: LUMPKIN

DATE/TIME OF TEST: 09/12/2021 23:45:58

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.079	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.134	23:41
Breath Volume	2.640	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:43
Subject Sample # 2	0.133	23:44
Breath Volume	2.687	23:44
Air Blank	0.000	23:45
Control Test	0.079	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 10/03/2021 22:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SANCHEZ,J
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: DISTEFANO

DATE/TIME OF TEST: 09/13/2021 04:26:43

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.079	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.227	04:22
Breath Volume	2.421	04:22
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.226	04:25
Breath Volume	2.136	04:25
Air Blank	0.000	04:25
Control Test	0.079	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 10/03/2021 22:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PONDER,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: PONDER,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: DEMPSEY

DATE/TIME OF TEST: 09/15/2021 20:56:43

Diagnostics Check	OK	20:48
Air Blank	0.000	20:49
Control Test	0.080	20:49
Air Blank	0.000	20:49
Subject Sample # 1	0.107	20:52
Breath Volume	1.382	20:52
Air Blank	0.000	20:52
Air Blank	0.000	20:54
Subject Sample # 2	0.103	20:55
Breath Volume	1.156	20:55
Air Blank	0.000	20:55
Control Test	0.079	20:56
Air Blank	0.000	20:56
Diagnostics Check	OK	20:56

Data Download Date/Time: 10/03/2021 22:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TEEHAN,MATTHEW
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: ROSSON

DATE/TIME OF TEST: 09/17/2021 01:21:12

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.079	01:15
Air Blank	0.000	01:15
Subject Sample # 1	0.226	01:16
Breath Volume	2.300	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.229	01:19
Breath Volume	2.410	01:19
Air Blank	0.000	01:20
Control Test	0.079	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 10/03/2021 22:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BRYAN,ADAM
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SICARD

DATE/TIME OF TEST: 09/17/2021 03:14:07

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:01	
Control Test	0.078	03:01	
Air Blank	0.000	03:02	
Subject Sample	# 1 0.163	03:04	
Breath Volume	1.308	03:04	
Air Blank	0.000	03:05	
Air Blank	0.000	03:06	
Subject Sample	# 2 0.129	03:10	Volume Not Met
Breath Volume	0.457	03:10	
Air Blank	0.000	03:10	
Air Blank	0.000	03:12	
Subject Sample	# 3	03:12	Subject Test Refused
Breath Volume	0.000	03:12	
Air Blank	0.000	03:13	
Control Test	0.078	03:13	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 10/03/2021 22:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,JASON,B
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: VANROMPA,VITO
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: MCCORD

DATE/TIME OF TEST: 09/19/2021 06:10:30

Diagnostics Check	OK	06:04
Air Blank	0.000	06:04
Control Test	0.078	06:05
Air Blank	0.000	06:05
Subject Sample # 1	0.252	06:06
Breath Volume	2.085	06:06
Air Blank	0.000	06:06
Air Blank	0.000	06:08
Subject Sample # 2	0.248	06:08
Breath Volume	2.082	06:08
Air Blank	0.000	06:09
Control Test	0.078	06:09
Air Blank	0.000	06:10
Diagnostics Check	OK	06:10

Data Download Date/Time: 10/03/2021 22:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 09/20/2021 03:42:42

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.221	03:38
Breath Volume	2.144	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.211	03:41
Breath Volume	1.175	03:41
Air Blank	0.000	03:41
Control Test	0.079	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 10/03/2021 22:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SAAVEDRA,JUAN
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 09/21/2021 01:00:47

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.079	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.213	00:56
Breath Volume	1.894	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:58
Subject Sample # 2	0.197	00:59
Breath Volume	1.351	00:59
Air Blank	0.000	00:59
Control Test	0.077	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 10/03/2021 22:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SMITH,MEGAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: ACKER

DATE/TIME OF TEST: 09/21/2021 21:49:08

Diagnostics Check	OK	21:43	
Air Blank	0.000	21:43	
Control Test	0.079	21:44	
Air Blank	0.000	21:44	
Subject Sample	# 1 0.280	21:45	
Breath Volume	1.785	21:45	
Air Blank	0.000	21:45	
Air Blank	0.000	21:47	
Subject Sample	# 2	21:47	Subject Test Refused
Breath Volume	0.000	21:47	
Air Blank	0.000	21:48	
Control Test	0.079	21:48	
Air Blank	0.000	21:49	
Diagnostics Check	OK	21:49	

Data Download Date/Time: 10/03/2021 22:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SMITH,MEGAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: ACKER

DATE/TIME OF TEST: 09/21/2021 22:12:28

Diagnostics Check	OK	22:06
Air Blank	0.000	22:06
Control Test	0.079	22:07
Air Blank	0.000	22:07
Subject Sample # 1	0.257	22:08
Breath Volume	1.835	22:08
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.263	22:10
Breath Volume	1.945	22:10
Air Blank	0.000	22:11
Control Test	0.078	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 10/03/2021 22:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARMAN,RICHARD,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HARMAN,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: HARDTER

DATE/TIME OF TEST: 09/24/2021 19:59:32

Diagnostics Check	OK	19:51
Air Blank	0.000	19:51
Control Test	0.079	19:52
Air Blank	0.000	19:52
Subject Sample # 1	0.147	19:53
Breath Volume	2.460	19:53
Air Blank	0.000	19:54
Air Blank	0.000	19:56
Subject Sample # 2	0.148	19:57
Breath Volume	3.121	19:57
Air Blank	0.000	19:58
Control Test	0.080	19:58
Air Blank	0.000	19:59
Diagnostics Check	OK	19:59

Data Download Date/Time: 10/03/2021 22:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HIDEROA,KEVIN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: PICERNE

DATE/TIME OF TEST: 09/25/2021 03:18:07

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:12	
Control Test	0.078	03:12	
Air Blank	0.000	03:13	
Subject Sample # 1	0.127	03:15	
Breath Volume	2.082	03:15	
Air Blank	0.000	03:15	
Air Blank		03:17	Ambient Fail
Air Blank	0.000	03:18	

Data Download Date/Time: 10/03/2021 22:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HIDEROA,KEVIN
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: PICERNE

DATE/TIME OF TEST: 09/25/2021 03:33:38

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:22	
Control Test	0.078	03:23	
Air Blank	0.000	03:23	
Subject Sample	# 1 0.104	03:26	Volume Not Met
Breath Volume	0.300	03:26	
Air Blank	0.000	03:27	
Air Blank	0.000	03:29	
Subject Sample	# 2	03:32	No Sample Provided
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Control Test	0.078	03:33	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 10/03/2021 22:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BENITEZ,STEVE
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: DAVID

DATE/TIME OF TEST: 09/26/2021 01:49:43

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.154	01:45
Breath Volume	3.933	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.154	01:48
Breath Volume	2.945	01:48
Air Blank	0.000	01:48
Control Test	0.079	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 10/03/2021 22:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SMITH,LOGAN
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: HEISE

DATE/TIME OF TEST: 09/26/2021 03:13:20

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.079	03:05
Air Blank	0.000	03:05
Subject Sample # 1	0.229	03:06
Breath Volume	1.906	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.253	03:08
Breath Volume	2.632	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 3	0.226	03:11
Breath Volume	1.546	03:11
Air Blank	0.000	03:12
Control Test	0.078	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 10/03/2021 22:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: UCELO,BRYAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: EBRIGHT

DATE/TIME OF TEST: 09/26/2021 04:47:44

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.079	04:42
Air Blank	0.000	04:42
Subject Sample # 1	0.133	04:43
Breath Volume	3.527	04:43
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.133	04:46
Breath Volume	3.121	04:46
Air Blank	0.000	04:46
Control Test	0.079	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 10/03/2021 22:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GUSTAFSEN,ZACHARY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 09/26/2021 05:12:06

Diagnostics Check	OK	05:06
Air Blank	0.000	05:06
Control Test	0.079	05:06
Air Blank	0.000	05:07
Subject Sample # 1	0.000	05:07
Breath Volume	4.019	05:07
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 2	0.000	05:10
Breath Volume	4.410	05:10
Air Blank	0.000	05:11
Control Test	0.078	05:11
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 10/03/2021 22:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PFANNERSTILL,JACOB
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 09/27/2021 00:17:01

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.080	00:11
Air Blank	0.000	00:11
Subject Sample # 1	0.158	00:12
Breath Volume	1.355	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.153	00:15
Breath Volume	1.250	00:15
Air Blank	0.000	00:16
Control Test	0.080	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:17

Data Download Date/Time: 10/03/2021 22:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: UCELO,BRYAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: MORENO CAMPOS

DATE/TIME OF TEST: 09/27/2021 04:35:21

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.080	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.203	04:27
Breath Volume	2.625	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:29
Subject Sample # 2	0.178	04:30
Breath Volume	1.859	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 3	0.191	04:33
Breath Volume	1.800	04:33
Air Blank	0.000	04:34
Control Test	0.079	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 10/03/2021 22:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODARD,DALLESSANDRO,D
BREATH TEST OPERATOR AGENCY: Winter Springs PD
ARRESTING OFFICER: ACENEDO-
PERALTA,EMMANUEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 08:30
GENDER: Male

OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: SCHICKER

DATE/TIME OF TEST: 09/28/2021 09:50:11

Diagnostics Check	OK	09:42	
Air Blank	0.000	09:43	
Control Test	0.078	09:43	
Air Blank	0.000	09:43	
Subject Sample	# 1 0.154	09:45	
Breath Volume	3.582	09:45	
Air Blank	0.000	09:46	
Air Blank	0.000	09:47	
Subject Sample	# 2 0.150	09:48	
Breath Volume	3.093	09:48	
Air Blank	0.000	09:49	
Control Test	0.079	09:49	
Air Blank		09:49	RFI Detect
Air Blank	0.000	09:50	

Data Download Date/Time: 10/03/2021 22:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENCH,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: COOK,J
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 08/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/29/2021 22:15:52

Diagnostics Check	OK	22:08
Air Blank	0.000	22:09
Control Test	0.079	22:09
Air Blank	0.000	22:09
Subject Sample # 1	0.201	22:11
Breath Volume	1.625	22:11
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.196	22:14
Breath Volume	1.929	22:14
Air Blank	0.000	22:14
Control Test	0.080	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 10/03/2021 22:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,ZOE,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: JULIANO

DATE/TIME OF TEST: 09/01/2021 00:45:42

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.080	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.124	00:41
Breath Volume	1.898	00:41
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.141	00:43
Breath Volume	2.312	00:43
Air Blank	0.000	00:44
Control Test	0.080	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 09/23/2021 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROBB,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: CURRINGTON

DATE/TIME OF TEST: 09/01/2021 05:25:10

Diagnostics Check	OK	05:19	
Air Blank	0.000	05:20	
Control Test	0.080	05:20	
Air Blank	0.000	05:21	
Subject Sample	# 1	05:23	Subject Test Refused
Breath Volume	0.000	05:23	
Air Blank	0.000	05:24	
Control Test	0.079	05:24	
Air Blank	0.000	05:25	
Diagnostics Check	OK	05:25	

Data Download Date/Time: 09/23/2021 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 09/01/2021 22:06:08

Diagnostics Check	OK	21:59
Air Blank	0.000	21:59
Control Test	0.079	22:00
Air Blank	0.000	22:00
Subject Sample # 1	0.237	22:01
Breath Volume	1.511	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:03
Subject Sample # 2	0.237	22:04
Breath Volume	1.941	22:04
Air Blank	0.000	22:05
Control Test	0.078	22:05
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 09/23/2021 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,KENNETH,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STONE,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: DUDECK

DATE/TIME OF TEST: 09/03/2021 02:57:00

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:54	
Control Test	0.079	02:54	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:55	Subject Test Refused
Breath Volume	0.000	02:55	
Air Blank	0.000	02:56	
Control Test	0.079	02:56	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:57	

Data Download Date/Time: 09/23/2021 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: MILES

DATE/TIME OF TEST: 09/04/2021 03:27:32

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.079	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.202	03:22
Breath Volume	1.167	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.197	03:25
Breath Volume	1.695	03:25
Air Blank	0.000	03:26
Control Test	0.078	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 09/23/2021 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: GERMAIN

DATE/TIME OF TEST: 09/04/2021 19:54:20

Diagnostics Check	OK	19:49	
Air Blank	0.000	19:50	
Control Test	0.080	19:50	
Air Blank	0.000	19:51	
Subject Sample	# 1	19:52	Subject Test Refused
Breath Volume	0.000	19:52	
Air Blank	0.000	19:53	
Control Test	0.080	19:53	
Air Blank	0.000	19:54	
Diagnostics Check	OK	19:54	

Data Download Date/Time: 09/23/2021 22:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RODRIGUEZ,B
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: BUI
ARREST TIME: 19:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 09/04/2021 20:52:29

Diagnostics Check	OK	20:45
Air Blank	0.000	20:46
Control Test	0.080	20:46
Air Blank	0.000	20:46
Subject Sample # 1	0.077	20:47
Breath Volume	1.261	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 2	0.078	20:50
Breath Volume	1.832	20:50
Air Blank	0.000	20:51
Control Test	0.079	20:51
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 09/23/2021 22:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,G
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: RADCLIFFE

DATE/TIME OF TEST: 09/04/2021 21:35:12

Diagnostics Check	OK	21:30	
Air Blank	0.000	21:31	
Control Test	0.079	21:31	
Air Blank	0.000	21:32	
Subject Sample	# 1	21:33	Subject Test Refused
Breath Volume	0.000	21:33	
Air Blank	0.000	21:34	
Control Test	0.078	21:34	
Air Blank	0.000	21:35	
Diagnostics Check	OK	21:35	

Data Download Date/Time: 09/23/2021 22:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEAVER,A
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: COHORST

DATE/TIME OF TEST: 09/04/2021 23:01:07

Diagnostics Check	OK	22:53
Air Blank	0.000	22:54
Control Test	0.080	22:54
Air Blank	0.000	22:55
Subject Sample # 1	0.178	22:56
Breath Volume	1.921	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:58
Subject Sample # 2	0.186	22:59
Breath Volume	1.968	22:59
Air Blank	0.000	23:00
Control Test	0.078	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 09/23/2021 22:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORO,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: FERNANDES

DATE/TIME OF TEST: 09/05/2021 00:09:31

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:01
Subject Sample # 1	0.187	00:03
Breath Volume	1.476	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.187	00:07
Breath Volume	1.140	00:07
Air Blank	0.000	00:08
Control Test	0.078	00:08
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

Data Download Date/Time: 09/23/2021 22:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HUGHES,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: DRISCOLL

DATE/TIME OF TEST: 09/05/2021 00:16:36

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:14	
Control Test	0.079	00:14	
Air Blank	0.000	00:15	
Subject Sample	# 1	00:15	Subject Test Refused
Breath Volume	0.000	00:15	
Air Blank	0.000	00:15	
Control Test	0.079	00:15	
Air Blank	0.000	00:16	
Diagnostics Check	OK	00:16	

Data Download Date/Time: 09/23/2021 22:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: MAIO

DATE/TIME OF TEST: 09/05/2021 01:24:19

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:20	
Control Test	0.079	01:20	
Air Blank	0.000	01:21	
Subject Sample	# 1	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:23	
Control Test	0.079	01:23	
Air Blank	0.000	01:24	
Diagnostics Check	OK	01:24	

Data Download Date/Time: 09/23/2021 22:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE, ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: PEROTTE-SHIM

DATE/TIME OF TEST: 09/05/2021 05:09:57

Diagnostics Check	OK	05:02
Air Blank	0.000	05:03
Control Test	0.079	05:03
Air Blank	0.000	05:03
Subject Sample # 1	0.127	05:04
Breath Volume	1.476	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.127	05:08
Breath Volume	1.343	05:08
Air Blank	0.000	05:08
Control Test	0.079	05:09
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 09/23/2021 22:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LAMBORGHINI,PETER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ALTER

DATE/TIME OF TEST: 09/06/2021 01:41:29

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.000	01:37
Breath Volume	3.367	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.000	01:40
Breath Volume	2.585	01:40
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 09/23/2021 22:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: TAURISANO

DATE/TIME OF TEST: 09/06/2021 03:24:21

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.079	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.136	03:19
Breath Volume	1.187	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.135	03:22
Breath Volume	1.148	03:22
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 09/23/2021 22:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SCHOMP,DALTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: BROOM

DATE/TIME OF TEST: 09/06/2021 21:00:22

Diagnostics Check	OK	20:53
Air Blank	0.000	20:54
Control Test	0.080	20:54
Air Blank	0.000	20:55
Subject Sample # 1	0.131	20:55
Breath Volume	2.304	20:55
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample # 2	0.129	20:58
Breath Volume	2.460	20:58
Air Blank	0.000	20:59
Control Test	0.077	20:59
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time: 09/23/2021 22:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBSON,JON,B
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUTH,TIMOTHY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 09:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:55

SUBJECT'S LAST NAME: SEARY

DATE/TIME OF TEST: 09/07/2021 10:23:45

Diagnostics Check	OK	10:17
Air Blank	0.000	10:18
Control Test	0.080	10:18
Air Blank	0.000	10:18
Subject Sample # 1	0.000	10:19
Breath Volume	3.148	10:19
Air Blank	0.000	10:19
Air Blank	0.000	10:21
Subject Sample # 2	0.000	10:22
Breath Volume	3.023	10:22
Air Blank	0.000	10:22
Control Test	0.079	10:23
Air Blank	0.000	10:23
Diagnostics Check	OK	10:23

Data Download Date/Time: 09/23/2021 22:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: NAUMANN

DATE/TIME OF TEST: 09/08/2021 20:28:57

Diagnostics Check	OK	20:26	
Air Blank	0.000	20:26	
Control Test	0.080	20:26	
Air Blank	0.000	20:27	
Subject Sample	# 1	20:27	Subject Test Refused
Breath Volume	0.000	20:27	
Air Blank	0.000	20:27	
Control Test	0.080	20:28	
Air Blank	0.000	20:28	
Diagnostics Check	OK	20:28	

Data Download Date/Time: 09/23/2021 22:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: RAVIS

DATE/TIME OF TEST: 09/08/2021 23:06:48

Diagnostics Check	OK	22:59
Air Blank	0.000	23:00
Control Test	0.079	23:00
Air Blank	0.000	23:01
Subject Sample # 1	0.000	23:02
Breath Volume	2.210	23:02
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.000	23:05
Breath Volume	2.273	23:05
Air Blank	0.000	23:05
Control Test	0.080	23:06
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 09/23/2021 22:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: CANTRELL

DATE/TIME OF TEST: 09/09/2021 03:29:41

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.000	03:25
Breath Volume	1.105	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.000	03:28
Breath Volume	2.097	03:28
Air Blank	0.000	03:28
Control Test	0.079	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 09/23/2021 22:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GROFF,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: GRASSIA

DATE/TIME OF TEST: 09/09/2021 04:00:53

Diagnostics Check	OK	03:58	
Air Blank	0.000	03:58	
Control Test	0.079	03:58	
Air Blank	0.000	03:59	
Subject Sample	# 1	03:59	Subject Test Refused
Breath Volume	0.000	03:59	
Air Blank	0.000	03:59	
Control Test	0.079	04:00	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

Data Download Date/Time: 09/23/2021 22:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SKALKO,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 09/09/2021 04:33:56

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:31	
Control Test	0.079	04:31	
Air Blank	0.000	04:32	
Subject Sample	# 1	04:32	Subject Test Refused
Breath Volume	0.000	04:32	
Air Blank	0.000	04:32	
Control Test	0.079	04:33	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 09/23/2021 22:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,K,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GILROY,SEAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 14:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:26

SUBJECT'S LAST NAME: MAGNO

DATE/TIME OF TEST: 09/10/2021 15:57:41

Diagnostics Check	OK	15:51
Air Blank	0.000	15:52
Control Test	0.079	15:52
Air Blank	0.000	15:52
Subject Sample # 1	0.000	15:53
Breath Volume	1.933	15:53
Air Blank	0.000	15:53
Air Blank	0.000	15:55
Subject Sample # 2	0.000	15:56
Breath Volume	2.375	15:56
Air Blank	0.000	15:56
Control Test	0.080	15:57
Air Blank	0.000	15:57
Diagnostics Check	OK	15:57

Data Download Date/Time: 09/23/2021 22:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/10/2021 18:13:55

Diagnostics Check	OK	18:07
Air Blank	0.000	18:08
Control Test	0.080	18:08
Air Blank	0.000	18:09
Subject Sample # 1	0.000	18:09
Breath Volume	2.496	18:09
Air Blank	0.000	18:10
Air Blank	0.000	18:12
Subject Sample # 2	0.000	18:12
Breath Volume	2.613	18:12
Air Blank	0.000	18:12
Control Test	0.080	18:13
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

Data Download Date/Time: 09/23/2021 22:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: MINER

DATE/TIME OF TEST: 09/10/2021 19:48:38

Diagnostics Check	OK	19:45	
Air Blank	0.000	19:45	
Control Test	0.079	19:46	
Air Blank	0.000	19:46	
Subject Sample	# 1	19:47	Subject Test Refused
Breath Volume	0.000	19:47	
Air Blank	0.000	19:47	
Control Test	0.079	19:48	
Air Blank	0.000	19:48	
Diagnostics Check	OK	19:48	

Data Download Date/Time: 09/23/2021 22:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEAVER,N
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: DEAL

DATE/TIME OF TEST: 09/11/2021 01:46:19

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.079	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.161	01:41
Breath Volume	1.789	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:43
Subject Sample # 2	0.159	01:44
Breath Volume	1.167	01:44
Air Blank	0.000	01:45
Control Test	0.079	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 09/23/2021 22:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: KELLY JR

DATE/TIME OF TEST: 09/11/2021 03:07:25

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.079	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.000	03:03
Breath Volume	1.218	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.000	03:06
Breath Volume	1.183	03:06
Air Blank	0.000	03:06
Control Test	0.079	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 09/23/2021 22:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: CRESPO

DATE/TIME OF TEST: 09/11/2021 04:55:42

Diagnostics Check	OK	04:52	
Air Blank	0.000	04:52	
Control Test	0.079	04:53	
Air Blank	0.000	04:53	
Subject Sample	# 1	04:54	Subject Test Refused
Breath Volume	0.000	04:54	
Air Blank	0.000	04:54	
Control Test	0.079	04:55	
Air Blank	0.000	04:55	
Diagnostics Check	OK	04:55	

Data Download Date/Time: 09/23/2021 22:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: STANTON,TRAVIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: DENITTO

DATE/TIME OF TEST: 09/11/2021 18:31:44

Diagnostics Check	OK	18:23
Air Blank	0.000	18:24
Control Test	0.079	18:24
Air Blank	0.000	18:25
Subject Sample # 1	0.277	18:26
Breath Volume	2.031	18:26
Air Blank	0.000	18:27
Air Blank	0.000	18:29
Subject Sample # 2	0.285	18:29
Breath Volume	1.890	18:29
Air Blank	0.000	18:30
Control Test	0.078	18:31
Air Blank	0.000	18:31
Diagnostics Check	OK	18:31

Data Download Date/Time: 09/23/2021 22:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONOHUE,LAWRENCE,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DONOHUE,LAWRENCE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/12/2021 03:10:42

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.079	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.159	03:06
Breath Volume	2.476	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.144	03:08
Breath Volume	1.332	03:08
Air Blank	0.000	03:09
Control Test	0.077	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 09/23/2021 22:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,ZOE,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAMBEL,ZOE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: NURGE

DATE/TIME OF TEST: 09/13/2021 22:12:05

Diagnostics Check	OK	22:05
Air Blank	0.000	22:06
Control Test	0.080	22:06
Air Blank	0.000	22:07
Subject Sample # 1	0.000	22:07
Breath Volume	1.523	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.000	22:10
Breath Volume	1.808	22:10
Air Blank	0.000	22:11
Control Test	0.079	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 09/23/2021 23:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILLIAMS,NATALIE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 09/14/2021 01:20:13

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.079	01:13
Air Blank	0.000	01:14
Subject Sample # 1	0.261	01:15
Breath Volume	1.773	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:17
Subject Sample # 2	0.262	01:18
Breath Volume	2.128	01:18
Air Blank	0.000	01:19
Control Test	0.077	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 09/23/2021 23:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IFFT,GARRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: IFFT,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: NONTHE DEZHA

DATE/TIME OF TEST: 09/14/2021 22:00:11

Diagnostics Check	OK	21:52
Air Blank	0.000	21:52
Control Test	0.079	21:52
Air Blank	0.000	21:53
Subject Sample # 1	0.112	21:54
Breath Volume	1.246	21:54
Air Blank	0.000	21:55
Air Blank	0.000	21:57
Subject Sample # 2	0.120	21:58
Breath Volume	2.453	21:58
Air Blank	0.000	21:59
Control Test	0.078	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 09/23/2021 23:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,DEANGELO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/15/2021 04:03:14

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:00	
Control Test	0.079	04:01	
Air Blank	0.000	04:01	
Subject Sample	# 1	04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:02	
Control Test	0.079	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 09/23/2021 23:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONOHUE,LAWRENCE,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DONOHUE,LAWRENCE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: YARBROUGH

DATE/TIME OF TEST: 09/16/2021 00:18:13

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:10
Subject Sample # 1	0.000	00:11
Breath Volume	1.390	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.000	00:16
Breath Volume	1.707	00:16
Air Blank	0.000	00:17
Control Test	0.080	00:17
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 09/23/2021 23:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: JAMES JR

DATE/TIME OF TEST: 09/17/2021 03:10:59

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.000	03:06
Breath Volume	1.781	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.000	03:09
Breath Volume	1.875	03:09
Air Blank	0.000	03:10
Control Test	0.079	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 09/23/2021 23:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MATTOX,DANIEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 13:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 09/17/2021 15:24:12

Diagnostics Check	OK	15:21	
Air Blank	0.000	15:21	
Control Test	0.080	15:22	
Air Blank	0.000	15:22	
Subject Sample	# 1	15:22	Subject Test Refused
Breath Volume	0.000	15:22	
Air Blank	0.000	15:23	
Control Test	0.079	15:23	
Air Blank	0.000	15:24	
Diagnostics Check	OK	15:24	

Data Download Date/Time: 09/23/2021 23:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: VOTTA

DATE/TIME OF TEST: 09/18/2021 02:43:56

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.150	02:36
Breath Volume	1.675	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.165	02:42
Breath Volume	1.828	02:42
Air Blank	0.000	02:42
Control Test	0.079	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 09/23/2021 23:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: GUIFARRO TURCIOS

DATE/TIME OF TEST: 09/18/2021 03:39:01

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.077	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.119	03:34
Breath Volume	3.152	03:34
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.120	03:37
Breath Volume	2.742	03:37
Air Blank	0.000	03:38
Control Test	0.078	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:39

Data Download Date/Time: 09/23/2021 23:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GROFF,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: TATAREVIC

DATE/TIME OF TEST: 09/18/2021 04:19:00

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.078	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.106	04:14
Breath Volume	2.125	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:16
Subject Sample # 2	0.108	04:17
Breath Volume	1.871	04:17
Air Blank	0.000	04:18
Control Test	0.079	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:19

Data Download Date/Time: 09/23/2021 23:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORT, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CRISWELL, JODY
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 07:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:28

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 09/18/2021 08:55:49

Diagnostics Check	OK	08:48
Air Blank	0.000	08:49
Control Test	0.078	08:49
Air Blank	0.000	08:50
Subject Sample # 1	0.231	08:50
Breath Volume	2.261	08:50
Air Blank	0.000	08:51
Air Blank	0.000	08:53
Subject Sample # 2	0.225	08:54
Breath Volume	1.933	08:54
Air Blank	0.000	08:54
Control Test	0.079	08:55
Air Blank	0.000	08:55
Diagnostics Check	OK	08:55

Data Download Date/Time: 09/23/2021 23:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ARLINGTON,N
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: WOLF

DATE/TIME OF TEST: 09/18/2021 21:10:04

Diagnostics Check	OK	21:07	
Air Blank	0.000	21:07	
Control Test	0.079	21:07	
Air Blank	0.000	21:08	
Subject Sample	# 1	21:08	Subject Test Refused
Breath Volume	0.000	21:08	
Air Blank	0.000	21:09	
Control Test	0.079	21:09	
Air Blank	0.000	21:09	
Diagnostics Check	OK	21:10	

Data Download Date/Time: 09/23/2021 23:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: MCINNES

DATE/TIME OF TEST: 09/19/2021 00:55:14

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:52	
Control Test	0.079	00:53	
Air Blank	0.000	00:53	
Subject Sample	# 1	00:53	Subject Test Refused
Breath Volume	0.000	00:53	
Air Blank	0.000	00:54	
Control Test	0.079	00:54	
Air Blank	0.000	00:55	
Diagnostics Check	OK	00:55	

Data Download Date/Time: 09/23/2021 23:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 09/19/2021 02:44:49

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.079	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.137	02:40
Breath Volume	1.734	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.133	02:43
Breath Volume	1.476	02:43
Air Blank	0.000	02:43
Control Test	0.078	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 09/23/2021 23:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBSEN,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 09/19/2021 04:24:03

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:21	
Control Test	0.079	04:21	
Air Blank	0.000	04:22	
Subject Sample	# 1	04:22	Subject Test Refused
Breath Volume	0.000	04:22	
Air Blank	0.000	04:23	
Control Test	0.079	04:23	
Air Blank	0.000	04:23	
Diagnostics Check	OK	04:24	

Data Download Date/Time: 09/23/2021 23:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,K,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BUSCH,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: FERRAND

DATE/TIME OF TEST: 09/21/2021 16:31:13

Diagnostics Check	OK	16:24
Air Blank	0.000	16:24
Control Test	0.080	16:25
Air Blank	0.000	16:25
Subject Sample # 1	0.338	16:26
Breath Volume	1.855	16:26
Air Blank	0.000	16:27
Air Blank	0.000	16:28
Subject Sample # 2	0.339	16:29
Breath Volume	2.000	16:29
Air Blank	0.000	16:30
Control Test	0.077	16:30
Air Blank	0.000	16:31
Diagnostics Check	OK	16:31

Data Download Date/Time: 09/23/2021 23:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,K,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COOK,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 17:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: FOWLER THOMAS

DATE/TIME OF TEST: 09/21/2021 17:55:43

Diagnostics Check	OK	17:52	
Air Blank	0.000	17:53	
Control Test	0.080	17:53	
Air Blank	0.000	17:54	
Subject Sample	# 1	17:54	Subject Test Refused
Breath Volume	0.000	17:54	
Air Blank	0.000	17:54	
Control Test	0.079	17:55	
Air Blank	0.000	17:55	
Diagnostics Check	OK	17:55	

Data Download Date/Time: 09/23/2021 23:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: BIRDSONG

DATE/TIME OF TEST: 09/22/2021 22:41:32

Diagnostics Check	OK	22:34
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:35
Subject Sample # 1	0.161	22:36
Breath Volume	1.492	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.153	22:39
Breath Volume	1.316	22:39
Air Blank	0.000	22:40
Control Test	0.078	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 09/23/2021 23:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: KUGLER

DATE/TIME OF TEST: 09/23/2021 00:25:48

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.079	00:19
Air Blank	0.000	00:19
Subject Sample # 1	0.000	00:21
Breath Volume	1.273	00:21
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.000	00:24
Breath Volume	1.574	00:24
Air Blank	0.000	00:24
Control Test	0.079	00:25
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 09/23/2021 23:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GROFF,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: GENOESE

DATE/TIME OF TEST: 09/23/2021 01:18:43

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.079	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.194	01:10
Breath Volume	1.878	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:12
Subject Sample # 2	0.171	01:13
Breath Volume	1.128	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:15
Subject Sample # 3	0.174	01:16
Breath Volume	1.703	01:16
Air Blank	0.000	01:17
Control Test	0.077	01:18
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 09/23/2021 23:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EALY,STEVEN,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FERGUSON,ROBERT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 12:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: TELLER

DATE/TIME OF TEST: 09/23/2021 14:28:46

Diagnostics Check	OK	14:21
Air Blank	0.000	14:21
Control Test	0.080	14:22
Air Blank	0.000	14:22
Subject Sample # 1	0.000	14:24
Breath Volume	2.390	14:24
Air Blank	0.000	14:24
Air Blank	0.000	14:26
Subject Sample # 2	0.000	14:27
Breath Volume	1.500	14:27
Air Blank	0.000	14:27
Control Test	0.080	14:28
Air Blank	0.000	14:28
Diagnostics Check	OK	14:28

Data Download Date/Time: 09/23/2021 23:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ARLINGTON,N
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: BOBER

DATE/TIME OF TEST: 09/23/2021 19:33:14

Diagnostics Check	OK	19:30	
Air Blank	0.000	19:30	
Control Test	0.080	19:31	
Air Blank	0.000	19:31	
Subject Sample	# 1	19:31	Subject Test Refused
Breath Volume	0.000	19:31	
Air Blank	0.000	19:32	
Control Test	0.079	19:32	
Air Blank	0.000	19:33	
Diagnostics Check	OK	19:33	

Data Download Date/Time: 09/23/2021 23:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DOUGHERTY,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: KESSLER

DATE/TIME OF TEST: 09/23/2021 20:48:25

Diagnostics Check	OK	20:45	
Air Blank	0.000	20:45	
Control Test	0.079	20:46	
Air Blank	0.000	20:46	
Subject Sample	# 1	20:47	Subject Test Refused
Breath Volume	0.000	20:47	
Air Blank	0.000	20:47	
Control Test	0.080	20:47	
Air Blank	0.000	20:48	
Diagnostics Check	OK	20:48	

Data Download Date/Time: 09/23/2021 23:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PETRACCO,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 09/24/2021 22:36:29

Diagnostics Check	OK	22:30
Air Blank	0.000	22:30
Control Test	0.079	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.000	22:32
Breath Volume	1.804	22:32
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.000	22:35
Breath Volume	1.757	22:35
Air Blank	0.000	22:35
Control Test	0.079	22:35
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 10/29/2021 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONOHUE,LAWRENCE,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: IVANOVICI,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 09/25/2021 02:36:02

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.080	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.138	02:31
Breath Volume	1.195	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:33
Subject Sample # 2	0.152	02:34
Breath Volume	1.351	02:34
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:36

Data Download Date/Time: 10/29/2021 19:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,K,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SMITH,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 17:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: DURATO

DATE/TIME OF TEST: 09/25/2021 17:56:24

Diagnostics Check	OK	17:48
Air Blank	0.000	17:48
Control Test	0.080	17:49
Air Blank	0.000	17:49
Subject Sample # 1	0.132	17:51
Breath Volume	2.753	17:51
Air Blank	0.000	17:52
Air Blank	0.000	17:53
Subject Sample # 2	0.122	17:54
Breath Volume	1.714	17:54
Air Blank	0.000	17:55
Control Test	0.078	17:55
Air Blank	0.000	17:56
Diagnostics Check	OK	17:56

Data Download Date/Time: 10/29/2021 19:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHRISTIE,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 09/25/2021 20:48:58

Diagnostics Check	OK	20:40
Air Blank	0.000	20:41
Control Test	0.079	20:41
Air Blank	0.000	20:42
Subject Sample # 1	0.101	20:43
Breath Volume	1.218	20:43
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.103	20:47
Breath Volume	1.894	20:47
Air Blank	0.000	20:48
Control Test	0.079	20:48
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

Data Download Date/Time: 10/29/2021 19:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,KENNETH,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STONE,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: MCLEAN

DATE/TIME OF TEST: 09/26/2021 01:56:30

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.179	01:51
Breath Volume	3.058	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.176	01:54
Breath Volume	2.925	01:54
Air Blank	0.000	01:55
Control Test	0.078	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 10/29/2021 19:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 09/26/2021 04:19:15

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:16	
Control Test	0.079	04:16	
Air Blank	0.000	04:17	
Subject Sample	# 1	04:17	Subject Test Refused
Breath Volume	0.000	04:17	
Air Blank	0.000	04:18	
Control Test	0.079	04:18	
Air Blank	0.000	04:19	
Diagnostics Check	OK	04:19	

Data Download Date/Time: 10/29/2021 19:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IFFT,GARRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: IFFT,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: SCHUMACHER

DATE/TIME OF TEST: 09/27/2021 23:03:50

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.080	22:54
Air Blank	0.000	22:54
Subject Sample # 1	0.175	22:55
Breath Volume	1.289	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.221	22:58
Breath Volume	2.015	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 3	0.209	23:02
Breath Volume	1.625	23:02
Air Blank	0.000	23:02
Control Test	0.079	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 10/29/2021 19:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,JOHN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 09/28/2021 04:25:09

Diagnostics Check	OK	04:17
Air Blank	0.000	04:18
Control Test	0.079	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.181	04:20
Breath Volume	2.457	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:22
Subject Sample # 2	0.176	04:23
Breath Volume	1.457	04:23
Air Blank	0.000	04:24
Control Test	0.079	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 10/29/2021 19:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EALY,S,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: LEFEBVRE

DATE/TIME OF TEST: 09/28/2021 06:52:39

Diagnostics Check	OK	06:45
Air Blank	0.000	06:46
Control Test	0.079	06:46
Air Blank	0.000	06:47
Subject Sample # 1	0.000	06:48
Breath Volume	1.750	06:48
Air Blank	0.000	06:48
Air Blank	0.000	06:50
Subject Sample # 2	0.000	06:51
Breath Volume	3.101	06:51
Air Blank	0.000	06:51
Control Test	0.080	06:52
Air Blank	0.000	06:52
Diagnostics Check	OK	06:52

Data Download Date/Time: 10/29/2021 19:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EALY, STEVEN, T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AUSTIN, JOSEPH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 15:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: WASILEWSKI

DATE/TIME OF TEST: 09/28/2021 16:22:44

Diagnostics Check	OK	16:16
Air Blank	0.000	16:16
Control Test	0.079	16:17
Air Blank	0.000	16:17
Subject Sample # 1	0.174	16:18
Breath Volume	2.660	16:18
Air Blank	0.000	16:18
Air Blank	0.000	16:20
Subject Sample # 2	0.170	16:21
Breath Volume	1.996	16:21
Air Blank	0.000	16:21
Control Test	0.079	16:22
Air Blank	0.000	16:22
Diagnostics Check	OK	16:22

Data Download Date/Time: 10/29/2021 19:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JOHNSON,CHESTER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 09/28/2021 19:45:24

Diagnostics Check	OK	19:42	
Air Blank	0.000	19:42	
Control Test	0.080	19:43	
Air Blank	0.000	19:43	
Subject Sample	# 1	19:44	Subject Test Refused
Breath Volume	0.000	19:44	
Air Blank	0.000	19:44	
Control Test	0.080	19:44	
Air Blank	0.000	19:45	
Diagnostics Check	OK	19:45	

Data Download Date/Time: 10/29/2021 19:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: HERRMANN

DATE/TIME OF TEST: 09/29/2021 03:09:26

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.080	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.098	03:04
Breath Volume	1.746	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.101	03:07
Breath Volume	1.613	03:07
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 10/29/2021 19:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMIN,N
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 09/30/2021 04:57:11

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:42	
Control Test	0.079	04:43	
Air Blank	0.000	04:43	
Subject Sample	# 1 0.107	04:46	
Breath Volume	1.808	04:46	
Air Blank	0.000	04:47	
Air Blank	0.000	04:48	
Subject Sample	# 2 0.104	04:51	Volume Not Met
Breath Volume	1.011	04:51	
Air Blank	0.000	04:52	
Air Blank	0.000	04:54	
Subject Sample	# 3 0.100	04:55	
Breath Volume	1.378	04:55	
Air Blank	0.000	04:56	
Control Test	0.077	04:56	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:57	

Data Download Date/Time: 10/29/2021 19:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,K,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUNDH,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 15:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:03

SUBJECT'S LAST NAME: MILLAN

DATE/TIME OF TEST: 09/30/2021 16:50:08

Diagnostics Check	OK	16:47	
Air Blank	0.000	16:47	
Control Test	0.079	16:48	
Air Blank	0.000	16:48	
Subject Sample	# 1	16:48	Subject Test Refused
Breath Volume	0.000	16:48	
Air Blank	0.000	16:49	
Control Test	0.079	16:49	
Air Blank	0.000	16:50	
Diagnostics Check	OK	16:50	

Data Download Date/Time: 10/29/2021 19:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,KENNETH,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STONE,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: SKIDD JR

DATE/TIME OF TEST: 09/30/2021 21:27:34

Diagnostics Check	OK	21:24	
Air Blank	0.000	21:25	
Control Test	0.079	21:25	
Air Blank	0.000	21:25	
Subject Sample	# 1	21:26	Subject Test Refused
Breath Volume	0.000	21:26	
Air Blank	0.000	21:26	
Control Test	0.079	21:26	
Air Blank	0.000	21:27	
Diagnostics Check	OK	21:27	

Data Download Date/Time: 10/29/2021 19:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: REIDY,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: MACKEY

DATE/TIME OF TEST: 09/01/2021 06:04:48

Diagnostics Check	OK	05:58
Air Blank	0.000	05:59
Control Test	0.079	05:59
Air Blank	0.000	05:59
Subject Sample # 1	0.201	06:00
Breath Volume	3.167	06:00
Air Blank	0.000	06:01
Air Blank	0.000	06:02
Subject Sample # 2	0.192	06:03
Breath Volume	2.316	06:03
Air Blank	0.000	06:03
Control Test	0.079	06:04
Air Blank	0.000	06:04
Diagnostics Check	OK	06:04

Data Download Date/Time: 09/24/2021 04:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: LITTLE III

DATE/TIME OF TEST: 09/02/2021 00:34:12

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:31	
Control Test	0.079	00:32	
Air Blank	0.000	00:32	
Subject Sample	# 1	00:32	Subject Test Refused
Breath Volume	0.000	00:32	
Air Blank	0.000	00:33	
Control Test	0.080	00:33	
Air Blank	0.000	00:34	
Diagnostics Check	OK	00:34	

Data Download Date/Time: 09/24/2021 04:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: BAEZ

DATE/TIME OF TEST: 09/02/2021 23:43:29

Diagnostics Check	OK	23:33
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.147	23:37
Breath Volume	1.351	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.139	23:41
Breath Volume	1.218	23:41
Air Blank	0.000	23:42
Control Test	0.080	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 09/24/2021 04:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PEREZ-GONZALEZ

DATE/TIME OF TEST: 09/03/2021 02:07:48

Diagnostics Check	OK	02:00
Air Blank	0.000	02:01
Control Test	0.079	02:01
Air Blank	0.000	02:02
Subject Sample # 1	0.141	02:03
Breath Volume	1.527	02:03
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.142	02:06
Breath Volume	2.394	02:06
Air Blank	0.000	02:06
Control Test	0.079	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 09/24/2021 04:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMIN,NAZIR
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: ALONSO

DATE/TIME OF TEST: 09/03/2021 03:00:36

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:56	
Control Test	0.079	02:56	
Air Blank	0.000	02:57	
Subject Sample	# 1	02:59	Subject Test Refused
Breath Volume	0.000	02:59	
Air Blank	0.000	02:59	
Control Test	0.080	03:00	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 09/24/2021 04:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: FOTEU

DATE/TIME OF TEST: 09/03/2021 04:30:23

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.079	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.068	04:25
Breath Volume	1.550	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.067	04:28
Breath Volume	2.203	04:28
Air Blank	0.000	04:29
Control Test	0.079	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 09/24/2021 04:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JORDAN,W
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SAPP,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: MEYER

DATE/TIME OF TEST: 09/03/2021 20:07:51

Diagnostics Check	OK	19:58
Air Blank	0.000	19:59
Control Test	0.080	19:59
Air Blank	0.000	20:00
Subject Sample # 1	0.180	20:01
Breath Volume	1.175	20:01
Air Blank	0.000	20:02
Air Blank	0.000	20:03
Subject Sample # 2	0.182	20:06
Breath Volume	2.308	20:06
Air Blank	0.000	20:06
Control Test	0.080	20:07
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 09/24/2021 04:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 09/04/2021 05:19:13

Diagnostics Check	OK	05:12
Air Blank	0.000	05:12
Control Test	0.079	05:12
Air Blank	0.000	05:13
Subject Sample # 1	0.072	05:14
Breath Volume	1.273	05:14
Air Blank	0.000	05:15
Air Blank	0.000	05:17
Subject Sample # 2	0.075	05:17
Breath Volume	1.128	05:17
Air Blank	0.000	05:18
Control Test	0.079	05:18
Air Blank	0.000	05:19
Diagnostics Check	OK	05:19

Data Download Date/Time: 09/24/2021 04:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SWETOKOS, MICHAEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 09/04/2021 06:06:53

Diagnostics Check	OK	05:58
Air Blank	0.000	05:58
Control Test	0.079	05:59
Air Blank	0.000	05:59
Subject Sample # 1	0.159	06:02
Breath Volume	1.589	06:02
Air Blank	0.000	06:03
Air Blank	0.000	06:04
Subject Sample # 2	0.155	06:05
Breath Volume	1.195	06:05
Air Blank	0.000	06:05
Control Test	0.077	06:06
Air Blank	0.000	06:06
Diagnostics Check	OK	06:06

Data Download Date/Time: 09/24/2021 04:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: FOLEY

DATE/TIME OF TEST: 09/06/2021 18:31:09

Diagnostics Check	OK	18:24
Air Blank	0.000	18:24
Control Test	0.080	18:25
Air Blank	0.000	18:25
Subject Sample # 1	0.153	18:26
Breath Volume	1.964	18:26
Air Blank	0.000	18:27
Air Blank	0.000	18:29
Subject Sample # 2	0.168	18:29
Breath Volume	1.750	18:29
Air Blank	0.000	18:30
Control Test	0.079	18:30
Air Blank	0.000	18:31
Diagnostics Check	OK	18:31

Data Download Date/Time: 09/24/2021 04:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 09/07/2021 02:34:12

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:31	
Control Test	0.079	02:32	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:32	Subject Test Refused
Breath Volume	0.000	02:32	
Air Blank	0.000	02:33	
Control Test	0.079	02:33	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:34	

Data Download Date/Time: 09/24/2021 04:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,MATTHEW,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SIMMONS,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ZUMBRUN

DATE/TIME OF TEST: 09/07/2021 21:58:06

Diagnostics Check	OK	21:51
Air Blank	0.000	21:51
Control Test	0.080	21:51
Air Blank	0.000	21:52
Subject Sample # 1	0.209	21:53
Breath Volume	3.679	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:55
Subject Sample # 2	0.207	21:56
Breath Volume	3.679	21:56
Air Blank	0.000	21:57
Control Test	0.080	21:57
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 09/24/2021 04:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,ZOE,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAMBEL,ZOE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: SAMMONS

DATE/TIME OF TEST: 09/12/2021 03:35:21

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.079	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.151	03:30
Breath Volume	2.316	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.153	03:33
Breath Volume	2.042	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 09/24/2021 04:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PASSMAN

DATE/TIME OF TEST: 09/12/2021 04:16:41

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.079	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.180	04:12
Breath Volume	1.902	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.176	04:15
Breath Volume	1.675	04:15
Air Blank	0.000	04:15
Control Test	0.080	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 09/24/2021 04:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: PETROVIC

DATE/TIME OF TEST: 09/12/2021 04:38:56

Diagnostics Check	OK	04:35	
Air Blank	0.000	04:35	
Control Test	0.079	04:35	
Air Blank	0.000	04:36	
Subject Sample	# 1	04:37	Subject Test Refused
Breath Volume	0.000	04:37	
Air Blank	0.000	04:38	
Control Test	0.080	04:38	
Air Blank	0.000	04:38	
Diagnostics Check	OK	04:38	

Data Download Date/Time: 09/24/2021 04:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONGMIRE,BENJAMIN,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE,RONNIE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: SALUD

DATE/TIME OF TEST: 09/13/2021 22:02:58

Diagnostics Check	OK	22:00	
Air Blank	0.000	22:00	
Control Test	0.080	22:00	
Air Blank	0.000	22:01	
Subject Sample	# 1	22:01	Subject Test Refused
Breath Volume	0.000	22:01	
Air Blank	0.000	22:02	
Control Test	0.080	22:02	
Air Blank	0.000	22:02	
Diagnostics Check	OK	22:02	

Data Download Date/Time: 09/24/2021 04:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 09/17/2021 20:05:12

Diagnostics Check	OK	19:58
Air Blank	0.000	19:59
Control Test	0.081	19:59
Air Blank	0.000	20:00
Subject Sample # 1	0.195	20:00
Breath Volume	1.328	20:00
Air Blank	0.000	20:01
Air Blank	0.000	20:03
Subject Sample # 2	0.187	20:03
Breath Volume	1.187	20:03
Air Blank	0.000	20:04
Control Test	0.080	20:04
Air Blank	0.000	20:05
Diagnostics Check	OK	20:05

Data Download Date/Time: 09/24/2021 04:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 09/18/2021 02:58:54

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.192	02:52
Breath Volume	1.257	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.175	02:57
Breath Volume	1.539	02:57
Air Blank	0.000	02:57
Control Test	0.080	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 09/24/2021 04:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/18/2021 04:15:05

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:12	
Control Test	0.080	04:13	
Air Blank	0.000	04:13	
Subject Sample	# 1	04:13	Subject Test Refused
Breath Volume	0.000	04:13	
Air Blank	0.000	04:14	
Control Test	0.081	04:14	
Air Blank	0.000	04:15	
Diagnostics Check	OK	04:15	

Data Download Date/Time: 09/24/2021 04:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,CLIFF,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FISHER,CLIFF
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: LARKIN

DATE/TIME OF TEST: 09/19/2021 03:50:31

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.079	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.141	03:46
Breath Volume	1.136	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.148	03:49
Breath Volume	2.242	03:49
Air Blank	0.000	03:49
Control Test	0.078	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 09/24/2021 04:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,SCOTT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUCKETT,JOHN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 10:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:12

SUBJECT'S LAST NAME: HAIGHT

DATE/TIME OF TEST: 09/20/2021 11:40:06

Diagnostics Check	OK	11:36	
Air Blank	0.000	11:36	
Control Test	0.079	11:37	
Air Blank	0.000	11:37	
Subject Sample	# 1	11:38	Subject Test Refused
Breath Volume	0.000	11:38	
Air Blank	0.000	11:39	
Control Test	0.080	11:39	
Air Blank	0.000	11:40	
Diagnostics Check	OK	11:40	

Data Download Date/Time: 09/24/2021 04:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,COURTNEY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: MECHURA

DATE/TIME OF TEST: 09/21/2021 04:27:00

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:24	
Control Test	0.079	04:24	
Air Blank	0.000	04:25	
Subject Sample	# 1	04:25	Subject Test Refused
Breath Volume	0.000	04:25	
Air Blank	0.000	04:26	
Control Test	0.080	04:26	
Air Blank	0.000	04:26	
Diagnostics Check	OK	04:27	

Data Download Date/Time: 09/24/2021 04:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROLAND,SINGSON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 16:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:38

SUBJECT'S LAST NAME: EHLERS

DATE/TIME OF TEST: 09/26/2021 18:01:51

Diagnostics Check	OK	17:59	
Air Blank	0.000	17:59	
Control Test	0.080	17:59	
Air Blank	0.000	18:00	
Subject Sample	# 1	18:00	Subject Test Refused
Breath Volume	0.000	18:00	
Air Blank	0.000	18:00	
Control Test	0.079	18:01	
Air Blank	0.000	18:01	
Diagnostics Check	OK	18:01	

Data Download Date/Time: 10/30/2021 00:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: ZAPOTE

DATE/TIME OF TEST: 09/27/2021 00:38:14

Diagnostics Check	OK	00:35	
Air Blank	0.000	00:35	
Control Test	0.080	00:36	
Air Blank	0.000	00:36	
Subject Sample	# 1	00:36	Subject Test Refused
Breath Volume	0.000	00:36	
Air Blank	0.000	00:37	
Control Test	0.079	00:37	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:38	

Data Download Date/Time: 10/30/2021 00:47