

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKER,ALEXA,C
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: ALZATE,MELENNI
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: ALFONSO

DATE/TIME OF TEST: 09/13/2021 19:07:31

Diagnostics Check	OK	19:01
Air Blank	0.000	19:01
Control Test	0.080	19:02
Air Blank	0.000	19:02
Subject Sample # 1	0.000	19:03
Breath Volume	1.210	19:03
Air Blank	0.000	19:03
Air Blank	0.000	19:05
Subject Sample # 2	0.000	19:06
Breath Volume	2.558	19:06
Air Blank	0.000	19:06
Control Test	0.079	19:06
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

Data Download Date/Time: 09/27/2021 18:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ESPINOSA, DAYNER,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: RIVERA, MICHAEL

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 23:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: MARICHAL

DATE/TIME OF TEST: 09/01/2021 00:36:23

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.082	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.120	00:28
Breath Volume	1.457	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:30
Subject Sample # 2	0.099	00:31
Breath Volume	1.183	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:33
Subject Sample # 3	0.116	00:34
Breath Volume	3.636	00:34
Air Blank	0.000	00:35
Control Test	0.082	00:35
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 09/27/2021 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,DANELIA,J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DIAZ,DANELIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: COURT ORD
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/02/2021 23:11:22

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.082	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.000	23:07
Breath Volume	1.777	23:07
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.000	23:09
Breath Volume	1.714	23:09
Air Blank	0.000	23:10
Control Test	0.081	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 09/27/2021 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRERA,DONNAVEN,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,CAMILA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 09/04/2021 02:55:58

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:53	
Control Test	0.082	02:53	
Air Blank	0.000	02:54	
Subject Sample	# 1	02:54	Subject Test Refused
Breath Volume	0.000	02:54	
Air Blank	0.000	02:55	
Control Test	0.081	02:55	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 09/27/2021 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUY,ENMANUEL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ALCE,JIMMITRY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 17:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:17

SUBJECT'S LAST NAME: ENNIS

DATE/TIME OF TEST: 09/04/2021 15:40:46

Diagnostics Check	OK	15:37	
Air Blank	0.000	15:38	
Control Test	0.082	15:38	
Air Blank	0.000	15:38	
Subject Sample	# 1	15:39	Subject Test Refused
Breath Volume	0.000	15:39	
Air Blank	0.000	15:39	
Control Test	0.082	15:40	
Air Blank	0.000	15:40	
Diagnostics Check	OK	15:40	

Data Download Date/Time: 09/27/2021 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,I,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,I
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: MACLAREN

DATE/TIME OF TEST: 09/05/2021 00:56:08

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:54	
Control Test	0.082	00:54	
Air Blank	0.000	00:55	
Subject Sample	# 1	00:55	RFI Detect
Breath Volume	0.000	00:55	
Air Blank		00:55	RFI Detect
Air Blank	0.000	00:56	

Data Download Date/Time: 09/27/2021 13:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,I,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,I
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: MACLAREN

DATE/TIME OF TEST: 09/05/2021 01:21:37

Diagnostics Check	OK	01:20	
Air Blank		01:21	RFI Detect
Air Blank	0.000	01:21	

Data Download Date/Time: 09/27/2021 13:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOSEPH,L,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,I
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: MACLAREN

DATE/TIME OF TEST: 09/05/2021 01:45:24

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:44	
Control Test	0.141	01:44	Control Outside Tolerance
Air Blank	0.000	01:45	

Data Download Date/Time: 09/27/2021 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOSEPH,L,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,I
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: MACLAREN

DATE/TIME OF TEST: 09/05/2021 02:14:04

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.081	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.127	02:09
Breath Volume	2.250	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:11
Subject Sample # 2	0.132	02:12
Breath Volume	2.085	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:14

Data Download Date/Time: 09/27/2021 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,REMY,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PAZ,VIRGINIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 06:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:32

SUBJECT'S LAST NAME: YORDI

DATE/TIME OF TEST: 09/05/2021 08:22:34

Diagnostics Check	OK	08:19	
Air Blank	0.000	08:20	
Control Test	0.081	08:20	
Air Blank	0.000	08:21	
Subject Sample	# 1	08:21	RFI Detect
Breath Volume	0.179	08:21	
Air Blank		08:21	RFI Detect
Air Blank	0.000	08:22	

Data Download Date/Time: 09/27/2021 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,REMY,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PAZ,VIRGINIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 06:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:32

SUBJECT'S LAST NAME: YORDI

DATE/TIME OF TEST: 09/05/2021 08:41:45

Diagnostics Check	OK	08:33
Air Blank	0.000	08:33
Control Test	0.081	08:33
Air Blank	0.000	08:34
Subject Sample # 1	0.103	08:37
Breath Volume	2.210	08:37
Air Blank	0.000	08:37
Air Blank	0.000	08:39
Subject Sample # 2	0.107	08:40
Breath Volume	1.789	08:40
Air Blank	0.000	08:40
Control Test	0.079	08:41
Air Blank	0.000	08:41
Diagnostics Check	OK	08:41

Data Download Date/Time: 09/27/2021 13:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD	Date or Time Changed: Yes
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1506449
BREATH TEST OPERATOR: NAVARRO TORRES,CHRISTIAN,A	DRY GAS EXPIRY DATE: 06/10/2024
BREATH TEST OPERATOR AGENCY: Miami PD	VIOLATION CODE: DUI
ARRESTING OFFICER: VILLEGAS,RICHARD	ARREST TIME: 22:49
ARRESTING OFFICER AGENCY: Miami PD	GENDER: Male
LAST AGENCY INSPECTION: 08/26/2021	OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: NAVARRO TORRES DATE/TIME OF TEST: 09/07/2021 00:11:23

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:09	
Control Test	0.081	00:09	
Air Blank	0.000	00:09	
Subject Sample	# 1	00:10	Subject Test Refused
Breath Volume	0.000	00:10	
Air Blank	0.000	00:10	
Control Test	0.081	00:10	
Air Blank	0.000	00:11	
Diagnostics Check	OK	00:11	

Data Download Date/Time: 09/27/2021 13:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILES,STEPHANIE,N
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: LORES,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: GRANERA

DATE/TIME OF TEST: 09/09/2021 04:58:44

Diagnostics Check	OK	04:55	
Air Blank	0.000	04:56	
Control Test	0.083	04:56	
Air Blank	0.000	04:57	
Subject Sample	# 1	04:57	Subject Test Refused
Breath Volume	0.000	04:57	
Air Blank	0.000	04:57	
Control Test	0.082	04:58	
Air Blank	0.000	04:58	
Diagnostics Check	OK	04:58	

Data Download Date/Time: 09/27/2021 13:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GONZALEZ,ANTONIO,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: GONZALEZ,ANTONIO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 02:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MATHEWS

DATE/TIME OF TEST: 09/10/2021 04:21:46

Diagnostics Check	OK	04:18	
Air Blank	0.000	04:19	
Control Test	0.082	04:19	
Air Blank	0.000	04:20	
Subject Sample	# 1	04:20	Subject Test Refused
Breath Volume	0.000	04:20	
Air Blank	0.000	04:20	
Control Test	0.082	04:21	
Air Blank	0.000	04:21	
Diagnostics Check	OK	04:21	

Data Download Date/Time: 09/27/2021 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GONZALEZ,CAMILA,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: RIOS,RICHARD

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: BUI

ARREST TIME: 20:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/11/2021 21:44:19

Diagnostics Check	OK	21:34	
Air Blank	0.000	21:35	
Control Test	0.082	21:35	
Air Blank	0.000	21:36	
Subject Sample	# 1	21:39	No Sample Provided
Breath Volume	0.000	21:39	
Air Blank	0.000	21:39	
Air Blank	0.000	21:41	
Subject Sample	# 2	21:42	Subject Test Refused
Breath Volume	0.000	21:42	
Air Blank	0.000	21:43	
Control Test	0.082	21:43	
Air Blank	0.000	21:44	
Diagnostics Check	OK	21:44	

Data Download Date/Time: 09/27/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LUIS,M
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ROMERO,BRANDON
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: BRYAN

DATE/TIME OF TEST: 09/12/2021 23:20:17

Diagnostics Check	OK	23:13
Air Blank	0.000	23:14
Control Test	0.083	23:14
Air Blank	0.000	23:15
Subject Sample # 1	0.214	23:15
Breath Volume	1.597	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:17
Subject Sample # 2	0.212	23:18
Breath Volume	2.250	23:18
Air Blank	0.000	23:19
Control Test	0.081	23:19
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 09/27/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORA,MANUEL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MORA,MANUEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: COURT ORD
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/16/2021 14:40:09

Diagnostics Check	OK	14:34
Air Blank	0.000	14:34
Control Test	0.081	14:34
Air Blank	0.000	14:35
Subject Sample # 1	0.000	14:35
Breath Volume	1.941	14:35
Air Blank	0.000	14:36
Air Blank	0.000	14:38
Subject Sample # 2	0.000	14:38
Breath Volume	1.699	14:38
Air Blank	0.000	14:39
Control Test	0.082	14:39
Air Blank	0.000	14:40
Diagnostics Check	OK	14:40

Data Download Date/Time: 09/27/2021 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NARDONI,J,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MERCADO,M
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: COLL

DATE/TIME OF TEST: 09/16/2021 15:38:54

Diagnostics Check	OK	15:31
Air Blank	0.000	15:31
Control Test	0.081	15:31
Air Blank	0.000	15:32
Subject Sample # 1	0.187	15:33
Breath Volume	1.578	15:33
Air Blank	0.000	15:34
Air Blank	0.000	15:35
Subject Sample # 2	0.188	15:36
Breath Volume	1.351	15:36
Air Blank	0.000	15:37
Control Test	0.080	15:38
Air Blank	0.000	15:38
Diagnostics Check	OK	15:38

Data Download Date/Time: 09/27/2021 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILES,STEPHANIE,N
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SILES,STEPHANIE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: MICHLER

DATE/TIME OF TEST: 09/18/2021 01:21:59

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:19	
Control Test	0.082	01:19	
Air Blank	0.000	01:20	
Subject Sample	# 1	01:20	Subject Test Refused
Breath Volume	0.000	01:20	
Air Blank	0.000	01:21	
Control Test	0.082	01:21	
Air Blank	0.000	01:21	
Diagnostics Check	OK	01:21	

Data Download Date/Time: 09/27/2021 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DEL COLLADO,ADRIAN,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: MARTINEZ,JORGE

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 09:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 09:38

SUBJECT'S LAST NAME: NARANJA

DATE/TIME OF TEST: 09/18/2021 10:05:07

Diagnostics Check	OK	09:59	
Air Blank	0.000	09:59	
Control Test	0.082	10:00	
Air Blank	0.000	10:00	
Subject Sample	# 1 0.287	10:03	Volume Not Met
Breath Volume	0.152	10:03	
Air Blank		10:04	Purge Fail
Air Blank	0.000	10:05	

Data Download Date/Time: 09/27/2021 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DEL COLLADO,ADRIAN,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: MARTINEZ,JORGE

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 09:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 09:39

SUBJECT'S LAST NAME: NARANJA

DATE/TIME OF TEST: 09/18/2021 10:12:06

Diagnostics Check	OK	10:07	
Air Blank	0.000	10:08	
Control Test	0.078	10:08	
Air Blank	0.000	10:09	
Subject Sample	# 1	10:10	Subject Test Refused
Breath Volume	0.148	10:10	
Air Blank	0.000	10:11	
Control Test	0.081	10:11	
Air Blank	0.000	10:12	
Diagnostics Check	OK	10:12	

Data Download Date/Time: 09/27/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINO,BRYAN,C
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PINO,BRYAN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: COURT ORD
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/21/2021 19:06:08

Diagnostics Check	OK	19:00
Air Blank	0.000	19:00
Control Test	0.082	19:01
Air Blank	0.000	19:01
Subject Sample # 1	0.000	19:02
Breath Volume	1.613	19:02
Air Blank	0.000	19:02
Air Blank	0.000	19:04
Subject Sample # 2	0.000	19:04
Breath Volume	1.386	19:04
Air Blank	0.000	19:05
Control Test	0.081	19:05
Air Blank	0.000	19:06
Diagnostics Check	OK	19:06

Data Download Date/Time: 09/27/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINO,BRYAN,C
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PINO,BRYAN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: BLANCO

DATE/TIME OF TEST: 09/21/2021 22:24:07

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.081	22:15
Air Blank	0.000	22:16
Subject Sample # 1	0.177	22:17
Breath Volume	1.378	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.176	22:22
Breath Volume	2.738	22:22
Air Blank	0.000	22:23
Control Test	0.080	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 09/27/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPAZ,TAMMY,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CAPAZ,TAMMY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: COURT ORD
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: WILLIAM

DATE/TIME OF TEST: 09/22/2021 21:56:26

Diagnostics Check	OK	21:50
Air Blank	0.000	21:51
Control Test	0.082	21:51
Air Blank	0.000	21:51
Subject Sample # 1	0.000	21:52
Breath Volume	1.640	21:52
Air Blank	0.000	21:52
Air Blank	0.000	21:54
Subject Sample # 2	0.000	21:55
Breath Volume	1.500	21:55
Air Blank	0.000	21:55
Control Test	0.082	21:55
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 09/27/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,J,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: COURT ORD
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/23/2021 16:12:36

Diagnostics Check	OK	16:06
Air Blank	0.000	16:07
Control Test	0.082	16:07
Air Blank	0.000	16:08
Subject Sample # 1	0.000	16:08
Breath Volume	1.761	16:08
Air Blank	0.000	16:08
Air Blank	0.000	16:10
Subject Sample # 2	0.000	16:11
Breath Volume	1.765	16:11
Air Blank	0.000	16:11
Control Test	0.083	16:12
Air Blank	0.000	16:12
Diagnostics Check	OK	16:12

Data Download Date/Time: 09/27/2021 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TESHAMARIAM,SEM,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: GONZALEZ,WILLIAMS

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 02:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: TESHAMARIAM

DATE/TIME OF TEST: 09/25/2021 03:37:50

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:35	
Control Test	0.082	03:35	
Air Blank	0.000	03:36	
Subject Sample	# 1	03:36	Subject Test Refused
Breath Volume	0.000	03:36	
Air Blank	0.000	03:36	
Control Test	0.083	03:37	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

Data Download Date/Time: 09/27/2021 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ANTONIO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,WILLIAMS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: TESFAMARIAM

DATE/TIME OF TEST: 09/25/2021 03:56:16

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:53	
Control Test	0.082	03:54	
Air Blank	0.000	03:54	
Subject Sample	# 1	03:54	Subject Test Refused
Breath Volume	0.000	03:54	
Air Blank	0.000	03:55	
Control Test	0.082	03:55	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 09/27/2021 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GONZALEZ,CAMILA,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: GONZALEZ,CAMILA

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 04:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: AVILA

DATE/TIME OF TEST: 09/25/2021 05:00:18

Diagnostics Check	OK	04:56	
Air Blank	0.000	04:57	
Control Test	0.081	04:57	
Air Blank	0.000	04:58	
Subject Sample	# 1	04:58	Subject Test Refused
Breath Volume	0.000	04:58	
Air Blank	0.000	04:59	
Control Test	0.082	04:59	
Air Blank	0.000	05:00	
Diagnostics Check	OK	05:00	

Data Download Date/Time: 09/27/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GONZALEZ,CAMILA,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CHARLES,BYANKA

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 03:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 09/25/2021 05:46:57

Diagnostics Check	OK	05:39
Air Blank	0.000	05:40
Control Test	0.081	05:40
Air Blank	0.000	05:41
Subject Sample # 1	0.086	05:42
Breath Volume	2.183	05:42
Air Blank	0.000	05:43
Air Blank	0.000	05:44
Subject Sample # 2	0.086	05:45
Breath Volume	2.019	05:45
Air Blank	0.000	05:46
Control Test	0.080	05:46
Air Blank	0.000	05:46
Diagnostics Check	OK	05:46

Data Download Date/Time: 09/27/2021 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,DANELIA,J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,YUNIER
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 15:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 09/26/2021 16:56:38

Diagnostics Check	OK	16:43	
Air Blank	0.000	16:44	
Control Test	0.081	16:44	
Air Blank	0.000	16:45	
Subject Sample	# 1	16:48	No Sample Provided
Breath Volume	0.000	16:48	
Air Blank	0.000	16:48	
Air Blank	0.000	16:50	
Subject Sample	# 2	0.205	16:51
Breath Volume	1.414	16:51	
Air Blank	0.000	16:52	
Air Blank	0.000	16:54	
Subject Sample	# 3	0.193	16:54
Breath Volume	1.941	16:54	
Air Blank	0.000	16:55	
Control Test	0.081	16:56	
Air Blank	0.000	16:56	
Diagnostics Check	OK	16:56	

Data Download Date/Time: 09/27/2021 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PASTOR,JOSE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: CARIAS ASCON

DATE/TIME OF TEST: 09/06/2021 20:06:21

Diagnostics Check	OK	19:57
Air Blank	0.000	19:57
Control Test	0.080	19:58
Air Blank	0.000	19:58
Subject Sample # 1	0.237	20:01
Breath Volume	1.644	20:01
Air Blank	0.000	20:02
Air Blank	0.000	20:04
Subject Sample # 2	0.230	20:04
Breath Volume	1.664	20:04
Air Blank	0.000	20:05
Control Test	0.076	20:05
Air Blank	0.000	20:06
Diagnostics Check	OK	20:06

Data Download Date/Time: 09/20/2021 16:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PASTOR,JOSE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: PEROZO PARRA

DATE/TIME OF TEST: 09/13/2021 06:59:45

Diagnostics Check	OK	06:52
Air Blank	0.000	06:53
Control Test	0.079	06:53
Air Blank	0.000	06:54
Subject Sample # 1	0.231	06:54
Breath Volume	1.359	06:54
Air Blank	0.000	06:55
Air Blank	0.000	06:57
Subject Sample # 2	0.235	06:58
Breath Volume	1.402	06:58
Air Blank	0.000	06:58
Control Test	0.079	06:59
Air Blank	0.000	06:59
Diagnostics Check	OK	06:59

Data Download Date/Time: 09/20/2021 16:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,JUSIN,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROSE,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/08/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 09/02/2021 00:14:53

Diagnostics Check		OK	00:08
Air Blank		0.000	00:09
Control Test		0.080	00:09
Air Blank		0.000	00:09
Subject Sample	# 1	0.102	00:10
Breath Volume		2.007	00:10
Air Blank		0.000	00:11
Air Blank		0.000	00:12
Subject Sample	# 2	0.097	00:13
Breath Volume		1.980	00:13
Air Blank		0.000	00:13
Control Test		0.080	00:14
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 09/10/2021 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHEETZ,JOHN,W
BREATH TEST OPERATOR AGENCY: Calhoun County SO
ARRESTING OFFICER: SCHEETZ,JOHN
ARRESTING OFFICER AGENCY: Calhoun County SO
LAST AGENCY INSPECTION: 09/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: PROB/PAROL
ARREST TIME: 12:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: KENYON

DATE/TIME OF TEST: 09/21/2021 12:45:54

Diagnostics Check	OK	12:40
Air Blank	0.000	12:40
Control Test	0.080	12:40
Air Blank	0.000	12:41
Subject Sample # 1	0.000	12:41
Breath Volume	1.925	12:41
Air Blank	0.000	12:42
Air Blank	0.000	12:44
Subject Sample # 2	0.000	12:44
Breath Volume	2.277	12:44
Air Blank	0.000	12:45
Control Test	0.080	12:45
Air Blank	0.000	12:45
Diagnostics Check	OK	12:45

Data Download Date/Time: 10/07/2021 08:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEJERA,MARTHA,M
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: TEJERA,MARTHA
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: CAMJI

DATE/TIME OF TEST: 09/22/2021 23:51:21

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:37	
Control Test	0.077	23:37	
Air Blank	0.000	23:38	
Subject Sample	# 1 0.144	23:41	
Breath Volume	1.308	23:41	
Air Blank	0.000	23:42	
Air Blank	0.000	23:44	
Subject Sample	# 2 0.132	23:47	Volume Not Met
Breath Volume	0.367	23:47	
Air Blank	0.000	23:47	
Air Blank	0.000	23:49	
Subject Sample	# 3	23:49	Subject Test Refused
Breath Volume	0.000	23:49	
Air Blank	0.000	23:50	
Control Test	0.077	23:50	
Air Blank	0.000	23:51	
Diagnostics Check	OK	23:51	

Data Download Date/Time: 10/14/2021 03:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IINO,RYUNOSUKE,
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: TILEY,JOHN
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 09/03/2021 19:43:24

Diagnostics Check	OK	19:37
Air Blank	0.000	19:37
Control Test	0.081	19:37
Air Blank	0.000	19:38
Subject Sample # 1	0.000	19:39
Breath Volume	1.695	19:39
Air Blank	0.000	19:39
Air Blank	0.000	19:41
Subject Sample # 2	0.000	19:42
Breath Volume	1.273	19:42
Air Blank	0.000	19:42
Control Test	0.080	19:42
Air Blank	0.000	19:43
Diagnostics Check	OK	19:43

Data Download Date/Time: 09/16/2021 08:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELFNER,ISAAC,S
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: BURNS,JEFFERY
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 11:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: IZAGUIRRE

DATE/TIME OF TEST: 09/12/2021 12:31:36

Diagnostics Check	OK	12:25
Air Blank	0.000	12:26
Control Test	0.082	12:26
Air Blank	0.000	12:26
Subject Sample # 1	0.000	12:27
Breath Volume	3.546	12:27
Air Blank	0.000	12:27
Air Blank	0.000	12:29
Subject Sample # 2	0.000	12:30
Breath Volume	2.195	12:30
Air Blank	0.000	12:30
Control Test	0.080	12:31
Air Blank	0.000	12:31
Diagnostics Check	OK	12:31

Data Download Date/Time: 09/16/2021 08:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TIRADO,KARENDANIELA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: CORDOVA

DATE/TIME OF TEST: 09/04/2021 20:03:59

Diagnostics Check	OK	20:01	
Air Blank	0.000	20:01	
Control Test	0.080	20:01	
Air Blank	0.000	20:02	
Subject Sample	# 1	20:02	Subject Test Refused
Breath Volume	0.000	20:02	
Air Blank	0.000	20:03	
Control Test	0.080	20:03	
Air Blank	0.000	20:03	
Diagnostics Check	OK	20:03	

Data Download Date/Time: 09/07/2021 06:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FHP TROOP A
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1327124
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 09/10/2021 05:05:39

Diagnostics Check		OK	04:52	
Air Blank		0.000	04:53	
Control Test		0.080	04:53	
Air Blank		0.000	04:54	
Subject Sample	# 1	0.074	04:57	Volume Not Met
Breath Volume		0.062	04:57	
Air Blank		0.000	04:57	
Air Blank		0.000	04:59	
Subject Sample	# 2	0.085	05:00	
Breath Volume		1.476	05:00	
Air Blank		0.000	05:00	
Air Blank		0.000	05:02	
Subject Sample	# 3	0.085	05:04	
Breath Volume		2.187	05:04	
Air Blank		0.000	05:04	
Control Test		0.079	05:05	
Air Blank		0.000	05:05	
Diagnostics Check		OK	05:05	

Data Download Date/Time: 10/08/2021 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FHP TROOP A
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1327124
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 09/11/2021 04:22:41

Diagnostics Check		OK	04:15
Air Blank		0.000	04:16
Control Test		0.080	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.173	04:18
Breath Volume		1.531	04:18
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.173	04:21
Breath Volume		1.160	04:21
Air Blank		0.000	04:21
Control Test		0.079	04:22
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 10/08/2021 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FHP TROOP A
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1327124
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 09/13/2021 02:10:44

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.079	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.000	02:06
Breath Volume	2.082	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.000	02:09
Breath Volume	1.410	02:09
Air Blank	0.000	02:09
Control Test	0.079	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 10/08/2021 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: LANG

DATE/TIME OF TEST: 09/03/2021 23:51:54

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.078	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.138	23:47
Breath Volume	2.238	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:49
Subject Sample # 2	0.131	23:50
Breath Volume	2.109	23:50
Air Blank	0.000	23:50
Control Test	0.077	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 09/24/2021 01:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 09/04/2021 03:23:25

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:21	
Control Test	0.077	03:21	
Air Blank	0.000	03:21	
Subject Sample	# 1	03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:22	
Control Test	0.077	03:22	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 09/24/2021 01:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: SANFORD

DATE/TIME OF TEST: 09/04/2021 23:23:45

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.077	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.049	23:19
Breath Volume	1.660	23:19
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.051	23:22
Breath Volume	2.082	23:22
Air Blank	0.000	23:22
Control Test	0.077	23:23
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 09/24/2021 01:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: FONTES III

DATE/TIME OF TEST: 09/05/2021 03:36:10

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:33	
Control Test	0.077	03:34	
Air Blank	0.000	03:34	
Subject Sample	# 1	03:34	Subject Test Refused
Breath Volume	0.000	03:34	
Air Blank	0.000	03:35	
Control Test	0.078	03:35	
Air Blank	0.000	03:36	
Diagnostics Check	OK	03:36	

Data Download Date/Time: 09/24/2021 01:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 09:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:38

SUBJECT'S LAST NAME: PINTA

DATE/TIME OF TEST: 09/08/2021 10:05:23

Diagnostics Check	OK	09:58
Air Blank	0.000	09:58
Control Test	0.077	09:59
Air Blank	0.000	09:59
Subject Sample # 1	0.114	10:00
Breath Volume	1.507	10:00
Air Blank	0.000	10:00
Air Blank	0.000	10:02
Subject Sample # 2	0.114	10:03
Breath Volume	1.593	10:03
Air Blank	0.000	10:04
Control Test	0.078	10:04
Air Blank	0.000	10:05
Diagnostics Check	OK	10:05

Data Download Date/Time: 09/24/2021 01:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:55

SUBJECT'S LAST NAME: BENITEZ

DATE/TIME OF TEST: 09/10/2021 23:48:48

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:46	
Control Test	0.077	23:46	
Air Blank	0.000	23:47	
Subject Sample	# 1	23:47	Subject Test Refused
Breath Volume	0.000	23:47	
Air Blank	0.000	23:47	
Control Test	0.078	23:48	
Air Blank	0.000	23:48	
Diagnostics Check	OK	23:48	

Data Download Date/Time: 09/24/2021 01:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,KENNETH,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STONE,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: DAVIDSON II

DATE/TIME OF TEST: 09/12/2021 01:45:04

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.078	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.124	01:39
Breath Volume	1.445	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.120	01:43
Breath Volume	1.191	01:43
Air Blank	0.000	01:44
Control Test	0.079	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:45

Data Download Date/Time: 09/24/2021 01:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MAHDER,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: WINDSCHAUER

DATE/TIME OF TEST: 09/19/2021 04:35:29

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.078	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.172	04:30
Breath Volume	1.507	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.173	04:33
Breath Volume	1.453	04:33
Air Blank	0.000	04:34
Control Test	0.078	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 09/24/2021 01:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZDIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: VUKELAJ

DATE/TIME OF TEST: 09/23/2021 20:13:54

Diagnostics Check	OK	20:11	
Air Blank	0.000	20:11	
Control Test	0.078	20:11	
Air Blank	0.000	20:12	
Subject Sample	# 1	20:12	Subject Test Refused
Breath Volume	0.000	20:12	
Air Blank	0.000	20:13	
Control Test	0.078	20:13	
Air Blank	0.000	20:13	
Diagnostics Check	OK	20:13	

Data Download Date/Time: 09/24/2021 01:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: THOMAS,E
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: MCCAIG

DATE/TIME OF TEST: 09/11/2021 20:31:38

Diagnostics Check	OK	20:24
Air Blank	0.000	20:25
Control Test	0.080	20:25
Air Blank	0.000	20:25
Subject Sample # 1	0.065	20:26
Breath Volume	1.601	20:26
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample # 2	0.061	20:30
Breath Volume	1.429	20:30
Air Blank	0.000	20:30
Control Test	0.080	20:31
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

Data Download Date/Time: 09/24/2021 01:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 09/12/2021 02:28:26

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:25	
Control Test	0.080	02:26	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:27	Subject Test Refused
Breath Volume	0.000	02:27	
Air Blank	0.000	02:27	
Control Test	0.080	02:27	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:28	

Data Download Date/Time: 09/24/2021 01:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: HINOJO

DATE/TIME OF TEST: 09/15/2021 22:01:29

Diagnostics Check	OK	21:54
Air Blank	0.000	21:55
Control Test	0.081	21:55
Air Blank	0.000	21:56
Subject Sample # 1	0.195	21:56
Breath Volume	1.382	21:56
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.196	21:59
Breath Volume	1.769	21:59
Air Blank	0.000	22:00
Control Test	0.082	22:00
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 09/24/2021 01:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: FOREMAN

DATE/TIME OF TEST: 09/18/2021 23:42:55

Diagnostics Check	OK	23:36
Air Blank	0.000	23:37
Control Test	0.081	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.000	23:38
Breath Volume	1.101	23:38
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.000	23:41
Breath Volume	1.640	23:41
Air Blank	0.000	23:41
Control Test	0.082	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 09/24/2021 01:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: STEINKE

DATE/TIME OF TEST: 09/30/2021 23:56:43

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:55	
Control Test	0.086	23:56	Control Outside Tolerance
Air Blank	0.000	23:56	

Data Download Date/Time: 10/30/2021 00:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLNAR,JARROD,T
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BARTLETT,SHELTON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: BUI
ARREST TIME: 16:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: HUFFMASTER
DATE/TIME OF TEST: 09/04/2021 17:01:14

Diagnostics Check	OK	16:53
Air Blank	0.000	16:54
Control Test	0.080	16:54
Air Blank	0.000	16:55
Subject Sample # 1	0.122	16:56
Breath Volume	1.203	16:56
Air Blank	0.000	16:57
Air Blank	0.000	16:59
Subject Sample # 2	0.121	16:59
Breath Volume	1.914	16:59
Air Blank	0.000	17:00
Control Test	0.078	17:00
Air Blank	0.000	17:01
Diagnostics Check	OK	17:01

Data Download Date/Time: 09/22/2021 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MALTAIS,ANDREW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: BUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:57

SUBJECT'S LAST NAME: CAMPOS SOLANO

DATE/TIME OF TEST: 09/04/2021 18:37:45

Diagnostics Check	OK	18:27	
Air Blank	0.000	18:27	
Control Test	0.079	18:27	
Air Blank	0.000	18:28	
Subject Sample	# 1	18:31	No Sample Provided
Breath Volume	0.000	18:31	
Air Blank	0.000	18:32	
Air Blank	0.000	18:34	
Subject Sample	# 2	18:36	Subject Test Refused
Breath Volume	0.000	18:36	
Air Blank	0.000	18:36	
Control Test	0.081	18:37	
Air Blank	0.000	18:37	
Diagnostics Check	OK	18:37	

Data Download Date/Time: 09/22/2021 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWNING,CADEN,T
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DOWNING,CADEN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: BUI
ARREST TIME: 17:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: SMITHEY
DATE/TIME OF TEST: 09/05/2021 18:03:21

Diagnostics Check	OK	17:51	
Air Blank	0.000	17:51	
Control Test	0.080	17:51	
Air Blank	0.000	17:52	
Subject Sample	# 1	17:55	No Sample Provided
Breath Volume	0.000	17:55	
Air Blank	0.000	17:56	
Air Blank	0.000	17:58	
Subject Sample	# 2	0.107	17:58
Breath Volume	2.347	17:58	
Air Blank	0.000	17:59	
Air Blank	0.000	18:01	
Subject Sample	# 3	0.109	18:01
Breath Volume	3.140	18:01	
Air Blank	0.000	18:02	
Control Test	0.080	18:02	
Air Blank	0.000	18:03	
Diagnostics Check	OK	18:03	

Data Download Date/Time: 09/22/2021 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWNING,CADEN,T
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DOWNING,CADEN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: BUI
ARREST TIME: 14:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:09

SUBJECT'S LAST NAME: LANCASTER
DATE/TIME OF TEST: 09/06/2021 14:46:03

Diagnostics Check	OK	14:43	
Air Blank	0.000	14:43	
Control Test	0.080	14:43	
Air Blank	0.000	14:44	
Subject Sample	# 1	14:44	Subject Test Refused
Breath Volume	0.000	14:44	
Air Blank	0.000	14:45	
Control Test	0.081	14:45	
Air Blank	0.000	14:45	
Diagnostics Check	OK	14:46	

Data Download Date/Time: 09/22/2021 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LECOCQ,MATTHEW,E
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: BUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: DALE

DATE/TIME OF TEST: 09/25/2021 17:32:56

Diagnostics Check	OK	17:26
Air Blank	0.000	17:26
Control Test	0.081	17:27
Air Blank	0.000	17:27
Subject Sample # 1	0.175	17:28
Breath Volume	2.625	17:28
Air Blank	0.000	17:29
Air Blank	0.000	17:30
Subject Sample # 2	0.181	17:31
Breath Volume	2.511	17:31
Air Blank	0.000	17:32
Control Test	0.078	17:32
Air Blank	0.000	17:32
Diagnostics Check	OK	17:32

Data Download Date/Time: 10/22/2021 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD,TRAVIS,L
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BASFORD,TRAVIS
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: BUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: MCCONNELL

DATE/TIME OF TEST: 09/05/2021 17:54:54

Diagnostics Check	OK	17:45
Air Blank	0.000	17:45
Control Test	0.080	17:45
Air Blank	0.000	17:46
Subject Sample # 1	0.206	17:47
Breath Volume	5.050	17:47
Air Blank	0.000	17:47
Air Blank	0.000	17:49
Subject Sample # 2	0.179	17:50
Breath Volume	2.582	17:50
Air Blank	0.000	17:50
Air Blank	0.000	17:52
Subject Sample # 3	0.187	17:53
Breath Volume	2.808	17:53
Air Blank	0.000	17:53
Control Test	0.079	17:54
Air Blank	0.000	17:54
Diagnostics Check	OK	17:54

Data Download Date/Time: 09/22/2021 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

INSTRUMENT REGISTERED TO: FFWCC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULASKI,DAMON,J
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: MOSCHIANO,DAVID
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 09/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: RUNYON

DATE/TIME OF TEST: 09/26/2021 21:14:36

Diagnostics Check	OK	21:04
Air Blank	0.000	21:04
Control Test	0.080	21:05
Air Blank	0.000	21:05
Subject Sample # 1	0.180	21:06
Breath Volume	1.769	21:06
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.202	21:09
Breath Volume	3.042	21:09
Air Blank	0.000	21:10
Air Blank	0.000	21:12
Subject Sample # 3	0.196	21:13
Breath Volume	2.925	21:13
Air Blank	0.000	21:13
Control Test	0.079	21:13
Air Blank	0.000	21:14
Diagnostics Check	OK	21:14

Data Download Date/Time: 10/15/2021 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTER,JESSICA,L
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: TRUEBLOOD,RYAN
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: BUI
ARREST TIME: 17:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 09/05/2021 18:22:31

Diagnostics Check	OK	18:15
Air Blank	0.000	18:16
Control Test	0.081	18:16
Air Blank	0.000	18:17
Subject Sample # 1	0.185	18:17
Breath Volume	1.535	18:17
Air Blank	0.000	18:18
Air Blank	0.000	18:20
Subject Sample # 2	0.170	18:20
Breath Volume	2.445	18:20
Air Blank	0.000	18:21
Control Test	0.081	18:21
Air Blank	0.000	18:22
Diagnostics Check	OK	18:22

Data Download Date/Time: 11/30/2021 15:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CASTILLO,CHRISTIAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 09/01/2021 23:38:50

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:27	
Control Test	0.080	23:28	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:31	No Sample Provided
Breath Volume	0.000	23:31	
Air Blank	0.000	23:32	
Air Blank	0.000	23:34	
Subject Sample	# 2	23:37	No Sample Provided
Breath Volume	0.000	23:37	
Air Blank	0.000	23:37	
Control Test	0.081	23:38	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

Data Download Date/Time: 09/29/2021 12:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,JOSE,L
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WHITE,NICHEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 06:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: LARGAESPDA

DATE/TIME OF TEST: 09/18/2021 07:40:58

Diagnostics Check		OK	07:35
Air Blank		0.000	07:35
Control Test		0.080	07:35
Air Blank		0.000	07:36
Subject Sample	# 1	0.200	07:36
Breath Volume		2.160	07:36
Air Blank		0.000	07:37
Air Blank		0.000	07:39
Subject Sample	# 2	0.201	07:39
Breath Volume		2.906	07:39
Air Blank		0.000	07:40
Control Test		0.078	07:40
Air Blank		0.000	07:40
Diagnostics Check		OK	07:40

Data Download Date/Time: 09/29/2021 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELSCHINGER,MICKAEL,L
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: ROSADO,CARLOS
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: ZUNIGA RAMIREZ

DATE/TIME OF TEST: 09/26/2021 00:50:32

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:38	
Control Test	0.080	00:38	
Air Blank	0.000	00:39	
Subject Sample	# 1	00:42	No Sample Provided
Breath Volume	0.000	00:42	
Air Blank	0.000	00:42	
Air Blank	0.000	00:44	
Subject Sample	# 2	0.157	00:45
Breath Volume	2.203	00:45	
Air Blank	0.000	00:46	
Air Blank	0.000	00:48	
Subject Sample	# 3	0.154	00:48
Breath Volume	1.968	00:48	
Air Blank	0.000	00:49	
Control Test	0.080	00:49	
Air Blank	0.000	00:50	
Diagnostics Check	OK	00:50	

Data Download Date/Time: 09/29/2021 12:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAHLKE,M,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUBOIS,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ABEND

DATE/TIME OF TEST: 09/02/2021 01:44:43

Diagnostics Check		OK	01:36
Air Blank		0.000	01:37
Control Test		0.079	01:37
Air Blank		0.000	01:38
Subject Sample	# 1	0.262	01:39
Breath Volume		2.910	01:39
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.244	01:43
Breath Volume		2.152	01:43
Air Blank		0.000	01:43
Control Test		0.077	01:44
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 09/28/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAHLKE,M,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: PAHLKE,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: JARQUIN SANTOS

DATE/TIME OF TEST: 09/02/2021 23:42:57

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.079	23:37
Air Blank	0.000	23:37
Subject Sample # 1	0.137	23:38
Breath Volume	1.527	23:38
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.143	23:41
Breath Volume	1.390	23:41
Air Blank	0.000	23:42
Control Test	0.078	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 09/28/2021 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MORRIS,HOWARD
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: CARILLO

DATE/TIME OF TEST: 09/04/2021 04:59:22

Diagnostics Check		OK	04:52
Air Blank		0.000	04:53
Control Test		0.079	04:53
Air Blank		0.000	04:54
Subject Sample	# 1	0.207	04:54
Breath Volume		2.597	04:54
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.190	04:57
Breath Volume		2.046	04:57
Air Blank		0.000	04:58
Control Test		0.078	04:58
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 09/28/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 09/04/2021 21:52:58

Diagnostics Check	OK	21:43	
Air Blank	0.000	21:43	
Control Test	0.079	21:43	
Air Blank	0.000	21:44	
Subject Sample	# 1	21:47	No Sample Provided
Breath Volume	0.000	21:47	
Air Blank	0.000	21:47	
Air Blank	0.000	21:50	
Subject Sample	# 2	21:51	Subject Test Refused
Breath Volume	0.000	21:51	
Air Blank	0.000	21:52	
Control Test	0.078	21:52	
Air Blank	0.000	21:52	
Diagnostics Check	OK	21:52	

Data Download Date/Time: 09/28/2021 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: SINGLETARY

DATE/TIME OF TEST: 09/05/2021 01:38:40

Diagnostics Check		OK	01:32
Air Blank		0.000	01:32
Control Test		0.078	01:32
Air Blank		0.000	01:33
Subject Sample	# 1	0.182	01:34
Breath Volume		1.199	01:34
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.180	01:37
Breath Volume		1.433	01:37
Air Blank		0.000	01:37
Control Test		0.077	01:38
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 09/28/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: GRUHL

DATE/TIME OF TEST: 09/06/2021 01:59:24

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.078	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.142	01:54
Breath Volume	1.691	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.139	01:57
Breath Volume	1.503	01:57
Air Blank	0.000	01:58
Control Test	0.078	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 09/28/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,ROHAN,G
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: MORGAN,ROHAN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: AVILA

DATE/TIME OF TEST: 09/06/2021 03:22:07

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.079	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.180	03:17
Breath Volume	1.445	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:19
Subject Sample # 2	0.178	03:20
Breath Volume	1.878	03:20
Air Blank	0.000	03:21
Control Test	0.078	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 09/28/2021 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAHLKE,M,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: NERI,C
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: MELVIN

DATE/TIME OF TEST: 09/07/2021 00:22:34

Diagnostics Check		OK	00:16
Air Blank		0.000	00:16
Control Test		0.079	00:16
Air Blank		0.000	00:17
Subject Sample	# 1	0.293	00:17
Breath Volume		3.003	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.298	00:20
Breath Volume		3.199	00:20
Air Blank		0.000	00:21
Control Test		0.077	00:21
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 09/28/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: BURRIER

DATE/TIME OF TEST: 09/08/2021 23:14:47

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.080	23:08
Air Blank	0.000	23:08
Subject Sample # 1	0.189	23:10
Breath Volume	1.707	23:10
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.182	23:13
Breath Volume	1.929	23:13
Air Blank	0.000	23:13
Control Test	0.078	23:14
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 09/28/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ADAMS,JAMES
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 09/09/2021 04:02:46

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.080	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.186	03:58
Breath Volume	2.730	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.180	04:01
Breath Volume	2.531	04:01
Air Blank	0.000	04:01
Control Test	0.078	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 09/28/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRUM,JACOB,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ARIZA,JOHN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: ALEJANDRE

DATE/TIME OF TEST: 09/11/2021 06:25:34

Diagnostics Check	OK	06:19
Air Blank	0.000	06:19
Control Test	0.080	06:20
Air Blank	0.000	06:20
Subject Sample # 1	0.131	06:21
Breath Volume	1.675	06:21
Air Blank	0.000	06:21
Air Blank	0.000	06:23
Subject Sample # 2	0.133	06:24
Breath Volume	1.226	06:24
Air Blank	0.000	06:24
Control Test	0.078	06:24
Air Blank	0.000	06:25
Diagnostics Check	OK	06:25

Data Download Date/Time: 09/28/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: KOLODNY

DATE/TIME OF TEST: 09/12/2021 19:37:12

Diagnostics Check	OK	19:30
Air Blank	0.000	19:31
Control Test	0.079	19:31
Air Blank	0.000	19:31
Subject Sample # 1	0.094	19:32
Breath Volume	2.589	19:32
Air Blank	0.000	19:33
Air Blank	0.000	19:35
Subject Sample # 2	0.089	19:35
Breath Volume	2.191	19:35
Air Blank	0.000	19:36
Control Test	0.079	19:36
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time: 09/28/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANPELT,J,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HULTS,D
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: GARRAPPA

DATE/TIME OF TEST: 09/12/2021 23:35:43

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.079	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.201	23:31
Breath Volume	2.289	23:31
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.200	23:34
Breath Volume	2.222	23:34
Air Blank	0.000	23:34
Control Test	0.079	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 09/28/2021 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LUTZ,D
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: SANCHEZ FLORES

DATE/TIME OF TEST: 09/13/2021 04:10:33

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.079	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.222	04:06
Breath Volume	2.136	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.221	04:08
Breath Volume	2.359	04:08
Air Blank	0.000	04:09
Control Test	0.078	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 09/28/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JAYDA,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: EUBANKS,AARON
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: WILLER

DATE/TIME OF TEST: 09/13/2021 18:40:46

Diagnostics Check	OK	18:34
Air Blank	0.000	18:35
Control Test	0.079	18:35
Air Blank	0.000	18:35
Subject Sample # 1	0.000	18:36
Breath Volume	2.589	18:36
Air Blank	0.000	18:36
Air Blank	0.000	18:38
Subject Sample # 2	0.000	18:39
Breath Volume	2.593	18:39
Air Blank	0.000	18:39
Control Test	0.078	18:40
Air Blank	0.000	18:40
Diagnostics Check	OK	18:40

Data Download Date/Time: 09/28/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 09/14/2021 22:59:32

Diagnostics Check	OK	22:52
Air Blank	0.000	22:53
Control Test	0.078	22:53
Air Blank	0.000	22:53
Subject Sample # 1	0.114	22:54
Breath Volume	1.421	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:56
Subject Sample # 2	0.118	22:57
Breath Volume	1.332	22:57
Air Blank	0.000	22:58
Control Test	0.079	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 09/28/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANPELT,J,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: VANPELT,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: APLICANO

DATE/TIME OF TEST: 09/16/2021 01:58:44

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.079	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.175	01:54
Breath Volume	1.421	01:54
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.180	01:57
Breath Volume	1.140	01:57
Air Blank	0.000	01:57
Control Test	0.079	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 09/28/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ADAMS,JAMES
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 09/16/2021 02:32:04

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.079	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.200	02:27
Breath Volume	2.164	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.199	02:30
Breath Volume	1.847	02:30
Air Blank	0.000	02:31
Control Test	0.077	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:32

Data Download Date/Time: 09/28/2021 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,J,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: ALVARADO GOMEZ

DATE/TIME OF TEST: 09/16/2021 04:40:57

Diagnostics Check	OK	04:34	
Air Blank	0.000	04:34	
Control Test	0.079	04:34	
Air Blank	0.000	04:35	
Subject Sample # 1	0.206	04:36	
Breath Volume	2.550	04:36	
Air Blank	0.000	04:37	
Air Blank	0.000	04:38	
Subject Sample # 2	0.203	04:39	
Breath Volume	2.140	04:39	
Air Blank	0.000	04:40	
Control Test	0.074	04:40	Control Outside Tolerance
Air Blank	0.000	04:40	

Data Download Date/Time: 09/28/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ADAMS,JAMES
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LEON VICTORIA

DATE/TIME OF TEST: 09/21/2021 04:29:23

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.079	04:23
Air Blank		0.000	04:24
Subject Sample	# 1	0.157	04:24
Breath Volume		1.593	04:24
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.155	04:27
Breath Volume		1.390	04:27
Air Blank		0.000	04:28
Control Test		0.078	04:28
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 09/28/2021 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,J,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/22/2021 01:57:26

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:55	
Control Test	0.079	01:55	
Air Blank	0.000	01:55	
Subject Sample	# 1	01:56	Subject Test Refused
Breath Volume	0.000	01:56	
Air Blank	0.000	01:56	
Control Test	0.077	01:56	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

Data Download Date/Time: 09/28/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,J,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: TORRES TORRES

DATE/TIME OF TEST: 09/23/2021 19:13:32

Diagnostics Check	OK	19:07
Air Blank	0.000	19:07
Control Test	0.079	19:07
Air Blank	0.000	19:08
Subject Sample # 1	0.347	19:08
Breath Volume	1.867	19:08
Air Blank	0.000	19:09
Air Blank	0.000	19:11
Subject Sample # 2	0.354	19:11
Breath Volume	2.722	19:11
Air Blank	0.000	19:12
Control Test	0.078	19:12
Air Blank	0.000	19:13
Diagnostics Check	OK	19:13

Data Download Date/Time: 09/28/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING, JONATHAN, L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: JOHNSTON, TORI
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: PAINTER

DATE/TIME OF TEST: 09/24/2021 03:08:34

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.080	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.000	03:04
Breath Volume	1.714	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.000	03:07
Breath Volume	1.308	03:07
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 09/28/2021 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,6
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: WETSCH ROBINSON

DATE/TIME OF TEST: 09/25/2021 00:57:20

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.154	00:52
Breath Volume	1.257	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.161	00:55
Breath Volume	1.984	00:55
Air Blank	0.000	00:56
Control Test	0.078	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 09/28/2021 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,J,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 09/25/2021 02:35:40

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:32	
Control Test	0.078	02:32	
Air Blank	0.000	02:32	
Subject Sample # 1	0.208	02:34	
Breath Volume	3.816	02:34	
Air Blank		02:35	Purge Fail
Air Blank	0.000	02:35	

Data Download Date/Time: 09/28/2021 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,J,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 09/25/2021 02:43:47

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.075	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.193	02:39
Breath Volume	2.269	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:41
Subject Sample # 2	0.181	02:42
Breath Volume	2.171	02:42
Air Blank	0.000	02:42
Control Test	0.075	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 09/28/2021 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAHLKE,M,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: PAHLKE,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: PATAKY

DATE/TIME OF TEST: 09/25/2021 22:07:15

Diagnostics Check	OK	21:59
Air Blank	0.000	22:00
Control Test	0.079	22:00
Air Blank	0.000	22:01
Subject Sample # 1	0.108	22:01
Breath Volume	2.125	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.108	22:05
Breath Volume	2.519	22:05
Air Blank	0.000	22:06
Control Test	0.079	22:06
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 09/28/2021 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAHLKE,M,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SOMMERS,E
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 09/26/2021 02:15:11

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.079	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.168	02:10
Breath Volume	3.789	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:12
Subject Sample # 2	0.161	02:13
Breath Volume	3.601	02:13
Air Blank	0.000	02:14
Control Test	0.078	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 09/28/2021 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ADAMS,JAMES
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: TREJO RAMIREZ

DATE/TIME OF TEST: 09/26/2021 02:27:57

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.078	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.179	02:23
Breath Volume	2.433	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.176	02:26
Breath Volume	2.023	02:26
Air Blank	0.000	02:27
Control Test	0.078	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 09/28/2021 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,J,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: TOAL

DATE/TIME OF TEST: 09/26/2021 04:24:03

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:21	
Control Test	0.079	04:21	
Air Blank	0.000	04:22	
Subject Sample	# 1	04:22	Subject Test Refused
Breath Volume	0.000	04:22	
Air Blank	0.000	04:23	
Control Test	0.078	04:23	
Air Blank	0.000	04:23	
Diagnostics Check	OK	04:24	

Data Download Date/Time: 09/28/2021 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLINGAME,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GRIFFIN,V
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:16

SUBJECT'S LAST NAME: LESTER

DATE/TIME OF TEST: 09/26/2021 14:21:24

Diagnostics Check	OK	14:12
Air Blank	0.000	14:13
Control Test	0.079	14:13
Air Blank	0.000	14:13
Subject Sample # 1	0.310	14:14
Breath Volume	1.175	14:14
Air Blank	0.000	14:15
Air Blank	0.000	14:17
Subject Sample # 2	0.303	14:19
Breath Volume	1.589	14:19
Air Blank	0.000	14:20
Control Test	0.079	14:20
Air Blank	0.000	14:21
Diagnostics Check	OK	14:21

Data Download Date/Time: 09/28/2021 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANPELT,J,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ECKERT,K
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 09/26/2021 21:39:19

Diagnostics Check	OK	21:32
Air Blank	0.000	21:33
Control Test	0.079	21:33
Air Blank	0.000	21:33
Subject Sample # 1	0.079	21:34
Breath Volume	1.218	21:34
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.083	21:37
Breath Volume	2.265	21:37
Air Blank	0.000	21:38
Control Test	0.079	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 09/28/2021 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: VERBIEST,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: ARELLANO

DATE/TIME OF TEST: 09/27/2021 01:22:04

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.099	01:17
Breath Volume		2.187	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.100	01:20
Breath Volume		2.554	01:20
Air Blank		0.000	01:21
Control Test		0.079	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:22

Data Download Date/Time: 09/28/2021 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JAYDA,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BELL,JOSEPH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: RIFE

DATE/TIME OF TEST: 09/27/2021 19:43:52

Diagnostics Check	OK	19:35
Air Blank	0.000	19:36
Control Test	0.079	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.000	19:39
Breath Volume	1.585	19:39
Air Blank	0.000	19:39
Air Blank	0.000	19:41
Subject Sample # 2	0.000	19:42
Breath Volume	1.402	19:42
Air Blank	0.000	19:42
Control Test	0.077	19:43
Air Blank	0.000	19:43
Diagnostics Check	OK	19:43

Data Download Date/Time: 09/28/2021 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JAYDA,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HILL,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: OSVALDO

DATE/TIME OF TEST: 09/28/2021 04:21:36

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.079	04:15
Air Blank	0.000	04:15
Subject Sample # 1	0.179	04:16
Breath Volume	2.125	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.179	04:19
Breath Volume	2.445	04:19
Air Blank	0.000	04:20
Control Test	0.078	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 09/28/2021 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAHLKE,M,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SOMMERS,E
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CHIPEL CHIVALAN

DATE/TIME OF TEST: 09/30/2021 03:59:01

Diagnostics Check		OK	03:44	
Air Blank		0.000	03:44	
Control Test		0.079	03:45	
Air Blank		0.000	03:45	
Subject Sample	# 1	0.171	03:48	Volume Not Met
Breath Volume		0.140	03:48	
Air Blank		0.000	03:49	
Air Blank		0.000	03:51	
Subject Sample	# 2	0.177	03:53	
Breath Volume		1.832	03:53	
Air Blank		0.000	03:53	
Air Blank		0.000	03:55	
Subject Sample	# 3		03:57	Subject Test Refused
Breath Volume		0.386	03:57	
Air Blank		0.000	03:58	
Control Test		0.076	03:58	
Air Blank		0.000	03:58	
Diagnostics Check		OK	03:59	

Data Download Date/Time: 10/20/2021 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,J,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: BAEZ TORRES

DATE/TIME OF TEST: 09/30/2021 23:17:57

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.079	23:12
Air Blank	0.000	23:12
Subject Sample # 1	0.137	23:13
Breath Volume	1.535	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:15
Subject Sample # 2	0.142	23:16
Breath Volume	2.828	23:16
Air Blank	0.000	23:17
Control Test	0.076	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 10/20/2021 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JUSTIN,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: REINHARDT,KYLE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: WARNER

DATE/TIME OF TEST: 09/01/2021 04:19:12

Diagnostics Check	OK	04:11
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.098	04:13
Breath Volume	1.519	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.097	04:17
Breath Volume	1.187	04:17
Air Blank	0.000	04:18
Control Test	0.078	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 09/21/2021 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, TYLER, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DEAL, BRYAN
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 09/03/2021 23:46:58

Diagnostics Check	OK	23:38	
Air Blank	0.000	23:39	
Control Test	0.080	23:39	
Air Blank	0.000	23:39	
Subject Sample	# 1	23:42	No Sample Provided
Breath Volume	0.000	23:42	
Air Blank	0.000	23:43	
Air Blank	0.000	23:45	
Subject Sample	# 2	23:45	Subject Test Refused
Breath Volume	0.000	23:45	
Air Blank	0.000	23:46	
Control Test	0.080	23:46	
Air Blank	0.000	23:46	
Diagnostics Check	OK	23:46	

Data Download Date/Time: 09/21/2021 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIDNER,CODY,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WALTERS,KENRICK
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 09/04/2021 00:00:33

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.000	23:55
Breath Volume	2.460	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.000	23:59
Breath Volume	3.285	23:59
Air Blank	0.000	23:59
Control Test	0.080	23:59
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 09/21/2021 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIDNER,CODY,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GREEN,KATELYN
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 09/04/2021 00:16:19

Diagnostics Check		OK	00:09
Air Blank		0.000	00:09
Control Test		0.080	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.216	00:11
Breath Volume		2.511	00:11
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.218	00:14
Breath Volume		3.082	00:14
Air Blank		0.000	00:15
Control Test		0.079	00:15
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 09/21/2021 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, TYLER, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GREEN, KATELYN
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 09/04/2021 20:11:22

Diagnostics Check	OK	20:04
Air Blank	0.000	20:04
Control Test	0.079	20:05
Air Blank	0.000	20:05
Subject Sample # 1	0.148	20:06
Breath Volume	4.000	20:06
Air Blank	0.000	20:07
Air Blank	0.000	20:08
Subject Sample # 2	0.152	20:09
Breath Volume	3.773	20:09
Air Blank	0.000	20:10
Control Test	0.079	20:10
Air Blank	0.000	20:11
Diagnostics Check	OK	20:11

Data Download Date/Time: 09/21/2021 10:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELIONE, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FINK, ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: CUMBUS

DATE/TIME OF TEST: 09/07/2021 21:16:27

Diagnostics Check	OK	21:09
Air Blank	0.000	21:09
Control Test	0.079	21:10
Air Blank	0.000	21:10
Subject Sample # 1	0.114	21:11
Breath Volume	3.222	21:11
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.111	21:14
Breath Volume	2.882	21:14
Air Blank	0.000	21:15
Control Test	0.078	21:15
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 09/21/2021 10:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIDNER,CODY,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ITRI,ANGELA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: WALDRON

DATE/TIME OF TEST: 09/08/2021 21:09:56

Diagnostics Check	OK	20:57	
Air Blank	0.000	20:58	
Control Test	0.079	20:58	
Air Blank	0.000	20:58	
Subject Sample	# 1	21:02	No Sample Provided
Breath Volume	0.000	21:02	
Air Blank	0.000	21:02	
Air Blank	0.000	21:04	
Subject Sample	# 2	0.309	21:05
Breath Volume	1.816	21:05	
Air Blank	0.000	21:06	
Air Blank	0.000	21:07	
Subject Sample	# 3	0.308	21:08
Breath Volume	2.316	21:08	
Air Blank	0.000	21:09	
Control Test	0.078	21:09	
Air Blank	0.000	21:09	
Diagnostics Check	OK	21:09	

Data Download Date/Time: 09/21/2021 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELIONE, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: JONES, DAKOTA
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 09/10/2021 20:38:44

Diagnostics Check	OK	20:31
Air Blank	0.000	20:31
Control Test	0.080	20:32
Air Blank	0.000	20:32
Subject Sample # 1	0.000	20:33
Breath Volume	1.789	20:33
Air Blank	0.000	20:34
Air Blank	0.000	20:36
Subject Sample # 2	0.000	20:37
Breath Volume	1.613	20:37
Air Blank	0.000	20:37
Control Test	0.079	20:38
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

Data Download Date/Time: 09/21/2021 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELIONE, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: TUFANO, MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: HANKO

DATE/TIME OF TEST: 09/11/2021 01:08:23

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.154	01:03
Breath Volume		2.062	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.143	01:06
Breath Volume		2.148	01:06
Air Blank		0.000	01:07
Control Test		0.078	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 09/21/2021 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELIONE, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: NUNEZ, ANTHONY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 09/11/2021 04:07:34

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.145	04:02
Breath Volume	2.832	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.134	04:05
Breath Volume	2.871	04:05
Air Blank	0.000	04:06
Control Test	0.078	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 09/21/2021 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEME,SANTIAGO,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GONZALES,EDUARDO
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: FRANCISO

DATE/TIME OF TEST: 09/11/2021 23:57:27

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.167	23:52
Breath Volume	2.105	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.166	23:55
Breath Volume	1.750	23:55
Air Blank	0.000	23:56
Control Test	0.079	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 09/21/2021 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANG, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DEFALCO, CHRISTOPHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: PROUDY

DATE/TIME OF TEST: 09/14/2021 22:15:02

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.080	22:09
Air Blank	0.000	22:09
Subject Sample # 1	0.102	22:10
Breath Volume	1.816	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.095	22:13
Breath Volume	1.597	22:13
Air Blank	0.000	22:14
Control Test	0.079	22:14
Air Blank	0.000	22:14
Diagnostics Check	OK	22:15

Data Download Date/Time: 09/21/2021 10:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANG, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CHRISTOFFEL, JAMES
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: LOPEZ JIMENEZ

DATE/TIME OF TEST: 09/23/2021 05:32:49

Diagnostics Check		OK	05:23
Air Blank		0.000	05:23
Control Test		0.080	05:24
Air Blank		0.000	05:24
Subject Sample	# 1	0.226	05:25
Breath Volume		1.378	05:25
Air Blank		0.000	05:25
Air Blank		0.000	05:27
Subject Sample	# 2	0.205	05:28
Breath Volume		1.617	05:28
Air Blank		0.000	05:29
Air Blank		0.000	05:30
Subject Sample	# 3	0.226	05:31
Breath Volume		1.593	05:31
Air Blank		0.000	05:31
Control Test		0.079	05:32
Air Blank		0.000	05:32
Diagnostics Check		OK	05:32

Data Download Date/Time: 10/28/2021 08:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIALOBRZESKI,M,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: IRIZARRY RODRIGUEZ,H
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: DELAQUIL

DATE/TIME OF TEST: 09/25/2021 23:11:46

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.079	23:05
Air Blank	0.000	23:05
Subject Sample # 1	0.194	23:06
Breath Volume	2.582	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.188	23:10
Breath Volume	2.593	23:10
Air Blank	0.000	23:10
Control Test	0.079	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 10/28/2021 08:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JASON,E
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HENDREN,CARSON
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 09/28/2021 10:52:20

Diagnostics Check	OK	10:45
Air Blank	0.000	10:45
Control Test	0.080	10:46
Air Blank	0.000	10:46
Subject Sample # 1	0.168	10:47
Breath Volume	1.269	10:47
Air Blank	0.000	10:48
Air Blank	0.000	10:50
Subject Sample # 2	0.161	10:50
Breath Volume	1.371	10:50
Air Blank	0.000	10:51
Control Test	0.079	10:51
Air Blank	0.000	10:52
Diagnostics Check	OK	10:52

Data Download Date/Time: 10/28/2021 08:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUIS-PIERRE,RUBEN,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: STAHL,SEAN
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/04/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 09/05/2021 01:16:52

Diagnostics Check		OK	01:10
Air Blank		0.000	01:11
Control Test		0.078	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.101	01:12
Breath Volume		2.972	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:14
Subject Sample	# 2	0.097	01:15
Breath Volume		2.726	01:15
Air Blank		0.000	01:15
Control Test		0.077	01:16
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 09/05/2021 06:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WESTON,LARRY,L
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: YOUNGMAN,TALON
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:04

SUBJECT'S LAST NAME: GRAFSTEIN

DATE/TIME OF TEST: 09/12/2021 13:45:05

Diagnostics Check	OK	13:38
Air Blank	0.000	13:39
Control Test	0.077	13:39
Air Blank	0.000	13:40
Subject Sample # 1	0.000	13:40
Breath Volume	1.320	13:40
Air Blank	0.000	13:41
Air Blank	0.000	13:43
Subject Sample # 2	0.000	13:43
Breath Volume	2.042	13:43
Air Blank	0.000	13:44
Control Test	0.079	13:44
Air Blank	0.000	13:44
Diagnostics Check	OK	13:45

Data Download Date/Time: 10/31/2021 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVINGSTON,JESSICA,L
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: DIAZ,R
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/04/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: RODRIGUEZ-GOMEZ

DATE/TIME OF TEST: 09/16/2021 23:47:42

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.079	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.224	23:43
Breath Volume	1.597	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.233	23:46
Breath Volume	3.007	23:46
Air Blank	0.000	23:46
Control Test	0.077	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 10/31/2021 11:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA, JOSUE, D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: DIAZ, R
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: RODRIGUEZ GOMEZ

DATE/TIME OF TEST: 09/17/2021 07:56:24

Diagnostics Check		OK	07:50
Air Blank		0.000	07:50
Control Test		0.078	07:51
Air Blank		0.000	07:51
Subject Sample	# 1	0.080	07:52
Breath Volume		2.042	07:52
Air Blank		0.000	07:52
Air Blank		0.000	07:54
Subject Sample	# 2	0.076	07:54
Breath Volume		2.687	07:54
Air Blank		0.000	07:55
Control Test		0.077	07:55
Air Blank		0.000	07:56
Diagnostics Check		OK	07:56

Data Download Date/Time: 10/31/2021 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,JOSUE,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: DIAZ,R
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: RODRIGUEZ-GOMEZ

DATE/TIME OF TEST: 09/17/2021 12:38:07

Diagnostics Check	OK	12:32
Air Blank	0.000	12:32
Control Test	0.078	12:33
Air Blank	0.000	12:33
Subject Sample # 1	0.000	12:34
Breath Volume	2.320	12:34
Air Blank	0.000	12:34
Air Blank	0.000	12:36
Subject Sample # 2	0.000	12:36
Breath Volume	1.730	12:36
Air Blank	0.000	12:37
Control Test	0.079	12:37
Air Blank	0.000	12:38
Diagnostics Check	OK	12:38

Data Download Date/Time: 10/31/2021 11:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVINGSTON,JESSICA,L
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: STAHL,SEAN
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: L563104910440

DATE/TIME OF TEST: 09/25/2021 04:32:16

Diagnostics Check	OK	04:29	
Air Blank	0.000	04:29	
Control Test	0.078	04:30	
Air Blank	0.000	04:30	
Subject Sample	# 1	04:30	Subject Test Refused
Breath Volume	0.000	04:30	
Air Blank	0.000	04:31	
Control Test	0.078	04:31	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:32	

Data Download Date/Time: 10/31/2021 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,H
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: JAMES,STEVE
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 08/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: VARNES

DATE/TIME OF TEST: 09/09/2021 21:35:53

Diagnostics Check	OK	21:29
Air Blank	0.000	21:30
Control Test	0.081	21:30
Air Blank	0.000	21:31
Subject Sample # 1	0.000	21:31
Breath Volume	3.222	21:31
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.000	21:34
Breath Volume	3.148	21:34
Air Blank	0.000	21:34
Control Test	0.080	21:35
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

Data Download Date/Time: 09/27/2021 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDEN,JAMES,H
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: SCARBIN,DOUG
ARRESTING OFFICER AGENCY: Carrabelle PD
LAST AGENCY INSPECTION: 08/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: WALDEN

DATE/TIME OF TEST: 09/15/2021 01:52:41

Diagnostics Check	OK	01:46
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.071	01:48
Breath Volume	3.800	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.066	01:51
Breath Volume	4.957	01:51
Air Blank	0.000	01:51
Control Test	0.079	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 09/27/2021 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAKER,RUEL,W
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: HARTZOG,TRISTAN
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 08/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: BUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 09/03/2021 22:29:10

Diagnostics Check	OK	22:22
Air Blank	0.000	22:23
Control Test	0.078	22:23
Air Blank	0.000	22:23
Subject Sample # 1	0.000	22:24
Breath Volume	3.433	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.000	22:27
Breath Volume	3.410	22:27
Air Blank	0.000	22:28
Control Test	0.080	22:28
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 09/27/2021 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMM,JAMES,L
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: GAY,TYLER
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 08/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: GUNDERSON

DATE/TIME OF TEST: 09/04/2021 19:50:39

Diagnostics Check	OK	19:39	
Air Blank	0.000	19:39	
Control Test	0.078	19:40	
Air Blank	0.000	19:40	
Subject Sample	# 1	19:43	No Sample Provided
Breath Volume	0.000	19:43	
Air Blank	0.000	19:44	
Air Blank	0.000	19:46	
Subject Sample	# 2	19:49	No Sample Provided
Breath Volume	0.000	19:49	
Air Blank	0.000	19:49	
Control Test	0.079	19:50	
Air Blank	0.000	19:50	
Diagnostics Check	OK	19:50	

Data Download Date/Time: 09/27/2021 12:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COPELAND, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: GILLEY, MICHAEL
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 08/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: SCHAAF

DATE/TIME OF TEST: 09/01/2021 22:50:52

Diagnostics Check	OK	22:47	
Air Blank	0.000	22:48	
Control Test	0.081	22:48	
Air Blank	0.000	22:49	
Subject Sample	# 1	22:49	Subject Test Refused
Breath Volume	0.000	22:49	
Air Blank	0.000	22:49	
Control Test	0.081	22:50	
Air Blank	0.000	22:50	
Diagnostics Check	OK	22:50	

Data Download Date/Time: 09/02/2021 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 09/02/2021 00:23:29

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.081	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.000	00:19
Breath Volume	1.714	00:19
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.000	00:22
Breath Volume	2.703	00:22
Air Blank	0.000	00:22
Control Test	0.082	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 09/02/2021 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HINDS,LOGAN
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 09/02/2021 02:18:09

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.185	02:13
Breath Volume	1.886	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.179	02:16
Breath Volume	1.628	02:16
Air Blank	0.000	02:17
Control Test	0.081	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 09/02/2021 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ELLINGTON,A
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 09/10/2021 22:00:13

Diagnostics Check	OK	21:53
Air Blank	0.000	21:54
Control Test	0.080	21:54
Air Blank	0.000	21:54
Subject Sample # 1	0.181	21:55
Breath Volume	3.863	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.172	21:58
Breath Volume	3.191	21:58
Air Blank	0.000	21:59
Control Test	0.081	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 10/07/2021 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUIS,BOBBY,G
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: DEAN,DONALD
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: MACKIN

DATE/TIME OF TEST: 09/12/2021 06:22:23

Diagnostics Check	OK	06:09	
Air Blank	0.000	06:10	
Control Test	0.081	06:10	
Air Blank	0.000	06:11	
Subject Sample	# 1 0.183	06:12	
Breath Volume	1.582	06:12	
Air Blank	0.000	06:12	
Air Blank	0.000	06:14	
Subject Sample	# 2 0.160	06:16	
Breath Volume	1.500	06:16	
Air Blank	0.000	06:17	
Air Blank	0.000	06:19	
Subject Sample	# 3	06:21	Subject Test Refused
Breath Volume	0.000	06:21	
Air Blank	0.000	06:21	
Control Test	0.078	06:21	
Air Blank	0.000	06:22	
Diagnostics Check	OK	06:22	

Data Download Date/Time: 10/07/2021 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATLOCK,LOGAN,R
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: CAMPBELL,SHAE
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: BEGAULT

DATE/TIME OF TEST: 09/15/2021 01:47:42

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.080	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.000	01:43
Breath Volume	2.425	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.000	01:46
Breath Volume	2.585	01:46
Air Blank	0.000	01:46
Control Test	0.082	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 10/07/2021 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: DINH,T
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: VANCOOTEN

DATE/TIME OF TEST: 09/18/2021 01:46:46

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.080	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.198	01:42
Breath Volume	2.187	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.210	01:45
Breath Volume	2.082	01:45
Air Blank	0.000	01:45
Control Test	0.081	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 10/07/2021 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SODERHOLM,CHELSEA,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: TACZAK,D
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 09/18/2021 03:02:13

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.175	02:57
Breath Volume	1.191	02:57
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.166	03:00
Breath Volume	2.175	03:00
Air Blank	0.000	03:01
Control Test	0.080	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 10/07/2021 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,TREVOR,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: JOHNSON,TREVOR
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:16

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 09/18/2021 07:44:08

Diagnostics Check	OK	07:41	
Air Blank	0.000	07:41	
Control Test	0.081	07:41	
Air Blank	0.000	07:42	
Subject Sample	# 1	07:42	Subject Test Refused
Breath Volume	0.000	07:42	
Air Blank	0.000	07:43	
Control Test	0.081	07:43	
Air Blank	0.000	07:44	
Diagnostics Check	OK	07:44	

Data Download Date/Time: 10/07/2021 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,BENJAMIN,Z
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: TACZAK,DAVID
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: PERALTA

DATE/TIME OF TEST: 09/19/2021 01:38:54

Diagnostics Check	OK	01:38	
Air Blank		01:38	RFI Detect
Air Blank		01:38	RFI Detect

Data Download Date/Time: 10/07/2021 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,BENJAMIN,Z
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: TACZAK,DAVID
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: PERALTA

DATE/TIME OF TEST: 09/19/2021 01:47:42

Diagnostics Check		OK	01:41
Air Blank		0.000	01:41
Control Test		0.081	01:41
Air Blank		0.000	01:42
Subject Sample	# 1	0.206	01:43
Breath Volume		2.691	01:43
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.204	01:46
Breath Volume		2.929	01:46
Air Blank		0.000	01:46
Control Test		0.079	01:47
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 10/07/2021 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,BENJAMIN,Z
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WESTMAN,PATRICK
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: STRATTON

DATE/TIME OF TEST: 09/19/2021 03:26:06

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:22	
Control Test	0.081	03:22	
Air Blank	0.000	03:23	
Subject Sample	# 1 0.194	03:24	
Breath Volume	2.351	03:24	
Air Blank		03:25	Purge Fail
Air Blank		03:26	Purge Fail

Data Download Date/Time: 10/07/2021 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,BENJAMIN,Z
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WESTMAN,PATRICK
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: STRATTON

DATE/TIME OF TEST: 09/19/2021 03:34:58

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:28	
Control Test	0.079	03:29	
Air Blank	0.000	03:29	
Subject Sample	# 1 0.194	03:30	
Breath Volume	2.675	03:30	
Air Blank	0.000	03:30	
Air Blank	0.000	03:32	
Subject Sample	# 2 0.203	03:33	
Breath Volume	2.460	03:33	
Air Blank		03:34	Purge Fail
Air Blank		03:34	Purge Fail

Data Download Date/Time: 10/07/2021 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,BENJAMIN,Z
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WESTMAN,PATRICK
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: STRATTON

DATE/TIME OF TEST: 09/19/2021 03:38:16

Diagnostics Check	OK	03:37	
Air Blank		03:37	RFI Detect
Air Blank	0.000	03:38	

Data Download Date/Time: 10/07/2021 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,BENJAMIN,Z
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WESTMAN,PATRICK
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: STRATTON

DATE/TIME OF TEST: 09/19/2021 03:41:27

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.056	03:40	Control Outside Tolerance
Air Blank	0.000	03:41	

Data Download Date/Time: 10/07/2021 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,BENJAMIN,Z
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WESTMAN,PATRICK
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: STRATTON

DATE/TIME OF TEST: 09/19/2021 03:49:22

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:48	
Control Test	0.058	03:48	Control Outside Tolerance
Air Blank	0.000	03:49	

Data Download Date/Time: 10/07/2021 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERESSE,DAVID,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: PEARSON,DAMIEN
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:57

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 09/19/2021 14:22:53

Diagnostics Check	OK	14:21	
Air Blank	0.000	14:21	
Control Test	0.074	14:22	Control Outside Tolerance
Air Blank	0.000	14:22	

Data Download Date/Time: 10/07/2021 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERESSE,DAVID,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: PEARSON,DAMIEN
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:57

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 09/19/2021 14:27:50

Diagnostics Check	OK	14:24	
Air Blank	0.000	14:24	
Control Test	0.077	14:24	
Air Blank	0.000	14:25	
Subject Sample # 1	0.223	14:26	
Breath Volume	2.789	14:26	
Air Blank		14:26	Purge Fail
Air Blank	0.000	14:27	

Data Download Date/Time: 10/07/2021 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERESSE,DAVID,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: PEARSON,DAMIEN
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:57

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 09/19/2021 14:30:21

Diagnostics Check	OK	14:28	
Air Blank	0.000	14:29	
Control Test	0.072	14:29	Control Outside Tolerance
Air Blank	0.000	14:30	

Data Download Date/Time: 10/07/2021 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/03/2021 15:48:34

Diagnostics Check	OK	15:42	
Air Blank	0.000	15:43	
Control Test	0.078	15:43	
Air Blank	0.000	15:44	
Subject Sample # 1	0.113	15:45	
Breath Volume	2.003	15:45	
Air Blank	0.000	15:45	
Air Blank	0.000	15:47	
Subject Sample # 2		15:47	RFI Detect
Breath Volume	1.050	15:47	
Air Blank	0.000	15:48	
		15:48	RFI Detect
Air Blank		15:48	RFI Detect

Data Download Date/Time: 10/11/2021 12:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/03/2021 15:51:32

Diagnostics Check	Fail	15:51	Temperature Regulation Test
Air Blank	0.000	15:51	

Data Download Date/Time: 10/11/2021 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 09/05/2021 02:26:30

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.106	02:21
Breath Volume	2.117	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.108	02:24
Breath Volume	1.824	02:24
Air Blank	0.000	02:25
Control Test	0.078	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 10/11/2021 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: GALLOWAY

DATE/TIME OF TEST: 09/05/2021 22:59:57

Diagnostics Check	OK	22:53
Air Blank	0.000	22:54
Control Test	0.076	22:54
Air Blank	0.000	22:55
Subject Sample # 1	0.187	22:55
Breath Volume	2.343	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.180	22:58
Breath Volume	2.531	22:58
Air Blank	0.000	22:59
Control Test	0.076	22:59
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 10/11/2021 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SAFFORD, STERLING
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 09/07/2021 02:01:50

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.081	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.159	01:57
Breath Volume	2.617	01:57
Air Blank	0.000	01:58
Air Blank	0.000	01:59
Subject Sample # 2	0.160	02:00
Breath Volume	2.503	02:00
Air Blank	0.000	02:00
Control Test	0.079	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 10/11/2021 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/03/2021 16:26:59

Diagnostics Check	OK	16:20
Air Blank	0.000	16:21
Control Test	0.082	16:21
Air Blank	0.000	16:22
Subject Sample # 1	0.113	16:22
Breath Volume	3.195	16:22
Air Blank	0.000	16:23
Air Blank	0.000	16:25
Subject Sample # 2	0.113	16:25
Breath Volume	3.136	16:25
Air Blank	0.000	16:26
Control Test	0.082	16:26
Air Blank	0.000	16:26
Diagnostics Check	OK	16:26

Data Download Date/Time: 09/07/2021 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUBOSE,ROGER,C
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BELL,MATT
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: BUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: TORO

DATE/TIME OF TEST: 09/04/2021 22:26:09

Diagnostics Check	OK	22:19
Air Blank	0.000	22:20
Control Test	0.082	22:20
Air Blank	0.000	22:20
Subject Sample # 1	0.205	22:21
Breath Volume	3.007	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.212	22:24
Breath Volume	2.894	22:24
Air Blank	0.000	22:25
Control Test	0.080	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 09/07/2021 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIDDLEBROOKS,ALAN,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MIDDLEBROOKS,ALAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 09/04/2021 04:36:12

Diagnostics Check	OK	04:32	
Air Blank	0.000	04:33	
Control Test	0.080	04:33	
Air Blank	0.000	04:34	
Subject Sample	# 1	04:34	Subject Test Refused
Breath Volume	0.000	04:34	
Air Blank	0.000	04:35	
Control Test	0.079	04:35	
Air Blank	0.000	04:36	
Diagnostics Check	OK	04:36	

Data Download Date/Time: 09/27/2021 12:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIDDLEBROOKS,ALAN,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MIDDLEBROOKS,ALAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 09/05/2021 01:43:05

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:40	
Control Test	0.080	01:41	
Air Blank	0.000	01:41	
Subject Sample	# 1	01:41	Subject Test Refused
Breath Volume	0.000	01:41	
Air Blank	0.000	01:42	
Control Test	0.079	01:42	
Air Blank	0.000	01:43	
Diagnostics Check	OK	01:43	

Data Download Date/Time: 09/27/2021 12:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,NATHAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HADDOCK,DAKOTA
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CAUSSEAUX

DATE/TIME OF TEST: 09/11/2021 01:32:50

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:30	
Control Test	0.080	01:30	
Air Blank	0.000	01:31	
Subject Sample	# 1	01:31	Subject Test Refused
Breath Volume	0.000	01:31	
Air Blank	0.000	01:31	
Control Test	0.079	01:32	
Air Blank	0.000	01:32	
Diagnostics Check	OK	01:32	

Data Download Date/Time: 09/27/2021 12:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LECKINGER, MICHAEL, E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: LECKINGER, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 09/19/2021 17:24:43

Diagnostics Check	OK	17:21	
Air Blank	0.000	17:22	
Control Test	0.078	17:22	
Air Blank	0.000	17:23	
Subject Sample	# 1	17:23	Subject Test Refused
Breath Volume	0.000	17:23	
Air Blank	0.000	17:23	
Control Test	0.079	17:24	
Air Blank	0.000	17:24	
Diagnostics Check	OK	17:24	

Data Download Date/Time: 09/27/2021 12:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE,ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: CHAVEZGOMEZ

DATE/TIME OF TEST: 09/04/2021 00:30:16

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.080	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.114	00:25
Breath Volume	1.148	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.106	00:28
Breath Volume	1.179	00:28
Air Blank	0.000	00:29
Control Test	0.078	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 09/08/2021 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOWNS,NATALIE
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: SAMAROO

DATE/TIME OF TEST: 09/04/2021 01:39:44

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:25	
Control Test		0.080	01:25	
Air Blank		0.000	01:26	
Subject Sample	# 1	0.210	01:26	
Breath Volume		2.054	01:26	
Air Blank		0.000	01:27	
Air Blank		0.000	01:29	
Subject Sample	# 2	0.151	01:32	Volume Not Met
Breath Volume		0.015	01:32	
Air Blank		0.000	01:33	
Air Blank		0.000	01:35	
Subject Sample	# 3	0.198	01:38	
Breath Volume		2.585	01:38	
Air Blank		0.000	01:38	
Control Test		0.078	01:39	
Air Blank		0.000	01:39	
Diagnostics Check		OK	01:39	

Data Download Date/Time: 09/08/2021 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: ARRIBA

DATE/TIME OF TEST: 09/04/2021 04:16:04

Diagnostics Check	OK	04:08
Air Blank	0.000	04:09
Control Test	0.079	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.149	04:11
Breath Volume	1.246	04:11
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.154	04:14
Breath Volume	1.179	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:16

Data Download Date/Time: 09/08/2021 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: OGLETREE,JERRELL
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: DIGGINS

DATE/TIME OF TEST: 09/04/2021 04:59:00

Diagnostics Check		OK	04:42	
Air Blank		0.000	04:43	
Control Test		0.079	04:43	
Air Blank		0.000	04:44	
Subject Sample	# 1	0.108	04:47	Volume Not Met
Breath Volume		0.214	04:47	
Air Blank		0.000	04:47	
Air Blank		0.000	04:49	
Subject Sample	# 2	0.098	04:52	
Breath Volume		1.601	04:52	
Air Blank		0.000	04:53	
Air Blank		0.000	04:55	
Subject Sample	# 3	0.107	04:57	
Breath Volume		1.226	04:57	
Air Blank		0.000	04:58	
Control Test		0.079	04:58	
Air Blank		0.000	04:58	
Diagnostics Check		OK	04:59	

Data Download Date/Time: 09/08/2021 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARARD,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/04/2021 05:22:16

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.079	05:17
Air Blank	0.000	05:17
Subject Sample # 1	0.135	05:18
Breath Volume	1.664	05:18
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.136	05:20
Breath Volume	3.027	05:20
Air Blank	0.000	05:21
Control Test	0.079	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 09/08/2021 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: INABNET,JUSTIN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: BABOOLALL

DATE/TIME OF TEST: 09/11/2021 02:55:31

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.079	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.198	02:51
Breath Volume	1.484	02:51
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.201	02:54
Breath Volume	1.523	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 10/14/2021 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: JIMENEZ RODRIGUEZ

DATE/TIME OF TEST: 09/12/2021 03:56:31

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.148	03:52
Breath Volume	3.550	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.150	03:55
Breath Volume	3.925	03:55
Air Blank	0.000	03:55
Control Test	0.079	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 10/14/2021 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PENA CARRION,MARLON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:32

SUBJECT'S LAST NAME: VACISKI BARBOSA

DATE/TIME OF TEST: 09/13/2021 06:09:54

Diagnostics Check	OK	06:00
Air Blank	0.000	06:01
Control Test	0.079	06:01
Air Blank	0.000	06:01
Subject Sample # 1	0.172	06:03
Breath Volume	1.425	06:03
Air Blank	0.000	06:04
Air Blank	0.000	06:06
Subject Sample # 2	0.167	06:08
Breath Volume	1.238	06:08
Air Blank	0.000	06:09
Control Test	0.079	06:09
Air Blank	0.000	06:09
Diagnostics Check	OK	06:09

Data Download Date/Time: 10/14/2021 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MELLA,MELISSA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: ALTAY

DATE/TIME OF TEST: 09/19/2021 00:15:57

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.081	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.110	00:11
Breath Volume	1.578	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.104	00:14
Breath Volume	1.250	00:14
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 10/14/2021 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PROFFER,KATIE
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: HODGE

DATE/TIME OF TEST: 09/19/2021 04:18:11

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.080	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.178	04:13
Breath Volume	1.386	04:13
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.164	04:16
Breath Volume	1.488	04:16
Air Blank	0.000	04:17
Control Test	0.080	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 10/14/2021 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HOUSTON,ISAAC
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: PHAM

DATE/TIME OF TEST: 09/19/2021 05:25:02

Diagnostics Check		OK	05:18
Air Blank		0.000	05:19
Control Test		0.079	05:19
Air Blank		0.000	05:20
Subject Sample	# 1	0.188	05:20
Breath Volume		2.203	05:20
Air Blank		0.000	05:21
Air Blank		0.000	05:23
Subject Sample	# 2	0.181	05:23
Breath Volume		1.710	05:23
Air Blank		0.000	05:24
Control Test		0.079	05:24
Air Blank		0.000	05:24
Diagnostics Check		OK	05:25

Data Download Date/Time: 10/14/2021 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GRIFFITHS,CODY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: PAULSWORTH

DATE/TIME OF TEST: 09/20/2021 00:09:53

Diagnostics Check		OK	00:04
Air Blank		0.000	00:04
Control Test		0.079	00:04
Air Blank		0.000	00:05
Subject Sample	# 1	0.207	00:05
Breath Volume		1.539	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.212	00:08
Breath Volume		1.687	00:08
Air Blank		0.000	00:08
Control Test		0.079	00:09
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 10/14/2021 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DEL CASTILLO,BRET
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: SIERRAHERNANDEZ

DATE/TIME OF TEST: 09/20/2021 21:33:37

Diagnostics Check	OK	21:27
Air Blank	0.000	21:28
Control Test	0.079	21:28
Air Blank	0.000	21:29
Subject Sample # 1	0.226	21:29
Breath Volume	1.250	21:29
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.229	21:32
Breath Volume	1.750	21:32
Air Blank	0.000	21:32
Control Test	0.080	21:33
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 10/14/2021 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMEY,MICHAEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 09/21/2021 02:13:07

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.177	02:08
Breath Volume	1.402	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.179	02:11
Breath Volume	1.949	02:11
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 10/14/2021 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOHNSON,JAMES
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: MENDEZSALES

DATE/TIME OF TEST: 09/25/2021 03:58:30

Diagnostics Check		OK	03:52
Air Blank		0.000	03:52
Control Test		0.080	03:53
Air Blank		0.000	03:53
Subject Sample	# 1	0.159	03:54
Breath Volume		2.062	03:54
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.161	03:57
Breath Volume		2.140	03:57
Air Blank		0.000	03:57
Control Test		0.079	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 10/14/2021 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IMPARATO,FRANK
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: ORTEGA

DATE/TIME OF TEST: 09/27/2021 03:26:55

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.000	03:22
Breath Volume	1.777	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.000	03:25
Breath Volume	1.726	03:25
Air Blank	0.000	03:26
Control Test	0.081	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 10/14/2021 09:09