Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,COLE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MCDONALD,COLE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: ZOOK

DATE/TIME OF TEST: 01/01/2022 02:51:13

Diagnostics Check		OK	02:48	
Air Blank		0.000	02:48	
Control Test		0.080	02:48	
Air Blank		0.000	02:49	
Subject Sample	# 1		02:49	Subject Test Refused
Breath Volume		0.000	02:49	
Air Blank		0.000	02:50	
Control Test		0.081	02:50	
Air Blank		0.000	02:51	
Diagnostics Check		OK	02:51	

Data Download Date/Time: 01/29/2022 14:13

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,COLE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: ELAFI,ALEX
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: SMEDLEY-MILLER

DATE/TIME OF TEST: 01/01/2022 04:05:06

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.080	03:58
Air Blank		0.000	03:58
Subject Sample	# 1	0.192	03:59
Breath Volume		2.050	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.184	04:03
Breath Volume		1.632	04:03
Air Blank		0.000	04:04
Control Test		0.080	04:04
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 01/29/2022 14:13

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBOLI,MELANIE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: TRIMBOLI,MELANIE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:04

SUBJECT'S LAST NAME: BAILES

DATE/TIME OF TEST: 01/02/2022 15:34:01

Diagnostics Check		OK	15:27
Air Blank		0.000	15:28
Control Test		0.080	15:28
Air Blank		0.000	15:29
Subject Sample	# 1	0.000	15:29
Breath Volume		1.960	15:29
Air Blank		0.000	15:30
Air Blank		0.000	15:32
Subject Sample	# 2	0.000	15:32
Breath Volume		2.148	15:32
Air Blank		0.000	15:33
Control Test		0.080	15:33
Air Blank		0.000	15:33
Diagnostics Check		OK	15:34

Data Download Date/Time: 01/29/2022 14:14

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: GIBBONS,L
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: BURTON

DATE/TIME OF TEST: 01/03/2022 19:07:41

Diagnostics Check		OK	19:01
Air Blank		0.000	19:02
Control Test		0.080	19:02
Air Blank		0.000	19:02
Subject Sample	# 1	0.000	19:03
Breath Volume		2.929	19:03
Air Blank		0.000	19:03
Air Blank		0.000	19:05
Subject Sample	# 2	0.000	19:06
Breath Volume		1.859	19:06
Air Blank		0.000	19:06
Control Test		0.081	19:07
Air Blank		0.000	19:07
Diagnostics Check		OK	19:07

Data Download Date/Time: 01/29/2022 14:14

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEILL,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: NEILL,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: CARSON

DATE/TIME OF TEST: 01/04/2022 00:17:11

Diagnostics Check		OK	00:08
Air Blank		0.000	00:09
Control Test		0.080	00:09
Air Blank		0.000	00:09
Subject Sample	# 1	0.157	00:11
Breath Volume		2.406	00:11
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.157	00:15
Breath Volume		2.410	00:15
Air Blank		0.000	00:16
Control Test		0.079	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 01/29/2022 14:14

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEILL,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: NEILL,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: CELLINI

DATE/TIME OF TEST: 01/05/2022 04:15:04

Diagnostics Check		OK	04:09
Air Blank		0.000	04:09
Control Test		0.081	04:09
Air Blank		0.000	04:10
Subject Sample	# 1	0.073	04:10
Breath Volume		2.136	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.073	04:13
Breath Volume		2.250	04:13
Air Blank		0.000	04:14
Control Test		0.080	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:15

Data Download Date/Time: 01/29/2022 14:14

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEILL,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: NEILL,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: MELGAR

DATE/TIME OF TEST: 01/09/2022 01:43:57

Diagnostics Check		OK	01:37
Air Blank		0.000	01:37
Control Test		0.080	01:38
Air Blank		0.000	01:38
Subject Sample	# 1	0.145	01:39
Breath Volume		1.925	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:41
Subject Sample	# 2	0.148	01:42
Breath Volume		2.652	01:42
Air Blank		0.000	01:43
Control Test		0.080	01:43
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 01/29/2022 14:14

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEILL,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: NEILL,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: FELIX

DATE/TIME OF TEST: 01/13/2022 01:01:10

Diagnostics Check		OK	00:54
Air Blank		0.000	00:55
Control Test		0.080	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.000	00:56
Breath Volume		4.117	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.007	00:59
Breath Volume		4.671	00:59
Air Blank		0.000	01:00
Control Test		0.080	01:00
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 01/29/2022 14:14

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLCROFT,AMANDA,L
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HOLLCROFT,AMANDA
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:07

SUBJECT'S LAST NAME: LIVINGSTON

DATE/TIME OF TEST: 01/15/2022 13:38:08

Diagnostics Check		OK	13:32
Air Blank		0.000	13:32
Control Test		0.081	13:33
Air Blank		0.000	13:33
Subject Sample	# 1	0.000	13:33
Breath Volume		1.964	13:33
Air Blank		0.000	13:34
Air Blank		0.000	13:36
Subject Sample	# 2	0.000	13:36
Breath Volume		2.371	13:36
Air Blank		0.000	13:37
Control Test		0.081	13:37
Air Blank		0.000	13:38
Diagnostics Check		OK	13:38

Data Download Date/Time: 01/29/2022 14:14

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: TEAGUE,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: HARTSFIELD

DATE/TIME OF TEST: 01/15/2022 22:28:11

Diagnostics Check		OK	22:21
Air Blank		0.000	22:21
Control Test		0.080	22:21
Air Blank		0.000	22:22
Subject Sample	# 1	0.000	22:22
Breath Volume		1.265	22:22
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.000	22:26
Breath Volume		1.183	22:26
Air Blank		0.000	22:27
Control Test		0.081	22:27
Air Blank		0.000	22:28
Diagnostics Check		OK	22:28

Data Download Date/Time: 01/29/2022 14:14

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLCROFT,AMANDA,L
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HOLLCROFT,AMANDA
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: RODRIGUEZ CRUZ

DATE/TIME OF TEST: 01/16/2022 07:17:18

Diagnostics Check		OK	07:10	
Air Blank		0.000	07:11	
Control Test		0.081	07:11	
Air Blank		0.000	07:12	
Subject Sample	# 1	0.125	07:12	
Breath Volume		1.484	07:12	
Air Blank		0.000	07:13	
Air Blank		0.000	07:15	
Subject Sample	# 2	0.123	07:15	
Breath Volume		1.441	07:15	
Air Blank		0.000	07:16	
Control Test		0.079	07:16	
Air Blank			07:16	RFI Detect
Air Blank		0.000	07:17	

Data Download Date/Time: 01/29/2022 14:14

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEILL,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: NEILL,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: SHAH

DATE/TIME OF TEST: 01/23/2022 02:04:30

Diagnostics Check		OK	01:58
Air Blank		0.000	01:58
Control Test		0.080	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.185	01:59
Breath Volume		1.511	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.192	02:02
Breath Volume		2.250	02:02
Air Blank		0.000	02:03
Control Test		0.080	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 01/29/2022 14:14

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEILL,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CHANEY,TONYA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: ELGHNIMI

DATE/TIME OF TEST: 01/23/2022 03:50:56

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.081	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.137	03:46
Breath Volume		2.421	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.134	03:49
Breath Volume		2.429	03:49
Air Blank		0.000	03:50
Control Test		0.080	03:50
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 01/29/2022 14:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,COLE,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: WEATHERMAN,BILLY
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: PETTEWAY

DATE/TIME OF TEST: 01/26/2022 04:06:31

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:03	
Control Test		0.081	04:03	
Air Blank		0.000	04:04	
Subject Sample	# 1		04:05	Subject Test Refused
Breath Volume		0.000	04:05	
Air Blank		0.000	04:05	
Control Test		0.080	04:05	
Air Blank		0.000	04:06	
Diagnostics Check		OK	04:06	

Data Download Date/Time: 01/29/2022 14:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: RANDOLPH

DATE/TIME OF TEST: 01/01/2022 01:45:00

Diagnostics Check		OK	01:38
Air Blank		0.000	01:39
Control Test		0.079	01:39
Air Blank		0.000	01:40
Subject Sample	# 1	0.226	01:40
Breath Volume		2.484	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.236	01:43
Breath Volume		2.187	01:43
Air Blank		0.000	01:44
Control Test		0.078	01:44
Air Blank		0.000	01:44
Diagnostics Check		OK	01:45

Data Download Date/Time: 01/19/2022 16:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: KUPLERSKI

DATE/TIME OF TEST: 01/01/2022 04:14:12

Diagnostics Check		OK	04:09	
Air Blank		0.000	04:10	
Control Test		0.079	04:10	
Air Blank		0.000	04:10	
Subject Sample	# 1		04:12	Subject Test Refused
Breath Volume		0.000	04:12	
Air Blank		0.000	04:13	
Control Test		0.078	04:13	
Air Blank		0.000	04:14	
Diagnostics Check		OK	04:14	

Data Download Date/Time: 01/19/2022 16:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/02/2022 02:19:30

Diagnostics Check		OK	02:16	
Air Blank		0.000	02:17	
Control Test		0.079	02:17	
Air Blank		0.000	02:17	
Subject Sample	# 1		02:18	Subject Test Refused
Breath Volume		0.000	02:18	
Air Blank		0.000	02:18	
Control Test		0.079	02:18	
Air Blank		0.000	02:19	
Diagnostics Check		OK	02:19	

Data Download Date/Time: 01/19/2022 16:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: SAMBWE

DATE/TIME OF TEST: 01/02/2022 03:18:06

Diagnostics Check		OK	03:04	
Air Blank		0.000	03:04	
Control Test		0.079	03:05	
Air Blank		0.000	03:05	
Subject Sample	# 1	0.236	03:07	
Breath Volume		1.832	03:07	
Air Blank		0.000	03:08	
Air Blank		0.000	03:10	
Subject Sample	# 2	0.240	03:13	Volume Not Met
Breath Volume		0.804	03:13	
Air Blank		0.000	03:14	
Air Blank		0.000	03:15	
Subject Sample	# 3	0.232	03:16	
Breath Volume		1.507	03:16	
Air Blank		0.000	03:17	
Control Test		0.075	03:17	
Air Blank		0.000	03:17	
Diagnostics Check		OK	03:18	

Data Download Date/Time: 01/19/2022 16:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 01/20/2022 03:14:44

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.079	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.129	03:10
Breath Volume		1.859	03:10
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.136	03:13
Breath Volume		2.003	03:13
Air Blank		0.000	03:13
Control Test		0.079	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 02/17/2022 15:28

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUFF,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HUFF,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MORROW

DATE/TIME OF TEST: 01/20/2022 22:13:33

Diagnostics Check		OK	22:10	
Air Blank		0.000	22:11	
Control Test		0.080	22:11	
Air Blank		0.000	22:12	
Subject Sample	# 1		22:12	Subject Test Refused
Breath Volume		0.000	22:12	
Air Blank		0.000	22:12	
Control Test		0.079	22:12	
Air Blank		0.000	22:13	
Diagnostics Check		OK	22:13	

Data Download Date/Time: 02/17/2022 15:28

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: DICKMANN

DATE/TIME OF TEST: 01/21/2022 02:56:01

Diagnostics Check		OK	02:53	
Air Blank		0.000	02:53	
Control Test		0.079	02:53	
Air Blank		0.000	02:54	
Subject Sample	# 1		02:54	Subject Test Refused
Breath Volume		0.000	02:54	
Air Blank		0.000	02:55	
Control Test		0.078	02:55	
Air Blank		0.000	02:55	
Diagnostics Check		OK	02:56	

Data Download Date/Time: 02/17/2022 15:28

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: MCSHAY

DATE/TIME OF TEST: 01/22/2022 03:57:25

Diagnostics Check		OK	03:54	
Air Blank		0.000	03:55	
Control Test		0.080	03:55	
Air Blank		0.000	03:55	
Subject Sample	# 1		03:56	Subject Test Refused
Breath Volume		0.000	03:56	
Air Blank		0.000	03:56	
Control Test		0.078	03:56	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 02/17/2022 15:28

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: DE LA CARIDAD ORTEGA

DATE/TIME OF TEST: 01/25/2022 23:35:21

Diagnostics Check		OK	23:32	
Air Blank		0.000	23:32	
Control Test		0.080	23:33	
Air Blank		0.000	23:33	
Subject Sample	# 1		23:33	Subject Test Refused
Breath Volume		0.000	23:33	
Air Blank		0.000	23:34	
Control Test		0.079	23:34	
Air Blank		0.000	23:35	
Diagnostics Check		OK	23:35	

Data Download Date/Time: 02/17/2022 15:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: PARKS

DATE/TIME OF TEST: 01/26/2022 18:49:33

Diagnostics Check		OK	18:46	
Air Blank		0.000	18:47	
Control Test		0.080	18:47	
Air Blank		0.000	18:47	
Subject Sample	# 1		18:48	Subject Test Refused
Breath Volume		0.000	18:48	
Air Blank		0.000	18:48	
Control Test		0.080	18:48	
Air Blank		0.000	18:49	
Diagnostics Check		OK	18:49	

Data Download Date/Time: 02/17/2022 15:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: WICKLINE

DATE/TIME OF TEST: 01/28/2022 03:10:42

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.080	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.192	03:05
Breath Volume		2.019	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.187	03:09
Breath Volume		2.308	03:09
Air Blank		0.000	03:09
Control Test		0.078	03:10
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 02/17/2022 15:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 01/28/2022 21:02:33

Diagnostics Check		OK	20:55
Air Blank		0.000	20:56
Control Test		0.080	20:56
Air Blank		0.000	20:56
Subject Sample	# 1	0.230	20:57
Breath Volume		2.070	20:57
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.223	21:00
Breath Volume		1.906	21:00
Air Blank		0.000	21:01
Control Test		0.077	21:01
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

Data Download Date/Time: 02/17/2022 15:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: GUITIERREZ

DATE/TIME OF TEST: 01/29/2022 04:12:48

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.080	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.163	04:08
Breath Volume		2.085	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:10
Subject Sample	# 2	0.160	04:11
Breath Volume		1.699	04:11
Air Blank		0.000	04:11
Control Test		0.080	04:12
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 02/17/2022 15:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REQUENA,CARLOS,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 01/31/2022 23:04:00

Diagnostics Check		OK	23:01	
Air Blank		0.000	23:01	
Control Test		0.081	23:01	
Air Blank		0.000	23:02	
Subject Sample	# 1		23:02	Subject Test Refused
Breath Volume		0.000	23:02	
Air Blank		0.000	23:03	
Control Test		0.081	23:03	
Air Blank		0.000	23:03	
Diagnostics Check		OK	23:04	

Data Download Date/Time: 02/17/2022 15:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JOHNSON,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/01/2022 01:28:58

Diagnostics Check		OK	01:22
Air Blank		0.000	01:22
Control Test		0.080	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.000	01:24
Breath Volume		2.117	01:24
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.000	01:27
Breath Volume		2.296	01:27
Air Blank		0.000	01:28
Control Test		0.080	01:28
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 01/31/2022 08:50

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK',
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SMITH,T
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: BUTCOSK

DATE/TIME OF TEST: 01/01/2022 02:29:38

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.080	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.156	02:25
Breath Volume		2.605	02:25
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.158	02:28
Breath Volume		2.136	02:28
Air Blank		0.000	02:28
Control Test		0.080	02:29
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 01/31/2022 08:50

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONN,CLARENCE,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WOELK,SGT
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: MORA-PARRA

DATE/TIME OF TEST: 01/01/2022 05:54:44

Diagnostics Check		OK	05:43	
Air Blank		0.000	05:44	
Control Test		0.080	05:44	
Air Blank		0.000	05:45	
Subject Sample	# 1	0.175	05:48	Volume Not Met
Breath Volume		0.882	05:48	
Air Blank		0.000	05:48	
Air Blank		0.000	05:50	
Subject Sample	# 2	0.158	05:53	Volume Not Met
Breath Volume		0.621	05:53	
Air Blank			05:54	Sequence Aborted
Air Blank		0.000	05:54	

Data Download Date/Time: 01/31/2022 08:50

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ANDERSON,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: GOUDREAU

DATE/TIME OF TEST: 01/02/2022 00:30:56

Diagnostics Check		OK	00:24
Air Blank		0.000	00:24
Control Test		0.081	00:24
Air Blank		0.000	00:25
Subject Sample	# 1	0.108	00:26
Breath Volume		2.074	00:26
Air Blank		0.000	00:26
Air Blank		0.000	00:28
Subject Sample	# 2	0.097	00:29
Breath Volume		1.125	00:29
Air Blank		0.000	00:30
Control Test		0.081	00:30
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 01/31/2022 08:50

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAMORA,DEP,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FEDUKE,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 01/02/2022 16:56:49

Diagnostics Check		OK	16:50
Air Blank		0.000	16:51
Control Test		0.081	16:51
Air Blank		0.000	16:51
Subject Sample	# 1	0.244	16:52
Breath Volume		1.730	16:52
Air Blank		0.000	16:52
Air Blank		0.000	16:54
Subject Sample	# 2	0.242	16:55
Breath Volume		2.093	16:55
Air Blank		0.000	16:55
Control Test		0.079	16:56
Air Blank		0.000	16:56
Diagnostics Check		OK	16:56

Data Download Date/Time: 01/31/2022 08:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONN,CLARENCE,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GANK,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:38

SUBJECT'S LAST NAME: GUEVARA

DATE/TIME OF TEST: 01/02/2022 18:05:43

Diagnostics Check		OK	17:58
Air Blank		0.000	17:59
Control Test		0.081	17:59
Air Blank		0.000	18:00
Subject Sample	# 1	0.009	18:00
Breath Volume		1.257	18:00
Air Blank		0.000	18:01
Air Blank		0.000	18:03
Subject Sample	# 2	0.007	18:04
Breath Volume		1.652	18:04
Air Blank		0.000	18:04
Control Test		0.081	18:05
Air Blank		0.000	18:05
Diagnostics Check		OK	18:05

Data Download Date/Time: 01/31/2022 08:50

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CCSO,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 01/03/2022 17:48:45

Diagnostics Check		OK	17:42
Air Blank		0.000	17:43
Control Test		0.078	17:43
Air Blank		0.000	17:44
Subject Sample	# 1	0.231	17:44
Breath Volume		3.101	17:44
Air Blank		0.000	17:45
Air Blank		0.000	17:46
Subject Sample	# 2	0.225	17:47
Breath Volume		2.820	17:47
Air Blank		0.000	17:47
Control Test		0.079	17:48
Air Blank		0.000	17:48
Diagnostics Check		OK	17:48

Data Download Date/Time: 01/31/2022 08:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENCIC,KENNETH,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,JONATHON
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: JOLLY

DATE/TIME OF TEST: 01/03/2022 23:17:42

Diagnostics Check		OK	23:10
Air Blank		0.000	23:10
Control Test		0.080	23:11
Air Blank		0.000	23:11
Subject Sample	# 1	0.241	23:12
Breath Volume		3.835	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.248	23:16
Breath Volume		2.230	23:16
Air Blank		0.000	23:16
Control Test		0.080	23:17
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 01/31/2022 08:51

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENCIC,KENNETH,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PENCIC,KENNETH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 01/04/2022 03:27:13

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.081	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.158	03:22
Breath Volume		2.632	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.171	03:25
Breath Volume		3.203	03:25
Air Blank		0.000	03:26
Control Test		0.080	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 01/31/2022 08:51

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DUNN,C
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 01/04/2022 07:50:40

Diagnostics Check		OK	07:44
Air Blank		0.000	07:45
Control Test		0.080	07:45
Air Blank		0.000	07:45
Subject Sample	# 1	0.000	07:46
Breath Volume		2.968	07:46
Air Blank		0.000	07:46
Air Blank		0.000	07:48
Subject Sample	# 2	0.000	07:49
Breath Volume		2.773	07:49
Air Blank		0.000	07:49
Control Test		0.079	07:50
Air Blank		0.000	07:50
Diagnostics Check		OK	07:50

Data Download Date/Time: 01/31/2022 08:51

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHANDLER,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 01/04/2022 09:26:13

Diagnostics Check		OK	09:20
Air Blank		0.000	09:20
Control Test		0.081	09:21
Air Blank		0.000	09:21
Subject Sample	# 1	0.050	09:21
Breath Volume		1.367	09:21
Air Blank		0.000	09:22
Air Blank		0.000	09:24
Subject Sample	# 2	0.051	09:24
Breath Volume		1.359	09:24
Air Blank		0.000	09:25
Control Test		0.080	09:25
Air Blank		0.000	09:26
Diagnostics Check		OK	09:26

Data Download Date/Time: 01/31/2022 08:51

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DUNN,D
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 01/04/2022 11:26:32

Diagnostics Check		OK	11:19
Air Blank		0.000	11:20
Control Test		0.081	11:20
Air Blank		0.000	11:20
Subject Sample	# 1	0.027	11:21
Breath Volume		1.113	11:21
Air Blank		0.000	11:22
Air Blank		0.000	11:24
Subject Sample	# 2	0.028	11:25
Breath Volume		2.425	11:25
Air Blank		0.000	11:25
Control Test		0.080	11:25
Air Blank		0.000	11:26
Diagnostics Check		OK	11:26

Data Download Date/Time: 01/31/2022 08:51

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DUNN,CDFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 16:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 01/04/2022 17:18:53

Diagnostics Check		OK	17:12
Air Blank		0.000	17:12
Control Test		0.080	17:13
Air Blank		0.000	17:13
Subject Sample	# 1	0.000	17:14
Breath Volume		1.941	17:14
Air Blank		0.000	17:14
Air Blank		0.000	17:16
Subject Sample	# 2	0.000	17:17
Breath Volume		1.886	17:17
Air Blank		0.000	17:17
Control Test		0.080	17:18
Air Blank		0.000	17:18
Diagnostics Check		OK	17:18

Data Download Date/Time: 01/31/2022 08:51

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENCIC,KENNETH,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PENCIC,KENNETH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 01/04/2022 21:33:36

Diagnostics Check		OK	21:27
Air Blank		0.000	21:28
Control Test		0.080	21:28
Air Blank		0.000	21:28
Subject Sample	# 1	0.000	21:29
Breath Volume		2.242	21:29
Air Blank		0.000	21:29
Air Blank		0.000	21:31
Subject Sample	# 2	0.000	21:32
Breath Volume		2.644	21:32
Air Blank		0.000	21:32
Control Test		0.080	21:33
Air Blank		0.000	21:33
Diagnostics Check		OK	21:33

Data Download Date/Time: 01/31/2022 08:51

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRITES,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CRITES,CHRISTOPHER
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 01/05/2022 04:01:14

Diagnostics Check		OK	03:54
Air Blank		0.000	03:54
Control Test		0.080	03:55
Air Blank		0.000	03:55
Subject Sample	# 1	0.193	03:56
Breath Volume		1.726	03:56
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.199	03:59
Breath Volume		3.597	03:59
Air Blank		0.000	04:00
Control Test		0.080	04:00
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 01/31/2022 08:51

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENCIC,KENNETH,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PENCIC,KENNETH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 01/05/2022 05:12:47

Diagnostics Check		OK	05:06
Air Blank		0.000	05:07
Control Test		0.081	05:07
Air Blank		0.000	05:08
Subject Sample	# 1	0.000	05:08
Breath Volume		2.933	05:08
Air Blank		0.000	05:08
Air Blank		0.000	05:10
Subject Sample	# 2	0.000	05:11
Breath Volume		2.324	05:11
Air Blank		0.000	05:11
Control Test		0.081	05:12
Air Blank		0.000	05:12
Diagnostics Check		OK	05:12

Data Download Date/Time: 01/31/2022 08:51

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,CELESTE,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CAMERON,CELESTE
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:01

SUBJECT'S LAST NAME: LUCCHESI

DATE/TIME OF TEST: 01/05/2022 08:31:25

Diagnostics Check		OK	08:25
Air Blank		0.000	08:25
Control Test		0.080	08:25
Air Blank		0.000	08:26
Subject Sample	# 1	0.111	08:27
Breath Volume		2.027	08:27
Air Blank		0.000	08:27
Air Blank		0.000	08:29
Subject Sample	# 2	0.112	08:29
Breath Volume		2.328	08:29
Air Blank		0.000	08:30
Control Test		0.080	08:30
Air Blank		0.000	08:31
Diagnostics Check		OK	08:31

Data Download Date/Time: 01/31/2022 08:51

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAMORA,DEP,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: IDK,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: LUCCHESI

DATE/TIME OF TEST: 01/05/2022 11:39:44

Diagnostics Check		OK	11:33
Air Blank		0.000	11:33
Control Test		0.080	11:34
Air Blank		0.000	11:34
Subject Sample	# 1	0.058	11:35
Breath Volume		3.101	11:35
Air Blank		0.000	11:35
Air Blank		0.000	11:37
Subject Sample	# 2	0.058	11:38
Breath Volume		4.011	11:38
Air Blank		0.000	11:38
Control Test		0.081	11:39
Air Blank		0.000	11:39
Diagnostics Check		OK	11:39

Data Download Date/Time: 01/31/2022 08:51

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CONANT,R
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: ALEIXO

DATE/TIME OF TEST: 01/05/2022 22:04:08

Diagnostics Check		OK	21:57
Air Blank		0.000	21:57
Control Test		0.080	21:57
Air Blank		0.000	21:58
Subject Sample	# 1	0.000	21:59
Breath Volume		2.503	21:59
Air Blank		0.000	21:59
Air Blank		0.000	22:01
Subject Sample	# 2	0.000	22:02
Breath Volume		2.039	22:02
Air Blank		0.000	22:03
Control Test		0.080	22:03
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 01/31/2022 08:51

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK',
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BALABAN,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: HEADLAM

DATE/TIME OF TEST: 01/06/2022 02:24:44

Diagnostics Check		OK	02:11	
Air Blank		0.000	02:12	
Control Test		0.081	02:12	
Air Blank		0.000	02:12	
Subject Sample	# 1	0.190	02:13	
Breath Volume		1.890	02:13	
Air Blank		0.000	02:14	
Air Blank		0.000	02:16	
Subject Sample	# 2	0.192	02:19	Volume Not Met
Breath Volume		0.167	02:19	
Air Blank		0.000	02:19	
Air Blank		0.000	02:21	
Subject Sample	# 3	0.196	02:23	
Breath Volume		1.449	02:23	
Air Blank		0.000	02:23	
Control Test		0.079	02:24	
Air Blank		0.000	02:24	
Diagnostics Check		OK	02:24	

Data Download Date/Time: 01/31/2022 08:51

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: NUNZIATO,C
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: SEARS

DATE/TIME OF TEST: 01/06/2022 04:19:26

Diagnostics Check		OK	04:06	
Air Blank		0.000	04:07	
Control Test		0.080	04:07	
Air Blank		0.000	04:08	
Subject Sample	# 1	0.185	04:11	Volume Not Met
Breath Volume		1.031	04:11	
Air Blank		0.000	04:11	
Air Blank		0.000	04:13	
Subject Sample	# 2	0.193	04:14	
Breath Volume		2.230	04:14	
Air Blank		0.000	04:15	
Air Blank		0.000	04:17	
Subject Sample	# 3	0.181	04:17	
Breath Volume		1.296	04:17	
Air Blank		0.000	04:18	
Control Test		0.081	04:18	
Air Blank		0.000	04:19	
Diagnostics Check		OK	04:19	

Data Download Date/Time: 01/31/2022 08:51

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENCIC,KENNETH,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PENCIC,KENNETH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 01/07/2022 19:17:50

Diagnostics Check		OK	19:11
Air Blank		0.000	19:12
Control Test		0.081	19:12
Air Blank		0.000	19:12
Subject Sample	# 1	0.063	19:13
Breath Volume		3.703	19:13
Air Blank		0.000	19:14
Air Blank		0.000	19:15
Subject Sample	# 2	0.061	19:16
Breath Volume		3.800	19:16
Air Blank		0.000	19:16
Control Test		0.080	19:17
Air Blank		0.000	19:17
Diagnostics Check		OK	19:17

Data Download Date/Time: 01/31/2022 08:51

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENCIC,KENNETH,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PENCIC,KENNETH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 01/08/2022 04:08:10

Diagnostics Check		OK	04:02
Air Blank		0.000	04:02
Control Test		0.081	04:02
Air Blank		0.000	04:03
Subject Sample	# 1	0.000	04:03
Breath Volume		2.937	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.000	04:06
Breath Volume		2.414	04:06
Air Blank		0.000	04:07
Control Test		0.080	04:07
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 01/31/2022 08:51

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KIMBERLYN,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: GAMBINO

DATE/TIME OF TEST: 01/08/2022 04:47:48

Diagnostics Check		OK	04:40
Air Blank		0.000	04:41
Control Test		0.080	04:41
Air Blank		0.000	04:41
Subject Sample	# 1	0.197	04:42
Breath Volume		2.425	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.194	04:46
Breath Volume		3.535	04:46
Air Blank		0.000	04:46
Control Test		0.079	04:47
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

Data Download Date/Time: 01/31/2022 08:52

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DUNN,CDFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 01/08/2022 08:34:31

Diagnostics Check		OK	08:28
Air Blank		0.000	08:28
Control Test		0.080	08:29
Air Blank		0.000	08:29
Subject Sample	# 1	0.000	08:30
Breath Volume		2.750	08:30
Air Blank		0.000	08:30
Air Blank		0.000	08:32
Subject Sample	# 2	0.000	08:33
Breath Volume		2.679	08:33
Air Blank		0.000	08:33
Control Test		0.080	08:33
Air Blank		0.000	08:34
Diagnostics Check		OK	08:34

Data Download Date/Time: 01/31/2022 08:52

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENCIC,KENNETH,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FRIEND,SEBASTIAN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: BURKARD

DATE/TIME OF TEST: 01/09/2022 04:22:21

Diagnostics Check		OK	04:19	
Air Blank		0.000	04:20	
Control Test		0.080	04:20	
Air Blank		0.000	04:20	
Subject Sample	# 1		04:21	Subject Test Refused
Breath Volume		0.000	04:21	
Air Blank		0.000	04:21	
Control Test		0.080	04:21	
Air Blank		0.000	04:22	
Diagnostics Check		OK	04:22	

Data Download Date/Time: 01/31/2022 08:52

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,BRADLEY,W
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CONANT,ROBERT
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:51

SUBJECT'S LAST NAME: SANCHEZ RUIZ

DATE/TIME OF TEST: 01/10/2022 17:19:27

Diagnostics Check		OK	17:13
Air Blank		0.000	17:13
Control Test		0.079	17:14
Air Blank		0.000	17:14
Subject Sample	# 1	0.000	17:15
Breath Volume		2.613	17:15
Air Blank		0.000	17:15
Air Blank		0.000	17:17
Subject Sample	# 2	0.000	17:18
Breath Volume		2.632	17:18
Air Blank		0.000	17:18
Control Test		0.080	17:18
Air Blank		0.000	17:19
Diagnostics Check		OK	17:19

Data Download Date/Time: 01/31/2022 08:52

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONN,CLARENCE,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LUPI,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: VINCENT

DATE/TIME OF TEST: 01/10/2022 21:23:25

Diagnostics Check		OK	21:15
Air Blank		0.000	21:16
Control Test		0.080	21:16
Air Blank		0.000	21:17
Subject Sample	# 1	0.052	21:18
Breath Volume		2.011	21:18
Air Blank		0.000	21:19
Air Blank		0.000	21:21
Subject Sample	# 2	0.052	21:21
Breath Volume		1.664	21:21
Air Blank		0.000	21:22
Control Test		0.079	21:22
Air Blank		0.000	21:23
Diagnostics Check		OK	21:23

Data Download Date/Time: 01/31/2022 08:52

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENCIC,KENNETH,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: THOMSEN

DATE/TIME OF TEST: 01/13/2022 03:41:09

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.080	03:35
Air Blank		0.000	03:35
Subject Sample	# 1	0.130	03:36
Breath Volume		1.609	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:38
Subject Sample	# 2	0.130	03:39
Breath Volume		1.898	03:39
Air Blank		0.000	03:40
Control Test		0.081	03:40
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 01/31/2022 08:52

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACEBO,ANDRES,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TEMS,N
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 01/15/2022 01:50:59

Diagnostics Check		OK	01:42
Air Blank		0.000	01:43
Control Test		0.081	01:43
Air Blank		0.000	01:43
Subject Sample	# 1	0.282	01:46
Breath Volume		1.566	01:46
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.299	01:49
Breath Volume		1.628	01:49
Air Blank		0.000	01:50
Control Test		0.079	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 01/31/2022 08:52

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MYRIE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: VALANTE

DATE/TIME OF TEST: 01/15/2022 02:49:31

Diagnostics Check		OK	02:42
Air Blank		0.000	02:42
Control Test		0.080	02:43
Air Blank		0.000	02:43
Subject Sample	# 1	0.187	02:44
Breath Volume		2.046	02:44
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.192	02:48
Breath Volume		1.660	02:48
Air Blank		0.000	02:48
Control Test		0.080	02:48
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 01/31/2022 08:52

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,CELESTE,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORA,R
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: SLOAN

DATE/TIME OF TEST: 01/15/2022 09:00:29

Diagnostics Check		OK	08:54
Air Blank		0.000	08:55
Control Test		0.079	08:55
Air Blank		0.000	08:55
Subject Sample	# 1	0.000	08:56
Breath Volume		1.781	08:56
Air Blank		0.000	08:56
Air Blank		0.000	08:58
Subject Sample	# 2	0.000	08:59
Breath Volume		1.343	08:59
Air Blank		0.000	08:59
Control Test		0.080	08:59
Air Blank		0.000	09:00
Diagnostics Check		OK	09:00

Data Download Date/Time: 01/31/2022 08:52

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MYRIE,K
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: DORSEY

DATE/TIME OF TEST: 01/17/2022 01:27:47

Diagnostics Check		OK	01:21
Air Blank		0.000	01:22
Control Test		0.079	01:22
Air Blank		0.000	01:22
Subject Sample	# 1	0.000	01:23
Breath Volume		2.085	01:23
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.000	01:26
Breath Volume		2.417	01:26
Air Blank		0.000	01:26
Control Test		0.080	01:27
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 01/31/2022 08:52

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ANDERSON,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/17/2022 20:47:56

Diagnostics Check		OK	20:41
Air Blank		0.000	20:41
Control Test		0.080	20:41
Air Blank		0.000	20:42
Subject Sample	# 1	0.074	20:43
Breath Volume		2.835	20:43
Air Blank		0.000	20:43
Air Blank		0.000	20:45
Subject Sample	# 2	0.073	20:46
Breath Volume		2.437	20:46
Air Blank		0.000	20:47
Control Test		0.080	20:47
Air Blank		0.000	20:47
Diagnostics Check		OK	20:47

Data Download Date/Time: 01/31/2022 08:52

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: LIQUORI

DATE/TIME OF TEST: 01/19/2022 00:20:00

Diagnostics Check		OK	00:13
Air Blank		0.000	00:13
Control Test		0.080	00:13
Air Blank		0.000	00:14
Subject Sample	# 1	0.108	00:15
Breath Volume		1.132	00:15
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.101	00:18
Breath Volume		1.812	00:18
Air Blank		0.000	00:19
Control Test		0.081	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:20

Data Download Date/Time: 01/31/2022 08:52

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTORFER,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: HADDOCK

DATE/TIME OF TEST: 01/19/2022 21:55:24

Diagnostics Check		OK	21:48
Air Blank		0.000	21:48
Control Test		0.080	21:49
Air Blank		0.000	21:49
Subject Sample	# 1	0.062	21:50
Breath Volume		1.707	21:50
Air Blank		0.000	21:50
Air Blank		0.000	21:52
Subject Sample	# 2	0.064	21:53
Breath Volume		1.968	21:53
Air Blank		0.000	21:54
Control Test		0.079	21:54
Air Blank		0.000	21:55
Diagnostics Check		OK	21:55

Data Download Date/Time: 01/31/2022 08:52

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CONANT,R
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/24/2022 19:16:16

Diagnostics Check		OK	19:04
Air Blank		0.000	19:04
Control Test		0.080	19:05
Air Blank		0.000	19:05
Subject Sample	# 1	0.267	19:06
Breath Volume		1.750	19:06
Air Blank		0.000	19:07
Air Blank		0.000	19:09
Subject Sample	# 2	0.242	19:10
Breath Volume		3.171	19:10
Air Blank		0.000	19:11
Air Blank		0.000	19:13
Subject Sample	# 3	0.258	19:14
Breath Volume		3.472	19:14
Air Blank		0.000	19:15
Control Test		0.078	19:15
Air Blank		0.000	19:16
Diagnostics Check		OK	19:16

Data Download Date/Time: 01/31/2022 08:52

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,ALAN,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ANDERSON,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: FURLEIGH

DATE/TIME OF TEST: 01/25/2022 17:08:53

Diagnostics Check		OK	17:02
Air Blank		0.000	17:03
Control Test		0.076	17:03
Air Blank		0.000	17:04
Subject Sample	# 1	0.270	17:04
Breath Volume		2.937	17:04
Air Blank		0.000	17:05
Air Blank		0.000	17:07
Subject Sample	# 2	0.259	17:07
Breath Volume		3.207	17:07
Air Blank		0.000	17:07
Control Test		0.078	17:08
Air Blank		0.000	17:08
Diagnostics Check		OK	17:08

Data Download Date/Time: 01/31/2022 08:52

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENCIC,KENNETH,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JANKOWSKI,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: VILLOPOTO

DATE/TIME OF TEST: 01/27/2022 22:45:01

Diagnostics Check		OK	22:38
Air Blank		0.000	22:39
Control Test		0.079	22:39
Air Blank		0.000	22:39
Subject Sample	# 1	0.000	22:40
Breath Volume		1.300	22:40
Air Blank		0.000	22:41
Air Blank		0.000	22:43
Subject Sample	# 2	0.000	22:43
Breath Volume		1.628	22:43
Air Blank		0.000	22:44
Control Test		0.079	22:44
Air Blank		0.000	22:44
Diagnostics Check		OK	22:45

Data Download Date/Time: 01/31/2022 08:53

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BRASCH,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: DRASE

DATE/TIME OF TEST: 01/29/2022 18:09:09

Diagnostics Check		OK	18:02
Air Blank		0.000	18:02
Control Test		0.080	18:02
Air Blank		0.000	18:03
Subject Sample	# 1	0.274	18:04
Breath Volume		3.023	18:04
Air Blank		0.000	18:05
Air Blank		0.000	18:07
Subject Sample	# 2	0.269	18:07
Breath Volume		2.511	18:07
Air Blank		0.000	18:08
Control Test		0.080	18:08
Air Blank		0.000	18:09
Diagnostics Check		OK	18:09

Data Download Date/Time: 01/31/2022 08:53

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CONTANT,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:29

SUBJECT'S LAST NAME: MILLION

DATE/TIME OF TEST: 01/30/2022 13:55:33

Diagnostics Check		OK	13:49
Air Blank		0.000	13:50
Control Test		0.080	13:50
Air Blank		0.000	13:50
Subject Sample	# 1	0.000	13:51
Breath Volume		3.109	13:51
Air Blank		0.000	13:51
Air Blank		0.000	13:53
Subject Sample	# 2	0.000	13:54
Breath Volume		3.187	13:54
Air Blank		0.000	13:54
Control Test		0.080	13:54
Air Blank		0.000	13:55
Diagnostics Check		OK	13:55

Data Download Date/Time: 01/31/2022 08:53

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONN,CLARENCE,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SMITH,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: DENLINGER

DATE/TIME OF TEST: 01/31/2022 02:15:18

Diagnostics Check		OK	02:08
Air Blank		0.000	02:08
Control Test		0.080	02:09
Air Blank		0.000	02:09
Subject Sample	# 1	0.000	02:10
Breath Volume		1.214	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.000	02:13
Breath Volume		1.621	02:13
Air Blank		0.000	02:14
Control Test		0.080	02:14
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 01/31/2022 08:53

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: COATES

DATE/TIME OF TEST: 01/01/2022 01:54:11

Diagnostics Check		OK	01:42	
Air Blank		0.000	01:43	
Control Test		0.077	01:43	
Air Blank		0.000	01:44	
Subject Sample	# 1		01:47	No Sample Provided
Breath Volume		0.000	01:47	
Air Blank		0.000	01:47	
Air Blank		0.000	01:49	
Subject Sample	# 2		01:52	No Sample Provided
Breath Volume		0.000	01:52	
Air Blank		0.000	01:53	
Control Test		0.077	01:53	
Air Blank		0.000	01:54	
Diagnostics Check		OK	01:54	

Data Download Date/Time: 01/20/2022 18:06

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIERS,S,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: STIERS,S
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: LEYENDECKER

DATE/TIME OF TEST: 01/01/2022 02:47:13

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:35	
Control Test		0.077	02:35	
Air Blank		0.000	02:36	
Subject Sample	# 1		02:39	No Sample Provided
Breath Volume		0.000	02:39	
Air Blank		0.000	02:39	
Air Blank		0.000	02:41	
Subject Sample	# 2	0.188	02:42	
Breath Volume		3.687	02:42	
Air Blank		0.000	02:43	
Air Blank		0.000	02:44	
Subject Sample	# 3	0.190	02:45	
Breath Volume		4.484	02:45	
Air Blank		0.000	02:46	
Control Test		0.077	02:46	
Air Blank		0.000	02:47	
Diagnostics Check		OK	02:47	

Data Download Date/Time: 01/20/2022 18:06

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,HENRY,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,HENRY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MELLO PASSOS

DATE/TIME OF TEST: 01/01/2022 03:13:23

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.078	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.062	03:08
Breath Volume		1.500	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:10
Subject Sample	# 2	0.065	03:11
Breath Volume		1.707	03:11
Air Blank		0.000	03:12
Control Test		0.077	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 01/20/2022 18:06

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: ANTELO-SUAREZ

DATE/TIME OF TEST: 01/01/2022 04:04:59

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.078	03:57
Air Blank		0.000	03:57
Subject Sample	# 1	0.120	03:59
Breath Volume		1.187	03:59
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.119	04:03
Breath Volume		2.222	04:03
Air Blank		0.000	04:04
Control Test		0.078	04:04
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 01/20/2022 18:06

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNSBROUGH,DREW,A
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: THORNSBROUGH,DREW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:14

SUBJECT'S LAST NAME: BOWERS

DATE/TIME OF TEST: 01/04/2022 18:36:03

Diagnostics Check		OK	18:29	
Air Blank		0.000	18:30	
Control Test		0.078	18:30	
Air Blank		0.000	18:31	
Subject Sample	# 1	0.173	18:31	
Breath Volume		2.953	18:31	
Air Blank		0.000	18:32	
Air Blank		0.000	18:34	
Subject Sample	# 2		18:34	Subject Test Refused
Breath Volume		0.000	18:34	
Air Blank		0.000	18:35	
Control Test		0.077	18:35	
Air Blank		0.000	18:35	
Diagnostics Check		OK	18:36	

Data Download Date/Time: 01/20/2022 18:06

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: COOPRIDER

DATE/TIME OF TEST: 01/08/2022 03:52:54

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.079	03:46
Air Blank		0.000	03:47
Subject Sample	# 1	0.203	03:48
Breath Volume		2.976	03:48
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.194	03:51
Breath Volume		1.625	03:51
Air Blank		0.000	03:51
Control Test		0.077	03:52
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 01/20/2022 18:07

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: ROTH

DATE/TIME OF TEST: 01/08/2022 21:12:36

Diagnostics Check		OK	21:09	
Air Blank		0.000	21:10	
Control Test		0.078	21:10	
Air Blank		0.000	21:10	
Subject Sample	# 1		21:11	Subject Test Refused
Breath Volume		0.000	21:11	
Air Blank		0.000	21:11	
Control Test		0.079	21:12	
Air Blank		0.000	21:12	
Diagnostics Check		OK	21:12	

Data Download Date/Time: 01/20/2022 18:07

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GOSIER,M
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: DUTTON

DATE/TIME OF TEST: 01/09/2022 20:55:51

Diagnostics Check		OK	20:48
Air Blank		0.000	20:49
Control Test		0.079	20:49
Air Blank		0.000	20:49
Subject Sample	# 1	0.216	20:51
Breath Volume		1.992	20:51
Air Blank		0.000	20:51
Air Blank		0.000	20:53
Subject Sample	# 2	0.211	20:54
Breath Volume		1.929	20:54
Air Blank		0.000	20:54
Control Test		0.077	20:55
Air Blank		0.000	20:55
Diagnostics Check		OK	20:55

Data Download Date/Time: 01/20/2022 18:07

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDMONDS,RICHARD,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REED,JOHNATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 01/11/2022 06:07:24

Diagnostics Check		OK	06:00
Air Blank		0.000	06:01
Control Test		0.078	06:01
Air Blank		0.000	06:02
Subject Sample	# 1	0.091	06:02
Breath Volume		1.937	06:02
Air Blank		0.000	06:03
Air Blank		0.000	06:05
Subject Sample	# 2	0.090	06:05
Breath Volume		2.242	06:05
Air Blank		0.000	06:06
Control Test		0.078	06:06
Air Blank		0.000	06:07
Diagnostics Check		OK	06:07

Data Download Date/Time: 01/20/2022 18:07

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIERS,S,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: STIERS,S
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 01/11/2022 09:38:51

Diagnostics Check		OK	09:31
Air Blank		0.000	09:32
Control Test		0.078	09:32
Air Blank		0.000	09:33
Subject Sample	# 1	0.132	09:34
Breath Volume		2.339	09:34
Air Blank		0.000	09:34
Air Blank		0.000	09:36
Subject Sample	# 2	0.132	09:37
Breath Volume		3.058	09:37
Air Blank		0.000	09:37
Control Test		0.077	09:38
Air Blank		0.000	09:38
Diagnostics Check		OK	09:38

Data Download Date/Time: 01/20/2022 18:07

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIE,LANGSTON,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GIANNAS,W
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/11/2022 20:20:06

Diagnostics Check		OK	20:13
Air Blank		0.000	20:13
Control Test		0.079	20:13
Air Blank		0.000	20:14
Subject Sample	# 1	0.203	20:15
Breath Volume		3.667	20:15
Air Blank		0.000	20:16
Air Blank		0.000	20:18
Subject Sample	# 2	0.211	20:18
Breath Volume		3.226	20:18
Air Blank		0.000	20:19
Control Test		0.076	20:19
Air Blank		0.000	20:20
Diagnostics Check		OK	20:20

Data Download Date/Time: 01/20/2022 18:07

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDMONDS,RICHARD,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: RACE,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/11/2022 23:27:26

Diagnostics Check		OK	23:20
Air Blank		0.000	23:21
Control Test		0.078	23:21
Air Blank		0.000	23:21
Subject Sample	# 1	0.305	23:22
Breath Volume		2.179	23:22
Air Blank		0.000	23:23
Air Blank		0.000	23:25
Subject Sample	# 2	0.309	23:25
Breath Volume		1.785	23:25
Air Blank		0.000	23:26
Control Test		0.075	23:26
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 01/20/2022 18:07

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIE,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: WOODIE,LANGSTON
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 01/12/2022 03:20:07

Diagnostics Check		OK	03:12
Air Blank		0.000	03:13
Control Test		0.079	03:13
Air Blank		0.000	03:14
Subject Sample	# 1	0.067	03:15
Breath Volume		3.406	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.064	03:18
Breath Volume		3.757	03:18
Air Blank		0.000	03:19
Control Test		0.077	03:19
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 01/20/2022 18:07

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: RANKIN

DATE/TIME OF TEST: 01/14/2022 19:08:31

Diagnostics Check		OK	19:01
Air Blank		0.000	19:01
Control Test		0.079	19:02
Air Blank		0.000	19:02
Subject Sample	# 1	0.000	19:03
Breath Volume		3.339	19:03
Air Blank		0.000	19:04
Air Blank		0.000	19:06
Subject Sample	# 2	0.000	19:07
Breath Volume		2.460	19:07
Air Blank		0.000	19:07
Control Test		0.079	19:07
Air Blank		0.000	19:08
Diagnostics Check		OK	19:08

Data Download Date/Time: 01/20/2022 18:07

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: PANTOJA

DATE/TIME OF TEST: 01/14/2022 19:39:14

Diagnostics Check		OK	19:32
Air Blank		0.000	19:32
Control Test		0.079	19:33
Air Blank		0.000	19:33
Subject Sample	# 1	0.215	19:34
Breath Volume		3.609	19:34
Air Blank		0.000	19:35
Air Blank		0.000	19:36
Subject Sample	# 2	0.217	19:37
Breath Volume		3.625	19:37
Air Blank		0.000	19:38
Control Test		0.076	19:38
Air Blank		0.000	19:39
Diagnostics Check		OK	19:39

Data Download Date/Time: 01/20/2022 18:07

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIE,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: WOODIE,LANGSTON
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: SPARLIN

DATE/TIME OF TEST: 01/15/2022 03:04:25

Diagnostics Check		OK	02:57
Air Blank		0.000	02:58
Control Test		0.079	02:58
Air Blank		0.000	02:59
Subject Sample	# 1	0.110	02:59
Breath Volume		1.785	02:59
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.117	03:02
Breath Volume		2.390	03:02
Air Blank		0.000	03:03
Control Test		0.077	03:03
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 01/20/2022 18:07

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 01/15/2022 19:32:11

Diagnostics Check		OK	19:25
Air Blank		0.000	19:26
Control Test		0.078	19:26
Air Blank		0.000	19:26
Subject Sample	# 1	0.000	19:27
Breath Volume		3.816	19:27
Air Blank		0.000	19:27
Air Blank		0.000	19:29
Subject Sample	# 2	0.000	19:30
Breath Volume		3.804	19:30
Air Blank		0.000	19:31
Control Test		0.079	19:31
Air Blank		0.000	19:32
Diagnostics Check		OK	19:32

Data Download Date/Time: 01/20/2022 18:07

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,J
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: ALSHEHRI

DATE/TIME OF TEST: 01/15/2022 22:53:02

Diagnostics Check		OK	22:46
Air Blank		0.000	22:46
Control Test		0.078	22:47
Air Blank		0.000	22:47
Subject Sample	# 1	0.165	22:48
Breath Volume		2.671	22:48
Air Blank		0.000	22:49
Air Blank		0.000	22:50
Subject Sample	# 2	0.165	22:51
Breath Volume		1.949	22:51
Air Blank		0.000	22:52
Control Test		0.078	22:52
Air Blank		0.000	22:52
Diagnostics Check		OK	22:53

Data Download Date/Time: 01/20/2022 18:07

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,LAMON,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REID,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: SPEED

DATE/TIME OF TEST: 01/19/2022 03:14:56

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.000	03:10
Breath Volume		1.304	03:10
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.000	03:13
Breath Volume		1.484	03:13
Air Blank		0.000	03:14
Control Test		0.078	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 01/20/2022 18:08

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: MORA RIVERA

DATE/TIME OF TEST: 01/20/2022 17:36:51

Diagnostics Check		OK	17:29
Air Blank		0.000	17:29
Control Test		0.078	17:30
Air Blank		0.000	17:30
Subject Sample	# 1	0.000	17:32
Breath Volume		1.539	17:32
Air Blank		0.000	17:32
Air Blank		0.000	17:34
Subject Sample	# 2	0.000	17:35
Breath Volume		3.472	17:35
Air Blank		0.000	17:35
Control Test		0.078	17:36
Air Blank		0.000	17:36
Diagnostics Check		OK	17:36

Data Download Date/Time: 01/20/2022 18:08

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,HENRY,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,HENRY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: WYNN

DATE/TIME OF TEST: 01/20/2022 23:10:26

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.078	23:02
Air Blank		0.000	23:03
Subject Sample	# 1	0.237	23:05
Breath Volume		1.492	23:05
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.234	23:08
Breath Volume		1.289	23:08
Air Blank		0.000	23:09
Control Test		0.077	23:09
Air Blank		0.000	23:10
Diagnostics Check		OK	23:10

Data Download Date/Time: 02/28/2022 09:57

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: DAY

DATE/TIME OF TEST: 01/21/2022 03:20:14

Diagnostics Check		OK	03:13
Air Blank		0.000	03:14
Control Test		0.078	03:14
Air Blank		0.000	03:15
Subject Sample	# 1	0.164	03:15
Breath Volume		1.945	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:17
Subject Sample	# 2	0.170	03:18
Breath Volume		1.667	03:18
Air Blank		0.000	03:19
Control Test		0.077	03:19
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 02/28/2022 09:57

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,OFFICER,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRUNEAU,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 01/21/2022 04:13:54

Diagnostics Check		OK	04:07
Air Blank		0.000	04:07
Control Test		0.077	04:07
Air Blank		0.000	04:08
Subject Sample	# 1	0.177	04:09
Breath Volume		3.402	04:09
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.176	04:12
Breath Volume		3.218	04:12
Air Blank		0.000	04:12
Control Test		0.077	04:13
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 02/28/2022 09:57

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: CHRISTOPHER

DATE/TIME OF TEST: 01/23/2022 02:09:35

Diagnostics Check		OK	01:58
Air Blank		0.000	01:59
Control Test		0.079	01:59
Air Blank		0.000	01:59
Subject Sample	# 1	0.158	02:03
Breath Volume		1.941	02:03
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.157	02:08
Breath Volume		1.652	02:08
Air Blank		0.000	02:08
Control Test		0.076	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 02/28/2022 09:57

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 01/23/2022 02:55:15

Diagnostics Check		OK	02:47
Air Blank		0.000	02:48
Control Test		0.077	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.108	02:50
Breath Volume		1.820	02:50
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.111	02:53
Breath Volume		1.777	02:53
Air Blank		0.000	02:54
Control Test		0.078	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 02/28/2022 09:58

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIARS,KELSEY,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VIARS,KELSEY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: BRADBURN

DATE/TIME OF TEST: 01/23/2022 03:58:19

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.079	03:50
Air Blank		0.000	03:51
Subject Sample	# 1	0.143	03:53
Breath Volume		1.632	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.136	03:56
Breath Volume		1.605	03:56
Air Blank		0.000	03:57
Control Test		0.075	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 02/28/2022 09:58

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HERNANDEZ,I
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: RODRIGUEZ-CORONA

DATE/TIME OF TEST: 01/27/2022 02:29:41

Diagnostics Check		OK	02:21
Air Blank		0.000	02:22
Control Test		0.078	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.034	02:25
Breath Volume		3.222	02:25
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.028	02:28
Breath Volume		1.523	02:28
Air Blank		0.000	02:28
Control Test		0.078	02:29
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 02/28/2022 09:58

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIE,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: WOODIE,LANGSTON
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: PINTO-VILLANUEVA

DATE/TIME OF TEST: 01/28/2022 21:28:34

Diagnostics Check		OK	21:22
Air Blank		0.000	21:22
Control Test		0.079	21:22
Air Blank		0.000	21:23
Subject Sample	# 1	0.151	21:23
Breath Volume		1.250	21:23
Air Blank		0.000	21:24
Air Blank		0.000	21:26
Subject Sample	# 2	0.147	21:26
Breath Volume		1.546	21:26
Air Blank		0.000	21:27
Control Test		0.079	21:27
Air Blank		0.000	21:28
Diagnostics Check		OK	21:28

Data Download Date/Time: 02/28/2022 09:58

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: NITSCHE

DATE/TIME OF TEST: 01/29/2022 01:11:30

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:00	
Control Test		0.078	01:00	
Air Blank		0.000	01:01	
Subject Sample	# 1		01:04	No Sample Provided
Breath Volume		0.000	01:04	
Air Blank		0.000	01:04	
Air Blank		0.000	01:06	
Subject Sample	# 2		01:10	No Sample Provided
Breath Volume		0.000	01:10	
Air Blank		0.000	01:10	
Control Test		0.078	01:10	
Air Blank		0.000	01:11	
Diagnostics Check		OK	01:11	

Data Download Date/Time: 02/28/2022 09:58

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: STOTS

DATE/TIME OF TEST: 01/29/2022 20:18:46

Diagnostics Check		OK	20:07
Air Blank		0.000	20:08
Control Test		0.079	20:08
Air Blank		0.000	20:08
Subject Sample	# 1	0.262	20:10
Breath Volume		2.703	20:10
Air Blank		0.000	20:10
Air Blank		0.000	20:12
Subject Sample	# 2	0.233	20:13
Breath Volume		1.363	20:13
Air Blank		0.000	20:14
Air Blank		0.000	20:16
Subject Sample	# 3	0.252	20:17
Breath Volume		2.640	20:17
Air Blank		0.000	20:17
Control Test		0.078	20:18
Air Blank		0.000	20:18
Diagnostics Check		OK	20:18

Data Download Date/Time: 02/28/2022 09:58

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: FABIN

DATE/TIME OF TEST: 01/29/2022 23:54:47

Diagnostics Check		OK	23:48
Air Blank		0.000	23:48
Control Test		0.079	23:49
Air Blank		0.000	23:49
Subject Sample	# 1	0.169	23:50
Breath Volume		2.277	23:50
Air Blank		0.000	23:50
Air Blank		0.000	23:52
Subject Sample	# 2	0.164	23:53
Breath Volume		1.589	23:53
Air Blank		0.000	23:53
Control Test		0.080	23:54
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 02/28/2022 09:58

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,PAUL,S
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GONZALES,PAUL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23621080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MUHAMMAD

DATE/TIME OF TEST: 01/30/2022 03:40:26

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.080	03:33
Air Blank		0.000	03:33
Subject Sample	# 1	0.119	03:35
Breath Volume		2.914	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.120	03:38
Breath Volume		2.925	03:38
Air Blank		0.000	03:39
Control Test		0.078	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 02/28/2022 09:58

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211943
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: CRISP

DATE/TIME OF TEST: 01/01/2022 02:35:04

Diagnostics Check		OK	02:26
Air Blank		0.000	02:27
Control Test		0.079	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.059	02:30
Breath Volume		1.468	02:30
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.061	02:33
Breath Volume		1.730	02:33
Air Blank		0.000	02:34
Control Test		0.078	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:35

Data Download Date/Time: 01/10/2022 13:51

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211943
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:32

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/01/2022 13:13:58

Diagnostics Check		OK	13:05
Air Blank		0.000	13:05
Control Test		0.080	13:05
Air Blank		0.000	13:06
Subject Sample	# 1	0.310	13:09
Breath Volume		1.191	13:09
Air Blank		0.000	13:10
Air Blank		0.000	13:11
Subject Sample	# 2	0.311	13:12
Breath Volume		1.300	13:12
Air Blank		0.000	13:13
Control Test		0.077	13:13
Air Blank		0.000	13:13
Diagnostics Check		OK	13:13

Data Download Date/Time: 01/10/2022 13:51

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211943
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: HORNER

DATE/TIME OF TEST: 01/04/2022 18:01:12

Diagnostics Check		OK	17:54
Air Blank		0.000	17:55
Control Test		0.079	17:55
Air Blank		0.000	17:56
Subject Sample	# 1	0.289	17:56
Breath Volume		2.500	17:56
Air Blank		0.000	17:57
Air Blank		0.000	17:59
Subject Sample	# 2	0.269	17:59
Breath Volume		1.335	17:59
Air Blank		0.000	18:00
Control Test		0.077	18:00
Air Blank		0.000	18:01
Diagnostics Check		OK	18:01

Data Download Date/Time: 01/10/2022 13:52

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211943
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: OVERSTREET

DATE/TIME OF TEST: 01/07/2022 22:04:40

Diagnostics Check		OK	21:58
Air Blank		0.000	21:58
Control Test		0.080	21:58
Air Blank		0.000	21:59
Subject Sample	# 1	0.263	21:59
Breath Volume		1.996	21:59
Air Blank		0.000	22:00
Air Blank		0.000	22:02
Subject Sample	# 2	0.265	22:03
Breath Volume		1.730	22:03
Air Blank		0.000	22:03
Control Test		0.079	22:04
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 01/10/2022 13:52

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211943
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: RAMOS PEREZ

DATE/TIME OF TEST: 01/09/2022 01:04:51

Diagnostics Check		OK	00:56	
Air Blank		0.000	00:56	
Control Test		0.080	00:56	
Air Blank		0.000	00:57	
Subject Sample	# 1	0.213	00:59	
Breath Volume		1.667	00:59	
Air Blank		0.000	01:00	
Air Blank		0.000	01:02	
Subject Sample	# 2	0.217	01:02	
Breath Volume		2.269	01:02	
Air Blank		0.000	01:03	
Control Test		0.072	01:04	Control Outside Tolerance
Air Blank		0.000	01:04	

Data Download Date/Time: 01/10/2022 13:52

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALMEIDA,CAMILO,J
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: ANTHONY,DEANGELO
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:53

SUBJECT'S LAST NAME: HALE

DATE/TIME OF TEST: 01/24/2022 10:26:50

Diagnostics Check		OK	10:20
Air Blank		0.000	10:20
Control Test		0.079	10:20
Air Blank		0.000	10:21
Subject Sample	# 1	0.000	10:22
Breath Volume		1.992	10:22
Air Blank		0.000	10:23
Air Blank		0.000	10:25
Subject Sample	# 2	0.000	10:25
Breath Volume		1.128	10:25
Air Blank		0.000	10:25
Control Test		0.079	10:26
Air Blank		0.000	10:26
Diagnostics Check		OK	10:26

Data Download Date/Time: 01/25/2022 14:55

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALMEIDA,CAMILO,J
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: BEHRENS,BRENT
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:17

SUBJECT'S LAST NAME: PATE

DATE/TIME OF TEST: 01/26/2022 12:59:21

Diagnostics Check		OK	12:53
Air Blank		0.000	12:53
Control Test		0.078	12:54
Air Blank		0.000	12:54
Subject Sample	# 1	0.000	12:55
Breath Volume		2.171	12:55
Air Blank		0.000	12:55
Air Blank		0.000	12:57
Subject Sample	# 2	0.000	12:58
Breath Volume		2.250	12:58
Air Blank		0.000	12:58
Control Test		0.079	12:58
Air Blank		0.000	12:59
Diagnostics Check		OK	12:59

Data Download Date/Time: 03/01/2022 16:13

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,SHANNON,N
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: O'CONNELL,SHANNON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 01/16/2022 02:31:19

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.080	02:25
Air Blank		0.000	02:25
Subject Sample	# 1	0.184	02:26
Breath Volume		1.359	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.182	02:29
Breath Volume		1.738	02:29
Air Blank		0.000	02:30
Control Test		0.080	02:30
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 01/29/2022 10:41

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAOLELLA,PATRICK,R
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: GEBBIA,SALVITORE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 01/03/2022 01:35:22

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:32	
Control Test		0.081	01:33	
Air Blank		0.000	01:33	
Subject Sample	# 1		01:33	Subject Test Refused
Breath Volume		0.000	01:33	
Air Blank		0.000	01:34	
Control Test		0.080	01:34	
Air Blank		0.000	01:35	
Diagnostics Check		OK	01:35	

Data Download Date/Time: 01/27/2022 08:17

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SHECKLES,M
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: PENDLETON

DATE/TIME OF TEST: 01/01/2022 02:08:42

Diagnostics Check		OK	02:01
Air Blank		0.000	02:02
Control Test		0.079	02:02
Air Blank		0.000	02:03
Subject Sample	# 1	0.225	02:04
Breath Volume		1.847	02:04
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.239	02:07
Breath Volume		1.417	02:07
Air Blank		0.000	02:07
Control Test		0.076	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 01/03/2022 13:17

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 01/05/2022 21:03:04

Diagnostics Check		OK	20:56
Air Blank		0.000	20:56
Control Test		0.079	20:57
Air Blank		0.000	20:57
Subject Sample	# 1	0.136	20:58
Breath Volume		3.992	20:58
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.130	21:01
Breath Volume		3.394	21:01
Air Blank		0.000	21:02
Control Test		0.079	21:02
Air Blank		0.000	21:02
Diagnostics Check		OK	21:03

Data Download Date/Time: 02/16/2022 09:28

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 01/12/2022 22:39:42

Diagnostics Check		OK	22:33
Air Blank		0.000	22:33
Control Test		0.079	22:33
Air Blank		0.000	22:34
Subject Sample	# 1	0.152	22:35
Breath Volume		2.480	22:35
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.150	22:38
Breath Volume		2.214	22:38
Air Blank		0.000	22:38
Control Test		0.080	22:39
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 02/16/2022 09:28

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,R,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: STEELE,R
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: CONLON

DATE/TIME OF TEST: 01/13/2022 00:47:45

Diagnostics Check		OK	00:39
Air Blank		0.000	00:40
Control Test		0.079	00:40
Air Blank		0.000	00:40
Subject Sample	# 1	0.286	00:41
Breath Volume		3.058	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:43
Subject Sample	# 2	0.280	00:46
Breath Volume		2.945	00:46
Air Blank		0.000	00:46
Control Test		0.079	00:47
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 02/16/2022 09:28

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: NEWTON

DATE/TIME OF TEST: 01/20/2022 23:44:41

Diagnostics Check		OK	23:35
Air Blank		0.000	23:36
Control Test		0.079	23:36
Air Blank		0.000	23:37
Subject Sample	# 1	0.178	23:39
Breath Volume		1.449	23:39
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.174	23:42
Breath Volume		1.406	23:42
Air Blank		0.000	23:43
Control Test		0.076	23:44
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 02/16/2022 09:28

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: THOMPSON,JEFFREY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: HOOVER

DATE/TIME OF TEST: 01/21/2022 00:35:42

Diagnostics Check		OK	00:24
Air Blank		0.000	00:25
Control Test		0.080	00:25
Air Blank		0.000	00:26
Subject Sample	# 1	0.283	00:27
Breath Volume		2.777	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:29
Subject Sample	# 2	0.248	00:30
Breath Volume		1.734	00:30
Air Blank		0.000	00:31
Air Blank		0.000	00:32
Subject Sample	# 3	0.265	00:33
Breath Volume		1.429	00:33
Air Blank		0.000	00:34
Control Test		0.076	00:35
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 02/16/2022 09:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 01/21/2022 23:27:48

Diagnostics Check		OK	23:21
Air Blank		0.000	23:22
Control Test		0.080	23:22
Air Blank		0.000	23:23
Subject Sample	# 1	0.000	23:23
Breath Volume		2.160	23:23
Air Blank		0.000	23:23
Air Blank		0.000	23:25
Subject Sample	# 2	0.000	23:26
Breath Volume		1.875	23:26
Air Blank		0.000	23:26
Control Test		0.080	23:27
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 02/16/2022 09:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,MARGARITA,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: GUASTO,NICHOLAS
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 12/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: TINOCO

DATE/TIME OF TEST: 01/01/2022 19:44:04

Diagnostics Check		OK	19:34
Air Blank		0.000	19:34
Control Test		0.080	19:35
Air Blank		0.000	19:35
Subject Sample	# 1	0.183	19:36
Breath Volume		2.089	19:36
Air Blank		0.000	19:37
Air Blank		0.000	19:38
Subject Sample	# 2	0.206	19:39
Breath Volume		2.417	19:39
Air Blank		0.000	19:40
Air Blank		0.000	19:41
Subject Sample	# 3	0.201	19:42
Breath Volume		1.343	19:42
Air Blank		0.000	19:43
Control Test		0.079	19:43
Air Blank		0.000	19:43
Diagnostics Check		OK	19:44

Data Download Date/Time: 01/26/2022 15:04

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,MARGARITA,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: SUAREZ,RONALD
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 12/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: LEMUS

DATE/TIME OF TEST: 01/02/2022 19:56:12

Diagnostics Check		OK	19:51	
Air Blank		0.000	19:51	
Control Test		0.079	19:51	
Air Blank		0.000	19:52	
Subject Sample	# 1		19:54	Subject Test Refused
Breath Volume		0.000	19:54	
Air Blank		0.000	19:55	
Control Test		0.078	19:55	
Air Blank		0.000	19:56	
Diagnostics Check		OK	19:56	

Data Download Date/Time: 01/26/2022 15:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOSA,JONATHAN,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: PICHS-PICADO,ALEXANDER
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 12/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: SALAMANCA

DATE/TIME OF TEST: 01/18/2022 19:58:30

Diagnostics Check		OK	19:42	
Air Blank		0.000	19:42	
Control Test		0.080	19:43	
Air Blank		0.000	19:43	
Subject Sample	# 1	0.189	19:47	Volume Not Met
Breath Volume		0.156	19:47	
Air Blank		0.000	19:47	
Air Blank		0.000	19:49	
Subject Sample	# 2	0.190	19:51	
Breath Volume		1.203	19:51	
Air Blank		0.000	19:52	
Air Blank		0.000	19:53	
Subject Sample	# 3	0.190	19:56	
Breath Volume		1.109	19:56	
Air Blank		0.000	19:57	
Control Test		0.077	19:57	
Air Blank		0.000	19:58	
Diagnostics Check		OK	19:58	

Data Download Date/Time: 01/26/2022 15:04

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE,DAVANO,L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: CONFESSORE,STEVE
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MANGUAL

DATE/TIME OF TEST: 01/29/2022 00:31:38

Diagnostics Check		OK	00:24
Air Blank		0.000	00:25
Control Test		0.080	00:25
Air Blank		0.000	00:26
Subject Sample	# 1	0.140	00:26
Breath Volume		3.519	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.141	00:29
Breath Volume		3.464	00:29
Air Blank		0.000	00:30
Control Test		0.080	00:30
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 02/24/2022 16:06

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE,DAVANO,L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: ALLEYNE,DAVANO
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 01/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: MENDOZA MONDRAGON

DATE/TIME OF TEST: 01/31/2022 02:43:37

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:41	
Control Test		0.081	02:41	
Air Blank		0.000	02:42	
Subject Sample	# 1		02:42	Subject Test Refused
Breath Volume		0.000	02:42	
Air Blank		0.000	02:42	
Control Test		0.079	02:42	
Air Blank		0.000	02:43	
Diagnostics Check		OK	02:43	

Data Download Date/Time: 02/24/2022 16:06

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN,WILLIAM,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MCLAUGHLIN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: CONROY

DATE/TIME OF TEST: 01/05/2022 23:12:45

Diagnostics Check		OK	23:05
Air Blank		0.000	23:06
Control Test		0.079	23:06
Air Blank		0.000	23:06
Subject Sample	# 1	0.254	23:07
Breath Volume		2.179	23:07
Air Blank		0.000	23:07
Air Blank		0.000	23:09
Subject Sample	# 2	0.255	23:11
Breath Volume		1.363	23:11
Air Blank		0.000	23:11
Control Test		0.079	23:12
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 01/07/2022 15:59

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,JESUS,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FIGUEROA,JESUS
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: DONALDSON

DATE/TIME OF TEST: 01/21/2022 04:29:32

Diagnostics Check		OK	04:23
Air Blank		0.000	04:23
Control Test		0.079	04:24
Air Blank		0.000	04:24
Subject Sample	# 1	0.258	04:25
Breath Volume		2.183	04:25
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.256	04:27
Breath Volume		1.984	04:27
Air Blank		0.000	04:28
Control Test		0.078	04:28
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 02/24/2022 18:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: BOODRAM

DATE/TIME OF TEST: 01/21/2022 05:51:17

Diagnostics Check		OK	05:39	
Air Blank		0.000	05:40	
Control Test		0.079	05:40	
Air Blank		0.000	05:41	
Subject Sample	# 1		05:44	No Sample Provided
Breath Volume		0.000	05:44	
Air Blank		0.000	05:44	
Air Blank		0.000	05:46	
Subject Sample	# 2		05:49	No Sample Provided
Breath Volume		0.000	05:49	
Air Blank		0.000	05:50	
Control Test		0.080	05:50	
Air Blank		0.000	05:51	
Diagnostics Check		OK	05:51	

Data Download Date/Time: 02/24/2022 18:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: MCNELLIS

DATE/TIME OF TEST: 01/28/2022 21:01:23

Diagnostics Check		OK	20:53
Air Blank		0.000	20:54
Control Test		0.080	20:54
Air Blank		0.000	20:55
Subject Sample	# 1	0.202	20:57
Breath Volume		1.589	20:57
Air Blank		0.000	20:57
Air Blank		0.000	20:59
Subject Sample	# 2	0.204	20:59
Breath Volume		2.531	20:59
Air Blank		0.000	21:00
Control Test		0.080	21:00
Air Blank		0.000	21:01
Diagnostics Check		OK	21:01

Data Download Date/Time: 02/24/2022 18:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: BOTH CAZUELA

DATE/TIME OF TEST: 01/29/2022 02:08:50

Diagnostics Check		OK	02:02
Air Blank		0.000	02:03
Control Test		0.080	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.213	02:04
Breath Volume		1.839	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:06
Subject Sample	# 2	0.214	02:07
Breath Volume		2.199	02:07
Air Blank		0.000	02:07
Control Test		0.080	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 02/24/2022 18:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STARLING,KELLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 01/29/2022 05:13:07

Diagnostics Check		OK	05:10	
Air Blank		0.000	05:10	
Control Test		0.080	05:11	
Air Blank		0.000	05:11	
Subject Sample	# 1		05:12	RFI Detect
Breath Volume		0.000	05:12	
Air Blank		0.000	05:12	
Control Test		0.080	05:12	
Air Blank			05:12	RFI Detect
Air Blank			05:13	RFI Detect

Data Download Date/Time: 02/24/2022 18:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STARLING,KELLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 01/29/2022 05:22:57

Diagnostics Check		OK	05:15
Air Blank		0.000	05:15
Control Test		0.079	05:16
Air Blank		0.000	05:16
Subject Sample	# 1	0.110	05:18
Breath Volume		2.007	05:18
Air Blank		0.000	05:18
Air Blank		0.000	05:20
Subject Sample	# 2	0.109	05:21
Breath Volume		2.003	05:21
Air Blank		0.000	05:22
Control Test		0.080	05:22
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 02/24/2022 18:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RODRIGUEZ,ALEXANDER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/29/2022 16:36:12

Diagnostics Check	OK	16:35	
Air Blank		16:35	RFI Detect
Air Blank		16:36	RFI Detect

Data Download Date/Time: 02/24/2022 18:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RODRIGUEZ,ALEXANDER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/29/2022 16:42:24

Diagnostics Check		OK	16:39	
Air Blank		0.000	16:39	
Control Test		0.081	16:40	
Air Blank		0.000	16:40	
Subject Sample	# 1		16:41	RFI Detect
Breath Volume		0.000	16:41	
Air Blank		0.000	16:41	
Control Test		0.081	16:41	
Air Blank		0.000	16:42	
Diagnostics Check		OK	16:42	

Data Download Date/Time: 02/24/2022 18:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RODRIGUEZ,ALEXANDER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 01/29/2022 22:06:31

Diagnostics Check		OK	22:04	
Air Blank		0.000	22:04	
Control Test		0.080	22:05	
Air Blank		0.000	22:05	
Subject Sample	# 1		22:05	RFI Detect
Breath Volume		0.000	22:05	
Air Blank			22:06	RFI Detect
Air Blank		0.000	22:06	

Data Download Date/Time: 02/24/2022 18:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STARLING,KELLEY,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STARLING,KELLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GALLOWAY

DATE/TIME OF TEST: 01/30/2022 04:07:43

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.080	04:02
Air Blank		0.000	04:02
Subject Sample	# 1	0.219	04:03
Breath Volume		2.644	04:03
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.208	04:06
Breath Volume		2.804	04:06
Air Blank		0.000	04:06
Control Test		0.080	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 02/24/2022 18:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDLER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GOFFENA,JOSHUA
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 09:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: HERMAN

DATE/TIME OF TEST: 01/30/2022 10:31:09

Diagnostics Check	OK	10:30	
Air Blank		10:30	RFI Detect
Air Blank	0.000	10:31	

Data Download Date/Time: 02/24/2022 18:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDLER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GOFFENA,JOSHUA
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 09:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: HERMAN

DATE/TIME OF TEST: 01/30/2022 10:41:28

Diagnostics Check		OK	10:34
Air Blank		0.000	10:34
Control Test		0.080	10:34
Air Blank		0.000	10:35
Subject Sample	# 1	0.241	10:35
Breath Volume		1.785	10:35
Air Blank		0.000	10:36
Air Blank		0.000	10:38
Subject Sample	# 2	0.222	10:39
Breath Volume		1.656	10:39
Air Blank		0.000	10:40
Control Test		0.079	10:40
Air Blank		0.000	10:41
Diagnostics Check		OK	10:41

Data Download Date/Time: 02/24/2022 18:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STARLING,KELLEY,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STARLNG,KELLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: HARRELSON

DATE/TIME OF TEST: 01/30/2022 21:19:13

Diagnostics Check		OK	21:05	
Air Blank		0.000	21:05	
Control Test		0.080	21:06	
Air Blank		0.000	21:06	
Subject Sample	# 1		21:09	No Sample Provided
Breath Volume		0.000	21:09	
Air Blank		0.000	21:10	
Air Blank		0.000	21:12	
Subject Sample	# 2	0.205	21:13	
Breath Volume		1.292	21:13	
Air Blank		0.000	21:13	
Air Blank		0.000	21:15	
Subject Sample	# 3	0.194	21:17	
Breath Volume		1.847	21:17	
Air Blank		0.000	21:18	
Control Test		0.080	21:18	
Air Blank		0.000	21:19	
Diagnostics Check		OK	21:19	

Data Download Date/Time: 02/24/2022 18:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMSON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:04

SUBJECT'S LAST NAME: MOROTTI

DATE/TIME OF TEST: 01/06/2022 17:31:30

Diagnostics Check		OK	17:24
Air Blank		0.000	17:25
Control Test		0.079	17:25
Air Blank		0.000	17:26
Subject Sample	# 1	0.000	17:26
Breath Volume		1.480	17:26
Air Blank		0.000	17:27
Air Blank		0.000	17:29
Subject Sample	# 2	0.000	17:30
Breath Volume		1.226	17:30
Air Blank		0.000	17:30
Control Test		0.079	17:30
Air Blank		0.000	17:31
Diagnostics Check		OK	17:31

Data Download Date/Time: 02/03/2022 13:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: TIUL COC

DATE/TIME OF TEST: 01/06/2022 20:19:34

Diagnostics Check		OK	20:08
Air Blank		0.000	20:08
Control Test		0.079	20:08
Air Blank		0.000	20:09
Subject Sample	# 1	0.189	20:10
Breath Volume		3.335	20:10
Air Blank		0.000	20:11
Air Blank		0.000	20:13
Subject Sample	# 2	0.159	20:14
Breath Volume		1.601	20:14
Air Blank		0.000	20:15
Air Blank		0.000	20:16
Subject Sample	# 3	0.182	20:17
Breath Volume		3.308	20:17
Air Blank		0.000	20:18
Control Test		0.078	20:18
Air Blank		0.000	20:19
Diagnostics Check		OK	20:19

Data Download Date/Time: 02/03/2022 13:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: COMM VEHIC
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: SAMMONS

DATE/TIME OF TEST: 01/07/2022 00:32:00

Diagnostics Check		OK	00:24
Air Blank		0.000	00:24
Control Test		0.079	00:25
Air Blank		0.000	00:25
Subject Sample	# 1	0.135	00:27
Breath Volume		2.425	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:29
Subject Sample	# 2	0.130	00:30
Breath Volume		2.140	00:30
Air Blank		0.000	00:31
Control Test		0.077	00:31
Air Blank		0.000	00:31
Diagnostics Check		OK	00:32

Data Download Date/Time: 02/03/2022 13:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: WALDEN

DATE/TIME OF TEST: 01/07/2022 02:35:18

Diagnostics Check		OK	02:23	
Air Blank		0.000	02:24	
Control Test		0.078	02:24	
Air Blank		0.000	02:24	
Subject Sample	# 1	0.158	02:28	Volume Not Met
Breath Volume		0.488	02:28	
Air Blank		0.000	02:28	
Air Blank		0.000	02:30	
Subject Sample	# 2	0.167	02:33	Volume Not Met
Breath Volume		1.015	02:33	
Air Blank		0.000	02:34	
Control Test		0.077	02:34	
Air Blank		0.000	02:35	
Diagnostics Check		OK	02:35	

Data Download Date/Time: 02/03/2022 13:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: WALDEN

DATE/TIME OF TEST: 01/07/2022 02:48:18

Diagnostics Check		OK	02:36	
Air Blank		0.000	02:37	
Control Test		0.077	02:37	
Air Blank		0.000	02:37	
Subject Sample	# 1	0.167	02:40	Volume Not Met
Breath Volume		0.593	02:40	
Air Blank		0.000	02:41	
Air Blank		0.000	02:43	
Subject Sample	# 2	0.162	02:46	Volume Not Met
Breath Volume		0.335	02:46	
Air Blank		0.000	02:47	
Control Test		0.076	02:47	
Air Blank		0.000	02:48	
Diagnostics Check		OK	02:48	

Data Download Date/Time: 02/03/2022 13:20

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: PROSS

DATE/TIME OF TEST: 01/07/2022 20:03:02

Diagnostics Check		OK	19:56
Air Blank		0.000	19:57
Control Test		0.079	19:57
Air Blank		0.000	19:58
Subject Sample	# 1	0.089	19:58
Breath Volume		1.765	19:58
Air Blank		0.000	19:59
Air Blank		0.000	20:01
Subject Sample	# 2	0.086	20:01
Breath Volume		1.324	20:01
Air Blank		0.000	20:02
Control Test		0.079	20:02
Air Blank		0.000	20:02
Diagnostics Check		OK	20:03

Data Download Date/Time: 02/03/2022 13:20

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: DOUGHERTY

DATE/TIME OF TEST: 01/08/2022 01:09:30

Diagnostics Check		OK	01:02
Air Blank		0.000	01:02
Control Test		0.080	01:02
Air Blank		0.000	01:03
Subject Sample	# 1	0.179	01:04
Breath Volume		1.570	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.177	01:07
Breath Volume		1.210	01:07
Air Blank		0.000	01:08
Control Test		0.077	01:08
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 02/03/2022 13:20

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:58

SUBJECT'S LAST NAME: HUGGINS 2

DATE/TIME OF TEST: 01/08/2022 07:25:20

Diagnostics Check		OK	07:18
Air Blank		0.000	07:19
Control Test		0.078	07:19
Air Blank		0.000	07:20
Subject Sample	# 1	0.121	07:20
Breath Volume		1.566	07:20
Air Blank		0.000	07:21
Air Blank		0.000	07:23
Subject Sample	# 2	0.128	07:23
Breath Volume		1.277	07:23
Air Blank		0.000	07:24
Control Test		0.078	07:24
Air Blank		0.000	07:25
Diagnostics Check		OK	07:25

Data Download Date/Time: 02/03/2022 13:20

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSEN,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: MCKINNEY

DATE/TIME OF TEST: 01/08/2022 23:55:46

Diagnostics Check		OK	23:48
Air Blank		0.000	23:48
Control Test		0.079	23:49
Air Blank		0.000	23:49
Subject Sample	# 1	0.161	23:51
Breath Volume		1.621	23:51
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 2	0.154	23:54
Breath Volume		1.238	23:54
Air Blank		0.000	23:54
Control Test		0.078	23:55
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 02/03/2022 13:20

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: LAKINS

DATE/TIME OF TEST: 01/09/2022 03:16:59

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.079	03:10
Air Blank		0.000	03:11
Subject Sample	# 1	0.204	03:12
Breath Volume		1.523	03:12
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.194	03:15
Breath Volume		2.222	03:15
Air Blank		0.000	03:16
Control Test		0.078	03:16
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 02/03/2022 13:20

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: SUMMERLINE

DATE/TIME OF TEST: 01/09/2022 04:00:35

Diagnostics Check		OK	03:53
Air Blank		0.000	03:54
Control Test		0.079	03:54
Air Blank		0.000	03:55
Subject Sample	# 1	0.119	03:56
Breath Volume		2.421	03:56
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.118	03:59
Breath Volume		1.773	03:59
Air Blank		0.000	03:59
Control Test		0.077	03:59
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 02/03/2022 13:20

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 01/09/2022 05:20:38

Diagnostics Check		OK	05:13
Air Blank		0.000	05:13
Control Test		0.079	05:14
Air Blank		0.000	05:14
Subject Sample	# 1	0.161	05:15
Breath Volume		1.406	05:15
Air Blank		0.000	05:16
Air Blank		0.000	05:18
Subject Sample	# 2	0.167	05:19
Breath Volume		1.535	05:19
Air Blank		0.000	05:19
Control Test		0.078	05:20
Air Blank		0.000	05:20
Diagnostics Check		OK	05:20

Data Download Date/Time: 02/03/2022 13:20

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MONROE,R
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:34

SUBJECT'S LAST NAME: PAZOS ROJAS

DATE/TIME OF TEST: 01/09/2022 08:01:29

Diagnostics Check		OK	07:54	
Air Blank		0.000	07:55	
Control Test		0.079	07:55	
Air Blank		0.000	07:56	
Subject Sample	# 1	0.072	07:56	
Breath Volume		1.121	07:56	
Air Blank		0.000	07:57	
Air Blank		0.000	07:58	
Subject Sample	# 2		07:59	RFI Detect
Breath Volume		0.093	07:59	
Air Blank		0.000	08:00	
Control Test		0.078	08:00	
Air Blank		0.000	08:01	
Diagnostics Check		OK	08:01	

Data Download Date/Time: 02/03/2022 13:20

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MONROE,R
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:34

SUBJECT'S LAST NAME: PAZOS ROJAS

DATE/TIME OF TEST: 01/09/2022 08:15:19

Diagnostics Check		OK	08:03	
Air Blank		0.000	08:04	
Control Test		0.079	08:04	
Air Blank		0.000	08:04	
Subject Sample	# 1	0.066	08:08	Volume Not Met
Breath Volume		1.042	08:08	
Air Blank		0.000	08:08	
Air Blank		0.000	08:10	
Subject Sample	# 2	0.061	08:13	Volume Not Met
Breath Volume		0.265	08:13	
Air Blank		0.000	08:14	
Control Test		0.078	08:14	
Air Blank		0.000	08:15	
Diagnostics Check		OK	08:15	

Data Download Date/Time: 02/03/2022 13:20

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MONROE,R
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:34

SUBJECT'S LAST NAME: PAZOS ROJAS

DATE/TIME OF TEST: 01/09/2022 08:26:44

Diagnostics Check		OK	08:17
Air Blank		0.000	08:18
Control Test		0.077	08:18
Air Blank		0.000	08:19
Subject Sample	# 1	0.062	08:20
Breath Volume		1.167	08:20
Air Blank		0.000	08:21
Air Blank		0.000	08:22
Subject Sample	# 2	0.062	08:25
Breath Volume		1.179	08:25
Air Blank		0.000	08:25
Control Test		0.079	08:26
Air Blank		0.000	08:26
Diagnostics Check		OK	08:26

Data Download Date/Time: 02/03/2022 13:21

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: NUNEZ NUNEZ

DATE/TIME OF TEST: 01/09/2022 22:02:48

Diagnostics Check		OK	21:56
Air Blank		0.000	21:56
Control Test		0.079	21:56
Air Blank		0.000	21:57
Subject Sample	# 1	0.203	21:58
Breath Volume		3.117	21:58
Air Blank		0.000	21:59
Air Blank		0.000	22:00
Subject Sample	# 2	0.198	22:01
Breath Volume		2.152	22:01
Air Blank		0.000	22:01
Control Test		0.078	22:02
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 02/03/2022 13:21

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/09/2022 23:42:17

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.079	23:36
Air Blank		0.000	23:36
Subject Sample	# 1	0.252	23:37
Breath Volume		1.347	23:37
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.263	23:40
Breath Volume		2.378	23:40
Air Blank		0.000	23:41
Control Test		0.078	23:41
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 02/03/2022 13:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 01/10/2022 01:20:09

Diagnostics Check		OK	01:07	
Air Blank		0.000	01:07	
Control Test		0.079	01:08	
Air Blank		0.000	01:08	
Subject Sample	# 1	0.123	01:11	Volume Not Met
Breath Volume		0.050	01:11	
Air Blank		0.000	01:12	
Air Blank		0.000	01:14	
Subject Sample	# 2	0.130	01:14	
Breath Volume		1.125	01:14	
Air Blank		0.000	01:15	
Air Blank		0.000	01:17	
Subject Sample	# 3	0.131	01:18	
Breath Volume		1.339	01:18	
Air Blank		0.000	01:19	
Control Test		0.078	01:19	
Air Blank		0.000	01:20	
Diagnostics Check		OK	01:20	

Data Download Date/Time: 02/03/2022 13:21

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: BERRY JR

DATE/TIME OF TEST: 01/10/2022 21:13:42

Diagnostics Check		OK	21:07
Air Blank		0.000	21:07
Control Test		0.080	21:08
Air Blank		0.000	21:08
Subject Sample	# 1	0.182	21:09
Breath Volume		2.773	21:09
Air Blank		0.000	21:09
Air Blank		0.000	21:11
Subject Sample	# 2	0.186	21:12
Breath Volume		2.882	21:12
Air Blank		0.000	21:12
Control Test		0.079	21:13
Air Blank		0.000	21:13
Diagnostics Check		OK	21:13

Data Download Date/Time: 02/03/2022 13:21

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: SANTOS ZAVALA

DATE/TIME OF TEST: 01/10/2022 22:15:15

Diagnostics Check		OK	22:07
Air Blank		0.000	22:08
Control Test		0.080	22:08
Air Blank		0.000	22:09
Subject Sample	# 1	0.235	22:10
Breath Volume		1.769	22:10
Air Blank		0.000	22:11
Air Blank		0.000	22:13
Subject Sample	# 2	0.222	22:13
Breath Volume		2.449	22:13
Air Blank		0.000	22:14
Control Test		0.078	22:14
Air Blank		0.000	22:15
Diagnostics Check		OK	22:15

Data Download Date/Time: 02/03/2022 13:21

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: JAEGER

DATE/TIME OF TEST: 01/13/2022 00:29:01

Diagnostics Check		OK	00:22
Air Blank		0.000	00:23
Control Test		0.079	00:23
Air Blank		0.000	00:24
Subject Sample	# 1	0.212	00:24
Breath Volume		2.664	00:24
Air Blank		0.000	00:25
Air Blank		0.000	00:26
Subject Sample	# 2	0.213	00:27
Breath Volume		2.699	00:27
Air Blank		0.000	00:28
Control Test		0.079	00:28
Air Blank		0.000	00:28
Diagnostics Check		OK	00:29

Data Download Date/Time: 02/03/2022 13:21

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 01/13/2022 03:46:41

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.079	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.117	03:42
Breath Volume		2.300	03:42
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.119	03:45
Breath Volume		2.382	03:45
Air Blank		0.000	03:45
Control Test		0.079	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 02/03/2022 13:22

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: HUTCHINS

DATE/TIME OF TEST: 01/13/2022 05:27:28

Diagnostics Check		OK	05:20
Air Blank		0.000	05:21
Control Test		0.080	05:21
Air Blank		0.000	05:22
Subject Sample	# 1	0.164	05:22
Breath Volume		2.929	05:22
Air Blank		0.000	05:23
Air Blank		0.000	05:25
Subject Sample	# 2	0.161	05:25
Breath Volume		2.343	05:25
Air Blank		0.000	05:26
Control Test		0.079	05:26
Air Blank		0.000	05:27
Diagnostics Check		OK	05:27

Data Download Date/Time: 02/03/2022 13:22

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROBERTS,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 01/13/2022 22:41:29

Diagnostics Check		OK	22:30	
Air Blank		0.000	22:30	
Control Test		0.079	22:30	
Air Blank		0.000	22:31	
Subject Sample	# 1		22:34	No Sample Provided
Breath Volume		0.000	22:34	
Air Blank		0.000	22:34	
Air Blank		0.000	22:36	
Subject Sample	# 2		22:40	No Sample Provided
Breath Volume		0.000	22:40	
Air Blank		0.000	22:40	
Control Test		0.079	22:40	
Air Blank		0.000	22:41	
Diagnostics Check		OK	22:41	

Data Download Date/Time: 02/03/2022 13:22

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BORG

DATE/TIME OF TEST: 01/14/2022 00:27:31

Diagnostics Check		OK	00:20
Air Blank		0.000	00:21
Control Test		0.079	00:21
Air Blank		0.000	00:22
Subject Sample	# 1	0.118	00:22
Breath Volume		4.222	00:22
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.111	00:25
Breath Volume		3.789	00:25
Air Blank		0.000	00:26
Control Test		0.078	00:26
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 02/03/2022 13:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: PRITCHETT

DATE/TIME OF TEST: 01/14/2022 20:54:13

Diagnostics Check		OK	20:47
Air Blank		0.000	20:48
Control Test		0.081	20:48
Air Blank		0.000	20:49
Subject Sample	# 1	0.117	20:49
Breath Volume		1.679	20:49
Air Blank		0.000	20:50
Air Blank		0.000	20:52
Subject Sample	# 2	0.117	20:52
Breath Volume		2.417	20:52
Air Blank		0.000	20:53
Control Test		0.080	20:53
Air Blank		0.000	20:54
Diagnostics Check		OK	20:54

Data Download Date/Time: 02/03/2022 13:22

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WINTER,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: DIGANGI

DATE/TIME OF TEST: 01/14/2022 23:28:11

Diagnostics Check		OK	23:21
Air Blank		0.000	23:22
Control Test		0.080	23:22
Air Blank		0.000	23:22
Subject Sample	# 1	0.000	23:23
Breath Volume		2.500	23:23
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.000	23:26
Breath Volume		2.031	23:26
Air Blank		0.000	23:27
Control Test		0.081	23:27
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 02/03/2022 13:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: SIERRA

DATE/TIME OF TEST: 01/15/2022 01:25:00

Diagnostics Check		OK	01:18
Air Blank		0.000	01:18
Control Test		0.080	01:19
Air Blank		0.000	01:19
Subject Sample	# 1	0.109	01:20
Breath Volume		3.718	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:22
Subject Sample	# 2	0.106	01:23
Breath Volume		1.527	01:23
Air Blank		0.000	01:24
Control Test		0.080	01:24
Air Blank		0.000	01:24
Diagnostics Check		OK	01:25

Data Download Date/Time: 02/03/2022 13:22

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: PELHAM

DATE/TIME OF TEST: 01/15/2022 03:40:02

Diagnostics Check		OK	03:28	
Air Blank		0.000	03:28	
Control Test		0.080	03:28	
Air Blank		0.000	03:29	
Subject Sample	# 1	0.141	03:32	Volume Not Met
Breath Volume		0.035	03:32	
Air Blank		0.000	03:33	
Air Blank		0.000	03:35	
Subject Sample	# 2	0.198	03:38	Volume Not Met
Breath Volume		0.117	03:38	
Air Blank		0.000	03:39	
Control Test		0.079	03:39	
Air Blank		0.000	03:39	
Diagnostics Check		OK	03:40	

Data Download Date/Time: 02/03/2022 13:22

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: MORO

DATE/TIME OF TEST: 01/15/2022 05:11:03

Diagnostics Check		OK	04:59	
Air Blank		0.000	04:59	
Control Test		0.079	05:00	
Air Blank		0.000	05:00	
Subject Sample	# 1		05:03	No Sample Provided
Breath Volume		0.000	05:03	
Air Blank		0.000	05:04	
Air Blank		0.000	05:06	
Subject Sample	# 2	0.177	05:09	Volume Not Met
Breath Volume		0.277	05:09	
Air Blank		0.000	05:10	
Control Test		0.080	05:10	
Air Blank		0.000	05:10	
Diagnostics Check		OK	05:11	

Data Download Date/Time: 02/03/2022 13:22

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: MORO

DATE/TIME OF TEST: 01/15/2022 05:20:55

Diagnostics Check		OK	05:13
Air Blank		0.000	05:13
Control Test		0.078	05:13
Air Blank		0.000	05:14
Subject Sample	# 1	0.190	05:15
Breath Volume		1.808	05:15
Air Blank		0.000	05:15
Air Blank		0.000	05:17
Subject Sample	# 2	0.191	05:19
Breath Volume		1.296	05:19
Air Blank		0.000	05:19
Control Test		0.078	05:20
Air Blank		0.000	05:20
Diagnostics Check		OK	05:20

Data Download Date/Time: 02/03/2022 13:22

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 01/15/2022 19:22:57

Diagnostics Check		OK	19:16
Air Blank		0.000	19:17
Control Test		0.079	19:17
Air Blank		0.000	19:17
Subject Sample	# 1	0.244	19:18
Breath Volume		2.371	19:18
Air Blank		0.000	19:19
Air Blank		0.000	19:20
Subject Sample	# 2	0.233	19:21
Breath Volume		2.507	19:21
Air Blank		0.000	19:22
Control Test		0.078	19:22
Air Blank		0.000	19:22
Diagnostics Check		OK	19:22

Data Download Date/Time: 02/03/2022 13:22

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: DONOHO

DATE/TIME OF TEST: 01/16/2022 02:00:38

Diagnostics Check		OK	01:54
Air Blank		0.000	01:54
Control Test		0.079	01:54
Air Blank		0.000	01:55
Subject Sample	# 1	0.314	01:55
Breath Volume		2.773	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.328	01:58
Breath Volume		2.601	01:58
Air Blank		0.000	01:59
Control Test		0.079	02:00
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

Data Download Date/Time: 02/03/2022 13:22

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: BOLANOS HERRERA

DATE/TIME OF TEST: 01/16/2022 02:31:19

Diagnostics Check		OK	02:19	
Air Blank		0.000	02:19	
Control Test		0.079	02:20	
Air Blank		0.000	02:20	
Subject Sample	# 1	0.210	02:23	Volume Not Met
Breath Volume		0.550	02:23	
Air Blank		0.000	02:24	
Air Blank		0.000	02:26	
Subject Sample	# 2	0.225	02:29	Volume Not Met
Breath Volume		0.054	02:29	
Air Blank		0.000	02:30	
Control Test		0.077	02:30	
Air Blank		0.000	02:31	
Diagnostics Check		OK	02:31	

Data Download Date/Time: 02/03/2022 13:22

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: DELAO

DATE/TIME OF TEST: 01/16/2022 02:59:18

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.077	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.132	02:54
Breath Volume		2.566	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.131	02:57
Breath Volume		2.519	02:57
Air Blank		0.000	02:58
Control Test		0.078	02:58
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 02/03/2022 13:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 01/16/2022 05:12:53

Diagnostics Check		OK	05:04	
Air Blank		0.000	05:04	
Control Test		0.079	05:05	
Air Blank		0.000	05:05	
Subject Sample	# 1		05:08	No Sample Provided
Breath Volume		0.000	05:08	
Air Blank		0.000	05:09	
Air Blank		0.000	05:11	
Subject Sample	# 2		05:11	Subject Test Refused
Breath Volume		0.000	05:11	
Air Blank		0.000	05:11	
Control Test		0.081	05:12	
Air Blank		0.000	05:12	
Diagnostics Check		OK	05:12	

Data Download Date/Time: 02/03/2022 13:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DURANCEAU,R
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: ESCALANTE MERIDA

DATE/TIME OF TEST: 01/16/2022 05:25:58

Diagnostics Check		OK	05:18
Air Blank		0.000	05:18
Control Test		0.079	05:19
Air Blank		0.000	05:19
Subject Sample	# 1	0.109	05:21
Breath Volume		1.343	05:21
Air Blank		0.000	05:22
Air Blank		0.000	05:23
Subject Sample	# 2	0.110	05:24
Breath Volume		1.609	05:24
Air Blank		0.000	05:25
Control Test		0.078	05:25
Air Blank		0.000	05:25
Diagnostics Check		OK	05:25

Data Download Date/Time: 02/03/2022 13:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHAW,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: MC CLOSKEY

DATE/TIME OF TEST: 01/16/2022 19:18:32

Diagnostics Check		OK	19:05	
Air Blank		0.000	19:06	
Control Test		0.080	19:06	
Air Blank		0.000	19:07	
Subject Sample	# 1		19:10	No Sample Provided
Breath Volume		0.000	19:10	
Air Blank		0.000	19:10	
Air Blank		0.000	19:12	
Subject Sample	# 2	0.167	19:13	
Breath Volume		2.269	19:13	
Air Blank		0.000	19:14	
Air Blank		0.000	19:16	
Subject Sample	# 3	0.161	19:16	
Breath Volume		1.890	19:16	
Air Blank		0.000	19:17	
Control Test		0.079	19:17	
Air Blank		0.000	19:18	
Diagnostics Check		OK	19:18	

Data Download Date/Time: 02/03/2022 13:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WINTER,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: LYN

DATE/TIME OF TEST: 01/16/2022 23:26:16

Diagnostics Check		OK	23:18
Air Blank		0.000	23:18
Control Test		0.080	23:19
Air Blank		0.000	23:19
Subject Sample	# 1	0.133	23:21
Breath Volume		2.914	23:21
Air Blank		0.000	23:21
Air Blank		0.000	23:23
Subject Sample	# 2	0.128	23:24
Breath Volume		2.207	23:24
Air Blank		0.000	23:25
Control Test		0.079	23:25
Air Blank		0.000	23:26
Diagnostics Check		OK	23:26

Data Download Date/Time: 02/03/2022 13:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: BARR III

DATE/TIME OF TEST: 01/16/2022 23:50:18

Diagnostics Check		OK	23:40
Air Blank		0.000	23:41
Control Test		0.081	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.123	23:43
Breath Volume		1.335	23:43
Air Blank		0.000	23:44
Air Blank		0.000	23:46
Subject Sample	# 2	0.130	23:48
Breath Volume		1.460	23:48
Air Blank		0.000	23:49
Control Test		0.079	23:49
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 02/03/2022 13:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 01/17/2022 03:39:31

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.080	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.194	03:35
Breath Volume		2.378	03:35
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.192	03:37
Breath Volume		2.535	03:37
Air Blank		0.000	03:38
Control Test		0.079	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 02/03/2022 13:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: SMITH JR

DATE/TIME OF TEST: 01/18/2022 02:52:50

Diagnostics Check		OK	02:46
Air Blank		0.000	02:46
Control Test		0.080	02:46
Air Blank		0.000	02:47
Subject Sample	# 1	0.248	02:48
Breath Volume		1.640	02:48
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.239	02:51
Breath Volume		1.484	02:51
Air Blank		0.000	02:51
Control Test		0.079	02:52
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 02/03/2022 13:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: HORNE

DATE/TIME OF TEST: 01/18/2022 03:40:04

Diagnostics Check		OK	03:33
Air Blank		0.000	03:34
Control Test		0.080	03:34
Air Blank		0.000	03:35
Subject Sample	# 1	0.248	03:35
Breath Volume		2.656	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:37
Subject Sample	# 2	0.254	03:38
Breath Volume		2.824	03:38
Air Blank		0.000	03:39
Control Test		0.079	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:40

Data Download Date/Time: 02/03/2022 13:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ARCHAMBEAU,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:26

SUBJECT'S LAST NAME: MARES

DATE/TIME OF TEST: 01/18/2022 14:54:11

Diagnostics Check		OK	14:46
Air Blank		0.000	14:47
Control Test		0.081	14:47
Air Blank		0.000	14:48
Subject Sample	# 1	0.238	14:48
Breath Volume		1.617	14:48
Air Blank		0.000	14:49
Air Blank		0.000	14:51
Subject Sample	# 2	0.228	14:52
Breath Volume		1.589	14:52
Air Blank		0.000	14:53
Control Test		0.080	14:53
Air Blank		0.000	14:54
Diagnostics Check		OK	14:54

Data Download Date/Time: 02/03/2022 13:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 01/19/2022 01:36:44

Diagnostics Check		OK	01:30
Air Blank		0.000	01:30
Control Test		0.081	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.101	01:32
Breath Volume		2.496	01:32
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 2	0.100	01:35
Breath Volume		2.011	01:35
Air Blank		0.000	01:35
Control Test		0.081	01:36
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 02/03/2022 13:24

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: LEWIS-HAUK

DATE/TIME OF TEST: 01/20/2022 02:59:35

Diagnostics Check		OK	02:52
Air Blank		0.000	02:53
Control Test		0.080	02:53
Air Blank		0.000	02:53
Subject Sample	# 1	0.000	02:55
Breath Volume		2.062	02:55
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.000	02:58
Breath Volume		1.375	02:58
Air Blank		0.000	02:58
Control Test		0.080	02:58
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 02/03/2022 13:24

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: CANDELARIA

DATE/TIME OF TEST: 01/21/2022 03:53:29

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:43	
Control Test		0.080	03:43	
Air Blank		0.000	03:44	
Subject Sample	# 1	0.143	03:47	Volume Not Met
Breath Volume		0.488	03:47	
Air Blank		0.000	03:48	
Air Blank		0.000	03:49	
Subject Sample	# 2	0.174	03:50	
Breath Volume		1.539	03:50	
Air Blank		0.000	03:51	
Air Blank		0.000	03:52	
Subject Sample	# 3		03:53	Sequence Aborted
Breath Volume		0.000	03:53	
Air Blank		0.000	03:53	

Data Download Date/Time: 02/03/2022 13:24

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 01/21/2022 05:15:20

Diagnostics Check		OK	05:08
Air Blank		0.000	05:09
Control Test		0.079	05:09
Air Blank		0.000	05:10
Subject Sample	# 1	0.161	05:10
Breath Volume		2.152	05:10
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.163	05:13
Breath Volume		2.101	05:13
Air Blank		0.000	05:14
Control Test		0.079	05:14
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

Data Download Date/Time: 02/03/2022 13:24

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: MILLINGTON

DATE/TIME OF TEST: 01/21/2022 23:39:50

Diagnostics Check		OK	23:25	
Air Blank		0.000	23:26	
Control Test		0.079	23:26	
Air Blank		0.000	23:27	
Subject Sample	# 1	0.171	23:29	
Breath Volume		1.148	23:29	
Air Blank		0.000	23:30	
Air Blank		0.000	23:32	
Subject Sample	# 2	0.131	23:35	Volume Not Met
Breath Volume		0.699	23:35	
Air Blank		0.000	23:36	
Air Blank		0.000	23:37	
Subject Sample	# 3	0.157	23:38	
Breath Volume		1.753	23:38	
Air Blank		0.000	23:38	
Control Test		0.076	23:39	
Air Blank		0.000	23:39	
Diagnostics Check		OK	23:39	

Data Download Date/Time: 02/03/2022 13:24

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: NORTON

DATE/TIME OF TEST: 01/22/2022 01:18:12

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.079	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.172	01:13
Breath Volume		1.203	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.174	01:16
Breath Volume		1.632	01:16
Air Blank		0.000	01:17
Control Test		0.078	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 02/03/2022 13:24

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: RAHUBA

DATE/TIME OF TEST: 01/22/2022 03:34:42

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.079	03:28
Air Blank		0.000	03:29
Subject Sample	# 1	0.178	03:29
Breath Volume		1.750	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.176	03:33
Breath Volume		2.429	03:33
Air Blank		0.000	03:33
Control Test		0.078	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 02/03/2022 13:24

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNYDER,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: GROGAN

DATE/TIME OF TEST: 01/22/2022 06:16:29

Diagnostics Check		OK	06:00	
Air Blank		0.000	06:01	
Control Test		0.080	06:01	
Air Blank		0.000	06:01	
Subject Sample	# 1	0.157	06:05	Volume Not Met
Breath Volume		0.660	06:05	
Air Blank		0.000	06:05	
Air Blank		0.000	06:07	
Subject Sample	# 2	0.182	06:09	
Breath Volume		1.390	06:09	
Air Blank		0.000	06:09	
Air Blank		0.000	06:11	
Subject Sample	# 3	0.163	06:14	Volume Not Met
Breath Volume		0.058	06:14	
Air Blank		0.000	06:15	
Control Test		0.077	06:15	
Air Blank		0.000	06:16	
Diagnostics Check		OK	06:16	

Data Download Date/Time: 02/03/2022 13:24

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: RINCHUSE III

DATE/TIME OF TEST: 01/22/2022 21:29:26

Diagnostics Check		OK	21:22
Air Blank		0.000	21:23
Control Test		0.079	21:23
Air Blank		0.000	21:24
Subject Sample	# 1	0.191	21:24
Breath Volume		1.523	21:24
Air Blank		0.000	21:25
Air Blank		0.000	21:27
Subject Sample	# 2	0.198	21:27
Breath Volume		1.687	21:27
Air Blank		0.000	21:28
Control Test		0.079	21:28
Air Blank		0.000	21:29
Diagnostics Check		OK	21:29

Data Download Date/Time: 02/03/2022 13:24

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: ROSS JR

DATE/TIME OF TEST: 01/22/2022 23:20:25

Diagnostics Check		OK	23:13
Air Blank		0.000	23:14
Control Test		0.080	23:14
Air Blank		0.000	23:15
Subject Sample	# 1	0.181	23:15
Breath Volume		1.464	23:15
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.178	23:18
Breath Volume		1.613	23:18
Air Blank		0.000	23:19
Control Test		0.079	23:19
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 02/03/2022 13:24

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: HUMMEL

DATE/TIME OF TEST: 01/23/2022 05:19:03

Diagnostics Check		OK	05:12
Air Blank		0.000	05:12
Control Test		0.080	05:13
Air Blank		0.000	05:13
Subject Sample	# 1	0.132	05:14
Breath Volume		2.664	05:14
Air Blank		0.000	05:15
Air Blank		0.000	05:16
Subject Sample	# 2	0.135	05:17
Breath Volume		2.105	05:17
Air Blank		0.000	05:18
Control Test		0.079	05:18
Air Blank		0.000	05:18
Diagnostics Check		OK	05:19

Data Download Date/Time: 02/03/2022 13:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: OLADAPO

DATE/TIME OF TEST: 01/24/2022 03:44:36

Diagnostics Check		OK	03:32	
Air Blank		0.000	03:32	
Control Test		0.080	03:33	
Air Blank		0.000	03:33	
Subject Sample	# 1	0.166	03:36	Volume Not Met
Breath Volume		0.484	03:36	
Air Blank		0.000	03:37	
Air Blank		0.000	03:39	
Subject Sample	# 2	0.164	03:40	
Breath Volume		4.730	03:40	
Air Blank		0.000	03:40	
Air Blank		0.000	03:42	
Subject Sample	# 3	0.151	03:42	
Breath Volume		2.503	03:42	
Air Blank		0.000	03:43	
Control Test		0.077	03:43	
Air Blank		0.000	03:44	
Diagnostics Check		OK	03:44	

Data Download Date/Time: 02/03/2022 13:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: WILCOX

DATE/TIME OF TEST: 01/26/2022 03:00:12

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.079	02:54
Air Blank		0.000	02:54
Subject Sample	# 1	0.214	02:55
Breath Volume		2.652	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:57
Subject Sample	# 2	0.222	02:58
Breath Volume		2.769	02:58
Air Blank		0.000	02:59
Control Test		0.079	02:59
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 02/03/2022 13:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ORTIZ,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 01/27/2022 21:32:11

Diagnostics Check		OK	21:25
Air Blank		0.000	21:26
Control Test		0.080	21:26
Air Blank		0.000	21:27
Subject Sample	# 1	0.081	21:27
Breath Volume		1.917	21:27
Air Blank		0.000	21:28
Air Blank		0.000	21:30
Subject Sample	# 2	0.083	21:30
Breath Volume		1.832	21:30
Air Blank		0.000	21:31
Control Test		0.079	21:31
Air Blank		0.000	21:32
Diagnostics Check		OK	21:32

Data Download Date/Time: 02/03/2022 13:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: KRAUSE

DATE/TIME OF TEST: 01/28/2022 03:58:53

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.224	03:53
Breath Volume		1.238	03:53
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.215	03:57
Breath Volume		1.484	03:57
Air Blank		0.000	03:57
Control Test		0.078	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 02/03/2022 13:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: RAMIREZ RUFINO

DATE/TIME OF TEST: 01/29/2022 00:54:23

Diagnostics Check		OK	00:47
Air Blank		0.000	00:48
Control Test		0.080	00:48
Air Blank		0.000	00:48
Subject Sample	# 1	0.162	00:49
Breath Volume		3.640	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.157	00:52
Breath Volume		3.117	00:52
Air Blank		0.000	00:53
Control Test		0.079	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 02/03/2022 13:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: GERMANO-SOUZA

DATE/TIME OF TEST: 01/29/2022 22:21:02

Diagnostics Check		OK	22:13
Air Blank		0.000	22:14
Control Test		0.080	22:14
Air Blank		0.000	22:14
Subject Sample	# 1	0.177	22:16
Breath Volume		1.914	22:16
Air Blank		0.000	22:17
Air Blank		0.000	22:18
Subject Sample	# 2	0.175	22:19
Breath Volume		1.128	22:19
Air Blank		0.000	22:20
Control Test		0.079	22:20
Air Blank		0.000	22:20
Diagnostics Check		OK	22:21

Data Download Date/Time: 02/03/2022 13:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DURANCEAU,R
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: DE LA CRUZ

DATE/TIME OF TEST: 01/30/2022 00:53:48

Diagnostics Check		OK	00:47
Air Blank		0.000	00:47
Control Test		0.080	00:47
Air Blank		0.000	00:48
Subject Sample	# 1	0.249	00:49
Breath Volume		3.207	00:49
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.250	00:52
Breath Volume		2.828	00:52
Air Blank		0.000	00:52
Control Test		0.079	00:53
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 02/03/2022 13:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: GERGERES

DATE/TIME OF TEST: 01/30/2022 02:22:37

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.080	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.186	02:17
Breath Volume		2.531	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.175	02:21
Breath Volume		2.562	02:21
Air Blank		0.000	02:21
Control Test		0.079	02:22
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 02/03/2022 13:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:11

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 01/30/2022 06:35:24

Diagnostics Check		OK	06:31	
Air Blank		0.000	06:32	
Control Test		0.081	06:32	
Air Blank		0.000	06:32	
Subject Sample	# 1		06:34	Subject Test Refused
Breath Volume		0.000	06:34	
Air Blank		0.000	06:34	
Control Test		0.081	06:34	
Air Blank		0.000	06:35	
Diagnostics Check		OK	06:35	

Data Download Date/Time: 02/03/2022 13:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 01/31/2022 00:20:22

Diagnostics Check		OK	00:12
Air Blank		0.000	00:13
Control Test		0.080	00:13
Air Blank		0.000	00:14
Subject Sample	# 1	0.129	00:15
Breath Volume		2.089	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.126	00:18
Breath Volume		2.710	00:18
Air Blank		0.000	00:19
Control Test		0.079	00:19
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 02/03/2022 13:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: MELECIO

DATE/TIME OF TEST: 01/31/2022 03:45:44

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.081	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.220	03:41
Breath Volume		1.246	03:41
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.214	03:44
Breath Volume		1.757	03:44
Air Blank		0.000	03:44
Control Test		0.080	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 02/03/2022 13:25

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: CARCAMO-POSADAS

DATE/TIME OF TEST: 01/31/2022 23:01:40

Diagnostics Check		OK	22:53
Air Blank		0.000	22:54
Control Test		0.081	22:54
Air Blank		0.000	22:55
Subject Sample	# 1	0.185	22:57
Breath Volume		1.765	22:57
Air Blank		0.000	22:57
Air Blank		0.000	22:59
Subject Sample	# 2	0.193	23:00
Breath Volume		1.359	23:00
Air Blank		0.000	23:00
Control Test		0.078	23:01
Air Blank		0.000	23:01
Diagnostics Check		OK	23:01

Data Download Date/Time: 02/03/2022 13:26

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: HARTNETT

DATE/TIME OF TEST: 01/07/2022 01:03:58

Diagnostics Check		OK	00:57
Air Blank		0.000	00:58
Control Test		0.080	00:58
Air Blank		0.000	00:58
Subject Sample	# 1	0.000	00:59
Breath Volume		3.054	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.000	01:02
Breath Volume		2.984	01:02
Air Blank		0.000	01:03
Control Test		0.079	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 02/03/2022 11:27

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: COMM VEHIC
ARREST TIME: 04:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: WEYER

DATE/TIME OF TEST: 01/07/2022 05:20:40

Diagnostics Check		OK	05:12
Air Blank		0.000	05:13
Control Test		0.080	05:13
Air Blank		0.000	05:14
Subject Sample	# 1	0.130	05:16
Breath Volume		1.640	05:16
Air Blank		0.000	05:16
Air Blank		0.000	05:18
Subject Sample	# 2	0.130	05:19
Breath Volume		1.453	05:19
Air Blank		0.000	05:19
Control Test		0.079	05:20
Air Blank		0.000	05:20
Diagnostics Check		OK	05:20

Data Download Date/Time: 02/03/2022 11:27

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 01/08/2022 01:22:37

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.081	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.179	01:17
Breath Volume		1.496	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.184	01:21
Breath Volume		1.656	01:21
Air Blank		0.000	01:21
Control Test		0.079	01:22
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 02/03/2022 11:27

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 01/08/2022 05:40:20

Diagnostics Check		OK	05:33
Air Blank		0.000	05:34
Control Test		0.079	05:34
Air Blank		0.000	05:34
Subject Sample	# 1	0.173	05:35
Breath Volume		3.378	05:35
Air Blank		0.000	05:36
Air Blank		0.000	05:38
Subject Sample	# 2	0.177	05:38
Breath Volume		3.894	05:38
Air Blank		0.000	05:39
Control Test		0.080	05:39
Air Blank		0.000	05:40
Diagnostics Check		OK	05:40

Data Download Date/Time: 02/03/2022 11:28

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: TAUSCHER

DATE/TIME OF TEST: 01/09/2022 00:36:26

Diagnostics Check		OK	00:30
Air Blank		0.000	00:30
Control Test		0.080	00:30
Air Blank		0.000	00:31
Subject Sample	# 1	0.000	00:31
Breath Volume		3.828	00:31
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.000	00:35
Breath Volume		3.679	00:35
Air Blank		0.000	00:35
Control Test		0.080	00:35
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 02/03/2022 11:28

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: SIMKINS

DATE/TIME OF TEST: 01/09/2022 00:54:25

Diagnostics Check		OK	00:44	
Air Blank		0.000	00:45	
Control Test		0.080	00:45	
Air Blank		0.000	00:46	
Subject Sample	# 1		00:49	No Sample Provided
Breath Volume		0.000	00:49	
Air Blank		0.000	00:49	
Air Blank		0.000	00:51	
Subject Sample	# 2		00:53	Subject Test Refused
Breath Volume		0.000	00:53	
Air Blank		0.000	00:53	
Control Test		0.081	00:53	
Air Blank		0.000	00:54	
Diagnostics Check		OK	00:54	

Data Download Date/Time: 02/03/2022 11:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:07

SUBJECT'S LAST NAME: TIRADO

DATE/TIME OF TEST: 01/09/2022 16:33:38

Diagnostics Check		OK	16:27
Air Blank		0.000	16:28
Control Test		0.079	16:28
Air Blank		0.000	16:28
Subject Sample	# 1	0.037	16:29
Breath Volume		1.960	16:29
Air Blank		0.000	16:29
Air Blank		0.000	16:31
Subject Sample	# 2	0.038	16:32
Breath Volume		2.007	16:32
Air Blank		0.000	16:32
Control Test		0.080	16:33
Air Blank		0.000	16:33
Diagnostics Check		OK	16:33

Data Download Date/Time: 02/03/2022 11:28

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: MENDOZA-SOTO

DATE/TIME OF TEST: 01/09/2022 22:07:57

Diagnostics Check		OK	22:00
Air Blank		0.000	22:01
Control Test		0.080	22:01
Air Blank		0.000	22:02
Subject Sample	# 1	0.115	22:03
Breath Volume		1.957	22:03
Air Blank		0.000	22:04
Air Blank		0.000	22:05
Subject Sample	# 2	0.111	22:06
Breath Volume		1.789	22:06
Air Blank		0.000	22:07
Control Test		0.079	22:07
Air Blank		0.000	22:07
Diagnostics Check		OK	22:07

Data Download Date/Time: 02/03/2022 11:28

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: VAZQUEZ JR

DATE/TIME OF TEST: 01/10/2022 20:34:00

Diagnostics Check		OK	20:26
Air Blank		0.000	20:27
Control Test		0.081	20:27
Air Blank		0.000	20:28
Subject Sample	# 1	0.172	20:29
Breath Volume		2.257	20:29
Air Blank		0.000	20:30
Air Blank		0.000	20:31
Subject Sample	# 2	0.171	20:32
Breath Volume		1.996	20:32
Air Blank		0.000	20:33
Control Test		0.080	20:33
Air Blank		0.000	20:33
Diagnostics Check		OK	20:34

Data Download Date/Time: 02/03/2022 11:28

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BROCATO,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: WHYTE

DATE/TIME OF TEST: 01/14/2022 00:44:28

Diagnostics Check		OK	00:37
Air Blank		0.000	00:38
Control Test		0.081	00:38
Air Blank		0.000	00:38
Subject Sample	# 1	0.000	00:40
Breath Volume		3.472	00:40
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.000	00:43
Breath Volume		2.984	00:43
Air Blank		0.000	00:43
Control Test		0.081	00:43
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 02/03/2022 11:28

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: BASSO

DATE/TIME OF TEST: 01/14/2022 01:01:58

Diagnostics Check		OK	00:54
Air Blank		0.000	00:54
Control Test		0.081	00:54
Air Blank		0.000	00:55
Subject Sample	# 1	0.154	00:56
Breath Volume		2.019	00:56
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.164	01:00
Breath Volume		3.015	01:00
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 02/03/2022 11:28

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: FARMEN

DATE/TIME OF TEST: 01/14/2022 23:58:42

Diagnostics Check		OK	23:47	
Air Blank		0.000	23:47	
Control Test		0.082	23:48	
Air Blank		0.000	23:48	
Subject Sample	# 1		23:51	No Sample Provided
Breath Volume		0.000	23:51	
Air Blank		0.000	23:52	
Air Blank		0.000	23:54	
Subject Sample	# 2		23:57	No Sample Provided
Breath Volume		0.000	23:57	
Air Blank		0.000	23:57	
Control Test		0.082	23:58	
Air Blank		0.000	23:58	
Diagnostics Check		OK	23:58	

Data Download Date/Time: 02/03/2022 11:28

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: SCIONTI

DATE/TIME OF TEST: 01/15/2022 00:19:50

Diagnostics Check		OK	00:12
Air Blank		0.000	00:12
Control Test		0.081	00:13
Air Blank		0.000	00:13
Subject Sample	# 1	0.000	00:14
Breath Volume		1.824	00:14
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.008	00:18
Breath Volume		2.644	00:18
Air Blank		0.000	00:18
Control Test		0.081	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 02/03/2022 11:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: HOPKINS

DATE/TIME OF TEST: 01/15/2022 05:30:59

Diagnostics Check		OK	05:23
Air Blank		0.000	05:24
Control Test		0.081	05:24
Air Blank		0.000	05:25
Subject Sample	# 1	0.159	05:26
Breath Volume		4.042	05:26
Air Blank		0.000	05:27
Air Blank		0.000	05:28
Subject Sample	# 2	0.159	05:29
Breath Volume		3.554	05:29
Air Blank		0.000	05:30
Control Test		0.081	05:30
Air Blank		0.000	05:30
Diagnostics Check		OK	05:30

Data Download Date/Time: 02/03/2022 11:28

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: RUTH JR

DATE/TIME OF TEST: 01/15/2022 05:43:35

Diagnostics Check		OK	05:39	
Air Blank		0.000	05:40	
Control Test		0.081	05:40	
Air Blank		0.000	05:41	
Subject Sample	# 1		05:42	Subject Test Refused
Breath Volume		0.000	05:42	
Air Blank		0.000	05:42	
Control Test		0.081	05:42	
Air Blank		0.000	05:43	
Diagnostics Check		OK	05:43	

Data Download Date/Time: 02/03/2022 11:28

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TABOADA,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: PEPPERACK

DATE/TIME OF TEST: 01/15/2022 19:53:51

Diagnostics Check		OK	19:47
Air Blank		0.000	19:47
Control Test		0.081	19:48
Air Blank		0.000	19:48
Subject Sample	# 1	0.199	19:49
Breath Volume		1.937	19:49
Air Blank		0.000	19:49
Air Blank		0.000	19:51
Subject Sample	# 2	0.200	19:52
Breath Volume		2.042	19:52
Air Blank		0.000	19:52
Control Test		0.081	19:53
Air Blank		0.000	19:53
Diagnostics Check		OK	19:53

Data Download Date/Time: 02/03/2022 11:28

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: VICKREY

DATE/TIME OF TEST: 01/15/2022 21:57:02

Diagnostics Check		OK	21:48
Air Blank		0.000	21:48
Control Test		0.080	21:48
Air Blank		0.000	21:49
Subject Sample	# 1	0.165	21:51
Breath Volume		1.414	21:51
Air Blank		0.000	21:52
Air Blank		0.000	21:54
Subject Sample	# 2	0.160	21:55
Breath Volume		1.332	21:55
Air Blank		0.000	21:56
Control Test		0.078	21:56
Air Blank		0.000	21:56
Diagnostics Check		OK	21:57

Data Download Date/Time: 02/03/2022 11:28

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VEGA,J
ARRESTING OFFICER AGENCY: Seminole Res (Hills)
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: RAMPERSAD

DATE/TIME OF TEST: 01/15/2022 22:48:11

Diagnostics Check		OK	22:35	
Air Blank		0.000	22:36	
Control Test		0.080	22:36	
Air Blank		0.000	22:37	
Subject Sample	# 1	0.181	22:40	Volume Not Met
Breath Volume		0.171	22:40	
Air Blank		0.000	22:40	
Air Blank		0.000	22:42	
Subject Sample	# 2	0.239	22:43	
Breath Volume		1.863	22:43	
Air Blank		0.000	22:44	
Air Blank		0.000	22:46	
Subject Sample	# 3	0.226	22:46	
Breath Volume		1.230	22:46	
Air Blank		0.000	22:47	
Control Test		0.077	22:47	
Air Blank		0.000	22:48	
Diagnostics Check		OK	22:48	

Data Download Date/Time: 02/03/2022 11:28

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 01/16/2022 04:36:07

Diagnostics Check		OK	04:29
Air Blank		0.000	04:29
Control Test		0.081	04:29
Air Blank		0.000	04:30
Subject Sample	# 1	0.143	04:31
Breath Volume		1.652	04:31
Air Blank		0.000	04:31
Air Blank		0.000	04:33
Subject Sample	# 2	0.150	04:34
Breath Volume		1.328	04:34
Air Blank		0.000	04:35
Control Test		0.079	04:35
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 02/03/2022 11:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: URIBE

DATE/TIME OF TEST: 01/16/2022 06:00:54

Diagnostics Check		OK	05:54
Air Blank		0.000	05:54
Control Test		0.080	05:55
Air Blank		0.000	05:55
Subject Sample	# 1	0.236	05:56
Breath Volume		1.101	05:56
Air Blank		0.000	05:56
Air Blank		0.000	05:58
Subject Sample	# 2	0.237	05:59
Breath Volume		1.410	05:59
Air Blank		0.000	05:59
Control Test		0.080	06:00
Air Blank		0.000	06:00
Diagnostics Check		OK	06:00

Data Download Date/Time: 02/03/2022 11:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: RAMOS VILLEGAS

DATE/TIME OF TEST: 01/16/2022 06:12:46

Diagnostics Check		OK	06:04
Air Blank		0.000	06:05
Control Test		0.080	06:05
Air Blank		0.000	06:05
Subject Sample	# 1	0.144	06:07
Breath Volume		2.070	06:07
Air Blank		0.000	06:08
Air Blank		0.000	06:10
Subject Sample	# 2	0.144	06:11
Breath Volume		1.718	06:11
Air Blank		0.000	06:11
Control Test		0.079	06:12
Air Blank		0.000	06:12
Diagnostics Check		OK	06:12

Data Download Date/Time: 02/03/2022 11:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: ZACARIAS

DATE/TIME OF TEST: 01/16/2022 20:37:09

Diagnostics Check		OK	20:30
Air Blank		0.000	20:31
Control Test		0.080	20:31
Air Blank		0.000	20:32
Subject Sample	# 1	0.167	20:32
Breath Volume		2.054	20:32
Air Blank		0.000	20:33
Air Blank		0.000	20:35
Subject Sample	# 2	0.163	20:35
Breath Volume		2.269	20:35
Air Blank		0.000	20:36
Control Test		0.080	20:36
Air Blank		0.000	20:37
Diagnostics Check		OK	20:37

Data Download Date/Time: 02/03/2022 11:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: BHUIYAN

DATE/TIME OF TEST: 01/17/2022 02:23:12

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.080	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.123	02:18
Breath Volume		1.468	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:20
Subject Sample	# 2	0.121	02:21
Breath Volume		1.511	02:21
Air Blank		0.000	02:22
Control Test		0.079	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 02/03/2022 11:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 01/17/2022 02:51:28

Diagnostics Check		OK	02:45
Air Blank		0.000	02:45
Control Test		0.080	02:45
Air Blank		0.000	02:46
Subject Sample	# 1	0.088	02:46
Breath Volume		3.308	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.087	02:49
Breath Volume		3.585	02:49
Air Blank		0.000	02:50
Control Test		0.079	02:50
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 02/03/2022 11:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: HAMMOND

DATE/TIME OF TEST: 01/17/2022 05:31:12

Diagnostics Check		OK	05:24
Air Blank		0.000	05:24
Control Test		0.081	05:24
Air Blank		0.000	05:25
Subject Sample	# 1	0.097	05:26
Breath Volume		2.324	05:26
Air Blank		0.000	05:27
Air Blank		0.000	05:29
Subject Sample	# 2	0.089	05:29
Breath Volume		1.878	05:29
Air Blank		0.000	05:30
Control Test		0.082	05:30
Air Blank		0.000	05:31
Diagnostics Check		OK	05:31

Data Download Date/Time: 02/03/2022 11:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: MCQUEEN

DATE/TIME OF TEST: 01/19/2022 04:09:54

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.081	04:03
Air Blank		0.000	04:04
Subject Sample	# 1	0.103	04:05
Breath Volume		1.750	04:05
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.103	04:08
Breath Volume		1.609	04:08
Air Blank		0.000	04:08
Control Test		0.081	04:09
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 02/03/2022 11:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PASS,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: BUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:39

SUBJECT'S LAST NAME: ALTERMAN

DATE/TIME OF TEST: 01/19/2022 18:06:28

Diagnostics Check		OK	18:00
Air Blank		0.000	18:00
Control Test		0.081	18:01
Air Blank		0.000	18:01
Subject Sample	# 1	0.098	18:02
Breath Volume		4.089	18:02
Air Blank		0.000	18:02
Air Blank		0.000	18:04
Subject Sample	# 2	0.098	18:04
Breath Volume		3.847	18:04
Air Blank		0.000	18:05
Control Test		0.080	18:05
Air Blank		0.000	18:06
Diagnostics Check		OK	18:06

Data Download Date/Time: 02/03/2022 11:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMSON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 01/20/2022 15:43:09

Diagnostics Check		OK	15:35
Air Blank		0.000	15:36
Control Test		0.082	15:36
Air Blank		0.000	15:37
Subject Sample	# 1	0.105	15:38
Breath Volume		2.699	15:38
Air Blank		0.000	15:39
Air Blank		0.000	15:40
Subject Sample	# 2	0.101	15:41
Breath Volume		2.839	15:41
Air Blank		0.000	15:42
Control Test		0.081	15:42
Air Blank		0.000	15:43
Diagnostics Check		OK	15:43

Data Download Date/Time: 02/03/2022 11:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: GRANATOWSKI

DATE/TIME OF TEST: 01/21/2022 02:10:12

Diagnostics Check		OK	02:03
Air Blank		0.000	02:04
Control Test		0.081	02:04
Air Blank		0.000	02:05
Subject Sample	# 1	0.139	02:05
Breath Volume		3.496	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.135	02:08
Breath Volume		3.093	02:08
Air Blank		0.000	02:09
Control Test		0.080	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 02/03/2022 11:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BROCATO,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: COMM VEHIC
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 01/21/2022 04:50:58

Diagnostics Check		OK	04:39	
Air Blank		0.000	04:40	
Control Test		0.081	04:40	
Air Blank		0.000	04:40	
Subject Sample	# 1		04:44	No Sample Provided
Breath Volume		0.000	04:44	
Air Blank		0.000	04:44	
Air Blank		0.000	04:46	
Subject Sample	# 2		04:49	No Sample Provided
Breath Volume		0.000	04:49	
Air Blank		0.000	04:50	
Control Test		0.080	04:50	
Air Blank		0.000	04:50	
Diagnostics Check		OK	04:50	

Data Download Date/Time: 02/03/2022 11:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROBERTS,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: DE LA CRUZ SOLIS

DATE/TIME OF TEST: 01/22/2022 01:50:11

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.080	01:44
Air Blank		0.000	01:44
Subject Sample	# 1	0.141	01:45
Breath Volume		1.812	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.140	01:48
Breath Volume		1.855	01:48
Air Blank		0.000	01:49
Control Test		0.079	01:49
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 02/03/2022 11:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: DEADY JR

DATE/TIME OF TEST: 01/22/2022 02:22:33

Diagnostics Check		OK	02:15
Air Blank		0.000	02:15
Control Test		0.079	02:15
Air Blank		0.000	02:16
Subject Sample	# 1	0.133	02:17
Breath Volume		1.867	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.140	02:20
Breath Volume		2.535	02:20
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 02/03/2022 11:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: CLARY

DATE/TIME OF TEST: 01/23/2022 01:25:43

Diagnostics Check		OK	01:18
Air Blank		0.000	01:19
Control Test		0.081	01:19
Air Blank		0.000	01:20
Subject Sample	# 1	0.113	01:20
Breath Volume		1.941	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.108	01:24
Breath Volume		1.695	01:24
Air Blank		0.000	01:24
Control Test		0.080	01:25
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 02/03/2022 11:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PARKER,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: MARTINEZ CORREA

DATE/TIME OF TEST: 01/23/2022 05:53:29

Diagnostics Check		OK	05:46
Air Blank		0.000	05:47
Control Test		0.080	05:47
Air Blank		0.000	05:47
Subject Sample	# 1	0.171	05:49
Breath Volume		1.347	05:49
Air Blank		0.000	05:49
Air Blank		0.000	05:51
Subject Sample	# 2	0.166	05:51
Breath Volume		1.554	05:51
Air Blank		0.000	05:52
Control Test		0.079	05:52
Air Blank		0.000	05:53
Diagnostics Check		OK	05:53

Data Download Date/Time: 02/03/2022 11:30

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: REICHERT,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 11:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:21

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 01/23/2022 12:50:50

Diagnostics Check		OK	12:42
Air Blank		0.000	12:43
Control Test		0.080	12:43
Air Blank		0.000	12:44
Subject Sample	# 1	0.000	12:46
Breath Volume		1.445	12:46
Air Blank		0.000	12:46
Air Blank		0.000	12:48
Subject Sample	# 2	0.000	12:49
Breath Volume		1.636	12:49
Air Blank		0.000	12:49
Control Test		0.081	12:50
Air Blank		0.000	12:50
Diagnostics Check		OK	12:50

Data Download Date/Time: 02/03/2022 11:30

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PRINCE,L
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/23/2022 19:14:20

Diagnostics Check		OK	19:07
Air Blank		0.000	19:07
Control Test		0.080	19:07
Air Blank		0.000	19:08
Subject Sample	# 1	0.176	19:09
Breath Volume		3.230	19:09
Air Blank		0.000	19:09
Air Blank		0.000	19:11
Subject Sample	# 2	0.191	19:12
Breath Volume		1.699	19:12
Air Blank		0.000	19:13
Control Test		0.081	19:13
Air Blank		0.000	19:14
Diagnostics Check		OK	19:14

Data Download Date/Time: 02/03/2022 11:30

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: MANTILLA

DATE/TIME OF TEST: 01/24/2022 03:57:17

Diagnostics Check		OK	03:44	
Air Blank		0.000	03:45	
Control Test		0.080	03:45	
Air Blank		0.000	03:46	
Subject Sample	# 1	0.189	03:49	Volume Not Met
Breath Volume		0.847	03:49	
Air Blank		0.000	03:50	
Air Blank		0.000	03:51	
Subject Sample	# 2	0.189	03:52	
Breath Volume		1.609	03:52	
Air Blank		0.000	03:52	
Air Blank		0.000	03:54	
Subject Sample	# 3	0.191	03:55	
Breath Volume		1.203	03:55	
Air Blank		0.000	03:56	
Control Test		0.077	03:56	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 02/03/2022 11:30

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: JACKSON JR

DATE/TIME OF TEST: 01/25/2022 20:00:41

Diagnostics Check		OK	19:54
Air Blank		0.000	19:54
Control Test		0.081	19:55
Air Blank		0.000	19:55
Subject Sample	# 1	0.143	19:56
Breath Volume		1.574	19:56
Air Blank		0.000	19:56
Air Blank		0.000	19:58
Subject Sample	# 2	0.143	19:59
Breath Volume		1.835	19:59
Air Blank		0.000	19:59
Control Test		0.080	20:00
Air Blank		0.000	20:00
Diagnostics Check		OK	20:00

Data Download Date/Time: 02/03/2022 11:30

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUGHEY,L
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: DIAZ FLORES

DATE/TIME OF TEST: 01/26/2022 02:09:17

Diagnostics Check		OK	02:02
Air Blank		0.000	02:03
Control Test		0.081	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.000	02:04
Breath Volume		2.042	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.000	02:07
Breath Volume		1.671	02:07
Air Blank		0.000	02:08
Control Test		0.081	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 02/03/2022 11:30

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: MENDEZ PEREZ

DATE/TIME OF TEST: 01/26/2022 23:43:46

Diagnostics Check		OK	23:31	
Air Blank		0.000	23:31	
Control Test		0.081	23:32	
Air Blank		0.000	23:32	
Subject Sample	# 1	0.310	23:33	No .020 agreement
Breath Volume		2.109	23:33	
Air Blank		0.000	23:34	
Air Blank		0.000	23:35	
Subject Sample	# 2	0.280	23:36	No .020 agreement
Breath Volume		1.203	23:36	
Air Blank		0.000	23:37	
Air Blank		0.000	23:39	
Subject Sample	# 3		23:42	No Sample Provided
Breath Volume		0.000	23:42	
Air Blank		0.000	23:42	
Control Test		0.080	23:43	
Air Blank		0.000	23:43	
Diagnostics Check		OK	23:43	

Data Download Date/Time: 02/03/2022 11:30

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: MENDEZ PEREZ

DATE/TIME OF TEST: 01/27/2022 00:06:32

Diagnostics Check		OK	23:59
Air Blank		0.000	23:59
Control Test		0.081	00:00
Air Blank		0.000	00:00
Subject Sample	# 1	0.303	00:01
Breath Volume		1.109	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.303	00:04
Breath Volume		2.332	00:04
Air Blank		0.000	00:05
Control Test		0.079	00:05
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

Data Download Date/Time: 02/03/2022 11:30

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BOUZOUBAA

DATE/TIME OF TEST: 01/28/2022 02:53:27

Diagnostics Check		OK	02:45
Air Blank		0.000	02:46
Control Test		0.081	02:46
Air Blank		0.000	02:46
Subject Sample	# 1	0.115	02:48
Breath Volume		1.550	02:48
Air Blank		0.000	02:49
Air Blank		0.000	02:51
Subject Sample	# 2	0.108	02:51
Breath Volume		1.671	02:51
Air Blank		0.000	02:52
Control Test		0.079	02:52
Air Blank		0.000	02:53
Diagnostics Check		OK	02:53

Data Download Date/Time: 02/03/2022 11:30

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: CRUZ PAZ

DATE/TIME OF TEST: 01/29/2022 02:34:59

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.079	02:28
Air Blank		0.000	02:28
Subject Sample	# 1	0.200	02:30
Breath Volume		2.179	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:32
Subject Sample	# 2	0.194	02:33
Breath Volume		1.910	02:33
Air Blank		0.000	02:34
Control Test		0.078	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 02/03/2022 11:30

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: GOMEZPEREZ

DATE/TIME OF TEST: 01/29/2022 04:49:59

Diagnostics Check		OK	04:42
Air Blank		0.000	04:43
Control Test		0.079	04:43
Air Blank		0.000	04:44
Subject Sample	# 1	0.176	04:45
Breath Volume		2.421	04:45
Air Blank		0.000	04:46
Air Blank		0.000	04:47
Subject Sample	# 2	0.169	04:48
Breath Volume		2.164	04:48
Air Blank		0.000	04:49
Control Test		0.078	04:49
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 02/03/2022 11:30

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: CHAPAS

DATE/TIME OF TEST: 01/30/2022 02:14:44

Diagnostics Check		OK	02:08
Air Blank		0.000	02:08
Control Test		0.081	02:09
Air Blank		0.000	02:09
Subject Sample	# 1	0.101	02:10
Breath Volume		1.769	02:10
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.103	02:13
Breath Volume		2.003	02:13
Air Blank		0.000	02:13
Control Test		0.080	02:14
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 02/03/2022 11:30

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTMAN,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: RUDRA

DATE/TIME OF TEST: 01/30/2022 03:12:16

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.081	03:05
Air Blank		0.000	03:06
Subject Sample	# 1	0.184	03:07
Breath Volume		1.992	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.182	03:10
Breath Volume		1.796	03:10
Air Blank		0.000	03:11
Control Test		0.080	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 02/03/2022 11:30

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 01/30/2022 04:21:34

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.080	04:16
Air Blank		0.000	04:16
Subject Sample	# 1	0.100	04:17
Breath Volume		1.980	04:17
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.101	04:20
Breath Volume		2.339	04:20
Air Blank		0.000	04:20
Control Test		0.080	04:20
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 02/03/2022 11:30

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: RESPESS

DATE/TIME OF TEST: 01/30/2022 22:12:48

Diagnostics Check		OK	22:06
Air Blank		0.000	22:06
Control Test		0.082	22:06
Air Blank		0.000	22:07
Subject Sample	# 1	0.094	22:08
Breath Volume		2.343	22:08
Air Blank		0.000	22:08
Air Blank		0.000	22:10
Subject Sample	# 2	0.092	22:11
Breath Volume		1.980	22:11
Air Blank		0.000	22:11
Control Test		0.081	22:12
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 02/03/2022 11:30

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: ESCALERA

DATE/TIME OF TEST: 01/31/2022 03:02:09

Diagnostics Check		OK	02:54
Air Blank		0.000	02:55
Control Test		0.081	02:55
Air Blank		0.000	02:56
Subject Sample	# 1	0.202	02:57
Breath Volume		2.203	02:57
Air Blank		0.000	02:58
Air Blank		0.000	02:59
Subject Sample	# 2	0.186	03:00
Breath Volume		1.421	03:00
Air Blank		0.000	03:01
Control Test		0.080	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 02/03/2022 11:31

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRAY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WRAY,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 16:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: DANIEL

DATE/TIME OF TEST: 01/29/2022 17:58:01

Diagnostics Check		OK	17:50
Air Blank		0.000	17:50
Control Test		0.082	17:51
Air Blank		0.000	17:51
Subject Sample	# 1	0.222	17:52
Breath Volume		1.699	17:52
Air Blank		0.000	17:53
Air Blank		0.000	17:55
Subject Sample	# 2	0.215	17:56
Breath Volume		1.652	17:56
Air Blank		0.000	17:57
Control Test		0.084	17:57
Air Blank		0.000	17:57
Diagnostics Check		OK	17:58

Data Download Date/Time: 02/28/2022 11:16

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBOSIEWICZ,ANTHONY,N
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULLIVAN,TABAIRE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: AGUIRRE

DATE/TIME OF TEST: 01/02/2022 02:48:21

Diagnostics Check		OK	02:42
Air Blank		0.000	02:42
Control Test		0.079	02:42
Air Blank		0.000	02:43
Subject Sample	# 1	0.233	02:43
Breath Volume		2.488	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.223	02:46
Breath Volume		1.890	02:46
Air Blank		0.000	02:47
Control Test		0.079	02:47
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 01/26/2022 21:04

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBOSIEWICZ,ANTHONY,N
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DOBOSIEWICZ,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 01/17/2022 01:35:20

Diagnostics Check	OK	01:35	
Air Blank		01:35	RFI Detect
Air Blank		01:35	RFI Detect

Data Download Date/Time: 01/26/2022 21:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBOSIEWICZ,ANTHONY,N
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DOBOSIEWICZ,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 01/17/2022 01:37:11

Diagnostics Check	OK	01:36	
Air Blank		01:37	RFI Detect
Air Blank		01:37	RFI Detect

Data Download Date/Time: 01/26/2022 21:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBOSIEWICZ,ANTHONY,N
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DOBOSIEWICZ,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 01/17/2022 01:39:08

Diagnostics Check	OK	01:38	
Air Blank		01:39	RFI Detect
Air Blank		01:39	RFI Detect

Data Download Date/Time: 01/26/2022 21:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBOSIEWICZ,ANTHONY,N
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DOBOSIEWICZ,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 01/17/2022 01:52:23

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.080	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.207	01:47
Breath Volume		1.722	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.212	01:50
Breath Volume		1.597	01:50
Air Blank		0.000	01:51
Control Test		0.081	01:51
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 01/26/2022 21:04

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBOSIEWICZ,ANTHONY,N
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DOKENDORF,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: KAHLER

DATE/TIME OF TEST: 01/24/2022 23:07:01

Diagnostics Check		OK	23:01
Air Blank		0.000	23:01
Control Test		0.080	23:01
Air Blank		0.000	23:02
Subject Sample	# 1	0.000	23:02
Breath Volume		2.667	23:02
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.000	23:05
Breath Volume		2.808	23:05
Air Blank		0.000	23:06
Control Test		0.081	23:06
Air Blank		0.000	23:06
Diagnostics Check		OK	23:07

Data Download Date/Time: 01/26/2022 21:04

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: DEVANEY

DATE/TIME OF TEST: 01/31/2022 20:16:30

Diagnostics Check		OK	20:10
Air Blank		0.000	20:10
Control Test		0.081	20:11
Air Blank		0.000	20:11
Subject Sample	# 1	0.174	20:12
Breath Volume		2.449	20:12
Air Blank		0.000	20:12
Air Blank		0.000	20:14
Subject Sample	# 2	0.168	20:14
Breath Volume		1.878	20:14
Air Blank		0.000	20:15
Control Test		0.080	20:15
Air Blank		0.000	20:16
Diagnostics Check		OK	20:16

Data Download Date/Time: 02/16/2022 22:33

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHOBREY,SPENCER,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WHOBREY,SPENCER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: RODRIGUEZ-HERNANDEZ

DATE/TIME OF TEST: 01/24/2022 01:20:52

Diagnostics Check		OK	01:13
Air Blank		0.000	01:13
Control Test		0.081	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.167	01:16
Breath Volume		2.167	01:16
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.162	01:19
Breath Volume		1.562	01:19
Air Blank		0.000	01:19
Control Test		0.081	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 01/28/2022 12:20

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHOBREY,SPENCER,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WHOBREY,SPENCER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: COE

DATE/TIME OF TEST: 01/26/2022 22:00:46

Diagnostics Check		OK	21:54
Air Blank		0.000	21:55
Control Test		0.080	21:55
Air Blank		0.000	21:55
Subject Sample	# 1	0.000	21:56
Breath Volume		3.144	21:56
Air Blank		0.000	21:56
Air Blank		0.000	21:59
Subject Sample	# 2	0.000	21:59
Breath Volume		2.273	21:59
Air Blank		0.000	21:59
Control Test		0.079	22:00
Air Blank		0.000	22:00
Diagnostics Check		OK	22:00

Data Download Date/Time: 01/28/2022 12:20

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: NABIAL

DATE/TIME OF TEST: 01/08/2022 04:10:14

Diagnostics Check		OK	04:03
Air Blank		0.000	04:04
Control Test		0.081	04:04
Air Blank		0.000	04:04
Subject Sample	# 1	0.135	04:05
Breath Volume		1.558	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.137	04:08
Breath Volume		1.496	04:08
Air Blank		0.000	04:09
Control Test		0.082	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 01/25/2022 15:18

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,COLLIN,J
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: BARBERI,F
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1192013
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 01/09/2022 01:24:07

Diagnostics Check		OK	01:16
Air Blank		0.000	01:17
Control Test		0.081	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.166	01:19
Breath Volume		1.523	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:21
Subject Sample	# 2	0.161	01:22
Breath Volume		1.667	01:22
Air Blank		0.000	01:23
Control Test		0.078	01:23
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 01/19/2022 12:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JIMENEZ,ADRIAN,M
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: JIMENEZ,ADRIAN
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1192013
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:52

SUBJECT'S LAST NAME: JIMENEZ-COLLAZO

DATE/TIME OF TEST: 01/16/2022 07:45:55

Diagnostics Check		OK	07:43	
Air Blank		0.000	07:43	
Control Test		0.082	07:43	
Air Blank		0.000	07:44	
Subject Sample	# 1		07:45	Sequence Aborted
Breath Volume		0.257	07:45	
Air Blank		0.000	07:45	

Data Download Date/Time: 01/19/2022 12:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JIMENEZ,ADRIAN,M
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: JIMENEZ,ADRIAN
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1192013
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:52

SUBJECT'S LAST NAME: JIMENEZ-COLLAZO

DATE/TIME OF TEST: 01/16/2022 07:56:32

Diagnostics Check		OK	07:44	
Air Blank		0.000	07:45	
Control Test		0.079	07:45	
Air Blank		0.000	07:46	
Subject Sample	# 1	0.124	07:49	Volume Not Met
Breath Volume		0.437	07:49	
Air Blank		0.000	07:50	
Air Blank		0.000	07:51	
Subject Sample	# 2	0.119	07:54	Volume Not Met
Breath Volume		0.015	07:54	
Air Blank		0.000	07:55	
Control Test		0.076	07:55	
Air Blank		0.000	07:56	
Diagnostics Check		OK	07:56	

Data Download Date/Time: 01/19/2022 12:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JIMENEZ,ADRIAN,M
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: JIMENEZ,ADRIAN
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/27/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1192013
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:52

SUBJECT'S LAST NAME: JIMENEZ-COLLAZO

DATE/TIME OF TEST: 01/16/2022 08:13:14

Diagnostics Check		OK	07:59	
Air Blank		0.000	07:59	
Control Test		0.077	08:00	
Air Blank		0.000	08:00	
Subject Sample	# 1	0.124	08:03	Volume Not Met
Breath Volume		0.074	08:03	
Air Blank		0.000	08:04	
Air Blank		0.000	08:06	
Subject Sample	# 2	0.133	08:06	
Breath Volume		1.433	08:06	
Air Blank		0.000	08:07	
Air Blank		0.000	08:09	
Subject Sample	# 3	0.122	08:11	
Breath Volume		1.179	08:11	
Air Blank		0.000	08:12	
Control Test		0.076	08:12	
Air Blank		0.000	08:13	
Diagnostics Check		OK	08:13	

Data Download Date/Time: 01/19/2022 12:09

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 01/01/2022 03:02:53

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.079	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.138	02:58
Breath Volume		2.390	02:58
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.147	03:01
Breath Volume		2.757	03:01
Air Blank		0.000	03:01
Control Test		0.077	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 01/30/2022 19:50

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: HRISTIJAN

DATE/TIME OF TEST: 01/20/2022 01:39:20

Diagnostics Check		OK	01:32
Air Blank		0.000	01:32
Control Test		0.080	01:33
Air Blank		0.000	01:33
Subject Sample	# 1	0.221	01:34
Breath Volume		1.207	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:36
Subject Sample	# 2	0.206	01:37
Breath Volume		1.300	01:37
Air Blank		0.000	01:38
Control Test		0.078	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 01/30/2022 19:50

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: HEPBURN

DATE/TIME OF TEST: 01/31/2022 01:52:10

Diagnostics Check		OK	01:45
Air Blank		0.000	01:45
Control Test		0.080	01:46
Air Blank		0.000	01:46
Subject Sample	# 1	0.115	01:47
Breath Volume		1.410	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:49
Subject Sample	# 2	0.130	01:50
Breath Volume		1.472	01:50
Air Blank		0.000	01:51
Control Test		0.079	01:51
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 02/13/2022 10:07

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: FERREIRA

DATE/TIME OF TEST: 01/01/2022 04:23:18

Diagnostics Check		OK	04:16
Air Blank		0.000	04:17
Control Test		0.079	04:17
Air Blank		0.000	04:18
Subject Sample	# 1	0.159	04:18
Breath Volume		3.617	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.153	04:21
Breath Volume		3.410	04:21
Air Blank		0.000	04:22
Control Test		0.078	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 01/06/2022 12:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MEARSE,RYAN
ARRESTING OFFICER AGENCY: Macdill Air Force Ba
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 01/01/2022 04:44:53

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.078	04:38
Air Blank		0.000	04:38
Subject Sample	# 1	0.120	04:40
Breath Volume		2.484	04:40
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.119	04:43
Breath Volume		2.527	04:43
Air Blank		0.000	04:43
Control Test		0.078	04:44
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 01/06/2022 12:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: GOMES ONEIL

DATE/TIME OF TEST: 01/01/2022 05:14:53

Diagnostics Check		OK	05:08
Air Blank		0.000	05:09
Control Test		0.078	05:09
Air Blank		0.000	05:10
Subject Sample	# 1	0.148	05:10
Breath Volume		2.234	05:10
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.133	05:13
Breath Volume		1.136	05:13
Air Blank		0.000	05:13
Control Test		0.078	05:14
Air Blank		0.000	05:14
Diagnostics Check		OK	05:14

Data Download Date/Time: 01/06/2022 12:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: FRIOL

DATE/TIME OF TEST: 01/02/2022 01:53:01

Diagnostics Check		OK	01:44
Air Blank		0.000	01:45
Control Test		0.078	01:45
Air Blank		0.000	01:46
Subject Sample	# 1	0.183	01:48
Breath Volume		1.625	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:50
Subject Sample	# 2	0.178	01:51
Breath Volume		1.218	01:51
Air Blank		0.000	01:52
Control Test		0.077	01:52
Air Blank		0.000	01:52
Diagnostics Check		OK	01:53

Data Download Date/Time: 01/06/2022 12:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: COKER

DATE/TIME OF TEST: 01/03/2022 03:20:00

Diagnostics Check		OK	03:13
Air Blank		0.000	03:14
Control Test		0.079	03:14
Air Blank		0.000	03:15
Subject Sample	# 1	0.177	03:15
Breath Volume		1.152	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.180	03:18
Breath Volume		1.925	03:18
Air Blank		0.000	03:19
Control Test		0.079	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:20

Data Download Date/Time: 01/06/2022 12:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: TOUISITHIPHONEXAY

DATE/TIME OF TEST: 01/01/2022 01:54:01

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.080	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.151	01:47
Breath Volume		1.636	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:49
Subject Sample	# 2	0.155	01:52
Breath Volume		1.437	01:52
Air Blank		0.000	01:53
Control Test		0.076	01:53
Air Blank		0.000	01:53
Diagnostics Check		OK	01:54

Data Download Date/Time: 01/07/2022 19:24

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SCHOMP,DALTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: EIGHNER

DATE/TIME OF TEST: 01/01/2022 02:25:55

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.080	02:20
Air Blank		0.000	02:20
Subject Sample	# 1	0.248	02:21
Breath Volume		1.843	02:21
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.252	02:24
Breath Volume		2.398	02:24
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 01/07/2022 19:24

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 05:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/01/2022 05:33:21

Diagnostics Check		OK	05:26
Air Blank		0.000	05:27
Control Test		0.078	05:27
Air Blank		0.000	05:28
Subject Sample	# 1	0.122	05:28
Breath Volume		2.402	05:28
Air Blank		0.000	05:29
Air Blank		0.000	05:31
Subject Sample	# 2	0.118	05:31
Breath Volume		2.039	05:31
Air Blank		0.000	05:32
Control Test		0.079	05:32
Air Blank		0.000	05:33
Diagnostics Check		OK	05:33

Data Download Date/Time: 01/07/2022 19:24

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUNDH,KAYLA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: SCHILD

DATE/TIME OF TEST: 01/01/2022 17:55:21

Diagnostics Check		OK	17:52	
Air Blank		0.000	17:52	
Control Test		0.081	17:53	
Air Blank		0.000	17:53	
Subject Sample	# 1		17:53	Subject Test Refused
Breath Volume		0.000	17:53	
Air Blank		0.000	17:54	
Control Test		0.080	17:54	
Air Blank		0.000	17:55	
Diagnostics Check		OK	17:55	

Data Download Date/Time: 01/07/2022 19:24

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,DEANGELO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: MIDDLETON

DATE/TIME OF TEST: 01/02/2022 01:00:24

Diagnostics Check	OK	00:59	
Air Blank	0.000	00:59	
Control Test	0.085	00:59	Interferent Detect
Air Blank	0.000	01:00	

Data Download Date/Time: 01/07/2022 19:24

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,DEANGELO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: MIDDLETON

DATE/TIME OF TEST: 01/02/2022 01:08:43

Diagnostics Check		OK	01:02
Air Blank		0.000	01:02
Control Test		0.084	01:02
Air Blank		0.000	01:03
Subject Sample	# 1	0.137	01:04
Breath Volume		1.726	01:04
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.133	01:07
Breath Volume		1.585	01:07
Air Blank		0.000	01:07
Control Test		0.082	01:08
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 01/07/2022 19:24

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: CZERNIAKOWSKI

DATE/TIME OF TEST: 01/02/2022 02:08:06

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:05	
Control Test		0.079	02:06	
Air Blank		0.000	02:06	
Subject Sample	# 1		02:06	Subject Test Refused
Breath Volume		0.000	02:06	
Air Blank		0.000	02:07	
Control Test		0.080	02:07	
Air Blank		0.000	02:08	
Diagnostics Check		OK	02:08	

Data Download Date/Time: 01/07/2022 19:24

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: WEIGOLD

DATE/TIME OF TEST: 01/13/2022 02:22:03

Diagnostics Check		OK	02:19	
Air Blank		0.000	02:19	
Control Test		0.081	02:20	
Air Blank		0.000	02:20	
Subject Sample	# 1		02:20	Subject Test Refused
Breath Volume		0.000	02:20	
Air Blank		0.000	02:21	
Control Test		0.082	02:21	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:22	

Data Download Date/Time: 02/16/2022 22:36

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: SIGNORE

DATE/TIME OF TEST: 01/14/2022 02:34:52

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.082	02:29
Air Blank		0.000	02:30
Subject Sample	# 1	0.000	02:30
Breath Volume		1.335	02:30
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.000	02:33
Breath Volume		1.160	02:33
Air Blank		0.000	02:33
Control Test		0.082	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 02/16/2022 22:36

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 01/14/2022 22:35:11

Diagnostics Check		OK	22:32	
Air Blank		0.000	22:32	
Control Test		0.082	22:33	
Air Blank		0.000	22:33	
Subject Sample	# 1		22:33	Subject Test Refused
Breath Volume		0.000	22:33	
Air Blank		0.000	22:34	
Control Test		0.082	22:34	
Air Blank		0.000	22:35	
Diagnostics Check		OK	22:35	

Data Download Date/Time: 02/16/2022 22:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: FISCHETTI

DATE/TIME OF TEST: 01/14/2022 23:44:52

Diagnostics Check		OK	23:42	
Air Blank		0.000	23:42	
Control Test		0.084	23:42	
Air Blank		0.000	23:43	
Subject Sample	# 1		23:43	Subject Test Refused
Breath Volume		0.000	23:43	
Air Blank		0.000	23:43	
Control Test		0.082	23:44	
Air Blank		0.000	23:44	
Diagnostics Check		OK	23:44	

Data Download Date/Time: 02/16/2022 22:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: FISCHETTI

DATE/TIME OF TEST: 01/15/2022 00:00:37

Diagnostics Check		OK	23:53
Air Blank		0.000	23:54
Control Test		0.083	23:54
Air Blank		0.000	23:55
Subject Sample	# 1	0.105	23:56
Breath Volume		1.722	23:56
Air Blank		0.000	23:56
Air Blank		0.000	23:58
Subject Sample	# 2	0.112	23:59
Breath Volume		1.464	23:59
Air Blank		0.000	23:59
Control Test		0.080	00:00
Air Blank		0.000	00:00
Diagnostics Check		OK	00:00

Data Download Date/Time: 02/16/2022 22:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: FARRIS

DATE/TIME OF TEST: 01/15/2022 01:48:20

Diagnostics Check		OK	01:41
Air Blank		0.000	01:41
Control Test		0.083	01:42
Air Blank		0.000	01:42
Subject Sample	# 1	0.153	01:43
Breath Volume		1.843	01:43
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.145	01:46
Breath Volume		1.453	01:46
Air Blank		0.000	01:47
Control Test		0.084	01:47
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 02/16/2022 22:36

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 01/15/2022 02:56:58

Diagnostics Check		OK	02:54	
Air Blank		0.000	02:54	
Control Test		0.083	02:54	
Air Blank		0.000	02:55	
Subject Sample	# 1		02:55	Subject Test Refused
Breath Volume		0.000	02:55	
Air Blank		0.000	02:56	
Control Test		0.081	02:56	
Air Blank		0.000	02:56	
Diagnostics Check		OK	02:56	

Data Download Date/Time: 02/16/2022 22:36

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/15/2022 03:45:21

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:43	
Control Test		0.082	03:43	
Air Blank		0.000	03:43	
Subject Sample	# 1		03:43	Subject Test Refused
Breath Volume		0.000	03:43	
Air Blank		0.000	03:44	
Control Test		0.082	03:44	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 02/16/2022 22:36

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: FARANO

DATE/TIME OF TEST: 01/15/2022 23:10:39

Diagnostics Check		OK	23:07	
Air Blank		0.000	23:08	
Control Test		0.083	23:08	
Air Blank		0.000	23:09	
Subject Sample	# 1		23:09	Subject Test Refused
Breath Volume		0.000	23:09	
Air Blank		0.000	23:09	
Control Test		0.082	23:10	
Air Blank		0.000	23:10	
Diagnostics Check		OK	23:10	

Data Download Date/Time: 02/16/2022 22:36

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: SLACK

DATE/TIME OF TEST: 01/15/2022 23:46:31

Diagnostics Check		OK	23:39
Air Blank		0.000	23:40
Control Test		0.080	23:40
Air Blank		0.000	23:41
Subject Sample	# 1	0.064	23:41
Breath Volume		1.207	23:41
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.068	23:44
Breath Volume		1.496	23:44
Air Blank		0.000	23:45
Control Test		0.084	23:45
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 02/16/2022 22:36

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 01/16/2022 01:20:50

Diagnostics Check		OK	01:18	
Air Blank		0.000	01:18	
Control Test		0.080	01:18	
Air Blank		0.000	01:19	
Subject Sample	# 1		01:19	Subject Test Refused
Breath Volume		0.000	01:19	
Air Blank		0.000	01:19	
Control Test		0.081	01:20	
Air Blank		0.000	01:20	
Diagnostics Check		OK	01:20	

Data Download Date/Time: 02/16/2022 22:36

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: GIRIMONTE

DATE/TIME OF TEST: 01/16/2022 01:51:19

Diagnostics Check		OK	01:44
Air Blank		0.000	01:45
Control Test		0.079	01:45
Air Blank		0.000	01:45
Subject Sample	# 1	0.166	01:46
Breath Volume		2.593	01:46
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.168	01:49
Breath Volume		1.890	01:49
Air Blank		0.000	01:50
Control Test		0.081	01:50
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 02/16/2022 22:36

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 01/23/2022 01:29:52

Diagnostics Check		OK	01:18	
Air Blank		0.000	01:18	
Control Test		0.081	01:19	
Air Blank		0.000	01:19	
Subject Sample	# 1	0.169	01:22	Volume Not Met
Breath Volume		0.300	01:22	
Air Blank		0.000	01:23	
Air Blank		0.000	01:25	
Subject Sample	# 2		01:28	No Sample Provided
Breath Volume		0.000	01:28	
Air Blank		0.000	01:28	
Control Test		0.081	01:29	
Air Blank		0.000	01:29	
Diagnostics Check		OK	01:29	

Data Download Date/Time: 02/16/2022 22:37

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 01/23/2022 01:58:15

Diagnostics Check		OK	01:50
Air Blank		0.000	01:51
Control Test		0.082	01:51
Air Blank		0.000	01:51
Subject Sample	# 1	0.203	01:52
Breath Volume		1.890	01:52
Air Blank		0.000	01:53
Air Blank		0.000	01:54
Subject Sample	# 2	0.207	01:56
Breath Volume		2.023	01:56
Air Blank		0.000	01:57
Control Test		0.079	01:57
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 02/16/2022 22:37

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: OCONNOR

DATE/TIME OF TEST: 01/27/2022 01:08:48

Diagnostics Check		OK	01:02
Air Blank		0.000	01:03
Control Test		0.081	01:03
Air Blank		0.000	01:03
Subject Sample	# 1	0.153	01:04
Breath Volume		2.781	01:04
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.153	01:07
Breath Volume		2.250	01:07
Air Blank		0.000	01:07
Control Test		0.083	01:08
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 02/16/2022 22:37

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: RICHMOND

DATE/TIME OF TEST: 01/29/2022 23:02:42

Diagnostics Check		OK	22:55
Air Blank		0.000	22:56
Control Test		0.083	22:56
Air Blank		0.000	22:57
Subject Sample	# 1	0.160	22:57
Breath Volume		2.367	22:57
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.169	23:01
Breath Volume		1.585	23:01
Air Blank		0.000	23:01
Control Test		0.085	23:02
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 02/16/2022 22:37

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 01/29/2022 23:26:01

Diagnostics Check		OK	23:23	
Air Blank		0.000	23:23	
Control Test		0.080	23:23	
Air Blank		0.000	23:24	
Subject Sample	# 1		23:24	Subject Test Refused
Breath Volume		0.000	23:24	
Air Blank		0.000	23:25	
Control Test		0.081	23:25	
Air Blank		0.000	23:25	
Diagnostics Check		OK	23:26	

Data Download Date/Time: 02/16/2022 22:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: REDONDO ACHOR

DATE/TIME OF TEST: 01/30/2022 01:21:02

Diagnostics Check		OK	01:18	
Air Blank		0.000	01:18	
Control Test		0.085	01:19	
Air Blank		0.000	01:19	
Subject Sample	# 1		01:19	Subject Test Refused
Breath Volume		0.000	01:19	
Air Blank		0.000	01:20	
Control Test		0.082	01:20	
Air Blank		0.000	01:20	
Diagnostics Check		OK	01:21	

Data Download Date/Time: 02/16/2022 22:37

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 01/30/2022 03:19:13

Diagnostics Check		OK	03:12
Air Blank		0.000	03:13
Control Test		0.084	03:13
Air Blank		0.000	03:14
Subject Sample	# 1	0.127	03:14
Breath Volume		1.761	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.124	03:17
Breath Volume		2.484	03:17
Air Blank		0.000	03:18
Control Test		0.085	03:18
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 02/16/2022 22:37

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,JOHN,S
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: SWINDLE,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:22

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 01/27/2022 13:51:36

Diagnostics Check		OK	13:44
Air Blank		0.000	13:45
Control Test		0.078	13:45
Air Blank		0.000	13:45
Subject Sample	# 1	0.000	13:46
Breath Volume		1.242	13:46
Air Blank		0.000	13:47
Air Blank		0.000	13:49
Subject Sample	# 2	0.000	13:50
Breath Volume		2.441	13:50
Air Blank		0.000	13:50
Control Test		0.078	13:51
Air Blank		0.000	13:51
Diagnostics Check		OK	13:51

Data Download Date/Time: 02/09/2022 11:23

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: RICKS

DATE/TIME OF TEST: 01/01/2022 19:26:34

Diagnostics Check		OK	19:18
Air Blank		0.000	19:18
Control Test		0.079	19:18
Air Blank		0.000	19:19
Subject Sample	# 1	0.168	19:21
Breath Volume		1.437	19:21
Air Blank		0.000	19:22
Air Blank		0.000	19:24
Subject Sample	# 2	0.157	19:24
Breath Volume		1.707	19:24
Air Blank		0.000	19:25
Control Test		0.078	19:25
Air Blank		0.000	19:26
Diagnostics Check		OK	19:26

Data Download Date/Time: 01/07/2022 11:50

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 01/02/2022 00:22:19

Diagnostics Check		OK	00:15
Air Blank		0.000	00:16
Control Test		0.078	00:16
Air Blank		0.000	00:17
Subject Sample	# 1	0.184	00:17
Breath Volume		1.601	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.195	00:20
Breath Volume		1.906	00:20
Air Blank		0.000	00:21
Control Test		0.078	00:21
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 01/07/2022 11:50

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: LANGSTON-GRAHAM

DATE/TIME OF TEST: 01/02/2022 02:04:03

Diagnostics Check		OK	01:55	
Air Blank		0.000	01:56	
Control Test		0.079	01:56	
Air Blank		0.000	01:56	
Subject Sample	# 1		02:00	No Sample Provided
Breath Volume		0.000	02:00	
Air Blank		0.000	02:00	
Air Blank		0.000	02:02	
Subject Sample	# 2		02:02	Subject Test Refused
Breath Volume		0.000	02:02	
Air Blank		0.000	02:03	
Control Test		0.078	02:03	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:04	

Data Download Date/Time: 01/07/2022 11:50

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HOUTEN,JONATHAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: BURTON

DATE/TIME OF TEST: 01/02/2022 03:39:32

Diagnostics Check		OK	03:28	
Air Blank		0.000	03:28	
Control Test		0.079	03:28	
Air Blank		0.000	03:29	
Subject Sample	# 1		03:32	No Sample Provided
Breath Volume		0.000	03:32	
Air Blank		0.000	03:33	
Air Blank		0.000	03:35	
Subject Sample	# 2		03:38	No Sample Provided
Breath Volume		0.000	03:38	
Air Blank		0.000	03:38	
Control Test		0.078	03:38	
Air Blank		0.000	03:39	
Diagnostics Check		OK	03:39	

Data Download Date/Time: 01/07/2022 11:50

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/02/2022 04:05:34

Diagnostics Check		OK	03:51	
Air Blank		0.000	03:52	
Control Test		0.078	03:52	
Air Blank		0.000	03:52	
Subject Sample	# 1	0.152	03:56	Volume Not Met
Breath Volume		0.070	03:56	
Air Blank		0.000	03:56	
Air Blank		0.000	03:58	
Subject Sample	# 2	0.137	03:59	
Breath Volume		1.164	03:59	
Air Blank		0.000	04:00	
Air Blank		0.000	04:01	
Subject Sample	# 3	0.143	04:03	
Breath Volume		1.203	04:03	
Air Blank		0.000	04:04	
Control Test		0.076	04:04	
Air Blank		0.000	04:05	
Diagnostics Check		OK	04:05	

Data Download Date/Time: 01/07/2022 11:50

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: HAKIM

DATE/TIME OF TEST: 01/04/2022 19:08:49

Diagnostics Check		OK	19:05	
Air Blank		0.000	19:06	
Control Test		0.081	19:06	
Air Blank		0.000	19:07	
Subject Sample	# 1		19:07	RFI Detect
Breath Volume		0.000	19:07	
Air Blank		0.000	19:07	
Control Test		0.081	19:08	
Air Blank		0.000	19:08	
Diagnostics Check		OK	19:08	

Data Download Date/Time: 01/07/2022 11:50

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: HAKIM

DATE/TIME OF TEST: 01/04/2022 19:16:05

Diagnostics Check		OK	19:10
Air Blank		0.000	19:10
Control Test		0.079	19:10
Air Blank		0.000	19:11
Subject Sample	# 1	0.133	19:11
Breath Volume		1.496	19:11
Air Blank		0.000	19:12
Air Blank		0.000	19:14
Subject Sample	# 2	0.140	19:14
Breath Volume		1.804	19:14
Air Blank		0.000	19:15
Control Test		0.079	19:15
Air Blank		0.000	19:16
Diagnostics Check		OK	19:16

Data Download Date/Time: 01/07/2022 11:50

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: MATHIS

DATE/TIME OF TEST: 01/05/2022 17:55:50

Diagnostics Check		OK	17:49
Air Blank		0.000	17:50
Control Test		0.081	17:50
Air Blank		0.000	17:51
Subject Sample	# 1	0.000	17:51
Breath Volume		2.214	17:51
Air Blank		0.000	17:52
Air Blank		0.000	17:54
Subject Sample	# 2	0.000	17:54
Breath Volume		2.511	17:54
Air Blank		0.000	17:54
Control Test		0.079	17:55
Air Blank		0.000	17:55
Diagnostics Check		OK	17:55

Data Download Date/Time: 01/07/2022 11:50

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 01/05/2022 23:07:38

Diagnostics Check		OK	22:59
Air Blank		0.000	23:00
Control Test		0.080	23:00
Air Blank		0.000	23:01
Subject Sample	# 1	0.143	23:02
Breath Volume		1.195	23:02
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.146	23:06
Breath Volume		1.406	23:06
Air Blank		0.000	23:06
Control Test		0.080	23:07
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 01/07/2022 11:50

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HOUTEN,JONATHAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: KEENER

DATE/TIME OF TEST: 01/07/2022 00:00:59

Diagnostics Check		OK	23:54
Air Blank		0.000	23:55
Control Test		0.079	23:55
Air Blank		0.000	23:56
Subject Sample	# 1	0.000	23:56
Breath Volume		2.347	23:56
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 2	0.000	23:59
Breath Volume		2.417	23:59
Air Blank		0.000	00:00
Control Test		0.079	00:00
Air Blank		0.000	00:00
Diagnostics Check		OK	00:00

Data Download Date/Time: 01/07/2022 11:50

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: MICHAELS

DATE/TIME OF TEST: 01/07/2022 21:45:57

Diagnostics Check		OK	21:39
Air Blank		0.000	21:39
Control Test		0.080	21:40
Air Blank		0.000	21:40
Subject Sample	# 1	0.140	21:41
Breath Volume		1.171	21:41
Air Blank		0.000	21:41
Air Blank		0.000	21:43
Subject Sample	# 2	0.143	21:44
Breath Volume		1.699	21:44
Air Blank		0.000	21:44
Control Test		0.079	21:45
Air Blank		0.000	21:45
Diagnostics Check		OK	21:45

Data Download Date/Time: 02/03/2022 10:53

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: LOPEZ GOMEZ

DATE/TIME OF TEST: 01/08/2022 00:08:58

Diagnostics Check		OK	00:02
Air Blank		0.000	00:03
Control Test		0.081	00:03
Air Blank		0.000	00:03
Subject Sample	# 1	0.170	00:04
Breath Volume		2.156	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:06
Subject Sample	# 2	0.167	00:07
Breath Volume		2.613	00:07
Air Blank		0.000	00:08
Control Test		0.080	00:08
Air Blank		0.000	00:08
Diagnostics Check		OK	00:08

Data Download Date/Time: 02/03/2022 10:53

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: RUTKOWSKI

DATE/TIME OF TEST: 01/08/2022 02:26:45

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.080	02:20
Air Blank		0.000	02:20
Subject Sample	# 1	0.115	02:21
Breath Volume		2.000	02:21
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.126	02:25
Breath Volume		2.695	02:25
Air Blank		0.000	02:25
Control Test		0.080	02:26
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 02/03/2022 10:53

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: CONSER

DATE/TIME OF TEST: 01/12/2022 20:11:14

Diagnostics Check		OK	20:05
Air Blank		0.000	20:05
Control Test		0.081	20:05
Air Blank		0.000	20:06
Subject Sample	# 1	0.020	20:06
Breath Volume		2.289	20:06
Air Blank		0.000	20:07
Air Blank		0.000	20:09
Subject Sample	# 2	0.019	20:09
Breath Volume		2.375	20:09
Air Blank		0.000	20:10
Control Test		0.082	20:10
Air Blank		0.000	20:11
Diagnostics Check		OK	20:11

Data Download Date/Time: 02/03/2022 10:53

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: CURLEY

DATE/TIME OF TEST: 01/12/2022 21:27:10

Diagnostics Check		OK	21:20
Air Blank		0.000	21:21
Control Test		0.080	21:21
Air Blank		0.000	21:22
Subject Sample	# 1	0.098	21:22
Breath Volume		2.429	21:22
Air Blank		0.000	21:23
Air Blank		0.000	21:25
Subject Sample	# 2	0.099	21:25
Breath Volume		1.546	21:25
Air Blank		0.000	21:26
Control Test		0.082	21:26
Air Blank		0.000	21:27
Diagnostics Check		OK	21:27

Data Download Date/Time: 02/03/2022 10:53

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 01/13/2022 01:12:38

Diagnostics Check		OK	01:04	
Air Blank		0.000	01:04	
Control Test		0.081	01:04	
Air Blank		0.000	01:05	
Subject Sample	# 1		01:08	No Sample Provided
Breath Volume		0.000	01:08	
Air Blank		0.000	01:08	
Air Blank		0.000	01:10	
Subject Sample	# 2		01:11	Subject Test Refused
Breath Volume		0.000	01:11	
Air Blank		0.000	01:11	
Control Test		0.080	01:12	
Air Blank		0.000	01:12	
Diagnostics Check		OK	01:12	

Data Download Date/Time: 02/03/2022 10:53

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: GIDIBMA

DATE/TIME OF TEST: 01/13/2022 23:59:18

Diagnostics Check		OK	23:52
Air Blank		0.000	23:53
Control Test		0.081	23:53
Air Blank		0.000	23:53
Subject Sample	# 1	0.178	23:54
Breath Volume		3.308	23:54
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.179	23:57
Breath Volume		3.210	23:57
Air Blank		0.000	23:58
Control Test		0.079	23:58
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 02/03/2022 10:54

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 01/14/2022 01:06:10

Diagnostics Check		OK	01:00
Air Blank		0.000	01:00
Control Test		0.081	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.000	01:01
Breath Volume		3.476	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.000	01:04
Breath Volume		2.878	01:04
Air Blank		0.000	01:05
Control Test		0.081	01:05
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 02/03/2022 10:54

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 01/15/2022 02:18:57

Diagnostics Check		OK	02:04	
Air Blank		0.000	02:04	
Control Test		0.081	02:05	
Air Blank		0.000	02:05	
Subject Sample	# 1		02:08	No Sample Provided
Breath Volume		0.000	02:08	
Air Blank		0.000	02:09	
Air Blank		0.000	02:11	
Subject Sample	# 2	0.147	02:13	
Breath Volume		2.808	02:13	
Air Blank		0.000	02:14	
Air Blank		0.000	02:16	
Subject Sample	# 3	0.138	02:17	
Breath Volume		2.875	02:17	
Air Blank		0.000	02:18	
Control Test		0.077	02:18	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 02/03/2022 10:54

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: MCCORMACK

DATE/TIME OF TEST: 01/15/2022 23:02:11

Diagnostics Check		OK	22:50	
Air Blank		0.000	22:51	
Control Test		0.081	22:51	
Air Blank		0.000	22:52	
Subject Sample	# 1	0.161	22:52	
Breath Volume		2.539	22:52	
Air Blank		0.000	22:53	
Air Blank		0.000	22:54	
Subject Sample	# 2		22:58	No Sample Provided
Breath Volume		0.000	22:58	
Air Blank		0.000	22:58	
Air Blank		0.000	23:00	
Subject Sample	# 3		23:00	Subject Test Refused
Breath Volume		0.000	23:00	
Air Blank		0.000	23:01	
Control Test		0.080	23:01	
Air Blank		0.000	23:02	
Diagnostics Check		OK	23:02	

Data Download Date/Time: 02/03/2022 10:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: PIHAYLIC

DATE/TIME OF TEST: 01/16/2022 01:08:47

Diagnostics Check		OK	01:01
Air Blank		0.000	01:02
Control Test		0.080	01:02
Air Blank		0.000	01:03
Subject Sample	# 1	0.127	01:04
Breath Volume		2.695	01:04
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.120	01:07
Breath Volume		2.003	01:07
Air Blank		0.000	01:07
Control Test		0.079	01:08
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 02/03/2022 10:54

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: RODRIGUEZ COSTA PASO

DATE/TIME OF TEST: 01/21/2022 00:35:08

Diagnostics Check		OK	00:26	
Air Blank		0.000	00:27	
Control Test		0.080	00:27	
Air Blank		0.000	00:27	
Subject Sample	# 1		00:31	No Sample Provided
Breath Volume		0.000	00:31	
Air Blank		0.000	00:31	
Air Blank		0.000	00:33	
Subject Sample	# 2		00:33	Subject Test Refused
Breath Volume		0.000	00:33	
Air Blank		0.000	00:34	
Control Test		0.080	00:34	
Air Blank		0.000	00:35	
Diagnostics Check		OK	00:35	

Data Download Date/Time: 02/03/2022 10:54

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GERRY,TABATHA,B
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MIZIC

DATE/TIME OF TEST: 01/23/2022 03:09:12

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.081	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.214	03:04
Breath Volume		1.855	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:06
Subject Sample	# 2	0.209	03:07
Breath Volume		1.628	03:07
Air Blank		0.000	03:08
Control Test		0.078	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 02/03/2022 10:54

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUSAM,MATTHEW,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GANNAWAY,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:18

SUBJECT'S LAST NAME: MAYFIELD

DATE/TIME OF TEST: 01/23/2022 06:48:00

Diagnostics Check		OK	06:39	
Air Blank		0.000	06:39	
Control Test		0.081	06:40	
Air Blank		0.000	06:40	
Subject Sample	# 1		06:43	No Sample Provided
Breath Volume		0.000	06:43	
Air Blank		0.000	06:44	
Air Blank		0.000	06:46	
Subject Sample	# 2		06:46	Subject Test Refused
Breath Volume		0.000	06:46	
Air Blank		0.000	06:47	
Control Test		0.080	06:47	
Air Blank		0.000	06:47	
Diagnostics Check		OK	06:48	

Data Download Date/Time: 02/03/2022 10:54

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,RONALD,W
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE,RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: UGLES

DATE/TIME OF TEST: 01/23/2022 14:16:49

Diagnostics Check		OK	14:10
Air Blank		0.000	14:10
Control Test		0.082	14:11
Air Blank		0.000	14:11
Subject Sample	# 1	0.000	14:12
Breath Volume		2.910	14:12
Air Blank		0.000	14:13
Air Blank		0.000	14:15
Subject Sample	# 2	0.000	14:15
Breath Volume		3.707	14:15
Air Blank		0.000	14:15
Control Test		0.079	14:16
Air Blank		0.000	14:16
Diagnostics Check		OK	14:16

Data Download Date/Time: 02/03/2022 10:54

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: DURANT JR

DATE/TIME OF TEST: 01/28/2022 00:29:43

Diagnostics Check		OK	00:21	
Air Blank		0.000	00:21	
Control Test		0.081	00:22	
Air Blank		0.000	00:22	
Subject Sample	# 1		00:25	No Sample Provided
Breath Volume		0.000	00:25	
Air Blank		0.000	00:26	
Air Blank		0.000	00:28	
Subject Sample	# 2		00:28	Subject Test Refused
Breath Volume		0.000	00:28	
Air Blank		0.000	00:28	
Control Test		0.081	00:29	
Air Blank		0.000	00:29	
Diagnostics Check		OK	00:29	

Data Download Date/Time: 02/03/2022 10:55

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 01/28/2022 02:56:12

Diagnostics Check		OK	02:46
Air Blank		0.000	02:47
Control Test		0.081	02:47
Air Blank		0.000	02:48
Subject Sample	# 1	0.216	02:50
Breath Volume		1.218	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.211	02:54
Breath Volume		1.519	02:54
Air Blank		0.000	02:55
Control Test		0.079	02:55
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 02/03/2022 10:55

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: HOLLANDSWORTH

DATE/TIME OF TEST: 01/29/2022 00:11:56

Diagnostics Check		OK	00:03	
Air Blank		0.000	00:03	
Control Test		0.081	00:04	
Air Blank		0.000	00:04	
Subject Sample	# 1		00:07	No Sample Provided
Breath Volume		0.000	00:07	
Air Blank		0.000	00:08	
Air Blank		0.000	00:10	
Subject Sample	# 2		00:10	Subject Test Refused
Breath Volume		0.000	00:10	
Air Blank		0.000	00:11	
Control Test		0.081	00:11	
Air Blank		0.000	00:11	
Diagnostics Check		OK	00:11	

Data Download Date/Time: 02/03/2022 10:55

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: MCDOWELL

DATE/TIME OF TEST: 01/29/2022 01:16:05

Diagnostics Check		OK	01:02	
Air Blank		0.000	01:03	
Control Test		0.081	01:03	
Air Blank		0.000	01:04	
Subject Sample	# 1	0.170	01:07	Volume Not Met
Breath Volume		0.535	01:07	
Air Blank		0.000	01:08	
Air Blank		0.000	01:09	
Subject Sample	# 2	0.183	01:10	
Breath Volume		1.117	01:10	
Air Blank		0.000	01:11	
Air Blank		0.000	01:13	
Subject Sample	# 3	0.199	01:14	
Breath Volume		1.105	01:14	
Air Blank		0.000	01:15	
Control Test		0.080	01:15	
Air Blank		0.000	01:16	
Diagnostics Check		OK	01:16	

Data Download Date/Time: 02/03/2022 10:55

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: RUIZ DELVIZO

DATE/TIME OF TEST: 01/29/2022 02:38:32

Diagnostics Check		OK	02:30	
Air Blank		0.000	02:30	
Control Test		0.081	02:30	
Air Blank		0.000	02:31	
Subject Sample	# 1		02:34	No Sample Provided
Breath Volume		0.000	02:34	
Air Blank		0.000	02:34	
Air Blank		0.000	02:36	
Subject Sample	# 2		02:37	Subject Test Refused
Breath Volume		0.000	02:37	
Air Blank		0.000	02:37	
Control Test		0.081	02:37	
Air Blank		0.000	02:38	
Diagnostics Check		OK	02:38	

Data Download Date/Time: 02/03/2022 10:55

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 01/29/2022 03:36:56

Diagnostics Check		OK	03:29
Air Blank		0.000	03:29
Control Test		0.081	03:30
Air Blank		0.000	03:30
Subject Sample	# 1	0.150	03:32
Breath Volume		1.300	03:32
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.169	03:35
Breath Volume		1.316	03:35
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 02/03/2022 10:55

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: SYLVESTRI

DATE/TIME OF TEST: 01/29/2022 04:14:39

Diagnostics Check		OK	04:06	
Air Blank		0.000	04:06	
Control Test		0.081	04:06	
Air Blank		0.000	04:07	
Subject Sample	# 1		04:10	No Sample Provided
Breath Volume		0.000	04:10	
Air Blank		0.000	04:11	
Air Blank		0.000	04:13	
Subject Sample	# 2		04:13	Subject Test Refused
Breath Volume		0.000	04:13	
Air Blank		0.000	04:13	
Control Test		0.082	04:14	
Air Blank		0.000	04:14	
Diagnostics Check		OK	04:14	

Data Download Date/Time: 02/03/2022 10:55

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,RONALD,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE,RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:25

SUBJECT'S LAST NAME: MICHAEL

DATE/TIME OF TEST: 01/29/2022 13:48:48

Diagnostics Check		OK	13:45	
Air Blank		0.000	13:46	
Control Test		0.082	13:46	
Air Blank		0.000	13:47	
Subject Sample	# 1		13:47	Subject Test Refused
Breath Volume		0.000	13:47	
Air Blank		0.000	13:47	
Control Test		0.082	13:48	
Air Blank		0.000	13:48	
Diagnostics Check		OK	13:48	

Data Download Date/Time: 02/03/2022 10:55

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KNOOP,DERRIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: BARRAZA BERMUDEZ

DATE/TIME OF TEST: 01/29/2022 18:08:27

Diagnostics Check		OK	18:05	
Air Blank		0.000	18:05	
Control Test		0.083	18:06	
Air Blank		0.000	18:06	
Subject Sample	# 1		18:06	Slope Not Met
Breath Volume		1.617	18:06	
Air Blank		0.000	18:07	
Control Test		0.083	18:07	
Air Blank		0.000	18:08	
Diagnostics Check		OK	18:08	

Data Download Date/Time: 02/03/2022 10:55

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KNOOP,DERRIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: BARRAZA BERMUDEZ

DATE/TIME OF TEST: 01/29/2022 18:18:03

Diagnostics Check		OK	18:10
Air Blank		0.000	18:11
Control Test		0.082	18:11
Air Blank		0.000	18:12
Subject Sample	# 1	0.368	18:13
Breath Volume		1.156	18:13
Air Blank		0.000	18:14
Air Blank		0.000	18:15
Subject Sample	# 2	0.383	18:16
Breath Volume		1.941	18:16
Air Blank		0.000	18:17
Control Test		0.080	18:17
Air Blank		0.000	18:17
Diagnostics Check		OK	18:18

Data Download Date/Time: 02/03/2022 10:55

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: AIELLO

DATE/TIME OF TEST: 01/30/2022 00:07:27

Diagnostics Check		OK	00:00
Air Blank		0.000	00:01
Control Test		0.083	00:01
Air Blank		0.000	00:02
Subject Sample	# 1	0.097	00:02
Breath Volume		1.753	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.097	00:05
Breath Volume		2.382	00:05
Air Blank		0.000	00:06
Control Test		0.080	00:06
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 02/03/2022 10:56

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: COLVAY

DATE/TIME OF TEST: 01/30/2022 02:03:29

Diagnostics Check		OK	01:54	
Air Blank		0.000	01:55	
Control Test		0.083	01:55	
Air Blank		0.000	01:56	
Subject Sample	# 1		01:59	No Sample Provided
Breath Volume		0.000	01:59	
Air Blank		0.000	01:59	
Air Blank		0.000	02:01	
Subject Sample	# 2		02:02	Subject Test Refused
Breath Volume		0.000	02:02	
Air Blank		0.000	02:02	
Control Test		0.084	02:02	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:03	

Data Download Date/Time: 02/03/2022 10:56

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: NJOROGE

DATE/TIME OF TEST: 01/30/2022 03:00:17

Diagnostics Check		OK	02:53
Air Blank		0.000	02:54
Control Test		0.083	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.187	02:55
Breath Volume		1.679	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.181	02:58
Breath Volume		1.949	02:58
Air Blank		0.000	02:59
Control Test		0.083	02:59
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 02/03/2022 10:56

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,CARLOS,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: STONE,JARED
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: EASON

DATE/TIME OF TEST: 01/31/2022 01:28:10

Diagnostics Check		OK	01:16	
Air Blank		0.000	01:17	
Control Test		0.083	01:17	
Air Blank		0.000	01:18	
Subject Sample	# 1		01:21	No Sample Provided
Breath Volume		0.000	01:21	
Air Blank		0.000	01:21	
Air Blank		0.000	01:23	
Subject Sample	# 2		01:26	No Sample Provided
Breath Volume		0.000	01:26	
Air Blank		0.000	01:27	
Control Test		0.083	01:27	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 02/03/2022 10:56

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,CARLOS,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: STONE,JARED
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: EASON

DATE/TIME OF TEST: 01/31/2022 01:34:55

Diagnostics Check		OK	01:31	
Air Blank		0.000	01:32	
Control Test		0.082	01:32	
Air Blank		0.000	01:32	
Subject Sample	# 1		01:33	Subject Test Refused
Breath Volume		0.000	01:33	
Air Blank		0.000	01:34	
Control Test		0.081	01:34	
Air Blank		0.000	01:34	
Diagnostics Check		OK	01:34	

Data Download Date/Time: 02/03/2022 10:56

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JAYDA,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WEISENBURGH,L
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: SCHWARTZ

DATE/TIME OF TEST: 01/08/2022 05:03:08

Diagnostics Check		OK	04:50	
Air Blank		0.000	04:50	
Control Test		0.081	04:50	
Air Blank		0.000	04:51	
Subject Sample	# 1	0.178	04:52	No .020 agreement
Breath Volume		1.242	04:52	
Air Blank		0.000	04:53	
Air Blank		0.000	04:54	
Subject Sample	# 2	0.155	04:55	No .020 agreement
Breath Volume		1.445	04:55	
Air Blank		0.000	04:56	
Air Blank		0.000	04:58	
Subject Sample	# 3	0.171	05:01	Volume Not Met
Breath Volume		0.015	05:01	
Air Blank		0.000	05:02	
Control Test		0.077	05:02	
Air Blank		0.000	05:03	
Diagnostics Check		OK	05:03	

Data Download Date/Time: 01/24/2022 09:58

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODAKOSKI,SARAH,R
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: PHILBIN,TOMMY
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: DUPONT

DATE/TIME OF TEST: 01/15/2022 01:40:18

Diagnostics Check		OK	01:33
Air Blank		0.000	01:34
Control Test		0.081	01:34
Air Blank		0.000	01:35
Subject Sample	# 1	0.217	01:35
Breath Volume		1.515	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.230	01:38
Breath Volume		2.019	01:38
Air Blank		0.000	01:39
Control Test		0.080	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 01/24/2022 09:58

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,YOEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GARCIA,YOEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:51

SUBJECT'S LAST NAME: JAMERSON

DATE/TIME OF TEST: 01/16/2022 16:42:51

Diagnostics Check		OK	16:36
Air Blank		0.000	16:36
Control Test		0.080	16:36
Air Blank		0.000	16:37
Subject Sample	# 1	0.304	16:37
Breath Volume		1.917	16:37
Air Blank		0.000	16:38
Air Blank		0.000	16:40
Subject Sample	# 2	0.302	16:40
Breath Volume		2.527	16:40
Air Blank		0.000	16:41
Control Test		0.079	16:42
Air Blank		0.000	16:42
Diagnostics Check		OK	16:42

Data Download Date/Time: 01/24/2022 09:59

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEOANRD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: MCHOLLAND

DATE/TIME OF TEST: 01/25/2022 02:45:48

Diagnostics Check		OK	02:39
Air Blank		0.000	02:40
Control Test		0.081	02:40
Air Blank		0.000	02:40
Subject Sample	# 1	0.181	02:41
Breath Volume		2.429	02:41
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.191	02:44
Breath Volume		2.859	02:44
Air Blank		0.000	02:44
Control Test		0.080	02:45
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 02/17/2022 10:32

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REILLY,S.,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: REILLY,S.
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: MCMILLAN-DALY

DATE/TIME OF TEST: 01/26/2022 20:22:31

Diagnostics Check		OK	20:15
Air Blank		0.000	20:16
Control Test		0.080	20:16
Air Blank		0.000	20:16
Subject Sample	# 1	0.150	20:17
Breath Volume		2.433	20:17
Air Blank		0.000	20:18
Air Blank		0.000	20:19
Subject Sample	# 2	0.141	20:20
Breath Volume		2.429	20:20
Air Blank		0.000	20:21
Control Test		0.079	20:21
Air Blank		0.000	20:22
Diagnostics Check		OK	20:22

Data Download Date/Time: 02/17/2022 10:32

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINS,KYLE,L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: ZUNIGA,STEVENS
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: TUM VELASCO

DATE/TIME OF TEST: 01/27/2022 22:43:57

Diagnostics Check		OK	22:35
Air Blank		0.000	22:36
Control Test		0.082	22:36
Air Blank		0.000	22:37
Subject Sample	# 1	0.251	22:38
Breath Volume		4.082	22:38
Air Blank		0.000	22:39
Air Blank		0.000	22:41
Subject Sample	# 2	0.253	22:42
Breath Volume		4.000	22:42
Air Blank		0.000	22:42
Control Test		0.079	22:43
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

Data Download Date/Time: 02/17/2022 10:32

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEWTON,JENNA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BYRD,B
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: O354541989170

DATE/TIME OF TEST: 01/30/2022 03:50:06

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.083	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.199	03:45
Breath Volume		2.375	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:47
Subject Sample	# 2	0.193	03:48
Breath Volume		2.664	03:48
Air Blank		0.000	03:49
Control Test		0.078	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 02/17/2022 10:32

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,JOSEPH,T
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: MILLIGAN,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 01/30/2022 06:47:43

Diagnostics Check		OK	06:40
Air Blank		0.000	06:40
Control Test		0.082	06:41
Air Blank		0.000	06:41
Subject Sample	# 1	0.077	06:42
Breath Volume		1.796	06:42
Air Blank		0.000	06:43
Air Blank		0.000	06:45
Subject Sample	# 2	0.070	06:45
Breath Volume		1.500	06:45
Air Blank		0.000	06:46
Control Test		0.082	06:47
Air Blank		0.000	06:47
Diagnostics Check		OK	06:47

Data Download Date/Time: 02/17/2022 10:32

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARTHY,K,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: ROAN,C
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 01/02/2022 03:21:59

Diagnostics Check		OK	03:14
Air Blank		0.000	03:14
Control Test		0.081	03:15
Air Blank		0.000	03:15
Subject Sample	# 1	0.171	03:17
Breath Volume		1.714	03:17
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.167	03:20
Breath Volume		2.304	03:20
Air Blank		0.000	03:21
Control Test		0.078	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 01/24/2022 16:29

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRABILL,JAMES,H
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: GRABILL,JAMES
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: SAPP

DATE/TIME OF TEST: 01/01/2022 02:59:55

Diagnostics Check		OK	02:52
Air Blank		0.000	02:53
Control Test		0.081	02:53
Air Blank		0.000	02:53
Subject Sample	# 1	0.000	02:55
Breath Volume		2.242	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.000	02:58
Breath Volume		1.839	02:58
Air Blank		0.000	02:58
Control Test		0.079	02:59
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 01/29/2022 09:18

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,STERLING,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: DONNELLY,SEAN
ARRESTING OFFICER AGENCY: Holly Hills PD
LAST AGENCY INSPECTION: 12/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: CANADY

DATE/TIME OF TEST: 01/04/2022 02:06:22

Diagnostics Check		OK	01:52
Air Blank		0.000	01:53
Control Test		0.081	01:53
Air Blank		0.000	01:54
Subject Sample	# 1	0.192	01:55
Breath Volume		1.234	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:57
Subject Sample	# 2	0.171	02:01
Breath Volume		1.222	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:03
Subject Sample	# 3	0.168	02:04
Breath Volume		1.230	02:04
Air Blank		0.000	02:05
Control Test		0.077	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 01/29/2022 09:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: OTHER
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 01/04/2022 20:02:47

Diagnostics Check		OK	19:53
Air Blank		0.000	19:53
Control Test		0.080	19:54
Air Blank		0.000	19:54
Subject Sample	# 1	0.173	19:55
Breath Volume		4.378	19:55
Air Blank		0.000	19:55
Air Blank		0.000	19:57
Subject Sample	# 2	0.138	19:58
Breath Volume		3.261	19:58
Air Blank		0.000	19:58
Air Blank		0.000	20:00
Subject Sample	# 3	0.158	20:01
Breath Volume		3.875	20:01
Air Blank		0.000	20:01
Control Test		0.080	20:02
Air Blank		0.000	20:02
Diagnostics Check		OK	20:02

Data Download Date/Time: 01/29/2022 09:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: OTHER
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 01/04/2022 20:09:39

Diagnostics Check		OK	20:03	
Air Blank		0.000	20:04	
Control Test		0.079	20:04	
Air Blank		0.000	20:05	
Subject Sample	# 1	0.117	20:05	
Breath Volume		2.710	20:05	
Air Blank		0.000	20:06	
Air Blank		0.000	20:08	
Subject Sample	# 2		20:08	Subject Test Refused
Breath Volume		0.000	20:08	
Air Blank		0.000	20:08	
Control Test		0.079	20:09	
Air Blank		0.000	20:09	
Diagnostics Check		OK	20:09	

Data Download Date/Time: 01/29/2022 09:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,STERLING,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: GALBREATH,NOAH
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 17:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: MONTEIRO

DATE/TIME OF TEST: 01/08/2022 18:44:54

Diagnostics Check		OK	18:38
Air Blank		0.000	18:39
Control Test		0.082	18:39
Air Blank		0.000	18:39
Subject Sample	# 1	0.000	18:40
Breath Volume		1.953	18:40
Air Blank		0.000	18:41
Air Blank		0.000	18:43
Subject Sample	# 2	0.000	18:43
Breath Volume		1.554	18:43
Air Blank		0.000	18:43
Control Test		0.080	18:44
Air Blank		0.000	18:44
Diagnostics Check		OK	18:44

Data Download Date/Time: 01/29/2022 09:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONLEY,SHAWNA,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: PRICE,MICHAEL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/16/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 07:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:55

SUBJECT'S LAST NAME: SLAUGHTER

DATE/TIME OF TEST: 01/12/2022 09:22:34

Diagnostics Check		OK	09:16
Air Blank		0.000	09:16
Control Test		0.081	09:16
Air Blank		0.000	09:17
Subject Sample	# 1	0.000	09:18
Breath Volume		2.179	09:18
Air Blank		0.000	09:18
Air Blank		0.000	09:20
Subject Sample	# 2	0.000	09:21
Breath Volume		2.632	09:21
Air Blank		0.000	09:21
Control Test		0.081	09:21
Air Blank		0.000	09:22
Diagnostics Check		OK	09:22

Data Download Date/Time: 01/29/2022 09:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRABILL,JAMES,H
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: GRABILL,JAMES
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 01/22/2022 02:45:03

Diagnostics Check		OK	02:38
Air Blank		0.000	02:39
Control Test		0.082	02:39
Air Blank		0.000	02:40
Subject Sample	# 1	0.137	02:40
Breath Volume		2.082	02:40
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.133	02:43
Breath Volume		2.085	02:43
Air Blank		0.000	02:44
Control Test		0.081	02:44
Air Blank		0.000	02:44
Diagnostics Check		OK	02:45

Data Download Date/Time: 01/29/2022 09:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWANDOWSKI,BRIAN,J
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: COLON,JOSEPH
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 07:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 01/25/2022 08:10:35

Diagnostics Check		OK	08:02	
Air Blank		0.000	08:02	
Control Test		0.081	08:02	
Air Blank		0.000	08:03	
Subject Sample	# 1		08:06	No Sample Provided
Breath Volume		0.000	08:06	
Air Blank		0.000	08:07	
Air Blank		0.000	08:09	
Subject Sample	# 2		08:09	Subject Test Refused
Breath Volume		0.000	08:09	
Air Blank		0.000	08:09	
Control Test		0.082	08:09	
Air Blank		0.000	08:10	
Diagnostics Check		OK	08:10	

Data Download Date/Time: 01/29/2022 09:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWANDOWSKI,BRIAN,J
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: COLON,JOSEPH
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 07:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 01/25/2022 08:20:38

Diagnostics Check		OK	08:13
Air Blank		0.000	08:13
Control Test		0.082	08:14
Air Blank		0.000	08:14
Subject Sample	# 1	0.142	08:15
Breath Volume		1.269	08:15
Air Blank		0.000	08:16
Air Blank		0.000	08:17
Subject Sample	# 2	0.146	08:18
Breath Volume		1.457	08:18
Air Blank		0.000	08:19
Control Test		0.081	08:20
Air Blank		0.000	08:20
Diagnostics Check		OK	08:20

Data Download Date/Time: 01/29/2022 09:19

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORREST,RYAN,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: FORREST,R
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/29/2022 18:26:31

Diagnostics Check		OK	18:19
Air Blank		0.000	18:20
Control Test		0.081	18:20
Air Blank		0.000	18:21
Subject Sample	# 1	0.140	18:21
Breath Volume		2.261	18:21
Air Blank		0.000	18:22
Air Blank		0.000	18:24
Subject Sample	# 2	0.135	18:24
Breath Volume		1.277	18:24
Air Blank		0.000	18:25
Control Test		0.081	18:25
Air Blank		0.000	18:26
Diagnostics Check		OK	18:26

Data Download Date/Time: 03/01/2022 08:43

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: VALDEZ

DATE/TIME OF TEST: 01/14/2022 20:30:37

Diagnostics Check		OK	20:24
Air Blank		0.000	20:24
Control Test		0.078	20:25
Air Blank		0.000	20:25
Subject Sample	# 1	0.233	20:26
Breath Volume		1.613	20:26
Air Blank		0.000	20:26
Air Blank		0.000	20:28
Subject Sample	# 2	0.231	20:28
Breath Volume		1.476	20:28
Air Blank		0.000	20:29
Control Test		0.078	20:30
Air Blank		0.000	20:30
Diagnostics Check		OK	20:30

Data Download Date/Time: 02/02/2022 03:42

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:47

SUBJECT'S LAST NAME: AVILES

DATE/TIME OF TEST: 01/15/2022 17:16:09

Diagnostics Check		OK	17:09
Air Blank		0.000	17:10
Control Test		0.079	17:10
Air Blank		0.000	17:11
Subject Sample	# 1	0.283	17:11
Breath Volume		1.753	17:11
Air Blank		0.000	17:12
Air Blank		0.000	17:14
Subject Sample	# 2	0.275	17:14
Breath Volume		1.203	17:14
Air Blank		0.000	17:15
Control Test		0.078	17:15
Air Blank		0.000	17:16
Diagnostics Check		OK	17:16

Data Download Date/Time: 02/02/2022 03:42

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 01/16/2022 00:12:59

Diagnostics Check		OK	00:10	
Air Blank		0.000	00:10	
Control Test		0.079	00:10	
Air Blank		0.000	00:11	
Subject Sample	# 1		00:11	Subject Test Refused
Breath Volume		0.000	00:11	
Air Blank		0.000	00:12	
Control Test		0.078	00:12	
Air Blank		0.000	00:12	
Diagnostics Check		OK	00:12	

Data Download Date/Time: 02/02/2022 03:42

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: DUKES

DATE/TIME OF TEST: 01/21/2022 23:45:35

Diagnostics Check		OK	23:38
Air Blank		0.000	23:38
Control Test		0.079	23:39
Air Blank		0.000	23:39
Subject Sample	# 1	0.172	23:40
Breath Volume		1.117	23:40
Air Blank		0.000	23:41
Air Blank		0.000	23:42
Subject Sample	# 2	0.179	23:43
Breath Volume		1.160	23:43
Air Blank		0.000	23:44
Control Test		0.079	23:44
Air Blank		0.000	23:45
Diagnostics Check		OK	23:45

Data Download Date/Time: 02/02/2022 03:42

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: MISHLER

DATE/TIME OF TEST: 01/26/2022 19:56:04

Diagnostics Check		OK	19:50
Air Blank		0.000	19:50
Control Test		0.079	19:50
Air Blank		0.000	19:51
Subject Sample	# 1	0.123	19:51
Breath Volume		1.683	19:51
Air Blank		0.000	19:52
Air Blank		0.000	19:54
Subject Sample	# 2	0.120	19:54
Breath Volume		2.074	19:54
Air Blank		0.000	19:55
Control Test		0.078	19:55
Air Blank		0.000	19:55
Diagnostics Check		OK	19:56

Data Download Date/Time: 02/02/2022 03:42

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: BRICENO MARQUEZ

DATE/TIME OF TEST: 01/14/2022 01:41:48

Diagnostics Check		OK	01:29	
Air Blank		0.000	01:29	
Control Test		0.082	01:30	
Air Blank		0.000	01:30	
Subject Sample	# 1	0.151	01:33	Volume Not Met
Breath Volume		0.074	01:33	
Air Blank		0.000	01:34	
Air Blank		0.000	01:36	
Subject Sample	# 2	0.159	01:39	Volume Not Met
Breath Volume		0.164	01:39	
Air Blank		0.000	01:40	
Control Test		0.079	01:40	
Air Blank		0.000	01:41	
Diagnostics Check		Fail	01:41	Voltage/Current Test
Air Blank		0.000	01:41	

Data Download Date/Time: 02/02/2022 03:39

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: BRICENO MARQUEZ

DATE/TIME OF TEST: 01/14/2022 01:48:47

Diagnostics Check	Fail	01:48	Voltage/Current Test
Air Blank	0.000	01:48	

Data Download Date/Time: 02/02/2022 03:39

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ORBETA,J
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD
LAST AGENCY INSPECTION: 01/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: CERDA-IBANEZ

DATE/TIME OF TEST: 01/23/2022 04:05:47

Diagnostics Check		OK	03:59
Air Blank		0.000	03:59
Control Test		0.079	04:00
Air Blank		0.000	04:00
Subject Sample	# 1	0.177	04:01
Breath Volume		1.667	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:03
Subject Sample	# 2	0.175	04:04
Breath Volume		1.656	04:04
Air Blank		0.000	04:04
Control Test		0.078	04:05
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 02/02/2022 03:34

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: KALMBACH

DATE/TIME OF TEST: 01/30/2022 03:45:57

Diagnostics Check		OK	03:37
Air Blank		0.000	03:38
Control Test		0.081	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.136	03:41
Breath Volume		1.234	03:41
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.146	03:44
Breath Volume		2.097	03:44
Air Blank		0.000	03:44
Control Test		0.082	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 02/02/2022 03:34

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HANSELL,ROBERT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: CARABALLO-SALDANA

DATE/TIME OF TEST: 01/31/2022 00:47:55

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.081	00:42
Air Blank		0.000	00:42
Subject Sample	# 1	0.215	00:43
Breath Volume		3.210	00:43
Air Blank		0.000	00:44
Air Blank		0.000	00:45
Subject Sample	# 2	0.197	00:46
Breath Volume		1.929	00:46
Air Blank		0.000	00:47
Control Test		0.077	00:47
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 02/07/2022 11:13

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTAYA,GUILLERMO,
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: CARTAYA,GUILLERMO
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 12/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: BUI
ARREST TIME: 17:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:31

SUBJECT'S LAST NAME: BARATUTE SOCARRAS

DATE/TIME OF TEST: 01/08/2022 18:14:29

Diagnostics Check		OK	18:08
Air Blank		0.000	18:09
Control Test		0.080	18:09
Air Blank		0.000	18:09
Subject Sample	# 1	0.015	18:10
Breath Volume		1.562	18:10
Air Blank		0.000	18:10
Air Blank		0.000	18:12
Subject Sample	# 2	0.015	18:12
Breath Volume		2.214	18:12
Air Blank		0.000	18:13
Control Test		0.079	18:13
Air Blank		0.000	18:14
Diagnostics Check		OK	18:14

Data Download Date/Time: 01/26/2022 14:58

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITT,JESSE,T
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: WHITT,JESSE
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 12/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: BUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: POROSOFF

DATE/TIME OF TEST: 01/21/2022 19:55:25

Diagnostics Check		OK	19:52	
Air Blank		0.000	19:52	
Control Test		0.081	19:53	
Air Blank		0.000	19:53	
Subject Sample	# 1		19:54	Subject Test Refused
Breath Volume		0.000	19:54	
Air Blank		0.000	19:54	
Control Test		0.080	19:54	
Air Blank		0.000	19:55	
Diagnostics Check		OK	19:55	

Data Download Date/Time: 01/26/2022 14:58

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 01/10/2022 21:50:14

Diagnostics Check		OK	21:43	
Air Blank		0.000	21:44	
Control Test		0.081	21:44	
Air Blank		0.000	21:45	
Subject Sample	# 1	0.278	21:45	
Breath Volume		2.605	21:45	
Air Blank		0.000	21:46	
Air Blank		0.000	21:48	
Subject Sample	# 2	0.291	21:48	
Breath Volume		2.382	21:48	
Air Blank		0.000	21:49	
Control Test		0.079	21:49	Interferent Detect
Air Blank		0.000	21:50	

Data Download Date/Time: 02/15/2022 18:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 01/10/2022 22:00:11

Diagnostics Check		OK	21:53	
Air Blank		0.000	21:53	
Control Test		0.080	21:53	
Air Blank		0.000	21:54	
Subject Sample	# 1	0.244	21:54	
Breath Volume		3.265	21:54	
Air Blank		0.000	21:55	
Air Blank		0.000	21:57	
Subject Sample	# 2	0.244	21:58	
Breath Volume		2.226	21:58	
Air Blank		0.000	21:58	
Control Test		0.080	21:59	
Air Blank		0.000	21:59	
Diagnostics Check		Fail	21:59	Temperature Regulation Test
Air Blank		0.000	22:00	

Data Download Date/Time: 02/15/2022 18:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: NUNES

DATE/TIME OF TEST: 01/19/2022 23:31:57

Diagnostics Check		OK	23:24
Air Blank		0.000	23:25
Control Test		0.082	23:25
Air Blank		0.000	23:25
Subject Sample	# 1	0.231	23:26
Breath Volume		2.730	23:26
Air Blank		0.000	23:27
Air Blank		0.000	23:29
Subject Sample	# 2	0.229	23:30
Breath Volume		2.984	23:30
Air Blank		0.000	23:31
Control Test		0.080	23:31
Air Blank		0.000	23:31
Diagnostics Check		OK	23:31

Data Download Date/Time: 02/15/2022 18:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: METCALF

DATE/TIME OF TEST: 01/21/2022 02:30:51

Diagnostics Check		OK	02:23
Air Blank		0.000	02:24
Control Test		0.080	02:24
Air Blank		0.000	02:25
Subject Sample	# 1	0.167	02:25
Breath Volume		2.253	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.173	02:29
Breath Volume		2.726	02:29
Air Blank		0.000	02:29
Control Test		0.080	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 02/15/2022 18:15

Printed: 15 March 2022

COBRA ©1997-2014 CMI, Inc.