

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ELMORE, BENJAMIN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: FOURNIER

DATE/TIME OF TEST: 03/07/2022 21:31:54

Diagnostics Check	OK	21:27	
Air Blank	0.000	21:28	
Control Test	0.079	21:28	
Air Blank	0.000	21:28	
Subject Sample	# 1	21:30	Subject Test Refused
Breath Volume	0.000	21:30	
Air Blank	0.000	21:30	
Control Test	0.079	21:31	
Air Blank	0.000	21:31	
Diagnostics Check	OK	21:31	

Data Download Date/Time: 04/08/2022 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,OBEL,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: VEGA,OBEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 02/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: GUERRA

DATE/TIME OF TEST: 03/04/2022 02:37:45

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.079	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.257	02:33
Breath Volume	1.117	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.262	02:36
Breath Volume	1.628	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 04/11/2022 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,OBEL,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: VEGA,OBEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 02/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: ERKHEMBAATAR

DATE/TIME OF TEST: 03/05/2022 02:52:23

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.080	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.179	02:47
Breath Volume	2.476	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.182	02:50
Breath Volume	3.777	02:50
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 04/11/2022 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 02/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: LEYVA LEYVA

DATE/TIME OF TEST: 03/07/2022 03:11:29

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.080	03:05
Air Blank	0.000	03:05
Subject Sample # 1	0.217	03:07
Breath Volume	2.621	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.212	03:09
Breath Volume	1.597	03:09
Air Blank	0.000	03:10
Control Test	0.079	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 04/11/2022 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNOZ,ROBERT,A
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: MUNOZ,ROBERT
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 02/27/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 03/07/2022 05:34:57

Diagnostics Check	OK	05:30	
Air Blank	0.000	05:31	
Control Test	0.078	05:31	
Air Blank	0.000	05:32	
Subject Sample	# 1	05:33	Subject Test Refused
Breath Volume	0.000	05:33	
Air Blank	0.000	05:34	
Control Test	0.080	05:34	
Air Blank	0.000	05:34	
Diagnostics Check	OK	05:34	

Data Download Date/Time: 04/11/2022 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,DANIEL,A
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: SMITH,DANIEL
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 02/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 11:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: SAMITH

DATE/TIME OF TEST: 03/28/2022 12:49:05

Diagnostics Check	OK	12:43
Air Blank	0.000	12:43
Control Test	0.080	12:43
Air Blank	0.000	12:44
Subject Sample # 1	0.000	12:44
Breath Volume	1.148	12:44
Air Blank	0.000	12:45
Air Blank	0.000	12:47
Subject Sample # 2	0.000	12:47
Breath Volume	1.433	12:47
Air Blank	0.000	12:48
Control Test	0.079	12:48
Air Blank	0.000	12:49
Diagnostics Check	OK	12:49

Data Download Date/Time: 04/11/2022 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE, DAVANO, L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: ALLEYNE, DAVANO
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 05:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: COLON TOLEDO

DATE/TIME OF TEST: 03/18/2022 06:28:36

Diagnostics Check	OK	06:25	
Air Blank	0.000	06:26	
Control Test	0.079	06:26	
Air Blank	0.000	06:27	
Subject Sample	# 1	06:27	Subject Test Refused
Breath Volume	0.000	06:27	
Air Blank	0.000	06:27	
Control Test	0.080	06:28	
Air Blank	0.000	06:28	
Diagnostics Check	OK	06:28	

Data Download Date/Time: 04/28/2022 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,PHILIP,R
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: BLACIO,GUSTAVO
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 16:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: DIFLORIO

DATE/TIME OF TEST: 03/19/2022 17:08:39

Diagnostics Check	OK	17:02
Air Blank	0.000	17:02
Control Test	0.080	17:02
Air Blank	0.000	17:03
Subject Sample # 1	0.112	17:04
Breath Volume	1.484	17:04
Air Blank	0.000	17:04
Air Blank	0.000	17:06
Subject Sample # 2	0.106	17:07
Breath Volume	3.234	17:07
Air Blank	0.000	17:07
Control Test	0.079	17:08
Air Blank	0.000	17:08
Diagnostics Check	OK	17:08

Data Download Date/Time: 04/28/2022 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: SERPA

DATE/TIME OF TEST: 03/25/2022 23:15:55

Diagnostics Check		OK	23:05	
Air Blank		0.000	23:06	
Control Test		0.080	23:06	
Air Blank		0.000	23:06	
Subject Sample	# 1	0.101	23:10	Volume Not Met
Breath Volume		0.066	23:10	
Air Blank		0.000	23:10	
Air Blank		0.000	23:12	
Subject Sample	# 2		23:14	Subject Test Refused
Breath Volume		0.015	23:14	
Air Blank		0.000	23:15	
Control Test		0.079	23:15	
Air Blank		0.000	23:15	
Diagnostics Check		OK	23:15	

Data Download Date/Time: 04/28/2022 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLAN,KEVIN,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: FAUL

DATE/TIME OF TEST: 03/26/2022 19:15:17

Diagnostics Check	OK	19:12	
Air Blank	0.000	19:12	
Control Test	0.080	19:13	
Air Blank	0.000	19:13	
Subject Sample	# 1	19:13	Subject Test Refused
Breath Volume	0.000	19:13	
Air Blank	0.000	19:14	
Control Test	0.080	19:14	
Air Blank	0.000	19:15	
Diagnostics Check	OK	19:15	

Data Download Date/Time: 04/28/2022 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAGOS,GUILLERMO,E
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: LAGOS,GUILLERMO
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: COITO

DATE/TIME OF TEST: 03/27/2022 04:11:42

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.080	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.262	04:07
Breath Volume	2.871	04:07
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.269	04:10
Breath Volume	3.316	04:10
Air Blank	0.000	04:10
Control Test	0.080	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 04/28/2022 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SAVINON,CARLOS
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:46

SUBJECT'S LAST NAME: REAMS

DATE/TIME OF TEST: 03/09/2022 07:16:40

Diagnostics Check	OK	07:10
Air Blank	0.000	07:10
Control Test	0.079	07:10
Air Blank	0.000	07:11
Subject Sample # 1	0.089	07:11
Breath Volume	3.187	07:11
Air Blank	0.000	07:12
Air Blank	0.000	07:14
Subject Sample # 2	0.088	07:15
Breath Volume	3.316	07:15
Air Blank	0.000	07:15
Control Test	0.079	07:16
Air Blank	0.000	07:16
Diagnostics Check	OK	07:16

Data Download Date/Time: 04/06/2022 09:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DIAZ,EDWARD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: LEONOR

DATE/TIME OF TEST: 03/10/2022 22:04:53

Diagnostics Check	OK	21:56
Air Blank	0.000	21:57
Control Test	0.080	21:57
Air Blank	0.000	21:58
Subject Sample # 1	0.202	22:00
Breath Volume	1.738	22:00
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.202	22:03
Breath Volume	3.453	22:03
Air Blank	0.000	22:03
Control Test	0.078	22:04
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 04/06/2022 09:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCEGUERA,YASMANI,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: OCEGUERA,YASMANI
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: PEREZ LEIVA

DATE/TIME OF TEST: 03/18/2022 22:26:59

Diagnostics Check		OK	22:18
Air Blank		0.000	22:18
Control Test		0.079	22:18
Air Blank		0.000	22:19
Subject Sample	# 1	0.204	22:20
Breath Volume		1.246	22:20
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.186	22:25
Breath Volume		1.136	22:25
Air Blank		0.000	22:26
Control Test		0.080	22:26
Air Blank		0.000	22:26
Diagnostics Check		OK	22:26

Data Download Date/Time: 04/06/2022 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEANITE,DUHAMEL,
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: BROWN,ROY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:16

SUBJECT'S LAST NAME: MARADIAGA BLANDON

DATE/TIME OF TEST: 03/09/2022 10:44:03

Diagnostics Check	OK	10:36
Air Blank	0.000	10:37
Control Test	0.080	10:37
Air Blank	0.000	10:37
Subject Sample # 1	0.171	10:39
Breath Volume	1.191	10:39
Air Blank	0.000	10:40
Air Blank	0.000	10:41
Subject Sample # 2	0.174	10:42
Breath Volume	1.683	10:42
Air Blank	0.000	10:43
Control Test	0.077	10:43
Air Blank	0.000	10:43
Diagnostics Check	OK	10:44

Data Download Date/Time: 04/20/2022 02:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH, MATTHEW, T
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: RUDOLPH, MATTHEW
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CONSOLO

DATE/TIME OF TEST: 03/13/2022 04:05:36

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:02	
Control Test	0.080	04:03	
Air Blank	0.000	04:03	
Subject Sample	# 1	04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:04	
Control Test	0.081	04:04	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:05	

Data Download Date/Time: 04/20/2022 02:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: JORDAN,S
ARRESTING OFFICER AGENCY: Miami Springs PD
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:11

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 03/04/2022 09:03:21

Diagnostics Check	OK	08:59	
Air Blank	0.000	09:00	
Control Test	0.082	09:00	
Air Blank	0.000	09:01	
Subject Sample	# 1	09:01	Subject Test Refused
Breath Volume	0.000	09:01	
Air Blank	0.000	09:02	
Control Test	0.081	09:02	
Air Blank	0.000	09:03	
Diagnostics Check	OK	09:03	

Data Download Date/Time: 04/06/2022 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGSTROM,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HARRIS,HEATHER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 06:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:38

SUBJECT'S LAST NAME: PEDROSO-SANTAMARIA

DATE/TIME OF TEST: 03/05/2022 07:23:26

Diagnostics Check	OK	07:22	
Air Blank	0.000	07:23	
		07:23	RFI Detect
Air Blank		07:23	RFI Detect

Data Download Date/Time: 03/23/2022 17:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGSTROM,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HARRIS,HEATHER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 06:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:38

SUBJECT'S LAST NAME: PEDROSO-SANTAMARIA

DATE/TIME OF TEST: 03/05/2022 07:33:19

Diagnostics Check		OK	07:25
Air Blank		0.000	07:25
Control Test		0.078	07:25
Air Blank		0.000	07:26
Subject Sample	# 1	0.026	07:28
Breath Volume		2.117	07:28
Air Blank		0.000	07:28
Air Blank		0.000	07:30
Subject Sample	# 2	0.026	07:31
Breath Volume		1.960	07:31
Air Blank		0.000	07:32
Control Test		0.078	07:32
Air Blank		0.000	07:33
Diagnostics Check		OK	07:33

Data Download Date/Time: 03/23/2022 17:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOTO,ISRAEL,
BREATH TEST OPERATOR AGENCY: FDLE
ARRESTING OFFICER: SOTO,ISRAEL
ARRESTING OFFICER AGENCY: FDLE
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: SYS CHECK
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:44

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 03/29/2022 07:10:47

Diagnostics Check	OK	07:04
Air Blank	0.000	07:05
Control Test	0.081	07:05
Air Blank	0.000	07:06
Subject Sample # 1	0.000	07:06
Breath Volume	2.941	07:06
Air Blank	0.000	07:06
Air Blank	0.000	07:08
Subject Sample # 2	0.000	07:09
Breath Volume	1.875	07:09
Air Blank	0.000	07:09
Control Test	0.082	07:10
Air Blank	0.000	07:10
Diagnostics Check	OK	07:10

Data Download Date/Time: 03/29/2022 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODALL,CORY,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WOODALL,CORY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 03/11/2022 04:24:30

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:15	
Control Test	0.080	04:16	
Air Blank	0.000	04:16	
Subject Sample	# 1 0.153	04:19	
Breath Volume	2.824	04:19	
Air Blank	0.000	04:20	
Air Blank	0.000	04:21	
Subject Sample	# 2	04:23	Sequence Aborted
Breath Volume	0.207	04:23	
Air Blank	0.000	04:24	

Data Download Date/Time: 04/19/2022 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARAFA,YUSUF,A
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: ARAFA,YUSUF
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: POSEY

DATE/TIME OF TEST: 03/20/2022 00:52:22

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.081	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.000	00:47
Breath Volume	1.515	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.000	00:51
Breath Volume	2.421	00:51
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 04/19/2022 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCKEEBY,WADE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ZAMORA

DATE/TIME OF TEST: 03/22/2022 03:00:26

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.215	02:55
Breath Volume	3.085	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.218	02:58
Breath Volume	3.136	02:58
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 04/19/2022 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GILMORE,RICHARD,L

BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)

ARRESTING OFFICER: BARRY,HEATHER

ARRESTING OFFICER AGENCY: FFWCC (Lake City)

LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 08121080A2

DRY GAS EXPIRY DATE: 05/05/2023

VIOLATION CODE: BUI

ARREST TIME: 13:41

GENDER: Male

OBSERVATION PERIOD BEGAN: 14:37

SUBJECT'S LAST NAME: ROENNSPIESS

DATE/TIME OF TEST: 03/02/2022 15:07:39

Diagnostics Check	OK	15:06	
Air Blank		15:07	RFI Detect
Air Blank	0.000	15:07	

Data Download Date/Time: 04/04/2022 12:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,OBEL,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: VEGA,OBEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: LEGON

DATE/TIME OF TEST: 03/12/2022 02:53:06

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:50	
Control Test	0.082	02:51	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:52	
Control Test	0.082	02:52	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 04/11/2022 04:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: South Miami PD
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: ZWIBELMAN

DATE/TIME OF TEST: 03/14/2022 04:13:10

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.082	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.139	04:08
Breath Volume	1.734	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:10
Subject Sample # 2	0.152	04:11
Breath Volume	3.476	04:11
Air Blank	0.000	04:12
Control Test	0.082	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 04/11/2022 04:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 12:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: QUINONEZ

DATE/TIME OF TEST: 03/18/2022 01:26:25

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:23	
Control Test	0.080	01:23	
Air Blank	0.000	01:24	
Subject Sample	# 1	01:24	RFI Detect
Breath Volume	2.554	01:24	
Air Blank	0.000	01:25	
Control Test	0.080	01:25	
Air Blank	0.000	01:26	
Diagnostics Check	OK	01:26	

Data Download Date/Time: 04/11/2022 04:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 12:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: QUINONEZ

DATE/TIME OF TEST: 03/18/2022 01:37:09

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.080	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.172	01:32
Breath Volume	3.105	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:34
Subject Sample # 2	0.176	01:35
Breath Volume	3.914	01:35
Air Blank	0.000	01:36
Control Test	0.080	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 04/11/2022 04:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,OBEL,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: VEGA,OBEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MIGNONE

DATE/TIME OF TEST: 03/18/2022 02:59:23

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.172	02:54
Breath Volume	2.414	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.176	02:57
Breath Volume	2.648	02:57
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 04/11/2022 04:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: South Miami PD
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: SACHEZ OCHOA

DATE/TIME OF TEST: 03/19/2022 04:00:59

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.081	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.174	03:56
Breath Volume	2.117	03:56
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.178	03:59
Breath Volume	2.031	03:59
Air Blank	0.000	04:00
Control Test	0.080	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 04/11/2022 04:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: South Miami PD
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: BAIDAL

DATE/TIME OF TEST: 03/20/2022 05:00:51

Diagnostics Check	OK	04:54
Air Blank	0.000	04:54
Control Test	0.080	04:54
Air Blank	0.000	04:55
Subject Sample # 1	0.204	04:56
Breath Volume	1.855	04:56
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.192	04:59
Breath Volume	1.621	04:59
Air Blank	0.000	04:59
Control Test	0.079	05:00
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 04/11/2022 04:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,CARLOS,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: VEGA,OBEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: VEGA CASTILLO

DATE/TIME OF TEST: 03/20/2022 06:17:28

Diagnostics Check	OK	06:17	
Air Blank		06:17	RFI Detect
Air Blank		06:17	RFI Detect

Data Download Date/Time: 04/11/2022 04:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,CARLOS,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: VEGA,OBEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: VEGA CASTILLO

DATE/TIME OF TEST: 03/20/2022 06:30:56

Diagnostics Check	OK	06:20	
Air Blank	0.000	06:21	
Control Test	0.080	06:21	
Air Blank	0.000	06:21	
Subject Sample	# 1 0.082	06:23	
Breath Volume	1.285	06:23	
Air Blank	0.000	06:24	
Air Blank	0.000	06:26	
Subject Sample	# 2	06:29	Subject Test Refused
Breath Volume	0.113	06:29	
Air Blank	0.000	06:30	
Control Test	0.078	06:30	
Air Blank	0.000	06:30	
Diagnostics Check	OK	06:30	

Data Download Date/Time: 04/11/2022 04:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,CARLOS,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: VEGA,OBEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: VEGA CASTILLO

DATE/TIME OF TEST: 03/20/2022 06:42:38

Diagnostics Check	OK	06:36
Air Blank	0.000	06:36
Control Test	0.079	06:36
Air Blank	0.000	06:37
Subject Sample # 1	0.085	06:37
Breath Volume	3.187	06:37
Air Blank	0.000	06:38
Air Blank	0.000	06:40
Subject Sample # 2	0.073	06:41
Breath Volume	2.015	06:41
Air Blank	0.000	06:41
Control Test	0.080	06:42
Air Blank	0.000	06:42
Diagnostics Check	OK	06:42

Data Download Date/Time: 04/11/2022 04:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,CARLOS,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: RESTREPO

DATE/TIME OF TEST: 03/21/2022 03:29:51

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:18	
Control Test		0.081	03:19	
Air Blank		0.000	03:19	
Subject Sample	# 1	0.099	03:22	Volume Not Met
Breath Volume		0.089	03:22	
Air Blank		0.000	03:23	
Air Blank		0.000	03:25	
Subject Sample	# 2	0.085	03:28	Volume Not Met
Breath Volume		0.042	03:28	
Air Blank		0.000	03:28	
Control Test		0.079	03:29	
Air Blank		0.000	03:29	
Diagnostics Check		OK	03:29	

Data Download Date/Time: 04/11/2022 04:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,CARLOS,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: RESTREPO

DATE/TIME OF TEST: 03/21/2022 03:40:31

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:36	
Control Test	0.079	03:36	
Air Blank	0.000	03:36	
Subject Sample	# 1	03:39	Subject Test Refused
Breath Volume	0.164	03:39	
Air Blank	0.000	03:39	
Control Test	0.081	03:39	
Air Blank	0.000	03:40	
Diagnostics Check	OK	03:40	

Data Download Date/Time: 04/11/2022 04:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: South Miami PD
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 03/22/2022 01:09:48

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.081	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.139	01:03
Breath Volume		1.175	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.137	01:08
Breath Volume		1.519	01:08
Air Blank		0.000	01:08
Control Test		0.080	01:09
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 04/11/2022 04:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,OBEL,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: VEGA,OBEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: MATOS GUTIERREZ

DATE/TIME OF TEST: 03/24/2022 04:27:37

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:25	
Control Test	0.081	04:25	
Air Blank	0.000	04:25	
Subject Sample	# 1	04:26	Subject Test Refused
Breath Volume	0.000	04:26	
Air Blank	0.000	04:26	
Control Test	0.081	04:27	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

Data Download Date/Time: 04/11/2022 04:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,OBEL,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: ROJAS SALAS

DATE/TIME OF TEST: 03/26/2022 02:32:54

Diagnostics Check	OK	02:32	
Air Blank		02:32	RFI Detect
Air Blank	0.000	02:32	

Data Download Date/Time: 04/11/2022 04:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,OBEL,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: ROJAS SALAS

DATE/TIME OF TEST: 03/26/2022 02:43:32

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.083	02:36
Air Blank		0.000	02:36
Subject Sample	# 1	0.212	02:37
Breath Volume		2.457	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.204	02:41
Breath Volume		2.777	02:41
Air Blank		0.000	02:42
Control Test		0.080	02:42
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 04/11/2022 04:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,OBEL,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ROJAS SALAS

DATE/TIME OF TEST: 03/26/2022 03:40:17

Diagnostics Check		OK	03:27	
Air Blank		0.000	03:27	
Control Test		0.082	03:28	
Air Blank		0.000	03:28	
Subject Sample	# 1	0.150	03:31	Volume Not Met
Breath Volume		0.230	03:31	
Air Blank		0.000	03:32	
Air Blank		0.000	03:34	
Subject Sample	# 2	0.176	03:34	
Breath Volume		1.726	03:34	
Air Blank		0.000	03:35	
Air Blank		0.000	03:37	
Subject Sample	# 3	0.186	03:38	
Breath Volume		2.433	03:38	
Air Blank		0.000	03:39	
Control Test		0.080	03:39	
Air Blank		0.000	03:40	
Diagnostics Check		OK	03:40	

Data Download Date/Time: 04/11/2022 04:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,OBEL,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: VEGA,OBEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: MCNAIR JR

DATE/TIME OF TEST: 03/27/2022 02:32:06

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.082	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.159	02:27
Breath Volume	2.976	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.162	02:30
Breath Volume	2.664	02:30
Air Blank	0.000	02:31
Control Test	0.081	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 04/11/2022 04:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: South Miami PD
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: KIERNAN

DATE/TIME OF TEST: 03/28/2022 04:34:28

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:32	
Control Test	0.081	04:32	
Air Blank	0.000	04:32	
Subject Sample	# 1	04:33	Subject Test Refused
Breath Volume	0.000	04:33	
Air Blank	0.000	04:33	
Control Test	0.081	04:33	
Air Blank	0.000	04:34	
Diagnostics Check	OK	04:34	

Data Download Date/Time: 04/11/2022 04:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNOZ,ROBERT,A
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: RODRIGUEZ,DARIEN
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: DUI
ARREST TIME: 06:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:54

SUBJECT'S LAST NAME: ARIAS CAMACHO

DATE/TIME OF TEST: 03/28/2022 07:29:15

Diagnostics Check	OK	07:21
Air Blank	0.000	07:22
Control Test	0.082	07:22
Air Blank	0.000	07:22
Subject Sample # 1	0.174	07:24
Breath Volume	2.765	07:24
Air Blank	0.000	07:24
Air Blank	0.000	07:26
Subject Sample # 2	0.176	07:27
Breath Volume	1.933	07:27
Air Blank	0.000	07:28
Control Test	0.082	07:28
Air Blank	0.000	07:29
Diagnostics Check	OK	07:29

Data Download Date/Time: 04/11/2022 04:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERONI, KRISTI,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: NERONI, KRISTI
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 15:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 03/03/2022 17:56:50

Diagnostics Check	OK	17:50
Air Blank	0.000	17:50
Control Test	0.078	17:51
Air Blank	0.000	17:51
Subject Sample # 1	0.059	17:52
Breath Volume	4.101	17:52
Air Blank	0.000	17:52
Air Blank	0.000	17:54
Subject Sample # 2	0.076	17:55
Breath Volume	2.738	17:55
Air Blank	0.000	17:55
Control Test	0.078	17:56
Air Blank	0.000	17:56
Diagnostics Check	OK	17:56

Data Download Date/Time: 03/23/2022 09:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,AUSTIN,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: SELLER,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: GASPAR

DATE/TIME OF TEST: 03/01/2022 00:42:36

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.077	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.128	00:38
Breath Volume	1.992	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.126	00:41
Breath Volume	2.652	00:41
Air Blank	0.000	00:41
Control Test	0.078	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 03/11/2022 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUAREZ,J.,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: WALKER,J.
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 03/03/2022 01:15:51

Diagnostics Check		OK	01:09
Air Blank		0.000	01:09
Control Test		0.078	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.202	01:11
Breath Volume		1.460	01:11
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.211	01:14
Breath Volume		2.234	01:14
Air Blank		0.000	01:14
Control Test		0.079	01:15
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 03/11/2022 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MELODY,B
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 02/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: LONGE

DATE/TIME OF TEST: 03/03/2022 21:12:53

Diagnostics Check	OK	21:04
Air Blank	0.000	21:05
Control Test	0.079	21:05
Air Blank	0.000	21:05
Subject Sample # 1	0.121	21:08
Breath Volume	1.671	21:08
Air Blank	0.000	21:09
Air Blank	0.000	21:10
Subject Sample # 2	0.115	21:11
Breath Volume	1.246	21:11
Air Blank	0.000	21:11
Control Test	0.075	21:12
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

Data Download Date/Time: 03/11/2022 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYRNIOS,AUSTIN,W
BREATH TEST OPERATOR AGENCY: NAPLES PD
ARRESTING OFFICER: MELODY,B
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 02/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: KILBURN

DATE/TIME OF TEST: 03/05/2022 23:54:57

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.078	23:45
Air Blank	0.000	23:46
Subject Sample # 1	0.227	23:47
Breath Volume	1.105	23:47
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.197	23:50
Breath Volume	1.597	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:52
Subject Sample # 3	0.202	23:53
Breath Volume	2.101	23:53
Air Blank	0.000	23:54
Control Test	0.079	23:54
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 03/11/2022 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: GOMEZLOPEZ

DATE/TIME OF TEST: 03/06/2022 00:37:59

Diagnostics Check		OK	00:31
Air Blank		0.000	00:32
Control Test		0.078	00:32
Air Blank		0.000	00:33
Subject Sample	# 1	0.259	00:33
Breath Volume		2.718	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:35
Subject Sample	# 2	0.257	00:36
Breath Volume		2.773	00:36
Air Blank		0.000	00:37
Control Test		0.078	00:37
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 03/11/2022 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: GEARY

DATE/TIME OF TEST: 03/10/2022 18:13:38

Diagnostics Check	OK	18:07
Air Blank	0.000	18:07
Control Test	0.078	18:07
Air Blank	0.000	18:08
Subject Sample # 1	0.165	18:09
Breath Volume	2.218	18:09
Air Blank	0.000	18:09
Air Blank	0.000	18:11
Subject Sample # 2	0.164	18:12
Breath Volume	2.164	18:12
Air Blank	0.000	18:12
Control Test	0.078	18:13
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

Data Download Date/Time: 03/11/2022 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GROSSKLAS,DENNIS
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/10/2022 20:33:34

Diagnostics Check	OK	20:26
Air Blank	0.000	20:27
Control Test	0.078	20:27
Air Blank	0.000	20:28
Subject Sample # 1	0.000	20:28
Breath Volume	2.296	20:28
Air Blank	0.000	20:29
Air Blank	0.000	20:31
Subject Sample # 2	0.000	20:32
Breath Volume	1.335	20:32
Air Blank	0.000	20:32
Control Test	0.078	20:32
Air Blank	0.000	20:33
Diagnostics Check	OK	20:33

Data Download Date/Time: 03/11/2022 12:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRICE,MARCUS,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: TARAZONA,T
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 03/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: GARCIADIAZ

DATE/TIME OF TEST: 03/13/2022 01:21:56

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:15	
Control Test	0.080	01:15	
Air Blank	0.000	01:16	
Subject Sample # 1	0.353	01:17	
Breath Volume	3.113	01:17	
Air Blank	0.000	01:17	
Air Blank	0.000	01:19	
Subject Sample # 2		01:20	Slope Not Met
Breath Volume	1.398	01:20	
Air Blank	0.000	01:21	
Control Test	0.080	01:21	
Air Blank	0.000	01:21	
Diagnostics Check	OK	01:21	

Data Download Date/Time: 04/08/2022 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLELLAND, WILLIAM, R
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: TARAZONA, B
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 03/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: GARCIADIAZ

DATE/TIME OF TEST: 03/13/2022 01:34:40

Diagnostics Check		OK	01:27
Air Blank		0.000	01:28
Control Test		0.079	01:28
Air Blank		0.000	01:29
Subject Sample	# 1	0.361	01:29
Breath Volume		2.210	01:29
Air Blank		0.000	01:30
Air Blank		0.000	01:31
Subject Sample	# 2	0.344	01:32
Breath Volume		2.160	01:32
Air Blank		0.000	01:33
Control Test		0.079	01:34
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 04/08/2022 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUONPANE,V,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BUONPANE,V
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 03/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: NONSANT

DATE/TIME OF TEST: 03/16/2022 19:37:45

Diagnostics Check	OK	19:30
Air Blank	0.000	19:31
Control Test	0.078	19:31
Air Blank	0.000	19:32
Subject Sample # 1	0.239	19:33
Breath Volume	1.410	19:33
Air Blank	0.000	19:33
Air Blank	0.000	19:35
Subject Sample # 2	0.239	19:36
Breath Volume	1.781	19:36
Air Blank	0.000	19:36
Control Test	0.079	19:37
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time: 04/08/2022 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLELLAND, WILLIAM, R
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: TARAZONA, B
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 03/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: LOPEZLOPEZ

DATE/TIME OF TEST: 03/17/2022 03:34:33

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.079	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.143	03:30
Breath Volume	3.226	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.141	03:32
Breath Volume	3.335	03:32
Air Blank	0.000	03:33
Control Test	0.079	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 04/08/2022 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,URIEL,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CAROZZA,A
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 03/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: DIAZ MARTIN

DATE/TIME OF TEST: 03/31/2022 22:25:24

Diagnostics Check	OK	22:18
Air Blank	0.000	22:19
Control Test	0.079	22:19
Air Blank	0.000	22:20
Subject Sample # 1	0.170	22:20
Breath Volume	3.027	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.175	22:23
Breath Volume	2.015	22:23
Air Blank	0.000	22:24
Control Test	0.079	22:24
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 04/08/2022 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMM,JAMES,L
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: YOUNG,ANDREW
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 02/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: CHASTAIN

DATE/TIME OF TEST: 03/04/2022 20:07:14

Diagnostics Check	OK	20:01	
Air Blank	0.000	20:01	
Control Test	0.080	20:02	
Air Blank	0.000	20:02	
Subject Sample	# 1 0.101	20:05	Volume Not Met
Breath Volume	0.574	20:05	
Air Blank		20:06	Purge Fail
Air Blank		20:07	Purge Fail

Data Download Date/Time: 03/29/2022 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMM,JAMES,L
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: YOUNG,ANDREW
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 02/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: CHASTAIN

DATE/TIME OF TEST: 03/04/2022 20:21:45

Diagnostics Check		OK	20:10	
Air Blank		0.000	20:10	
Control Test		0.079	20:10	
Air Blank		0.000	20:11	
Subject Sample	# 1	0.113	20:14	Volume Not Met
Breath Volume		0.046	20:14	
Air Blank		0.000	20:15	
Air Blank		0.000	20:17	
Subject Sample	# 2	0.084	20:20	Volume Not Met
Breath Volume		0.371	20:20	
Air Blank		0.000	20:20	
Control Test		0.081	20:21	
Air Blank		0.000	20:21	
Diagnostics Check		OK	20:21	

Data Download Date/Time: 03/29/2022 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,H
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: REGISTER,MIKEL
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 02/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: BRYAN

DATE/TIME OF TEST: 03/05/2022 16:59:19

Diagnostics Check	OK	16:51	
Air Blank	0.000	16:51	
Control Test	0.079	16:52	
Air Blank	0.000	16:52	
Subject Sample	# 1 0.245	16:52	
Breath Volume	1.214	16:52	
Air Blank	0.000	16:53	
Air Blank	0.000	16:55	
Subject Sample	# 2	16:57	Subject Test Refused
Breath Volume	0.148	16:57	
Air Blank	0.000	16:58	
Control Test	0.080	16:58	
Air Blank	0.000	16:59	
Diagnostics Check	OK	16:59	

Data Download Date/Time: 03/29/2022 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORDON,ROLF,E
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: PATERSON,JAKE
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 02/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: MAULDING

DATE/TIME OF TEST: 03/10/2022 01:56:44

Diagnostics Check		OK	01:48
Air Blank		0.000	01:48
Control Test		0.079	01:49
Air Blank		0.000	01:49
Subject Sample	# 1	0.091	01:51
Breath Volume		4.910	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.090	01:55
Breath Volume		4.445	01:55
Air Blank		0.000	01:55
Control Test		0.079	01:56
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 03/29/2022 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POSEY,COLE,J
BREATH TEST OPERATOR AGENCY: Carrabelle PD
ARRESTING OFFICER: PETERSON,LONI
ARRESTING OFFICER AGENCY: Carrabelle PD
LAST AGENCY INSPECTION: 02/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: POSEY

DATE/TIME OF TEST: 03/13/2022 16:22:04

Diagnostics Check	OK	16:16
Air Blank	0.000	16:16
Control Test	0.079	16:16
Air Blank	0.000	16:17
Subject Sample # 1	0.121	16:17
Breath Volume	3.437	16:17
Air Blank	0.000	16:18
Air Blank	0.000	16:20
Subject Sample # 2	0.116	16:20
Breath Volume	2.144	16:20
Air Blank	0.000	16:21
Control Test	0.080	16:21
Air Blank	0.000	16:21
Diagnostics Check	OK	16:22

Data Download Date/Time: 03/29/2022 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALZA, PAOLOGENE, R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ADAMCZYK, ANDY
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 03/08/2022 18:24:22

Diagnostics Check	OK	18:17	
Air Blank	0.000	18:18	
Control Test	0.077	18:18	
Air Blank	0.000	18:19	
Subject Sample # 1	0.097	18:19	
Breath Volume	1.773	18:19	
Air Blank	0.000	18:20	
Air Blank	0.000	18:22	
Subject Sample # 2	0.090	18:22	
Breath Volume	3.097	18:22	
Air Blank	0.000	18:23	
Control Test	0.073	18:23	Control Outside Tolerance
Air Blank	0.000	18:24	

Data Download Date/Time: 03/10/2022 14:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHEA, COLBY, R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BETHEA, COLBY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HARRELL JR

DATE/TIME OF TEST: 03/20/2022 03:10:39

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.076	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.164	03:06
Breath Volume	2.546	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.156	03:09
Breath Volume	2.468	03:09
Air Blank	0.000	03:09
Control Test	0.077	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 04/27/2022 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTON,JEREMY,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BARTON,JEREMY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SCHAFER

DATE/TIME OF TEST: 03/13/2022 04:31:36

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.081	04:25
Air Blank		0.000	04:26
Subject Sample	# 1	0.266	04:27
Breath Volume		2.730	04:27
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.257	04:30
Breath Volume		2.054	04:30
Air Blank		0.000	04:30
Control Test		0.082	04:30
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 04/14/2022 14:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILCOX,DAVID,B
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: MCCRAY,JAARON
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: BAUTISTA

DATE/TIME OF TEST: 03/23/2022 00:48:59

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:46	
Control Test	0.081	00:46	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:47	Subject Test Refused
Breath Volume	0.000	00:47	
Air Blank	0.000	00:48	
Control Test	0.080	00:48	
Air Blank	0.000	00:48	
Diagnostics Check	OK	00:48	

Data Download Date/Time: 04/14/2022 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,R,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: FISCHER,A
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: ODZE

DATE/TIME OF TEST: 03/28/2022 01:05:12

Diagnostics Check	OK	00:58
Air Blank	0.000	00:59
Control Test	0.081	00:59
Air Blank	0.000	01:00
Subject Sample # 1	0.089	01:00
Breath Volume	2.742	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.087	01:03
Breath Volume	2.046	01:03
Air Blank	0.000	01:04
Control Test	0.082	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 04/14/2022 14:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILCOX,DAVID,B
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: MCCRAY,JAARON
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: BAUTISTA

DATE/TIME OF TEST: 03/23/2022 00:22:38

Diagnostics Check	OK	00:19	
Air Blank	0.000	00:20	
Control Test	0.080	00:20	
Air Blank	0.000	00:21	
Subject Sample	# 1	00:21	Subject Test Refused
Breath Volume	0.000	00:21	
Air Blank	0.000	00:21	
Control Test	0.080	00:22	
Air Blank	0.000	00:22	
Diagnostics Check	OK	00:22	

Data Download Date/Time: 04/14/2022 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILCOX,DAVID,B
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: MCCRAY,JAARON
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: BAUTISTA

DATE/TIME OF TEST: 03/23/2022 00:28:50

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.081	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:27	Subject Test Refused
Breath Volume	0.000	00:27	
Air Blank	0.000	00:27	
Control Test	0.080	00:28	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:28	

Data Download Date/Time: 04/14/2022 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAUFFMAN, MITCHELL, P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: KAUFFMAN, MITCHELL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: RILES

DATE/TIME OF TEST: 03/12/2022 21:11:23

Diagnostics Check	OK	21:05
Air Blank	0.000	21:05
Control Test	0.083	21:06
Air Blank	0.000	21:06
Subject Sample # 1	0.245	21:06
Breath Volume	1.417	21:06
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.233	21:09
Breath Volume	1.855	21:09
Air Blank	0.000	21:10
Control Test	0.081	21:10
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

Data Download Date/Time: 04/05/2022 12:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO, ANTHONY, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: ROLLINS

DATE/TIME OF TEST: 03/18/2022 00:51:21

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.082	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.112	00:46
Breath Volume	2.187	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.109	00:49
Breath Volume	2.234	00:49
Air Blank	0.000	00:50
Control Test	0.082	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 04/05/2022 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPANGLER,GARY,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MATTHEWS,DUSTIN
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 02/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 03/08/2022 23:08:23

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:06	
Control Test	0.079	23:06	
Air Blank	0.000	23:06	
Subject Sample	# 1	23:07	Subject Test Refused
Breath Volume	0.000	23:07	
Air Blank	0.000	23:07	
Control Test	0.079	23:07	
Air Blank	0.000	23:08	
Diagnostics Check	OK	23:08	

Data Download Date/Time: 04/04/2022 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:22

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/17/2022 09:48:03

Diagnostics Check	OK	09:42
Air Blank	0.000	09:42
Control Test	0.079	09:43
Air Blank	0.000	09:43
Subject Sample # 1	0.000	09:43
Breath Volume	1.605	09:43
Air Blank	0.000	09:44
Air Blank	0.000	09:46
Subject Sample # 2	0.000	09:46
Breath Volume	2.066	09:46
Air Blank	0.000	09:47
Control Test	0.079	09:47
Air Blank	0.000	09:47
Diagnostics Check	OK	09:48

Data Download Date/Time: 04/04/2022 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ELLER,ASHLEY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: VASQUEZ FUENTES

DATE/TIME OF TEST: 03/05/2022 22:58:42

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.079	22:52
Air Blank	0.000	22:52
Subject Sample # 1	0.277	22:53
Breath Volume	1.382	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:55
Subject Sample # 2	0.295	22:57
Breath Volume	1.390	22:57
Air Blank	0.000	22:57
Control Test	0.080	22:58
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 03/10/2022 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PREDELUS,SMITH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: MOLINAPONCEDELEON

DATE/TIME OF TEST: 03/02/2022 23:07:40

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.081	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.170	23:03
Breath Volume	3.203	23:03
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.164	23:06
Breath Volume	3.097	23:06
Air Blank	0.000	23:06
Control Test	0.080	23:07
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 03/10/2022 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MERCADO,MELISSA
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 03/03/2022 02:46:08

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.081	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.085	02:40
Breath Volume	1.515	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.079	02:44
Breath Volume	1.535	02:44
Air Blank	0.000	02:45
Control Test	0.081	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 03/10/2022 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIDOMENICO,GINO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: BRUNOT

DATE/TIME OF TEST: 03/04/2022 02:12:47

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.078	02:05
Air Blank		0.000	02:06
Subject Sample	# 1	0.145	02:08
Breath Volume		1.367	02:08
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.150	02:11
Breath Volume		1.867	02:11
Air Blank		0.000	02:11
Control Test		0.078	02:12
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 03/10/2022 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS, TABITHA, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIDOMENICO, GINO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: GUIDA

DATE/TIME OF TEST: 03/04/2022 23:17:01

Diagnostics Check		OK	23:08	
Air Blank		0.000	23:09	
Control Test		0.080	23:09	
Air Blank		0.000	23:09	
Subject Sample	# 1	0.202	23:13	Volume Not Met
Breath Volume		0.218	23:13	
Air Blank		0.000	23:13	
Air Blank		0.000	23:15	
Subject Sample	# 2		23:15	Subject Test Refused
Breath Volume		0.000	23:15	
Air Blank		0.000	23:16	
Control Test		0.078	23:16	
Air Blank		0.000	23:16	
Diagnostics Check		OK	23:17	

Data Download Date/Time: 03/10/2022 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MEDINA,JAVIER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 03/05/2022 04:14:08

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.172	04:09
Breath Volume	1.535	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.177	04:12
Breath Volume	1.687	04:12
Air Blank	0.000	04:13
Control Test	0.078	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 03/10/2022 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CREWS,CURT
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 03/05/2022 05:44:17

Diagnostics Check	OK	05:41	
Air Blank	0.000	05:41	
Control Test	0.080	05:42	
Air Blank	0.000	05:42	
Subject Sample	# 1	05:42	Subject Test Refused
Breath Volume	0.000	05:42	
Air Blank	0.000	05:43	
Control Test	0.080	05:43	
Air Blank	0.000	05:44	
Diagnostics Check	OK	05:44	

Data Download Date/Time: 03/10/2022 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RODRIGUEZ,WILBERTO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: VEERANI

DATE/TIME OF TEST: 03/06/2022 00:41:18

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.079	00:36
Air Blank	0.000	00:36
Subject Sample # 1	0.203	00:37
Breath Volume	2.179	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.206	00:39
Breath Volume	2.226	00:39
Air Blank	0.000	00:40
Control Test	0.078	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 03/10/2022 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BERRIOS,MARK
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: DE JESUS VERAS

DATE/TIME OF TEST: 03/06/2022 01:37:07

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.056	01:32
Breath Volume	1.636	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.052	01:35
Breath Volume	1.191	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 03/10/2022 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCHOEN,ELLIS
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: GALEOTO

DATE/TIME OF TEST: 03/06/2022 23:22:03

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.080	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.178	23:17
Breath Volume	1.148	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.177	23:20
Breath Volume	1.804	23:20
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:22

Data Download Date/Time: 03/10/2022 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CREWS,CURT
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 03/07/2022 02:26:44

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.184	02:22
Breath Volume		2.441	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:24
Subject Sample	# 2	0.181	02:25
Breath Volume		2.343	02:25
Air Blank		0.000	02:25
Control Test		0.080	02:26
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 03/10/2022 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MELLA,MELLISA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ORELUS

DATE/TIME OF TEST: 03/07/2022 02:39:25

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.077	02:34
Breath Volume	2.378	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.071	02:37
Breath Volume	1.261	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 03/10/2022 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE,ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: COSTE MORALES

DATE/TIME OF TEST: 03/07/2022 04:02:24

Diagnostics Check		OK	03:56
Air Blank		0.000	03:56
Control Test		0.081	03:57
Air Blank		0.000	03:57
Subject Sample	# 1	0.110	03:58
Breath Volume		1.269	03:58
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.112	04:00
Breath Volume		1.179	04:00
Air Blank		0.000	04:01
Control Test		0.080	04:01
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 03/10/2022 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RODRIGUEZ,WILL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 03/08/2022 00:08:06

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.080	00:02
Air Blank	0.000	00:03
Subject Sample # 1	0.000	00:03
Breath Volume	2.246	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.000	00:06
Breath Volume	3.390	00:06
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 03/10/2022 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: BATISTA

DATE/TIME OF TEST: 03/09/2022 01:59:20

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.080	01:53
Air Blank	0.000	01:54
Subject Sample # 1	0.000	01:55
Breath Volume	3.140	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.000	01:57
Breath Volume	3.078	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 03/10/2022 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GRACE,JERMAINE
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: STONEY

DATE/TIME OF TEST: 03/09/2022 02:25:06

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.080	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.105	02:20
Breath Volume	1.257	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.104	02:23
Breath Volume	1.339	02:23
Air Blank	0.000	02:24
Control Test	0.080	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 03/10/2022 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FIGUEREDO,CHRISTIAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: MORTON

DATE/TIME OF TEST: 03/09/2022 23:15:13

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:11	
Control Test	0.079	23:11	
Air Blank	0.000	23:12	
Subject Sample	# 1	23:13	Subject Test Refused
Breath Volume	0.000	23:13	
Air Blank	0.000	23:14	
Control Test	0.080	23:14	
Air Blank	0.000	23:15	
Diagnostics Check	OK	23:15	

Data Download Date/Time: 03/10/2022 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATIAS,ROLANDO
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: SACARIAS GABRIEL

DATE/TIME OF TEST: 03/10/2022 02:17:25

Diagnostics Check		OK	02:04	
Air Blank		0.000	02:05	
Control Test		0.080	02:05	
Air Blank		0.000	02:06	
Subject Sample	# 1	0.166	02:09	Volume Not Met
Breath Volume		0.574	02:09	
Air Blank		0.000	02:09	
Air Blank		0.000	02:11	
Subject Sample	# 2	0.167	02:12	
Breath Volume		1.687	02:12	
Air Blank		0.000	02:13	
Air Blank		0.000	02:14	
Subject Sample	# 3	0.161	02:15	
Breath Volume		1.207	02:15	
Air Blank		0.000	02:16	
Control Test		0.078	02:16	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 03/10/2022 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NELSON,GREGG
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/10/2022 03:18:12

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.080	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.000	03:12
Breath Volume	1.187	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.000	03:16
Breath Volume	1.625	03:16
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 03/10/2022 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATIAS,ROLANDO
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: DIAZ LOPEZ

DATE/TIME OF TEST: 03/11/2022 00:48:08

Diagnostics Check	OK	00:44	
Air Blank	0.000	00:44	
Control Test	0.080	00:44	
Air Blank	0.000	00:45	
Subject Sample	# 1	00:46	Slope Not Met
Breath Volume	0.687	00:46	
Air Blank	0.000	00:47	
Control Test	0.081	00:47	
Air Blank	0.000	00:48	
Diagnostics Check	OK	00:48	

Data Download Date/Time: 04/07/2022 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATIAS,ROLANDO
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: DIAZ LOPEZ

DATE/TIME OF TEST: 03/11/2022 01:25:07

Diagnostics Check		OK	01:17
Air Blank		0.000	01:18
Control Test		0.079	01:18
Air Blank		0.000	01:19
Subject Sample	# 1	0.248	01:19
Breath Volume		1.597	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.267	01:23
Breath Volume		1.195	01:23
Air Blank		0.000	01:24
Control Test		0.080	01:24
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 04/07/2022 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 03/11/2022 19:15:58

Diagnostics Check	OK	19:10
Air Blank	0.000	19:10
Control Test	0.080	19:11
Air Blank	0.000	19:11
Subject Sample # 1	0.000	19:11
Breath Volume	2.402	19:11
Air Blank	0.000	19:12
Air Blank	0.000	19:14
Subject Sample # 2	0.000	19:14
Breath Volume	2.437	19:14
Air Blank	0.000	19:15
Control Test	0.080	19:15
Air Blank	0.000	19:15
Diagnostics Check	OK	19:15

Data Download Date/Time: 04/07/2022 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WATKING,ANTHONY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: LOPEZAGUSTIN

DATE/TIME OF TEST: 03/12/2022 15:30:30

Diagnostics Check	OK	15:19
Air Blank	0.000	15:20
Control Test	0.080	15:20
Air Blank	0.000	15:21
Subject Sample # 1	0.261	15:22
Breath Volume	1.847	15:22
Air Blank	0.000	15:22
Air Blank	0.000	15:24
Subject Sample # 2	0.238	15:25
Breath Volume	1.753	15:25
Air Blank	0.000	15:26
Air Blank	0.000	15:28
Subject Sample # 3	0.270	15:29
Breath Volume	2.800	15:29
Air Blank	0.000	15:29
Control Test	0.079	15:29
Air Blank	0.000	15:30
Diagnostics Check	OK	15:30

Data Download Date/Time: 04/07/2022 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JAMES,ANTONIO
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: MOSELEY

DATE/TIME OF TEST: 03/12/2022 20:37:51

Diagnostics Check	OK	20:24	
Air Blank	0.000	20:25	
Control Test	0.081	20:25	
Air Blank	0.000	20:26	
Subject Sample # 1	0.163	20:27	
Breath Volume	1.296	20:27	
Air Blank	0.000	20:28	
Air Blank	0.000	20:30	
Subject Sample # 2	0.143	20:33	Volume Not Met
Breath Volume	0.695	20:33	
Air Blank	0.000	20:33	
Air Blank	0.000	20:35	
Subject Sample # 3	0.153	20:36	
Breath Volume	1.226	20:36	
Air Blank	0.000	20:36	
Control Test	0.079	20:37	
Air Blank	0.000	20:37	
Diagnostics Check	OK	20:37	

Data Download Date/Time: 04/07/2022 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DANIEL,CONNELL
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: MATEO THOMAS

DATE/TIME OF TEST: 03/12/2022 23:49:11

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.079	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.177	23:44
Breath Volume	2.753	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.186	23:47
Breath Volume	2.804	23:47
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 04/07/2022 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DUFFUS,DAVID
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: FORNEY

DATE/TIME OF TEST: 03/14/2022 05:09:48

Diagnostics Check		OK	05:03
Air Blank		0.000	05:04
Control Test		0.080	05:04
Air Blank		0.000	05:05
Subject Sample	# 1	0.145	05:05
Breath Volume		1.656	05:05
Air Blank		0.000	05:06
Air Blank		0.000	05:07
Subject Sample	# 2	0.153	05:08
Breath Volume		2.027	05:08
Air Blank		0.000	05:08
Control Test		0.080	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 04/07/2022 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARRISON,ELIJAH
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: SANTIAGO NAZARIO

DATE/TIME OF TEST: 03/14/2022 06:09:06

Diagnostics Check	OK	06:02
Air Blank	0.000	06:03
Control Test	0.081	06:03
Air Blank	0.000	06:04
Subject Sample # 1	0.106	06:04
Breath Volume	2.171	06:04
Air Blank	0.000	06:05
Air Blank	0.000	06:07
Subject Sample # 2	0.100	06:07
Breath Volume	1.546	06:07
Air Blank	0.000	06:08
Control Test	0.080	06:08
Air Blank	0.000	06:09
Diagnostics Check	OK	06:09

Data Download Date/Time: 04/07/2022 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARSEILLE,GERRY
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: PLATT

DATE/TIME OF TEST: 03/15/2022 03:48:21

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.203	03:43
Breath Volume	3.695	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.204	03:46
Breath Volume	2.882	03:46
Air Blank	0.000	03:47
Control Test	0.079	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 04/07/2022 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CHARLES,DYON
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 03/18/2022 01:18:57

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.080	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.091	01:14
Breath Volume	2.550	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.087	01:17
Breath Volume	2.593	01:17
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 04/07/2022 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VALLE,OMAR
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: DELACRUZ

DATE/TIME OF TEST: 03/18/2022 02:07:06

Diagnostics Check		OK	01:51	
Air Blank		0.000	01:51	
Control Test		0.079	01:52	
Air Blank		0.000	01:52	
Subject Sample	# 1	0.087	01:55	Volume Not Met
Breath Volume		0.109	01:55	
Air Blank		0.000	01:56	
Air Blank		0.000	01:58	
Subject Sample	# 2	0.142	02:01	
Breath Volume		1.890	02:01	
Air Blank		0.000	02:01	
Air Blank		0.000	02:03	
Subject Sample	# 3	0.147	02:05	
Breath Volume		1.468	02:05	
Air Blank		0.000	02:06	
Control Test		0.079	02:06	
Air Blank		0.000	02:07	
Diagnostics Check		OK	02:07	

Data Download Date/Time: 04/07/2022 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SMITH, JAMES
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: MORAN VILLAFUERTE

DATE/TIME OF TEST: 03/18/2022 04:15:46

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.149	04:11
Breath Volume	3.046	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:13
Subject Sample # 2	0.145	04:14
Breath Volume	3.652	04:14
Air Blank	0.000	04:14
Control Test	0.079	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 04/07/2022 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LUCIANO MONTES, NICOLAS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 03/18/2022 04:44:59

Diagnostics Check		OK	04:39
Air Blank		0.000	04:39
Control Test		0.080	04:39
Air Blank		0.000	04:40
Subject Sample	# 1	0.158	04:40
Breath Volume		2.667	04:40
Air Blank		0.000	04:41
Air Blank		0.000	04:43
Subject Sample	# 2	0.154	04:43
Breath Volume		2.593	04:43
Air Blank		0.000	04:44
Control Test		0.079	04:44
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 04/07/2022 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: JEFFRIES

DATE/TIME OF TEST: 03/18/2022 05:17:37

Diagnostics Check		OK	05:11
Air Blank		0.000	05:11
Control Test		0.080	05:12
Air Blank		0.000	05:12
Subject Sample	# 1	0.226	05:13
Breath Volume		4.335	05:13
Air Blank		0.000	05:13
Air Blank		0.000	05:15
Subject Sample	# 2	0.230	05:16
Breath Volume		3.199	05:16
Air Blank		0.000	05:16
Control Test		0.079	05:17
Air Blank		0.000	05:17
Diagnostics Check		OK	05:17

Data Download Date/Time: 04/07/2022 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: CHANEY

DATE/TIME OF TEST: 03/18/2022 18:57:48

Diagnostics Check	OK	18:52
Air Blank	0.000	18:52
Control Test	0.079	18:52
Air Blank	0.000	18:53
Subject Sample # 1	0.230	18:53
Breath Volume	2.421	18:53
Air Blank	0.000	18:54
Air Blank	0.000	18:56
Subject Sample # 2	0.230	18:56
Breath Volume	2.046	18:56
Air Blank	0.000	18:56
Control Test	0.079	18:57
Air Blank	0.000	18:57
Diagnostics Check	OK	18:57

Data Download Date/Time: 04/07/2022 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATIAS,ROLANDO
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: CRUZ ROSALINO

DATE/TIME OF TEST: 03/19/2022 03:37:00

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.160	03:32
Breath Volume	1.925	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.160	03:35
Breath Volume	1.785	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:37

Data Download Date/Time: 04/07/2022 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LITTLE,ELAINA
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: JUVIER

DATE/TIME OF TEST: 03/19/2022 16:22:01

Diagnostics Check	OK	16:16
Air Blank	0.000	16:16
Control Test	0.080	16:16
Air Blank	0.000	16:17
Subject Sample # 1	0.118	16:17
Breath Volume	2.703	16:17
Air Blank	0.000	16:18
Air Blank	0.000	16:20
Subject Sample # 2	0.117	16:20
Breath Volume	3.066	16:20
Air Blank	0.000	16:21
Control Test	0.080	16:21
Air Blank	0.000	16:21
Diagnostics Check	OK	16:22

Data Download Date/Time: 04/07/2022 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARWOOD,ARMANDO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: REXHAJ

DATE/TIME OF TEST: 03/19/2022 20:29:34

Diagnostics Check	OK	20:23
Air Blank	0.000	20:24
Control Test	0.080	20:24
Air Blank	0.000	20:24
Subject Sample # 1	0.000	20:25
Breath Volume	1.535	20:25
Air Blank	0.000	20:25
Air Blank	0.000	20:27
Subject Sample # 2	0.000	20:28
Breath Volume	1.746	20:28
Air Blank	0.000	20:28
Control Test	0.080	20:28
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 04/07/2022 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MIRANDA LOPEZ

DATE/TIME OF TEST: 03/20/2022 02:22:36

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.080	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.200	02:18
Breath Volume		1.523	02:18
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.202	02:21
Breath Volume		1.191	02:21
Air Blank		0.000	02:21
Control Test		0.078	02:22
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 04/07/2022 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HEBEL,GRIFFIN
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: ALLIANO

DATE/TIME OF TEST: 03/20/2022 05:07:16

Diagnostics Check		OK	04:51	
Air Blank		0.000	04:51	
Control Test		0.080	04:51	
Air Blank		0.000	04:52	
Subject Sample	# 1	0.186	04:54	
Breath Volume		1.226	04:54	
Air Blank		0.000	04:55	
Air Blank		0.000	04:56	
Subject Sample	# 2	0.148	05:00	Volume Not Met
Breath Volume		0.316	05:00	
Air Blank		0.000	05:00	
Air Blank		0.000	05:02	
Subject Sample	# 3	0.184	05:05	Volume Not Met
Breath Volume		0.730	05:05	
Air Blank		0.000	05:06	
Control Test		0.079	05:06	
Air Blank		0.000	05:07	
Diagnostics Check		OK	05:07	

Data Download Date/Time: 04/07/2022 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TATTOLI,RYAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 03/23/2022 00:14:13

Diagnostics Check		OK	00:08
Air Blank		0.000	00:08
Control Test		0.079	00:09
Air Blank		0.000	00:09
Subject Sample	# 1	0.240	00:09
Breath Volume		1.550	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:12
Subject Sample	# 2	0.239	00:12
Breath Volume		1.675	00:12
Air Blank		0.000	00:13
Control Test		0.080	00:13
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 04/07/2022 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WILLS,DARREN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: NASRALLAH

DATE/TIME OF TEST: 03/23/2022 02:45:23

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.080	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.165	02:41
Breath Volume	1.605	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.166	02:43
Breath Volume	1.152	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 04/07/2022 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LIEBKNECHT,KEVIN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: FIGH

DATE/TIME OF TEST: 03/23/2022 03:42:54

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.158	03:38
Breath Volume	1.257	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.171	03:41
Breath Volume	1.566	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 04/07/2022 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KING,ARTHUR
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: GUTTENTAG

DATE/TIME OF TEST: 03/24/2022 03:42:56

Diagnostics Check		OK	03:36
Air Blank		0.000	03:37
Control Test		0.080	03:37
Air Blank		0.000	03:38
Subject Sample	# 1	0.000	03:38
Breath Volume		1.207	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.000	03:41
Breath Volume		1.441	03:41
Air Blank		0.000	03:42
Control Test		0.079	03:42
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 04/07/2022 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCHOEN,ELLIS
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 03/25/2022 02:47:00

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.080	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.106	02:42
Breath Volume		2.695	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.108	02:45
Breath Volume		2.234	02:45
Air Blank		0.000	02:46
Control Test		0.079	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:47

Data Download Date/Time: 04/07/2022 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/25/2022 04:05:37

Diagnostics Check		OK	03:59
Air Blank		0.000	03:59
Control Test		0.079	03:59
Air Blank		0.000	04:00
Subject Sample	# 1	0.000	04:01
Breath Volume		1.718	04:01
Air Blank		0.000	04:01
Air Blank		0.000	04:03
Subject Sample	# 2	0.000	04:04
Breath Volume		1.644	04:04
Air Blank		0.000	04:04
Control Test		0.080	04:05
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 04/07/2022 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DEBOTTIS,MARCUS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: RODRIGUEZ SUAREZ

DATE/TIME OF TEST: 03/25/2022 05:58:29

Diagnostics Check		OK	05:44	
Air Blank		0.000	05:44	
Control Test		0.080	05:45	
Air Blank		0.000	05:45	
Subject Sample	# 1	0.212	05:48	No .020 agreement
Breath Volume		1.265	05:48	
Air Blank		0.000	05:49	
Air Blank		0.000	05:50	
Subject Sample	# 2	0.217	05:54	Volume Not Met
Breath Volume		0.976	05:54	
Air Blank		0.000	05:54	
Air Blank		0.000	05:56	
Subject Sample	# 3	0.247	05:57	No .020 agreement
Breath Volume		1.363	05:57	
Air Blank		0.000	05:57	
Control Test		0.079	05:57	
Air Blank		0.000	05:58	
Diagnostics Check		OK	05:58	

Data Download Date/Time: 04/07/2022 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DEBOTTIS,MARCUS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: RODRIGUEZ SUAREZ

DATE/TIME OF TEST: 03/25/2022 06:07:58

Diagnostics Check		OK	06:00
Air Blank		0.000	06:00
Control Test		0.079	06:01
Air Blank		0.000	06:01
Subject Sample	# 1	0.205	06:03
Breath Volume		1.132	06:03
Air Blank		0.000	06:03
Air Blank		0.000	06:05
Subject Sample	# 2	0.208	06:06
Breath Volume		1.101	06:06
Air Blank		0.000	06:07
Control Test		0.079	06:07
Air Blank		0.000	06:07
Diagnostics Check		OK	06:07

Data Download Date/Time: 04/07/2022 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JONES JR.,EUGENE
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: HARLAN

DATE/TIME OF TEST: 03/26/2022 00:41:46

Diagnostics Check		OK	00:34
Air Blank		0.000	00:34
Control Test		0.080	00:34
Air Blank		0.000	00:35
Subject Sample	# 1	0.208	00:37
Breath Volume		1.949	00:37
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.213	00:40
Breath Volume		1.179	00:40
Air Blank		0.000	00:40
Control Test		0.080	00:41
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 04/07/2022 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DANIEL,CONNELL
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: ZACARIAS LOPEZ

DATE/TIME OF TEST: 03/26/2022 03:10:05

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.080	03:04
Air Blank		0.000	03:05
Subject Sample	# 1	0.236	03:05
Breath Volume		2.980	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.237	03:08
Breath Volume		2.992	03:08
Air Blank		0.000	03:09
Control Test		0.080	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 04/07/2022 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BLACKWELL, ROBERT
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: SOSA DELACRUZ

DATE/TIME OF TEST: 03/26/2022 04:15:40

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.147	04:11
Breath Volume	2.570	04:11
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.143	04:14
Breath Volume	1.855	04:14
Air Blank	0.000	04:14
Control Test	0.080	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 04/07/2022 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: MCCLAIN

DATE/TIME OF TEST: 03/26/2022 04:47:06

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.081	04:42
Air Blank	0.000	04:42
Subject Sample # 1	0.053	04:42
Breath Volume	1.324	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.051	04:45
Breath Volume	1.101	04:45
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 04/07/2022 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 03/26/2022 09:35:29

Diagnostics Check	OK	09:29
Air Blank	0.000	09:29
Control Test	0.080	09:30
Air Blank	0.000	09:30
Subject Sample # 1	0.141	09:31
Breath Volume	1.574	09:31
Air Blank	0.000	09:31
Air Blank	0.000	09:33
Subject Sample # 2	0.147	09:34
Breath Volume	1.269	09:34
Air Blank	0.000	09:34
Control Test	0.080	09:34
Air Blank	0.000	09:35
Diagnostics Check	OK	09:35

Data Download Date/Time: 04/07/2022 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BANOS,FRANK
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: GUERRA

DATE/TIME OF TEST: 03/27/2022 00:41:22

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.080	00:36
Air Blank	0.000	00:36
Subject Sample # 1	0.168	00:37
Breath Volume	2.730	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.154	00:39
Breath Volume	1.285	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 04/07/2022 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE,ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: AGAMASU

DATE/TIME OF TEST: 03/27/2022 03:38:31

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:36	
Control Test	0.080	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 1	03:37	Subject Test Refused
Breath Volume	0.000	03:37	
Air Blank	0.000	03:37	
Control Test	0.079	03:37	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 04/07/2022 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: OGLETREE,JERRELL
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: COLON ARROYO

DATE/TIME OF TEST: 03/27/2022 04:46:20

Diagnostics Check		OK	04:39
Air Blank		0.000	04:40
Control Test		0.081	04:40
Air Blank		0.000	04:41
Subject Sample	# 1	0.112	04:41
Breath Volume		1.980	04:41
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.116	04:44
Breath Volume		1.976	04:44
Air Blank		0.000	04:45
Control Test		0.081	04:45
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

Data Download Date/Time: 04/07/2022 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FELLOWS,BRENT
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 03/27/2022 07:59:23

Diagnostics Check		OK	07:52
Air Blank		0.000	07:52
Control Test		0.080	07:52
Air Blank		0.000	07:53
Subject Sample	# 1	0.156	07:53
Breath Volume		1.171	07:53
Air Blank		0.000	07:54
Air Blank		0.000	07:56
Subject Sample	# 2	0.165	07:57
Breath Volume		3.121	07:57
Air Blank		0.000	07:58
Control Test		0.081	07:58
Air Blank		0.000	07:59
Diagnostics Check		OK	07:59

Data Download Date/Time: 04/07/2022 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ORTIZ, PAUL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ROMAN JR

DATE/TIME OF TEST: 03/28/2022 01:31:12

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.081	01:25
Air Blank		0.000	01:26
Subject Sample	# 1	0.082	01:26
Breath Volume		2.390	01:26
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.080	01:29
Breath Volume		2.714	01:29
Air Blank		0.000	01:30
Control Test		0.081	01:30
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 04/07/2022 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HENRIQUEZ,ENRIQUE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: SANTIAGO SANTOS

DATE/TIME OF TEST: 03/28/2022 05:05:06

Diagnostics Check	OK	04:58
Air Blank	0.000	04:59
Control Test	0.081	04:59
Air Blank	0.000	05:00
Subject Sample # 1	0.164	05:00
Breath Volume	1.843	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.162	05:03
Breath Volume	1.328	05:03
Air Blank	0.000	05:04
Control Test	0.080	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 04/07/2022 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALL, ANDREW
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: SCHUESSLER

DATE/TIME OF TEST: 03/28/2022 05:17:30

Diagnostics Check		OK	05:11
Air Blank		0.000	05:11
Control Test		0.081	05:12
Air Blank		0.000	05:12
Subject Sample	# 1	0.134	05:12
Breath Volume		1.613	05:12
Air Blank		0.000	05:13
Air Blank		0.000	05:15
Subject Sample	# 2	0.132	05:16
Breath Volume		2.695	05:16
Air Blank		0.000	05:16
Control Test		0.081	05:16
Air Blank		0.000	05:17
Diagnostics Check		OK	05:17

Data Download Date/Time: 04/07/2022 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER, OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BARRGAN-GARCIA, ANTONIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 11:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: BARIN

DATE/TIME OF TEST: 03/29/2022 12:19:42

Diagnostics Check	OK	12:13
Air Blank	0.000	12:13
Control Test	0.079	12:14
Air Blank	0.000	12:14
Subject Sample # 1	0.197	12:15
Breath Volume	3.882	12:15
Air Blank	0.000	12:15
Air Blank	0.000	12:17
Subject Sample # 2	0.196	12:18
Breath Volume	4.046	12:18
Air Blank	0.000	12:18
Control Test	0.079	12:19
Air Blank	0.000	12:19
Diagnostics Check	OK	12:19

Data Download Date/Time: 04/07/2022 12:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SOSA,EMMANUEL
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PINILLOS MORA

DATE/TIME OF TEST: 03/30/2022 04:22:29

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.080	04:16
Air Blank	0.000	04:17
Subject Sample # 1	0.148	04:18
Breath Volume	1.316	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.145	04:21
Breath Volume	1.421	04:21
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 04/07/2022 12:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARC,VUKANSON
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: ONEIL

DATE/TIME OF TEST: 03/30/2022 22:27:15

Diagnostics Check	OK	22:20
Air Blank	0.000	22:21
Control Test	0.080	22:21
Air Blank	0.000	22:22
Subject Sample # 1	0.156	22:22
Breath Volume	2.488	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.146	22:25
Breath Volume	1.621	22:25
Air Blank	0.000	22:26
Control Test	0.080	22:26
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 04/07/2022 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: NORTON

DATE/TIME OF TEST: 03/31/2022 04:35:11

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.080	04:30
Air Blank	0.000	04:30
Subject Sample # 1	0.155	04:30
Breath Volume	1.285	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.159	04:33
Breath Volume	1.460	04:33
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 04/07/2022 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRISWOLD,DAVID,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: QUILES-DAVILA,DARIO
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 02/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: RODRIGUEZ RIVERA

DATE/TIME OF TEST: 03/02/2022 22:55:53

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.080	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.216	22:51
Breath Volume	2.324	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:53
Subject Sample # 2	0.208	22:54
Breath Volume	1.757	22:54
Air Blank	0.000	22:54
Control Test	0.080	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 03/11/2022 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES, RONNIE,
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: NUSCIS, NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: MINOR

DATE/TIME OF TEST: 03/05/2022 05:20:37

Diagnostics Check		OK	05:04	
Air Blank		0.000	05:04	
Control Test		0.081	05:04	
Air Blank		0.000	05:05	
Subject Sample	# 1	0.163	05:07	
Breath Volume		1.304	05:07	
Air Blank		0.000	05:08	
Air Blank		0.000	05:10	
Subject Sample	# 2	0.142	05:13	Volume Not Met
Breath Volume		0.015	05:13	
Air Blank		0.000	05:14	
Air Blank		0.000	05:15	
Subject Sample	# 3	0.140	05:18	Volume Not Met
Breath Volume		0.050	05:18	
Air Blank		0.000	05:19	
Control Test		0.078	05:20	
Air Blank		0.000	05:20	
Diagnostics Check		OK	05:20	

Data Download Date/Time: 03/11/2022 10:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES, RONNIE,
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: NUSCIS, NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: MINOR

DATE/TIME OF TEST: 03/05/2022 05:36:25

Diagnostics Check		OK	05:24	
Air Blank		0.000	05:24	
Control Test		0.079	05:25	
Air Blank		0.000	05:25	
Subject Sample	# 1	0.151	05:28	Volume Not Met
Breath Volume		0.636	05:28	
Air Blank		0.000	05:29	
Air Blank		0.000	05:31	
Subject Sample	# 2	0.157	05:31	
Breath Volume		1.562	05:31	
Air Blank		0.000	05:32	
Air Blank		0.000	05:34	
Subject Sample	# 3	0.151	05:34	
Breath Volume		1.296	05:34	
Air Blank		0.000	05:35	
Control Test		0.079	05:35	
Air Blank		0.000	05:36	
Diagnostics Check		OK	05:36	

Data Download Date/Time: 03/11/2022 10:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHOBREY,SPENCER,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: GONZALEZ-MALDONADO

DATE/TIME OF TEST: 03/06/2022 02:16:16

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.081	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.117	02:11
Breath Volume	1.414	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:13
Subject Sample # 2	0.119	02:14
Breath Volume	1.535	02:14
Air Blank	0.000	02:15
Control Test	0.081	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 03/11/2022 10:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,SHAWN,N
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: QUILES,DARIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 03/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:08

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 03/18/2022 06:41:08

Diagnostics Check	OK	06:34
Air Blank	0.000	06:35
Control Test	0.081	06:35
Air Blank	0.000	06:36
Subject Sample # 1	0.132	06:36
Breath Volume	1.976	06:36
Air Blank	0.000	06:37
Air Blank	0.000	06:39
Subject Sample # 2	0.127	06:39
Breath Volume	1.480	06:39
Air Blank	0.000	06:40
Control Test	0.081	06:40
Air Blank	0.000	06:41
Diagnostics Check	OK	06:41

Data Download Date/Time: 04/14/2022 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,BRIAN,G
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: WOOLLEY

DATE/TIME OF TEST: 03/19/2022 05:18:32

Diagnostics Check		OK	05:07	
Air Blank		0.000	05:07	
Control Test		0.081	05:07	
Air Blank		0.000	05:08	
Subject Sample	# 1	0.000	05:11	Volume Not Met
Breath Volume		0.210	05:11	
Air Blank		0.000	05:12	
Air Blank		0.000	05:14	
Subject Sample	# 2	0.000	05:17	Volume Not Met
Breath Volume		0.089	05:17	
Air Blank		0.000	05:17	
Control Test		0.081	05:17	
Air Blank		0.000	05:18	
Diagnostics Check		OK	05:18	

Data Download Date/Time: 04/14/2022 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: REYES,ISAAC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: AGUAYO COLON

DATE/TIME OF TEST: 03/25/2022 06:06:54

Diagnostics Check	OK	06:00
Air Blank	0.000	06:01
Control Test	0.080	06:01
Air Blank	0.000	06:01
Subject Sample # 1	0.177	06:02
Breath Volume	1.527	06:02
Air Blank	0.000	06:03
Air Blank	0.000	06:04
Subject Sample # 2	0.169	06:05
Breath Volume	1.796	06:05
Air Blank	0.000	06:06
Control Test	0.079	06:06
Air Blank	0.000	06:06
Diagnostics Check	OK	06:06

Data Download Date/Time: 04/14/2022 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHOBREY,SPENCER,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WHOBREY,SPENCER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: HERNANNDEZ

DATE/TIME OF TEST: 03/29/2022 23:35:57

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.081	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.249	23:31
Breath Volume	2.894	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:33
Subject Sample # 2	0.249	23:34
Breath Volume	2.980	23:34
Air Blank	0.000	23:35
Control Test	0.082	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 04/14/2022 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INGERSOLL,NICHOLAS,F
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: BYNUM,CHRISTIE
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 03/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: BUSTOS

DATE/TIME OF TEST: 03/04/2022 02:36:52

Diagnostics Check		OK	02:24	
Air Blank		0.000	02:25	
Control Test		0.081	02:25	
Air Blank		0.000	02:25	
Subject Sample	# 1	0.182	02:29	Volume Not Met
Breath Volume		0.210	02:29	
Air Blank		0.000	02:29	
Air Blank		0.000	02:31	
Subject Sample	# 2	0.200	02:32	
Breath Volume		2.523	02:32	
Air Blank		0.000	02:32	
Air Blank		0.000	02:34	
Subject Sample	# 3	0.185	02:35	
Breath Volume		1.457	02:35	
Air Blank		0.000	02:35	
Control Test		0.079	02:36	
Air Blank		0.000	02:36	
Diagnostics Check		OK	02:36	

Data Download Date/Time: 04/26/2022 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTMAN, MATTHEW, R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: BYNUM, C
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 03/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: HEWKIN

DATE/TIME OF TEST: 03/08/2022 04:34:21

Diagnostics Check		OK	04:27
Air Blank		0.000	04:28
Control Test		0.079	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.228	04:29
Breath Volume		1.660	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.245	04:32
Breath Volume		2.035	04:32
Air Blank		0.000	04:33
Control Test		0.080	04:33
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 04/26/2022 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSEPH,A
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: LAFLEUR,MATTHEW
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 03/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: ROUTTENBERG

DATE/TIME OF TEST: 03/15/2022 01:33:19

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.081	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.091	01:29
Breath Volume	2.718	01:29
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.091	01:31
Breath Volume	2.273	01:31
Air Blank	0.000	01:32
Control Test	0.080	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 04/26/2022 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTMAN, MATTHEW, R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: BYNUM, C
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 03/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: NELIS

DATE/TIME OF TEST: 03/17/2022 19:35:41

Diagnostics Check	OK	19:29
Air Blank	0.000	19:29
Control Test	0.081	19:29
Air Blank	0.000	19:30
Subject Sample # 1	0.113	19:31
Breath Volume	2.527	19:31
Air Blank	0.000	19:31
Air Blank	0.000	19:33
Subject Sample # 2	0.112	19:34
Breath Volume	1.601	19:34
Air Blank	0.000	19:34
Control Test	0.081	19:35
Air Blank	0.000	19:35
Diagnostics Check	OK	19:35

Data Download Date/Time: 04/26/2022 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: STRANGE,BOBBY
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 03/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 03/19/2022 21:48:45

Diagnostics Check		OK	21:41
Air Blank		0.000	21:41
Control Test		0.079	21:41
Air Blank		0.000	21:42
Subject Sample	# 1	0.206	21:43
Breath Volume		1.777	21:43
Air Blank		0.000	21:44
Air Blank		0.000	21:46
Subject Sample	# 2	0.216	21:47
Breath Volume		1.316	21:47
Air Blank		0.000	21:47
Control Test		0.079	21:48
Air Blank		0.000	21:48
Diagnostics Check		OK	21:48

Data Download Date/Time: 04/26/2022 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: BYNUM,CHRISTIE
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 03/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: WAYMIRE

DATE/TIME OF TEST: 03/27/2022 03:25:20

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.130	03:20
Breath Volume	2.980	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.117	03:23
Breath Volume	1.687	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 04/26/2022 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INGERSOLL,NICHOLAS,F
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: INGERSOLL,NICHOLAS
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 03/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: ABRAO

DATE/TIME OF TEST: 03/27/2022 21:38:32

Diagnostics Check	OK	21:32
Air Blank	0.000	21:32
Control Test	0.081	21:32
Air Blank	0.000	21:33
Subject Sample # 1	0.139	21:34
Breath Volume	3.531	21:34
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.136	21:37
Breath Volume	3.492	21:37
Air Blank	0.000	21:37
Control Test	0.079	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 04/26/2022 11:16