

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: GUSTAFSON, ZACHARY  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 03/08/2022 23:31:46

---

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.079	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.090	23:27
Breath Volume	1.382	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:29
Subject Sample # 2	0.092	23:30
Breath Volume	1.398	23:30
Air Blank	0.000	23:30
Control Test	0.079	23:31
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

---

Data Download Date/Time: 03/28/2022 22:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: ANDREWS, GARY  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 03/10/2022 04:20:05

---

Diagnostics Check		OK	04:14
Air Blank		0.000	04:14
Control Test		0.079	04:14
Air Blank		0.000	04:15
Subject Sample	# 1	0.183	04:15
Breath Volume		2.179	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.178	04:18
Breath Volume		2.195	04:18
Air Blank		0.000	04:19
Control Test		0.079	04:19
Air Blank		0.000	04:19
Diagnostics Check		OK	04:20

---

Data Download Date/Time: 03/28/2022 22:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: VANROMPA,VITO  
ARRESTING OFFICER AGENCY: Lake Mary PD  
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: ORTEGA MENENDEZ

DATE/TIME OF TEST: 03/11/2022 03:55:21

---

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.079	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.117	03:50
Breath Volume	1.546	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.123	03:53
Breath Volume	1.511	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

---

Data Download Date/Time: 03/28/2022 22:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: PFANNERSTILL,JACOB  
ARRESTING OFFICER AGENCY: Altamonte Springs PD  
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: MCQUAY

DATE/TIME OF TEST: 03/12/2022 03:49:57

---

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.080	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.145	03:45
Breath Volume	1.785	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.143	03:48
Breath Volume	1.476	03:48
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

---

Data Download Date/Time: 03/28/2022 22:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: UCELO,BRYAN  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: CAALPAAU

DATE/TIME OF TEST: 03/12/2022 04:28:25

---

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.080	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.159	04:23
Breath Volume	1.191	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.150	04:26
Breath Volume	2.136	04:26
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

---

Data Download Date/Time: 03/28/2022 22:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: SMITH,MEGAN  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: ROTAB

DATE/TIME OF TEST: 03/13/2022 05:02:42

---

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:51	
Control Test	0.081	04:51	
Air Blank	0.000	04:52	
Subject Sample	# 1 0.000	04:52	
Breath Volume	4.105	04:52	
Air Blank	0.000	04:53	
Air Blank	0.000	04:55	
Subject Sample	# 2	04:58	No Sample Provided
Breath Volume	0.000	04:58	
Air Blank	0.000	04:58	
Air Blank	0.000	05:00	
Subject Sample	# 3 0.000	05:01	
Breath Volume	3.570	05:01	
Air Blank	0.000	05:01	
Control Test	0.079	05:02	
Air Blank	0.000	05:02	
Diagnostics Check	OK	05:02	

---

Data Download Date/Time: 03/28/2022 22:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: BARNETT, AUSTIN  
ARRESTING OFFICER AGENCY: Altamonte Springs PD  
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: ARAGON TORRES

DATE/TIME OF TEST: 03/13/2022 23:06:11

---

Diagnostics Check	OK	23:00
Air Blank	0.000	23:00
Control Test	0.080	23:00
Air Blank	0.000	23:01
Subject Sample # 1	0.230	23:01
Breath Volume	3.253	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.227	23:04
Breath Volume	3.125	23:04
Air Blank	0.000	23:05
Control Test	0.080	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

---

Data Download Date/Time: 03/28/2022 22:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: UCELO, BRYAN  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 03/14/2022 01:51:33

---

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.080	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.169	01:47
Breath Volume	2.070	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.168	01:50
Breath Volume	2.332	01:50
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

---

Data Download Date/Time: 03/28/2022 22:26



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: RODRIGUEZ, EMMANUEL  
ARRESTING OFFICER AGENCY: Altamonte Springs PD  
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: ZAMMIT

DATE/TIME OF TEST: 03/14/2022 22:01:56

---

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.144	21:57
Breath Volume	3.390	21:57
Air Blank	0.000	21:58
Air Blank	0.000	21:59
Subject Sample # 2	0.138	22:00
Breath Volume	2.894	22:00
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

---

Data Download Date/Time: 03/28/2022 22:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: GRUNAT, JERRY  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: MACHUCA

DATE/TIME OF TEST: 03/15/2022 04:59:26

---

Diagnostics Check		OK	04:53
Air Blank		0.000	04:54
Control Test		0.080	04:54
Air Blank		0.000	04:54
Subject Sample	# 1	0.000	04:55
Breath Volume		2.695	04:55
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.000	04:58
Breath Volume		3.121	04:58
Air Blank		0.000	04:58
Control Test		0.080	04:58
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

---

Data Download Date/Time: 03/28/2022 22:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: MATTINGLY, BRIANNA  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: PALOMINO BENAVIDES

DATE/TIME OF TEST: 03/16/2022 23:22:30

---

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.080	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.000	23:18
Breath Volume	3.011	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.000	23:21
Breath Volume	3.812	23:21
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

---

Data Download Date/Time: 03/28/2022 22:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMENS,ADAM,J  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: SCHWARTZ,E  
ARRESTING OFFICER AGENCY: Winter Springs PD  
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: BARRIS

DATE/TIME OF TEST: 03/18/2022 21:44:14

---

Diagnostics Check	OK	21:37
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.126	21:38
Breath Volume	2.945	21:38
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.127	21:42
Breath Volume	1.238	21:42
Air Blank	0.000	21:43
Control Test	0.080	21:43
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

---

Data Download Date/Time: 03/28/2022 22:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: TEEHAN,MATTHEW  
ARRESTING OFFICER AGENCY: Altamonte Springs PD  
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 03/19/2022 02:04:32

---

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.115	01:59
Breath Volume	2.140	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.108	02:03
Breath Volume	1.691	02:03
Air Blank	0.000	02:03
Control Test	0.081	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

---

Data Download Date/Time: 03/28/2022 22:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: MUNN,SCOTT  
ARRESTING OFFICER AGENCY: Casselberry PD  
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 03/19/2022 02:32:39

---

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.213	02:27
Breath Volume	2.195	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.208	02:31
Breath Volume	2.699	02:31
Air Blank	0.000	02:31
Control Test	0.078	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

---

Data Download Date/Time: 03/28/2022 22:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: SCHWARTZ,ERIK  
ARRESTING OFFICER AGENCY: Winter Springs PD  
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: VALENZUELA

DATE/TIME OF TEST: 03/20/2022 01:33:37

---

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.081	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.130	01:29
Breath Volume	1.769	01:29
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.133	01:32
Breath Volume	2.308	01:32
Air Blank	0.000	01:32
Control Test	0.080	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

---

Data Download Date/Time: 03/28/2022 22:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: GUSTAFSON,ZACHARY  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: MOREHOUSE

DATE/TIME OF TEST: 03/20/2022 02:40:28

---

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.172	02:35
Breath Volume	1.378	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:37
Subject Sample # 2	0.172	02:38
Breath Volume	1.242	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

---

Data Download Date/Time: 03/28/2022 22:26



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: DIAL, VICTORIA  
ARRESTING OFFICER AGENCY: Altamonte Springs PD  
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: HOPF

DATE/TIME OF TEST: 03/21/2022 23:51:50

---

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:47
Subject Sample # 1	0.126	23:47
Breath Volume	1.781	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:49
Subject Sample # 2	0.125	23:50
Breath Volume	1.425	23:50
Air Blank	0.000	23:50
Control Test	0.080	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

---

Data Download Date/Time: 03/28/2022 22:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: SANCHEZ, JUBILEE  
ARRESTING OFFICER AGENCY: Casselberry PD  
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: SOTO CONTRERAS

DATE/TIME OF TEST: 03/23/2022 02:17:10

---

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.079	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.147	02:12
Breath Volume	1.617	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.147	02:15
Breath Volume	3.324	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

---

Data Download Date/Time: 03/28/2022 22:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: RIDDLE,JAMES  
ARRESTING OFFICER AGENCY: Lake Mary PD  
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 17:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: DENMAN

DATE/TIME OF TEST: 03/23/2022 21:59:13

---

Diagnostics Check	OK	21:52
Air Blank	0.000	21:52
Control Test	0.079	21:53
Air Blank	0.000	21:53
Subject Sample # 1	0.151	21:54
Breath Volume	5.308	21:54
Air Blank	0.000	21:55
Air Blank	0.000	21:56
Subject Sample # 2	0.132	21:57
Breath Volume	1.515	21:57
Air Blank	0.000	21:58
Control Test	0.079	21:58
Air Blank	0.000	21:59
Diagnostics Check	OK	21:59

---

Data Download Date/Time: 03/28/2022 22:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: SIMPSON,CARL  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: RICHARDVILLE

DATE/TIME OF TEST: 03/24/2022 01:16:17

---

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.079	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.167	01:11
Breath Volume	2.746	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.163	01:14
Breath Volume	1.953	01:14
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

---

Data Download Date/Time: 03/28/2022 22:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GERVAIS,ROSEMARY,C  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: LIEBKNECHT,K  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 03/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/31/2022 03:26:48

---

Diagnostics Check		OK	03:19
Air Blank		0.000	03:20
Control Test		0.080	03:20
Air Blank		0.000	03:21
Subject Sample	# 1	0.137	03:22
Breath Volume		2.996	03:22
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.130	03:25
Breath Volume		1.726	03:25
Air Blank		0.000	03:25
Control Test		0.080	03:26
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

---

Data Download Date/Time: 05/01/2022 22:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,MEGAN,N  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: SMITH,MEGAN  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 03/24/2022

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: KHORRAMIAN

DATE/TIME OF TEST: 03/31/2022 04:00:14

---

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.081	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.000	03:53
Breath Volume		1.304	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.000	03:58
Breath Volume		1.531	03:58
Air Blank		0.000	03:59
Control Test		0.080	03:59
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

---

Data Download Date/Time: 05/01/2022 22:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEILL,MICHAEL,A  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: WEBER,DYLAN  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: TAYLOE

DATE/TIME OF TEST: 03/02/2022 00:35:54

---

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.080	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.167	00:31
Breath Volume	1.421	00:31
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.184	00:34
Breath Volume	2.429	00:34
Air Blank	0.000	00:34
Control Test	0.080	00:35
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

---

Data Download Date/Time: 03/12/2022 16:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANDERSON,ERIC,G  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: ANDERSON,ERIC  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: PRITCHARD

DATE/TIME OF TEST: 03/04/2022 22:26:25

---

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.079	22:21
Air Blank	0.000	22:21
Subject Sample # 1	0.247	22:21
Breath Volume	3.242	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.254	22:24
Breath Volume	3.453	22:24
Air Blank	0.000	22:25
Control Test	0.079	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

---

Data Download Date/Time: 03/12/2022 16:37



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANDERSON,ERIC,G  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: MARTINEZ,PHILIP  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: AUCOIN

DATE/TIME OF TEST: 03/05/2022 23:18:01

---

Diagnostics Check		OK	23:11
Air Blank		0.000	23:11
Control Test		0.079	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.179	23:13
Breath Volume		2.265	23:13
Air Blank		0.000	23:14
Air Blank		0.000	23:15
Subject Sample	# 2	0.177	23:16
Breath Volume		2.093	23:16
Air Blank		0.000	23:17
Control Test		0.079	23:17
Air Blank		0.000	23:17
Diagnostics Check		OK	23:18

---

Data Download Date/Time: 03/12/2022 16:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANDERSON,ERIC,G  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: PARKER,MATTHEW  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: GREGG

DATE/TIME OF TEST: 03/06/2022 01:55:26

---

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.080	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.150	01:46
Breath Volume	2.046	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.174	01:50
Breath Volume	2.707	01:50
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 3	0.182	01:53
Breath Volume	4.191	01:53
Air Blank	0.000	01:54
Control Test	0.079	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

---

Data Download Date/Time: 03/12/2022 16:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEILL,MICHAEL,A  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: NEILL,MICHAEL  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: SNOWDEN

DATE/TIME OF TEST: 03/10/2022 23:07:51

---

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.080	23:02
Air Blank	0.000	23:02
Subject Sample # 1	0.292	23:03
Breath Volume	2.507	23:03
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.284	23:06
Breath Volume	2.417	23:06
Air Blank	0.000	23:06
Control Test	0.079	23:07
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

---

Data Download Date/Time: 03/12/2022 16:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: WHITE,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 03/12/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 12:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 03/15/2022 01:33:40

---

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.077	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.159	01:28
Breath Volume	1.406	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.168	01:32
Breath Volume	1.265	01:32
Air Blank	0.000	01:32
Control Test	0.078	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

---

Data Download Date/Time: 04/20/2022 04:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOUSMAN,JOHN,D  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: CLERMONT,ENITH  
ARRESTING OFFICER AGENCY: Indian Harbour Beach  
LAST AGENCY INSPECTION: 03/12/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: CORNELIUS

DATE/TIME OF TEST: 03/19/2022 05:42:32

---

Diagnostics Check	OK	05:36
Air Blank	0.000	05:36
Control Test	0.079	05:37
Air Blank	0.000	05:37
Subject Sample # 1	0.217	05:38
Breath Volume	2.117	05:38
Air Blank	0.000	05:38
Air Blank	0.000	05:40
Subject Sample # 2	0.208	05:41
Breath Volume	1.980	05:41
Air Blank	0.000	05:41
Control Test	0.079	05:41
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

---

Data Download Date/Time: 04/20/2022 04:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCRANN,SEAN,M  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: FERREIRA,ALEXANDER  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 03/12/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 03/26/2022 21:47:36

---

Diagnostics Check		OK	21:40
Air Blank		0.000	21:41
Control Test		0.079	21:41
Air Blank		0.000	21:42
Subject Sample	# 1	0.244	21:42
Breath Volume		1.527	21:42
Air Blank		0.000	21:43
Air Blank		0.000	21:45
Subject Sample	# 2	0.254	21:46
Breath Volume		2.261	21:46
Air Blank		0.000	21:46
Control Test		0.079	21:47
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

---

Data Download Date/Time: 04/20/2022 04:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY, RONALD, W  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: KINSEY, RONALD  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 03/12/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: DEITHORN

DATE/TIME OF TEST: 03/28/2022 20:09:54

---

Diagnostics Check	OK	20:02
Air Blank	0.000	20:03
Control Test	0.079	20:03
Air Blank	0.000	20:04
Subject Sample # 1	0.294	20:04
Breath Volume	2.136	20:04
Air Blank	0.000	20:05
Air Blank	0.000	20:07
Subject Sample # 2	0.279	20:08
Breath Volume	2.203	20:08
Air Blank	0.000	20:08
Control Test	0.079	20:09
Air Blank	0.000	20:09
Diagnostics Check	OK	20:09

---

Data Download Date/Time: 04/20/2022 04:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOUSMAN,JOHN,D  
BREATH TEST OPERATOR AGENCY: Satellite Beach PD  
ARRESTING OFFICER: WRIGHT,TIMOTHY  
ARRESTING OFFICER AGENCY: Satellite Beach PD  
LAST AGENCY INSPECTION: 03/12/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1497672  
DRY GAS EXPIRY DATE: 05/20/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: SOWERS

DATE/TIME OF TEST: 03/31/2022 02:54:15

---

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:50	
Control Test	0.079	02:50	
Air Blank	0.000	02:50	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Control Test	0.079	02:53	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

---

Data Download Date/Time: 04/20/2022 04:42



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPAIN,KAREN,H  
BREATH TEST OPERATOR AGENCY: Hardee County SO  
ARRESTING OFFICER: ANDREWS,BRITT  
ARRESTING OFFICER AGENCY: Hardee County SO  
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1076482  
DRY GAS EXPIRY DATE: 01/15/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 03/12/2022 02:14:51

---

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.080	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.223	02:10
Breath Volume	1.937	02:10
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.205	02:13
Breath Volume	1.921	02:13
Air Blank	0.000	02:13
Control Test	0.078	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

---

Data Download Date/Time: 04/08/2022 09:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVI,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAKE,LEVI  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:52

SUBJECT'S LAST NAME: KLEIN

DATE/TIME OF TEST: 03/03/2022 17:18:20

---

Diagnostics Check	OK	17:12
Air Blank	0.000	17:13
Control Test	0.081	17:13
Air Blank	0.000	17:13
Subject Sample # 1	0.000	17:14
Breath Volume	2.121	17:14
Air Blank	0.000	17:14
Air Blank	0.000	17:16
Subject Sample # 2	0.000	17:16
Breath Volume	1.695	17:16
Air Blank	0.000	17:17
Control Test	0.082	17:17
Air Blank	0.000	17:18
Diagnostics Check	OK	17:18

---

Data Download Date/Time: 03/26/2022 09:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOODNIGHT,MICHAEL,E  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GOODNIGHT,MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: HENDERSHOTT

DATE/TIME OF TEST: 03/04/2022 22:57:46

---

Diagnostics Check	OK	22:51
Air Blank	0.000	22:52
Control Test	0.080	22:52
Air Blank	0.000	22:52
Subject Sample # 1	0.225	22:53
Breath Volume	2.093	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:55
Subject Sample # 2	0.229	22:56
Breath Volume	2.023	22:56
Air Blank	0.000	22:56
Control Test	0.080	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

---

Data Download Date/Time: 03/26/2022 09:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOODNIGHT,MICHAEL,E  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GOODNIGHT,MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: SALVADOR

DATE/TIME OF TEST: 03/06/2022 00:43:27

---

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:40	
Control Test	0.081	00:40	
Air Blank	0.000	00:41	
Subject Sample # 1	0.169	00:41	
Breath Volume	1.363	00:41	
Air Blank		00:42	RFI Detect
Air Blank	0.000	00:42	
Control Test	0.081	00:42	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

---

Data Download Date/Time: 03/26/2022 09:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOODNIGHT,MICHAEL,E  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GOODNIGHT,MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: SALVADOR

DATE/TIME OF TEST: 03/06/2022 00:54:02

---

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.081	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.171	00:49
Breath Volume	1.109	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.168	00:52
Breath Volume	1.285	00:52
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:54

---

Data Download Date/Time: 03/26/2022 09:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVI,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAKE,LEVI  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: ROSTRAN SUAREZ

DATE/TIME OF TEST: 03/12/2022 04:49:44

---

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.081	04:44
Air Blank	0.000	04:44
Subject Sample # 1	0.199	04:45
Breath Volume	3.031	04:45
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.201	04:48
Breath Volume	2.472	04:48
Air Blank	0.000	04:48
Control Test	0.081	04:49
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

---

Data Download Date/Time: 03/26/2022 09:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,CRISTIAN,D  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MILLER,CRISTIAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MCGRATH

DATE/TIME OF TEST: 03/14/2022 03:03:35

---

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:01	
Control Test	0.082	03:01	
Air Blank	0.000	03:01	
Subject Sample	# 1	03:02	Subject Test Refused
Breath Volume	0.000	03:02	
Air Blank	0.000	03:02	
Control Test	0.082	03:03	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:03	

---

Data Download Date/Time: 03/26/2022 09:52

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,CRISTIAN,D  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MILLER,CRISTIAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 03/18/2022 03:14:20

---

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:11	
Control Test	0.081	03:12	
Air Blank	0.000	03:12	
Subject Sample	# 1	03:12	Subject Test Refused
Breath Volume	0.000	03:12	
Air Blank	0.000	03:13	
Control Test	0.082	03:13	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

---

Data Download Date/Time: 03/26/2022 09:52



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERGSTROM,STEPHEN,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CHURCHILL,CAMERON  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: GORTON

DATE/TIME OF TEST: 03/28/2022 07:12:12

---

Diagnostics Check		OK	07:00	
Air Blank		0.000	07:01	
Control Test		0.082	07:01	
Air Blank		0.000	07:01	
Subject Sample	# 1	0.242	07:04	Volume Not Met
Breath Volume		0.046	07:04	
Air Blank		0.000	07:05	
Air Blank		0.000	07:07	
Subject Sample	# 2	0.251	07:10	Volume Not Met
Breath Volume		0.835	07:10	
Air Blank		0.000	07:11	
Control Test		0.081	07:11	
Air Blank		0.000	07:12	
Diagnostics Check		OK	07:12	

---

Data Download Date/Time: 04/28/2022 03:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERGSTROM,STEPHEN,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: CHURCHILL,CAMERON  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: GORTON

DATE/TIME OF TEST: 03/28/2022 07:28:55

---

Diagnostics Check		OK	07:14	
Air Blank		0.000	07:14	
Control Test		0.080	07:15	
Air Blank		0.000	07:15	
Subject Sample	# 1	0.272	07:16	
Breath Volume		1.445	07:16	
Air Blank		0.000	07:16	
Air Blank		0.000	07:18	
Subject Sample	# 2	0.251	07:21	Volume Not Met
Breath Volume		0.835	07:21	
Air Blank		0.000	07:22	
Air Blank		0.000	07:24	
Subject Sample	# 3	0.230	07:27	Volume Not Met
Breath Volume		1.066	07:27	
Air Blank		0.000	07:27	
Control Test		0.081	07:28	
Air Blank		0.000	07:28	
Diagnostics Check		OK	07:28	

---

Data Download Date/Time: 04/28/2022 03:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: HIRSHMAN,BRIAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: STARKS

DATE/TIME OF TEST: 03/31/2022 02:10:10

---

Diagnostics Check	OK	02:00
Air Blank	0.000	02:01
Control Test	0.081	02:01
Air Blank	0.000	02:02
Subject Sample # 1	0.138	02:03
Breath Volume	2.347	02:03
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.131	02:08
Breath Volume	1.855	02:08
Air Blank	0.000	02:09
Control Test	0.083	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

---

Data Download Date/Time: 04/28/2022 03:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: COLLMAN,IAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG012501  
DRY GAS EXPIRY DATE: 05/04/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: CRUZ BERNAL

DATE/TIME OF TEST: 03/31/2022 02:42:32

---

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:40	
Control Test	0.082	02:40	
Air Blank	0.000	02:40	
Subject Sample	# 1	02:41	Subject Test Refused
Breath Volume	0.000	02:41	
Air Blank	0.000	02:41	
Control Test	0.081	02:41	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

---

Data Download Date/Time: 04/28/2022 03:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MITCHELL,ALEXANDR,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MITCHELL,ALEXANDRA  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 03/04/2022 23:34:33

---

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.079	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.133	23:30
Breath Volume	1.515	23:30
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.130	23:33
Breath Volume	1.386	23:33
Air Blank	0.000	23:33
Control Test	0.079	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

---

Data Download Date/Time: 03/23/2022 14:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVI,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAKE,LEVI  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/05/2022 02:31:21

---

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:28	
Control Test	0.080	02:29	
Air Blank	0.000	02:29	
Subject Sample	# 1	02:29	Subject Test Refused
Breath Volume	0.000	02:29	
Air Blank	0.000	02:30	
Control Test	0.079	02:30	
Air Blank	0.000	02:31	
Diagnostics Check	OK	02:31	

---

Data Download Date/Time: 03/23/2022 14:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVI,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAKE,LEVI  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 03/12/2022 21:38:32

---

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:36	
Control Test	0.078	21:36	
Air Blank	0.000	21:36	
Subject Sample	# 1	21:37	Subject Test Refused
Breath Volume	0.000	21:37	
Air Blank	0.000	21:37	
Control Test	0.079	21:37	
Air Blank	0.000	21:38	
Diagnostics Check	OK	21:38	

---

Data Download Date/Time: 03/23/2022 14:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVI,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAKE,LEVI  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: DERVISBEGOVIC

DATE/TIME OF TEST: 03/13/2022 00:18:13

---

Diagnostics Check	OK	00:15	
Air Blank	0.000	00:15	
Control Test	0.081	00:16	
Air Blank	0.000	00:16	
Subject Sample	# 1	00:16	Subject Test Refused
Breath Volume	0.000	00:16	
Air Blank	0.000	00:17	
Control Test	0.080	00:17	
Air Blank	0.000	00:18	
Diagnostics Check	OK	00:18	

---

Data Download Date/Time: 03/23/2022 14:13



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GAMBEL,ZOE,I  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: GAMBEL,ZOE  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: V626433810250

DATE/TIME OF TEST: 03/13/2022 00:57:53

---

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.081	00:52
Air Blank	0.000	00:52
Subject Sample # 1	0.027	00:53
Breath Volume	1.257	00:53
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.026	00:56
Breath Volume	1.238	00:56
Air Blank	0.000	00:56
Control Test	0.081	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

---

Data Download Date/Time: 03/23/2022 14:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DETRANO,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: DETRANO,JEFFREY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/13/2022 01:58:56

---

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.081	01:53
Air Blank		0.000	01:53
Subject Sample	# 1	0.188	01:54
Breath Volume		1.199	01:54
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.190	01:57
Breath Volume		1.910	01:57
Air Blank		0.000	01:57
Control Test		0.081	01:58
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

---

Data Download Date/Time: 03/23/2022 14:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVIS  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: TEMPLE,ELIZABETH  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: ZEKIC

DATE/TIME OF TEST: 03/13/2022 03:50:27

---

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.081	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.197	03:45
Breath Volume	1.906	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.197	03:48
Breath Volume	3.695	03:48
Air Blank	0.000	03:49
Control Test	0.081	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

---

Data Download Date/Time: 03/23/2022 14:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVIS  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: DORSEY

DATE/TIME OF TEST: 03/13/2022 04:34:12

---

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:31	
Control Test	0.081	04:32	
Air Blank	0.000	04:32	
Subject Sample	# 1	04:32	Subject Test Refused
Breath Volume	0.000	04:32	
Air Blank	0.000	04:33	
Control Test	0.081	04:33	
Air Blank	0.000	04:34	
Diagnostics Check	OK	04:34	

---

Data Download Date/Time: 03/23/2022 14:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVI,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAKE,LEVI  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 03/16/2022 20:09:26

---

Diagnostics Check	OK	20:03
Air Blank	0.000	20:03
Control Test	0.079	20:04
Air Blank	0.000	20:04
Subject Sample # 1	0.043	20:05
Breath Volume	3.265	20:05
Air Blank	0.000	20:05
Air Blank	0.000	20:07
Subject Sample # 2	0.043	20:07
Breath Volume	3.117	20:07
Air Blank	0.000	20:08
Control Test	0.079	20:08
Air Blank	0.000	20:09
Diagnostics Check	OK	20:09

---

Data Download Date/Time: 03/23/2022 14:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: MITCHELL,ALEXANDRA  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: MANSFIELD

DATE/TIME OF TEST: 03/16/2022 22:36:13

---

Diagnostics Check	OK	22:33	
Air Blank	0.000	22:33	
Control Test	0.079	22:34	
Air Blank	0.000	22:34	
Subject Sample	# 1	22:34	Subject Test Refused
Breath Volume	0.000	22:34	
Air Blank	0.000	22:35	
Control Test	0.079	22:35	
Air Blank	0.000	22:36	
Diagnostics Check	OK	22:36	

---

Data Download Date/Time: 03/23/2022 14:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVI,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAKE,LEVI  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/17/2022 01:30:07

---

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.079	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.000	01:25
Breath Volume	2.496	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.000	01:28
Breath Volume	2.292	01:28
Air Blank	0.000	01:29
Control Test	0.078	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

---

Data Download Date/Time: 03/23/2022 14:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVI,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAKE,LEVI  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: DANIEL IV

DATE/TIME OF TEST: 03/17/2022 02:57:51

---

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:55	
Control Test	0.078	02:55	
Air Blank	0.000	02:56	
Subject Sample	# 1	02:56	Subject Test Refused
Breath Volume	0.000	02:56	
Air Blank	0.000	02:56	
Control Test	0.079	02:57	
Air Blank	0.000	02:57	
Diagnostics Check	OK	02:57	

---

Data Download Date/Time: 03/23/2022 14:15



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: HIRSHMAN,BRIAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: ROMETO

DATE/TIME OF TEST: 03/17/2022 03:32:56

---

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:30	
Control Test	0.079	03:30	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:31	Subject Test Refused
Breath Volume	0.000	03:31	
Air Blank	0.000	03:32	
Control Test	0.080	03:32	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:32	

---

Data Download Date/Time: 03/23/2022 14:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVI,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAKE,LEVI  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: VALENTIN JR

DATE/TIME OF TEST: 03/17/2022 18:48:56

---

Diagnostics Check	OK	18:46	
Air Blank	0.000	18:46	
Control Test	0.080	18:46	
Air Blank	0.000	18:47	
Subject Sample	# 1	18:47	Subject Test Refused
Breath Volume	0.000	18:47	
Air Blank	0.000	18:48	
Control Test	0.079	18:48	
Air Blank	0.000	18:48	
Diagnostics Check	OK	18:48	

---

Data Download Date/Time: 03/23/2022 14:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVI,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAKE,LEVI  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: SERENA

DATE/TIME OF TEST: 03/17/2022 23:32:38

---

Diagnostics Check	OK	23:29	
Air Blank	0.000	23:30	
Control Test	0.079	23:30	
Air Blank	0.000	23:31	
Subject Sample	# 1	23:31	Subject Test Refused
Breath Volume	0.000	23:31	
Air Blank	0.000	23:31	
Control Test	0.079	23:32	
Air Blank	0.000	23:32	
Diagnostics Check	OK	23:32	

---

Data Download Date/Time: 03/23/2022 14:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVI,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAKE,LEVI  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: ROBINSON III

DATE/TIME OF TEST: 03/18/2022 01:16:11

---

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:13	
Control Test	0.079	01:14	
Air Blank	0.000	01:14	
Subject Sample	# 1	01:14	Subject Test Refused
Breath Volume	0.000	01:14	
Air Blank	0.000	01:15	
Control Test	0.079	01:15	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

---

Data Download Date/Time: 03/23/2022 14:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: HIRSHMAN,BRIAN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: SELIGMAN

DATE/TIME OF TEST: 03/18/2022 02:23:24

---

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:20	
Control Test	0.079	02:21	
Air Blank	0.000	02:21	
Subject Sample	# 1	02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:22	
Control Test	0.079	02:22	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

---

Data Download Date/Time: 03/23/2022 14:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVI,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAKE,LEVI  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 03/18/2022 03:11:14

---

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.079	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.214	03:06
Breath Volume	1.855	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.203	03:09
Breath Volume	2.476	03:09
Air Blank	0.000	03:10
Control Test	0.079	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

---

Data Download Date/Time: 03/23/2022 14:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIBLEY,TRAVIS,J  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: SIBLEY,TRAVIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SOJACK

DATE/TIME OF TEST: 03/18/2022 03:32:49

---

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:21	
Control Test		0.078	03:21	
Air Blank		0.000	03:21	
Subject Sample	# 1	0.081	03:25	Volume Not Met
Breath Volume		0.152	03:25	
Air Blank		0.000	03:25	
Air Blank		0.000	03:27	
Subject Sample	# 2	0.102	03:28	
Breath Volume		1.902	03:28	
Air Blank		0.000	03:28	
Air Blank		0.000	03:30	
Subject Sample	# 3		03:31	Subject Test Refused
Breath Volume		0.085	03:31	
Air Blank		0.000	03:31	
Control Test		0.078	03:32	
Air Blank		0.000	03:32	
Diagnostics Check		OK	03:32	

---

Data Download Date/Time: 03/23/2022 14:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVI,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: FIELDEN,ERNEST  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: BLAKE

DATE/TIME OF TEST: 03/18/2022 03:58:25

---

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.079	03:56	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Control Test	0.079	03:57	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

---

Data Download Date/Time: 03/23/2022 14:19



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVI,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAKE,LEVI  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 03/18/2022 22:00:55

---

Diagnostics Check	OK	21:58	
Air Blank	0.000	21:58	
Control Test	0.079	21:58	
Air Blank	0.000	21:59	
Subject Sample	# 1	21:59	Subject Test Refused
Breath Volume	0.000	21:59	
Air Blank	0.000	21:59	
Control Test	0.079	22:00	
Air Blank	0.000	22:00	
Diagnostics Check	OK	22:00	

---

Data Download Date/Time: 03/23/2022 14:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVI,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAKE,LEVI  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: DEVIA

DATE/TIME OF TEST: 03/18/2022 23:31:45

---

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.079	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.013	23:27
Breath Volume	3.085	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:29
Subject Sample # 2	0.012	23:30
Breath Volume	3.378	23:30
Air Blank	0.000	23:30
Control Test	0.079	23:31
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

---

Data Download Date/Time: 03/23/2022 14:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVI,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAKE,LEVI  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 03/19/2022 01:28:19

---

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:25	
Control Test	0.079	01:26	
Air Blank	0.000	01:26	
Subject Sample	# 1	01:26	Subject Test Refused
Breath Volume	0.000	01:26	
Air Blank	0.000	01:27	
Control Test	0.079	01:27	
Air Blank		01:27	RFI Detect
Air Blank	0.000	01:28	

---

Data Download Date/Time: 03/23/2022 14:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BLAKE,LEVI,S  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAKE,LEVI  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG106803  
DRY GAS EXPIRY DATE: 03/09/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 03/19/2022 01:32:20

---

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:29	
Control Test	0.079	01:30	
Air Blank	0.000	01:30	
Subject Sample	# 1	01:30	Subject Test Refused
Breath Volume	0.000	01:30	
Air Blank	0.000	01:31	
Control Test	0.079	01:31	
Air Blank	0.000	01:32	
Diagnostics Check	OK	01:32	

---

Data Download Date/Time: 03/23/2022 14:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNER,SEAN,R  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: BRUNER,SEAN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: MAYA

DATE/TIME OF TEST: 03/12/2022 00:21:47

---

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.082	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.000	00:17
Breath Volume	2.382	00:17
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.000	00:20
Breath Volume	1.691	00:20
Air Blank	0.000	00:20
Control Test	0.081	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

---

Data Download Date/Time: 03/30/2022 20:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNER,SEAN,R  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: TURNER,LATORIAN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: RYGH

DATE/TIME OF TEST: 03/13/2022 05:07:20

---

Diagnostics Check	OK	04:59
Air Blank	0.000	04:59
Control Test	0.082	04:59
Air Blank	0.000	05:00
Subject Sample # 1	0.070	05:02
Breath Volume	1.605	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:04
Subject Sample # 2	0.070	05:05
Breath Volume	1.285	05:05
Air Blank	0.000	05:06
Control Test	0.081	05:06
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

---

Data Download Date/Time: 03/30/2022 20:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNER,SEAN,R  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: BRUNER,SEAN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: PRESTON

DATE/TIME OF TEST: 03/26/2022 03:34:40

---

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.081	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.121	03:30
Breath Volume		1.867	03:30
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.116	03:33
Breath Volume		2.000	03:33
Air Blank		0.000	03:33
Control Test		0.081	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

---

Data Download Date/Time: 03/30/2022 20:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNER,SEAN,R  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: BRUNER,SEAN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1356262  
DRY GAS EXPIRY DATE: 08/03/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SCHULTZ

DATE/TIME OF TEST: 03/27/2022 02:55:40

---

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:53	
Control Test	0.081	02:53	
Air Blank	0.000	02:54	
Subject Sample	# 1	02:54	Subject Test Refused
Breath Volume	0.000	02:54	
Air Blank	0.000	02:54	
Control Test	0.081	02:55	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

---

Data Download Date/Time: 03/30/2022 20:06



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCLAUGHLIN, BRADLEY,  
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO  
ARRESTING OFFICER: MCLAUGHLIN, BRADLEY  
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO  
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 25721080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/22/2022 01:14:42

---

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:03	
Control Test	0.082	01:04	
Air Blank	0.000	01:04	
Subject Sample	# 1	01:07	No Sample Provided
Breath Volume	0.000	01:07	
Air Blank	0.000	01:08	
Air Blank	0.000	01:10	
Subject Sample	# 2	01:13	No Sample Provided
Breath Volume	0.000	01:13	
Air Blank	0.000	01:13	
Control Test	0.081	01:14	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

---

Data Download Date/Time: 03/28/2022 14:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBIN,GASTON,A  
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO  
ARRESTING OFFICER: CHAMBERLAIN,J  
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO  
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 25721080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: MISHRA

DATE/TIME OF TEST: 03/24/2022 21:17:09

---

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.082	21:11
Air Blank	0.000	21:11
Subject Sample # 1	0.167	21:12
Breath Volume	2.570	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.168	21:15
Breath Volume	2.542	21:15
Air Blank	0.000	21:16
Control Test	0.083	21:16
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

---

Data Download Date/Time: 03/28/2022 14:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCLAUGHLIN, BRADLEY,  
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO  
ARRESTING OFFICER: MCLAUGHLIN, BRADLEY  
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO  
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 25721080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 03/26/2022 02:50:17

---

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.082	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.147	02:45
Breath Volume	1.503	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:47
Subject Sample # 2	0.155	02:48
Breath Volume	1.468	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

---

Data Download Date/Time: 03/28/2022 14:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARRAGHER, THOMAS, J  
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO  
ARRESTING OFFICER: FARRAGHER, THOMAS  
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO  
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 25721080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: ESQUIVEL

DATE/TIME OF TEST: 03/28/2022 02:46:03

---

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.081	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.319	02:40
Breath Volume	3.164	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.318	02:44
Breath Volume	1.542	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:46

---

Data Download Date/Time: 03/28/2022 14:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: ULYSSE,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 03/02/2022 04:02:09

---

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.081	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.074	03:57
Breath Volume	1.589	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.073	04:00
Breath Volume	2.289	04:00
Air Blank	0.000	04:01
Control Test	0.081	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

---

Data Download Date/Time: 03/10/2022 03:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DURHAM,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: HEGGS

DATE/TIME OF TEST: 03/02/2022 23:11:48

---

Diagnostics Check	OK	23:01
Air Blank	0.000	23:02
Control Test	0.080	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.176	23:05
Breath Volume	1.632	23:05
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.173	23:09
Breath Volume	1.464	23:09
Air Blank	0.000	23:10
Control Test	0.079	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

---

Data Download Date/Time: 03/10/2022 03:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: SOSA

DATE/TIME OF TEST: 03/03/2022 01:15:25

---

Diagnostics Check		OK	01:08
Air Blank		0.000	01:09
Control Test		0.080	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.163	01:10
Breath Volume		3.972	01:10
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.167	01:13
Breath Volume		3.871	01:13
Air Blank		0.000	01:14
Control Test		0.080	01:14
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

---

Data Download Date/Time: 03/10/2022 03:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville Beach  
ARRESTING OFFICER: IKRAM,Y  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: PICKERING

DATE/TIME OF TEST: 03/04/2022 03:00:57

---

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:58	
Control Test	0.082	02:58	
Air Blank	0.000	02:59	
Subject Sample	# 1	02:59	Subject Test Refused
Breath Volume	0.000	02:59	
Air Blank	0.000	02:59	
Control Test	0.081	03:00	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

---

Data Download Date/Time: 03/10/2022 03:29



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: PARRISH,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: WILBER

DATE/TIME OF TEST: 03/05/2022 02:12:19

---

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:09	
Control Test	0.080	02:10	
Air Blank	0.000	02:10	
Subject Sample	# 1	02:10	Subject Test Refused
Breath Volume	0.000	02:10	
Air Blank	0.000	02:11	
Control Test	0.080	02:11	
Air Blank	0.000	02:12	
Diagnostics Check	OK	02:12	

---

Data Download Date/Time: 03/10/2022 03:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville Beach  
ARRESTING OFFICER: KEEN,C  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: REDING

DATE/TIME OF TEST: 03/05/2022 03:46:23

---

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:43	
Control Test	0.082	03:44	
Air Blank	0.000	03:44	
Subject Sample	# 1	03:44	Subject Test Refused
Breath Volume	0.000	03:44	
Air Blank	0.000	03:45	
Control Test	0.081	03:45	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

---

Data Download Date/Time: 03/10/2022 03:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SATHER,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/05/2022 04:48:51

---

Diagnostics Check		OK	04:36	
Air Blank		0.000	04:37	
Control Test		0.081	04:37	
Air Blank		0.000	04:38	
Subject Sample	# 1	0.138	04:41	Volume Not Met
Breath Volume		0.035	04:41	
Air Blank		0.000	04:42	
Air Blank		0.000	04:43	
Subject Sample	# 2	0.122	04:46	Volume Not Met
Breath Volume		0.015	04:46	
Air Blank		0.000	04:47	
Control Test		0.077	04:48	
Air Blank		0.000	04:48	
Diagnostics Check		OK	04:48	

---

Data Download Date/Time: 03/10/2022 03:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: GARCIA CARRILLO

DATE/TIME OF TEST: 03/06/2022 01:53:13

---

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:50	
Control Test	0.080	01:51	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:52	
Control Test	0.080	01:52	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

---

Data Download Date/Time: 03/10/2022 03:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HILLIER,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: ORTIZ MONTALVO

DATE/TIME OF TEST: 03/06/2022 03:31:19

---

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.080	03:23
Air Blank	0.000	03:23
Subject Sample # 1	0.047	03:26
Breath Volume	1.703	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.047	03:29
Breath Volume	1.355	03:29
Air Blank	0.000	03:30
Control Test	0.079	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

---

Data Download Date/Time: 03/10/2022 03:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HARRELL,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: ZGAMBO

DATE/TIME OF TEST: 03/06/2022 04:58:05

---

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.080	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.171	04:52
Breath Volume	1.152	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.168	04:56
Breath Volume	1.261	04:56
Air Blank	0.000	04:57
Control Test	0.080	04:57
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

---

Data Download Date/Time: 03/10/2022 03:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SIMMONS,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:32

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 03/06/2022 06:02:11

---

Diagnostics Check	OK	05:53
Air Blank	0.000	05:53
Control Test	0.080	05:54
Air Blank	0.000	05:54
Subject Sample # 1	0.153	05:56
Breath Volume	1.335	05:56
Air Blank	0.000	05:57
Air Blank	0.000	05:59
Subject Sample # 2	0.153	06:00
Breath Volume	1.683	06:00
Air Blank	0.000	06:01
Control Test	0.079	06:01
Air Blank	0.000	06:02
Diagnostics Check	OK	06:02

---

Data Download Date/Time: 03/10/2022 03:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LOKMIC,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 03/06/2022 22:21:51

---

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.081	22:15
Air Blank	0.000	22:15
Subject Sample # 1	0.205	22:16
Breath Volume	1.546	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.192	22:20
Breath Volume	2.726	22:20
Air Blank	0.000	22:20
Control Test	0.080	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

---

Data Download Date/Time: 03/10/2022 03:29



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HILLIER,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: TRAN

DATE/TIME OF TEST: 03/07/2022 01:50:06

---

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.013	01:44
Breath Volume	1.105	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.012	01:48
Breath Volume	1.468	01:48
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

---

Data Download Date/Time: 03/10/2022 03:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HARRELL,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 03/07/2022 05:01:18

---

Diagnostics Check	OK	04:54
Air Blank	0.000	04:54
Control Test	0.081	04:55
Air Blank	0.000	04:55
Subject Sample # 1	0.115	04:56
Breath Volume	1.429	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:58
Subject Sample # 2	0.122	04:59
Breath Volume	1.660	04:59
Air Blank	0.000	05:00
Control Test	0.081	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

---

Data Download Date/Time: 03/10/2022 03:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CLARK,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: SADA

DATE/TIME OF TEST: 03/07/2022 23:48:02

---

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:45	
Control Test	0.079	23:45	
Air Blank	0.000	23:46	
Subject Sample	# 1	23:46	Subject Test Refused
Breath Volume	0.000	23:46	
Air Blank	0.000	23:47	
Control Test	0.079	23:47	
Air Blank	0.000	23:47	
Diagnostics Check	OK	23:48	

---

Data Download Date/Time: 03/10/2022 03:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MUCA

DATE/TIME OF TEST: 03/08/2022 02:46:05

---

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:43	
Control Test	0.080	02:43	
Air Blank	0.000	02:44	
Subject Sample	# 1	02:44	Subject Test Refused
Breath Volume	0.000	02:44	
Air Blank	0.000	02:45	
Control Test	0.080	02:45	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

---

Data Download Date/Time: 03/10/2022 03:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Atlantic Beach PD  
ARRESTING OFFICER: JOCKERS,B  
ARRESTING OFFICER AGENCY: Atlantic Beach PD  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: LOGUE

DATE/TIME OF TEST: 03/08/2022 03:58:31

---

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:56	
Control Test	0.080	03:56	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Control Test	0.080	03:57	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

---

Data Download Date/Time: 03/10/2022 03:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLARIDA,K,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DIAZ,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 08:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:36

SUBJECT'S LAST NAME: COSTE

DATE/TIME OF TEST: 03/09/2022 11:03:10

---

Diagnostics Check	OK	10:56
Air Blank	0.000	10:57
Control Test	0.080	10:57
Air Blank	0.000	10:57
Subject Sample # 1	0.000	10:58
Breath Volume	4.117	10:58
Air Blank	0.000	10:59
Air Blank	0.000	11:01
Subject Sample # 2	0.000	11:01
Breath Volume	3.429	11:01
Air Blank	0.000	11:02
Control Test	0.081	11:02
Air Blank	0.000	11:03
Diagnostics Check	OK	11:03

---

Data Download Date/Time: 03/10/2022 03:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAGOMENI,G,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: IKRAM,Y  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: NEMETH

DATE/TIME OF TEST: 03/11/2022 00:33:12

---

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.095	00:28
Breath Volume	1.152	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:30
Subject Sample # 2	0.100	00:31
Breath Volume	1.179	00:31
Air Blank	0.000	00:32
Control Test	0.081	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

---

Data Download Date/Time: 03/23/2022 14:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAGOMENI,G,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LAQUIDARA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: PITH

DATE/TIME OF TEST: 03/11/2022 02:25:35

---

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.081	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.076	02:20
Breath Volume	1.339	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.078	02:23
Breath Volume	1.789	02:23
Air Blank	0.000	02:24
Control Test	0.080	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

---

Data Download Date/Time: 03/23/2022 14:47



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POWELL,EDWARD,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LUGO,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 03/12/2022 00:06:06

---

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:03	
Control Test	0.080	00:03	
Air Blank	0.000	00:04	
Subject Sample	# 1	00:04	Subject Test Refused
Breath Volume	0.000	00:04	
Air Blank	0.000	00:05	
Control Test	0.080	00:05	
Air Blank	0.000	00:06	
Diagnostics Check	OK	00:06	

---

Data Download Date/Time: 03/23/2022 14:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAGOMENI,G,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LAQUIDARA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CARMICHAEL

DATE/TIME OF TEST: 03/12/2022 01:53:28

---

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:51	
Control Test	0.081	01:51	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:52	
Control Test	0.080	01:52	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

---

Data Download Date/Time: 03/23/2022 14:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: PARRISH,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: DURRANCE

DATE/TIME OF TEST: 03/13/2022 01:22:43

---

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:19	
Control Test	0.081	01:20	
Air Blank	0.000	01:20	
Subject Sample	# 1	01:21	Subject Test Refused
Breath Volume	0.000	01:21	
Air Blank	0.000	01:21	
Control Test	0.082	01:22	
Air Blank	0.000	01:22	
Diagnostics Check	OK	01:22	

---

Data Download Date/Time: 03/23/2022 14:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HARRELL,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MANEY

DATE/TIME OF TEST: 03/13/2022 03:27:21

---

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.081	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.116	03:22
Breath Volume	1.472	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:24
Subject Sample # 2	0.113	03:25
Breath Volume	1.519	03:25
Air Blank	0.000	03:26
Control Test	0.082	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

---

Data Download Date/Time: 03/23/2022 14:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KEEN,C  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 03/13/2022 05:14:06

---

Diagnostics Check		OK	05:07
Air Blank		0.000	05:07
Control Test		0.082	05:08
Air Blank		0.000	05:08
Subject Sample	# 1	0.182	05:09
Breath Volume		1.945	05:09
Air Blank		0.000	05:10
Air Blank		0.000	05:11
Subject Sample	# 2	0.181	05:12
Breath Volume		2.046	05:12
Air Blank		0.000	05:13
Control Test		0.081	05:13
Air Blank		0.000	05:14
Diagnostics Check		OK	05:14

---

Data Download Date/Time: 03/23/2022 14:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FLORY,J  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/13/2022 07:14:33

---

Diagnostics Check	OK	07:11	
Air Blank	0.000	07:12	
Control Test	0.082	07:12	
Air Blank	0.000	07:12	
Subject Sample	# 1	07:13	Subject Test Refused
Breath Volume	0.000	07:13	
Air Blank	0.000	07:13	
Control Test	0.082	07:13	
Air Blank	0.000	07:14	
Diagnostics Check	OK	07:14	

---

Data Download Date/Time: 03/23/2022 14:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: PETERSON,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 03/14/2022 01:16:10

---

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:13	
Control Test	0.082	01:13	
Air Blank	0.000	01:14	
Subject Sample	# 1	01:14	Subject Test Refused
Breath Volume	0.000	01:14	
Air Blank	0.000	01:15	
Control Test	0.082	01:15	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

---

Data Download Date/Time: 03/23/2022 14:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HIXSON,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 03/14/2022 03:59:07

---

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.081	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 1	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:58	
Control Test	0.081	03:58	
Air Blank	0.000	03:59	
Diagnostics Check	OK	03:59	

---

Data Download Date/Time: 03/23/2022 14:48



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MATTIACE,R  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: BARELA

DATE/TIME OF TEST: 03/14/2022 22:11:26

---

Diagnostics Check	OK	22:08	
Air Blank	0.000	22:08	
Control Test	0.081	22:09	
Air Blank	0.000	22:09	
Subject Sample	# 1	22:10	Subject Test Refused
Breath Volume	0.000	22:10	
Air Blank	0.000	22:10	
Control Test	0.080	22:10	
Air Blank	0.000	22:11	
Diagnostics Check	OK	22:11	

---

Data Download Date/Time: 03/23/2022 14:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HARRELL,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 03/14/2022 23:40:34

---

Diagnostics Check	OK	23:33
Air Blank	0.000	23:34
Control Test	0.081	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.145	23:35
Breath Volume	1.472	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.136	23:38
Breath Volume	1.832	23:38
Air Blank	0.000	23:39
Control Test	0.081	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

---

Data Download Date/Time: 03/23/2022 14:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WILLIAMS,M  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: CHALCRAFT

DATE/TIME OF TEST: 03/15/2022 05:02:23

---

Diagnostics Check	OK	04:59	
Air Blank	0.000	04:59	
Control Test	0.082	05:00	
Air Blank	0.000	05:00	
Subject Sample	# 1	05:00	Subject Test Refused
Breath Volume	0.000	05:00	
Air Blank	0.000	05:01	
Control Test	0.081	05:01	
Air Blank	0.000	05:02	
Diagnostics Check	OK	05:02	

---

Data Download Date/Time: 03/23/2022 14:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COTHRAN,D,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FREY,G.B.  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 06:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:56

SUBJECT'S LAST NAME: ENLOE

DATE/TIME OF TEST: 03/15/2022 08:37:51

---

Diagnostics Check	OK	08:24	
Air Blank	0.000	08:24	
Control Test	0.081	08:25	
Air Blank	0.000	08:25	
Subject Sample	# 1 0.000	08:27	
Breath Volume	2.015	08:27	
Air Blank	0.000	08:28	
Air Blank	0.000	08:30	
Subject Sample	# 2	08:33	No Sample Provided
Breath Volume	0.000	08:33	
Air Blank	0.000	08:33	
Air Blank	0.000	08:35	
Subject Sample	# 3 0.000	08:36	
Breath Volume	2.097	08:36	
Air Blank	0.000	08:36	
Control Test	0.081	08:37	
Air Blank	0.000	08:37	
Diagnostics Check	OK	08:37	

---

Data Download Date/Time: 03/23/2022 14:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COTHRAN,D,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LOKMIC,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 15:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:08

SUBJECT'S LAST NAME: PIERUCKI

DATE/TIME OF TEST: 03/15/2022 17:36:51

---

Diagnostics Check	OK	17:28
Air Blank	0.000	17:29
Control Test	0.080	17:29
Air Blank	0.000	17:30
Subject Sample # 1	0.000	17:32
Breath Volume	4.117	17:32
Air Blank	0.000	17:32
Air Blank	0.000	17:34
Subject Sample # 2	0.000	17:35
Breath Volume	3.890	17:35
Air Blank	0.000	17:35
Control Test	0.080	17:36
Air Blank	0.000	17:36
Diagnostics Check	OK	17:36

---

Data Download Date/Time: 03/23/2022 14:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CLARK,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: SPARKS

DATE/TIME OF TEST: 03/15/2022 21:52:01

---

Diagnostics Check	OK	21:45
Air Blank	0.000	21:46
Control Test	0.081	21:46
Air Blank	0.000	21:47
Subject Sample # 1	0.000	21:47
Breath Volume	1.699	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 2	0.000	21:50
Breath Volume	1.535	21:50
Air Blank	0.000	21:51
Control Test	0.081	21:51
Air Blank	0.000	21:51
Diagnostics Check	OK	21:52

---

Data Download Date/Time: 03/23/2022 14:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COTHRAN,D,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LOKMIC,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 15:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: BERDUGO LOPEZ

DATE/TIME OF TEST: 03/16/2022 16:47:23

---

Diagnostics Check		OK	16:40
Air Blank		0.000	16:40
Control Test		0.081	16:41
Air Blank		0.000	16:41
Subject Sample	# 1	0.206	16:42
Breath Volume		1.714	16:42
Air Blank		0.000	16:43
Air Blank		0.000	16:44
Subject Sample	# 2	0.196	16:45
Breath Volume		1.226	16:45
Air Blank		0.000	16:46
Control Test		0.079	16:46
Air Blank		0.000	16:47
Diagnostics Check		OK	16:47

---

Data Download Date/Time: 03/23/2022 14:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEURER,D,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HARRELL,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: PRESCOTT

DATE/TIME OF TEST: 03/16/2022 23:02:32

---

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.081	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.052	22:57
Breath Volume	1.324	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.051	23:00
Breath Volume	1.273	23:00
Air Blank	0.000	23:01
Control Test	0.081	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

---

Data Download Date/Time: 03/23/2022 14:49



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HARRELL,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: TELFAIR

DATE/TIME OF TEST: 03/17/2022 03:53:00

---

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.081	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:52	
Control Test	0.081	03:52	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:53	

---

Data Download Date/Time: 03/23/2022 14:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAGOMENI,G,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: STANLEY,T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 03/17/2022 23:43:04

---

Diagnostics Check	OK	23:35
Air Blank	0.000	23:35
Control Test	0.081	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.145	23:37
Breath Volume	1.480	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:39
Subject Sample # 2	0.139	23:41
Breath Volume	1.769	23:41
Air Blank	0.000	23:42
Control Test	0.081	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:43

---

Data Download Date/Time: 03/23/2022 14:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAGOMENI,G,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/18/2022 01:30:40

---

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.080	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.198	01:23
Breath Volume	1.332	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.196	01:28
Breath Volume	1.714	01:28
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

---

Data Download Date/Time: 03/23/2022 14:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STANFORD,B,R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: ULYSSE,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 03/19/2022 03:35:37

---

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:33	
Control Test	0.080	03:33	
Air Blank	0.000	03:34	
Subject Sample	# 1	03:34	Subject Test Refused
Breath Volume	0.000	03:34	
Air Blank	0.000	03:34	
Control Test	0.080	03:34	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:35	

---

Data Download Date/Time: 03/23/2022 14:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POWELL, EDWARD, A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: PATTERSON, K  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: HONEYCUTT

DATE/TIME OF TEST: 03/19/2022 04:32:19

---

Diagnostics Check	OK	04:29	
Air Blank	0.000	04:29	
Control Test	0.080	04:30	
Air Blank	0.000	04:30	
Subject Sample	# 1	04:30	Subject Test Refused
Breath Volume	0.000	04:30	
Air Blank	0.000	04:31	
Control Test	0.080	04:31	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:32	

---

Data Download Date/Time: 03/23/2022 14:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAGOMENI,G,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: STRICKLAND,T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:26

SUBJECT'S LAST NAME: FELTS

DATE/TIME OF TEST: 03/19/2022 06:53:31

---

Diagnostics Check	OK	06:47
Air Blank	0.000	06:47
Control Test	0.081	06:47
Air Blank	0.000	06:48
Subject Sample # 1	0.064	06:48
Breath Volume	1.777	06:48
Air Blank	0.000	06:49
Air Blank	0.000	06:51
Subject Sample # 2	0.065	06:51
Breath Volume	1.585	06:51
Air Blank	0.000	06:52
Control Test	0.080	06:52
Air Blank	0.000	06:53
Diagnostics Check	OK	06:53

---

Data Download Date/Time: 03/23/2022 14:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STANFORD,B,R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WILLIAMS,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/20/2022 01:44:17

---

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.189	01:39
Breath Volume	2.281	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:41
Subject Sample # 2	0.187	01:42
Breath Volume	1.628	01:42
Air Blank	0.000	01:43
Control Test	0.079	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

---

Data Download Date/Time: 03/23/2022 14:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STANFORD,B,R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITMAR,V  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: TOMA

DATE/TIME OF TEST: 03/20/2022 03:49:58

---

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:47	
Control Test	0.080	03:47	
Air Blank	0.000	03:48	
Subject Sample	# 1	03:48	Subject Test Refused
Breath Volume	0.000	03:48	
Air Blank	0.000	03:49	
Control Test	0.080	03:49	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

---

Data Download Date/Time: 03/23/2022 14:54



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DURHAM,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: NICCO

DATE/TIME OF TEST: 03/20/2022 05:28:13

---

Diagnostics Check		OK	05:21
Air Blank		0.000	05:21
Control Test		0.081	05:21
Air Blank		0.000	05:22
Subject Sample	# 1	0.119	05:23
Breath Volume		1.414	05:23
Air Blank		0.000	05:23
Air Blank		0.000	05:25
Subject Sample	# 2	0.124	05:26
Breath Volume		1.101	05:26
Air Blank		0.000	05:27
Control Test		0.079	05:27
Air Blank		0.000	05:28
Diagnostics Check		OK	05:28

---

Data Download Date/Time: 03/23/2022 14:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STANFORD,B,R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KEEN,C  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: YORKOVITCH

DATE/TIME OF TEST: 03/20/2022 20:41:36

---

Diagnostics Check	OK	20:35
Air Blank	0.000	20:35
Control Test	0.081	20:35
Air Blank	0.000	20:36
Subject Sample # 1	0.111	20:36
Breath Volume	4.148	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.109	20:39
Breath Volume	3.386	20:39
Air Blank	0.000	20:40
Control Test	0.081	20:40
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

---

Data Download Date/Time: 03/23/2022 14:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DURHAM,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/21/2022 01:36:57

---

Diagnostics Check		OK	01:30
Air Blank		0.000	01:30
Control Test		0.081	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.000	01:32
Breath Volume		1.222	01:32
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 2	0.000	01:35
Breath Volume		1.582	01:35
Air Blank		0.000	01:36
Control Test		0.082	01:36
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

---

Data Download Date/Time: 03/23/2022 14:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STANFORD,B,R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LUGO,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: MILES

DATE/TIME OF TEST: 03/21/2022 02:51:16

---

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.082	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.220	02:46
Breath Volume	2.679	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:48
Subject Sample # 2	0.206	02:49
Breath Volume	3.042	02:49
Air Blank	0.000	02:50
Control Test	0.081	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

---

Data Download Date/Time: 03/23/2022 14:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DESOUZA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: JAMESON

DATE/TIME OF TEST: 03/21/2022 03:49:22

---

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:46	
Control Test	0.082	03:47	
Air Blank	0.000	03:47	
Subject Sample	# 1	03:47	Subject Test Refused
Breath Volume	0.000	03:47	
Air Blank	0.000	03:48	
Control Test	0.082	03:48	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

---

Data Download Date/Time: 03/23/2022 14:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STANFORD,B,R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FLANDERS,T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:32

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 03/21/2022 06:05:04

---

Diagnostics Check	OK	06:02	
Air Blank	0.000	06:02	
Control Test	0.081	06:02	
Air Blank	0.000	06:03	
Subject Sample	# 1	06:03	Subject Test Refused
Breath Volume	0.000	06:03	
Air Blank	0.000	06:04	
Control Test	0.082	06:04	
Air Blank	0.000	06:04	
Diagnostics Check	OK	06:05	

---

Data Download Date/Time: 03/23/2022 14:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNSON,G  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 03/22/2022 00:25:08

---

Diagnostics Check	OK	00:22	
Air Blank	0.000	00:22	
Control Test	0.081	00:23	
Air Blank	0.000	00:23	
Subject Sample	# 1	00:23	Subject Test Refused
Breath Volume	0.000	00:23	
Air Blank	0.000	00:24	
Control Test	0.081	00:24	
Air Blank	0.000	00:25	
Diagnostics Check	OK	00:25	

---

Data Download Date/Time: 03/23/2022 14:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STANFORD,B,R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNS,B  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: ALDRIDGE

DATE/TIME OF TEST: 03/22/2022 02:22:55

---

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.081	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.000	02:18
Breath Volume	1.921	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.000	02:21
Breath Volume	1.949	02:21
Air Blank	0.000	02:21
Control Test	0.080	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

---

Data Download Date/Time: 03/23/2022 14:55



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SIMMONS,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 03/22/2022 22:44:57

---

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:42	
Control Test	0.080	22:42	
Air Blank	0.000	22:43	
Subject Sample	# 1	22:43	Subject Test Refused
Breath Volume	0.000	22:43	
Air Blank	0.000	22:43	
Control Test	0.080	22:44	
Air Blank	0.000	22:44	
Diagnostics Check	OK	22:44	

---

Data Download Date/Time: 03/23/2022 14:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KEEN,C  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: ZITTROWER

DATE/TIME OF TEST: 03/23/2022 03:02:11

---

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.080	03:00	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:01	
Control Test	0.081	03:01	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

---

Data Download Date/Time: 03/23/2022 14:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HILLIER,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: GELO

DATE/TIME OF TEST: 03/24/2022 00:58:13

---

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.081	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.146	00:52
Breath Volume		1.597	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.151	00:56
Breath Volume		2.167	00:56
Air Blank		0.000	00:57
Control Test		0.079	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

---

Data Download Date/Time: 04/15/2022 16:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HIXSON,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 03/24/2022 04:22:42

---

Diagnostics Check		OK	04:16
Air Blank		0.000	04:16
Control Test		0.080	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.218	04:18
Breath Volume		1.605	04:18
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.217	04:20
Breath Volume		1.660	04:20
Air Blank		0.000	04:21
Control Test		0.079	04:22
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

---

Data Download Date/Time: 04/15/2022 16:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,G,S  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HIXSON,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: ANSPACH

DATE/TIME OF TEST: 03/25/2022 02:31:01

---

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.109	02:26
Breath Volume	2.800	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:28
Subject Sample # 2	0.110	02:29
Breath Volume	4.042	02:29
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:31

---

Data Download Date/Time: 04/15/2022 16:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SULLIVAN,JARROD,S  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MEJIA,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 03/25/2022 04:25:46

---

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:23	
Control Test	0.081	04:23	
Air Blank	0.000	04:24	
Subject Sample	# 1	04:24	Subject Test Refused
Breath Volume	0.000	04:24	
Air Blank	0.000	04:24	
Control Test	0.080	04:25	
Air Blank	0.000	04:25	
Diagnostics Check	OK	04:25	

---

Data Download Date/Time: 04/15/2022 16:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,G,  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: PETERSON,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: UNDERWOOD

DATE/TIME OF TEST: 03/26/2022 06:16:26

---

Diagnostics Check		OK	06:09
Air Blank		0.000	06:10
Control Test		0.082	06:10
Air Blank		0.000	06:11
Subject Sample	# 1	0.231	06:11
Breath Volume		1.558	06:11
Air Blank		0.000	06:12
Air Blank		0.000	06:14
Subject Sample	# 2	0.226	06:14
Breath Volume		1.300	06:14
Air Blank		0.000	06:15
Control Test		0.080	06:15
Air Blank		0.000	06:16
Diagnostics Check		OK	06:16

---

Data Download Date/Time: 04/15/2022 16:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STANFORD,B,R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: JOHNS,B  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 03/26/2022 22:28:03

---

Diagnostics Check	OK	22:21
Air Blank	0.000	22:21
Control Test	0.081	22:22
Air Blank	0.000	22:22
Subject Sample # 1	0.107	22:23
Breath Volume	1.570	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:25
Subject Sample # 2	0.111	22:26
Breath Volume	2.257	22:26
Air Blank	0.000	22:27
Control Test	0.081	22:27
Air Blank	0.000	22:27
Diagnostics Check	OK	22:28

---

Data Download Date/Time: 04/15/2022 16:50



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STANFORD,B,R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: ULYSSE,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 03/27/2022 01:43:48

---

Diagnostics Check		OK	01:36
Air Blank		0.000	01:36
Control Test		0.081	01:37
Air Blank		0.000	01:37
Subject Sample	# 1	0.203	01:38
Breath Volume		1.960	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.205	01:42
Breath Volume		1.863	01:42
Air Blank		0.000	01:42
Control Test		0.079	01:43
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

---

Data Download Date/Time: 04/15/2022 16:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STANFORD,B,R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DESOUZA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: SUNDBERG

DATE/TIME OF TEST: 03/27/2022 03:00:40

---

Diagnostics Check		OK	02:52
Air Blank		0.000	02:53
Control Test		0.081	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.203	02:55
Breath Volume		1.992	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:57
Subject Sample	# 2	0.201	02:58
Breath Volume		2.296	02:58
Air Blank		0.000	02:59
Control Test		0.079	03:00
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

---

Data Download Date/Time: 04/15/2022 16:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STANFORD,B,R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: O'TOOLE

DATE/TIME OF TEST: 03/27/2022 03:28:47

---

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:26	
Control Test	0.080	03:26	
Air Blank	0.000	03:27	
Subject Sample	# 1	03:27	Subject Test Refused
Breath Volume	0.000	03:27	
Air Blank	0.000	03:27	
Control Test	0.081	03:28	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:28	

---

Data Download Date/Time: 04/15/2022 16:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LAQUIDARA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 17:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/27/2022 19:41:26

---

Diagnostics Check	OK	19:37	
Air Blank	0.000	19:37	
Control Test	0.081	19:38	
Air Blank	0.000	19:38	
Subject Sample	# 1	19:40	Subject Test Refused
Breath Volume	0.000	19:40	
Air Blank	0.000	19:40	
Control Test	0.081	19:40	
Air Blank	0.000	19:41	
Diagnostics Check	OK	19:41	

---

Data Download Date/Time: 04/15/2022 16:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WHITE,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: HIXSON

DATE/TIME OF TEST: 03/27/2022 20:32:04

---

Diagnostics Check	OK	20:29	
Air Blank	0.000	20:29	
Control Test	0.081	20:29	
Air Blank	0.000	20:30	
Subject Sample	# 1	20:30	Subject Test Refused
Breath Volume	0.000	20:30	
Air Blank	0.000	20:31	
Control Test	0.080	20:31	
Air Blank	0.000	20:31	
Diagnostics Check	OK	20:32	

---

Data Download Date/Time: 04/15/2022 16:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STANFORD,B,R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DURHAM,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: JUAN-SALAS

DATE/TIME OF TEST: 03/27/2022 23:51:08

---

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.081	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.185	23:46
Breath Volume	1.425	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:48
Subject Sample # 2	0.189	23:49
Breath Volume	1.425	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

---

Data Download Date/Time: 04/15/2022 16:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DESOUZA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: NUGENT

DATE/TIME OF TEST: 03/28/2022 01:13:55

---

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:11	
Control Test	0.081	01:11	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:12	
Control Test	0.081	01:13	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

---

Data Download Date/Time: 04/15/2022 16:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STANFORD,B,R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: TURPIN,D  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: DURHAM

DATE/TIME OF TEST: 03/28/2022 02:38:53

---

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.081	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.124	02:33
Breath Volume	1.503	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.124	02:37
Breath Volume	1.742	02:37
Air Blank	0.000	02:37
Control Test	0.079	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

---

Data Download Date/Time: 04/15/2022 16:51



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRAGOMENI,G,D  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: RODRIGUEZ,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: LYNN

DATE/TIME OF TEST: 03/29/2022 00:25:05

---

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.081	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.000	00:20
Breath Volume	1.875	00:20
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.000	00:23
Breath Volume	1.105	00:23
Air Blank	0.000	00:24
Control Test	0.080	00:24
Air Blank	0.000	00:24
Diagnostics Check	OK	00:25

---

Data Download Date/Time: 04/15/2022 16:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STANFORD,B,R  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DESOUZA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/30/2022 01:20:31

---

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:18	
Control Test	0.080	01:18	
Air Blank	0.000	01:18	
Subject Sample	# 1	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:19	
Control Test	0.080	01:19	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

---

Data Download Date/Time: 04/15/2022 16:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KEEN,C  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/31/2022 00:05:19

---

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:02	
Control Test	0.081	00:03	
Air Blank	0.000	00:03	
Subject Sample	# 1	00:03	Subject Test Refused
Breath Volume	0.000	00:03	
Air Blank	0.000	00:04	
Control Test	0.081	00:04	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:05	

---

Data Download Date/Time: 04/15/2022 16:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KEEN,C  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 17:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: RAILEY

DATE/TIME OF TEST: 03/31/2022 19:34:49

---

Diagnostics Check	OK	19:31	
Air Blank	0.000	19:32	
Control Test	0.080	19:32	
Air Blank	0.000	19:33	
Subject Sample	# 1	19:33	Subject Test Refused
Breath Volume	0.000	19:33	
Air Blank	0.000	19:33	
Control Test	0.080	19:34	
Air Blank	0.000	19:34	
Diagnostics Check	OK	19:34	

---

Data Download Date/Time: 04/15/2022 16:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SEAWARD,MORGANA,A  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: MYERS,TAYLOR  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 14:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:02

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 03/05/2022 15:28:25

---

Diagnostics Check	OK	15:22
Air Blank	0.000	15:22
Control Test	0.078	15:23
Air Blank	0.000	15:23
Subject Sample # 1	0.097	15:24
Breath Volume	1.792	15:24
Air Blank	0.000	15:24
Air Blank	0.000	15:26
Subject Sample # 2	0.097	15:26
Breath Volume	1.996	15:26
Air Blank	0.000	15:27
Control Test	0.079	15:27
Air Blank	0.000	15:28
Diagnostics Check	OK	15:28

---

Data Download Date/Time: 03/23/2022 16:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROLLESTON,JACOB,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ROLLESTON,JACOB  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 03/05/2022 23:07:38

---

Diagnostics Check		OK	23:01
Air Blank		0.000	23:02
Control Test		0.077	23:02
Air Blank		0.000	23:02
Subject Sample	# 1	0.253	23:03
Breath Volume		2.199	23:03
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.251	23:06
Breath Volume		2.128	23:06
Air Blank		0.000	23:06
Control Test		0.076	23:07
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

---

Data Download Date/Time: 03/23/2022 16:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DETRANO,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: DETRANO,JEFFREY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: TERRELL

DATE/TIME OF TEST: 03/09/2022 00:27:59

---

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:25	
Control Test	0.076	00:25	
Air Blank	0.000	00:26	
Subject Sample	# 1	00:26	Subject Test Refused
Breath Volume	0.000	00:26	
Air Blank	0.000	00:27	
Control Test	0.077	00:27	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

---

Data Download Date/Time: 03/23/2022 16:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOYLE,THOMAS,E  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: WAGNER-ALLARD,MITCHELL  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 10:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: WINKLER

DATE/TIME OF TEST: 03/10/2022 11:56:46

---

Diagnostics Check	OK	11:51
Air Blank	0.000	11:51
Control Test	0.077	11:51
Air Blank	0.000	11:52
Subject Sample # 1	0.000	11:52
Breath Volume	3.519	11:52
Air Blank	0.000	11:53
Air Blank	0.000	11:55
Subject Sample # 2	0.000	11:55
Breath Volume	3.722	11:55
Air Blank	0.000	11:55
Control Test	0.078	11:56
Air Blank	0.000	11:56
Diagnostics Check	OK	11:56

---

Data Download Date/Time: 03/23/2022 16:37



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DETRANO,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: DETRANO,JEFFREY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: MENCL

DATE/TIME OF TEST: 03/11/2022 23:12:33

---

Diagnostics Check	OK	23:09	
Air Blank	0.000	23:09	
Control Test	0.076	23:10	
Air Blank	0.000	23:10	
Subject Sample	# 1	23:11	Subject Test Refused
Breath Volume	0.000	23:11	
Air Blank	0.000	23:11	
Control Test	0.076	23:11	
Air Blank	0.000	23:12	
Diagnostics Check	OK	23:12	

---

Data Download Date/Time: 03/23/2022 16:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMPKINS,BENJAMIN,F  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: FARRINGTON,C  
ARRESTING OFFICER AGENCY: Kenneth City PD  
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: W425112876500

DATE/TIME OF TEST: 03/12/2022 04:48:02

---

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:42	
Control Test	0.077	04:43	
Air Blank	0.000	04:43	
Subject Sample	# 1	04:46	Subject Test Refused
Breath Volume	0.000	04:46	
Air Blank	0.000	04:47	
Control Test	0.078	04:47	
Air Blank	0.000	04:47	
Diagnostics Check	OK	04:48	

---

Data Download Date/Time: 03/23/2022 16:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DETRANO,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: DETRANO,JEFFREY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 03/16/2022 20:54:57

---

Diagnostics Check	OK	20:48
Air Blank	0.000	20:49
Control Test	0.076	20:49
Air Blank	0.000	20:50
Subject Sample # 1	0.000	20:50
Breath Volume	2.464	20:50
Air Blank	0.000	20:50
Air Blank	0.000	20:53
Subject Sample # 2	0.000	20:53
Breath Volume	2.339	20:53
Air Blank	0.000	20:54
Control Test	0.076	20:54
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

---

Data Download Date/Time: 03/23/2022 16:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DETRANO,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: DETRANO,JEFFREY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/18/2022 01:16:32

---

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:14	
Control Test	0.077	01:14	
Air Blank	0.000	01:14	
Subject Sample	# 1	01:15	Subject Test Refused
Breath Volume	0.000	01:15	
Air Blank	0.000	01:15	
Control Test	0.077	01:15	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

---

Data Download Date/Time: 03/23/2022 16:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DETRANO,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: DETRANO,JEFFREY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 03/18/2022 04:25:21

---

Diagnostics Check		OK	04:18
Air Blank		0.000	04:18
Control Test		0.077	04:19
Air Blank		0.000	04:19
Subject Sample	# 1	0.152	04:20
Breath Volume		1.269	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.158	04:23
Breath Volume		1.988	04:23
Air Blank		0.000	04:24
Control Test		0.077	04:24
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

---

Data Download Date/Time: 03/23/2022 16:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DETRANO,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: DETRANO,JEFFREY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: LUNDY

DATE/TIME OF TEST: 03/19/2022 21:20:15

---

Diagnostics Check	OK	21:17	
Air Blank	0.000	21:17	
Control Test	0.076	21:18	
Air Blank	0.000	21:18	
Subject Sample	# 1	21:18	Subject Test Refused
Breath Volume	0.000	21:18	
Air Blank	0.000	21:19	
Control Test	0.076	21:19	
Air Blank	0.000	21:20	
Diagnostics Check	OK	21:20	

---

Data Download Date/Time: 03/23/2022 16:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON,RUSSELL,A  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ADAMS,STERLING  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SEJDIC

DATE/TIME OF TEST: 03/20/2022 04:32:21

---

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.075	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.188	04:27
Breath Volume	1.742	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:29
Subject Sample # 2	0.195	04:30
Breath Volume	1.355	04:30
Air Blank	0.000	04:31
Control Test	0.078	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

---

Data Download Date/Time: 03/23/2022 16:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DETRANO,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: DETRANO,JEFFREY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: SLOTHOWER

DATE/TIME OF TEST: 03/22/2022 22:45:11

---

Diagnostics Check	OK	22:39
Air Blank	0.000	22:39
Control Test	0.075	22:40
Air Blank	0.000	22:40
Subject Sample # 1	0.136	22:40
Breath Volume	3.062	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.147	22:43
Breath Volume	2.957	22:43
Air Blank	0.000	22:44
Control Test	0.076	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

---

Data Download Date/Time: 03/23/2022 16:37



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DETRANO,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: DETRANO,JEFFREY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: CALDERON NAVA

DATE/TIME OF TEST: 03/23/2022 03:57:44

---

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.075	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.249	03:53
Breath Volume		2.000	03:53
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.252	03:56
Breath Volume		2.417	03:56
Air Blank		0.000	03:56
Control Test		0.076	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

---

Data Download Date/Time: 03/23/2022 16:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DETRANO,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: DETRANO,JEFFREY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: TRAINER

DATE/TIME OF TEST: 03/24/2022 19:18:17

---

Diagnostics Check	OK	19:15	
Air Blank	0.000	19:15	
Control Test	0.076	19:16	
Air Blank	0.000	19:16	
Subject Sample	# 1	19:16	Subject Test Refused
Breath Volume	0.000	19:16	
Air Blank	0.000	19:17	
Control Test	0.076	19:17	
Air Blank	0.000	19:18	
Diagnostics Check	OK	19:18	

---

Data Download Date/Time: 04/04/2022 11:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROLLESTON,JACOB,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ROLLESTON,JACOB  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 03/25/2022 23:00:11

---

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.076	22:54
Air Blank	0.000	22:54
Subject Sample # 1	0.144	22:55
Breath Volume	2.058	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.151	22:58
Breath Volume	2.035	22:58
Air Blank	0.000	22:59
Control Test	0.077	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

---

Data Download Date/Time: 04/04/2022 11:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DETRANO,JEFFREY,J  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: DETRANO,JEFFREY  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: TULLY

DATE/TIME OF TEST: 03/26/2022 01:14:54

---

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:12	
Control Test	0.077	01:12	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:13	Subject Test Refused
Breath Volume	0.000	01:13	
Air Blank	0.000	01:14	
Control Test	0.077	01:14	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

---

Data Download Date/Time: 04/04/2022 11:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROLLESTON,JACOB,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ROLLESTON,JACOB  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 03/27/2022 01:52:38

---

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.077	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.181	01:47
Breath Volume	1.234	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.188	01:51
Breath Volume	1.539	01:51
Air Blank	0.000	01:51
Control Test	0.078	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

---

Data Download Date/Time: 04/04/2022 11:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON,RUSSELL,A  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ADAMS,S  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: PECK

DATE/TIME OF TEST: 03/28/2022 23:27:52

---

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:25	
Control Test	0.077	23:25	
Air Blank	0.000	23:26	
Subject Sample	# 1	23:26	Subject Test Refused
Breath Volume	0.000	23:26	
Air Blank	0.000	23:26	
Control Test	0.076	23:27	
Air Blank	0.000	23:27	
Diagnostics Check	OK	23:27	

---

Data Download Date/Time: 04/04/2022 11:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROLLESTON,JACOB,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ROLLESTON,JACOB  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 08121080A2  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: AUGER

DATE/TIME OF TEST: 03/31/2022 23:28:58

---

Diagnostics Check	OK	23:26	
Air Blank	0.000	23:26	
Control Test	0.077	23:26	
Air Blank	0.000	23:27	
Subject Sample	# 1	23:27	Subject Test Refused
Breath Volume	0.000	23:27	
Air Blank	0.000	23:28	
Control Test	0.076	23:28	
Air Blank	0.000	23:28	
Diagnostics Check	OK	23:28	

---

Data Download Date/Time: 04/04/2022 11:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY,GAVIN,O  
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf  
ARRESTING OFFICER: HENRY,GAVIN  
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf  
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02422080A1  
DRY GAS EXPIRY DATE: 02/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PFAFFENBERGER

DATE/TIME OF TEST: 03/03/2022 02:45:29

---

Diagnostics Check		OK	02:39
Air Blank		0.000	02:39
Control Test		0.079	02:40
Air Blank		0.000	02:40
Subject Sample	# 1	0.169	02:41
Breath Volume		3.746	02:41
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.167	02:44
Breath Volume		2.738	02:44
Air Blank		0.000	02:44
Control Test		0.077	02:44
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

---

Data Download Date/Time: 03/24/2022 12:47



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY,GAVIN,O  
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf  
ARRESTING OFFICER: HENRY,GAVIN  
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf  
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 02422080A1  
DRY GAS EXPIRY DATE: 02/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: MENTOR

DATE/TIME OF TEST: 03/16/2022 04:23:33

---

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:18
Air Blank	0.000	04:18
Subject Sample # 1	0.119	04:19
Breath Volume	2.289	04:19
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.117	04:22
Breath Volume	1.953	04:22
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

---

Data Download Date/Time: 03/24/2022 12:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEURER,D,E  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DURHAM,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 03/02/2022 03:15:16

---

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.081	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.120	03:10
Breath Volume		3.734	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.125	03:13
Breath Volume		4.003	03:13
Air Blank		0.000	03:14
Control Test		0.081	03:14
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

---

Data Download Date/Time: 03/10/2022 04:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TERRELL,J,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: DELBROCCO-BLUES,D  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: BLARE

DATE/TIME OF TEST: 03/03/2022 05:20:29

---

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.081	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.192	05:15
Breath Volume	2.671	05:15
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.186	05:18
Breath Volume	1.773	05:18
Air Blank	0.000	05:19
Control Test	0.081	05:19
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

---

Data Download Date/Time: 03/10/2022 04:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HILLIER,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/03/2022 22:04:17

---

Diagnostics Check		OK	21:52	
Air Blank		0.000	21:52	
Control Test		0.081	21:52	
Air Blank		0.000	21:53	
Subject Sample	# 1	0.048	21:56	Volume Not Met
Breath Volume		0.257	21:56	
Air Blank		0.000	21:57	
Air Blank		0.000	21:59	
Subject Sample	# 2	0.054	21:59	
Breath Volume		1.269	21:59	
Air Blank		0.000	22:00	
Air Blank		0.000	22:02	
Subject Sample	# 3	0.055	22:02	
Breath Volume		1.187	22:02	
Air Blank		0.000	22:03	
Control Test		0.081	22:03	
Air Blank		0.000	22:04	
Diagnostics Check		OK	22:04	

---

Data Download Date/Time: 03/10/2022 04:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HIXSON,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/04/2022 02:29:39

---

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:27	
Control Test	0.082	02:27	
Air Blank	0.000	02:28	
Subject Sample	# 1	02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:28	
Control Test	0.082	02:29	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

---

Data Download Date/Time: 03/10/2022 04:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SATHER,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: FLOYD

DATE/TIME OF TEST: 03/04/2022 06:01:59

---

Diagnostics Check	OK	05:59	
Air Blank	0.000	05:59	
Control Test	0.082	05:59	
Air Blank	0.000	06:00	
Subject Sample	# 1	06:00	Subject Test Refused
Breath Volume	0.000	06:00	
Air Blank	0.000	06:01	
Control Test	0.082	06:01	
Air Blank	0.000	06:01	
Diagnostics Check	OK	06:01	

---

Data Download Date/Time: 03/10/2022 04:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville Beach  
ARRESTING OFFICER: KEEN,C  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: RYMER

DATE/TIME OF TEST: 03/04/2022 20:49:23

---

Diagnostics Check	OK	20:46	
Air Blank	0.000	20:46	
Control Test	0.081	20:47	
Air Blank	0.000	20:47	
Subject Sample	# 1	20:48	Subject Test Refused
Breath Volume	0.000	20:48	
Air Blank	0.000	20:48	
Control Test	0.081	20:48	
Air Blank	0.000	20:49	
Diagnostics Check	OK	20:49	

---

Data Download Date/Time: 03/10/2022 04:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: PETERSON,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: MAINOR

DATE/TIME OF TEST: 03/04/2022 22:29:35

---

Diagnostics Check	OK	22:26	
Air Blank	0.000	22:27	
Control Test	0.081	22:27	
Air Blank	0.000	22:28	
Subject Sample	# 1	22:28	Subject Test Refused
Breath Volume	0.000	22:28	
Air Blank	0.000	22:28	
Control Test	0.081	22:28	
Air Blank	0.000	22:29	
Diagnostics Check	OK	22:29	

---

Data Download Date/Time: 03/10/2022 04:28



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HIXSON,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: MONTOYA

DATE/TIME OF TEST: 03/05/2022 02:06:01

---

Diagnostics Check	OK	01:59
Air Blank	0.000	02:00
Control Test	0.081	02:00
Air Blank	0.000	02:01
Subject Sample # 1	0.039	02:01
Breath Volume	2.648	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.038	02:04
Breath Volume	2.363	02:04
Air Blank	0.000	02:05
Control Test	0.081	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:06

---

Data Download Date/Time: 03/10/2022 04:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HILLIER,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 03/05/2022 03:06:45

---

Diagnostics Check	OK	03:03	
Air Blank	0.000	03:04	
Control Test	0.081	03:04	
Air Blank	0.000	03:05	
Subject Sample	# 1	03:05	Subject Test Refused
Breath Volume	0.000	03:05	
Air Blank	0.000	03:05	
Control Test	0.081	03:06	
Air Blank	0.000	03:06	
Diagnostics Check	OK	03:06	

---

Data Download Date/Time: 03/10/2022 04:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CINOTTI,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: BOWENS

DATE/TIME OF TEST: 03/05/2022 04:34:13

---

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.081	04:28
Air Blank	0.000	04:28
Subject Sample # 1	0.083	04:29
Breath Volume	2.242	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.079	04:32
Breath Volume	1.980	04:32
Air Blank	0.000	04:33
Control Test	0.080	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

---

Data Download Date/Time: 03/10/2022 04:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: OSBORNE,I  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: DE COSTA

DATE/TIME OF TEST: 03/05/2022 06:12:51

---

Diagnostics Check	OK	06:09	
Air Blank	0.000	06:10	
Control Test	0.081	06:10	
Air Blank	0.000	06:11	
Subject Sample	# 1	06:11	Subject Test Refused
Breath Volume	0.000	06:11	
Air Blank	0.000	06:11	
Control Test	0.081	06:12	
Air Blank	0.000	06:12	
Diagnostics Check	OK	06:12	

---

Data Download Date/Time: 03/10/2022 04:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSES,L,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: MILEY,L  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:11

SUBJECT'S LAST NAME: CERNAS

DATE/TIME OF TEST: 03/05/2022 07:35:02

---

Diagnostics Check	OK	07:31	
Air Blank	0.000	07:31	
Control Test	0.081	07:32	
Air Blank	0.000	07:32	
Subject Sample	# 1	07:33	Subject Test Refused
Breath Volume	0.000	07:33	
Air Blank	0.000	07:34	
Control Test	0.081	07:34	
Air Blank	0.000	07:34	
Diagnostics Check	OK	07:35	

---

Data Download Date/Time: 03/10/2022 04:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COTHRAN,D,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: TAYLOR,G.A.  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: SANCHEZ DAZA

DATE/TIME OF TEST: 03/05/2022 10:00:31

---

Diagnostics Check	OK	09:51
Air Blank	0.000	09:51
Control Test	0.081	09:52
Air Blank	0.000	09:52
Subject Sample # 1	0.059	09:53
Breath Volume	2.140	09:53
Air Blank	0.000	09:54
Air Blank	0.000	09:55
Subject Sample # 2	0.059	09:58
Breath Volume	3.628	09:58
Air Blank	0.000	09:59
Control Test	0.081	09:59
Air Blank	0.000	10:00
Diagnostics Check	OK	10:00

---

Data Download Date/Time: 03/10/2022 04:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COTHRAN,D,A  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LOKMIC,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: ELIEZER

DATE/TIME OF TEST: 03/05/2022 17:07:39

---

Diagnostics Check	OK	17:00
Air Blank	0.000	17:01
Control Test	0.081	17:01
Air Blank	0.000	17:02
Subject Sample # 1	0.212	17:03
Breath Volume	2.308	17:03
Air Blank	0.000	17:03
Air Blank	0.000	17:05
Subject Sample # 2	0.209	17:06
Breath Volume	1.835	17:06
Air Blank	0.000	17:06
Control Test	0.080	17:07
Air Blank	0.000	17:07
Diagnostics Check	OK	17:07

---

Data Download Date/Time: 03/10/2022 04:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:59

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 03/05/2022 19:26:30

---

Diagnostics Check	OK	19:20
Air Blank	0.000	19:20
Control Test	0.081	19:20
Air Blank	0.000	19:21
Subject Sample # 1	0.000	19:22
Breath Volume	1.878	19:22
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.000	19:25
Breath Volume	2.039	19:25
Air Blank	0.000	19:25
Control Test	0.081	19:25
Air Blank	0.000	19:26
Diagnostics Check	OK	19:26

---

Data Download Date/Time: 03/10/2022 04:29



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SIMMONS,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/05/2022 21:42:54

---

Diagnostics Check	OK	21:36
Air Blank	0.000	21:36
Control Test	0.081	21:36
Air Blank	0.000	21:37
Subject Sample # 1	0.123	21:38
Breath Volume	1.855	21:38
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.128	21:41
Breath Volume	1.742	21:41
Air Blank	0.000	21:41
Control Test	0.081	21:42
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

---

Data Download Date/Time: 03/10/2022 04:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HIXSON,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: FREDERICK

DATE/TIME OF TEST: 03/06/2022 00:07:27

---

Diagnostics Check		OK	00:00
Air Blank		0.000	00:00
Control Test		0.081	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.134	00:02
Breath Volume		2.625	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.129	00:05
Breath Volume		2.050	00:05
Air Blank		0.000	00:06
Control Test		0.080	00:06
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

---

Data Download Date/Time: 03/10/2022 04:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CLARK,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/06/2022 01:31:43

---

Diagnostics Check		OK	01:24
Air Blank		0.000	01:25
Control Test		0.080	01:25
Air Blank		0.000	01:26
Subject Sample	# 1	0.231	01:27
Breath Volume		2.613	01:27
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.223	01:30
Breath Volume		2.140	01:30
Air Blank		0.000	01:30
Control Test		0.080	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

---

Data Download Date/Time: 03/10/2022 04:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HIXSON,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: BRONSON

DATE/TIME OF TEST: 03/06/2022 02:42:03

---

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.081	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.072	02:35
Breath Volume	1.937	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.070	02:40
Breath Volume	1.828	02:40
Air Blank	0.000	02:41
Control Test	0.081	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:42

---

Data Download Date/Time: 03/10/2022 04:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: PARRISH,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/06/2022 04:49:13

---

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:46	
Control Test	0.081	04:47	
Air Blank	0.000	04:47	
Subject Sample	# 1	04:47	Subject Test Refused
Breath Volume	0.000	04:47	
Air Blank	0.000	04:48	
Control Test	0.081	04:48	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:49	

---

Data Download Date/Time: 03/10/2022 04:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CASCANTE,H  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:28

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 03/06/2022 06:55:19

---

Diagnostics Check	OK	06:52	
Air Blank	0.000	06:52	
Control Test	0.081	06:52	
Air Blank	0.000	06:53	
Subject Sample	# 1	06:53	Subject Test Refused
Breath Volume	0.000	06:53	
Air Blank	0.000	06:54	
Control Test	0.080	06:54	
Air Blank	0.000	06:55	
Diagnostics Check	OK	06:55	

---

Data Download Date/Time: 03/10/2022 04:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CURRY,D  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:28

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 03/06/2022 09:55:12

---

Diagnostics Check		OK	09:48
Air Blank		0.000	09:49
Control Test		0.080	09:49
Air Blank		0.000	09:49
Subject Sample	# 1	0.184	09:50
Breath Volume		2.257	09:50
Air Blank		0.000	09:51
Air Blank		0.000	09:53
Subject Sample	# 2	0.182	09:53
Breath Volume		2.574	09:53
Air Blank		0.000	09:54
Control Test		0.080	09:54
Air Blank		0.000	09:55
Diagnostics Check		OK	09:55

---

Data Download Date/Time: 03/10/2022 04:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HIXSON,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: PIANTADOSI

DATE/TIME OF TEST: 03/06/2022 22:28:26

---

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:25	
Control Test	0.081	22:26	
Air Blank	0.000	22:26	
Subject Sample	# 1	22:27	Subject Test Refused
Breath Volume	0.000	22:27	
Air Blank	0.000	22:27	
Control Test	0.081	22:27	
Air Blank	0.000	22:28	
Diagnostics Check	OK	22:28	

---

Data Download Date/Time: 03/10/2022 04:29



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: KELLY,G  
ARRESTING OFFICER AGENCY: Neptune Beach PD  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: SENNA

DATE/TIME OF TEST: 03/07/2022 00:10:19

---

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:07	
Control Test	0.082	00:08	
Air Blank	0.000	00:08	
Subject Sample	# 1	00:08	Subject Test Refused
Breath Volume	0.000	00:08	
Air Blank	0.000	00:09	
Control Test	0.081	00:09	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

---

Data Download Date/Time: 03/10/2022 04:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: HERNANDEZ RENDEROS

DATE/TIME OF TEST: 03/07/2022 01:54:52

---

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:52	
Control Test	0.081	01:52	
Air Blank	0.000	01:53	
Subject Sample	# 1	01:53	Subject Test Refused
Breath Volume	0.000	01:53	
Air Blank	0.000	01:53	
Control Test	0.081	01:54	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

---

Data Download Date/Time: 03/10/2022 04:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BORCHARDT,J,P  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: PETERSON,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 03/07/2022 05:18:23

---

Diagnostics Check	OK	05:15	
Air Blank	0.000	05:16	
Control Test	0.081	05:16	
Air Blank	0.000	05:16	
Subject Sample	# 1	05:17	Subject Test Refused
Breath Volume	0.000	05:17	
Air Blank	0.000	05:17	
Control Test	0.081	05:17	
Air Blank	0.000	05:18	
Diagnostics Check	OK	05:18	

---

Data Download Date/Time: 03/10/2022 04:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LOKMIC,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 03/07/2022 23:13:37

---

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.081	23:08
Air Blank	0.000	23:08
Subject Sample # 1	0.177	23:09
Breath Volume	3.843	23:09
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.178	23:12
Breath Volume	3.128	23:12
Air Blank	0.000	23:12
Control Test	0.080	23:13
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

---

Data Download Date/Time: 03/10/2022 04:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HARRELL,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: VAN

DATE/TIME OF TEST: 03/07/2022 23:54:09

---

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:51	
Control Test	0.081	23:52	
Air Blank	0.000	23:52	
Subject Sample	# 1	23:52	Subject Test Refused
Breath Volume	0.000	23:52	
Air Blank	0.000	23:53	
Control Test	0.081	23:53	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:54	

---

Data Download Date/Time: 03/10/2022 04:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: WILLIAMS,M  
ARRESTING OFFICER AGENCY: Jacksonville Beach  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: HUBER

DATE/TIME OF TEST: 03/08/2022 03:15:42

---

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.081	03:10
Air Blank	0.000	03:10
Subject Sample # 1	0.158	03:11
Breath Volume	3.152	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.160	03:14
Breath Volume	2.816	03:14
Air Blank	0.000	03:14
Control Test	0.080	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

---

Data Download Date/Time: 03/10/2022 04:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SIMMONS,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: MORRISSEY

DATE/TIME OF TEST: 03/08/2022 05:54:59

---

Diagnostics Check		OK	05:48
Air Blank		0.000	05:49
Control Test		0.081	05:49
Air Blank		0.000	05:50
Subject Sample	# 1	0.000	05:50
Breath Volume		1.601	05:50
Air Blank		0.000	05:51
Air Blank		0.000	05:53
Subject Sample	# 2	0.000	05:53
Breath Volume		1.402	05:53
Air Blank		0.000	05:54
Control Test		0.080	05:54
Air Blank		0.000	05:54
Diagnostics Check		OK	05:54

---

Data Download Date/Time: 03/10/2022 04:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PURVIS,M,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LAQUIDARA,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 03/08/2022 23:21:27

---

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.081	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.145	23:16
Breath Volume	1.625	23:16
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.154	23:19
Breath Volume	1.507	23:19
Air Blank	0.000	23:20
Control Test	0.080	23:20
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

---

Data Download Date/Time: 03/10/2022 04:30



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: RODRIGUEZ,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 03/11/2022 21:46:52

---

Diagnostics Check	OK	21:44	
Air Blank	0.000	21:44	
Control Test	0.081	21:44	
Air Blank	0.000	21:45	
Subject Sample	# 1	21:45	Subject Test Refused
Breath Volume	0.000	21:45	
Air Blank	0.000	21:45	
Control Test	0.081	21:46	
Air Blank	0.000	21:46	
Diagnostics Check	OK	21:46	

---

Data Download Date/Time: 03/23/2022 12:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CLARK,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: SASSNETT

DATE/TIME OF TEST: 03/12/2022 22:51:37

---

Diagnostics Check	OK	22:42
Air Blank	0.000	22:43
Control Test	0.082	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.149	22:45
Breath Volume	1.128	22:45
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.146	22:49
Breath Volume	1.253	22:49
Air Blank	0.000	22:50
Control Test	0.081	22:50
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

---

Data Download Date/Time: 03/23/2022 12:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HIXSON,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: CAMARILLO ALDACO

DATE/TIME OF TEST: 03/13/2022 01:30:32

---

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.082	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.163	01:25
Breath Volume	1.164	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.166	01:28
Breath Volume	1.339	01:28
Air Blank	0.000	01:29
Control Test	0.081	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

---

Data Download Date/Time: 03/23/2022 12:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: PETERSON,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/13/2022 03:47:06

---

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:44	
Control Test	0.081	03:45	
Air Blank	0.000	03:45	
Subject Sample	# 1	03:45	Subject Test Refused
Breath Volume	0.000	03:45	
Air Blank	0.000	03:46	
Control Test	0.082	03:46	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

---

Data Download Date/Time: 03/23/2022 12:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: SIMMONS,M  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:16

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 03/13/2022 06:45:36

---

Diagnostics Check	OK	06:36
Air Blank	0.000	06:37
Control Test	0.082	06:37
Air Blank	0.000	06:38
Subject Sample # 1	0.139	06:39
Breath Volume	1.433	06:39
Air Blank	0.000	06:40
Air Blank	0.000	06:41
Subject Sample # 2	0.126	06:43
Breath Volume	1.179	06:43
Air Blank	0.000	06:44
Control Test	0.082	06:44
Air Blank	0.000	06:45
Diagnostics Check	OK	06:45

---

Data Download Date/Time: 03/23/2022 12:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSES,L,M  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FRANI,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/13/2022 07:44:49

---

Diagnostics Check		OK	07:38
Air Blank		0.000	07:38
Control Test		0.082	07:38
Air Blank		0.000	07:39
Subject Sample	# 1	0.069	07:40
Breath Volume		1.804	07:40
Air Blank		0.000	07:40
Air Blank		0.000	07:42
Subject Sample	# 2	0.071	07:43
Breath Volume		2.089	07:43
Air Blank		0.000	07:43
Control Test		0.082	07:44
Air Blank		0.000	07:44
Diagnostics Check		OK	07:44

---

Data Download Date/Time: 03/23/2022 12:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LOKMIC,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: MCKEE

DATE/TIME OF TEST: 03/13/2022 23:09:29

---

Diagnostics Check		OK	23:02
Air Blank		0.000	23:03
Control Test		0.081	23:03
Air Blank		0.000	23:04
Subject Sample	# 1	0.286	23:04
Breath Volume		2.156	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.299	23:07
Breath Volume		2.250	23:07
Air Blank		0.000	23:08
Control Test		0.080	23:08
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

---

Data Download Date/Time: 03/23/2022 12:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: HART,A  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 03/15/2022 01:16:39

---

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.081	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.192	01:11
Breath Volume	1.734	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:13
Subject Sample # 2	0.185	01:15
Breath Volume	1.433	01:15
Air Blank	0.000	01:15
Control Test	0.081	01:16
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

---

Data Download Date/Time: 03/23/2022 12:36



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CAPPS,C  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 03/15/2022 05:49:40

---

Diagnostics Check		OK	05:42
Air Blank		0.000	05:43
Control Test		0.081	05:43
Air Blank		0.000	05:44
Subject Sample	# 1	0.113	05:44
Breath Volume		1.859	05:44
Air Blank		0.000	05:45
Air Blank		0.000	05:47
Subject Sample	# 2	0.113	05:48
Breath Volume		2.019	05:48
Air Blank		0.000	05:48
Control Test		0.082	05:49
Air Blank		0.000	05:49
Diagnostics Check		OK	05:49

---

Data Download Date/Time: 03/23/2022 13:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: VALERIO,E  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: HELTON

DATE/TIME OF TEST: 03/15/2022 22:45:19

---

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:42	
Control Test	0.081	22:43	
Air Blank	0.000	22:43	
Subject Sample	# 1	22:43	Subject Test Refused
Breath Volume	0.000	22:43	
Air Blank	0.000	22:44	
Control Test	0.081	22:44	
Air Blank	0.000	22:45	
Diagnostics Check	OK	22:45	

---

Data Download Date/Time: 03/23/2022 13:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: PARRISH,R  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: MARTINEZ PINEDA

DATE/TIME OF TEST: 03/16/2022 02:31:03

---

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.081	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.189	02:26
Breath Volume	1.378	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:28
Subject Sample # 2	0.185	02:29
Breath Volume	1.156	02:29
Air Blank	0.000	02:30
Control Test	0.081	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:31

---

Data Download Date/Time: 03/23/2022 13:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: CLARK,J  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 03/17/2022 05:12:03

---

Diagnostics Check	OK	05:03	
Air Blank	0.000	05:04	
Control Test	0.081	05:04	
Air Blank	0.000	05:04	
Subject Sample	# 1	05:07	No Sample Provided
Breath Volume	0.000	05:07	
Air Blank	0.000	05:08	
Air Blank	0.000	05:10	
Subject Sample	# 2	05:10	Subject Test Refused
Breath Volume	0.000	05:10	
Air Blank	0.000	05:11	
Control Test	0.081	05:11	
Air Blank	0.000	05:11	
Diagnostics Check	OK	05:12	

---

Data Download Date/Time: 03/23/2022 13:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLIDAY,E,C  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: FLANDERS,T  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 05:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: FANN

DATE/TIME OF TEST: 03/20/2022 07:17:15

---

Diagnostics Check		OK	07:10
Air Blank		0.000	07:11
Control Test		0.081	07:11
Air Blank		0.000	07:12
Subject Sample	# 1	0.141	07:12
Breath Volume		2.246	07:12
Air Blank		0.000	07:13
Air Blank		0.000	07:15
Subject Sample	# 2	0.134	07:15
Breath Volume		2.445	07:15
Air Blank		0.000	07:16
Control Test		0.080	07:16
Air Blank		0.000	07:17
Diagnostics Check		OK	07:17

---

Data Download Date/Time: 03/23/2022 13:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TROTIER,PAUL,F  
BREATH TEST OPERATOR AGENCY: Jacksonville SO  
ARRESTING OFFICER: LOKMIC,S  
ARRESTING OFFICER AGENCY: Jacksonville SO  
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 03/23/2022 00:09:33

---

Diagnostics Check		OK	00:02
Air Blank		0.000	00:03
Control Test		0.080	00:03
Air Blank		0.000	00:04
Subject Sample	# 1	0.200	00:04
Breath Volume		1.726	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:07
Subject Sample	# 2	0.194	00:07
Breath Volume		1.914	00:07
Air Blank		0.000	00:08
Control Test		0.081	00:08
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

---

Data Download Date/Time: 03/23/2022 13:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEUBAUER,J,  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: ADAMS,S  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 02/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: GRISHAM

DATE/TIME OF TEST: 03/01/2022 03:50:12

---

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.080	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.190	03:45
Breath Volume	3.292	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.196	03:48
Breath Volume	3.496	03:48
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

---

Data Download Date/Time: 03/28/2022 00:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLARIN,MISTY,D  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 02/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: POCOCK

DATE/TIME OF TEST: 03/03/2022 19:24:28

---

Diagnostics Check	OK	19:18
Air Blank	0.000	19:18
Control Test	0.081	19:19
Air Blank	0.000	19:19
Subject Sample # 1	0.276	19:19
Breath Volume	3.835	19:19
Air Blank	0.000	19:20
Air Blank	0.000	19:22
Subject Sample # 2	0.269	19:22
Breath Volume	2.546	19:22
Air Blank	0.000	19:23
Control Test	0.080	19:23
Air Blank	0.000	19:24
Diagnostics Check	OK	19:24

---

Data Download Date/Time: 03/28/2022 00:49



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VAN DYKE,T,M  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 02/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/04/2022 02:39:11

---

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:31	
Control Test	0.080	02:31	
Air Blank	0.000	02:31	
Subject Sample	# 1	02:35	No Sample Provided
Breath Volume	0.000	02:35	
Air Blank	0.000	02:35	
Air Blank	0.000	02:37	
Subject Sample	# 2	02:37	Subject Test Refused
Breath Volume	0.000	02:37	
Air Blank	0.000	02:38	
Control Test	0.080	02:38	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:39	

---

Data Download Date/Time: 03/28/2022 00:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEUBAUER,J,  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: ADAMS,S  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 02/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: LAMOT

DATE/TIME OF TEST: 03/05/2022 02:52:52

---

Diagnostics Check		OK	02:45
Air Blank		0.000	02:46
Control Test		0.079	02:46
Air Blank		0.000	02:46
Subject Sample	# 1	0.243	02:47
Breath Volume		1.320	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.248	02:51
Breath Volume		2.117	02:51
Air Blank		0.000	02:51
Control Test		0.079	02:52
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

---

Data Download Date/Time: 03/28/2022 00:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEUBAUER,J,  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: GELSOMINO,M  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 02/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: HARRELL STEWART

DATE/TIME OF TEST: 03/06/2022 01:31:24

---

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:23	
Control Test	0.079	01:23	
Air Blank	0.000	01:24	
Subject Sample	# 1	01:27	No Sample Provided
Breath Volume	0.000	01:27	
Air Blank	0.000	01:27	
Air Blank	0.000	01:29	
Subject Sample	# 2	01:30	Subject Test Refused
Breath Volume	0.000	01:30	
Air Blank	0.000	01:30	
Control Test	0.080	01:30	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:31	

---

Data Download Date/Time: 03/28/2022 00:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLARIN,MISTY,D  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 02/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: JOPLIN

DATE/TIME OF TEST: 03/08/2022 01:34:30

---

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:26	
Control Test	0.079	01:26	
Air Blank	0.000	01:27	
Subject Sample	# 1	01:30	No Sample Provided
Breath Volume	0.000	01:30	
Air Blank	0.000	01:30	
Air Blank	0.000	01:32	
Subject Sample	# 2	01:33	Subject Test Refused
Breath Volume	0.000	01:33	
Air Blank	0.000	01:33	
Control Test	0.079	01:33	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

---

Data Download Date/Time: 03/28/2022 00:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLARIN,MISTY,D  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 02/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 03/13/2022 00:49:05

---

Diagnostics Check		OK	00:41
Air Blank		0.000	00:42
Control Test		0.080	00:42
Air Blank		0.000	00:43
Subject Sample	# 1	0.204	00:44
Breath Volume		2.742	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:46
Subject Sample	# 2	0.203	00:47
Breath Volume		2.410	00:47
Air Blank		0.000	00:48
Control Test		0.079	00:48
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

---

Data Download Date/Time: 03/28/2022 00:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VAN DYKE,T,M  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 02/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 03/14/2022 02:44:59

---

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.080	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.175	02:40
Breath Volume	1.371	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.166	02:43
Breath Volume	2.125	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

---

Data Download Date/Time: 03/28/2022 00:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VAN DYKE,T,M  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 02/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: PRIDGEN JR

DATE/TIME OF TEST: 03/18/2022 02:46:24

---

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:38	
Control Test	0.080	02:38	
Air Blank	0.000	02:39	
Subject Sample	# 1	02:42	No Sample Provided
Breath Volume	0.000	02:42	
Air Blank	0.000	02:42	
Air Blank	0.000	02:44	
Subject Sample	# 2	02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:45	
Control Test	0.080	02:45	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

---

Data Download Date/Time: 03/28/2022 00:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAERTZ,MARK,J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: ADAMS,S  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 02/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: HONTZ

DATE/TIME OF TEST: 03/18/2022 21:46:27

---

Diagnostics Check	OK	21:38
Air Blank	0.000	21:39
Control Test	0.080	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.225	21:41
Breath Volume	2.589	21:41
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.221	21:44
Breath Volume	2.445	21:44
Air Blank	0.000	21:45
Control Test	0.080	21:45
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

---

Data Download Date/Time: 03/28/2022 00:50



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEUBAUER,J,  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: ADAMS,S  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 02/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: DOWNING

DATE/TIME OF TEST: 03/19/2022 04:31:39

---

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:23	
Control Test	0.080	04:23	
Air Blank	0.000	04:24	
Subject Sample	# 1	04:27	No Sample Provided
Breath Volume	0.000	04:27	
Air Blank	0.000	04:28	
Air Blank	0.000	04:30	
Subject Sample	# 2	04:30	Subject Test Refused
Breath Volume	0.000	04:30	
Air Blank	0.000	04:30	
Control Test	0.080	04:31	
Air Blank	0.000	04:31	
Diagnostics Check	OK	04:31	

---

Data Download Date/Time: 03/28/2022 00:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAERTZ,MARK,J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: WILLIAMSON,M  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 02/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/19/2022 04:52:50

---

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.079	04:46
Air Blank	0.000	04:46
Subject Sample # 1	0.131	04:47
Breath Volume	1.648	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.136	04:51
Breath Volume	2.601	04:51
Air Blank	0.000	04:51
Control Test	0.080	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

---

Data Download Date/Time: 03/28/2022 00:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NEUBAUER,J,  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: ADAMS,S  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 02/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 19:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: BARNES JR

DATE/TIME OF TEST: 03/19/2022 20:43:00

---

Diagnostics Check	OK	20:35
Air Blank	0.000	20:36
Control Test	0.079	20:36
Air Blank	0.000	20:37
Subject Sample # 1	0.336	20:37
Breath Volume	1.804	20:37
Air Blank	0.000	20:38
Air Blank	0.000	20:40
Subject Sample # 2	0.334	20:41
Breath Volume	2.078	20:41
Air Blank	0.000	20:42
Control Test	0.078	20:42
Air Blank	0.000	20:42
Diagnostics Check	OK	20:43

---

Data Download Date/Time: 03/28/2022 00:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAERTZ,MARK,J  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: GELSOMINO,M  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 02/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: POLLARD

DATE/TIME OF TEST: 03/19/2022 21:13:17

---

Diagnostics Check	OK	20:59
Air Blank	0.000	21:00
Control Test	0.079	21:00
Air Blank	0.000	21:01
Subject Sample # 1	0.187	21:03
Breath Volume	1.257	21:03
Air Blank	0.000	21:04
Air Blank	0.000	21:06
Subject Sample # 2	0.228	21:08
Breath Volume	3.472	21:08
Air Blank	0.000	21:09
Air Blank	0.000	21:11
Subject Sample # 3	0.196	21:11
Breath Volume	1.136	21:11
Air Blank	0.000	21:12
Control Test	0.078	21:12
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

---

Data Download Date/Time: 03/28/2022 00:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLARIN,MISTY,D  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: FLYNN,H  
ARRESTING OFFICER AGENCY: Green Cove Springs  
LAST AGENCY INSPECTION: 02/27/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: MAFFUCCI

DATE/TIME OF TEST: 03/27/2022 22:40:43

---

Diagnostics Check	OK	22:34
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:35
Subject Sample # 1	0.219	22:36
Breath Volume	3.597	22:36
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.217	22:39
Breath Volume	2.796	22:39
Air Blank	0.000	22:39
Control Test	0.081	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

---

Data Download Date/Time: 03/28/2022 00:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VILLARIN,MISTY,D  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 03/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CORLETO

DATE/TIME OF TEST: 03/31/2022 03:23:58

---

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:15	
Control Test	0.080	03:15	
Air Blank	0.000	03:16	
Subject Sample	# 1	03:19	No Sample Provided
Breath Volume	0.000	03:19	
Air Blank	0.000	03:19	
Air Blank	0.000	03:21	
Subject Sample	# 2	03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:23	
Control Test	0.080	03:23	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

---

Data Download Date/Time: 04/26/2022 18:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VAN DYKE,T,M  
BREATH TEST OPERATOR AGENCY: Clay County SO  
ARRESTING OFFICER: RILEY,R  
ARRESTING OFFICER AGENCY: Clay County SO  
LAST AGENCY INSPECTION: 03/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22620080A2  
DRY GAS EXPIRY DATE: 10/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: INSCOE

DATE/TIME OF TEST: 03/31/2022 23:08:43

---

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.080	23:03
Air Blank	0.000	23:03
Subject Sample # 1	0.000	23:04
Breath Volume	1.320	23:04
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.000	23:07
Breath Volume	2.289	23:07
Air Blank	0.000	23:07
Control Test	0.080	23:08
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

---

Data Download Date/Time: 04/26/2022 18:31

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STARLING,KELLEY,L  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: STARLING,KELLEY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11821080A2  
DRY GAS EXPIRY DATE: 06/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MEYER

DATE/TIME OF TEST: 03/12/2022 03:47:50

---

Diagnostics Check	OK	03:36		
Air Blank	0.000	03:36		
Control Test	0.080	03:36		
Air Blank	0.000	03:37		
Subject Sample	# 1	03:40	No Sample Provided	
Breath Volume	0.000	03:40		
Air Blank	0.000	03:40		
Air Blank	0.000	03:42		
Subject Sample	# 2	0.190	03:46	Volume Not Met
Breath Volume	0.355	03:46		
Air Blank	0.000	03:46		
Control Test	0.080	03:47		
Air Blank	0.000	03:47		
Diagnostics Check	OK	03:47		

---

Data Download Date/Time: 04/01/2022 12:39



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WELLS,AUSTIN,D  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: WELLS,AUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11821080A2  
DRY GAS EXPIRY DATE: 06/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: GOMEZ GALLEGOS

DATE/TIME OF TEST: 03/12/2022 04:25:30

---

Diagnostics Check		OK	04:18
Air Blank		0.000	04:18
Control Test		0.080	04:18
Air Blank		0.000	04:19
Subject Sample	# 1	0.128	04:19
Breath Volume		1.488	04:19
Air Blank		0.000	04:20
Air Blank		0.000	04:22
Subject Sample	# 2	0.131	04:23
Breath Volume		1.707	04:23
Air Blank		0.000	04:24
Control Test		0.079	04:24
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

---

Data Download Date/Time: 04/01/2022 12:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAUGHERTY,D,M  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: CALLAHAN,C  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 03/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11821080A2  
DRY GAS EXPIRY DATE: 06/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 08:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:44

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 03/19/2022 10:11:48

---

Diagnostics Check	OK	10:04
Air Blank	0.000	10:05
Control Test	0.080	10:05
Air Blank	0.000	10:06
Subject Sample # 1	0.172	10:07
Breath Volume	1.343	10:07
Air Blank	0.000	10:07
Air Blank	0.000	10:09
Subject Sample # 2	0.186	10:10
Breath Volume	1.339	10:10
Air Blank	0.000	10:10
Control Test	0.079	10:11
Air Blank	0.000	10:11
Diagnostics Check	OK	10:11

---

Data Download Date/Time: 04/01/2022 12:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WELLS,AUSTIN,D  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: WELLS,AUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11821080A2  
DRY GAS EXPIRY DATE: 06/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/22/2022 23:05:59

---

Diagnostics Check	OK	23:05	
Air Blank		23:05	Sequence Aborted
Air Blank	0.000	23:05	

---

Data Download Date/Time: 04/01/2022 12:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WELLS,AUSTIN,D  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: WELLS,AUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11821080A2  
DRY GAS EXPIRY DATE: 06/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/22/2022 23:34:34

---

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.078	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.139	23:29
Breath Volume	1.121	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.145	23:33
Breath Volume	1.914	23:33
Air Blank	0.000	23:33
Control Test	0.078	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

---

Data Download Date/Time: 04/01/2022 12:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,ALEXANDER,M  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: RODRIGUEZ,ALEXANDER  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11821080A2  
DRY GAS EXPIRY DATE: 06/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: CASANOVA

DATE/TIME OF TEST: 03/25/2022 00:19:02

---

Diagnostics Check	OK	00:15	
Air Blank	0.000	00:16	
Control Test	0.080	00:16	
Air Blank	0.000	00:17	
Subject Sample	# 1	00:17	Subject Test Refused
Breath Volume	0.000	00:17	
Air Blank	0.000	00:18	
Control Test	0.080	00:18	
Air Blank	0.000	00:18	
Diagnostics Check	OK	00:19	

---

Data Download Date/Time: 04/01/2022 12:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,ALEXANDER,M  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: RODRIGUEZ,ALEXANDER  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11821080A2  
DRY GAS EXPIRY DATE: 06/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 16:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: WHIPPLE

DATE/TIME OF TEST: 03/28/2022 18:18:34

---

Diagnostics Check	OK	18:12
Air Blank	0.000	18:12
Control Test	0.081	18:13
Air Blank	0.000	18:13
Subject Sample # 1	0.000	18:14
Breath Volume	1.562	18:14
Air Blank	0.000	18:14
Air Blank	0.000	18:16
Subject Sample # 2	0.000	18:17
Breath Volume	1.269	18:17
Air Blank	0.000	18:17
Control Test	0.080	18:17
Air Blank	0.000	18:18
Diagnostics Check	OK	18:18

---

Data Download Date/Time: 04/01/2022 12:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,ERIC,R  
BREATH TEST OPERATOR AGENCY: Coral Springs PD  
ARRESTING OFFICER: PEREZ,ERIC  
ARRESTING OFFICER AGENCY: Coral Springs PD  
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: HAAS

DATE/TIME OF TEST: 03/02/2022 01:53:01

---

Diagnostics Check	OK	01:46
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:48
Subject Sample # 1	0.229	01:48
Breath Volume	2.113	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:50
Subject Sample # 2	0.234	01:51
Breath Volume	2.535	01:51
Air Blank	0.000	01:52
Control Test	0.079	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:53

---

Data Download Date/Time: 04/01/2022 13:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,ERIC,R  
BREATH TEST OPERATOR AGENCY: Coral Springs PD  
ARRESTING OFFICER: PRATT,RICHARD  
ARRESTING OFFICER AGENCY: Coral Springs PD  
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: FAUST

DATE/TIME OF TEST: 03/05/2022 01:50:34

---

Diagnostics Check	OK	01:49	
Air Blank		01:50	RFI Detect
Air Blank	0.000	01:50	

---

Data Download Date/Time: 04/01/2022 13:46



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,ERIC,R  
BREATH TEST OPERATOR AGENCY: Coral Springs PD  
ARRESTING OFFICER: PRATT,RICHARD  
ARRESTING OFFICER AGENCY: Coral Springs PD  
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: FAUST

DATE/TIME OF TEST: 03/05/2022 02:00:17

---

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.081	01:53
Air Blank	0.000	01:54
Subject Sample # 1	0.132	01:54
Breath Volume	1.578	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.124	01:58
Breath Volume	2.566	01:58
Air Blank	0.000	01:59
Control Test	0.081	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

---

Data Download Date/Time: 04/01/2022 13:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,ERIC,R  
BREATH TEST OPERATOR AGENCY: Coral Springs PD  
ARRESTING OFFICER: PEREZ,ERIC  
ARRESTING OFFICER AGENCY: Coral Springs PD  
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PARISIEN

DATE/TIME OF TEST: 03/19/2022 02:48:57

---

Diagnostics Check		OK	02:41
Air Blank		0.000	02:42
Control Test		0.080	02:42
Air Blank		0.000	02:42
Subject Sample	# 1	0.207	02:44
Breath Volume		1.730	02:44
Air Blank		0.000	02:45
Air Blank		0.000	02:46
Subject Sample	# 2	0.194	02:47
Breath Volume		2.531	02:47
Air Blank		0.000	02:48
Control Test		0.078	02:48
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

---

Data Download Date/Time: 04/01/2022 13:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMAS,ALEXANDER,J  
BREATH TEST OPERATOR AGENCY: Coral Springs PD  
ARRESTING OFFICER: HINTZE,AIDAN  
ARRESTING OFFICER AGENCY: Coral Springs PD  
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06021080A2  
DRY GAS EXPIRY DATE: 04/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: ALAY

DATE/TIME OF TEST: 03/31/2022 06:16:51

---

Diagnostics Check		OK	06:10
Air Blank		0.000	06:10
Control Test		0.081	06:11
Air Blank		0.000	06:11
Subject Sample	# 1	0.166	06:12
Breath Volume		2.812	06:12
Air Blank		0.000	06:12
Air Blank		0.000	06:14
Subject Sample	# 2	0.160	06:15
Breath Volume		1.671	06:15
Air Blank		0.000	06:15
Control Test		0.080	06:16
Air Blank		0.000	06:16
Diagnostics Check		OK	06:16

---

Data Download Date/Time: 04/01/2022 13:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAYNE, ABIGAIL, D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BLACKMAN, GERIG  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: LEFRANC

DATE/TIME OF TEST: 03/02/2022 02:45:37

---

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.000	02:41
Breath Volume	2.140	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.000	02:44
Breath Volume	2.441	02:44
Air Blank	0.000	02:44
Control Test	0.081	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

---

Data Download Date/Time: 03/22/2022 08:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELCHER,ALEX,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BLACKMAN,GERIG  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: MATIAS

DATE/TIME OF TEST: 03/06/2022 00:30:53

---

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.080	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.146	00:25
Breath Volume	1.597	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.144	00:29
Breath Volume	2.480	00:29
Air Blank	0.000	00:29
Control Test	0.082	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

---

Data Download Date/Time: 03/22/2022 08:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELCHER,ALEX,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BLACKMAN,GERIG  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 06:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:24

SUBJECT'S LAST NAME: KENEMUTH

DATE/TIME OF TEST: 03/06/2022 06:56:02

---

Diagnostics Check	OK	06:49
Air Blank	0.000	06:50
Control Test	0.081	06:50
Air Blank	0.000	06:51
Subject Sample # 1	0.127	06:51
Breath Volume	3.132	06:51
Air Blank	0.000	06:52
Air Blank	0.000	06:54
Subject Sample # 2	0.130	06:54
Breath Volume	2.664	06:54
Air Blank	0.000	06:55
Control Test	0.080	06:55
Air Blank	0.000	06:55
Diagnostics Check	OK	06:56

---

Data Download Date/Time: 03/22/2022 08:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,ALBERT,J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WILLIAMS,ALBERT  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: REBERK

DATE/TIME OF TEST: 03/06/2022 19:07:28

---

Diagnostics Check	OK	18:58
Air Blank	0.000	18:58
Control Test	0.081	18:59
Air Blank	0.000	18:59
Subject Sample # 1	0.168	19:01
Breath Volume	2.074	19:01
Air Blank	0.000	19:02
Air Blank	0.000	19:04
Subject Sample # 2	0.163	19:05
Breath Volume	1.656	19:05
Air Blank	0.000	19:06
Control Test	0.080	19:06
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

---

Data Download Date/Time: 03/22/2022 08:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELCHER,ALEX,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BELCHER,ALEX  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: MONTES

DATE/TIME OF TEST: 03/07/2022 04:12:53

---

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:11	
Control Test	0.080	04:12	
Air Blank		04:12	Sequence Aborted
Air Blank	0.000	04:12	

---

Data Download Date/Time: 03/22/2022 08:10



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELCHER,ALEX,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BELCHER,ALEX  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: MONTES

DATE/TIME OF TEST: 03/07/2022 05:03:11

---

Diagnostics Check	OK	04:56
Air Blank	0.000	04:57
Control Test	0.080	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.100	04:58
Breath Volume	2.546	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.094	05:01
Breath Volume	2.191	05:01
Air Blank	0.000	05:02
Control Test	0.078	05:02
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

---

Data Download Date/Time: 03/22/2022 08:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COAKE,SPENCER,M  
BREATH TEST OPERATOR AGENCY: Winter Haven PD  
ARRESTING OFFICER: COAKE,SPENCER  
ARRESTING OFFICER AGENCY: Winter Haven PD  
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 06:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: HOLTZCLAW

DATE/TIME OF TEST: 03/06/2022 06:44:32

---

Diagnostics Check	OK	06:37
Air Blank	0.000	06:38
Control Test	0.081	06:38
Air Blank	0.000	06:39
Subject Sample # 1	0.195	06:39
Breath Volume	4.152	06:39
Air Blank	0.000	06:40
Air Blank	0.000	06:42
Subject Sample # 2	0.196	06:42
Breath Volume	4.742	06:42
Air Blank	0.000	06:43
Control Test	0.080	06:43
Air Blank	0.000	06:44
Diagnostics Check	OK	06:44

---

Data Download Date/Time: 03/24/2022 14:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HEBSON,BRET,  
BREATH TEST OPERATOR AGENCY: Winter Haven PD  
ARRESTING OFFICER: HEBSON,BRET  
ARRESTING OFFICER AGENCY: Winter Haven PD  
LAST AGENCY INSPECTION: 03/24/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: COURT ORD  
ARREST TIME: 15:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: CARRENO ANDRADE

DATE/TIME OF TEST: 03/24/2022 15:34:23

---

Diagnostics Check	OK	15:27
Air Blank	0.000	15:27
Control Test	0.081	15:28
Air Blank	0.000	15:28
Subject Sample # 1	0.000	15:29
Breath Volume	1.171	15:29
Air Blank	0.000	15:29
Air Blank	0.000	15:31
Subject Sample # 2	0.000	15:32
Breath Volume	1.570	15:32
Air Blank	0.000	15:33
Control Test	0.080	15:33
Air Blank	0.000	15:34
Diagnostics Check	OK	15:34

---

Data Download Date/Time: 04/29/2022 15:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DERIET,RAFAEL,  
BREATH TEST OPERATOR AGENCY: Hollywood PD  
ARRESTING OFFICER: DULCIO,PIERRE  
ARRESTING OFFICER AGENCY: Hollywood PD  
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1273831  
DRY GAS EXPIRY DATE: 01/31/2023  
VIOLATION CODE: DUI  
ARREST TIME: 05:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: GAUNTLETT

DATE/TIME OF TEST: 03/24/2022 06:14:25

---

Diagnostics Check	OK	06:08
Air Blank	0.000	06:08
Control Test	0.081	06:08
Air Blank	0.000	06:09
Subject Sample # 1	0.000	06:10
Breath Volume	1.402	06:10
Air Blank	0.000	06:10
Air Blank	0.000	06:12
Subject Sample # 2	0.000	06:13
Breath Volume	1.769	06:13
Air Blank	0.000	06:13
Control Test	0.080	06:13
Air Blank	0.000	06:14
Diagnostics Check	OK	06:14

---

Data Download Date/Time: 04/13/2022 11:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAZIME,CPL. A.,  
BREATH TEST OPERATOR AGENCY: New Port Richey PD  
ARRESTING OFFICER: CROWELL,FTO. D.  
ARRESTING OFFICER AGENCY: New Port Richey PD  
LAST AGENCY INSPECTION: 02/06/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A3  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: T420696618040

DATE/TIME OF TEST: 03/05/2022 18:35:41

---

Diagnostics Check	OK	18:28
Air Blank	0.000	18:29
Control Test	0.078	18:29
Air Blank	0.000	18:30
Subject Sample # 1	0.095	18:30
Breath Volume	1.542	18:30
Air Blank	0.000	18:31
Air Blank	0.000	18:33
Subject Sample # 2	0.095	18:34
Breath Volume	1.691	18:34
Air Blank	0.000	18:34
Control Test	0.078	18:35
Air Blank	0.000	18:35
Diagnostics Check	OK	18:35

---

Data Download Date/Time: 03/06/2022 10:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARAMIELLO,CA,A  
BREATH TEST OPERATOR AGENCY: Port Richey PD  
ARRESTING OFFICER: CARAMIELLO,CA  
ARRESTING OFFICER AGENCY: Port Richey PD  
LAST AGENCY INSPECTION: 03/06/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A3  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: VILLAFANE

DATE/TIME OF TEST: 03/16/2022 22:33:44

---

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.078	22:28
Air Blank	0.000	22:29
Subject Sample # 1	0.000	22:29
Breath Volume	2.949	22:29
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.000	22:32
Breath Volume	2.296	22:32
Air Blank	0.000	22:32
Control Test	0.079	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

---

Data Download Date/Time: 04/03/2022 09:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PENNELL,RUSSELL,M  
BREATH TEST OPERATOR AGENCY: New Port Richey PD  
ARRESTING OFFICER: MARTARANA,J  
ARRESTING OFFICER AGENCY: New Port Richey PD  
LAST AGENCY INSPECTION: 03/06/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A3  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: ANNAS

DATE/TIME OF TEST: 03/17/2022 02:00:07

---

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.078	01:54
Air Blank	0.000	01:54
Subject Sample # 1	0.232	01:55
Breath Volume	4.843	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.226	01:58
Breath Volume	4.660	01:58
Air Blank	0.000	01:59
Control Test	0.077	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

---

Data Download Date/Time: 04/03/2022 09:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMMELMAN, TYLER, R  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: BERRIOS, JOSE  
ARRESTING OFFICER AGENCY: New Port Richey PD  
LAST AGENCY INSPECTION: 03/06/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A3  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 03/19/2022 03:44:09

---

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.078	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.149	03:39
Breath Volume	2.925	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:41
Subject Sample # 2	0.149	03:42
Breath Volume	3.085	03:42
Air Blank	0.000	03:43
Control Test	0.077	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

---

Data Download Date/Time: 04/03/2022 09:27



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCONNELL,JEFFREY,W  
BREATH TEST OPERATOR AGENCY: Port Richey PD  
ARRESTING OFFICER: MCCONNELL,JEFFREY  
ARRESTING OFFICER AGENCY: Port Richey PD  
LAST AGENCY INSPECTION: 03/06/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A3  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: ASHLEY

DATE/TIME OF TEST: 03/21/2022 23:35:37

---

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:25	
Control Test	0.079	23:26	
Air Blank	0.000	23:26	
Subject Sample	# 1 0.234	23:29	Volume Not Met
Breath Volume	0.746	23:29	
Air Blank	0.000	23:30	
Air Blank	0.000	23:32	
Subject Sample	# 2	23:34	Subject Test Refused
Breath Volume	0.480	23:34	
Air Blank	0.000	23:34	
Control Test	0.077	23:35	
Air Blank	0.000	23:35	
Diagnostics Check	OK	23:35	

---

Data Download Date/Time: 04/03/2022 09:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PENNELL,RUSSELL,M  
BREATH TEST OPERATOR AGENCY: New Port Richey PD  
ARRESTING OFFICER: BERRIOS,JOSE  
ARRESTING OFFICER AGENCY: New Port Richey PD  
LAST AGENCY INSPECTION: 03/06/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A3  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: BECKMAN

DATE/TIME OF TEST: 03/26/2022 21:56:34

---

Diagnostics Check	OK	21:50
Air Blank	0.000	21:51
Control Test	0.079	21:51
Air Blank	0.000	21:51
Subject Sample # 1	0.180	21:52
Breath Volume	2.574	21:52
Air Blank	0.000	21:52
Air Blank	0.000	21:54
Subject Sample # 2	0.174	21:55
Breath Volume	1.652	21:55
Air Blank	0.000	21:55
Control Test	0.079	21:55
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

---

Data Download Date/Time: 04/03/2022 09:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCONNELL,JEFFREY,W  
BREATH TEST OPERATOR AGENCY: Port Richey PD  
ARRESTING OFFICER: MCCONNELL,JEFFREY  
ARRESTING OFFICER AGENCY: Port Richey PD  
LAST AGENCY INSPECTION: 03/06/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A3  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: DANYLUK

DATE/TIME OF TEST: 03/27/2022 03:04:49

---

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.079	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.160	03:00
Breath Volume	2.515	03:00
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.145	03:03
Breath Volume	2.964	03:03
Air Blank	0.000	03:03
Control Test	0.078	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

---

Data Download Date/Time: 04/03/2022 09:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PENNELL,RUSSELL,M  
BREATH TEST OPERATOR AGENCY: New Port Richey PD  
ARRESTING OFFICER: IVEY,B  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 03/06/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09820080A3  
DRY GAS EXPIRY DATE: 06/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: GARCELL

DATE/TIME OF TEST: 03/27/2022 03:41:35

---

Diagnostics Check		OK	03:31	
Air Blank		0.000	03:32	
Control Test		0.078	03:32	
Air Blank		0.000	03:32	
Subject Sample	# 1	0.142	03:36	Volume Not Met
Breath Volume		1.632	03:36	
Air Blank		0.000	03:36	
Air Blank		0.000	03:38	
Subject Sample	# 2		03:40	Subject Test Refused
Breath Volume		0.296	03:40	
Air Blank		0.000	03:40	
Control Test		0.076	03:40	
Air Blank		0.000	03:41	
Diagnostics Check		OK	03:41	

---

Data Download Date/Time: 04/03/2022 09:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GASSEN,STEVE,  
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD  
ARRESTING OFFICER: GASSEN,S  
ARRESTING OFFICER AGENCY: Tarpon Springs PD  
LAST AGENCY INSPECTION: 03/03/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00521080A2  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 14:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:17

SUBJECT'S LAST NAME: NAUMANN

DATE/TIME OF TEST: 03/14/2022 15:43:35

---

Diagnostics Check	OK	15:37
Air Blank	0.000	15:38
Control Test	0.080	15:38
Air Blank	0.000	15:38
Subject Sample # 1	0.000	15:39
Breath Volume	2.570	15:39
Air Blank	0.000	15:39
Air Blank	0.000	15:41
Subject Sample # 2	0.000	15:42
Breath Volume	2.484	15:42
Air Blank	0.000	15:42
Control Test	0.080	15:43
Air Blank	0.000	15:43
Diagnostics Check	OK	15:43

---

Data Download Date/Time: 05/02/2022 13:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAZIME,S,  
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD  
ARRESTING OFFICER: HAZIME,S  
ARRESTING OFFICER AGENCY: Tarpon Springs PD  
LAST AGENCY INSPECTION: 03/03/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00521080A2  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: SGROI

DATE/TIME OF TEST: 03/19/2022 02:30:43

---

Diagnostics Check		OK	02:23
Air Blank		0.000	02:24
Control Test		0.081	02:24
Air Blank		0.000	02:25
Subject Sample	# 1	0.287	02:26
Breath Volume		1.238	02:26
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.284	02:29
Breath Volume		1.652	02:29
Air Blank		0.000	02:29
Control Test		0.079	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

---

Data Download Date/Time: 05/02/2022 13:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOOD,NICHOLAS,W  
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD  
ARRESTING OFFICER: VANSUMEREN,LOGAN  
ARRESTING OFFICER AGENCY: Tarpon Springs PD  
LAST AGENCY INSPECTION: 03/03/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00521080A2  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/21/2022 20:44:34

---

Diagnostics Check	OK	20:38
Air Blank	0.000	20:38
Control Test	0.080	20:38
Air Blank	0.000	20:39
Subject Sample # 1	0.276	20:39
Breath Volume	2.546	20:39
Air Blank	0.000	20:40
Air Blank	0.000	20:42
Subject Sample # 2	0.262	20:43
Breath Volume	1.347	20:43
Air Blank	0.000	20:43
Control Test	0.079	20:43
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

---

Data Download Date/Time: 05/02/2022 13:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAZIME,S,  
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD  
ARRESTING OFFICER: HAZIME,S  
ARRESTING OFFICER AGENCY: Tarpon Springs PD  
LAST AGENCY INSPECTION: 03/03/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 00521080A2  
DRY GAS EXPIRY DATE: 02/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 12:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: RIOLO

DATE/TIME OF TEST: 03/30/2022 02:37:44

---

Diagnostics Check		OK	02:30
Air Blank		0.000	02:30
Control Test		0.080	02:30
Air Blank		0.000	02:31
Subject Sample	# 1	0.223	02:32
Breath Volume		2.535	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:34
Subject Sample	# 2	0.210	02:36
Breath Volume		1.851	02:36
Air Blank		0.000	02:36
Control Test		0.079	02:37
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

---

Data Download Date/Time: 05/02/2022 13:40



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHRISTENSEN,S,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: THOMAS,K  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 02/03/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 03/03/2022 01:21:32

---

Diagnostics Check		OK	01:13
Air Blank		0.000	01:14
Control Test		0.081	01:14
Air Blank		0.000	01:15
Subject Sample	# 1	0.026	01:17
Breath Volume		1.320	01:17
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.025	01:20
Breath Volume		1.464	01:20
Air Blank		0.000	01:20
Control Test		0.080	01:20
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

---

Data Download Date/Time: 03/08/2022 15:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WRIGHT,NAOMIE  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: SAIA,MICHAEL  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 02/03/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 10:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: MALKE

DATE/TIME OF TEST: 03/03/2022 11:26:36

---

Diagnostics Check	OK	11:20
Air Blank	0.000	11:20
Control Test	0.081	11:21
Air Blank	0.000	11:21
Subject Sample # 1	0.000	11:22
Breath Volume	3.535	11:22
Air Blank	0.000	11:22
Air Blank	0.000	11:24
Subject Sample # 2	0.000	11:25
Breath Volume	3.582	11:25
Air Blank	0.000	11:25
Control Test	0.081	11:26
Air Blank	0.000	11:26
Diagnostics Check	OK	11:26

---

Data Download Date/Time: 03/08/2022 15:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GERRY,TABATHA,B  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: LAMOUR,ZACKARY  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 02/03/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 03/03/2022 23:27:28

---

Diagnostics Check	OK	23:24	
Air Blank	0.000	23:25	
Control Test	0.080	23:25	
Air Blank	0.000	23:25	
Subject Sample	# 1	23:26	Subject Test Refused
Breath Volume	0.000	23:26	
Air Blank	0.000	23:26	
Control Test	0.080	23:26	
Air Blank	0.000	23:27	
Diagnostics Check	OK	23:27	

---

Data Download Date/Time: 03/08/2022 15:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHRISTENSEN,S,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: GODERICH,K  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 02/03/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: DACCHILLE

DATE/TIME OF TEST: 03/04/2022 03:29:23

---

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.081	03:18	
Air Blank	0.000	03:19	
Subject Sample	# 1	03:22	No Sample Provided
Breath Volume	0.000	03:22	
Air Blank	0.000	03:22	
Air Blank	0.000	03:24	
Subject Sample	# 2	03:28	No Sample Provided
Breath Volume	0.000	03:28	
Air Blank	0.000	03:28	
Control Test	0.080	03:28	
Air Blank	0.000	03:29	
Diagnostics Check	OK	03:29	

---

Data Download Date/Time: 03/08/2022 15:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,EMILY,K  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: SISTI,JORDAN  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 02/03/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24620080A2  
DRY GAS EXPIRY DATE: 11/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/05/2022 01:49:24

---

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:47	
Control Test	0.080	01:47	
Air Blank	0.000	01:47	
Subject Sample	# 1	01:48	Subject Test Refused
Breath Volume	0.000	01:48	
Air Blank	0.000	01:48	
Control Test	0.080	01:48	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

---

Data Download Date/Time: 03/08/2022 15:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 07220080A2

DRY GAS EXPIRY DATE: 05/05/2022

VIOLATION CODE: DUI

ARREST TIME: 00:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/04/2022 01:17:42

---

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:15	
Control Test	0.081	01:15	
Air Blank	0.000	01:16	
Subject Sample	# 1	01:16	Subject Test Refused
Breath Volume	0.000	01:16	
Air Blank	0.000	01:16	
Control Test	0.080	01:17	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

---

Data Download Date/Time: 03/21/2022 16:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSIER, MATTHEW, D  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: HICKS, JASON  
ARRESTING OFFICER AGENCY: Indian Shores PD  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MEMAX

DATE/TIME OF TEST: 03/04/2022 02:06:49

---

Diagnostics Check		OK	02:00
Air Blank		0.000	02:00
Control Test		0.081	02:01
Air Blank		0.000	02:01
Subject Sample	# 1	0.173	02:02
Breath Volume		2.617	02:02
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.168	02:05
Breath Volume		2.445	02:05
Air Blank		0.000	02:05
Control Test		0.081	02:06
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

---

Data Download Date/Time: 03/21/2022 16:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSIER, MATTHEW, D  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: GOSIER, MATTHEW  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: BANGER

DATE/TIME OF TEST: 03/04/2022 05:21:40

---

Diagnostics Check	OK	05:15
Air Blank	0.000	05:15
Control Test	0.081	05:16
Air Blank	0.000	05:16
Subject Sample # 1	0.129	05:17
Breath Volume	2.417	05:17
Air Blank	0.000	05:17
Air Blank	0.000	05:19
Subject Sample # 2	0.124	05:20
Breath Volume	2.269	05:20
Air Blank	0.000	05:20
Control Test	0.081	05:21
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

---

Data Download Date/Time: 03/21/2022 16:08



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 07220080A2

DRY GAS EXPIRY DATE: 05/05/2022

VIOLATION CODE: DUI

ARREST TIME: 21:24

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: FILBURN

DATE/TIME OF TEST: 03/04/2022 22:12:31

---

Diagnostics Check	OK	22:09	
Air Blank	0.000	22:10	
Control Test	0.082	22:10	
Air Blank	0.000	22:10	
Subject Sample	# 1	22:11	Subject Test Refused
Breath Volume	0.000	22:11	
Air Blank	0.000	22:11	
Control Test	0.081	22:11	
Air Blank	0.000	22:12	
Diagnostics Check	OK	22:12	

---

Data Download Date/Time: 03/21/2022 16:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: DIGMAN

DATE/TIME OF TEST: 03/05/2022 04:04:46

---

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:02	
Control Test	0.082	04:02	
Air Blank	0.000	04:03	
Subject Sample	# 1	04:03	Subject Test Refused
Breath Volume	0.000	04:03	
Air Blank	0.000	04:03	
Control Test	0.081	04:04	
Air Blank		04:04	RFI Detect
Air Blank	0.000	04:04	

---

Data Download Date/Time: 03/21/2022 16:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 07220080A2

DRY GAS EXPIRY DATE: 05/05/2022

VIOLATION CODE: DUI

ARREST TIME: 03:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: DIGMAN

DATE/TIME OF TEST: 03/05/2022 04:09:19

---

Diagnostics Check	OK	04:06	
Air Blank	0.000	04:06	
Control Test	0.081	04:07	
Air Blank	0.000	04:07	
Subject Sample	# 1	04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:08	
Control Test	0.081	04:08	
Air Blank	0.000	04:09	
Diagnostics Check	OK	04:09	

---

Data Download Date/Time: 03/21/2022 16:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CERMAK,OFC,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: HADLEY,OFC  
ARRESTING OFFICER AGENCY: Indian Shores PD  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: DIMOLA

DATE/TIME OF TEST: 03/06/2022 03:08:23

---

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:57	
Control Test		0.081	02:57	
Air Blank		0.000	02:58	
Subject Sample	# 1	0.105	03:01	Volume Not Met
Breath Volume		0.429	03:01	
Air Blank		0.000	03:02	
Air Blank		0.000	03:03	
Subject Sample	# 2		03:06	Subject Test Refused
Breath Volume		0.121	03:06	
Air Blank		0.000	03:07	
Control Test		0.081	03:07	
Air Blank		0.000	03:08	
Diagnostics Check		OK	03:08	

---

Data Download Date/Time: 03/21/2022 16:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSIER, MATTHEW, D  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: GOSIER, MATTHEW  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: RAMIREZ RAMIREZ

DATE/TIME OF TEST: 03/06/2022 03:53:54

---

Diagnostics Check		OK	03:44
Air Blank		0.000	03:45
Control Test		0.081	03:45
Air Blank		0.000	03:45
Subject Sample	# 1	0.217	03:46
Breath Volume		1.894	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.186	03:49
Breath Volume		1.789	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 3	0.216	03:52
Breath Volume		2.250	03:52
Air Blank		0.000	03:52
Control Test		0.081	03:53
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

---

Data Download Date/Time: 03/21/2022 16:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,DAVID,A  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: ROBERTS,DAVID  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 16:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:51

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 03/06/2022 17:21:54

---

Diagnostics Check	OK	17:14
Air Blank	0.000	17:14
Control Test	0.081	17:15
Air Blank	0.000	17:15
Subject Sample # 1	0.081	17:17
Breath Volume	2.558	17:17
Air Blank	0.000	17:17
Air Blank	0.000	17:19
Subject Sample # 2	0.081	17:20
Breath Volume	3.144	17:20
Air Blank	0.000	17:20
Control Test	0.081	17:21
Air Blank	0.000	17:21
Diagnostics Check	OK	17:21

---

Data Download Date/Time: 03/21/2022 16:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSIER, MATTHEW, D  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: GOSIER, MATTHEW  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 03/12/2022 02:56:47

---

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:54	
Control Test	0.080	02:54	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:55	Subject Test Refused
Breath Volume	0.000	02:55	
Air Blank	0.000	02:55	
Control Test	0.080	02:56	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

---

Data Download Date/Time: 03/21/2022 16:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/12/2022 03:05:42

---

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.080	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.200	03:01
Breath Volume	2.554	03:01
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.189	03:04
Breath Volume	2.195	03:04
Air Blank	0.000	03:04
Control Test	0.080	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

---

Data Download Date/Time: 03/21/2022 16:09



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSIER, MATTHEW, D  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: GOSIER, MATTHEW  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 04:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: MODRAK

DATE/TIME OF TEST: 03/13/2022 05:37:57

---

Diagnostics Check		OK	05:31
Air Blank		0.000	05:32
Control Test		0.082	05:32
Air Blank		0.000	05:33
Subject Sample	# 1	0.159	05:33
Breath Volume		3.203	05:33
Air Blank		0.000	05:33
Air Blank		0.000	05:35
Subject Sample	# 2	0.155	05:36
Breath Volume		3.734	05:36
Air Blank		0.000	05:37
Control Test		0.081	05:37
Air Blank		0.000	05:37
Diagnostics Check		OK	05:37

---

Data Download Date/Time: 03/21/2022 16:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HISA,NERITAN,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: HISA,NERITAN  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: BORDERS

DATE/TIME OF TEST: 03/15/2022 07:13:06

---

Diagnostics Check	OK	07:12	
Air Blank		07:12	Ambient Fail
Air Blank		07:13	RFI Detect

---

Data Download Date/Time: 03/21/2022 16:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HISA,NERITAN,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: HISA,NERITAN  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 06:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: BORDERS

DATE/TIME OF TEST: 03/15/2022 07:21:01

---

Diagnostics Check	OK	07:15	
Air Blank	0.000	07:16	
Control Test	0.081	07:16	
Air Blank	0.000	07:17	
Subject Sample	# 1 0.121	07:17	
Breath Volume	5.875	07:17	
Air Blank	0.000	07:18	
Air Blank	0.000	07:19	
Subject Sample	# 2 0.117	07:20	
Breath Volume	4.894	07:20	
Air Blank		07:20	RFI Detect
Air Blank	0.000	07:21	

---

Data Download Date/Time: 03/21/2022 16:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 07220080A2

DRY GAS EXPIRY DATE: 05/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:56

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: FALCON

DATE/TIME OF TEST: 03/17/2022 03:43:40

---

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:41	
Control Test	0.080	03:41	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:42	
Control Test	0.081	03:43	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

---

Data Download Date/Time: 03/21/2022 16:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICE,JOHN,B  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SINGSON,ROLAND  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 21:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: SCHNERR

DATE/TIME OF TEST: 03/17/2022 22:40:22

---

Diagnostics Check	OK	22:33
Air Blank	0.000	22:34
Control Test	0.081	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.260	22:35
Breath Volume	3.046	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.259	22:38
Breath Volume	3.199	22:38
Air Blank	0.000	22:39
Control Test	0.079	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

---

Data Download Date/Time: 03/21/2022 16:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KATZ,JUSTIN,T  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: MORALES CRUZ

DATE/TIME OF TEST: 03/18/2022 01:35:09

---

Diagnostics Check		OK	01:27
Air Blank		0.000	01:27
Control Test		0.081	01:28
Air Blank		0.000	01:28
Subject Sample	# 1	0.210	01:30
Breath Volume		2.539	01:30
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.200	01:33
Breath Volume		2.515	01:33
Air Blank		0.000	01:34
Control Test		0.080	01:34
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

---

Data Download Date/Time: 03/21/2022 16:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSIER, MATTHEW, D  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: GOSIER, MATTHEW  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BAUMAN

DATE/TIME OF TEST: 03/18/2022 01:47:44

---

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.169	01:43
Breath Volume	2.039	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.166	01:46
Breath Volume	2.000	01:46
Air Blank	0.000	01:46
Control Test	0.081	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

---

Data Download Date/Time: 03/21/2022 16:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 03/19/2022 03:09:55

---

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.081	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.175	03:05
Breath Volume	2.398	03:05
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.171	03:08
Breath Volume	2.210	03:08
Air Blank	0.000	03:09
Control Test	0.081	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

---

Data Download Date/Time: 03/21/2022 16:09



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSIER, MATTHEW, D  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: GOSIER, MATTHEW  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: WILBURN

DATE/TIME OF TEST: 03/19/2022 21:23:17

---

Diagnostics Check	OK	21:17
Air Blank	0.000	21:17
Control Test	0.080	21:18
Air Blank	0.000	21:18
Subject Sample # 1	0.000	21:18
Breath Volume	2.128	21:18
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample # 2	0.000	21:21
Breath Volume	2.324	21:21
Air Blank	0.000	21:22
Control Test	0.081	21:22
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

---

Data Download Date/Time: 03/21/2022 16:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSIER, MATTHEW, D  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: GOSIER, MATTHEW  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: SANTOS RESENDIZ

DATE/TIME OF TEST: 03/20/2022 04:26:42

---

Diagnostics Check		OK	04:18
Air Blank		0.000	04:18
Control Test		0.081	04:19
Air Blank		0.000	04:19
Subject Sample	# 1	0.156	04:21
Breath Volume		2.492	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.158	04:25
Breath Volume		1.878	04:25
Air Blank		0.000	04:25
Control Test		0.080	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

---

Data Download Date/Time: 03/21/2022 16:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSIER, MATTHEW, D  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: GOSIER, MATTHEW  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 03/20/2022 21:31:31

---

Diagnostics Check	OK	21:24
Air Blank	0.000	21:25
Control Test	0.081	21:25
Air Blank	0.000	21:26
Subject Sample # 1	0.139	21:26
Breath Volume	2.132	21:26
Air Blank	0.000	21:27
Air Blank	0.000	21:29
Subject Sample # 2	0.130	21:30
Breath Volume	1.832	21:30
Air Blank	0.000	21:30
Control Test	0.081	21:30
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

---

Data Download Date/Time: 03/21/2022 16:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,DAVID,A  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: ROBERTS,DAVID  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 07220080A2  
DRY GAS EXPIRY DATE: 05/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: ALBRECHT

DATE/TIME OF TEST: 03/21/2022 00:14:35

---

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.081	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.000	00:10
Breath Volume	3.507	00:10
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.000	00:13
Breath Volume	3.441	00:13
Air Blank	0.000	00:13
Control Test	0.081	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

---

Data Download Date/Time: 03/21/2022 16:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: GARCIA LIBRADO

DATE/TIME OF TEST: 03/23/2022 22:23:56

---

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.080	22:17
Air Blank	0.000	22:17
Subject Sample # 1	0.186	22:19
Breath Volume	2.367	22:19
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.173	22:22
Breath Volume	1.281	22:22
Air Blank	0.000	22:23
Control Test	0.080	22:23
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

---

Data Download Date/Time: 04/06/2022 10:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 01:22

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: MUTHIG

DATE/TIME OF TEST: 03/25/2022 02:02:25

---

Diagnostics Check	OK	01:59	
Air Blank	0.000	01:59	
Control Test	0.081	02:00	
Air Blank	0.000	02:00	
Subject Sample	# 1	02:01	Subject Test Refused
Breath Volume	0.000	02:01	
Air Blank	0.000	02:01	
Control Test	0.080	02:01	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

---

Data Download Date/Time: 04/06/2022 10:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 23:41

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: ROARK

DATE/TIME OF TEST: 03/25/2022 23:48:27

---

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.080	23:42
Air Blank	0.000	23:43
Subject Sample # 1	0.147	23:44
Breath Volume	2.042	23:44
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.138	23:46
Breath Volume	2.425	23:46
Air Blank	0.000	23:47
Control Test	0.079	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

---

Data Download Date/Time: 04/06/2022 10:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: HURTADO SANCHEZ

DATE/TIME OF TEST: 03/26/2022 05:42:28

---

Diagnostics Check	OK	05:36
Air Blank	0.000	05:36
Control Test	0.080	05:37
Air Blank	0.000	05:37
Subject Sample # 1	0.173	05:38
Breath Volume	1.835	05:38
Air Blank	0.000	05:38
Air Blank	0.000	05:40
Subject Sample # 2	0.165	05:40
Breath Volume	2.367	05:40
Air Blank	0.000	05:41
Control Test	0.080	05:41
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

---

Data Download Date/Time: 04/06/2022 10:02



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSIER, MATTHEW, D  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: HICKS, JASON  
ARRESTING OFFICER AGENCY: Indian Shores PD  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/28/2022 01:20:48

---

Diagnostics Check		OK	01:14
Air Blank		0.000	01:15
Control Test		0.080	01:15
Air Blank		0.000	01:15
Subject Sample	# 1	0.202	01:16
Breath Volume		2.316	01:16
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.197	01:19
Breath Volume		1.929	01:19
Air Blank		0.000	01:19
Control Test		0.080	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

---

Data Download Date/Time: 04/06/2022 10:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HISA,NERITAN,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: MARCOUX,HOPE  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 08:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:04

SUBJECT'S LAST NAME: SATTERWHITE

DATE/TIME OF TEST: 03/31/2022 09:27:21

---

Diagnostics Check	OK	09:25	
Air Blank	0.000	09:25	
Control Test	0.080	09:25	
Air Blank	0.000	09:26	
Subject Sample	# 1	09:26	RFI Detect
Breath Volume	0.000	09:26	
Air Blank		09:26	RFI Detect
Air Blank	0.000	09:27	

---

Data Download Date/Time: 04/06/2022 10:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HISA,NERITAN,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: MARCOUX,HOPE  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 08:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:04

SUBJECT'S LAST NAME: SATTERWHITE

DATE/TIME OF TEST: 03/31/2022 09:45:17

---

Diagnostics Check	OK	09:31	
Air Blank	0.000	09:32	
Control Test	0.081	09:32	
Air Blank	0.000	09:32	
Subject Sample	# 1	09:36	No Sample Provided
Breath Volume	0.121	09:36	
Air Blank	0.000	09:36	
Air Blank	0.000	09:38	
Subject Sample	# 2	09:40	
Breath Volume	2.050	09:40	
Air Blank	0.000	09:41	
Air Blank	0.000	09:43	
Subject Sample	# 3	09:43	
Breath Volume	2.019	09:43	
Air Blank	0.000	09:44	
Control Test	0.080	09:44	
Air Blank	0.000	09:45	
Diagnostics Check	OK	09:45	

---

Data Download Date/Time: 04/06/2022 10:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HISA,NERITAN,  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: MARCOUX,HOPE  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 08:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:04

SUBJECT'S LAST NAME: SATTERWHITE

DATE/TIME OF TEST: 03/31/2022 09:51:12

---

Diagnostics Check	OK	09:49	
Air Blank	0.000	09:49	
Control Test	0.081	09:49	
Air Blank	0.000	09:50	
Subject Sample	# 1	09:50	Sequence Aborted
Breath Volume	0.000	09:50	
Air Blank	0.000	09:51	

---

Data Download Date/Time: 04/06/2022 10:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 23621080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 19:57

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: SOTO ARIAS

DATE/TIME OF TEST: 03/31/2022 20:43:58

---

Diagnostics Check	OK	20:39	
Air Blank	0.000	20:40	
Control Test	0.080	20:40	
Air Blank	0.000	20:40	
Subject Sample	# 1	20:42	Subject Test Refused
Breath Volume	0.000	20:42	
Air Blank	0.000	20:42	
Control Test	0.081	20:43	
Air Blank		20:43	RFI Detect
Air Blank	0.000	20:43	

---

Data Download Date/Time: 04/06/2022 10:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13921080A2  
DRY GAS EXPIRY DATE: 07/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 03/26/2022 04:40:16

---

Diagnostics Check	OK	04:28	
Air Blank	0.000	04:29	
Control Test	0.078	04:29	
Air Blank	0.000	04:30	
Subject Sample	# 1	04:33	No Sample Provided
Breath Volume	0.000	04:33	
Air Blank	0.000	04:33	
Air Blank	0.000	04:35	
Subject Sample	# 2	04:38	No Sample Provided
Breath Volume	0.000	04:38	
Air Blank	0.000	04:39	
Control Test	0.079	04:39	
Air Blank	0.000	04:40	
Diagnostics Check	OK	04:40	

---

Data Download Date/Time: 04/22/2022 15:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DE LA CRUZ,LEONARDO,E  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: DE LA CRUZ,LEONARDO  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13921080A2  
DRY GAS EXPIRY DATE: 07/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 10:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:43

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 03/26/2022 12:25:32

---

Diagnostics Check	OK	12:21	
Air Blank	0.000	12:21	
Control Test	0.080	12:21	
Air Blank	0.000	12:22	
Subject Sample # 1	0.132	12:23	
Breath Volume	1.570	12:23	
Air Blank		12:24	RFI Detect
Air Blank	0.000	12:24	
		12:25	RFI Detect
Air Blank	0.000	12:25	

---

Data Download Date/Time: 04/22/2022 15:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DE LA CRUZ,LEONARDO,E

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: DE LA CRUZ,LEONARDO

ARRESTING OFFICER AGENCY: FHP Troop C

LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13921080A2

DRY GAS EXPIRY DATE: 07/05/2023

VIOLATION CODE: DUI

ARREST TIME: 10:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 11:43

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 03/26/2022 12:35:55

---

Diagnostics Check	OK	12:28
Air Blank	0.000	12:28
Control Test	0.076	12:29
Air Blank	0.000	12:29
Subject Sample # 1	0.127	12:30
Breath Volume	1.746	12:30
Air Blank	0.000	12:31
Air Blank	0.000	12:32
Subject Sample # 2	0.131	12:34
Breath Volume	1.757	12:34
Air Blank	0.000	12:35
Control Test	0.076	12:35
Air Blank	0.000	12:35
Diagnostics Check	OK	12:35

---

Data Download Date/Time: 04/22/2022 15:06



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELCHER,ALEX,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BELCHER,ALEX  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 13921080A2  
DRY GAS EXPIRY DATE: 07/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: CUCE

DATE/TIME OF TEST: 03/29/2022 03:00:38

---

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.078	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.247	02:55
Breath Volume	1.562	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.248	02:59
Breath Volume	1.906	02:59
Air Blank	0.000	02:59
Control Test	0.078	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

---

Data Download Date/Time: 04/22/2022 15:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DION,KEVIN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: DION,KEVIN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 03/02/2022 01:44:04

---

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.174	01:39
Breath Volume	2.656	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:41
Subject Sample # 2	0.174	01:42
Breath Volume	1.750	01:42
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:44

---

Data Download Date/Time: 03/07/2022 14:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUES, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: COOK, WILLIAM  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: GIL

DATE/TIME OF TEST: 03/04/2022 19:04:03

---

Diagnostics Check	OK	18:55
Air Blank	0.000	18:55
Control Test	0.080	18:56
Air Blank	0.000	18:56
Subject Sample # 1	0.156	18:58
Breath Volume	2.687	18:58
Air Blank	0.000	18:59
Air Blank	0.000	19:01
Subject Sample # 2	0.162	19:02
Breath Volume	2.484	19:02
Air Blank	0.000	19:03
Control Test	0.080	19:03
Air Blank	0.000	19:03
Diagnostics Check	OK	19:04

---

Data Download Date/Time: 03/07/2022 14:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIZZO,HAYLEY,J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: RIZZO,HAYLEY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: LIKUS

DATE/TIME OF TEST: 03/05/2022 01:16:58

---

Diagnostics Check		OK	01:08
Air Blank		0.000	01:09
Control Test		0.080	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.192	01:11
Breath Volume		2.207	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.191	01:15
Breath Volume		1.304	01:15
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

---

Data Download Date/Time: 03/07/2022 14:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: MOSCHIANO,DAVID  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: PUENTES

DATE/TIME OF TEST: 03/05/2022 04:09:29

---

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.081	04:04
Air Blank		0.000	04:04
Subject Sample	# 1	0.210	04:05
Breath Volume		2.292	04:05
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.194	04:07
Breath Volume		2.039	04:07
Air Blank		0.000	04:08
Control Test		0.080	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

---

Data Download Date/Time: 03/07/2022 14:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: WATSON,WILLIAM  
ARRESTING OFFICER AGENCY: Dade City PD  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: PEREZ REYNOSO

DATE/TIME OF TEST: 03/05/2022 04:24:49

---

Diagnostics Check		OK	04:17
Air Blank		0.000	04:18
Control Test		0.080	04:18
Air Blank		0.000	04:18
Subject Sample	# 1	0.201	04:20
Breath Volume		2.707	04:20
Air Blank		0.000	04:20
Air Blank		0.000	04:22
Subject Sample	# 2	0.202	04:23
Breath Volume		2.933	04:23
Air Blank		0.000	04:23
Control Test		0.079	04:24
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

---

Data Download Date/Time: 03/07/2022 14:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DION,KEVIN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: DION,KEVIN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: TROMBLEY

DATE/TIME OF TEST: 03/06/2022 01:56:33

---

Diagnostics Check		OK	01:49
Air Blank		0.000	01:50
Control Test		0.080	01:50
Air Blank		0.000	01:51
Subject Sample	# 1	0.139	01:51
Breath Volume		1.410	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.138	01:55
Breath Volume		1.289	01:55
Air Blank		0.000	01:55
Control Test		0.080	01:55
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

---

Data Download Date/Time: 03/07/2022 14:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIZZO,HAYLEY,J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: RIZZO,HAYLEY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 05:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: SOTO MARTINEZ

DATE/TIME OF TEST: 03/06/2022 06:24:15

---

Diagnostics Check	OK	06:21	
Air Blank	0.000	06:21	
Control Test	0.080	06:22	
Air Blank	0.000	06:22	
Subject Sample	# 1	06:22	Subject Test Refused
Breath Volume	0.000	06:22	
Air Blank	0.000	06:23	
Control Test	0.081	06:23	
Air Blank	0.000	06:24	
Diagnostics Check	OK	06:24	

---

Data Download Date/Time: 03/07/2022 14:56



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIZZO,HAYLEY,J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: RIZZO,HAYLEY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 20:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 03/06/2022 22:03:53

---

Diagnostics Check	OK	22:01	
Air Blank	0.000	22:01	
Control Test	0.079	22:01	
Air Blank	0.000	22:02	
Subject Sample	# 1	22:02	Subject Test Refused
Breath Volume	0.000	22:02	
Air Blank	0.000	22:02	
Control Test	0.079	22:03	
Air Blank	0.000	22:03	
Diagnostics Check	OK	22:03	

---

Data Download Date/Time: 03/07/2022 14:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHANSEN,LLOYD,D  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: JOHANSEN,LLOYD  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 00:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 03/07/2022 01:54:24

---

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.081	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.098	01:49
Breath Volume	3.675	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.103	01:52
Breath Volume	3.785	01:52
Air Blank	0.000	01:53
Control Test	0.080	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

---

Data Download Date/Time: 03/07/2022 14:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONG,JEFFERSON,D  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: LONG,JEFFERSON  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 03/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: WHELAN

DATE/TIME OF TEST: 03/09/2022 02:27:31

---

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.183	02:23
Breath Volume	3.695	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.188	02:25
Breath Volume	3.515	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

---

Data Download Date/Time: 04/01/2022 12:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIMARCO,AUSTIN,E  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: DIMARCO,AUSTIN  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 03/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 03/12/2022 00:11:30

---

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.081	00:05
Air Blank		0.000	00:06
Subject Sample	# 1	0.214	00:06
Breath Volume		3.203	00:06
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.217	00:09
Breath Volume		3.199	00:09
Air Blank		0.000	00:10
Control Test		0.079	00:10
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

---

Data Download Date/Time: 04/01/2022 12:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HERNANDEZ,ALAIN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: CAMPOS VALDEZ

DATE/TIME OF TEST: 03/14/2022 23:17:23

---

Diagnostics Check		OK	23:11
Air Blank		0.000	23:11
Control Test		0.081	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.181	23:12
Breath Volume		2.261	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.176	23:15
Breath Volume		2.160	23:15
Air Blank		0.000	23:16
Control Test		0.081	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

---

Data Download Date/Time: 04/01/2022 12:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DION,KEVIN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: DION,KEVIN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/15/2022 02:28:12

---

Diagnostics Check		OK	02:21
Air Blank		0.000	02:21
Control Test		0.080	02:22
Air Blank		0.000	02:22
Subject Sample	# 1	0.240	02:23
Breath Volume		1.738	02:23
Air Blank		0.000	02:24
Air Blank		0.000	02:25
Subject Sample	# 2	0.254	02:26
Breath Volume		1.511	02:26
Air Blank		0.000	02:27
Control Test		0.079	02:27
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

---

Data Download Date/Time: 04/01/2022 12:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DION,KEVIN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: DION,KEVIN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: GERST

DATE/TIME OF TEST: 03/19/2022 00:52:14

---

Diagnostics Check	OK	00:43	
Air Blank	0.000	00:44	
Control Test	0.081	00:44	
Air Blank	0.000	00:45	
Subject Sample	# 1	00:48	No Sample Provided
Breath Volume	0.000	00:48	
Air Blank	0.000	00:48	
Air Blank	0.000	00:50	
Subject Sample	# 2	00:50	Subject Test Refused
Breath Volume	0.000	00:50	
Air Blank	0.000	00:51	
Control Test	0.080	00:51	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

---

Data Download Date/Time: 04/01/2022 12:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIZZO,HAYLEY,J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: RIZZO,HAYLEY  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: MCCARTNEY

DATE/TIME OF TEST: 03/19/2022 03:09:21

---

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.080	03:03
Air Blank		0.000	03:04
Subject Sample	# 1	0.227	03:04
Breath Volume		2.992	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.231	03:07
Breath Volume		3.328	03:07
Air Blank		0.000	03:08
Control Test		0.079	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

---

Data Download Date/Time: 04/01/2022 12:43



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: HERNANDEZ,ALAIN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: COPELAND JR

DATE/TIME OF TEST: 03/19/2022 03:50:31

---

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.080	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.246	03:46
Breath Volume		3.230	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.244	03:48
Breath Volume		3.187	03:48
Air Blank		0.000	03:49
Control Test		0.079	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

---

Data Download Date/Time: 04/01/2022 12:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CONNELL,JACOB,B  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: CONNELL,JACOB  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 09:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:19

SUBJECT'S LAST NAME: FLEER

DATE/TIME OF TEST: 03/19/2022 09:50:28

---

Diagnostics Check	OK	09:43
Air Blank	0.000	09:44
Control Test	0.080	09:44
Air Blank	0.000	09:44
Subject Sample # 1	0.151	09:45
Breath Volume	1.558	09:45
Air Blank	0.000	09:46
Air Blank	0.000	09:48
Subject Sample # 2	0.152	09:48
Breath Volume	1.777	09:48
Air Blank	0.000	09:49
Control Test	0.079	09:49
Air Blank	0.000	09:50
Diagnostics Check	OK	09:50

---

Data Download Date/Time: 04/01/2022 12:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CONNELL,JACOB,B  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: CONNELL,JACOB  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 09:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: MERRITT

DATE/TIME OF TEST: 03/23/2022 11:10:43

---

Diagnostics Check	OK	11:04
Air Blank	0.000	11:05
Control Test	0.081	11:05
Air Blank	0.000	11:06
Subject Sample # 1	0.000	11:06
Breath Volume	2.800	11:06
Air Blank	0.000	11:06
Air Blank	0.000	11:08
Subject Sample # 2	0.000	11:09
Breath Volume	2.171	11:09
Air Blank	0.000	11:09
Control Test	0.082	11:10
Air Blank	0.000	11:10
Diagnostics Check	OK	11:10

---

Data Download Date/Time: 04/01/2022 12:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DION,KEVIN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: DION,KEVIN  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 15:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:52

SUBJECT'S LAST NAME: WILCZEWSKI

DATE/TIME OF TEST: 03/23/2022 16:40:57

---

Diagnostics Check	OK	16:32
Air Blank	0.000	16:32
Control Test	0.080	16:32
Air Blank	0.000	16:33
Subject Sample # 1	0.000	16:35
Breath Volume	2.011	16:35
Air Blank	0.000	16:36
Air Blank	0.000	16:38
Subject Sample # 2	0.000	16:39
Breath Volume	2.335	16:39
Air Blank	0.000	16:40
Control Test	0.081	16:40
Air Blank	0.000	16:40
Diagnostics Check	OK	16:40

---

Data Download Date/Time: 04/01/2022 12:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLOWAY,GRANT,D  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GALLOWAY,GRANT  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MARCEL

DATE/TIME OF TEST: 03/26/2022 04:22:27

---

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.081	04:16
Air Blank	0.000	04:17
Subject Sample # 1	0.152	04:17
Breath Volume	1.453	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.156	04:20
Breath Volume	1.707	04:20
Air Blank	0.000	04:21
Control Test	0.081	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

---

Data Download Date/Time: 04/01/2022 12:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CONNELL,JACOB,B  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: CONNELL,JACOB  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 12:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: AYERS

DATE/TIME OF TEST: 03/28/2022 13:29:32

---

Diagnostics Check	OK	13:21	
Air Blank	0.000	13:21	
Control Test	0.081	13:21	
Air Blank	0.000	13:22	
Subject Sample	# 1	13:25	No Sample Provided
Breath Volume	0.000	13:25	
Air Blank	0.000	13:25	
Air Blank	0.000	13:28	
Subject Sample	# 2	13:28	Subject Test Refused
Breath Volume	0.000	13:28	
Air Blank	0.000	13:28	
Control Test	0.081	13:28	
Air Blank	0.000	13:29	
Diagnostics Check	OK	13:29	

---

Data Download Date/Time: 04/01/2022 12:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: COLLINS,THOMAS  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 03/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 01:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: KERR

DATE/TIME OF TEST: 03/30/2022 02:11:23

---

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.081	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.228	02:06
Breath Volume		3.042	02:06
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.233	02:09
Breath Volume		3.328	02:09
Air Blank		0.000	02:10
Control Test		0.080	02:10
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

---

Data Download Date/Time: 04/01/2022 12:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WRIGHT,MARQUIS,A  
BREATH TEST OPERATOR AGENCY: Pasco County SO  
ARRESTING OFFICER: TUTTLE,FRANK  
ARRESTING OFFICER AGENCY: Pasco County SO  
LAST AGENCY INSPECTION: 03/07/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14020080A1  
DRY GAS EXPIRY DATE: 07/05/2022  
VIOLATION CODE: DUI  
ARREST TIME: 03:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: DAVIES

DATE/TIME OF TEST: 03/31/2022 05:18:11

---

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:13	
Control Test	0.080	05:14	
Air Blank	0.000	05:14	
Subject Sample	# 1	05:16	Subject Test Refused
Breath Volume	0.000	05:16	
Air Blank	0.000	05:17	
Control Test	0.081	05:17	
Air Blank	0.000	05:18	
Diagnostics Check	OK	05:18	

---

Data Download Date/Time: 04/01/2022 12:44



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHAHEEN,KEITH,B  
BREATH TEST OPERATOR AGENCY: Mayport Naval Base  
ARRESTING OFFICER: JILES,DARRYL  
ARRESTING OFFICER AGENCY: Mayport Naval Base  
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HUNSINGER

DATE/TIME OF TEST: 03/11/2022 04:06:17

---

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.078	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.197	04:01
Breath Volume	2.585	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.188	04:04
Breath Volume	2.210	04:04
Air Blank	0.000	04:05
Control Test	0.076	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

---

Data Download Date/Time: 03/24/2022 11:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOPEZ, RAYMOND, S  
BREATH TEST OPERATOR AGENCY: Mayport Naval Base  
ARRESTING OFFICER: MORAN, P  
ARRESTING OFFICER AGENCY: Mayport Naval Base  
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08121080A1  
DRY GAS EXPIRY DATE: 05/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: HAZWARD

DATE/TIME OF TEST: 03/16/2022 00:33:58

---

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:31	
Control Test	0.078	00:31	
Air Blank	0.000	00:32	
Subject Sample	# 1	00:32	Subject Test Refused
Breath Volume	0.000	00:32	
Air Blank	0.000	00:33	
Control Test	0.078	00:33	
Air Blank	0.000	00:33	
Diagnostics Check	OK	00:33	

---

Data Download Date/Time: 03/24/2022 11:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LITTLE,NICHOLAS,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: LITTLE,NICHOLAS  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 02/10/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 18:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/13/2022 19:45:51

---

Diagnostics Check	OK	19:42	
Air Blank	0.000	19:43	
Control Test	0.082	19:43	
Air Blank	0.000	19:43	
Subject Sample	# 1	19:44	Interferent Detect
Breath Volume	1.187	19:44	
Air Blank	0.000	19:44	
Control Test	0.081	19:45	
Air Blank	0.000	19:45	
Diagnostics Check	OK	19:45	

---

Data Download Date/Time: 04/05/2022 10:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LITTLE,NICHOLAS,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: LITTLE,NICHOLAS  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 02/10/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 18:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/13/2022 19:52:48

---

Diagnostics Check	OK	19:49	
Air Blank	0.000	19:50	
Control Test	0.081	19:50	
Air Blank	0.000	19:50	
Subject Sample	# 1	19:51	Interferent Detect
Breath Volume	1.230	19:51	
Air Blank	0.000	19:51	
Control Test	0.082	19:52	
Air Blank	0.000	19:52	
Diagnostics Check	OK	19:52	

---

Data Download Date/Time: 04/05/2022 10:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAGNER,BRETT,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: WAGNER,BRETT  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 02/10/2022

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1482048  
DRY GAS EXPIRY DATE: 04/21/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: JUNG

DATE/TIME OF TEST: 03/30/2022 21:15:11

---

Diagnostics Check	OK	21:08
Air Blank	0.000	21:08
Control Test	0.082	21:09
Air Blank	0.000	21:09
Subject Sample # 1	0.145	21:10
Breath Volume	1.460	21:10
Air Blank	0.000	21:10
Air Blank	0.000	21:12
Subject Sample # 2	0.151	21:13
Breath Volume	1.722	21:13
Air Blank	0.000	21:14
Control Test	0.082	21:14
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

---

Data Download Date/Time: 04/05/2022 10:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CODY,RICHARD,D  
BREATH TEST OPERATOR AGENCY: Deland PD  
ARRESTING OFFICER: LOWERY,DAISY  
ARRESTING OFFICER AGENCY: Deland PD  
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1433248  
DRY GAS EXPIRY DATE: 01/06/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/19/2022 23:22:16

---

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.080	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.194	23:17
Breath Volume	2.097	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.195	23:20
Breath Volume	1.851	23:20
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

---

Data Download Date/Time: 04/20/2022 09:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAYNE,ABIGAIL,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PAYNE,ABIGAIL  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: LANTIGUA

DATE/TIME OF TEST: 03/01/2022 02:24:00

---

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.082	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.000	02:19
Breath Volume	1.382	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.000	02:22
Breath Volume	1.507	02:22
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:24

---

Data Download Date/Time: 03/21/2022 16:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,ALBERT,J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WILLIAMS,ALBERT  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: ACIZ

DATE/TIME OF TEST: 03/02/2022 01:12:52

---

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:07
Subject Sample	# 1	0.199	01:08
Breath Volume		2.570	01:08
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.204	01:11
Breath Volume		3.109	01:11
Air Blank		0.000	01:11
Control Test		0.079	01:12
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

---

Data Download Date/Time: 03/21/2022 16:47



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAWSON JR, MICHAEL, E  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: STEPHENS, TONY  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: SIMSPON

DATE/TIME OF TEST: 03/02/2022 04:50:53

---

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:39	
Control Test	0.080	04:40	
Air Blank	0.000	04:40	
Subject Sample	# 1	04:43	No Sample Provided
Breath Volume	0.000	04:43	
Air Blank	0.000	04:44	
Air Blank	0.000	04:46	
Subject Sample	# 2	04:49	No Sample Provided
Breath Volume	0.000	04:49	
Air Blank	0.000	04:49	
Control Test	0.080	04:50	
Air Blank	0.000	04:50	
Diagnostics Check	OK	04:50	

---

Data Download Date/Time: 03/21/2022 16:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PROPST,JOHN,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PROPST,JOHN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 12:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: BIANCHI

DATE/TIME OF TEST: 03/02/2022 13:55:58

---

Diagnostics Check	OK	13:44	
Air Blank	0.000	13:45	
Control Test	0.081	13:45	
Air Blank	0.000	13:45	
Subject Sample	# 1	13:49	No Sample Provided
Breath Volume	0.000	13:49	
Air Blank	0.000	13:49	
Air Blank	0.000	13:51	
Subject Sample	# 2	13:54	No Sample Provided
Breath Volume	0.000	13:54	
Air Blank	0.000	13:55	
Control Test	0.080	13:55	
Air Blank	0.000	13:55	
Diagnostics Check	OK	13:55	

---

Data Download Date/Time: 03/21/2022 16:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: HODGES,MARCO  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 17:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: VILLEGAS

DATE/TIME OF TEST: 03/02/2022 19:28:24

---

Diagnostics Check	OK	19:20
Air Blank	0.000	19:20
Control Test	0.080	19:21
Air Blank	0.000	19:21
Subject Sample # 1	0.000	19:23
Breath Volume	1.261	19:23
Air Blank	0.000	19:24
Air Blank	0.000	19:26
Subject Sample # 2	0.000	19:27
Breath Volume	1.417	19:27
Air Blank	0.000	19:27
Control Test	0.080	19:27
Air Blank	0.000	19:28
Diagnostics Check	OK	19:28

---

Data Download Date/Time: 03/21/2022 16:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: GYPPTS

DATE/TIME OF TEST: 03/02/2022 21:23:34

---

Diagnostics Check	OK	21:12	
Air Blank	0.000	21:12	
Control Test	0.080	21:12	
Air Blank	0.000	21:13	
Subject Sample	# 1	21:16	No Sample Provided
Breath Volume	0.000	21:16	
Air Blank	0.000	21:17	
Air Blank	0.000	21:19	
Subject Sample	# 2	21:22	No Sample Provided
Breath Volume	0.000	21:22	
Air Blank	0.000	21:22	
Control Test	0.080	21:22	
Air Blank	0.000	21:23	
Diagnostics Check	OK	21:23	

---

Data Download Date/Time: 03/21/2022 16:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: RUDOLPH

DATE/TIME OF TEST: 03/03/2022 05:41:49

---

Diagnostics Check	OK	05:35
Air Blank	0.000	05:35
Control Test	0.080	05:36
Air Blank	0.000	05:36
Subject Sample # 1	0.116	05:37
Breath Volume	1.402	05:37
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample # 2	0.111	05:40
Breath Volume	1.648	05:40
Air Blank	0.000	05:40
Control Test	0.080	05:41
Air Blank	0.000	05:41
Diagnostics Check	OK	05:41

---

Data Download Date/Time: 03/21/2022 16:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: SALINAS,JOSE  
ARRESTING OFFICER AGENCY: Winter Haven PD  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: TUKES

DATE/TIME OF TEST: 03/03/2022 19:18:48

---

Diagnostics Check	OK	19:07	
Air Blank	0.000	19:07	
Control Test	0.080	19:08	
Air Blank	0.000	19:08	
Subject Sample	# 1	19:11	No Sample Provided
Breath Volume	0.000	19:11	
Air Blank	0.000	19:12	
Air Blank	0.000	19:14	
Subject Sample	# 2	19:17	No Sample Provided
Breath Volume	0.000	19:17	
Air Blank	0.000	19:17	
Control Test	0.080	19:18	
Air Blank	0.000	19:18	
Diagnostics Check	OK	19:18	

---

Data Download Date/Time: 03/21/2022 16:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 03/03/2022 21:09:31

---

Diagnostics Check		OK	20:57
Air Blank		0.000	20:58
Control Test		0.080	20:58
Air Blank		0.000	20:59
Subject Sample	# 1	0.218	21:00
Breath Volume		1.160	21:00
Air Blank		0.000	21:01
Air Blank		0.000	21:02
Subject Sample	# 2	0.195	21:04
Breath Volume		1.535	21:04
Air Blank		0.000	21:05
Air Blank		0.000	21:07
Subject Sample	# 3	0.194	21:08
Breath Volume		1.582	21:08
Air Blank		0.000	21:08
Control Test		0.079	21:08
Air Blank		0.000	21:09
Diagnostics Check		OK	21:09

---

Data Download Date/Time: 03/21/2022 16:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANDREWS, TYREE, J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: RIGGALL, NICHOLAS  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/05/2022 03:51:42

---

Diagnostics Check		OK	03:34	
Air Blank		0.000	03:35	
Control Test		0.080	03:35	
Air Blank		0.000	03:36	
Subject Sample	# 1	0.000	03:39	Volume Not Met
Breath Volume		0.015	03:39	
Air Blank		0.000	03:39	
Air Blank		0.000	03:41	
Subject Sample	# 2	0.073	03:44	
Breath Volume		1.222	03:44	
Air Blank		0.000	03:45	
Air Blank		0.000	03:47	
Subject Sample	# 3	0.074	03:50	
Breath Volume		1.988	03:50	
Air Blank		0.000	03:50	
Control Test		0.081	03:51	
Air Blank		0.000	03:51	
Diagnostics Check		OK	03:51	

---

Data Download Date/Time: 03/21/2022 16:48



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAWSON JR, MICHAEL, E  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: LEDFORD, ANN  
ARRESTING OFFICER AGENCY: Belleair Beach PD  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 03/06/2022 00:35:04

---

Diagnostics Check		OK	00:28
Air Blank		0.000	00:29
Control Test		0.081	00:29
Air Blank		0.000	00:29
Subject Sample	# 1	0.105	00:30
Breath Volume		1.667	00:30
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.101	00:33
Breath Volume		1.597	00:33
Air Blank		0.000	00:34
Control Test		0.080	00:34
Air Blank		0.000	00:34
Diagnostics Check		OK	00:35

---

Data Download Date/Time: 03/21/2022 16:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELCHER,ALEX,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BELCHER,ALEX  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: BUCK

DATE/TIME OF TEST: 03/06/2022 03:45:03

---

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:34	
Control Test	0.081	03:34	
Air Blank	0.000	03:34	
Subject Sample	# 1	03:38	No Sample Provided
Breath Volume	0.000	03:38	
Air Blank	0.000	03:38	
Air Blank	0.000	03:40	
Subject Sample	# 2	03:43	No Sample Provided
Breath Volume	0.000	03:43	
Air Blank	0.000	03:44	
Control Test	0.081	03:44	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:45	

---

Data Download Date/Time: 03/21/2022 16:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAYNE, ABIGAIL, D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PAYNE, ABIGAIL  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: MARTINEZ JUAREZ

DATE/TIME OF TEST: 03/06/2022 04:14:36

---

Diagnostics Check		OK	04:03	
Air Blank		0.000	04:03	
Control Test		0.081	04:03	
Air Blank		0.000	04:04	
Subject Sample	# 1	0.000	04:07	Volume Not Met
Breath Volume		0.000	04:07	
Air Blank		0.000	04:07	
Air Blank		0.000	04:09	
Subject Sample	# 2	0.233	04:13	Volume Not Met
Breath Volume		0.707	04:13	
Air Blank		0.000	04:13	
Control Test		0.081	04:14	
Air Blank		0.000	04:14	
Diagnostics Check		OK	04:14	

---

Data Download Date/Time: 03/21/2022 16:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAYNE,ABIGAIL,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PAYNE,ABIGAIL  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: HERNANDEZ CASTRO

DATE/TIME OF TEST: 03/06/2022 21:48:47

---

Diagnostics Check		OK	21:41
Air Blank		0.000	21:41
Control Test		0.081	21:42
Air Blank		0.000	21:42
Subject Sample	# 1	0.263	21:43
Breath Volume		1.605	21:43
Air Blank		0.000	21:44
Air Blank		0.000	21:46
Subject Sample	# 2	0.251	21:47
Breath Volume		1.769	21:47
Air Blank		0.000	21:47
Control Test		0.080	21:48
Air Blank		0.000	21:48
Diagnostics Check		OK	21:48

---

Data Download Date/Time: 03/21/2022 16:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAYNE, ABIGAIL, D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: STROUD, DILLON  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: LIETZ

DATE/TIME OF TEST: 03/06/2022 22:42:16

---

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.080	22:36
Air Blank	0.000	22:36
Subject Sample # 1	0.237	22:37
Breath Volume	1.199	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.237	22:40
Breath Volume	1.339	22:40
Air Blank	0.000	22:41
Control Test	0.079	22:41
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

---

Data Download Date/Time: 03/21/2022 16:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SYKES,STEVEN,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: SYKES,STEVEN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: MANKA

DATE/TIME OF TEST: 03/07/2022 22:19:47

---

Diagnostics Check	OK	22:08	
Air Blank	0.000	22:08	
Control Test	0.080	22:09	
Air Blank	0.000	22:09	
Subject Sample	# 1	22:12	No Sample Provided
Breath Volume	0.000	22:12	
Air Blank	0.000	22:13	
Air Blank	0.000	22:15	
Subject Sample	# 2	22:18	No Sample Provided
Breath Volume	0.000	22:18	
Air Blank	0.000	22:18	
Control Test	0.080	22:19	
Air Blank	0.000	22:19	
Diagnostics Check	OK	22:19	

---

Data Download Date/Time: 03/21/2022 16:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: VALERIO,ANTHONY  
ARRESTING OFFICER AGENCY: Bartow PD  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: JULIUS

DATE/TIME OF TEST: 03/07/2022 23:53:22

---

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.080	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.137	23:47
Breath Volume	1.460	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.136	23:51
Breath Volume	1.312	23:51
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

---

Data Download Date/Time: 03/21/2022 16:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: VALERIO,ANTHONY  
ARRESTING OFFICER AGENCY: Bartow PD  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: SYLVESTER

DATE/TIME OF TEST: 03/08/2022 00:27:06

---

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.080	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.129	00:22
Breath Volume	1.406	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:24
Subject Sample # 2	0.128	00:25
Breath Volume	1.445	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

---

Data Download Date/Time: 03/21/2022 16:48



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DE LA CRUZ,LEONARDO,E  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: MORALES,ORLANDO  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 16:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: RATTACASA

DATE/TIME OF TEST: 03/09/2022 17:41:31

---

Diagnostics Check	OK	17:35
Air Blank	0.000	17:36
Control Test	0.081	17:36
Air Blank	0.000	17:36
Subject Sample # 1	0.000	17:37
Breath Volume	3.179	17:37
Air Blank	0.000	17:37
Air Blank	0.000	17:39
Subject Sample # 2	0.000	17:40
Breath Volume	3.167	17:40
Air Blank	0.000	17:40
Control Test	0.079	17:40
Air Blank	0.000	17:41
Diagnostics Check	OK	17:41

---

Data Download Date/Time: 03/21/2022 16:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,ALBERT,J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BLACKMAN,GERIG  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 03/09/2022 19:44:50

---

Diagnostics Check	OK	19:32	
Air Blank	0.000	19:33	
Control Test	0.079	19:33	
Air Blank	0.000	19:34	
Subject Sample	# 1	19:37	No Sample Provided
Breath Volume	0.000	19:37	
Air Blank	0.000	19:37	
Air Blank	0.000	19:39	
Subject Sample	# 2	0.203	19:40
Breath Volume	2.382	19:40	
Air Blank	0.000	19:41	
Air Blank	0.000	19:42	
Subject Sample	# 3	0.205	19:43
Breath Volume	2.011	19:43	
Air Blank	0.000	19:43	
Control Test	0.079	19:44	
Air Blank	0.000	19:44	
Diagnostics Check	OK	19:44	

---

Data Download Date/Time: 03/21/2022 16:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAWSON JR, MICHAEL, E  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BLACKMAN, CHARLES  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 03/10/2022 21:50:41

---

Diagnostics Check	OK	21:39	
Air Blank	0.000	21:39	
Control Test	0.081	21:40	
Air Blank	0.000	21:40	
Subject Sample	# 1	21:43	No Sample Provided
Breath Volume	0.000	21:43	
Air Blank	0.000	21:44	
Air Blank	0.000	21:46	
Subject Sample	# 2	21:49	No Sample Provided
Breath Volume	0.000	21:49	
Air Blank	0.000	21:49	
Control Test	0.081	21:50	
Air Blank	0.000	21:50	
Diagnostics Check	OK	21:50	

---

Data Download Date/Time: 03/21/2022 16:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAYNE,ABIGAIL,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PAYNE,ABIGAIL  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: TRAVER

DATE/TIME OF TEST: 03/11/2022 05:32:54

---

Diagnostics Check	OK	05:26
Air Blank	0.000	05:27
Control Test	0.081	05:27
Air Blank	0.000	05:28
Subject Sample # 1	0.000	05:28
Breath Volume	1.699	05:28
Air Blank	0.000	05:29
Air Blank	0.000	05:31
Subject Sample # 2	0.000	05:31
Breath Volume	2.000	05:31
Air Blank	0.000	05:32
Control Test	0.081	05:32
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

---

Data Download Date/Time: 03/21/2022 16:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PROPST,JOHN,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PROPST,JOHN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 11:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:04

SUBJECT'S LAST NAME: WASHINGTON WHITE

DATE/TIME OF TEST: 03/11/2022 12:34:07

---

Diagnostics Check	OK	12:27
Air Blank	0.000	12:28
Control Test	0.080	12:28
Air Blank	0.000	12:28
Subject Sample # 1	0.184	12:29
Breath Volume	1.687	12:29
Air Blank	0.000	12:30
Air Blank	0.000	12:32
Subject Sample # 2	0.175	12:32
Breath Volume	1.839	12:32
Air Blank	0.000	12:33
Control Test	0.081	12:33
Air Blank	0.000	12:34
Diagnostics Check	OK	12:34

---

Data Download Date/Time: 03/21/2022 16:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLON,JOLNETH,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: COLON,JOLNETH  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: MARTINEZ CRUZ

DATE/TIME OF TEST: 03/11/2022 21:46:19

---

Diagnostics Check	OK	21:38
Air Blank	0.000	21:39
Control Test	0.080	21:39
Air Blank	0.000	21:40
Subject Sample # 1	0.093	21:41
Breath Volume	2.023	21:41
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.090	21:44
Breath Volume	1.527	21:44
Air Blank	0.000	21:45
Control Test	0.080	21:45
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

---

Data Download Date/Time: 03/21/2022 16:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: CONOVER

DATE/TIME OF TEST: 03/11/2022 22:07:47

---

Diagnostics Check		OK	21:56	
Air Blank		0.000	21:56	
Control Test		0.080	21:57	
Air Blank		0.000	21:57	
Subject Sample	# 1	0.138	22:00	Volume Not Met
Breath Volume		0.187	22:00	
Air Blank		0.000	22:01	
Air Blank		0.000	22:03	
Subject Sample	# 2	0.124	22:06	Volume Not Met
Breath Volume		0.097	22:06	
Air Blank		0.000	22:06	
Control Test		0.080	22:07	
Air Blank		0.000	22:07	
Diagnostics Check		OK	22:07	

---

Data Download Date/Time: 03/21/2022 16:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORSE,JOHN,C  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: AMAYA,PETER  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 09:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: MORSE

DATE/TIME OF TEST: 03/11/2022 23:52:32

---

Diagnostics Check	OK	23:49	
Air Blank	0.000	23:49	
Control Test	0.080	23:49	
Air Blank	0.000	23:50	
Subject Sample	# 1 0.232	23:50	
Breath Volume	2.839	23:50	
Air Blank		23:51	Improper Sample
Air Blank	0.000	23:51	
Control Test	0.080	23:51	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:52	

---

Data Download Date/Time: 03/21/2022 16:49



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: HODGES,MARCO  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: SANCHEZ-CORDOZA

DATE/TIME OF TEST: 03/12/2022 03:57:40

---

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:43	
Control Test		0.080	03:43	
Air Blank		0.000	03:43	
Subject Sample	# 1	0.184	03:47	Volume Not Met
Breath Volume		0.039	03:47	
Air Blank		0.000	03:47	
Air Blank		0.000	03:49	
Subject Sample	# 2	0.186	03:52	
Breath Volume		1.207	03:52	
Air Blank		0.000	03:53	
Air Blank		0.000	03:54	
Subject Sample	# 3	0.197	03:56	
Breath Volume		1.195	03:56	
Air Blank		0.000	03:56	
Control Test		0.078	03:57	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

---

Data Download Date/Time: 03/21/2022 16:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLON,JOLNETH,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: COLON,JOLNETH  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: ESCORCIA-GUZMAN

DATE/TIME OF TEST: 03/12/2022 22:48:51

---

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.081	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.147	22:44
Breath Volume	1.648	22:44
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.153	22:47
Breath Volume	2.792	22:47
Air Blank	0.000	22:47
Control Test	0.081	22:48
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

---

Data Download Date/Time: 03/21/2022 16:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KUMAR, CHRISTOPHER, A  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: MORALES, ORLANDO  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MILES

DATE/TIME OF TEST: 03/13/2022 00:28:18

---

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.081	00:21
Air Blank	0.000	00:22
Subject Sample # 1	0.088	00:23
Breath Volume	1.812	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.085	00:26
Breath Volume	1.964	00:26
Air Blank	0.000	00:27
Control Test	0.079	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

---

Data Download Date/Time: 03/21/2022 16:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PROPST,JOHN,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PROPST,JOHN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 13:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:07

SUBJECT'S LAST NAME: LABBRUZZO

DATE/TIME OF TEST: 03/13/2022 14:39:32

---

Diagnostics Check	OK	14:33
Air Blank	0.000	14:33
Control Test	0.082	14:33
Air Blank	0.000	14:34
Subject Sample # 1	0.000	14:35
Breath Volume	3.105	14:35
Air Blank	0.000	14:35
Air Blank	0.000	14:37
Subject Sample # 2	0.000	14:38
Breath Volume	1.691	14:38
Air Blank	0.000	14:38
Control Test	0.082	14:38
Air Blank	0.000	14:39
Diagnostics Check	OK	14:39

---

Data Download Date/Time: 03/21/2022 16:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: ORNELAS

DATE/TIME OF TEST: 03/13/2022 23:10:28

---

Diagnostics Check	OK	22:59	
Air Blank	0.000	22:59	
Control Test	0.080	22:59	
Air Blank	0.000	23:00	
Subject Sample	# 1	23:03	No Sample Provided
Breath Volume	0.000	23:03	
Air Blank	0.000	23:03	
Air Blank	0.000	23:05	
Subject Sample	# 2	23:09	No Sample Provided
Breath Volume	0.000	23:09	
Air Blank	0.000	23:09	
Control Test	0.080	23:09	
Air Blank	0.000	23:10	
Diagnostics Check	OK	23:10	

---

Data Download Date/Time: 03/21/2022 16:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FLETCHER,BRIANA,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: FLETCHER,BRIANA  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 03/14/2022 00:18:05

---

Diagnostics Check		OK	00:10
Air Blank		0.000	00:11
Control Test		0.081	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.208	00:13
Breath Volume		2.753	00:13
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.193	00:16
Breath Volume		1.707	00:16
Air Blank		0.000	00:17
Control Test		0.079	00:17
Air Blank		0.000	00:17
Diagnostics Check		OK	00:18

---

Data Download Date/Time: 03/21/2022 16:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAWSON JR, MICHAEL, E  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: STEPHENS, TONY  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: PULLARA

DATE/TIME OF TEST: 03/14/2022 23:03:57

---

Diagnostics Check	OK	22:57
Air Blank	0.000	22:58
Control Test	0.081	22:58
Air Blank	0.000	22:58
Subject Sample # 1	0.155	22:59
Breath Volume	1.617	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.156	23:02
Breath Volume	1.585	23:02
Air Blank	0.000	23:03
Control Test	0.080	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

---

Data Download Date/Time: 03/21/2022 16:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAYNE,ABIGAIL,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PAYNE,ABIGAIL  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: SPERO

DATE/TIME OF TEST: 03/15/2022 02:32:08

---

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.081	02:26
Air Blank		0.000	02:27
Subject Sample	# 1	0.107	02:27
Breath Volume		1.277	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.100	02:30
Breath Volume		1.769	02:30
Air Blank		0.000	02:31
Control Test		0.080	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

---

Data Download Date/Time: 03/21/2022 16:50



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAYNE, ABIGAIL, D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BLACKMAN, GERIG  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 01:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 03/15/2022 03:51:26

---

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.081	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.167	03:46
Breath Volume	1.519	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.160	03:49
Breath Volume	1.566	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

---

Data Download Date/Time: 03/21/2022 16:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRADLEY,KEVIN,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BRADLEY,KEVIN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 11:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:05

SUBJECT'S LAST NAME: RODRIGUEZ II

DATE/TIME OF TEST: 03/15/2022 12:47:00

---

Diagnostics Check	OK	12:33	
Air Blank	0.000	12:33	
Control Test	0.080	12:34	
Air Blank	0.000	12:34	
Subject Sample # 1	0.000	12:36	
Breath Volume	1.328	12:36	
Air Blank	0.000	12:37	
Air Blank	0.000	12:39	
Subject Sample # 2	0.000	12:42	Volume Not Met
Breath Volume	0.476	12:42	
Air Blank	0.000	12:42	
Air Blank	0.000	12:44	
Subject Sample # 3		12:45	Subject Test Refused
Breath Volume	0.000	12:45	
Air Blank	0.000	12:46	
Control Test	0.080	12:46	
Air Blank	0.000	12:46	
Diagnostics Check	OK	12:47	

---

Data Download Date/Time: 03/21/2022 16:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,ALBERT,J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WILLIAMS,ALBERT  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 03/15/2022 23:51:10

---

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.080	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.202	23:44
Breath Volume	1.371	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.197	23:49
Breath Volume	1.636	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

---

Data Download Date/Time: 03/21/2022 16:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAYNE,ABIGAIL,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PAYNE,ABIGAIL  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: VANCE

DATE/TIME OF TEST: 03/16/2022 02:34:20

---

Diagnostics Check	OK	02:27
Air Blank	0.000	02:28
Control Test	0.081	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.150	02:29
Breath Volume	1.488	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.158	02:32
Breath Volume	1.371	02:32
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

---

Data Download Date/Time: 03/21/2022 16:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: HODGES,MARCO  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: CAIN

DATE/TIME OF TEST: 03/16/2022 22:43:34

---

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:30	
Control Test	0.081	22:30	
Air Blank	0.000	22:31	
Subject Sample	# 1	22:34	No Sample Provided
Breath Volume	0.000	22:34	
Air Blank	0.000	22:34	
Air Blank	0.000	22:36	
Subject Sample	# 2	0.120	22:38
Breath Volume	1.117	22:38	
Air Blank	0.000	22:39	
Air Blank	0.000	22:40	
Subject Sample	# 3	0.113	22:42
Breath Volume	1.285	22:42	
Air Blank	0.000	22:42	
Control Test	0.080	22:42	
Air Blank	0.000	22:43	
Diagnostics Check	OK	22:43	

---

Data Download Date/Time: 03/21/2022 16:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: LIGHTFOOT

DATE/TIME OF TEST: 03/17/2022 01:29:47

---

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.082	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.000	01:25
Breath Volume	1.875	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.000	01:28
Breath Volume	1.968	01:28
Air Blank	0.000	01:28
Control Test	0.081	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

---

Data Download Date/Time: 03/21/2022 16:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLON,JOLNETH,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: COLON,JOLNETH  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/17/2022 03:19:10

---

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.081	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.089	03:14
Breath Volume	1.980	03:14
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.089	03:17
Breath Volume	1.554	03:17
Air Blank	0.000	03:18
Control Test	0.081	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

---

Data Download Date/Time: 03/21/2022 16:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLON,JOLNETH,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: COLON,JOLNETH  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: APPEL

DATE/TIME OF TEST: 03/17/2022 19:38:30

---

Diagnostics Check	OK	19:32
Air Blank	0.000	19:32
Control Test	0.081	19:32
Air Blank	0.000	19:33
Subject Sample # 1	0.244	19:34
Breath Volume	2.832	19:34
Air Blank	0.000	19:34
Air Blank	0.000	19:36
Subject Sample # 2	0.246	19:37
Breath Volume	1.683	19:37
Air Blank	0.000	19:37
Control Test	0.080	19:37
Air Blank	0.000	19:38
Diagnostics Check	OK	19:38

---

Data Download Date/Time: 03/21/2022 16:50



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WALMSLEY,ROBERT  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: RAMOS SANTIAGO

DATE/TIME OF TEST: 03/17/2022 22:39:16

---

Diagnostics Check	OK	22:32
Air Blank	0.000	22:33
Control Test	0.080	22:33
Air Blank	0.000	22:34
Subject Sample # 1	0.129	22:35
Breath Volume	1.140	22:35
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.126	22:37
Breath Volume	1.441	22:37
Air Blank	0.000	22:38
Control Test	0.080	22:38
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

---

Data Download Date/Time: 03/21/2022 16:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 03/18/2022 05:08:17

---

Diagnostics Check	OK	05:01
Air Blank	0.000	05:01
Control Test	0.080	05:01
Air Blank	0.000	05:02
Subject Sample # 1	0.146	05:03
Breath Volume	1.132	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:05
Subject Sample # 2	0.129	05:06
Breath Volume	1.113	05:06
Air Blank	0.000	05:07
Control Test	0.080	05:07
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

---

Data Download Date/Time: 03/21/2022 16:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAYNE,ABIGAIL,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PAYNE,ABIGAIL  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: REAVES

DATE/TIME OF TEST: 03/18/2022 23:03:01

---

Diagnostics Check	OK	22:56
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.199	22:58
Breath Volume	1.156	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:00
Subject Sample # 2	0.185	23:01
Breath Volume	1.308	23:01
Air Blank	0.000	23:02
Control Test	0.080	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:03

---

Data Download Date/Time: 03/21/2022 16:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,ALBERT,J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WILLIAMS,ALBERT  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: NAVARRO LEON

DATE/TIME OF TEST: 03/19/2022 00:59:37

---

Diagnostics Check		OK	00:53
Air Blank		0.000	00:53
Control Test		0.080	00:53
Air Blank		0.000	00:54
Subject Sample	# 1	0.211	00:55
Breath Volume		3.507	00:55
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.208	00:58
Breath Volume		3.417	00:58
Air Blank		0.000	00:58
Control Test		0.079	00:59
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

---

Data Download Date/Time: 03/21/2022 16:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOFFORD,CLINT,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WOFFORD,CLINT  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 12:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:46

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 03/19/2022 14:22:11

---

Diagnostics Check		OK	14:15
Air Blank		0.000	14:16
Control Test		0.080	14:16
Air Blank		0.000	14:17
Subject Sample	# 1	0.000	14:17
Breath Volume		2.667	14:17
Air Blank		0.000	14:18
Air Blank		0.000	14:20
Subject Sample	# 2	0.000	14:20
Breath Volume		2.816	14:20
Air Blank		0.000	14:21
Control Test		0.080	14:21
Air Blank		0.000	14:22
Diagnostics Check		OK	14:22

---

Data Download Date/Time: 03/21/2022 16:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAYNE,ABIGAIL,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PAYNE,ABIGAIL  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: DYKERS

DATE/TIME OF TEST: 03/19/2022 21:34:48

---

Diagnostics Check	OK	21:28
Air Blank	0.000	21:28
Control Test	0.080	21:28
Air Blank	0.000	21:29
Subject Sample # 1	0.000	21:30
Breath Volume	1.750	21:30
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample # 2	0.000	21:33
Breath Volume	1.476	21:33
Air Blank	0.000	21:33
Control Test	0.081	21:34
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

---

Data Download Date/Time: 03/21/2022 16:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,ALBERT,J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WILLIAMS,ALBERT  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 03/20/2022 01:38:51

---

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.080	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.090	01:34
Breath Volume	2.054	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.083	01:37
Breath Volume	2.445	01:37
Air Blank	0.000	01:37
Control Test	0.080	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

---

Data Download Date/Time: 03/21/2022 16:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FLETCHER,MELISSA,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BLACKMAN,GERIG  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 03/20/2022 04:10:54

---

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.081	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.208	04:04
Breath Volume		1.203	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.194	04:09
Breath Volume		1.449	04:09
Air Blank		0.000	04:10
Control Test		0.079	04:10
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

---

Data Download Date/Time: 03/21/2022 16:51



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,ALBERT,J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: WILLIAMS,ALBERT  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 18:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: BELING

DATE/TIME OF TEST: 03/20/2022 18:54:07

---

Diagnostics Check	OK	18:47
Air Blank	0.000	18:48
Control Test	0.080	18:48
Air Blank	0.000	18:48
Subject Sample # 1	0.289	18:49
Breath Volume	3.023	18:49
Air Blank	0.000	18:50
Air Blank	0.000	18:52
Subject Sample # 2	0.280	18:52
Breath Volume	2.464	18:52
Air Blank	0.000	18:53
Control Test	0.079	18:53
Air Blank	0.000	18:54
Diagnostics Check	OK	18:54

---

Data Download Date/Time: 03/21/2022 16:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANDREWS, TYREE, J  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: RIGGALL, NICHOLAS  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/21/2022 02:29:47

---

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:18	
Control Test	0.080	02:19	
Air Blank	0.000	02:19	
Subject Sample	# 1	02:22	No Sample Provided
Breath Volume	0.000	02:22	
Air Blank	0.000	02:23	
Air Blank	0.000	02:25	
Subject Sample	# 2	02:28	No Sample Provided
Breath Volume	0.000	02:28	
Air Blank	0.000	02:28	
Control Test	0.081	02:29	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

---

Data Download Date/Time: 03/21/2022 16:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTALBANO,JONATHAN,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: MONTALBANO,JONATHAN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 08:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: RAYNOR

DATE/TIME OF TEST: 03/21/2022 09:28:04

---

Diagnostics Check		OK	09:21
Air Blank		0.000	09:22
Control Test		0.081	09:22
Air Blank		0.000	09:22
Subject Sample	# 1	0.200	09:23
Breath Volume		1.617	09:23
Air Blank		0.000	09:24
Air Blank		0.000	09:26
Subject Sample	# 2	0.210	09:26
Breath Volume		1.109	09:26
Air Blank		0.000	09:27
Control Test		0.080	09:27
Air Blank		0.000	09:27
Diagnostics Check		OK	09:28

---

Data Download Date/Time: 03/21/2022 16:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLON,JOLNETH,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: COLON,JOLNETH  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: PINA

DATE/TIME OF TEST: 03/23/2022 01:04:20

---

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.081	00:59
Air Blank		0.000	00:59
Subject Sample	# 1	0.000	01:00
Breath Volume		2.273	01:00
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.000	01:02
Breath Volume		2.121	01:02
Air Blank		0.000	01:03
Control Test		0.081	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

---

Data Download Date/Time: 04/22/2022 15:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLON,JOLNETH,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: SALINAS,JOSE  
ARRESTING OFFICER AGENCY: Winter Haven PD  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MARINKE

DATE/TIME OF TEST: 03/23/2022 01:49:43

---

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.081	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.119	01:45
Breath Volume	2.511	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.118	01:48
Breath Volume	2.214	01:48
Air Blank	0.000	01:48
Control Test	0.081	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

---

Data Download Date/Time: 04/22/2022 15:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: HODGES,MARCO  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: GRANT-PERKINS

DATE/TIME OF TEST: 03/23/2022 02:08:13

---

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:57	
Control Test	0.080	01:57	
Air Blank	0.000	01:58	
Subject Sample	# 1	02:01	No Sample Provided
Breath Volume	0.000	02:01	
Air Blank	0.000	02:01	
Air Blank	0.000	02:03	
Subject Sample	# 2	02:06	No Sample Provided
Breath Volume	0.000	02:06	
Air Blank	0.000	02:07	
Control Test	0.080	02:07	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

---

Data Download Date/Time: 04/22/2022 15:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELCHER,ALEX,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BELCHER,ALEX  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 22:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: DOWDY

DATE/TIME OF TEST: 03/23/2022 23:38:20

---

Diagnostics Check	OK	23:26	
Air Blank	0.000	23:27	
Control Test	0.081	23:27	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:31	No Sample Provided
Breath Volume	0.000	23:31	
Air Blank	0.000	23:31	
Air Blank	0.000	23:33	
Subject Sample	# 2	23:36	No Sample Provided
Breath Volume	0.000	23:36	
Air Blank	0.000	23:37	
Control Test	0.081	23:37	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

---

Data Download Date/Time: 04/22/2022 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FLETCHER,MELISSA,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: STEPHENS,TONY  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 02:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: GIDDENS

DATE/TIME OF TEST: 03/24/2022 04:24:38

---

Diagnostics Check	OK	04:13	
Air Blank	0.000	04:13	
Control Test	0.081	04:14	
Air Blank	0.000	04:14	
Subject Sample	# 1	04:17	No Sample Provided
Breath Volume	0.000	04:17	
Air Blank	0.000	04:18	
Air Blank	0.000	04:20	
Subject Sample	# 2	04:23	No Sample Provided
Breath Volume	0.000	04:23	
Air Blank	0.000	04:23	
Control Test	0.081	04:24	
Air Blank	0.000	04:24	
Diagnostics Check	OK	04:24	

---

Data Download Date/Time: 04/22/2022 15:14



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELCHER,ALEX,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BELCHER,ALEX  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 07:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:14

SUBJECT'S LAST NAME: PATRICK

DATE/TIME OF TEST: 03/24/2022 08:54:11

---

Diagnostics Check	OK	08:42	
Air Blank	0.000	08:43	
Control Test	0.081	08:43	
Air Blank	0.000	08:44	
Subject Sample	# 1	08:47	No Sample Provided
Breath Volume	0.000	08:47	
Air Blank	0.000	08:47	
Air Blank	0.000	08:49	
Subject Sample	# 2	08:52	No Sample Provided
Breath Volume	0.000	08:52	
Air Blank	0.000	08:53	
Control Test	0.082	08:53	
Air Blank	0.000	08:54	
Diagnostics Check	OK	08:54	

---

Data Download Date/Time: 04/22/2022 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAYNE, ABIGAIL, D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: PAYNE, ABIGAIL  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 19:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: SCHWEITZER

DATE/TIME OF TEST: 03/24/2022 21:19:59

---

Diagnostics Check		OK	21:13
Air Blank		0.000	21:14
Control Test		0.081	21:14
Air Blank		0.000	21:14
Subject Sample	# 1	0.226	21:15
Breath Volume		1.539	21:15
Air Blank		0.000	21:15
Air Blank		0.000	21:17
Subject Sample	# 2	0.219	21:18
Breath Volume		1.445	21:18
Air Blank		0.000	21:19
Control Test		0.080	21:19
Air Blank		0.000	21:19
Diagnostics Check		OK	21:19

---

Data Download Date/Time: 04/22/2022 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAWSON JR, MICHAEL, E  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: MORALES, ORLANDO  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 03/25/2022 04:43:48

---

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.080	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.174	04:39
Breath Volume	1.593	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:41
Subject Sample # 2	0.165	04:42
Breath Volume	1.324	04:42
Air Blank	0.000	04:42
Control Test	0.080	04:43
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

---

Data Download Date/Time: 04/22/2022 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLON,JOLNETH,B  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: 4C07,JOLNETH  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DOBY

DATE/TIME OF TEST: 03/26/2022 02:14:18

---

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.080	02:05
Air Blank		0.000	02:06
Subject Sample	# 1	0.123	02:08
Breath Volume		1.105	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.120	02:12
Breath Volume		1.199	02:12
Air Blank		0.000	02:13
Control Test		0.080	02:13
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

---

Data Download Date/Time: 04/22/2022 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HODGES,MARCO,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: HODGES,MARCO  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 04:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: ANTONIO-ANTONIO

DATE/TIME OF TEST: 03/26/2022 05:29:09

---

Diagnostics Check	OK	05:21
Air Blank	0.000	05:21
Control Test	0.081	05:21
Air Blank	0.000	05:22
Subject Sample # 1	0.152	05:22
Breath Volume	1.605	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.168	05:27
Breath Volume	1.519	05:27
Air Blank	0.000	05:28
Control Test	0.081	05:28
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

---

Data Download Date/Time: 04/22/2022 15:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WALDRON,DAVID,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: HERMELBRACHT,DONALD  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 12:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:05

SUBJECT'S LAST NAME: BRADING

DATE/TIME OF TEST: 03/26/2022 13:25:26

---

Diagnostics Check	OK	13:18
Air Blank	0.000	13:19
Control Test	0.081	13:19
Air Blank	0.000	13:20
Subject Sample # 1	0.000	13:21
Breath Volume	1.519	13:21
Air Blank	0.000	13:21
Air Blank	0.000	13:23
Subject Sample # 2	0.000	13:24
Breath Volume	1.843	13:24
Air Blank	0.000	13:24
Control Test	0.081	13:24
Air Blank	0.000	13:25
Diagnostics Check	OK	13:25

---

Data Download Date/Time: 04/22/2022 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SYKES, STEVEN,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: SYKES, STEVEN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 20:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 03/26/2022 21:14:19

---

Diagnostics Check	OK	21:02	
Air Blank	0.000	21:03	
Control Test	0.082	21:03	
Air Blank	0.000	21:04	
Subject Sample	# 1	21:07	No Sample Provided
Breath Volume	0.000	21:07	
Air Blank	0.000	21:07	
Air Blank	0.000	21:09	
Subject Sample	# 2	21:12	No Sample Provided
Breath Volume	0.000	21:12	
Air Blank	0.000	21:13	
Control Test	0.082	21:13	
Air Blank	0.000	21:14	
Diagnostics Check	OK	21:14	

---

Data Download Date/Time: 04/22/2022 15:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SYKES,STEVEN,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: SYKES,STEVEN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/26/2022 22:31:36

---

Diagnostics Check	OK	22:24
Air Blank	0.000	22:24
Control Test	0.080	22:25
Air Blank	0.000	22:25
Subject Sample # 1	0.153	22:26
Breath Volume	1.367	22:26
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.153	22:30
Breath Volume	1.628	22:30
Air Blank	0.000	22:30
Control Test	0.080	22:31
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

---

Data Download Date/Time: 04/22/2022 15:14



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: COLLINS,JOSHUA  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 23:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: EVA

DATE/TIME OF TEST: 03/27/2022 01:26:43

---

Diagnostics Check		OK	01:18
Air Blank		0.000	01:19
Control Test		0.080	01:19
Air Blank		0.000	01:20
Subject Sample	# 1	0.188	01:20
Breath Volume		1.656	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.186	01:25
Breath Volume		1.550	01:25
Air Blank		0.000	01:25
Control Test		0.080	01:26
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

---

Data Download Date/Time: 04/22/2022 15:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALSTON,SONDRA,L  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: BELONY,ANATUS  
ARRESTING OFFICER AGENCY: Bartow PD  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 07:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: SMALL JR

DATE/TIME OF TEST: 03/27/2022 10:12:11

---

Diagnostics Check	OK	10:05
Air Blank	0.000	10:06
Control Test	0.081	10:06
Air Blank	0.000	10:07
Subject Sample # 1	0.218	10:07
Breath Volume	1.707	10:07
Air Blank	0.000	10:08
Air Blank	0.000	10:10
Subject Sample # 2	0.222	10:10
Breath Volume	2.101	10:10
Air Blank	0.000	10:11
Control Test	0.081	10:11
Air Blank	0.000	10:12
Diagnostics Check	OK	10:12

---

Data Download Date/Time: 04/22/2022 15:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: LAFONTANT,SEDONA  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SILVA SANABRIA

DATE/TIME OF TEST: 03/28/2022 02:06:33

---

Diagnostics Check	OK	01:59
Air Blank	0.000	02:00
Control Test	0.080	02:00
Air Blank	0.000	02:01
Subject Sample # 1	0.091	02:01
Breath Volume	3.222	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.090	02:05
Breath Volume	2.687	02:05
Air Blank	0.000	02:05
Control Test	0.081	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

---

Data Download Date/Time: 04/22/2022 15:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SYKES,STEVEN,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: SYKES,STEVEN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 03:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: STRAND

DATE/TIME OF TEST: 03/28/2022 04:21:31

---

Diagnostics Check	OK	04:10	
Air Blank	0.000	04:10	
Control Test	0.080	04:10	
Air Blank	0.000	04:11	
Subject Sample	# 1	04:14	No Sample Provided
Breath Volume	0.000	04:14	
Air Blank	0.000	04:15	
Air Blank	0.000	04:17	
Subject Sample	# 2	04:20	No Sample Provided
Breath Volume	0.000	04:20	
Air Blank	0.000	04:20	
Control Test	0.080	04:20	
Air Blank	0.000	04:21	
Diagnostics Check	OK	04:21	

---

Data Download Date/Time: 04/22/2022 15:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUCAS,NATHANIEL,  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: LUCAS,NATHANIEL  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23621080A2  
DRY GAS EXPIRY DATE: 11/05/2023  
VIOLATION CODE: DUI  
ARREST TIME: 08:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 03/28/2022 09:50:24

---

Diagnostics Check	OK	09:44
Air Blank	0.000	09:44
Control Test	0.081	09:44
Air Blank	0.000	09:45
Subject Sample # 1	0.068	09:46
Breath Volume	1.484	09:46
Air Blank	0.000	09:46
Air Blank	0.000	09:48
Subject Sample # 2	0.069	09:48
Breath Volume	1.644	09:48
Air Blank	0.000	09:49
Control Test	0.081	09:49
Air Blank	0.000	09:50
Diagnostics Check	OK	09:50

---

Data Download Date/Time: 04/22/2022 15:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WERFEL,JEFFREY,R  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: WERFEL,JEFFREY  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 03/03/2022

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 17:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: MOTTRAM

DATE/TIME OF TEST: 03/03/2022 18:21:34

---

Diagnostics Check	OK	18:12
Air Blank	0.000	18:13
Control Test	0.075	18:13
Air Blank	0.000	18:14
Subject Sample # 1	0.207	18:15
Breath Volume	1.167	18:15
Air Blank	0.000	18:15
Air Blank	0.000	18:17
Subject Sample # 2	0.219	18:19
Breath Volume	2.386	18:19
Air Blank	0.000	18:20
Control Test	0.076	18:20
Air Blank	0.000	18:21
Diagnostics Check	OK	18:21

---

Data Download Date/Time: 04/27/2022 20:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSARD,JOSHUA,D  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: HANSARD,JOSHUA  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 03/03/2022

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1206004  
DRY GAS EXPIRY DATE: 09/20/2022  
VIOLATION CODE: DUI  
ARREST TIME: 22:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: MCCORMICK

DATE/TIME OF TEST: 03/12/2022 22:52:39

---

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.078	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.198	22:48
Breath Volume	3.457	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.200	22:51
Breath Volume	3.558	22:51
Air Blank	0.000	22:51
Control Test	0.077	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

---

Data Download Date/Time: 04/27/2022 20:41